

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: TRIMBOLI,MELANIE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CLEMENTEBOTKE

DATE/TIME OF TEST: 09/03/2022 01:33:01

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.067	01:28
Breath Volume	1.390	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.066	01:31
Breath Volume	1.132	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:33

Data Download Date/Time: 09/18/2022 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCIS, MATTHEW, C
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SAMUELS, ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 09/10/2022 07:46:56

Diagnostics Check	OK	07:40
Air Blank	0.000	07:41
Control Test	0.079	07:41
Air Blank	0.000	07:41
Subject Sample # 1	0.158	07:42
Breath Volume	3.058	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:44
Subject Sample # 2	0.159	07:45
Breath Volume	3.398	07:45
Air Blank	0.000	07:46
Control Test	0.079	07:46
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

Data Download Date/Time: 09/18/2022 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN,JONATHAN,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: SALAZAR,V

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: LUTE JR

DATE/TIME OF TEST: 09/12/2022 01:04:45

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.170	00:56
Breath Volume	1.886	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.197	01:00
Breath Volume	4.691	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:02
Subject Sample # 3	0.167	01:03
Breath Volume	2.230	01:03
Air Blank	0.000	01:03
Control Test	0.078	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 09/18/2022 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD, COLE, M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: OCONNELL, WILLIAM
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 08/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 09/13/2022 00:33:58

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.079	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.247	00:26
Breath Volume	2.085	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.225	00:29
Breath Volume	1.660	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:31
Subject Sample # 3	0.228	00:32
Breath Volume	1.992	00:32
Air Blank	0.000	00:33
Control Test	0.077	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 09/18/2022 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI,MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:58

SUBJECT'S LAST NAME: BINDA

DATE/TIME OF TEST: 09/21/2022 10:25:26

Diagnostics Check	OK	10:18
Air Blank	0.000	10:19
Control Test	0.080	10:19
Air Blank	0.000	10:20
Subject Sample # 1	0.000	10:20
Breath Volume	1.398	10:20
Air Blank	0.000	10:21
Air Blank	0.000	10:23
Subject Sample # 2	0.000	10:24
Breath Volume	1.253	10:24
Air Blank	0.000	10:24
Control Test	0.080	10:24
Air Blank	0.000	10:25
Diagnostics Check	OK	10:25

Data Download Date/Time: 10/23/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: CROOKS,J
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: THORNTON JR

DATE/TIME OF TEST: 09/30/2022 02:32:36

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.195	02:27
Breath Volume	1.781	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.187	02:30
Breath Volume	1.347	02:30
Air Blank	0.000	02:31
Control Test	0.077	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 10/23/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MARTIN,C
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: RIGGINS

DATE/TIME OF TEST: 09/02/2022 16:04:57

Diagnostics Check	OK	15:54
Air Blank	0.000	15:55
Control Test	0.081	15:55
Air Blank	0.000	15:56
Subject Sample # 1	0.245	15:56
Breath Volume	2.238	15:56
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.218	15:59
Breath Volume	1.449	15:59
Air Blank	0.000	16:00
Air Blank	0.000	16:02
Subject Sample # 3	0.211	16:03
Breath Volume	1.339	16:03
Air Blank	0.000	16:04
Control Test	0.079	16:04
Air Blank	0.000	16:04
Diagnostics Check	OK	16:04

Data Download Date/Time: 09/06/2022 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: TIDORT

DATE/TIME OF TEST: 09/03/2022 02:46:31

Diagnostics Check	OK	02:45	
Air Blank		02:46	RFI Detect
Air Blank	0.000	02:46	

Data Download Date/Time: 09/06/2022 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: TIDORT

DATE/TIME OF TEST: 09/03/2022 02:52:01

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 09/06/2022 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: REMY,LUCKENSON
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 09/03/2022 04:36:38

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:28	
Control Test	0.079	04:28	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:32	No Sample Provided
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 2	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Control Test	0.080	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 09/06/2022 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY, RONALD, G
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: SELF, GRANT
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: BUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: SHIVEL

DATE/TIME OF TEST: 09/04/2022 19:19:11

Diagnostics Check	OK	19:12
Air Blank	0.000	19:13
Control Test	0.080	19:13
Air Blank	0.000	19:14
Subject Sample # 1	0.169	19:14
Breath Volume	1.714	19:14
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample # 2	0.161	19:17
Breath Volume	1.714	19:17
Air Blank	0.000	19:18
Control Test	0.078	19:18
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

Data Download Date/Time: 09/06/2022 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUM,JACOB,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WILSON,ALEX
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: GALLEGOS

DATE/TIME OF TEST: 09/05/2022 08:26:00

Diagnostics Check	OK	08:19
Air Blank	0.000	08:20
Control Test	0.080	08:20
Air Blank	0.000	08:21
Subject Sample # 1	0.000	08:21
Breath Volume	1.796	08:21
Air Blank	0.000	08:21
Air Blank	0.000	08:23
Subject Sample # 2	0.000	08:24
Breath Volume	2.066	08:24
Air Blank	0.000	08:25
Control Test	0.080	08:25
Air Blank	0.000	08:25
Diagnostics Check	OK	08:26

Data Download Date/Time: 09/06/2022 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CONDER

DATE/TIME OF TEST: 09/08/2022 02:58:16

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.127	02:53
Breath Volume	1.472	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.126	02:56
Breath Volume	1.488	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CHIAPPY,C
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/09/2022 00:39:24

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.150	00:34
Breath Volume	3.394	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.153	00:37
Breath Volume	2.714	00:37
Air Blank	0.000	00:38
Control Test	0.077	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/09/2022 22:27:11

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:24	
Control Test	0.080	22:25	
Air Blank	0.000	22:25	
Subject Sample	# 1	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/09/2022 22:45:52

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.081	22:32	
Air Blank	0.000	22:32	
Subject Sample	# 1 0.205	22:35	Volume Not Met
Breath Volume	0.140	22:35	
Air Blank	0.000	22:36	
Air Blank	0.000	22:38	
Subject Sample	# 2 0.211	22:39	
Breath Volume	1.769	22:39	
Air Blank	0.000	22:39	
Air Blank	0.000	22:41	
Subject Sample	# 3 0.212	22:44	
Breath Volume	1.820	22:44	
Air Blank	0.000	22:44	
Control Test	0.080	22:45	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BYRD,B
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: AGUILAR MARTINEZ

DATE/TIME OF TEST: 09/12/2022 00:50:38

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.187	00:45
Breath Volume	2.101	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.183	00:48
Breath Volume	2.781	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: SOMOHANO

DATE/TIME OF TEST: 09/13/2022 21:24:13

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.081	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.000	21:19
Breath Volume	4.253	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.000	21:22
Breath Volume	3.656	21:22
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CHAMLA,A
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/15/2022 21:56:53

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:45	
Control Test	0.080	21:45	
Air Blank	0.000	21:46	
Subject Sample	# 1 0.166	21:49	Volume Not Met
Breath Volume	0.617	21:49	
Air Blank	0.000	21:50	
Air Blank	0.000	21:52	
Subject Sample	# 2 0.167	21:55	Volume Not Met
Breath Volume	0.812	21:55	
Air Blank	0.000	21:55	
Control Test	0.077	21:56	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACH,B,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JAMES,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: FLEENOR

DATE/TIME OF TEST: 09/16/2022 16:00:04

Diagnostics Check	OK	15:50	
Air Blank	0.000	15:50	
Control Test	0.080	15:51	
Air Blank	0.000	15:51	
Subject Sample	# 1 0.288	15:52	No .020 agreement
Breath Volume	3.128	15:52	
Air Blank	0.000	15:53	
Air Blank	0.000	15:54	
Subject Sample	# 2 0.263	15:55	No .020 agreement
Breath Volume	2.578	15:55	
Air Blank	0.000	15:56	
Air Blank	0.000	15:57	
Subject Sample	# 3 0.241	15:58	No .020 agreement
Breath Volume	3.281	15:58	
Air Blank	0.000	15:59	
Control Test	0.075	15:59	
Air Blank	0.000	15:59	
Diagnostics Check	OK	16:00	

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 09/17/2022 01:48:23

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.139	01:43
Breath Volume	1.437	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.131	01:46
Breath Volume	1.289	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NCVEIGH,TRAVOR
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 09/17/2022 03:20:30

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.188	03:15
Breath Volume	2.125	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.169	03:18
Breath Volume	2.117	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: STOKER

DATE/TIME OF TEST: 09/17/2022 06:59:15

Diagnostics Check	OK	06:44	
Air Blank	0.000	06:45	
Control Test	0.080	06:45	
Air Blank	0.000	06:46	
Subject Sample # 1	0.131	06:48	
Breath Volume	1.406	06:48	
Air Blank	0.000	06:49	
Air Blank	0.000	06:51	
Subject Sample # 2	0.111	06:54	Volume Not Met
Breath Volume	0.753	06:54	
Air Blank	0.000	06:55	
Air Blank	0.000	06:56	
Subject Sample # 3	0.132	06:57	
Breath Volume	1.281	06:57	
Air Blank	0.000	06:58	
Control Test	0.077	06:58	
Air Blank	0.000	06:59	
Diagnostics Check	OK	06:59	

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 09/18/2022 02:32:16

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.126	02:26
Breath Volume	1.265	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.126	02:30
Breath Volume	1.136	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MATERA,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: ANAYA

DATE/TIME OF TEST: 09/18/2022 02:42:39

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.223	02:38
Breath Volume	3.312	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.226	02:40
Breath Volume	3.460	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 09/21/2022 02:17:02

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.263	02:12
Breath Volume	1.324	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.263	02:15
Breath Volume	1.570	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:17

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: CRUZ RAMIREZ

DATE/TIME OF TEST: 09/21/2022 19:17:46

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.081	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.262	19:12
Breath Volume	1.984	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.264	19:15
Breath Volume	1.613	19:15
Air Blank	0.000	19:16
Control Test	0.078	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: FULTON

DATE/TIME OF TEST: 09/22/2022 00:02:22

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.081	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.150	23:57
Breath Volume	1.968	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.164	00:00
Breath Volume	2.527	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GALLEGOS

DATE/TIME OF TEST: 09/24/2022 02:29:01

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.081	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.178	02:24
Breath Volume	2.109	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.171	02:27
Breath Volume	1.660	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:29

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SPIVY

DATE/TIME OF TEST: 09/25/2022 04:14:31

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.188	04:09
Breath Volume	3.707	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.200	04:12
Breath Volume	5.218	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: GALARZA ROJAS

DATE/TIME OF TEST: 09/26/2022 23:31:08

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.182	23:21
Breath Volume	2.796	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.206	23:26
Breath Volume	3.335	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:28
Subject Sample # 3	0.195	23:29
Breath Volume	2.339	23:29
Air Blank	0.000	23:30
Control Test	0.076	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENNIFER,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TORRES,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 09/28/2022 00:49:46

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.039	00:45
Breath Volume	2.992	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.041	00:48
Breath Volume	3.105	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: EVORA

DATE/TIME OF TEST: 09/03/2022 02:07:12

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.079	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 11/03/2022 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDA,CHELSY,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: LANDA,CHELSY
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: JARAMILLO

DATE/TIME OF TEST: 09/12/2022 16:13:57

Diagnostics Check	OK	16:06
Air Blank	0.000	16:07
Control Test	0.081	16:07
Air Blank	0.000	16:08
Subject Sample # 1	0.263	16:08
Breath Volume	3.105	16:08
Air Blank	0.000	16:10
Air Blank	0.000	16:11
Subject Sample # 2	0.246	16:11
Breath Volume	3.519	16:11
Air Blank	0.000	16:12
Control Test	0.076	16:13
Air Blank	0.000	16:13
Diagnostics Check	OK	16:13

Data Download Date/Time: 11/03/2022 18:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SEPULVEDA

DATE/TIME OF TEST: 09/25/2022 04:53:54

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.194	04:48
Breath Volume	2.570	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:50
Subject Sample # 2	0.201	04:51
Breath Volume	2.136	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 11/03/2022 18:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DARIEN,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: RODRIGUEZ,DARIEN
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: CORDOVES

DATE/TIME OF TEST: 09/26/2022 11:34:04

Diagnostics Check	OK	11:30	
Air Blank	0.000	11:31	
Control Test	0.080	11:31	
Air Blank	0.000	11:32	
Subject Sample	# 1	11:32	Subject Test Refused
Breath Volume	0.000	11:32	
Air Blank	0.000	11:33	
Control Test	0.080	11:33	
Air Blank	0.000	11:33	
Diagnostics Check	OK	11:34	

Data Download Date/Time: 11/03/2022 18:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 08/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: GELTZER

DATE/TIME OF TEST: 09/30/2022 21:53:31

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:48	
Control Test	0.080	21:49	
Air Blank	0.000	21:49	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Control Test	0.079	21:52	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 11/03/2022 18:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORITZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: SAVASUK-MALDONADO

DATE/TIME OF TEST: 09/01/2022 23:20:47

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.111	23:16
Breath Volume	1.199	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.113	23:19
Breath Volume	1.304	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CHANDLER,B
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 09/02/2022 04:28:11

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.000	04:23
Breath Volume	1.468	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.000	04:26
Breath Volume	1.468	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITMAN,JACELYN,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FREYTAG,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 09/04/2022 01:46:13

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.000	01:41
Breath Volume	1.539	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.000	01:44
Breath Volume	1.796	01:44
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: CUTRELL

DATE/TIME OF TEST: 09/04/2022 21:58:25

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.222	21:53
Breath Volume	3.269	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.225	21:56
Breath Volume	4.449	21:56
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,JOSEPH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BORGE

DATE/TIME OF TEST: 09/04/2022 23:31:31

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:23	
Control Test	0.080	23:23	
Air Blank	0.000	23:23	
Subject Sample	# 1 0.191	23:25	
Breath Volume	2.894	23:25	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 2	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Control Test	0.080	23:30	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BURNS,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CREUTZ

DATE/TIME OF TEST: 09/06/2022 00:32:14

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.000	00:27
Breath Volume	1.792	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.000	00:30
Breath Volume	1.964	00:30
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORGAN,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 09/06/2022 00:41:50

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.000	00:37
Breath Volume	1.605	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.000	00:40
Breath Volume	1.785	00:40
Air Blank	0.000	00:40
Control Test	0.081	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARVER,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MARTELL

DATE/TIME OF TEST: 09/07/2022 04:28:02

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.000	04:23
Breath Volume	2.191	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.000	04:26
Breath Volume	1.976	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:28

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWLIN,ERIC,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NEWLIN,CD
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 09/07/2022 13:07:24

Diagnostics Check	OK	13:01
Air Blank	0.000	13:01
Control Test	0.080	13:01
Air Blank	0.000	13:02
Subject Sample # 1	0.000	13:03
Breath Volume	2.300	13:03
Air Blank	0.000	13:03
Air Blank	0.000	13:05
Subject Sample # 2	0.000	13:06
Breath Volume	2.804	13:06
Air Blank	0.000	13:06
Control Test	0.080	13:06
Air Blank	0.000	13:07
Diagnostics Check	OK	13:07

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: SOUSA

DATE/TIME OF TEST: 09/07/2022 23:00:48

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.080	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.145	22:55
Breath Volume	1.683	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.139	22:59
Breath Volume	1.667	22:59
Air Blank	0.000	22:59
Control Test	0.080	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MYRIE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 09/09/2022 11:28:09

Diagnostics Check	OK	11:21
Air Blank	0.000	11:22
Control Test	0.080	11:22
Air Blank	0.000	11:23
Subject Sample # 1	0.000	11:23
Breath Volume	4.699	11:23
Air Blank	0.000	11:24
Air Blank	0.000	11:26
Subject Sample # 2	0.000	11:26
Breath Volume	4.078	11:26
Air Blank	0.000	11:27
Control Test	0.080	11:27
Air Blank	0.000	11:28
Diagnostics Check	OK	11:28

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SHEEN,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: CARNINE

DATE/TIME OF TEST: 09/09/2022 20:17:21

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.039	20:12
Breath Volume	4.312	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.034	20:15
Breath Volume	2.027	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 09/09/2022 22:50:22

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.081	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.231	22:45
Breath Volume	2.472	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.238	22:48
Breath Volume	2.675	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SAVIC,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 09/10/2022 00:58:29

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.136	00:53
Breath Volume	2.035	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.150	00:57
Breath Volume	1.835	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 09/27/2022 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARVER,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: DUTCHER

DATE/TIME OF TEST: 09/10/2022 05:20:40

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.080	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.224	05:16
Breath Volume	3.515	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.204	05:19
Breath Volume	2.394	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTOFER,CHRISTOPH
ER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Female

OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: ROCH

DATE/TIME OF TEST: 09/10/2022 16:19:36

Diagnostics Check	OK	16:13
Air Blank	0.000	16:14
Control Test	0.080	16:14
Air Blank	0.000	16:14
Subject Sample # 1	0.192	16:15
Breath Volume	2.332	16:15
Air Blank	0.000	16:15
Air Blank	0.000	16:17
Subject Sample # 2	0.192	16:18
Breath Volume	1.769	16:18
Air Blank	0.000	16:18
Control Test	0.080	16:19
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,HENRY,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CREWS,HENRY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: DUTCHER

DATE/TIME OF TEST: 09/10/2022 16:37:08

Diagnostics Check	OK	16:31
Air Blank	0.000	16:31
Control Test	0.080	16:32
Air Blank	0.000	16:32
Subject Sample # 1	0.000	16:32
Breath Volume	3.472	16:32
Air Blank	0.000	16:33
Air Blank	0.000	16:35
Subject Sample # 2	0.000	16:35
Breath Volume	3.000	16:35
Air Blank	0.000	16:36
Control Test	0.080	16:36
Air Blank	0.000	16:37
Diagnostics Check	OK	16:37

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEP,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: SOLOWAY

DATE/TIME OF TEST: 09/10/2022 18:42:15

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.080	18:36
Air Blank	0.000	18:37
Subject Sample # 1	0.000	18:37
Breath Volume	1.125	18:37
Air Blank	0.000	18:38
Air Blank	0.000	18:40
Subject Sample # 2	0.000	18:40
Breath Volume	1.265	18:40
Air Blank	0.000	18:41
Control Test	0.081	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SNOW,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ALAVACHE

DATE/TIME OF TEST: 09/10/2022 20:46:44

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.081	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.000	20:42
Breath Volume	2.207	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.000	20:45
Breath Volume	1.878	20:45
Air Blank	0.000	20:45
Control Test	0.081	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PATTERSON,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: CUFFE

DATE/TIME OF TEST: 09/10/2022 21:12:46

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.080	20:57	
Air Blank	0.000	20:57	
Subject Sample	# 1 0.233	21:00	Volume Not Met
Breath Volume	0.445	21:00	
Air Blank	0.000	21:01	
Air Blank	0.000	21:03	
Subject Sample	# 2 0.267	21:05	
Breath Volume	1.917	21:05	
Air Blank	0.000	21:06	
Air Blank	0.000	21:07	
Subject Sample	# 3 0.240	21:11	Volume Not Met
Breath Volume	0.167	21:11	
Air Blank	0.000	21:11	
Control Test	0.079	21:12	
Air Blank	0.000	21:12	
Diagnostics Check	OK	21:12	

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: DIAZ GONZALEZ

DATE/TIME OF TEST: 09/10/2022 22:04:55

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.085	22:00
Breath Volume	1.746	22:00
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.083	22:03
Breath Volume	1.339	22:03
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,OFC
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 09/10/2022 23:38:42

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.081	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.191	23:34
Breath Volume	2.222	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.194	23:37
Breath Volume	2.453	23:37
Air Blank	0.000	23:37
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PATTERSON,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: FURMAN

DATE/TIME OF TEST: 09/11/2022 04:49:50

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.141	04:45
Breath Volume	1.492	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.127	04:48
Breath Volume	1.613	04:48
Air Blank	0.000	04:48
Control Test	0.080	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,HENRY,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CREWS,HENRY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: FURMAN

DATE/TIME OF TEST: 09/11/2022 09:40:18

Diagnostics Check	OK	09:33
Air Blank	0.000	09:33
Control Test	0.080	09:34
Air Blank	0.000	09:34
Subject Sample # 1	0.034	09:35
Breath Volume	2.234	09:35
Air Blank	0.000	09:35
Air Blank	0.000	09:37
Subject Sample # 2	0.031	09:38
Breath Volume	1.562	09:38
Air Blank	0.000	09:39
Control Test	0.080	09:39
Air Blank	0.000	09:40
Diagnostics Check	OK	09:40

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: IRIS,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 09/11/2022 19:04:41

Diagnostics Check	OK	18:57
Air Blank	0.000	18:58
Control Test	0.081	18:58
Air Blank	0.000	18:59
Subject Sample # 1	0.000	19:00
Breath Volume	2.894	19:00
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.000	19:03
Breath Volume	1.980	19:03
Air Blank	0.000	19:03
Control Test	0.080	19:04
Air Blank	0.000	19:04
Diagnostics Check	OK	19:04

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GANK,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: VOYLES

DATE/TIME OF TEST: 09/11/2022 22:17:08

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.000	22:12
Breath Volume	3.894	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:15
Subject Sample # 2	0.000	22:15
Breath Volume	3.394	22:15
Air Blank	0.000	22:16
Control Test	0.081	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITMAN,JACELYN,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORITZ,JONATHAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 09/12/2022 22:18:15

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.081	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.112	22:13
Breath Volume	4.156	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.108	22:16
Breath Volume	3.546	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,HENRY,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WOELK,MATTHEW
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 09/14/2022 13:10:37

Diagnostics Check	OK	12:59	
Air Blank	0.000	12:59	
Control Test	0.080	12:59	
Air Blank	0.000	13:00	
Subject Sample	# 1	13:03	No Sample Provided
Breath Volume	0.000	13:03	
Air Blank	0.000	13:03	
Air Blank	0.000	13:05	
Subject Sample	# 2	0.311	13:06
Breath Volume	2.175	13:06	
Air Blank	0.000	13:06	
Air Blank	0.000	13:08	
Subject Sample	# 3	0.314	13:09
Breath Volume	1.507	13:09	
Air Blank	0.000	13:09	
Control Test	0.079	13:10	
Air Blank	0.000	13:10	
Diagnostics Check	OK	13:10	

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 09/14/2022 19:52:53

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.093	19:48
Breath Volume	2.156	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.086	19:51
Breath Volume	1.519	19:51
Air Blank	0.000	19:51
Control Test	0.080	19:52
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VETRI,OFC
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 09/14/2022 20:37:43

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.081	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.158	20:33
Breath Volume	1.816	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.162	20:36
Breath Volume	1.851	20:36
Air Blank	0.000	20:36
Control Test	0.081	20:37
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 09/27/2022 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,OFC
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MARCIANO

DATE/TIME OF TEST: 09/15/2022 01:06:22

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.024	01:01
Breath Volume	2.179	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.023	01:04
Breath Volume	1.597	01:04
Air Blank	0.000	01:05
Control Test	0.081	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: HOLSWORTH

DATE/TIME OF TEST: 09/16/2022 20:29:13

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.081	20:23
Air Blank	0.000	20:23
Subject Sample # 1	0.171	20:24
Breath Volume	2.496	20:24
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.161	20:27
Breath Volume	2.250	20:27
Air Blank	0.000	20:28
Control Test	0.081	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRITES,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PARRISH,GARRETT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 09/17/2022 00:16:16

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.137	00:11
Breath Volume	3.000	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.136	00:14
Breath Volume	2.898	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITMAN,JACELYN,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MCGILLIVRAY,STEPHANIE
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: DYKES

DATE/TIME OF TEST: 09/17/2022 23:15:13

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.081	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.128	23:10
Breath Volume	1.183	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.126	23:13
Breath Volume	1.148	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SIDES,RACHEL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: LUONG

DATE/TIME OF TEST: 09/18/2022 02:08:44

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.161	02:01
Breath Volume	1.605	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.149	02:07
Breath Volume	1.222	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 09/18/2022 09:40:41

Diagnostics Check	OK	09:34
Air Blank	0.000	09:34
Control Test	0.080	09:35
Air Blank	0.000	09:35
Subject Sample # 1	0.286	09:36
Breath Volume	2.425	09:36
Air Blank	0.000	09:36
Air Blank	0.000	09:38
Subject Sample # 2	0.275	09:39
Breath Volume	2.441	09:39
Air Blank	0.000	09:39
Control Test	0.079	09:40
Air Blank	0.000	09:40
Diagnostics Check	OK	09:40

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 09/18/2022 18:31:46

Diagnostics Check	OK	18:24
Air Blank	0.000	18:25
Control Test	0.080	18:25
Air Blank	0.000	18:26
Subject Sample # 1	0.164	18:27
Breath Volume	2.593	18:27
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.160	18:30
Breath Volume	2.539	18:30
Air Blank	0.000	18:30
Control Test	0.081	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITMAN,JACELYM,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LITTY,LITTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MARCH

DATE/TIME OF TEST: 09/19/2022 00:47:40

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.055	00:43
Breath Volume	2.632	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.053	00:46
Breath Volume	2.656	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LITTY,LIT'
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 09/19/2022 02:14:40

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.021	02:10
Breath Volume	2.796	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.020	02:13
Breath Volume	2.546	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MARCHMAN,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 09/19/2022 02:43:17

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.000	02:38
Breath Volume	1.546	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.000	02:41
Breath Volume	1.558	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SALES

DATE/TIME OF TEST: 09/19/2022 03:59:29

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.162	03:54
Breath Volume	2.343	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.164	03:58
Breath Volume	4.003	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TURNER,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 09/19/2022 19:45:52

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.080	19:40
Air Blank	0.000	19:40
Subject Sample # 1	0.000	19:41
Breath Volume	1.574	19:41
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.000	19:44
Breath Volume	1.570	19:44
Air Blank	0.000	19:44
Control Test	0.081	19:45
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MILLER,DET
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: ATWOOD

DATE/TIME OF TEST: 09/19/2022 20:31:00

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.081	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.000	20:26
Breath Volume	2.570	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.000	20:29
Breath Volume	2.472	20:29
Air Blank	0.000	20:30
Control Test	0.080	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:31

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MILLER,DET
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PREVITI

DATE/TIME OF TEST: 09/19/2022 20:39:44

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.080	20:34
Air Blank	0.000	20:35
Subject Sample # 1	0.000	20:35
Breath Volume	2.675	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.000	20:38
Breath Volume	2.539	20:38
Air Blank	0.000	20:38
Control Test	0.080	20:39
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,B
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SIMANDL

DATE/TIME OF TEST: 09/23/2022 03:38:49

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.119	03:33
Breath Volume	1.566	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.111	03:37
Breath Volume	2.761	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: JEWELL

DATE/TIME OF TEST: 09/23/2022 18:47:12

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.281	18:42
Breath Volume	3.484	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.288	18:45
Breath Volume	3.886	18:45
Air Blank	0.000	18:46
Control Test	0.079	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 09/27/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,KANADA,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: LEICHTY

DATE/TIME OF TEST: 09/23/2022 23:23:30

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.141	23:18
Breath Volume	1.851	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.138	23:22
Breath Volume	1.472	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 09/27/2022 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,M
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CALLAHAN

DATE/TIME OF TEST: 09/24/2022 03:04:12

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.081	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.061	02:59
Breath Volume	1.593	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.058	03:02
Breath Volume	1.710	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 09/27/2022 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,KANADA,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT,CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GURMEZA

DATE/TIME OF TEST: 09/24/2022 21:47:53

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.000	21:43
Breath Volume	2.941	21:43
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.000	21:46
Breath Volume	4.035	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 09/27/2022 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,KANADA,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BRASCH,CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: ABRAHAM JR

DATE/TIME OF TEST: 09/25/2022 02:41:41

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.000	02:37
Breath Volume	2.398	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.000	02:40
Breath Volume	2.269	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 09/27/2022 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MIR,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 09/27/2022 15:29:37

Diagnostics Check	OK	15:23
Air Blank	0.000	15:23
Control Test	0.080	15:24
Air Blank	0.000	15:24
Subject Sample # 1	0.180	15:25
Breath Volume	2.105	15:25
Air Blank	0.000	15:25
Air Blank	0.000	15:27
Subject Sample # 2	0.187	15:28
Breath Volume	2.027	15:28
Air Blank	0.000	15:28
Control Test	0.080	15:29
Air Blank	0.000	15:29
Diagnostics Check	OK	15:29

Data Download Date/Time: 11/02/2022 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,JONATHAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: ROBSON

DATE/TIME OF TEST: 09/27/2022 22:02:49

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.254	21:57
Breath Volume	1.718	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.240	22:01
Breath Volume	1.152	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 11/02/2022 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:57

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 09/01/2022 23:26:33

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.206	23:22
Breath Volume	2.062	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.212	23:25
Breath Volume	2.433	23:25
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 09/15/2022 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/03/2022 03:53:33

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.046	03:48
Breath Volume	4.421	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.046	03:52
Breath Volume	4.097	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 09/15/2022 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: CAMBRAY

DATE/TIME OF TEST: 09/04/2022 04:18:06

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.074	04:13
Breath Volume	1.789	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.074	04:16
Breath Volume	1.660	04:16
Air Blank	0.000	04:17
Control Test	0.078	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 09/15/2022 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: WILSON,KALEB

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 19:27

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: CARTMELL

DATE/TIME OF TEST: 09/04/2022 20:01:50

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.079	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.133	19:57
Breath Volume	1.488	19:57
Air Blank	0.000	19:58
Air Blank	0.000	19:59
Subject Sample # 2	0.136	20:00
Breath Volume	2.484	20:00
Air Blank	0.000	20:00
Control Test	0.079	20:01
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 09/15/2022 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 09/04/2022 23:02:44

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.172	22:58
Breath Volume	2.273	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.178	23:01
Breath Volume	2.742	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 09/15/2022 10:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 09/12/2022 20:59:02

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.060	20:54
Breath Volume	1.980	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.057	20:57
Breath Volume	1.386	20:57
Air Blank	0.000	20:58
Control Test	0.078	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:59

Data Download Date/Time: 09/15/2022 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 16:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 09/13/2022 17:30:43

Diagnostics Check	OK	17:24
Air Blank	0.000	17:24
Control Test	0.077	17:25
Air Blank	0.000	17:25
Subject Sample # 1	0.000	17:26
Breath Volume	3.031	17:26
Air Blank	0.000	17:26
Air Blank	0.000	17:28
Subject Sample # 2	0.000	17:29
Breath Volume	1.859	17:29
Air Blank	0.000	17:29
Control Test	0.078	17:30
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

Data Download Date/Time: 09/15/2022 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFONTANT, SEDONA, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: COLLINS, JOSHUA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 09/20/2022 23:25:54

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.128	23:21
Breath Volume	2.316	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:23
Subject Sample # 2	0.131	23:24
Breath Volume	2.671	23:24
Air Blank	0.000	23:25
Control Test	0.078	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 10/05/2022 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:54

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 09/22/2022 03:58:10

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.205	03:53
Breath Volume	2.089	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.191	03:56
Breath Volume	1.722	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 10/05/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: WILSON,KALEB

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 18:56

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/22/2022 20:06:10

Diagnostics Check	OK	19:59
Air Blank	0.000	19:59
Control Test	0.080	20:00
Air Blank	0.000	20:00
Subject Sample # 1	0.148	20:01
Breath Volume	2.937	20:01
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.165	20:04
Breath Volume	2.789	20:04
Air Blank	0.000	20:05
Control Test	0.079	20:05
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 10/05/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/22/2022 22:15:25

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:04	
Control Test	0.079	22:04	
Air Blank	0.000	22:05	
Subject Sample	# 1	22:08	No Sample Provided
Breath Volume	0.000	22:08	
Air Blank	0.000	22:08	
Air Blank	0.000	22:10	
Subject Sample	# 2	22:13	No Sample Provided
Breath Volume	0.000	22:13	
Air Blank	0.000	22:14	
Control Test	0.079	22:14	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 10/05/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHEK, LEONARD, H
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MUNOZ, CASANDRA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/03/2022 02:56:16

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:52	
Control Test	0.078	02:52	
Air Blank	0.000	02:52	
Subject Sample # 1	0.162	02:53	
Breath Volume	1.843	02:53	
Air Blank	0.000	02:53	
Air Blank		02:55	Ambient Fail
Air Blank	0.000	02:56	

Data Download Date/Time: 09/29/2022 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHEK, LEONARD, H
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MUNOZ, CASANDRA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/03/2022 03:04:56

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.077	02:59	
Air Blank	0.000	03:00	
Subject Sample # 1	0.136	03:00	
Breath Volume	1.789	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample # 2	0.151	03:03	
Breath Volume	2.054	03:03	
Air Blank	0.000	03:04	
Control Test	0.077	03:04	Interferent Detect
Air Blank	0.000	03:04	

Data Download Date/Time: 09/29/2022 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ROBERSON,ISRAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: GAJRAJ

DATE/TIME OF TEST: 09/12/2022 00:50:31

Diagnostics Check	OK	00:49	
Air Blank		00:50	Ambient Fail
Air Blank		00:50	Ambient Fail

Data Download Date/Time: 09/29/2022 04:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN',D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ROBERSON,ISRAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: GAJRAJ

DATE/TIME OF TEST: 09/12/2022 01:06:02

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:53	
Control Test	0.079	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 1 0.190	00:57	Volume Not Met
Breath Volume	0.609	00:57	
Air Blank	0.000	00:57	
Air Blank	0.000	00:59	
Subject Sample	# 2 0.224	01:00	
Breath Volume	1.894	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample	# 3 0.224	01:04	
Breath Volume	1.421	01:04	
Air Blank	0.000	01:05	
Control Test	0.078	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 09/29/2022 04:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETO,JOSHUA,W
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: JARRETT,WILLIAM
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: XILOJ CHUN

DATE/TIME OF TEST: 09/18/2022 18:09:52

Diagnostics Check	OK	18:03
Air Blank	0.000	18:03
Control Test	0.079	18:04
Air Blank	0.000	18:04
Subject Sample # 1	0.233	18:05
Breath Volume	2.390	18:05
Air Blank	0.000	18:05
Air Blank	0.000	18:07
Subject Sample # 2	0.225	18:08
Breath Volume	2.421	18:08
Air Blank	0.000	18:08
Control Test	0.078	18:09
Air Blank	0.000	18:09
Diagnostics Check	OK	18:09

Data Download Date/Time: 09/29/2022 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: JACOBS,DOMINIC
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/19/2022 21:46:13

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.225	21:41
Breath Volume	2.046	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.210	21:44
Breath Volume	1.367	21:44
Air Blank	0.000	21:45
Control Test	0.076	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 09/29/2022 04:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: THERIALT,TANNER
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 09/20/2022 00:55:12

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.046	00:50
Breath Volume	1.726	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.046	00:53
Breath Volume	1.347	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 09/29/2022 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: THOMPSON,BRIAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ACKON

DATE/TIME OF TEST: 09/21/2022 01:30:04

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.118	01:24
Breath Volume	1.957	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.118	01:28
Breath Volume	1.320	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:30

Data Download Date/Time: 09/29/2022 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCARDI,A,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RICCARDI,A
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 09/22/2022 02:06:26

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.177	02:01
Breath Volume	3.148	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.173	02:04
Breath Volume	3.500	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 09/29/2022 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,LEE,B
BREATH TEST OPERATOR AGENCY: NAS KEY WEST
ARRESTING OFFICER: ANDERSON,LEE
ARRESTING OFFICER AGENCY: NAS KEY WEST
LAST AGENCY INSPECTION: 08/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1514466
DRY GAS EXPIRY DATE: 06/18/2016
VIOLATION CODE: ADMIN
ARREST TIME: 09:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:49

SUBJECT'S LAST NAME: TUTOR

DATE/TIME OF TEST: 09/08/2022 10:19:40

Diagnostics Check	OK	10:13
Air Blank	0.000	10:13
Control Test	0.080	10:14
Air Blank	0.000	10:14
Subject Sample # 1	0.095	10:15
Breath Volume	2.519	10:15
Air Blank	0.000	10:15
Air Blank	0.000	10:17
Subject Sample # 2	0.095	10:18
Breath Volume	2.742	10:18
Air Blank	0.000	10:18
Control Test	0.078	10:19
Air Blank	0.000	10:19
Diagnostics Check	OK	10:19

Data Download Date/Time: 10/05/2022 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: PAIGE, LEON
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/03/2022 21:37:31

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.000	21:32
Breath Volume	2.089	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.000	21:36
Breath Volume	2.496	21:36
Air Blank	0.000	21:36
Control Test	0.081	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 09/22/2022 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1224982
BREATH TEST OPERATOR: BATTAGLIA,PAUL,M	DRY GAS EXPIRY DATE: 10/31/2022
BREATH TEST OPERATOR AGENCY: NORTH BAY VILLAGE PD	VIOLATION CODE: DUI
ARRESTING OFFICER: BATTAGLIA,PAUL	ARREST TIME: 02:58
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD	GENDER: Male
LAST AGENCY INSPECTION: 08/25/2022	OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: TORRES GONZALEZ DATE/TIME OF TEST: 09/04/2022 03:53:31

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 09/22/2022 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: DARIAS, SOFIA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HERNANDEZ DELGADO

DATE/TIME OF TEST: 09/11/2022 03:07:41

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.189	03:03
Breath Volume	2.167	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.188	03:06
Breath Volume	1.605	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 09/22/2022 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1224982
BREATH TEST OPERATOR: BATTAGLIA,PAUL,M	DRY GAS EXPIRY DATE: 10/31/2022
BREATH TEST OPERATOR AGENCY: NORTH BAY VILLAGE PD	VIOLATION CODE: DUI
ARRESTING OFFICER: NOWAK,KAMIL	ARREST TIME: 01:54
ARRESTING OFFICER AGENCY: MIAMI BEACH PD	GENDER: Female
LAST AGENCY INSPECTION: 09/22/2022	OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LALL-HARRY	DATE/TIME OF TEST: 09/25/2022	02:58:17
---------------------------------	-------------------------------	----------

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.170	02:53
Breath Volume	2.164	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.173	02:56
Breath Volume	2.121	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 10/20/2022 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1224982
BREATH TEST OPERATOR: BATTAGLIA,PAUL,M	DRY GAS EXPIRY DATE: 10/31/2022
BREATH TEST OPERATOR AGENCY: NORTH BAY VILLAGE PD	VIOLATION CODE: DUI
ARRESTING OFFICER: BATTAGLIA,PAUL	ARREST TIME: 23:43
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD	GENDER: Male
LAST AGENCY INSPECTION: 09/22/2022	OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: NUNEZ RODRIGUEZ	DATE/TIME OF TEST: 09/30/2022 00:39:27
--------------------------------------	--

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.079	00:37	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Control Test	0.080	00:38	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 10/20/2022 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,ALEXIS,
BREATH TEST OPERATOR AGENCY: HIALEAH PD
ARRESTING OFFICER: ARZOLA,KENNETH
ARRESTING OFFICER AGENCY: HIALEAH PD
LAST AGENCY INSPECTION: 08/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/24/2022 01:58:07

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:55	
Control Test	0.080	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:57	
Control Test	0.080	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 09/30/2022 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,KEVIN,X
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CONSUEGRA,ALEJANDRO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: AMAYA

DATE/TIME OF TEST: 09/21/2022 20:36:07

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.077	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.195	20:31
Breath Volume	1.761	20:31
Air Blank	0.000	20:32
Air Blank	0.000	20:33
Subject Sample # 2	0.186	20:34
Breath Volume	2.050	20:34
Air Blank	0.000	20:35
Control Test	0.075	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 09/29/2022 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: BATEY

DATE/TIME OF TEST: 09/02/2022 21:10:40

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.078	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.278	21:06
Breath Volume	2.210	21:06
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.273	21:08
Breath Volume	1.960	21:08
Air Blank	0.000	21:09
Control Test	0.077	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 09/20/2022 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: CALLOWAY

DATE/TIME OF TEST: 09/11/2022 00:00:49

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.078	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.121	23:56
Breath Volume	2.050	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.122	23:59
Breath Volume	1.968	23:59
Air Blank	0.000	23:59
Control Test	0.078	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 09/20/2022 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: VEADER

DATE/TIME OF TEST: 09/23/2022 22:19:52

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.078	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.211	22:15
Breath Volume	2.433	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.209	22:18
Breath Volume	1.839	22:18
Air Blank	0.000	22:18
Control Test	0.078	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 10/27/2022 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: ARRIA-DEVOE

DATE/TIME OF TEST: 09/24/2022 21:20:45

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.165	21:15
Breath Volume	1.792	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.168	21:19
Breath Volume	1.546	21:19
Air Blank	0.000	21:19
Control Test	0.079	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 10/27/2022 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: SCOGGINS

DATE/TIME OF TEST: 09/28/2022 21:19:18

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:16	
Control Test	0.080	21:17	
Air Blank	0.000	21:17	
Subject Sample	# 1	21:17	Subject Test Refused
Breath Volume	0.000	21:17	
Air Blank	0.000	21:18	
Control Test	0.080	21:18	
Air Blank	0.000	21:19	
Diagnostics Check	OK	21:19	

Data Download Date/Time: 10/27/2022 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLY,WILLIAM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/30/2022 02:49:42

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.078	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.166	02:44
Breath Volume	1.214	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.180	02:48
Breath Volume	1.105	02:48
Air Blank	0.000	02:48
Control Test	0.078	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 10/27/2022 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: DEPESTRE,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/03/2022 01:41:07

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.361	01:36
Breath Volume	2.046	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.351	01:39
Breath Volume	1.246	01:39
Air Blank	0.000	01:40
Control Test	0.078	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 10/26/2022 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LOPEZCASTANEDA

DATE/TIME OF TEST: 09/04/2022 01:29:39

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 1 0.116	01:25	
Breath Volume	2.257	01:25	
Air Blank	0.000	01:26	
Air Blank	0.000	01:28	
Subject Sample	# 2	01:28	RFI Detect
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.079	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 10/26/2022 08:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LOPEZCASTANEDA

DATE/TIME OF TEST: 09/04/2022 01:33:21

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
		01:32	RFI Detect
Air Blank	0.000	01:33	

Data Download Date/Time: 10/26/2022 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ, ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEA, ORLADO
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LOPEZCASTANEDA

DATE/TIME OF TEST: 09/04/2022 01:41:43

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.118	01:37
Breath Volume	2.464	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.115	01:40
Breath Volume	2.511	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 10/26/2022 08:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: RISCO,D
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 09/17/2022 03:16:04

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 10/26/2022 08:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: RIVERO,N
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RIVERAFLORES

DATE/TIME OF TEST: 09/30/2022 02:32:56

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.241	02:28
Breath Volume	1.457	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.249	02:31
Breath Volume	1.136	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 10/26/2022 08:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: MAYOTT

DATE/TIME OF TEST: 09/03/2022 02:17:53

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.135	02:13
Breath Volume	2.691	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.139	02:16
Breath Volume	2.742	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 10/13/2022 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CAPORICCI

DATE/TIME OF TEST: 09/03/2022 03:09:44

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.078	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1 0.113	02:56	
Breath Volume	1.390	02:56	
Air Blank	0.000	02:57	
Air Blank	0.000	02:59	
Subject Sample	# 2	03:02	No Sample Provided
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Air Blank	0.000	03:04	
Subject Sample	# 3	03:07	No Sample Provided
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	Fail	03:09	Analytical Stability Test
Air Blank	0.000	03:09	

Data Download Date/Time: 10/13/2022 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CAPORICCI

DATE/TIME OF TEST: 09/03/2022 03:28:54

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.106	03:24
Breath Volume	1.828	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.108	03:27
Breath Volume	1.292	03:27
Air Blank	0.000	03:27
Control Test	0.078	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 10/13/2022 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ROUSE JR

DATE/TIME OF TEST: 09/03/2022 04:02:18

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.193	03:57
Breath Volume	1.468	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.186	04:00
Breath Volume	1.804	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 10/13/2022 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 09/03/2022 05:16:07

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.079	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.217	05:11
Breath Volume	2.800	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.231	05:14
Breath Volume	3.214	05:14
Air Blank	0.000	05:15
Control Test	0.078	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 10/13/2022 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: YAT TOC

DATE/TIME OF TEST: 09/04/2022 02:59:17

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.258	02:54
Breath Volume	1.796	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.257	02:57
Breath Volume	1.394	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 10/13/2022 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: PACURAW

DATE/TIME OF TEST: 09/04/2022 05:35:33

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:29
Subject Sample # 1	0.178	05:31
Breath Volume	1.773	05:31
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.176	05:34
Breath Volume	1.292	05:34
Air Blank	0.000	05:34
Control Test	0.079	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 10/13/2022 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KROUSE,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 09/04/2022 22:59:27

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:44	
Control Test	0.078	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1 0.153	22:48	Slope Not Level
Breath Volume	1.164	22:48	
Air Blank	0.000	22:49	
Air Blank	0.000	22:51	
Subject Sample	# 2 0.142	22:53	
Breath Volume	1.246	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 3 0.147	22:57	
Breath Volume	1.226	22:57	
Air Blank	0.000	22:58	
Control Test	0.077	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 10/13/2022 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: RAMIREZ TREJOS

DATE/TIME OF TEST: 09/06/2022 03:00:55

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.135	02:54
Breath Volume	1.738	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:56
Subject Sample # 2	0.137	02:59
Breath Volume	1.890	02:59
Air Blank	0.000	02:59
Control Test	0.079	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 10/13/2022 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BABB

DATE/TIME OF TEST: 09/06/2022 04:36:39

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.181	04:28
Breath Volume	1.246	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.207	04:31
Breath Volume	1.289	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 3	0.210	04:35
Breath Volume	1.386	04:35
Air Blank	0.000	04:35
Control Test	0.078	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 10/13/2022 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: COMM VEHIC
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 09/10/2022 05:13:28

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.143	05:08
Breath Volume	2.210	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.136	05:11
Breath Volume	1.496	05:11
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: CHAVIANO

DATE/TIME OF TEST: 09/10/2022 23:50:56

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.150	23:46
Breath Volume	1.625	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.143	23:49
Breath Volume	1.847	23:49
Air Blank	0.000	23:50
Control Test	0.078	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: CUBELA

DATE/TIME OF TEST: 09/11/2022 03:56:19

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.191	03:51
Breath Volume	2.097	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.203	03:54
Breath Volume	2.566	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MAX

DATE/TIME OF TEST: 09/11/2022 04:48:02

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.199	04:39
Breath Volume	1.492	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.178	04:42
Breath Volume	1.457	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 3	0.191	04:46
Breath Volume	1.632	04:46
Air Blank	0.000	04:47
Control Test	0.078	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:48

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: CANON

DATE/TIME OF TEST: 09/12/2022 02:49:11

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.083	02:44
Breath Volume	1.203	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.082	02:47
Breath Volume	1.750	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HILL, B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/13/2022 19:10:05

Diagnostics Check	OK	19:03
Air Blank	0.000	19:03
Control Test	0.080	19:04
Air Blank	0.000	19:04
Subject Sample # 1	0.145	19:05
Breath Volume	1.496	19:05
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.147	19:08
Breath Volume	1.359	19:08
Air Blank	0.000	19:09
Control Test	0.079	19:09
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK, I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ARANGO

DATE/TIME OF TEST: 09/14/2022 00:43:41

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.061	00:38
Breath Volume	1.660	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.058	00:42
Breath Volume	1.359	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/16/2022 01:09:16

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.222	01:04
Breath Volume	1.992	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.226	01:07
Breath Volume	1.890	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HARDIN

DATE/TIME OF TEST: 09/16/2022 02:59:37

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1 0.119	02:58	
Breath Volume	2.320	02:58	
Air Blank		02:58	Purge Fail
Air Blank	0.000	02:59	

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HARDIN

DATE/TIME OF TEST: 09/16/2022 03:05:00

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.079	03:02	
Air Blank	0.000	03:03	
Subject Sample # 1	0.116	03:03	
Breath Volume	1.816	03:03	
Air Blank		03:04	Purge Fail
Air Blank	0.000	03:05	

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: NICOSIA

DATE/TIME OF TEST: 09/17/2022 22:32:10

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:21	
Air Blank	0.000	22:21	
Subject Sample	# 1 0.120	22:24	Volume Not Met
Breath Volume	0.531	22:24	
Air Blank	0.000	22:25	
Air Blank	0.000	22:27	
Subject Sample	# 2 0.111	22:30	Volume Not Met
Breath Volume	0.285	22:30	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: NICOSIA

DATE/TIME OF TEST: 09/17/2022 22:46:07

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.078	22:35	
Air Blank	0.000	22:35	
Subject Sample	# 1 0.133	22:38	Volume Not Met
Breath Volume	0.289	22:38	
Air Blank	0.000	22:39	
Air Blank	0.000	22:41	
Subject Sample	# 2 0.140	22:44	Volume Not Met
Breath Volume	0.160	22:44	
Air Blank	0.000	22:45	
Control Test	0.077	22:45	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BARCHAK JR

DATE/TIME OF TEST: 09/18/2022 04:16:01

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.194	04:11
Breath Volume	2.453	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.194	04:14
Breath Volume	1.785	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:16

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOBSON,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 09/18/2022 05:08:51

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.169	05:04
Breath Volume	1.445	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample # 2	0.178	05:07
Breath Volume	1.160	05:07
Air Blank	0.000	05:07
Control Test	0.079	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DORSETT

DATE/TIME OF TEST: 09/22/2022 00:22:12

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.069	00:17
Breath Volume	1.492	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.084	00:20
Breath Volume	2.359	00:20
Air Blank	0.000	00:21
Control Test	0.077	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: FEUER

DATE/TIME OF TEST: 09/23/2022 02:15:48

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.219	02:10
Breath Volume	1.417	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.217	02:14
Breath Volume	1.890	02:14
Air Blank	0.000	02:14
Control Test	0.078	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: BLUE BIRD

DATE/TIME OF TEST: 09/23/2022 20:04:23

Diagnostics Check	OK	19:57
Air Blank	0.000	19:58
Control Test	0.078	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.241	19:59
Breath Volume	2.367	19:59
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.250	20:02
Breath Volume	2.492	20:02
Air Blank	0.000	20:03
Control Test	0.079	20:03
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: ROQUE

DATE/TIME OF TEST: 09/24/2022 04:58:02

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.116	04:53
Breath Volume	2.988	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.108	04:56
Breath Volume	1.628	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:58

Data Download Date/Time: 10/13/2022 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY, JANICE, F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: BARROSO RODRIGUEZ

DATE/TIME OF TEST: 09/24/2022 05:20:35

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.078	05:14
Air Blank	0.000	05:14
Subject Sample # 1	0.103	05:15
Breath Volume	1.390	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.105	05:19
Breath Volume	2.820	05:19
Air Blank	0.000	05:19
Control Test	0.079	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HERNANDEZ GAMEZ

DATE/TIME OF TEST: 09/25/2022 03:08:09

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.000	03:03
Breath Volume	1.347	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.000	03:06
Breath Volume	1.566	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: JAVIER CEDENO

DATE/TIME OF TEST: 09/25/2022 04:23:30

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.131	04:18
Breath Volume	1.230	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.133	04:21
Breath Volume	1.539	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: MARTINEZ JR

DATE/TIME OF TEST: 09/25/2022 22:58:34

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.143	22:54
Breath Volume	3.777	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.145	22:57
Breath Volume	3.667	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAMPIASI,JAKE,L

BREATH TEST OPERATOR AGENCY: FHP Troop K

ARRESTING OFFICER: LAMPIASI,JAKE

ARRESTING OFFICER AGENCY: FHP Troop K

LAST AGENCY INSPECTION: 09/13/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1346146

DRY GAS EXPIRY DATE: 07/07/2023

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MAJAIL

DATE/TIME OF TEST: 09/24/2022 02:10:32

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.132	02:05
Breath Volume	1.605	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.127	02:08
Breath Volume	2.468	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 11/16/2022 15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,JACOB,W
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MEYER,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: SABARI

DATE/TIME OF TEST: 09/05/2022 22:01:17

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.000	21:55
Breath Volume	1.437	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.000	21:59
Breath Volume	1.957	21:59
Air Blank	0.000	22:00
Control Test	0.083	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 09/27/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: EDWARDS,ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/11/2022 23:33:05

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.081	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.161	23:28
Breath Volume	1.523	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.161	23:31
Breath Volume	1.332	23:31
Air Blank	0.000	23:32
Control Test	0.081	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 09/27/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANO,BRANDON,T
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CANO,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 09/20/2022 08:46:16

Diagnostics Check	OK	08:40
Air Blank	0.000	08:40
Control Test	0.080	08:40
Air Blank	0.000	08:41
Subject Sample # 1	0.000	08:41
Breath Volume	3.332	08:41
Air Blank	0.000	08:42
Air Blank	0.000	08:44
Subject Sample # 2	0.000	08:44
Breath Volume	2.859	08:44
Air Blank	0.000	08:45
Control Test	0.083	08:45
Air Blank	0.000	08:46
Diagnostics Check	OK	08:46

Data Download Date/Time: 09/27/2022 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: FARNSWORTH

DATE/TIME OF TEST: 09/02/2022 22:52:32

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.000	22:48
Breath Volume	1.875	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.000	22:51
Breath Volume	1.730	22:51
Air Blank	0.000	22:51
Control Test	0.078	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 09/28/2022 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 09/02/2022 23:02:11

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.080	23:01	
Air Blank		23:01	RFI Detect
Air Blank	0.000	23:02	

Data Download Date/Time: 09/28/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 09/02/2022 23:10:24

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.096	23:05
Breath Volume	3.031	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.097	23:08
Breath Volume	2.289	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 09/28/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: PERAZA

DATE/TIME OF TEST: 09/03/2022 01:16:17

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.177	01:11
Breath Volume	2.980	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.164	01:14
Breath Volume	2.152	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 09/28/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BERGANT

DATE/TIME OF TEST: 09/03/2022 01:46:26

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.000	01:42
Breath Volume	3.187	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.000	01:45
Breath Volume	2.757	01:45
Air Blank	0.000	01:45
Control Test	0.078	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 09/28/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KELLY,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WIMMLER

DATE/TIME OF TEST: 09/03/2022 22:26:46

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.082	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.099	22:22
Breath Volume	3.101	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.096	22:25
Breath Volume	3.148	22:25
Air Blank	0.000	22:25
Control Test	0.080	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 09/28/2022 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GALBREATH

DATE/TIME OF TEST: 09/04/2022 03:49:06

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.243	03:44
Breath Volume	1.812	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.252	03:47
Breath Volume	1.136	03:47
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:49

Data Download Date/Time: 09/28/2022 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO, EDGARDO,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD, CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: KELSEY

DATE/TIME OF TEST: 09/01/2022 03:37:01

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:32	No Sample Provided
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Air Blank	0.000	03:35	
Subject Sample	# 2	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:36	
Control Test	0.079	03:36	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 10/01/2022 04:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BEJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: INOA

DATE/TIME OF TEST: 09/02/2022 05:20:50

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:06	
Control Test	0.080	05:06	
Air Blank	0.000	05:07	
Subject Sample	# 1 0.199	05:08	
Breath Volume	1.292	05:08	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 2 0.177	05:13	Volume Not Met
Breath Volume	0.296	05:13	
Air Blank	0.000	05:14	
Air Blank	0.000	05:16	
Subject Sample	# 3	05:19	No Sample Provided
Breath Volume	0.000	05:19	
Air Blank	0.000	05:19	
Control Test	0.078	05:20	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 10/01/2022 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO, EDGARDO,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD, CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 09/05/2022 23:36:59

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.000	23:32
Breath Volume	2.023	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.000	23:35
Breath Volume	2.367	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 10/01/2022 04:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO, EDGARDO,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD, CARDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: SEAVOY

DATE/TIME OF TEST: 09/07/2022 05:34:15

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:22	
Control Test	0.080	05:22	
Air Blank	0.000	05:23	
Subject Sample	# 1	05:26	No Sample Provided
Breath Volume	0.000	05:26	
Air Blank	0.000	05:26	
Air Blank	0.000	05:28	
Subject Sample	# 2	0.165	05:29
Breath Volume	3.375	05:29	
Air Blank	0.000	05:30	
Air Blank	0.000	05:32	
Subject Sample	# 3	0.160	05:32
Breath Volume	3.011	05:32	
Air Blank	0.000	05:33	
Control Test	0.080	05:33	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 10/01/2022 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO, EDGARDO,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD, CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: BACON

DATE/TIME OF TEST: 09/08/2022 04:32:51

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:20	
Subject Sample # 1	0.061	04:22	
Breath Volume	1.234	04:22	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample # 2	0.053	04:28	Volume Not Met
Breath Volume	0.289	04:28	
Air Blank	0.000	04:28	
Air Blank	0.000	04:30	
Subject Sample # 3	0.066	04:31	
Breath Volume	1.484	04:31	
Air Blank	0.000	04:31	
Control Test	0.079	04:32	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 10/01/2022 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 09/17/2022 00:00:19

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.081	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Control Test	0.080	23:59	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 10/01/2022 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: MITCHELL III

DATE/TIME OF TEST: 09/17/2022 02:27:54

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Control Test	0.081	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 10/01/2022 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIRLING,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TWIDDY,LINDSEY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:11

SUBJECT'S LAST NAME: RIVERA PAZ

DATE/TIME OF TEST: 09/17/2022 10:37:58

Diagnostics Check	OK	10:31
Air Blank	0.000	10:32
Control Test	0.081	10:32
Air Blank	0.000	10:33
Subject Sample # 1	0.209	10:33
Breath Volume	3.285	10:33
Air Blank	0.000	10:34
Air Blank	0.000	10:36
Subject Sample # 2	0.200	10:36
Breath Volume	1.984	10:36
Air Blank	0.000	10:37
Control Test	0.079	10:37
Air Blank	0.000	10:37
Diagnostics Check	OK	10:37

Data Download Date/Time: 10/01/2022 04:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ELLSWORTH

DATE/TIME OF TEST: 09/25/2022 02:43:26

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 10/01/2022 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 09/25/2022 05:39:00

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.080	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.217	05:34
Breath Volume	2.023	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:36
Subject Sample # 2	0.210	05:37
Breath Volume	1.390	05:37
Air Blank	0.000	05:38
Control Test	0.079	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:39

Data Download Date/Time: 10/01/2022 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBURG,DUSTIN,L
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ROSENBURG,DUSTIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: ANDUS-KENNEDY

DATE/TIME OF TEST: 09/25/2022 16:36:37

Diagnostics Check	OK	16:29
Air Blank	0.000	16:29
Control Test	0.081	16:30
Air Blank	0.000	16:30
Subject Sample # 1	0.000	16:31
Breath Volume	1.496	16:31
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.000	16:35
Breath Volume	1.968	16:35
Air Blank	0.000	16:35
Control Test	0.078	16:36
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 10/01/2022 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: LIMARDO

DATE/TIME OF TEST: 09/25/2022 22:18:27

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.111	22:14
Breath Volume	3.621	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.113	22:16
Breath Volume	3.718	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 10/01/2022 04:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIECK,DAVID,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: BIECK,DAVID
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: MELLOTT

DATE/TIME OF TEST: 09/25/2022 22:23:49

Diagnostics Check	OK	22:23	
Air Blank		22:23	RFI Detect
Air Blank	0.000	22:23	

Data Download Date/Time: 10/01/2022 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIECK,DAVID,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: BIECK,DAVID
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: MELLOTT

DATE/TIME OF TEST: 09/25/2022 22:28:00

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.081	22:25	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.080	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 10/01/2022 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 09/26/2022 02:41:36

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.103	02:36
Breath Volume	3.582	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.099	02:40
Breath Volume	2.773	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 10/01/2022 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MCCRANEY

DATE/TIME OF TEST: 09/27/2022 00:46:02

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.000	00:41
Breath Volume	2.320	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.000	00:44
Breath Volume	2.292	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:46

Data Download Date/Time: 10/01/2022 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: LEASUM

DATE/TIME OF TEST: 09/28/2022 01:44:07

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.081	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Control Test	0.081	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 10/01/2022 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SUAZO

DATE/TIME OF TEST: 09/29/2022 00:08:14

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.081	00:06	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Control Test	0.080	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 10/01/2022 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CAIN,DANIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: RAMPY

DATE/TIME OF TEST: 09/29/2022 03:12:01

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.140	03:07
Breath Volume	4.113	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.139	03:10
Breath Volume	2.500	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:12

Data Download Date/Time: 10/01/2022 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIAGET,MISTY,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PIAGET,MISTY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/15/2022 01:49:23

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.000	01:44
Breath Volume	2.070	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.000	01:48
Breath Volume	1.253	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DOWNING

DATE/TIME OF TEST: 09/15/2022 22:36:26

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.079	22:25	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:29	No Sample Provided
Breath Volume	0.000	22:29	
Air Blank	0.000	22:29	
Air Blank	0.000	22:31	
Subject Sample	# 2	22:35	No Sample Provided
Breath Volume	0.000	22:35	
Air Blank	0.000	22:35	
Control Test	0.080	22:35	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MORTON,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 14:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: THORNBURG

DATE/TIME OF TEST: 09/16/2022 17:27:53

Diagnostics Check	OK	17:16	
Air Blank	0.000	17:16	
Control Test	0.080	17:17	
Air Blank	0.000	17:17	
Subject Sample	# 1	17:20	No Sample Provided
Breath Volume	0.000	17:20	
Air Blank	0.000	17:21	
Air Blank	0.000	17:23	
Subject Sample	# 2	17:26	No Sample Provided
Breath Volume	0.000	17:26	
Air Blank	0.000	17:26	
Control Test	0.079	17:27	
Air Blank	0.000	17:27	
Diagnostics Check	OK	17:27	

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: FAY

DATE/TIME OF TEST: 09/17/2022 02:05:29

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.178	02:00
Breath Volume	1.886	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.197	02:03
Breath Volume	1.738	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: MCCAY

DATE/TIME OF TEST: 09/17/2022 04:39:57

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.114	04:35
Breath Volume	1.238	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.119	04:38
Breath Volume	1.257	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: GAGAS

DATE/TIME OF TEST: 09/17/2022 20:24:01

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:16	
Control Test	0.080	20:16	
Air Blank	0.000	20:16	
Subject Sample	# 1	20:20	No Sample Provided
Breath Volume	0.000	20:20	
Air Blank	0.000	20:20	
Air Blank	0.000	20:22	
Subject Sample	# 2	20:22	Subject Test Refused
Breath Volume	0.000	20:22	
Air Blank	0.000	20:23	
Control Test	0.079	20:23	
Air Blank	0.000	20:23	
Diagnostics Check	OK	20:24	

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 09/18/2022 01:27:40

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:23	No Sample Provided
Breath Volume	0.000	01:23	
Air Blank	0.000	01:24	
Air Blank	0.000	01:26	
Subject Sample	# 2	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEWART, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEWART, WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-10
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: STODOLA

DATE/TIME OF TEST: 09/18/2022 03:26:44

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.095	03:22
Breath Volume	1.515	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.093	03:25
Breath Volume	1.863	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINT,PATRICK,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SWINT,PATRICK
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SEALE

DATE/TIME OF TEST: 09/22/2022 03:12:22

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.251	03:07
Breath Volume	1.734	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.254	03:10
Breath Volume	2.707	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CRUZ-MARTINEZ

DATE/TIME OF TEST: 09/23/2022 04:03:49

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.156	03:58
Breath Volume	1.550	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.141	04:02
Breath Volume	1.410	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: NUNEZ BAUTISTA

DATE/TIME OF TEST: 09/24/2022 02:55:10

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.114	02:50
Breath Volume	3.328	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.105	02:53
Breath Volume	1.574	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOONEY,KEVIN,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOONEY,KEVIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: GAMBOA

DATE/TIME OF TEST: 09/24/2022 07:49:39

Diagnostics Check	OK	07:42
Air Blank	0.000	07:42
Control Test	0.080	07:43
Air Blank	0.000	07:43
Subject Sample # 1	0.170	07:44
Breath Volume	3.707	07:44
Air Blank	0.000	07:45
Air Blank	0.000	07:47
Subject Sample # 2	0.169	07:48
Breath Volume	3.675	07:48
Air Blank	0.000	07:48
Control Test	0.079	07:49
Air Blank	0.000	07:49
Diagnostics Check	OK	07:49

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHAGAN,ANDREW,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HACKETT,TRAVIS
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 09/25/2022 05:48:08

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.080	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.057	05:43
Breath Volume	5.222	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.059	05:46
Breath Volume	5.238	05:46
Air Blank	0.000	05:47
Control Test	0.080	05:47
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES, ANTHONY, M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS, JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MURILLO RODRIGUEZ

DATE/TIME OF TEST: 09/26/2022 00:44:37

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.080	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:40	No Sample Provided
Breath Volume	0.000	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 2	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.080	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: CABALLERO HERNANDEZ

DATE/TIME OF TEST: 09/26/2022 22:45:22

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.081	22:33	
Air Blank	0.000	22:33	
Subject Sample	# 1 0.149	22:37	Volume Not Met
Breath Volume	0.179	22:37	
Air Blank	0.000	22:37	
Air Blank	0.000	22:39	
Subject Sample	# 2 0.149	22:40	
Breath Volume	2.164	22:40	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample	# 3 0.142	22:43	
Breath Volume	1.800	22:43	
Air Blank	0.000	22:44	
Control Test	0.079	22:44	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEWART, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEWART, WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: SHOREY

DATE/TIME OF TEST: 09/28/2022 02:16:23

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.081	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 10/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MILLER,JERRY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 09/30/2022 04:11:38

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.153	04:07
Breath Volume	3.675	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.140	04:10
Breath Volume	1.976	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 10/28/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUNLAP,GABRIEL,C

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: SPRY,NICHOLAS

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: DUI

ARREST TIME: 23:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 09/09/2022 00:10:35

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.082	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.167	00:06
Breath Volume	2.828	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.156	00:09
Breath Volume	1.652	00:09
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 09/22/2022 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PETERS,KRISTY,L

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: PETERS,KRISTY

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: DUI

ARREST TIME: 02:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 09/11/2022 03:44:14

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.082	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.048	03:39
Breath Volume	4.132	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.047	03:42
Breath Volume	4.078	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 09/22/2022 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: NEHER

DATE/TIME OF TEST: 09/11/2022 22:59:00

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.078	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.195	22:54
Breath Volume	1.636	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.189	22:57
Breath Volume	2.886	22:57
Air Blank	0.000	22:58
Control Test	0.078	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:59

Data Download Date/Time: 09/24/2022 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 09/24/2022 19:03:07

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.078	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.107	18:58
Breath Volume	1.445	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.109	19:01
Breath Volume	1.453	19:01
Air Blank	0.000	19:02
Control Test	0.079	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 10/23/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ANGELES LOPEZ

DATE/TIME OF TEST: 09/26/2022 03:50:56

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.161	03:46
Breath Volume	2.035	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.149	03:49
Breath Volume	2.058	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 10/23/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: MARCIN

DATE/TIME OF TEST: 09/01/2022 06:29:45

Diagnostics Check	OK	06:23
Air Blank	0.000	06:23
Control Test	0.080	06:23
Air Blank	0.000	06:24
Subject Sample # 1	0.000	06:25
Breath Volume	2.554	06:25
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.000	06:28
Breath Volume	1.953	06:28
Air Blank	0.000	06:28
Control Test	0.080	06:29
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

Data Download Date/Time: 09/01/2022 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: BERNHART

DATE/TIME OF TEST: 09/04/2022 01:17:15

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.024	01:12
Breath Volume	1.289	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.021	01:15
Breath Volume	1.984	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 10/13/2022 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: SOLIS MORENO

DATE/TIME OF TEST: 09/04/2022 05:10:20

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.080	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.207	05:05
Breath Volume	1.578	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.198	05:08
Breath Volume	1.289	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 10/13/2022 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 09/05/2022 03:55:14

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.136	03:49
Breath Volume	1.707	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.137	03:53
Breath Volume	1.589	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 10/13/2022 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: LWE

DATE/TIME OF TEST: 09/11/2022 04:42:21

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:29	
Control Test	0.080	04:29	
Air Blank	0.000	04:29	
Subject Sample	# 1 0.126	04:32	
Breath Volume	1.199	04:32	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 2 0.113	04:37	Volume Not Met
Breath Volume	0.296	04:37	
Air Blank	0.000	04:38	
Air Blank	0.000	04:40	
Subject Sample	# 3 0.134	04:40	
Breath Volume	1.324	04:40	
Air Blank	0.000	04:41	
Control Test	0.078	04:41	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 10/13/2022 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PEREZ CRUZ

DATE/TIME OF TEST: 09/15/2022 03:02:03

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.175	02:57
Breath Volume	2.109	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.173	03:00
Breath Volume	1.660	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:02

Data Download Date/Time: 10/13/2022 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HERNANDEZ-LOPEZ

DATE/TIME OF TEST: 09/17/2022 03:57:54

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.206	03:52
Breath Volume	2.003	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.209	03:56
Breath Volume	2.421	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 10/13/2022 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 09/18/2022 06:35:14

Diagnostics Check	OK	06:27
Air Blank	0.000	06:28
Control Test	0.080	06:28
Air Blank	0.000	06:29
Subject Sample # 1	0.157	06:30
Breath Volume	2.207	06:30
Air Blank	0.000	06:31
Air Blank	0.000	06:33
Subject Sample # 2	0.160	06:33
Breath Volume	1.527	06:33
Air Blank	0.000	06:34
Control Test	0.080	06:34
Air Blank	0.000	06:35
Diagnostics Check	OK	06:35

Data Download Date/Time: 10/13/2022 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: SANTERFEIT

DATE/TIME OF TEST: 09/19/2022 00:21:56

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.166	00:17
Breath Volume	2.460	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.169	00:20
Breath Volume	2.550	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 10/13/2022 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/01/2022 02:01:46

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.082	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.102	01:57
Breath Volume	1.148	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.107	02:00
Breath Volume	1.187	02:00
Air Blank	0.000	02:00
Control Test	0.082	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 09/01/2022 08:01:45

Diagnostics Check	OK	07:55
Air Blank	0.000	07:56
Control Test	0.082	07:56
Air Blank	0.000	07:57
Subject Sample # 1	0.000	07:57
Breath Volume	3.984	07:57
Air Blank	0.000	07:57
Air Blank	0.000	07:59
Subject Sample # 2	0.000	08:00
Breath Volume	3.113	08:00
Air Blank	0.000	08:00
Control Test	0.082	08:01
Air Blank	0.000	08:01
Diagnostics Check	OK	08:01

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARCINDAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: REIN

DATE/TIME OF TEST: 09/01/2022 20:09:59

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.082	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.000	20:03
Breath Volume	2.039	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.000	20:08
Breath Volume	1.562	20:08
Air Blank	0.000	20:09
Control Test	0.081	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: CASTILLO-PEREZ

DATE/TIME OF TEST: 09/01/2022 20:47:46

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.081	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.162	20:43
Breath Volume	2.355	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:45
Subject Sample # 2	0.158	20:46
Breath Volume	1.859	20:46
Air Blank	0.000	20:46
Control Test	0.081	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/02/2022 02:26:11

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:15	
Control Test	0.081	02:15	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:19	No Sample Provided
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Air Blank	0.000	02:21	
Subject Sample	# 2	02:24	No Sample Provided
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.082	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: VIGLUCCI BARTLETT

DATE/TIME OF TEST: 09/03/2022 02:42:39

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.189	02:37
Breath Volume	3.085	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.178	02:41
Breath Volume	2.523	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: CHMURA

DATE/TIME OF TEST: 09/03/2022 04:23:47

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.152	04:19
Breath Volume	1.906	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.144	04:22
Breath Volume	1.621	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/03/2022 19:26:34

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.081	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.069	19:22
Breath Volume	1.863	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.060	19:25
Breath Volume	1.269	19:25
Air Blank	0.000	19:25
Control Test	0.082	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MEREDITH

DATE/TIME OF TEST: 09/04/2022 20:26:39

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:23	
Control Test	0.083	20:24	
Air Blank	0.000	20:24	
Subject Sample	# 1	20:25	Subject Test Refused
Breath Volume	0.000	20:25	
Air Blank	0.000	20:25	
Control Test	0.082	20:26	
Air Blank	0.000	20:26	
Diagnostics Check	OK	20:26	

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 09/06/2022 01:33:38

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.082	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.266	01:29
Breath Volume	2.191	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.275	01:32
Breath Volume	2.375	01:32
Air Blank	0.000	01:32
Control Test	0.079	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TAMAN,KELLEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: GILES

DATE/TIME OF TEST: 09/06/2022 21:15:09

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.082	21:09
Air Blank	0.000	21:09
Subject Sample # 1	0.189	21:10
Breath Volume	1.691	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.178	21:13
Breath Volume	1.539	21:13
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWEIGER,J,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MATTHEWS,J
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: FLANAGAN

DATE/TIME OF TEST: 09/06/2022 22:51:42

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.371	22:47
Breath Volume	1.949	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.380	22:50
Breath Volume	1.863	22:50
Air Blank	0.000	22:50
Control Test	0.081	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWEIGER,J,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TAMAN,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 09/06/2022 23:06:06

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.081	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.253	23:04	
Air Blank	0.000	23:05	
Control Test	0.081	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TAMAN,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 09/06/2022 23:23:16

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:12	
Control Test	0.079	23:12	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:16	No Sample Provided
Breath Volume	0.000	23:16	
Air Blank	0.000	23:16	
Air Blank	0.000	23:18	
Subject Sample	# 2	23:21	No Sample Provided
Breath Volume	0.000	23:21	
Air Blank	0.000	23:22	
Control Test	0.081	23:22	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MADYUN

DATE/TIME OF TEST: 09/07/2022 03:36:35

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.082	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.134	03:32
Breath Volume	2.980	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.127	03:35
Breath Volume	2.511	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD, W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: LADA

DATE/TIME OF TEST: 09/07/2022 09:27:55

Diagnostics Check	OK	09:20
Air Blank	0.000	09:21
Control Test	0.082	09:21
Air Blank	0.000	09:22
Subject Sample # 1	0.013	09:23
Breath Volume	3.667	09:23
Air Blank	0.000	09:24
Air Blank	0.000	09:25
Subject Sample # 2	0.011	09:26
Breath Volume	3.878	09:26
Air Blank	0.000	09:27
Control Test	0.079	09:27
Air Blank	0.000	09:27
Diagnostics Check	OK	09:27

Data Download Date/Time: 09/19/2022 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: COSTON

DATE/TIME OF TEST: 09/07/2022 15:33:32

Diagnostics Check	OK	15:30	
Air Blank	0.000	15:31	
Control Test	0.081	15:31	
Air Blank	0.000	15:32	
Subject Sample	# 1	15:32	Subject Test Refused
Breath Volume	0.000	15:32	
Air Blank	0.000	15:32	
Control Test	0.081	15:32	
Air Blank	0.000	15:33	
Diagnostics Check	OK	15:33	

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 09/07/2022 21:50:40

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.081	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.000	21:46
Breath Volume	2.925	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.000	21:49
Breath Volume	3.210	21:49
Air Blank	0.000	21:49
Control Test	0.081	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRSTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: WHALEN

DATE/TIME OF TEST: 09/09/2022 00:11:36

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:04	No Sample Provided
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2	00:10	No Sample Provided
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Control Test	0.082	00:11	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GRAF

DATE/TIME OF TEST: 09/10/2022 00:56:30

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.082	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.000	00:52
Breath Volume	1.324	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.000	00:55
Breath Volume	1.421	00:55
Air Blank	0.000	00:55
Control Test	0.082	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MANESS

DATE/TIME OF TEST: 09/10/2022 03:35:59

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:29	No Sample Provided
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 2	03:34	No Sample Provided
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Control Test	0.082	03:35	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: LOWRY

DATE/TIME OF TEST: 09/10/2022 14:38:37

Diagnostics Check	OK	14:29	
Air Blank	0.000	14:30	
Control Test	0.081	14:30	
Air Blank	0.000	14:31	
Subject Sample	# 1	14:34	No Sample Provided
Breath Volume	0.000	14:34	
Air Blank	0.000	14:34	
Air Blank	0.000	14:36	
Subject Sample	# 2	14:37	Subject Test Refused
Breath Volume	0.000	14:37	
Air Blank	0.000	14:37	
Control Test	0.081	14:38	
Air Blank	0.000	14:38	
Diagnostics Check	OK	14:38	

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CALIFANO

DATE/TIME OF TEST: 09/11/2022 03:36:50

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.081	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:29	No Sample Provided
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2	03:35	No Sample Provided
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Control Test	0.080	03:36	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,PEREZ
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: CAGLIOTI

DATE/TIME OF TEST: 09/11/2022 04:13:01

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.000	04:08
Breath Volume	1.445	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.000	04:11
Breath Volume	1.734	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:13

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/12/2022 01:47:07

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.133	01:42
Breath Volume	1.503	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.131	01:45
Breath Volume	1.566	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GUNN

DATE/TIME OF TEST: 09/14/2022 02:17:06

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.168	02:12
Breath Volume	4.191	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.178	02:15
Breath Volume	2.742	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 09/15/2022 02:33:44

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:28	
Control Test	0.081	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1 0.185	02:32	Volume Not Met
Breath Volume	0.019	02:32	
Air Blank		02:32	Purge Fail
Air Blank		02:33	Purge Fail

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 09/15/2022 02:36:58

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.054	02:36	Control Outside Tolerance
Air Blank	0.000	02:36	

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 09/15/2022 02:40:00

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:39	
Control Test	0.061	02:39	Control Outside Tolerance
Air Blank	0.000	02:40	

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 09/15/2022 02:42:52

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.066	02:42	Control Outside Tolerance
Air Blank	0.000	02:42	

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISITNA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 09/15/2022 02:56:48

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.066	02:56	Control Outside Tolerance
Air Blank	0.000	02:56	

Data Download Date/Time: 09/19/2022 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ, S
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: FICKLEY

DATE/TIME OF TEST: 09/21/2022 23:42:44

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.224	23:38
Breath Volume	2.546	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.209	23:41
Breath Volume	1.757	23:41
Air Blank	0.000	23:41
Control Test	0.078	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 10/25/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 09/22/2022 05:09:50

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:07	
Control Test	0.080	05:07	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:08	Subject Test Refused
Breath Volume	0.000	05:08	
Air Blank	0.000	05:08	
Control Test	0.080	05:09	
Air Blank	0.000	05:09	
Diagnostics Check	OK	05:09	

Data Download Date/Time: 10/25/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 09/23/2022 00:42:32

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.138	00:36
Breath Volume	1.347	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.147	00:40
Breath Volume	2.152	00:40
Air Blank	0.000	00:41
Control Test	0.081	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 10/25/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FAMILIA ROSARIO

DATE/TIME OF TEST: 09/24/2022 02:10:25

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.081	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.159	02:05
Breath Volume	1.351	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 2	0.159	02:08
Breath Volume	2.089	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 10/25/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SISTI,JORDAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CHAN

DATE/TIME OF TEST: 09/24/2022 02:58:59

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.233	02:54
Breath Volume	1.617	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.233	02:57
Breath Volume	2.691	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 10/25/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MCKINNIE

DATE/TIME OF TEST: 09/24/2022 04:04:17

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.081	03:53	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 2	04:02	No Sample Provided
Breath Volume	0.000	04:02	
Air Blank	0.000	04:03	
Control Test	0.081	04:03	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 10/25/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: BUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: ABRAHAMS

DATE/TIME OF TEST: 09/24/2022 22:18:41

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.081	22:08	
Air Blank	0.000	22:08	
Subject Sample	# 1	22:11	No Sample Provided
Breath Volume	0.000	22:11	
Air Blank	0.000	22:12	
Air Blank	0.000	22:14	
Subject Sample	# 2	22:17	No Sample Provided
Breath Volume	0.000	22:17	
Air Blank	0.000	22:17	
Control Test	0.082	22:18	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 10/25/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 09/25/2022 00:47:57

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.081	00:37	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:40	No Sample Provided
Breath Volume	0.000	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 2	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.080	00:47	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 10/25/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: KOPTA

DATE/TIME OF TEST: 09/25/2022 03:15:39

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.162	03:11
Breath Volume	3.000	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.157	03:14
Breath Volume	3.054	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 10/25/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 09/25/2022 04:24:38

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.081	04:14	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:17	No Sample Provided
Breath Volume	0.000	04:17	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample	# 2	04:23	No Sample Provided
Breath Volume	0.000	04:23	
Air Blank	0.000	04:23	
Control Test	0.082	04:24	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 10/25/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:59

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 09/26/2022 13:37:53

Diagnostics Check	OK	13:31
Air Blank	0.000	13:32
Control Test	0.080	13:32
Air Blank	0.000	13:33
Subject Sample # 1	0.356	13:33
Breath Volume	1.386	13:33
Air Blank	0.000	13:34
Air Blank	0.000	13:35
Subject Sample # 2	0.371	13:36
Breath Volume	2.269	13:36
Air Blank	0.000	13:36
Control Test	0.079	13:37
Air Blank	0.000	13:37
Diagnostics Check	OK	13:37

Data Download Date/Time: 10/25/2022 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS, K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: DEGOUTTE

DATE/TIME OF TEST: 09/26/2022 23:51:07

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.081	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.000	23:46
Breath Volume	2.398	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.000	23:49
Breath Volume	2.324	23:49
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 10/25/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/27/2022 21:19:10

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.081	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.000	21:14
Breath Volume	1.261	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.000	21:17
Breath Volume	1.382	21:17
Air Blank	0.000	21:18
Control Test	0.081	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 10/25/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: FOLDS

DATE/TIME OF TEST: 09/28/2022 05:00:46

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.081	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.259	04:56
Breath Volume	2.921	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.264	04:59
Breath Volume	2.105	04:59
Air Blank	0.000	04:59
Control Test	0.077	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 10/25/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: FAVORS

DATE/TIME OF TEST: 09/29/2022 23:23:54

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.080	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.293	23:19
Breath Volume	1.785	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:21
Subject Sample # 2	0.291	23:22
Breath Volume	1.375	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 10/25/2022 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,S
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: VANCASSELE

DATE/TIME OF TEST: 09/30/2022 00:23:29

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.081	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.187	00:18
Breath Volume	3.023	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.188	00:21
Breath Volume	3.445	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 10/25/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: KIVUVA

DATE/TIME OF TEST: 09/30/2022 01:52:46

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.234	01:48
Breath Volume	1.667	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.241	01:51
Breath Volume	1.511	01:51
Air Blank	0.000	01:51
Control Test	0.081	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 10/25/2022 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: JESPERSEN

DATE/TIME OF TEST: 09/30/2022 03:35:18

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Air Blank	0.000	03:33	
Subject Sample	# 2	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Control Test	0.081	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 10/25/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINS,KYLE,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MARTINS,KYLE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 09/05/2022 00:54:05

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.082	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.247	00:48
Breath Volume	2.929	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.236	00:52
Breath Volume	2.300	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 10/25/2022 08:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON, TORI, D
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON, TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RUSSAW

DATE/TIME OF TEST: 09/10/2022 02:22:17

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.163	02:17
Breath Volume	2.339	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.172	02:20
Breath Volume	2.468	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 10/25/2022 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 09/10/2022 03:41:03

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.082	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.222	03:35
Breath Volume	2.109	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.218	03:39
Breath Volume	2.644	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:41

Data Download Date/Time: 10/25/2022 08:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON, TORI, D
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON, TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 09/11/2022 03:15:19

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:13	
Subject Sample # 1	0.287	03:13	
Breath Volume	3.738	03:13	
Air Blank		03:14	Purge Fail
Air Blank	0.000	03:15	

Data Download Date/Time: 10/25/2022 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON, TORI, D
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON, TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 09/11/2022 03:23:38

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.272	03:18
Breath Volume	2.417	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.264	03:21
Breath Volume	2.312	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 10/25/2022 08:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MAKSIMOVIC, DRAGOSLAV
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 09/11/2022 04:06:43

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.101	04:01
Breath Volume	3.527	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.097	04:04
Breath Volume	2.042	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 10/25/2022 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROSPER,BENSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CORNIER,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: NO

DATE/TIME OF TEST: 09/14/2022 11:01:17

Diagnostics Check	OK	10:58	
Air Blank	0.000	10:58	
Control Test	0.082	10:58	
Air Blank	0.000	10:59	
Subject Sample	# 1 0.313	10:59	
Breath Volume	2.914	10:59	
Air Blank		11:00	Purge Fail
Air Blank	0.000	11:01	

Data Download Date/Time: 10/25/2022 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROSPER,BENSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CORNIER,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: NO

DATE/TIME OF TEST: 09/14/2022 11:08:33

Diagnostics Check	OK	11:05	
Air Blank	0.000	11:06	
Control Test	0.078	11:06	
Air Blank	0.000	11:06	
Subject Sample	# 1 0.307	11:07	
Breath Volume	1.781	11:07	
Air Blank		11:08	Purge Fail
Air Blank	0.000	11:08	

Data Download Date/Time: 10/25/2022 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROSPER,BENSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CORNIER,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: NO

DATE/TIME OF TEST: 09/14/2022 11:20:48

Diagnostics Check	OK	11:12
Air Blank	0.000	11:13
Control Test	0.079	11:13
Air Blank	0.000	11:13
Subject Sample # 1	0.313	11:14
Breath Volume	1.925	11:14
Air Blank	0.000	11:15
Air Blank	0.000	11:17
Subject Sample # 2	0.306	11:18
Breath Volume	2.054	11:18
Air Blank	0.000	11:19
Control Test	0.078	11:20
Air Blank	0.000	11:20
Diagnostics Check	OK	11:20

Data Download Date/Time: 10/25/2022 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WITT,KELLY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: NAVAS LOPEZ

DATE/TIME OF TEST: 09/15/2022 05:50:44

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.082	05:44
Air Blank	0.000	05:45
Subject Sample # 1	0.079	05:46
Breath Volume	3.550	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.068	05:48
Breath Volume	3.003	05:48
Air Blank	0.000	05:49
Control Test	0.081	05:50
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 10/25/2022 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDWIN, DEAN,
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: BLACKARD, RAYMOND
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/22/2022 07:35:02

Diagnostics Check	OK	07:27
Air Blank	0.000	07:27
Control Test	0.080	07:28
Air Blank	0.000	07:28
Subject Sample # 1	0.089	07:29
Breath Volume	2.121	07:29
Air Blank	0.000	07:30
Air Blank	0.000	07:32
Subject Sample # 2	0.082	07:33
Breath Volume	1.757	07:33
Air Blank	0.000	07:34
Control Test	0.080	07:34
Air Blank	0.000	07:34
Diagnostics Check	OK	07:35

Data Download Date/Time: 10/18/2022 17:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: AMATRUDA,CHRIS,G

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: AMATRUDA,CHRIS

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/30/2022 00:07:19

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.163	00:02
Breath Volume	1.117	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.177	00:05
Breath Volume	2.222	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 10/18/2022 17:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 09/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: VALENZUELA

DATE/TIME OF TEST: 09/17/2022 02:05:56

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.250	02:00
Breath Volume	2.472	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.261	02:04
Breath Volume	2.812	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 10/06/2022 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 09/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 09/18/2022 22:51:05

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.216	22:46
Breath Volume	1.804	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.224	22:49
Breath Volume	1.839	22:49
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 10/06/2022 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 09/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ABBATE

DATE/TIME OF TEST: 09/19/2022 03:43:23

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.244	03:34
Breath Volume	1.871	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.221	03:38
Breath Volume	1.609	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:40
Subject Sample # 3	0.231	03:41
Breath Volume	1.371	03:41
Air Blank	0.000	03:42
Control Test	0.077	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 10/06/2022 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: BUI
ARREST TIME: 16:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 09/03/2022 17:52:14

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.076	17:46
Air Blank	0.000	17:47
Subject Sample # 1	0.126	17:47
Breath Volume	2.246	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.120	17:50
Breath Volume	2.296	17:50
Air Blank	0.000	17:51
Control Test	0.077	17:51
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 10/06/2022 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LARROSA,C
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BORGELLA

DATE/TIME OF TEST: 09/02/2022 03:57:19

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.077	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.158	03:52
Breath Volume	2.312	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.154	03:55
Breath Volume	2.320	03:55
Air Blank	0.000	03:56
Control Test	0.077	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 10/06/2022 04:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGLIA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: BOERATO

DATE/TIME OF TEST: 09/02/2022 04:11:57

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.077	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.084	04:07
Breath Volume	1.902	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.081	04:10
Breath Volume	2.578	04:10
Air Blank	0.000	04:10
Control Test	0.078	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 10/06/2022 04:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGLIA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ELHASSAN

DATE/TIME OF TEST: 09/03/2022 04:58:34

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.076	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.207	04:51
Breath Volume	1.312	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.205	04:56
Breath Volume	1.121	04:56
Air Blank	0.000	04:57
Control Test	0.077	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 10/06/2022 04:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: CARCAMO

DATE/TIME OF TEST: 09/03/2022 18:19:50

Diagnostics Check	OK	18:14	
Air Blank	0.000	18:15	
Control Test	0.075	18:15	
Air Blank	0.000	18:16	
Subject Sample	# 1	18:18	Subject Test Refused
Breath Volume	0.000	18:18	
Air Blank	0.000	18:18	
Control Test	0.077	18:19	
Air Blank	0.000	18:19	
Diagnostics Check	OK	18:19	

Data Download Date/Time: 10/06/2022 04:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGLIA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: SOMAROO

DATE/TIME OF TEST: 09/11/2022 01:03:00

Diagnostics Check	OK	00:59	
Air Blank	0.000	01:00	
Control Test	0.076	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.076	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 10/06/2022 04:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HINESTROZA

DATE/TIME OF TEST: 09/16/2022 00:54:52

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.077	00:50	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:53	
Control Test	0.078	00:54	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time: 10/06/2022 04:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTALGIA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: PICON DELGADO

DATE/TIME OF TEST: 09/16/2022 03:36:10

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.077	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.234	03:31
Breath Volume	2.335	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.235	03:34
Breath Volume	2.046	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 10/06/2022 04:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 09/17/2022 05:21:21

Diagnostics Check	OK	05:14
Air Blank	0.000	05:15
Control Test	0.078	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.148	05:16
Breath Volume	3.214	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:18
Subject Sample # 2	0.148	05:19
Breath Volume	3.382	05:19
Air Blank	0.000	05:20
Control Test	0.077	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 10/06/2022 04:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: DUQUE-RAMIREZ

DATE/TIME OF TEST: 09/02/2022 04:15:45

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.228	04:11
Breath Volume	4.492	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.215	04:14
Breath Volume	4.921	04:14
Air Blank	0.000	04:14
Control Test	0.077	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 10/06/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JULIO,C
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/11/2022 02:14:16

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.163	02:09
Breath Volume	1.179	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.162	02:12
Breath Volume	1.980	02:12
Air Blank	0.000	02:13
Control Test	0.077	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 10/06/2022 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRON,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,BRANDON
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: LOPEZ JR

DATE/TIME OF TEST: 09/04/2022 03:27:27

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.082	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.081	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 09/30/2022 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRON,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VAZQUEZ,RAUL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: ALBINO-QUILES

DATE/TIME OF TEST: 09/16/2022 22:09:47

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.081	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.109	22:05
Breath Volume	3.023	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:07
Subject Sample # 2	0.112	22:08
Breath Volume	3.007	22:08
Air Blank	0.000	22:08
Control Test	0.079	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 09/30/2022 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NERSTEN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: LOACH

DATE/TIME OF TEST: 09/08/2022 21:19:07

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:13
Subject Sample # 1	0.108	21:14
Breath Volume	1.238	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:16
Subject Sample # 2	0.109	21:17
Breath Volume	1.667	21:17
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 09/24/2022 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 09/13/2022 19:44:30

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.078	19:38
Air Blank	0.000	19:39
Subject Sample # 1	0.124	19:39
Breath Volume	1.949	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.122	19:42
Breath Volume	2.277	19:42
Air Blank	0.000	19:43
Control Test	0.078	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 09/24/2022 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: BURGOS TAPIA

DATE/TIME OF TEST: 09/02/2022 21:32:28

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.078	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.139	21:28
Breath Volume	2.937	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.131	21:30
Breath Volume	1.660	21:30
Air Blank	0.000	21:31
Control Test	0.076	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 09/16/2022 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DILLARD, DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 09/03/2022 00:26:41

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.176	00:22
Breath Volume	1.816	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.172	00:25
Breath Volume	1.792	00:25
Air Blank	0.000	00:25
Control Test	0.078	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 09/16/2022 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BRAMS

DATE/TIME OF TEST: 09/02/2022 21:50:09

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:45
Subject Sample # 1	0.000	21:45
Breath Volume	2.324	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.000	21:48
Breath Volume	3.046	21:48
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FESTA, JESSI
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/02/2022 23:23:44

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.081	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.116	23:19
Breath Volume	2.261	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.113	23:22
Breath Volume	1.601	23:22
Air Blank	0.000	23:22
Control Test	0.078	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS, VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: MINUTELLO

DATE/TIME OF TEST: 09/03/2022 00:00:49

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Subject Sample # 1	0.000	23:56
Breath Volume	1.238	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.000	23:59
Breath Volume	1.640	23:59
Air Blank	0.000	23:59
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PICARD, RONNIE
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: GODINEZ RAMIREZ

DATE/TIME OF TEST: 09/03/2022 00:45:15

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.161	00:40
Breath Volume	2.429	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.154	00:43
Breath Volume	2.621	00:43
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: LORA MARTINEZ

DATE/TIME OF TEST: 09/03/2022 02:17:35

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.000	02:13
Breath Volume	1.277	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.000	02:16
Breath Volume	2.289	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ALLESSADRINE

DATE/TIME OF TEST: 09/03/2022 03:23:01

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.000	03:18
Breath Volume	2.203	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.000	03:21
Breath Volume	3.046	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:23

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COSTALES, ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: GASPAR PEDRO

DATE/TIME OF TEST: 09/03/2022 06:52:59

Diagnostics Check	OK	06:46
Air Blank	0.000	06:46
Control Test	0.080	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.163	06:48
Breath Volume	1.238	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:50
Subject Sample # 2	0.169	06:51
Breath Volume	2.609	06:51
Air Blank	0.000	06:52
Control Test	0.079	06:52
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,REN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: FELIX JUAN

DATE/TIME OF TEST: 09/03/2022 22:19:42

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.167	22:15
Breath Volume	1.160	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.167	22:18
Breath Volume	1.375	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY, STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/05/2022 19:25:05

Diagnostics Check	OK	19:18
Air Blank	0.000	19:19
Control Test	0.081	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.000	19:20
Breath Volume	1.199	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.000	19:23
Breath Volume	2.015	19:23
Air Blank	0.000	19:24
Control Test	0.080	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORGAN, DAVID
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: OKKO

DATE/TIME OF TEST: 09/05/2022 19:45:30

Diagnostics Check	OK	19:38
Air Blank	0.000	19:39
Control Test	0.079	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.127	19:40
Breath Volume	3.179	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.125	19:43
Breath Volume	3.125	19:43
Air Blank	0.000	19:44
Control Test	0.078	19:44
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: JUAN PASCUAL

DATE/TIME OF TEST: 09/07/2022 00:14:24

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.148	00:09
Breath Volume	2.546	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.147	00:12
Breath Volume	3.457	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MATONTI, JOHN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: SHABER

DATE/TIME OF TEST: 09/07/2022 06:11:06

Diagnostics Check	OK	06:04
Air Blank	0.000	06:05
Control Test	0.081	06:05
Air Blank	0.000	06:06
Subject Sample # 1	0.183	06:06
Breath Volume	2.105	06:06
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample # 2	0.174	06:09
Breath Volume	1.144	06:09
Air Blank	0.000	06:10
Control Test	0.081	06:10
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 09/07/2022 21:39:41

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.239	21:35
Breath Volume	2.558	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.235	21:38
Breath Volume	1.460	21:38
Air Blank	0.000	21:38
Control Test	0.078	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HASSETT

DATE/TIME OF TEST: 09/07/2022 23:10:48

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.139	23:06
Breath Volume	1.230	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.143	23:09
Breath Volume	2.273	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BEATO

DATE/TIME OF TEST: 09/08/2022 00:07:37

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.080	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.210	00:02
Breath Volume	2.109	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample # 2	0.210	00:05
Breath Volume	1.953	00:05
Air Blank	0.000	00:06
Control Test	0.078	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: GRETCHYN

DATE/TIME OF TEST: 09/08/2022 01:06:13

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.000	01:01
Breath Volume	1.738	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.000	01:04
Breath Volume	1.667	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: LLERAS

DATE/TIME OF TEST: 09/08/2022 01:55:32

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.129	01:51
Breath Volume	2.437	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.118	01:53
Breath Volume	2.226	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: HERNANDEZ MORALES

DATE/TIME OF TEST: 09/09/2022 05:57:47

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.080	05:51
Air Blank	0.000	05:52
Subject Sample # 1	0.265	05:52
Breath Volume	1.738	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.255	05:56
Breath Volume	2.535	05:56
Air Blank	0.000	05:56
Control Test	0.080	05:57
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PRIETO

DATE/TIME OF TEST: 09/10/2022 02:00:37

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.173	01:55
Breath Volume	1.250	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:57
Subject Sample # 2	0.180	01:58
Breath Volume	1.464	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SIMONSON

DATE/TIME OF TEST: 09/10/2022 03:37:24

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.062	03:32
Breath Volume	1.351	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.058	03:35
Breath Volume	1.691	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 09/10/2022 05:46:28

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.080	05:41
Air Blank	0.000	05:41
Subject Sample # 1	0.157	05:42
Breath Volume	1.765	05:42
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.156	05:44
Breath Volume	2.238	05:44
Air Blank	0.000	05:45
Control Test	0.080	05:45
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MULLEN,GREGORY
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: HUVER RICCI

DATE/TIME OF TEST: 09/10/2022 13:00:29

Diagnostics Check	OK	12:53
Air Blank	0.000	12:53
Control Test	0.080	12:54
Air Blank	0.000	12:54
Subject Sample # 1	0.255	12:55
Breath Volume	1.625	12:55
Air Blank	0.000	12:56
Air Blank	0.000	12:58
Subject Sample # 2	0.252	12:58
Breath Volume	1.804	12:58
Air Blank	0.000	12:59
Control Test	0.079	12:59
Air Blank	0.000	13:00
Diagnostics Check	OK	13:00

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: FELIPE ACEYTUNO

DATE/TIME OF TEST: 09/11/2022 00:13:42

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.141	00:09
Breath Volume	2.328	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.133	00:12
Breath Volume	2.210	00:12
Air Blank	0.000	00:12
Control Test	0.078	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARULLO, D
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: REILLY

DATE/TIME OF TEST: 09/11/2022 01:06:10

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.000	01:01
Breath Volume	1.968	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.000	01:04
Breath Volume	1.781	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ADOLPHE, PATRICK
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 09/11/2022 03:17:04

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.181	03:12
Breath Volume	1.933	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.167	03:15
Breath Volume	1.730	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:17

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WOOD,DERRICK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: PERECHU SOHOM

DATE/TIME OF TEST: 09/11/2022 13:49:42

Diagnostics Check	OK	13:42
Air Blank	0.000	13:42
Control Test	0.080	13:42
Air Blank	0.000	13:43
Subject Sample # 1	0.167	13:45
Breath Volume	1.199	13:45
Air Blank	0.000	13:46
Air Blank	0.000	13:47
Subject Sample # 2	0.175	13:48
Breath Volume	1.281	13:48
Air Blank	0.000	13:48
Control Test	0.077	13:49
Air Blank	0.000	13:49
Diagnostics Check	OK	13:49

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,JOSE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: PASTOR OXLAJ

DATE/TIME OF TEST: 09/11/2022 18:49:55

Diagnostics Check	OK	18:42
Air Blank	0.000	18:43
Control Test	0.080	18:43
Air Blank	0.000	18:44
Subject Sample # 1	0.210	18:45
Breath Volume	1.675	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:47
Subject Sample # 2	0.199	18:48
Breath Volume	1.464	18:48
Air Blank	0.000	18:48
Control Test	0.079	18:49
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: GIFFELS

DATE/TIME OF TEST: 09/12/2022 01:01:00

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.093	00:56
Breath Volume	2.089	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.090	00:59
Breath Volume	2.167	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:01

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MICHAUD

DATE/TIME OF TEST: 09/12/2022 01:47:38

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.267	01:42
Breath Volume	2.210	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.281	01:45
Breath Volume	1.832	01:45
Air Blank	0.000	01:46
Control Test	0.078	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PICARD, RONNIE
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MORALES JORGE

DATE/TIME OF TEST: 09/12/2022 03:57:24

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.244	03:52
Breath Volume	1.792	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.248	03:55
Breath Volume	1.937	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HEWITT,SHANE
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: PLESCA

DATE/TIME OF TEST: 09/12/2022 22:24:00

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.000	22:19
Breath Volume	2.933	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.000	22:22
Breath Volume	3.496	22:22
Air Blank	0.000	22:23
Control Test	0.079	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:24

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: VUSTILLOS

DATE/TIME OF TEST: 09/13/2022 01:01:49

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.139	00:56
Breath Volume	1.277	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.139	01:00
Breath Volume	1.660	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 09/13/2022 03:50:57

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.160	03:46
Breath Volume	2.144	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.156	03:49
Breath Volume	2.023	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LEITON

DATE/TIME OF TEST: 09/13/2022 04:13:26

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.193	04:08
Breath Volume	1.980	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.182	04:11
Breath Volume	1.992	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: BARTOLON CAX

DATE/TIME OF TEST: 09/13/2022 20:53:15

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:47
Subject Sample # 1	0.159	20:48
Breath Volume	2.015	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.154	20:51
Breath Volume	2.457	20:51
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: GUETHLER

DATE/TIME OF TEST: 09/13/2022 22:25:20

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.081	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.189	22:20
Breath Volume	2.042	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.209	22:23
Breath Volume	2.464	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 09/14/2022 03:57:16

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.081	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1 0.184	03:49	Volume Not Met
Breath Volume	0.125	03:49	
Air Blank	0.000	03:50	
Air Blank	0.000	03:52	
Subject Sample	# 2 0.192	03:55	Volume Not Met
Breath Volume	0.886	03:55	
Air Blank	0.000	03:56	
Control Test	0.078	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/14/2022 20:54:27

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:48	
Control Test	0.081	20:48	
Air Blank	0.000	20:49	
Subject Sample	# 1 0.251	20:49	
Breath Volume	2.320	20:49	
Air Blank	0.000	20:50	
Air Blank	0.000	20:52	
Subject Sample	# 2	20:52	Interferent Detect
Breath Volume	2.027	20:52	
Air Blank	0.000	20:53	
Control Test	0.079	20:53	
Air Blank	0.000	20:54	
Diagnostics Check	OK	20:54	

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/14/2022 21:20:12

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.227	21:15
Breath Volume	2.289	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:17
Subject Sample # 2	0.230	21:18
Breath Volume	2.488	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: HAWTHORN

DATE/TIME OF TEST: 09/15/2022 01:46:59

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.000	01:42
Breath Volume	1.851	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.000	01:45
Breath Volume	2.183	01:45
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 09/15/2022 04:40:19

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.182	04:35
Breath Volume	3.917	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.181	04:38
Breath Volume	4.304	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 09/16/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: GILBERTO

DATE/TIME OF TEST: 09/16/2022 22:55:10

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.081	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.112	22:50
Breath Volume	2.792	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.113	22:53
Breath Volume	2.820	22:53
Air Blank	0.000	22:54
Control Test	0.079	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PICARD,RONNIE
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: ZUCKERBROD

DATE/TIME OF TEST: 09/17/2022 01:39:25

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.000	01:35
Breath Volume	1.250	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.000	01:38
Breath Volume	2.417	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 09/17/2022 04:47:23

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.165	04:42
Breath Volume	1.566	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.165	04:45
Breath Volume	1.683	04:45
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MCNAMEE JR

DATE/TIME OF TEST: 09/19/2022 00:27:35

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.081	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.180	00:23
Breath Volume	1.660	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.180	00:25
Breath Volume	2.304	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BURCHELL

DATE/TIME OF TEST: 09/20/2022 03:31:16

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.187	03:26
Breath Volume	1.785	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.186	03:29
Breath Volume	1.609	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: RAHME

DATE/TIME OF TEST: 09/20/2022 23:33:24

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.081	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.000	23:29
Breath Volume	1.968	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.000	23:32
Breath Volume	2.164	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEONARDI,KAILEY
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: DIGESU

DATE/TIME OF TEST: 09/22/2022 00:12:47

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.081	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.000	00:08
Breath Volume	2.097	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.000	00:11
Breath Volume	2.105	00:11
Air Blank	0.000	00:11
Control Test	0.081	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: DIAZ MALDONADO

DATE/TIME OF TEST: 09/22/2022 04:26:44

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.082	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.204	04:21
Breath Volume	2.214	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:23
Subject Sample # 2	0.201	04:25
Breath Volume	2.597	04:25
Air Blank	0.000	04:25
Control Test	0.081	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,KERWIN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 09/22/2022 22:24:20

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.164	22:19
Breath Volume	2.070	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:21
Subject Sample # 2	0.166	22:22
Breath Volume	1.867	22:22
Air Blank	0.000	22:23
Control Test	0.079	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: DIAZ SOTO

DATE/TIME OF TEST: 09/22/2022 23:41:44

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.081	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.162	23:36
Breath Volume	1.386	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.153	23:40
Breath Volume	1.609	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: SANCHEZ MORALES

DATE/TIME OF TEST: 09/23/2022 03:55:24

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.278	03:49
Breath Volume	1.539	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.271	03:53
Breath Volume	1.527	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: RAMIREZ MORALES

DATE/TIME OF TEST: 09/23/2022 22:15:42

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.081	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.333	22:11
Breath Volume	1.507	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.323	22:14
Breath Volume	1.851	22:14
Air Blank	0.000	22:14
Control Test	0.079	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RODRIGUEZ,ELIAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: AGUSTIN-MORALES

DATE/TIME OF TEST: 09/24/2022 01:48:02

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.191	01:43
Breath Volume	1.804	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample # 2	0.180	01:46
Breath Volume	2.335	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:48

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,KERWIN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 09/24/2022 04:31:39

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.106	04:27
Breath Volume	1.414	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.109	04:30
Breath Volume	1.789	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CAMPOSECO ROS

DATE/TIME OF TEST: 09/30/2022 01:28:56

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.081	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.163	01:24
Breath Volume	1.214	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.164	01:27
Breath Volume	2.535	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: SORKNESS

DATE/TIME OF TEST: 09/30/2022 05:46:52

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.081	05:41
Air Blank	0.000	05:41
Subject Sample # 1	0.137	05:42
Breath Volume	2.703	05:42
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.144	05:45
Breath Volume	3.367	05:45
Air Blank	0.000	05:45
Control Test	0.081	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: GEE

DATE/TIME OF TEST: 09/03/2022 01:09:55

Diagnostics Check	OK	01:08	
Air Blank		01:09	Ambient Fail
Air Blank		01:09	Purge Fail

Data Download Date/Time: 09/13/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: GEE

DATE/TIME OF TEST: 09/03/2022 01:17:41

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.000	01:13
Breath Volume	1.355	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.000	01:16
Breath Volume	1.570	01:16
Air Blank	0.000	01:16
Control Test	0.077	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 09/13/2022 15:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/04/2022 21:38:17

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.079	21:31	
Air Blank	0.000	21:32	
Subject Sample # 1	0.192	21:33	
Breath Volume	1.488	21:33	
Air Blank	0.000	21:33	
Air Blank	0.000	21:35	
Subject Sample # 2	0.182	21:36	
Breath Volume	1.808	21:36	
Air Blank	0.000	21:36	
Control Test	0.079	21:37	
Air Blank	0.000	21:37	
Diagnostics Check	Fail	21:37	Analytical Stability Test
Air Blank	0.000	21:38	

Data Download Date/Time: 09/13/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ORTIZ-DIAZ

DATE/TIME OF TEST: 09/10/2022 01:35:57

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.079	01:29	
Air Blank	0.000	01:30	
Subject Sample # 1	0.095	01:31	
Breath Volume	2.179	01:31	
Air Blank	0.000	01:31	
Air Blank	0.000	01:33	
Subject Sample # 2	0.100	01:34	
Breath Volume	2.992	01:34	
Air Blank		01:34	Purge Fail
Air Blank	0.000	01:35	

Data Download Date/Time: 09/13/2022 15:41