

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: PENA,MARISSA
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: RATHBURN

DATE/TIME OF TEST: 04/27/2023 18:22:09

Diagnostics Check	OK	18:10
Air Blank	0.000	18:11
Control Test	0.080	18:11
Air Blank	0.000	18:12
Subject Sample # 1	0.325	18:13
Breath Volume	2.308	18:13
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.297	18:16
Breath Volume	2.921	18:16
Air Blank	0.000	18:17
Air Blank	0.000	18:19
Subject Sample # 3	0.287	18:20
Breath Volume	3.812	18:20
Air Blank	0.000	18:21
Control Test	0.079	18:21
Air Blank	0.000	18:22
Diagnostics Check	OK	18:22

Data Download Date/Time: 05/04/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,JAVON,C
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LIVINGSTON,JAVON
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 04/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 04/30/2023 22:33:57

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.000	22:29
Breath Volume	3.406	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.000	22:32
Breath Volume	3.562	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 05/04/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,L
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: GRAHAM,EMMA
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BOOS

DATE/TIME OF TEST: 04/08/2023 00:52:33

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.078	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.089	00:47
Breath Volume	2.300	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 2	0.085	00:50
Breath Volume	3.160	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 05/01/2023 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: KAZEK,C
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 03/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: ACCORSO

DATE/TIME OF TEST: 04/21/2023 21:41:11

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:38	
Control Test	0.078	21:38	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:39	Subject Test Refused
Breath Volume	0.000	21:39	
Air Blank	0.000	21:40	
Control Test	0.079	21:40	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:41	

Data Download Date/Time: 05/01/2023 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: GRAHAM,EMMA
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/23/2023 05:38:38

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.079	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.000	05:34
Breath Volume	1.851	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.000	05:37
Breath Volume	1.953	05:37
Air Blank	0.000	05:37
Control Test	0.078	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 05/01/2023 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEWIS, BREEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 04/02/2023 05:49:55

Diagnostics Check	OK	05:43
Air Blank	0.000	05:43
Control Test	0.077	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.000	05:45
Breath Volume	2.238	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:48
Subject Sample # 2	0.000	05:48
Breath Volume	2.101	05:48
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOUNTAIN,DALE,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FOUNTAIN,DALE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: KENT

DATE/TIME OF TEST: 04/05/2023 02:37:51

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.078	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1 0.144	02:25	
Breath Volume	1.636	02:25	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2 0.140	02:30	Volume Not Met
Breath Volume	0.273	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 3 0.137	02:36	Volume Not Met
Breath Volume	0.015	02:36	
Air Blank	0.000	02:36	
Control Test	0.078	02:37	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: NOLET

DATE/TIME OF TEST: 04/06/2023 20:14:00

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:02	
Control Test	0.079	20:03	
Air Blank	0.000	20:03	
Subject Sample	# 1 0.243	20:06	Volume Not Met
Breath Volume	0.230	20:06	
Air Blank	0.000	20:07	
Air Blank	0.000	20:09	
Subject Sample	# 2 0.212	20:12	Volume Not Met
Breath Volume	0.261	20:12	
Air Blank	0.000	20:13	
Control Test	0.079	20:13	
Air Blank	0.000	20:13	
Diagnostics Check	OK	20:14	

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACHELLER,ALEXANDRIA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BACHELLER,ALEXANDRIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ROS-HERNANDEZ

DATE/TIME OF TEST: 04/08/2023 01:11:15

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.080	00:58	
Air Blank	0.000	00:59	
Subject Sample	# 1 0.178	01:02	Volume Not Met
Breath Volume	0.156	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:04	
Subject Sample	# 2 0.183	01:06	
Breath Volume	2.269	01:06	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample	# 3 0.176	01:09	
Breath Volume	2.250	01:09	
Air Blank	0.000	01:10	
Control Test	0.077	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACHELLER,ALEXANDRIA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ARTUS,WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 04/08/2023 02:32:05

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.079	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1 0.303	02:27	
Breath Volume	2.953	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample	# 2 0.290	02:30	
Breath Volume	2.308	02:30	
Air Blank	0.000	02:31	
Control Test	0.073	02:31	Control Outside Tolerance
Air Blank	0.000	02:32	

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HARTSELLE, AMBER
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: ROCHA

DATE/TIME OF TEST: 04/10/2023 22:48:34

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.097	22:43
Breath Volume	1.250	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.100	22:47
Breath Volume	1.269	22:47
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HARTSELLE, AMBER
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BLACKWELL

DATE/TIME OF TEST: 04/12/2023 00:10:16

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.159	00:05
Breath Volume	2.148	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:07
Subject Sample # 2	0.161	00:08
Breath Volume	2.792	00:08
Air Blank	0.000	00:09
Control Test	0.078	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: AMMERMAN,ERIC
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/16/2023 06:28:25

Diagnostics Check	OK	06:22	
Air Blank	0.000	06:22	
Control Test	0.078	06:23	
Air Blank	0.000	06:23	
Subject Sample # 1	0.135	06:24	
Breath Volume	2.800	06:24	
Air Blank	0.000	06:24	
Air Blank	0.000	06:26	
Subject Sample # 2	0.136	06:26	
Breath Volume	2.679	06:26	
Air Blank	0.000	06:27	
Control Test	0.073	06:27	Control Outside Tolerance
Air Blank	0.000	06:28	

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: AMMERMAN,ERIC
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/16/2023 06:47:18

Diagnostics Check	OK	06:41
Air Blank	0.000	06:41
Control Test	0.078	06:41
Air Blank	0.000	06:42
Subject Sample # 1	0.129	06:43
Breath Volume	2.574	06:43
Air Blank	0.000	06:43
Air Blank	0.000	06:45
Subject Sample # 2	0.128	06:45
Breath Volume	2.402	06:45
Air Blank	0.000	06:46
Control Test	0.077	06:46
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACHELLER,ALEXANDRIA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: PORTOCARRERO,CHRISTIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 04/17/2023 23:21:52

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.079	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.099	23:17
Breath Volume	1.253	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.100	23:20
Breath Volume	1.398	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACHELLER,ALEXANDRIA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SCHOENER,ANDREW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BURK

DATE/TIME OF TEST: 04/18/2023 00:19:33

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.160	00:14
Breath Volume	2.363	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.158	00:18
Breath Volume	2.167	00:18
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACHELLER,ALEXANDRIA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BACHELLER,ALEXANDRIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/19/2023 04:41:12

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.078	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.129	04:36
Breath Volume	1.957	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.131	04:39
Breath Volume	1.890	04:39
Air Blank	0.000	04:40
Control Test	0.077	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HARTSELLE, AMBER
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 04/21/2023 01:29:03

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.164	01:24
Breath Volume	2.167	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.154	01:27
Breath Volume	2.007	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:29

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCINI, JOSHUA, M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DULIN, H
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 04/22/2023 13:55:16

Diagnostics Check	OK	13:48
Air Blank	0.000	13:49
Control Test	0.079	13:49
Air Blank	0.000	13:49
Subject Sample # 1	0.059	13:50
Breath Volume	3.214	13:50
Air Blank	0.000	13:51
Air Blank	0.000	13:53
Subject Sample # 2	0.057	13:53
Breath Volume	3.484	13:53
Air Blank	0.000	13:54
Control Test	0.079	13:54
Air Blank	0.000	13:55
Diagnostics Check	OK	13:55

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACHELLER,ALEXANDRIA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BACHELLER,ALEXANDRIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: STICKEL

DATE/TIME OF TEST: 04/24/2023 02:51:25

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.000	02:47
Breath Volume	2.476	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.000	02:50
Breath Volume	2.445	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 04/28/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,JESSE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/28/2023 22:12:03

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.292	22:06
Breath Volume	3.976	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.302	22:10
Breath Volume	3.535	22:10
Air Blank	0.000	22:11
Control Test	0.078	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:12

Data Download Date/Time: 05/25/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 04/30/2023 05:55:56

Diagnostics Check	OK	05:55	
Air Blank		05:55	RFI Detect
Air Blank	0.000	05:55	

Data Download Date/Time: 05/25/2023 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WIGGINS, TYLER, D

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: WIGGINS, TYLER

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 04:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 04/30/2023 06:07:35

Diagnostics Check	OK	05:57
Air Blank	0.000	05:58
Control Test	0.080	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.145	05:59
Breath Volume	1.453	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.175	06:03
Breath Volume	2.097	06:03
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 3	0.146	06:06
Breath Volume	1.882	06:06
Air Blank	0.000	06:06
Control Test	0.079	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 05/25/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GARDNER, TREVOR
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 04/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/30/2023 22:59:13

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:48	
Control Test	0.079	22:48	
Air Blank	0.000	22:48	
Subject Sample	# 1 0.105	22:52	Volume Not Met
Breath Volume	0.332	22:52	
Air Blank	0.000	22:52	
Air Blank	0.000	22:54	
Subject Sample	# 2 0.092	22:57	Volume Not Met
Breath Volume	0.480	22:57	
Air Blank	0.000	22:58	
Control Test	0.078	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 05/25/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/21/2023 22:06:51

Diagnostics Check	OK	21:52	
Air Blank	0.000	21:52	
Control Test	0.079	21:53	
Air Blank	0.000	21:53	
Subject Sample # 1	0.037	21:54	
Breath Volume	2.988	21:54	
Air Blank	0.000	21:54	
Air Blank	0.000	21:56	
Subject Sample # 2		21:59	No Sample Provided
Breath Volume	0.000	21:59	
Air Blank	0.000	22:00	
Air Blank	0.000	22:02	
Subject Sample # 3		22:05	No Sample Provided
Breath Volume	0.000	22:05	
Air Blank	0.000	22:05	
Control Test	0.079	22:06	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 05/10/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/21/2023 22:22:15

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.027	22:17
Breath Volume	2.269	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.025	22:20
Breath Volume	1.886	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 05/10/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 04/22/2023 00:36:25

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:34	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.079	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 05/10/2023 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/22/2023 01:07:35

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.198	01:00
Breath Volume	2.324	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.173	01:03
Breath Volume	1.253	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 3	0.173	01:06
Breath Volume	1.214	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 05/10/2023 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 04/22/2023 03:28:51

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Control Test	0.079	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 05/10/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: OHORO

DATE/TIME OF TEST: 04/28/2023 23:31:56

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.000	23:27
Breath Volume	2.152	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.000	23:30
Breath Volume	1.113	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 05/10/2023 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,RANDI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 04/29/2023 00:56:15

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
Control Test	0.080	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 05/10/2023 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 04/29/2023 01:30:45

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.079	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1 0.201	01:19	Volume Not Met
Breath Volume	0.507	01:19	
Air Blank	0.000	01:20	
Air Blank	0.000	01:22	
Subject Sample	# 2 0.190	01:24	
Breath Volume	1.746	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 3 0.187	01:29	
Breath Volume	1.507	01:29	
Air Blank	0.000	01:29	
Control Test	0.078	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 05/10/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: RANUS

DATE/TIME OF TEST: 04/30/2023 03:45:20

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.275	03:40
Breath Volume	2.367	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.268	03:43
Breath Volume	2.347	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 05/10/2023 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: MCINTIRE SCHOMER

DATE/TIME OF TEST: 04/08/2023 22:16:47

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.079	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.000	22:12
Breath Volume	2.703	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.000	22:15
Breath Volume	1.871	22:15
Air Blank	0.000	22:15
Control Test	0.078	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 04/29/2023 19:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CAMERON,M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: LABOY,LONITA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: NATER

DATE/TIME OF TEST: 04/11/2023 18:13:23

Diagnostics Check	OK	18:06
Air Blank	0.000	18:06
Control Test	0.080	18:06
Air Blank	0.000	18:07
Subject Sample # 1	0.166	18:08
Breath Volume	3.539	18:08
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.165	18:11
Breath Volume	3.156	18:11
Air Blank	0.000	18:12
Control Test	0.077	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 04/29/2023 19:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORROW, ADAM, R
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: FEARNOW, SAMANTHA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 04/28/2023 23:01:06

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.079	22:58	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 04/29/2023 19:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,FELICIA,N
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: COOPER,FELICIA
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:28

SUBJECT'S LAST NAME: ZBLEWSKI

DATE/TIME OF TEST: 04/30/2023 12:20:58

Diagnostics Check	OK	12:15	
Air Blank	0.000	12:16	
Control Test	0.080	12:16	
Air Blank	0.000	12:16	
Subject Sample	# 1	12:19	Subject Test Refused
Breath Volume	0.222	12:19	
Air Blank	0.000	12:20	
Control Test	0.079	12:20	
Air Blank	0.000	12:20	
Diagnostics Check	OK	12:20	

Data Download Date/Time: 05/06/2023 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: CIFUENTES ROBLERO

DATE/TIME OF TEST: 04/02/2023 18:38:11

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.080	18:31
Air Blank	0.000	18:32
Subject Sample # 1	0.280	18:32
Breath Volume	1.101	18:32
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.287	18:36
Breath Volume	2.328	18:36
Air Blank	0.000	18:37
Control Test	0.080	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: DOMINGUEZ SUAREZ

DATE/TIME OF TEST: 04/05/2023 15:21:20

Diagnostics Check	OK	15:15
Air Blank	0.000	15:15
Control Test	0.081	15:16
Air Blank	0.000	15:16
Subject Sample # 1	0.000	15:17
Breath Volume	2.054	15:17
Air Blank	0.000	15:17
Air Blank	0.000	15:19
Subject Sample # 2	0.000	15:19
Breath Volume	1.855	15:19
Air Blank	0.000	15:20
Control Test	0.079	15:20
Air Blank	0.000	15:21
Diagnostics Check	OK	15:21

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GEORGE,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RAMOS,CARLOS
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: CORPORAN

DATE/TIME OF TEST: 04/07/2023 09:34:10

Diagnostics Check	OK	09:27
Air Blank	0.000	09:28
Control Test	0.079	09:28
Air Blank	0.000	09:29
Subject Sample # 1	0.201	09:29
Breath Volume	2.078	09:29
Air Blank	0.000	09:30
Air Blank	0.000	09:32
Subject Sample # 2	0.187	09:32
Breath Volume	1.914	09:32
Air Blank	0.000	09:33
Control Test	0.077	09:33
Air Blank	0.000	09:34
Diagnostics Check	OK	09:34

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: FOREMAN

DATE/TIME OF TEST: 04/09/2023 02:56:45

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.081	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.087	02:52
Breath Volume	1.718	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.081	02:55
Breath Volume	1.628	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: ANGLIN

DATE/TIME OF TEST: 04/14/2023 10:26:03

Diagnostics Check	OK	10:20
Air Blank	0.000	10:20
Control Test	0.080	10:21
Air Blank	0.000	10:21
Subject Sample # 1	0.000	10:21
Breath Volume	1.457	10:21
Air Blank	0.000	10:22
Air Blank	0.000	10:24
Subject Sample # 2	0.000	10:24
Breath Volume	1.945	10:24
Air Blank	0.000	10:25
Control Test	0.081	10:25
Air Blank	0.000	10:25
Diagnostics Check	OK	10:26

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 04/14/2023 11:59:06

Diagnostics Check	OK	11:53
Air Blank	0.000	11:53
Control Test	0.081	11:53
Air Blank	0.000	11:54
Subject Sample # 1	0.000	11:54
Breath Volume	1.558	11:54
Air Blank	0.000	11:55
Air Blank	0.000	11:57
Subject Sample # 2	0.000	11:57
Breath Volume	1.664	11:57
Air Blank	0.000	11:58
Control Test	0.081	11:58
Air Blank	0.000	11:59
Diagnostics Check	OK	11:59

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: FONSECA

DATE/TIME OF TEST: 04/14/2023 14:37:28

Diagnostics Check	OK	14:31
Air Blank	0.000	14:31
Control Test	0.081	14:32
Air Blank	0.000	14:32
Subject Sample # 1	0.000	14:33
Breath Volume	2.859	14:33
Air Blank	0.000	14:33
Air Blank	0.000	14:35
Subject Sample # 2	0.000	14:36
Breath Volume	1.878	14:36
Air Blank	0.000	14:36
Control Test	0.081	14:36
Air Blank	0.000	14:37
Diagnostics Check	OK	14:37

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: DOMINGUEZ-SUAREZ

DATE/TIME OF TEST: 04/14/2023 14:58:10

Diagnostics Check	OK	14:52
Air Blank	0.000	14:52
Control Test	0.081	14:52
Air Blank	0.000	14:53
Subject Sample # 1	0.000	14:53
Breath Volume	2.144	14:53
Air Blank	0.000	14:54
Air Blank	0.000	14:56
Subject Sample # 2	0.000	14:56
Breath Volume	2.429	14:56
Air Blank	0.000	14:57
Control Test	0.080	14:57
Air Blank	0.000	14:58
Diagnostics Check	OK	14:58

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: ARD

DATE/TIME OF TEST: 04/20/2023 11:11:34

Diagnostics Check	OK	11:05
Air Blank	0.000	11:05
Control Test	0.080	11:06
Air Blank	0.000	11:06
Subject Sample # 1	0.000	11:07
Breath Volume	2.726	11:07
Air Blank	0.000	11:07
Air Blank	0.000	11:09
Subject Sample # 2	0.000	11:10
Breath Volume	2.695	11:10
Air Blank	0.000	11:10
Control Test	0.079	11:10
Air Blank	0.000	11:11
Diagnostics Check	OK	11:11

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: REINKE

DATE/TIME OF TEST: 04/21/2023 10:22:34

Diagnostics Check	OK	10:16
Air Blank	0.000	10:17
Control Test	0.080	10:17
Air Blank	0.000	10:17
Subject Sample # 1	0.000	10:18
Breath Volume	1.992	10:18
Air Blank	0.000	10:18
Air Blank	0.000	10:20
Subject Sample # 2	0.000	10:21
Breath Volume	2.105	10:21
Air Blank	0.000	10:21
Control Test	0.081	10:21
Air Blank	0.000	10:22
Diagnostics Check	OK	10:22

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: VILLALONHERNANDEZ

DATE/TIME OF TEST: 04/21/2023 10:35:53

Diagnostics Check	OK	10:29
Air Blank	0.000	10:30
Control Test	0.081	10:30
Air Blank	0.000	10:30
Subject Sample # 1	0.000	10:31
Breath Volume	2.968	10:31
Air Blank	0.000	10:32
Air Blank	0.000	10:34
Subject Sample # 2	0.000	10:34
Breath Volume	3.039	10:34
Air Blank	0.000	10:34
Control Test	0.081	10:35
Air Blank	0.000	10:35
Diagnostics Check	OK	10:35

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 04/22/2023 03:57:31

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.164	03:53
Breath Volume	2.125	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.163	03:56
Breath Volume	1.769	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 04/24/2023 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: JEAN-SIMON, STIVE
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 03/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/02/2023 04:05:02

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.081	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.200	04:00
Breath Volume	1.914	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.196	04:03
Breath Volume	2.085	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:05

Data Download Date/Time: 04/06/2023 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VAZQUEZ,ANDREW,

BREATH TEST OPERATOR AGENCY: Melbourne PD

ARRESTING OFFICER: VAZQUEZ,ANDREW

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 33922080A1

DRY GAS EXPIRY DATE: 01/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 04/03/2023 21:35:36

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:32	
Control Test	0.081	21:33	
Air Blank	0.000	21:33	
Subject Sample	# 1	21:33	Slope Not Met
Breath Volume	1.496	21:33	
Air Blank	0.000	21:34	
Control Test	0.081	21:35	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 04/06/2023 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 04/03/2023 21:44:15

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.250	21:39
Breath Volume	1.531	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.251	21:42
Breath Volume	1.476	21:42
Air Blank	0.000	21:43
Control Test	0.078	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 04/06/2023 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPLAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MARTINEZ,OSUE
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CORDERO ROSA

DATE/TIME OF TEST: 04/08/2023 02:57:54

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.082	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.212	02:53
Breath Volume	3.582	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.212	02:56
Breath Volume	4.179	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 05/13/2023 08:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPLAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MARTINEZ,JOSUE
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: TORREYES

DATE/TIME OF TEST: 04/08/2023 21:46:21

Diagnostics Check	OK	21:39
Air Blank	0.000	21:40
Control Test	0.082	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.000	21:42
Breath Volume	1.488	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.000	21:44
Breath Volume	1.820	21:44
Air Blank	0.000	21:45
Control Test	0.081	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 05/13/2023 08:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY, THOMAS, A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CROOKS, JUSTIN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: JOHANSEN

DATE/TIME OF TEST: 04/15/2023 04:04:25

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.023	03:59
Breath Volume	1.468	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.021	04:02
Breath Volume	1.652	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 05/13/2023 08:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY, THOMAS, A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CROOKS, JUSTIN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: CEE

DATE/TIME OF TEST: 04/15/2023 18:57:03

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.082	18:51
Air Blank	0.000	18:52
Subject Sample # 1	0.133	18:52
Breath Volume	1.484	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.138	18:55
Breath Volume	1.582	18:55
Air Blank	0.000	18:56
Control Test	0.081	18:56
Air Blank	0.000	18:56
Diagnostics Check	OK	18:57

Data Download Date/Time: 05/13/2023 08:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPLAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: PAVLAKOS,PAVLOS
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 04/22/2023 03:26:31

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.081	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.082	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 05/13/2023 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPLAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: PAVLAKOS,PAVLOS
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 04/24/2023 02:35:10

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.081	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:30	No Sample Provided
Breath Volume	0.000	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 2	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:33	
Control Test	0.081	02:34	
Air Blank		02:34	RFI Detect
Air Blank	0.000	02:35	

Data Download Date/Time: 05/13/2023 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MARTINEZ, JOSUE
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 04/29/2023 03:02:15

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.082	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.082	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 05/13/2023 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MARTINEZ, JOSUE
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 04/29/2023 03:11:48

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.030	03:07
Breath Volume	1.210	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.031	03:10
Breath Volume	1.468	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 05/13/2023 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MARTINEZ, JOSUE
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 04/30/2023 03:22:03

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.082	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.167	03:17
Breath Volume	2.941	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.167	03:20
Breath Volume	3.121	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

Data Download Date/Time: 05/13/2023 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 08:14

GENDER: Female

OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: JECMEN

DATE/TIME OF TEST: 04/11/2023 08:26:56

Diagnostics Check	OK	08:20
Air Blank	0.000	08:20
Control Test	0.080	08:21
Air Blank	0.000	08:21
Subject Sample # 1	0.194	08:22
Breath Volume	1.597	08:22
Air Blank	0.000	08:22
Air Blank	0.000	08:24
Subject Sample # 2	0.185	08:25
Breath Volume	1.527	08:25
Air Blank	0.000	08:25
Control Test	0.078	08:26
Air Blank	0.000	08:26
Diagnostics Check	OK	08:26

Data Download Date/Time: 04/24/2023 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/08/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 17:08

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: LEINBACH

DATE/TIME OF TEST: 04/12/2023 17:29:02

Diagnostics Check	OK	17:22
Air Blank	0.000	17:23
Control Test	0.079	17:23
Air Blank	0.000	17:23
Subject Sample # 1	0.206	17:24
Breath Volume	2.824	17:24
Air Blank	0.000	17:25
Air Blank	0.000	17:26
Subject Sample # 2	0.202	17:27
Breath Volume	2.246	17:27
Air Blank	0.000	17:28
Control Test	0.078	17:28
Air Blank	0.000	17:28
Diagnostics Check	OK	17:29

Data Download Date/Time: 04/24/2023 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL, SHAQUANIA, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: MALONEY

DATE/TIME OF TEST: 04/03/2023 18:03:59

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:01	
Control Test	0.078	18:01	
Air Blank	0.000	18:02	
Subject Sample	# 1	18:02	Subject Test Refused
Breath Volume	0.000	18:02	
Air Blank	0.000	18:03	
Control Test	0.077	18:03	
Air Blank	0.000	18:03	
Diagnostics Check	OK	18:03	

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 04/04/2023 22:49:07

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.077	22:47	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:47	Subject Test Refused
Breath Volume	0.000	22:47	
Air Blank	0.000	22:48	
Control Test	0.077	22:48	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LYON,L
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SUCKLING

DATE/TIME OF TEST: 04/05/2023 02:27:07

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.077	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.197	02:22
Breath Volume	1.539	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.185	02:25
Breath Volume	1.507	02:25
Air Blank	0.000	02:26
Control Test	0.077	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/05/2023 22:04:46

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:02	
Control Test	0.077	22:02	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:03	
Control Test	0.077	22:04	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/06/2023 00:35:54

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.078	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.111	00:31
Breath Volume	3.375	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:33
Subject Sample # 2	0.114	00:34
Breath Volume	3.292	00:34
Air Blank	0.000	00:34
Control Test	0.078	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON, CONNER, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHILICKI, M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: ESCOBAR GRANADOS

DATE/TIME OF TEST: 04/06/2023 04:11:19

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.078	04:09	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Control Test	0.078	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: RIVERA HERNANDEZ

DATE/TIME OF TEST: 04/07/2023 03:51:17

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.094	03:46
Breath Volume	4.191	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.092	03:49
Breath Volume	4.230	03:49
Air Blank	0.000	03:50
Control Test	0.078	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: AUSTIN,N
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 04/08/2023 05:13:08

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:10	
Control Test	0.078	05:10	
Air Blank	0.000	05:11	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:12	
Control Test	0.078	05:12	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLIS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 04/08/2023 09:12:04

Diagnostics Check	OK	09:05
Air Blank	0.000	09:06
Control Test	0.079	09:06
Air Blank	0.000	09:07
Subject Sample # 1	0.148	09:07
Breath Volume	1.964	09:07
Air Blank	0.000	09:08
Air Blank	0.000	09:10
Subject Sample # 2	0.158	09:10
Breath Volume	3.003	09:10
Air Blank	0.000	09:11
Control Test	0.078	09:11
Air Blank	0.000	09:11
Diagnostics Check	OK	09:12

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CURRY,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:49

SUBJECT'S LAST NAME: HENRRIQUEZ

DATE/TIME OF TEST: 04/08/2023 11:16:43

Diagnostics Check	OK	11:09
Air Blank	0.000	11:10
Control Test	0.079	11:10
Air Blank	0.000	11:11
Subject Sample # 1	0.049	11:11
Breath Volume	1.777	11:11
Air Blank	0.000	11:12
Air Blank	0.000	11:14
Subject Sample # 2	0.049	11:15
Breath Volume	2.113	11:15
Air Blank	0.000	11:15
Control Test	0.077	11:16
Air Blank	0.000	11:16
Diagnostics Check	OK	11:16

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMS,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/09/2023 05:26:49

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.078	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.150	05:22
Breath Volume	1.472	05:22
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.159	05:25
Breath Volume	1.851	05:25
Air Blank	0.000	05:25
Control Test	0.078	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL, EDWARD, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE, S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SWEDBERG

DATE/TIME OF TEST: 04/13/2023 23:07:42

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.078	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.160	23:03
Breath Volume	1.753	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.160	23:06
Breath Volume	1.609	23:06
Air Blank	0.000	23:06
Control Test	0.077	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/14/2023 00:08:53

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.077	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.152	00:04
Breath Volume	3.140	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.155	00:07
Breath Volume	2.929	00:07
Air Blank	0.000	00:07
Control Test	0.078	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL, EDWARD, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA, A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/14/2023 02:36:16

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.078	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.000	02:32
Breath Volume	2.230	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.000	02:34
Breath Volume	2.203	02:34
Air Blank	0.000	02:35
Control Test	0.078	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: CRISTALES RUIZ

DATE/TIME OF TEST: 04/14/2023 23:13:35

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.078	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.064	23:09
Breath Volume	1.605	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.064	23:12
Breath Volume	1.218	23:12
Air Blank	0.000	23:12
Control Test	0.077	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: CALLEJAS FORERO

DATE/TIME OF TEST: 04/15/2023 02:09:06

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.151	02:04
Breath Volume	2.656	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.149	02:07
Breath Volume	2.285	02:07
Air Blank	0.000	02:08
Control Test	0.077	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: REICHART

DATE/TIME OF TEST: 04/15/2023 04:16:10

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.078	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.136	04:11
Breath Volume	2.007	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.133	04:14
Breath Volume	2.910	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 04/24/2023 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CHATMAN

DATE/TIME OF TEST: 04/16/2023 01:23:04

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.077	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.205	01:18
Breath Volume	1.539	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.197	01:21
Breath Volume	1.964	01:21
Air Blank	0.000	01:22
Control Test	0.077	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:23

Data Download Date/Time: 04/24/2023 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 04/16/2023 16:55:07

Diagnostics Check	OK	16:52	
Air Blank	0.000	16:52	
Control Test	0.078	16:53	
Air Blank	0.000	16:53	
Subject Sample	# 1	16:53	Subject Test Refused
Breath Volume	0.000	16:53	
Air Blank	0.000	16:54	
Control Test	0.078	16:54	
Air Blank	0.000	16:55	
Diagnostics Check	OK	16:55	

Data Download Date/Time: 04/24/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: GUZMAN CHACLAN

DATE/TIME OF TEST: 04/17/2023 03:16:06

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.078	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Control Test	0.078	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 04/24/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 04/21/2023 02:57:19

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.078	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:50	No Sample Provided
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2	0.049	02:52
Breath Volume	2.039	02:52	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 3	0.048	02:55
Breath Volume	1.769	02:55	
Air Blank	0.000	02:56	
Control Test	0.078	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 04/24/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 04/22/2023 01:16:28

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.078	01:14	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.078	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 04/24/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FUENTES ALVEREZ

DATE/TIME OF TEST: 04/22/2023 03:56:02

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.119	03:51
Breath Volume	1.867	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.118	03:54
Breath Volume	1.437	03:54
Air Blank	0.000	03:55
Control Test	0.077	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:56

Data Download Date/Time: 04/24/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DELEO,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/23/2023 05:33:45

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.078	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.204	05:29
Breath Volume	3.832	05:29
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.190	05:32
Breath Volume	2.632	05:32
Air Blank	0.000	05:32
Control Test	0.077	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 04/24/2023 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: LAND

DATE/TIME OF TEST: 04/23/2023 22:07:09

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.076	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.125	22:02
Breath Volume	2.445	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:04
Subject Sample # 2	0.114	22:05
Breath Volume	1.339	22:05
Air Blank	0.000	22:06
Control Test	0.077	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 04/24/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 04/24/2023 02:46:43

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.000	02:42
Breath Volume	2.367	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.000	02:45
Breath Volume	2.742	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 04/24/2023 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OCONNELL,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: RAGANS

DATE/TIME OF TEST: 04/24/2023 04:58:27

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:56	
Control Test	0.077	04:56	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:57	
Control Test	0.077	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 04/24/2023 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: O'BRIEN

DATE/TIME OF TEST: 04/25/2023 02:23:03

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.077	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.071	02:17
Breath Volume	1.921	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.067	02:21
Breath Volume	2.058	02:21
Air Blank	0.000	02:22
Control Test	0.077	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:23

Data Download Date/Time: 05/23/2023 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MERTZ

DATE/TIME OF TEST: 04/26/2023 02:07:40

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.078	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.078	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 05/23/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 04/26/2023 03:02:40

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.099	02:57
Breath Volume	2.164	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.095	03:01
Breath Volume	1.265	03:01
Air Blank	0.000	03:01
Control Test	0.077	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 05/23/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,E,V
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 04/27/2023 07:46:58

Diagnostics Check	OK	07:40
Air Blank	0.000	07:41
Control Test	0.078	07:41
Air Blank	0.000	07:41
Subject Sample # 1	0.148	07:42
Breath Volume	3.144	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:44
Subject Sample # 2	0.139	07:45
Breath Volume	1.531	07:45
Air Blank	0.000	07:46
Control Test	0.076	07:46
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

Data Download Date/Time: 05/23/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 04/28/2023 02:45:45

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.026	02:41
Breath Volume	1.765	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.025	02:44
Breath Volume	1.222	02:44
Air Blank	0.000	02:44
Control Test	0.077	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 05/23/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 04/28/2023 23:36:35

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:34	
Control Test	0.078	23:34	
Air Blank	0.000	23:35	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Control Test	0.078	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 05/23/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: ZARAGOZA CAMARILLO

DATE/TIME OF TEST: 04/29/2023 18:51:11

Diagnostics Check	OK	18:48	
Air Blank	0.000	18:48	
Control Test	0.078	18:48	
Air Blank	0.000	18:49	
Subject Sample	# 1	18:49	Subject Test Refused
Breath Volume	0.000	18:49	
Air Blank	0.000	18:50	
Control Test	0.078	18:50	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time: 05/23/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: KUUSELA,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/30/2023 01:39:46

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.077	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.167	01:35
Breath Volume	1.699	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.179	01:38
Breath Volume	1.449	01:38
Air Blank	0.000	01:38
Control Test	0.077	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 05/23/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: AVILA-ANGARITA

DATE/TIME OF TEST: 04/30/2023 04:01:05

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.078	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 05/23/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: BELTON

DATE/TIME OF TEST: 04/30/2023 23:19:57

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.077	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.079	23:15
Breath Volume	1.328	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:17
Subject Sample # 2	0.073	23:18
Breath Volume	1.207	23:18
Air Blank	0.000	23:19
Control Test	0.076	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 05/23/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MICHAEL,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MEROVEE

DATE/TIME OF TEST: 04/08/2023 04:16:33

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:14	
Control Test	0.078	04:14	
Air Blank	0.000	04:15	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Control Test	0.078	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 04/26/2023 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: GAVIN,HENRY
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MIAZOTI

DATE/TIME OF TEST: 04/13/2023 21:23:57

Diagnostics Check	OK	21:19	
Air Blank	0.000	21:20	
Control Test	0.078	21:20	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:22	Subject Test Refused
Breath Volume	0.542	21:22	
Air Blank	0.000	21:23	
Control Test	0.079	21:23	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 04/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: MAYORGA

DATE/TIME OF TEST: 04/18/2023 01:07:31

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.078	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.097	01:03
Breath Volume	2.746	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.091	01:06
Breath Volume	2.585	01:06
Air Blank	0.000	01:06
Control Test	0.078	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 04/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ELI

DATE/TIME OF TEST: 04/19/2023 01:51:02

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.078	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.102	01:46
Breath Volume	1.917	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.099	01:49
Breath Volume	1.242	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:51

Data Download Date/Time: 04/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/20/2023 23:26:58

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.078	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.219	23:22
Breath Volume	1.203	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.220	23:25
Breath Volume	1.296	23:25
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 04/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MICHAEL,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: DORIA

DATE/TIME OF TEST: 04/22/2023 20:45:15

Diagnostics Check	OK	20:44	
Air Blank		20:44	RFI Detect
Air Blank	0.000	20:45	

Data Download Date/Time: 04/26/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MICHAEL,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: DORIA

DATE/TIME OF TEST: 04/22/2023 20:50:21

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:48	
Control Test	0.078	20:48	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:49	Subject Test Refused
Breath Volume	0.000	20:49	
Air Blank	0.000	20:49	
Control Test	0.078	20:49	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time: 04/26/2023 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 04/27/2023 21:46:39

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:44	
Control Test	0.082	21:44	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:45	Subject Test Refused
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.082	21:46	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 05/31/2023 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: BOWEN PIAZZA

DATE/TIME OF TEST: 04/01/2023 02:33:53

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.165	02:28
Breath Volume	1.187	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.158	02:32
Breath Volume	1.675	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/01/2023 05:07:08

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:04	
Control Test	0.080	05:05	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:05	Subject Test Refused
Breath Volume	0.000	05:05	
Air Blank	0.000	05:06	
Control Test	0.080	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GUTH,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: LOPEZ MORALES

DATE/TIME OF TEST: 04/01/2023 07:00:40

Diagnostics Check	OK	06:53
Air Blank	0.000	06:54
Control Test	0.080	06:54
Air Blank	0.000	06:55
Subject Sample # 1	0.215	06:56
Breath Volume	2.597	06:56
Air Blank	0.000	06:56
Air Blank	0.000	06:58
Subject Sample # 2	0.208	06:58
Breath Volume	2.601	06:58
Air Blank	0.000	06:59
Control Test	0.080	07:00
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GIBSON,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: FUENTES LOPEZ

DATE/TIME OF TEST: 04/01/2023 20:42:01

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.079	20:39	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:41	
Control Test	0.079	20:41	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:42	

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: DEMPSEY

DATE/TIME OF TEST: 04/01/2023 23:46:34

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:44	
Control Test	0.081	23:44	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Control Test	0.080	23:45	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: GONZALEZ ZAVALA

DATE/TIME OF TEST: 04/02/2023 02:50:57

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.161	02:46
Breath Volume	2.394	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.147	02:49
Breath Volume	1.167	02:49
Air Blank	0.000	02:50
Control Test	0.077	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 04/03/2023 04:44:25

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.175	04:39
Breath Volume	2.171	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.168	04:42
Breath Volume	2.183	04:42
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: GOUKASSIAN

DATE/TIME OF TEST: 04/04/2023 00:23:14

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.039	00:18
Breath Volume	1.550	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.037	00:21
Breath Volume	1.609	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/04/2023 03:40:32

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.083	03:36
Breath Volume	2.343	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.086	03:38
Breath Volume	2.617	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,JARROD,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHIPP,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: MCKINSEY

DATE/TIME OF TEST: 04/04/2023 05:01:49

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.080	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.171	04:56
Breath Volume	1.425	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.176	05:00
Breath Volume	1.820	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 04/05/2023 01:18:15

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.100	01:13
Breath Volume	1.906	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.087	01:16
Breath Volume	1.136	01:16
Air Blank	0.000	01:17
Control Test	0.077	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 04/05/2023 03:05:59

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.081	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Control Test	0.080	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: URIAS ALBERTO

DATE/TIME OF TEST: 04/05/2023 21:44:25

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.081	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.094	21:39
Breath Volume	2.167	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.095	21:42
Breath Volume	2.324	21:42
Air Blank	0.000	21:43
Control Test	0.080	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 04/06/2023 01:02:37

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.000	00:58
Breath Volume	2.542	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.000	01:01
Breath Volume	2.152	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 04/20/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: VENN

DATE/TIME OF TEST: 04/06/2023 02:28:46

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.125	02:24
Breath Volume	1.230	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.125	02:27
Breath Volume	2.179	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 04/06/2023 03:32:42

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Control Test	0.079	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: LANGE

DATE/TIME OF TEST: 04/06/2023 22:29:03

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.081	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.194	22:24
Breath Volume	1.402	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:26
Subject Sample # 2	0.184	22:27
Breath Volume	1.289	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:29

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PEREZ FUENTES

DATE/TIME OF TEST: 04/07/2023 00:12:31

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.043	00:07
Breath Volume	1.601	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.046	00:10
Breath Volume	2.707	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MARTINO

DATE/TIME OF TEST: 04/07/2023 02:10:51

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.081	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.080	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHILICKI,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: JUDGE

DATE/TIME OF TEST: 04/07/2023 06:41:14

Diagnostics Check	OK	06:34
Air Blank	0.000	06:35
Control Test	0.080	06:35
Air Blank	0.000	06:36
Subject Sample # 1	0.127	06:36
Breath Volume	2.351	06:36
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.119	06:39
Breath Volume	2.925	06:39
Air Blank	0.000	06:40
Control Test	0.080	06:40
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 04/07/2023 21:19:32

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.081	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.000	21:15
Breath Volume	1.472	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.000	21:18
Breath Volume	1.332	21:18
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: SCHUMAN

DATE/TIME OF TEST: 04/08/2023 04:57:45

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.188	04:52
Breath Volume	1.191	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.172	04:56
Breath Volume	1.187	04:56
Air Blank	0.000	04:56
Control Test	0.079	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 04/08/2023 19:08:36

Diagnostics Check	OK	19:05	
Air Blank	0.000	19:06	
Control Test	0.080	19:06	
Air Blank	0.000	19:07	
Subject Sample	# 1	19:07	Subject Test Refused
Breath Volume	0.000	19:07	
Air Blank	0.000	19:07	
Control Test	0.079	19:08	
Air Blank	0.000	19:08	
Diagnostics Check	OK	19:08	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 04/08/2023 20:24:23

Diagnostics Check	OK	20:21	
Air Blank	0.000	20:21	
Control Test	0.080	20:22	
Air Blank	0.000	20:22	
Subject Sample	# 1	20:23	Subject Test Refused
Breath Volume	0.000	20:23	
Air Blank	0.000	20:23	
Control Test	0.080	20:23	
Air Blank	0.000	20:24	
Diagnostics Check	OK	20:24	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: DEWAN

DATE/TIME OF TEST: 04/09/2023 03:25:01

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.081	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DORNER,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/09/2023 04:13:07

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.111	04:08
Breath Volume	1.757	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.112	04:11
Breath Volume	1.753	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PORTILLO GALO

DATE/TIME OF TEST: 04/10/2023 02:34:54

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.259	02:29
Breath Volume	1.335	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.255	02:33
Breath Volume	1.750	02:33
Air Blank	0.000	02:33
Control Test	0.078	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARDWELL,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: FELDER

DATE/TIME OF TEST: 04/10/2023 19:44:46

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:42	
Control Test	0.080	19:42	
Air Blank	0.000	19:43	
Subject Sample	# 1	19:43	Subject Test Refused
Breath Volume	0.000	19:43	
Air Blank	0.000	19:43	
Control Test	0.080	19:44	
Air Blank	0.000	19:44	
Diagnostics Check	OK	19:44	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: MARTIN DIEGO

DATE/TIME OF TEST: 04/10/2023 23:04:24

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.072	22:59
Breath Volume	2.589	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.071	23:02
Breath Volume	3.089	23:02
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 04/11/2023 02:47:54

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.080	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WHELCHHEL

DATE/TIME OF TEST: 04/11/2023 21:27:43

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:25	
Control Test	0.081	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:26	
Control Test	0.080	21:27	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:27	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: HUX

DATE/TIME OF TEST: 04/12/2023 01:34:21

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.080	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1 0.204	01:26	Volume Not Met
Breath Volume	0.199	01:26	
Air Blank	0.000	01:27	
Air Blank	0.000	01:29	
Subject Sample	# 2 0.166	01:32	Volume Not Met
Breath Volume	0.363	01:32	
Air Blank	0.000	01:33	
Control Test	0.078	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 04/12/2023 03:08:41

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.080	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.080	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 04/13/2023 04:16:06

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.081	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.086	04:11
Breath Volume	1.652	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.088	04:14
Breath Volume	2.480	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 04/15/2023 03:39:55

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.081	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Control Test	0.080	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: TRAVALEE

DATE/TIME OF TEST: 04/15/2023 22:39:54

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.081	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.080	22:39	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 04/16/2023 04:28:51

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:26	
Control Test	0.080	04:26	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Control Test	0.081	04:28	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: ARLINE

DATE/TIME OF TEST: 04/16/2023 05:19:00

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.080	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:17	Subject Test Refused
Breath Volume	0.339	05:17	
Air Blank	0.000	05:18	
Control Test	0.081	05:18	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:19	

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAMAR,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: ZATINA

DATE/TIME OF TEST: 04/16/2023 06:39:35

Diagnostics Check	OK	06:33
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:34
Subject Sample # 1	0.084	06:34
Breath Volume	1.460	06:34
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample # 2	0.080	06:37
Breath Volume	1.402	06:37
Air Blank	0.000	06:38
Control Test	0.080	06:38
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BERNARD,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:54

SUBJECT'S LAST NAME: OBANNON

DATE/TIME OF TEST: 04/16/2023 09:21:50

Diagnostics Check	OK	09:14
Air Blank	0.000	09:15
Control Test	0.080	09:15
Air Blank	0.000	09:16
Subject Sample # 1	0.133	09:16
Breath Volume	1.761	09:16
Air Blank	0.000	09:17
Air Blank	0.000	09:19
Subject Sample # 2	0.131	09:20
Breath Volume	1.199	09:20
Air Blank	0.000	09:20
Control Test	0.080	09:21
Air Blank	0.000	09:21
Diagnostics Check	OK	09:21

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: ORDONEZ SOTO

DATE/TIME OF TEST: 04/16/2023 22:10:38

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.139	22:05
Breath Volume	2.082	22:05
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.128	22:08
Breath Volume	1.812	22:08
Air Blank	0.000	22:09
Control Test	0.080	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 04/20/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: BARRY

DATE/TIME OF TEST: 04/17/2023 02:46:31

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.081	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.081	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 04/20/2023 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: CHAMPION

DATE/TIME OF TEST: 04/17/2023 03:33:56

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.195	03:28
Breath Volume	1.992	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.201	03:32
Breath Volume	3.097	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 04/20/2023 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: GUTIERREZ MONTES

DATE/TIME OF TEST: 04/17/2023 23:37:16

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.080	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.112	23:32
Breath Volume	1.148	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.115	23:35
Breath Volume	1.496	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 04/20/2023 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUADARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 04/18/2023 00:25:13

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.081	00:23	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Control Test	0.080	00:24	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 04/20/2023 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HITE,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: KEEBLE

DATE/TIME OF TEST: 04/19/2023 01:44:04

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.157	01:39
Breath Volume	1.941	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.161	01:42
Breath Volume	1.804	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:44

Data Download Date/Time: 04/20/2023 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: ANCHEZ HERNANDEZ

DATE/TIME OF TEST: 04/19/2023 19:53:48

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:51	
Control Test	0.081	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:52	Subject Test Refused
Breath Volume	0.000	19:52	
Air Blank	0.000	19:52	
Control Test	0.081	19:53	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:53	

Data Download Date/Time: 04/20/2023 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: SABATER

DATE/TIME OF TEST: 04/19/2023 22:07:56

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.081	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.000	22:03
Breath Volume	2.113	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.000	22:06
Breath Volume	2.222	22:06
Air Blank	0.000	22:07
Control Test	0.081	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 04/20/2023 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: SIBBALD

DATE/TIME OF TEST: 04/19/2023 23:45:07

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.081	23:42	
Air Blank	0.000	23:43	
Subject Sample	# 1	23:43	Subject Test Refused
Breath Volume	0.000	23:43	
Air Blank	0.000	23:44	
Control Test	0.080	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 04/20/2023 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/20/2023 04:13:37

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.081	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.169	04:08
Breath Volume	2.488	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.160	04:12
Breath Volume	1.671	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 04/20/2023 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FUNEZ ALMENDAREZ

DATE/TIME OF TEST: 04/20/2023 23:28:48

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.085	23:24
Breath Volume	1.777	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.083	23:27
Breath Volume	1.878	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MIN

DATE/TIME OF TEST: 04/21/2023 04:27:44

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.079	04:23
Breath Volume	1.437	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.073	04:26
Breath Volume	1.312	04:26
Air Blank	0.000	04:26
Control Test	0.079	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: TENNYSON

DATE/TIME OF TEST: 04/21/2023 20:51:11

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:48	
Control Test	0.080	20:49	
Air Blank	0.000	20:49	
Subject Sample	# 1	20:49	Subject Test Refused
Breath Volume	0.000	20:49	
Air Blank	0.000	20:50	
Control Test	0.080	20:50	
Air Blank	0.000	20:51	
Diagnostics Check	OK	20:51	

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 04/21/2023 22:41:58

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.081	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.000	22:37
Breath Volume	2.269	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.000	22:40
Breath Volume	1.996	22:40
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: TAYLORSON

DATE/TIME OF TEST: 04/22/2023 00:44:01

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.133	00:39
Breath Volume	1.117	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.129	00:42
Breath Volume	1.710	00:42
Air Blank	0.000	00:43
Control Test	0.081	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:44

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: RATH

DATE/TIME OF TEST: 04/22/2023 20:42:08

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.081	20:40	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:41	
Control Test	0.081	20:41	
Air Blank	0.000	20:42	
Diagnostics Check	OK	20:42	

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: SUTER

DATE/TIME OF TEST: 04/22/2023 22:27:03

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:24	
Control Test	0.081	22:25	
Air Blank	0.000	22:25	
Subject Sample	# 1	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: RAYBURN

DATE/TIME OF TEST: 04/22/2023 23:34:16

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.000	23:29
Breath Volume	3.367	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.000	23:32
Breath Volume	3.230	23:32
Air Blank	0.000	23:33
Control Test	0.081	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: PATISTAN

DATE/TIME OF TEST: 04/23/2023 00:32:56

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.082	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.116	00:28
Breath Volume	2.859	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.102	00:31
Breath Volume	1.562	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 04/23/2023 02:37:27

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.081	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DRAGOMIR,I
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: FLOURNOY

DATE/TIME OF TEST: 04/23/2023 04:53:24

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.081	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:52	
Control Test	0.081	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 05/24/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: SNIDER

DATE/TIME OF TEST: 04/23/2023 05:29:38

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.082	05:24
Air Blank	0.000	05:24
Subject Sample # 1	0.145	05:25
Breath Volume	2.203	05:25
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.137	05:28
Breath Volume	2.296	05:28
Air Blank	0.000	05:28
Control Test	0.080	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GETZ,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: GARCIA ESTRADA

DATE/TIME OF TEST: 04/23/2023 06:31:33

Diagnostics Check	OK	06:28	
Air Blank	0.000	06:29	
Control Test	0.081	06:29	
Air Blank	0.000	06:29	
Subject Sample	# 1	06:30	Subject Test Refused
Breath Volume	0.000	06:30	
Air Blank	0.000	06:30	
Control Test	0.081	06:30	
Air Blank	0.000	06:31	
Diagnostics Check	OK	06:31	

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/23/2023 22:33:00

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.081	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.140	22:28
Breath Volume	1.375	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:30
Subject Sample # 2	0.144	22:31
Breath Volume	1.203	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:33

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: BRIDWELL

DATE/TIME OF TEST: 04/24/2023 19:34:53

Diagnostics Check	OK	19:28
Air Blank	0.000	19:29
Control Test	0.082	19:29
Air Blank	0.000	19:29
Subject Sample # 1	0.000	19:30
Breath Volume	1.648	19:30
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.000	19:33
Breath Volume	2.132	19:33
Air Blank	0.000	19:33
Control Test	0.082	19:34
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 04/25/2023 02:06:13

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.091	02:01
Breath Volume	1.566	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample # 2	0.086	02:04
Breath Volume	1.113	02:04
Air Blank	0.000	02:05
Control Test	0.081	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOBLEY,JAMES,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 04/25/2023 04:50:01

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.082	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:49	
Control Test	0.081	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: WEIR

DATE/TIME OF TEST: 04/25/2023 23:54:55

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.082	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.151	23:49
Breath Volume	1.445	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.140	23:53
Breath Volume	1.179	23:53
Air Blank	0.000	23:53
Control Test	0.081	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: RASO

DATE/TIME OF TEST: 04/26/2023 01:14:11

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.082	01:12	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: DIANGELO

DATE/TIME OF TEST: 04/26/2023 02:11:12

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.082	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.066	02:06
Breath Volume	2.507	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.066	02:09
Breath Volume	2.027	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MATTIACE,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 04/26/2023 03:09:28

Diagnostics Check	OK	03:08	
Air Blank		03:09	RFI Detect
Air Blank	0.000	03:09	

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MATTIACE,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 04/26/2023 03:14:45

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.082	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: PEREZ GODINEZ

DATE/TIME OF TEST: 04/26/2023 04:19:01

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.082	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.099	04:14
Breath Volume	1.226	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.106	04:17
Breath Volume	1.507	04:17
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:19

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 04/26/2023 05:25:10

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.082	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.123	05:19
Breath Volume	2.488	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.116	05:23
Breath Volume	2.054	05:23
Air Blank	0.000	05:24
Control Test	0.081	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 04/26/2023 21:53:30

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:51	
Control Test	0.082	21:51	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Control Test	0.082	21:52	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/27/2023 00:05:08

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.082	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:04	
Control Test	0.081	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: HELMS

DATE/TIME OF TEST: 04/27/2023 00:53:45

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.140	00:48
Breath Volume	1.519	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.143	00:52
Breath Volume	1.890	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CALDWELL,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: COCKFIELD

DATE/TIME OF TEST: 04/28/2023 02:42:48

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.082	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.239	02:36
Breath Volume	1.941	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.239	02:41
Breath Volume	1.781	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SAMAYOA-GUTIERREZ

DATE/TIME OF TEST: 04/29/2023 00:55:55

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.081	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.119	00:51
Breath Volume	2.101	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:53
Subject Sample # 2	0.110	00:54
Breath Volume	1.644	00:54
Air Blank	0.000	00:54
Control Test	0.081	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/29/2023 21:49:31

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.082	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.166	21:44
Breath Volume	1.296	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.155	21:47
Breath Volume	1.179	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/29/2023 22:30:39

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.081	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.141	22:26
Breath Volume	2.281	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.131	22:29
Breath Volume	1.582	22:29
Air Blank	0.000	22:29
Control Test	0.081	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/30/2023 01:32:25

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:30	
Control Test	0.082	01:30	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Control Test	0.081	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: PSHENICHNY

DATE/TIME OF TEST: 04/30/2023 02:28:35

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.082	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.081	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 04/30/2023 03:45:36

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.082	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.036	03:41
Breath Volume	2.199	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.036	03:44
Breath Volume	1.574	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: NOREN

DATE/TIME OF TEST: 04/30/2023 21:59:25

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:56	
Control Test	0.082	21:56	
Air Blank	0.000	21:57	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:58	
Control Test	0.081	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 05/24/2023 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ALIJAGIE,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: MONTFORT

DATE/TIME OF TEST: 04/01/2023 07:57:57

Diagnostics Check	OK	07:51
Air Blank	0.000	07:51
Control Test	0.079	07:52
Air Blank	0.000	07:52
Subject Sample # 1	0.164	07:53
Breath Volume	1.992	07:53
Air Blank	0.000	07:53
Air Blank	0.000	07:55
Subject Sample # 2	0.163	07:56
Breath Volume	1.832	07:56
Air Blank	0.000	07:57
Control Test	0.079	07:57
Air Blank	0.000	07:57
Diagnostics Check	OK	07:57

Data Download Date/Time: 04/28/2023 01:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TBER,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 04/08/2023 05:44:11

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:41	
Control Test	0.079	05:41	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:43	
Control Test	0.079	05:43	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 04/28/2023 01:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BOWLING

DATE/TIME OF TEST: 04/16/2023 03:19:43

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.212	03:14
Breath Volume	2.230	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.198	03:18
Breath Volume	1.269	03:18
Air Blank	0.000	03:18
Control Test	0.078	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 04/28/2023 01:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DASHER

DATE/TIME OF TEST: 04/19/2023 03:34:10

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.190	03:29
Breath Volume	2.144	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.177	03:32
Breath Volume	2.160	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 04/28/2023 01:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DOAN

DATE/TIME OF TEST: 04/20/2023 03:15:25

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.080	03:14	Interferent Detect
Air Blank	0.000	03:15	

Data Download Date/Time: 04/28/2023 01:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DOAN

DATE/TIME OF TEST: 04/20/2023 03:18:37

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.079	03:18	Interferent Detect
Air Blank	0.000	03:18	

Data Download Date/Time: 04/28/2023 01:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: OU

DATE/TIME OF TEST: 04/22/2023 02:35:21

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.234	02:30
Breath Volume	2.902	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.238	02:33
Breath Volume	3.050	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 04/28/2023 01:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/22/2023 20:04:32

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:56	
Control Test	0.079	19:56	
Air Blank	0.000	19:57	
Subject Sample	# 1	20:00	No Sample Provided
Breath Volume	0.000	20:00	
Air Blank	0.000	20:00	
Air Blank	0.000	20:02	
Subject Sample	# 2	20:03	Subject Test Refused
Breath Volume	0.000	20:03	
Air Blank	0.000	20:03	
Control Test	0.080	20:03	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 04/28/2023 01:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: PLUMPOWSKI

DATE/TIME OF TEST: 04/26/2023 23:47:47

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.000	23:43
Breath Volume	2.457	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.000	23:46
Breath Volume	2.804	23:46
Air Blank	0.000	23:46
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 04/28/2023 01:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HAIRE,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: GARCIA GARCIA

DATE/TIME OF TEST: 04/27/2023 23:42:08

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.000	23:37
Breath Volume	4.527	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.000	23:40
Breath Volume	4.484	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 04/28/2023 01:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/28/2023 03:36:30

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:35	Interferent Detect
Air Blank	0.000	03:36	

Data Download Date/Time: 05/26/2023 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/28/2023 03:41:07

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.078	03:40	Interferent Detect
Air Blank	0.000	03:41	

Data Download Date/Time: 05/26/2023 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: EGGIMANN

DATE/TIME OF TEST: 04/02/2023 21:25:51

Diagnostics Check	OK	21:17
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Subject Sample # 1	0.179	21:20
Breath Volume	1.453	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.193	21:24
Breath Volume	2.445	21:24
Air Blank	0.000	21:24
Control Test	0.080	21:25
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 05/03/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,D,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,D
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: GUARISCO

DATE/TIME OF TEST: 04/05/2023 18:50:34

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:45
Subject Sample # 1	0.000	18:46
Breath Volume	1.667	18:46
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 2	0.000	18:49
Breath Volume	2.632	18:49
Air Blank	0.000	18:49
Control Test	0.081	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 05/03/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: DELAURA

DATE/TIME OF TEST: 04/09/2023 21:06:44

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.080	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.172	21:02
Breath Volume	2.820	21:02
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.179	21:05
Breath Volume	3.125	21:05
Air Blank	0.000	21:05
Control Test	0.080	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 05/03/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: MAILISAND

DATE/TIME OF TEST: 04/12/2023 05:37:55

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.082	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.000	05:33
Breath Volume	2.312	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.000	05:36
Breath Volume	1.832	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 05/03/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: QUILLEN J

DATE/TIME OF TEST: 04/12/2023 23:51:03

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	RFI Detect
Air Blank	0.000	23:51	

Data Download Date/Time: 05/03/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: QUILLEN

DATE/TIME OF TEST: 04/12/2023 23:59:15

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.165	23:54
Breath Volume	2.492	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.164	23:57
Breath Volume	2.121	23:57
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 05/03/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: COLASUONNO

DATE/TIME OF TEST: 04/14/2023 02:58:16

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.155	02:53
Breath Volume	1.257	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.156	02:56
Breath Volume	1.312	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 05/03/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAUS,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KRAUS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 04/14/2023 04:18:42

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.159	04:13
Breath Volume	1.941	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.167	04:17
Breath Volume	2.082	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 05/03/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: BROWARD

DATE/TIME OF TEST: 04/15/2023 00:26:34

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:24	
Control Test	0.081	00:24	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:25	Subject Test Refused
Breath Volume	0.000	00:25	
Air Blank	0.000	00:25	
Control Test	0.082	00:25	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 05/03/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZO,HAYLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/17/2023 22:16:05

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.178	22:08
Breath Volume	2.222	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.154	22:11
Breath Volume	1.250	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 3	0.164	22:14
Breath Volume	1.859	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 05/03/2023 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE, GARRET
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 04/23/2023 04:27:48

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.000	04:23
Breath Volume	1.695	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.000	04:26
Breath Volume	2.968	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 05/03/2023 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAMOS, KEARSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: SUBHI

DATE/TIME OF TEST: 04/23/2023 05:22:22

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.000	05:18
Breath Volume	1.378	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.000	05:21
Breath Volume	2.664	05:21
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 05/03/2023 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: OIKNINE

DATE/TIME OF TEST: 04/25/2023 03:21:56

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.081	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 05/03/2023 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,DAVID,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: HEAPS

DATE/TIME OF TEST: 04/26/2023 18:56:32

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:54	
Control Test	0.081	18:54	
Air Blank	0.000	18:55	
Subject Sample	# 1	18:55	Subject Test Refused
Breath Volume	0.000	18:55	
Air Blank	0.000	18:55	
Control Test	0.080	18:55	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 05/03/2023 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/28/2023 01:46:41

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.080	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1 0.202	01:39	Volume Not Met
Breath Volume	0.015	01:39	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 05/03/2023 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ROWELLA

DATE/TIME OF TEST: 04/28/2023 22:51:23

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.080	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.153	22:46
Breath Volume	1.453	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:48
Subject Sample # 2	0.161	22:49
Breath Volume	1.953	22:49
Air Blank	0.000	22:50
Control Test	0.081	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 05/03/2023 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARQUIS,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: RAMOS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:19

SUBJECT'S LAST NAME: PILAWA

DATE/TIME OF TEST: 04/29/2023 12:45:50

Diagnostics Check	OK	12:39	
Air Blank	0.000	12:40	
Control Test	0.080	12:40	
Air Blank	0.000	12:41	
Subject Sample # 1	0.249	12:41	
Breath Volume	2.324	12:41	
Air Blank	0.000	12:42	
Air Blank	0.000	12:43	
Subject Sample # 2	0.262	12:44	
Breath Volume	2.652	12:44	
Air Blank	0.000	12:44	
Control Test	0.081	12:45	Interferent Detect
Air Blank	0.000	12:45	

Data Download Date/Time: 05/03/2023 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 04/30/2023 02:02:28

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.148	01:57
Breath Volume	2.871	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.149	02:00
Breath Volume	2.816	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 05/03/2023 17:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ERIC,R
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KIRK,ANDRAE
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COWVINS

DATE/TIME OF TEST: 04/16/2023 02:33:37

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.082	02:28	
Air Blank	0.000	02:28	
Subject Sample # 1	0.159	02:29	
Breath Volume	1.812	02:29	
Air Blank	0.000	02:29	
Air Blank	0.000	02:31	
Subject Sample # 2		02:32	Interferent Detect
Breath Volume	1.609	02:32	
Air Blank	0.000	02:32	
Control Test	0.081	02:33	Interferent Detect
Air Blank	0.000	02:33	

Data Download Date/Time: 05/30/2023 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ERIC,R
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KIRK,ANDRAE
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COWVINS

DATE/TIME OF TEST: 04/16/2023 02:42:34

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.153	02:38
Breath Volume	2.171	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.151	02:41
Breath Volume	2.046	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/30/2023 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,ALEXANDER,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: THOMAS,ALEXANDER
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 04/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: BORJA VALDOVINOS

DATE/TIME OF TEST: 04/30/2023 08:28:45

Diagnostics Check	OK	08:25	
Air Blank	0.000	08:26	
Control Test	0.081	08:26	
Air Blank	0.000	08:27	
Subject Sample	# 1	08:27	Subject Test Refused
Breath Volume	0.000	08:27	
Air Blank	0.000	08:27	
Control Test	0.081	08:28	
Air Blank	0.000	08:28	
Diagnostics Check	OK	08:28	

Data Download Date/Time: 05/30/2023 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 04/02/2023 03:06:16

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.079	03:04	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Control Test	0.078	03:05	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:06	

Data Download Date/Time: 04/25/2023 01:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,J,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GRIFFITH,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/02/2023 06:33:39

Diagnostics Check	OK	06:26
Air Blank	0.000	06:26
Control Test	0.079	06:26
Air Blank	0.000	06:27
Subject Sample # 1	0.118	06:28
Breath Volume	1.410	06:28
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.119	06:32
Breath Volume	2.085	06:32
Air Blank	0.000	06:32
Control Test	0.076	06:33
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

Data Download Date/Time: 04/25/2023 01:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BURNETTE,BRANDON
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HEMMINGWAY

DATE/TIME OF TEST: 04/04/2023 01:43:45

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.078	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.078	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 04/25/2023 01:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,B,T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: KUNKLE

DATE/TIME OF TEST: 04/05/2023 22:27:28

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.078	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.224	22:22
Breath Volume	2.375	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.227	22:25
Breath Volume	1.882	22:25
Air Blank	0.000	22:26
Control Test	0.078	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 04/25/2023 01:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,J,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ALBRECHT,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 04/06/2023 10:49:34

Diagnostics Check	OK	10:42
Air Blank	0.000	10:43
Control Test	0.079	10:43
Air Blank	0.000	10:43
Subject Sample # 1	0.000	10:45
Breath Volume	2.367	10:45
Air Blank	0.000	10:45
Air Blank	0.000	10:47
Subject Sample # 2	0.000	10:48
Breath Volume	1.445	10:48
Air Blank	0.000	10:48
Control Test	0.079	10:48
Air Blank	0.000	10:49
Diagnostics Check	OK	10:49

Data Download Date/Time: 04/25/2023 01:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,MATTHEW
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 04/08/2023 03:30:06

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.077	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.198	03:24
Breath Volume	3.031	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.184	03:28
Breath Volume	3.105	03:28
Air Blank	0.000	03:29
Control Test	0.076	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 04/25/2023 01:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: VAZQUEZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/09/2023 00:41:16

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.078	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.174	00:36
Breath Volume	3.195	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.174	00:39
Breath Volume	3.132	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 04/25/2023 01:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB',N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BURNETTE,BRANDON
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LEBRON AYALA

DATE/TIME OF TEST: 04/09/2023 01:18:26

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.101	01:12
Breath Volume	1.207	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.098	01:16
Breath Volume	1.804	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 04/25/2023 01:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,B,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CHILICK,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 04/14/2023 21:39:38

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.079	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.322	21:35
Breath Volume	2.105	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.323	21:38
Breath Volume	2.726	21:38
Air Blank	0.000	21:38
Control Test	0.078	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 04/25/2023 01:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,B,T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/20/2023 03:46:21

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.095	03:41
Breath Volume	2.371	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.097	03:44
Breath Volume	3.042	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 04/25/2023 01:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,B,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 04/21/2023 01:28:10

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.019	01:23
Breath Volume	4.000	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.018	01:26
Breath Volume	4.531	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 04/25/2023 01:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICKETT,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BURNETTE,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 04/22/2023 03:54:17

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.078	03:52	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Control Test	0.078	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 04/25/2023 01:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,B,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DILL,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 04/25/2023 20:40:29

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.079	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.208	20:34
Breath Volume	2.050	20:34
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.212	20:38
Breath Volume	1.761	20:38
Air Blank	0.000	20:39
Control Test	0.078	20:39
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 05/23/2023 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BURNETT,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: NETTLES

DATE/TIME OF TEST: 04/27/2023 03:22:25

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.079	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 05/23/2023 21:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,B,T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 04/29/2023 03:17:25

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.150	03:13
Breath Volume	2.527	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.158	03:15
Breath Volume	2.703	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 05/23/2023 21:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: VASQUEZ PORTILLO

DATE/TIME OF TEST: 04/02/2023 03:40:10

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.243	03:35
Breath Volume	4.414	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.236	03:38
Breath Volume	4.406	03:38
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 04/20/2023 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/08/2023 04:29:03

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.200	04:24
Breath Volume	2.746	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.189	04:27
Breath Volume	2.867	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:29

Data Download Date/Time: 04/20/2023 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 04/09/2023 00:24:22

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.157	00:19
Breath Volume	1.242	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.168	00:22
Breath Volume	1.894	00:22
Air Blank	0.000	00:23
Control Test	0.078	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 04/20/2023 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL, NICHOLAS, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: CARTER, JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: NATAL TORRES

DATE/TIME OF TEST: 04/23/2023 03:54:11

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.135	03:49
Breath Volume	1.648	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.131	03:52
Breath Volume	1.664	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 05/30/2023 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DENNEY

DATE/TIME OF TEST: 04/28/2023 04:04:31

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.169	03:59
Breath Volume	1.628	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.176	04:03
Breath Volume	2.441	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 05/30/2023 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MATIAS FERIA

DATE/TIME OF TEST: 04/30/2023 04:09:05

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.185	04:04
Breath Volume	2.898	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample # 2	0.192	04:07
Breath Volume	2.843	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 05/30/2023 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFONTANT, SEDONA, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: BAGGETT, TYLER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: MCDUFFIE

DATE/TIME OF TEST: 04/30/2023 20:46:45

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:35	
Control Test	0.079	20:36	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:39	No Sample Provided
Breath Volume	0.000	20:39	
Air Blank	0.000	20:40	
Air Blank	0.000	20:42	
Subject Sample	# 2	20:45	No Sample Provided
Breath Volume	0.000	20:45	
Air Blank	0.000	20:45	
Control Test	0.080	20:46	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

Data Download Date/Time: 05/30/2023 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE, AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: HESTAND

DATE/TIME OF TEST: 04/02/2023 03:25:28

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.078	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.166	03:21
Breath Volume	3.179	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.169	03:23
Breath Volume	2.785	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 04/24/2023 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 03/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SAPP

DATE/TIME OF TEST: 04/16/2023 02:32:19

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.080	02:27
Breath Volume	1.890	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:29
Subject Sample # 2	0.073	02:30
Breath Volume	1.324	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 04/24/2023 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: BENINCASA,VINCENT
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MARCHAN BATA

DATE/TIME OF TEST: 04/03/2023 02:32:49

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.077	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.115	02:27
Breath Volume	1.453	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.131	02:31
Breath Volume	3.355	02:31
Air Blank	0.000	02:31
Control Test	0.077	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 04/12/2023 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: BENINCASA,VINCENT
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: ROJAS LOPEZ

DATE/TIME OF TEST: 04/03/2023 05:44:41

Diagnostics Check	OK	05:37
Air Blank	0.000	05:38
Control Test	0.078	05:38
Air Blank	0.000	05:39
Subject Sample # 1	0.161	05:39
Breath Volume	1.921	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.172	05:43
Breath Volume	3.636	05:43
Air Blank	0.000	05:43
Control Test	0.078	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 04/12/2023 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: RICANO,CHRISTIAN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GONZALEZ MENDEZ

DATE/TIME OF TEST: 04/25/2023 00:59:33

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.078	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.221	00:55
Breath Volume	2.027	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.226	00:58
Breath Volume	1.863	00:58
Air Blank	0.000	00:58
Control Test	0.078	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 05/03/2023 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: BENINCASA,VINCENT
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 04/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: BATRES

DATE/TIME OF TEST: 04/26/2023 04:08:13

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.078	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.000	04:04
Breath Volume	2.363	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.000	04:06
Breath Volume	3.039	04:06
Air Blank	0.000	04:07
Control Test	0.078	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 05/03/2023 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CAMPANA,NICK
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 03/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: GIANOUTSOS

DATE/TIME OF TEST: 04/06/2023 02:35:20

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.211	02:30
Breath Volume	4.101	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.218	02:33
Breath Volume	4.464	02:33
Air Blank	0.000	02:34
Control Test	0.078	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 04/10/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIMMERMAN,J,
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: JASPER,A
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: BERGER

DATE/TIME OF TEST: 04/08/2023 21:32:17

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.172	21:27
Breath Volume	3.000	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.173	21:30
Breath Volume	4.039	21:30
Air Blank	0.000	21:31
Control Test	0.078	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 04/10/2023 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,THOMAS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: COLLINS,THOMAS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/13/2023 19:48:34

Diagnostics Check	OK	19:39
Air Blank	0.000	19:39
Control Test	0.078	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.192	19:41
Breath Volume	1.457	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:43
Subject Sample # 2	0.207	19:46
Breath Volume	1.730	19:46
Air Blank	0.000	19:47
Control Test	0.078	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 06/29/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,THOMAS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: STAHL,ASHLEY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/14/2023 03:32:51

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.077	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.206	03:28
Breath Volume	1.710	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.205	03:31
Breath Volume	1.500	03:31
Air Blank	0.000	03:31
Control Test	0.078	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 06/29/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CAMPANA,NICK
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/16/2023 00:08:28

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	00:01	No Sample Provided
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Air Blank	0.000	00:03	
Subject Sample	# 2	00:07	No Sample Provided
Breath Volume	0.000	00:07	
Air Blank	0.000	00:07	
Control Test	0.079	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 06/29/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: JASPER,A
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: WEATHERSTON

DATE/TIME OF TEST: 04/18/2023 02:02:27

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.164	01:57
Breath Volume	1.554	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.152	02:00
Breath Volume	1.730	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 06/29/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CURBELO, ANDREW
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CLEAVER

DATE/TIME OF TEST: 04/23/2023 01:09:39

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.204	01:05
Breath Volume	3.160	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.201	01:08
Breath Volume	3.195	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 06/29/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENNELL,RUSSELL,M
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PENNELL,RUSSELL
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 04/23/2023 17:44:45

Diagnostics Check	OK	17:38
Air Blank	0.000	17:38
Control Test	0.079	17:39
Air Blank	0.000	17:39
Subject Sample # 1	0.196	17:40
Breath Volume	4.761	17:40
Air Blank	0.000	17:40
Air Blank	0.000	17:42
Subject Sample # 2	0.197	17:43
Breath Volume	2.984	17:43
Air Blank	0.000	17:43
Control Test	0.077	17:44
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time: 06/29/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CAMPANA,N
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 04/26/2023 02:39:00

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.220	02:33
Breath Volume	1.664	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.232	02:37
Breath Volume	3.460	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 06/29/2023 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,Z
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: JASPER,ANDREW
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 04/28/2023 02:57:48

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.187	02:53
Breath Volume	2.050	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.183	02:56
Breath Volume	2.636	02:56
Air Blank	0.000	02:56
Control Test	0.077	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 06/29/2023 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,J,F
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: BROWN,J
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 04/07/2023 01:23:05

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.047	01:18
Breath Volume	1.949	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.043	01:21
Breath Volume	1.707	01:21
Air Blank	0.000	01:22
Control Test	0.081	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 04/28/2023 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: BRISENO,NATHANIEL
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: PEDRAZA CHARREZ

DATE/TIME OF TEST: 04/09/2023 05:38:02

Diagnostics Check	OK	05:27	
Air Blank	0.000	05:28	
Control Test	0.080	05:28	
Air Blank	0.000	05:28	
Subject Sample	# 1	05:31	No Sample Provided
Breath Volume	0.000	05:31	
Air Blank	0.000	05:32	
Air Blank	0.000	05:34	
Subject Sample	# 2	05:36	Subject Test Refused
Breath Volume	0.000	05:36	
Air Blank	0.000	05:37	
Control Test	0.081	05:37	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:38	

Data Download Date/Time: 04/28/2023 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: INSALACO,KAILI
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 04/18/2023 20:22:36

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:14	
Control Test	0.081	20:14	
Air Blank	0.000	20:15	
Subject Sample	# 1	20:18	No Sample Provided
Breath Volume	0.000	20:18	
Air Blank	0.000	20:19	
Air Blank	0.000	20:21	
Subject Sample	# 2	20:21	Subject Test Refused
Breath Volume	0.000	20:21	
Air Blank	0.000	20:21	
Control Test	0.081	20:22	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

Data Download Date/Time: 04/28/2023 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,J,F
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: BRISENO,N
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/21/2023 03:30:18

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.174	03:25
Breath Volume	1.207	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.175	03:28
Breath Volume	1.394	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 04/28/2023 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEWTON,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FUERETT

DATE/TIME OF TEST: 04/01/2023 01:17:40

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.081	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.080	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 04/19/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DELEO

DATE/TIME OF TEST: 04/01/2023 23:53:48

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.082	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.082	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 04/19/2023 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: DELEO

DATE/TIME OF TEST: 04/01/2023 23:59:45

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.082	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Control Test	0.082	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 04/19/2023 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PEREIRA,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/02/2023 01:17:17

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.082	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.172	01:12
Breath Volume	1.269	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.166	01:15
Breath Volume	1.101	01:15
Air Blank	0.000	01:16
Control Test	0.082	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 04/19/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BLACKSHEAR

DATE/TIME OF TEST: 04/02/2023 04:39:10

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.082	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.138	04:34
Breath Volume	2.167	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:36
Subject Sample # 2	0.135	04:37
Breath Volume	2.347	04:37
Air Blank	0.000	04:38
Control Test	0.082	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 04/19/2023 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 12:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:17

SUBJECT'S LAST NAME: AMAYA MARTINEZ

DATE/TIME OF TEST: 04/02/2023 12:46:32

Diagnostics Check	OK	12:39
Air Blank	0.000	12:40
Control Test	0.081	12:40
Air Blank	0.000	12:41
Subject Sample # 1	0.195	12:41
Breath Volume	1.691	12:41
Air Blank	0.000	12:42
Air Blank	0.000	12:44
Subject Sample # 2	0.203	12:45
Breath Volume	2.082	12:45
Air Blank	0.000	12:45
Control Test	0.082	12:45
Air Blank	0.000	12:46
Diagnostics Check	OK	12:46

Data Download Date/Time: 04/19/2023 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 04/06/2023 20:27:46

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.082	20:25	
Air Blank	0.000	20:26	
Subject Sample	# 1	20:26	Subject Test Refused
Breath Volume	0.000	20:26	
Air Blank	0.000	20:26	
Control Test	0.082	20:27	
Air Blank	0.000	20:27	
Diagnostics Check	OK	20:27	

Data Download Date/Time: 04/19/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: HEGARTY III

DATE/TIME OF TEST: 04/07/2023 04:35:01

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.082	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.160	04:30
Breath Volume	1.585	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:32
Subject Sample # 2	0.165	04:33
Breath Volume	1.585	04:33
Air Blank	0.000	04:34
Control Test	0.082	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:35

Data Download Date/Time: 04/19/2023 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THACKER, ADAM, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN, B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: THACKER

DATE/TIME OF TEST: 04/07/2023 04:57:56

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:55	
Control Test	0.082	04:55	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:57	
Control Test	0.082	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BANNISTER,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GILLIS

DATE/TIME OF TEST: 04/07/2023 20:26:49

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.082	20:21
Air Blank	0.000	20:21
Subject Sample # 1	0.141	20:22
Breath Volume	1.925	20:22
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.139	20:25
Breath Volume	1.535	20:25
Air Blank	0.000	20:25
Control Test	0.082	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: ROGERS III

DATE/TIME OF TEST: 04/08/2023 02:52:14

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.082	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.000	02:47
Breath Volume	2.308	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.000	02:50
Breath Volume	2.527	02:50
Air Blank	0.000	02:51
Control Test	0.082	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,OFC B,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: RODRIGUEZ,OFC B
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: MACIAS-RANGEL

DATE/TIME OF TEST: 04/08/2023 16:08:17

Diagnostics Check	OK	16:01
Air Blank	0.000	16:02
Control Test	0.083	16:02
Air Blank	0.000	16:02
Subject Sample # 1	0.000	16:03
Breath Volume	3.570	16:03
Air Blank	0.000	16:04
Air Blank	0.000	16:06
Subject Sample # 2	0.000	16:06
Breath Volume	3.582	16:06
Air Blank	0.000	16:07
Control Test	0.082	16:07
Air Blank	0.000	16:08
Diagnostics Check	OK	16:08

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: SCOLES

DATE/TIME OF TEST: 04/08/2023 22:20:34

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:17	
Control Test	0.083	22:18	
Air Blank	0.000	22:18	
Subject Sample	# 1	22:19	Subject Test Refused
Breath Volume	0.000	22:19	
Air Blank	0.000	22:19	
Control Test	0.082	22:19	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 04/09/2023 00:28:58

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.082	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:28	
Control Test	0.082	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORRIS,BRIDGETT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: PRODAN

DATE/TIME OF TEST: 04/10/2023 04:00:58

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.083	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.238	03:56
Breath Volume	1.484	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.232	03:59
Breath Volume	1.773	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS, DEP BRITT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: HAIRE

DATE/TIME OF TEST: 04/11/2023 01:55:48

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.083	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.164	01:51
Breath Volume	1.667	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.165	01:54
Breath Volume	1.386	01:54
Air Blank	0.000	01:54
Control Test	0.082	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: URDANETA CHACIN

DATE/TIME OF TEST: 04/11/2023 02:44:00

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.083	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:39	No Sample Provided
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.082	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 04/11/2023 03:16:51

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.082	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.126	03:12
Breath Volume	2.183	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.129	03:15
Breath Volume	1.109	03:15
Air Blank	0.000	03:15
Control Test	0.082	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TOWNES,FALANZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 11:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: GOLLOB

DATE/TIME OF TEST: 04/11/2023 12:16:29

Diagnostics Check	OK	12:10
Air Blank	0.000	12:10
Control Test	0.082	12:11
Air Blank	0.000	12:11
Subject Sample # 1	0.206	12:12
Breath Volume	2.480	12:12
Air Blank	0.000	12:12
Air Blank	0.000	12:14
Subject Sample # 2	0.211	12:14
Breath Volume	2.460	12:14
Air Blank	0.000	12:15
Control Test	0.082	12:15
Air Blank	0.000	12:16
Diagnostics Check	OK	12:16

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MARGIASSO

DATE/TIME OF TEST: 04/13/2023 02:17:34

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.083	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.000	02:13
Breath Volume	1.238	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.000	02:16
Breath Volume	1.218	02:16
Air Blank	0.000	02:16
Control Test	0.083	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUTH,TIMOTHY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: RAMBAUM

DATE/TIME OF TEST: 04/13/2023 06:25:19

Diagnostics Check	OK	06:18
Air Blank	0.000	06:19
Control Test	0.083	06:19
Air Blank	0.000	06:20
Subject Sample # 1	0.081	06:20
Breath Volume	1.882	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.083	06:23
Breath Volume	2.289	06:23
Air Blank	0.000	06:24
Control Test	0.082	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/14/2023 22:24:15

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.083	22:22	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Control Test	0.082	22:23	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 04/15/2023 01:51:42

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.082	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.000	01:47
Breath Volume	2.308	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.000	01:50
Breath Volume	2.460	01:50
Air Blank	0.000	01:50
Control Test	0.082	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BUCH

DATE/TIME OF TEST: 04/15/2023 02:30:58

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.083	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.126	02:26
Breath Volume	2.500	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.124	02:29
Breath Volume	2.082	02:29
Air Blank	0.000	02:30
Control Test	0.082	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GOODPASTOR,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 04/15/2023 04:07:12

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.083	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.000	04:02
Breath Volume	1.925	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.000	04:05
Breath Volume	2.398	04:05
Air Blank	0.000	04:06
Control Test	0.082	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: MICHALAK

DATE/TIME OF TEST: 04/15/2023 04:32:13

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:29	
Control Test	0.083	04:30	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.082	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: EDDINGS

DATE/TIME OF TEST: 04/15/2023 23:09:50

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.083	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.000	23:05
Breath Volume	1.621	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.000	23:08
Breath Volume	1.671	23:08
Air Blank	0.000	23:08
Control Test	0.082	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCOTT,A
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HOLLIS

DATE/TIME OF TEST: 04/16/2023 00:43:02

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.082	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Control Test	0.082	00:42	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRADY,JAMES
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 04/16/2023 06:51:57

Diagnostics Check	OK	06:45
Air Blank	0.000	06:46
Control Test	0.083	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.148	06:47
Breath Volume	2.632	06:47
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.146	06:50
Breath Volume	2.792	06:50
Air Blank	0.000	06:51
Control Test	0.082	06:51
Air Blank	0.000	06:51
Diagnostics Check	OK	06:51

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: STIDHAM

DATE/TIME OF TEST: 04/16/2023 21:12:40

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.083	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.109	21:08
Breath Volume	1.964	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.107	21:11
Breath Volume	1.457	21:11
Air Blank	0.000	21:11
Control Test	0.082	21:12
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HERNANDEZ,LUIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: RABICE

DATE/TIME OF TEST: 04/17/2023 05:10:05

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.083	05:08	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:08	Subject Test Refused
Breath Volume	0.000	05:08	
Air Blank	0.000	05:09	
Control Test	0.082	05:09	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

Data Download Date/Time: 04/19/2023 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: JANEY

DATE/TIME OF TEST: 04/27/2023 20:43:12

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.082	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.184	20:38
Breath Volume	1.511	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.180	20:41
Breath Volume	1.488	20:41
Air Blank	0.000	20:42
Control Test	0.081	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 05/10/2023 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG115904
DRY GAS EXPIRY DATE: 06/08/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/29/2023 02:53:24

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.082	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.081	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 05/10/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/01/2023 01:44:22

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.083	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.000	01:40
Breath Volume	1.753	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.000	01:42
Breath Volume	2.519	01:42
Air Blank	0.000	01:43
Control Test	0.083	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 04/01/2023 04:07:32

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.083	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.058	04:02
Breath Volume	1.910	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.058	04:05
Breath Volume	1.554	04:05
Air Blank	0.000	04:06
Control Test	0.082	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 04/01/2023 08:57:11

Diagnostics Check	OK	08:48
Air Blank	0.000	08:48
Control Test	0.083	08:49
Air Blank	0.000	08:49
Subject Sample # 1	0.134	08:52
Breath Volume	2.183	08:52
Air Blank	0.000	08:53
Air Blank	0.000	08:54
Subject Sample # 2	0.131	08:55
Breath Volume	1.773	08:55
Air Blank	0.000	08:56
Control Test	0.078	08:56
Air Blank	0.000	08:57
Diagnostics Check	OK	08:57

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPENCER,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:57

SUBJECT'S LAST NAME: GARDENER

DATE/TIME OF TEST: 04/01/2023 13:25:03

Diagnostics Check	OK	13:21	
Air Blank	0.000	13:22	
Control Test	0.083	13:22	
Air Blank	0.000	13:22	
Subject Sample	# 1	13:23	Subject Test Refused
Breath Volume	0.000	13:23	
Air Blank	0.000	13:23	
Control Test	0.081	13:23	
Air Blank	0.000	13:24	
Diagnostics Check	Fail	13:24	Analytical Stability Test
Air Blank	0.000	13:25	

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ECHEVERRIAVIDAL,X
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MARTINEZ RUIZ

DATE/TIME OF TEST: 04/02/2023 02:25:08

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.083	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.195	02:20
Breath Volume	2.468	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.188	02:23
Breath Volume	1.417	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/02/2023 03:13:00

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.082	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.080	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 04/02/2023 03:54:14

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.082	03:52	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Control Test	0.080	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MALDONADO ORTIZ

DATE/TIME OF TEST: 04/02/2023 20:20:37

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.082	20:14
Air Blank	0.000	20:15
Subject Sample # 1	0.165	20:15
Breath Volume	2.210	20:15
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.170	20:18
Breath Volume	2.574	20:18
Air Blank	0.000	20:19
Control Test	0.082	20:19
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: CRUZ ANTUN

DATE/TIME OF TEST: 04/02/2023 22:08:35

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.083	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 1	22:07	Subject Test Refused
Breath Volume	0.000	22:07	
Air Blank	0.000	22:07	
Control Test	0.081	22:07	
Air Blank	0.000	22:08	
Diagnostics Check	OK	22:08	

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ORDONEZ MENDEZ

DATE/TIME OF TEST: 04/03/2023 01:01:38

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.083	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.230	00:56
Breath Volume	1.683	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.230	00:59
Breath Volume	1.554	00:59
Air Blank	0.000	01:00
Control Test	0.082	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ERAM

DATE/TIME OF TEST: 04/04/2023 01:10:12

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.082	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Control Test	0.081	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RIVERA CRUZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: FINNEY

DATE/TIME OF TEST: 04/05/2023 02:29:32

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.194	02:24
Breath Volume	1.617	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.203	02:27
Breath Volume	1.390	02:27
Air Blank	0.000	02:28
Control Test	0.082	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: PRUITT

DATE/TIME OF TEST: 04/05/2023 03:45:59

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.083	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.082	03:41
Breath Volume	1.765	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.081	03:44
Breath Volume	1.511	03:44
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/06/2023 02:55:33

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.083	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.079	02:50
Breath Volume	2.035	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.077	02:53
Breath Volume	1.464	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DUNNEBACK,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/06/2023 18:57:42

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.083	18:51
Air Blank	0.000	18:52
Subject Sample # 1	0.327	18:52
Breath Volume	1.179	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.316	18:55
Breath Volume	1.234	18:55
Air Blank	0.000	18:56
Control Test	0.082	18:57
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITMAR,V
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/06/2023 22:36:14

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:33	
Control Test	0.083	22:34	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:34	Subject Test Refused
Breath Volume	0.000	22:34	
Air Blank	0.000	22:35	
Control Test	0.080	22:35	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: SALESCLAUDIO

DATE/TIME OF TEST: 04/07/2023 00:31:05

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.082	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.187	00:26
Breath Volume	1.835	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.190	00:29
Breath Volume	2.222	00:29
Air Blank	0.000	00:30
Control Test	0.082	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ARCHBOLD,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/07/2023 06:23:22

Diagnostics Check	OK	06:20	
Air Blank	0.000	06:20	
Control Test	0.084	06:21	
Air Blank	0.000	06:21	
Subject Sample	# 1	06:21	Subject Test Refused
Breath Volume	0.000	06:21	
Air Blank	0.000	06:22	
Control Test	0.080	06:22	
Air Blank	0.000	06:23	
Diagnostics Check	OK	06:23	

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: BELTRAN

DATE/TIME OF TEST: 04/07/2023 23:20:45

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:18	
Control Test	0.083	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.080	23:20	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MELO

DATE/TIME OF TEST: 04/08/2023 04:04:20

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.084	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.123	03:59
Breath Volume	1.417	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.130	04:02
Breath Volume	1.546	04:02
Air Blank	0.000	04:03
Control Test	0.082	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 04/21/2023 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: WOODSON

DATE/TIME OF TEST: 04/08/2023 23:36:33

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.083	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.061	23:32
Breath Volume	1.656	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.061	23:34
Breath Volume	1.699	23:34
Air Blank	0.000	23:35
Control Test	0.081	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: DENSON

DATE/TIME OF TEST: 04/09/2023 04:42:19

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.083	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.136	04:37
Breath Volume	1.429	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.135	04:40
Breath Volume	1.660	04:40
Air Blank	0.000	04:41
Control Test	0.084	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: NAVARRO BONILLA

DATE/TIME OF TEST: 04/12/2023 01:54:29

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.083	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.127	01:49
Breath Volume	1.808	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.124	01:52
Breath Volume	1.570	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: HANNA

DATE/TIME OF TEST: 04/12/2023 03:21:11

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.083	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: IGNACIO SILVESTRE

DATE/TIME OF TEST: 04/13/2023 01:26:58

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.083	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.000	01:22
Breath Volume	3.156	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.000	01:25
Breath Volume	2.535	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: FILLIO

DATE/TIME OF TEST: 04/13/2023 16:49:24

Diagnostics Check	OK	16:43
Air Blank	0.000	16:43
Control Test	0.082	16:43
Air Blank	0.000	16:44
Subject Sample # 1	0.025	16:44
Breath Volume	2.531	16:44
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.023	16:47
Breath Volume	2.917	16:47
Air Blank	0.000	16:48
Control Test	0.080	16:48
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/14/2023 23:38:02

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.083	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.176	23:32
Breath Volume	1.464	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.174	23:36
Breath Volume	1.671	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:38

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: DEUTSCH

DATE/TIME OF TEST: 04/15/2023 22:54:00

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.083	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.000	22:49
Breath Volume	3.054	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:52
Subject Sample # 2	0.000	22:52
Breath Volume	1.562	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:54

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 04/16/2023 00:59:00

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.083	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.032	00:54
Breath Volume	1.558	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:56
Subject Sample # 2	0.037	00:57
Breath Volume	1.691	00:57
Air Blank	0.000	00:58
Control Test	0.083	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:59

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/16/2023 02:35:35

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.082	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.081	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BOIUN

DATE/TIME OF TEST: 04/16/2023 03:48:42

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.083	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.171	03:43
Breath Volume	1.992	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.172	03:46
Breath Volume	1.265	03:46
Air Blank	0.000	03:47
Control Test	0.084	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 04/16/2023 09:18:45

Diagnostics Check	OK	09:15	
Air Blank	0.000	09:16	
Control Test	0.082	09:16	
Air Blank	0.000	09:17	
Subject Sample	# 1	09:17	Subject Test Refused
Breath Volume	0.000	09:17	
Air Blank	0.000	09:17	
Control Test	0.080	09:18	
Air Blank	0.000	09:18	
Diagnostics Check	OK	09:18	

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GIBSON,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: DARLING

DATE/TIME OF TEST: 04/16/2023 22:34:19

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.083	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.127	22:29
Breath Volume	2.335	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:31
Subject Sample # 2	0.132	22:32
Breath Volume	2.558	22:32
Air Blank	0.000	22:33
Control Test	0.082	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: CANTIN

DATE/TIME OF TEST: 04/17/2023 07:12:11

Diagnostics Check	OK	07:05
Air Blank	0.000	07:06
Control Test	0.082	07:06
Air Blank	0.000	07:07
Subject Sample # 1	0.084	07:07
Breath Volume	1.828	07:07
Air Blank	0.000	07:08
Air Blank	0.000	07:10
Subject Sample # 2	0.086	07:10
Breath Volume	2.046	07:10
Air Blank	0.000	07:11
Control Test	0.083	07:11
Air Blank	0.000	07:12
Diagnostics Check	OK	07:12

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: O'CONNELL,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: CALLICOTT

DATE/TIME OF TEST: 04/18/2023 23:36:22

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.083	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Control Test	0.080	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHIPP,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 04/19/2023 00:51:10

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.084	00:49	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:50	
Control Test	0.079	00:50	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: LYON,L
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: DI CARLO

DATE/TIME OF TEST: 04/20/2023 20:20:56

Diagnostics Check	OK	20:18	
Air Blank	0.000	20:18	
Control Test	0.083	20:18	
Air Blank	0.000	20:19	
Subject Sample	# 1	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:20	
Control Test	0.080	20:20	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 04/21/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:43

SUBJECT'S LAST NAME: PARISH

DATE/TIME OF TEST: 04/21/2023 13:10:03

Diagnostics Check	OK	13:03
Air Blank	0.000	13:04
Control Test	0.082	13:04
Air Blank	0.000	13:05
Subject Sample # 1	0.000	13:05
Breath Volume	2.355	13:05
Air Blank	0.000	13:06
Air Blank	0.000	13:08
Subject Sample # 2	0.000	13:08
Breath Volume	2.496	13:08
Air Blank	0.000	13:09
Control Test	0.083	13:09
Air Blank	0.000	13:09
Diagnostics Check	OK	13:10

Data Download Date/Time: 05/24/2023 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/22/2023 01:24:47

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.082	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:20	No Sample Provided
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 2	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.083	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 05/24/2023 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 04/22/2023 15:45:16

Diagnostics Check	OK	15:42	
Air Blank	0.000	15:42	
Control Test	0.083	15:42	
Air Blank	0.000	15:43	
Subject Sample	# 1	15:43	Subject Test Refused
Breath Volume	0.000	15:43	
Air Blank	0.000	15:44	
Control Test	0.081	15:44	
Air Blank	0.000	15:45	
Diagnostics Check	OK	15:45	

Data Download Date/Time: 05/24/2023 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: WALSTON

DATE/TIME OF TEST: 04/22/2023 22:12:36

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:10	
Control Test	0.083	22:10	
Air Blank	0.000	22:10	
Subject Sample	# 1	22:11	Subject Test Refused
Breath Volume	0.000	22:11	
Air Blank	0.000	22:11	
Control Test	0.082	22:11	
Air Blank	0.000	22:12	
Diagnostics Check	OK	22:12	

Data Download Date/Time: 05/24/2023 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MARTINEZ DE PAZ

DATE/TIME OF TEST: 04/22/2023 22:57:19

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.083	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.125	22:52
Breath Volume	1.707	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.120	22:55
Breath Volume	1.968	22:55
Air Blank	0.000	22:56
Control Test	0.083	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 05/24/2023 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAINES,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: LOPEZ MEJIA

DATE/TIME OF TEST: 04/23/2023 06:08:33

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:06	
Control Test	0.083	06:06	
Air Blank	0.000	06:06	
Subject Sample	# 1	06:07	Subject Test Refused
Breath Volume	0.000	06:07	
Air Blank	0.000	06:07	
Control Test	0.082	06:07	
Air Blank	0.000	06:08	
Diagnostics Check	OK	06:08	

Data Download Date/Time: 05/24/2023 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: RAMOS-PRATTS

DATE/TIME OF TEST: 04/23/2023 16:30:42

Diagnostics Check	OK	16:24
Air Blank	0.000	16:24
Control Test	0.083	16:24
Air Blank	0.000	16:25
Subject Sample # 1	0.000	16:26
Breath Volume	1.636	16:26
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.000	16:29
Breath Volume	1.761	16:29
Air Blank	0.000	16:29
Control Test	0.084	16:30
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 05/24/2023 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/23/2023 21:19:13

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.083	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.205	21:13
Breath Volume	1.375	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.194	21:17
Breath Volume	2.253	21:17
Air Blank	0.000	21:18
Control Test	0.078	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MANUKYAN,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: TAMAYO

DATE/TIME OF TEST: 04/23/2023 22:18:17

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.082	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:16	Subject Test Refused
Breath Volume	0.000	22:16	
Air Blank	0.000	22:17	
Control Test	0.082	22:17	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 04/23/2023 23:41:06

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:38	
Control Test	0.082	23:39	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:40	
Control Test	0.081	23:40	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SANTIAGO CORDOVA

DATE/TIME OF TEST: 04/24/2023 03:52:58

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.083	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.146	03:48
Breath Volume	3.656	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.148	03:51
Breath Volume	2.769	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/26/2023 16:02:33

Diagnostics Check	OK	15:55
Air Blank	0.000	15:55
Control Test	0.082	15:56
Air Blank	0.000	15:56
Subject Sample # 1	0.164	15:57
Breath Volume	3.210	15:57
Air Blank	0.000	15:58
Air Blank	0.000	16:00
Subject Sample # 2	0.171	16:00
Breath Volume	2.531	16:00
Air Blank	0.000	16:01
Control Test	0.079	16:01
Air Blank	0.000	16:02
Diagnostics Check	OK	16:02

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ANSLEY,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: MATTOCKS

DATE/TIME OF TEST: 04/26/2023 17:58:45

Diagnostics Check	OK	17:50
Air Blank	0.000	17:51
Control Test	0.082	17:51
Air Blank	0.000	17:51
Subject Sample # 1	0.218	17:52
Breath Volume	2.171	17:52
Air Blank	0.000	17:53
Air Blank	0.000	17:55
Subject Sample # 2	0.214	17:56
Breath Volume	2.320	17:56
Air Blank	0.000	17:57
Control Test	0.081	17:58
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,E,V
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALEXANDRE,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: MCCARLEY

DATE/TIME OF TEST: 04/27/2023 18:30:43

Diagnostics Check	OK	18:24
Air Blank	0.000	18:25
Control Test	0.080	18:25
Air Blank	0.000	18:25
Subject Sample # 1	0.103	18:26
Breath Volume	1.753	18:26
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.102	18:29
Breath Volume	1.246	18:29
Air Blank	0.000	18:29
Control Test	0.079	18:30
Air Blank	0.000	18:30
Diagnostics Check	OK	18:30

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL, SHAQUANIA, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON, TORRIE
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 04/29/2023 16:02:06

Diagnostics Check	OK	15:59	
Air Blank	0.000	15:59	
Control Test	0.082	16:00	
Air Blank	0.000	16:00	
Subject Sample	# 1	16:00	Subject Test Refused
Breath Volume	0.000	16:00	
Air Blank	0.000	16:01	
Control Test	0.080	16:01	
Air Blank	0.000	16:02	
Diagnostics Check	OK	16:02	

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL, SHAQUANIA, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: CLAUSON

DATE/TIME OF TEST: 04/29/2023 17:53:14

Diagnostics Check	OK	17:47
Air Blank	0.000	17:47
Control Test	0.083	17:47
Air Blank	0.000	17:48
Subject Sample # 1	0.000	17:48
Breath Volume	3.351	17:48
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.000	17:51
Breath Volume	2.898	17:51
Air Blank	0.000	17:52
Control Test	0.084	17:52
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 05/24/2023 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,E,V
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ACALA SWEAT,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: VILLANUEVA MARTINEZ

DATE/TIME OF TEST: 04/30/2023 09:51:29

Diagnostics Check	OK	09:45
Air Blank	0.000	09:45
Control Test	0.083	09:45
Air Blank	0.000	09:46
Subject Sample # 1	0.167	09:47
Breath Volume	3.550	09:47
Air Blank	0.000	09:47
Air Blank	0.000	09:49
Subject Sample # 2	0.165	09:49
Breath Volume	2.394	09:49
Air Blank	0.000	09:50
Control Test	0.080	09:50
Air Blank	0.000	09:51
Diagnostics Check	OK	09:51

Data Download Date/Time: 05/24/2023 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,E,V
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/30/2023 18:07:39

Diagnostics Check	OK	18:01
Air Blank	0.000	18:01
Control Test	0.082	18:01
Air Blank	0.000	18:02
Subject Sample # 1	0.282	18:02
Breath Volume	2.410	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.281	18:05
Breath Volume	2.757	18:05
Air Blank	0.000	18:06
Control Test	0.081	18:07
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

Data Download Date/Time: 05/24/2023 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WALKER, STEVEN
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 03/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: SYKORA

DATE/TIME OF TEST: 04/01/2023 23:00:01

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.075	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.144	22:55
Breath Volume	2.105	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:57
Subject Sample # 2	0.135	22:58
Breath Volume	1.976	22:58
Air Blank	0.000	22:59
Control Test	0.077	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	23:00

Data Download Date/Time: 04/27/2023 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SHEAHEN,BRENDEN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/03/2023 20:51:23

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.080	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.243	20:47
Breath Volume	2.445	20:47
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.247	20:49
Breath Volume	2.089	20:49
Air Blank	0.000	20:50
Control Test	0.080	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 05/04/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SHAVER

DATE/TIME OF TEST: 04/22/2023 02:27:48

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.188	02:22
Breath Volume	1.792	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.183	02:26
Breath Volume	1.660	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 05/04/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: CORONADO,ZACHARY
ARRESTING OFFICER AGENCY: Bowling Green PD
LAST AGENCY INSPECTION: 03/31/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/22/2023 04:14:16

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.079	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.080	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 05/04/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVELL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: CANO SALAZAR

DATE/TIME OF TEST: 04/27/2023 21:32:39

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.080	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.236	21:24
Breath Volume	1.902	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.211	21:28
Breath Volume	1.195	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 3	0.226	21:31
Breath Volume	2.679	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 05/04/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVELL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: RANGEL

DATE/TIME OF TEST: 04/28/2023 05:17:35

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:06	
Control Test	0.081	05:06	
Air Blank	0.000	05:07	
Subject Sample	# 1 0.116	05:10	Volume Not Met
Breath Volume	0.035	05:10	
Air Blank	0.000	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 2 0.115	05:16	Volume Not Met
Breath Volume	0.652	05:16	
Air Blank	0.000	05:16	
Control Test	0.080	05:17	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 05/04/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVELL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: RANGEL

DATE/TIME OF TEST: 04/28/2023 05:43:19

Diagnostics Check	OK	05:31	
Air Blank	0.000	05:32	
Control Test	0.080	05:32	
Air Blank	0.000	05:33	
Subject Sample	# 1 0.134	05:36	Volume Not Met
Breath Volume	0.199	05:36	
Air Blank	0.000	05:36	
Air Blank	0.000	05:38	
Subject Sample	# 2 0.128	05:41	Volume Not Met
Breath Volume	0.613	05:41	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 05/04/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: SKOGLUND

DATE/TIME OF TEST: 04/29/2023 02:15:02

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.078	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.097	02:10
Breath Volume	1.843	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.091	02:13
Breath Volume	1.808	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:15

Data Download Date/Time: 05/30/2023 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/30/2023 02:11:59

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.078	02:08	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Control Test	0.077	02:11	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 05/30/2023 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: HERN

DATE/TIME OF TEST: 04/30/2023 03:27:30

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.078	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.077	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 05/30/2023 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: KNIGHTON

DATE/TIME OF TEST: 04/01/2023 01:43:27

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1 0.126	01:36	Slope Not Level
Breath Volume	1.207	01:36	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample	# 2 0.121	01:41	Volume Not Met
Breath Volume	0.343	01:41	
Air Blank	0.000	01:42	
Control Test	0.079	01:42	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: KNIGHTON

DATE/TIME OF TEST: 04/01/2023 01:52:08

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.078	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.138	01:47
Breath Volume	2.085	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.137	01:50
Breath Volume	1.777	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GOSHORN

DATE/TIME OF TEST: 04/02/2023 01:31:36

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.045	01:27
Breath Volume	1.812	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.042	01:30
Breath Volume	1.777	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: STOCKING

DATE/TIME OF TEST: 04/02/2023 03:41:46

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.124	03:37
Breath Volume	1.179	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.128	03:40
Breath Volume	1.863	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SPOTTS

DATE/TIME OF TEST: 04/02/2023 05:34:24

Diagnostics Check	OK	05:30	
Air Blank	0.000	05:31	
Control Test	0.079	05:31	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:33	Subject Test Refused
Breath Volume	0.000	05:33	
Air Blank	0.000	05:33	
Control Test	0.078	05:33	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: STOVER

DATE/TIME OF TEST: 04/06/2023 22:26:08

Diagnostics Check	OK	22:19
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.213	22:21
Breath Volume	3.113	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.200	22:24
Breath Volume	1.640	22:24
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CERMAK,OFC,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GRIGGS,JACK

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/07/2023 03:23:04

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.078	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.138	03:18
Breath Volume	1.460	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.139	03:21
Breath Volume	1.664	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:23

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/08/2023 03:57:31

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.143	03:53
Breath Volume	2.750	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.144	03:56
Breath Volume	2.175	03:56
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: BARTOLOME

DATE/TIME OF TEST: 04/09/2023 23:12:33

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:03	
Control Test	0.079	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1 0.192	23:07	
Breath Volume	3.507	23:07	
Air Blank	0.000	23:07	
Air Blank	0.000	23:09	
Subject Sample	# 2	23:10	RFI Detect
Breath Volume	0.539	23:10	
Air Blank	0.000	23:11	
Control Test	0.078	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: BARTOLOME

DATE/TIME OF TEST: 04/09/2023 23:24:41

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.183	23:19
Breath Volume	3.031	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.183	23:23
Breath Volume	3.480	23:23
Air Blank	0.000	23:23
Control Test	0.078	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SINGSON,ROLAND

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 16:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: MOUNTS

DATE/TIME OF TEST: 04/10/2023 17:39:38

Diagnostics Check	OK	17:33
Air Blank	0.000	17:33
Control Test	0.079	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.183	17:35
Breath Volume	3.871	17:35
Air Blank	0.000	17:35
Air Blank	0.000	17:37
Subject Sample # 2	0.181	17:38
Breath Volume	3.195	17:38
Air Blank	0.000	17:38
Control Test	0.078	17:39
Air Blank	0.000	17:39
Diagnostics Check	OK	17:39

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CLARK,RYAN,N

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GRIGGS,JACK

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 04/11/2023 00:01:46

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:55
Subject Sample # 2	0.146	23:56
Breath Volume	2.035	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.147	00:00
Breath Volume	1.867	00:00
Air Blank	0.000	00:00
Control Test	0.078	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANSEN, TRAVIS, J

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GRIGGS, JACK

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: STYERS

DATE/TIME OF TEST: 04/12/2023 03:23:16

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1 0.254	03:19	
Breath Volume	1.722	03:19	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 2	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.077	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 04/18/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/12/2023 22:32:47

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.079	22:30	
Air Blank	0.000	22:30	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.078	22:32	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 04/18/2023 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:34

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: BRODIE

DATE/TIME OF TEST: 04/13/2023 01:21:07

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.079	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:20	
Control Test	0.078	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 04/18/2023 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: ESPAILLAT

DATE/TIME OF TEST: 04/14/2023 02:05:54

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.078	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.078	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 04/18/2023 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/15/2023 03:19:50

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.078	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:12	No Sample Provided
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 2	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.078	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 04/18/2023 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,SHANE,C
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: NEVERLINE,DAVID
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 04/16/2023 23:16:41

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.078	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.217	23:12
Breath Volume	2.304	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.226	23:15
Breath Volume	1.726	23:15
Air Blank	0.000	23:15
Control Test	0.078	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 04/18/2023 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SINGSON,ROLAND

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 18:09

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: KULIYEVA

DATE/TIME OF TEST: 04/17/2023 19:31:57

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.078	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.287	19:27
Breath Volume	1.292	19:27
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.283	19:30
Breath Volume	1.511	19:30
Air Blank	0.000	19:31
Control Test	0.078	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 04/18/2023 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CERMAK,OFC,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GRIGGS,JACK

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: MEHRA

DATE/TIME OF TEST: 04/19/2023 22:09:23

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.079	22:05	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:08	
Control Test	0.078	22:08	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CERMAK,OFC,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GRIGGS,JACK

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: TEBBETTS

DATE/TIME OF TEST: 04/21/2023 03:12:32

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.079	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.079	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 20:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: CORWIN

DATE/TIME OF TEST: 04/21/2023 20:49:43

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:47	
Control Test	0.079	20:47	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:48	Subject Test Refused
Breath Volume	0.000	20:48	
Air Blank	0.000	20:48	
Control Test	0.079	20:49	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANSEN,T,J

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GRIGGS,J

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: PLUMSTEAD

DATE/TIME OF TEST: 04/26/2023 00:55:30

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.162	00:51
Breath Volume	1.699	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.166	00:54
Breath Volume	2.222	00:54
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALAMEDA,JORGE,L
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SANTANA,S
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: NELLIS

DATE/TIME OF TEST: 04/26/2023 17:36:14

Diagnostics Check	OK	17:33	
Air Blank	0.000	17:33	
Control Test	0.080	17:34	
Air Blank	0.000	17:34	
Subject Sample	# 1	17:34	Subject Test Refused
Breath Volume	0.000	17:34	
Air Blank	0.000	17:35	
Control Test	0.079	17:35	
Air Blank	0.000	17:36	
Diagnostics Check	OK	17:36	

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ALTAMIRANO

DATE/TIME OF TEST: 04/27/2023 01:08:56

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.079	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	01:01	No Sample Provided
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 2	01:07	Volume Not Met
Breath Volume	0.144	01:07	
Air Blank	0.000	01:08	
Control Test	0.079	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ALTAMIRANO

DATE/TIME OF TEST: 04/27/2023 01:20:56

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.117	01:14
Breath Volume	1.535	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.099	01:19
Breath Volume	1.269	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN, TRAVIS, J
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HANSEN, T
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: JAGGER

DATE/TIME OF TEST: 04/27/2023 02:20:46

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.079	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Control Test	0.079	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: HAMROCK

DATE/TIME OF TEST: 04/28/2023 02:07:07

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.202	02:02
Breath Volume	2.070	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.185	02:05
Breath Volume	1.898	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 04/29/2023 01:30:43

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.151	01:25
Breath Volume	1.718	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.150	01:29
Breath Volume	1.695	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: CARRILLO CALZADA

DATE/TIME OF TEST: 04/29/2023 04:53:44

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.228	04:49
Breath Volume	1.664	04:49
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.228	04:52
Breath Volume	1.214	04:52
Air Blank	0.000	04:52
Control Test	0.078	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: JIMENEZ GARCIA

DATE/TIME OF TEST: 04/29/2023 19:09:30

Diagnostics Check	OK	19:03
Air Blank	0.000	19:03
Control Test	0.079	19:03
Air Blank	0.000	19:04
Subject Sample # 1	0.202	19:04
Breath Volume	2.070	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.194	19:08
Breath Volume	1.832	19:08
Air Blank	0.000	19:08
Control Test	0.079	19:08
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 05/15/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: CASTANEDA CABALLERO

DATE/TIME OF TEST: 04/01/2023 03:46:09

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.050	03:40
Breath Volume	1.265	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.047	03:44
Breath Volume	3.199	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 04/11/2023 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BELCHER,ALEX,B

BREATH TEST OPERATOR AGENCY: Auburndale PD

ARRESTING OFFICER: BELCHER,ALEX

ARRESTING OFFICER AGENCY: Auburndale PD

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 04:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: SAINTJUSTE

DATE/TIME OF TEST: 04/01/2023 05:03:35

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.078	04:50	
Air Blank	0.000	04:50	
Subject Sample	# 1 0.163	04:53	Volume Not Met
Breath Volume	0.031	04:53	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample	# 2 0.179	04:57	
Breath Volume	1.312	04:57	
Air Blank	0.000	04:58	
Air Blank	0.000	05:00	
Subject Sample	# 3 0.178	05:02	
Breath Volume	2.085	05:02	
Air Blank	0.000	05:02	
Control Test	0.077	05:02	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 04/11/2023 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: HOOYMAN,DAVID

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 03/14/2023

VIOLATION CODE: DUI

ARREST TIME: 00:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: MCMULLEN

DATE/TIME OF TEST: 04/02/2023 01:09:01

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.078	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	01:02	No Sample Provided
Breath Volume	0.000	01:02	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 2	01:07	No Sample Provided
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Control Test	0.078	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 04/11/2023 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MONCION,CHRISTIAN,L

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: MONCION,CHRISTIAN

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 03:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: FORLEMU

DATE/TIME OF TEST: 04/02/2023 05:27:17

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:18	
Control Test	0.078	05:18	
Air Blank	0.000	05:19	
Subject Sample	# 1 0.128	05:22	Volume Not Met
Breath Volume	0.023	05:22	
Air Blank	0.000	05:23	
Air Blank	0.000	05:25	
Subject Sample	# 2	05:25	Subject Test Refused
Breath Volume	0.000	05:25	
Air Blank	0.000	05:26	
Control Test	0.077	05:26	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

Data Download Date/Time: 04/11/2023 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTOS,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUSTOS,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 04/04/2023 02:05:55

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:56	
Control Test	0.078	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1	02:00	No Sample Provided
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample	# 2	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Control Test	0.079	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 04/11/2023 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: WALLISER,SHAUN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 23:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: VALENZUELA

DATE/TIME OF TEST: 04/05/2023 00:57:30

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.176	00:52
Breath Volume	1.429	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.175	00:55
Breath Volume	1.687	00:55
Air Blank	0.000	00:56
Control Test	0.076	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 04/11/2023 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALDRON,DAVID,A

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: WOFFORD,CLINT

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 14:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: JANSEN

DATE/TIME OF TEST: 04/05/2023 15:49:09

Diagnostics Check	OK	15:42
Air Blank	0.000	15:42
Control Test	0.078	15:42
Air Blank	0.000	15:43
Subject Sample # 1	0.188	15:44
Breath Volume	1.449	15:44
Air Blank	0.000	15:44
Air Blank	0.000	15:46
Subject Sample # 2	0.194	15:47
Breath Volume	2.984	15:47
Air Blank	0.000	15:48
Control Test	0.078	15:48
Air Blank	0.000	15:49
Diagnostics Check	OK	15:49

Data Download Date/Time: 04/11/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:34

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 04/06/2023 23:35:42

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.079	23:30	
Air Blank	0.000	23:30	
Subject Sample	# 1 0.000	23:31	
Breath Volume	1.269	23:31	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 2 0.000	23:34	
Breath Volume	1.484	23:34	
Air Blank	0.000	23:34	
Control Test	0.079	23:34	
Air Blank		23:35	RFI Detect
Air Blank	0.000	23:35	

Data Download Date/Time: 04/11/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: WALLISER,SHAUN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/07/2023 23:16:14

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.184	23:11
Breath Volume	1.386	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.189	23:14
Breath Volume	2.105	23:14
Air Blank	0.000	23:15
Control Test	0.076	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 04/11/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOLNETH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: VITAL

DATE/TIME OF TEST: 04/08/2023 05:24:00

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:13	
Control Test	0.077	05:13	
Air Blank	0.000	05:13	
Subject Sample	# 1	05:17	No Sample Provided
Breath Volume	0.000	05:17	
Air Blank	0.000	05:17	
Air Blank	0.000	05:19	
Subject Sample	# 2	05:22	No Sample Provided
Breath Volume	0.000	05:22	
Air Blank	0.000	05:23	
Control Test	0.079	05:23	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 04/11/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BELCHER,ALEX,B

BREATH TEST OPERATOR AGENCY: Auburndale PD

ARRESTING OFFICER: BELCHER,ALEX

ARRESTING OFFICER AGENCY: Auburndale PD

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 19:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: DIAZ-GUTIERREZ

DATE/TIME OF TEST: 04/08/2023 19:49:46

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.078	19:44
Air Blank	0.000	19:44
Subject Sample # 1	0.067	19:45
Breath Volume	1.378	19:45
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.069	19:48
Breath Volume	2.468	19:48
Air Blank	0.000	19:48
Control Test	0.079	19:49
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 04/11/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: WALLISER,SHAUN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 21:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 04/08/2023 22:35:10

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.078	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.152	22:29
Breath Volume	1.722	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.145	22:33
Breath Volume	1.367	22:33
Air Blank	0.000	22:34
Control Test	0.077	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 04/11/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: STANCIL

DATE/TIME OF TEST: 04/09/2023 02:26:39

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.000	02:22
Breath Volume	3.164	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.000	02:25
Breath Volume	2.070	02:25
Air Blank	0.000	02:25
Control Test	0.077	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 04/11/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: WALLISER,SHAUN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ASSINI

DATE/TIME OF TEST: 04/10/2023 00:45:55

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.078	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.113	00:40
Breath Volume	1.750	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.124	00:44
Breath Volume	1.781	00:44
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 04/11/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRON,DAVID,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALFRED
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:42

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/10/2023 12:29:43

Diagnostics Check	OK	12:23	
Air Blank	0.000	12:24	
Control Test	0.078	12:24	
Air Blank	0.000	12:25	
Subject Sample	# 1 0.159	12:25	
Breath Volume	3.363	12:25	
Air Blank	0.000	12:26	
Air Blank	0.000	12:28	
Subject Sample	# 2 0.156	12:29	
Breath Volume	2.855	12:29	
Air Blank		12:29	RFI Detect
Air Blank	0.000	12:29	

Data Download Date/Time: 04/11/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 03/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 21:21

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 04/10/2023 22:25:30

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.078	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1 0.263	22:15	Volume Not Met
Breath Volume	0.855	22:15	
Air Blank	0.000	22:16	
Air Blank	0.000	22:18	
Subject Sample	# 2 0.290	22:20	
Breath Volume	2.042	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:23	
Subject Sample	# 3 0.282	22:23	
Breath Volume	1.261	22:23	
Air Blank	0.000	22:24	
Control Test	0.075	22:24	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 04/11/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BELCHER,ALEX,B

BREATH TEST OPERATOR AGENCY: Auburndale PD

ARRESTING OFFICER: GRAYES,BOBBY

ARRESTING OFFICER AGENCY: Auburndale PD

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SHINAFELT

DATE/TIME OF TEST: 04/13/2023 03:57:26

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.124	03:52
Breath Volume	1.152	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.124	03:55
Breath Volume	1.625	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 05/18/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: WALLISER,SHAUN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 19:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: JASSO BERNAL

DATE/TIME OF TEST: 04/13/2023 20:57:03

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.078	20:51
Air Blank	0.000	20:51
Subject Sample # 1	0.028	20:52
Breath Volume	2.289	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:54
Subject Sample # 2	0.027	20:55
Breath Volume	2.070	20:55
Air Blank	0.000	20:56
Control Test	0.078	20:56
Air Blank	0.000	20:56
Diagnostics Check	OK	20:57

Data Download Date/Time: 05/18/2023 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DAWSON JR,MICHAEL,E

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1592208A3

DRY GAS EXPIRY DATE: 08/25/2024

VIOLATION CODE: DUI

ARREST TIME: 20:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/14/2023 20:59:55

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.079	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.178	20:55
Breath Volume	1.203	20:55
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.198	20:58
Breath Volume	1.300	20:58
Air Blank	0.000	20:59
Control Test	0.077	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 05/18/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERRELL,AUSTIN,G

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 04:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: PEREZ-RAMIREZ

DATE/TIME OF TEST: 04/15/2023 07:07:25

Diagnostics Check	OK	07:00
Air Blank	0.000	07:01
Control Test	0.079	07:01
Air Blank	0.000	07:02
Subject Sample # 1	0.103	07:02
Breath Volume	2.875	07:02
Air Blank	0.000	07:03
Air Blank	0.000	07:05
Subject Sample # 2	0.104	07:05
Breath Volume	3.003	07:05
Air Blank	0.000	07:06
Control Test	0.076	07:06
Air Blank	0.000	07:07
Diagnostics Check	OK	07:07

Data Download Date/Time: 05/18/2023 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARBANAS,SHANE,M

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: ARBANAS,SHANE

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 04/16/2023 03:09:16

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.079	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 05/18/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:19

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: DE LA PAZ OLIVA

DATE/TIME OF TEST: 04/16/2023 05:00:01

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.138	04:55
Breath Volume	1.199	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.141	04:58
Breath Volume	1.273	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	05:00

Data Download Date/Time: 05/18/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: JOHNSTON,CHRISTOPHER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 16:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: HIRSH

DATE/TIME OF TEST: 04/17/2023 17:51:26

Diagnostics Check	OK	17:45
Air Blank	0.000	17:45
Control Test	0.079	17:45
Air Blank	0.000	17:46
Subject Sample # 1	0.169	17:46
Breath Volume	2.007	17:46
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.163	17:49
Breath Volume	1.933	17:49
Air Blank	0.000	17:50
Control Test	0.079	17:50
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time: 05/18/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: GONCALVES

DATE/TIME OF TEST: 04/17/2023 21:55:37

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:44	
Control Test	0.080	21:44	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:48	No Sample Provided
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 2	21:54	No Sample Provided
Breath Volume	0.000	21:54	
Air Blank	0.000	21:54	
Control Test	0.080	21:55	
Air Blank	0.000	21:55	
Diagnostics Check	OK	21:55	

Data Download Date/Time: 05/18/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: COLON,JOLNETH

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 18:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: TEEHAN

DATE/TIME OF TEST: 04/18/2023 20:32:24

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.077	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.265	20:23
Breath Volume	3.695	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.216	20:27
Breath Volume	1.667	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 3	0.260	20:30
Breath Volume	3.929	20:30
Air Blank	0.000	20:31
Control Test	0.077	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 05/18/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 21:57

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: SWARTZ

DATE/TIME OF TEST: 04/19/2023 22:38:25

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.078	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.215	22:33
Breath Volume	1.269	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.213	22:36
Breath Volume	1.496	22:36
Air Blank	0.000	22:37
Control Test	0.078	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 05/18/2023 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 04/20/2023 01:52:24

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.078	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.080	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 05/18/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TRONSEN

DATE/TIME OF TEST: 04/20/2023 23:52:28

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.078	23:50	
Air Blank	0.000	23:50	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Control Test	0.079	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 05/18/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 159220080A

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: CLAVELL

DATE/TIME OF TEST: 04/21/2023 00:02:30

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.078	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.125	23:58
Breath Volume	1.414	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.127	00:00
Breath Volume	1.578	00:00
Air Blank	0.000	00:01
Control Test	0.078	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 05/18/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DAWSON JR,MICHAEL,E

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: VICENTE

DATE/TIME OF TEST: 04/21/2023 02:55:46

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.078	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.126	02:50
Breath Volume	1.488	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.124	02:54
Breath Volume	2.199	02:54
Air Blank	0.000	02:54
Control Test	0.076	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 05/18/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 04:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/21/2023 04:31:00

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.133	04:26
Breath Volume	1.269	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.133	04:29
Breath Volume	1.531	04:29
Air Blank	0.000	04:30
Control Test	0.077	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:31

Data Download Date/Time: 05/18/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: PALMORN

DATE/TIME OF TEST: 04/21/2023 17:00:33

Diagnostics Check	OK	16:54
Air Blank	0.000	16:54
Control Test	0.078	16:55
Air Blank	0.000	16:55
Subject Sample # 1	0.041	16:56
Breath Volume	1.613	16:56
Air Blank	0.000	16:56
Air Blank	0.000	16:58
Subject Sample # 2	0.042	16:59
Breath Volume	1.875	16:59
Air Blank	0.000	16:59
Control Test	0.079	16:59
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

Data Download Date/Time: 05/18/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: COLON,JOLNETH

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ZIGLER

DATE/TIME OF TEST: 04/22/2023 03:01:29

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.113	02:56
Breath Volume	2.093	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.108	02:59
Breath Volume	1.621	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 05/18/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: TINER,NICHOLAS

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 20:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: GAGNON

DATE/TIME OF TEST: 04/22/2023 22:31:26

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.078	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:24	No Sample Provided
Breath Volume	0.000	22:24	
Air Blank	0.000	22:24	
Air Blank	0.000	22:26	
Subject Sample	# 2	22:30	No Sample Provided
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Control Test	0.079	22:30	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 05/18/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: AGUILAR,KIMBERLY

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 21:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: CORTEZ RODRIGUEZ

DATE/TIME OF TEST: 04/22/2023 23:20:29

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.191	23:15
Breath Volume	2.351	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.188	23:18
Breath Volume	2.000	23:18
Air Blank	0.000	23:19
Control Test	0.077	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: BUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: MAYORGA

DATE/TIME OF TEST: 04/22/2023 23:48:19

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:41	No Sample Provided
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 2	23:46	No Sample Provided
Breath Volume	0.000	23:46	
Air Blank	0.000	23:47	
Control Test	0.078	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: DEVILLE

DATE/TIME OF TEST: 04/23/2023 00:14:34

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.079	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.163	00:10
Breath Volume	2.167	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.174	00:13
Breath Volume	2.433	00:13
Air Blank	0.000	00:13
Control Test	0.078	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HARE,LAKAZE
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LAFFIN

DATE/TIME OF TEST: 04/23/2023 02:36:22

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.164	02:31
Breath Volume	2.597	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.168	02:34
Breath Volume	2.074	02:34
Air Blank	0.000	02:35
Control Test	0.078	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: COLON,JOLNETH

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 05:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: CASTRO-MARTINEZ

DATE/TIME OF TEST: 04/23/2023 06:35:59

Diagnostics Check	OK	06:29
Air Blank	0.000	06:29
Control Test	0.079	06:30
Air Blank	0.000	06:30
Subject Sample # 1	0.000	06:31
Breath Volume	1.238	06:31
Air Blank	0.000	06:31
Air Blank	0.000	06:33
Subject Sample # 2	0.000	06:34
Breath Volume	1.820	06:34
Air Blank	0.000	06:35
Control Test	0.079	06:35
Air Blank	0.000	06:35
Diagnostics Check	OK	06:35

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JUDAH,PATTY,A

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: WALLISER,SHAUN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 10:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 11:37

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/23/2023 12:10:54

Diagnostics Check	OK	12:04
Air Blank	0.000	12:04
Control Test	0.079	12:05
Air Blank	0.000	12:05
Subject Sample # 1	0.201	12:06
Breath Volume	2.968	12:06
Air Blank	0.000	12:06
Air Blank	0.000	12:08
Subject Sample # 2	0.203	12:09
Breath Volume	3.117	12:09
Air Blank	0.000	12:09
Control Test	0.079	12:10
Air Blank	0.000	12:10
Diagnostics Check	OK	12:10

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 04/25/2023 17:19:15

Diagnostics Check	OK	17:09
Air Blank	0.000	17:10
Control Test	0.078	17:10
Air Blank	0.000	17:10
Subject Sample # 1	0.228	17:11
Breath Volume	1.468	17:11
Air Blank	0.000	17:12
Air Blank	0.000	17:13
Subject Sample # 2	0.195	17:14
Breath Volume	1.351	17:14
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 3	0.223	17:17
Breath Volume	1.703	17:17
Air Blank	0.000	17:18
Control Test	0.077	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: METELLUS,SAM,S

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2004

VIOLATION CODE: DUI

ARREST TIME: 20:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MARTINEZ-SALGADO

DATE/TIME OF TEST: 04/25/2023 21:52:05

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:37	
Control Test	0.079	21:37	
Air Blank	0.000	21:38	
Subject Sample	# 1 0.140	21:41	Volume Not Met
Breath Volume	0.972	21:41	
Air Blank	0.000	21:42	
Air Blank	0.000	21:44	
Subject Sample	# 2 0.170	21:46	
Breath Volume	2.003	21:46	
Air Blank	0.000	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 3 0.153	21:50	
Breath Volume	2.011	21:50	
Air Blank	0.000	21:51	
Control Test	0.075	21:51	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:52	

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BELEN,MATTHEW,P

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: BELEN,MATTHEW

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 23:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: TIRADO DELGADO

DATE/TIME OF TEST: 04/26/2023 00:12:53

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.136	00:04
Breath Volume	1.117	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.161	00:08
Breath Volume	1.695	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample # 3	0.148	00:11
Breath Volume	1.546	00:11
Air Blank	0.000	00:11
Control Test	0.075	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NELSON,VICTORIA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 04/27/2023 05:46:29

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.079	05:41
Air Blank	0.000	05:41
Subject Sample # 1	0.120	05:42
Breath Volume	1.644	05:42
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.126	05:44
Breath Volume	1.746	05:44
Air Blank	0.000	05:45
Control Test	0.079	05:45
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: GIRGIS,JOHN
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CAMACHO TREJO

DATE/TIME OF TEST: 04/27/2023 20:30:04

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.235	20:24
Breath Volume	1.132	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.229	20:28
Breath Volume	1.117	20:28
Air Blank	0.000	20:29
Control Test	0.077	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:30

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILLS, MARK, A
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: SILLS, MARK
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: STOW

DATE/TIME OF TEST: 04/27/2023 23:05:42

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.078	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:58	No Sample Provided
Breath Volume	0.000	22:58	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample	# 2	23:04	No Sample Provided
Breath Volume	0.000	23:04	
Air Blank	0.000	23:04	
Control Test	0.078	23:05	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JEUDY,RINALDO,

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: SMITH,PARKER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LIMA

DATE/TIME OF TEST: 04/28/2023 03:21:18

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.079	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.238	03:16
Breath Volume	1.746	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.240	03:19
Breath Volume	2.316	03:19
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PROPST,JOHN,B

BREATH TEST OPERATOR AGENCY: Auburndale PD

ARRESTING OFFICER: PROPST,JOHN

ARRESTING OFFICER AGENCY: Auburndale PD

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 19:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: VANCOL

DATE/TIME OF TEST: 04/28/2023 20:41:02

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.079	20:33
Air Blank	0.000	20:34
Subject Sample # 1	0.000	20:36
Breath Volume	1.253	20:36
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.000	20:39
Breath Volume	1.253	20:39
Air Blank	0.000	20:40
Control Test	0.078	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:41

Data Download Date/Time: 05/18/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GARCIA-SANTIAGO

DATE/TIME OF TEST: 04/29/2023 03:39:43

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.076	03:35
Breath Volume	1.312	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.074	03:38
Breath Volume	1.121	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 05/18/2023 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,ANN,R
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: LEDFORD,ANN
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/29/2023 04:00:54

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.079	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 2	03:59	No Sample Provided
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.079	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 05/18/2023 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 04/29/2023 22:02:04

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.079	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.235	21:56
Breath Volume	1.582	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:58
Subject Sample # 2	0.225	22:00
Breath Volume	1.390	22:00
Air Blank	0.000	22:01
Control Test	0.076	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:02

Data Download Date/Time: 05/18/2023 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/30/2023 03:13:05

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.074	03:12	Control Outside Tolerance
Air Blank	0.000	03:13	

Data Download Date/Time: 05/18/2023 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 04/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A3

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/30/2023 03:22:24

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.165	03:16
Breath Volume	1.433	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.153	03:20
Breath Volume	1.140	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 05/18/2023 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/22/2023 04:02:48

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.078	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.000	03:58
Breath Volume	2.011	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.000	04:01
Breath Volume	1.296	04:01
Air Blank	0.000	04:01
Control Test	0.077	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 05/03/2023 00:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BLAND

DATE/TIME OF TEST: 04/23/2023 04:37:39

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.078	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.193	04:33
Breath Volume	1.621	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.198	04:36
Breath Volume	3.238	04:36
Air Blank	0.000	04:36
Control Test	0.077	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 05/03/2023 00:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HUSBANDS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 04/27/2023 22:30:52

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.078	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:29	
Control Test	0.078	22:30	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 05/03/2023 00:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CHANG

DATE/TIME OF TEST: 04/28/2023 04:15:31

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.078	04:13	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Control Test	0.078	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 05/03/2023 00:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDY,JOE,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARDY,JOE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 04/30/2023 15:03:28

Diagnostics Check	OK	14:54	
Air Blank	0.000	14:54	
Control Test	0.081	14:54	
Air Blank	0.000	14:55	
Subject Sample	# 1 0.203	14:58	Volume Not Met
Breath Volume	0.070	14:58	
Air Blank	0.000	14:59	
Air Blank	0.000	15:01	
Subject Sample	# 2	15:02	Subject Test Refused
Breath Volume	0.019	15:02	
Air Blank	0.000	15:02	
Control Test	0.079	15:02	
Air Blank	0.000	15:03	
Diagnostics Check	OK	15:03	

Data Download Date/Time: 05/01/2023 16:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES, AMAURY, A
BREATH TEST OPERATOR AGENCY: FHP TROOP K
ARRESTING OFFICER: QUINONES, AMAURY
ARRESTING OFFICER AGENCY: FHP TROOP K
LAST AGENCY INSPECTION: 04/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: QUILES

DATE/TIME OF TEST: 04/29/2023 11:08:13

Diagnostics Check	OK	10:55	
Air Blank	0.000	10:56	
Control Test	0.083	10:56	
Air Blank	0.000	10:57	
Subject Sample	# 1 0.136	11:00	Volume Not Met
Breath Volume	0.472	11:00	
Air Blank	0.000	11:00	
Air Blank	0.000	11:02	
Subject Sample	# 2 0.137	11:03	
Breath Volume	1.777	11:03	
Air Blank	0.000	11:04	
Air Blank	0.000	11:05	
Subject Sample	# 3 0.142	11:06	
Breath Volume	1.585	11:06	
Air Blank	0.000	11:07	
Control Test	0.081	11:07	
Air Blank	0.000	11:08	
Diagnostics Check	OK	11:08	

Data Download Date/Time: 07/01/2023 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,RICHARD,S
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: TISDALE,RICHARD
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MAUK

DATE/TIME OF TEST: 04/02/2023 04:38:14

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.118	04:33
Breath Volume	3.476	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.116	04:36
Breath Volume	2.949	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 04/28/2023 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEIMAN,ERIC,M
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: BYRD,ERVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/10/2023 01:39:55

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.108	01:35
Breath Volume	1.832	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 2	0.106	01:38
Breath Volume	1.714	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 04/28/2023 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,ERVIN,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BYRD,ERVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/12/2023 23:36:41

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.204	23:32
Breath Volume	2.039	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.201	23:35
Breath Volume	1.242	23:35
Air Blank	0.000	23:35
Control Test	0.078	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 04/28/2023 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,RICHARD,S
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: TISDALE,RICHARD
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: CATARINO

DATE/TIME OF TEST: 04/16/2023 19:47:58

Diagnostics Check	OK	19:39	
Air Blank	0.000	19:40	
Control Test	0.080	19:40	
Air Blank	0.000	19:40	
Subject Sample	# 1	19:43	No Sample Provided
Breath Volume	0.000	19:43	
Air Blank	0.000	19:44	
Air Blank	0.000	19:46	
Subject Sample	# 2	19:46	Subject Test Refused
Breath Volume	0.000	19:46	
Air Blank	0.000	19:47	
Control Test	0.080	19:47	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

Data Download Date/Time: 04/28/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,RICHARD,S
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: TISDALE,RICHARD
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 03/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: CATARINO

DATE/TIME OF TEST: 04/16/2023 19:54:01

Diagnostics Check	OK	19:49	
Air Blank	0.000	19:49	
Control Test	0.080	19:49	
Air Blank	0.000	19:50	
Subject Sample	# 1	19:52	Subject Test Refused
Breath Volume	0.000	19:52	
Air Blank	0.000	19:53	
Control Test	0.080	19:53	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:54	

Data Download Date/Time: 04/28/2023 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MORGAN,MICHAEL
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 04/16/2023 21:40:24

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.080	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.193	21:35
Breath Volume	1.230	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.205	21:38
Breath Volume	1.171	21:38
Air Blank	0.000	21:39
Control Test	0.079	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 04/21/2023 16:09