

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SANTANA GUADALUPE

DATE/TIME OF TEST: 07/01/2023 00:20:38

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.007	00:15
Breath Volume	3.878	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.007	00:19
Breath Volume	3.925	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 07/20/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: QUIROZ JUAREZ

DATE/TIME OF TEST: 07/01/2023 02:54:21

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.175	02:49
Breath Volume	1.253	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.174	02:52
Breath Volume	2.574	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 07/20/2023 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 07/01/2023 04:08:33

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.068	04:02
Breath Volume	1.781	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.065	04:06
Breath Volume	1.242	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 07/20/2023 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/01/2023 22:43:36

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.080	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1 0.121	22:32	
Breath Volume	1.109	22:32	
Air Blank	0.000	22:33	
Air Blank	0.000	22:35	
Subject Sample	# 2 0.114	22:38	Volume Not Met
Breath Volume	0.082	22:38	
Air Blank	0.000	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 3 0.121	22:42	
Breath Volume	1.273	22:42	
Air Blank	0.000	22:42	
Control Test	0.078	22:42	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 07/20/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 07/02/2023 20:37:16

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:34	
Control Test	0.081	20:35	
Air Blank	0.000	20:35	
Subject Sample	# 1	20:35	Subject Test Refused
Breath Volume	0.000	20:35	
Air Blank	0.000	20:36	
Control Test	0.080	20:36	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

Data Download Date/Time: 07/20/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE, GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE, GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 07/02/2023 23:11:16

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.200	23:05
Breath Volume	2.152	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample # 2	0.188	23:09
Breath Volume	2.324	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 07/20/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 07:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:58

SUBJECT'S LAST NAME: TEREZIU

DATE/TIME OF TEST: 07/04/2023 08:25:52

Diagnostics Check	OK	08:19
Air Blank	0.000	08:19
Control Test	0.080	08:20
Air Blank	0.000	08:20
Subject Sample # 1	0.150	08:21
Breath Volume	2.035	08:21
Air Blank	0.000	08:22
Air Blank	0.000	08:23
Subject Sample # 2	0.150	08:24
Breath Volume	2.078	08:24
Air Blank	0.000	08:24
Control Test	0.080	08:25
Air Blank	0.000	08:25
Diagnostics Check	OK	08:25

Data Download Date/Time: 07/20/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,ELIER,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SANTANA,ELIER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:39

SUBJECT'S LAST NAME: KUMMERFELDT

DATE/TIME OF TEST: 07/05/2023 09:06:33

Diagnostics Check	OK	08:59
Air Blank	0.000	09:00
Control Test	0.079	09:00
Air Blank	0.000	09:01
Subject Sample # 1	0.200	09:01
Breath Volume	3.027	09:01
Air Blank	0.000	09:02
Air Blank	0.000	09:04
Subject Sample # 2	0.204	09:04
Breath Volume	2.570	09:04
Air Blank	0.000	09:05
Control Test	0.078	09:05
Air Blank	0.000	09:06
Diagnostics Check	OK	09:06

Data Download Date/Time: 07/20/2023 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BULL

DATE/TIME OF TEST: 07/07/2023 00:26:55

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.116	00:22
Breath Volume	1.789	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.119	00:25
Breath Volume	2.074	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 07/20/2023 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE, GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE, GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 07/12/2023 02:27:27

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.173	02:22
Breath Volume	1.593	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.170	02:25
Breath Volume	1.500	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 07/20/2023 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 07/13/2023 07:33:40

Diagnostics Check	OK	07:21	
Air Blank	0.000	07:22	
Control Test	0.079	07:22	
Air Blank	0.000	07:23	
Subject Sample	# 1 0.114	07:26	Volume Not Met
Breath Volume	0.082	07:26	
Air Blank	0.000	07:26	
Air Blank	0.000	07:28	
Subject Sample	# 2 0.109	07:31	Volume Not Met
Breath Volume	0.074	07:31	
Air Blank	0.000	07:32	
Control Test	0.079	07:33	
Air Blank	0.000	07:33	
Diagnostics Check	OK	07:33	

Data Download Date/Time: 07/20/2023 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ALTUCHOFF,JOSEPH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 07/16/2023 04:05:02

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.127	04:00
Breath Volume	2.195	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.130	04:03
Breath Volume	3.074	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:05

Data Download Date/Time: 07/20/2023 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KOSAKOWSKI,PETER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 07/16/2023 05:27:13

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.080	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.180	05:22
Breath Volume	2.203	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.179	05:25
Breath Volume	1.140	05:25
Air Blank	0.000	05:26
Control Test	0.079	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 07/20/2023 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE, GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN, JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SANCHEZ PEREZ

DATE/TIME OF TEST: 07/17/2023 02:26:36

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.234	02:21
Breath Volume	1.308	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.251	02:24
Breath Volume	1.312	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 07/20/2023 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: NANTAIS

DATE/TIME OF TEST: 07/17/2023 02:35:11

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.080	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Control Test	0.081	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 07/20/2023 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: CASTRO SERRANO

DATE/TIME OF TEST: 07/19/2023 03:26:15

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.144	03:20
Breath Volume	1.769	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.135	03:24
Breath Volume	1.417	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 07/20/2023 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 11:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: BADOUR

DATE/TIME OF TEST: 07/01/2023 13:22:23

Diagnostics Check	OK	13:15
Air Blank	0.000	13:16
Control Test	0.080	13:16
Air Blank	0.000	13:17
Subject Sample # 1	0.246	13:17
Breath Volume	1.597	13:17
Air Blank	0.000	13:18
Air Blank	0.000	13:20
Subject Sample # 2	0.248	13:20
Breath Volume	1.476	13:20
Air Blank	0.000	13:21
Control Test	0.080	13:21
Air Blank	0.000	13:22
Diagnostics Check	OK	13:22

Data Download Date/Time: 07/20/2023 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: REYNA DIAZ

DATE/TIME OF TEST: 07/08/2023 03:23:41

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.082	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.128	03:18
Breath Volume	1.183	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.114	03:21
Breath Volume	1.406	03:21
Air Blank	0.000	03:22
Control Test	0.078	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 07/20/2023 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: RIGGSBEE

DATE/TIME OF TEST: 07/08/2023 03:40:48

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.081	03:35
Breath Volume	1.167	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:37
Subject Sample # 2	0.075	03:39
Breath Volume	1.167	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 07/20/2023 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/11/2023 17:58:00

Diagnostics Check	OK	17:51
Air Blank	0.000	17:52
Control Test	0.081	17:52
Air Blank	0.000	17:52
Subject Sample # 1	0.000	17:53
Breath Volume	2.718	17:53
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample # 2	0.000	17:56
Breath Volume	2.347	17:56
Air Blank	0.000	17:57
Control Test	0.081	17:57
Air Blank	0.000	17:57
Diagnostics Check	OK	17:58

Data Download Date/Time: 07/20/2023 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BLAISE

DATE/TIME OF TEST: 07/14/2023 03:01:33

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.233	02:56
Breath Volume	1.289	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.240	02:59
Breath Volume	1.589	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 07/20/2023 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: ROBLERO ESCOBAR

DATE/TIME OF TEST: 07/16/2023 18:45:06

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.081	18:38
Air Blank	0.000	18:39
Subject Sample # 1	0.205	18:40
Breath Volume	1.597	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:42
Subject Sample # 2	0.202	18:43
Breath Volume	1.667	18:43
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

Data Download Date/Time: 07/20/2023 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CORSO

DATE/TIME OF TEST: 07/18/2023 03:10:55

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.164	03:04
Breath Volume	1.250	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample # 2	0.162	03:09
Breath Volume	1.480	03:09
Air Blank	0.000	03:09
Control Test	0.077	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 07/20/2023 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WORDEN

DATE/TIME OF TEST: 07/20/2023 02:05:37

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.081	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1 0.193	02:03	
Breath Volume	2.796	02:03	
Air Blank	0.000	02:03	
Air Blank		02:05	RFI Detect
Air Blank		02:05	RFI Detect
Air Blank		02:05	RFI Detect

Data Download Date/Time: 07/20/2023 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE, GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE, GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WORDEN

DATE/TIME OF TEST: 07/20/2023 02:13:22

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.181	02:08
Breath Volume	2.066	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.189	02:11
Breath Volume	2.285	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 07/20/2023 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: TORIBIO

DATE/TIME OF TEST: 07/20/2023 02:30:26

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.187	02:25
Breath Volume	1.511	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.183	02:28
Breath Volume	2.460	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 07/20/2023 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COFFMAN,D
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 07/01/2023 03:41:07

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1 0.198	03:35	
Breath Volume	1.652	03:35	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2 0.192	03:39	
Breath Volume	1.730	03:39	
Air Blank	0.000	03:39	
Control Test	0.077	03:40	
Air Blank		03:40	RFI Detect
Air Blank	0.000	03:41	

Data Download Date/Time: 07/14/2023 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHMIDT,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: SIMAS

DATE/TIME OF TEST: 07/04/2023 23:01:48

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:58	
Control Test	0.080	22:58	
Air Blank	0.000	22:59	
Subject Sample	# 1	23:00	Subject Test Refused
Breath Volume	0.000	23:00	
Air Blank	0.000	23:00	
Control Test	0.081	23:01	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 07/14/2023 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:37

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 07/11/2023 14:04:59

Diagnostics Check	OK	13:57
Air Blank	0.000	13:58
Control Test	0.081	13:58
Air Blank	0.000	13:59
Subject Sample # 1	0.000	13:59
Breath Volume	2.792	13:59
Air Blank	0.000	14:00
Air Blank	0.000	14:02
Subject Sample # 2	0.000	14:03
Breath Volume	2.296	14:03
Air Blank	0.000	14:04
Control Test	0.081	14:04
Air Blank	0.000	14:04
Diagnostics Check	OK	14:04

Data Download Date/Time: 07/14/2023 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAUN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ZALAK,JONATHAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: CRAIN

DATE/TIME OF TEST: 07/15/2023 02:30:45

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:28	
Control Test	0.081	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.081	02:30	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 08/29/2023 08:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAUN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 07/16/2023 00:11:49

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.154	00:06
Breath Volume	1.625	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.155	00:09
Breath Volume	2.136	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 08/29/2023 08:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICHMAN,MICHELLE,K
BREATH TEST OPERATOR AGENCY: Bunnell PD
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/16/2023 04:06:37

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.000	04:02
Breath Volume	2.894	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.000	04:05
Breath Volume	3.316	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 08/29/2023 08:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICHMAN,MICHELLE,K
BREATH TEST OPERATOR AGENCY: Bunnell PD
ARRESTING OFFICER: JENKINS,LAURA
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: DOWELL

DATE/TIME OF TEST: 07/21/2023 00:32:30

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.081	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.449	00:30	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LICHTY,DAVID
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: LUGGER

DATE/TIME OF TEST: 07/21/2023 22:34:44

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:32	
Control Test	0.081	22:32	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:33	Subject Test Refused
Breath Volume	0.000	22:33	
Air Blank	0.000	22:33	
Control Test	0.080	22:34	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BARLOTTA,V
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: PROSSER

DATE/TIME OF TEST: 07/23/2023 00:37:16

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.081	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.157	00:32
Breath Volume	2.125	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.156	00:35
Breath Volume	1.882	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SANDERS,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 07/23/2023 05:35:02

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.155	05:30
Breath Volume	1.515	05:30
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.145	05:33
Breath Volume	1.117	05:33
Air Blank	0.000	05:34
Control Test	0.080	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:35

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: ROOM

DATE/TIME OF TEST: 07/26/2023 19:48:18

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.080	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.245	19:43
Breath Volume	2.847	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.244	19:46
Breath Volume	1.847	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHRAGER, OLIVIA
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: RIGGLE

DATE/TIME OF TEST: 07/26/2023 20:52:49

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.081	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.000	20:47
Breath Volume	1.796	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.000	20:51
Breath Volume	1.929	20:51
Air Blank	0.000	20:51
Control Test	0.081	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BARLOTTA,V
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ELMER

DATE/TIME OF TEST: 07/27/2023 01:54:51

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.082	01:52	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.080	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,KELVIN,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VEGA,KELVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/29/2023 05:38:54

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.081	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.158	05:34
Breath Volume	2.218	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:36
Subject Sample # 2	0.152	05:37
Breath Volume	2.160	05:37
Air Blank	0.000	05:37
Control Test	0.080	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICHMAN,MICHELLE,K
BREATH TEST OPERATOR AGENCY: Bunnell PD
ARRESTING OFFICER: JENKINS,LAURA
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BEACH

DATE/TIME OF TEST: 07/30/2023 00:14:32

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1 0.165	00:08	
Breath Volume	1.304	00:08	
Air Blank	0.000	00:09	
Air Blank	0.000	00:10	
Subject Sample	# 2	00:13	Subject Test Refused
Breath Volume	0.097	00:13	
Air Blank		00:14	Purge Fail
Air Blank	0.000	00:14	

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICHMAN,MICHELLE,K
BREATH TEST OPERATOR AGENCY: Bunnell PD
ARRESTING OFFICER: JENKINS,LAURA
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BEACH

DATE/TIME OF TEST: 07/30/2023 00:20:14

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.075	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.035	00:18	
Air Blank	0.000	00:19	
Control Test	0.078	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RAIMUNDO

DATE/TIME OF TEST: 07/30/2023 04:07:23

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.011	04:02
Breath Volume	2.207	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.012	04:05
Breath Volume	2.128	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICHMAN,MICHELLE,K
BREATH TEST OPERATOR AGENCY: Bunnell PD
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 07/30/2023 04:16:39

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:12	
Control Test	0.081	04:12	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.132	04:14	
Air Blank	0.000	04:15	
Control Test	0.081	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICHMAN,MICHELLE,K
BREATH TEST OPERATOR AGENCY: Bunnell PD
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 07/30/2023 04:22:00

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.078	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:20	Subject Test Refused
Breath Volume	0.289	04:20	
Air Blank	0.000	04:21	
Control Test	0.080	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: DEZI

DATE/TIME OF TEST: 07/30/2023 19:45:59

Diagnostics Check	OK	19:39
Air Blank	0.000	19:39
Control Test	0.080	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.174	19:41
Breath Volume	1.949	19:41
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.174	19:44
Breath Volume	1.839	19:44
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 07/31/2023 03:08:23

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.081	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1 0.116	03:00	Volume Not Met
Breath Volume	0.589	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 2 0.121	03:03	
Breath Volume	1.164	03:03	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 3 0.125	03:06	
Breath Volume	1.460	03:06	
Air Blank	0.000	03:07	
Control Test	0.079	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTLEY, BRANDEN, C
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: JACKSON, BRIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: EKTOV

DATE/TIME OF TEST: 07/31/2023 16:43:09

Diagnostics Check	OK	16:36
Air Blank	0.000	16:37
Control Test	0.081	16:37
Air Blank	0.000	16:37
Subject Sample # 1	0.155	16:38
Breath Volume	1.355	16:38
Air Blank	0.000	16:39
Air Blank	0.000	16:40
Subject Sample # 2	0.154	16:41
Breath Volume	1.429	16:41
Air Blank	0.000	16:42
Control Test	0.079	16:42
Air Blank	0.000	16:43
Diagnostics Check	OK	16:43

Data Download Date/Time: 08/29/2023 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: IBANEZ,CHLOE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: LACERTOSA

DATE/TIME OF TEST: 07/02/2023 18:30:20

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:27	
Control Test	0.079	18:28	
Air Blank	0.000	18:28	
Subject Sample	# 1	18:28	Subject Test Refused
Breath Volume	0.000	18:28	
Air Blank	0.000	18:29	
Control Test	0.079	18:29	
Air Blank	0.000	18:30	
Diagnostics Check	OK	18:30	

Data Download Date/Time: 07/11/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: IBANEZ,CHLOE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: SOWINSKI

DATE/TIME OF TEST: 07/03/2023 04:21:40

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.197	04:16
Breath Volume	2.164	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.183	04:19
Breath Volume	1.625	04:19
Air Blank	0.000	04:20
Control Test	0.077	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 07/11/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BAUMGARTEN,CHAYA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/10/2023 02:13:00

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.207	02:07
Breath Volume	1.816	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.212	02:11
Breath Volume	2.402	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:13

Data Download Date/Time: 07/11/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIMONEY,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: GARI

DATE/TIME OF TEST: 07/16/2023 04:34:42

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.175	04:29
Breath Volume	2.125	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.170	04:32
Breath Volume	1.687	04:32
Air Blank	0.000	04:33
Control Test	0.078	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: IBANEZ,CHLOE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/19/2023 02:40:01

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.079	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:39	
Control Test	0.080	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASAS,JAVIER,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: NETTLETON

DATE/TIME OF TEST: 07/19/2023 03:31:07

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/19/2023 21:46:41

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:42	
Control Test	0.080	21:42	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:45	Subject Test Refused
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.080	21:46	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIMONEY,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: IBANEZ,CHLOE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 07/20/2023 01:58:16

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.196	01:53
Breath Volume	2.726	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 2	0.196	01:56
Breath Volume	2.812	01:56
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: IBANEZ,CHLOE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/21/2023 02:53:12

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.000	02:48
Breath Volume	3.488	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.000	02:51
Breath Volume	3.050	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNE,ASHTON,D
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 07/22/2023 19:57:16

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.000	19:53
Breath Volume	3.113	19:53
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.000	19:55
Breath Volume	1.242	19:55
Air Blank	0.000	19:56
Control Test	0.075	19:56
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: MCMAHON

DATE/TIME OF TEST: 07/28/2023 05:13:13

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:08	
Control Test	0.079	05:08	
Air Blank	0.000	05:09	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.144	05:11	
Air Blank	0.000	05:12	
Control Test	0.080	05:12	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIMONEY,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ROSARIO,JONATHAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/29/2023 21:13:59

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:11	
Control Test	0.079	21:11	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:13	
Control Test	0.080	21:13	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HORNE,ASHTON
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: ALMONTE

DATE/TIME OF TEST: 07/30/2023 23:44:18

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.127	23:39
Breath Volume	1.800	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.125	23:42
Breath Volume	1.710	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BUFFE

DATE/TIME OF TEST: 07/31/2023 03:00:45

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.079	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,EDWARD
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: BALENA

DATE/TIME OF TEST: 07/31/2023 19:12:40

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:10	
Control Test	0.079	19:10	
Air Blank	0.000	19:11	
Subject Sample	# 1	19:11	Subject Test Refused
Breath Volume	0.000	19:11	
Air Blank	0.000	19:11	
Control Test	0.080	19:12	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

Data Download Date/Time: 08/22/2023 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: TEMA ARDEANO

DATE/TIME OF TEST: 07/03/2023 08:08:09

Diagnostics Check	OK	07:58
Air Blank	0.000	07:58
Control Test	0.081	07:59
Air Blank	0.000	07:59
Subject Sample # 1	0.243	08:00
Breath Volume	2.179	08:00
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.211	08:03
Breath Volume	1.878	08:03
Air Blank	0.000	08:04
Air Blank	0.000	08:06
Subject Sample # 3	0.205	08:06
Breath Volume	1.429	08:06
Air Blank	0.000	08:07
Control Test	0.081	08:07
Air Blank	0.000	08:08
Diagnostics Check	OK	08:08

Data Download Date/Time: 08/01/2023 00:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODDEN,DANIEL,R
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BODDEN,DANIEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:31

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 07/04/2023 16:59:03

Diagnostics Check	OK	16:52
Air Blank	0.000	16:52
Control Test	0.082	16:53
Air Blank	0.000	16:53
Subject Sample # 1	0.182	16:54
Breath Volume	1.742	16:54
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 2	0.190	16:57
Breath Volume	2.386	16:57
Air Blank	0.000	16:58
Control Test	0.081	16:58
Air Blank	0.000	16:58
Diagnostics Check	OK	16:59

Data Download Date/Time: 08/01/2023 00:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEIRONG,SHU,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MORALES,RICHARD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/01/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BEIRONG

DATE/TIME OF TEST: 07/08/2023 00:27:22

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.082	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1 0.110	00:16	Slope Not Level
Breath Volume	1.949	00:16	
Air Blank	0.000	00:17	
Air Blank	0.000	00:19	
Subject Sample	# 2 0.130	00:19	
Breath Volume	1.511	00:19	
Air Blank	0.000	00:20	
Air Blank	0.000	00:22	
Subject Sample	# 3 0.103	00:25	Volume Not Met
Breath Volume	0.515	00:25	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 08/01/2023 00:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,RICHARD,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MORALES,RICHARD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BEIRONG

DATE/TIME OF TEST: 07/08/2023 00:15:56

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:14	
Control Test	0.080	00:14	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:15	Sequence Aborted
Breath Volume	0.000	00:15	
Air Blank	0.000	00:15	

Data Download Date/Time: 08/01/2023 00:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODDEN,DANIEL,R
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BODDEN,DANIEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/01/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: ABENDANO

DATE/TIME OF TEST: 07/16/2023 04:11:45

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.082	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.210	04:07
Breath Volume	1.433	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.224	04:10
Breath Volume	2.652	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 08/01/2023 00:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ROBINSON,HAROLD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 07/17/2023 22:18:53

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:16	
Control Test	0.081	22:16	
Air Blank	0.000	22:17	
Subject Sample	# 1	22:17	Subject Test Refused
Breath Volume	0.000	22:17	
Air Blank	0.000	22:17	
Control Test	0.081	22:18	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 08/01/2023 00:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEVEDO VALDES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ACEVEDO VALDES,CRISTIAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MENENDEZ

DATE/TIME OF TEST: 07/20/2023 04:46:51

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.082	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.152	04:42
Breath Volume	2.722	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.147	04:45
Breath Volume	2.363	04:45
Air Blank	0.000	04:45
Control Test	0.081	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 08/01/2023 00:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,SLOANE,
BREATH TEST OPERATOR AGENCY: FHP TROOP K
ARRESTING OFFICER: GOMEZ,SLOANE
ARRESTING OFFICER AGENCY: FHP TROOP K
LAST AGENCY INSPECTION: 07/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: SWEETING

DATE/TIME OF TEST: 07/29/2023 17:09:37

Diagnostics Check	OK	17:03
Air Blank	0.000	17:03
Control Test	0.081	17:04
Air Blank	0.000	17:04
Subject Sample # 1	0.274	17:05
Breath Volume	1.410	17:05
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.267	17:08
Breath Volume	1.816	17:08
Air Blank	0.000	17:08
Control Test	0.081	17:09
Air Blank	0.000	17:09
Diagnostics Check	OK	17:09

Data Download Date/Time: 08/01/2023 00:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/01/2023 20:13:07

Diagnostics Check	OK	20:06
Air Blank	0.000	20:06
Control Test	0.078	20:06
Air Blank	0.000	20:07
Subject Sample # 1	0.180	20:08
Breath Volume	2.941	20:08
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.189	20:11
Breath Volume	3.328	20:11
Air Blank	0.000	20:12
Control Test	0.079	20:12
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 07/20/2023 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 07/02/2023 21:55:22

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.078	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.095	21:50
Breath Volume	1.546	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.091	21:53
Breath Volume	1.402	21:53
Air Blank	0.000	21:54
Control Test	0.078	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 07/20/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BUI

DATE/TIME OF TEST: 07/06/2023 01:36:44

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.124	01:31
Breath Volume	1.273	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.136	01:35
Breath Volume	2.101	01:35
Air Blank	0.000	01:35
Control Test	0.077	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 07/20/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: FARLEY

DATE/TIME OF TEST: 07/06/2023 16:51:14

Diagnostics Check	OK	16:42
Air Blank	0.000	16:42
Control Test	0.078	16:43
Air Blank	0.000	16:43
Subject Sample # 1	0.167	16:44
Breath Volume	1.347	16:44
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.172	16:49
Breath Volume	1.261	16:49
Air Blank	0.000	16:50
Control Test	0.078	16:50
Air Blank	0.000	16:51
Diagnostics Check	OK	16:51

Data Download Date/Time: 07/20/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MANION

DATE/TIME OF TEST: 07/21/2023 02:35:14

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.079	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:27	No Sample Provided
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample	# 2	0.128	02:33
Breath Volume	0.226	02:33	Volume Not Met
Air Blank	0.000	02:34	
Control Test	0.080	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 08/14/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MANION

DATE/TIME OF TEST: 07/21/2023 03:02:26

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.165	02:57
Breath Volume	4.066	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.160	03:00
Breath Volume	4.839	03:00
Air Blank	0.000	03:01
Control Test	0.076	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 08/14/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DONALD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: ROUSSEAU,DMITRY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/22/2023 00:25:02

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:20	No Sample Provided
Breath Volume	0.000	00:20	
Air Blank	0.000	00:21	
Air Blank	0.000	00:23	
Subject Sample	# 2	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Control Test	0.079	00:24	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 08/14/2023 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 07/22/2023 01:26:39

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.077	01:24	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Control Test	0.076	01:26	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 08/14/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,SHELBY,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MENDES,KAVEN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: C515432875030

DATE/TIME OF TEST: 07/30/2023 02:16:21

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.181	02:11
Breath Volume	1.902	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.178	02:14
Breath Volume	1.675	02:14
Air Blank	0.000	02:15
Control Test	0.077	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 08/14/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 07/22/2023 23:33:19

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.223	23:28
Breath Volume	1.558	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:30
Subject Sample # 2	0.220	23:31
Breath Volume	1.511	23:31
Air Blank	0.000	23:32
Control Test	0.078	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 08/14/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SAMAYOA MARTINEZ

DATE/TIME OF TEST: 07/28/2023 01:34:23

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.080	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 08/14/2023 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: RANCE

DATE/TIME OF TEST: 07/27/2023 01:39:44

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.075	01:33	
Air Blank	0.000	01:34	
Subject Sample # 1	0.117	01:35	
Breath Volume	1.863	01:35	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample # 2	0.116	01:38	
Breath Volume	1.746	01:38	
Air Blank		01:39	Purge Fail
Air Blank	0.000	01:39	

Data Download Date/Time: 07/31/2023 01:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: RANCE

DATE/TIME OF TEST: 07/27/2023 01:49:12

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.110	01:43
Breath Volume	1.117	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.106	01:47
Breath Volume	1.207	01:47
Air Blank	0.000	01:48
Control Test	0.077	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 07/31/2023 01:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: VLTRES RIVERA

DATE/TIME OF TEST: 07/30/2023 02:42:19

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.078	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1 0.072	02:37	
Breath Volume	1.664	02:37	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2 0.063	02:40	
Breath Volume	1.156	02:40	
Air Blank	0.000	02:41	
Control Test	0.076	02:41	
Air Blank		02:41	RFI Detect
Air Blank	0.000	02:42	

Data Download Date/Time: 07/31/2023 01:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: MONTGOMERY,KENNETH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24229080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:09

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: BILKA

DATE/TIME OF TEST: 07/01/2023 02:52:16

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.082	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.074	02:47
Breath Volume	2.683	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.073	02:50
Breath Volume	2.359	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 07/18/2023 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: YUNGER

DATE/TIME OF TEST: 07/03/2023 22:29:42

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.080	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.096	22:22
Breath Volume	1.570	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.092	22:28
Breath Volume	2.144	22:28
Air Blank	0.000	22:28
Control Test	0.081	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 07/18/2023 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FIGUEIRDO

DATE/TIME OF TEST: 07/04/2023 02:52:13

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.165	02:46
Breath Volume	1.378	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.164	02:50
Breath Volume	1.527	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 07/18/2023 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/04/2023 23:11:10

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.102	23:06
Breath Volume	2.195	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.094	23:09
Breath Volume	1.863	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 07/18/2023 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SCADUTO

DATE/TIME OF TEST: 07/08/2023 23:46:18

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.016	23:41
Breath Volume	2.960	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.017	23:44
Breath Volume	2.714	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 07/18/2023 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: ZACCARINO

DATE/TIME OF TEST: 07/18/2023 00:38:06

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.000	00:33
Breath Volume	3.035	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.000	00:36
Breath Volume	3.406	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:38

Data Download Date/Time: 07/18/2023 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 06/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ALFREDO

DATE/TIME OF TEST: 07/08/2023 04:04:23

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.078	04:03	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 07/13/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 06/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 07/09/2023 04:02:26

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.195	03:57
Breath Volume	3.421	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.191	04:00
Breath Volume	2.546	04:00
Air Blank	0.000	04:01
Control Test	0.077	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 07/13/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 06/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: HUBBARD VII

DATE/TIME OF TEST: 07/13/2023 04:06:15

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.076	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.116	04:01
Breath Volume	2.152	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.113	04:04
Breath Volume	1.695	04:04
Air Blank	0.000	04:05
Control Test	0.076	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 07/13/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: B652298710130

DATE/TIME OF TEST: 07/01/2023 01:41:41

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.188	01:36
Breath Volume	2.753	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.187	01:40
Breath Volume	3.308	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 07/05/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,ALEXANDER,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GIVENS,ETHAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: BUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: TRACY

DATE/TIME OF TEST: 07/01/2023 18:45:55

Diagnostics Check	OK	18:37	
Air Blank	0.000	18:37	
Control Test	0.078	18:38	
Air Blank	0.000	18:38	
Subject Sample	# 1	18:41	No Sample Provided
Breath Volume	0.000	18:41	
Air Blank	0.000	18:42	
Air Blank	0.000	18:44	
Subject Sample	# 2	18:44	Subject Test Refused
Breath Volume	0.000	18:44	
Air Blank	0.000	18:45	
Control Test	0.078	18:45	
Air Blank	0.000	18:45	
Diagnostics Check	OK	18:45	

Data Download Date/Time: 07/05/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICK,JESSE,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: GRUBA,JACK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: WICK

DATE/TIME OF TEST: 07/01/2023 19:02:55

Diagnostics Check	OK	18:51	
Air Blank	0.000	18:51	
Control Test	0.078	18:51	
Air Blank	0.000	18:52	
Subject Sample	# 1 0.287	18:55	Volume Not Met
Breath Volume	0.156	18:55	
Air Blank	0.000	18:56	
Air Blank	0.000	18:58	
Subject Sample	# 2 0.245	19:01	Volume Not Met
Breath Volume	0.429	19:01	
Air Blank	0.000	19:01	
Control Test	0.076	19:02	
Air Blank	0.000	19:02	
Diagnostics Check	OK	19:02	

Data Download Date/Time: 07/05/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: GARVEY

DATE/TIME OF TEST: 07/02/2023 04:37:18

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.098	04:32
Breath Volume	1.187	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.097	04:35
Breath Volume	1.460	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 07/05/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MULTY,TANISHA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: SOEHNGEN

DATE/TIME OF TEST: 07/02/2023 04:56:24

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:45	
Control Test	0.078	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:49	No Sample Provided
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	
Air Blank	0.000	04:51	
Subject Sample	# 2	04:55	No Sample Provided
Breath Volume	0.000	04:55	
Air Blank	0.000	04:55	
Control Test	0.079	04:55	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 07/05/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LADD,JACOB
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 07/03/2023 03:21:15

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.276	03:16
Breath Volume	3.535	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.268	03:19
Breath Volume	3.574	03:19
Air Blank	0.000	03:20
Control Test	0.077	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 07/05/2023 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TITKO,NICOLE,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PIKE,TREVOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: MOREJON

DATE/TIME OF TEST: 07/04/2023 23:57:41

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.286	23:53
Breath Volume	1.691	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.289	23:56
Breath Volume	2.152	23:56
Air Blank	0.000	23:56
Control Test	0.077	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 07/05/2023 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/05/2023 02:37:32

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.079	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.015	02:35	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 07/05/2023 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TITKO,NICOLE,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TITKO,NICOLE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 07/05/2023 02:49:29

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.078	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.076	02:44
Breath Volume	1.527	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.087	02:47
Breath Volume	2.839	02:47
Air Blank	0.000	02:48
Control Test	0.078	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 07/05/2023 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ANDRADE,ESTEBAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: FRITZ

DATE/TIME OF TEST: 07/05/2023 06:12:58

Diagnostics Check	OK	06:10	
Air Blank	0.000	06:10	
Control Test	0.079	06:10	
Air Blank	0.000	06:11	
Subject Sample	# 1	06:11	Subject Test Refused
Breath Volume	0.000	06:11	
Air Blank	0.000	06:12	
Control Test	0.079	06:12	
Air Blank	0.000	06:12	
Diagnostics Check	OK	06:12	

Data Download Date/Time: 07/05/2023 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOWELL,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 07/05/2023 17:45:49

Diagnostics Check	OK	17:42	
Air Blank	0.000	17:42	
Control Test	0.079	17:43	
Air Blank	0.000	17:43	
Subject Sample	# 1 0.225	17:44	
Breath Volume	1.941	17:44	
Air Blank		17:44	RFI Detect
Air Blank	0.000	17:44	
Control Test	0.079	17:45	
Air Blank	0.000	17:45	
Diagnostics Check	OK	17:45	

Data Download Date/Time: 08/02/2023 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOWELL,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 07/05/2023 17:53:32

Diagnostics Check	OK	17:47
Air Blank	0.000	17:47
Control Test	0.079	17:48
Air Blank	0.000	17:48
Subject Sample # 1	0.236	17:49
Breath Volume	1.648	17:49
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.233	17:51
Breath Volume	1.812	17:51
Air Blank	0.000	17:52
Control Test	0.077	17:52
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 08/02/2023 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: ONSGARD

DATE/TIME OF TEST: 07/07/2023 00:44:01

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.167	00:39
Breath Volume	2.800	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.153	00:42
Breath Volume	3.199	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:44

Data Download Date/Time: 08/02/2023 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GARRETT III

DATE/TIME OF TEST: 07/08/2023 04:37:52

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.079	04:30	
Air Blank	0.000	04:31	
Subject Sample # 1	0.000	04:32	
Breath Volume	1.910	04:32	
Air Blank	0.000	04:33	
Air Blank	0.000	04:35	
Subject Sample # 2		04:36	Subject Test Refused
Breath Volume	0.031	04:36	
Air Blank	0.000	04:36	
Control Test	0.079	04:37	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 08/02/2023 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRICKE,KEN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FRICKE,KEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 07/08/2023 05:31:45

Diagnostics Check	OK	05:25
Air Blank	0.000	05:26
Control Test	0.079	05:26
Air Blank	0.000	05:26
Subject Sample # 1	0.125	05:27
Breath Volume	1.738	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:29
Subject Sample # 2	0.129	05:30
Breath Volume	2.472	05:30
Air Blank	0.000	05:30
Control Test	0.078	05:31
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: OWENS,ROY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: HANDY

DATE/TIME OF TEST: 07/11/2023 02:46:06

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.181	02:41
Breath Volume	2.347	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.169	02:44
Breath Volume	1.832	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KETCHAM,JENNIFER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: TREMINO GOMEZ

DATE/TIME OF TEST: 07/12/2023 04:32:05

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.079	04:21	
Air Blank	0.000	04:21	
Subject Sample	# 1 0.203	04:24	Volume Not Met
Breath Volume	0.132	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample	# 2 0.169	04:30	Volume Not Met
Breath Volume	0.437	04:30	
Air Blank	0.000	04:31	
Control Test	0.076	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KETCHAM,JENNIFER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: TREMINO GOMEZ

DATE/TIME OF TEST: 07/12/2023 04:41:47

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.077	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.193	04:36
Breath Volume	1.484	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.200	04:40
Breath Volume	1.843	04:40
Air Blank	0.000	04:40
Control Test	0.077	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: LA MATTINA

DATE/TIME OF TEST: 07/15/2023 03:32:08

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.120	03:27
Breath Volume	1.402	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.116	03:30
Breath Volume	1.398	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,JEDIDIAH,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: GRUBA,JACK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 07/15/2023 04:18:03

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.079	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.078	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TITKO,NICOLE,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TITKO,NICOLE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: MCKAIN

DATE/TIME OF TEST: 07/17/2023 23:15:35

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.201	23:10
Breath Volume	1.507	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.213	23:13
Breath Volume	3.269	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MAURER

DATE/TIME OF TEST: 07/20/2023 01:46:56

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.079	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:03

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 07/20/2023 15:35:06

Diagnostics Check	OK	15:28
Air Blank	0.000	15:28
Control Test	0.079	15:29
Air Blank	0.000	15:29
Subject Sample # 1	0.173	15:30
Breath Volume	1.746	15:30
Air Blank	0.000	15:30
Air Blank	0.000	15:32
Subject Sample # 2	0.180	15:33
Breath Volume	1.898	15:33
Air Blank	0.000	15:34
Control Test	0.079	15:34
Air Blank	0.000	15:35
Diagnostics Check	OK	15:35

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,LEONARDO,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HERNANDEZ,LEONARDO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: AGUIRRE GUTIERREZ

DATE/TIME OF TEST: 07/21/2023 06:56:56

Diagnostics Check	OK	06:47	
Air Blank	0.000	06:48	
Control Test	0.079	06:48	
Air Blank	0.000	06:48	
Subject Sample	# 1 0.203	06:52	Volume Not Met
Breath Volume	0.136	06:52	
Air Blank	0.000	06:52	
Air Blank	0.000	06:54	
Subject Sample	# 2	06:55	Subject Test Refused
Breath Volume	0.210	06:55	
Air Blank	0.000	06:56	
Control Test	0.077	06:56	
Air Blank	0.000	06:56	
Diagnostics Check	OK	06:56	

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: PENT

DATE/TIME OF TEST: 07/21/2023 23:24:09

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.079	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:23	
Control Test	0.079	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ADORNO,S
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/22/2023 00:47:46

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.216	00:43
Breath Volume	2.101	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.209	00:46
Breath Volume	2.023	00:46
Air Blank	0.000	00:46
Control Test	0.078	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/23/2023 03:04:54

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.201	02:59
Breath Volume	2.394	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.201	03:03
Breath Volume	2.589	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TITKO,NICOLE,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BUSCEMI,HUNTER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 07/24/2023 01:50:02

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.216	01:44
Breath Volume	1.179	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.206	01:48
Breath Volume	1.757	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:50

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: NEZUH

DATE/TIME OF TEST: 07/25/2023 00:34:11

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.000	00:29
Breath Volume	3.609	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:32
Breath Volume	3.332	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: DALEY DIXON

DATE/TIME OF TEST: 07/29/2023 01:09:17

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Control Test	0.080	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: JEANTINE

DATE/TIME OF TEST: 07/29/2023 22:53:23

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.079	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.202	22:48
Breath Volume	1.277	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.206	22:51
Breath Volume	2.066	22:51
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: ISABELLE

DATE/TIME OF TEST: 07/31/2023 03:17:02

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.079	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 08/02/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 07/28/2023 18:47:23

Diagnostics Check	OK	18:44	
Air Blank	0.000	18:44	
Control Test	0.077	18:45	
Air Blank	0.000	18:45	
Subject Sample	# 1	18:45	Subject Test Refused
Breath Volume	0.000	18:45	
Air Blank	0.000	18:46	
Control Test	0.078	18:46	
Air Blank	0.000	18:47	
Diagnostics Check	OK	18:47	

Data Download Date/Time: 08/14/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GONZALEZ-LOPEZ

DATE/TIME OF TEST: 07/30/2023 02:56:40

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.078	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.015	02:52
Breath Volume	1.500	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.012	02:55
Breath Volume	1.652	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 08/14/2023 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 07/13/2023 00:51:18

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.236	00:46
Breath Volume	3.246	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.251	00:49
Breath Volume	3.367	00:49
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 07/20/2023 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: CLOUD

DATE/TIME OF TEST: 07/01/2023 17:55:46

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:53	
Control Test	0.078	17:53	
Air Blank	0.000	17:53	
Subject Sample	# 1	17:54	Subject Test Refused
Breath Volume	0.000	17:54	
Air Blank	0.000	17:54	
Control Test	0.078	17:55	
Air Blank	0.000	17:55	
Diagnostics Check	OK	17:55	

Data Download Date/Time: 07/03/2023 19:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 07/01/2023 02:25:24

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.230	02:20
Breath Volume	2.472	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.229	02:23
Breath Volume	2.046	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 07/18/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 03:10
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 06/01/2023	OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: PABON	DATE/TIME OF TEST: 07/02/2023	03:24:04
----------------------------	-------------------------------	----------

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.127	03:19
Breath Volume	1.953	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.119	03:22
Breath Volume	1.574	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/18/2023 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: BARONE

DATE/TIME OF TEST: 07/14/2023 02:52:48

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 1 0.277	02:44	Volume Not Met
Breath Volume	0.046	02:44	
Air Blank	0.000	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 2 0.290	02:47	
Breath Volume	1.542	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 3 0.276	02:51	
Breath Volume	1.816	02:51	
Air Blank	0.000	02:51	
Control Test	0.078	02:52	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 07/18/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 02:25
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 06/01/2023	OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: TOCASUCHE	DATE/TIME OF TEST: 07/16/2023	02:42:13
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.190	02:37
Breath Volume	1.785	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.176	02:40
Breath Volume	1.367	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 07/18/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: THOMAS,ROGER

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ERWIN III

DATE/TIME OF TEST: 07/16/2023 04:44:56

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.232	04:40
Breath Volume	3.429	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:42
Subject Sample # 2	0.233	04:43
Breath Volume	3.503	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 07/18/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 01:20
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 07/18/2023	OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BOLTON	DATE/TIME OF TEST: 07/28/2023 01:32:38
-----------------------------	--

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.124	01:28
Breath Volume	1.312	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.127	01:31
Breath Volume	1.542	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 09/05/2023 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,MATTHEW,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DAVIS,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024023571
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: JEROME

DATE/TIME OF TEST: 07/04/2023 19:41:01

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:31	
Control Test	0.081	19:31	
Air Blank	0.000	19:32	
Subject Sample	# 1 0.281	19:35	Volume Not Met
Breath Volume	0.082	19:35	
Air Blank	0.000	19:36	
Air Blank	0.000	19:37	
Subject Sample	# 2	19:39	Subject Test Refused
Breath Volume	0.121	19:39	
Air Blank	0.000	19:40	
Control Test	0.078	19:40	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:41	

Data Download Date/Time: 08/01/2023 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,MATTHEW,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DAVIS,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024023571
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: JEROME

DATE/TIME OF TEST: 07/04/2023 20:09:50

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:05	
Control Test	0.079	20:05	
Air Blank	0.000	20:05	
Subject Sample	# 1	20:08	Subject Test Refused
Breath Volume	0.042	20:08	
Air Blank		20:09	Purge Fail
Air Blank	0.000	20:09	

Data Download Date/Time: 08/01/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REILLY,S.,M

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: REILLY,S.

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 3024023571

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 01:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MENDOZA ARGUELLO

DATE/TIME OF TEST: 07/05/2023 01:52:07

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.124	01:47
Breath Volume	2.675	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.130	01:50
Breath Volume	3.890	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 08/01/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,MATTHEW,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GARCIA,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024023571
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MARTINEZ VARELA

DATE/TIME OF TEST: 07/10/2023 05:07:00

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.164	05:02
Breath Volume	1.996	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:04
Subject Sample # 2	0.161	05:05
Breath Volume	1.687	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:07

Data Download Date/Time: 08/01/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,MATTHEW,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GARCIA,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024023571
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: ESCOBAR GIL

DATE/TIME OF TEST: 07/22/2023 21:20:01

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.163	21:15
Breath Volume	2.355	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.158	21:18
Breath Volume	2.175	21:18
Air Blank	0.000	21:19
Control Test	0.078	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:20

Data Download Date/Time: 08/01/2023 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: KOKOREV

DATE/TIME OF TEST: 07/16/2023 05:21:47

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:19	
Control Test	0.078	05:19	
Air Blank	0.000	05:19	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:20	
Control Test	0.077	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 08/17/2023 04:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FUJITA-MORENO

DATE/TIME OF TEST: 07/20/2023 02:55:17

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.079	02:53	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank	0.000	02:54	
Control Test	0.077	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 08/17/2023 04:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DIAZ ORAN,A
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: FICHMAN

DATE/TIME OF TEST: 07/21/2023 04:45:01

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.079	04:42	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Control Test	0.076	04:44	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 08/17/2023 04:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: SEGURA

DATE/TIME OF TEST: 07/22/2023 06:11:39

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:57	
Control Test	0.079	05:58	
Air Blank	0.000	05:58	
Subject Sample	# 1 0.098	06:01	Volume Not Met
Breath Volume	0.074	06:01	
Air Blank	0.000	06:02	
Air Blank	0.000	06:04	
Subject Sample	# 2 0.108	06:04	
Breath Volume	1.128	06:04	
Air Blank	0.000	06:05	
Air Blank	0.000	06:07	
Subject Sample	# 3 0.106	06:10	
Breath Volume	1.527	06:10	
Air Blank	0.000	06:10	
Control Test	0.076	06:11	
Air Blank	0.000	06:11	
Diagnostics Check	OK	06:11	

Data Download Date/Time: 08/17/2023 04:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: BALLESTER MONTESINO

DATE/TIME OF TEST: 07/27/2023 01:13:53

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.097	01:09
Breath Volume	2.343	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:11
Subject Sample # 2	0.104	01:12
Breath Volume	2.632	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 08/17/2023 04:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,S
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 07/29/2023 00:06:02

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.079	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Control Test	0.078	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 08/17/2023 04:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: VILLON

DATE/TIME OF TEST: 07/29/2023 03:10:31

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.078	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.131	03:05
Breath Volume	2.656	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.131	03:08
Breath Volume	2.660	03:08
Air Blank	0.000	03:09
Control Test	0.078	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 08/17/2023 04:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HALES,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 07/01/2023 19:19:49

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:17	
Control Test	0.081	19:17	
Air Blank	0.000	19:18	
Subject Sample	# 1	19:18	Subject Test Refused
Breath Volume	0.000	19:18	
Air Blank	0.000	19:18	
Control Test	0.081	19:19	
Air Blank	0.000	19:19	
Diagnostics Check	OK	19:19	

Data Download Date/Time: 07/05/2023 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 07/02/2023 04:28:06

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1 0.217	04:20	Volume Not Met
Breath Volume	0.101	04:20	
Air Blank	0.000	04:21	
Air Blank	0.000	04:23	
Subject Sample	# 2 0.223	04:26	Volume Not Met
Breath Volume	0.203	04:26	
Air Blank	0.000	04:27	
Control Test	0.078	04:27	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 07/05/2023 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALTUNAY,MUHAMMED
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MC GINLEY

DATE/TIME OF TEST: 07/02/2023 04:46:39

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.078	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.000	04:42
Breath Volume	1.332	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.000	04:45
Breath Volume	1.476	04:45
Air Blank	0.000	04:45
Control Test	0.078	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 07/05/2023 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: YEPES

DATE/TIME OF TEST: 07/04/2023 04:29:01

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.081	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.112	04:24
Breath Volume	2.554	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.111	04:27
Breath Volume	2.527	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:29

Data Download Date/Time: 07/05/2023 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COPELAND, MICHAEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: HEARN

DATE/TIME OF TEST: 07/26/2023 00:59:26

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Subject Sample # 1	0.212	00:50	
Breath Volume	1.601	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:52	
Subject Sample # 2		00:56	No Sample Provided
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample # 3		00:58	Sequence Aborted
Breath Volume	0.000	00:58	
Air Blank	0.000	00:59	

Data Download Date/Time: 08/25/2023 05:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: PAGANO

DATE/TIME OF TEST: 07/03/2023 01:25:11

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.077	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:24	
Control Test	0.078	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 07/30/2023 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/12/2023 22:55:39

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.213	22:50
Breath Volume	3.679	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.213	22:53
Breath Volume	3.183	22:53
Air Blank	0.000	22:54
Control Test	0.077	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 07/30/2023 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: CECCHINI

DATE/TIME OF TEST: 07/18/2023 23:44:08

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.078	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.042	23:39
Breath Volume	1.980	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:41
Subject Sample # 2	0.045	23:42
Breath Volume	1.332	23:42
Air Blank	0.000	23:43
Control Test	0.078	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 07/30/2023 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/08/2023 00:54:41

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.176	00:50
Breath Volume	2.695	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.174	00:53
Breath Volume	2.562	00:53
Air Blank	0.000	00:53
Control Test	0.079	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCNIN,DAVE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 07/09/2023 01:48:11

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:45	
Subject Sample # 1	0.134	01:46	
Breath Volume	2.179	01:46	
Air Blank		01:46	RFI Detect
Air Blank	0.000	01:47	
Control Test	0.081	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 07/09/2023 01:56:22

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.138	01:51
Breath Volume	1.753	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.142	01:54
Breath Volume	2.156	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: NOLASCO

DATE/TIME OF TEST: 07/10/2023 01:36:27

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.252	01:28
Breath Volume	3.625	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.222	01:31
Breath Volume	3.226	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 3	0.241	01:34
Breath Volume	2.765	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STERN,ADAM
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: OSORIO

DATE/TIME OF TEST: 07/10/2023 04:52:37

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.134	04:48
Breath Volume	2.437	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.132	04:51
Breath Volume	2.093	04:51
Air Blank	0.000	04:51
Control Test	0.080	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RIOSECO,LUIS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: MCSHERRY

DATE/TIME OF TEST: 07/10/2023 23:23:11

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.204	23:18
Breath Volume	1.511	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.200	23:21
Breath Volume	3.894	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: ELEAZARD

DATE/TIME OF TEST: 07/11/2023 03:00:39

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.150	02:56
Breath Volume	1.535	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.151	02:59
Breath Volume	1.773	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DENARDO,RYAN
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: NORTHROP

DATE/TIME OF TEST: 07/11/2023 22:39:56

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.290	22:35
Breath Volume	2.074	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.293	22:38
Breath Volume	2.410	22:38
Air Blank	0.000	22:39
Control Test	0.078	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SANTIAGO,LEONEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: NAVES

DATE/TIME OF TEST: 07/14/2023 03:39:55

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.136	03:35
Breath Volume	1.804	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.138	03:38
Breath Volume	2.070	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALTUNAY,MUHAMMED
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SHROPSHIRE

DATE/TIME OF TEST: 07/14/2023 04:43:51

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:33	
Subject Sample	# 1	04:36	No Sample Provided
Breath Volume	0.000	04:36	
Air Blank	0.000	04:37	
Air Blank	0.000	04:38	
Subject Sample	# 2	04:42	No Sample Provided
Breath Volume	0.000	04:42	
Air Blank	0.000	04:42	
Control Test	0.079	04:43	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALTUNAY,MUHAMMED
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:38

SUBJECT'S LAST NAME: YANEZ

DATE/TIME OF TEST: 07/14/2023 09:16:29

Diagnostics Check	OK	09:10
Air Blank	0.000	09:10
Control Test	0.080	09:11
Air Blank	0.000	09:11
Subject Sample # 1	0.000	09:12
Breath Volume	2.500	09:12
Air Blank	0.000	09:12
Air Blank	0.000	09:14
Subject Sample # 2	0.000	09:15
Breath Volume	3.007	09:15
Air Blank	0.000	09:15
Control Test	0.080	09:15
Air Blank	0.000	09:16
Diagnostics Check	OK	09:16

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FRATICHELLI,RICARDO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: HUMPHREYS

DATE/TIME OF TEST: 07/20/2023 20:47:55

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.080	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.292	20:43
Breath Volume	1.605	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.300	20:46
Breath Volume	1.757	20:46
Air Blank	0.000	20:47
Control Test	0.079	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VALENTIN,ENRIQUE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: MELTON II

DATE/TIME OF TEST: 07/22/2023 18:13:50

Diagnostics Check	OK	18:07
Air Blank	0.000	18:07
Control Test	0.079	18:08
Air Blank	0.000	18:08
Subject Sample # 1	0.184	18:09
Breath Volume	4.242	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.189	18:12
Breath Volume	4.144	18:12
Air Blank	0.000	18:12
Control Test	0.080	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SANTIAGO,LEONEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: PARBOOSINGH

DATE/TIME OF TEST: 07/23/2023 02:24:07

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.087	02:19
Breath Volume	1.906	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.092	02:22
Breath Volume	1.867	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SCOTT,HUNTER
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: PRABHAKAR

DATE/TIME OF TEST: 07/23/2023 03:55:07

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.139	03:50
Breath Volume	1.449	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.131	03:53
Breath Volume	2.238	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLEURANT,RODSON
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 07/23/2023 05:07:43

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.155	05:02
Breath Volume	1.664	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:04
Subject Sample # 2	0.151	05:06
Breath Volume	1.757	05:06
Air Blank	0.000	05:06
Control Test	0.078	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/26/2023 02:20:00

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.240	02:14
Breath Volume	2.527	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.239	02:18
Breath Volume	1.683	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:20

Data Download Date/Time: 08/02/2023 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 07/29/2023 04:37:25

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.218	04:32
Breath Volume	2.378	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.221	04:35
Breath Volume	3.101	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 08/02/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LOBBAN

DATE/TIME OF TEST: 07/31/2023 01:53:50

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.125	01:47
Breath Volume	1.570	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.116	01:52
Breath Volume	1.191	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 08/02/2023 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BELTON

DATE/TIME OF TEST: 07/03/2023 03:19:38

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.076	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.184	03:15
Breath Volume	1.414	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.183	03:18
Breath Volume	1.644	03:18
Air Blank	0.000	03:18
Control Test	0.077	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/30/2023 20:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GUERIN

DATE/TIME OF TEST: 07/14/2023 02:15:35

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.077	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.099	02:10
Breath Volume	1.363	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.099	02:14
Breath Volume	2.175	02:14
Air Blank	0.000	02:14
Control Test	0.077	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 07/30/2023 20:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENRRIQUEZ,A,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: APONTE,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: TORRES ROMERO

DATE/TIME OF TEST: 07/02/2023 04:12:27

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.078	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.178	04:07
Breath Volume	1.425	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.181	04:10
Breath Volume	1.816	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 07/13/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAKUBEK, THOMAS, T
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GRISALES, S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/02/2023 21:47:12

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.079	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.201	21:42
Breath Volume	1.804	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:44
Subject Sample # 2	0.210	21:45
Breath Volume	1.531	21:45
Air Blank	0.000	21:46
Control Test	0.076	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 07/13/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNSBERGER,L,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: VELEZ,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 07/03/2023 22:47:14

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.202	22:42
Breath Volume	4.179	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.202	22:45
Breath Volume	4.027	22:45
Air Blank	0.000	22:46
Control Test	0.078	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 07/13/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ROHRAFF,HARRY
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: CRUZ ALONZO

DATE/TIME OF TEST: 07/04/2023 21:18:00

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.122	21:13
Breath Volume	1.476	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.138	21:16
Breath Volume	1.464	21:16
Air Blank	0.000	21:17
Control Test	0.077	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:18

Data Download Date/Time: 07/13/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GULOTTA,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: NARDELLA

DATE/TIME OF TEST: 07/06/2023 00:36:14

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.185	00:31
Breath Volume	2.324	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.181	00:34
Breath Volume	3.316	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 07/13/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CAMPBELL,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/07/2023 01:26:05

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.148	01:21
Breath Volume	2.855	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.142	01:24
Breath Volume	2.652	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 07/13/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,DEAN,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PARKS,DEAN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: KUTNER

DATE/TIME OF TEST: 07/09/2023 15:43:12

Diagnostics Check	OK	15:37
Air Blank	0.000	15:37
Control Test	0.079	15:37
Air Blank	0.000	15:38
Subject Sample # 1	0.000	15:38
Breath Volume	3.308	15:38
Air Blank	0.000	15:39
Air Blank	0.000	15:41
Subject Sample # 2	0.000	15:41
Breath Volume	3.636	15:41
Air Blank	0.000	15:42
Control Test	0.079	15:42
Air Blank	0.000	15:43
Diagnostics Check	OK	15:43

Data Download Date/Time: 07/13/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLOAT,LUCAS,G
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GULOTTA,JUSTIN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HOLMAN

DATE/TIME OF TEST: 07/11/2023 01:24:18

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.125	01:19
Breath Volume	3.113	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.118	01:22
Breath Volume	2.769	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 07/13/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENRRIQUEZ,A,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SILVA,J
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GARCIAGUZMAN

DATE/TIME OF TEST: 07/11/2023 02:56:43

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.178	02:51
Breath Volume	1.621	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.167	02:55
Breath Volume	1.769	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/13/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAHAGIAN,TARA,K
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LEVICE

DATE/TIME OF TEST: 07/11/2023 20:36:22

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.079	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.212	20:31
Breath Volume	1.449	20:31
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.221	20:34
Breath Volume	1.726	20:34
Air Blank	0.000	20:35
Control Test	0.077	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 07/13/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LESNIEWSKI,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: ANGELILLI

DATE/TIME OF TEST: 07/11/2023 22:41:36

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.000	22:37
Breath Volume	1.250	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.000	22:40
Breath Volume	2.164	22:40
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 07/13/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GUTIERREZ,I
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DREW

DATE/TIME OF TEST: 07/13/2023 02:56:03

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.173	02:51
Breath Volume	2.781	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.184	02:54
Breath Volume	2.539	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/13/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCE,MICHAEL,O
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ALLEN,TRISTIN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:57

SUBJECT'S LAST NAME: RUMMER

DATE/TIME OF TEST: 07/15/2023 10:33:27

Diagnostics Check	OK	10:25	
Air Blank	0.000	10:26	
Control Test	0.079	10:26	
Air Blank	0.000	10:27	
Subject Sample	# 1 0.210	10:28	
Breath Volume	2.078	10:28	
Air Blank	0.000	10:29	
Air Blank	0.000	10:31	
Subject Sample	# 2 0.215	10:31	
Breath Volume	2.230	10:31	
Air Blank	0.000	10:32	
Control Test	0.074	10:32	Control Outside Tolerance
Air Blank	0.000	10:33	

Data Download Date/Time: 08/17/2023 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ACCIME,L
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: GASPAR LORENZO

DATE/TIME OF TEST: 07/29/2023 18:49:33

Diagnostics Check	OK	18:42
Air Blank	0.000	18:42
Control Test	0.079	18:43
Air Blank	0.000	18:43
Subject Sample # 1	0.221	18:44
Breath Volume	2.121	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.214	18:47
Breath Volume	1.988	18:47
Air Blank	0.000	18:48
Control Test	0.078	18:48
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 08/17/2023 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD,JAMES,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GANK,DEP.
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MARADIAGA VALLE

DATE/TIME OF TEST: 07/01/2023 00:01:00

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.126	23:56
Breath Volume	3.394	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.121	23:59
Breath Volume	4.386	23:59
Air Blank	0.000	00:00
Control Test	0.078	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:01

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD,JAMES,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GANK,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 07/01/2023 03:20:09

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1 0.171	03:10	Volume Not Met
Breath Volume	0.078	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 2 0.175	03:14	
Breath Volume	2.500	03:14	
Air Blank	0.000	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 3 0.177	03:18	
Breath Volume	2.003	03:18	
Air Blank	0.000	03:19	
Control Test	0.077	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD,JAMES,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GOFF,CPL.
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CORRIVEAU

DATE/TIME OF TEST: 07/01/2023 04:11:44

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.108	04:07
Breath Volume	2.269	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.109	04:10
Breath Volume	3.296	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAZZARESE,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: DRAPOLA

DATE/TIME OF TEST: 07/01/2023 21:14:09

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.080	21:07
Air Blank	0.000	21:07
Subject Sample # 1	0.305	21:09
Breath Volume	1.503	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.294	21:12
Breath Volume	1.542	21:12
Air Blank	0.000	21:13
Control Test	0.076	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANO,MARDEE,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MANO,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/02/2023 09:54:03

Diagnostics Check	OK	09:47
Air Blank	0.000	09:48
Control Test	0.079	09:48
Air Blank	0.000	09:49
Subject Sample # 1	0.000	09:49
Breath Volume	1.324	09:49
Air Blank	0.000	09:50
Air Blank	0.000	09:52
Subject Sample # 2	0.000	09:52
Breath Volume	1.433	09:52
Air Blank	0.000	09:53
Control Test	0.078	09:53
Air Blank	0.000	09:53
Diagnostics Check	OK	09:54

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITMAN,JACELYN,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LITMAN,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 07/03/2023 19:15:20

Diagnostics Check	OK	19:08
Air Blank	0.000	19:08
Control Test	0.081	19:09
Air Blank	0.000	19:09
Subject Sample # 1	0.000	19:10
Breath Volume	1.183	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.000	19:13
Breath Volume	1.464	19:13
Air Blank	0.000	19:14
Control Test	0.080	19:14
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITMAN,JACELYN,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ACEBO,ANDRES
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: AT

DATE/TIME OF TEST: 07/03/2023 20:41:15

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.081	20:34
Air Blank	0.000	20:35
Subject Sample # 1	0.184	20:36
Breath Volume	2.230	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:38
Subject Sample # 2	0.187	20:39
Breath Volume	2.261	20:39
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WILES,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 07/04/2023 18:18:22

Diagnostics Check	OK	18:12
Air Blank	0.000	18:12
Control Test	0.080	18:12
Air Blank	0.000	18:13
Subject Sample # 1	0.000	18:14
Breath Volume	1.878	18:14
Air Blank	0.000	18:14
Air Blank	0.000	18:16
Subject Sample # 2	0.000	18:17
Breath Volume	1.496	18:17
Air Blank	0.000	18:17
Control Test	0.080	18:17
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY K,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COLON,OFFICER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: FRIG

DATE/TIME OF TEST: 07/06/2023 00:01:54

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.105	23:54	Volume Not Met
Breath Volume	0.296	23:54	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 1 0.111	00:00	Volume Not Met
Breath Volume	0.343	00:00	
Air Blank	0.000	00:00	
Control Test	0.077	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY K,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,OFFICER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 07/06/2023 01:38:26

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.095	01:33
Breath Volume	1.386	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample # 2	0.097	01:36
Breath Volume	2.054	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PERRO,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: DELPH

DATE/TIME OF TEST: 07/06/2023 08:27:42

Diagnostics Check	OK	08:21	
Air Blank	0.000	08:21	
Control Test	0.081	08:22	
Air Blank	0.000	08:22	
Subject Sample	# 1 0.170	08:23	
Breath Volume	1.687	08:23	
Air Blank	0.000	08:23	
Air Blank	0.000	08:25	
Subject Sample	# 2 0.175	08:26	
Breath Volume	3.265	08:26	
Air Blank	0.000	08:26	
Control Test	0.078	08:27	
Air Blank		08:27	RFI Detect
Air Blank	0.000	08:27	

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WONGA,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 07/09/2023 14:43:08

Diagnostics Check	OK	14:36
Air Blank	0.000	14:37
Control Test	0.080	14:37
Air Blank	0.000	14:37
Subject Sample # 1	0.000	14:38
Breath Volume	1.550	14:38
Air Blank	0.000	14:39
Air Blank	0.000	14:41
Subject Sample # 2	0.000	14:41
Breath Volume	1.484	14:41
Air Blank	0.000	14:42
Control Test	0.080	14:42
Air Blank	0.000	14:43
Diagnostics Check	OK	14:43

Data Download Date/Time: 07/31/2023 14:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,VANESSA,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,MYKELL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SOMARRIBA

DATE/TIME OF TEST: 07/03/2023 00:45:16

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.078	00:37	
Air Blank	0.000	00:37	
Subject Sample	# 1 0.181	00:39	
Breath Volume	2.660	00:39	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.077	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 07/14/2023 01:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,VANESSA,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,MYKELL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 07/04/2023 01:24:37

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.079	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 07/14/2023 01:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ANDRADE,LESLIET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: BALBUENA GUTIERREZ

DATE/TIME OF TEST: 07/12/2023 22:47:31

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:45	
Control Test	0.081	22:45	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Control Test	0.081	22:46	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 07/14/2023 01:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/31/2023 04:52:16

Diagnostics Check	Fail	04:51	Internal Printer Test
Air Blank	0.000	04:52	

Data Download Date/Time: 08/02/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/31/2023 04:55:36

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:55	
Control Test	0.069	04:55	Control Outside Tolerance
Air Blank		04:55	RFI Detect

Data Download Date/Time: 08/02/2023 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/31/2023 05:01:16

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:58	
Control Test	0.080	04:58	
Air Blank	0.000	04:59	
Subject Sample	# 1	04:59	RFI Detect
Breath Volume	0.000	04:59	
Air Blank	0.000	04:59	
Control Test	0.079	05:00	
Air Blank	0.000	05:00	
Diagnostics Check	Fail	05:00	Internal Printer Test
Air Blank	0.000	05:01	

Data Download Date/Time: 08/02/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/31/2023 05:09:11

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.079	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.090	05:04
Breath Volume	1.335	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample # 2	0.093	05:07
Breath Volume	1.570	05:07
Air Blank	0.000	05:08
Control Test	0.079	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 08/02/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GARCIA,MELANIE
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 07/01/2023 01:23:16

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.082	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.000	01:18
Breath Volume	1.597	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.000	01:21
Breath Volume	2.300	01:21
Air Blank	0.000	01:22
Control Test	0.082	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 07/05/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,FRANK
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 07/01/2023 01:39:33

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.082	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.202	01:34
Breath Volume	2.644	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.204	01:37
Breath Volume	2.328	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 07/05/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: OLIVERA,RAUL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: YALDO

DATE/TIME OF TEST: 07/02/2023 19:43:56

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.083	19:38
Air Blank	0.000	19:39
Subject Sample # 1	0.000	19:39
Breath Volume	2.667	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.000	19:42
Breath Volume	2.906	19:42
Air Blank	0.000	19:43
Control Test	0.083	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 07/05/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: EMERENCIANO,MATHEUS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/02/2023 22:03:43

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.082	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.187	21:59
Breath Volume	2.511	21:59
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.192	22:02
Breath Volume	2.066	22:02
Air Blank	0.000	22:02
Control Test	0.082	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 07/05/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PAYNE,BRIANNA
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 07/02/2023 22:45:16

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.082	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.191	22:40
Breath Volume	1.507	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.200	22:43
Breath Volume	1.617	22:43
Air Blank	0.000	22:44
Control Test	0.081	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 07/05/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BONDS,JAMES
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 07/02/2023 23:32:10

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.082	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.130	23:27
Breath Volume	3.375	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.126	23:30
Breath Volume	2.929	23:30
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 07/05/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: GONZALEZ BAUZA

DATE/TIME OF TEST: 07/03/2023 18:03:41

Diagnostics Check	OK	17:56
Air Blank	0.000	17:57
Control Test	0.082	17:57
Air Blank	0.000	17:58
Subject Sample # 1	0.162	17:59
Breath Volume	1.210	17:59
Air Blank	0.000	17:59
Air Blank	0.000	18:01
Subject Sample # 2	0.167	18:02
Breath Volume	1.437	18:02
Air Blank	0.000	18:02
Control Test	0.082	18:03
Air Blank	0.000	18:03
Diagnostics Check	OK	18:03

Data Download Date/Time: 07/05/2023 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PEREZ MARTINEZ

DATE/TIME OF TEST: 07/05/2023 03:30:31

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.083	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.200	03:25
Breath Volume	3.042	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.200	03:28
Breath Volume	2.468	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 07/05/2023 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 07/01/2023 03:08:06

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.053	03:03
Breath Volume	2.152	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.052	03:06
Breath Volume	1.324	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 07/05/2023 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MANGOL

DATE/TIME OF TEST: 07/02/2023 01:37:39

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.351	01:30
Breath Volume	1.785	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.351	01:35
Breath Volume	1.132	01:35
Air Blank	0.000	01:36
Control Test	0.077	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 07/05/2023 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: RUCCI CRUZ

DATE/TIME OF TEST: 07/02/2023 02:55:43

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.078	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1 0.130	02:48	Slope Not Level
Breath Volume	1.562	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 2 0.118	02:54	Volume Not Met
Breath Volume	0.730	02:54	
Air Blank	0.000	02:54	
Control Test	0.079	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 07/05/2023 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: RUCCI CRUZ

DATE/TIME OF TEST: 07/02/2023 03:08:57

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.078	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1 0.130	03:01	Volume Not Met
Breath Volume	0.812	03:01	
Air Blank	0.000	03:02	
Air Blank	0.000	03:04	
Subject Sample	# 2 0.122	03:07	Volume Not Met
Breath Volume	0.574	03:07	
Air Blank	0.000	03:08	
Control Test	0.078	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 07/05/2023 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: YZQUIERDO

DATE/TIME OF TEST: 07/02/2023 22:42:26

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.135	22:37
Breath Volume	3.015	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.135	22:40
Breath Volume	2.796	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 07/05/2023 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: FLORES HERNANDEZ

DATE/TIME OF TEST: 07/03/2023 05:31:11

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:19	
Control Test	0.080	05:19	
Air Blank	0.000	05:20	
Subject Sample	# 1 0.172	05:23	Volume Not Met
Breath Volume	0.601	05:23	
Air Blank	0.000	05:23	
Air Blank	0.000	05:25	
Subject Sample	# 2 0.154	05:26	No .020 agreement
Breath Volume	1.394	05:26	
Air Blank	0.000	05:26	
Air Blank	0.000	05:28	
Subject Sample	# 3 0.178	05:29	No .020 agreement
Breath Volume	1.746	05:29	
Air Blank	0.000	05:30	
Control Test	0.079	05:30	
Air Blank	0.000	05:31	
Diagnostics Check	OK	05:31	

Data Download Date/Time: 07/05/2023 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: FLORES HERNANDEZ

DATE/TIME OF TEST: 07/03/2023 05:39:15

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.078	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.170	05:34
Breath Volume	1.718	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.168	05:37
Breath Volume	1.406	05:37
Air Blank	0.000	05:38
Control Test	0.078	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 07/05/2023 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: VIVERETTE

DATE/TIME OF TEST: 07/04/2023 23:30:12

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.079	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.144	23:25
Breath Volume	3.144	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.141	23:28
Breath Volume	3.203	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 07/05/2023 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 07/05/2023 00:19:14

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.079	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1 0.238	00:13	
Breath Volume	1.367	00:13	
Air Blank	0.000	00:13	
Air Blank	0.000	00:15	
Subject Sample	# 2	00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank	0.000	00:18	
Control Test	0.080	00:18	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

Data Download Date/Time: 07/05/2023 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: MEJILLA AMALLA

DATE/TIME OF TEST: 07/05/2023 04:52:24

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.241	04:47
Breath Volume	1.566	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.242	04:50
Breath Volume	1.578	04:50
Air Blank	0.000	04:51
Control Test	0.078	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 07/05/2023 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 07/06/2023 05:05:59

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.109	05:00
Breath Volume	1.351	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.114	05:04
Breath Volume	1.355	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: PIMENTEL

DATE/TIME OF TEST: 07/07/2023 19:32:04

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.080	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.181	19:27
Breath Volume	3.476	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:29
Subject Sample # 2	0.187	19:30
Breath Volume	2.875	19:30
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:32

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 07/08/2023 02:04:02

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.152	01:58
Breath Volume	1.886	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.145	02:02
Breath Volume	1.296	02:02
Air Blank	0.000	02:03
Control Test	0.076	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:04

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: GOMEZ AGUILAR

DATE/TIME OF TEST: 07/08/2023 02:23:02

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.212	02:18
Breath Volume	2.250	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.195	02:21
Breath Volume	1.281	02:21
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:23

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/08/2023 03:27:59

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.092	03:23
Breath Volume	1.929	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.086	03:26
Breath Volume	2.171	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEWIS,I
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ENCALADE

DATE/TIME OF TEST: 07/08/2023 04:16:26

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.150	04:11
Breath Volume	1.589	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.165	04:14
Breath Volume	2.761	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 07/09/2023 03:17:37

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.201	03:11
Breath Volume	1.507	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.213	03:15
Breath Volume	2.652	03:15
Air Blank	0.000	03:16
Control Test	0.078	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO, E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: CASTILLO MARTINEZ

DATE/TIME OF TEST: 07/09/2023 04:11:24

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.078	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.141	04:06
Breath Volume	1.335	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.133	04:09
Breath Volume	1.507	04:09
Air Blank	0.000	04:10
Control Test	0.078	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/09/2023 04:32:13

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.160	04:27
Breath Volume	3.691	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.161	04:30
Breath Volume	1.550	04:30
Air Blank	0.000	04:31
Control Test	0.078	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 07/13/2023 04:09:53

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.234	04:05
Breath Volume	3.406	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.236	04:08
Breath Volume	3.312	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: STABLEY-VAUGHN

DATE/TIME OF TEST: 07/14/2023 06:25:29

Diagnostics Check	OK	06:18
Air Blank	0.000	06:19
Control Test	0.079	06:19
Air Blank	0.000	06:20
Subject Sample # 1	0.171	06:20
Breath Volume	1.589	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.171	06:23
Breath Volume	1.308	06:23
Air Blank	0.000	06:24
Control Test	0.078	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Temple Terrace PD
ARRESTING OFFICER: MOLITOR,M
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/14/2023 23:48:36

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.196	23:42
Breath Volume	1.792	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.189	23:46
Breath Volume	1.523	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: PARAMORE

DATE/TIME OF TEST: 07/15/2023 00:38:43

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.155	00:34
Breath Volume	1.812	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.173	00:37
Breath Volume	1.906	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: JONES IV

DATE/TIME OF TEST: 07/15/2023 03:26:18

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.137	03:21
Breath Volume	2.644	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.137	03:24
Breath Volume	1.835	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 08/01/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 07/16/2023 02:54:03

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.241	02:49
Breath Volume	2.515	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.239	02:52
Breath Volume	1.964	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:54

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SERRANO COLLAZO

DATE/TIME OF TEST: 07/16/2023 03:04:56

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.078	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.094	03:00
Breath Volume	1.992	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample # 2	0.092	03:03
Breath Volume	2.355	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GILLIAM

DATE/TIME OF TEST: 07/22/2023 03:08:41

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.143	03:03
Breath Volume	1.890	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.124	03:07
Breath Volume	1.257	03:07
Air Blank	0.000	03:07
Control Test	0.078	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: DE LOS SANTOS AMARAN

DATE/TIME OF TEST: 07/23/2023 03:35:53

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.065	03:30
Breath Volume	1.285	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.070	03:34
Breath Volume	1.187	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: PLEASANT

DATE/TIME OF TEST: 07/23/2023 04:24:50

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.161	04:20
Breath Volume	3.015	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.157	04:23
Breath Volume	2.835	04:23
Air Blank	0.000	04:23
Control Test	0.078	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NORMAN,ROBERT
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 07/23/2023 18:34:06

Diagnostics Check	OK	18:16	
Air Blank	0.000	18:17	
Control Test	0.079	18:17	
Air Blank	0.000	18:17	
Subject Sample	# 1 0.238	18:21	
Breath Volume	1.386	18:21	
Air Blank	0.000	18:21	
Air Blank	0.000	18:23	
Subject Sample	# 2 0.217	18:26	Volume Not Met
Breath Volume	0.875	18:26	
Air Blank	0.000	18:27	
Air Blank	0.000	18:29	
Subject Sample	# 3 0.188	18:32	Volume Not Met
Breath Volume	0.046	18:32	
Air Blank	0.000	18:33	
Control Test	0.078	18:33	
Air Blank	0.000	18:34	
Diagnostics Check	OK	18:34	

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NORMAN,ROBERT
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 07/23/2023 18:51:08

Diagnostics Check	OK	18:36	
Air Blank	0.000	18:37	
Control Test	0.077	18:37	
Air Blank	0.000	18:37	
Subject Sample	# 1 0.233	18:41	Volume Not Met
Breath Volume	0.144	18:41	
Air Blank	0.000	18:41	
Air Blank	0.000	18:43	
Subject Sample	# 2 0.231	18:45	No .020 agreement
Breath Volume	1.738	18:45	
Air Blank	0.000	18:46	
Air Blank	0.000	18:48	
Subject Sample	# 3 0.204	18:49	No .020 agreement
Breath Volume	1.464	18:49	
Air Blank	0.000	18:50	
Control Test	0.077	18:50	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NORMAN,ROBERT
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 07/23/2023 19:15:35

Diagnostics Check	OK	19:00	
Air Blank	0.000	19:00	
Control Test	0.078	19:01	
Air Blank	0.000	19:01	
Subject Sample	# 1 0.179	19:04	Volume Not Met
Breath Volume	0.085	19:04	
Air Blank	0.000	19:05	
Air Blank	0.000	19:07	
Subject Sample	# 2 0.204	19:08	
Breath Volume	2.109	19:08	
Air Blank	0.000	19:08	
Air Blank	0.000	19:10	
Subject Sample	# 3 0.190	19:13	Volume Not Met
Breath Volume	0.148	19:13	
Air Blank	0.000	19:14	
Control Test	0.078	19:14	
Air Blank	0.000	19:15	
Diagnostics Check	OK	19:15	

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 07/24/2023 23:51:04

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.202	23:46
Breath Volume	1.121	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.209	23:49
Breath Volume	1.675	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:51

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: ALUPURI

DATE/TIME OF TEST: 07/25/2023 05:54:13

Diagnostics Check	OK	05:45
Air Blank	0.000	05:46
Control Test	0.080	05:46
Air Blank	0.000	05:47
Subject Sample # 1	0.173	05:49
Breath Volume	1.511	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.167	05:52
Breath Volume	1.542	05:52
Air Blank	0.000	05:53
Control Test	0.077	05:53
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 07/26/2023 05:47:19

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.106	05:42
Breath Volume	3.187	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.107	05:45
Breath Volume	3.324	05:45
Air Blank	0.000	05:46
Control Test	0.079	05:46
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: PETITFRERE

DATE/TIME OF TEST: 07/27/2023 04:29:34

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.179	04:23
Breath Volume	1.179	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.182	04:27
Breath Volume	1.539	04:27
Air Blank	0.000	04:28
Control Test	0.077	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MOREDO

DATE/TIME OF TEST: 07/28/2023 01:29:25

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.175	01:24
Breath Volume	1.191	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.171	01:27
Breath Volume	1.542	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/28/2023 04:16:35

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.200	04:09
Breath Volume	1.343	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:11
Subject Sample # 2	0.190	04:14
Breath Volume	1.476	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: ERVIN

DATE/TIME OF TEST: 07/29/2023 03:24:05

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 1 0.106	03:15	Volume Not Met
Breath Volume	0.082	03:15	
Air Blank	0.000	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 2 0.098	03:18	
Breath Volume	1.484	03:18	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 3 0.100	03:22	
Breath Volume	1.210	03:22	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: SIMMS

DATE/TIME OF TEST: 07/29/2023 03:56:49

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.079	03:45	
Air Blank	0.000	03:46	
Subject Sample	# 1 0.000	03:49	Volume Not Met
Breath Volume	0.000	03:49	
Air Blank	0.000	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 2 0.102	03:52	
Breath Volume	2.449	03:52	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 3 0.096	03:55	
Breath Volume	1.242	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RESZKO

DATE/TIME OF TEST: 07/29/2023 04:12:36

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.199	04:08
Breath Volume	2.257	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.197	04:11
Breath Volume	2.093	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 07/29/2023 04:37:22

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.197	04:32
Breath Volume	2.039	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.196	04:35
Breath Volume	1.871	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEWIS,I
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ACKMAN

DATE/TIME OF TEST: 07/30/2023 01:46:00

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.263	01:41
Breath Volume	2.464	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:43
Subject Sample # 2	0.261	01:44
Breath Volume	2.289	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:46

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: VILTRES RIVERA

DATE/TIME OF TEST: 07/30/2023 03:59:13

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.051	03:54
Breath Volume	1.367	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.052	03:57
Breath Volume	1.937	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 07/30/2023 04:10:02

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.112	04:05
Breath Volume	2.167	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.115	04:08
Breath Volume	2.242	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:10

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: ADAMS JR

DATE/TIME OF TEST: 07/30/2023 04:28:02

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.117	04:23
Breath Volume	1.339	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.121	04:26
Breath Volume	2.199	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:28

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEWIS,I
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: FELDMAN

DATE/TIME OF TEST: 07/30/2023 04:48:43

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.078	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.197	04:43
Breath Volume	1.644	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.197	04:47
Breath Volume	2.269	04:47
Air Blank	0.000	04:47
Control Test	0.079	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 08/01/2023 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: JOHNSON,DARIUS
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ALAMIA

DATE/TIME OF TEST: 07/22/2023 03:24:58

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.169	03:20
Breath Volume	2.960	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:22
Subject Sample # 2	0.169	03:23
Breath Volume	2.785	03:23
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/28/2023 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 07/01/2023 00:36:33

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.079	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:29	No Sample Provided
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 2	00:35	No Sample Provided
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.079	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/01/2023 03:17:30

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.177	03:12
Breath Volume	1.527	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.190	03:15
Breath Volume	1.820	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: STEADMAN

DATE/TIME OF TEST: 07/01/2023 22:03:06

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:51	
Control Test	0.080	21:51	
Air Blank	0.000	21:52	
Subject Sample	# 1 0.174	21:55	Volume Not Met
Breath Volume	0.714	21:55	
Air Blank	0.000	21:56	
Air Blank	0.000	21:57	
Subject Sample	# 2 0.214	21:58	
Breath Volume	1.949	21:58	
Air Blank	0.000	21:59	
Air Blank	0.000	22:00	
Subject Sample	# 3 0.217	22:01	
Breath Volume	1.484	22:01	
Air Blank	0.000	22:02	
Control Test	0.079	22:02	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:03	

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/01/2023 22:20:28

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.079	22:09	
Air Blank	0.000	22:09	
Subject Sample	# 1 0.190	22:09	
Breath Volume	1.238	22:09	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample	# 2 0.165	22:15	Volume Not Met
Breath Volume	0.105	22:15	
Air Blank	0.000	22:16	
Air Blank	0.000	22:18	
Subject Sample	# 3 0.180	22:18	
Breath Volume	1.578	22:18	
Air Blank	0.000	22:19	
Control Test	0.077	22:19	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: SAABEDRA

DATE/TIME OF TEST: 07/02/2023 00:19:33

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.225	00:14
Breath Volume	1.929	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.234	00:17
Breath Volume	2.300	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: RODRIGUEZ GARCIA

DATE/TIME OF TEST: 07/02/2023 03:31:57

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.238	03:27
Breath Volume	1.230	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.256	03:30
Breath Volume	2.066	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: DOMINGUEZ CASAREZ

DATE/TIME OF TEST: 07/02/2023 04:24:17

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.094	04:19
Breath Volume	1.863	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.093	04:22
Breath Volume	1.593	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 07/02/2023 04:55:29

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:44	
Control Test	0.079	04:44	
Air Blank	0.000	04:45	
Subject Sample	# 1 0.175	04:48	Volume Not Met
Breath Volume	0.296	04:48	
Air Blank	0.000	04:48	
Air Blank	0.000	04:50	
Subject Sample	# 2 0.153	04:53	Volume Not Met
Breath Volume	0.144	04:53	
Air Blank	0.000	04:54	
Control Test	0.077	04:54	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 07/02/2023 05:27:24

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:19	
Control Test	0.078	05:19	
Air Blank	0.000	05:20	
Subject Sample	# 1 0.156	05:23	Volume Not Met
Breath Volume	0.382	05:23	
Air Blank	0.000	05:24	
Air Blank	0.000	05:25	
Subject Sample	# 2	05:26	Subject Test Refused
Breath Volume	0.000	05:26	
Air Blank	0.000	05:26	
Control Test	0.076	05:26	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 07/02/2023 07:36:43

Diagnostics Check	OK	07:30
Air Blank	0.000	07:30
Control Test	0.079	07:31
Air Blank	0.000	07:31
Subject Sample # 1	0.124	07:32
Breath Volume	1.410	07:32
Air Blank	0.000	07:32
Air Blank	0.000	07:34
Subject Sample # 2	0.122	07:35
Breath Volume	2.265	07:35
Air Blank	0.000	07:35
Control Test	0.079	07:36
Air Blank	0.000	07:36
Diagnostics Check	OK	07:36

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOLMSTROM,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: KIMBLE

DATE/TIME OF TEST: 07/02/2023 21:52:31

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.079	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.281	21:46
Breath Volume	1.550	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:48
Subject Sample # 2	0.273	21:50
Breath Volume	2.171	21:50
Air Blank	0.000	21:51
Control Test	0.077	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 07/02/2023 22:45:34

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.078	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.178	22:40
Breath Volume	2.832	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.178	22:43
Breath Volume	2.988	22:43
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 07/02/2023 23:47:49

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.078	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.111	23:42
Breath Volume	1.285	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.120	23:46
Breath Volume	1.554	23:46
Air Blank	0.000	23:46
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: LOPEZ CAPELLAN

DATE/TIME OF TEST: 07/03/2023 03:07:58

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.200	03:01
Breath Volume	1.312	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.189	03:06
Breath Volume	1.625	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 07/05/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: QUESADA CALA

DATE/TIME OF TEST: 07/03/2023 03:34:39

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.141	03:28
Breath Volume	1.656	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.140	03:32
Breath Volume	2.234	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 07/03/2023 03:49:19

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:38	
Control Test	0.078	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:42	No Sample Provided
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Air Blank	0.000	03:44	
Subject Sample	# 2	03:47	No Sample Provided
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MESA CLEMENTE

DATE/TIME OF TEST: 07/03/2023 04:06:18

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.108	04:01
Breath Volume	1.628	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.108	04:04
Breath Volume	1.285	04:04
Air Blank	0.000	04:05
Control Test	0.078	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: DELY

DATE/TIME OF TEST: 07/03/2023 05:39:11

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.079	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.154	05:33
Breath Volume	1.652	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.157	05:37
Breath Volume	1.675	05:37
Air Blank	0.000	05:38
Control Test	0.078	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 07/04/2023 02:05:35

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.079	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.155	02:00
Breath Volume	1.449	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.158	02:03
Breath Volume	1.304	02:03
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 07/04/2023 02:14:44

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.173	02:10
Breath Volume	1.246	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.176	02:13
Breath Volume	1.371	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BROSS

DATE/TIME OF TEST: 07/04/2023 03:22:35

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.173	03:17
Breath Volume	2.835	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.175	03:20
Breath Volume	1.890	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GALINANES

DATE/TIME OF TEST: 07/04/2023 03:58:51

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.143	03:54
Breath Volume	1.132	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.146	03:57
Breath Volume	2.292	03:57
Air Blank	0.000	03:57
Control Test	0.077	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: REINA MACHADO

DATE/TIME OF TEST: 07/04/2023 05:51:30

Diagnostics Check	OK	05:47	
Air Blank	0.000	05:47	
Control Test	0.078	05:48	
Air Blank	0.000	05:48	
Subject Sample	# 1	05:49	Subject Test Refused
Breath Volume	0.070	05:49	
Air Blank	0.000	05:50	
Control Test	0.079	05:50	
Air Blank	0.000	05:51	
Diagnostics Check	OK	05:51	

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: ALBERTERIS PAEZ

DATE/TIME OF TEST: 07/04/2023 22:02:50

Diagnostics Check	OK	21:54
Air Blank	0.000	21:55
Control Test	0.078	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.149	21:58
Breath Volume	2.625	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.150	22:01
Breath Volume	2.019	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: WIAN

DATE/TIME OF TEST: 07/04/2023 23:13:50

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.223	23:09
Breath Volume	1.371	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.227	23:12
Breath Volume	2.000	23:12
Air Blank	0.000	23:12
Control Test	0.078	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS, J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: RUDDY

DATE/TIME OF TEST: 07/04/2023 23:42:32

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.172	23:37
Breath Volume	1.386	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.172	23:40
Breath Volume	1.718	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: LOPEZ MILLER

DATE/TIME OF TEST: 07/05/2023 04:48:43

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.079	04:37	
Air Blank	0.000	04:37	
Subject Sample	# 1 0.169	04:40	Volume Not Met
Breath Volume	0.210	04:40	
Air Blank	0.000	04:41	
Air Blank	0.000	04:43	
Subject Sample	# 2 0.176	04:43	
Breath Volume	1.214	04:43	
Air Blank	0.000	04:44	
Air Blank	0.000	04:46	
Subject Sample	# 3 0.187	04:46	
Breath Volume	1.144	04:46	
Air Blank	0.000	04:47	
Control Test	0.078	04:48	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: MONTES DE OCA BLANCO

DATE/TIME OF TEST: 07/05/2023 05:10:34

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.180	05:05
Breath Volume	2.035	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.185	05:08
Breath Volume	1.457	05:08
Air Blank	0.000	05:09
Control Test	0.077	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 07/05/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/09/2023 04:07:28

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.192	04:02
Breath Volume	1.972	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.192	04:05
Breath Volume	2.312	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 08/01/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/15/2023 00:46:52

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.167	00:42
Breath Volume	2.417	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.175	00:45
Breath Volume	2.183	00:45
Air Blank	0.000	00:45
Control Test	0.078	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 08/01/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RESTO

DATE/TIME OF TEST: 07/15/2023 03:32:40

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.098	03:28
Breath Volume	2.945	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.097	03:31
Breath Volume	3.410	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 08/01/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: LAWSON III

DATE/TIME OF TEST: 07/15/2023 04:29:09

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.018	04:24
Breath Volume	1.656	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.021	04:27
Breath Volume	2.476	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 08/01/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: GONZALEZ RANGEL

DATE/TIME OF TEST: 07/23/2023 03:36:39

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.143	03:32
Breath Volume	1.945	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.138	03:35
Breath Volume	1.523	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 08/01/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MURRY JR

DATE/TIME OF TEST: 07/23/2023 04:24:39

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.123	04:19
Breath Volume	1.414	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.125	04:23
Breath Volume	1.328	04:23
Air Blank	0.000	04:23
Control Test	0.079	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 08/01/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: LONG JR

DATE/TIME OF TEST: 07/28/2023 01:15:16

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.359	01:10
Breath Volume	1.777	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.358	01:13
Breath Volume	1.507	01:13
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 08/01/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 07/28/2023 03:24:49

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.112	03:19
Breath Volume	1.449	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.113	03:23
Breath Volume	1.230	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 08/01/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: WHITTINGTON

DATE/TIME OF TEST: 07/29/2023 04:13:09

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.064	04:08
Breath Volume	2.585	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.062	04:11
Breath Volume	2.726	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 08/01/2023 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 07/04/2023 20:36:15

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.081	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.000	20:31
Breath Volume	1.308	20:31
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.000	20:34
Breath Volume	1.445	20:34
Air Blank	0.000	20:35
Control Test	0.081	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 07/06/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/05/2023 03:39:36

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.210	03:33
Breath Volume	1.703	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.209	03:37
Breath Volume	1.710	03:37
Air Blank	0.000	03:38
Control Test	0.077	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 07/06/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,MEGHAN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DULLY,STEPHEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: NORIEGA

DATE/TIME OF TEST: 07/01/2023 22:17:14

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.081	22:15	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:15	Subject Test Refused
Breath Volume	0.000	22:15	
Air Blank	0.000	22:16	
Control Test	0.081	22:16	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

Data Download Date/Time: 07/05/2023 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ZAMBRANO MENDOZA

DATE/TIME OF TEST: 07/04/2023 21:53:13

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.080	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.185	21:45
Breath Volume	3.683	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:47
Subject Sample # 2	0.164	21:48
Breath Volume	2.050	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 3	0.177	21:51
Breath Volume	1.898	21:51
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 07/05/2023 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: COLES

DATE/TIME OF TEST: 07/05/2023 02:20:23

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1 0.043	02:10	Volume Not Met
Breath Volume	0.730	02:10	
Air Blank	0.000	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 2 0.041	02:14	
Breath Volume	1.578	02:14	
Air Blank	0.000	02:15	
Air Blank	0.000	02:16	
Subject Sample	# 3 0.041	02:18	
Breath Volume	1.250	02:18	
Air Blank	0.000	02:19	
Control Test	0.080	02:19	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 07/05/2023 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKONOUGH,ROBERT,F
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VENTRY,WILLIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 07/02/2023 05:23:53

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.082	05:17
Air Blank	0.000	05:18
Subject Sample # 1	0.315	05:19
Breath Volume	2.753	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.315	05:22
Breath Volume	3.101	05:22
Air Blank	0.000	05:22
Control Test	0.082	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 07/03/2023 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/05/2023 04:24:47

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.082	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.060	04:19
Breath Volume	2.429	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.058	04:23
Breath Volume	2.429	04:23
Air Blank	0.000	04:23
Control Test	0.084	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 07/27/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ, JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 07/07/2023 01:59:03

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.179	01:54
Breath Volume	2.769	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.183	01:57
Breath Volume	3.316	01:57
Air Blank	0.000	01:58
Control Test	0.083	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:59

Data Download Date/Time: 07/27/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAYA,ZACHARY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 07/11/2023 09:49:16

Diagnostics Check	OK	09:43	
Air Blank	0.000	09:43	
Control Test	0.082	09:43	
Air Blank	0.000	09:44	
Subject Sample # 1	0.000	09:44	
Breath Volume	3.820	09:44	
Air Blank	0.000	09:45	
Air Blank	0.000	09:47	
Subject Sample # 2	0.000	09:47	
Breath Volume	2.519	09:47	
Air Blank	0.000	09:48	
Control Test	0.086	09:48	Control Outside Tolerance
Air Blank	0.000	09:49	

Data Download Date/Time: 07/27/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAYA,ZACHARY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 07/11/2023 09:59:59

Diagnostics Check	OK	09:53
Air Blank	0.000	09:54
Control Test	0.080	09:54
Air Blank	0.000	09:55
Subject Sample # 1	0.000	09:55
Breath Volume	3.957	09:55
Air Blank	0.000	09:56
Air Blank	0.000	09:58
Subject Sample # 2	0.000	09:58
Breath Volume	3.210	09:58
Air Blank	0.000	09:59
Control Test	0.085	09:59
Air Blank	0.000	09:59
Diagnostics Check	OK	09:59

Data Download Date/Time: 07/27/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS, BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ALCALA VEGA

DATE/TIME OF TEST: 07/16/2023 01:17:08

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.082	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.173	01:12
Breath Volume	2.648	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:14
Subject Sample # 2	0.176	01:15
Breath Volume	3.695	01:15
Air Blank	0.000	01:16
Control Test	0.085	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 07/27/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KOUCHNIR,EUGENE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 07/17/2023 21:11:51

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.082	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.163	21:07
Breath Volume	2.628	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.171	21:10
Breath Volume	2.203	21:10
Air Blank	0.000	21:10
Control Test	0.083	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 07/27/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDENAS, ADAM,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CARDENAS, A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:53

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 07/19/2023 10:32:14

Diagnostics Check	OK	10:26
Air Blank	0.000	10:26
Control Test	0.082	10:27
Air Blank	0.000	10:27
Subject Sample # 1	0.000	10:28
Breath Volume	3.562	10:28
Air Blank	0.000	10:28
Air Blank	0.000	10:30
Subject Sample # 2	0.000	10:30
Breath Volume	1.371	10:30
Air Blank	0.000	10:31
Control Test	0.083	10:31
Air Blank	0.000	10:32
Diagnostics Check	OK	10:32

Data Download Date/Time: 07/27/2023 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAYA,ZACHARY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 07/24/2023 07:20:19

Diagnostics Check	OK	07:14
Air Blank	0.000	07:14
Control Test	0.083	07:15
Air Blank	0.000	07:15
Subject Sample # 1	0.000	07:16
Breath Volume	2.519	07:16
Air Blank	0.000	07:16
Air Blank	0.000	07:18
Subject Sample # 2	0.000	07:18
Breath Volume	2.562	07:18
Air Blank	0.000	07:19
Control Test	0.084	07:19
Air Blank	0.000	07:20
Diagnostics Check	OK	07:20

Data Download Date/Time: 07/27/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 03:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MIMS

DATE/TIME OF TEST: 07/02/2023 04:17:54

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.105	04:13
Breath Volume	1.425	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.104	04:16
Breath Volume	1.519	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 21:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: GUTHRIE

DATE/TIME OF TEST: 07/06/2023 21:58:34

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.078	21:47	
Air Blank	0.000	21:47	
Subject Sample	# 1 0.280	21:51	Volume Not Met
Breath Volume	0.152	21:51	
Air Blank	0.000	21:52	
Air Blank	0.000	21:53	
Subject Sample	# 2 0.284	21:56	Volume Not Met
Breath Volume	0.351	21:56	
Air Blank	0.000	21:57	
Control Test	0.073	21:57	Control Outside Tolerance
Air Blank	0.000	21:58	

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COTTON,PIERCE,A

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COTTON,PIERCE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 03:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: UPCHURCH

DATE/TIME OF TEST: 07/07/2023 03:45:29

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:39	
Control Test	0.080	03:39	
Air Blank	0.000	03:39	
Subject Sample	# 1 0.267	03:40	
Breath Volume	2.410	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 2 0.273	03:43	
Breath Volume	2.289	03:43	
Air Blank	0.000	03:44	
Control Test	0.079	03:44	
Air Blank		03:45	RFI Detect
Air Blank	0.000	03:45	

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY-IVERSEN,MIYA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAY-IVERSEN,MIYA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: MARACLE

DATE/TIME OF TEST: 07/08/2023 18:41:15

Diagnostics Check	OK	18:31
Air Blank	0.000	18:32
Control Test	0.081	18:32
Air Blank	0.000	18:33
Subject Sample # 1	0.277	18:33
Breath Volume	3.144	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.254	18:36
Breath Volume	1.921	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 3	0.271	18:39
Breath Volume	3.261	18:39
Air Blank	0.000	18:40
Control Test	0.080	18:40
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY-IVERSEN,MIYA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAY-IVERSEN,MIYA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: GOLLAKNER

DATE/TIME OF TEST: 07/08/2023 21:19:37

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.068	21:19	Control Outside Tolerance
Air Blank	0.000	21:19	

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY-IVERSEN,MIYA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAY-IVERSEN,MIYA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: GOLLAKNER

DATE/TIME OF TEST: 07/08/2023 21:22:40

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:21	
Control Test	0.071	21:22	Control Outside Tolerance
Air Blank	0.000	21:22	

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY-IVERSEN,MIYA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAY-IVERSEN,MIYA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: GOLLAKNER

DATE/TIME OF TEST: 07/08/2023 21:25:53

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:24	
Control Test	0.072	21:25	Control Outside Tolerance
Air Blank	0.000	21:25	

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY-IVERSEN,MIYA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAY-IVERSEN,MIYA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: GOLLAKNER

DATE/TIME OF TEST: 07/08/2023 21:33:14

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.075	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.000	21:28
Breath Volume	2.203	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.000	21:31
Breath Volume	2.656	21:31
Air Blank	0.000	21:32
Control Test	0.078	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUFOUR,NOAH,I

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: DUFOUR,NOAH

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 02:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: G230102974080

DATE/TIME OF TEST: 07/09/2023 03:18:58

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.086	03:14
Breath Volume	1.292	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.082	03:17
Breath Volume	1.363	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: BECKER,JONATHAN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 03:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/16/2023 03:51:52

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.134	03:47
Breath Volume	2.718	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.138	03:50
Breath Volume	2.894	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WERNER,D
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/20/2023 01:55:13

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.082	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.168	01:50
Breath Volume	3.019	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.161	01:53
Breath Volume	2.589	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BRAY,K
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 07/20/2023 23:39:09

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.250	23:33
Breath Volume	2.785	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.253	23:37
Breath Volume	2.156	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASWELL,AUSTIN,J

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: CUNNINGHAM,BENJAMIN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 04:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: MUHAMMAD

DATE/TIME OF TEST: 07/23/2023 05:02:33

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.081	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:01	RFI Detect
Breath Volume	0.000	05:01	
Air Blank	0.000	05:01	
Control Test	0.081	05:01	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASWELL,AUSTIN,J

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: CUNNINGHAM,BENJAMIN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 04:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: MUHAMMAD

DATE/TIME OF TEST: 07/23/2023 05:18:27

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:05	
Control Test	0.081	05:05	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:09	No Sample Provided
Breath Volume	0.000	05:09	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample	# 2	0.226	05:12
Breath Volume	1.585	05:12	
Air Blank	0.000	05:12	
Air Blank	0.000	05:14	
Subject Sample	# 3	05:16	Subject Test Refused
Breath Volume	0.843	05:16	
Air Blank	0.000	05:17	
Control Test	0.079	05:17	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:18	

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: KNIFFEN

DATE/TIME OF TEST: 07/23/2023 16:21:34

Diagnostics Check	OK	16:10	
Air Blank	0.000	16:10	
Control Test	0.082	16:10	
Air Blank	0.000	16:11	
Subject Sample	# 1	16:14	No Sample Provided
Breath Volume	0.000	16:14	
Air Blank	0.000	16:15	
Air Blank	0.000	16:17	
Subject Sample	# 2	16:20	No Sample Provided
Breath Volume	0.000	16:20	
Air Blank	0.000	16:20	
Control Test	0.082	16:20	
Air Blank	0.000	16:21	
Diagnostics Check	OK	16:21	

Data Download Date/Time: 07/31/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PINO,RICHARD,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: DELL,CHRISTOPHER

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 21:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SAMUEL

DATE/TIME OF TEST: 07/25/2023 22:01:44

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.081	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.294	21:57
Breath Volume	2.093	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.283	22:00
Breath Volume	1.816	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 07/31/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COTTON, PIERCE, A

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: CUNNINGHAM, BENJAMIN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 18:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: MORTENSEN

DATE/TIME OF TEST: 07/27/2023 18:34:06

Diagnostics Check	OK	18:33	
Air Blank	0.000	18:33	
		18:33	RFI Detect
Air Blank		18:34	RFI Detect

Data Download Date/Time: 07/31/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1408275
BREATH TEST OPERATOR: COTTON, PIERCE, A	DRY GAS EXPIRY DATE: 11/19/2023
BREATH TEST OPERATOR AGENCY: Pensacola PD	VIOLATION CODE: DUI
ARRESTING OFFICER: CUNNINGHAM, BENJAMIN	ARREST TIME: 18:00
ARRESTING OFFICER AGENCY: Pensacola PD	GENDER: Male
LAST AGENCY INSPECTION: 06/27/2023	OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: MORTENSEN	DATE/TIME OF TEST: 07/27/2023	18:44:44
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.080	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.204	18:38
Breath Volume	2.273	18:38
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.206	18:43
Breath Volume	1.597	18:43
Air Blank	0.000	18:43
Control Test	0.080	18:44
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 07/31/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: WERNER,DANIEL

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 03:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: ALFT

DATE/TIME OF TEST: 07/29/2023 04:18:32

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.213	04:13
Breath Volume	3.214	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.212	04:16
Breath Volume	2.851	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 07/31/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: WERNER,DANIEL

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1408275

DRY GAS EXPIRY DATE: 11/19/2023

VIOLATION CODE: DUI

ARREST TIME: 03:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: ALFT

DATE/TIME OF TEST: 07/29/2023 04:43:10

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.209	04:38
Breath Volume	2.843	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:40
Subject Sample # 2	0.208	04:41
Breath Volume	2.980	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 07/31/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: MICHAELS

DATE/TIME OF TEST: 07/01/2023 02:39:19

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.109	02:34
Breath Volume	1.132	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.111	02:37
Breath Volume	1.601	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 07/26/2023 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: MCCANDLESS,MATTHEW

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 22:04

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BEKAVAC

DATE/TIME OF TEST: 07/06/2023 22:42:13

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:36	
Control Test	0.077	22:36	
Air Blank	0.000	22:37	
Subject Sample	# 1 0.171	22:37	
Breath Volume	2.082	22:37	
Air Blank	0.000	22:38	
Air Blank	0.000	22:40	
Subject Sample	# 2 0.172	22:40	
Breath Volume	2.304	22:40	
Air Blank	0.000	22:41	
Control Test	0.083	22:41	Interferent Detect
Air Blank	0.000	22:42	

Data Download Date/Time: 07/26/2023 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BEKAVAC

DATE/TIME OF TEST: 07/06/2023 22:44:47

Diagnostics Check	Fail	22:44	Internal Printer Test
Air Blank	0.000	22:44	

Data Download Date/Time: 07/26/2023 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BEKAVAC

DATE/TIME OF TEST: 07/06/2023 22:46:36

Diagnostics Check	Fail	22:46	Internal Printer Test
Air Blank	0.000	22:46	

Data Download Date/Time: 07/26/2023 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: MCCANDLESS,MATTHEW

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 22:04

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BEKAVAC

DATE/TIME OF TEST: 07/06/2023 22:54:46

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.077	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.169	22:50
Breath Volume	1.878	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.175	22:53
Breath Volume	2.800	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 07/26/2023 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: RAMANO SMITH

DATE/TIME OF TEST: 07/07/2023 00:11:10

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.161	00:05
Breath Volume	1.800	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:07
Subject Sample # 2	0.164	00:09
Breath Volume	1.402	00:09
Air Blank	0.000	00:10
Control Test	0.082	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 07/26/2023 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: TAMBURO

DATE/TIME OF TEST: 07/08/2023 01:35:13

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.206	01:30
Breath Volume	2.714	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.206	01:33
Breath Volume	2.937	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 07/26/2023 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 07/22/2023 01:51:46

Diagnostics Check	Fail	01:51	Internal Printer Test
Air Blank	0.000	01:51	

Data Download Date/Time: 07/26/2023 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 01:46

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 07/22/2023 02:09:12

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:53	
Control Test	0.077	01:53	
Air Blank	0.000	01:54	
Subject Sample	# 1 0.188	01:57	Volume Not Met
Breath Volume	0.000	01:57	
Air Blank	0.000	01:58	
Air Blank	0.000	02:00	
Subject Sample	# 2 0.208	02:03	
Breath Volume	1.328	02:03	
Air Blank	0.000	02:03	
Air Blank	0.000	02:05	
Subject Sample	# 3 0.200	02:07	
Breath Volume	1.488	02:07	
Air Blank	0.000	02:08	
Control Test	0.080	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 07/26/2023 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,JASON,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: SHARRAH

DATE/TIME OF TEST: 07/27/2023 22:13:04

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.076	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.154	22:08
Breath Volume	2.179	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.152	22:11
Breath Volume	2.125	22:11
Air Blank	0.000	22:12
Control Test	0.079	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:13

Data Download Date/Time: 08/23/2023 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARRERA,ALEXANDER,R

BREATH TEST OPERATOR AGENCY: St. Augustine PD

ARRESTING OFFICER: BARRERA,ALEXANDER

ARRESTING OFFICER AGENCY: St. Augustine PD

LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 01:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: TYER

DATE/TIME OF TEST: 07/29/2023 01:34:59

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.129	01:30
Breath Volume	3.554	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.135	01:33
Breath Volume	3.785	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 08/23/2023 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,JASON,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 02:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: CANELOS

DATE/TIME OF TEST: 07/29/2023 02:35:20

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.076	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.106	02:30
Breath Volume	1.773	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample # 2	0.105	02:33
Breath Volume	2.394	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 08/23/2023 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GATCHELL,T

ARRESTING OFFICER AGENCY: St. Augustine PD

LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 01:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 07/30/2023 02:24:20

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.076	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.233	02:19
Breath Volume	1.437	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.232	02:22
Breath Volume	1.484	02:22
Air Blank	0.000	02:23
Control Test	0.077	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 08/23/2023 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: ELLIS,ALEXANDER,	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ELLIS,ALEXANDER	ARREST TIME: 00:30
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/28/2023	OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SANSBURY	DATE/TIME OF TEST: 07/07/2023 01:05:21
-------------------------------	--

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.077	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.000	01:01
Breath Volume	2.625	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.000	01:03
Breath Volume	3.089	01:03
Air Blank	0.000	01:04
Control Test	0.076	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 07/26/2023 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: MCCANDLESS,MATTHEW

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211543

DRY GAS EXPIRY DATE: 01/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: BLANKENSHIP

DATE/TIME OF TEST: 07/11/2023 01:29:51

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.075	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1 0.110	01:24	
Breath Volume	2.457	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 2 0.109	01:27	
Breath Volume	2.714	01:27	
Air Blank	0.000	01:28	
Control Test	0.076	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	Fail	01:29	Internal Printer Test
Air Blank	0.000	01:29	

Data Download Date/Time: 07/26/2023 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/14/2023 01:31:21

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.077	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.123	01:25
Breath Volume	1.140	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.123	01:29
Breath Volume	1.269	01:29
Air Blank	0.000	01:30
Control Test	0.077	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 07/26/2023 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: ELLIS,ALEXANDER,	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ELLIS,ALEXANDER	ARREST TIME: 03:10
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/28/2023	OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: WRIGHT	DATE/TIME OF TEST: 07/14/2023 03:23:42
-----------------------------	--

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.111	03:19
Breath Volume	2.511	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.111	03:22
Breath Volume	2.050	03:22
Air Blank	0.000	03:22
Control Test	0.076	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 07/26/2023 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: YOCIUS

DATE/TIME OF TEST: 07/14/2023 16:51:36

Diagnostics Check	OK	16:45
Air Blank	0.000	16:45
Control Test	0.080	16:46
Air Blank	0.000	16:46
Subject Sample # 1	0.000	16:47
Breath Volume	2.320	16:47
Air Blank	0.000	16:47
Air Blank	0.000	16:49
Subject Sample # 2	0.000	16:50
Breath Volume	2.078	16:50
Air Blank	0.000	16:50
Control Test	0.076	16:51
Air Blank	0.000	16:51
Diagnostics Check	OK	16:51

Data Download Date/Time: 07/26/2023 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: ANDREU

DATE/TIME OF TEST: 07/28/2023 22:38:05

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.079	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.150	22:33
Breath Volume	2.343	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.149	22:36
Breath Volume	2.304	22:36
Air Blank	0.000	22:37
Control Test	0.078	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 08/23/2023 23:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORREA,JONATHAN,E

BREATH TEST OPERATOR AGENCY: St. Augustine PD

ARRESTING OFFICER: CORREA,JONATHAN

ARRESTING OFFICER AGENCY: St. Augustine PD

LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211543

DRY GAS EXPIRY DATE: 01/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: OTTILO

DATE/TIME OF TEST: 07/28/2023 22:46:35

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.078	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.175	22:42
Breath Volume	2.457	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.169	22:45
Breath Volume	1.546	22:45
Air Blank	0.000	22:45
Control Test	0.078	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 08/23/2023 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CANAAN,AUSTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 07/30/2023 23:00:23

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.079	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.196	22:55
Breath Volume	1.550	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.194	22:58
Breath Volume	1.160	22:58
Air Blank	0.000	22:59
Control Test	0.078	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 08/23/2023 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ETTEL,M
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: RIDDLE

DATE/TIME OF TEST: 07/31/2023 21:33:33

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.079	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.000	21:29
Breath Volume	1.144	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.000	21:32
Breath Volume	1.593	21:32
Air Blank	0.000	21:32
Control Test	0.079	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 08/23/2023 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: QUIROGA,DIEGO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: OLMOS

DATE/TIME OF TEST: 07/01/2023 00:54:58

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.080	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:50	No Sample Provided
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 2	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:54	
Control Test	0.079	00:54	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 07/01/2023 19:40:49

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.079	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.156	19:35
Breath Volume	4.132	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.153	19:39
Breath Volume	3.335	19:39
Air Blank	0.000	19:39
Control Test	0.079	19:40
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: RAYMOUNDO CEDILLO

DATE/TIME OF TEST: 07/01/2023 21:37:43

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:29	
Control Test	0.080	21:29	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:33	No Sample Provided
Breath Volume	0.000	21:33	
Air Blank	0.000	21:34	
Air Blank	0.000	21:36	
Subject Sample	# 2	21:36	Subject Test Refused
Breath Volume	0.000	21:36	
Air Blank	0.000	21:36	
Control Test	0.079	21:37	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 07/12/2023 04:00:16

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PEDRO CRISTOBAL

DATE/TIME OF TEST: 07/15/2023 02:07:29

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.081	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:03	No Sample Provided
Breath Volume	0.000	02:03	
Air Blank	0.000	02:03	
Air Blank	0.000	02:05	
Subject Sample	# 2	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: DIAZ GARCIA

DATE/TIME OF TEST: 07/15/2023 19:30:34

Diagnostics Check	OK	19:24	
Air Blank	0.000	19:25	
Control Test	0.080	19:25	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:28	Subject Test Refused
Breath Volume	0.054	19:28	
Air Blank	0.000	19:29	
Control Test	0.079	19:29	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CLAYBOURNE,R
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: JONES,CLAYBOURNE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GERMAIN

DATE/TIME OF TEST: 07/20/2023 03:17:31

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.081	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANECA,DANILO,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,CARLOS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: DOOMINGOCLAUDIO

DATE/TIME OF TEST: 07/20/2023 23:23:07

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.080	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1 0.138	23:21	Volume Not Met
Breath Volume	0.875	23:21	
Air Blank		23:22	Purge Fail
Air Blank	0.000	23:23	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANECA,DANILO,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,CARLOS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: DOMINGOCLAUDIO

DATE/TIME OF TEST: 07/20/2023 23:30:39

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:28	
Control Test	0.076	23:28	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.078	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGOSTO,VICTOR,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RAYMUNDO HERNANDEZ

DATE/TIME OF TEST: 07/23/2023 00:15:58

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.080	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:11	No Sample Provided
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample	# 2	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROCA,LUIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/23/2023 16:34:52

Diagnostics Check	OK	16:31	
Air Blank	0.000	16:31	
Control Test	0.080	16:31	
Air Blank	0.000	16:32	
Subject Sample	# 1	16:33	Subject Test Refused
Breath Volume	0.000	16:33	
Air Blank	0.000	16:33	
Control Test	0.080	16:34	
Air Blank	0.000	16:34	
Diagnostics Check	OK	16:34	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: LAMPKIN,JARKKAR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: HERNANDEZ MENDEZ

DATE/TIME OF TEST: 07/23/2023 19:44:28

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:41	
Control Test	0.080	19:41	
Air Blank	0.000	19:42	
Subject Sample	# 1	19:43	Subject Test Refused
Breath Volume	0.000	19:43	
Air Blank	0.000	19:43	
Control Test	0.081	19:43	
Air Blank	0.000	19:44	
Diagnostics Check	OK	19:44	

Data Download Date/Time: 07/26/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOLDBERG,JOSHUA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: CHANEY

DATE/TIME OF TEST: 07/06/2023 06:24:49

Diagnostics Check	OK	06:17
Air Blank	0.000	06:18
Control Test	0.081	06:18
Air Blank	0.000	06:19
Subject Sample # 1	0.327	06:19
Breath Volume	1.191	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample # 2	0.319	06:22
Breath Volume	1.398	06:22
Air Blank	0.000	06:23
Control Test	0.079	06:24
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PAYNE,BRIANNA
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: QUILES

DATE/TIME OF TEST: 07/08/2023 19:40:03

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.081	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.168	19:35
Breath Volume	1.613	19:35
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.168	19:38
Breath Volume	1.867	19:38
Air Blank	0.000	19:39
Control Test	0.080	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:40

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHARNECO,TAYLOR
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: VINACCIA

DATE/TIME OF TEST: 07/09/2023 01:07:01

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.191	01:02
Breath Volume	1.304	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:04
Subject Sample # 2	0.183	01:05
Breath Volume	1.703	01:05
Air Blank	0.000	01:06
Control Test	0.078	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:07

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VALDES,VICTOR
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LOPEZ SANCHEZ

DATE/TIME OF TEST: 07/13/2023 01:06:12

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.179	01:00
Breath Volume	2.902	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.167	01:04
Breath Volume	3.222	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MARCONI

DATE/TIME OF TEST: 07/15/2023 02:30:58

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.000	02:26
Breath Volume	1.593	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.000	02:29
Breath Volume	1.554	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARTINEZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 07/15/2023 03:49:17

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.167	03:44
Breath Volume	2.328	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.158	03:47
Breath Volume	1.914	03:47
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 07/15/2023 23:14:14

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.081	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.350	23:09
Breath Volume	1.437	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample # 2	0.364	23:12
Breath Volume	1.351	23:12
Air Blank	0.000	23:13
Control Test	0.077	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: JIMENEZ NAVARRO

DATE/TIME OF TEST: 07/16/2023 01:42:36

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.175	01:36
Breath Volume	2.414	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.167	01:40
Breath Volume	1.632	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FERNANDEZ,JONATHAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 07/18/2023 21:46:43

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.081	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.194	21:40
Breath Volume	1.300	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:42
Subject Sample # 2	0.191	21:44
Breath Volume	1.257	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SKELLETT

DATE/TIME OF TEST: 07/24/2023 23:22:36

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.162	23:17
Breath Volume	1.300	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.179	23:20
Breath Volume	2.246	23:20
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: YASKO,JOE
ARRESTING OFFICER AGENCY: Lauderdale PD
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:23

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 07/25/2023 07:52:38

Diagnostics Check	OK	07:45
Air Blank	0.000	07:46
Control Test	0.082	07:46
Air Blank	0.000	07:47
Subject Sample # 1	0.119	07:47
Breath Volume	3.148	07:47
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.122	07:50
Breath Volume	3.082	07:50
Air Blank	0.000	07:51
Control Test	0.080	07:52
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ABRAM,JUSTIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 07/29/2023 15:53:46

Diagnostics Check	OK	15:47
Air Blank	0.000	15:47
Control Test	0.082	15:48
Air Blank	0.000	15:48
Subject Sample # 1	0.000	15:49
Breath Volume	2.726	15:49
Air Blank	0.000	15:49
Air Blank	0.000	15:51
Subject Sample # 2	0.000	15:52
Breath Volume	2.671	15:52
Air Blank	0.000	15:52
Control Test	0.080	15:53
Air Blank	0.000	15:53
Diagnostics Check	OK	15:53

Data Download Date/Time: 08/02/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CALEB,A
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: CRAVEN,J
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: CHROME

DATE/TIME OF TEST: 07/09/2023 06:13:43

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.078	06:07
Air Blank	0.000	06:08
Subject Sample # 1	0.205	06:08
Breath Volume	2.269	06:08
Air Blank	0.000	06:09
Air Blank	0.000	06:11
Subject Sample # 2	0.209	06:11
Breath Volume	1.347	06:11
Air Blank	0.000	06:12
Control Test	0.076	06:13
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 07/22/2023 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRELAND,LELAND,J
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: MEDLOCK,G
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: IRELAND

DATE/TIME OF TEST: 07/20/2023 01:07:17

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.138	01:02
Breath Volume	3.406	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.126	01:05
Breath Volume	1.421	01:05
Air Blank	0.000	01:06
Control Test	0.078	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 07/22/2023 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALLEN,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BARLEY

DATE/TIME OF TEST: 07/04/2023 01:38:21

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.078	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 07/29/2023 17:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: KNOUS

DATE/TIME OF TEST: 07/22/2023 05:21:55

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:19	
Control Test	0.078	05:19	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:21	
Control Test	0.080	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 07/29/2023 17:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LEEP

DATE/TIME OF TEST: 07/23/2023 01:40:58

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.078	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1 0.109	01:39	
Breath Volume	1.847	01:39	
Air Blank		01:39	RFI Detect
Air Blank	0.000	01:40	
Control Test	0.079	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 07/29/2023 17:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LEEP

DATE/TIME OF TEST: 07/23/2023 01:50:08

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.104	01:45
Breath Volume	1.558	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:47
Subject Sample # 2	0.111	01:48
Breath Volume	1.964	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/29/2023 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: SMITH,MATTHEW,M	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: MUSIAL,CHRISTOPHER	ARREST TIME: 09:13
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 07/07/2023	OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: KOVACS	DATE/TIME OF TEST: 07/21/2023	10:17:06
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	10:06	
Air Blank	0.000	10:06	
Control Test	0.080	10:07	
Air Blank	0.000	10:07	
Subject Sample	# 1 0.359	10:08	
Breath Volume	2.621	10:08	
Air Blank	0.000	10:09	
Air Blank	0.000	10:10	
Subject Sample	# 2 0.332	10:12	
Breath Volume	2.156	10:12	
Air Blank	0.000	10:12	
Air Blank	0.000	10:14	
Subject Sample	# 3 0.353	10:15	
Breath Volume	2.597	10:15	
Air Blank	0.000	10:15	
Control Test	0.080	10:16	
Air Blank	0.000	10:16	
Diagnostics Check	Fail	10:16	Internal Printer Test
Air Blank		10:17	RFI Detect

Data Download Date/Time: 08/19/2023 09:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,MATTHEW,M

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: MUSIAL,CHRISTOPHER

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 07/07/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 09:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: KOVACS

DATE/TIME OF TEST: 07/21/2023 10:32:37

Diagnostics Check	OK	10:21	
Air Blank	0.000	10:21	
Control Test	0.080	10:21	
Air Blank	0.000	10:22	
Subject Sample	# 1 0.337	10:23	
Breath Volume	1.972	10:23	
Air Blank	0.000	10:23	
Air Blank	0.000	10:25	
Subject Sample	# 2 0.286	10:28	Volume Not Met
Breath Volume	0.339	10:28	
Air Blank	0.000	10:29	
Air Blank	0.000	10:31	
Subject Sample	# 3 0.297	10:31	
Breath Volume	1.378	10:31	
Air Blank		10:31	Improper Sample
Air Blank	0.000	10:32	

Data Download Date/Time: 08/19/2023 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGGINS,ALAN,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 07/08/2023 23:10:23

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.077	23:06	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:09	Subject Test Refused
Breath Volume	0.015	23:09	
Air Blank		23:09	Ambient Fail
Air Blank	0.000	23:10	

Data Download Date/Time: 08/01/2023 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: DIAZ DIAZ

DATE/TIME OF TEST: 07/23/2023 03:52:06

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.075	03:45	
Air Blank	0.000	03:46	
Subject Sample # 1	0.182	03:47	
Breath Volume	1.566	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:49	
Subject Sample # 2	0.202	03:50	
Breath Volume	1.957	03:50	
Air Blank	0.000	03:51	
Control Test	0.074	03:51	Control Outside Tolerance
Air Blank	0.000	03:52	

Data Download Date/Time: 08/01/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: DIAZ DIAZ

DATE/TIME OF TEST: 07/23/2023 04:03:09

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.076	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.204	03:58
Breath Volume	1.566	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.207	04:01
Breath Volume	2.437	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 08/01/2023 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: PEREZ ROBLERO

DATE/TIME OF TEST: 07/28/2023 00:13:07

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.075	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1 0.170	00:08	
Breath Volume	2.875	00:08	
Air Blank	0.000	00:09	
Air Blank	0.000	00:10	
Subject Sample	# 2 0.171	00:11	
Breath Volume	1.914	00:11	
Air Blank	0.000	00:12	
Control Test	0.073	00:12	Control Outside Tolerance
Air Blank	0.000	00:13	

Data Download Date/Time: 08/01/2023 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY',
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: PEREZ ROBLERO

DATE/TIME OF TEST: 07/28/2023 00:21:16

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.076	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.172	00:16
Breath Volume	2.402	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:18
Subject Sample # 2	0.169	00:19
Breath Volume	1.835	00:19
Air Blank	0.000	00:20
Control Test	0.077	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 08/01/2023 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BROWNING

DATE/TIME OF TEST: 07/11/2023 00:45:02

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.154	00:40
Breath Volume	1.562	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.153	00:43
Breath Volume	1.917	00:43
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:45

Data Download Date/Time: 08/01/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CEDREZ JR

DATE/TIME OF TEST: 07/15/2023 03:36:10

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.125	03:31
Breath Volume	4.132	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.122	03:34
Breath Volume	2.917	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 08/01/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: CHIPANA RICARDI

DATE/TIME OF TEST: 07/16/2023 02:16:24

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.072	02:11
Breath Volume	1.476	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.075	02:14
Breath Volume	2.046	02:14
Air Blank	0.000	02:15
Control Test	0.077	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 08/01/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/21/2023 02:19:16

Diagnostics Check	OK	02:18	
Air Blank		02:18	RFI Detect
Air Blank	0.000	02:19	

Data Download Date/Time: 08/01/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/21/2023 02:27:13

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.222	02:22
Breath Volume	2.113	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.224	02:25
Breath Volume	1.691	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 08/01/2023 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 07/29/2023 22:53:53

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.097	22:49
Breath Volume	2.519	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.095	22:52
Breath Volume	2.488	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 08/01/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/30/2023 06:10:39

Diagnostics Check	OK	06:04
Air Blank	0.000	06:04
Control Test	0.079	06:04
Air Blank	0.000	06:05
Subject Sample # 1	0.246	06:06
Breath Volume	3.320	06:06
Air Blank	0.000	06:06
Air Blank	0.000	06:08
Subject Sample # 2	0.233	06:09
Breath Volume	2.925	06:09
Air Blank	0.000	06:09
Control Test	0.077	06:10
Air Blank	0.000	06:10
Diagnostics Check	OK	06:10

Data Download Date/Time: 08/01/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: VOLERO CARMENATE

DATE/TIME OF TEST: 07/31/2023 03:11:52

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.195	03:07
Breath Volume	1.828	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.186	03:10
Breath Volume	1.394	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 08/01/2023 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BERG,TIM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: PACI

DATE/TIME OF TEST: 07/01/2023 19:36:44

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.081	19:30
Air Blank	0.000	19:31
Subject Sample # 1	0.124	19:32
Breath Volume	2.128	19:32
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.128	19:35
Breath Volume	2.539	19:35
Air Blank	0.000	19:35
Control Test	0.081	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 07/05/2023 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BERG,TIM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: BUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 07/03/2023 16:09:12

Diagnostics Check	OK	16:01
Air Blank	0.000	16:02
Control Test	0.080	16:02
Air Blank	0.000	16:03
Subject Sample # 1	0.083	16:04
Breath Volume	2.503	16:04
Air Blank	0.000	16:05
Air Blank	0.000	16:07
Subject Sample # 2	0.081	16:07
Breath Volume	2.542	16:07
Air Blank	0.000	16:08
Control Test	0.079	16:08
Air Blank	0.000	16:09
Diagnostics Check	OK	16:09

Data Download Date/Time: 07/05/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WILKINS,MALACHI
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 07/04/2023 17:59:50

Diagnostics Check	OK	17:53
Air Blank	0.000	17:53
Control Test	0.078	17:54
Air Blank	0.000	17:54
Subject Sample # 1	0.043	17:55
Breath Volume	3.835	17:55
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.043	17:58
Breath Volume	3.531	17:58
Air Blank	0.000	17:58
Control Test	0.080	17:59
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

Data Download Date/Time: 07/05/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: RIVERA GUERRA

DATE/TIME OF TEST: 07/09/2023 02:06:33

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.082	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.127	02:01
Breath Volume	1.296	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.126	02:04
Breath Volume	2.242	02:04
Air Blank	0.000	02:05
Control Test	0.082	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 08/01/2023 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MAGANA JACOBO

DATE/TIME OF TEST: 07/09/2023 06:01:00

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.080	05:54
Air Blank	0.000	05:55
Subject Sample # 1	0.191	05:56
Breath Volume	3.339	05:56
Air Blank	0.000	05:57
Air Blank	0.000	05:58
Subject Sample # 2	0.191	05:59
Breath Volume	2.324	05:59
Air Blank	0.000	06:00
Control Test	0.081	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:01

Data Download Date/Time: 08/01/2023 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: GAYNOR

DATE/TIME OF TEST: 07/14/2023 04:33:13

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.082	04:24	
Air Blank	0.000	04:24	
Subject Sample # 1	0.165	04:27	
Breath Volume	1.769	04:27	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample # 2	0.152	04:31	
Breath Volume	2.433	04:31	
Air Blank	0.000	04:32	
Control Test	0.074	04:32	Interferent Detect
Air Blank	0.000	04:33	

Data Download Date/Time: 08/01/2023 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: GAYNOR

DATE/TIME OF TEST: 07/14/2023 04:42:55

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.082	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.156	04:38
Breath Volume	2.652	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.156	04:41
Breath Volume	2.683	04:41
Air Blank	0.000	04:42
Control Test	0.083	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 08/01/2023 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: CLARKSTON

DATE/TIME OF TEST: 07/19/2023 01:00:48

Diagnostics Check	Fail	01:00	Internal Printer Test
Air Blank	0.000	01:00	

Data Download Date/Time: 08/01/2023 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/22/2023 02:45:58

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.081	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.108	02:41
Breath Volume	2.679	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.104	02:44
Breath Volume	1.949	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 08/01/2023 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BEDOYA

DATE/TIME OF TEST: 07/27/2023 03:30:19

Diagnostics Check	Fail	03:29	Internal Printer Test
Air Blank		03:30	Ambient Fail

Data Download Date/Time: 08/01/2023 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BEDOYA

DATE/TIME OF TEST: 07/27/2023 03:33:05

Diagnostics Check	OK	03:32	
Air Blank		03:32	Ambient Fail
Air Blank	0.000	03:33	

Data Download Date/Time: 08/01/2023 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBE,BRIAN,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: RILEY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: PENA GONZALEZ

DATE/TIME OF TEST: 07/03/2023 18:40:38

Diagnostics Check	OK	18:37	
Air Blank	0.000	18:38	
Control Test	0.080	18:38	
Air Blank	0.000	18:39	
Subject Sample	# 1	18:39	Subject Test Refused
Breath Volume	0.000	18:39	
Air Blank	0.000	18:39	
Control Test	0.079	18:40	
Air Blank	0.000	18:40	
Diagnostics Check	OK	18:40	

Data Download Date/Time: 07/28/2023 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JACOB,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HYLTON,ROBERT
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 06/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 07/09/2023 23:00:19

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.180	22:55
Breath Volume	4.156	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:57
Subject Sample # 2	0.182	22:58
Breath Volume	3.855	22:58
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 07/28/2023 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRALSON,SAMUEL,E
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: WALKER,KYLE
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 06/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: SHOUP

DATE/TIME OF TEST: 07/15/2023 16:01:54

Diagnostics Check	OK	15:55
Air Blank	0.000	15:55
Control Test	0.080	15:56
Air Blank	0.000	15:56
Subject Sample # 1	0.273	15:57
Breath Volume	2.082	15:57
Air Blank	0.000	15:58
Air Blank	0.000	15:59
Subject Sample # 2	0.275	16:00
Breath Volume	1.316	16:00
Air Blank	0.000	16:00
Control Test	0.076	16:01
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

Data Download Date/Time: 07/28/2023 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JACOB,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: FRANKLIN,JACOB
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/23/2023 01:41:52

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.176	01:37
Breath Volume	3.351	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.179	01:40
Breath Volume	2.246	01:40
Air Blank	0.000	01:40
Control Test	0.077	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 07/28/2023 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEDESCH,LAURA,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: DIEDESCH,LAURA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MAUK

DATE/TIME OF TEST: 07/26/2023 01:52:06

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.167	01:47
Breath Volume	2.851	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.160	01:50
Breath Volume	1.683	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 07/28/2023 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFFMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: CRISERA,MATTHEW
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 07/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/29/2023 21:31:09

Diagnostics Check	OK	21:24
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.206	21:26
Breath Volume	1.707	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.204	21:29
Breath Volume	1.339	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 08/25/2023 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILKINS,MALACHI

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: BUI

ARREST TIME: 15:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/02/2023 16:35:10

Diagnostics Check	OK	16:33	
Air Blank	0.000	16:34	
Control Test	0.073	16:34	Control Outside Tolerance
Air Blank	0.000	16:35	

Data Download Date/Time: 07/28/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILKINS,MALACHI

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: BUI

ARREST TIME: 15:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/02/2023 16:38:17

Diagnostics Check	OK	16:37	
Air Blank		16:37	RFI Detect
Air Blank	0.000	16:38	

Data Download Date/Time: 07/28/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILKINS,MALACHI

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: BUI

ARREST TIME: 15:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/02/2023 16:41:28

Diagnostics Check

Fail

16:40

Internal Printer Test

Air Blank

0.000

16:41

Data Download Date/Time: 07/28/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WILKINS,MALACHI
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: BUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/02/2023 16:53:07

Diagnostics Check	OK	16:45	
Air Blank	0.000	16:45	
Control Test	0.078	16:46	
Air Blank	0.000	16:46	
Subject Sample	# 1 0.059	16:47	
Breath Volume	1.730	16:47	
Air Blank	0.000	16:48	
Air Blank	0.000	16:50	
Subject Sample	# 2 0.056	16:50	
Breath Volume	2.656	16:50	
Air Blank	0.000	16:51	
Control Test	0.078	16:51	
Air Blank	0.000	16:52	
Diagnostics Check	Fail	16:52	Internal Printer Test
Air Blank	0.000	16:53	

Data Download Date/Time: 07/28/2023 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,CLINT,R

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILLIAMS,CLINT

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: BUI

ARREST TIME: 18:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: GUSTAFSON

DATE/TIME OF TEST: 07/02/2023 19:02:06

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.077	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.138	18:57
Breath Volume	3.554	18:57
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.138	19:00
Breath Volume	3.398	19:00
Air Blank	0.000	19:01
Control Test	0.077	19:01
Air Blank	0.000	19:01
Diagnostics Check	OK	19:02

Data Download Date/Time: 07/28/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, OFC B,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: RODRIGUEZ, OFC B
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: BUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: SWAILES

DATE/TIME OF TEST: 07/03/2023 17:32:35

Diagnostics Check	OK	17:25
Air Blank	0.000	17:25
Control Test	0.076	17:26
Air Blank	0.000	17:26
Subject Sample # 1	0.008	17:27
Breath Volume	2.031	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.008	17:31
Breath Volume	1.968	17:31
Air Blank	0.000	17:31
Control Test	0.078	17:31
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

Data Download Date/Time: 07/28/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,CLINT,R

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: HOPPE,AUSTIN

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 19:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 07/04/2023 20:12:03

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.077	20:07	
Air Blank	0.000	20:07	
Subject Sample	# 1 0.199	20:08	
Breath Volume	3.445	20:08	
Air Blank	0.000	20:08	
Air Blank	0.000	20:10	
Subject Sample	# 2 0.194	20:11	
Breath Volume	3.007	20:11	
Air Blank		20:11	RFI Detect
Air Blank		20:12	RFI Detect

Data Download Date/Time: 07/28/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CLINT,R
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PETTIFER,JAMES
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MARTINS

DATE/TIME OF TEST: 07/04/2023 22:56:58

Diagnostics Check	Fail	22:56	Internal Printer Test
Air Blank	0.000	22:56	

Data Download Date/Time: 07/28/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORTH,DUANE,S

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: VANAS,EZEKIAL

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 06/20/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 211143

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 17:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: GALLIFORD

DATE/TIME OF TEST: 07/01/2023 17:50:20

Diagnostics Check	OK	17:43
Air Blank	0.000	17:44
Control Test	0.078	17:44
Air Blank	0.000	17:45
Subject Sample # 1	0.109	17:45
Breath Volume	2.328	17:45
Air Blank	0.000	17:46
Air Blank	0.000	17:48
Subject Sample # 2	0.112	17:48
Breath Volume	2.476	17:48
Air Blank	0.000	17:49
Control Test	0.078	17:49
Air Blank	0.000	17:50
Diagnostics Check	OK	17:50

Data Download Date/Time: 07/24/2023 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: JARRELL,CHRISTIAN,P	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: JARRELL,CHRISTIAN	ARREST TIME: 19:02
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/12/2023	OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: VANAERNAM	DATE/TIME OF TEST: 07/03/2023 00:35:25
--------------------------------	--

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.078	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.146	00:30
Breath Volume	5.605	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.136	00:33
Breath Volume	2.437	00:33
Air Blank	0.000	00:34
Control Test	0.077	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 07/24/2023 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: BASFORD, TRAVIS, L	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BASFORD, TRAVIS	ARREST TIME: 16:31
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/12/2023	OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: AYALA	DATE/TIME OF TEST: 07/03/2023 22:09:54
----------------------------	--

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:04	
Control Test	0.078	22:05	
Air Blank	0.000	22:05	
Subject Sample	# 1	22:08	No Sample Provided
Breath Volume	0.000	22:08	
Air Blank	0.000	22:09	
Air Blank	0.000	22:09	

Data Download Date/Time: 07/24/2023 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD, TRAVIS, L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD, TRAVIS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 07/03/2023 22:17:42

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.078	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.078	22:13
Breath Volume	1.449	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.073	22:16
Breath Volume	2.281	22:16
Air Blank	0.000	22:16
Control Test	0.077	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 07/24/2023 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BASFORD,NICOLE,M

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: JARRELL,CHRISTIAN

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/03/2023 19:33:27

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:29	
Control Test	0.077	19:29	
Air Blank	0.000	19:30	
Subject Sample	# 1	19:32	Sequence Aborted
Breath Volume	0.000	19:32	
Air Blank	0.000	19:33	

Data Download Date/Time: 07/24/2023 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JARRELL,CHRISTIAN,P

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: JARRELL,CHRISTIAN

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/03/2023 19:40:26

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:36	
Control Test	0.078	19:37	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:39	Subject Test Refused
Breath Volume	0.000	19:39	
Air Blank	0.000	19:39	
Control Test	0.077	19:39	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 07/24/2023 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JARRELL,CHRISTIAN,P

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: JARRELL,CHRISTIAN

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 16:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 07/04/2023 17:25:43

Diagnostics Check	OK	17:22	
Air Blank	0.000	17:23	
Control Test	0.078	17:23	
Air Blank	0.000	17:24	
Subject Sample	# 1	17:24	Subject Test Refused
Breath Volume	0.000	17:24	
Air Blank	0.000	17:24	
Control Test	0.077	17:25	
Air Blank	0.000	17:25	
Diagnostics Check	OK	17:25	

Data Download Date/Time: 07/24/2023 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BASFORD,NICOLE	ARREST TIME: 18:03
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/12/2023	OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: YOUNGBLOOD	DATE/TIME OF TEST: 07/04/2023 18:42:56
---------------------------------	--

Diagnostics Check	OK	18:36
Air Blank	0.000	18:37
Control Test	0.078	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.105	18:38
Breath Volume	2.238	18:38
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.102	18:41
Breath Volume	2.531	18:41
Air Blank	0.000	18:42
Control Test	0.077	18:42
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 07/24/2023 19:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOLNAR,JARROD,T

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: MIDOLO,GABRIELLE

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 16:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: RICHMOND

DATE/TIME OF TEST: 07/07/2023 17:56:40

Diagnostics Check	OK	17:53	
Air Blank	0.000	17:54	
Control Test	0.078	17:54	
Air Blank	0.000	17:55	
Subject Sample	# 1	17:55	Subject Test Refused
Breath Volume	0.000	17:55	
Air Blank	0.000	17:55	
Control Test	0.078	17:56	
Air Blank	0.000	17:56	
Diagnostics Check	OK	17:56	

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PRITCHARD, THOMAS, D

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: PRITCHARD, THOMAS

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 16:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 07/07/2023 18:17:07

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:11	
Control Test	0.077	18:11	
Air Blank	0.000	18:11	
Subject Sample	# 1 0.128	18:13	
Breath Volume	3.058	18:13	
Air Blank	0.000	18:13	
Air Blank	0.000	18:15	
Subject Sample	# 2 0.124	18:16	
Breath Volume	2.636	18:16	
Air Blank		18:16	RFI Detect
Air Blank		18:17	RFI Detect

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 07/07/2023 18:28:04

Diagnostics Check	OK	18:21
Air Blank	0.000	18:22
Control Test	0.076	18:22
Air Blank	0.000	18:22
Subject Sample # 1	0.153	18:23
Breath Volume	1.320	18:23
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.143	18:26
Breath Volume	2.367	18:26
Air Blank	0.000	18:27
Control Test	0.076	18:27
Air Blank	0.000	18:27
Diagnostics Check	OK	18:28

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: HAHR,DANIEL,E	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: HAHR,DANIEL	ARREST TIME: 19:09
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/12/2023	OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: DORMAN	DATE/TIME OF TEST: 07/07/2023	19:54:03
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.155	19:48
Breath Volume	2.781	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.162	19:52
Breath Volume	2.078	19:52
Air Blank	0.000	19:53
Control Test	0.077	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:54

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PRITCHARD,THOMAS,D

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: PRITCHARD,THOMAS

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 13:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: ARDITO

DATE/TIME OF TEST: 07/08/2023 14:43:47

Diagnostics Check	OK	14:40	
Air Blank	0.000	14:41	
Control Test	0.078	14:41	
Air Blank	0.000	14:42	
Subject Sample	# 1	14:42	Subject Test Refused
Breath Volume	0.000	14:42	
Air Blank	0.000	14:42	
Control Test	0.078	14:43	
Air Blank	0.000	14:43	
Diagnostics Check	OK	14:43	

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKHEAD,ELI,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BURKHEAD,ELI
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: BLAZIER

DATE/TIME OF TEST: 07/08/2023 16:21:04

Diagnostics Check	OK	16:19	
Air Blank	0.000	16:19	
Control Test	0.078	16:20	
Air Blank		16:20	RFI Detect
Air Blank	0.000	16:21	

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKHEAD,ELI,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BURKHEAD,ELI
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: BLAZIER

DATE/TIME OF TEST: 07/08/2023 16:32:39

Diagnostics Check	OK	16:23
Air Blank	0.000	16:24
Control Test	0.078	16:24
Air Blank	0.000	16:24
Subject Sample # 1	0.109	16:27
Breath Volume	2.476	16:27
Air Blank	0.000	16:28
Air Blank	0.000	16:30
Subject Sample # 2	0.117	16:31
Breath Volume	3.289	16:31
Air Blank	0.000	16:31
Control Test	0.076	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NICHOLS,HEATH,C

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: GRAY,SETH

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 15:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:47

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/08/2023 16:47:32

Diagnostics Check	OK	16:44	
Air Blank	0.000	16:45	
Control Test	0.077	16:45	
Air Blank	0.000	16:45	
Subject Sample	# 1	16:46	Subject Test Refused
Breath Volume	0.000	16:46	
Air Blank	0.000	16:46	
Control Test	0.077	16:46	
Air Blank	0.000	16:47	
Diagnostics Check	OK	16:47	

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAHR,DANIEL,E

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: HAHR,DANIEL

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 16:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 07/08/2023 17:12:30

Diagnostics Check	OK	17:07	
Air Blank	0.000	17:08	
Control Test	0.077	17:08	
Air Blank	0.000	17:09	
Subject Sample	# 1	17:11	Subject Test Refused
Breath Volume	0.000	17:11	
Air Blank	0.000	17:11	
Control Test	0.077	17:11	
Air Blank	0.000	17:12	
Diagnostics Check	OK	17:12	

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WEBB,ANDREW,H

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: WEBB,ANDREW

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 17:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: YOUNGBLOOD

DATE/TIME OF TEST: 07/08/2023 17:42:25

Diagnostics Check	OK	17:38	
Air Blank	0.000	17:39	
Control Test	0.077	17:39	
Air Blank	0.000	17:40	
Subject Sample	# 1	17:41	Subject Test Refused
Breath Volume	0.000	17:41	
Air Blank	0.000	17:41	
Control Test	0.077	17:41	
Air Blank	0.000	17:42	
Diagnostics Check	OK	17:42	

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: PRITCHARD, THOMAS, D	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: PRITCHARD, THOMAS	ARREST TIME: 17:20
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/12/2023	OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: LATOURELLE	DATE/TIME OF TEST: 07/08/2023 18:24:05
---------------------------------	--

Diagnostics Check	OK	18:21	
Air Blank	0.000	18:21	
Control Test	0.078	18:22	
Air Blank	0.000	18:22	
Subject Sample	# 1	18:22	Subject Test Refused
Breath Volume	0.000	18:22	
Air Blank	0.000	18:23	
Control Test	0.077	18:23	
Air Blank	0.000	18:24	
Diagnostics Check	OK	18:24	

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWER,RACHEL,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOWER,RACHEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 07/08/2023 18:42:10

Diagnostics Check	OK	18:41	
Air Blank		18:42	RFI Detect
Air Blank		18:42	RFI Detect

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWER,RACHEL,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOWER,RACHEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 07/08/2023 18:46:06

Diagnostics Check	OK	18:45	
Air Blank		18:45	RFI Detect
Air Blank		18:46	RFI Detect

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODGERS,LOGAN,K
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOWER,RACHEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 07/08/2023 18:48:51

Diagnostics Check	OK	18:48	
Air Blank		18:48	RFI Detect
Air Blank		18:48	RFI Detect

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BOWER,RACHEL,A

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BOWER,RACHEL

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 07/08/2023 18:55:58

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:53	
Control Test	0.078	18:53	
Air Blank	0.000	18:54	
Subject Sample	# 1	18:54	Subject Test Refused
Breath Volume	0.000	18:54	
Air Blank	0.000	18:55	
Control Test	0.077	18:55	
Air Blank	0.000	18:55	
Diagnostics Check	OK	18:55	

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARTLETT,SHELTON,T

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BARTLETT,SHELTON

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 08/05/2023

VIOLATION CODE: BUI

ARREST TIME: 19:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: GRAYSON

DATE/TIME OF TEST: 07/08/2023 19:41:16

Diagnostics Check	OK	19:38	
Air Blank	0.000	19:38	
Control Test	0.078	19:39	
Air Blank	0.000	19:39	
Subject Sample	# 1	19:39	Subject Test Refused
Breath Volume	0.000	19:39	
Air Blank	0.000	19:40	
Control Test	0.078	19:40	
Air Blank	0.000	19:41	
Diagnostics Check	OK	19:41	

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 15421080A1
BREATH TEST OPERATOR: Hahr,Daniel,E	DRY GAS EXPIRY DATE: 08/05/2023
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: Hahr,Daniel	ARREST TIME: 20:09
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/12/2023	OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MACGREGOR	DATE/TIME OF TEST: 07/08/2023 20:34:35
--------------------------------	--

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:26	
Control Test	0.078	20:26	
Air Blank	0.000	20:27	
Subject Sample	# 1	20:30	No Sample Provided
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Air Blank	0.000	20:32	
Subject Sample	# 2	20:33	Subject Test Refused
Breath Volume	0.000	20:33	
Air Blank	0.000	20:33	
Control Test	0.078	20:33	
Air Blank	0.000	20:34	
Diagnostics Check	OK	20:34	

Data Download Date/Time: 07/24/2023 19:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/01/2023 17:01:31

Diagnostics Check	OK	16:55
Air Blank	0.000	16:55
Control Test	0.079	16:56
Air Blank	0.000	16:56
Subject Sample # 1	0.200	16:57
Breath Volume	2.882	16:57
Air Blank	0.000	16:57
Air Blank	0.000	16:59
Subject Sample # 2	0.207	16:59
Breath Volume	2.218	16:59
Air Blank	0.000	17:00
Control Test	0.077	17:00
Air Blank	0.000	17:01
Diagnostics Check	OK	17:01

Data Download Date/Time: 07/24/2023 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MCVICAR

DATE/TIME OF TEST: 07/02/2023 01:13:53

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.195	01:09
Breath Volume	2.539	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.201	01:12
Breath Volume	2.832	01:12
Air Blank	0.000	01:12
Control Test	0.076	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 07/24/2023 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402469736
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 06/16/2025
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BASFORD,NICOLE	ARREST TIME: 18:37
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/12/2023	OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: MCGUIGAN	DATE/TIME OF TEST: 07/02/2023 19:37:22
-------------------------------	--

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.078	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.119	19:32
Breath Volume	2.566	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.118	19:35
Breath Volume	2.171	19:35
Air Blank	0.000	19:36
Control Test	0.071	19:36
Air Blank	0.000	19:37

Interferent Detect

Data Download Date/Time: 07/24/2023 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402469736
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 06/16/2025
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BASFORD,NICOLE	ARREST TIME: 18:37
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/12/2023	OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: MCGUIGAN	DATE/TIME OF TEST: 07/02/2023 19:47:53
-------------------------------	--

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.078	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.111	19:43
Breath Volume	2.503	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.109	19:46
Breath Volume	2.964	19:46
Air Blank	0.000	19:46
Control Test	0.078	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 07/24/2023 19:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOLNAR,JARROD,T

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BARRETT,DAVID

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 17:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 07/08/2023 18:49:32

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:47	
Control Test	0.079	18:47	
Air Blank	0.000	18:47	
Subject Sample	# 1	18:48	Subject Test Refused
Breath Volume	0.000	18:48	
Air Blank	0.000	18:48	
Control Test	0.079	18:48	
Air Blank	0.000	18:49	
Diagnostics Check	OK	18:49	

Data Download Date/Time: 07/24/2023 19:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOLNAR,JARROD,T

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BROOKS,JOSEPH

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402469736

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 17:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: ZOMERSCHOE

DATE/TIME OF TEST: 07/08/2023 19:01:27

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:59	
Control Test	0.079	18:59	
Air Blank	0.000	18:59	
Subject Sample	# 1	19:00	Subject Test Refused
Breath Volume	0.000	19:00	
Air Blank	0.000	19:00	
Control Test	0.079	19:00	
Air Blank	0.000	19:01	
Diagnostics Check	OK	19:01	

Data Download Date/Time: 07/24/2023 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,JOSEPH,P
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: PRITCHARD,THOMAS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: BUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 07/08/2023 20:49:58

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.095	20:44
Breath Volume	1.824	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.096	20:48
Breath Volume	2.097	20:48
Air Blank	0.000	20:49
Control Test	0.078	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 07/24/2023 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYCHOFF, COLLIN, T
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: BURNSED, DAVID
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/02/2023 20:31:51

Diagnostics Check	OK	20:19	
Air Blank	0.000	20:19	
Control Test	0.075	20:20	
Air Blank	0.000	20:20	
Subject Sample	# 1	20:23	No Sample Provided
Breath Volume	0.000	20:23	
Air Blank	0.000	20:24	
Air Blank	0.000	20:26	
Subject Sample	# 2	0.076	20:26
Breath Volume	1.570	20:26	
Air Blank	0.000	20:27	
Air Blank	0.000	20:29	
Subject Sample	# 3	0.076	20:30
Breath Volume	2.179	20:30	
Air Blank	0.000	20:30	
Control Test	0.076	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 07/24/2023 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUSTIN,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: LENTZ,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: BRODERICK JR

DATE/TIME OF TEST: 07/02/2023 18:34:57

Diagnostics Check	Fail	18:34	Internal Printer Test
Air Blank	0.000	18:34	

Data Download Date/Time: 07/24/2023 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUSTIN,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: SAPP,WAYNE
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: DRESSER

DATE/TIME OF TEST: 07/02/2023 18:48:56

Diagnostics Check	OK	18:42
Air Blank	0.000	18:42
Control Test	0.080	18:42
Air Blank	0.000	18:43
Subject Sample # 1	0.177	18:44
Breath Volume	1.507	18:44
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.177	18:47
Breath Volume	2.945	18:47
Air Blank	0.000	18:47
Control Test	0.081	18:48
Air Blank	0.000	18:48
Diagnostics Check	OK	18:48

Data Download Date/Time: 07/24/2023 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILMORE,RICHARD,L
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: HAUSLER,TREVOR
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: DEPRATTER

DATE/TIME OF TEST: 07/02/2023 20:11:43

Diagnostics Check	OK	20:08	
Air Blank	0.000	20:09	
Control Test	0.079	20:09	
Air Blank	0.000	20:10	
Subject Sample	# 1	20:10	Subject Test Refused
Breath Volume	0.000	20:10	
Air Blank	0.000	20:10	
Control Test	0.080	20:10	
Air Blank	0.000	20:11	
Diagnostics Check	OK	20:11	

Data Download Date/Time: 07/24/2023 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILMORE,RICHARD,L
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: TYLER,MASON
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: MCPEAK

DATE/TIME OF TEST: 07/03/2023 19:44:50

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:42	
Control Test	0.079	19:42	
Air Blank	0.000	19:43	
Subject Sample	# 1	19:43	Subject Test Refused
Breath Volume	0.000	19:43	
Air Blank	0.000	19:43	
Control Test	0.079	19:44	
Air Blank	0.000	19:44	
Diagnostics Check	OK	19:44	

Data Download Date/Time: 07/24/2023 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILMORE,RICHARD,L
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: WAYCHOFF,COLLIN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: WHITWORTH

DATE/TIME OF TEST: 07/04/2023 21:18:06

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:15	
Control Test	0.080	21:15	
Air Blank	0.000	21:16	
Subject Sample	# 1	21:16	Subject Test Refused
Breath Volume	0.000	21:16	
Air Blank	0.000	21:17	
Control Test	0.079	21:17	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:18	

Data Download Date/Time: 07/24/2023 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITT,JESSE,T
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: WHITT,JESSE
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: MEREDITH

DATE/TIME OF TEST: 07/08/2023 12:05:52

Diagnostics Check	OK	11:59
Air Blank	0.000	11:59
Control Test	0.081	12:00
Air Blank	0.000	12:00
Subject Sample # 1	0.000	12:01
Breath Volume	2.488	12:01
Air Blank	0.000	12:02
Air Blank	0.000	12:04
Subject Sample # 2	0.000	12:04
Breath Volume	1.929	12:04
Air Blank	0.000	12:04
Control Test	0.079	12:05
Air Blank	0.000	12:05
Diagnostics Check	OK	12:05

Data Download Date/Time: 07/18/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:08

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 07/13/2023 09:41:18

Diagnostics Check	OK	09:35
Air Blank	0.000	09:35
Control Test	0.080	09:35
Air Blank	0.000	09:36
Subject Sample # 1	0.000	09:37
Breath Volume	2.792	09:37
Air Blank	0.000	09:37
Air Blank	0.000	09:39
Subject Sample # 2	0.000	09:39
Breath Volume	2.414	09:39
Air Blank	0.000	09:40
Control Test	0.080	09:40
Air Blank	0.000	09:41
Diagnostics Check	OK	09:41

Data Download Date/Time: 08/28/2023 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,A,K
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALMERTON,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:49

SUBJECT'S LAST NAME: TOLBERT

DATE/TIME OF TEST: 07/02/2023 14:18:28

Diagnostics Check	OK	14:09
Air Blank	0.000	14:09
Control Test	0.080	14:10
Air Blank	0.000	14:10
Subject Sample # 1	0.000	14:13
Breath Volume	1.148	14:13
Air Blank	0.000	14:14
Air Blank	0.000	14:16
Subject Sample # 2	0.000	14:17
Breath Volume	1.582	14:17
Air Blank	0.000	14:17
Control Test	0.080	14:17
Air Blank	0.000	14:18
Diagnostics Check	OK	14:18

Data Download Date/Time: 07/10/2023 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTE,KYLE,M
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GORMAN,H
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/03/2023 04:54:36

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.081	04:49
Breath Volume	1.800	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.080	04:52
Breath Volume	1.480	04:52
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 07/10/2023 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN, SHON, W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RICHARDSON, SARAH
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: ETHRIDGE

DATE/TIME OF TEST: 07/03/2023 23:53:21

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.233	23:48
Breath Volume	1.835	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.231	23:51
Breath Volume	2.203	23:51
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 07/10/2023 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,C,H
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SERGEYEV,P
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ESPINOZA HERNANDEZ

DATE/TIME OF TEST: 07/07/2023 00:03:52

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.081	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.080	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 07/10/2023 06:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,DALE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: GLUCK

DATE/TIME OF TEST: 07/14/2023 16:14:41

Diagnostics Check	OK	16:08
Air Blank	0.000	16:08
Control Test	0.081	16:09
Air Blank	0.000	16:09
Subject Sample # 1	0.000	16:10
Breath Volume	3.195	16:10
Air Blank	0.000	16:10
Air Blank	0.000	16:12
Subject Sample # 2	0.000	16:13
Breath Volume	2.750	16:13
Air Blank	0.000	16:13
Control Test	0.079	16:14
Air Blank	0.000	16:14
Diagnostics Check	OK	16:14

Data Download Date/Time: 08/28/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ALLRED,A
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 07/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: TELESKO

DATE/TIME OF TEST: 07/15/2023 04:53:34

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.147	04:48
Breath Volume	1.417	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.141	04:51
Breath Volume	1.687	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 08/28/2023 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROBLES,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BOCHNER

DATE/TIME OF TEST: 07/17/2023 00:13:12

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.081	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Control Test	0.080	00:12	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 08/28/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,J,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 12:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: TOLBERT

DATE/TIME OF TEST: 07/17/2023 13:11:31

Diagnostics Check	OK	13:04
Air Blank	0.000	13:05
Control Test	0.081	13:05
Air Blank	0.000	13:06
Subject Sample # 1	0.000	13:07
Breath Volume	1.917	13:07
Air Blank	0.000	13:07
Air Blank	0.000	13:09
Subject Sample # 2	0.000	13:10
Breath Volume	1.820	13:10
Air Blank	0.000	13:10
Control Test	0.080	13:10
Air Blank	0.000	13:11
Diagnostics Check	OK	13:11

Data Download Date/Time: 08/28/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RICHARDSON,S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BACON

DATE/TIME OF TEST: 07/17/2023 21:53:39

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.081	21:51	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Control Test	0.080	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 08/28/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WALKER,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 07/28/2023 02:40:42

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.118	02:35
Breath Volume	1.699	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.118	02:38
Breath Volume	2.179	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 08/28/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALMERTON,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MONTENEGRO-RODRIGUEZ

DATE/TIME OF TEST: 07/28/2023 23:42:50

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.196	23:38
Breath Volume	2.125	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.181	23:41
Breath Volume	2.148	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 08/28/2023 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALMERTON,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 07/29/2023 19:16:39

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:13	
Control Test	0.081	19:13	
Air Blank	0.000	19:14	
Subject Sample	# 1	19:15	Subject Test Refused
Breath Volume	0.000	19:15	
Air Blank	0.000	19:15	
Control Test	0.081	19:16	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:16	

Data Download Date/Time: 08/28/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDSON,A,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SERGEYEV,OFFICER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ALONZO

DATE/TIME OF TEST: 07/30/2023 02:14:05

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.080	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 08/28/2023 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIZARDO, RAILY, J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: GARZA, DAVID
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: S360475691780

DATE/TIME OF TEST: 07/01/2023 03:42:07

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.080	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	RFI Detect
Breath Volume	0.000	03:41	
Air Blank		03:42	Improper Sample
Air Blank		03:42	RFI Detect

Data Download Date/Time: 07/28/2023 19:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIZARDO, RAILY, J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: GARZA, DAVID
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: STEYER

DATE/TIME OF TEST: 07/01/2023 03:46:27

Diagnostics Check	OK	03:45	
Air Blank		03:45	Ambient Fail
Air Blank	0.000	03:46	

Data Download Date/Time: 07/28/2023 19:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIZARDO,RAILY,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: GARZA,DAVID
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: STEYER

DATE/TIME OF TEST: 07/01/2023 04:00:15

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:49	
Control Test	0.079	03:49	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2	03:58	No Sample Provided
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.080	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 07/28/2023 19:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISCOCK,MANUELA,E
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: ULLRICH,LAURA
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:23

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 07/09/2023 13:51:53

Diagnostics Check	OK	13:44
Air Blank	0.000	13:45
Control Test	0.080	13:45
Air Blank	0.000	13:46
Subject Sample # 1	0.202	13:47
Breath Volume	2.199	13:47
Air Blank	0.000	13:48
Air Blank	0.000	13:49
Subject Sample # 2	0.192	13:50
Breath Volume	1.347	13:50
Air Blank	0.000	13:50
Control Test	0.076	13:51
Air Blank	0.000	13:51
Diagnostics Check	OK	13:51

Data Download Date/Time: 07/28/2023 19:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: SMITH,SAUL
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: GRECKY

DATE/TIME OF TEST: 07/16/2023 03:53:25

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 07/28/2023 19:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLFLOWER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: MCHALE,CASSIE
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: BARO

DATE/TIME OF TEST: 07/26/2023 17:33:03

Diagnostics Check	OK	17:30	
Air Blank	0.000	17:30	
Control Test	0.079	17:30	
Air Blank	0.000	17:31	
Subject Sample	# 1	17:31	Subject Test Refused
Breath Volume	0.000	17:31	
Air Blank	0.000	17:32	
Control Test	0.079	17:32	
Air Blank	0.000	17:32	
Diagnostics Check	OK	17:33	

Data Download Date/Time: 07/28/2023 19:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BILLON,LAURENT
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 07/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: GONZALEZREYES

DATE/TIME OF TEST: 07/30/2023 00:04:59

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.078	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 2 0.232	23:59	
Breath Volume	1.191	23:59	
Air Blank	0.000	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.238	00:02	
Breath Volume	2.156	00:02	
Air Blank	0.000	00:03	
Control Test	0.076	00:03	
Air Blank	0.000	00:04	
Diagnostics Check	Fail	00:04	Internal Printer Test
Air Blank	0.000	00:04	

Data Download Date/Time: 08/10/2023 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIZARDO, RAILY, J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: NUNEZ, NIURVIS
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 07/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: LACROIX

DATE/TIME OF TEST: 07/30/2023 05:17:27

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:14	
Control Test	0.079	05:15	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:16	Subject Test Refused
Breath Volume	0.000	05:16	
Air Blank	0.000	05:16	
Control Test	0.079	05:16	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 08/10/2023 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 07/04/2023 00:07:59

Diagnostics Check	Fail	00:07	Internal Printer Test
Air Blank	0.000	00:07	

Data Download Date/Time: 07/26/2023 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 07/04/2023 00:09:34

Diagnostics Check	Fail	00:09	Internal Printer Test
Air Blank	0.000	00:09	

Data Download Date/Time: 07/26/2023 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 07/04/2023 00:11:32

Diagnostics Check	Fail	00:11	Internal Printer Test
Air Blank	0.000	00:11	

Data Download Date/Time: 07/26/2023 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 07/04/2023 00:18:54

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.077	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.261	00:14
Breath Volume	2.253	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:16
Subject Sample # 2	0.254	00:17
Breath Volume	2.015	00:17
Air Blank	0.000	00:17
Control Test	0.077	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 07/26/2023 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BURNUM

DATE/TIME OF TEST: 07/14/2023 01:50:33

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.077	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.164	01:46
Breath Volume	2.281	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.151	01:48
Breath Volume	2.015	01:48
Air Blank	0.000	01:49
Control Test	0.076	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/26/2023 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/27/2023 22:00:45

Diagnostics Check	Fail	22:00	Internal Printer Test
Air Blank	0.000	22:00	

Data Download Date/Time: 08/23/2023 22:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/27/2023 22:03:09

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
		22:02	Sequence Aborted
Air Blank	0.000	22:03	

Data Download Date/Time: 08/23/2023 22:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/27/2023 22:23:48

Diagnostics Check	Fail	22:23	Internal Printer Test
Air Blank	0.000	22:23	

Data Download Date/Time: 08/23/2023 22:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/27/2023 22:25:21

Diagnostics Check	Fail	22:24	Internal Printer Test
Air Blank	0.000	22:25	

Data Download Date/Time: 08/23/2023 22:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/27/2023 22:33:17

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.077	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.157	22:28
Breath Volume	1.613	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.155	22:31
Breath Volume	1.390	22:31
Air Blank	0.000	22:32
Control Test	0.077	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 08/23/2023 22:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RUBIO,ROBERT
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MURILLO MURILLO

DATE/TIME OF TEST: 07/09/2023 01:14:20

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.079	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:13	
Control Test	0.080	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 07/31/2023 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ESPINOSA,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 07/24/2023 00:13:26

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:11	
Control Test	0.079	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:12	
Control Test	0.080	00:12	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 07/31/2023 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 03:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: HERNANDEZ BARAHONA

DATE/TIME OF TEST: 07/30/2023 04:11:30

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.078	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Control Test	0.079	04:10	
Air Blank	0.000	04:10	
Diagnostics Check	Fail	04:11	Internal Printer Test
Air Blank	0.000	04:11	

Data Download Date/Time: 07/31/2023 13:31