

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 08/03/2023 03:12:04

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.142	03:07
Breath Volume	2.566	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.145	03:10
Breath Volume	3.023	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:12

Data Download Date/Time: 08/08/2023 19:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,KEVIN,E
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MORRIS,KEVIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:04

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 08/03/2023 13:32:19

Diagnostics Check	OK	13:25
Air Blank	0.000	13:26
Control Test	0.080	13:26
Air Blank	0.000	13:27
Subject Sample # 1	0.342	13:27
Breath Volume	2.199	13:27
Air Blank	0.000	13:28
Air Blank	0.000	13:30
Subject Sample # 2	0.332	13:30
Breath Volume	2.136	13:30
Air Blank	0.000	13:31
Control Test	0.080	13:31
Air Blank	0.000	13:32
Diagnostics Check	OK	13:32

Data Download Date/Time: 08/08/2023 19:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: GIANGIACOMO

DATE/TIME OF TEST: 08/05/2023 22:47:37

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.081	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.229	22:43
Breath Volume	1.492	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.229	22:46
Breath Volume	1.542	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 08/08/2023 19:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MONROY

DATE/TIME OF TEST: 08/06/2023 00:10:12

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.176	00:05
Breath Volume	3.410	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.174	00:08
Breath Volume	2.312	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 08/08/2023 19:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPIRITO,GEORGE,W
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SPIRITO,GEORGE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: BUI
ARREST TIME: 17:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 08/06/2023 21:00:28

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.081	20:54
Air Blank	0.000	20:54
Subject Sample # 1	0.000	20:55
Breath Volume	2.875	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.000	20:59
Breath Volume	2.820	20:59
Air Blank	0.000	20:59
Control Test	0.080	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 08/08/2023 19:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: COFFINDAFFER

DATE/TIME OF TEST: 08/09/2023 21:00:42

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.080	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.267	20:56
Breath Volume	1.796	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.260	20:59
Breath Volume	1.875	20:59
Air Blank	0.000	20:59
Control Test	0.080	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHAM,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SANTAMARIA,LEONARDO
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: CASHMAN

DATE/TIME OF TEST: 08/09/2023 23:41:14

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.146	23:36
Breath Volume	1.640	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.146	23:39
Breath Volume	1.324	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GREIS,BRANDON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 08/11/2023 21:01:46

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.080	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.154	20:57
Breath Volume	2.707	20:57
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.147	21:00
Breath Volume	3.128	21:00
Air Blank	0.000	21:00
Control Test	0.080	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SCHAJATOVIC

DATE/TIME OF TEST: 08/13/2023 00:31:23

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.081	00:20	
Air Blank	0.000	00:20	
Subject Sample	# 1 0.218	00:23	Slope Not Level
Breath Volume	1.273	00:23	
Air Blank	0.000	00:24	
Air Blank	0.000	00:26	
Subject Sample	# 2 0.256	00:26	
Breath Volume	2.257	00:26	
Air Blank	0.000	00:27	
Air Blank	0.000	00:29	
Subject Sample	# 3 0.258	00:29	
Breath Volume	1.312	00:29	
Air Blank	0.000	00:30	
Control Test	0.079	00:30	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: AUVIL

DATE/TIME OF TEST: 08/14/2023 21:40:50

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.080	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:35	No Sample Provided
Breath Volume	0.000	21:35	
Air Blank	0.000	21:35	
Air Blank	0.000	21:37	
Subject Sample	# 2	21:39	Subject Test Refused
Breath Volume	0.000	21:39	
Air Blank	0.000	21:39	
Control Test	0.081	21:40	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: CIFUENTES

DATE/TIME OF TEST: 08/15/2023 00:51:47

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.081	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.303	00:47
Breath Volume	1.960	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.300	00:50
Breath Volume	2.875	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FATZINGER

DATE/TIME OF TEST: 08/15/2023 01:27:33

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.148	01:23
Breath Volume	1.699	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.146	01:26
Breath Volume	2.128	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: KOZLOWSKI

DATE/TIME OF TEST: 08/16/2023 03:10:39

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.212	03:05
Breath Volume	1.382	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:07
Subject Sample # 2	0.194	03:09
Breath Volume	1.781	03:09
Air Blank	0.000	03:09
Control Test	0.078	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 08/21/2023 00:37:32

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.081	00:33
Breath Volume	1.496	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.079	00:36
Breath Volume	1.800	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: LOPEZ VICENTE

DATE/TIME OF TEST: 08/21/2023 02:05:15

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.130	02:00
Breath Volume	1.933	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.132	02:03
Breath Volume	1.718	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: XIVIR-ZABALA

DATE/TIME OF TEST: 08/22/2023 22:48:49

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.246	22:44
Breath Volume	3.277	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.245	22:47
Breath Volume	3.429	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: DESMOND

DATE/TIME OF TEST: 08/26/2023 03:16:47

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.141	03:12
Breath Volume	3.796	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.131	03:15
Breath Volume	2.679	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: LINMAN

DATE/TIME OF TEST: 08/26/2023 20:42:56

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.114	20:38
Breath Volume	3.042	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.107	20:41
Breath Volume	2.031	20:41
Air Blank	0.000	20:41
Control Test	0.080	20:42
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: KNOOP,KENNETH
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: DA SILVA CARVALHEDO

DATE/TIME OF TEST: 08/27/2023 02:02:21

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.241	01:57
Breath Volume	1.714	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.240	02:00
Breath Volume	2.089	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 08/28/2023 01:25:43

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.208	01:21
Breath Volume	1.789	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.204	01:24
Breath Volume	1.808	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SPORTELLI

DATE/TIME OF TEST: 08/30/2023 02:01:14

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.158	01:56
Breath Volume	2.609	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.151	01:59
Breath Volume	1.218	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: ABBEY

DATE/TIME OF TEST: 08/30/2023 23:21:21

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:13	
Control Test	0.081	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 1	23:17	No Sample Provided
Breath Volume	0.000	23:17	
Air Blank	0.000	23:17	
Air Blank	0.000	23:19	
Subject Sample	# 2	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:20	
Control Test	0.080	23:20	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 09/08/2023 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY, WAYNE, G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERRY, WAYNE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 08/11/2023 18:52:23

Diagnostics Check	OK	18:41
Air Blank	0.000	18:42
Control Test	0.079	18:42
Air Blank	0.000	18:42
Subject Sample # 1	0.254	18:43
Breath Volume	1.808	18:43
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.275	18:47
Breath Volume	2.386	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:49
Subject Sample # 3	0.264	18:50
Breath Volume	1.773	18:50
Air Blank	0.000	18:51
Control Test	0.076	18:51
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

Data Download Date/Time: 08/30/2023 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAKEMAN,TAYLOR,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LAKEMAN,TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: ADMIN
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HAAS

DATE/TIME OF TEST: 08/12/2023 03:34:33

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.000	03:30
Breath Volume	3.128	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.000	03:33
Breath Volume	1.734	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 08/15/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIAS, HUNTER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 08/24/2023 03:16:21

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.079	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.124	03:11
Breath Volume	1.367	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.120	03:14
Breath Volume	1.539	03:14
Air Blank	0.000	03:15
Control Test	0.077	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 09/21/2023 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: DEMOUGIN

DATE/TIME OF TEST: 08/29/2023 01:06:22

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.018	01:01
Breath Volume	1.707	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.017	01:04
Breath Volume	1.832	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 09/21/2023 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: TRIGUERO LOPEZ

DATE/TIME OF TEST: 08/20/2023 21:47:46

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.078	21:42	
Air Blank	0.000	21:42	
Subject Sample # 1	0.243	21:43	
Breath Volume	1.882	21:43	
Air Blank	0.000	21:44	
Air Blank	0.000	21:45	
Subject Sample # 2	0.224	21:46	
Breath Volume	1.292	21:46	
Air Blank	0.000	21:46	
Control Test	0.066	21:47	Control Outside Tolerance
Air Blank	0.000	21:47	

Data Download Date/Time: 09/18/2023 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: TRIGUERO LOPEZ

DATE/TIME OF TEST: 08/20/2023 21:56:37

Diagnostics Check	OK	21:49
Air Blank	0.000	21:50
Control Test	0.077	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.245	21:52
Breath Volume	1.914	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.225	21:54
Breath Volume	1.300	21:54
Air Blank	0.000	21:55
Control Test	0.077	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 09/18/2023 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYES JR,DARRON,O
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MAYES JR,DARRON
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 08/21/2023 18:17:03

Diagnostics Check	OK	18:09
Air Blank	0.000	18:10
Control Test	0.078	18:10
Air Blank	0.000	18:11
Subject Sample # 1	0.209	18:11
Breath Volume	1.839	18:11
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.198	18:15
Breath Volume	1.996	18:15
Air Blank	0.000	18:16
Control Test	0.078	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:17

Data Download Date/Time: 09/18/2023 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 08/22/2023 10:56:07

Diagnostics Check	OK	10:47
Air Blank	0.000	10:48
Control Test	0.078	10:48
Air Blank	0.000	10:49
Subject Sample # 1	0.000	10:51
Breath Volume	2.382	10:51
Air Blank	0.000	10:51
Air Blank	0.000	10:53
Subject Sample # 2	0.000	10:54
Breath Volume	1.937	10:54
Air Blank	0.000	10:55
Control Test	0.078	10:55
Air Blank	0.000	10:56
Diagnostics Check	OK	10:56

Data Download Date/Time: 09/18/2023 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,ELIER,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SANTANA,ELIER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: OGILVIE

DATE/TIME OF TEST: 08/22/2023 14:01:19

Diagnostics Check	OK	13:54	
Air Blank	0.000	13:55	
Control Test	0.078	13:55	
Air Blank	0.000	13:56	
Subject Sample	# 1 0.269	13:56	
Breath Volume	2.609	13:56	
Air Blank	0.000	13:57	
Air Blank	0.000	13:59	
Subject Sample	# 2 0.251	13:59	
Breath Volume	2.250	13:59	
Air Blank	0.000	14:00	
Control Test	0.074	14:00	Control Outside Tolerance
Air Blank	0.000	14:01	

Data Download Date/Time: 09/18/2023 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,ELIER,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SANTANA,ELIER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: OGILVIE

DATE/TIME OF TEST: 08/22/2023 14:37:53

Diagnostics Check	OK	14:31
Air Blank	0.000	14:32
Control Test	0.077	14:32
Air Blank	0.000	14:32
Subject Sample # 1	0.247	14:33
Breath Volume	2.316	14:33
Air Blank	0.000	14:34
Air Blank	0.000	14:35
Subject Sample # 2	0.242	14:36
Breath Volume	2.179	14:36
Air Blank	0.000	14:36
Control Test	0.077	14:37
Air Blank	0.000	14:37
Diagnostics Check	OK	14:37

Data Download Date/Time: 09/18/2023 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYES JR,DARRON,O
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MAYES JR,DARRON
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 08/24/2023 00:44:01

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.078	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1 0.174	00:38	Volume Not Met
Breath Volume	0.015	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:40	
Subject Sample	# 2	00:42	Subject Test Refused
Breath Volume	0.082	00:42	
Air Blank	0.000	00:43	
Control Test	0.075	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 09/18/2023 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DOWDY

DATE/TIME OF TEST: 08/24/2023 02:20:39

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.076	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.189	02:15
Breath Volume	2.535	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.169	02:18
Breath Volume	1.273	02:18
Air Blank	0.000	02:19
Control Test	0.075	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 09/18/2023 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYES JR,DARRON,O
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: FERRIGINE,JOSEPH
ARRESTING OFFICER AGENCY: Longboat Key PD
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DINNIE

DATE/TIME OF TEST: 08/24/2023 02:27:00

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.076	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:26	
Control Test	0.077	02:26	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 09/18/2023 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYES JR,DARRON,O
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MILLER,JORDAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VICKERS

DATE/TIME OF TEST: 08/24/2023 03:07:05

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.137	03:02
Breath Volume	1.199	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.137	03:05
Breath Volume	1.300	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 09/18/2023 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: CLASS

DATE/TIME OF TEST: 08/25/2023 00:38:15

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.079	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:37	
Control Test	0.078	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 09/18/2023 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SANCHEZ-MARTINEZ

DATE/TIME OF TEST: 08/27/2023 01:24:03

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.156	01:19
Breath Volume	2.488	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.165	01:22
Breath Volume	3.222	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:24

Data Download Date/Time: 09/18/2023 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SEWELL

DATE/TIME OF TEST: 08/28/2023 02:52:56

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.038	02:48
Breath Volume	2.628	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.042	02:51
Breath Volume	2.898	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 09/18/2023 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 08/31/2023 02:19:45

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.193	02:14
Breath Volume	2.156	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.189	02:18
Breath Volume	1.628	02:18
Air Blank	0.000	02:18
Control Test	0.076	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 09/18/2023 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,JOHN,P
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEEMAN,NICOLAS
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: LASALLE AVILES

DATE/TIME OF TEST: 08/31/2023 04:18:15

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.136	04:13
Breath Volume	3.394	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.137	04:16
Breath Volume	3.242	04:16
Air Blank	0.000	04:17
Control Test	0.077	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 09/18/2023 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/16/2023 02:42:22

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.078	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.000	02:38
Breath Volume	1.359	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.000	02:40
Breath Volume	1.453	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 09/07/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: SCHWARZER

DATE/TIME OF TEST: 08/24/2023 00:09:12

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.182	00:04
Breath Volume	1.843	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.178	00:07
Breath Volume	1.734	00:07
Air Blank	0.000	00:08
Control Test	0.077	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 09/07/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 08/25/2023 01:39:31

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.077	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.000	01:35
Breath Volume	1.496	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.000	01:38
Breath Volume	1.652	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 09/07/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PERRIE

DATE/TIME OF TEST: 08/31/2023 01:36:42

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.078	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:35	RFI Detect
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.079	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 09/07/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PERRIE

DATE/TIME OF TEST: 08/31/2023 01:45:06

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.077	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.120	01:40
Breath Volume	2.054	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.118	01:43
Breath Volume	1.921	01:43
Air Blank	0.000	01:44
Control Test	0.077	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 09/07/2023 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,MESHELLE,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NAYLOR,MESHELLE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 08/27/2023 23:18:49

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.081	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.183	23:14
Breath Volume	1.562	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:16
Subject Sample # 2	0.184	23:17
Breath Volume	2.253	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 09/12/2023 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,MESHELLE,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NAYLOR,MESHELLE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/29/2023 23:29:09

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.082	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.188	23:24
Breath Volume	2.039	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:26
Subject Sample # 2	0.183	23:27
Breath Volume	2.046	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 09/12/2023 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 08/08/2023 01:32:17

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.062	01:27
Breath Volume	1.835	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.061	01:30
Breath Volume	2.062	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 08/27/2023 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIRADO,FREDERID,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TIRADO,FREDERID
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: PEDERSEN

DATE/TIME OF TEST: 08/12/2023 21:19:02

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.079	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1 0.223	21:15	Volume Not Met
Breath Volume	0.074	21:15	
Air Blank	0.000	21:15	
Air Blank	0.000	21:17	
Subject Sample	# 2	21:17	RFI Detect
Breath Volume	0.000	21:17	
Air Blank	0.000	21:18	
Control Test	0.076	21:18	
Air Blank	0.000	21:18	
Diagnostics Check	OK	21:19	

Data Download Date/Time: 08/27/2023 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIRADO,FREDERID,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TIRADO,FREDERID
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: PEDERSEN

DATE/TIME OF TEST: 08/12/2023 21:28:43

Diagnostics Check	OK	21:21
Air Blank	0.000	21:22
Control Test	0.077	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.224	21:23
Breath Volume	1.859	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.231	21:26
Breath Volume	1.480	21:26
Air Blank	0.000	21:27
Control Test	0.078	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 08/27/2023 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BOSTICK,JESSE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: CALAF VAZQUEZ

DATE/TIME OF TEST: 08/20/2023 01:11:34

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.271	01:06
Breath Volume	1.179	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.272	01:09
Breath Volume	1.214	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 08/27/2023 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: COWART

DATE/TIME OF TEST: 08/26/2023 00:45:14

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.078	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:40	No Sample Provided
Breath Volume	0.000	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 2	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.079	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 08/27/2023 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LARRY,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RODRIGUEZ,LARRY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PINERO

DATE/TIME OF TEST: 08/01/2023 21:54:22

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:51	
Control Test	0.080	21:52	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:53	
Control Test	0.080	21:53	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:54	

Data Download Date/Time: 08/04/2023 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RAMIREZ GARCIA, HERNAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: AYALA RIOS

DATE/TIME OF TEST: 08/02/2023 02:14:17

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.099	02:09
Breath Volume	1.546	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.103	02:12
Breath Volume	2.421	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 08/04/2023 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, LAURA,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: ROSADO, KIARA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: MONROY MALVAES

DATE/TIME OF TEST: 08/06/2023 18:25:07

Diagnostics Check	OK	18:18
Air Blank	0.000	18:19
Control Test	0.080	18:19
Air Blank	0.000	18:19
Subject Sample # 1	0.302	18:20
Breath Volume	3.136	18:20
Air Blank	0.000	18:21
Air Blank	0.000	18:22
Subject Sample # 2	0.297	18:23
Breath Volume	2.531	18:23
Air Blank	0.000	18:24
Control Test	0.076	18:24
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

Data Download Date/Time: 09/07/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LARRY,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RODRIGUEZ,LARRY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 08/08/2023 06:06:16

Diagnostics Check	OK	05:59
Air Blank	0.000	05:59
Control Test	0.080	06:00
Air Blank	0.000	06:00
Subject Sample # 1	0.147	06:01
Breath Volume	3.777	06:01
Air Blank	0.000	06:02
Air Blank	0.000	06:03
Subject Sample # 2	0.148	06:04
Breath Volume	2.773	06:04
Air Blank	0.000	06:05
Control Test	0.079	06:05
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 09/07/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: ESTIVERNE,JAMES
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 08/11/2023 03:14:23

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.099	03:09
Breath Volume	3.535	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.098	03:12
Breath Volume	3.238	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 09/07/2023 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEIL,DARRIN,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANEIL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/13/2023 04:17:57

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.058	04:13
Breath Volume	2.160	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.057	04:16
Breath Volume	1.558	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 09/07/2023 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,MATTHEW,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GARCIA,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 08/19/2023 00:19:15

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.080	00:14
Breath Volume	3.007	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.080	00:17
Breath Volume	2.871	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 09/07/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, LAURA,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GOLDBERG, JUSTIN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 10:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:21

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 08/23/2023 11:52:33

Diagnostics Check	OK	11:45
Air Blank	0.000	11:46
Control Test	0.080	11:46
Air Blank	0.000	11:47
Subject Sample # 1	0.178	11:47
Breath Volume	2.863	11:47
Air Blank	0.000	11:48
Air Blank	0.000	11:50
Subject Sample # 2	0.171	11:50
Breath Volume	2.445	11:50
Air Blank	0.000	11:51
Control Test	0.077	11:51
Air Blank	0.000	11:52
Diagnostics Check	OK	11:52

Data Download Date/Time: 09/07/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS-PIERRE,RUBEN,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: LOUIS-PIERE,RUBEN
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: FRIDAY

DATE/TIME OF TEST: 08/27/2023 02:37:43

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.205	02:32
Breath Volume	1.523	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.217	02:35
Breath Volume	3.046	02:35
Air Blank	0.000	02:36
Control Test	0.075	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 09/07/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS-PIERRE,RUBEN,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: BELL,AUSTIN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BURKITT

DATE/TIME OF TEST: 08/27/2023 03:54:25

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.079	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1 0.141	03:52	
Breath Volume	1.296	03:52	
Air Blank		03:52	Improper Sample
Air Blank	0.000	03:53	
Control Test	0.080	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 09/07/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS-PIERRE,RUBEN,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: BELL,AUSTIN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BURKITT

DATE/TIME OF TEST: 08/27/2023 04:05:05

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.078	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.150	04:00
Breath Volume	1.390	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.144	04:03
Breath Volume	1.121	04:03
Air Blank	0.000	04:04
Control Test	0.077	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 09/07/2023 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLANDA, RICARDO, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HOLANDA, RICARDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 08/20/2023 00:12:00

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1 0.210	00:05	
Breath Volume	1.382	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:08	
Subject Sample	# 2 0.189	00:10	
Breath Volume	1.101	00:10	
Air Blank		00:11	Purge Fail
Air Blank		00:12	Purge Fail

Data Download Date/Time: 09/13/2023 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLTON,JUSTIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOLTON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: KENT

DATE/TIME OF TEST: 08/05/2023 16:56:52

Diagnostics Check	OK	16:50
Air Blank	0.000	16:51
Control Test	0.080	16:51
Air Blank	0.000	16:51
Subject Sample # 1	0.268	16:52
Breath Volume	1.265	16:52
Air Blank	0.000	16:53
Air Blank	0.000	16:54
Subject Sample # 2	0.257	16:55
Breath Volume	1.355	16:55
Air Blank	0.000	16:55
Control Test	0.078	16:56
Air Blank	0.000	16:56
Diagnostics Check	OK	16:56

Data Download Date/Time: 08/15/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MCMULLEN,GRANT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/09/2023 18:58:25

Diagnostics Check	OK	18:49	
Air Blank	0.000	18:50	
Control Test	0.081	18:50	
Air Blank	0.000	18:51	
Subject Sample	# 1	18:54	No Sample Provided
Breath Volume	0.000	18:54	
Air Blank	0.000	18:54	
Air Blank	0.000	18:56	
Subject Sample	# 2	18:57	Subject Test Refused
Breath Volume	0.000	18:57	
Air Blank	0.000	18:57	
Control Test	0.081	18:57	
Air Blank	0.000	18:58	
Diagnostics Check	OK	18:58	

Data Download Date/Time: 08/15/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIRES, DAVID,
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: CUTCHIN, HARBEY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 08:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 08/10/2023 10:08:18

Diagnostics Check	OK	09:56	
Air Blank	0.000	09:57	
Control Test	0.081	09:57	
Air Blank	0.000	09:57	
Subject Sample	# 1 0.166	10:01	Volume Not Met
Breath Volume	0.699	10:01	
Air Blank	0.000	10:01	
Air Blank	0.000	10:03	
Subject Sample	# 2 0.144	10:06	Volume Not Met
Breath Volume	0.777	10:06	
Air Blank	0.000	10:07	
Control Test	0.078	10:07	
Air Blank	0.000	10:08	
Diagnostics Check	OK	10:08	

Data Download Date/Time: 08/15/2023 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: Key Biscayne PD
ARRESTING OFFICER: LORENZO,FRANCISCO
ARRESTING OFFICER AGENCY: Key Biscayne PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MEDINA VARGAS

DATE/TIME OF TEST: 08/02/2023 01:51:22

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.079	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Control Test	0.079	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 08/18/2023 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,DANELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: WILSON FELLERS

DATE/TIME OF TEST: 08/05/2023 16:16:13

Diagnostics Check	OK	16:08	
Air Blank	0.000	16:08	
Control Test	0.080	16:09	
Air Blank	0.000	16:09	
Subject Sample	# 1 0.000	16:11	
Breath Volume	1.535	16:11	
Air Blank	0.000	16:11	
Air Blank	0.000	16:13	
Subject Sample	# 2 0.000	16:15	
Breath Volume	1.855	16:15	
Air Blank	0.000	16:15	
Control Test	0.000	16:15	RFI Detect
Air Blank	0.000	16:16	

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,DANELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 08/06/2023 01:57:49

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.144	01:51
Breath Volume	1.519	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.140	01:56
Breath Volume	2.636	01:56
Air Blank	0.000	01:56
Control Test	0.077	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,ANGEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LARA,JOSHUA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 02:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: GONZALEZ-OBANDO

DATE/TIME OF TEST: 08/06/2023 03:21:30

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.061	03:16
Breath Volume	1.476	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.059	03:19
Breath Volume	1.558	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOUSSA,ELIAS,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MOUSSA,ELIAS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: VEHAR

DATE/TIME OF TEST: 08/08/2023 04:53:34

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.080	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:52	
Control Test	0.080	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAMIREZ,CRISTIAN,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: SIMMONS,DAMIAN

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 04:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: DEL TORO GUERRA

DATE/TIME OF TEST: 08/10/2023 06:09:15

Diagnostics Check	OK	06:02
Air Blank	0.000	06:02
Control Test	0.080	06:03
Air Blank	0.000	06:03
Subject Sample # 1	0.212	06:04
Breath Volume	2.175	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:06
Subject Sample # 2	0.218	06:07
Breath Volume	2.445	06:07
Air Blank	0.000	06:08
Control Test	0.076	06:08
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HERNANDEZ,FERNANDO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 05:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: AREAS ALTAMIRANO

DATE/TIME OF TEST: 08/12/2023 06:21:29

Diagnostics Check	OK	06:12	
Air Blank	0.000	06:13	
Control Test	0.079	06:13	
Air Blank	0.000	06:14	
Subject Sample	# 1 0.169	06:17	Volume Not Met
Breath Volume	0.234	06:17	
Air Blank	0.000	06:17	
Air Blank	0.000	06:19	
Subject Sample	# 2	06:20	Subject Test Refused
Breath Volume	0.000	06:20	
Air Blank	0.000	06:20	
Control Test	0.078	06:20	
Air Blank		06:21	RFI Detect
Air Blank	0.000	06:21	

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ,MICHEAL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RODRIGUEZ,MICHEAL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 05:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: AREAS ALTAMIRANO

DATE/TIME OF TEST: 08/12/2023 06:28:45

Diagnostics Check	OK	06:24	
Air Blank	0.000	06:24	
Control Test	0.079	06:24	
Air Blank	0.000	06:25	
Subject Sample	# 1	06:27	Subject Test Refused
Breath Volume	0.000	06:27	
Air Blank	0.000	06:27	
Control Test	0.079	06:28	
Air Blank	0.000	06:28	
Diagnostics Check	OK	06:28	

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: UMANA

DATE/TIME OF TEST: 08/12/2023 20:01:50

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:59	
Control Test	0.080	19:59	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:00	Subject Test Refused
Breath Volume	0.000	20:00	
Air Blank	0.000	20:00	
Control Test	0.079	20:01	
Air Blank	0.000	20:01	
Diagnostics Check	OK	20:01	

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,DANELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PINTO CARDENAS

DATE/TIME OF TEST: 08/13/2023 02:19:39

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.204	02:15
Breath Volume	3.253	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.205	02:18
Breath Volume	2.281	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,DANELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: VASQUEZ GARCIA

DATE/TIME OF TEST: 08/13/2023 05:17:51

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.078	05:14	
Air Blank	0.000	05:14	
Subject Sample	# 1 0.258	05:15	
Breath Volume	1.429	05:15	
Air Blank		05:16	Improper Sample
Air Blank	0.000	05:16	
Control Test	0.079	05:17	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,DANELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: VASQUEZ GARCIA

DATE/TIME OF TEST: 08/13/2023 05:28:54

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.079	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.266	05:24
Breath Volume	2.152	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:26
Subject Sample # 2	0.257	05:27
Breath Volume	2.242	05:27
Air Blank	0.000	05:27
Control Test	0.077	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELGUERA,L,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,A
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: SAMPAIO

DATE/TIME OF TEST: 08/14/2023 21:28:56

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:26	
Control Test	0.079	21:26	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:28	
Control Test	0.079	21:28	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 08/18/2023 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JOSE,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RUBIO,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ACOSTA FIGUEROA

DATE/TIME OF TEST: 08/20/2023 01:35:02

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.096	01:30
Breath Volume	2.359	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.087	01:33
Breath Volume	1.933	01:33
Air Blank	0.000	01:34
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:35

Data Download Date/Time: 09/01/2023 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,P
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,CAMILA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MULLENDER

DATE/TIME OF TEST: 08/22/2023 01:48:24

Diagnostics Check	OK	01:47	
Air Blank		01:47	Ambient Fail
Air Blank	0.000	01:48	

Data Download Date/Time: 09/01/2023 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,P
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,CAMILA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MULLENDER

DATE/TIME OF TEST: 08/22/2023 01:55:30

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.079	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:53	Subject Test Refused
Breath Volume	0.210	01:53	
Air Blank	0.000	01:54	
Control Test	0.079	01:54	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 09/01/2023 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MESA,ERIC,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HERRER-RUBIO,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 02:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MARTINEZ HERRERA

DATE/TIME OF TEST: 08/27/2023 03:15:46

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.079	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Control Test	0.079	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 09/01/2023 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,ERIC,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERRERA-RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ESCOBAR TORREZ

DATE/TIME OF TEST: 08/27/2023 06:10:32

Diagnostics Check	OK	06:02
Air Blank	0.000	06:02
Control Test	0.080	06:03
Air Blank	0.000	06:03
Subject Sample # 1	0.127	06:05
Breath Volume	1.636	06:05
Air Blank	0.000	06:06
Air Blank	0.000	06:08
Subject Sample # 2	0.122	06:08
Breath Volume	1.496	06:08
Air Blank	0.000	06:09
Control Test	0.078	06:09
Air Blank	0.000	06:10
Diagnostics Check	OK	06:10

Data Download Date/Time: 09/01/2023 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSCHINGER,MICKAEL,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WELSCHINGER,MICKAEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MONTES DE OCA

DATE/TIME OF TEST: 08/27/2023 23:53:51

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.081	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 09/01/2023 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 07/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SIRIA MARTINEZ

DATE/TIME OF TEST: 08/01/2023 23:41:55

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.029	23:37
Breath Volume	2.312	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:39
Subject Sample # 2	0.030	23:40
Breath Volume	2.839	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 08/02/2023 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN JACQUES, WEDLET,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JEAN JACQUES, WEDLET
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SLOSAR

DATE/TIME OF TEST: 08/05/2023 03:17:36

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.079	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2	0.187	
Breath Volume	3.871	03:12	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 3	0.193	
Breath Volume	2.023	03:15	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: FEREIRA,SALISHA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: GONZALEZ PEREZ

DATE/TIME OF TEST: 08/05/2023 05:48:27

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.080	05:42
Air Blank	0.000	05:42
Subject Sample # 1	0.162	05:43
Breath Volume	1.312	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.157	05:46
Breath Volume	1.769	05:46
Air Blank	0.000	05:47
Control Test	0.080	05:47
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ONEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ONEILL,CONNOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: MARTINEZ VERA

DATE/TIME OF TEST: 08/08/2023 17:00:39

Diagnostics Check	OK	16:56	
Air Blank	0.000	16:56	
Control Test	0.080	16:57	
Air Blank	0.000	16:57	
Subject Sample	# 1	16:59	Subject Test Refused
Breath Volume	0.000	16:59	
Air Blank	0.000	16:59	
Control Test	0.079	17:00	
Air Blank	0.000	17:00	
Diagnostics Check	OK	17:00	

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ESPINOZA,OSCAR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MACEDO

DATE/TIME OF TEST: 08/08/2023 23:48:08

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.080	23:40	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:43	No Sample Provided
Breath Volume	0.000	23:43	
Air Blank	0.000	23:44	
Air Blank	0.000	23:46	
Subject Sample	# 2	23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GARCIA,JULIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 08/09/2023 19:26:04

Diagnostics Check	OK	19:22	
Air Blank	0.000	19:23	
Control Test	0.080	19:23	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:24	Subject Test Refused
Breath Volume	0.000	19:24	
Air Blank	0.000	19:25	
Control Test	0.080	19:25	
Air Blank	0.000	19:25	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/15/2023 19:33:53

Diagnostics Check	OK	19:31	
Air Blank	0.000	19:31	
Control Test	0.080	19:31	
Air Blank	0.000	19:32	
Subject Sample	# 1	19:32	Subject Test Refused
Breath Volume	0.000	19:32	
Air Blank	0.000	19:32	
Control Test	0.080	19:33	
Air Blank	0.000	19:33	
Diagnostics Check	OK	19:33	

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: FEREIRA,SALISHA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: SAROTTE

DATE/TIME OF TEST: 08/15/2023 21:34:44

Diagnostics Check	OK	21:24
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Subject Sample # 1	0.220	21:29
Breath Volume	1.480	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.223	21:33
Breath Volume	1.914	21:33
Air Blank	0.000	21:33
Control Test	0.079	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN JACQUES, WEDLET,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JEAN JACQUES, WEDLET
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CLAIREMONT

DATE/TIME OF TEST: 08/15/2023 22:58:16

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.204	22:53
Breath Volume	2.417	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.205	22:56
Breath Volume	2.500	22:56
Air Blank	0.000	22:57
Control Test	0.078	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TSOI,YO-QUEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 08/18/2023 03:52:57

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GARCIA,JULIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BASIC

DATE/TIME OF TEST: 08/19/2023 00:10:02

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.000	00:05
Breath Volume	3.210	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.000	00:08
Breath Volume	3.199	00:08
Air Blank	0.000	00:09
Control Test	0.081	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:10

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN JACQUES, WEDLET,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JEAN JACQUES, WEDLET
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/19/2023 00:25:11

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.080	00:23	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Control Test	0.079	00:24	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 09/13/2023 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JORGE,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: BIEGAJ

DATE/TIME OF TEST: 08/21/2023 23:01:08

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:58	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.079	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 09/13/2023 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CORTES,CHRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 08/22/2023 20:16:12

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:13	
Control Test	0.080	20:13	
Air Blank	0.000	20:14	
Subject Sample	# 1	20:14	Subject Test Refused
Breath Volume	0.000	20:14	
Air Blank	0.000	20:15	
Control Test	0.080	20:15	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:16	

Data Download Date/Time: 09/13/2023 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,DANIEL,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: VALDES,DANIEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 12:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:19

SUBJECT'S LAST NAME: CONRADO

DATE/TIME OF TEST: 08/26/2023 12:42:25

Diagnostics Check	OK	12:39	
Air Blank	0.000	12:40	
Control Test	0.080	12:40	
Air Blank	0.000	12:40	
Subject Sample	# 1	12:41	Subject Test Refused
Breath Volume	0.000	12:41	
Air Blank	0.000	12:41	
Control Test	0.080	12:41	
Air Blank	0.000	12:42	
Diagnostics Check	OK	12:42	

Data Download Date/Time: 09/13/2023 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: JOE

DATE/TIME OF TEST: 08/29/2023 16:28:41

Diagnostics Check	OK	16:24	
Air Blank	0.000	16:25	
Control Test	0.080	16:25	
Air Blank	0.000	16:26	
Subject Sample	# 1	16:27	Subject Test Refused
Breath Volume	0.000	16:27	
Air Blank	0.000	16:27	
Control Test	0.080	16:28	
Air Blank	0.000	16:28	
Diagnostics Check	OK	16:28	

Data Download Date/Time: 09/13/2023 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 08/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: RABIDOU

DATE/TIME OF TEST: 08/29/2023 22:40:16

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.081	22:38	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:39	
Control Test	0.080	22:39	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 09/13/2023 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DARIEN,
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: RODRIGUEZ,DARIEN
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 07/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: PUIG GARCIA

DATE/TIME OF TEST: 08/16/2023 22:54:17

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.081	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.298	22:46
Breath Volume	1.539	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:48
Subject Sample # 2	0.281	22:52
Breath Volume	2.335	22:52
Air Blank	0.000	22:53
Control Test	0.078	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 08/24/2023 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: HOLM

DATE/TIME OF TEST: 08/02/2023 01:39:03

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.120	01:31
Breath Volume	1.773	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.146	01:34
Breath Volume	4.335	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 3	0.139	01:37
Breath Volume	1.960	01:37
Air Blank	0.000	01:38
Control Test	0.077	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:39

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER, CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: RODRIGUEZ HEREDIA

DATE/TIME OF TEST: 08/02/2023 01:58:29

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.078	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.105	01:53
Breath Volume	1.175	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.105	01:56
Breath Volume	1.703	01:56
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: O REGGIO

DATE/TIME OF TEST: 08/02/2023 04:04:46

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.104	03:59
Breath Volume	1.734	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.095	04:03
Breath Volume	1.113	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HADLEY, NICHOLAS
ARRESTING OFFICER AGENCY: Manalapan Pub Safety
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/02/2023 22:32:56

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:18	
Control Test	0.080	22:18	
Air Blank	0.000	22:19	
Subject Sample	# 1 0.000	22:20	
Breath Volume	1.320	22:20	
Air Blank	0.000	22:20	
Air Blank	0.000	22:22	
Subject Sample	# 2 0.000	22:25	Volume Not Met
Breath Volume	0.031	22:25	
Air Blank	0.000	22:26	
Air Blank	0.000	22:28	
Subject Sample	# 3 0.000	22:31	Volume Not Met
Breath Volume	0.394	22:31	
Air Blank	0.000	22:32	
Control Test	0.079	22:32	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE, RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: ELLIOT

DATE/TIME OF TEST: 08/03/2023 15:47:06

Diagnostics Check	OK	15:39
Air Blank	0.000	15:40
Control Test	0.079	15:40
Air Blank	0.000	15:40
Subject Sample # 1	0.132	15:41
Breath Volume	1.203	15:41
Air Blank	0.000	15:42
Air Blank	0.000	15:43
Subject Sample # 2	0.130	15:45
Breath Volume	1.718	15:45
Air Blank	0.000	15:46
Control Test	0.078	15:46
Air Blank	0.000	15:47
Diagnostics Check	OK	15:47

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER, MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: VALUS

DATE/TIME OF TEST: 08/03/2023 22:29:39

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.000	22:25
Breath Volume	3.347	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.000	22:28
Breath Volume	3.328	22:28
Air Blank	0.000	22:28
Control Test	0.080	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REID,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: TOMAS

DATE/TIME OF TEST: 08/05/2023 03:03:35

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.239	02:58
Breath Volume	2.218	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.247	03:01
Breath Volume	2.398	03:01
Air Blank	0.000	03:02
Control Test	0.076	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN,JOHNATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: SMART

DATE/TIME OF TEST: 08/05/2023 05:32:41

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.078	05:27
Air Blank	0.000	05:27
Subject Sample # 1	0.120	05:28
Breath Volume	2.484	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.123	05:31
Breath Volume	1.898	05:31
Air Blank	0.000	05:31
Control Test	0.079	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: SMART

DATE/TIME OF TEST: 08/05/2023 06:07:27

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.078	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.104	06:03
Breath Volume	2.656	06:03
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.102	06:05
Breath Volume	2.230	06:05
Air Blank	0.000	06:06
Control Test	0.078	06:06
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: REZANIA

DATE/TIME OF TEST: 08/05/2023 23:22:07

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:08	
Control Test	0.079	23:08	
Air Blank	0.000	23:08	
Subject Sample	# 1 0.165	23:12	Volume Not Met
Breath Volume	0.035	23:12	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample	# 2 0.160	23:16	
Breath Volume	1.476	23:16	
Air Blank	0.000	23:17	
Air Blank	0.000	23:19	
Subject Sample	# 3 0.156	23:20	
Breath Volume	1.750	23:20	
Air Blank	0.000	23:21	
Control Test	0.077	23:21	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU, YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MARTIN RODRIGUEZ

DATE/TIME OF TEST: 08/06/2023 01:25:46

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.078	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1 0.141	01:21	Volume Not Met
Breath Volume	0.070	01:21	
Air Blank	0.000	01:22	
Air Blank	0.000	01:24	
Subject Sample	# 2	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Control Test	0.077	01:25	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: CARRILLO PABLO

DATE/TIME OF TEST: 08/06/2023 01:46:13

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.077	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.181	01:41
Breath Volume	2.261	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.177	01:44
Breath Volume	1.253	01:44
Air Blank	0.000	01:45
Control Test	0.077	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: MENDEZ MALDONADO

DATE/TIME OF TEST: 08/06/2023 22:16:59

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.078	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.220	22:12
Breath Volume	2.136	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.215	22:15
Breath Volume	2.132	22:15
Air Blank	0.000	22:16
Control Test	0.078	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALKER,JODI
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 08/07/2023 01:05:23

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.079	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1 0.121	00:58	Volume Not Met
Breath Volume	0.046	00:58	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample	# 2 0.178	01:03	Volume Not Met
Breath Volume	0.503	01:03	
Air Blank	0.000	01:04	
Control Test	0.078	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALKER,JODI
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 08/07/2023 01:20:15

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.077	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1 0.193	01:09	
Breath Volume	1.925	01:09	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample	# 2 0.158	01:15	Volume Not Met
Breath Volume	1.078	01:15	
Air Blank	0.000	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 3 0.184	01:18	
Breath Volume	1.859	01:18	
Air Blank	0.000	01:19	
Control Test	0.076	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: QUINTANILLA

DATE/TIME OF TEST: 08/09/2023 02:00:30

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.207	01:55
Breath Volume	1.414	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.203	01:58
Breath Volume	1.148	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: BALTAZAR ZACARIA

DATE/TIME OF TEST: 08/10/2023 01:03:45

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.079	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1 0.142	00:58	Volume Not Met
Breath Volume	0.050	00:58	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample	# 2	01:02	Subject Test Refused
Breath Volume	0.039	01:02	
Air Blank	0.000	01:02	
Control Test	0.077	01:03	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COUNTS,RUSSELL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 08/11/2023 00:12:39

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.077	00:07
Breath Volume	1.324	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.079	00:11
Breath Volume	1.519	00:11
Air Blank	0.000	00:11
Control Test	0.078	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: JIMENEZ GARCIA

DATE/TIME OF TEST: 08/11/2023 01:11:27

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.078	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.194	01:06
Breath Volume	2.175	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.204	01:09
Breath Volume	2.414	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: PEREZ JOYA

DATE/TIME OF TEST: 08/11/2023 05:38:11

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.078	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.152	05:33
Breath Volume	4.359	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.150	05:36
Breath Volume	4.015	05:36
Air Blank	0.000	05:37
Control Test	0.079	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 08/11/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCDONOUGH, MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ENGLER

DATE/TIME OF TEST: 08/05/2023 03:28:20

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1 0.158	03:22	Volume Not Met
Breath Volume	0.320	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2	03:26	Subject Test Refused
Breath Volume	0.402	03:26	
Air Blank	0.000	03:27	
Control Test	0.075	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 08/11/2023 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PASTRANA,XAVIER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: ALDEN

DATE/TIME OF TEST: 08/05/2023 10:28:12

Diagnostics Check	OK	10:21
Air Blank	0.000	10:21
Control Test	0.079	10:21
Air Blank	0.000	10:22
Subject Sample # 1	0.000	10:23
Breath Volume	1.464	10:23
Air Blank	0.000	10:23
Air Blank	0.000	10:25
Subject Sample # 2	0.000	10:26
Breath Volume	1.589	10:26
Air Blank	0.000	10:27
Control Test	0.076	10:27
Air Blank	0.000	10:28
Diagnostics Check	OK	10:28

Data Download Date/Time: 08/11/2023 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: LOPEZ MARTINEZ

DATE/TIME OF TEST: 08/06/2023 19:03:12

Diagnostics Check	OK	18:56
Air Blank	0.000	18:57
Control Test	0.079	18:57
Air Blank	0.000	18:57
Subject Sample # 1	0.256	18:58
Breath Volume	2.046	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:00
Subject Sample # 2	0.254	19:01
Breath Volume	1.820	19:01
Air Blank	0.000	19:02
Control Test	0.077	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 08/11/2023 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,DAYRON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 08/07/2023 10:44:29

Diagnostics Check	OK	10:36
Air Blank	0.000	10:36
Control Test	0.078	10:36
Air Blank	0.000	10:37
Subject Sample # 1	0.170	10:38
Breath Volume	1.277	10:38
Air Blank	0.000	10:39
Air Blank	0.000	10:41
Subject Sample # 2	0.169	10:42
Breath Volume	1.425	10:42
Air Blank	0.000	10:43
Control Test	0.077	10:43
Air Blank	0.000	10:44
Diagnostics Check	OK	10:44

Data Download Date/Time: 08/11/2023 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: ODDO

DATE/TIME OF TEST: 08/10/2023 22:48:10

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.080	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:47	
Control Test	0.081	22:47	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 08/11/2023 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 08/13/2023 01:26:20

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.250	01:19
Breath Volume	2.429	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:21
Subject Sample # 2	0.245	01:24
Breath Volume	2.042	01:24
Air Blank	0.000	01:25
Control Test	0.076	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 08/13/2023 02:19:09

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.077	02:07	
Air Blank	0.000	02:07	
Subject Sample # 1	0.133	02:08	
Breath Volume	1.589	02:08	
Air Blank	0.000	02:09	
Air Blank	0.000	02:10	
Subject Sample # 2	0.089	02:14	Volume Not Met
Breath Volume	0.136	02:14	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample # 3	0.124	02:17	
Breath Volume	1.656	02:17	
Air Blank	0.000	02:18	
Control Test	0.074	02:18	Control Outside Tolerance
Air Blank	0.000	02:19	

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU, YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 08/13/2023 02:51:09

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.077	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.120	02:46
Breath Volume	1.765	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.124	02:49
Breath Volume	1.773	02:49
Air Blank	0.000	02:50
Control Test	0.077	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: FERNANDEZ PEREZ

DATE/TIME OF TEST: 08/13/2023 05:30:29

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.179	05:25
Breath Volume	1.160	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.182	05:28
Breath Volume	2.070	05:28
Air Blank	0.000	05:29
Control Test	0.078	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 08/14/2023 06:03:35

Diagnostics Check	OK	05:57
Air Blank	0.000	05:57
Control Test	0.079	05:57
Air Blank	0.000	05:58
Subject Sample # 1	0.157	05:58
Breath Volume	2.890	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.160	06:01
Breath Volume	3.750	06:01
Air Blank	0.000	06:02
Control Test	0.078	06:02
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:53

SUBJECT'S LAST NAME: MONEY

DATE/TIME OF TEST: 08/16/2023 15:33:54

Diagnostics Check	OK	15:27
Air Blank	0.000	15:27
Control Test	0.079	15:27
Air Blank	0.000	15:28
Subject Sample # 1	0.299	15:29
Breath Volume	1.542	15:29
Air Blank	0.000	15:29
Air Blank	0.000	15:31
Subject Sample # 2	0.302	15:32
Breath Volume	1.769	15:32
Air Blank	0.000	15:32
Control Test	0.077	15:33
Air Blank	0.000	15:33
Diagnostics Check	OK	15:33

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE, RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 08/17/2023 22:05:16

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.208	22:00
Breath Volume	1.824	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.204	22:03
Breath Volume	1.550	22:03
Air Blank	0.000	22:04
Control Test	0.077	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALES LOPEZ,KEVIN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DELEON RAMIREZ

DATE/TIME OF TEST: 08/18/2023 03:34:06

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.055	03:29
Breath Volume	2.539	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.050	03:32
Breath Volume	1.593	03:32
Air Blank	0.000	03:33
Control Test	0.077	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: RICKARDS

DATE/TIME OF TEST: 08/18/2023 22:26:25

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.187	22:21
Breath Volume	2.132	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.187	22:24
Breath Volume	2.332	22:24
Air Blank	0.000	22:25
Control Test	0.075	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LALJIE,SEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: ANEQ

DATE/TIME OF TEST: 08/19/2023 12:18:19

Diagnostics Check	OK	12:11
Air Blank	0.000	12:12
Control Test	0.078	12:12
Air Blank	0.000	12:13
Subject Sample # 1	0.008	12:13
Breath Volume	2.414	12:13
Air Blank	0.000	12:14
Air Blank	0.000	12:16
Subject Sample # 2	0.000	12:16
Breath Volume	2.078	12:16
Air Blank	0.000	12:17
Control Test	0.075	12:17
Air Blank	0.000	12:18
Diagnostics Check	OK	12:18

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REA,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: PAYTON

DATE/TIME OF TEST: 08/19/2023 14:53:38

Diagnostics Check	OK	14:47
Air Blank	0.000	14:48
Control Test	0.078	14:48
Air Blank	0.000	14:48
Subject Sample # 1	0.000	14:49
Breath Volume	3.558	14:49
Air Blank	0.000	14:49
Air Blank	0.000	14:51
Subject Sample # 2	0.000	14:52
Breath Volume	4.343	14:52
Air Blank	0.000	14:52
Control Test	0.077	14:52
Air Blank	0.000	14:53
Diagnostics Check	OK	14:53

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: ROBLERO GONZALEZ

DATE/TIME OF TEST: 08/20/2023 00:51:26

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.240	00:46
Breath Volume	1.980	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.237	00:49
Breath Volume	1.605	00:49
Air Blank	0.000	00:50
Control Test	0.076	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: CAMPOSECO CARMELO

DATE/TIME OF TEST: 08/20/2023 05:08:06

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.025	05:03
Breath Volume	2.101	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.022	05:06
Breath Volume	2.039	05:06
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DE COSKEY

DATE/TIME OF TEST: 08/24/2023 01:24:39

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.219	01:18
Breath Volume	3.328	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.218	01:22
Breath Volume	2.898	01:22
Air Blank	0.000	01:23
Control Test	0.077	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/26/2023 01:54:40

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.117	01:49
Breath Volume	2.457	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.118	01:52
Breath Volume	2.714	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: CRISTOBAL NICOLAS

DATE/TIME OF TEST: 08/26/2023 21:39:14

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.079	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.200	21:34
Breath Volume	2.164	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:36
Subject Sample # 2	0.202	21:37
Breath Volume	2.023	21:37
Air Blank	0.000	21:38
Control Test	0.077	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MOLNAR

DATE/TIME OF TEST: 08/31/2023 00:52:06

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Subject Sample	# 1 0.275	00:38	
Breath Volume	1.289	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2 0.234	00:44	Volume Not Met
Breath Volume	0.746	00:44	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 3 0.255	00:50	Volume Not Met
Breath Volume	0.464	00:50	
Air Blank	0.000	00:51	
Control Test	0.075	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 09/15/2023 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 08/31/2023 22:19:38

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:08	
Control Test	0.079	22:08	
Air Blank	0.000	22:09	
Subject Sample	# 1 0.052	22:12	Volume Not Met
Breath Volume	0.347	22:12	
Air Blank	0.000	22:13	
Air Blank	0.000	22:14	
Subject Sample	# 2 0.045	22:17	Volume Not Met
Breath Volume	0.707	22:17	
Air Blank	0.000	22:18	
Control Test	0.076	22:19	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 09/15/2023 15:57