

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARNETT,JAMES,C

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BARNETT,JAMES

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: GROTH

DATE/TIME OF TEST: 09/01/2023 02:22:39

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.080	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 1 0.209	02:14	Volume Not Met
Breath Volume	0.320	02:14	
Air Blank	0.000	02:15	
Air Blank	0.000	02:16	
Subject Sample	# 2 0.212	02:18	No .020 agreement
Breath Volume	2.035	02:18	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample	# 3 0.170	02:21	No .020 agreement
Breath Volume	1.226	02:21	
Air Blank	0.000	02:21	
Control Test	0.077	02:22	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 09/27/2023 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERINO, WESLEY, G

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FERINO, WESLEY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 20:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 09/01/2023 21:09:59

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.080	21:04
Air Blank	0.000	21:04
Subject Sample # 1	0.000	21:05
Breath Volume	1.238	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.000	21:08
Breath Volume	1.261	21:08
Air Blank	0.000	21:09
Control Test	0.080	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 09/27/2023 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERINO, WESLEY, G

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FERINO, WESLEY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 09/03/2023 00:04:03

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.080	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.120	23:59
Breath Volume	2.652	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.122	00:02
Breath Volume	3.304	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:04

Data Download Date/Time: 09/27/2023 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERINO, WESLEY, G

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FERINO, WESLEY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 09/08/2023 02:13:12

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.219	02:07
Breath Volume	1.187	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.204	02:11
Breath Volume	1.343	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 09/27/2023 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: COON,D

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ELDEMIRE

DATE/TIME OF TEST: 09/14/2023 00:45:12

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.111	00:39
Breath Volume	1.289	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.112	00:43
Breath Volume	1.339	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 09/27/2023 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: SMITH,J

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:59

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 09/22/2023 02:36:56

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.115	02:32
Breath Volume	1.281	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.115	02:35
Breath Volume	1.281	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 09/27/2023 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEPP,V,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: STEPP,V

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 19:53

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: STALEY

DATE/TIME OF TEST: 09/23/2023 20:45:55

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.250	20:41
Breath Volume	2.937	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:43
Subject Sample # 2	0.251	20:44
Breath Volume	2.523	20:44
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 09/27/2023 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,T

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 09/25/2023 04:31:35

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.278	04:26
Breath Volume	2.261	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.283	04:29
Breath Volume	2.007	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 09/27/2023 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DELIGNY,JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GIL RAMIREZ

DATE/TIME OF TEST: 09/01/2023 01:31:30

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.229	01:26
Breath Volume	1.496	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.234	01:29
Breath Volume	1.746	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 09/18/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,JILLIAN,F
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KNAPP,MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: REYES JUAREZ

DATE/TIME OF TEST: 09/01/2023 03:58:56

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.007	03:54
Breath Volume	3.429	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.000	03:57
Breath Volume	4.269	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 09/18/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HURST,THOMAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:22

SUBJECT'S LAST NAME: LUNKEN

DATE/TIME OF TEST: 09/03/2023 11:18:17

Diagnostics Check	OK	11:14	
Air Blank	0.000	11:15	
Control Test	0.079	11:15	
Air Blank	0.000	11:16	
Subject Sample	# 1	11:16	Slope Not Met
Breath Volume	0.761	11:16	
Air Blank	0.000	11:17	
Control Test	0.080	11:17	
Air Blank	0.000	11:18	
Diagnostics Check	OK	11:18	

Data Download Date/Time: 09/18/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HURST,THOMAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:22

SUBJECT'S LAST NAME: LUNKEN

DATE/TIME OF TEST: 09/03/2023 11:27:14

Diagnostics Check	OK	11:20
Air Blank	0.000	11:20
Control Test	0.079	11:21
Air Blank	0.000	11:21
Subject Sample # 1	0.363	11:22
Breath Volume	1.789	11:22
Air Blank	0.000	11:23
Air Blank	0.000	11:24
Subject Sample # 2	0.362	11:25
Breath Volume	1.390	11:25
Air Blank	0.000	11:26
Control Test	0.078	11:26
Air Blank	0.000	11:27
Diagnostics Check	OK	11:27

Data Download Date/Time: 09/18/2023 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KAYLEE,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WALKER,KAYLEE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/03/2023 19:58:18

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.081	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.000	19:53
Breath Volume	2.414	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.000	19:56
Breath Volume	3.398	19:56
Air Blank	0.000	19:57
Control Test	0.080	19:57
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 09/18/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN-PAUL,M,J
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MORALES,C
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: GUTIERREZ MORENO

DATE/TIME OF TEST: 09/04/2023 06:59:43

Diagnostics Check	OK	06:53
Air Blank	0.000	06:53
Control Test	0.081	06:54
Air Blank	0.000	06:54
Subject Sample # 1	0.159	06:55
Breath Volume	1.625	06:55
Air Blank	0.000	06:55
Air Blank	0.000	06:57
Subject Sample # 2	0.152	06:58
Breath Volume	1.824	06:58
Air Blank	0.000	06:58
Control Test	0.081	06:59
Air Blank	0.000	06:59
Diagnostics Check	OK	06:59

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGES,JARRETT,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BRIDGES,JARRETT
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:36

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 09/04/2023 09:58:53

Diagnostics Check	OK	09:57	
Air Blank	0.000	09:58	
Control Test	0.080	09:58	RFI Detect
Air Blank	0.000	09:58	

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGES,JARRETT,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BRIDGES,JARRETT
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:36

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 09/04/2023 10:02:24

Diagnostics Check	OK	10:00	
Air Blank	0.000	10:01	
Control Test	0.152	10:01	RFI Detect
Air Blank	0.000	10:02	

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: RIVERA GRAU

DATE/TIME OF TEST: 09/05/2023 04:27:01

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.166	04:22
Breath Volume	1.855	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.161	04:25
Breath Volume	2.277	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:27

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 09/05/2023 23:22:23

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.153	23:17
Breath Volume	1.882	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample # 2	0.167	23:20
Breath Volume	2.714	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SOZIO

DATE/TIME OF TEST: 09/06/2023 01:41:19

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.134	01:36
Breath Volume	2.570	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.131	01:39
Breath Volume	2.195	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: DUNANN

DATE/TIME OF TEST: 09/07/2023 22:23:17

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.081	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.148	22:18
Breath Volume	2.265	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.145	22:21
Breath Volume	2.070	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWTON,JENNA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: NEWTON,JENNA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: DELAURENTIS

DATE/TIME OF TEST: 09/08/2023 18:42:04

Diagnostics Check	OK	18:35
Air Blank	0.000	18:35
Control Test	0.081	18:35
Air Blank	0.000	18:36
Subject Sample # 1	0.064	18:37
Breath Volume	2.046	18:37
Air Blank	0.000	18:38
Air Blank	0.000	18:39
Subject Sample # 2	0.061	18:40
Breath Volume	2.593	18:40
Air Blank	0.000	18:41
Control Test	0.078	18:41
Air Blank	0.000	18:41
Diagnostics Check	OK	18:42

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/08/2023 22:54:41

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.081	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.126	22:49
Breath Volume	1.910	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.122	22:53
Breath Volume	1.703	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY,JORDAN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: COHEN,ZACH
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: KETCHMARK

DATE/TIME OF TEST: 09/10/2023 00:24:45

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.297	00:19
Breath Volume	2.585	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.294	00:22
Breath Volume	2.835	00:22
Air Blank	0.000	00:23
Control Test	0.077	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,LINDSAY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,ASHLEY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:31

SUBJECT'S LAST NAME: TOJIN

DATE/TIME OF TEST: 09/10/2023 09:03:36

Diagnostics Check	OK	08:53
Air Blank	0.000	08:54
Control Test	0.080	08:54
Air Blank	0.000	08:55
Subject Sample # 1	0.139	08:58
Breath Volume	1.640	08:58
Air Blank	0.000	08:58
Air Blank	0.000	09:00
Subject Sample # 2	0.141	09:01
Breath Volume	1.765	09:01
Air Blank	0.000	09:02
Control Test	0.077	09:02
Air Blank	0.000	09:03
Diagnostics Check	OK	09:03

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES DE OCA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CLEMENTS,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: MAIN

DATE/TIME OF TEST: 09/12/2023 14:19:59

Diagnostics Check	OK	14:19	
Air Blank		14:19	RFI Detect
Air Blank	0.000	14:19	

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES DE OCA, ALEXANDER,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CLEMENTS, T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: MAIN

DATE/TIME OF TEST: 09/12/2023 14:27:08

Diagnostics Check	OK	14:24	
Air Blank	0.000	14:24	
Control Test	0.080	14:24	
Air Blank	0.000	14:25	
Subject Sample	# 1	14:25	RFI Detect
Breath Volume	0.000	14:25	
Air Blank	0.000	14:26	
Control Test	0.080	14:26	
Air Blank	0.000	14:27	
Diagnostics Check	OK	14:27	

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMEN,LOGAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CLEMENTS,T.
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: MAIN

DATE/TIME OF TEST: 09/12/2023 14:35:01

Diagnostics Check	OK	14:28
Air Blank	0.000	14:29
Control Test	0.081	14:29
Air Blank	0.000	14:29
Subject Sample # 1	0.000	14:30
Breath Volume	2.480	14:30
Air Blank	0.000	14:31
Air Blank	0.000	14:33
Subject Sample # 2	0.000	14:33
Breath Volume	2.566	14:33
Air Blank	0.000	14:34
Control Test	0.080	14:34
Air Blank	0.000	14:34
Diagnostics Check	OK	14:35

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: KING,JONATHAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: HUFFMAN

DATE/TIME OF TEST: 09/13/2023 22:14:29

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.081	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.000	22:10
Breath Volume	3.410	22:10
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.000	22:13
Breath Volume	2.679	22:13
Air Blank	0.000	22:13
Control Test	0.079	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: KING,JONATHAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: HUFFMAN

DATE/TIME OF TEST: 09/13/2023 22:22:17

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.081	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.000	22:17
Breath Volume	2.347	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.000	22:20
Breath Volume	2.316	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FISHMAN,ALEX
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/14/2023 23:46:20

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.228	23:36
Breath Volume	1.402	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:38
Subject Sample # 2	0.267	23:40
Breath Volume	1.921	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 3	0.264	23:44
Breath Volume	2.160	23:44
Air Blank	0.000	23:45
Control Test	0.078	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 09/18/2023 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: GILLILAND

DATE/TIME OF TEST: 09/18/2023 23:47:15

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.185	23:42
Breath Volume	3.929	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.186	23:45
Breath Volume	4.421	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 10/18/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUM,JACOB,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LIMDEMANN,ANTHONY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: DUBY

DATE/TIME OF TEST: 09/19/2023 05:14:08

Diagnostics Check	OK	05:07
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.118	05:09
Breath Volume	1.406	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.121	05:12
Breath Volume	1.863	05:12
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 10/18/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,JILLIAN,F
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KNAPP,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: ALTON

DATE/TIME OF TEST: 09/20/2023 01:04:12

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.000	00:59
Breath Volume	3.847	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.000	01:02
Breath Volume	4.035	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 10/18/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEOANRD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WEBDALE

DATE/TIME OF TEST: 09/23/2023 02:37:46

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Subject Sample # 1	0.217	02:33	
Breath Volume	2.878	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
		02:36	RFI Detect
Air Blank	0.000	02:36	
Control Test	0.082	02:37	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 10/18/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEOANRD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WEBDALE

DATE/TIME OF TEST: 09/23/2023 02:45:12

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.196	02:40
Breath Volume	2.074	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.202	02:43
Breath Volume	2.488	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 10/18/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY,JORDAN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VULICEVIC,RADOSAV
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HOMOLKA

DATE/TIME OF TEST: 09/23/2023 22:51:10

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.082	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.135	22:46
Breath Volume	2.578	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.128	22:49
Breath Volume	2.710	22:49
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 10/18/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: JUAN

DATE/TIME OF TEST: 09/24/2023 01:00:04

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.195	00:55
Breath Volume	4.023	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.194	00:58
Breath Volume	4.417	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	01:00

Data Download Date/Time: 10/18/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KARSNICK,KEVIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ANDRES JOSE

DATE/TIME OF TEST: 09/24/2023 03:13:34

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.163	03:08
Breath Volume	1.996	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.160	03:11
Breath Volume	1.984	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 10/18/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY, JORDAN, T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CRUZ, WINSTON
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 09/24/2023 04:11:01

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.187	04:06
Breath Volume	1.953	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:08
Subject Sample # 2	0.191	04:09
Breath Volume	1.125	04:09
Air Blank	0.000	04:10
Control Test	0.078	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:11

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 09/25/2023 03:57:41

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.197	03:53
Breath Volume	2.355	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.194	03:56
Breath Volume	1.953	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: VAN METER

DATE/TIME OF TEST: 09/26/2023 03:09:28

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.145	03:04
Breath Volume	1.722	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.132	03:07
Breath Volume	1.382	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PARTRIGE,LANDON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: TOP

DATE/TIME OF TEST: 09/26/2023 20:06:07

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:03	
Control Test	0.080	20:03	
Air Blank	0.000	20:03	
Subject Sample	# 1	20:04	Slope Not Met
Breath Volume	1.171	20:04	
Air Blank	0.000	20:05	
Control Test	0.080	20:05	
Air Blank	0.000	20:06	
Diagnostics Check	OK	20:06	

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PARTRIDGE,L
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: TOP

DATE/TIME OF TEST: 09/26/2023 21:12:31

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.081	21:07
Air Blank	0.000	21:07
Subject Sample # 1	0.234	21:08
Breath Volume	1.765	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.233	21:10
Breath Volume	2.105	21:10
Air Blank	0.000	21:11
Control Test	0.080	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,L
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MOUTON

DATE/TIME OF TEST: 09/26/2023 21:25:18

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:17	
Control Test	0.080	21:17	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:21	No Sample Provided
Breath Volume	0.000	21:21	
Air Blank	0.000	21:21	
Air Blank	0.000	21:23	
Subject Sample	# 2	21:23	Subject Test Refused
Breath Volume	0.000	21:23	
Air Blank	0.000	21:24	
Control Test	0.080	21:24	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 09/27/2023 23:03:30

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.000	22:59
Breath Volume	2.152	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.000	23:02
Breath Volume	1.984	23:02
Air Blank	0.000	23:02
Control Test	0.081	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: APARICIO

DATE/TIME OF TEST: 09/28/2023 01:48:00

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.104	01:43
Breath Volume	3.261	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample # 2	0.104	01:46
Breath Volume	4.285	01:46
Air Blank	0.000	01:47
Control Test	0.081	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:48

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: CAPEZZUTO

DATE/TIME OF TEST: 09/29/2023 22:23:01

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:22	
Control Test	0.080	22:22	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JACKSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WILL

DATE/TIME OF TEST: 09/29/2023 23:12:22

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.081	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.121	23:07
Breath Volume	2.734	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.113	23:10
Breath Volume	2.195	23:10
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SETTER

DATE/TIME OF TEST: 09/30/2023 03:22:22

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:14	No Sample Provided
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Air Blank	0.000	03:17	
Subject Sample	# 2	0.188	03:17
Breath Volume	2.500	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 3	0.190	03:20
Breath Volume	1.996	03:20	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: LINARES

DATE/TIME OF TEST: 09/30/2023 04:36:04

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.141	04:31
Breath Volume	2.269	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.128	04:34
Breath Volume	2.421	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:36

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: SANTAELLA DELGADO

DATE/TIME OF TEST: 09/30/2023 05:15:44

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.137	05:11
Breath Volume	1.593	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.146	05:14
Breath Volume	2.695	05:14
Air Blank	0.000	05:14
Control Test	0.080	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 10/18/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BERNIER,MAGDIEL
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SERGEANT

DATE/TIME OF TEST: 09/30/2023 05:51:41

Diagnostics Check	OK	05:45
Air Blank	0.000	05:45
Control Test	0.080	05:45
Air Blank	0.000	05:46
Subject Sample # 1	0.123	05:47
Breath Volume	4.187	05:47
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.119	05:50
Breath Volume	4.492	05:50
Air Blank	0.000	05:50
Control Test	0.079	05:51
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 10/18/2023 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 09/02/2023 22:28:35

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.079	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:27	Subject Test Refused
Breath Volume	0.000	22:27	
Air Blank	0.000	22:27	
Control Test	0.078	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 09/25/2023 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,GARY,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CALDWELL,GARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FARIA

DATE/TIME OF TEST: 09/03/2023 02:16:36

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.078	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 09/25/2023 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: RONGE,D
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:11

SUBJECT'S LAST NAME: MORIN

DATE/TIME OF TEST: 09/03/2023 14:38:57

Diagnostics Check	OK	14:32
Air Blank	0.000	14:32
Control Test	0.079	14:33
Air Blank	0.000	14:33
Subject Sample # 1	0.000	14:34
Breath Volume	2.792	14:34
Air Blank	0.000	14:35
Air Blank	0.000	14:37
Subject Sample # 2	0.000	14:37
Breath Volume	2.132	14:37
Air Blank	0.000	14:38
Control Test	0.079	14:38
Air Blank	0.000	14:38
Diagnostics Check	OK	14:38

Data Download Date/Time: 09/25/2023 08:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SKRZPEK

DATE/TIME OF TEST: 09/16/2023 03:02:18

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.078	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 09/25/2023 08:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOEDE, BRANDON, W
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: RONGE, DEREK
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/16/2023 08:25:02

Diagnostics Check	OK	08:20	
Air Blank	0.000	08:20	
Control Test	0.079	08:21	
Air Blank	0.000	08:21	
Subject Sample	# 1	08:23	Subject Test Refused
Breath Volume	0.113	08:23	
Air Blank	0.000	08:24	
Control Test	0.079	08:24	
Air Blank	0.000	08:24	
Diagnostics Check	OK	08:25	

Data Download Date/Time: 09/25/2023 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOEDE, BRANDON, W
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: RONGE, DEREK
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/16/2023 08:38:08

Diagnostics Check	OK	08:26	
Air Blank	0.000	08:27	
Control Test	0.078	08:27	
Air Blank	0.000	08:28	
Subject Sample	# 1 0.000	08:31	Volume Not Met
Breath Volume	0.363	08:31	
Air Blank	0.000	08:31	
Air Blank	0.000	08:33	
Subject Sample	# 2 0.000	08:36	Volume Not Met
Breath Volume	0.136	08:36	
Air Blank	0.000	08:37	
Control Test	0.078	08:37	
Air Blank	0.000	08:38	
Diagnostics Check	OK	08:38	

Data Download Date/Time: 09/25/2023 08:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/16/2023 20:08:56

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.079	20:03
Air Blank	0.000	20:03
Subject Sample # 1	0.222	20:04
Breath Volume	2.734	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:06
Subject Sample # 2	0.208	20:07
Breath Volume	2.539	20:07
Air Blank	0.000	20:08
Control Test	0.078	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 09/25/2023 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/17/2023 00:51:55

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.078	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 09/25/2023 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MEJIA ORTEGA

DATE/TIME OF TEST: 09/17/2023 04:26:38

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.078	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Control Test	0.078	04:26	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 09/25/2023 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/23/2023 04:20:04

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.119	04:15
Breath Volume	1.976	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.114	04:18
Breath Volume	2.289	04:18
Air Blank	0.000	04:19
Control Test	0.078	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:20

Data Download Date/Time: 09/25/2023 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: CADENA

DATE/TIME OF TEST: 09/23/2023 19:10:15

Diagnostics Check	OK	19:01
Air Blank	0.000	19:01
Control Test	0.079	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.251	19:05
Breath Volume	1.605	19:05
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.249	19:08
Breath Volume	2.058	19:08
Air Blank	0.000	19:09
Control Test	0.075	19:09
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 09/25/2023 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: QUESADA

DATE/TIME OF TEST: 09/24/2023 05:49:33

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.079	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.203	05:44
Breath Volume	1.515	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.195	05:47
Breath Volume	1.597	05:47
Air Blank	0.000	05:48
Control Test	0.077	05:48
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 09/25/2023 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 09/24/2023 18:24:04

Diagnostics Check	OK	18:12	
Air Blank	0.000	18:13	
Control Test	0.078	18:13	
Air Blank	0.000	18:13	
Subject Sample	# 1	18:17	No Sample Provided
Breath Volume	0.000	18:17	
Air Blank	0.000	18:17	
Air Blank	0.000	18:19	
Subject Sample	# 2	18:22	No Sample Provided
Breath Volume	0.000	18:22	
Air Blank	0.000	18:23	
Control Test	0.078	18:23	
Air Blank	0.000	18:23	
Diagnostics Check	OK	18:24	

Data Download Date/Time: 09/25/2023 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,TAYLOR,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HERNANDEZ,JONATHAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 09/24/2023 21:49:02

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:38	
Control Test	0.079	21:38	
Air Blank	0.000	21:38	
Subject Sample	# 1	21:42	No Sample Provided
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Air Blank	0.000	21:44	
Subject Sample	# 2	21:47	No Sample Provided
Breath Volume	0.000	21:47	
Air Blank	0.000	21:48	
Control Test	0.078	21:48	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 09/25/2023 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESSLER,GREYSON,L
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GRIGGS,J
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: KOVACS

DATE/TIME OF TEST: 09/26/2023 01:04:16

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.169	00:59
Breath Volume	1.437	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.172	01:02
Breath Volume	2.707	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 10/23/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PAULSEN

DATE/TIME OF TEST: 09/28/2023 04:16:58

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.000	04:12
Breath Volume	2.621	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.000	04:15
Breath Volume	1.828	04:15
Air Blank	0.000	04:16
Control Test	0.078	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 10/23/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 09/29/2023 03:45:51

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.078	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.079	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 10/23/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,GARY,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CALDWELL,GARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/30/2023 01:09:26

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.078	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.150	01:04
Breath Volume	2.421	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.152	01:07
Breath Volume	1.210	01:07
Air Blank	0.000	01:08
Control Test	0.076	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 10/23/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 09/30/2023 03:31:33

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.000	03:27
Breath Volume	2.273	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.000	03:30
Breath Volume	2.218	03:30
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 10/23/2023 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS, STEWART, A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: PHILLIPS, STEWART
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: BASTIDA

DATE/TIME OF TEST: 09/03/2023 01:28:43

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.112	01:23
Breath Volume	2.519	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.113	01:27
Breath Volume	2.539	01:27
Air Blank	0.000	01:27
Control Test	0.078	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,SETH,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RODRIGUEZ,XAVIER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: HERNANDEZ-CABANAS

DATE/TIME OF TEST: 09/05/2023 23:47:31

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.000	23:42
Breath Volume	3.000	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.000	23:46
Breath Volume	2.476	23:46
Air Blank	0.000	23:46
Control Test	0.077	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS, STEWART, A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: PHILLIPS, STEWART
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ROMNEY

DATE/TIME OF TEST: 09/08/2023 03:10:03

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 2	03:08	No Sample Provided
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMA,N,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RODRIGUEZ,X
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 09/08/2023 20:02:52

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.176	19:58
Breath Volume	2.121	19:58
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.157	20:01
Breath Volume	2.363	20:01
Air Blank	0.000	20:01
Control Test	0.078	20:02
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS, STEWART, A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: PHILLIPS, STEWART
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: FISHEL

DATE/TIME OF TEST: 09/09/2023 02:45:09

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.168	02:40
Breath Volume	1.265	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.149	02:43
Breath Volume	1.242	02:43
Air Blank	0.000	02:44
Control Test	0.077	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMA,N,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: PALOMA,N
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: KINNA

DATE/TIME OF TEST: 09/09/2023 17:06:19

Diagnostics Check	OK	16:59
Air Blank	0.000	17:00
Control Test	0.079	17:00
Air Blank	0.000	17:00
Subject Sample # 1	0.170	17:01
Breath Volume	2.582	17:01
Air Blank	0.000	17:02
Air Blank	0.000	17:04
Subject Sample # 2	0.169	17:04
Breath Volume	2.484	17:04
Air Blank	0.000	17:05
Control Test	0.077	17:05
Air Blank	0.000	17:06
Diagnostics Check	OK	17:06

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: HEHMANN

DATE/TIME OF TEST: 09/13/2023 22:19:55

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.078	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.203	22:15
Breath Volume	2.484	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.204	22:18
Breath Volume	2.824	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMA,N,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REICH,O
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/14/2023 22:04:51

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.209	21:59
Breath Volume	1.285	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:01
Subject Sample # 2	0.199	22:03
Breath Volume	1.242	22:03
Air Blank	0.000	22:03
Control Test	0.075	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CARTAGENA

DATE/TIME OF TEST: 09/16/2023 01:58:51

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.141	01:54
Breath Volume	1.902	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.145	01:57
Breath Volume	1.792	01:57
Air Blank	0.000	01:57
Control Test	0.078	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: RIVAS SOSA

DATE/TIME OF TEST: 09/23/2023 21:20:05

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.022	21:15
Breath Volume	3.691	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.021	21:18
Breath Volume	3.250	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MANDY

DATE/TIME OF TEST: 09/24/2023 02:56:53

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.054	02:52
Breath Volume	1.937	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.050	02:55
Breath Volume	2.296	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELATORRE-CASTELLANO,OFC,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MEJIA TORRES

DATE/TIME OF TEST: 09/26/2023 01:06:30

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.225	01:01
Breath Volume	1.800	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.229	01:04
Breath Volume	1.945	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,SETH,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,SETH
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BILLING

DATE/TIME OF TEST: 09/27/2023 00:06:39

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.157	00:01
Breath Volume	2.320	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.165	00:05
Breath Volume	2.667	00:05
Air Blank	0.000	00:05
Control Test	0.078	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 09/27/2023 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: DACKMAN

DATE/TIME OF TEST: 09/29/2023 22:27:41

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.080	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1 0.000	22:23	
Breath Volume	1.394	22:23	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample	# 2	22:26	RFI Detect
Breath Volume	0.000	22:26	
Air Blank	0.000	22:26	
Control Test	0.078	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 10/27/2023 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: DACKMAN

DATE/TIME OF TEST: 09/29/2023 22:35:07

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.078	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.000	22:30
Breath Volume	1.378	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.000	22:33
Breath Volume	1.496	22:33
Air Blank	0.000	22:34
Control Test	0.078	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 10/27/2023 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REICH,OWEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 09/29/2023 23:33:53

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.000	23:29
Breath Volume	1.539	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.000	23:32
Breath Volume	1.140	23:32
Air Blank	0.000	23:32
Control Test	0.078	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 10/27/2023 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS, STEWART, A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH, DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/30/2023 22:30:08

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.079	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.000	22:25
Breath Volume	3.511	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.000	22:28
Breath Volume	3.730	22:28
Air Blank	0.000	22:29
Control Test	0.078	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 10/27/2023 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ALRIDGE

DATE/TIME OF TEST: 09/01/2023 00:50:30

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.077	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.114	00:46
Breath Volume	3.757	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.114	00:49
Breath Volume	3.519	00:49
Air Blank	0.000	00:49
Control Test	0.078	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 09/20/2023 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: MCNELLY

DATE/TIME OF TEST: 09/01/2023 23:30:02

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.079	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:29	
Control Test	0.078	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 09/20/2023 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LYON,L
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: VOSS

DATE/TIME OF TEST: 09/02/2023 03:58:47

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.078	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.078	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 09/20/2023 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: CRUZ LOPEZ

DATE/TIME OF TEST: 09/02/2023 21:13:55

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.077	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.139	21:09
Breath Volume	3.300	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample # 2	0.135	21:12
Breath Volume	2.835	21:12
Air Blank	0.000	21:12
Control Test	0.077	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 09/20/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CORREA RIVERA

DATE/TIME OF TEST: 09/03/2023 00:44:36

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.078	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.201	00:37
Breath Volume	2.289	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.179	00:40
Breath Volume	1.785	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 3	0.194	00:43
Breath Volume	2.656	00:43
Air Blank	0.000	00:43
Control Test	0.078	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 09/20/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: REAGAN-TEDFORD

DATE/TIME OF TEST: 09/03/2023 03:18:04

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.077	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.088	03:13
Breath Volume	1.250	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.086	03:16
Breath Volume	1.906	03:16
Air Blank	0.000	03:17
Control Test	0.077	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:18

Data Download Date/Time: 09/20/2023 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: HERNANDEZ-ANTUNEZ

DATE/TIME OF TEST: 09/03/2023 05:47:29

Diagnostics Check	OK	05:41
Air Blank	0.000	05:41
Control Test	0.078	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.095	05:42
Breath Volume	2.433	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.096	05:45
Breath Volume	2.984	05:45
Air Blank	0.000	05:46
Control Test	0.078	05:46
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

Data Download Date/Time: 09/20/2023 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/03/2023 21:10:06

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.077	21:04
Air Blank	0.000	21:04
Subject Sample # 1	0.152	21:05
Breath Volume	2.652	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.144	21:08
Breath Volume	2.031	21:08
Air Blank	0.000	21:09
Control Test	0.078	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/04/2023 02:54:13

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.078	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.077	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: ABRIL FLORIAN

DATE/TIME OF TEST: 09/04/2023 04:44:23

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.077	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.000	04:39
Breath Volume	1.160	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.000	04:43
Breath Volume	1.425	04:43
Air Blank	0.000	04:43
Control Test	0.078	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 09/05/2023 04:26:09

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.077	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.212	04:21
Breath Volume	3.781	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.207	04:24
Breath Volume	3.214	04:24
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: WOODY

DATE/TIME OF TEST: 09/08/2023 01:05:05

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.077	01:03	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Control Test	0.077	01:04	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: CALLAZO

DATE/TIME OF TEST: 09/08/2023 04:09:32

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.077	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.141	04:05
Breath Volume	1.753	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.138	04:08
Breath Volume	1.300	04:08
Air Blank	0.000	04:08
Control Test	0.078	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Atlantic Beach PD
ARRESTING OFFICER: MIMS,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: DETLEFS

DATE/TIME OF TEST: 09/08/2023 21:33:53

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:22	
Control Test	0.078	21:23	
Air Blank	0.000	21:23	
Subject Sample	# 1 0.139	21:26	Volume Not Met
Breath Volume	0.089	21:26	
Air Blank	0.000	21:27	
Air Blank	0.000	21:29	
Subject Sample	# 2 0.123	21:32	Volume Not Met
Breath Volume	0.726	21:32	
Air Blank	0.000	21:32	
Control Test	0.077	21:33	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: IDARRAGA

DATE/TIME OF TEST: 09/08/2023 23:46:21

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.018	23:41
Breath Volume	2.796	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.017	23:44
Breath Volume	2.660	23:44
Air Blank	0.000	23:45
Control Test	0.078	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/10/2023 00:08:33

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.077	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.000	00:04
Breath Volume	2.308	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.000	00:07
Breath Volume	2.085	00:07
Air Blank	0.000	00:07
Control Test	0.077	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 09/10/2023 01:38:11

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.077	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Control Test	0.077	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 09/10/2023 04:31:25

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.108	04:26
Breath Volume	2.636	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.110	04:29
Breath Volume	3.488	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 09/20/2023 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CUARTERO,L
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/10/2023 04:59:31

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:57	
Control Test	0.078	04:57	
Air Blank	0.000	04:57	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:58	
Control Test	0.077	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 09/20/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: QOSJA

DATE/TIME OF TEST: 09/10/2023 22:59:18

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.077	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.185	22:54
Breath Volume	1.492	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.191	22:57
Breath Volume	1.453	22:57
Air Blank	0.000	22:58
Control Test	0.077	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 09/20/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: HORDGE

DATE/TIME OF TEST: 09/11/2023 21:03:21

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.078	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:58
Breath Volume	1.468	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.000	21:01
Breath Volume	1.738	21:01
Air Blank	0.000	21:02
Control Test	0.076	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 09/20/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: HIGGINBOTHAM

DATE/TIME OF TEST: 09/13/2023 01:15:30

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.078	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.000	01:11
Breath Volume	3.191	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.000	01:14
Breath Volume	2.496	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 09/20/2023 12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 09/13/2023 04:24:09

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.079	04:22	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:23	
Control Test	0.079	04:23	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 09/20/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARPE,W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/13/2023 17:00:04

Diagnostics Check	OK	16:53
Air Blank	0.000	16:53
Control Test	0.078	16:53
Air Blank	0.000	16:54
Subject Sample # 1	0.000	16:55
Breath Volume	2.636	16:55
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 2	0.000	16:58
Breath Volume	2.453	16:58
Air Blank	0.000	16:59
Control Test	0.077	16:59
Air Blank	0.000	16:59
Diagnostics Check	OK	17:00

Data Download Date/Time: 09/20/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: TUPIAK

DATE/TIME OF TEST: 09/13/2023 21:35:17

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.078	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.000	21:30
Breath Volume	1.515	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.000	21:33
Breath Volume	1.398	21:33
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 09/20/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: GOODRICH

DATE/TIME OF TEST: 09/15/2023 02:29:08

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.078	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.167	02:24
Breath Volume	1.343	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:26
Subject Sample # 2	0.163	02:27
Breath Volume	1.300	02:27
Air Blank	0.000	02:28
Control Test	0.078	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 09/20/2023 12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,JAMES,G
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 09/15/2023 16:04:46

Diagnostics Check	OK	15:57
Air Blank	0.000	15:58
Control Test	0.079	15:58
Air Blank	0.000	15:59
Subject Sample # 1	0.091	15:59
Breath Volume	1.554	15:59
Air Blank	0.000	16:00
Air Blank	0.000	16:02
Subject Sample # 2	0.096	16:03
Breath Volume	3.507	16:03
Air Blank	0.000	16:03
Control Test	0.079	16:04
Air Blank	0.000	16:04
Diagnostics Check	OK	16:04

Data Download Date/Time: 09/20/2023 12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: CORDER

DATE/TIME OF TEST: 09/15/2023 22:32:56

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.000	22:28
Breath Volume	1.726	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.000	22:31
Breath Volume	1.417	22:31
Air Blank	0.000	22:31
Control Test	0.079	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 09/20/2023 12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 09/16/2023 02:32:36

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.303	02:27
Breath Volume	2.195	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.311	02:31
Breath Volume	2.203	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 09/20/2023 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CHOWDHURY,P
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: SHEETS

DATE/TIME OF TEST: 09/16/2023 04:48:27

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:46	
Control Test	0.079	04:46	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:47	
Control Test	0.079	04:47	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 09/20/2023 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Atlantic Beach PD
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: BRAMEL

DATE/TIME OF TEST: 09/16/2023 21:47:59

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:45	
Control Test	0.079	21:45	
Air Blank	0.000	21:46	
Subject Sample	# 1	21:46	Subject Test Refused
Breath Volume	0.000	21:46	
Air Blank	0.000	21:47	
Control Test	0.079	21:47	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

Data Download Date/Time: 09/20/2023 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MATTIACE,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: CHISHOLM

DATE/TIME OF TEST: 09/17/2023 01:18:07

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.153	01:13
Breath Volume	1.851	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.154	01:16
Breath Volume	2.320	01:16
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 09/20/2023 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 09/17/2023 03:08:45

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.079	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 09/20/2023 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: AVILA FONSECA

DATE/TIME OF TEST: 09/18/2023 01:33:14

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.157	01:28
Breath Volume	2.320	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.141	01:31
Breath Volume	1.234	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 09/20/2023 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 09/18/2023 06:18:59

Diagnostics Check	OK	06:16	
Air Blank	0.000	06:16	
Control Test	0.080	06:16	
Air Blank	0.000	06:17	
Subject Sample	# 1	06:17	Subject Test Refused
Breath Volume	0.000	06:17	
Air Blank	0.000	06:18	
Control Test	0.080	06:18	
Air Blank	0.000	06:18	
Diagnostics Check	OK	06:18	

Data Download Date/Time: 09/20/2023 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 09/18/2023 22:34:00

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.078	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.066	22:29
Breath Volume	1.851	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.068	22:32
Breath Volume	1.625	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:34

Data Download Date/Time: 09/20/2023 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,JUAN,A
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: FERRER,YONY
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: FROMETA MATOS

DATE/TIME OF TEST: 09/23/2023 23:33:27

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.078	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.198	23:26
Breath Volume	1.792	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:28
Subject Sample # 2	0.198	23:31
Breath Volume	2.027	23:31
Air Blank	0.000	23:32
Control Test	0.077	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 10/31/2023 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUER,S,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: HERNANDEZ,L
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: LOPEZ BATISTA

DATE/TIME OF TEST: 09/07/2023 23:27:20

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.074	23:26	Control Outside Tolerance
Air Blank	0.000	23:27	

Data Download Date/Time: 09/21/2023 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUER,STEVEN,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: HERANDEZ,LOPEZ
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: LOPEZ BATISTA

DATE/TIME OF TEST: 09/07/2023 23:34:36

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:33	RFI Detect
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.079	23:34	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 09/21/2023 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUER,STEVEN,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: HERANDEZ,LOPEZ
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 08/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: LOPEZ BATISTA

DATE/TIME OF TEST: 09/07/2023 23:47:02

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:38	
Air Blank	0.000	23:38	
Subject Sample	# 1 0.311	23:38	No .020 agreement
Breath Volume	1.992	23:38	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample	# 2 0.283	23:41	No .020 agreement
Breath Volume	2.515	23:41	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample	# 3 0.258	23:45	No .020 agreement
Breath Volume	3.125	23:45	
Air Blank	0.000	23:46	
Control Test	0.075	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 09/21/2023 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATARENO,BRIANNE,S
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: GONZALEZ,EARLES
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PROSPER

DATE/TIME OF TEST: 09/29/2023 01:35:22

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.078	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1 0.220	01:30	Volume Not Met
Breath Volume	0.070	01:30	
Air Blank	0.000	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 2	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.075	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 10/31/2023 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A2
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/06/2023 23:29:30

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.077	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.169	23:24
Breath Volume	1.105	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.170	23:27
Breath Volume	1.410	23:27
Air Blank	0.000	23:28
Control Test	0.077	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 09/14/2023 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BIDDLE,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A2
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/15/2023 23:04:55

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.200	23:00
Breath Volume	3.121	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.192	23:03
Breath Volume	2.675	23:03
Air Blank	0.000	23:03
Control Test	0.077	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 10/30/2023 09:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A2
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: JUAN PEDRO

DATE/TIME OF TEST: 09/30/2023 00:20:19

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.080	00:13	
Air Blank	0.000	00:14	
Subject Sample # 1	0.181	00:14	
Breath Volume	3.046	00:14	
Air Blank	0.000	00:15	
Air Blank	0.000	00:17	
Subject Sample # 2	0.188	00:18	
Breath Volume	2.914	00:18	
Air Blank	0.000	00:18	
Control Test	0.079	00:19	
Air Blank	0.000	00:19	
Diagnostics Check	Fail	00:19	Analytical Stability Test
Air Blank	0.000	00:20	

Data Download Date/Time: 10/30/2023 09:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A2
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: JUAN PEDRO

DATE/TIME OF TEST: 09/30/2023 00:29:04

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.077	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.195	00:24
Breath Volume	2.832	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:26
Subject Sample # 2	0.193	00:27
Breath Volume	2.921	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:29

Data Download Date/Time: 10/30/2023 09:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,KEVIN,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: COSIC,ELVIS
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: ALVAREZ BOGAERT

DATE/TIME OF TEST: 09/08/2023 02:10:58

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.125	02:05
Breath Volume	1.445	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.118	02:09
Breath Volume	1.960	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 09/21/2023 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ALLEYNE, DAVANO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: FIGUEROA CANTE

DATE/TIME OF TEST: 09/09/2023 23:06:13

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.080	23:03	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Control Test	0.079	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 09/21/2023 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: TUCCI,JOSE
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/16/2023 04:41:16

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:39	
Subject Sample	# 1	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:40	
Control Test	0.080	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 09/21/2023 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-
ROSQUETE,STEPHANNE,J
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: QUINTERO,REY
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024

VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: BAKALAR

DATE/TIME OF TEST: 09/18/2023 05:46:17

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:39	
Control Test	0.080	05:39	
Air Blank	0.000	05:40	
Subject Sample # 1	0.225	05:40	
Breath Volume	1.433	05:40	
Air Blank	0.000	05:41	
Air Blank	0.000	05:43	
Subject Sample # 2	0.218	05:44	
Breath Volume	1.648	05:44	
Air Blank	0.000	05:45	
Control Test	0.070	05:45	Control Outside Tolerance
Air Blank	0.000	05:46	

Data Download Date/Time: 09/21/2023 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAGOS, GUILLERMO, E
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: PORTILLO, SCARLYN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: KUZYAK

DATE/TIME OF TEST: 09/28/2023 22:40:58

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.080	22:38	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 10/19/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ALLEYNE, DAVANO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: DA SILVA

DATE/TIME OF TEST: 09/30/2023 18:15:06

Diagnostics Check	OK	18:07
Air Blank	0.000	18:08
Control Test	0.080	18:08
Air Blank	0.000	18:09
Subject Sample # 1	0.136	18:10
Breath Volume	2.871	18:10
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.135	18:13
Breath Volume	2.656	18:13
Air Blank	0.000	18:14
Control Test	0.080	18:14
Air Blank	0.000	18:15
Diagnostics Check	OK	18:15

Data Download Date/Time: 10/19/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/03/2023 04:25:36

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.171	04:21
Breath Volume	2.878	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.173	04:24
Breath Volume	1.890	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 09/16/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASPER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BUONICONTI,ALEXI
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: STEPHAN

DATE/TIME OF TEST: 09/05/2023 05:23:39

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.000	05:19
Breath Volume	1.667	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.000	05:22
Breath Volume	1.996	05:22
Air Blank	0.000	05:22
Control Test	0.080	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 09/16/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,L
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MATTEA

DATE/TIME OF TEST: 09/11/2023 03:46:18

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:41	No Sample Provided
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Air Blank	0.000	03:44	
Subject Sample	# 2	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.080	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 09/16/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZIZ, DAVID,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: AZIZ, DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/28/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 09/14/2023 03:50:29

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.039	03:46
Breath Volume	1.527	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.039	03:49
Breath Volume	2.363	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 09/16/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ROBERT,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNYDER,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SCHEERER

DATE/TIME OF TEST: 09/17/2023 02:39:40

Diagnostics Check	OK	02:38	
Air Blank		02:39	RFI Detect
Air Blank	0.000	02:39	

Data Download Date/Time: 10/13/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ROBERT,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNYDER,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SCHEERER

DATE/TIME OF TEST: 09/17/2023 02:48:11

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.186	02:43
Breath Volume	2.757	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.185	02:46
Breath Volume	2.679	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 10/13/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZIZ,DAVID,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: AZIZ,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: DIXON PERALES

DATE/TIME OF TEST: 09/18/2023 15:15:06

Diagnostics Check	OK	15:08
Air Blank	0.000	15:09
Control Test	0.080	15:09
Air Blank	0.000	15:09
Subject Sample # 1	0.145	15:10
Breath Volume	2.011	15:10
Air Blank	0.000	15:11
Air Blank	0.000	15:12
Subject Sample # 2	0.131	15:13
Breath Volume	1.722	15:13
Air Blank	0.000	15:14
Control Test	0.080	15:14
Air Blank	0.000	15:15
Diagnostics Check	OK	15:15

Data Download Date/Time: 10/13/2023 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: BAIM

DATE/TIME OF TEST: 09/23/2023 21:44:41

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.156	21:40
Breath Volume	2.015	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.159	21:43
Breath Volume	2.226	21:43
Air Blank	0.000	21:43
Control Test	0.080	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 10/13/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIELE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MIELE,ROBERT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 09/30/2023 03:18:55

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Subject Sample # 1	0.197	03:15	
Breath Volume	1.210	03:15	
Air Blank	0.000	03:15	
Air Blank	0.000	03:17	
Subject Sample # 2	0.199	03:18	
Breath Volume	1.988	03:18	
Air Blank		03:18	RFI Detect
Air Blank		03:18	RFI Detect

Data Download Date/Time: 10/13/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIELE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MCALLISTER,NICK
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 09/30/2023 03:30:11

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.205	03:25
Breath Volume	2.027	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.201	03:28
Breath Volume	2.582	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 10/13/2023 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SKALA

DATE/TIME OF TEST: 09/07/2023 03:53:40

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.155	03:48
Breath Volume	1.109	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.169	03:52
Breath Volume	2.109	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 09/08/2023 04:12:02

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.209	04:07
Breath Volume	1.339	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.218	04:10
Breath Volume	1.601	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:12

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 09/08/2023 04:37:22

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.303	04:31
Breath Volume	1.644	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.305	04:35
Breath Volume	1.769	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: DAO

DATE/TIME OF TEST: 09/08/2023 22:48:58

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.175	22:44
Breath Volume	2.921	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:46
Subject Sample # 2	0.173	22:47
Breath Volume	2.351	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 09/09/2023 01:59:20

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.174	01:54
Breath Volume	2.210	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.175	01:57
Breath Volume	1.960	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: TINDALL

DATE/TIME OF TEST: 09/09/2023 02:51:16

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.151	02:46
Breath Volume	4.414	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.148	02:49
Breath Volume	3.960	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: ZIETZ

DATE/TIME OF TEST: 09/09/2023 05:28:14

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.081	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.169	05:23
Breath Volume	2.093	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.178	05:26
Breath Volume	3.484	05:26
Air Blank	0.000	05:27
Control Test	0.075	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: HERNANDEZ GARCIA

DATE/TIME OF TEST: 09/09/2023 22:17:54

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.080	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.216	22:13
Breath Volume	1.847	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.222	22:16
Breath Volume	2.023	22:16
Air Blank	0.000	22:16
Control Test	0.081	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 09/09/2023 23:37:53

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.176	23:33
Breath Volume	2.269	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:35
Subject Sample # 2	0.194	23:36
Breath Volume	2.058	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: SANDS

DATE/TIME OF TEST: 09/10/2023 01:33:58

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.156	01:29
Breath Volume	2.542	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.154	01:32
Breath Volume	2.097	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ORELLANA DOMINGUEZ

DATE/TIME OF TEST: 09/10/2023 02:37:02

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.186	02:32
Breath Volume	1.765	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.190	02:35
Breath Volume	1.304	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:37

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: AGUILAR GOMEZ

DATE/TIME OF TEST: 09/10/2023 03:10:14

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.080	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.191	03:05
Breath Volume	1.953	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.190	03:08
Breath Volume	1.292	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 10/02/2023 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: DISHMEY MONTES DE OC

DATE/TIME OF TEST: 09/10/2023 03:42:43

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:39	
Control Test	0.079	03:39	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Control Test	0.080	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: TELLEZ DIAZ

DATE/TIME OF TEST: 09/10/2023 04:02:31

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.164	03:58
Breath Volume	2.027	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.166	04:00
Breath Volume	2.070	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: DOMINGUEZ REYES

DATE/TIME OF TEST: 09/10/2023 06:04:19

Diagnostics Check	OK	05:58
Air Blank	0.000	05:58
Control Test	0.081	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.099	05:59
Breath Volume	1.757	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.101	06:02
Breath Volume	2.679	06:02
Air Blank	0.000	06:03
Control Test	0.081	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 09/10/2023 18:32:49

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.081	18:26
Air Blank	0.000	18:27
Subject Sample # 1	0.164	18:27
Breath Volume	1.671	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.159	18:31
Breath Volume	2.414	18:31
Air Blank	0.000	18:31
Control Test	0.080	18:32
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GONZALEZ BATIZ

DATE/TIME OF TEST: 09/11/2023 00:22:28

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.081	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:15	No Sample Provided
Breath Volume	0.000	00:15	
Air Blank	0.000	00:15	
Air Blank	0.000	00:17	
Subject Sample	# 2	00:21	No Sample Provided
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: DODGE

DATE/TIME OF TEST: 09/11/2023 02:34:19

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.082	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.205	02:29
Breath Volume	1.472	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.202	02:32
Breath Volume	2.156	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 09/11/2023 19:00:00

Diagnostics Check	OK	18:52
Air Blank	0.000	18:53
Control Test	0.080	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.234	18:54
Breath Volume	1.976	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.233	18:58
Breath Volume	2.039	18:58
Air Blank	0.000	18:59
Control Test	0.081	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	19:00

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 09/11/2023 23:41:45

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.287	23:37
Breath Volume	2.089	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.287	23:40
Breath Volume	1.625	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SAMOS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:19

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 09/12/2023 11:45:29

Diagnostics Check	OK	11:39
Air Blank	0.000	11:40
Control Test	0.081	11:40
Air Blank	0.000	11:40
Subject Sample # 1	0.000	11:41
Breath Volume	1.355	11:41
Air Blank	0.000	11:41
Air Blank	0.000	11:43
Subject Sample # 2	0.000	11:44
Breath Volume	1.417	11:44
Air Blank	0.000	11:44
Control Test	0.081	11:44
Air Blank	0.000	11:45
Diagnostics Check	OK	11:45

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 09/13/2023 02:21:13

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.156	02:16
Breath Volume	1.105	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.147	02:19
Breath Volume	1.175	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: LANIER

DATE/TIME OF TEST: 09/13/2023 10:19:28

Diagnostics Check	OK	10:13
Air Blank	0.000	10:13
Control Test	0.081	10:13
Air Blank	0.000	10:14
Subject Sample # 1	0.000	10:14
Breath Volume	2.738	10:14
Air Blank	0.000	10:15
Air Blank	0.000	10:17
Subject Sample # 2	0.000	10:18
Breath Volume	1.980	10:18
Air Blank	0.000	10:18
Control Test	0.082	10:18
Air Blank	0.000	10:19
Diagnostics Check	OK	10:19

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUSWELL,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: ECKLUND

DATE/TIME OF TEST: 09/14/2023 16:30:24

Diagnostics Check	OK	16:21
Air Blank	0.000	16:22
Control Test	0.082	16:22
Air Blank	0.000	16:22
Subject Sample # 1	0.114	16:25
Breath Volume	1.171	16:25
Air Blank	0.000	16:25
Air Blank	0.000	16:27
Subject Sample # 2	0.115	16:28
Breath Volume	1.468	16:28
Air Blank	0.000	16:29
Control Test	0.081	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: FULTON

DATE/TIME OF TEST: 09/15/2023 03:45:28

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.160	03:40
Breath Volume	2.457	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.148	03:43
Breath Volume	2.476	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: ASHE

DATE/TIME OF TEST: 09/15/2023 10:47:24

Diagnostics Check	OK	10:33	
Air Blank	0.000	10:33	
Control Test	0.081	10:33	
Air Blank	0.000	10:34	
Subject Sample	# 1 0.037	10:34	
Breath Volume	1.960	10:34	
Air Blank	0.000	10:35	
Air Blank	0.000	10:37	
Subject Sample	# 2	10:40	No Sample Provided
Breath Volume	0.000	10:40	
Air Blank	0.000	10:40	
Air Blank	0.000	10:42	
Subject Sample	# 3	10:46	No Sample Provided
Breath Volume	0.000	10:46	
Air Blank	0.000	10:46	
Control Test	0.081	10:46	
Air Blank	0.000	10:47	
Diagnostics Check	OK	10:47	

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIIH,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 09/15/2023 14:15:00

Diagnostics Check	OK	14:08
Air Blank	0.000	14:09
Control Test	0.081	14:09
Air Blank	0.000	14:10
Subject Sample # 1	0.000	14:10
Breath Volume	3.203	14:10
Air Blank	0.000	14:11
Air Blank	0.000	14:13
Subject Sample # 2	0.000	14:13
Breath Volume	2.980	14:13
Air Blank	0.000	14:14
Control Test	0.082	14:14
Air Blank	0.000	14:14
Diagnostics Check	OK	14:15

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 09/15/2023 23:34:52

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.081	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.177	23:28
Breath Volume	1.628	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.166	23:33
Breath Volume	1.492	23:33
Air Blank	0.000	23:33
Control Test	0.081	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LORENZO,E
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 09/16/2023 01:49:07

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.081	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.204	01:44
Breath Volume	2.363	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:46
Subject Sample # 2	0.204	01:47
Breath Volume	2.152	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: AVILA CASTRO

DATE/TIME OF TEST: 09/16/2023 02:42:02

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Control Test	0.081	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: AGUIRRE-GARCIA

DATE/TIME OF TEST: 09/16/2023 05:17:11

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.081	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.176	05:12
Breath Volume	1.191	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:14
Subject Sample # 2	0.186	05:15
Breath Volume	2.085	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 09/16/2023 05:28:37

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.148	05:24
Breath Volume	1.121	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.151	05:27
Breath Volume	1.480	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: DOMINGUEZ DOMINGUEZ

DATE/TIME OF TEST: 09/16/2023 22:09:44

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.224	22:05
Breath Volume	1.468	22:05
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.212	22:08
Breath Volume	1.546	22:08
Air Blank	0.000	22:08
Control Test	0.079	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CONSUEGRA PEDRAZA

DATE/TIME OF TEST: 09/16/2023 22:43:30

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.164	22:38
Breath Volume	2.976	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.162	22:41
Breath Volume	2.371	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: GIRON SALAS

DATE/TIME OF TEST: 09/17/2023 01:54:57

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.098	01:50
Breath Volume	1.484	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:52
Subject Sample # 2	0.096	01:53
Breath Volume	2.164	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: DACEY

DATE/TIME OF TEST: 09/17/2023 02:08:37

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.187	02:04
Breath Volume	2.867	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.173	02:07
Breath Volume	2.421	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: PEREZ NARANJO

DATE/TIME OF TEST: 09/17/2023 03:07:53

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.100	03:03
Breath Volume	3.863	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.100	03:06
Breath Volume	3.136	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFAMY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS, J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SHIREY

DATE/TIME OF TEST: 09/17/2023 03:21:19

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.176	03:15
Breath Volume	1.593	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.168	03:19
Breath Volume	1.957	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: GONZALEZ GONZALEZ

DATE/TIME OF TEST: 09/17/2023 04:13:28

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.118	04:08
Breath Volume	1.308	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.121	04:11
Breath Volume	2.191	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 10/02/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: LOWO

DATE/TIME OF TEST: 09/17/2023 05:20:03

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:08	
Control Test	0.080	05:09	
Air Blank	0.000	05:09	
Subject Sample	# 1 0.000	05:12	Volume Not Met
Breath Volume	0.105	05:12	
Air Blank	0.000	05:13	
Air Blank	0.000	05:15	
Subject Sample	# 2 0.148	05:18	Volume Not Met
Breath Volume	0.203	05:18	
Air Blank	0.000	05:19	
Control Test	0.080	05:19	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEGGETT,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:24

SUBJECT'S LAST NAME: LEDEZMA JR

DATE/TIME OF TEST: 09/17/2023 08:52:21

Diagnostics Check	OK	08:45
Air Blank	0.000	08:46
Control Test	0.081	08:46
Air Blank	0.000	08:47
Subject Sample # 1	0.187	08:47
Breath Volume	1.953	08:47
Air Blank	0.000	08:48
Air Blank	0.000	08:50
Subject Sample # 2	0.194	08:50
Breath Volume	2.042	08:50
Air Blank	0.000	08:51
Control Test	0.082	08:51
Air Blank	0.000	08:52
Diagnostics Check	OK	08:52

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ORTIZ PEREZ

DATE/TIME OF TEST: 09/17/2023 23:11:54

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.081	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.237	23:07
Breath Volume	1.628	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:09
Subject Sample # 2	0.247	23:10
Breath Volume	2.582	23:10
Air Blank	0.000	23:10
Control Test	0.082	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/18/2023 01:20:57

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1 0.111	01:08	
Breath Volume	1.113	01:08	
Air Blank	0.000	01:08	
Air Blank	0.000	01:10	
Subject Sample	# 2 0.101	01:13	Volume Not Met
Breath Volume	0.156	01:13	
Air Blank	0.000	01:14	
Air Blank	0.000	01:16	
Subject Sample	# 3 0.100	01:19	Volume Not Met
Breath Volume	0.195	01:19	
Air Blank	0.000	01:20	
Control Test	0.079	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUSWELL,MARK
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: SOBERANES PALMA

DATE/TIME OF TEST: 09/18/2023 11:19:11

Diagnostics Check	OK	11:12
Air Blank	0.000	11:12
Control Test	0.081	11:13
Air Blank	0.000	11:13
Subject Sample # 1	0.257	11:14
Breath Volume	2.058	11:14
Air Blank	0.000	11:15
Air Blank	0.000	11:17
Subject Sample # 2	0.260	11:17
Breath Volume	2.175	11:17
Air Blank	0.000	11:18
Control Test	0.080	11:18
Air Blank	0.000	11:19
Diagnostics Check	OK	11:19

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHRISTMAN,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: PALACIOS MARTINEZ

DATE/TIME OF TEST: 09/18/2023 17:29:57

Diagnostics Check	OK	17:21
Air Blank	0.000	17:22
Control Test	0.082	17:22
Air Blank	0.000	17:23
Subject Sample # 1	0.074	17:25
Breath Volume	1.679	17:25
Air Blank	0.000	17:26
Air Blank	0.000	17:27
Subject Sample # 2	0.069	17:28
Breath Volume	1.250	17:28
Air Blank	0.000	17:29
Control Test	0.081	17:29
Air Blank	0.000	17:29
Diagnostics Check	OK	17:29

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LOGAN,Z
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: TIDROW

DATE/TIME OF TEST: 09/19/2023 22:12:22

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.082	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.086	22:07
Breath Volume	3.855	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.085	22:10
Breath Volume	3.558	22:10
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BILOBROVKA

DATE/TIME OF TEST: 09/19/2023 23:42:09

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.249	23:37
Breath Volume	2.128	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.269	23:40
Breath Volume	2.519	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: GOMEZ DIAZ

DATE/TIME OF TEST: 09/20/2023 01:46:22

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.144	01:41
Breath Volume	2.035	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.149	01:44
Breath Volume	2.062	01:44
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/20/2023 03:10:55

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.251	03:06
Breath Volume	1.835	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.247	03:09
Breath Volume	1.808	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CHAPARRO-VASCONSELLO

DATE/TIME OF TEST: 09/21/2023 01:31:27

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.048	01:27
Breath Volume	2.234	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.049	01:29
Breath Volume	1.949	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: OQUENDO

DATE/TIME OF TEST: 09/21/2023 02:32:06

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.170	02:27
Breath Volume	1.335	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.172	02:30
Breath Volume	1.425	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: COMM VEHIC
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/21/2023 02:55:06

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.000	02:50
Breath Volume	2.851	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.000	02:53
Breath Volume	1.960	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LAROSE,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: KRISHNASAMY

DATE/TIME OF TEST: 09/21/2023 04:06:45

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.158	04:02
Breath Volume	1.355	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.169	04:05
Breath Volume	1.648	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: ELLENBERGER

DATE/TIME OF TEST: 09/21/2023 19:31:14

Diagnostics Check	OK	19:24
Air Blank	0.000	19:24
Control Test	0.081	19:25
Air Blank	0.000	19:25
Subject Sample # 1	0.000	19:26
Breath Volume	2.242	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.000	19:29
Breath Volume	1.851	19:29
Air Blank	0.000	19:30
Control Test	0.080	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: MENA

DATE/TIME OF TEST: 09/21/2023 22:55:55

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.081	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.213	22:50
Breath Volume	2.480	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.204	22:54
Breath Volume	1.902	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: PESHEK

DATE/TIME OF TEST: 09/21/2023 23:10:37

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.000	23:06
Breath Volume	1.617	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.000	23:09
Breath Volume	1.414	23:09
Air Blank	0.000	23:09
Control Test	0.081	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: DAAS

DATE/TIME OF TEST: 09/22/2023 00:33:06

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.000	00:28
Breath Volume	1.589	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.000	00:31
Breath Volume	1.574	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: WARE

DATE/TIME OF TEST: 09/22/2023 03:10:00

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.099	03:05
Breath Volume	1.328	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.105	03:08
Breath Volume	2.191	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:10

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SNIPES

DATE/TIME OF TEST: 09/22/2023 03:50:31

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.077	03:45
Breath Volume	1.390	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.080	03:48
Breath Volume	1.585	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: FONSECA

DATE/TIME OF TEST: 09/23/2023 04:33:25

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.166	04:28
Breath Volume	1.417	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.162	04:31
Breath Volume	1.777	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: SAMAS

DATE/TIME OF TEST: 09/23/2023 05:21:39

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.166	05:16
Breath Volume	2.109	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.165	05:20
Breath Volume	1.687	05:20
Air Blank	0.000	05:20
Control Test	0.081	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUSWELL,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:14

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 09/23/2023 09:41:58

Diagnostics Check	OK	09:35
Air Blank	0.000	09:35
Control Test	0.081	09:36
Air Blank	0.000	09:36
Subject Sample # 1	0.161	09:37
Breath Volume	2.019	09:37
Air Blank	0.000	09:37
Air Blank	0.000	09:39
Subject Sample # 2	0.161	09:40
Breath Volume	1.792	09:40
Air Blank	0.000	09:41
Control Test	0.081	09:41
Air Blank	0.000	09:41
Diagnostics Check	OK	09:41

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ORTIZ SANCHEZ

DATE/TIME OF TEST: 09/23/2023 23:47:47

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.081	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:39	No Sample Provided
Breath Volume	0.000	23:39	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 2	0.213	23:42
Breath Volume	1.925	23:42	
Air Blank	0.000	23:43	
Air Blank	0.000	23:45	
Subject Sample	# 3	0.212	23:46
Breath Volume	2.242	23:46	
Air Blank	0.000	23:46	
Control Test	0.080	23:47	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 10/02/2023 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 09/24/2023 01:56:52

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.081	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.114	01:52
Breath Volume	1.367	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.115	01:55
Breath Volume	1.562	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CARDENAL REYES

DATE/TIME OF TEST: 09/24/2023 02:41:27

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.082	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.103	02:36
Breath Volume	2.742	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.099	02:39
Breath Volume	1.812	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: BENSCH

DATE/TIME OF TEST: 09/24/2023 03:04:21

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.081	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.110	02:59
Breath Volume	2.242	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.101	03:02
Breath Volume	2.539	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: ADAIR

DATE/TIME OF TEST: 09/24/2023 03:28:05

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.127	03:23
Breath Volume	2.308	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.120	03:26
Breath Volume	1.714	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: HEIM

DATE/TIME OF TEST: 09/24/2023 05:13:30

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.082	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.072	05:08
Breath Volume	1.832	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:10
Subject Sample # 2	0.073	05:11
Breath Volume	2.511	05:11
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MORRAMREDDY VEERA

DATE/TIME OF TEST: 09/24/2023 05:47:25

Diagnostics Check	OK	05:30	
Air Blank	0.000	05:31	
Control Test	0.081	05:31	
Air Blank	0.000	05:32	
Subject Sample # 1	0.182	05:34	
Breath Volume	1.312	05:34	
Air Blank	0.000	05:35	
Air Blank	0.000	05:37	
Subject Sample # 2	0.158	05:40	Volume Not Met
Breath Volume	0.675	05:40	
Air Blank	0.000	05:40	
Air Blank	0.000	05:42	
Subject Sample # 3	0.167	05:45	Volume Not Met
Breath Volume	0.082	05:45	
Air Blank	0.000	05:46	
Control Test	0.078	05:46	
Air Blank	0.000	05:47	
Diagnostics Check	OK	05:47	

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: CAJOLIN

DATE/TIME OF TEST: 09/24/2023 07:08:48

Diagnostics Check	OK	06:53	
Air Blank	0.000	06:54	
Control Test	0.081	06:54	
Air Blank	0.000	06:55	
Subject Sample	# 1 0.146	06:56	
Breath Volume	1.347	06:56	
Air Blank	0.000	06:56	
Air Blank	0.000	06:58	
Subject Sample	# 2	07:01	No Sample Provided
Breath Volume	0.000	07:01	
Air Blank	0.000	07:02	
Air Blank	0.000	07:04	
Subject Sample	# 3	07:07	No Sample Provided
Breath Volume	0.000	07:07	
Air Blank	0.000	07:07	
Control Test	0.081	07:08	
Air Blank	0.000	07:08	
Diagnostics Check	OK	07:08	

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GRECO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: CHEBOTAROV

DATE/TIME OF TEST: 09/24/2023 18:21:41

Diagnostics Check	OK	18:15
Air Blank	0.000	18:15
Control Test	0.081	18:16
Air Blank	0.000	18:16
Subject Sample # 1	0.245	18:17
Breath Volume	2.894	18:17
Air Blank	0.000	18:17
Air Blank	0.000	18:19
Subject Sample # 2	0.237	18:20
Breath Volume	2.871	18:20
Air Blank	0.000	18:20
Control Test	0.080	18:21
Air Blank	0.000	18:21
Diagnostics Check	OK	18:21

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MCDONALD-SMITH

DATE/TIME OF TEST: 09/24/2023 20:20:22

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.081	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.389	20:15
Breath Volume	2.156	20:15
Air Blank	0.000	20:16
Air Blank	0.000	20:17
Subject Sample # 2	0.400	20:18
Breath Volume	2.000	20:18
Air Blank	0.000	20:19
Control Test	0.080	20:19
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/24/2023 22:16:12

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.175	22:10
Breath Volume	1.636	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.170	22:14
Breath Volume	1.535	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: DURBAN

DATE/TIME OF TEST: 09/25/2023 02:55:02

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.160	02:50
Breath Volume	2.093	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.154	02:53
Breath Volume	1.570	02:53
Air Blank	0.000	02:54
Control Test	0.078	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOTHEM,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: HENDRY

DATE/TIME OF TEST: 09/25/2023 18:44:59

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.081	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.117	18:40
Breath Volume	1.480	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:42
Subject Sample # 2	0.117	18:43
Breath Volume	1.210	18:43
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: RUIZ HERNANDEZ

DATE/TIME OF TEST: 09/25/2023 22:18:47

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.081	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.224	22:14
Breath Volume	2.468	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.228	22:17
Breath Volume	2.601	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BARRIENTOS

DATE/TIME OF TEST: 09/25/2023 23:58:03

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:55	
Control Test	0.081	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:56	Subject Test Refused
Breath Volume	0.000	23:56	
Air Blank	0.000	23:57	
Control Test	0.081	23:57	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: MCKAMEY

DATE/TIME OF TEST: 09/26/2023 01:25:54

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.081	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.000	01:21
Breath Volume	1.945	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.000	01:24
Breath Volume	2.207	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/26/2023 01:43:24

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.138	01:38
Breath Volume	1.453	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.131	01:41
Breath Volume	1.468	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIRALDO,J
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 09/26/2023 02:31:02

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.098	02:26
Breath Volume	2.324	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:28
Subject Sample # 2	0.099	02:29
Breath Volume	2.574	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:31

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: TORREZ ICO

DATE/TIME OF TEST: 09/27/2023 01:56:50

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.051	01:52
Breath Volume	1.480	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.050	01:55
Breath Volume	1.699	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: POINT DU JOUR

DATE/TIME OF TEST: 09/27/2023 04:38:45

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.160	04:34
Breath Volume	1.800	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.146	04:37
Breath Volume	1.253	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS, J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 09/28/2023 00:07:09

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 3 0.145	23:57	Volume Not Met
Breath Volume	0.285	23:57	
Air Blank	0.000	23:58	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.145	00:01	
Breath Volume	1.515	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:04	
Subject Sample	# 2 0.144	00:05	
Breath Volume	1.156	00:05	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: PERDOMO

DATE/TIME OF TEST: 09/28/2023 08:45:06

Diagnostics Check	OK	08:37
Air Blank	0.000	08:38
Control Test	0.081	08:38
Air Blank	0.000	08:38
Subject Sample # 1	0.117	08:40
Breath Volume	1.128	08:40
Air Blank	0.000	08:41
Air Blank	0.000	08:42
Subject Sample # 2	0.112	08:43
Breath Volume	1.113	08:43
Air Blank	0.000	08:44
Control Test	0.080	08:44
Air Blank	0.000	08:45
Diagnostics Check	OK	08:45

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON,B
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: EARLL

DATE/TIME OF TEST: 09/30/2023 01:15:23

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.152	01:10
Breath Volume	2.144	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.150	01:13
Breath Volume	1.492	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BURRELL

DATE/TIME OF TEST: 09/30/2023 02:00:53

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.082	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.106	01:56
Breath Volume	2.136	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.101	01:59
Breath Volume	2.238	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: NOSWORTHY

DATE/TIME OF TEST: 09/30/2023 02:57:37

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.110	02:53
Breath Volume	2.601	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.109	02:56
Breath Volume	1.996	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/30/2023 04:43:07

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.081	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.149	04:38
Breath Volume	3.304	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.141	04:41
Breath Volume	2.457	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 10/02/2023 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: FONTENELLE

DATE/TIME OF TEST: 09/24/2023 00:48:15

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.082	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.282	00:43
Breath Volume	1.882	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.280	00:46
Breath Volume	1.199	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 10/02/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FELIZ RUIZ

DATE/TIME OF TEST: 09/24/2023 03:55:27

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.081	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.097	03:51
Breath Volume	2.554	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.100	03:53
Breath Volume	1.839	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 10/02/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: HEMMINGWAY JR

DATE/TIME OF TEST: 09/25/2023 03:14:56

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.145	03:09
Breath Volume	2.042	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.136	03:13
Breath Volume	1.898	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 10/02/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA,ARMANDO,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA,ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ROMERO GARCIA

DATE/TIME OF TEST: 09/26/2023 04:45:58

Diagnostics Check	OK	04:45	
Air Blank		04:45	RFI Detect
Air Blank	0.000	04:45	

Data Download Date/Time: 10/19/2023 22:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ROMERO GARCIA

DATE/TIME OF TEST: 09/26/2023 05:01:16

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.081	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.106	04:56
Breath Volume	2.320	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.114	04:59
Breath Volume	1.828	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 10/19/2023 22:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA,ARMANDO,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA,ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CARRILLO ECHEVARRIA

DATE/TIME OF TEST: 09/30/2023 00:34:19

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.198	00:26
Breath Volume	1.687	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.173	00:29
Breath Volume	1.554	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 3	0.182	00:32
Breath Volume	1.609	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 10/19/2023 22:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZOG,TRISTAN,H
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HARTZOG,TRISTAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: BUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 09/17/2023 20:27:00

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:24	
Control Test	0.082	20:24	
Air Blank	0.000	20:25	
Subject Sample	# 1	20:25	Subject Test Refused
Breath Volume	0.000	20:25	
Air Blank	0.000	20:26	
Control Test	0.082	20:26	
Air Blank	0.000	20:26	
Diagnostics Check	OK	20:27	

Data Download Date/Time: 09/21/2023 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: OSECHAS-CARRILLO

DATE/TIME OF TEST: 09/27/2023 04:30:22

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.052	04:26
Breath Volume	2.082	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.053	04:28
Breath Volume	2.195	04:28
Air Blank	0.000	04:29
Control Test	0.081	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 10/23/2023 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: GURDON

DATE/TIME OF TEST: 09/30/2023 01:12:12

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.081	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.175	01:07
Breath Volume	1.488	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.170	01:10
Breath Volume	1.703	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 10/23/2023 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: LISBOA

DATE/TIME OF TEST: 09/01/2023 19:54:34

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.081	19:48
Air Blank	0.000	19:48
Subject Sample # 1	0.156	19:49
Breath Volume	1.171	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:51
Subject Sample # 2	0.141	19:52
Breath Volume	1.355	19:52
Air Blank	0.000	19:53
Control Test	0.081	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 09/25/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/02/2023 00:56:02

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.082	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.039	00:51
Breath Volume	2.125	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.038	00:54
Breath Volume	1.683	00:54
Air Blank	0.000	00:55
Control Test	0.081	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:56

Data Download Date/Time: 09/25/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 09/02/2023 02:19:22

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.015	02:13
Breath Volume	1.640	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.016	02:17
Breath Volume	1.730	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 09/25/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: RODRIGUEZ-SILVA

DATE/TIME OF TEST: 09/03/2023 02:31:04

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.082	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.091	02:25
Breath Volume	2.843	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.091	02:29
Breath Volume	2.609	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:31

Data Download Date/Time: 09/25/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: RIGGALL,NICHOLAS

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: DELONG

DATE/TIME OF TEST: 09/03/2023 22:44:21

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.081	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.130	22:38
Breath Volume	1.492	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:40
Subject Sample # 2	0.129	22:42
Breath Volume	1.167	22:42
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 09/25/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFONTANT, SEDONA, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: LAFONTANT, SEDONA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: JEFFREY

DATE/TIME OF TEST: 09/08/2023 02:17:46

Diagnostics Check	OK	02:17	
Air Blank		02:17	RFI Detect
Air Blank		02:17	RFI Detect

Data Download Date/Time: 09/25/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFONTANT, SEDONA, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: LAFONTANT, SEDONA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: JEFFREY

DATE/TIME OF TEST: 09/08/2023 02:27:08

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.000	02:22
Breath Volume	1.156	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.000	02:25
Breath Volume	1.660	02:25
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 09/25/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: PAXTON

DATE/TIME OF TEST: 09/09/2023 02:18:26

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.199	02:13
Breath Volume	1.480	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.199	02:16
Breath Volume	1.558	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 09/25/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: FAIRCLOTH

DATE/TIME OF TEST: 09/17/2023 00:41:29

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.222	00:36
Breath Volume	1.375	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.231	00:39
Breath Volume	1.792	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 09/25/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ARROYO,ALEJANDRO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 407397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: CUBERO

DATE/TIME OF TEST: 09/14/2023 21:04:23

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:01	
Control Test	0.078	21:02	
Air Blank	0.000	21:02	
Subject Sample	# 1	21:03	RFI Detect
Breath Volume	0.000	21:03	
Air Blank	0.000	21:03	
Control Test	0.078	21:03	
Air Blank	0.000	21:04	
Diagnostics Check	OK	21:04	

Data Download Date/Time: 10/23/2023 16:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ARROYO,ALEJANDRO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 407397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 09/14/2023 21:11:28

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:06	
Control Test	0.077	21:06	
Air Blank	0.000	21:07	
Subject Sample	# 1 0.093	21:07	
Breath Volume	3.226	21:07	
Air Blank	0.000	21:08	
Air Blank	0.000	21:10	
Subject Sample	# 2 0.096	21:10	
Breath Volume	1.941	21:10	
Air Blank		21:10	RFI Detect
Air Blank	0.000	21:11	

Data Download Date/Time: 10/23/2023 16:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ARROYO,ALEJANDRO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 407397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/20/2023 02:20:44

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.079	02:18	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:19	RFI Detect
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Control Test	0.079	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 10/23/2023 16:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ARROYO,ALEJANDRO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 407397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/20/2023 02:25:38

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.078	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:24	RFI Detect
Breath Volume	2.480	02:24	
Air Blank	0.000	02:24	
Control Test	0.078	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 10/23/2023 16:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ARROYO,ALEJANDRO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 407397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/20/2023 02:33:17

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.130	02:28
Breath Volume	1.902	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.130	02:31
Breath Volume	2.136	02:31
Air Blank	0.000	02:32
Control Test	0.078	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 10/23/2023 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ARROYO,ALEJANDRO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 407397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: WAITS

DATE/TIME OF TEST: 09/23/2023 01:40:12

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.149	01:35
Breath Volume	2.832	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.143	01:38
Breath Volume	2.019	01:38
Air Blank	0.000	01:39
Control Test	0.078	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 10/23/2023 16:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/05/2023 03:40:09

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.082	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.131	03:35
Breath Volume	2.207	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.138	03:38
Breath Volume	2.027	03:38
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 09/29/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, SETH, D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: LARGEN, J
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: COLIN

DATE/TIME OF TEST: 09/10/2023 03:41:53

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.082	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.143	03:37
Breath Volume	1.683	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.143	03:40
Breath Volume	1.589	03:40
Air Blank	0.000	03:40
Control Test	0.081	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 09/29/2023 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIRLING,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MOONEY,PATRICK
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 09/16/2023 22:10:39

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.081	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 1	22:09	Subject Test Refused
Breath Volume	0.000	22:09	
Air Blank	0.000	22:09	
Control Test	0.081	22:10	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

Data Download Date/Time: 09/29/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MOTLEY

DATE/TIME OF TEST: 09/17/2023 04:14:17

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.113	04:09
Breath Volume	1.351	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.122	04:12
Breath Volume	1.398	04:12
Air Blank	0.000	04:13
Control Test	0.081	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 09/29/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIECK, DAVID, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: BIECK, DAVID
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: MANZANO DIAZ

DATE/TIME OF TEST: 09/23/2023 17:24:41

Diagnostics Check	OK	17:12	
Air Blank	0.000	17:13	
Control Test	0.081	17:13	
Air Blank	0.000	17:14	
Subject Sample	# 1 0.247	17:17	Volume Not Met
Breath Volume	0.296	17:17	
Air Blank	0.000	17:17	
Air Blank	0.000	17:19	
Subject Sample	# 2 0.243	17:22	Volume Not Met
Breath Volume	0.242	17:22	
Air Blank	0.000	17:23	
Control Test	0.081	17:23	
Air Blank	0.000	17:24	
Diagnostics Check	OK	17:24	

Data Download Date/Time: 09/29/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: SWEARINGEN

DATE/TIME OF TEST: 09/24/2023 04:23:25

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.155	04:18
Breath Volume	1.488	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.152	04:21
Breath Volume	2.054	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 09/29/2023 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: BORN

DATE/TIME OF TEST: 09/26/2023 03:37:41

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.081	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.167	03:33
Breath Volume	1.921	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.162	03:36
Breath Volume	2.238	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 09/29/2023 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RODRUNGSEE

DATE/TIME OF TEST: 09/26/2023 22:27:08

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:24	
Control Test	0.082	22:25	
Air Blank	0.000	22:25	
Subject Sample	# 1	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:26	
Control Test	0.082	22:26	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 09/29/2023 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 09/27/2023 03:57:00

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.228	03:52
Breath Volume	2.757	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.213	03:55
Breath Volume	2.644	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

Data Download Date/Time: 09/29/2023 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,S,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOHLIN,M
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/30/2023 01:56:40

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:54	
Control Test	0.082	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:55	
Control Test	0.082	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 10/25/2023 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ORDOQUI,DANIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211122
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WHITEHEAD

DATE/TIME OF TEST: 09/30/2023 21:56:51

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.082	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.300	21:52
Breath Volume	1.789	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:54
Subject Sample # 2	0.289	21:55
Breath Volume	1.511	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 10/25/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WOOD,TIMOTHY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MCWHIRTER

DATE/TIME OF TEST: 09/01/2023 03:59:07

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.244	03:54
Breath Volume	2.386	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.249	03:57
Breath Volume	5.179	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEWART, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEWART, WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: HIGHERS JR

DATE/TIME OF TEST: 09/01/2023 22:36:33

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:34	
Control Test	0.080	22:34	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:35	
Control Test	0.080	22:35	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MORALES RODRIGUEZ

DATE/TIME OF TEST: 09/03/2023 04:29:11

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.165	04:24
Breath Volume	2.183	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.162	04:27
Breath Volume	2.121	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT, ADAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SCRUGGS, JADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:34

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 09/03/2023 09:01:08

Diagnostics Check	OK	08:55	
Air Blank	0.000	08:55	
Control Test	0.082	08:55	
Air Blank	0.000	08:56	
Subject Sample # 1	0.000	08:57	
Breath Volume	1.937	08:57	
Air Blank	0.000	08:58	
Air Blank		08:59	RFI Detect
Air Blank	0.000	09:00	
Control Test	0.082	09:00	
Air Blank	0.000	09:01	
Diagnostics Check	OK	09:01	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT, ADAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SCRUGGS, JADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:34

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 09/03/2023 09:09:47

Diagnostics Check	OK	09:03
Air Blank	0.000	09:03
Control Test	0.081	09:04
Air Blank	0.000	09:04
Subject Sample # 1	0.000	09:05
Breath Volume	2.300	09:05
Air Blank	0.000	09:05
Air Blank	0.000	09:07
Subject Sample # 2	0.000	09:08
Breath Volume	2.191	09:08
Air Blank	0.000	09:08
Control Test	0.081	09:09
Air Blank	0.000	09:09
Diagnostics Check	OK	09:09

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SAENZ,MELISSA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 09/03/2023 23:55:28

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.081	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:54	
Control Test	0.081	23:54	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: CHERAMIE

DATE/TIME OF TEST: 09/04/2023 23:02:27

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:54	
Control Test	0.081	22:54	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:58	No Sample Provided
Breath Volume	0.000	22:58	
Air Blank	0.000	22:58	
Air Blank	0.000	23:00	
Subject Sample	# 2	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:01	
Control Test	0.081	23:01	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: GAIO

DATE/TIME OF TEST: 09/05/2023 20:00:40

Diagnostics Check	OK	19:52	
Air Blank	0.000	19:52	
Control Test	0.081	19:52	
Air Blank	0.000	19:53	
Subject Sample	# 1	19:56	No Sample Provided
Breath Volume	0.000	19:56	
Air Blank	0.000	19:57	
Air Blank	0.000	19:59	
Subject Sample	# 2	19:59	Subject Test Refused
Breath Volume	0.000	19:59	
Air Blank	0.000	19:59	
Control Test	0.081	20:00	
Air Blank	0.000	20:00	
Diagnostics Check	OK	20:00	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 09/08/2023 00:33:22

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.125	00:28
Breath Volume	1.578	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.124	00:31
Breath Volume	1.554	00:31
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNIS,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HENNIS,TIMOTHY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DOS SANTOS

DATE/TIME OF TEST: 09/10/2023 04:00:41

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.152	03:56
Breath Volume	2.085	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.155	03:59
Breath Volume	3.441	03:59
Air Blank	0.000	03:59
Control Test	0.081	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WOOD,TIMOTHY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GAMBLE

DATE/TIME OF TEST: 09/10/2023 21:44:51

Diagnostics Check	OK	21:33	
Air Blank	0.000	21:33	
Control Test	0.081	21:34	
Air Blank	0.000	21:34	
Subject Sample	# 1	21:37	No Sample Provided
Breath Volume	0.000	21:37	
Air Blank	0.000	21:38	
Air Blank	0.000	21:40	
Subject Sample	# 2	21:43	No Sample Provided
Breath Volume	0.000	21:43	
Air Blank	0.000	21:43	
Control Test	0.080	21:44	
Air Blank	0.000	21:44	
Diagnostics Check	OK	21:44	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELTON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELTON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 09/10/2023 22:15:48

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.081	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.206	22:11
Breath Volume	3.171	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.209	22:14
Breath Volume	2.449	22:14
Air Blank	0.000	22:14
Control Test	0.080	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUN DORIA,NORMA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAUN DORIA,NORMA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 09/13/2023 03:54:22

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:47	No Sample Provided
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Air Blank	0.000	03:49	
Subject Sample	# 2	03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.080	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MURPHY,SEAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: SALAZAR DUBON

DATE/TIME OF TEST: 09/14/2023 04:41:32

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.155	04:37
Breath Volume	2.675	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.163	04:40
Breath Volume	3.492	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RILEY,DEJA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEARCE,GRAHAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/16/2023 03:51:16

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.081	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:49	Subject Test Refused
Breath Volume	0.000	03:49	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 09/17/2023 00:52:33

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.082	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Control Test	0.081	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RILEY,DEJA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RILEY,DEJA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: JACO-GARCIA

DATE/TIME OF TEST: 09/17/2023 03:57:32

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.081	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1 0.102	03:50	Volume Not Met
Breath Volume	0.261	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 2 0.097	03:55	Volume Not Met
Breath Volume	1.074	03:55	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RILEY,DEJA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RILEY,DEJA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: JACO-GARCIA

DATE/TIME OF TEST: 09/17/2023 04:12:45

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.108	04:05	Slope Not Level
Breath Volume	1.156	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 2 0.104	04:11	Volume Not Met
Breath Volume	0.113	04:11	
Air Blank	0.000	04:11	
Control Test	0.078	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RILEY,DEJA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RILEY,DEJA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JACO-GARCIA

DATE/TIME OF TEST: 09/17/2023 04:32:26

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.080	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1 0.098	04:25	Volume Not Met
Breath Volume	0.117	04:25	
Air Blank	0.000	04:26	
Air Blank	0.000	04:27	
Subject Sample	# 2 0.038	04:30	Volume Not Met
Breath Volume	0.031	04:30	
Air Blank	0.000	04:31	
Control Test	0.080	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RILEY,DEJA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RILEY,DEJA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JACO-GARCIA

DATE/TIME OF TEST: 09/17/2023 04:39:13

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.081	04:37	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:38	
Control Test	0.081	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: SCHAFF

DATE/TIME OF TEST: 09/19/2023 23:29:57

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.081	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:29	
Control Test	0.081	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: LABRECQUE

DATE/TIME OF TEST: 09/21/2023 05:30:32

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.082	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.144	05:26
Breath Volume	1.652	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.144	05:29
Breath Volume	1.464	05:29
Air Blank	0.000	05:29
Control Test	0.081	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBO,GREGORY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUBO,GREGORY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JORAN-VARGAS

DATE/TIME OF TEST: 09/24/2023 01:24:26

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.082	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.081	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MURPHY,SEAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MOLINA CERRANO

DATE/TIME OF TEST: 09/24/2023 03:11:10

Diagnostics Check	OK	03:10	
Air Blank		03:10	RFI Detect
Air Blank	0.000	03:11	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MURPHY,SEAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MOLINA CERRANO

DATE/TIME OF TEST: 09/24/2023 03:17:58

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.081	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Control Test	0.082	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARTER,LARRY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: QUINONEZ

DATE/TIME OF TEST: 09/24/2023 04:13:29

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.081	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:06	No Sample Provided
Breath Volume	0.000	04:06	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2	04:12	No Sample Provided
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Control Test	0.082	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBO,GREGORY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUBO,GREGORY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/24/2023 04:26:39

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.082	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.181	04:22
Breath Volume	2.800	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.176	04:25
Breath Volume	2.660	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: RIVERA RIVERA

DATE/TIME OF TEST: 09/24/2023 04:38:35

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.137	04:34
Breath Volume	2.933	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.136	04:37
Breath Volume	3.062	04:37
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KITTRELL, WILLIAM, G
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KITTRELL, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/24/2023 04:52:14

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.081	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.115	04:47
Breath Volume	1.425	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.111	04:50
Breath Volume	1.445	04:50
Air Blank	0.000	04:51
Control Test	0.082	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DURAN,EDDIE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 09/24/2023 05:18:16

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.082	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.183	05:13
Breath Volume	2.015	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.173	05:16
Breath Volume	1.625	05:16
Air Blank	0.000	05:17
Control Test	0.081	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ESCRIBANO,DAVID
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 09/24/2023 13:47:38

Diagnostics Check	OK	13:44	
Air Blank	0.000	13:45	
Control Test	0.080	13:45	
Air Blank	0.000	13:46	
Subject Sample	# 1	13:46	Subject Test Refused
Breath Volume	0.000	13:46	
Air Blank	0.000	13:46	
Control Test	0.081	13:47	
Air Blank	0.000	13:47	
Diagnostics Check	OK	13:47	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEWART, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEWART, WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: KNIGHTON

DATE/TIME OF TEST: 09/26/2023 21:45:56

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.080	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:44	Subject Test Refused
Breath Volume	0.000	21:44	
Air Blank	0.000	21:45	
Control Test	0.081	21:45	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: THRAILKILL

DATE/TIME OF TEST: 09/28/2023 03:28:07

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.081	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:23	No Sample Provided
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 2	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.081	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 09/28/2023 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEWART, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEWART, WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: EACHOLES

DATE/TIME OF TEST: 09/30/2023 03:49:11

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.224	03:44
Breath Volume	1.152	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.216	03:47
Breath Volume	1.304	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 10/30/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVER, WILLIAM, H
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: OLIVER, WILLIAM
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-4
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: BUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: HUDGINS

DATE/TIME OF TEST: 09/30/2023 18:54:57

Diagnostics Check	OK	18:49
Air Blank	0.000	18:49
Control Test	0.081	18:49
Air Blank	0.000	18:50
Subject Sample # 1	0.085	18:50
Breath Volume	2.964	18:50
Air Blank	0.000	18:51
Air Blank	0.000	18:53
Subject Sample # 2	0.083	18:53
Breath Volume	3.207	18:53
Air Blank	0.000	18:54
Control Test	0.081	18:54
Air Blank	0.000	18:54
Diagnostics Check	OK	18:54

Data Download Date/Time: 10/30/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BURGESS,JOSHUA,L

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: MINNIFIELD,SAM

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 08/31/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402590345

DRY GAS EXPIRY DATE: 11/04/2025

VIOLATION CODE: DUI

ARREST TIME: 01:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ORELLANA

DATE/TIME OF TEST: 09/12/2023 02:24:36

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.082	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.183	02:19
Breath Volume	1.238	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.179	02:22
Breath Volume	1.484	02:22
Air Blank	0.000	02:23
Control Test	0.078	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 09/29/2023 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRIE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: CURRIE,MATTHEW
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 08/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/12/2023 23:15:08

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.083	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.233	23:10
Breath Volume	1.417	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.225	23:13
Breath Volume	3.781	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 09/29/2023 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: WALTER,KATELAND
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 08/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 09/17/2023 14:13:08

Diagnostics Check	OK	14:10	
Air Blank	0.000	14:11	
Control Test	0.083	14:11	
Air Blank	0.000	14:12	
Subject Sample	# 1	14:12	Subject Test Refused
Breath Volume	0.000	14:12	
Air Blank		14:12	RFI Detect
Air Blank	0.000	14:13	

Data Download Date/Time: 09/29/2023 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRY,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: PHILLIPS,MATHEW
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 08/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CALLOWAY

DATE/TIME OF TEST: 09/21/2023 22:35:00

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.081	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:33	Subject Test Refused
Breath Volume	0.000	22:33	
Air Blank	0.000	22:34	
Control Test	0.081	22:34	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 09/29/2023 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FLINN,AARON,D

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: FLINN,AARON

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 08/31/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 402590345

DRY GAS EXPIRY DATE: 11/04/2025

VIOLATION CODE: DUI

ARREST TIME: 23:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 09/22/2023 00:15:46

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.082	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.198	00:11
Breath Volume	3.414	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.199	00:14
Breath Volume	1.972	00:14
Air Blank	0.000	00:14
Control Test	0.081	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 09/29/2023 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AULTMAN,BRENDON,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: MINNIFIELD,SAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 08/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 09/23/2023 02:01:27

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.083	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.121	01:56
Breath Volume	1.675	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.110	01:59
Breath Volume	1.429	01:59
Air Blank	0.000	02:00
Control Test	0.082	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 09/29/2023 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AULTMAN,BRENDON,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: MINNIFIELD,SAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 08/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: TYE

DATE/TIME OF TEST: 09/23/2023 23:52:20

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.084	23:51	
Air Blank		23:52	RFI Detect
Air Blank		23:52	RFI Detect

Data Download Date/Time: 09/29/2023 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AULTMAN,BRENDON,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: MINNIFIELD,SAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 08/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #:]
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: TYE

DATE/TIME OF TEST: 09/23/2023 23:58:03

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.084	23:56	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:56	Subject Test Refused
Breath Volume	0.000	23:56	
Air Blank	0.000	23:57	
Control Test	0.084	23:57	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 09/29/2023 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 09/04/2023 18:28:55

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.079	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.000	18:24
Breath Volume	3.945	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.000	18:27
Breath Volume	3.500	18:27
Air Blank	0.000	18:28
Control Test	0.080	18:28
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 09/25/2023 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 09/05/2023 22:29:56

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.250	22:25
Breath Volume	2.359	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:27
Subject Sample # 2	0.252	22:28
Breath Volume	2.125	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 09/25/2023 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: FLEEK

DATE/TIME OF TEST: 09/09/2023 03:26:18

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.271	03:21
Breath Volume	3.906	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.252	03:24
Breath Volume	2.679	03:24
Air Blank	0.000	03:25
Control Test	0.078	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 09/25/2023 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MATTISON

DATE/TIME OF TEST: 09/10/2023 02:24:23

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.090	02:20
Breath Volume	3.371	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.084	02:22
Breath Volume	3.171	02:22
Air Blank	0.000	02:23
Control Test	0.078	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 09/25/2023 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: HARRAH

DATE/TIME OF TEST: 09/02/2023 21:45:19

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.082	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.032	21:40
Breath Volume	1.542	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.030	21:43
Breath Volume	1.523	21:43
Air Blank	0.000	21:44
Control Test	0.082	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 09/05/2023 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: JIMENEZ DE LA CRUZ

DATE/TIME OF TEST: 09/02/2023 22:47:50

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.082	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.197	22:43
Breath Volume	1.546	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:45
Subject Sample # 2	0.185	22:46
Breath Volume	2.523	22:46
Air Blank	0.000	22:46
Control Test	0.082	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 09/05/2023 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HOLOWAY

DATE/TIME OF TEST: 09/03/2023 03:47:02

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.082	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.058	03:42
Breath Volume	2.171	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.057	03:45
Breath Volume	2.183	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:47

Data Download Date/Time: 09/05/2023 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: LANTIGUA

DATE/TIME OF TEST: 09/03/2023 04:04:19

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.074	03:59
Breath Volume	3.828	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.076	04:02
Breath Volume	3.718	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 09/05/2023 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: GARCIA PEREZ

DATE/TIME OF TEST: 09/04/2023 04:15:13

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.082	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1 0.165	04:07	Volume Not Met
Breath Volume	0.371	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2 0.174	04:10	
Breath Volume	1.984	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 3 0.169	04:13	
Breath Volume	1.562	04:13	
Air Blank	0.000	04:14	
Control Test	0.081	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 09/05/2023 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: BENZRIHEM

DATE/TIME OF TEST: 09/04/2023 23:48:13

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.082	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.000	23:43
Breath Volume	1.312	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.000	23:46
Breath Volume	1.996	23:46
Air Blank	0.000	23:47
Control Test	0.082	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 09/05/2023 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CATANZARO JR

DATE/TIME OF TEST: 09/26/2023 03:38:31

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.258	03:30
Breath Volume	1.953	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.233	03:33
Breath Volume	1.574	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 3	0.253	03:36
Breath Volume	2.070	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 10/02/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: WILES

DATE/TIME OF TEST: 09/26/2023 12:23:51

Diagnostics Check	OK	12:17
Air Blank	0.000	12:17
Control Test	0.081	12:18
Air Blank	0.000	12:18
Subject Sample # 1	0.000	12:19
Breath Volume	1.886	12:19
Air Blank	0.000	12:19
Air Blank	0.000	12:21
Subject Sample # 2	0.000	12:22
Breath Volume	2.167	12:22
Air Blank	0.000	12:22
Control Test	0.081	12:23
Air Blank	0.000	12:23
Diagnostics Check	OK	12:23

Data Download Date/Time: 10/02/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MAKSUTI

DATE/TIME OF TEST: 09/27/2023 01:46:17

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.082	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.193	01:41
Breath Volume	1.835	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.188	01:44
Breath Volume	2.191	01:44
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 10/02/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI, H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ARANDA

DATE/TIME OF TEST: 09/27/2023 02:49:49

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.081	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.036	02:45
Breath Volume	1.539	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.034	02:48
Breath Volume	2.226	02:48
Air Blank	0.000	02:48
Control Test	0.082	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 10/02/2023 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: BOOZE JR

DATE/TIME OF TEST: 09/28/2023 01:54:34

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.082	01:40	
Air Blank	0.000	01:40	
Subject Sample	# 1 0.091	01:43	Volume Not Met
Breath Volume	0.132	01:43	
Air Blank	0.000	01:44	
Air Blank	0.000	01:46	
Subject Sample	# 2 0.131	01:48	
Breath Volume	1.972	01:48	
Air Blank	0.000	01:48	
Air Blank	0.000	01:50	
Subject Sample	# 3 0.129	01:53	
Breath Volume	1.839	01:53	
Air Blank	0.000	01:53	
Control Test	0.080	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 10/02/2023 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 09/29/2023 02:29:37

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.158	02:25
Breath Volume	1.582	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.167	02:28
Breath Volume	1.910	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 10/02/2023 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: D'ARCY

DATE/TIME OF TEST: 09/29/2023 23:54:31

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:47	No Sample Provided
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Air Blank	0.000	23:49	
Subject Sample	# 2	0.226	23:52
Breath Volume	0.496	23:52	Volume Not Met
Air Blank	0.000	23:53	
Control Test	0.079	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 10/02/2023 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: D'ARCY

DATE/TIME OF TEST: 09/30/2023 00:03:18

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.078	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.254	23:58
Breath Volume	1.156	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.255	00:01
Breath Volume	1.425	00:01
Air Blank	0.000	00:02
Control Test	0.078	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 10/02/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 09/30/2023 01:14:36

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.152	01:09
Breath Volume	1.601	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.149	01:13
Breath Volume	2.191	01:13
Air Blank	0.000	01:13
Control Test	0.084	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 10/02/2023 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: LEON CUEVA

DATE/TIME OF TEST: 09/30/2023 02:51:07

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.127	02:46
Breath Volume	1.519	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.136	02:49
Breath Volume	1.902	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 10/02/2023 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 09/30/2023 03:46:35

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.080	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Control Test	0.080	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 10/02/2023 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BEJAR

DATE/TIME OF TEST: 09/30/2023 04:33:41

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.207	04:29
Breath Volume	1.941	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.203	04:32
Breath Volume	1.207	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 10/02/2023 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: MCDONALD-SMITH

DATE/TIME OF TEST: 09/30/2023 20:14:54

Diagnostics Check	OK	20:07
Air Blank	0.000	20:08
Control Test	0.080	20:08
Air Blank	0.000	20:09
Subject Sample # 1	0.335	20:10
Breath Volume	2.191	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:12
Subject Sample # 2	0.336	20:13
Breath Volume	1.671	20:13
Air Blank	0.000	20:13
Control Test	0.079	20:14
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 10/02/2023 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WORKMAN

DATE/TIME OF TEST: 09/01/2023 04:13:11

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.083	04:11	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.083	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 09/06/2023 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 09/02/2023 23:27:49

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.084	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.152	23:23
Breath Volume	1.484	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:25
Subject Sample # 2	0.155	23:26
Breath Volume	1.183	23:26
Air Blank	0.000	23:26
Control Test	0.083	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 09/06/2023 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RYAN,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: SANCHEZ GALATUAR

DATE/TIME OF TEST: 09/04/2023 05:12:31

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:10	
Control Test	0.083	05:10	
Air Blank	0.000	05:10	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:11	
Control Test	0.083	05:11	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 09/06/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ADAMS,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 09/05/2023 18:16:46

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.083	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.153	18:12
Breath Volume	1.613	18:12
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.157	18:15
Breath Volume	1.683	18:15
Air Blank	0.000	18:15
Control Test	0.081	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 09/06/2023 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,D
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 13:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:34

SUBJECT'S LAST NAME: KELLERMAN

DATE/TIME OF TEST: 09/26/2023 15:22:07

Diagnostics Check	OK	15:16
Air Blank	0.000	15:16
Control Test	0.081	15:16
Air Blank	0.000	15:17
Subject Sample # 1	0.000	15:17
Breath Volume	1.718	15:17
Air Blank	0.000	15:18
Air Blank	0.000	15:20
Subject Sample # 2	0.000	15:20
Breath Volume	2.492	15:20
Air Blank	0.000	15:21
Control Test	0.080	15:21
Air Blank	0.000	15:22
Diagnostics Check	OK	15:22

Data Download Date/Time: 10/18/2023 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JASON,A
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MORALES,BRIAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 09/01/2023 23:25:05

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.204	23:20
Breath Volume	1.410	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.192	23:23
Breath Volume	1.132	23:23
Air Blank	0.000	23:24
Control Test	0.078	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 09/29/2023 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHNIBBEN,BRIAN,K
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MCDOWELL,ADRIAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: BARTSCH

DATE/TIME OF TEST: 09/02/2023 17:05:27

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:52	
Control Test	0.081	16:52	
Air Blank	0.000	16:53	
Subject Sample	# 1 0.138	16:54	
Breath Volume	1.988	16:54	
Air Blank	0.000	16:55	
Air Blank	0.000	16:57	
Subject Sample	# 2 0.125	17:00	Volume Not Met
Breath Volume	1.042	17:00	
Air Blank	0.000	17:01	
Air Blank	0.000	17:03	
Subject Sample	# 3 0.146	17:03	
Breath Volume	1.558	17:03	
Air Blank	0.000	17:04	
Control Test	0.080	17:04	
Air Blank	0.000	17:05	
Diagnostics Check	OK	17:05	

Data Download Date/Time: 09/29/2023 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOFE,DEREK,T
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: BACON,ZAKARIAH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 09/04/2023 04:33:42

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.143	04:29
Breath Volume	1.742	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.150	04:32
Breath Volume	2.007	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 09/29/2023 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JASON,A
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: BRASHIER,PHILLIP
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: JARREAU

DATE/TIME OF TEST: 09/09/2023 01:40:06

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.082	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.000	01:35
Breath Volume	1.457	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.000	01:38
Breath Volume	1.648	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 09/29/2023 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH, THOMAS,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: COATS, OFFICER
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: CRAFT

DATE/TIME OF TEST: 09/10/2023 06:46:13

Diagnostics Check	OK	06:40
Air Blank	0.000	06:40
Control Test	0.081	06:40
Air Blank	0.000	06:41
Subject Sample # 1	0.000	06:42
Breath Volume	2.902	06:42
Air Blank	0.000	06:42
Air Blank	0.000	06:44
Subject Sample # 2	0.000	06:44
Breath Volume	3.062	06:44
Air Blank	0.000	06:45
Control Test	0.080	06:45
Air Blank	0.000	06:46
Diagnostics Check	OK	06:46

Data Download Date/Time: 09/29/2023 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VINES,KIRSTEN,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: PHILLIPS,JAMES
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: SANSEVERE

DATE/TIME OF TEST: 09/13/2023 22:01:07

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.000	21:56
Breath Volume	1.378	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.000	21:59
Breath Volume	1.816	21:59
Air Blank	0.000	22:00
Control Test	0.081	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 09/29/2023 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH, THOMAS,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MENDOZA, JONATHAN
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 08/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 11:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:03

SUBJECT'S LAST NAME: HATCHETT

DATE/TIME OF TEST: 09/23/2023 11:38:50

Diagnostics Check	OK	11:30	
Air Blank	0.000	11:30	
Control Test	0.081	11:31	
Air Blank	0.000	11:31	
Subject Sample	# 1	11:34	No Sample Provided
Breath Volume	0.000	11:34	
Air Blank	0.000	11:35	
Air Blank	0.000	11:37	
Subject Sample	# 2	11:37	Subject Test Refused
Breath Volume	0.000	11:37	
Air Blank	0.000	11:37	
Control Test	0.081	11:38	
Air Blank	0.000	11:38	
Diagnostics Check	OK	11:38	

Data Download Date/Time: 09/29/2023 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORRIS,CASSANDRA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: CORDOVA CARRIO

DATE/TIME OF TEST: 09/02/2023 01:57:47

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.079	01:48	
Air Blank	0.000	01:49	
Subject Sample # 1	0.240	01:50	
Breath Volume	1.777	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:52	
Subject Sample # 2	0.219	01:53	
Breath Volume	3.421	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample # 3	0.175	01:56	
Breath Volume	2.851	01:56	
Air Blank		01:56	Improper Sample
Air Blank	0.000	01:57	

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORRIS,CASSANDRA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: CORDOVA CARRIO

DATE/TIME OF TEST: 09/02/2023 02:09:45

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.076	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.222	02:04
Breath Volume	1.750	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.228	02:07
Breath Volume	5.164	02:07
Air Blank	0.000	02:08
Control Test	0.077	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: KRILAVICIUTE

DATE/TIME OF TEST: 09/03/2023 03:35:19

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.207	03:30
Breath Volume	2.386	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.199	03:33
Breath Volume	2.335	03:33
Air Blank	0.000	03:34
Control Test	0.077	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGES,JARRETT,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BRIDGES,JARRETT
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:36

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 09/04/2023 10:14:34

Diagnostics Check	OK	10:08
Air Blank	0.000	10:08
Control Test	0.078	10:09
Air Blank	0.000	10:09
Subject Sample # 1	0.159	10:09
Breath Volume	2.761	10:09
Air Blank	0.000	10:10
Air Blank	0.000	10:12
Subject Sample # 2	0.161	10:12
Breath Volume	3.093	10:12
Air Blank	0.000	10:13
Control Test	0.077	10:13
Air Blank	0.000	10:14
Diagnostics Check	OK	10:14

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFER,BROOKS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: DEMPSEY,ALEXIS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: KOSTYSHYN

DATE/TIME OF TEST: 09/07/2023 22:43:56

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.079	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.143	22:39
Breath Volume	2.277	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.145	22:42
Breath Volume	2.652	22:42
Air Blank	0.000	22:42
Control Test	0.076	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,TORI,D
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON,TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 09/10/2023 23:00:10

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.217	22:54
Breath Volume	2.535	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.226	22:58
Breath Volume	2.027	22:58
Air Blank	0.000	22:59
Control Test	0.078	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GARGANO

DATE/TIME OF TEST: 09/15/2023 00:40:58

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.000	00:35
Breath Volume	1.406	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.000	00:39
Breath Volume	1.937	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: WOJCINSKI JR

DATE/TIME OF TEST: 09/15/2023 03:08:38

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.126	03:03
Breath Volume	2.871	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.123	03:06
Breath Volume	2.371	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: DONALD

DATE/TIME OF TEST: 09/17/2023 02:01:19

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.099	01:56
Breath Volume	1.812	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.099	01:59
Breath Volume	2.074	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: VEGA,JAHMIL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: SAN MARTIN RUIZ

DATE/TIME OF TEST: 09/17/2023 05:26:06

Diagnostics Check	OK	05:18
Air Blank	0.000	05:19
Control Test	0.079	05:19
Air Blank	0.000	05:20
Subject Sample # 1	0.136	05:21
Breath Volume	2.429	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:23
Subject Sample # 2	0.134	05:24
Breath Volume	2.117	05:24
Air Blank	0.000	05:25
Control Test	0.077	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: SEAMAN

DATE/TIME OF TEST: 09/17/2023 23:34:43

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.078	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.278	23:29
Breath Volume	3.218	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.267	23:32
Breath Volume	3.019	23:32
Air Blank	0.000	23:33
Control Test	0.078	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,TORL,D
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: HERNANDEZ,P
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: SEBASTIAN

DATE/TIME OF TEST: 09/18/2023 20:40:40

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.080	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.230	20:36
Breath Volume	2.031	20:36
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.238	20:38
Breath Volume	1.968	20:38
Air Blank	0.000	20:39
Control Test	0.079	20:39
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,T,D
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: HERNANDEZ,P
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: SEBASTIAN

DATE/TIME OF TEST: 09/18/2023 20:52:11

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.078	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.241	20:47
Breath Volume	2.464	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:49
Subject Sample # 2	0.225	20:50
Breath Volume	1.417	20:50
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 09/19/2023 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ESPINOZA SALAZAR

DATE/TIME OF TEST: 09/02/2023 03:14:19

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.083	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1 0.131	03:05	Volume Not Met
Breath Volume	0.304	03:05	
Air Blank	0.000	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 2 0.136	03:08	
Breath Volume	1.484	03:08	
Air Blank	0.000	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 3 0.147	03:12	
Breath Volume	1.695	03:12	
Air Blank	0.000	03:13	
Control Test	0.082	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 09/25/2023 01:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,C,D
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: PLYMAL

DATE/TIME OF TEST: 09/03/2023 03:07:24

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.083	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.220	03:02
Breath Volume	1.851	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.212	03:05
Breath Volume	1.636	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 09/25/2023 01:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/04/2023 01:24:02

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.083	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.076	01:19
Breath Volume	1.238	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.075	01:22
Breath Volume	1.347	01:22
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:24

Data Download Date/Time: 09/25/2023 01:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: TILLMAN

DATE/TIME OF TEST: 09/23/2023 16:42:28

Diagnostics Check	OK	16:39	
Air Blank	0.000	16:39	
Control Test	0.084	16:40	
Air Blank	0.000	16:40	
Subject Sample	# 1	16:41	Subject Test Refused
Breath Volume	0.000	16:41	
Air Blank	0.000	16:41	
Control Test	0.083	16:41	
Air Blank	0.000	16:42	
Diagnostics Check	OK	16:42	

Data Download Date/Time: 09/25/2023 01:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: SHERWOOD

DATE/TIME OF TEST: 09/23/2023 19:48:04

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.084	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.166	19:43
Breath Volume	1.601	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.167	19:46
Breath Volume	1.539	19:46
Air Blank	0.000	19:47
Control Test	0.082	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:48

Data Download Date/Time: 09/25/2023 01:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:58

SUBJECT'S LAST NAME: ALVARADOESPINOZA

DATE/TIME OF TEST: 09/03/2023 11:33:24

Diagnostics Check	OK	11:27
Air Blank	0.000	11:27
Control Test	0.079	11:27
Air Blank	0.000	11:28
Subject Sample # 1	0.092	11:28
Breath Volume	1.679	11:28
Air Blank	0.000	11:29
Air Blank	0.000	11:31
Subject Sample # 2	0.099	11:31
Breath Volume	2.054	11:31
Air Blank	0.000	11:32
Control Test	0.079	11:32
Air Blank	0.000	11:33
Diagnostics Check	OK	11:33

Data Download Date/Time: 10/05/2023 04:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: O'CONNOR

DATE/TIME OF TEST: 09/03/2023 19:04:46

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.078	18:52
Air Blank	0.000	18:53
Subject Sample # 1	0.174	18:56
Breath Volume	1.621	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:58
Subject Sample # 2	0.210	18:59
Breath Volume	4.816	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 3	0.206	19:03
Breath Volume	4.921	19:03
Air Blank	0.000	19:03
Control Test	0.075	19:04
Air Blank	0.000	19:04
Diagnostics Check	OK	19:04

Data Download Date/Time: 10/05/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIEEN,BOB,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAINTILIEEN,BOB
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/04/2023 09:19:24

Diagnostics Check	OK	09:11
Air Blank	0.000	09:11
Control Test	0.079	09:12
Air Blank	0.000	09:12
Subject Sample # 1	0.000	09:13
Breath Volume	1.593	09:13
Air Blank	0.000	09:14
Air Blank	0.000	09:16
Subject Sample # 2	0.000	09:18
Breath Volume	1.355	09:18
Air Blank	0.000	09:18
Control Test	0.079	09:18
Air Blank	0.000	09:19
Diagnostics Check	OK	09:19

Data Download Date/Time: 10/05/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RODRIGUEZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 09/09/2023 00:45:18

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.078	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.078	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 10/05/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYALA,RAUDEL,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: PINO,ALEJANDRO
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:34

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 09/09/2023 08:58:10

Diagnostics Check	OK	08:54	
Air Blank	0.000	08:55	
Control Test	0.078	08:55	
Air Blank	0.000	08:56	
Subject Sample	# 1	08:56	Subject Test Refused
Breath Volume	0.000	08:56	
Air Blank	0.000	08:57	
Control Test	0.078	08:57	
Air Blank	0.000	08:58	
Diagnostics Check	OK	08:58	

Data Download Date/Time: 10/05/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JORDAN,C
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: RIVERA,JORDAN
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 09/13/2023 23:04:30

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:02	
Control Test	0.079	23:02	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:03	
Control Test	0.078	23:03	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 10/05/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: MALE

DATE/TIME OF TEST: 09/16/2023 04:51:26

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.077	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.149	04:47
Breath Volume	1.207	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.152	04:49
Breath Volume	1.312	04:49
Air Blank	0.000	04:50
Control Test	0.076	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 10/05/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MONTEZUMA

DATE/TIME OF TEST: 09/23/2023 03:35:39

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:34	
Control Test	0.078	03:35	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 10/05/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: TORRES-ECHEVARRIA

DATE/TIME OF TEST: 09/28/2023 04:08:53

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.078	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.077	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 10/05/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PALACIOS,L
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: GAINES

DATE/TIME OF TEST: 09/03/2023 07:38:45

Diagnostics Check	OK	07:29
Air Blank	0.000	07:29
Control Test	0.077	07:30
Air Blank	0.000	07:30
Subject Sample # 1	0.215	07:33
Breath Volume	1.675	07:33
Air Blank	0.000	07:33
Air Blank	0.000	07:35
Subject Sample # 2	0.217	07:36
Breath Volume	1.375	07:36
Air Blank	0.000	07:37
Control Test	0.077	07:38
Air Blank	0.000	07:38
Diagnostics Check	OK	07:38

Data Download Date/Time: 10/05/2023 04:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGILA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: TRIVELLA

DATE/TIME OF TEST: 09/09/2023 04:10:19

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.078	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:08	Subject Test Refused
Breath Volume	0.000	04:08	
Air Blank	0.000	04:09	
Control Test	0.079	04:09	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 10/05/2023 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 09/10/2023 02:59:50

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.150	02:54
Breath Volume	2.386	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.145	02:58
Breath Volume	2.097	02:58
Air Blank	0.000	02:58
Control Test	0.077	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 10/05/2023 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MYERS,W
ARRESTING OFFICER AGENCY: BAL HARBOUR VIL PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: GALAN

DATE/TIME OF TEST: 09/16/2023 05:24:06

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.078	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.129	05:19
Breath Volume	1.203	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:21
Subject Sample # 2	0.121	05:22
Breath Volume	1.167	05:22
Air Blank	0.000	05:23
Control Test	0.077	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:24

Data Download Date/Time: 10/05/2023 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RODRIGUEZ,J
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: VARGAS RODRIGUEZ

DATE/TIME OF TEST: 09/17/2023 00:01:54

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.077	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.140	23:56
Breath Volume	2.429	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.141	00:00
Breath Volume	1.437	00:00
Air Blank	0.000	00:00
Control Test	0.076	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 10/05/2023 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CEBALLOS

DATE/TIME OF TEST: 09/30/2023 04:53:36

Diagnostics Check	OK	04:52	
Air Blank		04:52	Ambient Fail
Air Blank	0.000	04:53	

Data Download Date/Time: 10/05/2023 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CEBALLOS

DATE/TIME OF TEST: 09/30/2023 04:57:36

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:56	
Control Test	0.076	04:56	Interferent Detect
Air Blank	0.000	04:57	

Data Download Date/Time: 10/05/2023 04:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: BELIDOR

DATE/TIME OF TEST: 09/18/2023 01:27:19

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.080	01:24	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 10/05/2023 04:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 09/23/2023 04:04:21

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.079	03:52	
Air Blank	0.000	03:52	
Subject Sample	# 1 0.244	03:54	
Breath Volume	1.386	03:54	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 2 0.239	03:59	Volume Not Met
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 3 0.251	04:02	
Breath Volume	1.664	04:02	
Air Blank	0.000	04:03	
Control Test	0.078	04:03	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 10/05/2023 04:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 09/01/2023 03:07:32

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.156	03:02
Breath Volume	2.980	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.151	03:05
Breath Volume	3.484	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 09/15/2023 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 09/01/2023 03:42:21

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.199	03:37
Breath Volume	1.871	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.193	03:40
Breath Volume	1.554	03:40
Air Blank	0.000	03:41
Control Test	0.077	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 09/15/2023 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: RODRIGUEZ ALVAREZ

DATE/TIME OF TEST: 09/01/2023 22:52:33

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:48	
Control Test	0.079	22:49	
Air Blank	0.000	22:49	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.079	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 09/15/2023 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/02/2023 01:54:18

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.114	01:49
Breath Volume	1.515	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.107	01:52
Breath Volume	1.316	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 09/15/2023 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DILLARD,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: BENITEZ CASTILLO

DATE/TIME OF TEST: 09/02/2023 03:18:07

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.153	03:13
Breath Volume	1.382	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.151	03:16
Breath Volume	2.394	03:16
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 09/15/2023 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: OH

DATE/TIME OF TEST: 09/02/2023 05:20:47

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.078	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.142	05:16
Breath Volume	4.925	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.144	05:19
Breath Volume	4.625	05:19
Air Blank	0.000	05:19
Control Test	0.078	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 09/15/2023 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: OERTH

DATE/TIME OF TEST: 09/02/2023 19:55:13

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.078	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.168	19:50
Breath Volume	1.531	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.175	19:53
Breath Volume	1.625	19:53
Air Blank	0.000	19:54
Control Test	0.079	19:54
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: AMSTER

DATE/TIME OF TEST: 09/02/2023 20:39:57

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.079	20:33
Air Blank	0.000	20:34
Subject Sample # 1	0.204	20:35
Breath Volume	1.359	20:35
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.213	20:38
Breath Volume	2.230	20:38
Air Blank	0.000	20:39
Control Test	0.078	20:39
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: AMAYA BONILLA

DATE/TIME OF TEST: 09/02/2023 23:24:22

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.151	23:19
Breath Volume	3.843	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.153	23:22
Breath Volume	2.988	23:22
Air Blank	0.000	23:23
Control Test	0.077	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: ESCOBAR TORRES

DATE/TIME OF TEST: 09/02/2023 23:48:21

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.077	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.141	23:43
Breath Volume	3.695	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.140	23:46
Breath Volume	3.304	23:46
Air Blank	0.000	23:47
Control Test	0.077	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,JUAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 09/03/2023 00:39:56

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.162	00:34
Breath Volume	1.164	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.167	00:38
Breath Volume	1.929	00:38
Air Blank	0.000	00:39
Control Test	0.077	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 09/03/2023 01:02:53

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.078	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.000	00:58
Breath Volume	2.148	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.000	01:01
Breath Volume	2.304	01:01
Air Blank	0.000	01:01
Control Test	0.078	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: AMORIELLO

DATE/TIME OF TEST: 09/03/2023 02:25:46

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.238	02:21
Breath Volume	2.804	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.234	02:24
Breath Volume	2.906	02:24
Air Blank	0.000	02:24
Control Test	0.077	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLACKMAN, VICTOR
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JULIEN

DATE/TIME OF TEST: 09/03/2023 03:53:06

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.078	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.271	03:48
Breath Volume	2.601	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.276	03:51
Breath Volume	1.125	03:51
Air Blank	0.000	03:52
Control Test	0.075	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: CHAMU CRISTOBAL

DATE/TIME OF TEST: 09/03/2023 04:10:44

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.078	04:03	
Air Blank	0.000	04:03	
Subject Sample # 1	0.245	04:05	
Breath Volume	1.816	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample # 2	0.247	04:09	
Breath Volume	2.015	04:09	
Air Blank	0.000	04:09	
Control Test	0.074	04:10	Control Outside Tolerance
Air Blank	0.000	04:10	

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CHAMU CRISTOBAL

DATE/TIME OF TEST: 09/03/2023 04:53:50

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.078	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.232	04:48
Breath Volume	1.839	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.225	04:52
Breath Volume	1.976	04:52
Air Blank	0.000	04:52
Control Test	0.077	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BAGDZIUNAS,BRYAN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: RAMIREZ PASCUAL

DATE/TIME OF TEST: 09/03/2023 06:04:47

Diagnostics Check	OK	05:58
Air Blank	0.000	05:58
Control Test	0.078	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.191	06:00
Breath Volume	3.210	06:00
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.190	06:03
Breath Volume	3.496	06:03
Air Blank	0.000	06:03
Control Test	0.078	06:04
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: QUINONEZ PEREZ

DATE/TIME OF TEST: 09/03/2023 19:38:12

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.079	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.276	19:29
Breath Volume	2.261	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.254	19:33
Breath Volume	1.898	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:35
Subject Sample # 3	0.274	19:36
Breath Volume	2.175	19:36
Air Blank	0.000	19:37
Control Test	0.077	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 09/03/2023 20:32:00

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.078	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.025	20:27
Breath Volume	3.027	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:29
Subject Sample # 2	0.024	20:30
Breath Volume	2.500	20:30
Air Blank	0.000	20:31
Control Test	0.078	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:32

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: CECERE

DATE/TIME OF TEST: 09/03/2023 22:25:42

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.192	22:21
Breath Volume	2.828	22:21
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.177	22:24
Breath Volume	2.722	22:24
Air Blank	0.000	22:24
Control Test	0.078	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 09/03/2023 22:41:24

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.077	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.212	22:36
Breath Volume	1.664	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:38
Subject Sample # 2	0.219	22:39
Breath Volume	1.824	22:39
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/04/2023 01:18:20

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.139	01:13
Breath Volume	2.457	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.137	01:16
Breath Volume	2.292	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: MENDOZA CAMPOSECO

DATE/TIME OF TEST: 09/04/2023 04:50:51

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.152	04:46
Breath Volume	1.300	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.158	04:49
Breath Volume	2.394	04:49
Air Blank	0.000	04:49
Control Test	0.078	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: ROCHA ARZE

DATE/TIME OF TEST: 09/04/2023 22:24:49

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.174	22:20
Breath Volume	2.828	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:22
Subject Sample # 2	0.168	22:23
Breath Volume	2.257	22:23
Air Blank	0.000	22:23
Control Test	0.076	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: TALAVERA

DATE/TIME OF TEST: 09/05/2023 01:03:47

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.103	00:59
Breath Volume	2.093	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.104	01:02
Breath Volume	1.726	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: CAULEY

DATE/TIME OF TEST: 09/05/2023 01:30:18

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.189	01:25
Breath Volume	3.023	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.181	01:28
Breath Volume	3.097	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: DOMINGO MALDONADO

DATE/TIME OF TEST: 09/06/2023 00:17:45

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.238	00:13
Breath Volume	2.445	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.236	00:16
Breath Volume	1.976	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRENTMAN,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:44

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/06/2023 10:18:18

Diagnostics Check	OK	10:11
Air Blank	0.000	10:12
Control Test	0.079	10:12
Air Blank	0.000	10:12
Subject Sample # 1	0.219	10:13
Breath Volume	1.605	10:13
Air Blank	0.000	10:14
Air Blank	0.000	10:15
Subject Sample # 2	0.219	10:16
Breath Volume	1.957	10:16
Air Blank	0.000	10:17
Control Test	0.079	10:17
Air Blank	0.000	10:18
Diagnostics Check	OK	10:18

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: CARBO

DATE/TIME OF TEST: 09/07/2023 02:02:35

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.164	01:57
Breath Volume	1.867	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.167	02:00
Breath Volume	2.753	02:00
Air Blank	0.000	02:01
Control Test	0.078	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU, YASMINE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: VIGOA RODRIGUEZ

DATE/TIME OF TEST: 09/07/2023 02:26:34

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1 0.094	02:18	Volume Not Met
Breath Volume	0.058	02:18	
Air Blank	0.000	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 2 0.105	02:21	
Breath Volume	1.972	02:21	
Air Blank	0.000	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 3 0.104	02:24	
Breath Volume	2.433	02:24	
Air Blank	0.000	02:25	
Control Test	0.078	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 09/07/2023 03:24:04

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.078	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:19	No Sample Provided
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample	# 2	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE, DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/07/2023 03:45:22

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.205	03:40
Breath Volume	2.941	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.205	03:43
Breath Volume	2.917	03:43
Air Blank	0.000	03:44
Control Test	0.076	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SELAH

DATE/TIME OF TEST: 09/08/2023 00:24:51

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.031	00:20
Breath Volume	1.246	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.030	00:23
Breath Volume	1.125	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS, VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: BERMAN

DATE/TIME OF TEST: 09/08/2023 01:19:19

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.206	01:14
Breath Volume	2.375	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:16
Subject Sample # 2	0.201	01:17
Breath Volume	2.648	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ALVARADO SEPULVEDA

DATE/TIME OF TEST: 09/08/2023 01:49:47

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.151	01:44
Breath Volume	1.722	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.149	01:48
Breath Volume	1.761	01:48
Air Blank	0.000	01:48
Control Test	0.078	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALES-LOPEZ,KEVIN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: DIAS LOPES

DATE/TIME OF TEST: 09/08/2023 03:07:07

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.236	02:59
Breath Volume	2.296	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.207	03:02
Breath Volume	2.152	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 3	0.208	03:05
Breath Volume	1.839	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 09/08/2023 23:22:40

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.156	23:17
Breath Volume	2.027	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.159	23:21
Breath Volume	1.253	23:21
Air Blank	0.000	23:21
Control Test	0.083	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 09/09/2023 01:16:03

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.088	01:11
Breath Volume	1.722	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:13
Subject Sample # 2	0.087	01:14
Breath Volume	1.214	01:14
Air Blank	0.000	01:15
Control Test	0.077	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:16

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ACOSTA,NERELYS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: LIMERICK

DATE/TIME OF TEST: 09/09/2023 03:26:32

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.202	03:19	Volume Not Met
Breath Volume	1.023	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 2 0.194	03:24	Volume Not Met
Breath Volume	0.582	03:24	
Air Blank	0.000	03:25	
Control Test	0.077	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ACOSTA,NERELYS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: LIMERICK

DATE/TIME OF TEST: 09/09/2023 03:40:53

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:29	
Control Test	0.077	03:29	
Air Blank	0.000	03:30	
Subject Sample # 1	0.193	03:31	
Breath Volume	1.394	03:31	
Air Blank	0.000	03:31	
Air Blank	0.000	03:33	
Subject Sample # 2	0.178	03:36	Volume Not Met
Breath Volume	0.195	03:36	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample # 3		03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:39	
Control Test	0.077	03:40	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SPANGLER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:23

SUBJECT'S LAST NAME: BROWNE

DATE/TIME OF TEST: 09/09/2023 15:34:43

Diagnostics Check	OK	15:28
Air Blank	0.000	15:28
Control Test	0.079	15:29
Air Blank	0.000	15:29
Subject Sample # 1	0.250	15:30
Breath Volume	1.769	15:30
Air Blank	0.000	15:30
Air Blank	0.000	15:32
Subject Sample # 2	0.257	15:33
Breath Volume	1.621	15:33
Air Blank	0.000	15:33
Control Test	0.078	15:34
Air Blank	0.000	15:34
Diagnostics Check	OK	15:34

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CISNEROS LOPEZ

DATE/TIME OF TEST: 09/10/2023 03:32:54

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.255	03:28
Breath Volume	1.578	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.250	03:31
Breath Volume	2.183	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOOHEY,DOUGLAS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: OSORIO QUEVEDO

DATE/TIME OF TEST: 09/10/2023 03:59:14

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.208	03:54
Breath Volume	1.367	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.199	03:57
Breath Volume	1.800	03:57
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAGET,HORLKINS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:19

SUBJECT'S LAST NAME: CALLON

DATE/TIME OF TEST: 09/10/2023 15:25:14

Diagnostics Check	OK	15:16	
Air Blank	0.000	15:17	
Control Test	0.079	15:17	
Air Blank	0.000	15:18	
Subject Sample	# 1 0.171	15:21	Volume Not Met
Breath Volume	0.050	15:21	
Air Blank	0.000	15:22	
Air Blank	0.000	15:23	
Subject Sample	# 2	15:23	Subject Test Refused
Breath Volume	0.000	15:23	
Air Blank	0.000	15:24	
Control Test	0.075	15:24	
Air Blank	0.000	15:25	
Diagnostics Check	OK	15:25	

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: BROFF

DATE/TIME OF TEST: 09/10/2023 22:55:57

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.142	22:49
Breath Volume	1.617	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.139	22:54
Breath Volume	1.738	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MORALES MARTINEZ

DATE/TIME OF TEST: 09/10/2023 23:18:52

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.078	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.140	23:13
Breath Volume	1.753	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.135	23:17
Breath Volume	1.632	23:17
Air Blank	0.000	23:17
Control Test	0.078	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 09/11/2023 00:04:46

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.078	23:58
Air Blank	0.000	23:59
Subject Sample # 2	0.133	23:59
Breath Volume	1.269	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.135	00:03
Breath Volume	1.730	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEHENY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ZULETA SANCHEZ

DATE/TIME OF TEST: 09/11/2023 01:08:32

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.079	00:55	
Air Blank	0.000	00:56	
Subject Sample # 1	0.244	00:57	
Breath Volume	1.410	00:57	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample # 2	0.226	01:03	Volume Not Met
Breath Volume	0.320	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:05	
Subject Sample # 3	0.250	01:06	
Breath Volume	1.203	01:06	
Air Blank	0.000	01:07	
Control Test	0.076	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: US GONZALEZ

DATE/TIME OF TEST: 09/11/2023 03:08:00

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.145	03:03
Breath Volume	1.257	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.145	03:06
Breath Volume	2.566	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:08

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MAIONE,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/11/2023 18:08:25

Diagnostics Check	OK	17:58
Air Blank	0.000	17:58
Control Test	0.080	17:59
Air Blank	0.000	17:59
Subject Sample # 1	0.198	18:00
Breath Volume	2.234	18:00
Air Blank	0.000	18:00
Air Blank	0.000	18:02
Subject Sample # 2	0.174	18:03
Breath Volume	1.242	18:03
Air Blank	0.000	18:04
Air Blank	0.000	18:05
Subject Sample # 3	0.204	18:06
Breath Volume	3.000	18:06
Air Blank	0.000	18:07
Control Test	0.078	18:07
Air Blank	0.000	18:08
Diagnostics Check	OK	18:08

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: ARDAJI

DATE/TIME OF TEST: 09/12/2023 00:23:53

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.079	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.161	00:19
Breath Volume	3.621	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:21
Subject Sample # 2	0.162	00:22
Breath Volume	3.601	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD,EDWARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SEDLER

DATE/TIME OF TEST: 09/13/2023 00:41:04

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.136	00:36
Breath Volume	2.074	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:38
Subject Sample # 2	0.138	00:39
Breath Volume	2.921	00:39
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:41

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE, DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 09/13/2023 02:59:29

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.238	02:54
Breath Volume	1.277	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.258	02:57
Breath Volume	2.601	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO,ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: ORTA

DATE/TIME OF TEST: 09/13/2023 20:39:37

Diagnostics Check	OK	20:32
Air Blank	0.000	20:33
Control Test	0.079	20:33
Air Blank	0.000	20:34
Subject Sample # 1	0.174	20:34
Breath Volume	1.507	20:34
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.166	20:37
Breath Volume	1.937	20:37
Air Blank	0.000	20:38
Control Test	0.079	20:39
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: PEREZ-LOPEZ

DATE/TIME OF TEST: 09/14/2023 01:46:33

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.132	01:41
Breath Volume	1.664	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.123	01:44
Breath Volume	2.179	01:44
Air Blank	0.000	01:45
Control Test	0.078	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXANDER, MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 09/14/2023 05:56:09

Diagnostics Check	OK	05:48
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:50
Subject Sample # 1	0.218	05:51
Breath Volume	1.246	05:51
Air Blank	0.000	05:52
Air Blank	0.000	05:53
Subject Sample # 2	0.213	05:54
Breath Volume	1.347	05:54
Air Blank	0.000	05:55
Control Test	0.077	05:55
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: GALVEZ GONZALEZ

DATE/TIME OF TEST: 09/14/2023 22:42:13

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.285	22:36
Breath Volume	1.550	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.305	22:40
Breath Volume	2.437	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: OLIVERA

DATE/TIME OF TEST: 09/15/2023 03:56:09

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.052	03:51
Breath Volume	1.527	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.053	03:54
Breath Volume	1.691	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 09/15/2023 05:22:57

Diagnostics Check	OK	05:16
Air Blank	0.000	05:17
Control Test	0.080	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.180	05:18
Breath Volume	2.742	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:20
Subject Sample # 2	0.183	05:21
Breath Volume	2.550	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 09/15/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 09/15/2023 22:02:15

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.158	21:57
Breath Volume	1.726	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.153	22:00
Breath Volume	1.515	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BURKHARDT

DATE/TIME OF TEST: 09/16/2023 04:26:46

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.173	04:22
Breath Volume	2.027	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.176	04:25
Breath Volume	2.371	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 09/16/2023 23:48:37

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.062	23:44
Breath Volume	1.574	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.059	23:47
Breath Volume	1.789	23:47
Air Blank	0.000	23:47
Control Test	0.079	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ALBARRAN

DATE/TIME OF TEST: 09/17/2023 04:35:39

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.080	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.118	04:31
Breath Volume	1.828	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.121	04:34
Breath Volume	2.339	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ACOSTA,NERELYS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 09/17/2023 23:17:06

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.214	23:11
Breath Volume	2.343	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.209	23:15
Breath Volume	1.507	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: FIALLOS

DATE/TIME OF TEST: 09/18/2023 01:56:29

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.113	01:52
Breath Volume	2.273	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.118	01:54
Breath Volume	1.230	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: SANTANA DE LEON

DATE/TIME OF TEST: 09/18/2023 04:20:12

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.080	04:08	
Air Blank	0.000	04:08	
Subject Sample	# 1 0.143	04:12	Volume Not Met
Breath Volume	0.457	04:12	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 2 0.169	04:15	No .020 agreement
Breath Volume	1.285	04:15	
Air Blank	0.000	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 3 0.190	04:18	No .020 agreement
Breath Volume	2.867	04:18	
Air Blank	0.000	04:19	
Control Test	0.079	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: SANTANA DE LEON

DATE/TIME OF TEST: 09/18/2023 04:36:25

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.208	04:31
Breath Volume	4.730	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.199	04:34
Breath Volume	2.515	04:34
Air Blank	0.000	04:35
Control Test	0.077	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: FIALLOS

DATE/TIME OF TEST: 09/20/2023 02:03:14

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.156	01:58
Breath Volume	3.351	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.153	02:01
Breath Volume	3.136	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HAYLES,R
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: VERENICH

DATE/TIME OF TEST: 09/20/2023 21:59:23

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.000	21:55
Breath Volume	2.222	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.000	21:58
Breath Volume	2.277	21:58
Air Blank	0.000	21:58
Control Test	0.079	21:58
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/21/2023 02:32:42

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.169	02:28
Breath Volume	3.066	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.153	02:31
Breath Volume	1.585	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCRECIU, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: ROSENBLUM

DATE/TIME OF TEST: 09/21/2023 19:19:56

Diagnostics Check	OK	19:12
Air Blank	0.000	19:12
Control Test	0.080	19:13
Air Blank	0.000	19:13
Subject Sample # 1	0.000	19:14
Breath Volume	1.535	19:14
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.000	19:18
Breath Volume	1.210	19:18
Air Blank	0.000	19:19
Control Test	0.080	19:19
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 09/22/2023 01:57:55

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.146	01:52
Breath Volume	1.171	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.149	01:56
Breath Volume	2.429	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,JAIME
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PARISE

DATE/TIME OF TEST: 09/22/2023 03:22:01

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.145	03:17
Breath Volume	1.519	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.139	03:20
Breath Volume	1.304	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KOZLOWSKI,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: ALONZO ALVAREZ

DATE/TIME OF TEST: 09/23/2023 03:24:03

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.079	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.143	03:19
Breath Volume	1.921	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:21
Subject Sample # 2	0.149	03:22
Breath Volume	1.917	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:24

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ORDONEZ

DATE/TIME OF TEST: 09/23/2023 22:12:26

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.196	22:07
Breath Volume	2.273	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.182	22:10
Breath Volume	1.789	22:10
Air Blank	0.000	22:11
Control Test	0.077	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRANT,JEROME
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: VELASQUEZ LOPEZ

DATE/TIME OF TEST: 09/24/2023 00:06:15

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.220	00:01
Breath Volume	2.207	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.215	00:04
Breath Volume	2.492	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: VELASQUEZ SALES

DATE/TIME OF TEST: 09/24/2023 01:20:04

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.078	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.222	01:18	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/24/2023 04:39:23

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.134	04:34
Breath Volume	3.195	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.145	04:37
Breath Volume	2.843	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ABBOTT,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GARG

DATE/TIME OF TEST: 09/25/2023 02:41:15

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.080	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1 0.239	02:29	Volume Not Met
Breath Volume	0.117	02:29	
Air Blank	0.000	02:30	
Air Blank	0.000	02:32	
Subject Sample	# 2 0.251	02:34	
Breath Volume	1.843	02:34	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 3 0.245	02:39	
Breath Volume	2.261	02:39	
Air Blank	0.000	02:40	
Control Test	0.077	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MORENO NAVARRO

DATE/TIME OF TEST: 09/28/2023 04:06:42

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.215	04:01
Breath Volume	1.507	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.204	04:05
Breath Volume	2.347	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAPA,DAIANI
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 09/28/2023 08:12:18

Diagnostics Check	OK	08:05
Air Blank	0.000	08:06
Control Test	0.081	08:06
Air Blank	0.000	08:07
Subject Sample # 1	0.138	08:07
Breath Volume	1.285	08:07
Air Blank	0.000	08:08
Air Blank	0.000	08:10
Subject Sample # 2	0.135	08:10
Breath Volume	2.925	08:10
Air Blank	0.000	08:11
Control Test	0.081	08:11
Air Blank	0.000	08:12
Diagnostics Check	OK	08:12

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: SALAND

DATE/TIME OF TEST: 09/28/2023 23:32:24

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.186	23:24
Breath Volume	1.347	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.210	23:27
Breath Volume	2.046	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 3	0.213	23:30
Breath Volume	3.039	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 09/30/2023 03:24:16

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.192	03:18
Breath Volume	1.792	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.199	03:22
Breath Volume	1.585	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 10/20/2023 17:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTERO,NICHOLAS,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: MONTERO,NICHOLAS
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: BUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/03/2023 21:40:23

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:37	
Control Test	0.078	21:37	
Air Blank	0.000	21:38	
Subject Sample	# 1	21:38	Subject Test Refused
Breath Volume	0.000	21:38	
Air Blank	0.000	21:39	
Control Test	0.079	21:39	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

Data Download Date/Time: 09/21/2023 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAR,ANG,A
BREATH TEST OPERATOR AGENCY: NAVAL LAB PANA CITY
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: BAY COUNTY SO
LAST AGENCY INSPECTION: 08/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: CAR

DATE/TIME OF TEST: 09/07/2023 09:33:13

Diagnostics Check	OK	09:26
Air Blank	0.000	09:27
Control Test	0.079	09:27
Air Blank	0.000	09:28
Subject Sample # 1	0.000	09:28
Breath Volume	1.464	09:28
Air Blank	0.000	09:29
Air Blank	0.000	09:31
Subject Sample # 2	0.000	09:31
Breath Volume	2.464	09:31
Air Blank	0.000	09:32
Control Test	0.078	09:32
Air Blank	0.000	09:33
Diagnostics Check	OK	09:33

Data Download Date/Time: 09/21/2023 11:45