

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: PORTILLO CARDONA

DATE/TIME OF TEST: 10/01/2023 04:31:08

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.179	04:26
Breath Volume	2.410	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.183	04:29
Breath Volume	2.515	04:29
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 10/09/2023 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HARBIN

DATE/TIME OF TEST: 10/01/2023 21:55:20

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.107	21:50
Breath Volume	1.542	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.109	21:53
Breath Volume	1.292	21:53
Air Blank	0.000	21:54
Control Test	0.078	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 10/09/2023 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,SHELBY,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: OLIVAS,PHILLIPA
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 10/02/2023 01:27:30

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.124	01:23
Breath Volume	2.171	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.123	01:25
Breath Volume	1.972	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 10/09/2023 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/09/2023 01:23:58

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:09	
Control Test	0.081	01:10	
Air Blank	0.000	01:10	
Subject Sample	# 1 0.119	01:13	Volume Not Met
Breath Volume	0.050	01:13	
Air Blank	0.000	01:14	
Air Blank	0.000	01:16	
Subject Sample	# 2 0.150	01:17	
Breath Volume	1.355	01:17	
Air Blank	0.000	01:18	
Air Blank	0.000	01:20	
Subject Sample	# 3 0.144	01:22	
Breath Volume	1.960	01:22	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 10/09/2023 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: NOVO MESA

DATE/TIME OF TEST: 10/09/2023 03:53:51

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.103	03:48
Breath Volume	2.132	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.109	03:52
Breath Volume	3.390	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 10/09/2023 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SUBOYU,DOMINIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: ITLER

DATE/TIME OF TEST: 10/01/2023 04:59:10

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:56	
Control Test	0.080	04:57	
Air Blank	0.000	04:57	
Subject Sample	# 1	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:58	
Control Test	0.081	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 10/09/2023 16:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIMONEY,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 10/08/2023 03:13:08

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.188	03:08
Breath Volume	2.308	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.170	03:11
Breath Volume	2.167	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 10/09/2023 16:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: HOLE

DATE/TIME OF TEST: 10/11/2023 00:43:42

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.080	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Control Test	0.080	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 11/14/2023 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNE,ASHTON,D
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: MADOTTA,LIONEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BREUER

DATE/TIME OF TEST: 10/14/2023 23:02:33

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.000	22:58
Breath Volume	2.273	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.000	23:01
Breath Volume	1.917	23:01
Air Blank	0.000	23:01
Control Test	0.076	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 11/14/2023 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/16/2023 11:10:26

Diagnostics Check	OK	11:07	
Air Blank	0.000	11:08	
Control Test	0.080	11:08	
Air Blank	0.000	11:08	
Subject Sample	# 1	11:09	Subject Test Refused
Breath Volume	0.000	11:09	
Air Blank	0.000	11:09	
Control Test	0.080	11:09	
Air Blank	0.000	11:10	
Diagnostics Check	OK	11:10	

Data Download Date/Time: 11/14/2023 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: COFER

DATE/TIME OF TEST: 10/16/2023 19:38:40

Diagnostics Check	OK	19:27
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.273	19:30
Breath Volume	1.757	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.250	19:33
Breath Volume	1.867	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 3	0.252	19:36
Breath Volume	2.003	19:36
Air Blank	0.000	19:37
Control Test	0.077	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 11/14/2023 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/17/2023 02:54:09

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.079	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 11/14/2023 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ROSARIO,JONATHAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 10/22/2023 23:54:07

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:49	
Control Test	0.079	23:49	
Air Blank	0.000	23:50	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 11/14/2023 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: FELIU

DATE/TIME OF TEST: 10/23/2023 03:44:59

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.080	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.080	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 11/14/2023 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PIEDRA

DATE/TIME OF TEST: 10/25/2023 02:29:28

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.079	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.081	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 11/14/2023 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: EISENBERG

DATE/TIME OF TEST: 10/28/2023 23:12:21

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.116	23:07
Breath Volume	4.066	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.107	23:10
Breath Volume	3.496	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 11/14/2023 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: ALVAREZ SILVA

DATE/TIME OF TEST: 10/09/2023 00:10:19

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.082	00:02	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:05	No Sample Provided
Breath Volume	0.000	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:08	
Subject Sample	# 2	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 10/31/2023 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,DAVID,B
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JUDAH,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: MANTARAS PARADA

DATE/TIME OF TEST: 10/13/2023 01:08:00

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.138	01:03
Breath Volume	1.847	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.133	01:06
Breath Volume	2.023	01:06
Air Blank	0.000	01:07
Control Test	0.081	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:08

Data Download Date/Time: 10/31/2023 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,VIJAY,N
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GARCIA,HENRY
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: BRITO LOPEZ

DATE/TIME OF TEST: 10/14/2023 20:21:55

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.156	20:17
Breath Volume	4.019	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:19
Subject Sample # 2	0.158	20:20
Breath Volume	3.859	20:20
Air Blank	0.000	20:20
Control Test	0.080	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 10/31/2023 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: TANQUERO-BERNAL

DATE/TIME OF TEST: 10/15/2023 00:51:02

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.000	00:46
Breath Volume	1.476	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.000	00:49
Breath Volume	1.125	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:51

Data Download Date/Time: 10/31/2023 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: PEREZ DIAZ

DATE/TIME OF TEST: 10/15/2023 05:25:56

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:15	
Control Test	0.082	05:15	
Air Blank	0.000	05:16	
Subject Sample	# 1 0.165	05:19	Volume Not Met
Breath Volume	0.101	05:19	
Air Blank	0.000	05:20	
Air Blank	0.000	05:21	
Subject Sample	# 2	05:24	Subject Test Refused
Breath Volume	0.066	05:24	
Air Blank	0.000	05:25	
Control Test	0.079	05:25	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 10/31/2023 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: PEREZ DIAZ

DATE/TIME OF TEST: 10/15/2023 05:42:40

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:29	
Control Test	0.081	05:29	
Air Blank	0.000	05:30	
Subject Sample	# 1 0.160	05:31	
Breath Volume	1.175	05:31	
Air Blank	0.000	05:32	
Air Blank	0.000	05:34	
Subject Sample	# 2 0.162	05:37	Volume Not Met
Breath Volume	0.000	05:37	
Air Blank	0.000	05:38	
Air Blank	0.000	05:39	
Subject Sample	# 3 0.171	05:41	
Breath Volume	1.394	05:41	
Air Blank	0.000	05:41	
Control Test	0.079	05:42	
Air Blank	0.000	05:42	
Diagnostics Check	OK	05:42	

Data Download Date/Time: 10/31/2023 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: IGLESIAS,JORGE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ARAUJOBARRETO

DATE/TIME OF TEST: 10/15/2023 20:19:46

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.081	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.265	20:15
Breath Volume	1.890	20:15
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.268	20:18
Breath Volume	2.195	20:18
Air Blank	0.000	20:18
Control Test	0.081	20:19
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 10/31/2023 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 10/16/2023 03:09:07

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.082	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.000	03:04
Breath Volume	3.519	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.000	03:07
Breath Volume	3.722	03:07
Air Blank	0.000	03:08
Control Test	0.082	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 10/31/2023 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: PEREZ,A
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 10/20/2023 00:12:43

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.159	00:08
Breath Volume	2.796	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.163	00:11
Breath Volume	3.000	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 10/31/2023 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: OLIVA DURAN

DATE/TIME OF TEST: 10/22/2023 03:50:12

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.082	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.063	03:44
Breath Volume	1.347	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.059	03:48
Breath Volume	1.585	03:48
Air Blank	0.000	03:49
Control Test	0.082	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 10/31/2023 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 10/23/2023 00:18:07

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.080	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.226	00:10
Breath Volume	1.167	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:12
Subject Sample # 2	0.252	00:13
Breath Volume	2.253	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 3	0.244	00:16
Breath Volume	1.808	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 10/31/2023 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: MASLOV

DATE/TIME OF TEST: 10/23/2023 03:55:47

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.082	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.148	03:50
Breath Volume	1.812	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.149	03:54
Breath Volume	1.929	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 10/31/2023 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,JEFFREY,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MONTE DE OCA
VIERA,ALEXANDER
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male

OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: SARMIENTO

DATE/TIME OF TEST: 10/26/2023 18:21:41

Diagnostics Check	OK	18:13	
Air Blank	0.000	18:13	
Control Test	0.082	18:14	
Air Blank	0.000	18:14	
Subject Sample	# 1 0.000	18:17	Volume Not Met
Breath Volume	0.257	18:17	
Air Blank	0.000	18:18	
Air Blank	0.000	18:20	
Subject Sample	# 2	18:20	Subject Test Refused
Breath Volume	0.000	18:20	
Air Blank	0.000	18:20	
Control Test	0.081	18:21	
Air Blank	0.000	18:21	
Diagnostics Check	OK	18:21	

Data Download Date/Time: 10/31/2023 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,JEFFREY,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MONTE DE OCA
VIERA,ALEXANDER
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male

OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: VALDES ACOSTA

DATE/TIME OF TEST: 10/29/2023 17:03:14

Diagnostics Check	OK	16:57
Air Blank	0.000	16:57
Control Test	0.081	16:57
Air Blank	0.000	16:58
Subject Sample # 1	0.000	16:59
Breath Volume	1.644	16:59
Air Blank	0.000	16:59
Air Blank	0.000	17:01
Subject Sample # 2	0.000	17:01
Breath Volume	1.636	17:01
Air Blank	0.000	17:02
Control Test	0.082	17:02
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

Data Download Date/Time: 10/31/2023 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARARD,WILLIAM,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 10/20/2023 02:54:52

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.208	02:49
Breath Volume	1.675	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.209	02:53
Breath Volume	1.886	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 12/29/2023 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: MELI

DATE/TIME OF TEST: 10/09/2023 22:41:48

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.077	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.319	22:37
Breath Volume	4.472	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.311	22:40
Breath Volume	2.410	22:40
Air Blank	0.000	22:40
Control Test	0.077	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: WORY

DATE/TIME OF TEST: 10/14/2023 00:14:26

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1 0.103	00:07	Volume Not Met
Breath Volume	0.113	00:07	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 2 0.097	00:12	Volume Not Met
Breath Volume	0.050	00:12	
Air Blank	0.000	00:13	
Control Test	0.079	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARI,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GARI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: WORY

DATE/TIME OF TEST: 10/14/2023 00:27:47

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1 0.109	00:20	Volume Not Met
Breath Volume	0.132	00:20	
Air Blank	0.000	00:21	
Air Blank	0.000	00:23	
Subject Sample	# 2 0.096	00:26	Volume Not Met
Breath Volume	0.222	00:26	
Air Blank	0.000	00:26	
Control Test	0.078	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 10/14/2023 00:42:16

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.078	00:39	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:40	RFI Detect
Breath Volume	0.000	00:40	
Air Blank	0.000	00:41	
Control Test	0.078	00:41	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 10/14/2023 00:55:46

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.079	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.204	00:50
Breath Volume	1.253	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.203	00:54
Breath Volume	1.140	00:54
Air Blank	0.000	00:54
Control Test	0.078	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: EASON

DATE/TIME OF TEST: 10/14/2023 01:19:14

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.079	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.189	01:12
Breath Volume	2.527	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.185	01:17
Breath Volume	2.828	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 10/15/2023 21:06:59

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.238	21:02
Breath Volume	1.285	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:04
Subject Sample # 2	0.236	21:05
Breath Volume	1.437	21:05
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSAKOWSKI,PETER,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KOSAKOWSKI,PETER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: MATEO PEDRO

DATE/TIME OF TEST: 10/18/2023 04:48:07

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.191	04:41
Breath Volume	1.152	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.181	04:46
Breath Volume	1.906	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/19/2023 00:52:25

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.079	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.258	00:47
Breath Volume	1.800	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.260	00:50
Breath Volume	1.738	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,SHELBY,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: REPPL,JADEN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: RODRIGUEZ PEREZ

DATE/TIME OF TEST: 10/21/2023 23:36:04

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.128	23:31
Breath Volume	1.996	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.143	23:34
Breath Volume	1.875	23:34
Air Blank	0.000	23:35
Control Test	0.077	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:36

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: HOLEWA

DATE/TIME OF TEST: 10/21/2023 23:54:32

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.078	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.126	23:49
Breath Volume	1.566	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.119	23:52
Breath Volume	1.289	23:52
Air Blank	0.000	23:53
Control Test	0.078	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 10/23/2023 00:29:33

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.177	00:24
Breath Volume	1.851	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.175	00:27
Breath Volume	3.078	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: ORTIZ CARRASQUILLO

DATE/TIME OF TEST: 10/24/2023 22:30:33

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.081	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.132	22:25
Breath Volume	1.101	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.131	22:28
Breath Volume	1.468	22:28
Air Blank	0.000	22:29
Control Test	0.078	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI, MARK, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JABLONSKI, MARK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: WESTBERRY BARKER

DATE/TIME OF TEST: 10/28/2023 04:51:31

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.147	04:46
Breath Volume	3.183	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.146	04:49
Breath Volume	3.113	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI, MARK, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JABLONSKI, MARK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: CAMUS

DATE/TIME OF TEST: 10/29/2023 04:15:19

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:04	
Control Test	0.079	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1 0.186	04:07	Volume Not Met
Breath Volume	0.277	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2 0.186	04:13	Volume Not Met
Breath Volume	0.000	04:13	
Air Blank	0.000	04:14	
Control Test	0.076	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI, MARK, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JABLONSKI, MARK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: CAMUS

DATE/TIME OF TEST: 10/29/2023 04:32:54

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.076	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 1 0.179	04:22	Volume Not Met
Breath Volume	0.589	04:22	
Air Blank	0.000	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 2 0.165	04:25	
Breath Volume	2.195	04:25	
Air Blank	0.000	04:26	
Air Blank	0.000	04:28	
Subject Sample	# 3 0.170	04:31	Volume Not Met
Breath Volume	0.195	04:31	
Air Blank	0.000	04:31	
Control Test	0.078	04:32	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: MORIARTY,QUINN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: ALMOA

DATE/TIME OF TEST: 10/30/2023 04:33:49

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.079	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.059	04:29
Breath Volume	1.968	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.056	04:32
Breath Volume	2.832	04:32
Air Blank	0.000	04:32
Control Test	0.078	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUSSEAU,DMITRY,A
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: ROUSSEAU,DMITRY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: STULTS

DATE/TIME OF TEST: 10/30/2023 04:41:56

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:39	
Control Test	0.079	04:39	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:41	
Control Test	0.078	04:41	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 11/06/2023 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: HUSBANDS, TYLER, J	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: HUSBANDS, TYLER	ARREST TIME: 23:50
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Female
LAST AGENCY INSPECTION: 09/05/2023	OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: GREENE	DATE/TIME OF TEST: 10/18/2023 00:32:18
-----------------------------	--

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.249	00:27
Breath Volume	3.062	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.233	00:30
Breath Volume	3.023	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 10/30/2023 17:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: HUSBANDS, TYLER, J	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: GONZALEZ, CHRISTOPHER	ARREST TIME: 04:01
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Female
LAST AGENCY INSPECTION: 09/05/2023	OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SCIAMMETTA	DATE/TIME OF TEST: 10/21/2023 04:29:53
---------------------------------	--

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.155	04:25
Breath Volume	1.601	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:27
Subject Sample # 2	0.148	04:28
Breath Volume	2.535	04:28
Air Blank	0.000	04:28
Control Test	0.078	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 10/30/2023 17:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SANABRIA

DATE/TIME OF TEST: 10/19/2023 22:37:30

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.263	22:32
Breath Volume	2.492	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.274	22:35
Breath Volume	2.703	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 11/06/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: HOLEWA

DATE/TIME OF TEST: 10/21/2023 22:47:17

Diagnostics Check	OK	22:46	
Air Blank		22:46	RFI Detect
Air Blank	0.000	22:47	

Data Download Date/Time: 11/06/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: HOLEWA

DATE/TIME OF TEST: 10/21/2023 22:54:38

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:49	
Control Test	0.080	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1 0.140	22:50	
Breath Volume	1.519	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 2	22:53	RFI Detect
Breath Volume	0.000	22:53	
Air Blank	0.000	22:53	
Control Test	0.078	22:54	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 11/06/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: FLAGLER

DATE/TIME OF TEST: 10/01/2023 03:22:22

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.053	03:18
Breath Volume	1.476	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.050	03:20
Breath Volume	1.796	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 10/19/2023 19:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: FRIDY

DATE/TIME OF TEST: 10/01/2023 20:41:03

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.081	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.191	20:36
Breath Volume	3.328	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:38
Subject Sample # 2	0.200	20:39
Breath Volume	3.273	20:39
Air Blank	0.000	20:40
Control Test	0.081	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:41

Data Download Date/Time: 10/19/2023 19:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: DANNER

DATE/TIME OF TEST: 10/02/2023 00:26:00

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.000	00:21
Breath Volume	2.316	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.000	00:24
Breath Volume	2.945	00:24
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:26

Data Download Date/Time: 10/19/2023 19:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: MONTGOMERY,KENNETH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24229080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 18:54

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 10/09/2023 19:16:48

Diagnostics Check	OK	19:05
Air Blank	0.000	19:05
Control Test	0.082	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.228	19:08
Breath Volume	2.746	19:08
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.206	19:11
Breath Volume	1.113	19:11
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 3	0.191	19:15
Breath Volume	1.500	19:15
Air Blank	0.000	19:15
Control Test	0.080	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 10/19/2023 19:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 10/15/2023 02:43:02

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.253	02:38
Breath Volume	2.066	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.249	02:41
Breath Volume	2.574	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:43

Data Download Date/Time: 10/19/2023 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY, KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: GUFFEY

DATE/TIME OF TEST: 10/19/2023 23:28:28

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.082	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.182	23:23
Breath Volume	2.066	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.170	23:26
Breath Volume	1.984	23:26
Air Blank	0.000	23:27
Control Test	0.081	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 11/12/2023 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: LOONEY

DATE/TIME OF TEST: 10/27/2023 23:21:34

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.082	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.124	23:17
Breath Volume	1.281	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.125	23:20
Breath Volume	1.339	23:20
Air Blank	0.000	23:20
Control Test	0.081	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 11/12/2023 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WOFFORD,JOSEPH,T

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: ALIJAGIC,SAMIR

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24229080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: DUNAWAY

DATE/TIME OF TEST: 10/29/2023 04:14:42

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.237	04:09
Breath Volume	3.332	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.251	04:13
Breath Volume	3.222	04:13
Air Blank	0.000	04:13
Control Test	0.081	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 11/12/2023 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 09/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MOONE

DATE/TIME OF TEST: 10/01/2023 01:31:56

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.078	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.176	01:27
Breath Volume	2.492	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.166	01:30
Breath Volume	1.558	01:30
Air Blank	0.000	01:31
Control Test	0.076	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 10/18/2023 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 09/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: PADILLA NAJERA

DATE/TIME OF TEST: 10/01/2023 04:36:50

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.145	04:30
Breath Volume	1.242	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.142	04:35
Breath Volume	1.488	04:35
Air Blank	0.000	04:35
Control Test	0.078	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 10/18/2023 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 10/07/2023 02:50:26

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.078	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.199	02:45
Breath Volume	1.894	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.185	02:48
Breath Volume	1.269	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 10/09/2023 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,DARIN,L
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ABREU CABRERA

DATE/TIME OF TEST: 10/07/2023 23:54:45

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.079	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.116	23:50
Breath Volume	3.292	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.106	23:53
Breath Volume	3.304	23:53
Air Blank	0.000	23:53
Control Test	0.077	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 10/09/2023 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/07/2023 04:55:32

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.078	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.114	04:51
Breath Volume	2.187	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.116	04:53
Breath Volume	2.410	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 10/09/2023 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ALBIN

DATE/TIME OF TEST: 10/08/2023 03:46:03

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.130	03:41
Breath Volume	2.351	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.131	03:44
Breath Volume	2.394	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:46

Data Download Date/Time: 10/09/2023 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HANEWINCKEL

DATE/TIME OF TEST: 10/26/2023 04:07:35

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.164	04:03
Breath Volume	3.914	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.159	04:06
Breath Volume	4.000	04:06
Air Blank	0.000	04:06
Control Test	0.077	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 11/06/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/09/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: MULLINEX

DATE/TIME OF TEST: 10/31/2023 21:32:37

Diagnostics Check	OK	21:28	
Air Blank	0.000	21:28	
Control Test	0.079	21:28	
Air Blank	0.000	21:29	
Subject Sample	# 1	21:31	Subject Test Refused
Breath Volume	0.000	21:31	
Air Blank	0.000	21:31	
Control Test	0.079	21:31	
Air Blank		21:32	RFI Detect
Air Blank	0.000	21:32	

Data Download Date/Time: 11/06/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,MATTHEW,M

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: SMITH,MATTHEW

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 23:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: DESAI

DATE/TIME OF TEST: 10/20/2023 23:49:07

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.141	23:44
Breath Volume	1.710	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:46
Subject Sample # 2	0.126	23:47
Breath Volume	1.273	23:47
Air Blank	0.000	23:48
Control Test	0.078	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 12/04/2023 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 03:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 10/05/2023 03:47:01

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.158	03:39
Breath Volume	2.242	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.137	03:42
Breath Volume	1.179	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 3	0.149	03:45
Breath Volume	1.535	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:47

Data Download Date/Time: 10/19/2023 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: KEEN,CHARLES
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: GRAY III

DATE/TIME OF TEST: 10/05/2023 23:51:51

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.194	23:46
Breath Volume	1.742	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.198	23:50
Breath Volume	2.074	23:50
Air Blank	0.000	23:50
Control Test	0.078	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 10/19/2023 19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 03:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CRUZ PEREZ

DATE/TIME OF TEST: 10/07/2023 03:49:04

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.081	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	RFI Detect
Breath Volume	0.359	03:47	
Air Blank	0.000	03:48	
Control Test	0.080	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 10/19/2023 19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 03:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PEREZ CRUZ

DATE/TIME OF TEST: 10/07/2023 03:57:47

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.181	03:53
Breath Volume	1.390	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.182	03:56
Breath Volume	1.324	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 10/19/2023 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,MATTHEW,M

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: SMITH,MATTHEW

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 00:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: PAWAR

DATE/TIME OF TEST: 10/15/2023 01:00:26

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.081	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.141	00:55
Breath Volume	1.472	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.138	00:58
Breath Volume	1.414	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 10/19/2023 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 02:18
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 09/07/2023	OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: OCTEUS	DATE/TIME OF TEST: 10/15/2023 02:34:03
-----------------------------	--

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.138	02:29
Breath Volume	2.652	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.143	02:32
Breath Volume	1.566	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:34

Data Download Date/Time: 10/19/2023 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 01:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MCFARLAND

DATE/TIME OF TEST: 10/19/2023 01:28:56

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.081	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.207	01:23
Breath Volume	1.218	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.201	01:27
Breath Volume	2.925	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 10/19/2023 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: NORDEEN

DATE/TIME OF TEST: 10/20/2023 02:36:18

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 1 0.203	02:31	
Breath Volume	1.296	02:31	
Air Blank	0.000	02:32	
Air Blank	0.000	02:34	
Subject Sample	# 2 0.215	02:34	
Breath Volume	1.750	02:34	
Air Blank	0.000	02:35	
Control Test	0.081	02:35	
Air Blank		02:35	RFI Detect
Air Blank	0.000	02:36	

Data Download Date/Time: 11/13/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 02:21
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 10/19/2023	OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: NORDEEN	DATE/TIME OF TEST: 10/20/2023 02:45:14
------------------------------	--

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.213	02:40
Breath Volume	2.523	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.223	02:43
Breath Volume	2.281	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 11/13/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:08

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/28/2023 02:22:51

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.133	02:18
Breath Volume	1.812	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.127	02:21
Breath Volume	1.269	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 11/13/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 18:55
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 10/19/2023	OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: LINARES FLORES	DATE/TIME OF TEST: 10/28/2023 19:08:00
-------------------------------------	--

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.081	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.080	19:03
Breath Volume	1.585	19:03
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.075	19:06
Breath Volume	2.769	19:06
Air Blank	0.000	19:07
Control Test	0.080	19:07
Air Blank	0.000	19:07
Diagnostics Check	OK	19:08

Data Download Date/Time: 11/13/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MORAN,MAUREEN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: BUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/28/2023 21:29:37

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.080	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.118	21:25
Breath Volume	1.753	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.118	21:28
Breath Volume	1.378	21:28
Air Blank	0.000	21:28
Control Test	0.081	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 11/13/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,MATTHEW,M

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: LAQUIDARA,ANTHONY

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 01:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: DOMINGUEZ ORTIZ

DATE/TIME OF TEST: 10/29/2023 03:58:43

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.139	03:54
Breath Volume	2.144	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.138	03:57
Breath Volume	2.312	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 11/13/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ECHEVERRIAVIDAL,XAVIER
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 10/29/2023 04:13:34

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.077	04:08
Breath Volume	1.589	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.079	04:11
Breath Volume	1.515	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 11/13/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: FARLEY,JOSEPH,E	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: FARLEY,JOSEPH	ARREST TIME: 05:39
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Female
LAST AGENCY INSPECTION: 10/19/2023	OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: MARQUEZ	DATE/TIME OF TEST: 10/29/2023 05:53:36
------------------------------	--

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.081	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.183	05:47
Breath Volume	1.722	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:49
Subject Sample # 2	0.170	05:51
Breath Volume	1.410	05:51
Air Blank	0.000	05:52
Control Test	0.078	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 11/13/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLIFFORD,C
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: ASHBY

DATE/TIME OF TEST: 10/01/2023 00:41:59

Diagnostics Check	OK	00:41	
Air Blank		00:41	RFI Detect
Air Blank	0.000	00:41	

Data Download Date/Time: 10/05/2023 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLIFFORD,C
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: ASHBY

DATE/TIME OF TEST: 10/01/2023 00:50:51

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.078	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.189	00:46
Breath Volume	2.070	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.190	00:49
Breath Volume	1.363	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 10/05/2023 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RODRIGUEZ,G
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: LOZADA HERNANDEZ

DATE/TIME OF TEST: 10/01/2023 02:29:51

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.079	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1 0.183	02:24	
Breath Volume	1.824	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 2 0.169	02:27	
Breath Volume	1.539	02:27	
Air Blank	0.000	02:28	
Control Test	0.077	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	Fail	02:29	Internal Printer Test
Air Blank	0.000	02:29	

Data Download Date/Time: 10/05/2023 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RODRIGUEZ,G
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: LOZADA HERNANDEZ

DATE/TIME OF TEST: 10/01/2023 02:39:04

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.172	02:34
Breath Volume	1.519	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.171	02:37
Breath Volume	1.585	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 10/05/2023 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ROMERO,D
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FLORIAN ESCAVEDUQUE

DATE/TIME OF TEST: 10/01/2023 03:02:21

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.079	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.078	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 10/05/2023 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: JARQUIN-HURTADO

DATE/TIME OF TEST: 10/20/2023 00:32:34

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.165	00:27
Breath Volume	3.109	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.164	00:30
Breath Volume	2.449	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 11/03/2023 05:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: NAVARRO,R
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: COLMAN

DATE/TIME OF TEST: 10/28/2023 06:43:49

Diagnostics Check	OK	06:40	
Air Blank	0.000	06:41	
Control Test	0.080	06:41	
Air Blank	0.000	06:42	
Subject Sample	# 1	06:42	Subject Test Refused
Breath Volume	0.000	06:42	
Air Blank	0.000	06:42	
Control Test	0.078	06:43	
Air Blank	0.000	06:43	
Diagnostics Check	OK	06:43	

Data Download Date/Time: 11/03/2023 05:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: COLMAN

DATE/TIME OF TEST: 10/28/2023 06:55:54

Diagnostics Check	OK	06:48
Air Blank	0.000	06:48
Control Test	0.079	06:49
Air Blank	0.000	06:49
Subject Sample # 1	0.198	06:50
Breath Volume	1.671	06:50
Air Blank	0.000	06:51
Air Blank	0.000	06:53
Subject Sample # 2	0.192	06:54
Breath Volume	1.722	06:54
Air Blank	0.000	06:54
Control Test	0.079	06:55
Air Blank	0.000	06:55
Diagnostics Check	OK	06:55

Data Download Date/Time: 11/03/2023 05:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AMPARO,IAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 10/06/2023 19:23:03

Diagnostics Check	OK	19:16
Air Blank	0.000	19:17
Control Test	0.080	19:17
Air Blank	0.000	19:18
Subject Sample # 1	0.249	19:18
Breath Volume	1.910	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:20
Subject Sample # 2	0.238	19:21
Breath Volume	1.687	19:21
Air Blank	0.000	19:22
Control Test	0.079	19:22
Air Blank	0.000	19:22
Diagnostics Check	OK	19:23

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALONSO,BERNIE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: KESHEN

DATE/TIME OF TEST: 10/06/2023 19:58:43

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.081	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.000	19:54
Breath Volume	1.636	19:54
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.000	19:57
Breath Volume	2.363	19:57
Air Blank	0.000	19:57
Control Test	0.081	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RYAN,JASON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 10/07/2023 03:25:18

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.082	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.171	03:20
Breath Volume	2.425	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.170	03:23
Breath Volume	1.492	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 10/09/2023 00:56:56

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.202	00:52
Breath Volume	1.468	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:54
Subject Sample # 2	0.198	00:55
Breath Volume	1.828	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 10/09/2023 05:22:28

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.189	05:18
Breath Volume	1.808	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.188	05:20
Breath Volume	1.562	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUERRA,ALVARO
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: SANCHEZ ZAMORANO

DATE/TIME OF TEST: 10/13/2023 20:29:12

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.081	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.280	20:23
Breath Volume	1.902	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:25
Subject Sample # 2	0.269	20:27
Breath Volume	1.878	20:27
Air Blank	0.000	20:28
Control Test	0.080	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/14/2023 01:46:58

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.208	01:42
Breath Volume	1.902	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.212	01:45
Breath Volume	2.082	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/14/2023 04:23:50

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.158	04:18
Breath Volume	1.957	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.164	04:22
Breath Volume	2.351	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHAUVIN,AUSTIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 09:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: KETOR

DATE/TIME OF TEST: 10/15/2023 11:18:18

Diagnostics Check	OK	11:11
Air Blank	0.000	11:12
Control Test	0.081	11:12
Air Blank	0.000	11:13
Subject Sample # 1	0.194	11:13
Breath Volume	1.667	11:13
Air Blank	0.000	11:14
Air Blank	0.000	11:16
Subject Sample # 2	0.201	11:16
Breath Volume	1.785	11:16
Air Blank	0.000	11:17
Control Test	0.080	11:17
Air Blank	0.000	11:18
Diagnostics Check	OK	11:18

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VEGA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: VICENTE

DATE/TIME OF TEST: 10/15/2023 22:53:31

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.188	22:48
Breath Volume	1.640	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:50
Subject Sample # 2	0.189	22:51
Breath Volume	1.625	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CESPEDES GARCIA

DATE/TIME OF TEST: 10/16/2023 00:55:34

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.202	00:50
Breath Volume	1.726	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.201	00:53
Breath Volume	1.679	00:53
Air Blank	0.000	00:54
Control Test	0.081	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: PALACIOS LOPEZ

DATE/TIME OF TEST: 10/16/2023 04:33:41

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.271	04:29
Breath Volume	3.164	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.272	04:32
Breath Volume	3.550	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ADAMSON,ANDREW
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/16/2023 14:49:47

Diagnostics Check	OK	14:43
Air Blank	0.000	14:43
Control Test	0.081	14:44
Air Blank	0.000	14:44
Subject Sample # 1	0.237	14:45
Breath Volume	1.476	14:45
Air Blank	0.000	14:45
Air Blank	0.000	14:47
Subject Sample # 2	0.238	14:48
Breath Volume	1.269	14:48
Air Blank	0.000	14:48
Control Test	0.080	14:49
Air Blank	0.000	14:49
Diagnostics Check	OK	14:49

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SIERRA,NICHOLAS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 13:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 10/16/2023 16:39:03

Diagnostics Check	OK	16:32
Air Blank	0.000	16:33
Control Test	0.080	16:33
Air Blank	0.000	16:34
Subject Sample # 1	0.000	16:34
Breath Volume	1.468	16:34
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.000	16:37
Breath Volume	1.554	16:37
Air Blank	0.000	16:38
Control Test	0.081	16:38
Air Blank	0.000	16:38
Diagnostics Check	OK	16:39

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/22/2023 02:41:32

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.081	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:33	No Sample Provided
Breath Volume	0.000	02:33	
Air Blank	0.000	02:33	
Air Blank	0.000	02:35	
Subject Sample	# 2	0.010	02:37
Breath Volume	1.445	02:37	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 3	0.008	02:40
Breath Volume	1.128	02:40	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUC,NORBENS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ST.GERARD

DATE/TIME OF TEST: 10/22/2023 04:09:44

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.083	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.240	04:04
Breath Volume	1.355	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.243	04:08
Breath Volume	1.558	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 10/23/2023 22:41:44

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.081	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.272	22:36
Breath Volume	2.089	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.269	22:40
Breath Volume	2.226	22:40
Air Blank	0.000	22:40
Control Test	0.080	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/28/2023 19:33:20

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.081	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.239	19:28
Breath Volume	1.441	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.247	19:31
Breath Volume	1.398	19:31
Air Blank	0.000	19:32
Control Test	0.077	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALTUNAY,MUHAMMED
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: MARTIN JR

DATE/TIME OF TEST: 10/29/2023 06:47:04

Diagnostics Check	OK	06:39
Air Blank	0.000	06:39
Control Test	0.080	06:40
Air Blank	0.000	06:40
Subject Sample # 1	0.000	06:42
Breath Volume	1.535	06:42
Air Blank	0.000	06:42
Air Blank	0.000	06:44
Subject Sample # 2	0.000	06:45
Breath Volume	2.558	06:45
Air Blank	0.000	06:46
Control Test	0.080	06:46
Air Blank	0.000	06:46
Diagnostics Check	OK	06:47

Data Download Date/Time: 11/08/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,GERALD,N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON,GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 10/08/2023 00:36:25

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.081	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1 0.202	00:30	
Breath Volume	1.527	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample	# 2 0.203	00:34	
Breath Volume	1.742	00:34	
Air Blank	0.000	00:34	
Control Test	0.085	00:35	
Air Blank	0.000	00:35	
Diagnostics Check	Fail	00:35	Analytical Stability Test
Air Blank	0.000	00:36	

Data Download Date/Time: 10/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,GERALD,N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON,GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 10/14/2023 18:18:23

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.082	18:11
Air Blank	0.000	18:12
Subject Sample # 1	0.240	18:12
Breath Volume	1.781	18:12
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.234	18:16
Breath Volume	1.675	18:16
Air Blank	0.000	18:17
Control Test	0.082	18:17
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

Data Download Date/Time: 10/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: VASILOPOULOS

DATE/TIME OF TEST: 10/28/2023 23:17:25

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.081	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.147	23:12
Breath Volume	3.707	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:14
Subject Sample # 2	0.148	23:15
Breath Volume	3.160	23:15
Air Blank	0.000	23:16
Control Test	0.082	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 10/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 10/23/2023 02:52:05

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.270	02:46
Breath Volume	1.566	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.271	02:50
Breath Volume	1.488	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 10/24/2023 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MOREJON,DANIEL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LAKE

DATE/TIME OF TEST: 10/08/2023 01:47:01

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.108	01:42
Breath Volume	1.878	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.110	01:45
Breath Volume	2.542	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:47

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: PACK

DATE/TIME OF TEST: 10/08/2023 05:10:53

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.207	05:03
Breath Volume	1.222	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.204	05:09
Breath Volume	2.007	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: PITTALUGA

DATE/TIME OF TEST: 10/15/2023 02:00:55

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.146	01:55
Breath Volume	2.718	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.148	01:59
Breath Volume	3.464	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RUSSELL,AMON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MBOUP

DATE/TIME OF TEST: 10/15/2023 02:31:18

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.252	02:26
Breath Volume	1.628	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.263	02:29
Breath Volume	2.183	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: WIMBERLY

DATE/TIME OF TEST: 10/15/2023 03:55:01

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.093	03:50
Breath Volume	1.359	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.096	03:53
Breath Volume	1.140	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:55

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ENRIQUE,ALEX
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 10/15/2023 04:10:45

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.202	04:06
Breath Volume	2.675	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.212	04:09
Breath Volume	2.429	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HALES,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 10/18/2023 19:31:42

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.081	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.294	19:27
Breath Volume	2.210	19:27
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.297	19:30
Breath Volume	2.320	19:30
Air Blank	0.000	19:30
Control Test	0.080	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,BRIAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GALLIGAN

DATE/TIME OF TEST: 10/19/2023 01:18:53

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1 0.173	01:11	Volume Not Met
Breath Volume	0.843	01:11	
Air Blank	0.000	01:12	
Air Blank	0.000	01:14	
Subject Sample	# 2 0.192	01:17	Volume Not Met
Breath Volume	0.269	01:17	
Air Blank	0.000	01:17	
Control Test	0.079	01:18	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,BRIAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GALLIGAN

DATE/TIME OF TEST: 10/19/2023 01:36:05

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.079	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1 0.157	01:25	Volume Not Met
Breath Volume	0.613	01:25	
Air Blank	0.000	01:26	
Air Blank	0.000	01:28	
Subject Sample	# 2 0.206	01:29	
Breath Volume	1.437	01:29	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 3 0.192	01:34	
Breath Volume	1.296	01:34	
Air Blank	0.000	01:35	
Control Test	0.078	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: LUSTER

DATE/TIME OF TEST: 10/20/2023 21:19:02

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.080	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.231	21:14
Breath Volume	3.351	21:14
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.226	21:17
Breath Volume	2.894	21:17
Air Blank	0.000	21:18
Control Test	0.079	21:18
Air Blank	0.000	21:18
Diagnostics Check	OK	21:19

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: MCCAFFREY

DATE/TIME OF TEST: 10/21/2023 07:11:07

Diagnostics Check	OK	07:02
Air Blank	0.000	07:03
Control Test	0.081	07:03
Air Blank	0.000	07:03
Subject Sample # 1	0.225	07:05
Breath Volume	1.988	07:05
Air Blank	0.000	07:05
Air Blank	0.000	07:07
Subject Sample # 2	0.220	07:09
Breath Volume	1.527	07:09
Air Blank	0.000	07:10
Control Test	0.079	07:10
Air Blank	0.000	07:11
Diagnostics Check	OK	07:11

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,FRANK
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:21

SUBJECT'S LAST NAME: CLERGER

DATE/TIME OF TEST: 10/21/2023 08:51:18

Diagnostics Check	OK	08:44
Air Blank	0.000	08:45
Control Test	0.080	08:45
Air Blank	0.000	08:46
Subject Sample # 1	0.129	08:46
Breath Volume	2.734	08:46
Air Blank	0.000	08:47
Air Blank	0.000	08:49
Subject Sample # 2	0.134	08:49
Breath Volume	3.199	08:49
Air Blank	0.000	08:50
Control Test	0.079	08:50
Air Blank	0.000	08:51
Diagnostics Check	OK	08:51

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCKIN,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: ZUBAK

DATE/TIME OF TEST: 10/22/2023 20:06:44

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.080	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.131	20:00
Breath Volume	1.437	20:00
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.117	20:05
Breath Volume	1.191	20:05
Air Blank	0.000	20:05
Control Test	0.081	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: DIONNE

DATE/TIME OF TEST: 10/23/2023 04:06:27

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.224	04:01
Breath Volume	1.785	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.210	04:04
Breath Volume	1.160	04:04
Air Blank	0.000	04:05
Control Test	0.078	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 11/08/2023 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 10/25/2023 02:15:59

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.081	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.177	02:11
Breath Volume	2.449	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.175	02:14
Breath Volume	1.378	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 11/08/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DOMINGUEZ PEREZ,ALAIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 10/25/2023 21:53:14

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.238	21:48
Breath Volume	2.421	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.224	21:51
Breath Volume	2.250	21:51
Air Blank	0.000	21:52
Control Test	0.079	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 11/08/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALTUNES,MUHAMMED
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: NHOMME

DATE/TIME OF TEST: 10/26/2023 00:39:00

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.000	00:34
Breath Volume	3.417	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.000	00:37
Breath Volume	3.253	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:39

Data Download Date/Time: 11/08/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PETERKIN,MARSHANE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 10/26/2023 06:09:25

Diagnostics Check	OK	06:02
Air Blank	0.000	06:02
Control Test	0.080	06:03
Air Blank	0.000	06:03
Subject Sample # 1	0.117	06:04
Breath Volume	1.558	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample # 2	0.109	06:07
Breath Volume	1.847	06:07
Air Blank	0.000	06:08
Control Test	0.080	06:08
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 11/08/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REDFEARN,JAMES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: RUIZ RIVERA

DATE/TIME OF TEST: 10/29/2023 09:28:20

Diagnostics Check	OK	09:20
Air Blank	0.000	09:21
Control Test	0.080	09:21
Air Blank	0.000	09:22
Subject Sample # 1	0.114	09:23
Breath Volume	1.742	09:23
Air Blank	0.000	09:24
Air Blank	0.000	09:25
Subject Sample # 2	0.105	09:26
Breath Volume	1.835	09:26
Air Blank	0.000	09:27
Control Test	0.081	09:27
Air Blank	0.000	09:28
Diagnostics Check	OK	09:28

Data Download Date/Time: 11/08/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: CASTILLO AVENTIN

DATE/TIME OF TEST: 10/29/2023 21:39:45

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.182	21:35
Breath Volume	2.035	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.164	21:38
Breath Volume	2.378	21:38
Air Blank	0.000	21:38
Control Test	0.080	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 11/08/2023 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/30/2023 00:33:07

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.124	00:27
Breath Volume	1.410	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.124	00:31
Breath Volume	2.074	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 11/08/2023 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO',DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ELLIS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: GODREAU

DATE/TIME OF TEST: 10/06/2023 09:17:03

Diagnostics Check	OK	09:10
Air Blank	0.000	09:10
Control Test	0.080	09:10
Air Blank	0.000	09:11
Subject Sample # 1	0.310	09:12
Breath Volume	2.109	09:12
Air Blank	0.000	09:12
Air Blank	0.000	09:14
Subject Sample # 2	0.302	09:15
Breath Volume	1.265	09:15
Air Blank	0.000	09:16
Control Test	0.079	09:16
Air Blank	0.000	09:16
Diagnostics Check	OK	09:17

Data Download Date/Time: 10/25/2023 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT,ROBERT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: KUCHE

DATE/TIME OF TEST: 10/06/2023 18:41:38

Diagnostics Check	OK	18:33
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:35
Subject Sample # 1	0.113	18:37
Breath Volume	1.445	18:37
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 2	0.124	18:40
Breath Volume	1.433	18:40
Air Blank	0.000	18:40
Control Test	0.079	18:41
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

Data Download Date/Time: 10/25/2023 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DZUGAN,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DZUGAN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GARCIA MARTE

DATE/TIME OF TEST: 10/07/2023 02:57:18

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.028	02:52
Breath Volume	1.500	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.026	02:55
Breath Volume	2.339	02:55
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 10/25/2023 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAZZARESE,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GIARDINA,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MCCURRY

DATE/TIME OF TEST: 10/07/2023 03:11:50

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.097	03:06
Breath Volume	1.582	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.102	03:10
Breath Volume	1.785	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 10/25/2023 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARCHMAN,MARCHMAN,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: E,E
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/07/2023 23:35:20

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.080	23:30	
Air Blank	0.000	23:30	
Subject Sample # 1	0.147	23:31	
Breath Volume	2.214	23:31	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample # 2		23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:34	
Control Test	0.079	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 10/25/2023 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SLIVKA

DATE/TIME OF TEST: 10/08/2023 00:23:29

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.311	00:19
Breath Volume	2.679	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:21
Subject Sample # 2	0.314	00:21
Breath Volume	3.515	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 10/25/2023 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANO,MARDEE,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MANO,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: TIERNEY

DATE/TIME OF TEST: 10/08/2023 10:21:22

Diagnostics Check	OK	10:15
Air Blank	0.000	10:15
Control Test	0.080	10:15
Air Blank	0.000	10:16
Subject Sample # 1	0.000	10:16
Breath Volume	1.468	10:16
Air Blank	0.000	10:17
Air Blank	0.000	10:19
Subject Sample # 2	0.000	10:19
Breath Volume	1.679	10:19
Air Blank	0.000	10:20
Control Test	0.080	10:20
Air Blank	0.000	10:21
Diagnostics Check	OK	10:21

Data Download Date/Time: 10/25/2023 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,D,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RONGA,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/09/2023 15:17:47

Diagnostics Check	OK	15:10
Air Blank	0.000	15:11
Control Test	0.080	15:11
Air Blank	0.000	15:11
Subject Sample # 1	0.041	15:12
Breath Volume	1.242	15:12
Air Blank	0.000	15:13
Air Blank	0.000	15:15
Subject Sample # 2	0.038	15:16
Breath Volume	1.414	15:16
Air Blank	0.000	15:16
Control Test	0.079	15:17
Air Blank	0.000	15:17
Diagnostics Check	OK	15:17

Data Download Date/Time: 10/25/2023 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORETON,MYKELL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,MYKELL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: M655426684270

DATE/TIME OF TEST: 10/03/2023 18:48:17

Diagnostics Check	OK	18:45	
Air Blank	0.000	18:45	
Control Test	0.081	18:46	
Air Blank	0.000	18:46	
Subject Sample	# 1	18:46	Subject Test Refused
Breath Volume	0.000	18:46	
Air Blank	0.000	18:47	
Control Test	0.081	18:47	
Air Blank	0.000	18:48	
Diagnostics Check	OK	18:48	

Data Download Date/Time: 10/05/2023 01:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LEON, SERGIO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DE LEON, SERGIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/04/2023 19:51:22

Diagnostics Check	OK	19:50	
Air Blank		19:50	RFI Detect
Air Blank	0.000	19:51	

Data Download Date/Time: 10/05/2023 01:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LEON, SERGIO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DE LEON, SERGIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/04/2023 19:58:51

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:55	
Control Test	0.081	19:55	
Air Blank	0.000	19:56	
Subject Sample	# 1	19:57	Subject Test Refused
Breath Volume	0.000	19:57	
Air Blank	0.000	19:57	
Control Test	0.081	19:58	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:58	

Data Download Date/Time: 10/05/2023 01:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORETON,MYKELL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,MYKELL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CORRALES CRUZ

DATE/TIME OF TEST: 10/08/2023 03:34:39

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.081	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.263	03:30
Breath Volume	1.812	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.252	03:32
Breath Volume	2.164	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 11/03/2023 02:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GONZALEZ,MADELIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:58

SUBJECT'S LAST NAME: FLORES TEJADA

DATE/TIME OF TEST: 10/11/2023 09:24:41

Diagnostics Check	OK	09:18
Air Blank	0.000	09:19
Control Test	0.080	09:19
Air Blank	0.000	09:19
Subject Sample # 1	0.119	09:20
Breath Volume	1.355	09:20
Air Blank	0.000	09:21
Air Blank	0.000	09:22
Subject Sample # 2	0.119	09:23
Breath Volume	1.406	09:23
Air Blank	0.000	09:23
Control Test	0.079	09:24
Air Blank	0.000	09:24
Diagnostics Check	OK	09:24

Data Download Date/Time: 11/03/2023 02:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORETON,MYKELL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,MYKELL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: PARADASUAREZ

DATE/TIME OF TEST: 10/22/2023 04:10:54

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.055	04:06
Breath Volume	1.816	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.044	04:09
Breath Volume	1.835	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 11/03/2023 02:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORETON,MYKELL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: PARADA-SUAREZ

DATE/TIME OF TEST: 10/22/2023 06:05:52

Diagnostics Check	OK	05:59
Air Blank	0.000	06:00
Control Test	0.081	06:00
Air Blank	0.000	06:01
Subject Sample # 1	0.027	06:01
Breath Volume	1.980	06:01
Air Blank	0.000	06:02
Air Blank	0.000	06:03
Subject Sample # 2	0.023	06:04
Breath Volume	1.800	06:04
Air Blank	0.000	06:04
Control Test	0.079	06:05
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time: 11/03/2023 02:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: ALVAREZ LOPEZ

DATE/TIME OF TEST: 10/22/2023 23:35:37

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.154	23:30
Breath Volume	2.656	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.148	23:33
Breath Volume	2.093	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 11/03/2023 02:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUADRA,MARVIN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SIRIANI,FRANCISCO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/27/2023 23:37:50

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:28	
Control Test	0.080	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:32	No Sample Provided
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Air Blank	0.000	23:34	
Subject Sample	# 2	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:36	
Control Test	0.079	23:37	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 11/03/2023 02:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COCHRAN,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 10/05/2023 01:49:59

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.000	01:45
Breath Volume	1.429	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.000	01:48
Breath Volume	1.808	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 10/25/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY-IVERSEN,MIYA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAY-IVERSEN,MIYA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HOBBLEY

DATE/TIME OF TEST: 10/06/2023 01:32:00

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.049	01:27
Breath Volume	1.777	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.048	01:30
Breath Volume	1.421	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:32

Data Download Date/Time: 10/25/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: BECKER,JONATHAN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 02:53

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MCVAY

DATE/TIME OF TEST: 10/09/2023 03:37:07

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.200	03:32
Breath Volume	2.902	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.200	03:35
Breath Volume	2.507	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 10/25/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLLINS,CHASE,M

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COLLINS,CHASE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 12:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:24

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 10/09/2023 13:51:11

Diagnostics Check	OK	13:44
Air Blank	0.000	13:45
Control Test	0.082	13:45
Air Blank	0.000	13:45
Subject Sample # 1	0.033	13:46
Breath Volume	1.472	13:46
Air Blank	0.000	13:47
Air Blank	0.000	13:49
Subject Sample # 2	0.032	13:49
Breath Volume	1.335	13:49
Air Blank	0.000	13:50
Control Test	0.081	13:50
Air Blank	0.000	13:51
Diagnostics Check	OK	13:51

Data Download Date/Time: 10/25/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HASWELL,AUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: HENNINGTON

DATE/TIME OF TEST: 10/11/2023 17:20:53

Diagnostics Check	OK	17:14
Air Blank	0.000	17:14
Control Test	0.081	17:15
Air Blank	0.000	17:15
Subject Sample # 1	0.000	17:16
Breath Volume	2.195	17:16
Air Blank	0.000	17:16
Air Blank	0.000	17:18
Subject Sample # 2	0.000	17:19
Breath Volume	2.148	17:19
Air Blank	0.000	17:19
Control Test	0.081	17:20
Air Blank	0.000	17:20
Diagnostics Check	OK	17:20

Data Download Date/Time: 10/25/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTNER,STEVEN,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTNER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: NORWOOD

DATE/TIME OF TEST: 10/14/2023 04:22:08

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.082	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.000	04:17
Breath Volume	3.742	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.000	04:20
Breath Volume	3.964	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 10/25/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: BECKER,JONATHAN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 03:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MUZAHM

DATE/TIME OF TEST: 10/17/2023 04:02:54

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.082	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.140	03:58
Breath Volume	1.667	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.134	04:01
Breath Volume	4.269	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 10/25/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GORDON,GREGORY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:39

SUBJECT'S LAST NAME: SHELBY

DATE/TIME OF TEST: 10/20/2023 11:10:25

Diagnostics Check	OK	11:04
Air Blank	0.000	11:04
Control Test	0.082	11:04
Air Blank	0.000	11:05
Subject Sample # 1	0.000	11:05
Breath Volume	2.343	11:05
Air Blank	0.000	11:06
Air Blank	0.000	11:08
Subject Sample # 2	0.000	11:09
Breath Volume	1.148	11:09
Air Blank	0.000	11:09
Control Test	0.081	11:09
Air Blank	0.000	11:10
Diagnostics Check	OK	11:10

Data Download Date/Time: 10/25/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 23:18

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 10/21/2023 23:57:28

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.082	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.224	23:52
Breath Volume	1.765	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.216	23:55
Breath Volume	2.359	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 10/25/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SUSTAITA,CHRISTIAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: DORANTES

DATE/TIME OF TEST: 10/01/2023 03:35:27

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.191	03:30
Breath Volume	1.636	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.178	03:33
Breath Volume	1.441	03:33
Air Blank	0.000	03:34
Control Test	0.075	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 10/22/2023 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SUSTAITA,CRISTIAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/02/2023 00:27:49

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.080	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:26	
Control Test	0.081	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 10/22/2023 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: MORT,BROOKE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: CORTES JIMENEZ

DATE/TIME OF TEST: 10/10/2023 22:59:16

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.246	22:54
Breath Volume	1.808	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.235	22:57
Breath Volume	1.847	22:57
Air Blank	0.000	22:58
Control Test	0.076	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 10/22/2023 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 11:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BUSBY

DATE/TIME OF TEST: 10/19/2023 04:35:39

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:33	
Control Test	0.079	04:33	
Air Blank	0.000	04:33	
Subject Sample	# 1	04:34	Subject Test Refused
Breath Volume	0.000	04:34	
Air Blank	0.000	04:34	
Control Test	0.081	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 10/22/2023 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 10/20/2023 20:35:56

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.081	20:29
Air Blank	0.000	20:29
Subject Sample # 1	0.296	20:30
Breath Volume	1.289	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:32
Subject Sample # 2	0.298	20:34
Breath Volume	1.289	20:34
Air Blank	0.000	20:34
Control Test	0.079	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 10/22/2023 21:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANDY,JAMIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SUSTAITA,CRISTIAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 09/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/22/2023 06:25:33

Diagnostics Check	OK	06:09	
Air Blank	0.000	06:09	
Control Test	0.081	06:09	
Air Blank	0.000	06:10	
Subject Sample	# 1 0.120	06:13	Volume Not Met
Breath Volume	0.234	06:13	
Air Blank	0.000	06:14	
Air Blank	0.000	06:16	
Subject Sample	# 2 0.163	06:18	
Breath Volume	1.730	06:18	
Air Blank	0.000	06:19	
Air Blank	0.000	06:21	
Subject Sample	# 3 0.165	06:23	
Breath Volume	1.230	06:23	
Air Blank	0.000	06:24	
Control Test	0.079	06:24	
Air Blank	0.000	06:25	
Diagnostics Check	OK	06:25	

Data Download Date/Time: 10/22/2023 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVELL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/26/2023 22:01:39

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:54	
Control Test	0.080	21:54	
Air Blank	0.000	21:55	
Subject Sample	# 1 0.131	21:56	
Breath Volume	1.339	21:56	
Air Blank	0.000	21:57	
Air Blank	0.000	21:59	
Subject Sample	# 2	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.076	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 12/05/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SMITH,DUSTIN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/29/2023 00:51:23

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.081	00:49	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:50	
Control Test	0.081	00:50	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 12/05/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVELL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SAKIHAMA

DATE/TIME OF TEST: 10/31/2023 02:21:08

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.087	02:16
Breath Volume	1.167	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.081	02:19
Breath Volume	1.394	02:19
Air Blank	0.000	02:20
Control Test	0.076	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 12/05/2023 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: PINO VAZQUEZ

DATE/TIME OF TEST: 10/01/2023 04:42:47

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.182	04:38
Breath Volume	3.386	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.180	04:41
Breath Volume	1.972	04:41
Air Blank	0.000	04:41
Control Test	0.079	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 10/02/2023 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: PATISHTAN MENDEZ

DATE/TIME OF TEST: 10/01/2023 05:10:32

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.080	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.167	05:05
Breath Volume	3.175	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.181	05:08
Breath Volume	3.164	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 10/02/2023 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: DESAI

DATE/TIME OF TEST: 10/02/2023 02:54:59

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.141	02:49
Breath Volume	1.140	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.128	02:53
Breath Volume	1.371	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 10/02/2023 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: DULIN

DATE/TIME OF TEST: 10/04/2023 00:42:49

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.093	00:37
Breath Volume	1.117	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.088	00:41
Breath Volume	1.816	00:41
Air Blank	0.000	00:41
Control Test	0.082	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 11/08/2023 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WIERZBA,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:12

SUBJECT'S LAST NAME: AVALOS JR

DATE/TIME OF TEST: 10/04/2023 08:39:25

Diagnostics Check	OK	08:32
Air Blank	0.000	08:32
Control Test	0.080	08:33
Air Blank	0.000	08:33
Subject Sample # 1	0.090	08:34
Breath Volume	1.527	08:34
Air Blank	0.000	08:35
Air Blank	0.000	08:37
Subject Sample # 2	0.089	08:37
Breath Volume	2.933	08:37
Air Blank	0.000	08:38
Control Test	0.079	08:38
Air Blank	0.000	08:39
Diagnostics Check	OK	08:39

Data Download Date/Time: 11/08/2023 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: FERNANDEZ RAMOS

DATE/TIME OF TEST: 10/05/2023 02:41:36

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.127	02:36
Breath Volume	1.492	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.128	02:39
Breath Volume	1.660	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 11/08/2023 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO, E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 10/07/2023 03:52:45

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.133	03:47
Breath Volume	1.851	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.142	03:51
Breath Volume	1.714	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 11/08/2023 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: PRYOR

DATE/TIME OF TEST: 10/07/2023 04:51:40

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.118	04:46
Breath Volume	1.316	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.109	04:50
Breath Volume	1.507	04:50
Air Blank	0.000	04:50
Control Test	0.079	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 11/08/2023 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHANNON,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:16

SUBJECT'S LAST NAME: LATORRE

DATE/TIME OF TEST: 10/07/2023 11:42:38

Diagnostics Check	OK	11:36
Air Blank	0.000	11:36
Control Test	0.080	11:37
Air Blank	0.000	11:37
Subject Sample # 1	0.000	11:38
Breath Volume	2.058	11:38
Air Blank	0.000	11:38
Air Blank	0.000	11:40
Subject Sample # 2	0.000	11:41
Breath Volume	2.121	11:41
Air Blank	0.000	11:41
Control Test	0.080	11:42
Air Blank	0.000	11:42
Diagnostics Check	OK	11:42

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WEST,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: HESSELTON

DATE/TIME OF TEST: 10/07/2023 19:05:38

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:54	
Control Test	0.081	18:54	
Air Blank	0.000	18:55	
Subject Sample	# 1	18:58	No Sample Provided
Breath Volume	0.000	18:58	
Air Blank	0.000	18:58	
Air Blank	0.000	19:00	
Subject Sample	# 2	19:03	Volume Not Met
Breath Volume	0.050	19:03	
Air Blank	0.000	19:04	
Control Test	0.079	19:05	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WEST,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: HESSELTON

DATE/TIME OF TEST: 10/07/2023 19:22:38

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:07	
Control Test	0.079	19:08	
Air Blank	0.000	19:08	
Subject Sample	# 1 0.193	19:11	Volume Not Met
Breath Volume	0.667	19:11	
Air Blank	0.000	19:12	
Air Blank	0.000	19:14	
Subject Sample	# 2 0.206	19:17	
Breath Volume	1.500	19:17	
Air Blank	0.000	19:18	
Air Blank	0.000	19:19	
Subject Sample	# 3 0.216	19:20	
Breath Volume	1.402	19:20	
Air Blank	0.000	19:21	
Control Test	0.077	19:22	
Air Blank	0.000	19:22	
Diagnostics Check	OK	19:22	

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 10/08/2023 02:17:18

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.294	02:12
Breath Volume	2.976	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.294	02:15
Breath Volume	2.492	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: BESON

DATE/TIME OF TEST: 10/08/2023 05:04:04

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.179	04:59
Breath Volume	3.312	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.184	05:02
Breath Volume	3.257	05:02
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:04

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WURSTER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:57

SUBJECT'S LAST NAME: PROENZA

DATE/TIME OF TEST: 10/08/2023 09:29:15

Diagnostics Check	OK	09:17	
Air Blank	0.000	09:18	
Control Test	0.080	09:18	
Air Blank	0.000	09:19	
Subject Sample	# 1 0.000	09:22	Volume Not Met
Breath Volume	0.625	09:22	
Air Blank	0.000	09:22	
Air Blank	0.000	09:24	
Subject Sample	# 2 0.000	09:27	Volume Not Met
Breath Volume	0.585	09:27	
Air Blank	0.000	09:28	
Control Test	0.082	09:28	
Air Blank	0.000	09:29	
Diagnostics Check	OK	09:29	

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WURSTER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:57

SUBJECT'S LAST NAME: PROENZA

DATE/TIME OF TEST: 10/08/2023 09:40:29

Diagnostics Check	OK	09:31
Air Blank	0.000	09:31
Control Test	0.080	09:32
Air Blank	0.000	09:32
Subject Sample # 1	0.000	09:34
Breath Volume	1.335	09:34
Air Blank	0.000	09:35
Air Blank	0.000	09:37
Subject Sample # 2	0.000	09:39
Breath Volume	1.144	09:39
Air Blank	0.000	09:39
Control Test	0.081	09:39
Air Blank	0.000	09:40
Diagnostics Check	OK	09:40

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PIERRE,C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 10/09/2023 00:00:30

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.081	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.156	23:55
Breath Volume	2.792	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.160	23:58
Breath Volume	2.972	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: VANG

DATE/TIME OF TEST: 10/09/2023 03:28:03

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.106	03:23
Breath Volume	3.210	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.109	03:26
Breath Volume	3.203	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: BROOKS III

DATE/TIME OF TEST: 10/09/2023 19:41:03

Diagnostics Check	OK	19:35
Air Blank	0.000	19:35
Control Test	0.080	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.000	19:36
Breath Volume	2.671	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.000	19:39
Breath Volume	2.605	19:39
Air Blank	0.000	19:40
Control Test	0.081	19:40
Air Blank	0.000	19:40
Diagnostics Check	OK	19:41

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BANG

DATE/TIME OF TEST: 10/13/2023 01:01:57

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.148	00:57
Breath Volume	1.488	00:57
Air Blank	0.000	00:58
Air Blank	0.000	00:59
Subject Sample # 2	0.149	01:00
Breath Volume	1.175	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/13/2023 04:52:42

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:40	
Control Test	0.080	04:40	
Air Blank	0.000	04:41	
Subject Sample	# 1 0.165	04:44	Volume Not Met
Breath Volume	0.632	04:44	
Air Blank	0.000	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 2 0.173	04:47	
Breath Volume	1.578	04:47	
Air Blank	0.000	04:48	
Air Blank	0.000	04:50	
Subject Sample	# 3 0.173	04:51	
Breath Volume	1.207	04:51	
Air Blank	0.000	04:51	
Control Test	0.079	04:52	
Air Blank	0.000	04:52	
Diagnostics Check	OK	04:52	

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: SHAHNARAD

DATE/TIME OF TEST: 10/13/2023 05:33:28

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.079	05:25	
Air Blank	0.000	05:26	
Subject Sample	# 1 0.069	05:29	Volume Not Met
Breath Volume	0.082	05:29	
Air Blank	0.000	05:30	
Air Blank	0.000	05:31	
Subject Sample	# 2	05:32	Subject Test Refused
Breath Volume	0.000	05:32	
Air Blank	0.000	05:32	
Control Test	0.078	05:32	
Air Blank	0.000	05:33	
Diagnostics Check	OK	05:33	

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: WALDEN

DATE/TIME OF TEST: 10/14/2023 05:30:45

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.080	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.137	05:26
Breath Volume	1.531	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.138	05:29
Breath Volume	2.011	05:29
Air Blank	0.000	05:29
Control Test	0.080	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: BURROUGHS

DATE/TIME OF TEST: 10/14/2023 05:50:18

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.079	05:43
Air Blank	0.000	05:44
Subject Sample # 1	0.145	05:45
Breath Volume	1.792	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.144	05:48
Breath Volume	1.929	05:48
Air Blank	0.000	05:49
Control Test	0.080	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:59

SUBJECT'S LAST NAME: RAYO NOLASCO

DATE/TIME OF TEST: 10/14/2023 14:29:00

Diagnostics Check	OK	14:22
Air Blank	0.000	14:23
Control Test	0.080	14:23
Air Blank	0.000	14:23
Subject Sample # 1	0.000	14:24
Breath Volume	3.511	14:24
Air Blank	0.000	14:25
Air Blank	0.000	14:27
Subject Sample # 2	0.000	14:27
Breath Volume	1.675	14:27
Air Blank	0.000	14:28
Control Test	0.080	14:28
Air Blank	0.000	14:28
Diagnostics Check	OK	14:29

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: HERNANDEZ ESCOBAR

DATE/TIME OF TEST: 10/15/2023 00:25:38

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.079	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.237	00:20
Breath Volume	1.449	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.240	00:23
Breath Volume	2.925	00:23
Air Blank	0.000	00:24
Control Test	0.078	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO, E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 10/15/2023 03:20:24

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.159	03:15
Breath Volume	1.953	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.143	03:18
Breath Volume	1.414	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACEO,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:47

SUBJECT'S LAST NAME: SANCHEZ BERDUO

DATE/TIME OF TEST: 10/15/2023 08:21:55

Diagnostics Check	OK	08:08	
Air Blank	0.000	08:09	
Control Test	0.080	08:09	
Air Blank	0.000	08:09	
Subject Sample	# 1 0.152	08:13	Volume Not Met
Breath Volume	0.628	08:13	
Air Blank	0.000	08:13	
Air Blank	0.000	08:15	
Subject Sample	# 2 0.184	08:17	
Breath Volume	1.851	08:17	
Air Blank	0.000	08:17	
Air Blank	0.000	08:19	
Subject Sample	# 3 0.191	08:20	
Breath Volume	2.789	08:20	
Air Blank	0.000	08:20	
Control Test	0.078	08:21	
Air Blank	0.000	08:21	
Diagnostics Check	OK	08:21	

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 10/15/2023 22:40:27

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.080	22:38	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:39	
Control Test	0.080	22:39	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HERNANDEZ GUADARAMMA

DATE/TIME OF TEST: 10/17/2023 00:55:59

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.180	00:51
Breath Volume	2.378	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:53
Subject Sample # 2	0.175	00:54
Breath Volume	2.105	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: GLASGOW

DATE/TIME OF TEST: 10/18/2023 03:10:18

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.109	03:05
Breath Volume	2.343	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.105	03:08
Breath Volume	3.335	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: ORJUELA

DATE/TIME OF TEST: 10/18/2023 04:32:39

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.076	04:28
Breath Volume	4.058	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.076	04:31
Breath Volume	3.941	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CHEW

DATE/TIME OF TEST: 10/18/2023 23:46:49

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.184	23:42
Breath Volume	2.125	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.176	23:45
Breath Volume	1.730	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/19/2023 03:31:17

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.135	03:26
Breath Volume	2.097	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.137	03:29
Breath Volume	1.828	03:29
Air Blank	0.000	03:30
Control Test	0.084	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: JIMENEZ PEREZ

DATE/TIME OF TEST: 10/20/2023 01:55:08

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.161	01:50
Breath Volume	3.339	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:52
Subject Sample # 2	0.156	01:53
Breath Volume	2.476	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 10/21/2023 01:42:21

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.090	01:37
Breath Volume	1.414	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.087	01:40
Breath Volume	1.531	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUEVARA, C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: MINKO

DATE/TIME OF TEST: 10/21/2023 03:22:04

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.179	03:15
Breath Volume	1.917	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.180	03:20
Breath Volume	1.609	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 10/21/2023 19:55:46

Diagnostics Check	OK	19:48
Air Blank	0.000	19:49
Control Test	0.081	19:49
Air Blank	0.000	19:49
Subject Sample # 1	0.229	19:51
Breath Volume	1.539	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.227	19:54
Breath Volume	1.195	19:54
Air Blank	0.000	19:54
Control Test	0.079	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOLITOR,M
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GREAVES

DATE/TIME OF TEST: 10/21/2023 22:08:52

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.183	22:04
Breath Volume	2.136	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.184	22:07
Breath Volume	1.691	22:07
Air Blank	0.000	22:07
Control Test	0.080	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/22/2023 00:54:32

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.080	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.257	00:49
Breath Volume	1.250	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.265	00:52
Breath Volume	1.292	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: KITCHELL

DATE/TIME OF TEST: 10/22/2023 01:16:31

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.073	01:12
Breath Volume	2.972	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.074	01:14
Breath Volume	3.156	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MAX PAAU DE LUCAS

DATE/TIME OF TEST: 10/22/2023 03:28:41

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.149	03:24
Breath Volume	1.695	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.157	03:27
Breath Volume	1.855	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 10/22/2023 04:06:20

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.160	04:01
Breath Volume	1.222	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.167	04:04
Breath Volume	1.480	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 11/08/2023 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: FONSECA MACIAS

DATE/TIME OF TEST: 10/22/2023 05:08:34

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.101	05:03
Breath Volume	1.484	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:05
Subject Sample # 2	0.101	05:06
Breath Volume	1.339	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 10/23/2023 00:22:50

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.231	00:18
Breath Volume	2.250	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.234	00:21
Breath Volume	2.312	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ESCALANTE LUCAS

DATE/TIME OF TEST: 10/23/2023 20:17:32

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.233	20:12
Breath Volume	2.371	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.231	20:15
Breath Volume	2.421	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: CUBAS RAMOS

DATE/TIME OF TEST: 10/24/2023 01:44:28

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.117	01:39
Breath Volume	1.867	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.115	01:42
Breath Volume	2.753	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSELLES, F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MEZA GUZMAN

DATE/TIME OF TEST: 10/26/2023 02:11:07

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.175	02:06
Breath Volume	1.488	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.181	02:09
Breath Volume	1.468	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: CARABALLO VARGAS

DATE/TIME OF TEST: 10/27/2023 00:16:37

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.080	00:06	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:09	No Sample Provided
Breath Volume	0.000	00:09	
Air Blank	0.000	00:10	
Air Blank	0.000	00:12	
Subject Sample	# 2	00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:15	
Control Test	0.081	00:16	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HILALI

DATE/TIME OF TEST: 10/27/2023 01:17:59

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.080	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.081	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/27/2023 22:22:51

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.081	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.172	22:16
Breath Volume	1.625	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.176	22:21
Breath Volume	1.800	22:21
Air Blank	0.000	22:21
Control Test	0.081	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: CALISTRO

DATE/TIME OF TEST: 10/28/2023 03:18:19

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.134	03:13
Breath Volume	1.183	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.147	03:16
Breath Volume	1.562	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHIRELING,S
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: GALICIA MERIDA

DATE/TIME OF TEST: 10/28/2023 03:42:10

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:29	
Control Test	0.080	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1 0.215	03:31	No .020 agreement
Breath Volume	1.453	03:31	
Air Blank	0.000	03:31	
Air Blank	0.000	03:33	
Subject Sample	# 2 0.220	03:36	Volume Not Met
Breath Volume	0.464	03:36	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 3 0.249	03:40	No .020 agreement
Breath Volume	2.761	03:40	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHIRELING,S
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: GALICIA MERIDA

DATE/TIME OF TEST: 10/28/2023 03:52:35

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.219	03:47
Breath Volume	1.621	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.223	03:50
Breath Volume	1.609	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 10/28/2023 09:57:23

Diagnostics Check	OK	09:50
Air Blank	0.000	09:50
Control Test	0.080	09:50
Air Blank	0.000	09:51
Subject Sample # 1	0.424	09:52
Breath Volume	2.175	09:52
Air Blank	0.000	09:53
Air Blank	0.000	09:54
Subject Sample # 2	0.405	09:55
Breath Volume	2.199	09:55
Air Blank	0.000	09:56
Control Test	0.080	09:56
Air Blank	0.000	09:57
Diagnostics Check	OK	09:57

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUEVARA, C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/28/2023 22:57:18

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.129	22:52
Breath Volume	2.050	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.131	22:55
Breath Volume	1.730	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MAYFIELD

DATE/TIME OF TEST: 10/28/2023 23:22:23

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.117	23:17
Breath Volume	2.886	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.117	23:20
Breath Volume	3.570	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 10/29/2023 02:05:51

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.108	02:01
Breath Volume	1.710	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample # 2	0.099	02:04
Breath Volume	1.804	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: O'KELLY

DATE/TIME OF TEST: 10/29/2023 02:42:48

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.156	02:38
Breath Volume	1.968	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.150	02:41
Breath Volume	1.894	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 10/29/2023 03:49:35

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.196	03:44
Breath Volume	2.761	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.181	03:47
Breath Volume	1.628	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RIDENOUR

DATE/TIME OF TEST: 10/29/2023 04:19:44

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.079	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1 0.161	04:12	Volume Not Met
Breath Volume	0.015	04:12	
Air Blank	0.000	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 2 0.140	04:18	Slope Not Level
Breath Volume	2.226	04:18	
Air Blank	0.000	04:18	
Control Test	0.078	04:19	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RIDENOUR

DATE/TIME OF TEST: 10/29/2023 04:30:46

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.078	04:22	
Air Blank	0.000	04:22	
Subject Sample	# 1 0.158	04:25	Volume Not Met
Breath Volume	0.414	04:25	
Air Blank	0.000	04:26	
Air Blank	0.000	04:28	
Subject Sample	# 2	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.078	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: VARGAS SALAS

DATE/TIME OF TEST: 10/29/2023 11:02:09

Diagnostics Check	OK	10:55
Air Blank	0.000	10:55
Control Test	0.080	10:56
Air Blank	0.000	10:56
Subject Sample # 1	0.193	10:57
Breath Volume	2.109	10:57
Air Blank	0.000	10:58
Air Blank	0.000	11:00
Subject Sample # 2	0.182	11:00
Breath Volume	1.472	11:00
Air Blank	0.000	11:01
Control Test	0.079	11:01
Air Blank	0.000	11:02
Diagnostics Check	OK	11:02

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/31/2023 01:34:29

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.206	01:29
Breath Volume	1.914	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.203	01:32
Breath Volume	2.062	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: THREATT JR

DATE/TIME OF TEST: 10/31/2023 23:43:19

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:35	
Control Test	0.081	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:39	No Sample Provided
Breath Volume	0.000	23:39	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample	# 2	23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:42	
Control Test	0.080	23:42	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 11/08/2023 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/04/2023 03:28:57

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.081	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.081	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 10/04/2023 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESILVA,JOEY,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ESILVA,JOEY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: LEBO

DATE/TIME OF TEST: 10/09/2023 16:57:39

Diagnostics Check	OK	16:46	
Air Blank	0.000	16:46	
Control Test	0.080	16:47	
Air Blank	0.000	16:47	
Subject Sample	# 1	16:50	No Sample Provided
Breath Volume	0.000	16:50	
Air Blank	0.000	16:51	
Air Blank	0.000	16:53	
Subject Sample	# 2	16:56	No Sample Provided
Breath Volume	0.000	16:56	
Air Blank	0.000	16:56	
Control Test	0.081	16:57	
Air Blank	0.000	16:57	
Diagnostics Check	OK	16:57	

Data Download Date/Time: 11/08/2023 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRON,DAVID,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: WINTER,PARIS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HERRON

DATE/TIME OF TEST: 10/09/2023 19:48:07

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.081	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.195	19:43
Breath Volume	1.652	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:45
Subject Sample # 2	0.203	19:46
Breath Volume	2.421	19:46
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 11/08/2023 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/15/2023 02:23:58

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.081	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 11/08/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MYHRE

DATE/TIME OF TEST: 10/15/2023 04:04:30

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.230	03:59
Breath Volume	1.976	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.233	04:02
Breath Volume	2.937	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 11/08/2023 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 10/15/2023 23:24:36

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.082	23:22	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:23	
Control Test	0.081	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 11/08/2023 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,BRIAN,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SARGENT,RIANA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: BUI
ARREST TIME: 13:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/19/2023 14:43:12

Diagnostics Check	OK	14:39	
Air Blank	0.000	14:39	
Control Test	0.080	14:40	
Air Blank	0.000	14:40	
Subject Sample	# 1	14:41	Subject Test Refused
Breath Volume	0.000	14:41	
Air Blank	0.000	14:42	
Control Test	0.080	14:42	
Air Blank	0.000	14:43	
Diagnostics Check	OK	14:43	

Data Download Date/Time: 11/08/2023 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN, ENOCH, A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ESILVA, JOEY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: BUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 10/19/2023 23:19:14

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:08	
Control Test	0.081	23:08	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:12	No Sample Provided
Breath Volume	0.000	23:12	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample	# 2	23:17	No Sample Provided
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Control Test	0.081	23:18	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 11/08/2023 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: TILLERY

DATE/TIME OF TEST: 10/21/2023 01:04:14

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 11/08/2023 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: REIDY

DATE/TIME OF TEST: 10/21/2023 01:54:22

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.080	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 11/08/2023 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BENTURA LOPEZ

DATE/TIME OF TEST: 10/22/2023 03:31:27

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.080	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 11/08/2023 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 10/25/2023 00:22:05

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.081	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 11/08/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: STONELY

DATE/TIME OF TEST: 10/25/2023 01:54:31

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.081	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.085	01:49
Breath Volume	1.761	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.081	01:52
Breath Volume	1.464	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 11/08/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUBA,JACK,Z
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CLARK,THOMAS
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: GRIGGS

DATE/TIME OF TEST: 10/26/2023 01:05:51

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.177	01:01
Breath Volume	2.406	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:03
Subject Sample # 2	0.160	01:04
Breath Volume	1.300	01:04
Air Blank	0.000	01:04
Control Test	0.079	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 11/08/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB-HOCKMAN,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: COBB-HOCKMAN,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: BUI
ARREST TIME: 15:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 10/26/2023 17:01:06

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:52	
Control Test	0.080	16:52	
Air Blank	0.000	16:52	
Subject Sample	# 1	16:56	No Sample Provided
Breath Volume	0.000	16:56	
Air Blank	0.000	16:56	
Air Blank	0.000	16:58	
Subject Sample	# 2	16:59	Subject Test Refused
Breath Volume	0.000	16:59	
Air Blank	0.000	17:00	
Control Test	0.079	17:00	
Air Blank	0.000	17:01	
Diagnostics Check	OK	17:01	

Data Download Date/Time: 11/08/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: LITE

DATE/TIME OF TEST: 10/27/2023 21:48:59

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.081	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.092	21:44
Breath Volume	1.257	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:46
Subject Sample # 2	0.089	21:47
Breath Volume	1.570	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 11/08/2023 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CVHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PASTELLA

DATE/TIME OF TEST: 10/30/2023 02:15:00

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.081	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.081	02:14	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 11/08/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIERRO,ALEXANDER,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FIERRO,ALEXANDER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 10/30/2023 04:03:59

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.137	03:59
Breath Volume	1.839	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.131	04:02
Breath Volume	1.464	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 11/08/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BETZ,FRANK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 06:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 10/31/2023 07:46:14

Diagnostics Check	OK	07:39
Air Blank	0.000	07:39
Control Test	0.080	07:40
Air Blank	0.000	07:40
Subject Sample # 1	0.143	07:41
Breath Volume	1.964	07:41
Air Blank	0.000	07:42
Air Blank	0.000	07:43
Subject Sample # 2	0.148	07:44
Breath Volume	1.789	07:44
Air Blank	0.000	07:45
Control Test	0.080	07:45
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

Data Download Date/Time: 11/08/2023 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTERO,NICHOLAS,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MONTERO,NICHOLAS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: BUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PICADO

DATE/TIME OF TEST: 10/05/2023 21:09:09

Diagnostics Check	OK	21:06	
Air Blank	0.000	21:06	
Control Test	0.080	21:07	
Air Blank	0.000	21:07	
Subject Sample	# 1	21:07	Subject Test Refused
Breath Volume	0.000	21:07	
Air Blank	0.000	21:08	
Control Test	0.080	21:08	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

Data Download Date/Time: 10/24/2023 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADLER,COREY,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ADLER,COREY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: BUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 10/21/2023 23:08:12

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.080	22:57	
Air Blank	0.000	22:57	
Subject Sample	# 1 0.099	23:00	Volume Not Met
Breath Volume	0.015	23:00	
Air Blank	0.000	23:01	
Air Blank	0.000	23:03	
Subject Sample	# 2 0.087	23:06	Volume Not Met
Breath Volume	0.070	23:06	
Air Blank	0.000	23:07	
Control Test	0.078	23:07	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 10/24/2023 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FLANEGAN,BROOKE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: LYSIAK

DATE/TIME OF TEST: 10/01/2023 00:39:55

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:36	
Control Test	0.080	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.079	00:39	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 10/05/2023 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DULLY,STEPHEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MARTINEZ AVALOS

DATE/TIME OF TEST: 10/03/2023 00:43:15

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.079	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Control Test	0.079	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 10/05/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGE,DANIEL,T
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GIVENS,ETHAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: BUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/07/2023 21:12:26

Diagnostics Check	OK	21:11	
Air Blank		21:11	Ambient Fail
Air Blank	0.000	21:12	

Data Download Date/Time: 11/09/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGE,DANIEL,T
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GIVENS,ETHAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: BUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 10/07/2023 21:23:55

Diagnostics Check	OK	21:15
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:16
Subject Sample # 1	0.071	21:18
Breath Volume	1.746	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.081	21:22
Breath Volume	1.871	21:22
Air Blank	0.000	21:22
Control Test	0.078	21:23
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 11/09/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,MEGHAN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: NEPIUS,LENS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: KEEGAN

DATE/TIME OF TEST: 10/08/2023 03:04:13

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.116	02:59
Breath Volume	2.187	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.120	03:02
Breath Volume	2.562	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 11/09/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,MEGHAN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: RODRIGUEZ MOYA

DATE/TIME OF TEST: 10/08/2023 03:38:29

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.119	03:33
Breath Volume	2.628	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.124	03:36
Breath Volume	2.183	03:36
Air Blank	0.000	03:37
Control Test	0.077	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 11/09/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILROY,ZACHARY,I
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRERO-MARTINEZ,KEVEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: LEMOINE

DATE/TIME OF TEST: 10/12/2023 23:45:36

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Control Test	0.080	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 11/09/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/14/2023 02:15:35

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 11/09/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: AXUME VASQUEZ

DATE/TIME OF TEST: 10/20/2023 03:27:17

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.079	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.081	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 11/09/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,MEGHAN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 10/22/2023 04:43:25

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.170	04:37
Breath Volume	1.257	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.162	04:41
Breath Volume	2.429	04:41
Air Blank	0.000	04:42
Control Test	0.076	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 11/09/2023 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FOSTER,MEGHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CHIL SIRET

DATE/TIME OF TEST: 10/22/2023 22:58:08

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:55	
Control Test	0.079	22:56	
Air Blank	0.000	22:56	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Control Test	0.080	22:57	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 11/09/2023 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DEL CUETO,ERYKA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: BALBUENA

DATE/TIME OF TEST: 10/24/2023 20:40:48

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:38	
Control Test	0.080	20:38	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:39	Subject Test Refused
Breath Volume	0.000	20:39	
Air Blank	0.000	20:39	
Control Test	0.080	20:40	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:40	

Data Download Date/Time: 11/09/2023 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PEREZ-LOPEZ

DATE/TIME OF TEST: 10/05/2023 23:06:26

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.050	23:01
Breath Volume	1.472	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.050	23:04
Breath Volume	1.910	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 10/08/2023 01:33:10

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.161	01:27
Breath Volume	1.453	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.176	01:31
Breath Volume	1.367	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BARBOZA,JONATHAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: PENUEL

DATE/TIME OF TEST: 10/09/2023 00:48:24

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:33	
Control Test	0.081	00:33	
Air Blank	0.000	00:34	
Subject Sample	# 1 0.207	00:35	
Breath Volume	2.210	00:35	
Air Blank	0.000	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 2	00:40	No Sample Provided
Breath Volume	0.000	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 3 0.202	00:46	
Breath Volume	2.312	00:46	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: MASIELLO

DATE/TIME OF TEST: 10/11/2023 21:28:04

Diagnostics Check	OK	21:27	
Air Blank		21:27	RFI Detect
Air Blank		21:28	RFI Detect

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: MASIELLO

DATE/TIME OF TEST: 10/11/2023 21:33:30

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:30	
Control Test	0.079	21:31	
Air Blank	0.000	21:31	
Subject Sample	# 1	21:32	RFI Detect
Breath Volume	0.000	21:32	
Air Blank	0.000	21:32	
Control Test	0.079	21:32	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: MASIELLO

DATE/TIME OF TEST: 10/11/2023 21:41:38

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.214	21:36
Breath Volume	1.300	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.212	21:39
Breath Volume	1.492	21:39
Air Blank	0.000	21:40
Control Test	0.078	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MENDEZ COLON

DATE/TIME OF TEST: 10/14/2023 23:50:35

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.146	23:46
Breath Volume	2.824	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.147	23:48
Breath Volume	2.851	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/17/2023 01:08:49

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.136	01:04
Breath Volume	1.503	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.144	01:07
Breath Volume	1.378	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: DIAZ RIOS

DATE/TIME OF TEST: 10/18/2023 03:21:12

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.177	03:16
Breath Volume	1.777	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.170	03:19
Breath Volume	1.609	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FULFORD

DATE/TIME OF TEST: 10/18/2023 03:36:14

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.191	03:31
Breath Volume	2.207	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.187	03:34
Breath Volume	2.312	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 10/21/2023 03:07:16

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.207	03:02
Breath Volume	2.523	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.190	03:05
Breath Volume	2.296	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 10/25/2023 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 10/01/2023 02:37:27

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.126	02:32
Breath Volume	2.570	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.115	02:35
Breath Volume	2.273	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUILLARD,ALEXANDER,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GUILLARD,ALEXANDER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: CLASS

DATE/TIME OF TEST: 10/01/2023 08:32:16

Diagnostics Check	OK	08:25
Air Blank	0.000	08:26
Control Test	0.079	08:26
Air Blank	0.000	08:27
Subject Sample # 1	0.193	08:27
Breath Volume	3.976	08:27
Air Blank	0.000	08:28
Air Blank	0.000	08:30
Subject Sample # 2	0.200	08:30
Breath Volume	3.113	08:30
Air Blank	0.000	08:31
Control Test	0.078	08:31
Air Blank	0.000	08:32
Diagnostics Check	OK	08:32

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSNYAK,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOSNYAK,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 10:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: SMOKE

DATE/TIME OF TEST: 10/02/2023 12:20:45

Diagnostics Check	OK	12:12
Air Blank	0.000	12:12
Control Test	0.079	12:13
Air Blank	0.000	12:13
Subject Sample # 1	0.175	12:14
Breath Volume	2.191	12:14
Air Blank	0.000	12:15
Air Blank	0.000	12:16
Subject Sample # 2	0.162	12:18
Breath Volume	1.796	12:18
Air Blank	0.000	12:19
Control Test	0.078	12:20
Air Blank	0.000	12:20
Diagnostics Check	OK	12:20

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENCHAUTEQUI,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ENCHAUTEQUI,JUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2023
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: BRANNEN

DATE/TIME OF TEST: 10/02/2023 18:41:07

Diagnostics Check	OK	18:38	
Air Blank	0.000	18:38	
Control Test	0.079	18:38	
Air Blank	0.000	18:39	
Subject Sample	# 1	18:39	Subject Test Refused
Breath Volume	0.000	18:39	
Air Blank	0.000	18:40	
Control Test	0.080	18:40	
Air Blank	0.000	18:41	
Diagnostics Check	OK	18:41	

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EAGLEN,LIVIU,N
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: EAGLEN,LIVIU
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BENAC

DATE/TIME OF TEST: 10/03/2023 21:22:57

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.078	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.142	21:18
Breath Volume	2.253	21:18
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.141	21:21
Breath Volume	1.886	21:21
Air Blank	0.000	21:21
Control Test	0.078	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERTY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 10/04/2023 01:15:20

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.214	01:10
Breath Volume	1.558	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.227	01:13
Breath Volume	1.878	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DILLON,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,DILLON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 10/06/2023 03:18:55

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.181	03:13
Breath Volume	2.718	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.179	03:17
Breath Volume	2.199	03:17
Air Blank	0.000	03:17
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/07/2023 03:08:28

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.178	03:02
Breath Volume	1.375	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.183	03:06
Breath Volume	2.394	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERT,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PEREZ,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 10/07/2023 21:05:22

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.080	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.284	20:56
Breath Volume	1.542	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:58
Subject Sample # 2	0.260	21:00
Breath Volume	1.386	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 3	0.263	21:03
Breath Volume	1.476	21:03
Air Blank	0.000	21:04
Control Test	0.076	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MAYS,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BOLDUC

DATE/TIME OF TEST: 10/07/2023 23:24:20

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.079	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.131	23:17
Breath Volume	1.121	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.139	23:22
Breath Volume	1.226	23:22
Air Blank	0.000	23:23
Control Test	0.078	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 11:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DIBERARDINO

DATE/TIME OF TEST: 10/09/2023 01:23:32

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.080	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFENBERG,ASHTON,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WELFENBERG,ASHTON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/12/2023 22:23:10

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.010	22:18
Breath Volume	1.574	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.008	22:21
Breath Volume	1.246	22:21
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 10/13/2023 23:08:05

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.136	23:03
Breath Volume	2.226	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:05
Subject Sample # 2	0.136	23:06
Breath Volume	1.867	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEMLER,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SEMLER,JOSEPH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 10/13/2023 23:46:59

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.251	23:41
Breath Volume	4.101	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.248	23:44
Breath Volume	2.257	23:44
Air Blank	0.000	23:46
Control Test	0.076	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 10/14/2023 01:45:13

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.114	01:40
Breath Volume	2.574	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.111	01:43
Breath Volume	1.839	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: URONU,PATRICK,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: URONU,PATRICK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: DORTCH

DATE/TIME OF TEST: 10/14/2023 06:54:03

Diagnostics Check	OK	06:46
Air Blank	0.000	06:47
Control Test	0.079	06:47
Air Blank	0.000	06:48
Subject Sample # 1	0.283	06:48
Breath Volume	1.710	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:51
Subject Sample # 2	0.280	06:52
Breath Volume	1.628	06:52
Air Blank	0.000	06:53
Control Test	0.077	06:53
Air Blank	0.000	06:53
Diagnostics Check	OK	06:54

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,BRYAN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOWMAN,BRYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:34

SUBJECT'S LAST NAME: ACHESON

DATE/TIME OF TEST: 10/14/2023 09:05:44

Diagnostics Check	OK	08:58
Air Blank	0.000	08:58
Control Test	0.079	08:58
Air Blank	0.000	08:59
Subject Sample # 1	0.074	09:00
Breath Volume	2.558	09:00
Air Blank	0.000	09:01
Air Blank	0.000	09:03
Subject Sample # 2	0.070	09:03
Breath Volume	2.074	09:03
Air Blank	0.000	09:04
Control Test	0.078	09:05
Air Blank	0.000	09:05
Diagnostics Check	OK	09:05

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZEZUSKI,JIMMY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZEZUSKI,JIMMY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PAZ RODAS

DATE/TIME OF TEST: 10/15/2023 01:41:08

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.147	01:36
Breath Volume	1.875	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample # 2	0.144	01:39
Breath Volume	1.718	01:39
Air Blank	0.000	01:40
Control Test	0.077	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MONES

DATE/TIME OF TEST: 10/15/2023 02:46:14

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.079	02:41	
Air Blank	0.000	02:41	
Subject Sample # 1	0.160	02:42	
Breath Volume	1.257	02:42	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample # 2		02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Control Test	0.077	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: MONES

DATE/TIME OF TEST: 10/15/2023 03:08:39

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:05	
Control Test	0.079	03:05	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.080	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DILLON,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,DILLON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 10/15/2023 04:49:36

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.172	04:42
Breath Volume	2.511	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.180	04:47
Breath Volume	3.925	04:47
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONDANK, BLAKE, P
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PATTERSON, HANNAH
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: CARBAJAL

DATE/TIME OF TEST: 10/15/2023 23:10:51

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.219	23:04
Breath Volume	1.960	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.202	23:08
Breath Volume	1.785	23:08
Air Blank	0.000	23:09
Control Test	0.076	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 10/25/2023 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABBE,KYLER,S
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LABBE,KYLER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PRINCE-WEST

DATE/TIME OF TEST: 10/16/2023 02:54:33

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1 0.111	02:45	Volume Not Met
Breath Volume	0.253	02:45	
Air Blank	0.000	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 2 0.116	02:49	
Breath Volume	1.734	02:49	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 3 0.123	02:52	
Breath Volume	1.132	02:52	
Air Blank	0.000	02:53	
Control Test	0.078	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 10/25/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLANSON,JACE,W
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: DAVIS,D
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CLEARWATER

DATE/TIME OF TEST: 10/20/2023 23:47:56

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.038	23:43
Breath Volume	2.800	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample # 2	0.038	23:46
Breath Volume	2.503	23:46
Air Blank	0.000	23:46
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 10/25/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLANSON,JACE,W
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: ALLANSON,JACE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DALES

DATE/TIME OF TEST: 10/21/2023 00:02:25

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.139	23:57
Breath Volume	2.261	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.139	00:00
Breath Volume	1.863	00:00
Air Blank	0.000	00:01
Control Test	0.077	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 10/25/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:08

SUBJECT'S LAST NAME: ESCOBAR ORTEGA

DATE/TIME OF TEST: 10/21/2023 10:38:28

Diagnostics Check	OK	10:32	
Air Blank	0.000	10:33	
Control Test	0.080	10:33	
Air Blank	0.000	10:33	
Subject Sample # 1	0.214	10:34	
Breath Volume	3.437	10:34	
Air Blank	0.000	10:35	
Air Blank	0.000	10:37	
Subject Sample # 2	0.203	10:37	
Breath Volume	3.226	10:37	
Air Blank		10:37	RFI Detect
Air Blank	0.000	10:38	

Data Download Date/Time: 10/25/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:08

SUBJECT'S LAST NAME: ESCOBAR ORTEGA

DATE/TIME OF TEST: 10/21/2023 10:46:40

Diagnostics Check	OK	10:40
Air Blank	0.000	10:40
Control Test	0.077	10:40
Air Blank	0.000	10:41
Subject Sample # 1	0.206	10:41
Breath Volume	3.542	10:41
Air Blank	0.000	10:42
Air Blank	0.000	10:44
Subject Sample # 2	0.203	10:44
Breath Volume	3.175	10:44
Air Blank	0.000	10:45
Control Test	0.077	10:46
Air Blank	0.000	10:46
Diagnostics Check	OK	10:46

Data Download Date/Time: 10/25/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,DENVER,H
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WATSON,DENVER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 10/21/2023 22:00:24

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.051	21:55
Breath Volume	2.074	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.055	21:58
Breath Volume	2.550	21:58
Air Blank	0.000	21:59
Control Test	0.078	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 10/25/2023 15:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAMY,ALLIE,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BELLAMY,ALLIE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: WALBEY

DATE/TIME OF TEST: 10/22/2023 04:27:45

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.128	04:23
Breath Volume	2.535	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.125	04:26
Breath Volume	2.238	04:26
Air Blank	0.000	04:26
Control Test	0.078	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 10/25/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERT,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PEREZ,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2028
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 10/22/2023 22:51:12

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:47	
Control Test	0.080	22:47	
Air Blank	0.000	22:48	
Subject Sample	# 1	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:50	
Control Test	0.079	22:50	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time: 10/25/2023 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SIMMONS,JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:16

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 10/04/2023 09:54:32

Diagnostics Check	OK	09:48
Air Blank	0.000	09:49
Control Test	0.082	09:49
Air Blank	0.000	09:49
Subject Sample # 1	0.000	09:50
Breath Volume	3.539	09:50
Air Blank	0.000	09:50
Air Blank	0.000	09:52
Subject Sample # 2	0.000	09:53
Breath Volume	3.347	09:53
Air Blank	0.000	09:53
Control Test	0.082	09:53
Air Blank	0.000	09:54
Diagnostics Check	OK	09:54

Data Download Date/Time: 10/30/2023 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 10/07/2023 03:20:07

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.000	03:15
Breath Volume	1.273	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.000	03:18
Breath Volume	1.429	03:18
Air Blank	0.000	03:19
Control Test	0.084	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 10/30/2023 21:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PEDROTTI,BRYAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: MAYORGA

DATE/TIME OF TEST: 10/08/2023 02:48:59

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.082	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.130	02:42
Breath Volume	1.535	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.119	02:47
Breath Volume	1.492	02:47
Air Blank	0.000	02:47
Control Test	0.082	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 10/30/2023 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SIRIPHOKHA,HATSADY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:23

SUBJECT'S LAST NAME: CEDILLO TORRES

DATE/TIME OF TEST: 10/08/2023 11:03:51

Diagnostics Check	OK	10:52	
Air Blank	0.000	10:52	
Control Test	0.082	10:53	
Air Blank	0.000	10:53	
Subject Sample	# 1 0.208	10:56	Volume Not Met
Breath Volume	0.000	10:56	
Air Blank	0.000	10:57	
Air Blank	0.000	10:59	
Subject Sample	# 2	11:02	No Sample Provided
Breath Volume	0.000	11:02	
Air Blank	0.000	11:02	
Control Test	0.081	11:03	
Air Blank	0.000	11:03	
Diagnostics Check	OK	11:03	

Data Download Date/Time: 10/30/2023 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SIRIPHOKHA,HATSADY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:23

SUBJECT'S LAST NAME: CEDILLO TORRES

DATE/TIME OF TEST: 10/08/2023 11:11:32

Diagnostics Check	OK	11:08	
Air Blank	0.000	11:09	
Control Test	0.082	11:09	
Air Blank	0.000	11:09	
Subject Sample	# 1	11:10	Subject Test Refused
Breath Volume	0.000	11:10	
Air Blank	0.000	11:10	
Control Test	0.082	11:10	
Air Blank	0.000	11:11	
Diagnostics Check	OK	11:11	

Data Download Date/Time: 10/30/2023 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SIMMONS,JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: STRANGE

DATE/TIME OF TEST: 10/09/2023 09:39:51

Diagnostics Check	OK	09:33
Air Blank	0.000	09:34
Control Test	0.081	09:34
Air Blank	0.000	09:35
Subject Sample # 1	0.000	09:35
Breath Volume	4.433	09:35
Air Blank	0.000	09:36
Air Blank	0.000	09:38
Subject Sample # 2	0.000	09:38
Breath Volume	3.187	09:38
Air Blank	0.000	09:38
Control Test	0.082	09:39
Air Blank	0.000	09:39
Diagnostics Check	OK	09:39

Data Download Date/Time: 10/30/2023 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDENAS, ADAM,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A, N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 10/12/2023 10:57:48

Diagnostics Check	OK	10:51
Air Blank	0.000	10:52
Control Test	0.082	10:52
Air Blank	0.000	10:53
Subject Sample # 1	0.000	10:53
Breath Volume	2.011	10:53
Air Blank	0.000	10:53
Air Blank	0.000	10:55
Subject Sample # 2	0.000	10:56
Breath Volume	2.085	10:56
Air Blank	0.000	10:56
Control Test	0.082	10:57
Air Blank	0.000	10:57
Diagnostics Check	OK	10:57

Data Download Date/Time: 10/30/2023 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHMIDT,KATHLYNN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: MILLIGAN

DATE/TIME OF TEST: 10/12/2023 22:45:18

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.159	22:40
Breath Volume	1.488	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.162	22:43
Breath Volume	1.437	22:43
Air Blank	0.000	22:44
Control Test	0.083	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 10/30/2023 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: GARCIA-JAIMES

DATE/TIME OF TEST: 10/18/2023 19:33:58

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.080	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.207	19:29
Breath Volume	2.472	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:31
Subject Sample # 2	0.206	19:32
Breath Volume	2.550	19:32
Air Blank	0.000	19:33
Control Test	0.081	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 10/30/2023 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: COURT ORD
ARREST TIME: 12:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:24

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 10/21/2023 12:54:00

Diagnostics Check	OK	12:47	
Air Blank	0.000	12:48	
Control Test	0.082	12:48	
Air Blank	0.000	12:48	
Subject Sample	# 1 0.000	12:49	
Breath Volume	2.761	12:49	
Air Blank	0.000	12:50	
Air Blank	0.000	12:52	
Subject Sample	# 2 0.000	12:52	
Breath Volume	3.355	12:52	
Air Blank	0.000	12:52	
Control Test	0.080	12:53	
Air Blank		12:53	RFI Detect
Air Blank	0.000	12:54	

Data Download Date/Time: 10/30/2023 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JOHNSON,JC
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 10/22/2023 02:17:17

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.081	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1 0.266	02:06	
Breath Volume	1.675	02:06	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample	# 2	02:12	No Sample Provided
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Air Blank	0.000	02:15	
Subject Sample	# 3 0.261	02:15	
Breath Volume	1.867	02:15	
Air Blank	0.000	02:16	
Control Test	0.082	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 10/30/2023 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: EASTER

DATE/TIME OF TEST: 10/23/2023 03:13:55

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.079	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.082	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 10/30/2023 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLINGS, ALLEN, C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WEBSTER, CHRISTOPHER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: CASTRO MENCHU

DATE/TIME OF TEST: 10/24/2023 14:23:07

Diagnostics Check	OK	14:15
Air Blank	0.000	14:16
Control Test	0.080	14:16
Air Blank	0.000	14:17
Subject Sample # 1	0.433	14:17
Breath Volume	2.906	14:17
Air Blank	0.000	14:18
Air Blank	0.000	14:20
Subject Sample # 2	0.426	14:21
Breath Volume	1.777	14:21
Air Blank	0.000	14:22
Control Test	0.079	14:22
Air Blank	0.000	14:23
Diagnostics Check	OK	14:23

Data Download Date/Time: 10/30/2023 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAYA,ZACHARY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: COURT ORD
ARREST TIME: 12:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 10/26/2023 13:00:00

Diagnostics Check	OK	12:53
Air Blank	0.000	12:54
Control Test	0.082	12:54
Air Blank	0.000	12:55
Subject Sample # 1	0.000	12:55
Breath Volume	2.527	12:55
Air Blank	0.000	12:55
Air Blank	0.000	12:57
Subject Sample # 2	0.000	12:58
Breath Volume	2.511	12:58
Air Blank	0.000	12:59
Control Test	0.084	12:59
Air Blank	0.000	12:59
Diagnostics Check	OK	13:00

Data Download Date/Time: 10/30/2023 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAYA,ZACHARY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302415
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: COURT ORD
ARREST TIME: 13:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 10/30/2023 14:00:59

Diagnostics Check	OK	13:54
Air Blank	0.000	13:55
Control Test	0.081	13:55
Air Blank	0.000	13:56
Subject Sample # 1	0.000	13:56
Breath Volume	2.652	13:56
Air Blank	0.000	13:56
Air Blank	0.000	13:58
Subject Sample # 2	0.000	13:59
Breath Volume	2.734	13:59
Air Blank	0.000	14:00
Control Test	0.083	14:00
Air Blank	0.000	14:00
Diagnostics Check	OK	14:00

Data Download Date/Time: 10/30/2023 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,E
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:41

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/22/2023 08:08:27

Diagnostics Check	OK	08:02
Air Blank	0.000	08:02
Control Test	0.078	08:02
Air Blank	0.000	08:03
Subject Sample # 1	0.076	08:03
Breath Volume	1.601	08:03
Air Blank	0.000	08:04
Air Blank	0.000	08:06
Subject Sample # 2	0.074	08:06
Breath Volume	1.992	08:06
Air Blank	0.000	08:07
Control Test	0.078	08:07
Air Blank	0.000	08:08
Diagnostics Check	OK	08:08

Data Download Date/Time: 11/13/2023 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: MARTINEK

DATE/TIME OF TEST: 10/26/2023 20:05:02

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:51	
Control Test	0.078	19:52	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:55	No Sample Provided
Breath Volume	0.000	19:55	
Air Blank	0.000	19:56	
Air Blank	0.000	19:58	
Subject Sample	# 2	0.186	20:00
Breath Volume	1.527	20:00	
Air Blank	0.000	20:01	
Air Blank	0.000	20:02	
Subject Sample	# 3	0.184	20:03
Breath Volume	2.578	20:03	
Air Blank	0.000	20:04	
Control Test	0.076	20:04	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 11/13/2023 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MAVES,JEFFREY
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: CORREA CLAUDIO

DATE/TIME OF TEST: 10/27/2023 23:25:39

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.078	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.125	23:20
Breath Volume	1.617	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.126	23:24
Breath Volume	2.656	23:24
Air Blank	0.000	23:24
Control Test	0.078	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 11/13/2023 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COTNER,STEVEN,A

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COTNER,STEVEN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 01:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: HOLLERS

DATE/TIME OF TEST: 10/28/2023 02:22:48

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.121	02:17
Breath Volume	1.289	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.128	02:21
Breath Volume	1.636	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 11/28/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 07:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: SEWELL JR

DATE/TIME OF TEST: 10/28/2023 08:06:36

Diagnostics Check	OK	08:00
Air Blank	0.000	08:00
Control Test	0.080	08:00
Air Blank	0.000	08:01
Subject Sample # 1	0.195	08:02
Breath Volume	3.433	08:02
Air Blank	0.000	08:02
Air Blank	0.000	08:04
Subject Sample # 2	0.177	08:05
Breath Volume	2.078	08:05
Air Blank	0.000	08:05
Control Test	0.079	08:06
Air Blank	0.000	08:06
Diagnostics Check	OK	08:06

Data Download Date/Time: 11/28/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASWELL,AUSTIN,J

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: POTTS,MICHAEL

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 00:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 10/29/2023 01:15:04

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.039	01:10
Breath Volume	2.554	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.041	01:13
Breath Volume	3.714	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:15

Data Download Date/Time: 11/28/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLLINS,CHASE,M

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COLLINS,CHASE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 16:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: MARACLE

DATE/TIME OF TEST: 10/30/2023 17:00:01

Diagnostics Check	OK	16:48	
Air Blank	0.000	16:49	
Control Test	0.080	16:49	
Air Blank	0.000	16:49	
Subject Sample	# 1	16:53	No Sample Provided
Breath Volume	0.000	16:53	
Air Blank	0.000	16:53	
Air Blank	0.000	16:55	
Subject Sample	# 2	16:58	No Sample Provided
Breath Volume	0.000	16:58	
Air Blank	0.000	16:59	
Control Test	0.082	16:59	
Air Blank	0.000	16:59	
Diagnostics Check	OK	17:00	

Data Download Date/Time: 11/28/2023 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: HAYDEN

DATE/TIME OF TEST: 10/28/2023 03:04:50

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.135	03:00
Breath Volume	2.351	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.126	03:03
Breath Volume	2.140	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 11/01/2023 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: SCHEEL

DATE/TIME OF TEST: 10/28/2023 21:48:18

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:43	
Air Blank	0.000	21:44	
Subject Sample	# 1 0.169	21:44	
Breath Volume	1.410	21:44	
Air Blank	0.000	21:45	
Air Blank	0.000	21:47	
Subject Sample	# 2	21:47	RFI Detect
Breath Volume	0.328	21:47	
Air Blank		21:47	RFI Detect
Air Blank	0.000	21:48	

Data Download Date/Time: 11/01/2023 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 21:13

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: SCHEEL

DATE/TIME OF TEST: 10/28/2023 21:55:31

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.184	21:51
Breath Volume	1.632	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.175	21:54
Breath Volume	1.714	21:54
Air Blank	0.000	21:54
Control Test	0.078	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 11/01/2023 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/29/2023 01:05:17

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.159	01:00
Breath Volume	1.863	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.150	01:03
Breath Volume	2.199	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 11/01/2023 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: GREEN,J,A	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HINES,M	ARREST TIME: 02:33
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/18/2023	OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HOVERSEN	DATE/TIME OF TEST: 10/29/2023	03:01:36
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.147	02:57
Breath Volume	1.851	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.154	03:00
Breath Volume	2.109	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 11/01/2023 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: LOPEZ MARTA

DATE/TIME OF TEST: 10/01/2023 01:43:38

Diagnostics Check	Fail	01:43	Internal Printer Test
Air Blank	0.000	01:43	

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: LOPEZ MARTA

DATE/TIME OF TEST: 10/01/2023 01:51:57

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.078	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.187	01:47
Breath Volume	1.687	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.198	01:50
Breath Volume	1.871	01:50
Air Blank	0.000	01:50
Control Test	0.076	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: BEST

DATE/TIME OF TEST: 10/02/2023 22:31:16

Diagnostics Check	Fail	22:30	Internal Printer Test
Air Blank	0.000	22:31	

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: BEST

DATE/TIME OF TEST: 10/02/2023 22:40:17

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.079	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.095	22:35
Breath Volume	2.359	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.102	22:38
Breath Volume	2.437	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: HOWARD,J,G	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HOWARD,J	ARREST TIME: 00:02
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/18/2023	OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: MCDANIEL	DATE/TIME OF TEST: 10/04/2023	00:17:53
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.078	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.156	00:13
Breath Volume	2.718	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:15
Subject Sample # 2	0.163	00:16
Breath Volume	3.757	00:16
Air Blank	0.000	00:16
Control Test	0.077	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: HOWARD,J,G	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HOWARD,J	ARREST TIME: 18:58
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/18/2023	OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: FROUMY	DATE/TIME OF TEST: 10/04/2023	19:10:39
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	19:03
Air Blank	0.000	19:04
Control Test	0.077	19:04
Air Blank	0.000	19:05
Subject Sample # 1	0.173	19:06
Breath Volume	2.214	19:06
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.169	19:09
Breath Volume	1.976	19:09
Air Blank	0.000	19:09
Control Test	0.077	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: WHIDDON

DATE/TIME OF TEST: 10/04/2023 21:59:50

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:53	
Control Test	0.078	21:53	
Air Blank	0.000	21:54	
Subject Sample	# 1 0.035	21:54	
Breath Volume	2.457	21:54	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2 0.035	21:57	
Breath Volume	3.429	21:57	
Air Blank	0.000	21:58	
Control Test	0.077	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	Fail	21:59	Internal Printer Test
Air Blank	0.000	21:59	

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: HOWARD,J,G	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HOWARD,J	ARREST TIME: 14:58
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/18/2023	OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: BOWLBY	DATE/TIME OF TEST: 10/16/2023	16:38:28
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	16:32
Air Blank	0.000	16:32
Control Test	0.080	16:33
Air Blank	0.000	16:33
Subject Sample # 1	0.000	16:34
Breath Volume	1.632	16:34
Air Blank	0.000	16:34
Air Blank	0.000	16:36
Subject Sample # 2	0.000	16:37
Breath Volume	2.460	16:37
Air Blank	0.000	16:37
Control Test	0.077	16:37
Air Blank	0.000	16:38
Diagnostics Check	OK	16:38

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 10/16/2023 22:13:16

Diagnostics Check	Fail	22:12	Internal Printer Test
Air Blank	0.000	22:13	

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 10/16/2023 22:24:25

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:15	
Control Test	0.080	22:15	
Air Blank	0.000	22:15	
Subject Sample	# 1 0.221	22:17	
Breath Volume	1.382	22:17	
Air Blank	0.000	22:18	
Air Blank	0.000	22:19	
Subject Sample	# 2 0.225	22:22	
Breath Volume	2.359	22:22	
Air Blank	0.000	22:22	
Control Test	0.080	22:23	
Air Blank	0.000	22:23	
Diagnostics Check	Fail	22:23	Internal Printer Test
Air Blank	0.000	22:24	

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, HOWARD, G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD, J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 10/16/2023 22:26:45

Diagnostics Check	Fail	22:26	Internal Printer Test
Air Blank	0.000	22:26	

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: HOWARD,J,G	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HOWARD,J	ARREST TIME: 22:08
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Female
LAST AGENCY INSPECTION: 09/18/2023	OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GRIFFIN	DATE/TIME OF TEST: 10/16/2023	22:42:38
------------------------------	-------------------------------	----------

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:29	
Control Test	0.079	22:29	
Air Blank	0.000	22:29	
Subject Sample	# 1 0.202	22:33	Slope Not Level
Breath Volume	1.656	22:33	
Air Blank	0.000	22:33	
Air Blank	0.000	22:35	
Subject Sample	# 2 0.212	22:36	
Breath Volume	1.753	22:36	
Air Blank	0.000	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 3 0.207	22:40	
Breath Volume	2.484	22:40	
Air Blank	0.000	22:41	
Control Test	0.075	22:42	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MORAN,S
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 10/17/2023 20:51:23

Diagnostics Check	Fail	20:50	Internal Printer Test
Air Blank	0.000	20:51	

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MORAN,S
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 10/17/2023 20:59:26

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.080	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.285	20:54
Breath Volume	4.515	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.265	20:57
Breath Volume	3.714	20:57
Air Blank	0.000	20:58
Control Test	0.077	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCCANDLESS,MATTHEW	ARREST TIME: 23:45
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Female
LAST AGENCY INSPECTION: 09/18/2023	OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: BEEBE	DATE/TIME OF TEST: 10/18/2023	00:30:54
----------------------------	-------------------------------	----------

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.080	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.119	00:26
Breath Volume	1.675	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.112	00:29
Breath Volume	1.433	00:29
Air Blank	0.000	00:29
Control Test	0.077	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: HOWARD,J,G	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GREEN,L	ARREST TIME: 18:54
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/18/2023	OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BAIRD	DATE/TIME OF TEST: 10/18/2023	19:37:18
----------------------------	-------------------------------	----------

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.244	19:32
Breath Volume	2.015	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:34
Subject Sample # 2	0.241	19:35
Breath Volume	1.843	19:35
Air Blank	0.000	19:36
Control Test	0.078	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: HOWARD,J,G	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HOWARD,J	ARREST TIME: 00:49
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/18/2023	OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: ASSELTA	DATE/TIME OF TEST: 10/19/2023	01:00:59
------------------------------	-------------------------------	----------

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.000	00:56
Breath Volume	2.734	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.000	00:59
Breath Volume	3.179	00:59
Air Blank	0.000	01:00
Control Test	0.077	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: HOWARD,J,G	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HOWARD,J	ARREST TIME: 17:45
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/18/2023	OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: JARQUIN MARTINEZ	DATE/TIME OF TEST: 10/22/2023	17:56:45
---------------------------------------	-------------------------------	----------

Diagnostics Check	OK	17:47
Air Blank	0.000	17:48
Control Test	0.079	17:48
Air Blank	0.000	17:48
Subject Sample # 1	0.200	17:50
Breath Volume	1.292	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.211	17:54
Breath Volume	2.050	17:54
Air Blank	0.000	17:55
Control Test	0.076	17:56
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211543
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 10/22/2023 19:35:49

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.078	19:30
Air Blank	0.000	19:30
Subject Sample # 1	0.102	19:31
Breath Volume	2.480	19:31
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.104	19:34
Breath Volume	3.468	19:34
Air Blank	0.000	19:34
Control Test	0.077	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 211543
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 01/05/2024
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCCANDLESS,MATTHEW	ARREST TIME: 21:37
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 09/18/2023	OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: STROUP	DATE/TIME OF TEST: 10/22/2023	21:50:40
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.078	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.000	21:46
Breath Volume	2.128	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.000	21:49
Breath Volume	2.468	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 11/01/2023 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCORE,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: SINCORE,MICHAEL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 10:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: CASTRO ARIAS

DATE/TIME OF TEST: 10/01/2023 11:38:44

Diagnostics Check	OK	11:32
Air Blank	0.000	11:32
Control Test	0.080	11:32
Air Blank	0.000	11:33
Subject Sample # 1	0.201	11:33
Breath Volume	1.609	11:33
Air Blank	0.000	11:34
Air Blank	0.000	11:36
Subject Sample # 2	0.207	11:36
Breath Volume	1.519	11:36
Air Blank	0.000	11:37
Control Test	0.079	11:38
Air Blank	0.000	11:38
Diagnostics Check	OK	11:38

Data Download Date/Time: 10/31/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILPIN, SCHELISA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GILPIN, S
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 10/04/2023 22:31:28

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.081	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Control Test	0.081	22:30	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 10/31/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ARREAGA,ALDO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: MENDEZ DIAZ

DATE/TIME OF TEST: 10/07/2023 20:08:17

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.081	19:59	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:03	No Sample Provided
Breath Volume	0.000	20:03	
Air Blank	0.000	20:04	
Air Blank	0.000	20:06	
Subject Sample	# 2	20:06	Subject Test Refused
Breath Volume	0.000	20:06	
Air Blank	0.000	20:07	
Control Test	0.081	20:07	
Air Blank	0.000	20:08	
Diagnostics Check	OK	20:08	

Data Download Date/Time: 10/31/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40235711
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: PEREZ GUTIERREZ

DATE/TIME OF TEST: 10/28/2023 19:39:16

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.081	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.233	19:34
Breath Volume	1.875	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.241	19:37
Breath Volume	1.449	19:37
Air Blank	0.000	19:38
Control Test	0.080	19:38
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 10/31/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INNOCENT,L,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: AGUILAR,W
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MARTINEZ GUTIERREZ

DATE/TIME OF TEST: 10/31/2023 00:53:49

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.081	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:52	Subject Test Refused
Breath Volume	0.000	00:52	
Air Blank	0.000	00:52	
Control Test	0.081	00:53	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 10/31/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: RODRIGUES SANTOS

DATE/TIME OF TEST: 10/01/2023 04:13:15

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.007	04:08
Breath Volume	2.035	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.007	04:11
Breath Volume	1.296	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 10/04/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANTONINI,DANIEL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 13:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/01/2023 15:41:37

Diagnostics Check	OK	15:35
Air Blank	0.000	15:35
Control Test	0.081	15:35
Air Blank	0.000	15:36
Subject Sample # 1	0.152	15:36
Breath Volume	3.464	15:36
Air Blank	0.000	15:37
Air Blank	0.000	15:39
Subject Sample # 2	0.154	15:39
Breath Volume	3.421	15:39
Air Blank	0.000	15:40
Control Test	0.080	15:40
Air Blank	0.000	15:41
Diagnostics Check	OK	15:41

Data Download Date/Time: 10/04/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: DEDECKER

DATE/TIME OF TEST: 10/02/2023 00:18:22

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.081	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.229	00:13
Breath Volume	1.339	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:15
Subject Sample # 2	0.229	00:16
Breath Volume	1.453	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 10/04/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 09/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/02/2023 01:17:18

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.161	01:12
Breath Volume	1.539	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:14
Subject Sample # 2	0.159	01:15
Breath Volume	1.687	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 10/04/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: PETRONE

DATE/TIME OF TEST: 10/27/2023 17:58:09

Diagnostics Check	OK	17:51
Air Blank	0.000	17:51
Control Test	0.081	17:51
Air Blank	0.000	17:52
Subject Sample # 1	0.242	17:53
Breath Volume	1.121	17:53
Air Blank	0.000	17:54
Air Blank	0.000	17:55
Subject Sample # 2	0.242	17:56
Breath Volume	1.792	17:56
Air Blank	0.000	17:57
Control Test	0.080	17:57
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 12/03/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GUNN

DATE/TIME OF TEST: 10/28/2023 01:32:31

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.082	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.237	01:27
Breath Volume	2.406	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.222	01:30
Breath Volume	1.761	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 12/03/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ORTIZ RIOS

DATE/TIME OF TEST: 10/28/2023 05:20:44

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.077	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.164	05:16
Breath Volume	2.031	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.161	05:18
Breath Volume	1.683	05:18
Air Blank	0.000	05:19
Control Test	0.075	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 11/20/2023 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 10/29/2023 03:06:38

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.077	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.178	03:02
Breath Volume	1.812	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.188	03:04
Breath Volume	1.445	03:04
Air Blank	0.000	03:05
Control Test	0.075	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 11/20/2023 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: DORADO MONTERO

DATE/TIME OF TEST: 10/08/2023 23:49:09

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.082	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.116	23:44
Breath Volume	2.218	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.115	23:47
Breath Volume	3.097	23:47
Air Blank	0.000	23:48
Control Test	0.082	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 11/08/2023 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOI, HYUNSOO,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI, HYUNSOO
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/01/2023 02:52:35

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.077	02:49	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.078	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 10/02/2023 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SANCHEZ MATOS

DATE/TIME OF TEST: 10/01/2023 04:23:27

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.077	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.226	04:18
Breath Volume	2.144	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.218	04:21
Breath Volume	1.769	04:21
Air Blank	0.000	04:22
Control Test	0.075	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 10/02/2023 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MARTINEZ VALDES

DATE/TIME OF TEST: 10/02/2023 00:16:13

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.076	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.218	00:11
Breath Volume	2.582	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample # 2	0.214	00:14
Breath Volume	2.085	00:14
Air Blank	0.000	00:15
Control Test	0.076	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 10/02/2023 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: WOFFORD

DATE/TIME OF TEST: 10/09/2023 19:41:31

Diagnostics Check	OK	19:34
Air Blank	0.000	19:35
Control Test	0.076	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.089	19:36
Breath Volume	2.707	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.088	19:39
Breath Volume	2.945	19:39
Air Blank	0.000	19:40
Control Test	0.079	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 12/13/2023 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SUTTON JR

DATE/TIME OF TEST: 10/15/2023 04:46:34

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.076	04:39	
Air Blank	0.000	04:40	
Subject Sample # 1	0.181	04:41	
Breath Volume	2.101	04:41	
Air Blank	0.000	04:42	
Air Blank	0.000	04:43	
Subject Sample # 2	0.170	04:44	
Breath Volume	1.863	04:44	
Air Blank	0.000	04:45	
Control Test	0.073	04:45	Control Outside Tolerance
Air Blank	0.000	04:46	

Data Download Date/Time: 12/13/2023 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SUTTON JR

DATE/TIME OF TEST: 10/15/2023 04:57:05

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.077	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.176	04:51
Breath Volume	1.734	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.170	04:55
Breath Volume	1.863	04:55
Air Blank	0.000	04:56
Control Test	0.076	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 12/13/2023 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOI, HYUNSOO,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI, HYUNSOO
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: DE LA CRUZ PEREZ

DATE/TIME OF TEST: 10/19/2023 22:23:54

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.076	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.184	22:18
Breath Volume	1.593	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.204	22:22
Breath Volume	1.468	22:22
Air Blank	0.000	22:22
Control Test	0.078	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 12/13/2023 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:46

SUBJECT'S LAST NAME: SLONE

DATE/TIME OF TEST: 10/05/2023 11:10:30

Diagnostics Check	OK	11:07	
Air Blank	0.000	11:07	
Control Test	0.077	11:07	
Air Blank	0.000	11:08	
Subject Sample	# 1	11:09	Subject Test Refused
Breath Volume	0.000	11:09	
Air Blank	0.000	11:09	
Control Test	0.078	11:09	
Air Blank	0.000	11:10	
Diagnostics Check	OK	11:10	

Data Download Date/Time: 11/08/2023 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBE,BRIAN,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: NEILY,STEVEN
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: BRIONES

DATE/TIME OF TEST: 10/01/2023 00:25:31

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.108	00:20
Breath Volume	1.785	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.102	00:23
Breath Volume	1.742	00:23
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 10/24/2023 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBE,BRIAN,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: CRISERA,MATTHEW
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DESOYE

DATE/TIME OF TEST: 10/01/2023 02:17:15

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.226	02:12
Breath Volume	1.914	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.219	02:15
Breath Volume	1.394	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 10/24/2023 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESLEY,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: HYLTON,ROBERT
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/11/2023 00:03:10

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.079	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.211	23:54	Volume Not Met
Breath Volume	1.093	23:54	
Air Blank	0.000	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 3 0.223	23:57	
Breath Volume	2.343	23:57	
Air Blank	0.000	23:58	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.230	00:01	
Breath Volume	2.457	00:01	
Air Blank	0.000	00:02	
Control Test	0.076	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 10/24/2023 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESLEY,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: CRISERA,MATTHEW
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 10/15/2023 00:52:36

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.208	00:48
Breath Volume	1.843	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.197	00:50
Breath Volume	1.785	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 10/24/2023 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESLEY, JONATHAN, R
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: NEILY, STEVEN
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 10/20/2023 01:15:34

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.165	01:10
Breath Volume	1.824	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.176	01:13
Breath Volume	1.578	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 10/24/2023 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFFMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: NEILY,STEVEN
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SCHADENFROH

DATE/TIME OF TEST: 10/25/2023 01:49:43

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.191	01:45
Breath Volume	1.207	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.203	01:48
Breath Volume	2.261	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 11/02/2023 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARROQUIN,RICHARD,J

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: RIVARD,MCKENZIE

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG222203

DRY GAS EXPIRY DATE: 08/10/2024

VIOLATION CODE: BUI

ARREST TIME: 17:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 10/15/2023 18:47:37

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.081	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.141	18:42
Breath Volume	3.500	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.151	18:46
Breath Volume	4.941	18:46
Air Blank	0.000	18:46
Control Test	0.082	18:47
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 10/31/2023 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,W,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN,W
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 10:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: GOPHER

DATE/TIME OF TEST: 10/02/2023 11:19:03

Diagnostics Check	OK	11:12
Air Blank	0.000	11:12
Control Test	0.081	11:13
Air Blank	0.000	11:13
Subject Sample # 1	0.000	11:14
Breath Volume	2.234	11:14
Air Blank	0.000	11:14
Air Blank	0.000	11:16
Subject Sample # 2	0.000	11:17
Breath Volume	2.226	11:17
Air Blank	0.000	11:18
Control Test	0.081	11:18
Air Blank	0.000	11:18
Diagnostics Check	OK	11:19

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/05/2023 02:37:58

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.000	02:33
Breath Volume	1.757	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.000	02:36
Breath Volume	4.730	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,A,K
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,R
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: MERKEY

DATE/TIME OF TEST: 10/07/2023 12:07:39

Diagnostics Check	OK	12:00
Air Blank	0.000	12:01
Control Test	0.081	12:01
Air Blank	0.000	12:01
Subject Sample # 1	0.195	12:02
Breath Volume	2.203	12:02
Air Blank	0.000	12:03
Air Blank	0.000	12:04
Subject Sample # 2	0.195	12:05
Breath Volume	1.636	12:05
Air Blank	0.000	12:06
Control Test	0.078	12:07
Air Blank	0.000	12:07
Diagnostics Check	OK	12:07

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RICHARDSON,S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: SOOY

DATE/TIME OF TEST: 10/07/2023 20:49:14

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:44
Subject Sample # 1	0.097	20:44
Breath Volume	1.636	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.095	20:47
Breath Volume	1.976	20:47
Air Blank	0.000	20:48
Control Test	0.081	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: COTTO,G
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 10/11/2023 02:02:44

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.080	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WALKER, DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 10/11/2023 06:34:02

Diagnostics Check	OK	06:30	
Air Blank	0.000	06:31	
Control Test	0.080	06:31	
Air Blank	0.000	06:32	
Subject Sample	# 1	06:32	Subject Test Refused
Breath Volume	0.000	06:32	
Air Blank	0.000	06:33	
Control Test	0.080	06:33	
Air Blank	0.000	06:33	
Diagnostics Check	OK	06:34	

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HODGE,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/11/2023 21:49:58

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:47	
Control Test	0.081	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Control Test	0.081	21:49	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAY,JOEL,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FAY,JOEL
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:03

SUBJECT'S LAST NAME: FERRETTI

DATE/TIME OF TEST: 10/14/2023 13:29:19

Diagnostics Check	OK	13:23
Air Blank	0.000	13:23
Control Test	0.080	13:24
Air Blank	0.000	13:24
Subject Sample # 1	0.000	13:24
Breath Volume	3.859	13:24
Air Blank	0.000	13:25
Air Blank	0.000	13:27
Subject Sample # 2	0.000	13:27
Breath Volume	4.531	13:27
Air Blank	0.000	13:28
Control Test	0.080	13:28
Air Blank	0.000	13:29
Diagnostics Check	OK	13:29

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RIVERA,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 10/18/2023 05:04:29

Diagnostics Check	OK	05:01	
Air Blank	0.000	05:01	
Control Test	0.082	05:02	
Air Blank	0.000	05:02	
Subject Sample	# 1	05:03	Subject Test Refused
Breath Volume	0.000	05:03	
Air Blank	0.000	05:03	
Control Test	0.081	05:03	
Air Blank	0.000	05:04	
Diagnostics Check	OK	05:04	

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,R,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECCERA,J
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: FLESHER

DATE/TIME OF TEST: 10/22/2023 02:37:52

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.081	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:36	
Control Test	0.081	02:37	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARTINEZ,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 10/24/2023 18:13:43

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:10	
Control Test	0.081	18:11	
Air Blank	0.000	18:11	
Subject Sample	# 1	18:12	Subject Test Refused
Breath Volume	0.000	18:12	
Air Blank	0.000	18:12	
Control Test	0.081	18:13	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GORMAN,H
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: RACKI

DATE/TIME OF TEST: 10/26/2023 23:29:23

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.081	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:28	
Control Test	0.082	23:28	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 10/27/2023 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: COTTO,G
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DELEON-MACARIO

DATE/TIME OF TEST: 10/28/2023 02:45:12

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.081	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.081	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 11/29/2023 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIZARDO,RAILY,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: GARZA,DAVID
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HERNANDEZ-GOMEZ

DATE/TIME OF TEST: 10/07/2023 04:15:49

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.147	04:11
Breath Volume	1.226	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.151	04:14
Breath Volume	1.359	04:14
Air Blank	0.000	04:14
Control Test	0.078	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 10/30/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAQUETTE,AARON,Y
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: THUREK,MAX
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MAGANA

DATE/TIME OF TEST: 10/15/2023 02:48:22

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.079	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 10/30/2023 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFRESNE,BRANDY,M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: RAY,DYLAN
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 10/19/2023 00:10:11

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.080	23:55	
Air Blank	0.000	23:55	
Subject Sample	# 3 0.000	23:58	Volume Not Met
Breath Volume	0.214	23:58	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.000	00:04	
Breath Volume	1.125	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2 0.000	00:08	
Breath Volume	1.175	00:08	
Air Blank	0.000	00:09	
Control Test	0.079	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 10/30/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAQUETTE,AARON,Y
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: GAMMAD,GARY
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GONZALEZREYES

DATE/TIME OF TEST: 10/20/2023 02:20:24

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.243	02:15
Breath Volume	2.394	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.248	02:18
Breath Volume	2.242	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 10/30/2023 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNING,SEAN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: LEATIOTA,JIMMY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: KAVURI

DATE/TIME OF TEST: 10/01/2023 01:05:49

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.157	01:01
Breath Volume	2.593	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:03
Subject Sample # 2	0.154	01:04
Breath Volume	2.632	01:04
Air Blank	0.000	01:04
Control Test	0.076	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNING,SEAN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,JOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/01/2023 03:14:20

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.142	03:09
Breath Volume	2.707	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.136	03:12
Breath Volume	2.261	03:12
Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNING,SEAN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CANNING,SEAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MARINO-KOHNEN

DATE/TIME OF TEST: 10/02/2023 04:12:12

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.079	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1 0.141	04:01	Volume Not Met
Breath Volume	1.085	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 2 0.138	04:06	
Breath Volume	1.230	04:06	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 3 0.134	04:10	
Breath Volume	1.253	04:10	
Air Blank	0.000	04:11	
Control Test	0.076	04:11	
Air Blank		04:11	RFI Detect
Air Blank	0.000	04:12	

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFFORD,JOSEPH,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 12:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: KEATING

DATE/TIME OF TEST: 10/04/2023 13:14:16

Diagnostics Check	OK	13:07
Air Blank	0.000	13:08
Control Test	0.078	13:08
Air Blank	0.000	13:08
Subject Sample # 1	0.322	13:09
Breath Volume	3.167	13:09
Air Blank	0.000	13:10
Air Blank	0.000	13:11
Subject Sample # 2	0.333	13:12
Breath Volume	1.574	13:12
Air Blank	0.000	13:13
Control Test	0.076	13:13
Air Blank	0.000	13:14
Diagnostics Check	OK	13:14

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 10/06/2023 00:01:51

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.079	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 2 0.045	23:50	
Breath Volume	2.042	23:50	
Air Blank	0.000	23:51	
Air Blank	0.000	23:53	
Subject Sample	# 3 0.036	23:56	Volume Not Met
Breath Volume	0.988	23:56	
Air Blank	0.000	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.045	00:00	
Breath Volume	2.320	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LOGAN,MARCUS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 10/06/2023 03:38:25

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.134	03:32
Breath Volume	1.875	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.127	03:36
Breath Volume	1.566	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 10/06/2023 19:56:32

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:55	
Control Test	0.000	19:56	RFI Detect
Air Blank		19:56	RFI Detect

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 10/06/2023 20:10:49

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.079	19:59	
Air Blank	0.000	20:00	
Subject Sample	# 1 0.000	20:03	Volume Not Met
Breath Volume	0.000	20:03	
Air Blank	0.000	20:04	
Air Blank	0.000	20:06	
Subject Sample	# 2 0.000	20:09	Volume Not Met
Breath Volume	0.000	20:09	
Air Blank	0.000	20:09	
Control Test	0.078	20:10	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 10/07/2023 01:16:22

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.079	01:11
Breath Volume	3.082	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.078	01:14
Breath Volume	2.722	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/07/2023 03:44:48

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.152	03:40
Breath Volume	1.425	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.160	03:43
Breath Volume	2.429	03:43
Air Blank	0.000	03:43
Control Test	0.078	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MCGUIRE,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: GALEAS SANCHEZ

DATE/TIME OF TEST: 10/07/2023 06:16:42

Diagnostics Check	OK	06:10
Air Blank	0.000	06:10
Control Test	0.079	06:11
Air Blank	0.000	06:11
Subject Sample # 1	0.201	06:12
Breath Volume	3.195	06:12
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.197	06:15
Breath Volume	2.851	06:15
Air Blank	0.000	06:15
Control Test	0.078	06:16
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/08/2023 00:23:35

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.127	00:18
Breath Volume	2.121	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.131	00:21
Breath Volume	3.402	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 07:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: GEDEON

DATE/TIME OF TEST: 10/08/2023 08:46:56

Diagnostics Check	OK	08:40
Air Blank	0.000	08:41
Control Test	0.078	08:41
Air Blank	0.000	08:41
Subject Sample # 1	0.199	08:42
Breath Volume	1.574	08:42
Air Blank	0.000	08:43
Air Blank	0.000	08:44
Subject Sample # 2	0.201	08:45
Breath Volume	1.667	08:45
Air Blank	0.000	08:46
Control Test	0.078	08:46
Air Blank	0.000	08:46
Diagnostics Check	OK	08:46

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMANS,MORNE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: PARKER,TAYLOR
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 10/08/2023 20:16:59

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.294	20:12
Breath Volume	1.652	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:14
Subject Sample # 2	0.295	20:15
Breath Volume	2.101	20:15
Air Blank	0.000	20:16
Control Test	0.078	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RICHARDSON,NATHAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 10/08/2023 23:50:07

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.164	23:45
Breath Volume	2.097	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.167	23:48
Breath Volume	1.585	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNING,SEAN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MANDILE,JEREMY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: SOUD

DATE/TIME OF TEST: 10/09/2023 22:09:05

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:55	
Control Test	0.079	21:55	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:59	No Sample Provided
Breath Volume	0.000	21:59	
Air Blank	0.000	21:59	
Air Blank	0.000	22:01	
Subject Sample	# 2	0.049	22:04
Breath Volume	1.625	22:04	
Air Blank	0.000	22:05	
Air Blank	0.000	22:07	
Subject Sample	# 3	0.054	22:07
Breath Volume	2.136	22:07	
Air Blank	0.000	22:08	
Control Test	0.079	22:08	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: ARDINGER

DATE/TIME OF TEST: 10/11/2023 02:48:04

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.057	02:43
Breath Volume	1.175	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.060	02:46
Breath Volume	1.195	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:48

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: HINES,MIKO
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: FULWIDER

DATE/TIME OF TEST: 10/13/2023 18:28:08

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.079	18:22
Air Blank	0.000	18:22
Subject Sample # 1	0.133	18:23
Breath Volume	1.800	18:23
Air Blank	0.000	18:24
Air Blank	0.000	18:25
Subject Sample # 2	0.131	18:26
Breath Volume	1.800	18:26
Air Blank	0.000	18:27
Control Test	0.078	18:27
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRONSON,TAYLOR
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GLICA

DATE/TIME OF TEST: 10/13/2023 23:01:34

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.000	22:57
Breath Volume	1.882	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.000	23:00
Breath Volume	1.539	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGOLD,K,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RICHWINE,C
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ZISI

DATE/TIME OF TEST: 10/19/2023 03:19:00

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.203	03:14
Breath Volume	2.355	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:16
Subject Sample # 2	0.208	03:17
Breath Volume	2.316	03:17
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:19

Data Download Date/Time: 11/01/2023 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/19/2023 23:52:59

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.190	23:48
Breath Volume	2.191	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.189	23:51
Breath Volume	1.863	23:51
Air Blank	0.000	23:52
Control Test	0.077	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGOLD,K,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RICHWINE,C
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DE LA CRUZ DIAZ

DATE/TIME OF TEST: 10/20/2023 02:48:00

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.124	02:43
Breath Volume	2.359	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.130	02:46
Breath Volume	1.656	02:46
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:48

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: LOGGINS

DATE/TIME OF TEST: 10/20/2023 03:02:16

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.000	02:57
Breath Volume	3.671	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.000	03:00
Breath Volume	2.652	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: UTLEY,BRETT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 10/20/2023 22:19:03

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.079	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.026	22:14
Breath Volume	1.949	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:16
Subject Sample # 2	0.028	22:17
Breath Volume	2.058	22:17
Air Blank	0.000	22:18
Control Test	0.079	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:19

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: MAGRE

DATE/TIME OF TEST: 10/21/2023 00:48:34

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.183	00:43
Breath Volume	2.015	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.189	00:46
Breath Volume	2.140	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: ABINET

DATE/TIME OF TEST: 10/21/2023 00:58:49

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.078	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.120	00:54
Breath Volume	1.250	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.121	00:57
Breath Volume	1.761	00:57
Air Blank	0.000	00:57
Control Test	0.079	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 10/22/2023 02:06:10

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.177	02:01
Breath Volume	1.949	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.177	02:04
Breath Volume	2.406	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: MORLEY

DATE/TIME OF TEST: 10/22/2023 04:09:10

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.079	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Control Test	0.079	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: HAILE

DATE/TIME OF TEST: 10/22/2023 04:39:12

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.121	04:34
Breath Volume	1.968	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.121	04:37
Breath Volume	1.980	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MYERS,BENJAMIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: MCLEAN

DATE/TIME OF TEST: 10/22/2023 21:05:36

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.080	21:00
Air Blank	0.000	21:00
Subject Sample # 1	0.182	21:01
Breath Volume	2.332	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.179	21:03
Breath Volume	1.781	21:03
Air Blank	0.000	21:04
Control Test	0.076	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: TREIBER

DATE/TIME OF TEST: 10/23/2023 02:51:43

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.204	02:47
Breath Volume	2.519	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.201	02:50
Breath Volume	2.359	02:50
Air Blank	0.000	02:50
Control Test	0.078	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYERS,SPENCER,O
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: AYERS,S
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/23/2023 07:54:46

Diagnostics Check	OK	07:54	
Air Blank		07:54	RFI Detect
Air Blank	0.000	07:54	

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYERS,SPENCER,O
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: AYERS,S
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/23/2023 08:03:29

Diagnostics Check	OK	07:56
Air Blank	0.000	07:57
Control Test	0.078	07:57
Air Blank	0.000	07:57
Subject Sample # 1	0.314	07:58
Breath Volume	1.406	07:58
Air Blank	0.000	07:59
Air Blank	0.000	08:01
Subject Sample # 2	0.334	08:01
Breath Volume	2.355	08:01
Air Blank	0.000	08:02
Control Test	0.075	08:02
Air Blank	0.000	08:03
Diagnostics Check	OK	08:03

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,CHRIS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 10/24/2023 19:40:27

Diagnostics Check	OK	19:26	
Air Blank	0.000	19:26	
Control Test	0.079	19:27	
Air Blank	0.000	19:27	
Subject Sample	# 1 0.150	19:30	Slope Not Level
Breath Volume	1.406	19:30	
Air Blank	0.000	19:31	
Air Blank	0.000	19:33	
Subject Sample	# 2 0.158	19:35	
Breath Volume	1.156	19:35	
Air Blank	0.000	19:35	
Air Blank	0.000	19:37	
Subject Sample	# 3 0.170	19:38	
Breath Volume	1.218	19:38	
Air Blank	0.000	19:39	
Control Test	0.077	19:39	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MOYERS,MALLORY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: NARDUCCI

DATE/TIME OF TEST: 10/24/2023 21:02:36

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:54	
Control Test	0.079	20:54	
Air Blank	0.000	20:54	
Subject Sample	# 1 0.054	20:58	Volume Not Met
Breath Volume	0.230	20:58	
Air Blank	0.000	20:58	
Air Blank	0.000	21:00	
Subject Sample	# 2	21:01	RFI Detect
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Control Test	0.078	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MOYERS,MALLORY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: NARDUCCI

DATE/TIME OF TEST: 10/24/2023 21:20:03

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.079	21:05	
Air Blank	0.000	21:05	
Subject Sample	# 1 0.065	21:09	Volume Not Met
Breath Volume	0.707	21:09	
Air Blank	0.000	21:09	
Air Blank	0.000	21:11	
Subject Sample	# 2 0.063	21:12	
Breath Volume	1.300	21:12	
Air Blank	0.000	21:13	
Air Blank	0.000	21:15	
Subject Sample	# 3 0.062	21:18	
Breath Volume	1.277	21:18	
Air Blank	0.000	21:19	
Control Test	0.079	21:19	
Air Blank	0.000	21:19	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: CRUZ MALDONADO

DATE/TIME OF TEST: 10/26/2023 13:33:36

Diagnostics Check	OK	13:27
Air Blank	0.000	13:27
Control Test	0.079	13:28
Air Blank	0.000	13:28
Subject Sample # 1	0.000	13:29
Breath Volume	2.246	13:29
Air Blank	0.000	13:29
Air Blank	0.000	13:31
Subject Sample # 2	0.000	13:32
Breath Volume	2.070	13:32
Air Blank	0.000	13:32
Control Test	0.079	13:33
Air Blank	0.000	13:33
Diagnostics Check	OK	13:33

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 10:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:51

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 10/26/2023 23:12:06

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.079	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:07
Breath Volume	2.941	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.000	23:10
Breath Volume	2.027	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 10/27/2023 01:29:58

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.077	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.233	01:25
Breath Volume	2.281	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.220	01:28
Breath Volume	1.804	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/27/2023 04:31:55

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.126	04:27
Breath Volume	1.851	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.125	04:30
Breath Volume	1.914	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODFREY,JACOB,M
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: ETTEL,MICHAEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: NOWLIN

DATE/TIME OF TEST: 10/27/2023 17:47:13

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.080	17:41
Air Blank	0.000	17:42
Subject Sample # 1	0.000	17:43
Breath Volume	2.519	17:43
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.000	17:45
Breath Volume	1.289	17:45
Air Blank	0.000	17:46
Control Test	0.078	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: KING,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: MONTAGUE

DATE/TIME OF TEST: 10/28/2023 20:05:51

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:51	
Control Test	0.080	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1 0.000	19:53	
Breath Volume	1.121	19:53	
Air Blank	0.000	19:53	
Air Blank	0.000	19:55	
Subject Sample	# 2 0.000	19:58	Volume Not Met
Breath Volume	1.082	19:58	
Air Blank	0.000	19:59	
Air Blank	0.000	20:01	
Subject Sample	# 3 0.000	20:04	Volume Not Met
Breath Volume	0.250	20:04	
Air Blank	0.000	20:04	
Control Test	0.078	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAHNSEN,ANDREW,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: KING,N
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/29/2023 23:17:14

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.220	23:12
Breath Volume	1.847	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.226	23:15
Breath Volume	2.429	23:15
Air Blank	0.000	23:16
Control Test	0.078	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 11/01/2023 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,DANELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 10/08/2023 04:49:45

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.079	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.079	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 10/27/2023 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LARA,JOSHUA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 05:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: SILVA CENTENO

DATE/TIME OF TEST: 10/15/2023 06:16:48

Diagnostics Check	OK	06:15	
Air Blank		06:16	Ambient Fail
Air Blank	0.000	06:16	

Data Download Date/Time: 10/27/2023 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LARA,JOSHUA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 05:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: SILVA CENTENO

DATE/TIME OF TEST: 10/15/2023 06:32:32

Diagnostics Check	OK	06:19	
Air Blank	0.000	06:19	
Control Test	0.078	06:20	
Air Blank	0.000	06:20	
Subject Sample	# 1 0.124	06:23	Volume Not Met
Breath Volume	0.234	06:23	
Air Blank	0.000	06:24	
Air Blank	0.000	06:26	
Subject Sample	# 2 0.131	06:27	
Breath Volume	1.511	06:27	
Air Blank	0.000	06:28	
Air Blank	0.000	06:30	
Subject Sample	# 3 0.147	06:30	
Breath Volume	1.140	06:30	
Air Blank	0.000	06:31	
Control Test	0.075	06:31	
Air Blank	0.000	06:32	
Diagnostics Check	OK	06:32	

Data Download Date/Time: 10/27/2023 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: MOSQUERA,SANTIAGO,D	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)	VIOLATION CODE: DUI
ARRESTING OFFICER: MOSQUERA,SANTIAGO	ARREST TIME: 20:35
ARRESTING OFFICER AGENCY: FFWCC (West Palm)	GENDER: Male
LAST AGENCY INSPECTION: 09/01/2023	OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: RAMIREZ	DATE/TIME OF TEST: 10/18/2023 22:25:25
------------------------------	--

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:23	
Control Test	0.079	22:23	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:24	Subject Test Refused
Breath Volume	0.000	22:24	
Air Blank	0.000	22:24	
Control Test	0.079	22:24	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 10/27/2023 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO	ARREST TIME: 02:58
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Male
LAST AGENCY INSPECTION: 09/01/2023	OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: RODRIGUEZ VEGA	DATE/TIME OF TEST: 10/27/2023 04:00:15
-------------------------------------	--

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.233	03:54
Breath Volume	2.066	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.239	03:58
Breath Volume	1.765	03:58
Air Blank	0.000	03:59
Control Test	0.076	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 10/27/2023 12:11