

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: NOLAND,THOMAS	ARREST TIME: 23:29
ARRESTING OFFICER AGENCY: Lake Mary PD	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MILLER	DATE/TIME OF TEST: 12/06/2023 00:32:23
-----------------------------	--

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.100	00:27
Breath Volume	1.531	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.106	00:30
Breath Volume	1.804	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 12/15/2023 04:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SCHNIEDER,NICOLAS	ARREST TIME: 01:37
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Female
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: CAMERON	DATE/TIME OF TEST: 12/06/2023 03:55:24
------------------------------	--

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.162	03:50
Breath Volume	2.183	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.165	03:53
Breath Volume	2.078	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 12/15/2023 04:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BETHAM,KEITH,D	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GENTRUP,MEGAN	ARREST TIME: 01:34
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: VASSELL	DATE/TIME OF TEST: 12/08/2023 02:32:26
------------------------------	--

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.184	02:27
Breath Volume	3.402	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.185	02:30
Breath Volume	2.925	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 12/15/2023 04:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: FERGUSON,GREGORY	ARREST TIME: 01:45
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CURTIS	DATE/TIME OF TEST: 12/09/2023 04:12:12
-----------------------------	--

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.256	04:07
Breath Volume	3.039	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.265	04:10
Breath Volume	2.082	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 12/15/2023 04:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: KAREEM,DEVIN

ARRESTING OFFICER AGENCY: Seminole County SO

LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 01:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GOMEZ-FERNANDEZ

DATE/TIME OF TEST: 12/10/2023 02:18:20

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.080	02:15	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:16	Slope Not Met
Breath Volume	1.859	02:16	
Air Blank	0.000	02:17	
Control Test	0.080	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 12/15/2023 04:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: KAREEM,DEVIN	ARREST TIME: 01:19
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GOMEZ-FERNANDEZ	DATE/TIME OF TEST: 12/10/2023 02:42:37
--------------------------------------	--

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.035	02:41	
Air Blank	0.000	02:41	
Control Test	0.080	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 12/15/2023 04:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: PACHECO,MANUEL

ARRESTING OFFICER AGENCY: Seminole County SO

LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 02:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BELLINATO

DATE/TIME OF TEST: 12/10/2023 03:56:40

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.161	03:52
Breath Volume	3.046	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.162	03:55
Breath Volume	3.152	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 12/15/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: COOK,ANTHONY	ARREST TIME: 03:01
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Female
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: ABRANTE	DATE/TIME OF TEST: 12/10/2023 05:03:33
------------------------------	--

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.172	04:59
Breath Volume	2.800	04:59
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.176	05:02
Breath Volume	2.769	05:02
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 12/15/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: MORRISSEY,ZACHARY,M	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD	VIOLATION CODE: DUI
ARRESTING OFFICER: BELT,WILLIAM	ARREST TIME: 06:53
ARRESTING OFFICER AGENCY: Altamonte Springs PD	GENDER: Female
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: BELALCAZAR SANCHEZ	DATE/TIME OF TEST: 12/10/2023	08:36:06
---	-------------------------------	----------

Diagnostics Check	OK	08:30	
Air Blank	0.000	08:30	
Control Test	0.079	08:31	
Air Blank	0.000	08:31	
Subject Sample	# 1 0.119	08:34	
Breath Volume	1.175	08:34	
Air Blank		08:34	RFI Detect
Air Blank	0.000	08:35	
Control Test	0.081	08:35	
Air Blank	0.000	08:36	
Diagnostics Check	OK	08:36	

Data Download Date/Time: 12/15/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: MORRISSEY,ZACHARY,M	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD	VIOLATION CODE: DUI
ARRESTING OFFICER: BELT,WILLIAM	ARREST TIME: 06:53
ARRESTING OFFICER AGENCY: Altamonte Springs PD	GENDER: Female
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: BELALCAZAR SANCHEZ	DATE/TIME OF TEST: 12/10/2023 08:49:28
---	--

Diagnostics Check	OK	08:37	
Air Blank	0.000	08:38	
Control Test	0.078	08:38	
Air Blank	0.000	08:39	
Subject Sample	# 1 0.105	08:42	Volume Not Met
Breath Volume	0.015	08:42	
Air Blank	0.000	08:43	
Air Blank	0.000	08:44	
Subject Sample	# 2 0.109	08:47	Volume Not Met
Breath Volume	0.507	08:47	
Air Blank	0.000	08:48	
Control Test	0.078	08:48	
Air Blank	0.000	08:49	
Diagnostics Check	OK	08:49	

Data Download Date/Time: 12/15/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: MORRISSEY,ZACHARY,M	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD	VIOLATION CODE: DUI
ARRESTING OFFICER: TRAN,THOMAS	ARREST TIME: 07:33
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 07:28

SUBJECT'S LAST NAME: DOMINGO	DATE/TIME OF TEST: 12/10/2023 09:02:55
------------------------------	--

Diagnostics Check	OK	08:56
Air Blank	0.000	08:56
Control Test	0.079	08:56
Air Blank	0.000	08:57
Subject Sample # 1	0.098	08:58
Breath Volume	1.671	08:58
Air Blank	0.000	08:59
Air Blank	0.000	09:00
Subject Sample # 2	0.096	09:01
Breath Volume	1.792	09:01
Air Blank	0.000	09:02
Control Test	0.078	09:02
Air Blank	0.000	09:02
Diagnostics Check	OK	09:02

Data Download Date/Time: 12/15/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GUSTAVE,L	ARREST TIME: 20:51
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BRAVO BERNARDO	DATE/TIME OF TEST: 12/11/2023 00:14:35
-------------------------------------	--

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.080	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.217	00:09
Breath Volume	1.675	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.208	00:13
Breath Volume	1.574	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 12/15/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: RIOS,E	ARREST TIME: 03:06
ARRESTING OFFICER AGENCY: Longwood PD	GENDER: Female
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CLEMENTS	DATE/TIME OF TEST: 12/13/2023 04:09:24
-------------------------------	--

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.172	04:04
Breath Volume	1.363	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.177	04:07
Breath Volume	1.753	04:07
Air Blank	0.000	04:08
Control Test	0.078	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 12/15/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VANROMPA,V,

BREATH TEST OPERATOR AGENCY: Lake Mary PD

ARRESTING OFFICER: SNYDER,B

ARRESTING OFFICER AGENCY: Longwood PD

LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 17:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/13/2023 18:45:18

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.080	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.000	18:40
Breath Volume	2.519	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.000	18:43
Breath Volume	2.535	18:43
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

Data Download Date/Time: 12/15/2023 04:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PRICE,CHASE	ARREST TIME: 04:31
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Female
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: SANTIAGO MALDONADO	DATE/TIME OF TEST: 12/16/2023 05:19:36
---	--

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.079	05:09
Air Blank	0.000	05:09
Subject Sample # 1	0.179	05:11
Breath Volume	1.265	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.200	05:14
Breath Volume	1.507	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 3	0.207	05:18
Breath Volume	1.281	05:18
Air Blank	0.000	05:18
Control Test	0.079	05:19
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BETHAM,KEITH,D	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: BYRD,RYAN	ARREST TIME: 21:23
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: MITCHELL	DATE/TIME OF TEST: 12/16/2023 22:31:45
-------------------------------	--

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.080	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:30	No Sample Provided
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Air Blank	0.000	22:31	

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BETHAM,KEITH,D	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HIDALGO,ALFREDO	ARREST TIME: 21:38
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: VANDERBERG	DATE/TIME OF TEST: 12/16/2023 23:05:38
---------------------------------	--

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.203	23:01
Breath Volume	2.730	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.199	23:04
Breath Volume	2.429	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BETHAM,KEITH,D	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: BYRD,RYAN	ARREST TIME: 21:23
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MITCHELL	DATE/TIME OF TEST: 12/16/2023	23:30:06
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.137	23:25
Breath Volume	1.519	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.136	23:28
Breath Volume	1.621	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: UTZIG,N	ARREST TIME: 03:55
ARRESTING OFFICER AGENCY: Altamonte Springs PD	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: WOODARD	DATE/TIME OF TEST: 12/18/2023 05:53:38
------------------------------	--

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.080	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.155	05:48
Breath Volume	1.441	05:48
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.174	05:52
Breath Volume	1.937	05:52
Air Blank	0.000	05:52
Control Test	0.080	05:53
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: KELLER,KOBE	ARREST TIME: 03:46
ARRESTING OFFICER AGENCY: Lake Mary PD	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: GUTIERREZ	DATE/TIME OF TEST: 12/19/2023 05:05:30
--------------------------------	--

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.244	05:00
Breath Volume	1.375	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.226	05:04
Breath Volume	1.335	05:04
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HORNE,CLINTEN	ARREST TIME: 01:12
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: STROUSE	DATE/TIME OF TEST: 12/20/2023 04:20:32
------------------------------	--

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Subject Sample # 1	0.101	04:09	Volume Not Met
Breath Volume	0.320	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample # 2	0.101	04:14	
Breath Volume	1.679	04:14	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample # 3	0.078	04:19	
Breath Volume	1.117	04:19	
Air Blank	0.000	04:19	
Control Test	0.066	04:19	Control Outside Tolerance
Air Blank	0.000	04:20	

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HORNE,CLINTEN	ARREST TIME: 01:12
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: STROUSE	DATE/TIME OF TEST: 12/20/2023 04:41:53
------------------------------	--

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.080	04:28	
Air Blank	0.000	04:29	
Subject Sample	# 1 0.086	04:32	Volume Not Met
Breath Volume	0.867	04:32	
Air Blank	0.000	04:33	
Air Blank	0.000	04:34	
Subject Sample	# 2 0.090	04:37	
Breath Volume	1.417	04:37	
Air Blank	0.000	04:37	
Air Blank	0.000	04:39	
Subject Sample	# 3 0.079	04:40	
Breath Volume	1.343	04:40	
Air Blank	0.000	04:41	
Control Test	0.072	04:41	Control Outside Tolerance
Air Blank	0.000	04:41	

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BETHAM,KEITH,D

BREATH TEST OPERATOR AGENCY: Seminole County SO

ARRESTING OFFICER: WATT,ANTONIO

ARRESTING OFFICER AGENCY: Altamonte Springs PD

LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 02:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 12/22/2023 03:37:53

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.259	03:33
Breath Volume	1.773	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.254	03:36
Breath Volume	1.640	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POLEY,KYLE,C

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: WILD,JACOB

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 11:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/23/2023 13:04:28

Diagnostics Check	OK	12:57
Air Blank	0.000	12:58
Control Test	0.079	12:58
Air Blank	0.000	12:59
Subject Sample # 1	0.190	12:59
Breath Volume	2.726	12:59
Air Blank	0.000	13:00
Air Blank	0.000	13:02
Subject Sample # 2	0.199	13:02
Breath Volume	2.847	13:02
Air Blank	0.000	13:03
Control Test	0.078	13:03
Air Blank	0.000	13:04
Diagnostics Check	OK	13:04

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GRUNAT,J	ARREST TIME: 02:12
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CAMARAO	DATE/TIME OF TEST: 12/25/2023 03:11:44
------------------------------	--

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.115	03:06
Breath Volume	2.601	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.114	03:10
Breath Volume	1.476	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ESTRADA,BRETT	ARREST TIME: 19:28
ARRESTING OFFICER AGENCY: Casselberry PD	GENDER: Female
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: HILL	DATE/TIME OF TEST: 12/25/2023	21:41:31
---------------------------	-------------------------------	----------

Diagnostics Check	OK	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.221	21:37
Breath Volume	1.925	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.210	21:40
Breath Volume	1.679	21:40
Air Blank	0.000	21:40
Control Test	0.079	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ADAMS,J	ARREST TIME: 20:13
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Female
LAST AGENCY INSPECTION: 11/22/2023	OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DONALDSON	DATE/TIME OF TEST: 12/25/2023	22:19:33
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.196	22:15
Breath Volume	1.863	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.188	22:18
Breath Volume	1.500	22:18
Air Blank	0.000	22:18
Control Test	0.079	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 12/27/2023 05:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: KOURI,TAMY,	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ARMOUR,RACHAL	ARREST TIME: 01:01
ARRESTING OFFICER AGENCY: Casselberry PD	GENDER: Male
LAST AGENCY INSPECTION: 12/27/2023	OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: KAMM	DATE/TIME OF TEST: 12/28/2023 02:05:19
---------------------------	--

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.216	02:00
Breath Volume	1.480	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.201	02:03
Breath Volume	1.636	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 12/28/2023 03:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: KRUPA,KYLE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 12/29/2023 01:46:05

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.173	01:41
Breath Volume	1.621	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.178	01:44
Breath Volume	2.906	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 01/11/2024 05:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BETHAM,KEITH,D	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: PRICE,CHASE	ARREST TIME: 20:41
ARRESTING OFFICER AGENCY: Seminole County SO	GENDER: Male
LAST AGENCY INSPECTION: 12/27/2023	OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: MORENO LARA	DATE/TIME OF TEST: 12/29/2023 21:46:49
----------------------------------	--

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.218	21:41
Breath Volume	1.753	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.221	21:45
Breath Volume	2.375	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 01/11/2024 05:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: AJUILAR,JHONNY	ARREST TIME: 00:50
ARRESTING OFFICER AGENCY: Lake Mary PD	GENDER: Female
LAST AGENCY INSPECTION: 12/27/2023	OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: KEIGER	DATE/TIME OF TEST: 12/30/2023 02:02:37
-----------------------------	--

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.139	01:58
Breath Volume	2.347	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.138	02:01
Breath Volume	2.417	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 01/11/2024 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: RODRIGUEZ,EMMANUEL	ARREST TIME: 03:17
ARRESTING OFFICER AGENCY: Altamonte Springs PD	GENDER: Male
LAST AGENCY INSPECTION: 12/27/2023	OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CASTRO	DATE/TIME OF TEST: 12/31/2023 04:26:57
-----------------------------	--

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.088	04:22
Breath Volume	3.183	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.087	04:25
Breath Volume	3.105	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 01/11/2024 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1433248
BREATH TEST OPERATOR: BLAZINA,MEGAN,N	DRY GAS EXPIRY DATE: 01/06/2024
BREATH TEST OPERATOR AGENCY: Seminole County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: DELORBE,DOMINICK	ARREST TIME: 19:30
ARRESTING OFFICER AGENCY: Sanford PD	GENDER: Male
LAST AGENCY INSPECTION: 12/27/2023	OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: ROSARIO	DATE/TIME OF TEST: 12/31/2023 20:37:31
------------------------------	--

Diagnostics Check	OK	20:21	
Air Blank	0.000	20:21	
Control Test	0.079	20:21	
Air Blank	0.000	20:22	
Subject Sample	# 1 0.236	20:25	No .020 agreement
Breath Volume	1.656	20:25	
Air Blank	0.000	20:25	
Air Blank	0.000	20:27	
Subject Sample	# 2 0.259	20:30	No .020 agreement
Breath Volume	2.199	20:30	
Air Blank	0.000	20:30	
Air Blank	0.000	20:32	
Subject Sample	# 3 0.228	20:35	Volume Not Met
Breath Volume	0.367	20:35	
Air Blank	0.000	20:36	
Control Test	0.077	20:36	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

Data Download Date/Time: 01/11/2024 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/10/2023 02:52:59

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.077	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.163	02:48
Breath Volume	2.906	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.147	02:51
Breath Volume	1.242	02:51
Air Blank	0.000	02:52
Control Test	0.077	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 12/20/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402623828

DRY GAS EXPIRY DATE: 10/22/2025

VIOLATION CODE: DUI

ARREST TIME: 23:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MABINI

DATE/TIME OF TEST: 12/14/2023 23:58:52

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.080	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.152	23:54
Breath Volume	2.226	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.141	23:57
Breath Volume	1.503	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 12/20/2023 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: ESQUIVEL,MISEI,	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ESQUIVEL,MISEI	ARREST TIME: 02:32
ARRESTING OFFICER AGENCY: Okeechobee County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/28/2023	OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: APONTE	DATE/TIME OF TEST: 12/02/2023 03:19:20
-----------------------------	--

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.081	03:06	
Air Blank	0.000	03:06	
Subject Sample # 1	0.208	03:07	
Breath Volume	2.648	03:07	
Air Blank	0.000	03:08	
Air Blank	0.000	03:10	
Subject Sample # 2		03:13	No Sample Provided
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample # 3		03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Control Test	0.080	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBERTS,MARK,K

BREATH TEST OPERATOR AGENCY: Okeechobee County SO

ARRESTING OFFICER: FAIRTRACE,JENNIFER

ARRESTING OFFICER AGENCY: Okeechobee County SO

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1474747

DRY GAS EXPIRY DATE: 04/12/2024

VIOLATION CODE: DUI

ARREST TIME: 01:01

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/05/2023 01:54:41

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.080	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: ESQUIVEL,MISEI,	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ESQUIVEL,MISEI	ARREST TIME: 20:05
ARRESTING OFFICER AGENCY: Okeechobee County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/28/2023	OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: SHRUM	DATE/TIME OF TEST: 12/08/2023 20:45:35
----------------------------	--

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.081	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.102	20:41
Breath Volume	4.796	20:41
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.098	20:44
Breath Volume	4.941	20:44
Air Blank	0.000	20:44
Control Test	0.080	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: ROBERTS,MARK,K	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GONZALEZ,VICTOR	ARREST TIME: 21:29
ARRESTING OFFICER AGENCY: Okeechobee County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/28/2023	OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: PINEDA-GALVIN	DATE/TIME OF TEST: 12/09/2023 22:20:25
------------------------------------	--

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.123	22:15
Breath Volume	2.898	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.121	22:18
Breath Volume	2.648	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: ROBERTS,MARK,K	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: VARNADORE,RICHARD	ARREST TIME: 22:31
ARRESTING OFFICER AGENCY: Okeechobee County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/28/2023	OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: ARROYO-HERNANDEZ	DATE/TIME OF TEST: 12/09/2023 23:42:38
---------------------------------------	--

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.081	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.142	23:37
Breath Volume	2.105	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.140	23:41
Breath Volume	2.367	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: ROBERTS,MARK,K	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: VARNADORE,RICHARD	ARREST TIME: 00:31
ARRESTING OFFICER AGENCY: Okeechobee County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/28/2023	OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: YOUNG	DATE/TIME OF TEST: 12/10/2023 01:15:57
----------------------------	--

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.196	01:11
Breath Volume	3.476	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:13
Subject Sample # 2	0.200	01:14
Breath Volume	3.687	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: WRIGHT, MARK, A	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: WRIGHT, M	ARREST TIME: 19:44
ARRESTING OFFICER AGENCY: Okeechobee County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/28/2023	OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ARNETT	DATE/TIME OF TEST: 12/22/2023 20:38:30
-----------------------------	--

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:36	
Control Test	0.080	20:36	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	
Control Test	0.081	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASTANEDA,RAFAEL,

BREATH TEST OPERATOR AGENCY: Okeechobee PD

ARRESTING OFFICER: CASTANEDA,RAFAEL

ARRESTING OFFICER AGENCY: Okeechobee PD

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1474747

DRY GAS EXPIRY DATE: 04/12/2024

VIOLATION CODE: DUI

ARREST TIME: 23:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: LUNA CERVANTES

DATE/TIME OF TEST: 12/26/2023 00:48:45

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.235	00:43
Breath Volume	2.039	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.243	00:47
Breath Volume	1.875	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: ESQUIVEL,MISEI,	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ESQUIVEL,MISEI	ARREST TIME: 06:38
ARRESTING OFFICER AGENCY: Okeechobee County SO	GENDER: Female
LAST AGENCY INSPECTION: 11/28/2023	OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: CURTIS	DATE/TIME OF TEST: 12/26/2023 07:25:12
-----------------------------	--

Diagnostics Check	OK	07:22	
Air Blank	0.000	07:22	
Control Test	0.081	07:23	
Air Blank	0.000	07:23	
Subject Sample	# 1	07:23	Subject Test Refused
Breath Volume	0.000	07:23	
Air Blank	0.000	07:24	
Control Test	0.081	07:24	
Air Blank	0.000	07:25	
Diagnostics Check	OK	07:25	

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: MURTAGH,ROBERT,J	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MURTAGH,ROBERT	ARREST TIME: 19:13
ARRESTING OFFICER AGENCY: Okeechobee County SO	GENDER: Female
LAST AGENCY INSPECTION: 11/28/2023	OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: DAMES	DATE/TIME OF TEST: 12/29/2023 20:19:43
----------------------------	--

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
Control Test	0.080	20:10	
Air Blank	0.000	20:11	
Subject Sample # 1	0.171	20:14	Volume Not Met
Breath Volume	0.578	20:14	
Air Blank	0.000	20:15	
Air Blank	0.000	20:16	
Subject Sample # 2		20:18	RFI Detect
Breath Volume	0.718	20:18	
Air Blank	0.000	20:18	
Control Test	0.079	20:19	
Air Blank	0.000	20:19	
Diagnostics Check	OK	20:19	

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: MURTAGH,ROBERT,J	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MURTAGH,ROBERT	ARREST TIME: 19:13
ARRESTING OFFICER AGENCY: Okeechobee County SO	GENDER: Female
LAST AGENCY INSPECTION: 11/28/2023	OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: DAMES	DATE/TIME OF TEST: 12/29/2023 20:30:41
----------------------------	--

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.080	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.170	20:25
Breath Volume	1.582	20:25
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.174	20:29
Breath Volume	2.062	20:29
Air Blank	0.000	20:29
Control Test	0.080	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 12/30/2023 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MURTAGH,ROBERT,J

BREATH TEST OPERATOR AGENCY: Okeechobee County SO

ARRESTING OFFICER: FAIRTRACE,JENNIFER

ARRESTING OFFICER AGENCY: Okeechobee County SO

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1474747

DRY GAS EXPIRY DATE: 04/12/2024

VIOLATION CODE: DUI

ARREST TIME: 22:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 12/29/2023 23:27:59

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.080	23:25	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:27	Sequence Aborted
Breath Volume	0.000	23:27	
Air Blank	0.000	23:27	

Data Download Date/Time: 12/30/2023 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: MURTAGH,ROBERT,J	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: FAIRTRACE,JENNIFER	ARREST TIME: 22:43
ARRESTING OFFICER AGENCY: Bay County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/28/2023	OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: VALDES	DATE/TIME OF TEST: 12/29/2023 23:46:10
-----------------------------	--

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank		23:45	Improper Sample
Air Blank	0.000	23:46	

Data Download Date/Time: 12/30/2023 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1474747
BREATH TEST OPERATOR: MURTAGH,ROBERT,J	DRY GAS EXPIRY DATE: 04/12/2024
BREATH TEST OPERATOR AGENCY: Okeechobee County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MURTAGH,ROBERT	ARREST TIME: 00:03
ARRESTING OFFICER AGENCY: Okeechobee County SO	GENDER: Male
LAST AGENCY INSPECTION: 12/30/2023	OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MONTIEL GUTIERREZ	DATE/TIME OF TEST: 12/31/2023 00:57:42
--	--

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.126	00:52
Breath Volume	2.140	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.120	00:56
Breath Volume	1.250	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 12/31/2023 09:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHAELS,GLEN,A
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: MICHAELS,GLEN
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 12/13/2023 19:10:44

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:07	
Control Test	0.080	19:07	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:09	Subject Test Refused
Breath Volume	0.000	19:09	
Air Blank	0.000	19:09	
Control Test	0.080	19:10	
Air Blank	0.000	19:10	
Diagnostics Check	OK	19:10	

Data Download Date/Time: 01/16/2024 00:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA, ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: PORTILLO PEREZ

DATE/TIME OF TEST: 12/01/2023 18:53:03

Diagnostics Check	OK	18:43
Air Blank	0.000	18:43
Control Test	0.079	18:44
Air Blank	0.000	18:44
Subject Sample # 1	0.214	18:45
Breath Volume	1.312	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:47
Subject Sample # 2	0.239	18:48
Breath Volume	1.687	18:48
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 3	0.235	18:51
Breath Volume	1.519	18:51
Air Blank	0.000	18:52
Control Test	0.080	18:52
Air Blank	0.000	18:52
Diagnostics Check	OK	18:53

Data Download Date/Time: 12/18/2023 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 12/08/2023 00:37:26

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.177	00:32
Breath Volume	1.152	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:34
Subject Sample # 2	0.174	00:35
Breath Volume	1.886	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 12/18/2023 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: OATES

DATE/TIME OF TEST: 12/12/2023 22:35:18

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.081	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1 0.088	22:26	Volume Not Met
Breath Volume	0.207	22:26	
Air Blank	0.000	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 2 0.098	22:30	
Breath Volume	1.488	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 3 0.094	22:33	
Breath Volume	1.292	22:33	
Air Blank	0.000	22:34	
Control Test	0.078	22:34	
Air Blank	0.000	22:35	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 12/18/2023 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,TIMYRON,N
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: YOUNG,TIMYRON
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/13/2023 00:23:11

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.000	00:18
Breath Volume	1.222	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.000	00:21
Breath Volume	1.695	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 12/18/2023 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: BURDESS

DATE/TIME OF TEST: 12/13/2023 23:38:12

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.138	23:33
Breath Volume	2.863	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.147	23:36
Breath Volume	1.804	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 12/18/2023 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: ZIEGLER

DATE/TIME OF TEST: 12/03/2023 07:18:09

Diagnostics Check	OK	07:10
Air Blank	0.000	07:10
Control Test	0.081	07:10
Air Blank	0.000	07:11
Subject Sample # 1	0.163	07:12
Breath Volume	2.031	07:12
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample # 2	0.163	07:16
Breath Volume	1.132	07:16
Air Blank	0.000	07:17
Control Test	0.080	07:17
Air Blank	0.000	07:18
Diagnostics Check	OK	07:18

Data Download Date/Time: 12/13/2023 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MINAYA,AYANIS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 12/05/2023 03:09:58

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.153	03:04
Breath Volume	1.496	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.154	03:08
Breath Volume	1.328	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 12/13/2023 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 12/09/2023 02:41:13

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1 0.167	02:33	Volume Not Met
Breath Volume	0.398	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 2 0.180	02:39	Volume Not Met
Breath Volume	0.753	02:39	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 12/13/2023 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 12/09/2023 02:55:08

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.079	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1 0.192	02:47	Volume Not Met
Breath Volume	0.019	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample	# 2 0.176	02:53	Volume Not Met
Breath Volume	0.527	02:53	
Air Blank	0.000	02:54	
Control Test	0.080	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 12/13/2023 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LUCKEY

DATE/TIME OF TEST: 12/05/2023 02:16:06

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.078	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Control Test	0.079	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 12/25/2023 02:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,NATHAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SUMNER,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/09/2023 08:34:11

Diagnostics Check	OK	08:27
Air Blank	0.000	08:27
Control Test	0.079	08:28
Air Blank	0.000	08:28
Subject Sample # 1	0.141	08:29
Breath Volume	2.601	08:29
Air Blank	0.000	08:30
Air Blank	0.000	08:32
Subject Sample # 2	0.133	08:32
Breath Volume	2.781	08:32
Air Blank	0.000	08:33
Control Test	0.077	08:33
Air Blank	0.000	08:34
Diagnostics Check	OK	08:34

Data Download Date/Time: 12/25/2023 02:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: RECCA,M
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 11/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 12/10/2023 01:09:25

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.186	01:04
Breath Volume	2.687	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.189	01:07
Breath Volume	2.753	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 12/25/2023 02:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 11/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/16/2023 04:33:12

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.080	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 12/25/2023 02:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 11/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 12/17/2023 19:36:22

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.079	19:30
Air Blank	0.000	19:31
Subject Sample # 1	0.108	19:31
Breath Volume	3.000	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.115	19:34
Breath Volume	1.750	19:34
Air Blank	0.000	19:35
Control Test	0.078	19:35
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 12/25/2023 02:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL SASSO,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: DEL SASSO,MICHAEL
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 11/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402469736
DRY GAS EXPIRY DATE: 06/16/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LINARES MARTINEZ

DATE/TIME OF TEST: 12/23/2023 02:34:41

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.206	02:30
Breath Volume	3.570	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.207	02:32
Breath Volume	4.566	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 12/25/2023 02:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIAGET,PHILIP,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PIAGET,PHILIP
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 12/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: RUTHVEN-BUSBY

DATE/TIME OF TEST: 12/06/2023 05:40:07

Diagnostics Check	OK	05:27	
Air Blank	0.000	05:28	
Control Test	0.079	05:28	
Air Blank	0.000	05:28	
Subject Sample	# 1 0.129	05:30	
Breath Volume	1.171	05:30	
Air Blank	0.000	05:30	
Air Blank	0.000	05:32	
Subject Sample	# 2 0.110	05:35	Volume Not Met
Breath Volume	0.378	05:35	
Air Blank	0.000	05:36	
Air Blank	0.000	05:38	
Subject Sample	# 3 0.127	05:38	
Breath Volume	3.300	05:38	
Air Blank	0.000	05:39	
Control Test	0.076	05:39	
Air Blank	0.000	05:40	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 01/08/2024 05:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,DYLAN,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HART,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: MUSSO

DATE/TIME OF TEST: 12/06/2023 20:17:25

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.081	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.028	20:13
Breath Volume	2.125	20:13
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.024	20:15
Breath Volume	1.550	20:15
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 01/08/2024 05:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 12/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SCHUNDELMEIR

DATE/TIME OF TEST: 12/09/2023 00:42:50

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.150	00:37
Breath Volume	1.839	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.148	00:41
Breath Volume	1.507	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 01/08/2024 05:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACKENZIE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 12/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/09/2023 02:06:40

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank		02:06	RFI Detect
Air Blank	0.000	02:06	

Data Download Date/Time: 01/08/2024 05:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACKENZIE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 12/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/09/2023 02:19:47

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:12	No Sample Provided
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Air Blank	0.000	02:15	
Subject Sample	# 2	02:18	No Sample Provided
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Control Test	0.081	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 01/08/2024 05:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HISAW,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 12/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: SMOLKO

DATE/TIME OF TEST: 12/09/2023 05:23:37

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:12	
Control Test	0.080	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 1	05:16	No Sample Provided
Breath Volume	0.000	05:16	
Air Blank	0.000	05:16	
Air Blank	0.000	05:18	
Subject Sample	# 2	0.209	05:19
Breath Volume	1.894	05:19	
Air Blank	0.000	05:19	
Air Blank	0.000	05:21	
Subject Sample	# 3	0.211	05:22
Breath Volume	1.957	05:22	
Air Blank	0.000	05:22	
Control Test	0.079	05:23	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:23	

Data Download Date/Time: 01/08/2024 05:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JESSICA,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NELSON,JESSICA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 12/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 12/10/2023 00:11:09

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.117	00:06
Breath Volume	2.417	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.118	00:09
Breath Volume	1.371	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 01/08/2024 05:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACKENZIE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 12/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CLEVELAND

DATE/TIME OF TEST: 12/10/2023 03:30:41

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:23	No Sample Provided
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2	0.130	03:26
Breath Volume	2.074	03:26	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 3	0.142	03:29
Breath Volume	1.679	03:29	
Air Blank	0.000	03:29	
Control Test	0.079	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 12/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: REYNER

DATE/TIME OF TEST: 12/10/2023 04:54:46

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.081	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.150	04:49
Breath Volume	1.425	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.131	04:53
Breath Volume	1.589	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LOPEZ,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: ANDREW

DATE/TIME OF TEST: 12/11/2023 02:35:51

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.000	02:31
Breath Volume	2.238	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.000	02:34
Breath Volume	1.777	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACKENZIE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HORAK

DATE/TIME OF TEST: 12/15/2023 01:58:19

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.114	01:53
Breath Volume	1.773	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.126	01:56
Breath Volume	1.277	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COOLEY,LOGAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 12/16/2023 03:12:46

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.146	03:08
Breath Volume	2.726	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.139	03:11
Breath Volume	2.851	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: BARBAREE

DATE/TIME OF TEST: 12/16/2023 21:19:01

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.080	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:14	No Sample Provided
Breath Volume	0.000	21:14	
Air Blank	0.000	21:15	
Air Blank	0.000	21:17	
Subject Sample	# 2	21:17	Subject Test Refused
Breath Volume	0.000	21:17	
Air Blank	0.000	21:18	
Control Test	0.079	21:18	
Air Blank	0.000	21:18	
Diagnostics Check	OK	21:19	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISAIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ODDO,ISAIAH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/16/2023 21:28:19

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.078	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:27	
Control Test	0.078	21:27	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MAPOLES

DATE/TIME OF TEST: 12/16/2023 23:24:34

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.208	23:19
Breath Volume	1.199	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.206	23:23
Breath Volume	1.175	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISAIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: RIVERA RAMOS

DATE/TIME OF TEST: 12/17/2023 03:28:39

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Control Test	0.080	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JESSICA,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BLACKWOOD,ADAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 12/19/2023 04:02:07

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.080	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:55	No Sample Provided
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 2	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.081	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: HILLARD

DATE/TIME OF TEST: 12/21/2023 21:01:22

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.081	20:56	
Air Blank	0.000	20:57	
Subject Sample	# 1	21:00	Subject Test Refused
Breath Volume	0.000	21:00	
Air Blank	0.000	21:00	
Control Test	0.081	21:00	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 12/21/2023 21:19:12

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.082	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.000	21:14
Breath Volume	3.128	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.000	21:17
Breath Volume	2.550	21:17
Air Blank	0.000	21:18
Control Test	0.082	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BRACKEEN

DATE/TIME OF TEST: 12/22/2023 00:28:50

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.081	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1 0.132	00:21	Volume Not Met
Breath Volume	0.507	00:21	
Air Blank	0.000	00:22	
Air Blank	0.000	00:24	
Subject Sample	# 2 0.159	00:27	Volume Not Met
Breath Volume	0.164	00:27	
Air Blank	0.000	00:27	
Control Test	0.082	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BRACKEEN

DATE/TIME OF TEST: 12/22/2023 00:41:31

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1 0.162	00:34	Volume Not Met
Breath Volume	0.347	00:34	
Air Blank	0.000	00:34	
Air Blank	0.000	00:36	
Subject Sample	# 2 0.159	00:39	Volume Not Met
Breath Volume	0.000	00:39	
Air Blank	0.000	00:40	
Control Test	0.079	00:40	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BRACKEEN

DATE/TIME OF TEST: 12/22/2023 00:56:43

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:43	
Subject Sample	# 1 0.162	00:47	Volume Not Met
Breath Volume	1.050	00:47	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2 0.154	00:50	
Breath Volume	1.523	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 3 0.148	00:55	
Breath Volume	1.433	00:55	
Air Blank	0.000	00:55	
Control Test	0.080	00:56	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACKENZIE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: CARLISLE

DATE/TIME OF TEST: 12/22/2023 21:22:25

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.080	21:18	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:21	Subject Test Refused
Breath Volume	0.000	21:21	
Air Blank	0.000	21:21	
Control Test	0.082	21:21	
Air Blank	0.000	21:22	
Diagnostics Check	OK	21:22	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,BOBBY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARRISON,BOBBY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 12/22/2023 21:41:17

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.226	21:36
Breath Volume	1.679	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.225	21:39
Breath Volume	1.101	21:39
Air Blank	0.000	21:40
Control Test	0.080	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,DYLAN,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HARTER,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: KOSSIK

DATE/TIME OF TEST: 12/26/2023 03:02:24

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.123	02:57
Breath Volume	1.125	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.140	03:00
Breath Volume	1.582	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RODRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CEJAS

DATE/TIME OF TEST: 12/27/2023 01:04:33

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.081	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 1	01:00	No Sample Provided
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.081	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: ARSENAULT

DATE/TIME OF TEST: 12/29/2023 21:46:36

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.135	21:41
Breath Volume	3.035	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:43
Subject Sample # 2	0.120	21:44
Breath Volume	1.757	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,DYLAN,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: GOMILLION,LANCE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HEYWARD

DATE/TIME OF TEST: 12/30/2023 02:07:18

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.166	02:02
Breath Volume	2.121	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.171	02:05
Breath Volume	2.796	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BODIE

DATE/TIME OF TEST: 12/30/2023 04:39:47

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.155	04:34
Breath Volume	1.230	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.156	04:38
Breath Volume	1.203	04:38
Air Blank	0.000	04:38
Control Test	0.081	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,LONNIE,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ODDO,ISAIAH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 12/31/2023 03:53:48

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.081	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODDO,ISAIAH,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELCH,BLAKE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 12/31/2023 05:35:38

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.125	05:30
Breath Volume	3.902	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.121	05:34
Breath Volume	3.476	05:34
Air Blank	0.000	05:34
Control Test	0.078	05:35
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 01/08/2024 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LORD,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/03/2023 02:41:14

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.077	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.251	02:36
Breath Volume	3.101	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:38
Subject Sample # 2	0.248	02:39
Breath Volume	3.449	02:39
Air Blank	0.000	02:40
Control Test	0.076	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 12/29/2023 17:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: GUERRA,JOSE
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 12/09/2023 09:57:13

Diagnostics Check	OK	09:50
Air Blank	0.000	09:51
Control Test	0.077	09:51
Air Blank	0.000	09:51
Subject Sample # 1	0.109	09:52
Breath Volume	1.218	09:52
Air Blank	0.000	09:53
Air Blank	0.000	09:54
Subject Sample # 2	0.118	09:55
Breath Volume	2.410	09:55
Air Blank	0.000	09:56
Control Test	0.077	09:56
Air Blank	0.000	09:57
Diagnostics Check	OK	09:57

Data Download Date/Time: 12/29/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUMES,HEIDI,N
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HUMES,HEIDI
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 12/12/2023 21:41:52

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.048	21:37
Breath Volume	1.128	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.047	21:40
Breath Volume	2.832	21:40
Air Blank	0.000	21:40
Control Test	0.076	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 12/29/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARAFA,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 12/18/2023 21:52:21

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.078	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.157	21:47
Breath Volume	1.359	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.173	21:50
Breath Volume	1.843	21:50
Air Blank	0.000	21:51
Control Test	0.076	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 12/29/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURR,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DURR,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: TLEHUACTLE CALIHUA

DATE/TIME OF TEST: 12/23/2023 00:47:27

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.158	00:42
Breath Volume	1.589	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.161	00:45
Breath Volume	1.910	00:45
Air Blank	0.000	00:46
Control Test	0.076	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 12/29/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A3
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MERRIT

DATE/TIME OF TEST: 12/23/2023 03:45:55

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.136	03:41
Breath Volume	2.871	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.137	03:44
Breath Volume	2.949	03:44
Air Blank	0.000	03:45
Control Test	0.077	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 12/29/2023 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 12/02/2023 04:10:47

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.083	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.121	04:06
Breath Volume	2.632	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.126	04:08
Breath Volume	2.562	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 12/11/2023 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,DALLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: DELATOUR

DATE/TIME OF TEST: 12/10/2023 05:29:11

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:19	
Control Test	0.082	05:19	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:23	No Sample Provided
Breath Volume	0.000	05:23	
Air Blank	0.000	05:23	
Air Blank	0.000	05:25	
Subject Sample	# 2	05:27	Subject Test Refused
Breath Volume	0.000	05:27	
Air Blank	0.000	05:28	
Control Test	0.082	05:28	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:29	

Data Download Date/Time: 12/11/2023 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 12/14/2023 02:15:06

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.082	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.000	02:10
Breath Volume	1.230	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.000	02:13
Breath Volume	1.800	02:13
Air Blank	0.000	02:14
Control Test	0.082	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 01/04/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,DALLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: GRESHAM JR

DATE/TIME OF TEST: 12/15/2023 01:06:56

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.124	01:01
Breath Volume	1.289	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.126	01:05
Breath Volume	1.757	01:05
Air Blank	0.000	01:05
Control Test	0.077	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 01/04/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,DALLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: BOBER

DATE/TIME OF TEST: 12/16/2023 02:24:17

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.128	02:19
Breath Volume	1.355	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.125	02:22
Breath Volume	1.164	02:22
Air Blank	0.000	02:23
Control Test	0.078	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 01/04/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 12/16/2023 03:13:08

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.161	03:07
Breath Volume	2.179	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.173	03:11
Breath Volume	1.339	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 01/04/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BRISCOE

DATE/TIME OF TEST: 12/21/2023 01:26:58

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.082	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.212	01:22
Breath Volume	1.507	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.201	01:25
Breath Volume	1.355	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 01/04/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,CALEB,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WILLIS,CALEB
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/23/2023 17:53:42

Diagnostics Check	OK	17:47
Air Blank	0.000	17:48
Control Test	0.082	17:48
Air Blank	0.000	17:48
Subject Sample # 1	0.000	17:49
Breath Volume	2.550	17:49
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.000	17:52
Breath Volume	2.664	17:52
Air Blank	0.000	17:52
Control Test	0.082	17:53
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 01/04/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEROYT,TAD,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCAVUZZO,NICHOLAS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/02/2023 07:30:14

Diagnostics Check	OK	07:23
Air Blank	0.000	07:24
Control Test	0.081	07:24
Air Blank	0.000	07:25
Subject Sample # 1	0.130	07:25
Breath Volume	4.824	07:25
Air Blank	0.000	07:26
Air Blank	0.000	07:28
Subject Sample # 2	0.129	07:28
Breath Volume	4.460	07:28
Air Blank	0.000	07:29
Control Test	0.081	07:29
Air Blank	0.000	07:30
Diagnostics Check	OK	07:30

Data Download Date/Time: 12/15/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LEHNER

DATE/TIME OF TEST: 12/09/2023 01:12:21

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.083	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.212	01:07
Breath Volume	1.195	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.222	01:10
Breath Volume	2.054	01:10
Air Blank	0.000	01:11
Control Test	0.082	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 12/15/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: KITCHEN

DATE/TIME OF TEST: 12/09/2023 20:57:43

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.082	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.194	20:53
Breath Volume	1.667	20:53
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.183	20:56
Breath Volume	1.292	20:56
Air Blank	0.000	20:56
Control Test	0.082	20:57
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 12/15/2023 16:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ROQUE

DATE/TIME OF TEST: 12/14/2023 03:38:48

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.084	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.175	03:33
Breath Volume	2.734	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.167	03:37
Breath Volume	1.984	03:37
Air Blank	0.000	03:37
Control Test	0.083	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 12/15/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMMER,DALTON,J
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: SWIFT,MICHAEL
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/01/2023 21:56:01

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.080	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.149	21:51
Breath Volume	1.546	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.144	21:54
Breath Volume	2.046	21:54
Air Blank	0.000	21:55
Control Test	0.079	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:56

Data Download Date/Time: 12/04/2023 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINDHAM,AUSTIN,K
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CEPHUS,JAMIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: TRACHTENBERG

DATE/TIME OF TEST: 12/16/2023 00:43:55

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.078	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.000	00:39
Breath Volume	2.140	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.000	00:42
Breath Volume	2.636	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 01/23/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: HODGE,MITCHELL
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: MINA

DATE/TIME OF TEST: 12/23/2023 00:26:44

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.190	00:21
Breath Volume	1.750	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.189	00:25
Breath Volume	1.246	00:25
Air Blank	0.000	00:25
Control Test	0.078	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 01/23/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CEPHUS,JAMIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: TWONS

DATE/TIME OF TEST: 12/29/2023 23:18:48

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.083	23:14
Breath Volume	3.003	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.084	23:17
Breath Volume	3.125	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 01/23/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEEKS,JOHN,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,JOHN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: SOUSA

DATE/TIME OF TEST: 12/01/2023 23:29:56

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:18	
Control Test	0.079	23:19	
Air Blank	0.000	23:19	
Subject Sample	# 1 0.156	23:22	Volume Not Met
Breath Volume	0.031	23:22	
Air Blank	0.000	23:23	
Air Blank	0.000	23:25	
Subject Sample	# 2 0.146	23:28	Volume Not Met
Breath Volume	0.035	23:28	
Air Blank	0.000	23:28	
Control Test	0.078	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 12/25/2023 20:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,ROGER,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MORALES MORALES

DATE/TIME OF TEST: 12/02/2023 01:30:31

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.078	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.132	01:26
Breath Volume	2.875	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.128	01:28
Breath Volume	2.046	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 12/25/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: POLIGO,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: TILLMAN

DATE/TIME OF TEST: 12/02/2023 05:08:56

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:06	
Control Test	0.079	05:06	
Air Blank	0.000	05:07	
Subject Sample	# 1	05:07	Subject Test Refused
Breath Volume	0.000	05:07	
Air Blank	0.000	05:08	
Control Test	0.078	05:08	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 12/25/2023 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,ROGER,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,ROGER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: LOMBARDI

DATE/TIME OF TEST: 12/02/2023 22:26:11

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.077	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.159	22:20
Breath Volume	1.468	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.156	22:24
Breath Volume	1.523	22:24
Air Blank	0.000	22:25
Control Test	0.078	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 12/25/2023 20:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEEKS,JOHN,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,JOHN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: ARNDT

DATE/TIME OF TEST: 12/03/2023 02:05:15

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.078	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:04	
Control Test	0.079	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 12/25/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,MITCHELL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HORN

DATE/TIME OF TEST: 12/08/2023 01:59:35

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.080	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1 0.000	01:50	Volume Not Met
Breath Volume	0.214	01:50	
Air Blank	0.000	01:50	
Air Blank	0.000	01:52	
Subject Sample	# 2 0.000	01:53	
Breath Volume	1.386	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 3 0.000	01:58	
Breath Volume	1.437	01:58	
Air Blank	0.000	01:58	
Control Test	0.081	01:58	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 12/25/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 12/14/2023 01:25:21

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.068	01:20
Breath Volume	1.863	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.070	01:23
Breath Volume	2.507	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 12/25/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ANDRUSS,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/16/2023 02:53:47

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.078	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.213	02:48
Breath Volume	1.121	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.225	02:52
Breath Volume	1.562	02:52
Air Blank	0.000	02:52
Control Test	0.077	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 12/25/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,JAMES,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: GABRIELS

DATE/TIME OF TEST: 12/17/2023 16:06:28

Diagnostics Check	OK	15:59
Air Blank	0.000	16:00
Control Test	0.078	16:00
Air Blank	0.000	16:01
Subject Sample # 1	0.138	16:01
Breath Volume	1.984	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.137	16:04
Breath Volume	2.214	16:04
Air Blank	0.000	16:05
Control Test	0.078	16:05
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 12/25/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEEKS,JOHN,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,JOHN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 12/20/2023 23:16:51

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.079	23:09	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:12	No Sample Provided
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:15	
Subject Sample	# 2	23:15	Subject Test Refused
Breath Volume	0.000	23:15	
Air Blank	0.000	23:15	
Control Test	0.079	23:16	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 12/25/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEEKS,JOHN,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROAN,ELIZABETH
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: HOISINGTON

DATE/TIME OF TEST: 12/21/2023 00:41:42

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.166	00:36
Breath Volume	2.617	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.164	00:40
Breath Volume	2.347	00:40
Air Blank	0.000	00:40
Control Test	0.078	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 12/25/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ELDRIDGE

DATE/TIME OF TEST: 12/23/2023 23:52:48

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.191	23:48
Breath Volume	1.968	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.192	23:51
Breath Volume	2.589	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 12/25/2023 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30822080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: JACKLYN

DATE/TIME OF TEST: 12/24/2023 23:12:12

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.224	23:07
Breath Volume	2.191	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.231	23:10
Breath Volume	1.171	23:10
Air Blank	0.000	23:11
Control Test	0.078	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 12/25/2023 20:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MARSZALEK,B
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: OMAR

DATE/TIME OF TEST: 12/02/2023 00:47:32

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:45	
Control Test	0.079	00:45	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Control Test	0.079	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAMOS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:47

SUBJECT'S LAST NAME: COLEMAN DIALO

DATE/TIME OF TEST: 12/02/2023 14:14:20

Diagnostics Check	OK	14:07
Air Blank	0.000	14:07
Control Test	0.079	14:08
Air Blank	0.000	14:08
Subject Sample # 1	0.150	14:09
Breath Volume	1.667	14:09
Air Blank	0.000	14:10
Air Blank	0.000	14:12
Subject Sample # 2	0.158	14:12
Breath Volume	1.851	14:12
Air Blank	0.000	14:13
Control Test	0.079	14:13
Air Blank	0.000	14:14
Diagnostics Check	OK	14:14

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,MARCUS,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FLOYD,D
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: COUTAIN

DATE/TIME OF TEST: 12/04/2023 01:05:24

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.079	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.126	00:59
Breath Volume	1.921	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.124	01:03
Breath Volume	1.304	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,RODERICK,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HILLMAN,E
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: COLON COLON

DATE/TIME OF TEST: 12/04/2023 20:55:50

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:53	
Control Test	0.079	20:53	
Air Blank	0.000	20:54	
Subject Sample	# 1	20:54	Subject Test Refused
Breath Volume	0.000	20:54	
Air Blank	0.000	20:54	
Control Test	0.080	20:55	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSARIO,F
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: IAMS

DATE/TIME OF TEST: 12/05/2023 20:35:49

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.080	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.175	20:30
Breath Volume	1.753	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.162	20:34
Breath Volume	1.148	20:34
Air Blank	0.000	20:34
Control Test	0.080	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 12/07/2023 00:00:14

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.184	23:55
Breath Volume	3.582	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.173	23:58
Breath Volume	3.660	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSARIO,F
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MCKNIGHT

DATE/TIME OF TEST: 12/08/2023 22:38:37

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.080	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.030	22:33
Breath Volume	1.144	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.028	22:37
Breath Volume	1.167	22:37
Air Blank	0.000	22:37
Control Test	0.079	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MALESKE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 11/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BOGER

DATE/TIME OF TEST: 12/09/2023 03:53:04

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 12/21/2023 19:56:13

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.081	19:50
Air Blank	0.000	19:50
Subject Sample # 1	0.135	19:51
Breath Volume	1.746	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:53
Subject Sample # 2	0.126	19:54
Breath Volume	1.167	19:54
Air Blank	0.000	19:55
Control Test	0.080	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOHERIG,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ROACH

DATE/TIME OF TEST: 12/22/2023 23:47:22

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.079	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.111	23:42
Breath Volume	1.703	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.106	23:45
Breath Volume	1.726	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MALESKE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 12/24/2023 20:31:15

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.079	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.156	20:26
Breath Volume	1.574	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.165	20:29
Breath Volume	2.128	20:29
Air Blank	0.000	20:30
Control Test	0.079	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOGGS,ZACHARY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 12/27/2023 23:52:28

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:41	
Control Test	0.080	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1 0.077	23:45	Volume Not Met
Breath Volume	0.906	23:45	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample	# 2 0.080	23:50	Volume Not Met
Breath Volume	0.941	23:50	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOGGS,ZACHARY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 12/28/2023 00:01:26

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.079	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:59	Subject Test Refused
Breath Volume	0.910	23:59	
Air Blank	0.000	00:00	
Control Test	0.079	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,RODERICK,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DE MATOS

DATE/TIME OF TEST: 12/28/2023 22:28:19

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.080	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.167	22:23
Breath Volume	1.324	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.166	22:26
Breath Volume	1.386	22:26
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,RODERICK,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NORDLE,L
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/29/2023 01:06:41

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.080	01:03	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:05	
Control Test	0.079	01:06	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 12/30/2023 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 12/03/2023 16:08:07

Diagnostics Check	OK	15:55	
Air Blank	0.000	15:56	
Control Test	0.080	15:56	
Air Blank	0.000	15:57	
Subject Sample	# 1 0.000	16:00	Volume Not Met
Breath Volume	0.214	16:00	
Air Blank	0.000	16:00	
Air Blank	0.000	16:02	
Subject Sample	# 2 0.000	16:03	
Breath Volume	1.527	16:03	
Air Blank	0.000	16:03	
Air Blank	0.000	16:05	
Subject Sample	# 3 0.000	16:06	
Breath Volume	1.425	16:06	
Air Blank	0.000	16:07	
Control Test	0.080	16:07	
Air Blank	0.000	16:08	
Diagnostics Check	OK	16:08	

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CENTENO,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BOWENS

DATE/TIME OF TEST: 12/04/2023 02:05:51

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.041	02:01
Breath Volume	3.937	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.043	02:04
Breath Volume	3.796	02:04
Air Blank	0.000	02:04
Control Test	0.081	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,TIM
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: SEISE

DATE/TIME OF TEST: 12/06/2023 03:33:29

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.081	03:30	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ,G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 12/06/2023 22:50:10

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.185	22:44
Breath Volume	1.246	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.173	22:48
Breath Volume	1.496	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BEETOW,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LUTFER

DATE/TIME OF TEST: 12/07/2023 01:25:08

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.124	01:20
Breath Volume	3.050	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample # 2	0.128	01:23
Breath Volume	3.167	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ,G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 12/07/2023 21:04:04

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:01	
Control Test	0.081	21:02	
Air Blank	0.000	21:02	
Subject Sample	# 1	21:02	Subject Test Refused
Breath Volume	0.000	21:02	
Air Blank	0.000	21:03	
Control Test	0.082	21:03	
Air Blank	0.000	21:03	
Diagnostics Check	OK	21:04	

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: NEPHEW

DATE/TIME OF TEST: 12/08/2023 19:44:38

Diagnostics Check	OK	19:34
Air Blank	0.000	19:34
Control Test	0.081	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.246	19:36
Breath Volume	2.539	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:38
Subject Sample # 2	0.209	19:39
Breath Volume	1.410	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 3	0.228	19:42
Breath Volume	3.062	19:42
Air Blank	0.000	19:43
Control Test	0.079	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,ANTHONY,T
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,TIM
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 12/08/2023 22:59:19

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.081	22:56	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:58	
Control Test	0.079	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,L
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SCHIAVONE

DATE/TIME OF TEST: 12/11/2023 22:29:33

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:28	Subject Test Refused
Breath Volume	0.000	22:28	
Air Blank	0.000	22:28	
Control Test	0.081	22:28	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 12/14/2023 20:14:39

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.080	20:08
Air Blank	0.000	20:09
Subject Sample # 1	0.162	20:10
Breath Volume	2.437	20:10
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.162	20:13
Breath Volume	1.238	20:13
Air Blank	0.000	20:13
Control Test	0.079	20:14
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,SHAWN,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BANNON,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: SEIMS

DATE/TIME OF TEST: 12/16/2023 04:18:13

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.081	04:16	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.081	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 12/18/2023 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUBICEK,JORDAN,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ARROYO,JIANNI
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: REYES DE LA CRUZ

DATE/TIME OF TEST: 12/01/2023 00:16:16

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.081	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.204	00:11
Breath Volume	2.562	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.207	00:14
Breath Volume	3.257	00:14
Air Blank	0.000	00:15
Control Test	0.081	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 12/01/2023 23:52:13

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:39	
Control Test	0.081	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:43	No Sample Provided
Breath Volume	0.000	23:43	
Air Blank	0.000	23:43	
Air Blank	0.000	23:45	
Subject Sample	# 2	0.184	23:47
Breath Volume	1.812	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:49	
Subject Sample	# 3	0.187	23:50
Breath Volume	1.285	23:50	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CODY, DEONTA, M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CODY, DEONTA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GATSON J

DATE/TIME OF TEST: 12/02/2023 04:00:15

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.080	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDELL, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: JONES, DENZEL
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GRATEROL MONTILLA

DATE/TIME OF TEST: 12/03/2023 02:59:06

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.188	02:54
Breath Volume	2.667	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.187	02:57
Breath Volume	1.632	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CODY,DEONTA,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CODY,DEONTA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/03/2023 03:14:09

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDELL,M,S
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: TISDELL,M
ARRESTING OFFICER AGENCY: Astatula PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 12/07/2023 20:58:41

Diagnostics Check	OK	20:57	
Air Blank		20:58	RFI Detect
Air Blank	0.000	20:58	

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDELL,M,S
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: TISDELL,M
ARRESTING OFFICER AGENCY: Astatula PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 12/07/2023 21:15:25

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:00	
Control Test	0.081	21:01	
Air Blank	0.000	21:01	
Subject Sample	# 1 0.085	21:04	Volume Not Met
Breath Volume	0.109	21:04	
Air Blank	0.000	21:05	
Air Blank	0.000	21:07	
Subject Sample	# 2 0.083	21:09	
Breath Volume	1.152	21:09	
Air Blank	0.000	21:10	
Air Blank	0.000	21:12	
Subject Sample	# 3 0.082	21:13	
Breath Volume	2.296	21:13	
Air Blank	0.000	21:14	
Control Test	0.078	21:14	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,L,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: EASTRIDGE,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 12/09/2023 21:01:47

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.081	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.279	20:57
Breath Volume	1.628	20:57
Air Blank	0.000	20:58
Air Blank	0.000	20:59
Subject Sample # 2	0.271	21:00
Breath Volume	1.347	21:00
Air Blank	0.000	21:00
Control Test	0.081	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,L,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: WILSON,S
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: GOMEZ SANTIZ

DATE/TIME OF TEST: 12/09/2023 22:32:57

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.081	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.199	22:28
Breath Volume	2.609	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:30
Subject Sample # 2	0.195	22:31
Breath Volume	2.457	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,L,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: ANTHONY,M
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: JAMES HERNANDEZ

DATE/TIME OF TEST: 12/10/2023 01:06:41

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.154	01:02
Breath Volume	1.898	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.155	01:05
Breath Volume	2.164	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ARROYO,JIANNI
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 12/10/2023 21:29:51

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:27	
Control Test	0.080	21:27	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:28	
Control Test	0.080	21:29	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,L,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: SILVER,I
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 12/15/2023 02:10:30

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.081	02:08	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILLIAMS,AUSTIN
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 12/18/2023 03:07:40

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.081	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1 0.136	03:00	Volume Not Met
Breath Volume	1.015	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 2 0.138	03:06	Volume Not Met
Breath Volume	0.144	03:06	
Air Blank	0.000	03:06	
Control Test	0.082	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILLIAMS,AUSTIN
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 12/18/2023 03:17:16

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.150	03:12
Breath Volume	1.593	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.146	03:15
Breath Volume	1.265	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 12/19/2023 09:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: HOBSON,B
ARRESTING OFFICER AGENCY: Umatilla PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 12/20/2023 18:45:54

Diagnostics Check	OK	18:43	
Air Blank	0.000	18:43	
Control Test	0.082	18:43	
Air Blank	0.000	18:44	
Subject Sample	# 1	18:44	Subject Test Refused
Breath Volume	0.000	18:44	
Air Blank	0.000	18:45	
Control Test	0.080	18:45	
Air Blank	0.000	18:45	
Diagnostics Check	OK	18:45	

Data Download Date/Time: 01/23/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: FRIEND

DATE/TIME OF TEST: 12/22/2023 00:43:21

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.081	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.322	00:38
Breath Volume	2.792	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.331	00:41
Breath Volume	2.882	00:41
Air Blank	0.000	00:42
Control Test	0.082	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 01/23/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILLIAMS,AUSTIN
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 12/22/2023 01:17:15

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.180	01:12
Breath Volume	3.589	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.179	01:15
Breath Volume	3.121	01:15
Air Blank	0.000	01:16
Control Test	0.082	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 01/23/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUBICEK,JORDAN,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ROUBICEK,JORDAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: LIGHTNER

DATE/TIME OF TEST: 12/22/2023 18:38:29

Diagnostics Check	OK	18:35	
Air Blank	0.000	18:36	
Control Test	0.081	18:36	
Air Blank	0.000	18:36	
Subject Sample	# 1	18:37	Subject Test Refused
Breath Volume	0.000	18:37	
Air Blank	0.000	18:37	
Control Test	0.081	18:37	
Air Blank	0.000	18:38	
Diagnostics Check	OK	18:38	

Data Download Date/Time: 01/23/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: PARRY,T
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GILLILAND

DATE/TIME OF TEST: 12/23/2023 00:36:44

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.082	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1 0.065	00:32	
Breath Volume	3.003	00:32	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample	# 2 0.060	00:35	
Breath Volume	2.042	00:35	
Air Blank		00:36	RFI Detect
Air Blank		00:36	RFI Detect

Data Download Date/Time: 01/23/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: VARELA,KEVIN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: NOVOTNY

DATE/TIME OF TEST: 12/27/2023 02:07:38

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.082	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 01/23/2024 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: RAPP,ERIC
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: KNOWLES

DATE/TIME OF TEST: 12/28/2023 01:46:55

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.081	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Control Test	0.082	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 01/23/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,NATASH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/30/2023 00:49:58

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.081	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 01/23/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURTZLAFF, SYDNEE, M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF, SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: ESCALANTE ESCALANTE

DATE/TIME OF TEST: 12/31/2023 03:08:20

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.082	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.222	03:03
Breath Volume	1.406	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.220	03:06
Breath Volume	3.207	03:06
Air Blank	0.000	03:07
Control Test	0.082	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 01/23/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN, DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 12/08/2023 23:56:10

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.134	23:51
Breath Volume	3.574	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.127	23:54
Breath Volume	2.960	23:54
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 12/26/2023 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOERING,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: NELL,ALEXANDER
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GONZALEZ DIAZ

DATE/TIME OF TEST: 12/13/2023 03:01:00

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.081	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1 0.155	02:56	
Breath Volume	3.816	02:56	
Air Blank	0.000	02:57	
Air Blank	0.000	02:59	
Subject Sample	# 2 0.147	02:59	
Breath Volume	3.632	02:59	
Air Blank	0.000	03:00	
Control Test	0.073	03:00	Control Outside Tolerance
Air Blank	0.000	03:01	

Data Download Date/Time: 12/26/2023 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: HILLIKER, COREY
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/23/2023 04:26:23

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.101	04:22
Breath Volume	1.535	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.098	04:24
Breath Volume	1.328	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 12/26/2023 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI, H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: HERMANSON

DATE/TIME OF TEST: 12/01/2023 22:03:43

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.082	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.180	21:58
Breath Volume	1.464	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.165	22:02
Breath Volume	2.187	22:02
Air Blank	0.000	22:02
Control Test	0.077	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 01/03/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 12/02/2023 02:45:44

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.081	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1 0.158	02:35	Volume Not Met
Breath Volume	0.605	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 2 0.166	02:38	
Breath Volume	1.324	02:38	
Air Blank	0.000	02:39	
Air Blank	0.000	02:40	
Subject Sample	# 3 0.144	02:44	Volume Not Met
Breath Volume	0.328	02:44	
Air Blank	0.000	02:44	
Control Test	0.080	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 01/03/2024 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: KEENS-DUMAS

DATE/TIME OF TEST: 12/03/2023 00:01:48

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.081	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 01/03/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: HERNANDEZ MORALES

DATE/TIME OF TEST: 12/03/2023 03:15:28

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.153	03:10
Breath Volume	2.449	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.160	03:13
Breath Volume	1.585	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 01/03/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: JIMENEZ MARTINEZ

DATE/TIME OF TEST: 12/03/2023 03:30:47

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.160	03:26
Breath Volume	2.179	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.154	03:29
Breath Volume	1.230	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 01/03/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANAYA,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/03/2023 04:06:51

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.218	04:02
Breath Volume	2.730	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.207	04:05
Breath Volume	2.714	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: LOPEZ DOMINGO

DATE/TIME OF TEST: 12/03/2023 04:21:26

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.050	04:16
Breath Volume	1.808	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.048	04:19
Breath Volume	3.046	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: DORANTES LILLO

DATE/TIME OF TEST: 12/03/2023 04:30:26

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:28	Slope Not Met
Breath Volume	2.027	04:28	
Air Blank	0.000	04:29	
Control Test	0.080	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: RASSO ARROYO

DATE/TIME OF TEST: 12/03/2023 04:47:47

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.165	04:40
Breath Volume	1.156	04:40
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.190	04:43
Breath Volume	1.546	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 3	0.185	04:46
Breath Volume	1.839	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: DORANTES LILLO

DATE/TIME OF TEST: 12/03/2023 04:57:32

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.305	04:52
Breath Volume	1.972	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.300	04:55
Breath Volume	2.351	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: MORENO SOTOLONGO

DATE/TIME OF TEST: 12/03/2023 06:44:52

Diagnostics Check	OK	06:35
Air Blank	0.000	06:35
Control Test	0.080	06:35
Air Blank	0.000	06:36
Subject Sample # 1	0.221	06:39
Breath Volume	1.660	06:39
Air Blank	0.000	06:40
Air Blank	0.000	06:41
Subject Sample # 2	0.241	06:43
Breath Volume	1.656	06:43
Air Blank	0.000	06:43
Control Test	0.077	06:44
Air Blank	0.000	06:44
Diagnostics Check	OK	06:44

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 12/03/2023 23:10:07

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:07	
Control Test	0.081	23:07	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BRIGHT

DATE/TIME OF TEST: 12/04/2023 03:21:22

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.081	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.238	03:12
Breath Volume	1.871	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.214	03:15
Breath Volume	1.199	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 3	0.221	03:19
Breath Volume	2.222	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGENS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PULLEN

DATE/TIME OF TEST: 12/05/2023 00:02:19

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.082	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.046	23:57
Breath Volume	1.437	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.048	00:00
Breath Volume	2.554	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: LARA SARAVIA

DATE/TIME OF TEST: 12/05/2023 21:25:39

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:12	
Control Test	0.082	21:13	
Air Blank	0.000	21:13	
Subject Sample	# 1 0.141	21:16	Slope Not Level
Breath Volume	1.273	21:16	
Air Blank	0.000	21:17	
Air Blank	0.000	21:19	
Subject Sample	# 2 0.151	21:20	
Breath Volume	1.578	21:20	
Air Blank	0.000	21:21	
Air Blank	0.000	21:23	
Subject Sample	# 3 0.160	21:24	
Breath Volume	1.230	21:24	
Air Blank	0.000	21:24	
Control Test	0.079	21:25	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/06/2023 03:23:33

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.181	03:17
Breath Volume	1.640	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.170	03:21
Breath Volume	1.437	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 12/06/2023 05:26:33

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.081	05:19
Air Blank	0.000	05:20
Subject Sample # 1	0.234	05:21
Breath Volume	1.609	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.214	05:24
Breath Volume	2.574	05:24
Air Blank	0.000	05:25
Control Test	0.079	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: TRACY

DATE/TIME OF TEST: 12/08/2023 00:01:17

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.155	23:56
Breath Volume	3.722	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 2	0.148	23:59
Breath Volume	1.183	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CLUTE

DATE/TIME OF TEST: 12/08/2023 04:07:17

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.176	04:02
Breath Volume	2.148	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.171	04:05
Breath Volume	2.113	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: BARRINER

DATE/TIME OF TEST: 12/08/2023 05:01:28

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.082	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.152	04:55
Breath Volume	1.289	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.151	04:59
Breath Volume	1.417	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 12/08/2023 20:08:14

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.081	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.337	20:02
Breath Volume	1.648	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.317	20:06
Breath Volume	1.246	20:06
Air Blank	0.000	20:07
Control Test	0.080	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HERNANDEZ CORRALES

DATE/TIME OF TEST: 12/10/2023 01:29:40

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.081	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:22	No Sample Provided
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample	# 2	01:28	No Sample Provided
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.080	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: THORSEN,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: CASTELLON

DATE/TIME OF TEST: 12/10/2023 02:50:10

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.182	02:44
Breath Volume	2.023	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.194	02:48
Breath Volume	1.421	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: AGOSTINI

DATE/TIME OF TEST: 12/10/2023 04:15:59

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:12	
Control Test	0.081	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:15	
Control Test	0.080	04:15	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: TAMAYO AGUILERA

DATE/TIME OF TEST: 12/10/2023 23:26:31

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.081	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.202	23:21
Breath Volume	2.058	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.214	23:24
Breath Volume	1.308	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: ROCHA SEGURA

DATE/TIME OF TEST: 12/11/2023 06:19:13

Diagnostics Check	OK	06:12
Air Blank	0.000	06:12
Control Test	0.082	06:12
Air Blank	0.000	06:13
Subject Sample # 1	0.203	06:14
Breath Volume	1.265	06:14
Air Blank	0.000	06:15
Air Blank	0.000	06:16
Subject Sample # 2	0.210	06:17
Breath Volume	1.761	06:17
Air Blank	0.000	06:18
Control Test	0.080	06:18
Air Blank	0.000	06:19
Diagnostics Check	OK	06:19

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAPP, D
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/11/2023 20:41:35

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.081	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.000	20:37
Breath Volume	2.011	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.000	20:40
Breath Volume	2.132	20:40
Air Blank	0.000	20:40
Control Test	0.081	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MEEKS

DATE/TIME OF TEST: 12/13/2023 04:15:19

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.081	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.131	04:10
Breath Volume	2.601	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.133	04:13
Breath Volume	2.250	04:13
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BROWNLEE,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 12/14/2023 16:23:24

Diagnostics Check	OK	16:11	
Air Blank	0.000	16:12	
Control Test	0.082	16:12	
Air Blank	0.000	16:13	
Subject Sample	# 1 0.000	16:16	Volume Not Met
Breath Volume	0.730	16:16	
Air Blank	0.000	16:16	
Air Blank	0.000	16:18	
Subject Sample	# 2 0.000	16:22	Volume Not Met
Breath Volume	0.503	16:22	
Air Blank	0.000	16:22	
Control Test	0.081	16:22	
Air Blank	0.000	16:23	
Diagnostics Check	OK	16:23	

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BROWNLEE,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 12/14/2023 16:37:29

Diagnostics Check	OK	16:26	
Air Blank	0.000	16:26	
Control Test	0.080	16:26	
Air Blank	0.000	16:27	
Subject Sample	# 1 0.000	16:30	Volume Not Met
Breath Volume	0.015	16:30	
Air Blank	0.000	16:30	
Air Blank	0.000	16:32	
Subject Sample	# 2 0.000	16:36	Volume Not Met
Breath Volume	0.757	16:36	
Air Blank	0.000	16:36	
Control Test	0.081	16:36	
Air Blank	0.000	16:37	
Diagnostics Check	OK	16:37	

Data Download Date/Time: 01/03/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CROSBY, J
ARRESTING OFFICER AGENCY: Seminole Res (Hills)
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 12/14/2023 21:38:03

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:30	
Control Test	0.081	21:30	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:34	No Sample Provided
Breath Volume	0.000	21:34	
Air Blank	0.000	21:34	
Air Blank	0.000	21:36	
Subject Sample	# 2	21:36	Subject Test Refused
Breath Volume	0.000	21:36	
Air Blank	0.000	21:37	
Control Test	0.081	21:37	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: HASTING

DATE/TIME OF TEST: 12/14/2023 23:01:19

Diagnostics Check	OK	23:00	
Air Blank		23:00	RFI Detect
Air Blank	0.000	23:01	

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: HASTING

DATE/TIME OF TEST: 12/14/2023 23:10:19

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.081	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.212	23:05
Breath Volume	3.578	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.217	23:08
Breath Volume	3.062	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GIANNARIS

DATE/TIME OF TEST: 12/16/2023 00:32:28

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.096	00:28
Breath Volume	3.726	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.092	00:30
Breath Volume	2.304	00:30
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: YARA CORDOBA

DATE/TIME OF TEST: 12/16/2023 01:46:04

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.157	01:41
Breath Volume	1.304	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.150	01:44
Breath Volume	1.207	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:46

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: TREMBLAY

DATE/TIME OF TEST: 12/16/2023 02:25:29

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.214	02:20
Breath Volume	1.859	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.224	02:23
Breath Volume	1.253	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:07

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 12/16/2023 15:39:32

Diagnostics Check	OK	15:28	
Air Blank	0.000	15:28	
Control Test	0.080	15:28	
Air Blank	0.000	15:29	
Subject Sample	# 1 0.151	15:32	Volume Not Met
Breath Volume	0.175	15:32	
Air Blank	0.000	15:33	
Air Blank	0.000	15:35	
Subject Sample	# 2	15:38	No Sample Provided
Breath Volume	0.000	15:38	
Air Blank	0.000	15:38	
Control Test	0.078	15:38	
Air Blank	0.000	15:39	
Diagnostics Check	OK	15:39	

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 12/16/2023 16:02:53

Diagnostics Check	OK	16:00	
Air Blank	0.000	16:00	
Control Test	0.080	16:00	
Air Blank	0.000	16:01	
Subject Sample	# 1	16:01	Subject Test Refused
Breath Volume	0.000	16:01	
Air Blank	0.000	16:01	
Control Test	0.080	16:02	
Air Blank	0.000	16:02	
Diagnostics Check	OK	16:02	

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HEINTZ,J
ARRESTING OFFICER AGENCY: Seminole Res (Hills)
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:14

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 12/17/2023 08:42:07

Diagnostics Check	OK	08:35
Air Blank	0.000	08:35
Control Test	0.082	08:35
Air Blank	0.000	08:36
Subject Sample # 1	0.153	08:37
Breath Volume	2.378	08:37
Air Blank	0.000	08:37
Air Blank	0.000	08:39
Subject Sample # 2	0.148	08:40
Breath Volume	2.441	08:40
Air Blank	0.000	08:41
Control Test	0.081	08:41
Air Blank	0.000	08:42
Diagnostics Check	OK	08:42

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 12/17/2023 20:38:33

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:35	
Control Test	0.081	20:35	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	
Control Test	0.081	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VELEZ YAUCE

DATE/TIME OF TEST: 12/18/2023 03:52:36

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.150	03:47
Breath Volume	4.027	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.147	03:50
Breath Volume	2.535	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 12/19/2023 00:02:50

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.082	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.145	23:58
Breath Volume	2.121	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.155	00:01
Breath Volume	3.425	00:01
Air Blank	0.000	00:01
Control Test	0.081	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI, H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: CHILLURA

DATE/TIME OF TEST: 12/19/2023 20:44:07

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.082	20:38
Air Blank	0.000	20:38
Subject Sample # 1	0.060	20:39
Breath Volume	2.542	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:41
Subject Sample # 2	0.059	20:42
Breath Volume	2.660	20:42
Air Blank	0.000	20:43
Control Test	0.081	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NONE,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 12/20/2023 11:00:02

Diagnostics Check	OK	10:53
Air Blank	0.000	10:54
Control Test	0.082	10:54
Air Blank	0.000	10:55
Subject Sample # 1	0.000	10:55
Breath Volume	2.113	10:55
Air Blank	0.000	10:56
Air Blank	0.000	10:58
Subject Sample # 2	0.000	10:58
Breath Volume	1.367	10:58
Air Blank	0.000	10:59
Control Test	0.081	10:59
Air Blank	0.000	10:59
Diagnostics Check	OK	11:00

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/20/2023 19:33:33

Diagnostics Check	OK	19:26
Air Blank	0.000	19:27
Control Test	0.082	19:27
Air Blank	0.000	19:28
Subject Sample # 1	0.275	19:28
Breath Volume	1.675	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.273	19:31
Breath Volume	1.496	19:31
Air Blank	0.000	19:32
Control Test	0.080	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIS, K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 12/20/2023 21:10:55

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.081	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.085	21:05
Breath Volume	1.761	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.082	21:09
Breath Volume	1.742	21:09
Air Blank	0.000	21:09
Control Test	0.080	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARQUEZ DIAZ

DATE/TIME OF TEST: 12/21/2023 04:29:08

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.081	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.148	04:24
Breath Volume	1.203	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.145	04:27
Breath Volume	1.332	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KELLY,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: CAPELLAN

DATE/TIME OF TEST: 12/21/2023 16:01:58

Diagnostics Check	OK	15:55
Air Blank	0.000	15:55
Control Test	0.081	15:55
Air Blank	0.000	15:56
Subject Sample # 1	0.202	15:57
Breath Volume	3.218	15:57
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.202	16:00
Breath Volume	2.535	16:00
Air Blank	0.000	16:01
Control Test	0.080	16:01
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: DE LOS SANTOS DIAZ

DATE/TIME OF TEST: 12/21/2023 23:51:58

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.095	23:47
Breath Volume	1.136	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.093	23:50
Breath Volume	1.324	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TABOADA,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: POUNDS

DATE/TIME OF TEST: 12/22/2023 18:38:13

Diagnostics Check	OK	18:31
Air Blank	0.000	18:32
Control Test	0.081	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.228	18:33
Breath Volume	2.625	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.235	18:36
Breath Volume	2.472	18:36
Air Blank	0.000	18:37
Control Test	0.080	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: OLIVER III

DATE/TIME OF TEST: 12/23/2023 06:39:27

Diagnostics Check	OK	06:32
Air Blank	0.000	06:32
Control Test	0.081	06:33
Air Blank	0.000	06:33
Subject Sample # 1	0.079	06:34
Breath Volume	1.328	06:34
Air Blank	0.000	06:35
Air Blank	0.000	06:36
Subject Sample # 2	0.082	06:37
Breath Volume	1.699	06:37
Air Blank	0.000	06:38
Control Test	0.080	06:38
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: CARDOSO CRUZ

DATE/TIME OF TEST: 12/23/2023 20:48:50

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.082	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.176	20:44
Breath Volume	1.941	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.183	20:47
Breath Volume	2.035	20:47
Air Blank	0.000	20:47
Control Test	0.080	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/24/2023 00:29:24

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.081	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.199	00:24
Breath Volume	1.238	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.209	00:27
Breath Volume	2.472	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 01/03/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Temple Terrace PD
ARRESTING OFFICER: LORENZO,E
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SPEIGHTS

DATE/TIME OF TEST: 12/24/2023 01:17:15

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.000	01:12
Breath Volume	2.593	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.000	01:15
Breath Volume	1.460	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: CROWTHERS

DATE/TIME OF TEST: 12/24/2023 21:02:40

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.081	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.178	20:55
Breath Volume	1.882	20:55
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.157	20:57
Breath Volume	1.105	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 3	0.170	21:00
Breath Volume	1.441	21:00
Air Blank	0.000	21:01
Control Test	0.081	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: FINDLAY

DATE/TIME OF TEST: 12/25/2023 05:42:02

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.082	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.161	05:36
Breath Volume	2.207	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.166	05:40
Breath Volume	1.589	05:40
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:42

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CERRO REBOLLEDO

DATE/TIME OF TEST: 12/26/2023 03:09:44

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.224	03:05
Breath Volume	1.640	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.241	03:08
Breath Volume	1.503	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: LEITE

DATE/TIME OF TEST: 12/26/2023 05:30:22

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:18	
Control Test	0.080	05:18	
Air Blank	0.000	05:19	
Subject Sample	# 1	05:22	No Sample Provided
Breath Volume	0.000	05:22	
Air Blank	0.000	05:22	
Air Blank	0.000	05:24	
Subject Sample	# 2	0.208 05:25	
Breath Volume	1.515	05:25	
Air Blank	0.000	05:26	
Air Blank	0.000	05:28	
Subject Sample	# 3	0.210 05:28	
Breath Volume	1.492	05:28	
Air Blank	0.000	05:29	
Control Test	0.080	05:29	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:30	

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: MARTINEZ CAPOTE

DATE/TIME OF TEST: 12/26/2023 21:15:31

Diagnostics Check	OK	21:06
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:08
Subject Sample # 1	0.157	21:09
Breath Volume	1.796	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample # 2	0.167	21:13
Breath Volume	1.820	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 12/28/2023 00:31:41

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.049	00:27
Breath Volume	1.601	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.048	00:30
Breath Volume	1.707	00:30
Air Blank	0.000	00:30
Control Test	0.081	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SAHLMAN III

DATE/TIME OF TEST: 12/28/2023 00:50:50

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.096	00:46
Breath Volume	3.261	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.099	00:49
Breath Volume	3.531	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: LOUISSAINT

DATE/TIME OF TEST: 12/28/2023 01:27:57

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.081	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1 0.173	01:18	Volume Not Met
Breath Volume	0.503	01:18	
Air Blank	0.000	01:19	
Air Blank	0.000	01:21	
Subject Sample	# 2 0.195	01:22	
Breath Volume	2.652	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 3	01:26	Subject Test Refused
Breath Volume	0.367	01:26	
Air Blank	0.000	01:27	
Control Test	0.078	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEBOLIAC,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 12/28/2023 08:46:27

Diagnostics Check	OK	08:40
Air Blank	0.000	08:40
Control Test	0.081	08:40
Air Blank	0.000	08:41
Subject Sample # 1	0.120	08:42
Breath Volume	1.378	08:42
Air Blank	0.000	08:42
Air Blank	0.000	08:44
Subject Sample # 2	0.116	08:44
Breath Volume	2.906	08:44
Air Blank	0.000	08:45
Control Test	0.080	08:45
Air Blank	0.000	08:46
Diagnostics Check	OK	08:46

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: ALARCON RONDON

DATE/TIME OF TEST: 12/29/2023 00:19:41

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.081	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:18	
Control Test	0.081	00:19	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS, J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 12/30/2023 02:49:31

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.082	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.243	02:44
Breath Volume	3.277	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.236	02:47
Breath Volume	2.160	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: RUIZ LOPEZ

DATE/TIME OF TEST: 12/30/2023 03:34:39

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.141	03:29
Breath Volume	3.332	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.138	03:33
Breath Volume	2.398	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: JARAMILLO

DATE/TIME OF TEST: 12/30/2023 04:08:17

Diagnostics Check	OK	04:07	
Air Blank		04:07	RFI Detect
Air Blank	0.000	04:08	

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: JARAMILLO

DATE/TIME OF TEST: 12/30/2023 04:19:41

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.082	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.166	04:14
Breath Volume	1.347	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.162	04:18
Breath Volume	1.968	04:18
Air Blank	0.000	04:18
Control Test	0.081	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RUIZ,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:43

SUBJECT'S LAST NAME: WING

DATE/TIME OF TEST: 12/30/2023 13:11:49

Diagnostics Check	OK	13:05
Air Blank	0.000	13:05
Control Test	0.081	13:06
Air Blank	0.000	13:06
Subject Sample # 1	0.000	13:07
Breath Volume	1.203	13:07
Air Blank	0.000	13:07
Air Blank	0.000	13:09
Subject Sample # 2	0.000	13:10
Breath Volume	1.191	13:10
Air Blank	0.000	13:10
Control Test	0.081	13:11
Air Blank	0.000	13:11
Diagnostics Check	OK	13:11

Data Download Date/Time: 01/03/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: AMORA

DATE/TIME OF TEST: 12/16/2023 02:44:00

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.107	02:39
Breath Volume	1.542	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.111	02:42
Breath Volume	3.015	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:44

Data Download Date/Time: 01/05/2024 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 12/05/2023 21:06:49

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.080	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.215	21:02
Breath Volume	3.910	21:02
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.220	21:05
Breath Volume	4.203	21:05
Air Blank	0.000	21:05
Control Test	0.079	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 12/19/2023 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: DECARLO

DATE/TIME OF TEST: 12/10/2023 03:14:07

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.079	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1 0.378	03:08	
Breath Volume	1.183	03:08	
Air Blank	0.000	03:09	
Air Blank	0.000	03:11	
Subject Sample	# 2 0.399	03:11	
Breath Volume	1.160	03:11	
Air Blank	0.000	03:12	
Air Blank		03:13	RFI Detect
Air Blank		03:13	RFI Detect
Air Blank		03:14	RFI Detect

Data Download Date/Time: 12/19/2023 12:24