

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: MOYER

DATE/TIME OF TEST: 07/02/2024 02:11:42

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.121	02:07
Breath Volume	1.750	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.124	02:10
Breath Volume	2.656	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/08/2024 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,SHELBY,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GARDNER,SHELBY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: ALDRIDGE

DATE/TIME OF TEST: 07/04/2024 00:15:57

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.229	00:11
Breath Volume	1.617	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample # 2	0.229	00:14
Breath Volume	1.296	00:14
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 07/08/2024 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HELCHEN

DATE/TIME OF TEST: 07/12/2024 01:17:32

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.123	01:12
Breath Volume	1.328	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.123	01:15
Breath Volume	1.414	01:15
Air Blank	0.000	01:16
Control Test	0.078	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 08/12/2024 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VAUGHN,RYAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: ADMIN
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/15/2024 17:02:04

Diagnostics Check	OK	16:55
Air Blank	0.000	16:56
Control Test	0.081	16:56
Air Blank	0.000	16:56
Subject Sample # 1	0.223	16:57
Breath Volume	2.523	16:57
Air Blank	0.000	16:58
Air Blank	0.000	16:59
Subject Sample # 2	0.227	17:00
Breath Volume	2.597	17:00
Air Blank	0.000	17:01
Control Test	0.080	17:01
Air Blank	0.000	17:01
Diagnostics Check	OK	17:02

Data Download Date/Time: 08/12/2024 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/16/2024 01:57:09

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.086	01:52
Breath Volume	3.511	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.085	01:55
Breath Volume	3.792	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 08/12/2024 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: RUGGIERO

DATE/TIME OF TEST: 07/27/2024 19:39:57

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.081	19:34
Air Blank	0.000	19:34
Subject Sample # 1	0.176	19:35
Breath Volume	1.437	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:37
Subject Sample # 2	0.172	19:38
Breath Volume	1.335	19:38
Air Blank	0.000	19:39
Control Test	0.081	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 08/12/2024 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: RAMIREZ AGUSTIN

DATE/TIME OF TEST: 07/28/2024 02:48:56

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.171	02:44
Breath Volume	1.867	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.169	02:47
Breath Volume	1.863	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 08/12/2024 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: GARCIA CARTEGENA

DATE/TIME OF TEST: 07/28/2024 23:11:09

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.082	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.127	23:06
Breath Volume	1.210	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.125	23:09
Breath Volume	1.511	23:09
Air Blank	0.000	23:10
Control Test	0.081	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 08/12/2024 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LORING

DATE/TIME OF TEST: 07/03/2024 02:37:26

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.000	02:33
Breath Volume	1.367	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.000	02:36
Breath Volume	1.179	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 07/08/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GALINDO FUNEZ

DATE/TIME OF TEST: 07/06/2024 01:23:54

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.111	01:19
Breath Volume	2.855	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:21
Subject Sample # 2	0.108	01:22
Breath Volume	2.140	01:22
Air Blank	0.000	01:22
Control Test	0.077	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 07/08/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 07/07/2024 18:06:49

Diagnostics Check	OK	18:00
Air Blank	0.000	18:00
Control Test	0.080	18:01
Air Blank	0.000	18:01
Subject Sample # 1	0.211	18:02
Breath Volume	3.230	18:02
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.207	18:05
Breath Volume	3.433	18:05
Air Blank	0.000	18:05
Control Test	0.078	18:06
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

Data Download Date/Time: 07/08/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SANCHEZ PULIDO

DATE/TIME OF TEST: 07/11/2024 00:57:21

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.166	00:52
Breath Volume	1.769	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.170	00:55
Breath Volume	1.753	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ZAVALA XIQUIN

DATE/TIME OF TEST: 07/11/2024 04:19:54

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.191	04:15
Breath Volume	2.207	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:17
Subject Sample # 2	0.201	04:18
Breath Volume	3.031	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: VILLATORO MOLINA

DATE/TIME OF TEST: 07/12/2024 01:35:04

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.176	01:30
Breath Volume	1.203	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.194	01:33
Breath Volume	1.882	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:35

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 07/12/2024 04:46:44

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.079	04:38	
Air Blank	0.000	04:39	
Subject Sample	# 1 0.166	04:42	Volume Not Met
Breath Volume	0.730	04:42	
Air Blank	0.000	04:43	
Air Blank	0.000	04:45	
Subject Sample	# 2	04:45	Subject Test Refused
Breath Volume	0.000	04:45	
Air Blank	0.000	04:45	
Control Test	0.078	04:46	
Air Blank	0.000	04:46	
Diagnostics Check	OK	04:46	

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,ROBERT,T
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: HUDDELSON,CHASE
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/14/2024 01:06:50

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.250	01:02
Breath Volume	1.734	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:04
Subject Sample # 2	0.236	01:05
Breath Volume	1.125	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HUNTZINGER

DATE/TIME OF TEST: 07/15/2024 21:26:26

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.000	21:21
Breath Volume	1.207	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.000	21:25
Breath Volume	1.460	21:25
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: COREY

DATE/TIME OF TEST: 07/17/2024 04:52:46

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.193	04:47
Breath Volume	1.601	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:49
Subject Sample # 2	0.191	04:51
Breath Volume	2.121	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/18/2024 16:40:31

Diagnostics Check	OK	16:33
Air Blank	0.000	16:33
Control Test	0.081	16:34
Air Blank	0.000	16:34
Subject Sample # 1	0.084	16:35
Breath Volume	1.734	16:35
Air Blank	0.000	16:36
Air Blank	0.000	16:38
Subject Sample # 2	0.081	16:38
Breath Volume	2.238	16:38
Air Blank	0.000	16:39
Control Test	0.080	16:39
Air Blank	0.000	16:40
Diagnostics Check	OK	16:40

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FARES,RAMEZ
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: VASQUEZ-VELASQUEZ

DATE/TIME OF TEST: 07/25/2024 02:46:52

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.129	02:42
Breath Volume	1.175	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample # 2	0.130	02:45
Breath Volume	2.402	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: TAYLOR,J
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/25/2024 03:10:37

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.207	03:06
Breath Volume	1.859	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.196	03:08
Breath Volume	2.343	03:08
Air Blank	0.000	03:09
Control Test	0.078	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DEANS

DATE/TIME OF TEST: 07/25/2024 03:28:50

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.186	03:23
Breath Volume	1.636	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.185	03:27
Breath Volume	2.472	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: KLOC

DATE/TIME OF TEST: 07/25/2024 23:41:52

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:30	
Control Test	0.080	23:30	
Air Blank	0.000	23:31	
Subject Sample	# 1 0.138	23:34	Volume Not Met
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2 0.130	23:40	Volume Not Met
Breath Volume	0.777	23:40	
Air Blank	0.000	23:40	
Control Test	0.077	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: KLOC

DATE/TIME OF TEST: 07/25/2024 23:48:21

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.078	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:46	Subject Test Refused
Breath Volume	0.636	23:46	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: REYES SANCHEZ

DATE/TIME OF TEST: 07/28/2024 01:09:01

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.093	01:04
Breath Volume	1.226	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.093	01:07
Breath Volume	1.160	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:09

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GARCIA MELCHOR

DATE/TIME OF TEST: 07/28/2024 03:51:29

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.111	03:46
Breath Volume	1.839	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.124	03:49
Breath Volume	3.300	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MARTIN ROJAS

DATE/TIME OF TEST: 07/30/2024 01:32:14

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.167	01:27
Breath Volume	2.003	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.169	01:30
Breath Volume	1.531	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: TRAYLOR

DATE/TIME OF TEST: 07/30/2024 04:08:52

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.169	04:03
Breath Volume	1.847	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.171	04:07
Breath Volume	1.617	04:07
Air Blank	0.000	04:07
Control Test	0.076	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: DACEY

DATE/TIME OF TEST: 07/30/2024 05:05:52

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:03	
Control Test	0.080	05:03	
Air Blank	0.000	05:03	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Control Test	0.080	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: TAYLOR,JONATHAN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: CASTILLO CALLES

DATE/TIME OF TEST: 07/30/2024 20:55:03

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.080	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.217	20:50
Breath Volume	2.578	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:52
Subject Sample # 2	0.211	20:53
Breath Volume	3.082	20:53
Air Blank	0.000	20:54
Control Test	0.078	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:55

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FARES,RAMEZ
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ROMIH

DATE/TIME OF TEST: 07/30/2024 21:36:50

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:30
Subject Sample # 1	0.174	21:32
Breath Volume	1.984	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:34
Subject Sample # 2	0.170	21:35
Breath Volume	1.777	21:35
Air Blank	0.000	21:35
Control Test	0.076	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 08/12/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 07/04/2024 02:49:42

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.147	02:44
Breath Volume	2.187	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.145	02:48
Breath Volume	2.136	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 08/06/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: DUCHARME

DATE/TIME OF TEST: 07/04/2024 04:32:27

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.141	04:27
Breath Volume	1.136	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.151	04:30
Breath Volume	1.242	04:30
Air Blank	0.000	04:31
Control Test	0.078	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 08/06/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: KIRLEW JR

DATE/TIME OF TEST: 07/04/2024 23:21:58

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.068	23:17
Breath Volume	1.625	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.067	23:20
Breath Volume	1.375	23:20
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 08/06/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: BODDEN

DATE/TIME OF TEST: 07/05/2024 04:48:50

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.057	04:44
Breath Volume	2.113	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample # 2	0.053	04:47
Breath Volume	1.398	04:47
Air Blank	0.000	04:47
Control Test	0.080	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 08/06/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 07/05/2024 21:50:49

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:44	
Control Test	0.080	21:45	
Air Blank	0.000	21:45	
Subject Sample # 1	0.132	21:46	
Breath Volume	2.722	21:46	
Air Blank	0.000	21:47	
Air Blank	0.000	21:48	
Subject Sample # 2		21:49	Slope Not Met
Breath Volume	2.019	21:49	
Air Blank	0.000	21:49	
Control Test	0.000	21:50	Control Outside Tolerance
Air Blank	0.000	21:50	

Data Download Date/Time: 08/06/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 07/05/2024 22:19:27

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.113	22:14
Breath Volume	2.714	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.121	22:17
Breath Volume	2.781	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 08/06/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: MCAFFEE

DATE/TIME OF TEST: 07/06/2024 05:36:35

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Subject Sample # 1	0.341	05:31
Breath Volume	2.589	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.346	05:34
Breath Volume	1.785	05:34
Air Blank	0.000	05:35
Control Test	0.079	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 08/06/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: RIQUENES GINARTE

DATE/TIME OF TEST: 07/07/2024 00:23:22

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.200	00:18
Breath Volume	1.316	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.184	00:21
Breath Volume	1.640	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 08/06/2024 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CARMONA OTRIZ

DATE/TIME OF TEST: 07/07/2024 02:03:51

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.239	01:58
Breath Volume	2.195	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.238	02:02
Breath Volume	1.910	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: QUARTERMAN JR

DATE/TIME OF TEST: 07/07/2024 02:25:05

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.079	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:23	Interferent Detect
Breath Volume	3.777	02:23	
Air Blank	0.000	02:24	
Control Test	0.051	02:24	Interferent Detect
Air Blank	0.000	02:25	

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: QUARTERMAN JR

DATE/TIME OF TEST: 07/07/2024 02:34:11

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.161	02:29
Breath Volume	4.093	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.165	02:32
Breath Volume	3.695	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: TORRES LOPEZ

DATE/TIME OF TEST: 07/07/2024 05:05:52

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.080	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.162	05:01
Breath Volume	2.550	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.165	05:04
Breath Volume	2.328	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: LOPEZ JIMENEZ

DATE/TIME OF TEST: 07/07/2024 19:03:17

Diagnostics Check	OK	18:56
Air Blank	0.000	18:57
Control Test	0.080	18:57
Air Blank	0.000	18:57
Subject Sample # 1	0.275	18:58
Breath Volume	3.257	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.270	19:01
Breath Volume	3.355	19:01
Air Blank	0.000	19:02
Control Test	0.080	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: LANGSTON

DATE/TIME OF TEST: 07/07/2024 22:10:56

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.180	22:06
Breath Volume	2.277	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.175	22:09
Breath Volume	1.761	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: SALAZAR-SANCHEZ

DATE/TIME OF TEST: 07/07/2024 23:09:18

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.209	23:04
Breath Volume	1.777	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:06
Subject Sample # 2	0.196	23:07
Breath Volume	1.226	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: EARLE

DATE/TIME OF TEST: 07/08/2024 22:34:41

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.080	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 1 0.243	22:30	
Breath Volume	1.519	22:30	
Air Blank	0.000	22:30	
Air Blank	0.000	22:32	
Subject Sample	# 2	22:33	Slope Not Met
Breath Volume	0.726	22:33	
Air Blank	0.000	22:33	
Control Test	0.081	22:34	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: EARLE

DATE/TIME OF TEST: 07/08/2024 23:06:21

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.224	23:01
Breath Volume	1.441	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.231	23:04
Breath Volume	1.660	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: PATINO

DATE/TIME OF TEST: 07/10/2024 01:35:01

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.081	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.081	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: MONTEMARANO

DATE/TIME OF TEST: 07/10/2024 05:50:47

Diagnostics Check	OK	05:44
Air Blank	0.000	05:45
Control Test	0.081	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.138	05:46
Breath Volume	2.820	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.138	05:49
Breath Volume	2.464	05:49
Air Blank	0.000	05:49
Control Test	0.080	05:50
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,LELA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: SIMMERS

DATE/TIME OF TEST: 07/10/2024 17:54:16

Diagnostics Check	OK	17:48
Air Blank	0.000	17:48
Control Test	0.080	17:48
Air Blank	0.000	17:49
Subject Sample # 1	0.151	17:49
Breath Volume	2.406	17:49
Air Blank	0.000	17:50
Air Blank	0.000	17:52
Subject Sample # 2	0.154	17:52
Breath Volume	2.277	17:52
Air Blank	0.000	17:53
Control Test	0.081	17:53
Air Blank	0.000	17:54
Diagnostics Check	OK	17:54

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/11/2024 01:05:53

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:04	Interferent Detect
Breath Volume	1.328	01:04	
Air Blank	0.000	01:04	
Control Test	0.081	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/11/2024 01:12:06

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:08	
Control Test	0.079	01:08	
Air Blank	0.000	01:09	
Subject Sample	# 1	01:10	Interferent Detect
Breath Volume	1.273	01:10	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARRASCO,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: SANCHEZ DAMIAN

DATE/TIME OF TEST: 07/11/2024 03:36:31

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.081	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1 0.184	03:24	
Breath Volume	1.406	03:24	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2 0.158	03:30	Slope Not Level
Breath Volume	1.171	03:30	
Air Blank	0.000	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 3 0.190	03:34	
Breath Volume	2.265	03:34	
Air Blank	0.000	03:35	
Control Test	0.078	03:35	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: LOPEZ CARMONA

DATE/TIME OF TEST: 07/11/2024 05:41:36

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.080	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.163	05:34
Breath Volume	1.453	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.169	05:39
Breath Volume	1.339	05:39
Air Blank	0.000	05:40
Control Test	0.079	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/11/2024 06:59:47

Diagnostics Check	OK	06:53
Air Blank	0.000	06:54
Control Test	0.079	06:54
Air Blank	0.000	06:54
Subject Sample # 1	0.000	06:55
Breath Volume	2.207	06:55
Air Blank	0.000	06:55
Air Blank	0.000	06:57
Subject Sample # 2	0.000	06:58
Breath Volume	1.812	06:58
Air Blank	0.000	06:58
Control Test	0.081	06:59
Air Blank	0.000	06:59
Diagnostics Check	OK	06:59

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: RUIZ NUNEZ

DATE/TIME OF TEST: 07/12/2024 01:06:37

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.080	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1 0.166	00:56	Volume Not Met
Breath Volume	0.152	00:56	
Air Blank	0.000	00:57	
Air Blank	0.000	00:59	
Subject Sample	# 2 0.169	01:01	
Breath Volume	1.566	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 3 0.164	01:04	
Breath Volume	1.183	01:04	
Air Blank	0.000	01:05	
Control Test	0.078	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: STEFAN

DATE/TIME OF TEST: 07/12/2024 04:33:00

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.164	04:28
Breath Volume	2.019	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:30
Subject Sample # 2	0.167	04:31
Breath Volume	1.722	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:33

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: TANKERSLEY JR

DATE/TIME OF TEST: 07/12/2024 23:11:10

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.081	23:02	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:06	No Sample Provided
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Air Blank	0.000	23:08	
Subject Sample	# 2	23:09	Subject Test Refused
Breath Volume	0.000	23:09	
Air Blank	0.000	23:10	
Control Test	0.080	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 08/06/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: HERNANDEZ ALVAREZ

DATE/TIME OF TEST: 07/13/2024 02:10:19

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.142	02:05
Breath Volume	2.496	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.139	02:08
Breath Volume	2.046	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: TOMICH

DATE/TIME OF TEST: 07/13/2024 03:36:39

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.144	03:32
Breath Volume	2.679	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.144	03:35
Breath Volume	1.613	03:35
Air Blank	0.000	03:35
Control Test	0.081	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ESQUERRA HURTADO

DATE/TIME OF TEST: 07/13/2024 03:54:03

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.081	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Control Test	0.081	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: PATISAUL

DATE/TIME OF TEST: 07/13/2024 04:38:18

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.093	04:33
Breath Volume	1.878	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.089	04:36
Breath Volume	2.585	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/13/2024 05:13:56

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.161	05:09
Breath Volume	2.664	05:09
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.157	05:12
Breath Volume	2.617	05:12
Air Blank	0.000	05:13
Control Test	0.081	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HACKETT,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:44

SUBJECT'S LAST NAME: CANTRELL

DATE/TIME OF TEST: 07/13/2024 12:08:29

Diagnostics Check	OK	12:04	
Air Blank	0.000	12:04	
Control Test	0.081	12:05	
Air Blank	0.000	12:05	
Subject Sample	# 1	12:06	Interferent Detect
Breath Volume	1.921	12:06	
Air Blank	0.000	12:07	
Control Test	0.081	12:07	
Air Blank	0.000	12:08	
Diagnostics Check	OK	12:08	

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS,JAY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 07/13/2024 18:31:18

Diagnostics Check	OK	18:24
Air Blank	0.000	18:25
Control Test	0.081	18:25
Air Blank	0.000	18:26
Subject Sample # 1	0.094	18:26
Breath Volume	1.261	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.094	18:29
Breath Volume	1.949	18:29
Air Blank	0.000	18:30
Control Test	0.081	18:30
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BOTHE,T
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: BUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 07/13/2024 22:30:03

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:28	Subject Test Refused
Breath Volume	0.000	22:28	
Air Blank	0.000	22:29	
Control Test	0.081	22:29	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: SPRADLEY

DATE/TIME OF TEST: 07/13/2024 23:06:24

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.080	23:01
Breath Volume	2.085	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.076	23:04
Breath Volume	2.089	23:04
Air Blank	0.000	23:05
Control Test	0.081	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUG,Z
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/14/2024 00:26:45

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.017	00:22
Breath Volume	2.714	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.017	00:25
Breath Volume	1.847	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: JIMENEZ MOLINA

DATE/TIME OF TEST: 07/14/2024 03:13:23

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.242	03:08
Breath Volume	2.312	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.240	03:11
Breath Volume	2.109	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 07/14/2024 04:16:45

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.089	04:12
Breath Volume	2.167	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.086	04:15
Breath Volume	1.953	04:15
Air Blank	0.000	04:15
Control Test	0.081	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 07/14/2024 05:38:37

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.115	05:34
Breath Volume	1.394	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.127	05:37
Breath Volume	1.558	05:37
Air Blank	0.000	05:37
Control Test	0.080	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUG,Z
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: ADDERLEY

DATE/TIME OF TEST: 07/14/2024 06:54:32

Diagnostics Check	OK	06:47
Air Blank	0.000	06:48
Control Test	0.080	06:48
Air Blank	0.000	06:49
Subject Sample # 1	0.154	06:50
Breath Volume	1.304	06:50
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample # 2	0.149	06:52
Breath Volume	1.312	06:52
Air Blank	0.000	06:53
Control Test	0.080	06:53
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: COMM VEHIC
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: FUMERO HERNANDEZ

DATE/TIME OF TEST: 07/14/2024 20:38:41

Diagnostics Check	OK	20:32
Air Blank	0.000	20:32
Control Test	0.081	20:32
Air Blank	0.000	20:33
Subject Sample # 1	0.245	20:33
Breath Volume	2.941	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.255	20:37
Breath Volume	2.390	20:37
Air Blank	0.000	20:37
Control Test	0.081	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: FARABEE

DATE/TIME OF TEST: 07/14/2024 23:29:03

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.082	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.119	23:24
Breath Volume	3.453	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.124	23:27
Breath Volume	2.375	23:27
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:29

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: VERA BALLESTERO

DATE/TIME OF TEST: 07/15/2024 03:41:46

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.263	03:36
Breath Volume	1.976	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.259	03:40
Breath Volume	2.585	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: MADLOCK

DATE/TIME OF TEST: 07/15/2024 05:26:02

Diagnostics Check	OK	05:16
Air Blank	0.000	05:17
Control Test	0.081	05:17
Air Blank	0.000	05:18
Subject Sample # 1	0.086	05:20
Breath Volume	1.222	05:20
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.089	05:24
Breath Volume	2.128	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:25
Diagnostics Check	OK	05:26

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CATON

DATE/TIME OF TEST: 07/15/2024 23:56:23

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.081	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.061	23:51
Breath Volume	2.304	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.058	23:54
Breath Volume	1.460	23:54
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: STEWART, JR

DATE/TIME OF TEST: 07/16/2024 00:53:28

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.082	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.000	00:49
Breath Volume	3.242	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.000	00:52
Breath Volume	3.218	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MAROCOLA,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 07/16/2024 22:24:01

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.082	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.152	22:19
Breath Volume	1.980	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:21
Subject Sample # 2	0.141	22:22
Breath Volume	1.808	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:24

Data Download Date/Time: 08/06/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/18/2024 01:23:56

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.082	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.131	01:18
Breath Volume	1.394	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.127	01:22
Breath Volume	2.007	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/18/2024 22:00:37

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.082	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.293	21:56
Breath Volume	2.285	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.297	21:58
Breath Volume	2.531	21:58
Air Blank	0.000	21:59
Control Test	0.080	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 07/19/2024 04:01:26

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.000	03:57
Breath Volume	1.457	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.000	04:00
Breath Volume	1.859	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: DIAZ GOMEZ

DATE/TIME OF TEST: 07/19/2024 23:00:35

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.087	22:55
Breath Volume	2.265	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.086	22:59
Breath Volume	1.773	22:59
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BETANCOURT SILVA

DATE/TIME OF TEST: 07/20/2024 02:24:16

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.183	02:19
Breath Volume	1.640	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.192	02:22
Breath Volume	1.269	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: RAMOS RUIZ

DATE/TIME OF TEST: 07/20/2024 04:37:40

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.207	04:33
Breath Volume	1.136	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.216	04:36
Breath Volume	1.640	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 07/20/2024 04:54:10

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.080	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.189	04:49
Breath Volume	1.929	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.201	04:52
Breath Volume	1.917	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BREWTON

DATE/TIME OF TEST: 07/20/2024 05:26:57

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.072	05:22
Breath Volume	1.703	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.070	05:25
Breath Volume	1.679	05:25
Air Blank	0.000	05:26
Control Test	0.080	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEWART, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: JASSO HERNANDEZ

DATE/TIME OF TEST: 07/20/2024 19:08:26

Diagnostics Check	OK	19:02
Air Blank	0.000	19:02
Control Test	0.080	19:02
Air Blank	0.000	19:03
Subject Sample # 1	0.124	19:03
Breath Volume	1.453	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.112	19:06
Breath Volume	1.312	19:06
Air Blank	0.000	19:07
Control Test	0.080	19:07
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MASOCOLA, D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: CEPERO LABRADA

DATE/TIME OF TEST: 07/21/2024 02:45:32

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.205	02:40
Breath Volume	2.394	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.206	02:43
Breath Volume	2.460	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GUERRA ACOSTA

DATE/TIME OF TEST: 07/21/2024 04:48:42

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.263	04:44
Breath Volume	2.761	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.270	04:47
Breath Volume	2.531	04:47
Air Blank	0.000	04:47
Control Test	0.079	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 07/21/2024 05:33:42

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.147	05:29
Breath Volume	2.140	05:29
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.150	05:32
Breath Volume	2.074	05:32
Air Blank	0.000	05:32
Control Test	0.080	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRRIZARY,KATTIE,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: DAVILA ALAMO

DATE/TIME OF TEST: 07/21/2024 06:33:59

Diagnostics Check	OK	06:27
Air Blank	0.000	06:28
Control Test	0.080	06:28
Air Blank	0.000	06:28
Subject Sample # 1	0.078	06:29
Breath Volume	2.335	06:29
Air Blank	0.000	06:30
Air Blank	0.000	06:31
Subject Sample # 2	0.075	06:32
Breath Volume	1.812	06:32
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAPPER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:16

SUBJECT'S LAST NAME: BLUBAUGH

DATE/TIME OF TEST: 07/21/2024 08:43:07

Diagnostics Check	OK	08:36
Air Blank	0.000	08:36
Control Test	0.080	08:37
Air Blank	0.000	08:37
Subject Sample # 1	0.243	08:38
Breath Volume	3.183	08:38
Air Blank	0.000	08:39
Air Blank	0.000	08:40
Subject Sample # 2	0.250	08:41
Breath Volume	3.011	08:41
Air Blank	0.000	08:42
Control Test	0.080	08:42
Air Blank	0.000	08:43
Diagnostics Check	OK	08:43

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: ESCOTO

DATE/TIME OF TEST: 07/22/2024 01:41:55

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.129	01:37
Breath Volume	1.839	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.130	01:40
Breath Volume	1.488	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FUSCO,M
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 07/22/2024 03:06:33

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.209	02:59
Breath Volume	4.558	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.181	03:02
Breath Volume	2.027	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 3	0.172	03:04
Breath Volume	1.156	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CUCUL XHO

DATE/TIME OF TEST: 07/22/2024 23:38:56

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.081	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.253	23:34
Breath Volume	2.539	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.240	23:37
Breath Volume	3.531	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: GUZMAN GOMEZ

DATE/TIME OF TEST: 07/24/2024 01:25:09

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.257	01:20
Breath Volume	2.363	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample # 2	0.259	01:23
Breath Volume	1.667	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: VILLEGAS

DATE/TIME OF TEST: 07/24/2024 02:20:26

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.000	02:16
Breath Volume	2.824	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.000	02:19
Breath Volume	2.324	02:19
Air Blank	0.000	02:19
Control Test	0.082	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/24/2024 04:59:48

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.116	04:54
Breath Volume	2.355	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.112	04:58
Breath Volume	1.855	04:58
Air Blank	0.000	04:58
Control Test	0.080	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 08/06/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MASOCOLA, D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: OECHSNER

DATE/TIME OF TEST: 07/24/2024 22:08:24

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.081	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.175	22:03
Breath Volume	1.718	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.170	22:06
Breath Volume	1.488	22:06
Air Blank	0.000	22:07
Control Test	0.081	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANAYA,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MEZA NAVA

DATE/TIME OF TEST: 07/25/2024 00:31:11

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.191	00:26
Breath Volume	3.007	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.186	00:29
Breath Volume	2.500	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/25/2024 02:05:07

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.081	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:04	
Control Test	0.081	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: SABORIT SANCHEZ

DATE/TIME OF TEST: 07/25/2024 20:43:31

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.081	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.197	20:38
Breath Volume	1.402	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.197	20:41
Breath Volume	1.515	20:41
Air Blank	0.000	20:42
Control Test	0.080	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRRIZARY,KATTIE,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARRASCO,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: CEPEDA CARDOZO

DATE/TIME OF TEST: 07/25/2024 22:05:54

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.067	22:01
Breath Volume	2.617	22:01
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.068	22:04
Breath Volume	2.593	22:04
Air Blank	0.000	22:04
Control Test	0.080	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: GAPP

DATE/TIME OF TEST: 07/26/2024 05:28:07

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.156	05:23
Breath Volume	2.812	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:25
Subject Sample # 2	0.151	05:26
Breath Volume	2.761	05:26
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRRIZARY,KATTIE,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: KITCHEN

DATE/TIME OF TEST: 07/26/2024 23:26:47

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.142	23:22
Breath Volume	1.406	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.142	23:25
Breath Volume	1.128	23:25
Air Blank	0.000	23:25
Control Test	0.079	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/27/2024 02:52:16

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.081	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:48	No Sample Provided
Breath Volume	0.000	02:48	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample	# 2	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: HOLLY

DATE/TIME OF TEST: 07/27/2024 03:59:26

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.111	03:54
Breath Volume	1.523	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.107	03:57
Breath Volume	1.898	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: SANCHEZ ALARCON

DATE/TIME OF TEST: 07/27/2024 04:40:45

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.208	04:35
Breath Volume	2.371	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.209	04:39
Breath Volume	4.152	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: OLDHAM

DATE/TIME OF TEST: 07/27/2024 06:27:28

Diagnostics Check	OK	06:19
Air Blank	0.000	06:20
Control Test	0.080	06:20
Air Blank	0.000	06:20
Subject Sample # 1	0.220	06:22
Breath Volume	1.527	06:22
Air Blank	0.000	06:23
Air Blank	0.000	06:24
Subject Sample # 2	0.224	06:25
Breath Volume	2.171	06:25
Air Blank	0.000	06:26
Control Test	0.079	06:26
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: CORTES FLORES

DATE/TIME OF TEST: 07/27/2024 20:34:34

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.080	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.000	20:30
Breath Volume	2.273	20:30
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.000	20:33
Breath Volume	1.207	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:33
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLIPIN,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: GAVIRIA JARAMILLO

DATE/TIME OF TEST: 07/27/2024 22:02:59

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.081	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.036	21:58
Breath Volume	1.113	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.038	22:01
Breath Volume	1.648	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS, K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: DE LA CRUZ LOPEZ

DATE/TIME OF TEST: 07/27/2024 23:43:53

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.193	23:39
Breath Volume	2.796	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.196	23:42
Breath Volume	2.675	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MERRIWEATHER

DATE/TIME OF TEST: 07/27/2024 23:55:50

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.080	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.210	23:54	
Air Blank	0.000	23:54	
Control Test	0.080	23:55	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MERRIWEATHER

DATE/TIME OF TEST: 07/28/2024 00:11:46

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.079	00:04	
Air Blank	0.000	00:04	
Subject Sample	# 1 0.155	00:07	Volume Not Met
Breath Volume	0.199	00:07	
Air Blank	0.000	00:08	
Air Blank	0.000	00:10	
Subject Sample	# 2	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Control Test	0.079	00:11	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MANCHESTER

DATE/TIME OF TEST: 07/28/2024 00:31:06

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.163	00:26
Breath Volume	4.238	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.158	00:29
Breath Volume	3.851	00:29
Air Blank	0.000	00:30
Control Test	0.082	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: VASQUEZ PEREZ

DATE/TIME OF TEST: 07/28/2024 00:58:51

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.218	00:54
Breath Volume	3.148	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.227	00:57
Breath Volume	3.750	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: RICARDO GUTIERREZ

DATE/TIME OF TEST: 07/28/2024 04:19:38

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.140	04:15
Breath Volume	1.281	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.132	04:18
Breath Volume	1.660	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/28/2024 05:12:16

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.152	05:07
Breath Volume	1.148	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.158	05:10
Breath Volume	2.058	05:10
Air Blank	0.000	05:11
Control Test	0.080	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: ESPICHAN

DATE/TIME OF TEST: 07/28/2024 06:01:05

Diagnostics Check	OK	05:54
Air Blank	0.000	05:55
Control Test	0.080	05:55
Air Blank	0.000	05:56
Subject Sample # 1	0.021	05:56
Breath Volume	1.781	05:56
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.022	05:59
Breath Volume	2.265	05:59
Air Blank	0.000	06:00
Control Test	0.080	06:00
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 08/06/2024 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/28/2024 23:24:01

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.109	23:19
Breath Volume	4.394	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:21
Subject Sample # 2	0.104	23:22
Breath Volume	2.769	23:22
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:24

Data Download Date/Time: 08/06/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 07/29/2024 04:40:03

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.080	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:39	
Control Test	0.080	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 08/06/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,KEVIN,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MARQUEZ,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 07/29/2024 14:47:18

Diagnostics Check	OK	14:37
Air Blank	0.000	14:38
Control Test	0.080	14:38
Air Blank	0.000	14:39
Subject Sample # 1	0.328	14:39
Breath Volume	1.792	14:39
Air Blank	0.000	14:40
Air Blank	0.000	14:42
Subject Sample # 2	0.300	14:42
Breath Volume	1.128	14:42
Air Blank	0.000	14:43
Air Blank	0.000	14:45
Subject Sample # 3	0.313	14:45
Breath Volume	1.500	14:45
Air Blank	0.000	14:46
Control Test	0.080	14:46
Air Blank	0.000	14:47
Diagnostics Check	OK	14:47

Data Download Date/Time: 08/06/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANDCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: AL-MASRI

DATE/TIME OF TEST: 07/30/2024 21:58:34

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.081	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.185	21:53
Breath Volume	1.871	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.187	21:56
Breath Volume	2.777	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 08/06/2024 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: MCQUAY

DATE/TIME OF TEST: 07/31/2024 00:02:57

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.081	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.328	00:01	
Air Blank	0.000	00:02	
Control Test	0.080	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 08/06/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRRIZARY,KATTIE,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: PEDRAZZOLI

DATE/TIME OF TEST: 07/31/2024 01:02:25

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 08/06/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRRIZARY,KATTIE,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: PEDRAZZOLI

DATE/TIME OF TEST: 07/31/2024 01:30:18

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.123	01:25
Breath Volume	1.789	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.117	01:28
Breath Volume	1.726	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 08/06/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: MILWOOD

DATE/TIME OF TEST: 07/31/2024 22:37:14

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.080	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.240	22:32
Breath Volume	1.691	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.224	22:35
Breath Volume	1.125	22:35
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 08/06/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIMONEY,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CHANQUET,NICOLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: RONAN

DATE/TIME OF TEST: 07/02/2024 23:19:36

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:17	
Control Test	0.077	23:17	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:18	Subject Test Refused
Breath Volume	0.000	23:18	
Air Blank	0.000	23:18	
Control Test	0.078	23:19	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 07/23/2024 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 07/03/2024 08:23:39

Diagnostics Check	OK	08:17
Air Blank	0.000	08:17
Control Test	0.078	08:17
Air Blank	0.000	08:18
Subject Sample # 1	0.094	08:18
Breath Volume	2.261	08:18
Air Blank	0.000	08:19
Air Blank	0.000	08:21
Subject Sample # 2	0.096	08:21
Breath Volume	3.207	08:21
Air Blank	0.000	08:22
Control Test	0.077	08:23
Air Blank	0.000	08:23
Diagnostics Check	OK	08:23

Data Download Date/Time: 07/23/2024 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ROSARIO,JONATHAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SORAGHAN

DATE/TIME OF TEST: 07/09/2024 02:27:38

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.357	02:22
Breath Volume	2.957	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.348	02:25
Breath Volume	3.164	02:25
Air Blank	0.000	02:26
Control Test	0.075	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 07/23/2024 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: DE LUCA

DATE/TIME OF TEST: 07/12/2024 12:12:17

Diagnostics Check	OK	12:06
Air Blank	0.000	12:06
Control Test	0.077	12:07
Air Blank	0.000	12:07
Subject Sample # 1	0.000	12:08
Breath Volume	3.148	12:08
Air Blank	0.000	12:08
Air Blank	0.000	12:10
Subject Sample # 2	0.000	12:10
Breath Volume	3.031	12:10
Air Blank	0.000	12:11
Control Test	0.077	12:11
Air Blank	0.000	12:12
Diagnostics Check	OK	12:12

Data Download Date/Time: 07/23/2024 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GRUBBS,ALEC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: MCINTYRE

DATE/TIME OF TEST: 07/27/2024 20:53:37

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.078	20:47
Air Blank	0.000	20:47
Subject Sample # 1	0.230	20:48
Breath Volume	1.414	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.230	20:51
Breath Volume	1.605	20:51
Air Blank	0.000	20:52
Control Test	0.078	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 08/15/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANSSEN,JACOB,E
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JANSSEN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 07:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: WHITTAKER

DATE/TIME OF TEST: 07/13/2024 09:20:56

Diagnostics Check	OK	09:10
Air Blank	0.000	09:11
Control Test	0.079	09:11
Air Blank	0.000	09:12
Subject Sample # 1	0.118	09:12
Breath Volume	3.921	09:12
Air Blank	0.000	09:13
Air Blank	0.000	09:15
Subject Sample # 2	0.097	09:16
Breath Volume	1.914	09:16
Air Blank	0.000	09:16
Air Blank	0.000	09:18
Subject Sample # 3	0.106	09:19
Breath Volume	2.531	09:19
Air Blank	0.000	09:20
Control Test	0.078	09:20
Air Blank	0.000	09:20
Diagnostics Check	OK	09:20

Data Download Date/Time: 08/12/2024 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: URUCHIMA,ANGEL,R
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: URUCHIMA,ANGEL
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ANTUNEZ SALGADO

DATE/TIME OF TEST: 07/14/2024 04:24:00

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.245	04:19
Breath Volume	1.394	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.257	04:22
Breath Volume	1.617	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:24

Data Download Date/Time: 08/12/2024 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI, MARK, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JABLONSKI, MARK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: CAMACHO MARTINEZ

DATE/TIME OF TEST: 07/20/2024 04:59:04

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.161	04:52
Breath Volume	1.171	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.154	04:57
Breath Volume	1.546	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:59

Data Download Date/Time: 08/12/2024 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOW,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FLOW,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: BENAVIDES MENESES

DATE/TIME OF TEST: 07/20/2024 17:12:02

Diagnostics Check	OK	17:00	
Air Blank	0.000	17:00	
Control Test	0.079	17:00	
Air Blank	0.000	17:01	
Subject Sample	# 1 0.250	17:04	Volume Not Met
Breath Volume	0.285	17:04	
Air Blank	0.000	17:05	
Air Blank	0.000	17:07	
Subject Sample	# 2 0.255	17:10	Volume Not Met
Breath Volume	0.121	17:10	
Air Blank	0.000	17:11	
Control Test	0.073	17:11	Control Outside Tolerance
Air Blank	0.000	17:12	

Data Download Date/Time: 08/12/2024 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOW,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FLOW,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: BENAVIDES MENESES

DATE/TIME OF TEST: 07/20/2024 17:26:03

Diagnostics Check	OK	17:14	
Air Blank	0.000	17:14	
Control Test	0.076	17:14	
Air Blank	0.000	17:15	
Subject Sample	# 1 0.232	17:18	Volume Not Met
Breath Volume	0.476	17:18	
Air Blank	0.000	17:19	
Air Blank	0.000	17:21	
Subject Sample	# 2 0.225	17:24	Volume Not Met
Breath Volume	0.179	17:24	
Air Blank	0.000	17:25	
Control Test	0.074	17:25	Control Outside Tolerance
Air Blank	0.000	17:26	

Data Download Date/Time: 08/12/2024 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: CRUZ MEDINA

DATE/TIME OF TEST: 07/21/2024 05:29:45

Diagnostics Check	OK	05:24	
Air Blank	0.000	05:25	
Control Test	0.078	05:25	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:28	Subject Test Refused
Breath Volume	0.101	05:28	
Air Blank	0.000	05:28	
Control Test	0.078	05:29	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:29	

Data Download Date/Time: 08/12/2024 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MEDINA BERNAL

DATE/TIME OF TEST: 07/21/2024 23:05:16

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.078	23:03	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:04	
Control Test	0.079	23:04	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 08/12/2024 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUSSEAU,DMITRY,A
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: ROUSSEAU,DMITRY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: FAISON

DATE/TIME OF TEST: 07/22/2024 00:38:41

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.079	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:31	No Sample Provided
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2	00:37	No Sample Provided
Breath Volume	0.000	00:37	
Air Blank	0.000	00:37	
Control Test	0.078	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 08/12/2024 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GOMEZ MORALES

DATE/TIME OF TEST: 07/26/2024 03:32:56

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.210	03:23
Breath Volume	1.105	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.185	03:28
Breath Volume	2.019	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 3	0.188	03:31
Breath Volume	4.007	03:31
Air Blank	0.000	03:32
Control Test	0.076	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 08/12/2024 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANSSEN,JACOB,E
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JANSSEN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 07/28/2024 05:44:24

Diagnostics Check	OK	05:37
Air Blank	0.000	05:37
Control Test	0.079	05:38
Air Blank	0.000	05:38
Subject Sample # 1	0.075	05:39
Breath Volume	3.597	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.074	05:42
Breath Volume	2.890	05:42
Air Blank	0.000	05:43
Control Test	0.079	05:43
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 08/12/2024 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: DEDHO DURAN

DATE/TIME OF TEST: 07/12/2024 22:43:53

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.076	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.046	22:39
Breath Volume	2.878	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.046	22:42
Breath Volume	2.093	22:42
Air Blank	0.000	22:42
Control Test	0.078	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 08/27/2024 23:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: RAHUL

DATE/TIME OF TEST: 07/16/2024 01:31:13

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.076	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.109	01:25
Breath Volume	1.566	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.103	01:29
Breath Volume	2.019	01:29
Air Blank	0.000	01:30
Control Test	0.077	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 08/27/2024 23:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: OROZCO VASQUEZ

DATE/TIME OF TEST: 07/17/2024 23:22:21

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.074	23:21	Interferent Detect
Air Blank		23:22	Ambient Fail

Data Download Date/Time: 08/27/2024 23:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: OROZCO VASQUEZ

DATE/TIME OF TEST: 07/17/2024 23:29:26

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:25	
Control Test	0.085	23:25	
Air Blank	0.000	23:25	
Subject Sample	# 1 0.181	23:26	
Breath Volume	2.078	23:26	
Air Blank	0.000	23:27	
Air Blank		23:28	Ambient Fail
Air Blank		23:29	Ambient Fail

Data Download Date/Time: 08/27/2024 23:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 07/03/2024 01:41:20

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:37
Breath Volume	2.812	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.000	01:39
Breath Volume	3.125	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 08/12/2024 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: WALL

DATE/TIME OF TEST: 07/07/2024 03:06:15

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.000	03:01
Breath Volume	2.894	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.000	03:04
Breath Volume	2.937	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 08/12/2024 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BENNICI

DATE/TIME OF TEST: 07/26/2024 02:43:20

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1 0.087	02:34	Volume Not Met
Breath Volume	0.203	02:34	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 2 0.101	02:38	
Breath Volume	1.828	02:38	
Air Blank	0.000	02:39	
Air Blank	0.000	02:40	
Subject Sample	# 3 0.104	02:41	
Breath Volume	1.285	02:41	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 08/12/2024 20:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: JASON,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 07/05/2024 01:49:32

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.079	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1 0.048	01:40	Volume Not Met
Breath Volume	0.902	01:40	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 2 0.052	01:44	
Breath Volume	1.667	01:44	
Air Blank	0.000	01:44	
Air Blank	0.000	01:46	
Subject Sample	# 3 0.049	01:48	
Breath Volume	2.207	01:48	
Air Blank	0.000	01:48	
Control Test	0.078	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 07/08/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 07/05/2024 03:28:06

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.076	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.247	03:23
Breath Volume	1.230	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.251	03:26
Breath Volume	2.128	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 07/08/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OLGUIN-MARTINEZ,F
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 07/05/2024 18:10:27

Diagnostics Check	OK	18:04
Air Blank	0.000	18:04
Control Test	0.082	18:04
Air Blank	0.000	18:05
Subject Sample # 1	0.186	18:05
Breath Volume	1.242	18:05
Air Blank	0.000	18:06
Air Blank	0.000	18:08
Subject Sample # 2	0.190	18:08
Breath Volume	1.351	18:08
Air Blank	0.000	18:09
Control Test	0.081	18:09
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 07/26/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: SEABROOKS

DATE/TIME OF TEST: 07/26/2024 22:12:29

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:04	
Control Test	0.081	22:04	
Air Blank	0.000	22:05	
Subject Sample	# 1 0.198	22:08	Volume Not Met
Breath Volume	0.160	22:08	
Air Blank	0.000	22:09	
Air Blank	0.000	22:10	
Subject Sample	# 2	22:11	Subject Test Refused
Breath Volume	0.000	22:11	
Air Blank	0.000	22:11	
Control Test	0.080	22:11	
Air Blank	0.000	22:12	
Diagnostics Check	OK	22:12	

Data Download Date/Time: 08/21/2024 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: TOBOUL

DATE/TIME OF TEST: 07/14/2024 22:21:07

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.079	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.000	22:16
Breath Volume	2.175	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.000	22:19
Breath Volume	1.105	22:19
Air Blank	0.000	22:20
Control Test	0.079	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SSANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ALCIME

DATE/TIME OF TEST: 07/04/2024 03:18:41

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.236	03:13
Breath Volume	1.527	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.243	03:17
Breath Volume	2.035	03:17
Air Blank	0.000	03:17
Control Test	0.079	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LOPEZ PADILLA

DATE/TIME OF TEST: 07/04/2024 03:39:59

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.140	03:35
Breath Volume	2.636	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.131	03:38
Breath Volume	2.109	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FAUL,JOHN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ROTSCH

DATE/TIME OF TEST: 07/05/2024 01:55:07

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.000	01:50
Breath Volume	2.121	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.000	01:53
Breath Volume	1.621	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 07/05/2024 02:19:53

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.233	02:12
Breath Volume	1.171	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.240	02:18
Breath Volume	1.253	02:18
Air Blank	0.000	02:18
Control Test	0.077	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PIERRE-LOUIS,LUDNA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 07/06/2024 06:41:37

Diagnostics Check	OK	06:35
Air Blank	0.000	06:35
Control Test	0.079	06:35
Air Blank	0.000	06:36
Subject Sample # 1	0.130	06:37
Breath Volume	1.996	06:37
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.129	06:40
Breath Volume	1.664	06:40
Air Blank	0.000	06:40
Control Test	0.080	06:41
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JENNINGS,LUKE
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: CAJAL

DATE/TIME OF TEST: 07/06/2024 21:41:33

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.079	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.261	21:37
Breath Volume	1.429	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.267	21:39
Breath Volume	1.156	21:39
Air Blank	0.000	21:40
Control Test	0.077	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REDFEARN,JAMES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/07/2024 01:59:14

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.231	01:54
Breath Volume	1.867	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.222	01:57
Breath Volume	1.406	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/07/2024 03:57:28

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.127	03:53
Breath Volume	1.312	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.123	03:55
Breath Volume	1.839	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: LAGUERRE

DATE/TIME OF TEST: 07/07/2024 05:00:19

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.079	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.116	04:55
Breath Volume	1.179	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.118	04:58
Breath Volume	2.363	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: VERNA

DATE/TIME OF TEST: 07/10/2024 04:32:30

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.078	04:27
Breath Volume	1.457	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.075	04:30
Breath Volume	1.304	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOPEZ,BRYAN
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: CASTILLO MONTERO

DATE/TIME OF TEST: 07/11/2024 11:45:07

Diagnostics Check	OK	11:38
Air Blank	0.000	11:38
Control Test	0.080	11:39
Air Blank	0.000	11:39
Subject Sample # 1	0.203	11:40
Breath Volume	1.277	11:40
Air Blank	0.000	11:41
Air Blank	0.000	11:43
Subject Sample # 2	0.204	11:43
Breath Volume	2.093	11:43
Air Blank	0.000	11:44
Control Test	0.080	11:44
Air Blank	0.000	11:45
Diagnostics Check	OK	11:45

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MAJOR,BERNARD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 10:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:46

SUBJECT'S LAST NAME: DORSAINVILLE

DATE/TIME OF TEST: 07/13/2024 13:24:16

Diagnostics Check	OK	13:17
Air Blank	0.000	13:17
Control Test	0.080	13:17
Air Blank	0.000	13:18
Subject Sample # 1	0.079	13:19
Breath Volume	2.269	13:19
Air Blank	0.000	13:19
Air Blank	0.000	13:21
Subject Sample # 2	0.075	13:22
Breath Volume	2.062	13:22
Air Blank	0.000	13:23
Control Test	0.080	13:23
Air Blank	0.000	13:24
Diagnostics Check	OK	13:24

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/14/2024 03:15:32

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.186	03:11
Breath Volume	1.136	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.193	03:13
Breath Volume	1.312	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CHEREVKO

DATE/TIME OF TEST: 07/14/2024 03:57:48

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.204	03:53
Breath Volume	2.156	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.199	03:56
Breath Volume	1.945	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 07/14/2024 06:46:36

Diagnostics Check	OK	06:38
Air Blank	0.000	06:39
Control Test	0.080	06:39
Air Blank	0.000	06:39
Subject Sample # 1	0.000	06:42
Breath Volume	1.347	06:42
Air Blank	0.000	06:42
Air Blank	0.000	06:44
Subject Sample # 2	0.000	06:45
Breath Volume	1.132	06:45
Air Blank	0.000	06:45
Control Test	0.078	06:45
Air Blank	0.000	06:46
Diagnostics Check	OK	06:46

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUC,NORBENS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 08:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: YUDITSKY

DATE/TIME OF TEST: 07/15/2024 10:55:52

Diagnostics Check	OK	10:48
Air Blank	0.000	10:49
Control Test	0.080	10:49
Air Blank	0.000	10:50
Subject Sample # 1	0.000	10:50
Breath Volume	1.121	10:50
Air Blank	0.000	10:51
Air Blank	0.000	10:53
Subject Sample # 2	0.000	10:54
Breath Volume	2.535	10:54
Air Blank	0.000	10:54
Control Test	0.081	10:55
Air Blank	0.000	10:55
Diagnostics Check	OK	10:55

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RYAN,JASON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: DUQUE CORREA

DATE/TIME OF TEST: 07/17/2024 23:08:37

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.188	23:03
Breath Volume	2.257	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.200	23:07
Breath Volume	2.160	23:07
Air Blank	0.000	23:07
Control Test	0.077	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: ZAPATA RENTERIA

DATE/TIME OF TEST: 07/21/2024 04:26:15

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.167	04:21
Breath Volume	1.816	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.172	04:24
Breath Volume	2.035	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BENNEFIELD

DATE/TIME OF TEST: 07/27/2024 01:46:25

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.223	01:41
Breath Volume	1.773	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.220	01:44
Breath Volume	1.300	01:44
Air Blank	0.000	01:45
Control Test	0.078	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 08/07/2024 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FERIA,MARIA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SCANLON

DATE/TIME OF TEST: 07/27/2024 03:36:24

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.115	03:31
Breath Volume	2.996	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.112	03:34
Breath Volume	2.906	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 08/07/2024 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: HOZANNA

DATE/TIME OF TEST: 07/27/2024 04:39:48

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.195	04:35
Breath Volume	1.781	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.205	04:38
Breath Volume	1.730	04:38
Air Blank	0.000	04:38
Control Test	0.080	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 08/07/2024 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: SORIANO

DATE/TIME OF TEST: 07/27/2024 06:21:05

Diagnostics Check	OK	06:14
Air Blank	0.000	06:15
Control Test	0.080	06:15
Air Blank	0.000	06:16
Subject Sample # 1	0.184	06:16
Breath Volume	1.332	06:16
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.190	06:19
Breath Volume	1.261	06:19
Air Blank	0.000	06:20
Control Test	0.080	06:20
Air Blank	0.000	06:21
Diagnostics Check	OK	06:21

Data Download Date/Time: 08/07/2024 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MAJOR,BERNARD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 07/28/2024 09:24:00

Diagnostics Check	OK	09:17
Air Blank	0.000	09:17
Control Test	0.080	09:18
Air Blank	0.000	09:18
Subject Sample # 1	0.104	09:19
Breath Volume	1.433	09:19
Air Blank	0.000	09:19
Air Blank	0.000	09:21
Subject Sample # 2	0.095	09:22
Breath Volume	1.402	09:22
Air Blank	0.000	09:23
Control Test	0.080	09:23
Air Blank	0.000	09:23
Diagnostics Check	OK	09:24

Data Download Date/Time: 08/07/2024 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOWE,RACHEL
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: DE ALMAGRO

DATE/TIME OF TEST: 07/28/2024 22:46:39

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.079	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.204	22:42
Breath Volume	1.859	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.205	22:45
Breath Volume	1.476	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 08/07/2024 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: HOLDING

DATE/TIME OF TEST: 07/04/2024 15:23:02

Diagnostics Check	OK	15:15
Air Blank	0.000	15:16
Control Test	0.078	15:16
Air Blank	0.000	15:17
Subject Sample # 1	0.277	15:18
Breath Volume	1.445	15:18
Air Blank	0.000	15:18
Air Blank	0.000	15:20
Subject Sample # 2	0.268	15:21
Breath Volume	1.871	15:21
Air Blank	0.000	15:22
Control Test	0.077	15:22
Air Blank	0.000	15:22
Diagnostics Check	OK	15:23

Data Download Date/Time: 07/30/2024 22:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,TREVOR,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: THOMAS,TREVOR
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/25/2024 01:19:13

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.196	01:14
Breath Volume	2.148	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.179	01:17
Breath Volume	2.535	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 07/30/2024 22:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: HORKY

DATE/TIME OF TEST: 07/26/2024 04:04:50

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.078	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.207	04:00
Breath Volume	1.773	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.213	04:03
Breath Volume	1.984	04:03
Air Blank	0.000	04:03
Control Test	0.077	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 07/30/2024 22:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ TRIANA,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SPEICH,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: GONZALEZ JIMENEZ

DATE/TIME OF TEST: 07/04/2024 00:08:41

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.225	00:03
Breath Volume	1.679	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.220	00:07
Breath Volume	2.445	00:07
Air Blank	0.000	00:07
Control Test	0.078	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 08/12/2024 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ODELL,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ELLINGTON

DATE/TIME OF TEST: 07/04/2024 02:16:30

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.167	02:11
Breath Volume	2.132	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.160	02:14
Breath Volume	1.160	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 08/12/2024 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ TRIANA,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VERREAULT,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RUIZ GARCIA

DATE/TIME OF TEST: 07/05/2024 02:04:48

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.256	01:59
Breath Volume	1.406	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.252	02:03
Breath Volume	2.574	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 08/12/2024 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDASH, MATTHEW,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESTES, S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: JULES FILS

DATE/TIME OF TEST: 07/05/2024 20:44:09

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.247	20:39
Breath Volume	1.476	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.244	20:42
Breath Volume	1.976	20:42
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 08/12/2024 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGE,L,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LUKAS,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/07/2024 17:44:50

Diagnostics Check	OK	17:33	
Air Blank	0.000	17:33	
Control Test	0.078	17:34	
Air Blank	0.000	17:34	
Subject Sample	# 1 0.066	17:37	Volume Not Met
Breath Volume	0.035	17:37	
Air Blank	0.000	17:38	
Air Blank	0.000	17:40	
Subject Sample	# 2 0.000	17:43	Volume Not Met
Breath Volume	0.082	17:43	
Air Blank	0.000	17:43	
Control Test	0.079	17:44	
Air Blank	0.000	17:44	
Diagnostics Check	OK	17:44	

Data Download Date/Time: 08/12/2024 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ABALOS VIDAL

DATE/TIME OF TEST: 07/08/2024 02:18:25

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.079	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1 0.101	02:07	Volume Not Met
Breath Volume	0.343	02:07	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample	# 2 0.148	02:11	
Breath Volume	2.000	02:11	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 3 0.144	02:16	Volume Not Met
Breath Volume	0.035	02:16	
Air Blank	0.000	02:17	
Control Test	0.078	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 08/12/2024 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ABALOS VIDAL

DATE/TIME OF TEST: 07/08/2024 02:37:21

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.077	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1 0.137	02:24	
Breath Volume	1.191	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 2 0.133	02:29	Volume Not Met
Breath Volume	0.097	02:29	
Air Blank	0.000	02:30	
Air Blank	0.000	02:32	
Subject Sample	# 3 0.136	02:35	Volume Not Met
Breath Volume	0.609	02:35	
Air Blank	0.000	02:36	
Control Test	0.075	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 08/12/2024 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEATON,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HILL,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BELSITO

DATE/TIME OF TEST: 07/08/2024 02:48:41

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.075	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.000	02:43
Breath Volume	2.761	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.000	02:47
Breath Volume	1.921	02:47
Air Blank	0.000	02:47
Control Test	0.077	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 08/12/2024 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ABALOS VIDAL

DATE/TIME OF TEST: 07/08/2024 03:06:33

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.077	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1 0.143	02:55	Volume Not Met
Breath Volume	0.156	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2 0.122	02:59	
Breath Volume	1.277	02:59	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 3 0.120	03:04	
Breath Volume	1.679	03:04	
Air Blank	0.000	03:05	
Control Test	0.076	03:05	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:06	

Data Download Date/Time: 08/12/2024 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,AUSTIN,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GUTIERREZ,I
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: RANDOL

DATE/TIME OF TEST: 07/11/2024 22:44:53

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.079	22:32	
Air Blank	0.000	22:32	
Subject Sample	# 1 0.173	22:35	Volume Not Met
Breath Volume	0.843	22:35	
Air Blank	0.000	22:36	
Air Blank	0.000	22:38	
Subject Sample	# 2 0.184	22:38	
Breath Volume	1.214	22:38	
Air Blank	0.000	22:39	
Air Blank	0.000	22:41	
Subject Sample	# 3 0.183	22:43	
Breath Volume	1.179	22:43	
Air Blank	0.000	22:43	
Control Test	0.077	22:44	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAHAGIAN,TARA,K
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: EMANS,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: GALARZA

DATE/TIME OF TEST: 07/13/2024 03:13:13

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.197	03:08
Breath Volume	1.628	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.194	03:11
Breath Volume	1.253	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BERMUDEZ,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: LOPEZ-BAUTISTA

DATE/TIME OF TEST: 07/13/2024 03:53:11

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.229	03:48
Breath Volume	1.503	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.247	03:51
Breath Volume	1.628	03:51
Air Blank	0.000	03:52
Control Test	0.077	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL, ADAM, J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WENNLUND, F
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 07/13/2024 04:27:32

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.077	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 1 0.243	04:17	Volume Not Met
Breath Volume	0.128	04:17	
Air Blank	0.000	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 2 0.235	04:21	
Breath Volume	2.066	04:21	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample	# 3 0.233	04:25	
Breath Volume	1.558	04:25	
Air Blank	0.000	04:26	
Control Test	0.075	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOMANDO,TINA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GRISALES,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SILHOMME

DATE/TIME OF TEST: 07/14/2024 02:20:30

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.189	02:15
Breath Volume	1.140	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.193	02:18
Breath Volume	2.699	02:18
Air Blank	0.000	02:19
Control Test	0.076	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTT,RYAN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WENNLUND,F
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 07/14/2024 05:14:58

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.134	05:10
Breath Volume	1.648	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:12
Subject Sample # 2	0.134	05:13
Breath Volume	1.804	05:13
Air Blank	0.000	05:14
Control Test	0.076	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: CRUZ LEAL

DATE/TIME OF TEST: 07/15/2024 03:57:25

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.198	03:52
Breath Volume	1.347	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.192	03:55
Breath Volume	1.328	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,KADE,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HILL,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: LOSURDO

DATE/TIME OF TEST: 07/16/2024 00:46:28

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.079	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 1 0.137	00:42	Volume Not Met
Breath Volume	0.406	00:42	
Air Blank	0.000	00:43	
Air Blank	0.000	00:44	
Subject Sample	# 2	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Control Test	0.077	00:45	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,KADE,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HILL,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: LOSURDO

DATE/TIME OF TEST: 07/16/2024 00:58:00

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.078	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.160	00:51
Breath Volume	1.605	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.159	00:56
Breath Volume	1.300	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:58

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,JUAN,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MEDINA,JUAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MAGOS CORONA

DATE/TIME OF TEST: 07/16/2024 04:16:58

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.078	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.064	04:12
Breath Volume	2.484	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.060	04:15
Breath Volume	2.566	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHNEISS,J,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,KURT
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: NOONE

DATE/TIME OF TEST: 07/17/2024 14:49:32

Diagnostics Check	OK	14:40
Air Blank	0.000	14:40
Control Test	0.079	14:41
Air Blank	0.000	14:41
Subject Sample # 1	0.357	14:43
Breath Volume	2.824	14:43
Air Blank	0.000	14:44
Air Blank	0.000	14:45
Subject Sample # 2	0.354	14:47
Breath Volume	2.375	14:47
Air Blank	0.000	14:48
Control Test	0.076	14:48
Air Blank	0.000	14:49
Diagnostics Check	OK	14:49

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEISE,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 07/17/2024 19:39:47

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:28	
Control Test	0.079	19:28	
Air Blank	0.000	19:29	
Subject Sample	# 1 0.131	19:32	Volume Not Met
Breath Volume	0.417	19:32	
Air Blank	0.000	19:33	
Air Blank	0.000	19:34	
Subject Sample	# 2 0.130	19:38	Volume Not Met
Breath Volume	0.320	19:38	
Air Blank	0.000	19:38	
Control Test	0.077	19:39	
Air Blank	0.000	19:39	
Diagnostics Check	OK	19:39	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEISE,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 07/17/2024 20:00:52

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:44	
Control Test	0.076	19:45	
Air Blank	0.000	19:45	
Subject Sample	# 1 0.130	19:47	
Breath Volume	1.132	19:47	
Air Blank	0.000	19:48	
Air Blank	0.000	19:50	
Subject Sample	# 2 0.124	19:53	Volume Not Met
Breath Volume	0.695	19:53	
Air Blank	0.000	19:54	
Air Blank	0.000	19:56	
Subject Sample	# 3 0.104	19:59	Volume Not Met
Breath Volume	0.609	19:59	
Air Blank	0.000	19:59	
Control Test	0.076	20:00	
Air Blank	0.000	20:00	
Diagnostics Check	OK	20:00	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEISE,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 07/17/2024 20:17:08

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.076	20:10
Air Blank	0.000	20:10
Subject Sample # 1	0.118	20:12
Breath Volume	2.417	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.115	20:15
Breath Volume	2.257	20:15
Air Blank	0.000	20:16
Control Test	0.076	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,T,H
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GRISALES,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 07/18/2024 00:57:19

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.197	00:52
Breath Volume	2.375	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.201	00:55
Breath Volume	2.742	00:55
Air Blank	0.000	00:56
Control Test	0.078	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HARRISON,Z
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: LASCAREZ BALTAZAR

DATE/TIME OF TEST: 07/18/2024 18:56:32

Diagnostics Check	OK	18:45	
Air Blank	0.000	18:45	
Control Test	0.078	18:45	
Air Blank	0.000	18:46	
Subject Sample	# 1	18:49	No Sample Provided
Breath Volume	0.000	18:49	
Air Blank	0.000	18:50	
Air Blank	0.000	18:52	
Subject Sample	# 2	18:55	No Sample Provided
Breath Volume	0.000	18:55	
Air Blank	0.000	18:55	
Control Test	0.079	18:55	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HARRISON,Z
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: LASCAREZ BALTAZAR

DATE/TIME OF TEST: 07/18/2024 19:01:07

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:58	
Control Test	0.078	18:58	
Air Blank	0.000	18:59	
Subject Sample	# 1	18:59	Subject Test Refused
Breath Volume	0.000	18:59	
Air Blank	0.000	19:00	
Control Test	0.078	19:00	
Air Blank	0.000	19:01	
Diagnostics Check	OK	19:01	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BUILES,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/19/2024 21:38:43

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.079	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:37	
Control Test	0.079	21:38	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LING,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GARCIA ALVAREZ

DATE/TIME OF TEST: 07/20/2024 01:46:15

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.185	01:41
Breath Volume	1.429	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:43
Subject Sample # 2	0.180	01:44
Breath Volume	2.554	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDASH, MATTHEW,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LANG, J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: JOSNAIDA

DATE/TIME OF TEST: 07/20/2024 05:20:14

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:06	
Control Test	0.079	05:07	
Air Blank	0.000	05:07	
Subject Sample # 1	0.112	05:08	
Breath Volume	1.531	05:08	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample # 2	0.098	05:14	Volume Not Met
Breath Volume	0.382	05:14	
Air Blank	0.000	05:15	
Air Blank	0.000	05:16	
Subject Sample # 3	0.108	05:18	
Breath Volume	2.476	05:18	
Air Blank	0.000	05:19	
Control Test	0.077	05:19	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEATON,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BUILES,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/20/2024 05:36:07

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.078	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.120	05:31
Breath Volume	1.921	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:33
Subject Sample # 2	0.120	05:34
Breath Volume	1.917	05:34
Air Blank	0.000	05:35
Control Test	0.078	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:23

SUBJECT'S LAST NAME: FLORES REYNALDO

DATE/TIME OF TEST: 07/21/2024 14:11:35

Diagnostics Check	OK	13:58	
Air Blank	0.000	13:59	
Control Test	0.079	13:59	
Air Blank	0.000	14:00	
Subject Sample	# 1 0.063	14:03	Volume Not Met
Breath Volume	0.363	14:03	
Air Blank	0.000	14:04	
Air Blank	0.000	14:05	
Subject Sample	# 2 0.060	14:06	
Breath Volume	2.796	14:06	
Air Blank	0.000	14:07	
Air Blank	0.000	14:09	
Subject Sample	# 3 0.063	14:09	
Breath Volume	1.339	14:09	
Air Blank	0.000	14:10	
Control Test	0.077	14:10	
Air Blank	0.000	14:11	
Diagnostics Check	OK	14:11	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,KADE,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEGRON,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: GONZALEZJOSE

DATE/TIME OF TEST: 07/21/2024 23:32:33

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.079	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1 0.152	23:31	Slope Not Level
Breath Volume	1.203	23:31	
Air Blank	0.000	23:31	
Air Blank	0.000	23:32	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,KADE,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEGRON,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: GONZALEZJOSE

DATE/TIME OF TEST: 07/21/2024 23:41:32

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.077	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.164	23:36
Breath Volume	1.371	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.167	23:39
Breath Volume	1.425	23:39
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,DEAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PARKS,DEAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: NOONE

DATE/TIME OF TEST: 07/23/2024 17:04:53

Diagnostics Check	OK	16:52	
Air Blank	0.000	16:53	
Control Test	0.079	16:53	
Air Blank	0.000	16:54	
Subject Sample	# 1	16:57	No Sample Provided
Breath Volume	0.000	16:57	
Air Blank	0.000	16:57	
Air Blank	0.000	16:59	
Subject Sample	# 2	0.330	17:02
Breath Volume	0.140	17:02	Volume Not Met
Air Blank	0.000	17:03	
Control Test	0.079	17:04	
Air Blank	0.000	17:04	
Diagnostics Check	OK	17:04	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,KADE,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEISE,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/24/2024 23:19:02

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.080	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.162	23:14
Breath Volume	1.824	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.159	23:17
Breath Volume	1.414	23:17
Air Blank	0.000	23:18
Control Test	0.078	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:19

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROELL,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MISSAL

DATE/TIME OF TEST: 07/24/2024 23:36:03

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.158	23:31
Breath Volume	1.250	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.161	23:34
Breath Volume	1.820	23:34
Air Blank	0.000	23:35
Control Test	0.077	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:36

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BUILES,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: RODRIGUEZ TORNE

DATE/TIME OF TEST: 07/25/2024 05:26:22

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.079	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.177	05:21
Breath Volume	1.609	05:21
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.171	05:24
Breath Volume	2.417	05:24
Air Blank	0.000	05:25
Control Test	0.078	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KELLEY,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: VELASQUEZ VELASQUEZ

DATE/TIME OF TEST: 07/26/2024 04:04:58

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.237	04:00
Breath Volume	2.414	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.242	04:03
Breath Volume	2.191	04:03
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,NICKOLAUS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAREY,E
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: BRECHNITZ

DATE/TIME OF TEST: 07/26/2024 17:14:47

Diagnostics Check	OK	17:02	
Air Blank	0.000	17:03	
Control Test	0.079	17:03	
Air Blank	0.000	17:04	
Subject Sample	# 1 0.042	17:07	Volume Not Met
Breath Volume	0.140	17:07	
Air Blank	0.000	17:07	
Air Blank	0.000	17:09	
Subject Sample	# 2 0.036	17:10	
Breath Volume	1.460	17:10	
Air Blank	0.000	17:10	
Air Blank	0.000	17:12	
Subject Sample	# 3 0.036	17:13	
Breath Volume	1.433	17:13	
Air Blank	0.000	17:13	
Control Test	0.077	17:14	
Air Blank	0.000	17:14	
Diagnostics Check	OK	17:14	

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL, ADAM, J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GUZMAN, JOSE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/27/2024 03:29:11

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.083	03:24
Breath Volume	1.765	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.084	03:27
Breath Volume	2.222	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILAR,J,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: AGUILAR,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BOYJOONAUTH

DATE/TIME OF TEST: 07/30/2024 02:32:00

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.079	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1 0.179	02:26	
Breath Volume	1.562	02:26	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample	# 2 0.171	02:30	
Breath Volume	2.167	02:30	
Air Blank	0.000	02:31	
Control Test	0.037	02:31	RFI Detect
Air Blank		02:32	RFI Detect

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILAR,J,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: AGUILAR,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BOYJOONAUTH

DATE/TIME OF TEST: 07/30/2024 02:45:21

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.190	02:40
Breath Volume	1.132	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.172	02:43
Breath Volume	1.402	02:43
Air Blank	0.000	02:44
Control Test	0.077	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PURSLEY,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CORNELIO SOTO

DATE/TIME OF TEST: 07/31/2024 01:18:44

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.154	01:14
Breath Volume	1.437	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.155	01:17
Breath Volume	1.367	01:17
Air Blank	0.000	01:17
Control Test	0.079	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,AUSTIN,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 07/31/2024 03:18:46

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.168	03:14
Breath Volume	1.835	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.172	03:17
Breath Volume	1.609	03:17
Air Blank	0.000	03:17
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 08/12/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GULOTTA,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: LOPEZSALES

DATE/TIME OF TEST: 07/03/2024 21:12:00

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.081	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.230	21:07
Breath Volume	2.519	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:09
Subject Sample # 2	0.233	21:10
Breath Volume	2.906	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:12

Data Download Date/Time: 07/04/2024 17:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: AGUAYO

DATE/TIME OF TEST: 07/03/2024 05:49:48

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:47	
Control Test	0.081	05:47	
Air Blank	0.000	05:48	
Subject Sample	# 1	05:48	Subject Test Refused
Breath Volume	0.000	05:48	
Air Blank	0.000	05:48	
Control Test	0.082	05:49	
Air Blank	0.000	05:49	
Diagnostics Check	OK	05:49	

Data Download Date/Time: 07/11/2024 03:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGGINS,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: IRIAS MONCADA,J
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: TABARES

DATE/TIME OF TEST: 07/03/2024 17:33:14

Diagnostics Check	OK	17:29	
Air Blank	0.000	17:30	
Control Test	0.079	17:30	
Air Blank	0.000	17:30	
Subject Sample	# 1	17:31	Subject Test Refused
Breath Volume	0.000	17:31	
Air Blank	0.000	17:32	
Control Test	0.079	17:32	
Air Blank	0.000	17:33	
Diagnostics Check	OK	17:33	

Data Download Date/Time: 07/11/2024 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,CRISTIAN,X
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SANTOS,CRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:04

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 07/06/2024 11:28:08

Diagnostics Check	OK	11:24	
Air Blank	0.000	11:25	
Control Test	0.080	11:25	
Air Blank	0.000	11:26	
Subject Sample	# 1	11:26	Subject Test Refused
Breath Volume	0.000	11:26	
Air Blank	0.000	11:27	
Control Test	0.081	11:27	
Air Blank	0.000	11:28	
Diagnostics Check	OK	11:28	

Data Download Date/Time: 07/11/2024 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGGINS,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ACOSTA,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: OSORIO

DATE/TIME OF TEST: 07/25/2024 20:53:59

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.081	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 1	20:52	Subject Test Refused
Breath Volume	0.000	20:52	
Air Blank	0.000	20:53	
Control Test	0.081	20:53	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

Data Download Date/Time: 08/07/2024 02:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY-IVERSEN, MIYA, K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: POTTS, MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: BROOK

DATE/TIME OF TEST: 07/03/2024 04:30:47

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.082	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.152	04:26
Breath Volume	2.464	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.154	04:29
Breath Volume	1.664	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: WYNN

DATE/TIME OF TEST: 07/03/2024 19:52:15

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.081	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.247	19:47
Breath Volume	3.335	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:49
Subject Sample # 2	0.239	19:50
Breath Volume	2.937	19:50
Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: TORGERSON

DATE/TIME OF TEST: 07/04/2024 00:15:59

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.081	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.189	00:11
Breath Volume	1.746	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.190	00:14
Breath Volume	2.578	00:14
Air Blank	0.000	00:15
Control Test	0.081	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENT,ROBERT,T
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: KENT,ROBERT
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: KAMINSKI

DATE/TIME OF TEST: 07/05/2024 04:51:16

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.082	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.215	04:46
Breath Volume	2.000	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.209	04:49
Breath Volume	2.203	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: NUNEZ AISPURO

DATE/TIME OF TEST: 07/09/2024 04:31:04

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.192	04:26
Breath Volume	2.933	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.185	04:29
Breath Volume	2.062	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:31

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENT,ROBERT,T
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LEDBETTER,CALEB
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: ARCENEAX

DATE/TIME OF TEST: 07/09/2024 22:41:03

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.082	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.191	22:36
Breath Volume	1.464	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:38
Subject Sample # 2	0.196	22:39
Breath Volume	1.273	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:41

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COTTON,PIERCE,A

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: BECKER,JONATHAN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221783

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 23:35

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 07/10/2024 01:03:45

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.082	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.000	00:58
Breath Volume	1.660	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.000	01:02
Breath Volume	2.125	01:02
Air Blank	0.000	01:02
Control Test	0.082	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY-IVERSEN, MIYA, K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SIPHER, VICTORIA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: WICKERSHAM

DATE/TIME OF TEST: 07/11/2024 19:50:21

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.082	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.075	19:45
Breath Volume	2.117	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.071	19:48
Breath Volume	1.675	19:48
Air Blank	0.000	19:49
Control Test	0.081	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: THOMPSON,GREGORY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: STOVALL

DATE/TIME OF TEST: 07/12/2024 23:07:51

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.082	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.000	23:02
Breath Volume	2.804	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.000	23:06
Breath Volume	3.437	23:06
Air Blank	0.000	23:06
Control Test	0.082	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASWELL,AUSTIN,J

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: MAYFIELD,MATTHEW

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221783

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 01:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 07/20/2024 02:59:21

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.274	02:51
Breath Volume	2.121	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.243	02:54
Breath Volume	1.257	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 3	0.254	02:57
Breath Volume	1.343	02:57
Air Blank	0.000	02:58
Control Test	0.078	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: POTTS,MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: KRAUSE

DATE/TIME OF TEST: 07/21/2024 03:23:08

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.082	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.149	03:18
Breath Volume	1.792	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.146	03:21
Breath Volume	1.933	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 07/31/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 07/04/2024 21:52:13

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.175	21:47
Breath Volume	2.199	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:49
Subject Sample # 2	0.168	21:50
Breath Volume	2.683	21:50
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 08/06/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 07/05/2024 02:09:15

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.167	02:04
Breath Volume	1.703	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:06
Subject Sample # 2	0.159	02:07
Breath Volume	1.460	02:07
Air Blank	0.000	02:08
Control Test	0.078	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 08/06/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK, I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/05/2024 02:25:10

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.219	02:20
Breath Volume	1.566	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.223	02:23
Breath Volume	1.214	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 08/06/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,B,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/07/2024 01:06:19

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.078	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.062	01:01
Breath Volume	1.664	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.062	01:04
Breath Volume	2.070	01:04
Air Blank	0.000	01:05
Control Test	0.077	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 08/06/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SIMPSON,JAMES
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: DICKINSON

DATE/TIME OF TEST: 07/01/2024 00:58:31

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1 0.157	00:46	
Breath Volume	1.750	00:46	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2 0.158	00:52	Volume Not Met
Breath Volume	0.257	00:52	
Air Blank	0.000	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 3 0.167	00:56	
Breath Volume	1.585	00:56	
Air Blank	0.000	00:57	
Control Test	0.077	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 07/03/2024 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTI,CHRISTIAN,P
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ROSS,CHRISTIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: BUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: GATTI

DATE/TIME OF TEST: 07/04/2024 17:34:07

Diagnostics Check	OK	17:25	
Air Blank	0.000	17:25	
Control Test	0.080	17:26	
Air Blank	0.000	17:26	
Subject Sample	# 1 0.144	17:29	
Breath Volume	1.558	17:29	
Air Blank	0.000	17:30	
Air Blank	0.000	17:31	
Subject Sample	# 2	17:32	Subject Test Refused
Breath Volume	0.000	17:32	
Air Blank	0.000	17:33	
Control Test	0.076	17:33	
Air Blank	0.000	17:34	
Diagnostics Check	OK	17:34	

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: FALCON

DATE/TIME OF TEST: 07/04/2024 19:21:43

Diagnostics Check	OK	19:15
Air Blank	0.000	19:15
Control Test	0.081	19:15
Air Blank	0.000	19:16
Subject Sample # 1	0.167	19:17
Breath Volume	3.308	19:17
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.156	19:20
Breath Volume	1.800	19:20
Air Blank	0.000	19:20
Control Test	0.079	19:21
Air Blank	0.000	19:21
Diagnostics Check	OK	19:21

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARGENT,RIANA,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SARGENT,RIANA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: BUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SOCARRAS MEDEROS

DATE/TIME OF TEST: 07/04/2024 19:34:59

Diagnostics Check	OK	19:34	
Air Blank		19:34	RFI Detect
Air Blank		19:34	RFI Detect

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARGENT,RIANA,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SARGENT,RIANA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: BUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: SOCARRAS MEDEROS

DATE/TIME OF TEST: 07/04/2024 19:37:09

Diagnostics Check	OK	19:36	
Air Blank		19:37	RFI Detect
Air Blank		19:37	RFI Detect

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARGENT,RIANA,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SARGENT,RIANA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: BUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: SOCARRAS MEDEROS

DATE/TIME OF TEST: 07/04/2024 19:46:12

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.081	19:40
Air Blank	0.000	19:41
Subject Sample # 1	0.140	19:41
Breath Volume	1.484	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.136	19:44
Breath Volume	1.527	19:44
Air Blank	0.000	19:45
Control Test	0.079	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HERNANDEZ,MATTHEW
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BARANDIARAN GOMEZ

DATE/TIME OF TEST: 07/04/2024 23:35:32

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:33	
Control Test	0.080	23:33	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI, RYAN, P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BOURCIER, ALEXIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: FITZPATRICK

DATE/TIME OF TEST: 07/06/2024 01:22:50

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:20	
Control Test	0.081	01:20	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:21	
Control Test	0.081	01:22	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SEGARIA

DATE/TIME OF TEST: 07/06/2024 03:07:13

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.151	03:01
Breath Volume	1.378	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.146	03:05
Breath Volume	1.691	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 07/10/2024 23:36:37

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:34	
Control Test	0.080	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Control Test	0.080	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: PROUT

DATE/TIME OF TEST: 07/13/2024 22:20:38

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:18	
Control Test	0.081	22:18	
Air Blank	0.000	22:18	
Subject Sample	# 1	22:19	Subject Test Refused
Breath Volume	0.000	22:19	
Air Blank	0.000	22:19	
Control Test	0.081	22:20	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BETZ,FRANK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: SCHELLBERG

DATE/TIME OF TEST: 07/14/2024 16:17:50

Diagnostics Check	OK	16:10
Air Blank	0.000	16:10
Control Test	0.080	16:11
Air Blank	0.000	16:11
Subject Sample # 1	0.264	16:12
Breath Volume	1.464	16:12
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.267	16:16
Breath Volume	3.453	16:16
Air Blank	0.000	16:16
Control Test	0.077	16:17
Air Blank	0.000	16:17
Diagnostics Check	OK	16:17

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/16/2024 04:12:22

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.031	04:07
Breath Volume	1.210	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.028	04:10
Breath Volume	1.386	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 08/01/2024 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ANTONOVA,KRISTINA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 07/16/2024 21:05:28

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.081	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.253	21:00
Breath Volume	2.277	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:02
Subject Sample # 2	0.262	21:03
Breath Volume	2.960	21:03
Air Blank	0.000	21:04
Control Test	0.079	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JAMES
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MANDEL

DATE/TIME OF TEST: 07/18/2024 02:52:29

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.161	02:47
Breath Volume	4.464	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.156	02:50
Breath Volume	2.761	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: VILCHEZ

DATE/TIME OF TEST: 07/19/2024 22:19:59

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:17	
Control Test	0.081	22:17	
Air Blank	0.000	22:18	
Subject Sample	# 1	22:18	Subject Test Refused
Breath Volume	0.000	22:18	
Air Blank	0.000	22:19	
Control Test	0.081	22:19	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/20/2024 04:05:42

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.080	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: GLADDEN

DATE/TIME OF TEST: 07/20/2024 21:46:22

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:43	
Control Test	0.081	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:44	Subject Test Refused
Breath Volume	0.000	21:44	
Air Blank	0.000	21:45	
Control Test	0.080	21:45	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BARDALES

DATE/TIME OF TEST: 07/20/2024 22:23:58

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.081	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.011	22:19
Breath Volume	1.917	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.010	22:22
Breath Volume	1.515	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE',JOSHUA',C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: UA,CAMERON
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: MIRASHOV

DATE/TIME OF TEST: 07/22/2024 06:07:27

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.081	06:01
Air Blank	0.000	06:01
Subject Sample # 1	0.141	06:02
Breath Volume	1.273	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:04
Subject Sample # 2	0.137	06:05
Breath Volume	3.308	06:05
Air Blank	0.000	06:06
Control Test	0.080	06:06
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE',JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: AGEE

DATE/TIME OF TEST: 07/24/2024 23:35:55

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:33	
Control Test	0.081	23:33	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:35	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE',JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 07/25/2024 22:56:08

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.081	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.115	22:51
Breath Volume	2.210	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:53
Subject Sample # 2	0.113	22:54
Breath Volume	2.152	22:54
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BRABLC,MICHAL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 08:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: QUICHE LOPEZ

DATE/TIME OF TEST: 07/27/2024 09:19:14

Diagnostics Check	OK	09:12
Air Blank	0.000	09:12
Control Test	0.080	09:13
Air Blank	0.000	09:13
Subject Sample # 1	0.182	09:14
Breath Volume	2.265	09:14
Air Blank	0.000	09:15
Air Blank	0.000	09:17
Subject Sample # 2	0.184	09:17
Breath Volume	1.855	09:17
Air Blank	0.000	09:18
Control Test	0.078	09:18
Air Blank	0.000	09:19
Diagnostics Check	OK	09:19

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ROBINSON,TYREE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: TIPPS

DATE/TIME OF TEST: 07/27/2024 18:47:26

Diagnostics Check	OK	18:44	
Air Blank	0.000	18:44	
Control Test	0.080	18:45	
Air Blank	0.000	18:45	
Subject Sample	# 1	18:46	Subject Test Refused
Breath Volume	0.000	18:46	
Air Blank	0.000	18:46	
Control Test	0.079	18:46	
Air Blank	0.000	18:47	
Diagnostics Check	OK	18:47	

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ANTONOVA,KRISTINA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: RIVERA ZEPEDA

DATE/TIME OF TEST: 07/30/2024 01:10:02

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Control Test	0.080	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 08/01/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO-MARTINEZ, KEVEN,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ARDREY-GIGUERE, DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: THARP

DATE/TIME OF TEST: 07/04/2024 23:54:38

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.081	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:53	
Control Test	0.081	23:54	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 08/05/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODDY,AARON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RODDY,AARON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 13:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: PIENSCHKE

DATE/TIME OF TEST: 07/05/2024 14:11:44

Diagnostics Check	OK	13:59	
Air Blank	0.000	13:59	
Control Test	0.080	13:59	
Air Blank	0.000	14:00	
Subject Sample	# 1 0.276	14:03	Volume Not Met
Breath Volume	0.640	14:03	
Air Blank	0.000	14:04	
Air Blank	0.000	14:05	
Subject Sample	# 2 0.327	14:06	
Breath Volume	1.285	14:06	
Air Blank	0.000	14:07	
Air Blank	0.000	14:09	
Subject Sample	# 3 0.329	14:09	
Breath Volume	1.878	14:09	
Air Blank	0.000	14:10	
Control Test	0.076	14:11	
Air Blank	0.000	14:11	
Diagnostics Check	OK	14:11	

Data Download Date/Time: 08/05/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MUMPER,REBECCA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MODICA

DATE/TIME OF TEST: 07/05/2024 23:36:59

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:34	
Air Blank	0.000	23:35	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:36	
Control Test	0.081	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 08/05/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: LUACES ABREU

DATE/TIME OF TEST: 07/06/2024 01:25:07

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.080	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 08/05/2024 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: RUDY

DATE/TIME OF TEST: 07/06/2024 20:05:17

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:02	
Control Test	0.081	20:03	
Air Blank	0.000	20:03	
Subject Sample	# 1	20:03	Subject Test Refused
Breath Volume	0.000	20:03	
Air Blank	0.000	20:04	
Control Test	0.080	20:04	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 08/05/2024 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MUMPER,REBECCA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: NEALIS

DATE/TIME OF TEST: 07/06/2024 21:35:17

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:32	
Control Test	0.080	21:33	
Air Blank	0.000	21:33	
Subject Sample	# 1	21:33	Subject Test Refused
Breath Volume	0.000	21:33	
Air Blank	0.000	21:34	
Control Test	0.080	21:34	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 08/05/2024 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MAURY

DATE/TIME OF TEST: 07/07/2024 03:53:28

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 08/05/2024 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRAW,KAELEE,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MCGRAW,KAELEE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: SALIGA

DATE/TIME OF TEST: 07/11/2024 23:30:10

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.080	23:28	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:29	
Control Test	0.080	23:29	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 08/05/2024 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RODDA, TIFFANY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DICKERSON

DATE/TIME OF TEST: 07/11/2024 23:48:08

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.215	23:43
Breath Volume	3.863	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.215	23:46
Breath Volume	3.105	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 08/05/2024 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: PETTORINI

DATE/TIME OF TEST: 07/12/2024 03:47:37

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.138	03:43
Breath Volume	3.589	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.133	03:45
Breath Volume	3.101	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 08/05/2024 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL CUETO, ERYKA, L
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DEL CUETO, ERYKA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/20/2024 17:13:53

Diagnostics Check	OK	17:07
Air Blank	0.000	17:07
Control Test	0.081	17:08
Air Blank	0.000	17:08
Subject Sample # 1	0.331	17:09
Breath Volume	2.835	17:09
Air Blank	0.000	17:09
Air Blank	0.000	17:11
Subject Sample # 2	0.337	17:12
Breath Volume	2.582	17:12
Air Blank	0.000	17:12
Control Test	0.079	17:13
Air Blank	0.000	17:13
Diagnostics Check	OK	17:13

Data Download Date/Time: 08/05/2024 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: CARDONA QUINTANA

DATE/TIME OF TEST: 07/21/2024 03:21:53

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.080	03:15
Breath Volume	1.339	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.087	03:20
Breath Volume	2.066	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 08/05/2024 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/21/2024 04:27:39

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.080	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Control Test	0.080	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 08/05/2024 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 07/22/2024 03:07:14

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.018	03:02
Breath Volume	2.218	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.016	03:05
Breath Volume	1.539	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 08/05/2024 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DULLY,STEPHEN,W
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ARDREY-GIGUERE,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: GAINERY

DATE/TIME OF TEST: 07/23/2024 00:07:34

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.081	00:05	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 08/05/2024 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 07/25/2024 01:36:47

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.081	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 08/05/2024 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: LINDBERG

DATE/TIME OF TEST: 07/28/2024 04:27:44

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.066	04:23
Breath Volume	1.175	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.063	04:26
Breath Volume	1.437	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 08/05/2024 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREITAS,MARCIA,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FREITAS,MARCIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/26/2024 05:47:40

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.143	05:42
Breath Volume	2.187	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.142	05:45
Breath Volume	2.957	05:45
Air Blank	0.000	05:46
Control Test	0.077	05:47
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

Data Download Date/Time: 08/25/2024 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN-FRANCOIS,LEEF,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: JEAN-FRANCOIS,LEEF
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 09:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: BROKERING

DATE/TIME OF TEST: 07/27/2024 10:59:50

Diagnostics Check	OK	10:53
Air Blank	0.000	10:53
Control Test	0.079	10:53
Air Blank	0.000	10:54
Subject Sample # 1	0.240	10:54
Breath Volume	3.531	10:54
Air Blank	0.000	10:55
Air Blank	0.000	10:57
Subject Sample # 2	0.242	10:57
Breath Volume	2.804	10:57
Air Blank	0.000	10:58
Control Test	0.077	10:59
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

Data Download Date/Time: 08/25/2024 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JESSE,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BROWN,JESSE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 07/28/2024 02:17:11

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.135	02:12
Breath Volume	1.617	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.139	02:15
Breath Volume	3.652	02:15
Air Blank	0.000	02:16
Control Test	0.077	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 08/25/2024 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: KOSKO

DATE/TIME OF TEST: 07/30/2024 00:25:26

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.080	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1 0.181	00:24	
Breath Volume	1.933	00:24	
Air Blank		00:24	Purge Fail
Air Blank	0.000	00:25	

Data Download Date/Time: 08/25/2024 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KOSKO

DATE/TIME OF TEST: 07/30/2024 00:34:02

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.077	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.173	00:29
Breath Volume	1.222	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:31
Subject Sample # 2	0.166	00:32
Breath Volume	1.523	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:34

Data Download Date/Time: 08/25/2024 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTIAN,O
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FITCH,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: KISH

DATE/TIME OF TEST: 07/03/2024 01:18:19

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.143	01:13
Breath Volume	2.011	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.127	01:16
Breath Volume	1.171	01:16
Air Blank	0.000	01:17
Control Test	0.077	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERT,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PEREZ,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/04/2024 02:07:34

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.135	02:01
Breath Volume	1.453	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample # 2	0.123	02:05
Breath Volume	1.609	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSTANCE,WADE,O
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: CONSTANCE,WADE
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: BUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: DREADIN

DATE/TIME OF TEST: 07/05/2024 01:18:28

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.000	01:17	RFI Detect
Air Blank	0.000	01:18	

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSTANCE,WADE,O
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: CONSTANCE,WADE
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: BUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: DREADIN

DATE/TIME OF TEST: 07/05/2024 01:27:26

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.085	01:22
Breath Volume	2.851	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.085	01:25
Breath Volume	4.312	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 07/06/2024 01:35:01

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.201	01:30
Breath Volume	2.179	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.202	01:33
Breath Volume	1.707	01:33
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:35

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,V,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: COUGHLIN,M
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BOWLER

DATE/TIME OF TEST: 07/06/2024 03:57:43

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.145	03:52
Breath Volume	2.386	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.151	03:55
Breath Volume	2.246	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETTS,J,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 07/06/2024 20:44:06

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.079	20:31	
Air Blank	0.000	20:32	
Subject Sample	# 1 0.124	20:35	Volume Not Met
Breath Volume	0.023	20:35	
Air Blank	0.000	20:36	
Air Blank	0.000	20:37	
Subject Sample	# 2 0.156	20:39	
Breath Volume	1.953	20:39	
Air Blank	0.000	20:40	
Air Blank	0.000	20:41	
Subject Sample	# 3 0.143	20:42	
Breath Volume	1.753	20:42	
Air Blank	0.000	20:43	
Control Test	0.075	20:43	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETTS,J,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ASHER

DATE/TIME OF TEST: 07/07/2024 02:49:48

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.079	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1 0.099	02:42	Volume Not Met
Breath Volume	0.058	02:42	
Air Blank	0.000	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 2 0.111	02:48	Volume Not Met
Breath Volume	0.074	02:48	
Air Blank	0.000	02:48	
Control Test	0.078	02:49	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: URONU,PATRICK,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SAN MIGUEL,FRANCISCO
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/07/2024 08:56:29

Diagnostics Check	OK	08:49
Air Blank	0.000	08:49
Control Test	0.079	08:50
Air Blank	0.000	08:50
Subject Sample # 1	0.128	08:51
Breath Volume	3.636	08:51
Air Blank	0.000	08:52
Air Blank	0.000	08:53
Subject Sample # 2	0.120	08:54
Breath Volume	3.687	08:54
Air Blank	0.000	08:55
Control Test	0.078	08:55
Air Blank	0.000	08:56
Diagnostics Check	OK	08:56

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSTANCE,WADE,O
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: CONSTANCE,WADE
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: BUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/07/2024 18:33:34

Diagnostics Check	OK	18:32	
Air Blank	0.000	18:32	
Control Test	0.079	18:32	
Air Blank		18:33	RFI Detect
Air Blank	0.000	18:33	

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZEZUSKI,JIMMY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZEZUSKI,JIMMY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 07/07/2024 20:18:28

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.078	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.153	20:13
Breath Volume	1.519	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.148	20:16
Breath Volume	1.167	20:16
Air Blank	0.000	20:17
Control Test	0.076	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCELDFOWNEY,STEPHEN,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MCELDFOWNEY,STEPHEN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BERMUDEZ

DATE/TIME OF TEST: 07/08/2024 02:13:36

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.078	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.191	02:08
Breath Volume	1.109	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.180	02:11
Breath Volume	1.175	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BREANA,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: GASKEY

DATE/TIME OF TEST: 07/11/2024 00:47:26

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:36	
Air Blank	0.000	00:36	
Subject Sample	# 1 0.149	00:39	Volume Not Met
Breath Volume	0.035	00:39	
Air Blank	0.000	00:40	
Air Blank	0.000	00:42	
Subject Sample	# 2 0.161	00:45	Volume Not Met
Breath Volume	0.066	00:45	
Air Blank	0.000	00:46	
Control Test	0.077	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BREANA,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: GASKEY

DATE/TIME OF TEST: 07/11/2024 01:02:58

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.075	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 1 0.144	00:55	Volume Not Met
Breath Volume	0.148	00:55	
Air Blank	0.000	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 2 0.134	01:01	Volume Not Met
Breath Volume	0.082	01:01	
Air Blank	0.000	01:01	
Control Test	0.076	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BREANA,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: GASKEY

DATE/TIME OF TEST: 07/11/2024 01:17:18

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.076	01:06	
Air Blank	0.000	01:06	
Subject Sample	# 1 0.116	01:09	Volume Not Met
Breath Volume	0.613	01:09	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample	# 2 0.137	01:15	Volume Not Met
Breath Volume	0.093	01:15	
Air Blank	0.000	01:16	
Control Test	0.074	01:16	Control Outside Tolerance
Air Blank	0.000	01:17	

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCELLOWNEY,STEPHEN,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MCELLOWNEY,STEPHEN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/11/2024 04:48:48

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.183	04:44
Breath Volume	3.039	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.186	04:47
Breath Volume	3.761	04:47
Air Blank	0.000	04:47
Control Test	0.079	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREITAS,MARCIA,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FREITAS,MARCIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: VANPELT

DATE/TIME OF TEST: 07/11/2024 21:02:31

Diagnostics Check	OK	20:59	
Air Blank	0.000	21:00	
Control Test	0.080	21:00	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:01	RFI Detect
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Control Test	0.079	21:01	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREITAS,MARCIA,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FREITAS,MARCIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: VANPELT

DATE/TIME OF TEST: 07/11/2024 21:05:43

Diagnostics Check	OK	21:04	
Air Blank		21:05	RFI Detect
Air Blank	0.000	21:05	

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREITAS,MARCIA\,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FREITAS,MARCIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: VANPELT

DATE/TIME OF TEST: 07/11/2024 21:14:10

Diagnostics Check	OK	21:07
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Subject Sample # 1	0.104	21:09
Breath Volume	1.242	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample # 2	0.104	21:12
Breath Volume	1.136	21:12
Air Blank	0.000	21:13
Control Test	0.078	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEASEL,KELSEY,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PEREZ,ALEXANDER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 07/13/2024 19:53:25

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.244	19:48
Breath Volume	1.851	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.240	19:51
Breath Volume	1.371	19:51
Air Blank	0.000	19:52
Control Test	0.077	19:52
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUTRAL,WILLIAM,Z
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FUTRAL,WILLIAM
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 07/14/2024 01:52:30

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.184	01:47
Breath Volume	4.296	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.183	01:50
Breath Volume	1.964	01:50
Air Blank	0.000	01:51
Control Test	0.077	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,SPENSER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KENNEDY,SPENSER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SELLER

DATE/TIME OF TEST: 07/14/2024 02:53:41

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.080	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,BRYAN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOWMAN,BRYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 10:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:57

SUBJECT'S LAST NAME: BELLACH

DATE/TIME OF TEST: 07/15/2024 11:25:28

Diagnostics Check	OK	11:18
Air Blank	0.000	11:18
Control Test	0.079	11:19
Air Blank	0.000	11:19
Subject Sample # 1	0.212	11:20
Breath Volume	1.613	11:20
Air Blank	0.000	11:21
Air Blank	0.000	11:22
Subject Sample # 2	0.222	11:23
Breath Volume	1.390	11:23
Air Blank	0.000	11:24
Control Test	0.078	11:24
Air Blank	0.000	11:25
Diagnostics Check	OK	11:25

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,BRANDON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: ROBERTS,BRANDON
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: BUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 07/18/2024 01:20:44

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.165	01:15
Breath Volume	2.269	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.171	01:18
Breath Volume	2.199	01:18
Air Blank	0.000	01:19
Control Test	0.078	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRINGER,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: STRINGER,MICHAEL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: LEANDRI DE JESUS

DATE/TIME OF TEST: 07/18/2024 19:42:36

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.079	19:38
Breath Volume	2.972	19:38
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.079	19:40
Breath Volume	2.785	19:40
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,JOSHUA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PARKER,JOSHUA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KATZ

DATE/TIME OF TEST: 07/20/2024 00:53:20

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.200	00:48
Breath Volume	2.035	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:50
Subject Sample # 2	0.198	00:51
Breath Volume	1.269	00:51
Air Blank	0.000	00:52
Control Test	0.077	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY, RONALD, J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BAILEY, RONALD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CAVE

DATE/TIME OF TEST: 07/20/2024 03:24:36

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.167	03:19
Breath Volume	1.968	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.162	03:22
Breath Volume	2.808	03:22
Air Blank	0.000	03:23
Control Test	0.077	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/25/2024 17:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AULTMAN,ANDREW,A
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: LAMBERT,CODY
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: BUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/20/2024 22:16:03

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.170	22:10
Breath Volume	4.503	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.164	22:14
Breath Volume	5.089	22:14
Air Blank	0.000	22:15
Control Test	0.077	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:16

Data Download Date/Time: 07/25/2024 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CASTELLANOS CARUS

DATE/TIME OF TEST: 07/20/2024 23:19:08

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.252	23:14
Breath Volume	3.175	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:16
Subject Sample # 2	0.249	23:17
Breath Volume	1.968	23:17
Air Blank	0.000	23:18
Control Test	0.076	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 07/25/2024 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,RYAN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEAVER,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 07/21/2024 04:04:53

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:53	
Control Test	0.079	03:53	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 2	04:03	No Sample Provided
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 07/25/2024 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY, RONALD, J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BAILEY, RONALD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: VETTER

DATE/TIME OF TEST: 07/21/2024 23:01:47

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.232	22:56
Breath Volume	3.078	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.225	22:59
Breath Volume	3.507	22:59
Air Blank	0.000	23:00
Control Test	0.077	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 07/25/2024 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUELLER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SCHUELLER,ZACHARY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 07/23/2024 00:19:41

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.109	00:14
Breath Volume	1.269	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.109	00:17
Breath Volume	2.253	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 07/25/2024 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/24/2024 03:52:20

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.191	03:47
Breath Volume	2.167	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.184	03:50
Breath Volume	2.664	03:50
Air Blank	0.000	03:51
Control Test	0.076	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/25/2024 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,JUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TISDALE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/03/2024 03:35:41

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.082	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:34	
Control Test	0.083	03:35	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 07/10/2024 01:42:36

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.169	01:37
Breath Volume	3.242	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.157	01:40
Breath Volume	1.699	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GALLOWAY

DATE/TIME OF TEST: 07/12/2024 22:11:52

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.082	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.000	22:07
Breath Volume	1.359	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.000	22:10
Breath Volume	1.738	22:10
Air Blank	0.000	22:10
Control Test	0.082	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WHEELER,KYRA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/13/2024 00:04:33

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.082	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.225	23:57
Breath Volume	1.726	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 3	0.204	23:59
Breath Volume	1.851	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.203	00:02
Breath Volume	1.414	00:02
Air Blank	0.000	00:03
Control Test	0.082	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: FLORES GUADARRAMA

DATE/TIME OF TEST: 07/14/2024 05:02:31

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.082	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.177	04:58
Breath Volume	2.738	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.179	05:00
Breath Volume	1.566	05:00
Air Blank	0.000	05:01
Control Test	0.082	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: TRUJILLO HURTADO

DATE/TIME OF TEST: 07/15/2024 03:32:54

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.082	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.194	03:26
Breath Volume	2.578	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.187	03:31
Breath Volume	1.726	03:31
Air Blank	0.000	03:31
Control Test	0.081	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCALIA, ADAM, C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCALIA, ADAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: S530515542909

DATE/TIME OF TEST: 07/16/2024 18:37:22

Diagnostics Check	OK	18:34	
Air Blank	0.000	18:34	
Control Test	0.082	18:35	
Air Blank	0.000	18:35	
Subject Sample	# 1	18:35	Subject Test Refused
Breath Volume	0.000	18:35	
Air Blank	0.000	18:36	
Control Test	0.082	18:36	
Air Blank	0.000	18:37	
Diagnostics Check	OK	18:37	

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDENAS,A,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCMILLIAN,R
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/18/2024 11:02:12

Diagnostics Check	OK	10:55
Air Blank	0.000	10:56
Control Test	0.082	10:56
Air Blank	0.000	10:56
Subject Sample # 1	0.224	10:57
Breath Volume	3.207	10:57
Air Blank	0.000	10:58
Air Blank	0.000	10:59
Subject Sample # 2	0.207	11:00
Breath Volume	2.679	11:00
Air Blank	0.000	11:01
Control Test	0.083	11:01
Air Blank	0.000	11:02
Diagnostics Check	OK	11:02

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,JUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TISDALE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: KINNAIRD

DATE/TIME OF TEST: 07/19/2024 00:18:11

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.083	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.177	00:12
Breath Volume	1.183	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.179	00:16
Breath Volume	1.394	00:16
Air Blank	0.000	00:17
Control Test	0.083	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: NADERA,ANTONIO
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: GARCIA VEGA

DATE/TIME OF TEST: 07/22/2024 16:57:33

Diagnostics Check	OK	16:50
Air Blank	0.000	16:51
Control Test	0.082	16:51
Air Blank	0.000	16:51
Subject Sample # 1	0.020	16:52
Breath Volume	2.308	16:52
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.018	16:55
Breath Volume	2.515	16:55
Air Blank	0.000	16:56
Control Test	0.084	16:56
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 07/23/2024 20:52:19

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:49	
Control Test	0.083	20:50	
Air Blank	0.000	20:50	
Subject Sample	# 1	20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:51	
Control Test	0.082	20:51	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 07/26/2024 23:14:03

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.083	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.083	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SANTOS,PAUL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SNETMAN

DATE/TIME OF TEST: 07/28/2024 01:49:39

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.083	01:43	
Air Blank	0.000	01:43	
Subject Sample	# 1 0.105	01:44	
Breath Volume	1.429	01:44	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample	# 2 0.101	01:47	
Breath Volume	1.511	01:47	
Air Blank	0.000	01:48	
Control Test	0.086	01:48	Control Outside Tolerance
Air Blank	0.000	01:49	

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: ROBLEDO VAZQUEZ

DATE/TIME OF TEST: 07/28/2024 18:35:53

Diagnostics Check	OK	18:32	
Air Blank	0.000	18:33	
Control Test	0.083	18:33	
Air Blank	0.000	18:34	
Subject Sample	# 1	18:34	Subject Test Refused
Breath Volume	0.000	18:34	
Air Blank	0.000	18:34	
Control Test	0.082	18:35	
Air Blank	0.000	18:35	
Diagnostics Check	OK	18:35	

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: PETER FRANCIS

DATE/TIME OF TEST: 07/29/2024 04:32:19

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.082	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.223	04:27
Breath Volume	1.863	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.216	04:30
Breath Volume	1.867	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 07/30/2024 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELS,JESSICA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: BALLADARES,A
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: DEBROSSE

DATE/TIME OF TEST: 07/01/2024 11:22:44

Diagnostics Check	OK	11:15	
Air Blank	0.000	11:15	
Control Test	0.079	11:16	
Air Blank	0.000	11:16	
Subject Sample	# 1 0.184	11:17	
Breath Volume	1.496	11:17	
Air Blank	0.000	11:18	
Air Blank	0.000	11:19	
Subject Sample	# 2 0.187	11:20	
Breath Volume	1.972	11:20	
Air Blank	0.000	11:21	
Control Test	0.079	11:21	
Air Blank		11:22	RFI Detect
Air Blank	0.000	11:22	

Data Download Date/Time: 07/23/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: GIGUERE

DATE/TIME OF TEST: 07/05/2024 02:30:52

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 07/23/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,ZAHIRYD,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: RODRIGUEZ,Y
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: MCGREGOR

DATE/TIME OF TEST: 07/06/2024 22:43:59

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:32	
Control Test	0.079	22:33	
Air Blank	0.000	22:33	
Subject Sample	# 1 0.171	22:36	Volume Not Met
Breath Volume	0.082	22:36	
Air Blank	0.000	22:37	
Air Blank	0.000	22:39	
Subject Sample	# 2 0.156	22:42	Volume Not Met
Breath Volume	0.445	22:42	
Air Blank	0.000	22:43	
Control Test	0.077	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 07/23/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SOUDERS,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: VILSON LAYNE

DATE/TIME OF TEST: 07/09/2024 17:17:54

Diagnostics Check	OK	17:03	
Air Blank	0.000	17:04	
Control Test	0.080	17:04	
Air Blank	0.000	17:05	
Subject Sample	# 1 0.316	17:08	Volume Not Met
Breath Volume	0.140	17:08	
Air Blank	0.000	17:09	
Air Blank	0.000	17:10	
Subject Sample	# 2 0.336	17:12	
Breath Volume	1.871	17:12	
Air Blank	0.000	17:13	
Air Blank	0.000	17:15	
Subject Sample	# 3 0.326	17:16	
Breath Volume	1.558	17:16	
Air Blank	0.000	17:16	
Control Test	0.075	17:17	
Air Blank	0.000	17:17	
Diagnostics Check	OK	17:17	

Data Download Date/Time: 07/23/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: GOETZ

DATE/TIME OF TEST: 07/09/2024 17:28:53

Diagnostics Check	OK	17:26	
Air Blank	0.000	17:26	
Control Test	0.078	17:26	
Air Blank	0.000	17:27	
Subject Sample	# 1	17:27	Subject Test Refused
Breath Volume	0.000	17:27	
Air Blank	0.000	17:27	
Control Test	0.078	17:28	
Air Blank	0.000	17:28	
Diagnostics Check	OK	17:28	

Data Download Date/Time: 07/23/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,ZAHIRYD,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: HENDRICKS,B
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 07/11/2024 22:58:59

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.079	22:56	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:58	
Control Test	0.079	22:58	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 07/23/2024 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,J
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: JENSEN,K
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: ESCAMILLA

DATE/TIME OF TEST: 07/15/2024 03:37:33

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 07/23/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COTTON,PIERCE,A

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COTTON,PIERCE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222522

DRY GAS EXPIRY DATE: 05/19/2026

VIOLATION CODE: DUI

ARREST TIME: 22:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: TOOLEY

DATE/TIME OF TEST: 07/31/2024 23:36:56

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.000	23:32
Breath Volume	2.382	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.000	23:35
Breath Volume	2.335	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 08/30/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: ELLIS,ALEXANDER,	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ELLIS,ALEXANDER	ARREST TIME: 23:06
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/26/2024	OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WILLIFORD	DATE/TIME OF TEST: 07/03/2024	23:25:47
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:13	
Control Test	0.078	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 1 0.135	23:17	
Breath Volume	1.261	23:17	
Air Blank	0.000	23:17	
Air Blank	0.000	23:19	
Subject Sample	# 2 0.110	23:22	Volume Not Met
Breath Volume	0.488	23:22	
Air Blank	0.000	23:23	
Air Blank		23:25	Ambient Fail
Air Blank	0.000	23:25	

Data Download Date/Time: 07/10/2024 22:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: ELLIS,ALEXANDER,	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ELLIS,ALEXANDER	ARREST TIME: 23:06
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/26/2024	OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WILLIFORD	DATE/TIME OF TEST: 07/03/2024 23:31:17
--------------------------------	--

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.077	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:30	
Control Test	0.076	23:30	
Air Blank		23:31	RFI Detect
Air Blank		23:31	RFI Detect

Data Download Date/Time: 07/10/2024 22:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALIJAGIC,SAMIR,

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: ALIJAGIC,SAMIR

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 00:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 07/11/2024 01:10:35

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.100	01:06
Breath Volume	1.789	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.100	01:09
Breath Volume	2.203	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 08/07/2024 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: KOCHER,MEGHAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/16/2024 04:38:03

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.000	04:33
Breath Volume	3.972	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.000	04:36
Breath Volume	2.480	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:38

Data Download Date/Time: 08/07/2024 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 07/17/2024 20:48:02

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.080	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.172	20:43
Breath Volume	2.003	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:45
Subject Sample # 2	0.181	20:46
Breath Volume	2.121	20:46
Air Blank	0.000	20:47
Control Test	0.079	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:48

Data Download Date/Time: 08/07/2024 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: MILLS,SCOTT,T	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MILLS,SCOTT	ARREST TIME: 01:09
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Female
LAST AGENCY INSPECTION: 07/10/2024	OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: PARRISH DATE/TIME OF TEST: 07/19/2024 02:11:59

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.121	02:06
Breath Volume	1.593	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.121	02:10
Breath Volume	1.707	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 08/07/2024 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: RAPPERPORT

DATE/TIME OF TEST: 07/20/2024 04:22:00

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.080	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:20	RFI Detect
Breath Volume	0.000	04:20	
Air Blank	0.000	04:21	
Control Test	0.000	04:21	RFI Detect
Air Blank	0.000	04:22	

Data Download Date/Time: 08/07/2024 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SHERROD

DATE/TIME OF TEST: 07/23/2024 22:13:20

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.107	22:08
Breath Volume	2.753	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.100	22:11
Breath Volume	1.925	22:11
Air Blank	0.000	22:12
Control Test	0.079	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 08/07/2024 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/28/2024 04:13:20

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.109	04:08
Breath Volume	3.359	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.111	04:11
Breath Volume	3.203	04:11
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 08/07/2024 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/29/2024 00:50:41

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.181	00:46
Breath Volume	2.523	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.184	00:49
Breath Volume	2.628	00:49
Air Blank	0.000	00:49
Control Test	0.078	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 08/07/2024 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: ALIJAGIC,SAMIR	ARREST TIME: 21:20
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 07/10/2024	OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: RICHARD	DATE/TIME OF TEST: 07/29/2024	22:54:17
------------------------------	-------------------------------	----------

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.234	22:49
Breath Volume	2.167	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.233	22:52
Breath Volume	1.515	22:52
Air Blank	0.000	22:53
Control Test	0.078	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 08/07/2024 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SLOCUM,T
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BOWDEN

DATE/TIME OF TEST: 07/06/2024 02:21:32

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.076	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.000	02:17
Breath Volume	3.238	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.000	02:20
Breath Volume	2.484	02:20
Air Blank	0.000	02:20
Control Test	0.076	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 07/10/2024 22:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: AUCH,DOMINIK
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: LEBER

DATE/TIME OF TEST: 07/07/2024 00:35:20

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.076	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1 0.183	00:29	
Breath Volume	1.492	00:29	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 2 0.202	00:33	
Breath Volume	3.687	00:33	
Air Blank	0.000	00:33	
Control Test	0.076	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	Fail	00:34	Internal Printer Test
Air Blank	0.000	00:35	

Data Download Date/Time: 07/10/2024 22:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: AUCH,DOMINIK
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: LEBER

DATE/TIME OF TEST: 07/07/2024 00:43:59

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.077	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.195	00:39
Breath Volume	2.101	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.183	00:42
Breath Volume	1.949	00:42
Air Blank	0.000	00:43
Control Test	0.075	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 07/10/2024 22:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/19/2024 01:33:08

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.076	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.142	01:28
Breath Volume	2.945	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.137	01:31
Breath Volume	2.832	01:31
Air Blank	0.000	01:32
Control Test	0.077	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 08/07/2024 21:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402816921

DRY GAS EXPIRY DATE: 08/09/2026

VIOLATION CODE: DUI

ARREST TIME: 01:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: HARBERT

DATE/TIME OF TEST: 07/25/2024 02:09:27

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.075	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 1 0.227	02:02	
Breath Volume	3.503	02:02	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample	# 2 0.192	02:05	
Breath Volume	1.742	02:05	
Air Blank	0.000	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 3 0.238	02:08	
Breath Volume	3.617	02:08	
Air Blank		02:08	RFI Detect
Air Blank	0.000	02:09	

Data Download Date/Time: 08/07/2024 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: HARBERT

DATE/TIME OF TEST: 07/25/2024 02:17:51

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.076	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.228	02:12
Breath Volume	1.398	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.233	02:16
Breath Volume	2.808	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 08/07/2024 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BLOUNT

DATE/TIME OF TEST: 07/27/2024 02:40:00

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.075	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.000	02:35
Breath Volume	1.332	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.000	02:38
Breath Volume	1.894	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:40

Data Download Date/Time: 08/07/2024 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,DANIEL,D
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: RAMIREZ,ARMANDO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: TADEO

DATE/TIME OF TEST: 07/21/2024 18:13:16

Diagnostics Check	OK	17:59	
Air Blank	0.000	17:59	
Control Test	0.080	18:00	
Air Blank	0.000	18:00	
Subject Sample	# 1 0.249	18:03	Volume Not Met
Breath Volume	0.144	18:03	
Air Blank	0.000	18:04	
Air Blank	0.000	18:06	
Subject Sample	# 2 0.248	18:08	
Breath Volume	1.582	18:08	
Air Blank	0.000	18:09	
Air Blank	0.000	18:11	
Subject Sample	# 3 0.264	18:11	
Breath Volume	1.753	18:11	
Air Blank	0.000	18:12	
Control Test	0.075	18:12	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

Data Download Date/Time: 07/31/2024 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,DANIEL,D
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: VARGAS,DANIEL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: CEDILLO LOPEZ

DATE/TIME OF TEST: 07/28/2024 22:39:34

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.080	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1 0.236	22:30	Volume Not Met
Breath Volume	1.070	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 2 0.231	22:34	
Breath Volume	3.148	22:34	
Air Blank	0.000	22:35	
Air Blank	0.000	22:37	
Subject Sample	# 3 0.235	22:37	
Breath Volume	3.152	22:37	
Air Blank	0.000	22:38	
Control Test	0.076	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 07/31/2024 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MAJOR,BERNARD
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:33

SUBJECT'S LAST NAME: KAYE

DATE/TIME OF TEST: 07/06/2024 10:06:56

Diagnostics Check	OK	10:00
Air Blank	0.000	10:00
Control Test	0.080	10:01
Air Blank	0.000	10:01
Subject Sample # 1	0.209	10:02
Breath Volume	3.308	10:02
Air Blank	0.000	10:03
Air Blank	0.000	10:04
Subject Sample # 2	0.206	10:05
Breath Volume	3.218	10:05
Air Blank	0.000	10:05
Control Test	0.078	10:06
Air Blank	0.000	10:06
Diagnostics Check	OK	10:06

Data Download Date/Time: 08/07/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ESPINOSA-OCHOA

DATE/TIME OF TEST: 07/07/2024 23:10:09

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.266	23:05
Breath Volume	2.101	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample # 2	0.280	23:08
Breath Volume	2.589	23:08
Air Blank	0.000	23:09
Control Test	0.078	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 08/07/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 07/07/2024 23:37:36

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.079	23:25	
Air Blank	0.000	23:26	
Subject Sample	# 1 0.148	23:29	Slope Not Level
Breath Volume	1.253	23:29	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2 0.159	23:32	
Breath Volume	1.113	23:32	
Air Blank	0.000	23:33	
Air Blank	0.000	23:35	
Subject Sample	# 3 0.161	23:35	
Breath Volume	1.328	23:35	
Air Blank	0.000	23:36	
Control Test	0.078	23:36	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 08/07/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LISOVA,INNA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: SERRA

DATE/TIME OF TEST: 07/10/2024 16:33:33

Diagnostics Check	OK	16:26
Air Blank	0.000	16:27
Control Test	0.080	16:27
Air Blank	0.000	16:27
Subject Sample # 1	0.000	16:29
Breath Volume	1.785	16:29
Air Blank	0.000	16:29
Air Blank	0.000	16:31
Subject Sample # 2	0.000	16:32
Breath Volume	1.187	16:32
Air Blank	0.000	16:32
Control Test	0.079	16:32
Air Blank	0.000	16:33
Diagnostics Check	OK	16:33

Data Download Date/Time: 08/07/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: AVILA HERNANDEZ

DATE/TIME OF TEST: 07/11/2024 02:28:55

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.162	02:24
Breath Volume	2.843	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.156	02:27
Breath Volume	3.097	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 08/07/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANDRADE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: PRUITT

DATE/TIME OF TEST: 07/12/2024 00:04:08

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.080	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.168	23:59
Breath Volume	3.085	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.162	00:02
Breath Volume	3.117	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 08/07/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/12/2024 00:32:20

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.079	00:18	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:21	No Sample Provided
Breath Volume	0.000	00:21	
Air Blank	0.000	00:22	
Air Blank	0.000	00:24	
Subject Sample	# 2	0.102	00:26
Breath Volume	1.277	00:26	
Air Blank	0.000	00:27	
Air Blank	0.000	00:28	
Subject Sample	# 3	0.097	00:30
Breath Volume	1.179	00:30	
Air Blank	0.000	00:31	
Control Test	0.079	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 08/07/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROMANO,FABIOLA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 07/12/2024 23:55:22

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.236	23:50
Breath Volume	1.902	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.239	23:53
Breath Volume	2.023	23:53
Air Blank	0.000	23:54
Control Test	0.078	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 08/07/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: LEVASSEUR

DATE/TIME OF TEST: 07/13/2024 02:38:43

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.187	02:33
Breath Volume	1.664	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.181	02:36
Breath Volume	1.484	02:36
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 08/07/2024 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/15/2024 00:09:28

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.080	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.000	00:01	Volume Not Met
Breath Volume	0.128	00:01	
Air Blank	0.000	00:01	
Air Blank	0.000	00:03	
Subject Sample	# 2 0.342	00:04	
Breath Volume	1.394	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 3 0.343	00:07	
Breath Volume	1.207	00:07	
Air Blank	0.000	00:08	
Control Test	0.076	00:08	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: LAPORTE

DATE/TIME OF TEST: 07/15/2024 05:46:00

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.182	05:41
Breath Volume	1.390	05:41
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.183	05:44
Breath Volume	1.875	05:44
Air Blank	0.000	05:45
Control Test	0.076	05:45
Air Blank	0.000	05:45
Diagnostics Check	OK	05:46

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: LANSCHÉ

DATE/TIME OF TEST: 07/18/2024 23:48:15

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.212	23:40
Breath Volume	1.785	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:42
Subject Sample # 2	0.235	23:43
Breath Volume	1.406	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample # 3	0.211	23:46
Breath Volume	1.222	23:46
Air Blank	0.000	23:47
Control Test	0.077	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: ARNETT

DATE/TIME OF TEST: 07/19/2024 00:41:46

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.079	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1 0.020	00:30	
Breath Volume	1.218	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample	# 2 0.015	00:36	Volume Not Met
Breath Volume	0.480	00:36	
Air Blank	0.000	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 3 0.016	00:40	
Breath Volume	1.578	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:41	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOGAN,JOHNATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 07/19/2024 05:24:55

Diagnostics Check	OK	05:18
Air Blank	0.000	05:18
Control Test	0.080	05:19
Air Blank	0.000	05:19
Subject Sample # 1	0.088	05:20
Breath Volume	1.707	05:20
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.084	05:23
Breath Volume	1.253	05:23
Air Blank	0.000	05:23
Control Test	0.079	05:24
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DUQUE,BRYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: CARUTHERS

DATE/TIME OF TEST: 07/19/2024 22:04:05

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.212	21:58
Breath Volume	1.332	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.222	22:02
Breath Volume	2.308	22:02
Air Blank	0.000	22:03
Control Test	0.079	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:04

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: RAMOS-IZQUIERDO SOSA

DATE/TIME OF TEST: 07/20/2024 01:53:54

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.080	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1 0.130	01:41	Volume Not Met
Breath Volume	0.605	01:41	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 2 0.135	01:46	
Breath Volume	1.230	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 3 0.139	01:52	Volume Not Met
Breath Volume	0.929	01:52	
Air Blank	0.000	01:52	
Control Test	0.077	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: RAMOS-IZQUIERDO SOSA

DATE/TIME OF TEST: 07/20/2024 02:03:55

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.077	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.147	01:59
Breath Volume	1.945	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:01
Subject Sample # 2	0.136	02:02
Breath Volume	1.316	02:02
Air Blank	0.000	02:02
Control Test	0.077	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MORONI

DATE/TIME OF TEST: 07/23/2024 03:53:41

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1 0.147	03:41	
Breath Volume	1.707	03:41	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 2 0.102	03:46	Volume Not Met
Breath Volume	0.542	03:46	
Air Blank	0.000	03:47	
Air Blank	0.000	03:49	
Subject Sample	# 3	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.077	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WOOLLEY,KYLE
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BACCHUS

DATE/TIME OF TEST: 07/23/2024 22:08:54

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.175	22:04
Breath Volume	2.761	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.172	22:07
Breath Volume	2.878	22:07
Air Blank	0.000	22:07
Control Test	0.078	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 07/28/2024 02:30:12

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:15	
Air Blank	0.000	02:15	
Subject Sample	# 1 0.254	02:18	Volume Not Met
Breath Volume	0.523	02:18	
Air Blank	0.000	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 2 0.300	02:24	
Breath Volume	1.339	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 3 0.298	02:28	
Breath Volume	1.550	02:28	
Air Blank	0.000	02:29	
Control Test	0.072	02:29	Control Outside Tolerance
Air Blank	0.000	02:30	

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 07/28/2024 02:41:06

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.076	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.294	02:35
Breath Volume	1.578	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.282	02:39
Breath Volume	1.367	02:39
Air Blank	0.000	02:40
Control Test	0.075	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LYON,ALVARO
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: ST JUSTE

DATE/TIME OF TEST: 07/31/2024 15:12:34

Diagnostics Check	OK	15:05
Air Blank	0.000	15:06
Control Test	0.081	15:06
Air Blank	0.000	15:07
Subject Sample # 1	0.229	15:07
Breath Volume	2.671	15:07
Air Blank	0.000	15:08
Air Blank	0.000	15:10
Subject Sample # 2	0.237	15:10
Breath Volume	2.101	15:10
Air Blank	0.000	15:11
Control Test	0.079	15:11
Air Blank	0.000	15:12
Diagnostics Check	OK	15:12

Data Download Date/Time: 08/07/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTINO,ERIK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRET
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402649340
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: TOUNTAS

DATE/TIME OF TEST: 07/21/2024 02:49:58

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.225	02:45
Breath Volume	1.656	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.227	02:48
Breath Volume	1.804	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 08/27/2024 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SVIDENKO

DATE/TIME OF TEST: 07/06/2024 02:05:37

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.079	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.187	02:00
Breath Volume	1.207	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.189	02:03
Breath Volume	3.035	02:03
Air Blank	0.000	02:04
Control Test	0.077	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 08/28/2024 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 07/13/2024 01:42:31

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:35	
Control Test	0.079	01:35	
Air Blank	0.000	01:36	
Subject Sample # 1	0.185	01:37	
Breath Volume	2.160	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample # 2	0.178	01:40	
Breath Volume	2.097	01:40	
Air Blank	0.000	01:41	
Control Test	0.077	01:41	
Air Blank	0.000	01:41	
Diagnostics Check	Fail	01:42	Voltage/Current Test
Air Blank	0.000	01:42	

Data Download Date/Time: 08/28/2024 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 07/13/2024 01:50:13

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.186	01:45
Breath Volume	1.730	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.179	01:48
Breath Volume	1.828	01:48
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 08/28/2024 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 07/15/2024 23:49:51

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.081	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.164	23:45
Breath Volume	2.238	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:47
Subject Sample # 2	0.163	23:48
Breath Volume	2.089	23:48
Air Blank	0.000	23:48
Control Test	0.078	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 08/06/2024 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: WHITE,MICAH	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop H	VIOLATION CODE: DUI
ARRESTING OFFICER: WHITE,MICAH	ARREST TIME: 22:06
ARRESTING OFFICER AGENCY: FHP Troop H	GENDER: Male
LAST AGENCY INSPECTION: 06/26/2024	OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: WARREN	DATE/TIME OF TEST: 07/17/2024 22:33:28
-----------------------------	--

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.081	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.103	22:28
Breath Volume	3.054	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.108	22:31
Breath Volume	4.261	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 08/06/2024 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 07/18/2024 21:53:10

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.081	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.127	21:48
Breath Volume	1.453	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:50
Subject Sample # 2	0.130	21:51
Breath Volume	1.339	21:51
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 08/06/2024 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WHARTON

DATE/TIME OF TEST: 07/02/2024 23:06:32

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.018	23:02
Breath Volume	1.746	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.017	23:04
Breath Volume	1.742	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 08/20/2024 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: KEMPF

DATE/TIME OF TEST: 07/06/2024 02:30:53

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.269	02:26
Breath Volume	3.484	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:28
Subject Sample # 2	0.251	02:29
Breath Volume	2.406	02:29
Air Blank	0.000	02:29
Control Test	0.076	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 08/20/2024 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,B,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SANTOS GUARDADO

DATE/TIME OF TEST: 07/06/2024 03:35:53

Diagnostics Check	Fail	03:35	Internal Printer Test
Air Blank	0.000	03:35	

Data Download Date/Time: 08/06/2024 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,B,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: PEREDA

DATE/TIME OF TEST: 07/25/2024 04:36:03

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.232	04:31
Breath Volume	1.980	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.212	04:34
Breath Volume	1.917	04:34
Air Blank	0.000	04:35
Control Test	0.078	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:36

Data Download Date/Time: 08/06/2024 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIBIASE,JUSTIN,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: MEDRANO CABRERA

DATE/TIME OF TEST: 07/01/2024 21:04:55

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.079	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.151	21:00
Breath Volume	1.226	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.147	21:03
Breath Volume	1.144	21:03
Air Blank	0.000	21:03
Control Test	0.078	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 07/03/2024 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILILLO,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,MICHAEL
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BELLIDO DE LUNA

DATE/TIME OF TEST: 07/01/2024 23:30:22

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.163	23:25
Breath Volume	2.152	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.160	23:28
Breath Volume	2.390	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 07/03/2024 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BODDEN

DATE/TIME OF TEST: 07/05/2024 03:57:55

Diagnostics Check	Fail	03:57	Internal Printer Test
Air Blank		03:57	RFI Detect

Data Download Date/Time: 08/12/2024 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,STEVEN,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,STEVEN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 07/05/2024 21:43:54

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.157	21:39
Breath Volume	1.468	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.156	21:42
Breath Volume	1.230	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 08/12/2024 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: DIAZ DIAZ

DATE/TIME OF TEST: 07/06/2024 03:53:43

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.080	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1 0.171	03:43	Volume Not Met
Breath Volume	0.078	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2 0.175	03:48	
Breath Volume	1.156	03:48	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample	# 3 0.193	03:52	
Breath Volume	1.582	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 08/12/2024 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SMYTHE JR

DATE/TIME OF TEST: 07/07/2024 03:48:21

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.161	03:43
Breath Volume	1.636	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.161	03:46
Breath Volume	1.914	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 08/12/2024 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENDON,D,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: GARCIA DELABRA

DATE/TIME OF TEST: 07/08/2024 00:51:53

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.167	00:47
Breath Volume	1.277	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.165	00:50
Breath Volume	1.429	00:50
Air Blank	0.000	00:50
Control Test	0.077	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 08/12/2024 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DE LA CRUZ-JACINTO

DATE/TIME OF TEST: 07/25/2024 01:56:58

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:45	
Subject Sample # 1	0.135	01:47	
Breath Volume	1.406	01:47	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample # 2		01:52	No Sample Provided
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample # 3		01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 08/12/2024 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: LAMBERT II

DATE/TIME OF TEST: 07/31/2024 00:43:36

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.000	00:39
Breath Volume	2.812	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.000	00:42
Breath Volume	2.937	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 08/12/2024 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILY,STEVEN,B
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: CRISERA,MATTHEW
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 07/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: SINCLAIR

DATE/TIME OF TEST: 07/26/2024 02:14:42

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.036	02:10
Breath Volume	1.347	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.038	02:13
Breath Volume	1.878	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 08/30/2024 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBE,BRIAN,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: GRANITO,ALEXIS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/26/2024 04:09:51

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.143	04:05
Breath Volume	2.457	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.145	04:08
Breath Volume	2.167	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 08/30/2024 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOVANOVA, MICHAEL, N
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: JOVANOVA, MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: BUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 07/12/2024 22:44:22

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.078	22:37	
Air Blank	0.000	22:38	
Subject Sample # 1	0.136	22:39	
Breath Volume	1.730	22:39	
Air Blank	0.000	22:39	
Air Blank	0.000	22:41	
Subject Sample # 2	0.140	22:42	
Breath Volume	1.695	22:42	
Air Blank	0.000	22:42	
Control Test	0.078	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	Fail	22:43	Internal Printer Test
Air Blank	0.000	22:44	

Data Download Date/Time: 07/22/2024 18:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: AG310302
BREATH TEST OPERATOR: WILD,JACOB,G	DRY GAS EXPIRY DATE: 04/13/2025
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: WILD,JACOB	ARREST TIME: 20:28
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Female
LAST AGENCY INSPECTION: 06/28/2024	OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: TOMAS	DATE/TIME OF TEST: 07/12/2024	22:57:57
----------------------------	-------------------------------	----------

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.000	22:53
Breath Volume	1.722	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.000	22:56
Breath Volume	2.203	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 07/22/2024 18:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,CHARLES,B
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: REEVES,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: BUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: WATTERSON

DATE/TIME OF TEST: 07/19/2024 19:25:52

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.078	19:20
Air Blank	0.000	19:20
Subject Sample # 1	0.181	19:21
Breath Volume	2.488	19:21
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.188	19:24
Breath Volume	2.230	19:24
Air Blank	0.000	19:24
Control Test	0.077	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 08/02/2024 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVER, WILLIAM, H
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: OLIVER, WILLIAM
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: BUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 07/20/2024 20:40:03

Diagnostics Check	OK	20:36	
Air Blank	0.000	20:36	
Control Test	0.078	20:37	
Air Blank	0.000	20:37	
Subject Sample	# 1	20:38	Subject Test Refused
Breath Volume	0.000	20:38	
Air Blank	0.000	20:39	
Control Test	0.077	20:39	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:40	

Data Download Date/Time: 08/02/2024 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402648278
BREATH TEST OPERATOR: PHILLIPS,CHARLES,B	DRY GAS EXPIRY DATE: 01/27/2026
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: PHILLIPS,CHARLES	ARREST TIME: 17:36
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/26/2024	OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: MCINTEE	DATE/TIME OF TEST: 07/21/2024 18:09:17
------------------------------	--

Diagnostics Check	OK	18:01
Air Blank	0.000	18:02
Control Test	0.078	18:02
Air Blank	0.000	18:03
Subject Sample # 1	0.198	18:04
Breath Volume	3.738	18:04
Air Blank	0.000	18:05
Air Blank	0.000	18:06
Subject Sample # 2	0.197	18:07
Breath Volume	4.101	18:07
Air Blank	0.000	18:08
Control Test	0.076	18:08
Air Blank	0.000	18:09
Diagnostics Check	OK	18:09

Data Download Date/Time: 08/02/2024 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: STINSON,E
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: BUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: WALDROP

DATE/TIME OF TEST: 07/23/2024 17:52:10

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.078	17:46
Air Blank	0.000	17:47
Subject Sample # 1	0.175	17:47
Breath Volume	2.371	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.176	17:50
Breath Volume	2.828	17:50
Air Blank	0.000	17:51
Control Test	0.076	17:51
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 08/02/2024 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVER, WILLIAM, H
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: OLIVER, WILLIAM
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: BUI
ARREST TIME: 13:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:47

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 07/27/2024 14:17:19

Diagnostics Check	OK	14:10
Air Blank	0.000	14:10
Control Test	0.078	14:11
Air Blank	0.000	14:11
Subject Sample # 1	0.266	14:12
Breath Volume	1.808	14:12
Air Blank	0.000	14:13
Air Blank	0.000	14:14
Subject Sample # 2	0.269	14:15
Breath Volume	1.664	14:15
Air Blank	0.000	14:16
Control Test	0.077	14:16
Air Blank	0.000	14:17
Diagnostics Check	OK	14:17

Data Download Date/Time: 08/02/2024 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: BUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/27/2024 16:15:15

Diagnostics Check	OK	16:12	
Air Blank	0.000	16:12	
Control Test	0.078	16:12	
Air Blank	0.000	16:13	
Subject Sample	# 1	16:13	Subject Test Refused
Breath Volume	0.000	16:13	
Air Blank	0.000	16:14	
Control Test	0.075	16:14	
Air Blank	0.000	16:15	
Diagnostics Check	OK	16:15	

Data Download Date/Time: 08/02/2024 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: TAYLOR,A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: DALEY

DATE/TIME OF TEST: 07/27/2024 03:49:58

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:47	
Control Test	0.081	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.081	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 08/22/2024 01:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALMERTON,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SANTOYO

DATE/TIME OF TEST: 07/28/2024 00:30:51

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.082	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.258	00:26
Breath Volume	2.003	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.254	00:29
Breath Volume	1.550	00:29
Air Blank	0.000	00:29
Control Test	0.080	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 08/22/2024 01:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ARMSTEAD,ARMSTEAD
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 07/01/2024 19:03:17

Diagnostics Check	OK	18:57
Air Blank	0.000	18:57
Control Test	0.081	18:58
Air Blank	0.000	18:58
Subject Sample # 1	0.000	18:59
Breath Volume	2.863	18:59
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.000	19:01
Breath Volume	2.480	19:01
Air Blank	0.000	19:02
Control Test	0.080	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN, WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 07/03/2024 12:41:53

Diagnostics Check	OK	12:35
Air Blank	0.000	12:35
Control Test	0.081	12:36
Air Blank	0.000	12:36
Subject Sample # 1	0.000	12:37
Breath Volume	2.425	12:37
Air Blank	0.000	12:37
Air Blank	0.000	12:39
Subject Sample # 2	0.000	12:40
Breath Volume	2.933	12:40
Air Blank	0.000	12:40
Control Test	0.081	12:41
Air Blank	0.000	12:41
Diagnostics Check	OK	12:41

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDSON,DEPUTY,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: JONES,OFFICER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GARCIA-PEREZ

DATE/TIME OF TEST: 07/04/2024 23:31:33

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.081	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.135	23:26
Breath Volume	1.496	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.135	23:29
Breath Volume	1.398	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,RODNEY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 07/05/2024 02:05:21

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.082	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.200	02:00
Breath Volume	1.972	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.199	02:03
Breath Volume	3.070	02:03
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,J,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: RACKI

DATE/TIME OF TEST: 07/05/2024 12:28:13

Diagnostics Check	OK	12:21
Air Blank	0.000	12:22
Control Test	0.081	12:22
Air Blank	0.000	12:23
Subject Sample # 1	0.000	12:23
Breath Volume	1.554	12:23
Air Blank	0.000	12:24
Air Blank	0.000	12:26
Subject Sample # 2	0.000	12:26
Breath Volume	1.230	12:26
Air Blank	0.000	12:27
Control Test	0.082	12:27
Air Blank	0.000	12:28
Diagnostics Check	OK	12:28

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILES,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLARD,TYLER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CARRICO

DATE/TIME OF TEST: 07/06/2024 02:15:15

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.081	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.081	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN, SHON, W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE, RODNEY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/06/2024 03:11:05

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.150	03:06
Breath Volume	1.261	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.150	03:09
Breath Volume	1.156	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:11

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILES,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MINSHEW,JOE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GAYLE

DATE/TIME OF TEST: 07/06/2024 04:12:58

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:06	
Control Test	0.081	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1 0.077	04:08	
Breath Volume	1.875	04:08	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2 0.081	04:11	
Breath Volume	2.257	04:11	
Air Blank	0.000	04:12	
Control Test	0.070	04:12	Interferent Detect
Air Blank	0.000	04:12	

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILES,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GRAZA,CLARISSA
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 07/06/2024 21:56:07

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.082	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.206	21:51
Breath Volume	3.328	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:53
Subject Sample # 2	0.207	21:54
Breath Volume	2.644	21:54
Air Blank	0.000	21:55
Control Test	0.079	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WISE,K,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WISE,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: GOPHER

DATE/TIME OF TEST: 07/08/2024 10:27:14

Diagnostics Check	OK	10:21
Air Blank	0.000	10:21
Control Test	0.082	10:22
Air Blank	0.000	10:22
Subject Sample # 1	0.000	10:22
Breath Volume	1.957	10:22
Air Blank	0.000	10:23
Air Blank	0.000	10:25
Subject Sample # 2	0.000	10:25
Breath Volume	1.800	10:25
Air Blank	0.000	10:26
Control Test	0.083	10:26
Air Blank	0.000	10:27
Diagnostics Check	OK	10:27

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,R,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HALL,R
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/08/2024 19:39:40

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.081	19:34
Air Blank	0.000	19:34
Subject Sample # 1	0.000	19:35
Breath Volume	3.167	19:35
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.000	19:38
Breath Volume	3.312	19:38
Air Blank	0.000	19:38
Control Test	0.082	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DUPONT,N
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 07/12/2024 20:10:41

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:08	
Control Test	0.082	20:08	
Air Blank	0.000	20:09	
Subject Sample	# 1	20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:09	
Control Test	0.081	20:10	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: JACOBS,S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: FELIPE-ZUNIGA

DATE/TIME OF TEST: 07/13/2024 02:01:33

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.081	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.130	01:56
Breath Volume	2.246	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.134	01:59
Breath Volume	1.449	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,R
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: MOREE

DATE/TIME OF TEST: 07/13/2024 23:56:46

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.081	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.174	23:51
Breath Volume	1.492	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.167	23:55
Breath Volume	1.445	23:55
Air Blank	0.000	23:55
Control Test	0.081	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOTT,CALEB
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: RACKI

DATE/TIME OF TEST: 07/18/2024 14:17:12

Diagnostics Check	OK	14:11
Air Blank	0.000	14:11
Control Test	0.082	14:11
Air Blank	0.000	14:12
Subject Sample # 1	0.000	14:12
Breath Volume	2.968	14:12
Air Blank	0.000	14:13
Air Blank	0.000	14:15
Subject Sample # 2	0.000	14:15
Breath Volume	3.062	14:15
Air Blank	0.000	14:16
Control Test	0.081	14:16
Air Blank	0.000	14:17
Diagnostics Check	OK	14:17

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FORBES,Z
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:48

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 07/19/2024 13:16:42

Diagnostics Check	OK	13:08
Air Blank	0.000	13:09
Control Test	0.082	13:09
Air Blank	0.000	13:09
Subject Sample # 1	0.135	13:11
Breath Volume	1.574	13:11
Air Blank	0.000	13:12
Air Blank	0.000	13:13
Subject Sample # 2	0.130	13:15
Breath Volume	1.644	13:15
Air Blank	0.000	13:15
Control Test	0.081	13:16
Air Blank	0.000	13:16
Diagnostics Check	OK	13:16

Data Download Date/Time: 07/22/2024 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,LIEUTENANT,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,LIEUTENANT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 12:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: GOPHER

DATE/TIME OF TEST: 07/22/2024 13:39:34

Diagnostics Check	OK	13:33
Air Blank	0.000	13:33
Control Test	0.082	13:33
Air Blank	0.000	13:34
Subject Sample # 1	0.000	13:35
Breath Volume	2.097	13:35
Air Blank	0.000	13:35
Air Blank	0.000	13:37
Subject Sample # 2	0.000	13:38
Breath Volume	2.195	13:38
Air Blank	0.000	13:38
Control Test	0.081	13:38
Air Blank	0.000	13:39
Diagnostics Check	OK	13:39

Data Download Date/Time: 08/22/2024 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DELOERA,PATRICIA
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: TORRENTE

DATE/TIME OF TEST: 07/29/2024 22:10:46

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:01	
Control Test	0.082	22:01	
Air Blank	0.000	22:02	
Subject Sample	# 1 0.170	22:05	Volume Not Met
Breath Volume	0.835	22:05	
Air Blank	0.000	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 2	22:08	Subject Test Refused
Breath Volume	0.570	22:08	
Air Blank	0.000	22:09	
Control Test	0.080	22:10	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

Data Download Date/Time: 08/22/2024 00:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOVAR,LUIS,D
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: CERON,JOSEPH
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: KAHL

DATE/TIME OF TEST: 07/05/2024 02:01:32

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.082	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.081	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 08/01/2024 22:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: Arcadia PD
ARRESTING OFFICER: MATOS,FRANK
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MOTLEY

DATE/TIME OF TEST: 07/07/2024 03:36:41

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.110	03:32
Breath Volume	1.371	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.112	03:35
Breath Volume	1.300	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 08/01/2024 22:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOVAR,LUIS,D
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: TRUJILLO,JOSE
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/12/2024 21:47:09

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.190	21:42
Breath Volume	3.519	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:44
Subject Sample # 2	0.183	21:45
Breath Volume	2.156	21:45
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 08/01/2024 22:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: SUTTON,V
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: GUZMAN-PEREZ

DATE/TIME OF TEST: 07/22/2024 02:01:27

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.051	01:56
Breath Volume	1.308	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.052	01:59
Breath Volume	1.378	01:59
Air Blank	0.000	02:00
Control Test	0.078	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 08/01/2024 22:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOVAR,LUIS,D
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: NUNEZ,NIURVIS
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: ALBRITTON

DATE/TIME OF TEST: 07/28/2024 04:53:50

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.094	04:48
Breath Volume	2.398	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.093	04:52
Breath Volume	1.191	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 08/01/2024 22:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: WALKER,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: KOLOGINSKI

DATE/TIME OF TEST: 07/29/2024 23:29:59

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.079	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:29	
Control Test	0.080	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 08/01/2024 22:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,ERIC,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MESA,ERIC
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BENITEZ BENAVIDES

DATE/TIME OF TEST: 07/04/2024 02:36:00

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.000	02:31
Breath Volume	2.394	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.000	02:34
Breath Volume	3.718	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:36

Data Download Date/Time: 07/31/2024 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEGERA,JERRY,J
BREATH TEST OPERATOR AGENCY: OTHER State
ARRESTING OFFICER: SALAZAR,JUSTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: BUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: VITON

DATE/TIME OF TEST: 07/04/2024 23:55:59

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.081	23:54	
Air Blank	0.000	23:55	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank		23:55	RFI Detect
Air Blank		23:55	RFI Detect

Data Download Date/Time: 07/31/2024 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: AGUILA,OSVALDO,L

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RUIZ,HERMES

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: DUI

ARREST TIME: 17:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: SEGOVIA

DATE/TIME OF TEST: 07/05/2024 18:13:17

Diagnostics Check	OK	18:04
Air Blank	0.000	18:05
Control Test	0.081	18:05
Air Blank	0.000	18:06
Subject Sample # 1	0.211	18:06
Breath Volume	4.000	18:06
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.208	18:11
Breath Volume	3.203	18:11
Air Blank	0.000	18:12
Control Test	0.080	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 07/31/2024 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNAUD,GEORGE,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: VEIGA,RICARDO
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: BUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 07/06/2024 21:55:21

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.126	21:47
Breath Volume	2.214	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.157	21:50
Breath Volume	2.570	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 3	0.154	21:53
Breath Volume	2.082	21:53
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 07/31/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADILLA,A,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: NOLASCO,C
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:23

SUBJECT'S LAST NAME: HERMIDAHONDARES

DATE/TIME OF TEST: 07/13/2024 08:00:32

Diagnostics Check	OK	07:53
Air Blank	0.000	07:53
Control Test	0.082	07:54
Air Blank	0.000	07:54
Subject Sample # 1	0.203	07:55
Breath Volume	1.925	07:55
Air Blank	0.000	07:56
Air Blank	0.000	07:57
Subject Sample # 2	0.205	07:58
Breath Volume	1.785	07:58
Air Blank	0.000	07:59
Control Test	0.080	07:59
Air Blank	0.000	08:00
Diagnostics Check	OK	08:00

Data Download Date/Time: 07/31/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARANA,MARLON,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 07/13/2024 23:20:41

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.113	23:16
Breath Volume	2.476	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.109	23:19
Breath Volume	1.792	23:19
Air Blank	0.000	23:19
Control Test	0.078	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 07/31/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARANA,MARLON,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: ALFARO TALAVERA

DATE/TIME OF TEST: 07/14/2024 02:28:41

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.171	02:24
Breath Volume	2.605	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.180	02:27
Breath Volume	3.082	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/31/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEGERA,JERRY,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BEGERA,JERRY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: BUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 07/14/2024 19:10:18

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:07	
Control Test	0.081	19:08	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:08	Subject Test Refused
Breath Volume	0.000	19:08	
Air Blank	0.000	19:09	
Control Test	0.081	19:09	
Air Blank	0.000	19:10	
Diagnostics Check	OK	19:10	

Data Download Date/Time: 07/31/2024 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ESPINOSA,MIGUEL,A

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: MONTERO,NICHOLAS

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: BUI

ARREST TIME: 23:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/15/2024 00:33:46

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.081	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Control Test	0.081	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 07/31/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSCHINGER,MICKAEL,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WATTLEY,EMILIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 07/15/2024 21:21:31

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:20	Subject Test Refused
Breath Volume	0.000	21:20	
Air Blank	0.000	21:20	
Control Test	0.080	21:20	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 07/31/2024 16:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEJARZA,LUIS,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RINCON,SEBASTIAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: FRAGA

DATE/TIME OF TEST: 07/17/2024 05:25:22

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:22	
Control Test	0.082	05:23	
Air Blank	0.000	05:23	
Subject Sample	# 1	05:23	Subject Test Refused
Breath Volume	0.000	05:23	
Air Blank	0.000	05:24	
Control Test	0.081	05:24	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 07/31/2024 16:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ECHEVERRI,DANIEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: SARMIENTO,VAITIERE

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: DUI

ARREST TIME: 03:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NOSSA

DATE/TIME OF TEST: 07/20/2024 04:31:21

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.081	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:27	No Sample Provided
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 2	04:30	Sequence Aborted
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	

Data Download Date/Time: 07/31/2024 16:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARANA,MARLON,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARANA,MARLON
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SANCHEZ RODRIGUEZ

DATE/TIME OF TEST: 07/21/2024 00:49:22

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.082	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:47	Subject Test Refused
Breath Volume	0.000	00:47	
Air Blank	0.000	00:48	
Control Test	0.081	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 07/31/2024 16:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GALVEZ

DATE/TIME OF TEST: 07/24/2024 23:14:32

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.081	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:13	
Control Test	0.082	23:13	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 07/31/2024 16:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,ELSTON,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARZOLA,KENNETH
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: QUILCA

DATE/TIME OF TEST: 07/26/2024 04:14:38

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.081	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:07	No Sample Provided
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2	04:13	No Sample Provided
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Control Test	0.081	04:14	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 07/31/2024 16:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ECHEVERRI,DANIEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: SARMIENTO,VAITIERE

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: DUI

ARREST TIME: 02:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: GUADAMUZ

DATE/TIME OF TEST: 07/28/2024 04:08:35

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.181	04:01	Volume Not Met
Breath Volume	0.128	04:01	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2 0.189	04:06	Volume Not Met
Breath Volume	0.207	04:06	
Air Blank	0.000	04:07	
Control Test	0.078	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 07/31/2024 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ECHEVERRI,DANIEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: SARMIENTO,VAITIERE

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: DUI

ARREST TIME: 02:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: GUADAMUZ

DATE/TIME OF TEST: 07/28/2024 04:21:38

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.077	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:17	No Sample Provided
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample	# 2	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Control Test	0.079	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 07/31/2024 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,R,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: NAVARRO,I
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 08:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:37

SUBJECT'S LAST NAME: C234780861830

DATE/TIME OF TEST: 07/28/2024 10:01:19

Diagnostics Check	OK	09:59	
Air Blank	0.000	10:00	
Control Test	0.080	10:00	
Air Blank		10:00	RFI Detect
Air Blank	0.000	10:01	

Data Download Date/Time: 07/31/2024 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,R,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: NAVARRO,I
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 08:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:37

SUBJECT'S LAST NAME: CASTILLO SANCHEZ

DATE/TIME OF TEST: 07/28/2024 10:07:55

Diagnostics Check	OK	10:05	
Air Blank	0.000	10:05	
Control Test	0.081	10:05	
Air Blank	0.000	10:06	
Subject Sample	# 1	10:06	Subject Test Refused
Breath Volume	0.000	10:06	
Air Blank	0.000	10:06	
Control Test	0.080	10:07	
Air Blank	0.000	10:07	
Diagnostics Check	OK	10:07	

Data Download Date/Time: 07/31/2024 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABALLERO,A,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CABALLERO,A
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/29/2024 02:13:13

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.081	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 07/31/2024 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,ANGEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FERNANDEZ,ANGEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: DUI

ARREST TIME: 08:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:16

SUBJECT'S LAST NAME: JIMENEZA-AZUM

DATE/TIME OF TEST: 07/31/2024 08:49:18

Diagnostics Check	OK	08:42
Air Blank	0.000	08:43
Control Test	0.081	08:43
Air Blank	0.000	08:44
Subject Sample # 1	0.000	08:44
Breath Volume	2.437	08:44
Air Blank	0.000	08:45
Air Blank	0.000	08:47
Subject Sample # 2	0.000	08:47
Breath Volume	2.312	08:47
Air Blank	0.000	08:48
Control Test	0.080	08:48
Air Blank	0.000	08:49
Diagnostics Check	OK	08:49

Data Download Date/Time: 07/31/2024 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,BRIAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GUERRA,BRIAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: POUNDS

DATE/TIME OF TEST: 07/31/2024 20:20:35

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:18	
Control Test	0.081	20:18	
Air Blank	0.000	20:18	
Subject Sample	# 1	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:19	
Control Test	0.081	20:19	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 08/30/2024 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,ELSTON,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WALLACE,ELSTON
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: IKRAMELAHAI

DATE/TIME OF TEST: 07/04/2024 04:49:20

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.159	04:44
Breath Volume	3.152	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.154	04:47
Breath Volume	3.378	04:47
Air Blank	0.000	04:48
Control Test	0.078	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 07/31/2024 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEGERA,JERRY,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SALAZAR,JUSTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: BUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: VITON

DATE/TIME OF TEST: 07/05/2024 00:01:55

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.079	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 07/31/2024 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARANA,MARLON,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ZAMBRANA

DATE/TIME OF TEST: 07/14/2024 03:10:26

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.105	03:05
Breath Volume	1.398	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.106	03:08
Breath Volume	1.687	03:08
Air Blank	0.000	03:09
Control Test	0.078	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 07/31/2024 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,MANUEL,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LAGRANGE

DATE/TIME OF TEST: 07/21/2024 02:59:04

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.079	02:57	
Air Blank	0.000	02:57	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	
Control Test	0.079	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 07/31/2024 16:43