

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: WHITE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 08/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402783528  
DRY GAS EXPIRY DATE: 06/27/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/23/2024 01:44:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Control Test       | 0.078 | 01:39 |
| Air Blank          | 0.000 | 01:39 |
| Subject Sample # 1 | 0.180 | 01:40 |
| Breath Volume      | 2.363 | 01:40 |
| Air Blank          | 0.000 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 2 | 0.174 | 01:43 |
| Breath Volume      | 2.195 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Control Test       | 0.076 | 01:44 |
| Air Blank          | 0.000 | 01:44 |
| Diagnostics Check  | OK    | 01:44 |

---

Data Download Date/Time: 08/24/2024 12:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: WHITE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 08/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402783528  
DRY GAS EXPIRY DATE: 06/27/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: DOAK

DATE/TIME OF TEST: 08/24/2024 01:55:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:49 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.079 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Subject Sample # 1 | 0.109 | 01:51 |
| Breath Volume      | 1.867 | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Air Blank          | 0.000 | 01:53 |
| Subject Sample # 2 | 0.098 | 01:54 |
| Breath Volume      | 1.488 | 01:54 |
| Air Blank          | 0.000 | 01:54 |
| Control Test       | 0.078 | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Diagnostics Check  | OK    | 01:55 |

---

Data Download Date/Time: 08/24/2024 12:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRIEBELL,DON,R  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: ABRAHAMSEN,ERIK  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 08/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402783528  
DRY GAS EXPIRY DATE: 06/27/2026  
VIOLATION CODE: DUI  
ARREST TIME: 15:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 08/30/2024 16:29:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:22 |
| Air Blank          | 0.000 | 16:22 |
| Control Test       | 0.078 | 16:22 |
| Air Blank          | 0.000 | 16:23 |
| Subject Sample # 1 | 0.233 | 16:23 |
| Breath Volume      | 1.437 | 16:23 |
| Air Blank          | 0.000 | 16:24 |
| Air Blank          | 0.000 | 16:26 |
| Subject Sample # 2 | 0.230 | 16:27 |
| Breath Volume      | 1.398 | 16:27 |
| Air Blank          | 0.000 | 16:28 |
| Control Test       | 0.077 | 16:28 |
| Air Blank          | 0.000 | 16:28 |
| Diagnostics Check  | OK    | 16:29 |

---

Data Download Date/Time: 09/25/2024 01:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEILL, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MUELLER, WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402783528  
DRY GAS EXPIRY DATE: 06/27/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MELONE

DATE/TIME OF TEST: 08/30/2024 21:13:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:07 |
| Air Blank          | 0.000 | 21:08 |
| Control Test       | 0.077 | 21:08 |
| Air Blank          | 0.000 | 21:08 |
| Subject Sample # 1 | 0.309 | 21:09 |
| Breath Volume      | 1.750 | 21:09 |
| Air Blank          | 0.000 | 21:09 |
| Air Blank          | 0.000 | 21:11 |
| Subject Sample # 2 | 0.302 | 21:12 |
| Breath Volume      | 2.117 | 21:12 |
| Air Blank          | 0.000 | 21:12 |
| Control Test       | 0.077 | 21:13 |
| Air Blank          | 0.000 | 21:13 |
| Diagnostics Check  | OK    | 21:13 |

---

Data Download Date/Time: 09/25/2024 01:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOOTH,JUSTIN,E  
BREATH TEST OPERATOR AGENCY: Indialantic PD  
ARRESTING OFFICER: RIVERA,EDGAR  
ARRESTING OFFICER AGENCY: Melbourne Beach PD  
LAST AGENCY INSPECTION: 08/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/15/2024 20:19:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:11 |
| Air Blank          | 0.000 | 20:12 |
| Control Test       | 0.080 | 20:12 |
| Air Blank          | 0.000 | 20:12 |
| Subject Sample # 1 | 0.101 | 20:13 |
| Breath Volume      | 1.417 | 20:13 |
| Air Blank          | 0.000 | 20:14 |
| Air Blank          | 0.000 | 20:16 |
| Subject Sample # 2 | 0.100 | 20:17 |
| Breath Volume      | 1.230 | 20:17 |
| Air Blank          | 0.000 | 20:18 |
| Control Test       | 0.079 | 20:18 |
| Air Blank          | 0.000 | 20:19 |
| Diagnostics Check  | OK    | 20:19 |

---

Data Download Date/Time: 10/01/2024 11:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PINSKER,JOSEPH,A  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: PINSKER,JOSEPH  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 08/10/2024 22:47:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Control Test       | 0.082 | 22:42 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 1 | 0.157 | 22:43 |
| Breath Volume      | 1.140 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 2 | 0.156 | 22:46 |
| Breath Volume      | 1.585 | 22:46 |
| Air Blank          | 0.000 | 22:46 |
| Control Test       | 0.080 | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Diagnostics Check  | OK    | 22:47 |

---

Data Download Date/Time: 09/23/2024 03:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: FARRAGHER,THOMAS  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 08/11/2024 04:07:18

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 03:55 |                           |
| Air Blank         | 0.000     | 03:56 |                           |
| Control Test      | 0.081     | 03:56 |                           |
| Air Blank         | 0.000     | 03:56 |                           |
| Subject Sample    | # 1 0.174 | 04:00 | Volume Not Met            |
| Breath Volume     | 0.082     | 04:00 |                           |
| Air Blank         | 0.000     | 04:00 |                           |
| Air Blank         | 0.000     | 04:02 |                           |
| Subject Sample    | # 2 0.159 | 04:05 | Volume Not Met            |
| Breath Volume     | 0.105     | 04:05 |                           |
| Air Blank         | 0.000     | 04:06 |                           |
| Control Test      | 0.071     | 04:06 | Control Outside Tolerance |
| Air Blank         | 0.000     | 04:07 |                           |

Data Download Date/Time: 09/23/2024 03:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: SCHAAD, VINCENT  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: AMICANGELO

DATE/TIME OF TEST: 08/16/2024 22:17:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:11 |
| Air Blank          | 0.000 | 22:12 |
| Control Test       | 0.081 | 22:12 |
| Air Blank          | 0.000 | 22:12 |
| Subject Sample # 1 | 0.035 | 22:13 |
| Breath Volume      | 2.523 | 22:13 |
| Air Blank          | 0.000 | 22:13 |
| Air Blank          | 0.000 | 22:15 |
| Subject Sample # 2 | 0.033 | 22:16 |
| Breath Volume      | 3.226 | 22:16 |
| Air Blank          | 0.000 | 22:16 |
| Control Test       | 0.080 | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Diagnostics Check  | OK    | 22:17 |

---

Data Download Date/Time: 09/23/2024 03:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,DYLAN,T  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WEBER,DYLAN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: ARGUETA REYES

DATE/TIME OF TEST: 08/17/2024 03:02:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Control Test       | 0.081 | 02:57 |
| Air Blank          | 0.000 | 02:57 |
| Subject Sample # 1 | 0.131 | 02:58 |
| Breath Volume      | 3.316 | 02:58 |
| Air Blank          | 0.000 | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 2 | 0.125 | 03:01 |
| Breath Volume      | 3.300 | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Control Test       | 0.079 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Diagnostics Check  | OK    | 03:02 |

---

Data Download Date/Time: 09/23/2024 03:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: RICHARDS, JAIDON  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BURKHART

DATE/TIME OF TEST: 08/18/2024 00:37:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.081 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Subject Sample # 1 | 0.160 | 00:32 |
| Breath Volume      | 4.078 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Air Blank          | 0.000 | 00:35 |
| Subject Sample # 2 | 0.161 | 00:35 |
| Breath Volume      | 3.519 | 00:35 |
| Air Blank          | 0.000 | 00:36 |
| Control Test       | 0.079 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Diagnostics Check  | OK    | 00:37 |

---

Data Download Date/Time: 09/23/2024 03:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: SYMEON, S  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: MICHAELS

DATE/TIME OF TEST: 08/18/2024 19:06:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:00 |
| Air Blank          | 0.000 | 19:01 |
| Control Test       | 0.081 | 19:01 |
| Air Blank          | 0.000 | 19:02 |
| Subject Sample # 1 | 0.096 | 19:02 |
| Breath Volume      | 1.839 | 19:02 |
| Air Blank          | 0.000 | 19:03 |
| Air Blank          | 0.000 | 19:04 |
| Subject Sample # 2 | 0.093 | 19:05 |
| Breath Volume      | 1.882 | 19:05 |
| Air Blank          | 0.000 | 19:06 |
| Control Test       | 0.080 | 19:06 |
| Air Blank          | 0.000 | 19:06 |
| Diagnostics Check  | OK    | 19:06 |

---

Data Download Date/Time: 09/23/2024 03:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:53

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 08/06/2024 22:10:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:02 |
| Air Blank          | 0.000 | 22:03 |
| Control Test       | 0.079 | 22:03 |
| Air Blank          | 0.000 | 22:03 |
| Subject Sample # 1 | 0.184 | 22:05 |
| Breath Volume      | 1.593 | 22:05 |
| Air Blank          | 0.000 | 22:06 |
| Air Blank          | 0.000 | 22:08 |
| Subject Sample # 2 | 0.184 | 22:08 |
| Breath Volume      | 1.558 | 22:08 |
| Air Blank          | 0.000 | 22:09 |
| Control Test       | 0.078 | 22:09 |
| Air Blank          | 0.000 | 22:10 |
| Diagnostics Check  | OK    | 22:10 |

---

Data Download Date/Time: 08/28/2024 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 18:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: DIAZ HERNANDEZ

DATE/TIME OF TEST: 08/07/2024 18:41:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:32 |
| Air Blank          | 0.000 | 18:33 |
| Control Test       | 0.079 | 18:33 |
| Air Blank          | 0.000 | 18:34 |
| Subject Sample # 1 | 0.210 | 18:34 |
| Breath Volume      | 1.523 | 18:34 |
| Air Blank          | 0.000 | 18:35 |
| Air Blank          | 0.000 | 18:37 |
| Subject Sample # 2 | 0.198 | 18:39 |
| Breath Volume      | 1.328 | 18:39 |
| Air Blank          | 0.000 | 18:40 |
| Control Test       | 0.078 | 18:40 |
| Air Blank          | 0.000 | 18:41 |
| Diagnostics Check  | OK    | 18:41 |

---

Data Download Date/Time: 08/28/2024 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SABO

DATE/TIME OF TEST: 08/07/2024 23:51:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Control Test       | 0.078 | 23:46 |
| Air Blank          | 0.000 | 23:47 |
| Subject Sample # 1 | 0.000 | 23:47 |
| Breath Volume      | 2.394 | 23:47 |
| Air Blank          | 0.000 | 23:47 |
| Air Blank          | 0.000 | 23:49 |
| Subject Sample # 2 | 0.000 | 23:50 |
| Breath Volume      | 2.136 | 23:50 |
| Air Blank          | 0.000 | 23:50 |
| Control Test       | 0.079 | 23:51 |
| Air Blank          | 0.000 | 23:51 |
| Diagnostics Check  | OK    | 23:51 |

---

Data Download Date/Time: 08/28/2024 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 20:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: JOHANSEN

DATE/TIME OF TEST: 08/13/2024 20:41:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:35 |
| Air Blank          | 0.000 | 20:36 |
| Control Test       | 0.079 | 20:36 |
| Air Blank          | 0.000 | 20:37 |
| Subject Sample # 1 | 0.076 | 20:37 |
| Breath Volume      | 2.093 | 20:37 |
| Air Blank          | 0.000 | 20:38 |
| Air Blank          | 0.000 | 20:39 |
| Subject Sample # 2 | 0.075 | 20:40 |
| Breath Volume      | 1.441 | 20:40 |
| Air Blank          | 0.000 | 20:40 |
| Control Test       | 0.079 | 20:41 |
| Air Blank          | 0.000 | 20:41 |
| Diagnostics Check  | OK    | 20:41 |

---

Data Download Date/Time: 08/28/2024 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 19:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: BEAM

DATE/TIME OF TEST: 08/23/2024 19:25:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:19 |
| Air Blank          | 0.000 | 19:19 |
| Control Test       | 0.079 | 19:20 |
| Air Blank          | 0.000 | 19:20 |
| Subject Sample # 1 | 0.134 | 19:21 |
| Breath Volume      | 1.324 | 19:21 |
| Air Blank          | 0.000 | 19:21 |
| Air Blank          | 0.000 | 19:23 |
| Subject Sample # 2 | 0.135 | 19:24 |
| Breath Volume      | 1.902 | 19:24 |
| Air Blank          | 0.000 | 19:24 |
| Control Test       | 0.079 | 19:25 |
| Air Blank          | 0.000 | 19:25 |
| Diagnostics Check  | OK    | 19:25 |

---

Data Download Date/Time: 08/28/2024 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 22:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: WILKE

DATE/TIME OF TEST: 08/23/2024 22:24:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:18 |
| Air Blank          | 0.000 | 22:19 |
| Control Test       | 0.079 | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Subject Sample # 1 | 0.228 | 22:20 |
| Breath Volume      | 2.324 | 22:20 |
| Air Blank          | 0.000 | 22:21 |
| Air Blank          | 0.000 | 22:22 |
| Subject Sample # 2 | 0.232 | 22:23 |
| Breath Volume      | 2.585 | 22:23 |
| Air Blank          | 0.000 | 22:23 |
| Control Test       | 0.079 | 22:24 |
| Air Blank          | 0.000 | 22:24 |
| Diagnostics Check  | OK    | 22:24 |

---

Data Download Date/Time: 08/28/2024 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 08/28/2024 21:56:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Control Test       | 0.080 | 21:48 |
| Air Blank          | 0.000 | 21:49 |
| Subject Sample # 1 | 0.125 | 21:50 |
| Breath Volume      | 1.632 | 21:50 |
| Air Blank          | 0.000 | 21:50 |
| Air Blank          | 0.000 | 21:52 |
| Subject Sample # 2 | 0.122 | 21:55 |
| Breath Volume      | 1.621 | 21:55 |
| Air Blank          | 0.000 | 21:55 |
| Control Test       | 0.080 | 21:56 |
| Air Blank          | 0.000 | 21:56 |
| Diagnostics Check  | OK    | 21:56 |

---

Data Download Date/Time: 10/29/2024 10:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: VOL.,VOL.

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 00:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 08/31/2024 00:05:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Control Test       | 0.080 | 23:59 |
| Air Blank          | 0.000 | 23:59 |
| Subject Sample # 1 | 0.096 | 00:00 |
| Breath Volume      | 1.906 | 00:00 |
| Air Blank          | 0.000 | 00:01 |
| Air Blank          | 0.000 | 00:02 |
| Subject Sample # 2 | 0.102 | 00:04 |
| Breath Volume      | 1.847 | 00:04 |
| Air Blank          | 0.000 | 00:04 |
| Control Test       | 0.079 | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Diagnostics Check  | OK    | 00:05 |

---

Data Download Date/Time: 10/29/2024 10:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAMBERLAIN,JOSHUA,L  
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO  
ARRESTING OFFICER: CHAMBERLAIN,JOSHUA  
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO  
LAST AGENCY INSPECTION: 07/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A3  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: OZEBEK

DATE/TIME OF TEST: 08/17/2024 21:26:16

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 21:13 |                |
| Air Blank         | 0.000     | 21:13 |                |
| Control Test      | 0.081     | 21:14 |                |
| Air Blank         | 0.000     | 21:14 |                |
| Subject Sample    | # 1 0.210 | 21:15 |                |
| Breath Volume     | 2.601     | 21:15 |                |
| Air Blank         | 0.000     | 21:16 |                |
| Air Blank         | 0.000     | 21:18 |                |
| Subject Sample    | # 2 0.188 | 21:21 | Volume Not Met |
| Breath Volume     | 0.179     | 21:21 |                |
| Air Blank         | 0.000     | 21:22 |                |
| Air Blank         | 0.000     | 21:23 |                |
| Subject Sample    | # 3 0.191 | 21:24 |                |
| Breath Volume     | 2.183     | 21:24 |                |
| Air Blank         | 0.000     | 21:25 |                |
| Control Test      | 0.076     | 21:25 |                |
| Air Blank         | 0.000     | 21:26 |                |
| Diagnostics Check | OK        | 21:26 |                |

---

Data Download Date/Time: 08/22/2024 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TARR,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO  
ARRESTING OFFICER: SWATCHICK,TYLER  
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24923080A3  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/24/2024 02:56:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.082 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 1 | 0.257 | 02:51 |
| Breath Volume      | 1.351 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 2 | 0.253 | 02:54 |
| Breath Volume      | 1.179 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.081 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Diagnostics Check  | OK    | 02:56 |

---

Data Download Date/Time: 09/18/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DAVIS,P  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MCNEIL

DATE/TIME OF TEST: 08/01/2024 01:48:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:39 |
| Air Blank          | 0.000 | 01:40 |
| Control Test       | 0.078 | 01:40 |
| Air Blank          | 0.000 | 01:41 |
| Subject Sample # 1 | 0.127 | 01:43 |
| Breath Volume      | 1.863 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Air Blank          | 0.000 | 01:45 |
| Subject Sample # 2 | 0.126 | 01:47 |
| Breath Volume      | 3.175 | 01:47 |
| Air Blank          | 0.000 | 01:47 |
| Control Test       | 0.078 | 01:48 |
| Air Blank          | 0.000 | 01:48 |
| Diagnostics Check  | OK    | 01:48 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: DIAZ BORRAYO

DATE/TIME OF TEST: 08/06/2024 03:55:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Control Test       | 0.078 | 03:49 |
| Air Blank          | 0.000 | 03:50 |
| Subject Sample # 1 | 0.153 | 03:50 |
| Breath Volume      | 2.117 | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Air Blank          | 0.000 | 03:53 |
| Subject Sample # 2 | 0.143 | 03:53 |
| Breath Volume      | 1.410 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Control Test       | 0.079 | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Diagnostics Check  | OK    | 03:55 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 08/08/2024 20:16:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:09 |
| Air Blank          | 0.000 | 20:09 |
| Control Test       | 0.078 | 20:09 |
| Air Blank          | 0.000 | 20:10 |
| Subject Sample # 1 | 0.115 | 20:11 |
| Breath Volume      | 1.613 | 20:11 |
| Air Blank          | 0.000 | 20:11 |
| Air Blank          | 0.000 | 20:13 |
| Subject Sample # 2 | 0.122 | 20:14 |
| Breath Volume      | 2.070 | 20:14 |
| Air Blank          | 0.000 | 20:15 |
| Control Test       | 0.078 | 20:15 |
| Air Blank          | 0.000 | 20:16 |
| Diagnostics Check  | OK    | 20:16 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CARLSON,P  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/09/2024 02:24:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:22 |                      |
| Air Blank         | 0.000 | 02:22 |                      |
| Control Test      | 0.079 | 02:22 |                      |
| Air Blank         | 0.000 | 02:23 |                      |
| Subject Sample    | # 1   | 02:23 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:23 |                      |
| Air Blank         | 0.000 | 02:24 |                      |
| Control Test      | 0.078 | 02:24 |                      |
| Air Blank         | 0.000 | 02:24 |                      |
| Diagnostics Check | OK    | 02:24 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: THOMPSON,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BLACKWELL

DATE/TIME OF TEST: 08/09/2024 03:13:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:06 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.078 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Subject Sample # 1 | 0.111 | 03:07 |
| Breath Volume      | 1.339 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Air Blank          | 0.000 | 03:10 |
| Subject Sample # 2 | 0.116 | 03:11 |
| Breath Volume      | 3.382 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Control Test       | 0.078 | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Diagnostics Check  | OK    | 03:13 |

---

Data Download Date/Time: 08/22/2024 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ROBINSON,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 08/09/2024 04:14:24

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:11 |                      |
| Air Blank         | 0.000 | 04:12 |                      |
| Control Test      | 0.079 | 04:12 |                      |
| Air Blank         | 0.000 | 04:12 |                      |
| Subject Sample    | # 1   | 04:13 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:13 |                      |
| Air Blank         | 0.000 | 04:13 |                      |
| Control Test      | 0.078 | 04:13 |                      |
| Air Blank         | 0.000 | 04:14 |                      |
| Diagnostics Check | OK    | 04:14 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SPENCER,K  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 08/09/2024 07:25:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 07:22 |                      |
| Air Blank         | 0.000 | 07:23 |                      |
| Control Test      | 0.078 | 07:23 |                      |
| Air Blank         | 0.000 | 07:24 |                      |
| Subject Sample    | # 1   | 07:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:24 |                      |
| Air Blank         | 0.000 | 07:24 |                      |
| Control Test      | 0.078 | 07:24 |                      |
| Air Blank         | 0.000 | 07:25 |                      |
| Diagnostics Check | OK    | 07:25 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 08/09/2024 22:27:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:24 |                      |
| Air Blank         | 0.000 | 22:25 |                      |
| Control Test      | 0.078 | 22:25 |                      |
| Air Blank         | 0.000 | 22:26 |                      |
| Subject Sample    | # 1   | 22:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:26 |                      |
| Air Blank         | 0.000 | 22:26 |                      |
| Control Test      | 0.079 | 22:26 |                      |
| Air Blank         | 0.000 | 22:27 |                      |
| Diagnostics Check | OK    | 22:27 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BAEZ,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/10/2024 03:53:45

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:50 |                      |
| Air Blank         | 0.000 | 03:51 |                      |
| Control Test      | 0.078 | 03:51 |                      |
| Air Blank         | 0.000 | 03:52 |                      |
| Subject Sample    | # 1   | 03:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:52 |                      |
| Air Blank         | 0.000 | 03:52 |                      |
| Control Test      | 0.078 | 03:53 |                      |
| Air Blank         | 0.000 | 03:53 |                      |
| Diagnostics Check | OK    | 03:53 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHISMAN,N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: VILCHEZ

DATE/TIME OF TEST: 08/10/2024 05:50:04

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:47 |                      |
| Air Blank         | 0.000 | 05:47 |                      |
| Control Test      | 0.079 | 05:48 |                      |
| Air Blank         | 0.000 | 05:48 |                      |
| Subject Sample    | # 1   | 05:48 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:48 |                      |
| Air Blank         | 0.000 | 05:49 |                      |
| Control Test      | 0.078 | 05:49 |                      |
| Air Blank         | 0.000 | 05:49 |                      |
| Diagnostics Check | OK    | 05:50 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSON,G  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: CUMMINS

DATE/TIME OF TEST: 08/10/2024 06:55:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:48 |
| Air Blank          | 0.000 | 06:49 |
| Control Test       | 0.079 | 06:49 |
| Air Blank          | 0.000 | 06:49 |
| Subject Sample # 1 | 0.214 | 06:50 |
| Breath Volume      | 1.933 | 06:50 |
| Air Blank          | 0.000 | 06:51 |
| Air Blank          | 0.000 | 06:53 |
| Subject Sample # 2 | 0.222 | 06:53 |
| Breath Volume      | 1.214 | 06:53 |
| Air Blank          | 0.000 | 06:54 |
| Control Test       | 0.078 | 06:54 |
| Air Blank          | 0.000 | 06:55 |
| Diagnostics Check  | OK    | 06:55 |

---

Data Download Date/Time: 08/22/2024 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHISMAN,N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/10/2024 22:59:09

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:56 |                      |
| Air Blank         | 0.000 | 22:56 |                      |
| Control Test      | 0.078 | 22:57 |                      |
| Air Blank         | 0.000 | 22:57 |                      |
| Subject Sample    | # 1   | 22:57 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:57 |                      |
| Air Blank         | 0.000 | 22:58 |                      |
| Control Test      | 0.079 | 22:58 |                      |
| Air Blank         | 0.000 | 22:59 |                      |
| Diagnostics Check | OK    | 22:59 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HART,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/11/2024 03:19:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:16 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Control Test      | 0.078 | 03:17 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Subject Sample    | # 1   | 03:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:18 |                      |
| Air Blank         | 0.000 | 03:18 |                      |
| Control Test      | 0.078 | 03:18 |                      |
| Air Blank         | 0.000 | 03:19 |                      |
| Diagnostics Check | OK    | 03:19 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,J,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SMYTHERS,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: FERREBUS

DATE/TIME OF TEST: 08/11/2024 09:14:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:06 |
| Air Blank          | 0.000 | 09:07 |
| Control Test       | 0.078 | 09:07 |
| Air Blank          | 0.000 | 09:08 |
| Subject Sample # 1 | 0.136 | 09:08 |
| Breath Volume      | 1.605 | 09:08 |
| Air Blank          | 0.000 | 09:09 |
| Air Blank          | 0.000 | 09:11 |
| Subject Sample # 2 | 0.132 | 09:12 |
| Breath Volume      | 2.023 | 09:12 |
| Air Blank          | 0.000 | 09:13 |
| Control Test       | 0.079 | 09:13 |
| Air Blank          | 0.000 | 09:13 |
| Diagnostics Check  | OK    | 09:14 |

---

Data Download Date/Time: 08/22/2024 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: ALVAREZ MANZANO

DATE/TIME OF TEST: 08/11/2024 22:03:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:00 |                      |
| Air Blank         | 0.000 | 22:01 |                      |
| Control Test      | 0.078 | 22:01 |                      |
| Air Blank         | 0.000 | 22:02 |                      |
| Subject Sample    | # 1   | 22:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:02 |                      |
| Air Blank         | 0.000 | 22:02 |                      |
| Control Test      | 0.078 | 22:03 |                      |
| Air Blank         | 0.000 | 22:03 |                      |
| Diagnostics Check | OK    | 22:03 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MORETA,C  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 08/12/2024 00:23:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:20 |                      |
| Air Blank         | 0.000 | 00:21 |                      |
| Control Test      | 0.078 | 00:21 |                      |
| Air Blank         | 0.000 | 00:22 |                      |
| Subject Sample    | # 1   | 00:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:22 |                      |
| Air Blank         | 0.000 | 00:22 |                      |
| Control Test      | 0.078 | 00:23 |                      |
| Air Blank         | 0.000 | 00:23 |                      |
| Diagnostics Check | OK    | 00:23 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: GARCIA,Y  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: MARTINEZ OCHOA

DATE/TIME OF TEST: 08/15/2024 03:11:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:04 |
| Air Blank          | 0.000 | 03:04 |
| Control Test       | 0.079 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Subject Sample # 1 | 0.077 | 03:06 |
| Breath Volume      | 1.421 | 03:06 |
| Air Blank          | 0.000 | 03:06 |
| Air Blank          | 0.000 | 03:08 |
| Subject Sample # 2 | 0.071 | 03:09 |
| Breath Volume      | 1.152 | 03:09 |
| Air Blank          | 0.000 | 03:10 |
| Control Test       | 0.079 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Diagnostics Check  | OK    | 03:11 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HAWKINS,A  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BEARD

DATE/TIME OF TEST: 08/15/2024 21:36:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:33 |                      |
| Air Blank         | 0.000 | 21:34 |                      |
| Control Test      | 0.078 | 21:34 |                      |
| Air Blank         | 0.000 | 21:35 |                      |
| Subject Sample    | # 1   | 21:35 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:35 |                      |
| Air Blank         | 0.000 | 21:35 |                      |
| Control Test      | 0.078 | 21:36 |                      |
| Air Blank         | 0.000 | 21:36 |                      |
| Diagnostics Check | OK    | 21:36 |                      |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKHAM,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DAVIS,P  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 08/16/2024 01:34:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:27 |
| Air Blank          | 0.000 | 01:27 |
| Control Test       | 0.079 | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Subject Sample # 1 | 0.142 | 01:29 |
| Breath Volume      | 1.132 | 01:29 |
| Air Blank          | 0.000 | 01:29 |
| Air Blank          | 0.000 | 01:31 |
| Subject Sample # 2 | 0.139 | 01:33 |
| Breath Volume      | 1.726 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Control Test       | 0.079 | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Diagnostics Check  | OK    | 01:34 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SOLIS SANTIZO

DATE/TIME OF TEST: 08/16/2024 23:29:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:21 |
| Air Blank          | 0.000 | 23:22 |
| Control Test       | 0.078 | 23:22 |
| Air Blank          | 0.000 | 23:23 |
| Subject Sample # 1 | 0.292 | 23:24 |
| Breath Volume      | 3.917 | 23:24 |
| Air Blank          | 0.000 | 23:25 |
| Air Blank          | 0.000 | 23:26 |
| Subject Sample # 2 | 0.278 | 23:27 |
| Breath Volume      | 3.539 | 23:27 |
| Air Blank          | 0.000 | 23:28 |
| Control Test       | 0.079 | 23:28 |
| Air Blank          | 0.000 | 23:29 |
| Diagnostics Check  | OK    | 23:29 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/17/2024 00:48:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.078 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 1 | 0.073 | 00:44 |
| Breath Volume      | 1.828 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Air Blank          | 0.000 | 00:47 |
| Subject Sample # 2 | 0.070 | 00:47 |
| Breath Volume      | 1.312 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Control Test       | 0.078 | 00:48 |
| Air Blank          | 0.000 | 00:48 |
| Diagnostics Check  | OK    | 00:48 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 08/17/2024 04:16:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.078 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 1 | 0.141 | 04:11 |
| Breath Volume      | 3.398 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Air Blank          | 0.000 | 04:14 |
| Subject Sample # 2 | 0.139 | 04:15 |
| Breath Volume      | 2.218 | 04:15 |
| Air Blank          | 0.000 | 04:15 |
| Control Test       | 0.078 | 04:16 |
| Air Blank          | 0.000 | 04:16 |
| Diagnostics Check  | OK    | 04:16 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DAVIS,P  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: OLIVERA IBARRA

DATE/TIME OF TEST: 08/18/2024 01:47:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Control Test       | 0.079 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 1 | 0.020 | 01:43 |
| Breath Volume      | 2.519 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Subject Sample # 2 | 0.017 | 01:46 |
| Breath Volume      | 2.386 | 01:46 |
| Air Blank          | 0.000 | 01:46 |
| Control Test       | 0.078 | 01:47 |
| Air Blank          | 0.000 | 01:47 |
| Diagnostics Check  | OK    | 01:47 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GLADDEN,A,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TAYLOR,G  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: HENDRY

DATE/TIME OF TEST: 08/18/2024 09:28:25

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 09:22 |            |
| Air Blank         | 0.000     | 09:23 |            |
| Control Test      | 0.079     | 09:23 |            |
| Air Blank         | 0.000     | 09:23 |            |
| Subject Sample    | # 1 0.143 | 09:24 |            |
| Breath Volume     | 3.140     | 09:24 |            |
| Air Blank         | 0.000     | 09:25 |            |
| Air Blank         |           | 09:27 | RFI Detect |
| Air Blank         | 0.000     | 09:27 |            |
| Control Test      | 0.077     | 09:27 |            |
| Air Blank         | 0.000     | 09:28 |            |
| Diagnostics Check | OK        | 09:28 |            |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GLADDEN,A,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TAYLOR,G  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: HENDRY

DATE/TIME OF TEST: 08/18/2024 09:38:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:32 |
| Air Blank          | 0.000 | 09:32 |
| Control Test       | 0.078 | 09:32 |
| Air Blank          | 0.000 | 09:33 |
| Subject Sample # 1 | 0.135 | 09:34 |
| Breath Volume      | 3.304 | 09:34 |
| Air Blank          | 0.000 | 09:34 |
| Air Blank          | 0.000 | 09:36 |
| Subject Sample # 2 | 0.134 | 09:37 |
| Breath Volume      | 2.847 | 09:37 |
| Air Blank          | 0.000 | 09:37 |
| Control Test       | 0.078 | 09:37 |
| Air Blank          | 0.000 | 09:38 |
| Diagnostics Check  | OK    | 09:38 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DAVIS,P  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 08/18/2024 21:48:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Control Test       | 0.078 | 21:42 |
| Air Blank          | 0.000 | 21:43 |
| Subject Sample # 1 | 0.134 | 21:43 |
| Breath Volume      | 1.105 | 21:43 |
| Air Blank          | 0.000 | 21:44 |
| Air Blank          | 0.000 | 21:46 |
| Subject Sample # 2 | 0.139 | 21:47 |
| Breath Volume      | 1.867 | 21:47 |
| Air Blank          | 0.000 | 21:47 |
| Control Test       | 0.078 | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Diagnostics Check  | OK    | 21:48 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSON,I  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/21/2024 02:23:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Control Test       | 0.079 | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Subject Sample # 1 | 0.240 | 02:18 |
| Breath Volume      | 1.867 | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 2 | 0.239 | 02:21 |
| Breath Volume      | 2.042 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Control Test       | 0.078 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Diagnostics Check  | OK    | 02:23 |

---

Data Download Date/Time: 08/22/2024 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LE,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 08/22/2024 21:48:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:45 |                      |
| Air Blank         | 0.000 | 21:46 |                      |
| Control Test      | 0.079 | 21:46 |                      |
| Air Blank         | 0.000 | 21:47 |                      |
| Subject Sample    | # 1   | 21:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:47 |                      |
| Air Blank         | 0.000 | 21:47 |                      |
| Control Test      | 0.079 | 21:48 |                      |
| Air Blank         | 0.000 | 21:48 |                      |
| Diagnostics Check | OK    | 21:48 |                      |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LE,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: GARCIA,Y  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: VILLALOBOS HERNANDEZ

DATE/TIME OF TEST: 08/23/2024 03:08:22

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:05 |                      |
| Air Blank         | 0.000 | 03:05 |                      |
| Control Test      | 0.079 | 03:06 |                      |
| Air Blank         | 0.000 | 03:06 |                      |
| Subject Sample    | # 1   | 03:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:07 |                      |
| Air Blank         | 0.000 | 03:07 |                      |
| Control Test      | 0.079 | 03:07 |                      |
| Air Blank         | 0.000 | 03:08 |                      |
| Diagnostics Check | OK    | 03:08 |                      |

---

Data Download Date/Time: 09/16/2024 22:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWELL,EDWARD,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/23/2024 04:24:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Control Test       | 0.079 | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Subject Sample # 1 | 0.174 | 04:19 |
| Breath Volume      | 3.312 | 04:19 |
| Air Blank          | 0.000 | 04:20 |
| Air Blank          | 0.000 | 04:22 |
| Subject Sample # 2 | 0.190 | 04:22 |
| Breath Volume      | 3.386 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Control Test       | 0.079 | 04:23 |
| Air Blank          | 0.000 | 04:24 |
| Diagnostics Check  | OK    | 04:24 |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALE,J,K  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 13:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 08/23/2024 15:21:37

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 15:10 |                |
| Air Blank         | 0.000     | 15:10 |                |
| Control Test      | 0.079     | 15:10 |                |
| Air Blank         | 0.000     | 15:11 |                |
| Subject Sample    | # 1 0.000 | 15:14 | Volume Not Met |
| Breath Volume     | 0.480     | 15:14 |                |
| Air Blank         | 0.000     | 15:15 |                |
| Air Blank         | 0.000     | 15:17 |                |
| Subject Sample    | # 2 0.000 | 15:20 | Volume Not Met |
| Breath Volume     | 0.707     | 15:20 |                |
| Air Blank         | 0.000     | 15:20 |                |
| Control Test      | 0.079     | 15:21 |                |
| Air Blank         | 0.000     | 15:21 |                |
| Diagnostics Check | OK        | 15:21 |                |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALE,J,K  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 13:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 08/23/2024 15:32:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:24 |
| Air Blank          | 0.000 | 15:25 |
| Control Test       | 0.078 | 15:25 |
| Air Blank          | 0.000 | 15:25 |
| Subject Sample # 1 | 0.000 | 15:27 |
| Breath Volume      | 1.687 | 15:27 |
| Air Blank          | 0.000 | 15:28 |
| Air Blank          | 0.000 | 15:30 |
| Subject Sample # 2 | 0.000 | 15:30 |
| Breath Volume      | 1.222 | 15:30 |
| Air Blank          | 0.000 | 15:31 |
| Control Test       | 0.078 | 15:31 |
| Air Blank          | 0.000 | 15:32 |
| Diagnostics Check  | OK    | 15:32 |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWELL,EDWARD,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LAGODNY

DATE/TIME OF TEST: 08/23/2024 23:33:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:30 |                      |
| Air Blank         | 0.000 | 23:31 |                      |
| Control Test      | 0.079 | 23:31 |                      |
| Air Blank         | 0.000 | 23:32 |                      |
| Subject Sample    | # 1   | 23:32 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:32 |                      |
| Air Blank         | 0.000 | 23:32 |                      |
| Control Test      | 0.078 | 23:32 |                      |
| Air Blank         | 0.000 | 23:33 |                      |
| Diagnostics Check | OK    | 23:33 |                      |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWELL,EDWARD,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HART,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: JACINTO BRAVO

DATE/TIME OF TEST: 08/24/2024 02:02:25

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:59 |                      |
| Air Blank         | 0.000 | 02:00 |                      |
| Control Test      | 0.079 | 02:00 |                      |
| Air Blank         | 0.000 | 02:00 |                      |
| Subject Sample    | # 1   | 02:01 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:01 |                      |
| Air Blank         | 0.000 | 02:01 |                      |
| Control Test      | 0.078 | 02:01 |                      |
| Air Blank         | 0.000 | 02:02 |                      |
| Diagnostics Check | OK    | 02:02 |                      |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LE,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: GARCIA,Y  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 08/24/2024 03:58:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.078 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Subject Sample # 1 | 0.257 | 03:53 |
| Breath Volume      | 2.835 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Subject Sample # 2 | 0.244 | 03:56 |
| Breath Volume      | 1.734 | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Control Test       | 0.079 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Diagnostics Check  | OK    | 03:58 |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWELL,EDWARD,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: FORSHEE

DATE/TIME OF TEST: 08/24/2024 05:20:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:13 |
| Air Blank          | 0.000 | 05:14 |
| Control Test       | 0.079 | 05:14 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 1 | 0.104 | 05:16 |
| Breath Volume      | 2.363 | 05:16 |
| Air Blank          | 0.000 | 05:16 |
| Air Blank          | 0.000 | 05:18 |
| Subject Sample # 2 | 0.108 | 05:18 |
| Breath Volume      | 2.929 | 05:18 |
| Air Blank          | 0.000 | 05:19 |
| Control Test       | 0.079 | 05:19 |
| Air Blank          | 0.000 | 05:20 |
| Diagnostics Check  | OK    | 05:20 |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: JOHNSON,G  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 08/25/2024 00:51:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:45 |
| Air Blank          | 0.000 | 00:45 |
| Control Test       | 0.079 | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 1 | 0.114 | 00:46 |
| Breath Volume      | 1.128 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Air Blank          | 0.000 | 00:49 |
| Subject Sample # 2 | 0.115 | 00:50 |
| Breath Volume      | 1.550 | 00:50 |
| Air Blank          | 0.000 | 00:50 |
| Control Test       | 0.078 | 00:50 |
| Air Blank          | 0.000 | 00:51 |
| Diagnostics Check  | OK    | 00:51 |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: THOMPSON,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: RODRIGUEZ SANDOVAL

DATE/TIME OF TEST: 08/25/2024 05:17:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.078 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Subject Sample # 1 | 0.204 | 05:12 |
| Breath Volume      | 1.226 | 05:12 |
| Air Blank          | 0.000 | 05:13 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 2 | 0.204 | 05:15 |
| Breath Volume      | 1.496 | 05:15 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.079 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Diagnostics Check  | OK    | 05:17 |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON,B,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: THOMPSON,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: JOHNSON JR

DATE/TIME OF TEST: 08/26/2024 03:59:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:56 |                      |
| Air Blank         | 0.000 | 03:57 |                      |
| Control Test      | 0.079 | 03:57 |                      |
| Air Blank         | 0.000 | 03:57 |                      |
| Subject Sample    | # 1   | 03:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:58 |                      |
| Air Blank         | 0.000 | 03:58 |                      |
| Control Test      | 0.078 | 03:58 |                      |
| Air Blank         | 0.000 | 03:59 |                      |
| Diagnostics Check | OK    | 03:59 |                      |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLIDAY,E,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WORDEN

DATE/TIME OF TEST: 08/26/2024 22:36:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:30 |
| Air Blank          | 0.000 | 22:30 |
| Control Test       | 0.079 | 22:31 |
| Air Blank          | 0.000 | 22:31 |
| Subject Sample # 1 | 0.112 | 22:32 |
| Breath Volume      | 2.378 | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Air Blank          | 0.000 | 22:34 |
| Subject Sample # 2 | 0.114 | 22:35 |
| Breath Volume      | 2.472 | 22:35 |
| Air Blank          | 0.000 | 22:35 |
| Control Test       | 0.079 | 22:36 |
| Air Blank          | 0.000 | 22:36 |
| Diagnostics Check  | OK    | 22:36 |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DAVIS,P  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MARTINEZ MORENO

DATE/TIME OF TEST: 08/28/2024 02:24:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:18 |
| Air Blank          | 0.000 | 02:18 |
| Control Test       | 0.078 | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Subject Sample # 1 | 0.081 | 02:20 |
| Breath Volume      | 2.007 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Subject Sample # 2 | 0.076 | 02:23 |
| Breath Volume      | 2.636 | 02:23 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.078 | 02:24 |
| Air Blank          | 0.000 | 02:24 |
| Diagnostics Check  | OK    | 02:24 |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CHAVEZ,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 08/28/2024 22:10:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:03 |
| Air Blank          | 0.000 | 22:04 |
| Control Test       | 0.078 | 22:04 |
| Air Blank          | 0.000 | 22:04 |
| Subject Sample # 1 | 0.143 | 22:05 |
| Breath Volume      | 1.503 | 22:05 |
| Air Blank          | 0.000 | 22:06 |
| Air Blank          | 0.000 | 22:08 |
| Subject Sample # 2 | 0.145 | 22:09 |
| Breath Volume      | 1.582 | 22:09 |
| Air Blank          | 0.000 | 22:09 |
| Control Test       | 0.079 | 22:10 |
| Air Blank          | 0.000 | 22:10 |
| Diagnostics Check  | OK    | 22:10 |

---

Data Download Date/Time: 09/16/2024 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CHAVEZ,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 08/29/2024 03:06:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:04 |                      |
| Air Blank         | 0.000 | 03:04 |                      |
| Control Test      | 0.079 | 03:04 |                      |
| Air Blank         | 0.000 | 03:05 |                      |
| Subject Sample    | # 1   | 03:05 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:05 |                      |
| Air Blank         | 0.000 | 03:05 |                      |
| Control Test      | 0.079 | 03:06 |                      |
| Air Blank         | 0.000 | 03:06 |                      |
| Diagnostics Check | OK    | 03:06 |                      |

---

Data Download Date/Time: 09/16/2024 22:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALE,J,K  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 14:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: PECK

DATE/TIME OF TEST: 08/29/2024 16:30:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:22 |
| Air Blank          | 0.000 | 16:23 |
| Control Test       | 0.078 | 16:23 |
| Air Blank          | 0.000 | 16:23 |
| Subject Sample # 1 | 0.147 | 16:25 |
| Breath Volume      | 1.437 | 16:25 |
| Air Blank          | 0.000 | 16:25 |
| Air Blank          | 0.000 | 16:27 |
| Subject Sample # 2 | 0.159 | 16:28 |
| Breath Volume      | 2.097 | 16:28 |
| Air Blank          | 0.000 | 16:29 |
| Control Test       | 0.078 | 16:29 |
| Air Blank          | 0.000 | 16:30 |
| Diagnostics Check  | OK    | 16:30 |

---

Data Download Date/Time: 09/16/2024 22:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTTER, JONAH, L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HART, A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/30/2024 03:18:26

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:15 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Control Test      | 0.078 | 03:16 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Subject Sample    | # 1   | 03:17 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:17 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Control Test      | 0.079 | 03:17 |                      |
| Air Blank         | 0.000 | 03:18 |                      |
| Diagnostics Check | OK    | 03:18 |                      |

Data Download Date/Time: 09/16/2024 22:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LE,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: IEZZI,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/30/2024 22:58:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:51 |
| Air Blank          | 0.000 | 22:52 |
| Control Test       | 0.079 | 22:52 |
| Air Blank          | 0.000 | 22:52 |
| Subject Sample # 1 | 0.000 | 22:54 |
| Breath Volume      | 2.433 | 22:54 |
| Air Blank          | 0.000 | 22:54 |
| Air Blank          | 0.000 | 22:56 |
| Subject Sample # 2 | 0.000 | 22:57 |
| Breath Volume      | 2.082 | 22:57 |
| Air Blank          | 0.000 | 22:57 |
| Control Test       | 0.079 | 22:57 |
| Air Blank          | 0.000 | 22:58 |
| Diagnostics Check  | OK    | 22:58 |

---

Data Download Date/Time: 09/16/2024 22:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LE,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHISMAN,N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/31/2024 05:06:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:59 |
| Air Blank          | 0.000 | 04:59 |
| Control Test       | 0.079 | 05:00 |
| Air Blank          | 0.000 | 05:00 |
| Subject Sample # 1 | 0.125 | 05:01 |
| Breath Volume      | 2.550 | 05:01 |
| Air Blank          | 0.000 | 05:02 |
| Air Blank          | 0.000 | 05:04 |
| Subject Sample # 2 | 0.130 | 05:05 |
| Breath Volume      | 2.046 | 05:05 |
| Air Blank          | 0.000 | 05:05 |
| Control Test       | 0.078 | 05:06 |
| Air Blank          | 0.000 | 05:06 |
| Diagnostics Check  | OK    | 05:06 |

---

Data Download Date/Time: 09/16/2024 22:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWSON,K,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MARTIN,R  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: BLATHERS

DATE/TIME OF TEST: 08/31/2024 06:03:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:56 |
| Air Blank          | 0.000 | 05:56 |
| Control Test       | 0.078 | 05:57 |
| Air Blank          | 0.000 | 05:57 |
| Subject Sample # 1 | 0.142 | 05:58 |
| Breath Volume      | 2.300 | 05:58 |
| Air Blank          | 0.000 | 05:59 |
| Air Blank          | 0.000 | 06:01 |
| Subject Sample # 2 | 0.144 | 06:01 |
| Breath Volume      | 2.285 | 06:01 |
| Air Blank          | 0.000 | 06:02 |
| Control Test       | 0.078 | 06:02 |
| Air Blank          | 0.000 | 06:03 |
| Diagnostics Check  | OK    | 06:03 |

---

Data Download Date/Time: 09/16/2024 22:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALE,J,K  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: IEZZI,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 15:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MONTAGUE

DATE/TIME OF TEST: 08/31/2024 17:23:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:21 |                      |
| Air Blank         | 0.000 | 17:21 |                      |
| Control Test      | 0.080 | 17:21 |                      |
| Air Blank         | 0.000 | 17:22 |                      |
| Subject Sample    | # 1   | 17:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:22 |                      |
| Air Blank         | 0.000 | 17:22 |                      |
| Control Test      | 0.080 | 17:23 |                      |
| Air Blank         | 0.000 | 17:23 |                      |
| Diagnostics Check | OK    | 17:23 |                      |

---

Data Download Date/Time: 09/16/2024 22:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALE,J,K  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LITTLE,J  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 08/31/2024 18:11:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:05 |
| Air Blank          | 0.000 | 18:05 |
| Control Test       | 0.080 | 18:06 |
| Air Blank          | 0.000 | 18:06 |
| Subject Sample # 1 | 0.234 | 18:07 |
| Breath Volume      | 2.562 | 18:07 |
| Air Blank          | 0.000 | 18:07 |
| Air Blank          | 0.000 | 18:09 |
| Subject Sample # 2 | 0.246 | 18:10 |
| Breath Volume      | 1.539 | 18:10 |
| Air Blank          | 0.000 | 18:10 |
| Control Test       | 0.078 | 18:11 |
| Air Blank          | 0.000 | 18:11 |
| Diagnostics Check  | OK    | 18:11 |

---

Data Download Date/Time: 09/16/2024 22:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LE,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SAPP,W  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: BUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 08/31/2024 21:37:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:30 |
| Air Blank          | 0.000 | 21:30 |
| Control Test       | 0.079 | 21:31 |
| Air Blank          | 0.000 | 21:31 |
| Subject Sample # 1 | 0.031 | 21:32 |
| Breath Volume      | 1.851 | 21:32 |
| Air Blank          | 0.000 | 21:33 |
| Air Blank          | 0.000 | 21:35 |
| Subject Sample # 2 | 0.031 | 21:35 |
| Breath Volume      | 2.171 | 21:35 |
| Air Blank          | 0.000 | 21:36 |
| Control Test       | 0.078 | 21:36 |
| Air Blank          | 0.000 | 21:37 |
| Diagnostics Check  | OK    | 21:37 |

---

Data Download Date/Time: 09/16/2024 22:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: SMITH,JAYSON  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 07/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 08/07/2024 04:09:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.080 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 1 | 0.120 | 04:01 |
| Breath Volume      | 1.250 | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 2 | 0.148 | 04:05 |
| Breath Volume      | 2.835 | 04:05 |
| Air Blank          | 0.000 | 04:05 |
| Air Blank          | 0.000 | 04:07 |
| Subject Sample # 3 | 0.144 | 04:08 |
| Breath Volume      | 1.984 | 04:08 |
| Air Blank          | 0.000 | 04:08 |
| Control Test       | 0.079 | 04:09 |
| Air Blank          | 0.000 | 04:09 |
| Diagnostics Check  | OK    | 04:09 |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CALDWELL,GARY,D  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: CALDWELL,GARY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 07/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:33

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 08/09/2024 14:56:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 14:53 |                      |
| Air Blank         | 0.000 | 14:54 |                      |
| Control Test      | 0.080 | 14:54 |                      |
| Air Blank         | 0.000 | 14:55 |                      |
| Subject Sample    | # 1   | 14:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 14:55 |                      |
| Air Blank         | 0.000 | 14:55 |                      |
| Control Test      | 0.079 | 14:56 |                      |
| Air Blank         | 0.000 | 14:56 |                      |
| Diagnostics Check | OK    | 14:56 |                      |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 07/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 08/09/2024 20:31:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:24 |
| Air Blank          | 0.000 | 20:24 |
| Control Test       | 0.079 | 20:24 |
| Air Blank          | 0.000 | 20:25 |
| Subject Sample # 1 | 0.252 | 20:26 |
| Breath Volume      | 1.382 | 20:26 |
| Air Blank          | 0.000 | 20:26 |
| Air Blank          | 0.000 | 20:28 |
| Subject Sample # 2 | 0.250 | 20:30 |
| Breath Volume      | 1.230 | 20:30 |
| Air Blank          | 0.000 | 20:30 |
| Control Test       | 0.078 | 20:31 |
| Air Blank          | 0.000 | 20:31 |
| Diagnostics Check  | OK    | 20:31 |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 07/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 08/09/2024 23:56:25

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:53 |                      |
| Air Blank         | 0.000 | 23:54 |                      |
| Control Test      | 0.080 | 23:54 |                      |
| Air Blank         | 0.000 | 23:54 |                      |
| Subject Sample    | # 1   | 23:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:55 |                      |
| Air Blank         | 0.000 | 23:55 |                      |
| Control Test      | 0.079 | 23:55 |                      |
| Air Blank         | 0.000 | 23:56 |                      |
| Diagnostics Check | OK    | 23:56 |                      |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: SMITH,JAYSON  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 07/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MOORHEAD

DATE/TIME OF TEST: 08/11/2024 01:25:39

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:22 |                      |
| Air Blank         | 0.000 | 01:22 |                      |
| Control Test      | 0.080 | 01:23 |                      |
| Air Blank         | 0.000 | 01:23 |                      |
| Subject Sample    | # 1   | 01:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:24 |                      |
| Air Blank         | 0.000 | 01:24 |                      |
| Control Test      | 0.079 | 01:25 |                      |
| Air Blank         | 0.000 | 01:25 |                      |
| Diagnostics Check | OK    | 01:25 |                      |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: SMITH,JAYSON  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 07/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 08/11/2024 03:54:08

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:50 |                      |
| Air Blank         | 0.000 | 03:51 |                      |
| Control Test      | 0.081 | 03:51 |                      |
| Air Blank         | 0.000 | 03:51 |                      |
| Subject Sample    | # 1   | 03:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:52 |                      |
| Air Blank         | 0.000 | 03:53 |                      |
| Control Test      | 0.079 | 03:53 |                      |
| Air Blank         | 0.000 | 03:54 |                      |
| Diagnostics Check | OK    | 03:54 |                      |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUY,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: SMITH,JAYSON  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 07/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MINNEAR

DATE/TIME OF TEST: 08/13/2024 21:38:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:35 |                      |
| Air Blank         | 0.000 | 21:35 |                      |
| Control Test      | 0.081 | 21:36 |                      |
| Air Blank         | 0.000 | 21:36 |                      |
| Subject Sample    | # 1   | 21:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:36 |                      |
| Air Blank         | 0.000 | 21:37 |                      |
| Control Test      | 0.079 | 21:37 |                      |
| Air Blank         | 0.000 | 21:38 |                      |
| Diagnostics Check | OK    | 21:38 |                      |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 08/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 08/16/2024 21:26:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:20 |
| Air Blank          | 0.000 | 21:20 |
| Control Test       | 0.080 | 21:21 |
| Air Blank          | 0.000 | 21:21 |
| Subject Sample # 1 | 0.137 | 21:21 |
| Breath Volume      | 1.257 | 21:21 |
| Air Blank          | 0.000 | 21:22 |
| Air Blank          | 0.000 | 21:24 |
| Subject Sample # 2 | 0.128 | 21:25 |
| Breath Volume      | 1.902 | 21:25 |
| Air Blank          | 0.000 | 21:25 |
| Control Test       | 0.079 | 21:25 |
| Air Blank          | 0.000 | 21:26 |
| Diagnostics Check  | OK    | 21:26 |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CALDWELL,GARY,D  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: CALDWELL,GARY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 08/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: PEEK

DATE/TIME OF TEST: 08/16/2024 22:39:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Control Test       | 0.080 | 22:34 |
| Air Blank          | 0.000 | 22:34 |
| Subject Sample # 1 | 0.000 | 22:35 |
| Breath Volume      | 3.652 | 22:35 |
| Air Blank          | 0.000 | 22:35 |
| Air Blank          | 0.000 | 22:37 |
| Subject Sample # 2 | 0.000 | 22:38 |
| Breath Volume      | 3.347 | 22:38 |
| Air Blank          | 0.000 | 22:38 |
| Control Test       | 0.080 | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Diagnostics Check  | OK    | 22:39 |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CALDWELL,GARY,D  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: CALDWELL,GARY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 08/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: WALASON

DATE/TIME OF TEST: 08/17/2024 01:23:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.080 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Subject Sample # 1 | 0.112 | 01:19 |
| Breath Volume      | 1.746 | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Air Blank          | 0.000 | 01:22 |
| Subject Sample # 2 | 0.111 | 01:22 |
| Breath Volume      | 1.335 | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Control Test       | 0.079 | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Diagnostics Check  | OK    | 01:23 |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 08/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SCHUMPERT

DATE/TIME OF TEST: 08/17/2024 02:11:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:09 |                      |
| Air Blank         | 0.000 | 02:09 |                      |
| Control Test      | 0.080 | 02:09 |                      |
| Air Blank         | 0.000 | 02:10 |                      |
| Subject Sample    | # 1   | 02:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:10 |                      |
| Air Blank         | 0.000 | 02:10 |                      |
| Control Test      | 0.079 | 02:11 |                      |
| Air Blank         | 0.000 | 02:11 |                      |
| Diagnostics Check | OK    | 02:11 |                      |

---

Data Download Date/Time: 08/17/2024 18:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CALDWELL,GARY,D  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: CALDWELL,GARY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 08/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: MEHAFFEY

DATE/TIME OF TEST: 08/29/2024 15:46:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:43 |                      |
| Air Blank         | 0.000 | 15:44 |                      |
| Control Test      | 0.081 | 15:44 |                      |
| Air Blank         | 0.000 | 15:45 |                      |
| Subject Sample    | # 1   | 15:45 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:45 |                      |
| Air Blank         | 0.000 | 15:46 |                      |
| Control Test      | 0.079 | 15:46 |                      |
| Air Blank         | 0.000 | 15:46 |                      |
| Diagnostics Check | OK    | 15:46 |                      |

---

Data Download Date/Time: 10/02/2024 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUY,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: GUY,JOSHUA  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 08/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A4  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 08/30/2024 23:22:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:16 |
| Air Blank          | 0.000 | 23:17 |
| Control Test       | 0.081 | 23:17 |
| Air Blank          | 0.000 | 23:17 |
| Subject Sample # 1 | 0.176 | 23:18 |
| Breath Volume      | 3.011 | 23:18 |
| Air Blank          | 0.000 | 23:18 |
| Air Blank          | 0.000 | 23:20 |
| Subject Sample # 2 | 0.171 | 23:21 |
| Breath Volume      | 3.410 | 23:21 |
| Air Blank          | 0.000 | 23:21 |
| Control Test       | 0.080 | 23:22 |
| Air Blank          | 0.000 | 23:22 |
| Diagnostics Check  | OK    | 23:22 |

---

Data Download Date/Time: 10/02/2024 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: DE LA CAMPA,BRYANT  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 07/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: GONZALEZ VELASQUEZ

DATE/TIME OF TEST: 08/10/2024 15:00:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:50 |
| Air Blank          | 0.000 | 14:51 |
| Control Test       | 0.080 | 14:51 |
| Air Blank          | 0.000 | 14:51 |
| Subject Sample # 1 | 0.233 | 14:52 |
| Breath Volume      | 1.308 | 14:52 |
| Air Blank          | 0.000 | 14:53 |
| Air Blank          | 0.000 | 14:55 |
| Subject Sample # 2 | 0.203 | 14:56 |
| Breath Volume      | 2.078 | 14:56 |
| Air Blank          | 0.000 | 14:56 |
| Air Blank          | 0.000 | 14:58 |
| Subject Sample # 3 | 0.221 | 14:59 |
| Breath Volume      | 2.324 | 14:59 |
| Air Blank          | 0.000 | 14:59 |
| Control Test       | 0.079 | 15:00 |
| Air Blank          | 0.000 | 15:00 |
| Diagnostics Check  | OK    | 15:00 |

---

Data Download Date/Time: 08/26/2024 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,GAVIN,O  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: HENRY,GAVIN  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 07/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 08/11/2024 01:37:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.080 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 1 | 0.184 | 01:33 |
| Breath Volume      | 1.468 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.196 | 01:35 |
| Breath Volume      | 1.804 | 01:35 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.081 | 01:36 |
| Air Blank          | 0.000 | 01:37 |
| Diagnostics Check  | OK    | 01:37 |

---

Data Download Date/Time: 08/26/2024 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: DE LA CAMPA,BRYANT  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 07/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 08/11/2024 14:35:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:29 |
| Air Blank          | 0.000 | 14:29 |
| Control Test       | 0.080 | 14:30 |
| Air Blank          | 0.000 | 14:30 |
| Subject Sample # 1 | 0.000 | 14:31 |
| Breath Volume      | 3.843 | 14:31 |
| Air Blank          | 0.000 | 14:31 |
| Air Blank          | 0.000 | 14:33 |
| Subject Sample # 2 | 0.000 | 14:33 |
| Breath Volume      | 4.003 | 14:33 |
| Air Blank          | 0.000 | 14:34 |
| Control Test       | 0.081 | 14:34 |
| Air Blank          | 0.000 | 14:35 |
| Diagnostics Check  | OK    | 14:35 |

---

Data Download Date/Time: 08/26/2024 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ESCOBAR,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 07/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DUMELLE

DATE/TIME OF TEST: 08/14/2024 02:10:00

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:07 |                      |
| Air Blank         | 0.000 | 02:07 |                      |
| Control Test      | 0.080 | 02:08 |                      |
| Air Blank         | 0.000 | 02:08 |                      |
| Subject Sample    | # 1   | 02:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:08 |                      |
| Air Blank         | 0.000 | 02:09 |                      |
| Control Test      | 0.081 | 02:09 |                      |
| Air Blank         | 0.000 | 02:09 |                      |
| Diagnostics Check | OK    | 02:10 |                      |

Data Download Date/Time: 08/26/2024 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORALES,DAVID,J  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: HENNINGSEN,KENNETH  
ARRESTING OFFICER AGENCY: Coral Springs PD  
LAST AGENCY INSPECTION: 07/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 08:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:29

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/18/2024 09:01:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:55 |
| Air Blank          | 0.000 | 08:55 |
| Control Test       | 0.079 | 08:56 |
| Air Blank          | 0.000 | 08:56 |
| Subject Sample # 1 | 0.101 | 08:57 |
| Breath Volume      | 2.488 | 08:57 |
| Air Blank          | 0.000 | 08:58 |
| Air Blank          | 0.000 | 08:59 |
| Subject Sample # 2 | 0.087 | 09:00 |
| Breath Volume      | 1.183 | 09:00 |
| Air Blank          | 0.000 | 09:01 |
| Control Test       | 0.080 | 09:01 |
| Air Blank          | 0.000 | 09:01 |
| Diagnostics Check  | OK    | 09:01 |

---

Data Download Date/Time: 08/26/2024 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORALES,DAVID,J  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: BESADA,JOHN  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 07/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:46

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 08/18/2024 16:28:37

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 16:24 |                      |
| Air Blank         | 0.000 | 16:24 |                      |
| Control Test      | 0.080 | 16:25 |                      |
| Air Blank         | 0.000 | 16:25 |                      |
| Subject Sample    | # 1   | 16:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 16:27 |                      |
| Air Blank         | 0.000 | 16:27 |                      |
| Control Test      | 0.082 | 16:28 |                      |
| Air Blank         | 0.000 | 16:28 |                      |
| Diagnostics Check | OK    | 16:28 |                      |

---

Data Download Date/Time: 08/26/2024 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ESCOBAR,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 07/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 08/20/2024 23:09:35

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:56 |                |
| Air Blank         | 0.000     | 22:57 |                |
| Control Test      | 0.080     | 22:57 |                |
| Air Blank         | 0.000     | 22:58 |                |
| Subject Sample    | # 1 0.227 | 23:01 | Volume Not Met |
| Breath Volume     | 1.050     | 23:01 |                |
| Air Blank         | 0.000     | 23:01 |                |
| Air Blank         | 0.000     | 23:03 |                |
| Subject Sample    | # 2 0.236 | 23:05 |                |
| Breath Volume     | 1.652     | 23:05 |                |
| Air Blank         | 0.000     | 23:05 |                |
| Air Blank         | 0.000     | 23:07 |                |
| Subject Sample    | # 3 0.235 | 23:08 |                |
| Breath Volume     | 1.917     | 23:08 |                |
| Air Blank         | 0.000     | 23:08 |                |
| Control Test      | 0.079     | 23:09 |                |
| Air Blank         | 0.000     | 23:09 |                |
| Diagnostics Check | OK        | 23:09 |                |

---

Data Download Date/Time: 08/26/2024 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,GAVIN,O  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: HENRY,GAVIN  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 07/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MYRTIL

DATE/TIME OF TEST: 08/25/2024 01:56:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:49 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.080 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Subject Sample # 1 | 0.165 | 01:51 |
| Breath Volume      | 1.128 | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Air Blank          | 0.000 | 01:53 |
| Subject Sample # 2 | 0.171 | 01:54 |
| Breath Volume      | 1.156 | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Control Test       | 0.080 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Diagnostics Check  | OK    | 01:56 |

---

Data Download Date/Time: 08/26/2024 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CHAVEZ,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/01/2024 03:45:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:42 |                      |
| Air Blank         | 0.000 | 03:42 |                      |
| Control Test      | 0.081 | 03:43 |                      |
| Air Blank         | 0.000 | 03:43 |                      |
| Subject Sample    | # 1   | 03:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:43 |                      |
| Air Blank         | 0.000 | 03:44 |                      |
| Control Test      | 0.081 | 03:44 |                      |
| Air Blank         | 0.000 | 03:45 |                      |
| Diagnostics Check | OK    | 03:45 |                      |

---

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SWEET

DATE/TIME OF TEST: 08/05/2024 01:57:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:50 |
| Air Blank          | 0.000 | 01:50 |
| Control Test       | 0.081 | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Subject Sample # 1 | 0.175 | 01:52 |
| Breath Volume      | 2.398 | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 2 | 0.177 | 01:55 |
| Breath Volume      | 1.855 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.079 | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Diagnostics Check  | OK    | 01:57 |

---

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHEY,A,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LIGHTSEY,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 09:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 08/07/2024 11:45:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:38 |
| Air Blank          | 0.000 | 11:38 |
| Control Test       | 0.081 | 11:39 |
| Air Blank          | 0.000 | 11:39 |
| Subject Sample # 1 | 0.000 | 11:40 |
| Breath Volume      | 2.093 | 11:40 |
| Air Blank          | 0.000 | 11:41 |
| Air Blank          | 0.000 | 11:43 |
| Subject Sample # 2 | 0.000 | 11:44 |
| Breath Volume      | 1.410 | 11:44 |
| Air Blank          | 0.000 | 11:44 |
| Control Test       | 0.081 | 11:44 |
| Air Blank          | 0.000 | 11:45 |
| Diagnostics Check  | OK    | 11:45 |

---

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHEY,A,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MILLER,K  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 11:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/07/2024 14:12:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:05 |
| Air Blank          | 0.000 | 14:06 |
| Control Test       | 0.081 | 14:06 |
| Air Blank          | 0.000 | 14:07 |
| Subject Sample # 1 | 0.179 | 14:07 |
| Breath Volume      | 1.710 | 14:07 |
| Air Blank          | 0.000 | 14:08 |
| Air Blank          | 0.000 | 14:10 |
| Subject Sample # 2 | 0.179 | 14:10 |
| Breath Volume      | 1.664 | 14:10 |
| Air Blank          | 0.000 | 14:11 |
| Control Test       | 0.079 | 14:11 |
| Air Blank          | 0.000 | 14:12 |
| Diagnostics Check  | OK    | 14:12 |

---

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHEY,A,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: KAROLEWICZ

DATE/TIME OF TEST: 08/07/2024 18:38:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:31 |
| Air Blank          | 0.000 | 18:32 |
| Control Test       | 0.081 | 18:32 |
| Air Blank          | 0.000 | 18:33 |
| Subject Sample # 1 | 0.137 | 18:33 |
| Breath Volume      | 1.187 | 18:33 |
| Air Blank          | 0.000 | 18:34 |
| Air Blank          | 0.000 | 18:36 |
| Subject Sample # 2 | 0.137 | 18:36 |
| Breath Volume      | 1.527 | 18:36 |
| Air Blank          | 0.000 | 18:37 |
| Control Test       | 0.080 | 18:37 |
| Air Blank          | 0.000 | 18:38 |
| Diagnostics Check  | OK    | 18:38 |

---

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DAVIS,P  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 08/09/2024 02:38:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.081 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Subject Sample # 1 | 0.108 | 02:32 |
| Breath Volume      | 1.816 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 2 | 0.099 | 02:36 |
| Breath Volume      | 1.570 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Control Test       | 0.079 | 02:37 |
| Air Blank          | 0.000 | 02:37 |
| Diagnostics Check  | OK    | 02:38 |

---

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WHITTINGTON

DATE/TIME OF TEST: 08/15/2024 03:58:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.081 | 03:52 |
| Air Blank          | 0.000 | 03:53 |
| Subject Sample # 1 | 0.000 | 03:54 |
| Breath Volume      | 2.160 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 2 | 0.000 | 03:57 |
| Breath Volume      | 2.089 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Control Test       | 0.082 | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Diagnostics Check  | OK    | 03:58 |

---

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 08/18/2024 01:41:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:36 |                      |
| Air Blank         | 0.000 | 01:36 |                      |
| Control Test      | 0.081 | 01:36 |                      |
| Air Blank         | 0.000 | 01:37 |                      |
| Subject Sample    | # 1   | 01:40 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:40 |                      |
| Air Blank         | 0.000 | 01:40 |                      |
| Control Test      | 0.082 | 01:41 |                      |
| Air Blank         | 0.000 | 01:41 |                      |
| Diagnostics Check | OK    | 01:41 |                      |

---

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHEY,A,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MORLEY,R  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 07:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 08/18/2024 08:52:27

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 08:48 |                      |
| Air Blank         | 0.000 | 08:49 |                      |
| Control Test      | 0.082 | 08:49 |                      |
| Air Blank         | 0.000 | 08:50 |                      |
| Subject Sample    | # 1   | 08:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 08:51 |                      |
| Air Blank         | 0.000 | 08:51 |                      |
| Control Test      | 0.080 | 08:51 |                      |
| Air Blank         | 0.000 | 08:52 |                      |
| Diagnostics Check | OK    | 08:52 |                      |

---

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GLADDEN,A,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ROBINSON JR,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 17:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: DOMINGO MARTIN

DATE/TIME OF TEST: 08/18/2024 18:50:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:44 |
| Air Blank          | 0.000 | 18:44 |
| Control Test       | 0.081 | 18:44 |
| Air Blank          | 0.000 | 18:45 |
| Subject Sample # 1 | 0.287 | 18:46 |
| Breath Volume      | 1.207 | 18:46 |
| Air Blank          | 0.000 | 18:46 |
| Air Blank          | 0.000 | 18:48 |
| Subject Sample # 2 | 0.276 | 18:49 |
| Breath Volume      | 2.023 | 18:49 |
| Air Blank          | 0.000 | 18:49 |
| Control Test       | 0.079 | 18:50 |
| Air Blank          | 0.000 | 18:50 |
| Diagnostics Check  | OK    | 18:50 |

---

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/18/2024 23:03:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:00 |                      |
| Air Blank         | 0.000 | 23:00 |                      |
| Control Test      | 0.081 | 23:00 |                      |
| Air Blank         | 0.000 | 23:01 |                      |
| Subject Sample    | # 1   | 23:01 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:01 |                      |
| Air Blank         | 0.000 | 23:02 |                      |
| Control Test      | 0.081 | 23:02 |                      |
| Air Blank         | 0.000 | 23:03 |                      |
| Diagnostics Check | OK    | 23:03 |                      |

---

Data Download Date/Time: 08/22/2024 13:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,J,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CHILIKI,M  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15924080A2  
DRY GAS EXPIRY DATE: 07/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: JAKKIDI

DATE/TIME OF TEST: 08/19/2024 04:44:26

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:41 |                      |
| Air Blank         | 0.000 | 04:41 |                      |
| Control Test      | 0.082 | 04:42 |                      |
| Air Blank         | 0.000 | 04:42 |                      |
| Subject Sample    | # 1   | 04:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:43 |                      |
| Air Blank         | 0.000 | 04:43 |                      |
| Control Test      | 0.081 | 04:43 |                      |
| Air Blank         | 0.000 | 04:44 |                      |
| Diagnostics Check | OK    | 04:44 |                      |

Data Download Date/Time: 08/22/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTON,GARY,W  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: JONES,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/01/2024 04:17:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:11 |
| Air Blank          | 0.000 | 04:11 |
| Control Test       | 0.079 | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Subject Sample # 1 | 0.131 | 04:13 |
| Breath Volume      | 1.585 | 04:13 |
| Air Blank          | 0.000 | 04:13 |
| Air Blank          | 0.000 | 04:15 |
| Subject Sample # 2 | 0.126 | 04:16 |
| Breath Volume      | 3.093 | 04:16 |
| Air Blank          | 0.000 | 04:16 |
| Control Test       | 0.079 | 04:16 |
| Air Blank          | 0.000 | 04:17 |
| Diagnostics Check  | OK    | 04:17 |

---

Data Download Date/Time: 08/25/2024 10:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: SNOW,C  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A1  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 08/02/2024 18:23:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:15 |
| Air Blank          | 0.000 | 18:16 |
| Control Test       | 0.079 | 18:16 |
| Air Blank          | 0.000 | 18:17 |
| Subject Sample # 1 | 0.235 | 18:17 |
| Breath Volume      | 1.968 | 18:17 |
| Air Blank          | 0.000 | 18:18 |
| Air Blank          | 0.000 | 18:20 |
| Subject Sample # 2 | 0.221 | 18:21 |
| Breath Volume      | 2.484 | 18:21 |
| Air Blank          | 0.000 | 18:22 |
| Control Test       | 0.079 | 18:22 |
| Air Blank          | 0.000 | 18:23 |
| Diagnostics Check  | OK    | 18:23 |

---

Data Download Date/Time: 08/25/2024 10:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,M,T  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 08/03/2024 01:49:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.079 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 1 | 0.105 | 01:44 |
| Breath Volume      | 4.058 | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Air Blank          | 0.000 | 01:47 |
| Subject Sample # 2 | 0.101 | 01:47 |
| Breath Volume      | 3.550 | 01:47 |
| Air Blank          | 0.000 | 01:48 |
| Control Test       | 0.079 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Diagnostics Check  | OK    | 01:49 |

---

Data Download Date/Time: 08/25/2024 10:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A1  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 08/03/2024 04:28:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:20 |                      |
| Air Blank         | 0.000 | 04:20 |                      |
| Control Test      | 0.080 | 04:20 |                      |
| Air Blank         | 0.000 | 04:21 |                      |
| Subject Sample    | # 1   | 04:24 | No Sample Provided   |
| Breath Volume     | 0.000 | 04:24 |                      |
| Air Blank         | 0.000 | 04:25 |                      |
| Air Blank         | 0.000 | 04:27 |                      |
| Subject Sample    | # 2   | 04:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:27 |                      |
| Air Blank         | 0.000 | 04:27 |                      |
| Control Test      | 0.079 | 04:28 |                      |
| Air Blank         | 0.000 | 04:28 |                      |
| Diagnostics Check | OK    | 04:28 |                      |

---

Data Download Date/Time: 08/25/2024 10:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,M,T  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: WILLOUGHBY

DATE/TIME OF TEST: 08/05/2024 00:58:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:51 |
| Air Blank          | 0.000 | 00:52 |
| Control Test       | 0.079 | 00:52 |
| Air Blank          | 0.000 | 00:52 |
| Subject Sample # 1 | 0.000 | 00:53 |
| Breath Volume      | 2.410 | 00:53 |
| Air Blank          | 0.000 | 00:54 |
| Air Blank          | 0.000 | 00:56 |
| Subject Sample # 2 | 0.000 | 00:57 |
| Breath Volume      | 2.050 | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Control Test       | 0.080 | 00:58 |
| Air Blank          | 0.000 | 00:58 |
| Diagnostics Check  | OK    | 00:58 |

---

Data Download Date/Time: 08/25/2024 10:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GELSOMINO,M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: MCCARY

DATE/TIME OF TEST: 08/06/2024 22:14:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:06 |                      |
| Air Blank         | 0.000 | 22:06 |                      |
| Control Test      | 0.079 | 22:07 |                      |
| Air Blank         | 0.000 | 22:07 |                      |
| Subject Sample    | # 1   | 22:10 | No Sample Provided   |
| Breath Volume     | 0.000 | 22:10 |                      |
| Air Blank         | 0.000 | 22:11 |                      |
| Air Blank         | 0.000 | 22:13 |                      |
| Subject Sample    | # 2   | 22:13 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:13 |                      |
| Air Blank         | 0.000 | 22:13 |                      |
| Control Test      | 0.079 | 22:14 |                      |
| Air Blank         | 0.000 | 22:14 |                      |
| Diagnostics Check | OK    | 22:14 |                      |

---

Data Download Date/Time: 08/25/2024 10:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A1  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: FORSYTH

DATE/TIME OF TEST: 08/08/2024 00:07:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:00 |
| Air Blank          | 0.000 | 00:00 |
| Control Test       | 0.079 | 00:00 |
| Air Blank          | 0.000 | 00:01 |
| Subject Sample # 1 | 0.098 | 00:02 |
| Breath Volume      | 2.132 | 00:02 |
| Air Blank          | 0.000 | 00:02 |
| Air Blank          | 0.000 | 00:04 |
| Subject Sample # 2 | 0.096 | 00:05 |
| Breath Volume      | 2.386 | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Control Test       | 0.079 | 00:06 |
| Air Blank          | 0.000 | 00:07 |
| Diagnostics Check  | OK    | 00:07 |

---

Data Download Date/Time: 08/25/2024 10:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: MORANDO,B  
ARRESTING OFFICER AGENCY: Green Cove Springs  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A1  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HERSHKOPF

DATE/TIME OF TEST: 08/09/2024 01:28:50

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:20 |                      |
| Air Blank         | 0.000 | 01:20 |                      |
| Control Test      | 0.080 | 01:21 |                      |
| Air Blank         | 0.000 | 01:21 |                      |
| Subject Sample    | # 1   | 01:24 | No Sample Provided   |
| Breath Volume     | 0.000 | 01:24 |                      |
| Air Blank         | 0.000 | 01:25 |                      |
| Air Blank         | 0.000 | 01:27 |                      |
| Subject Sample    | # 2   | 01:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:27 |                      |
| Air Blank         | 0.000 | 01:27 |                      |
| Control Test      | 0.080 | 01:28 |                      |
| Air Blank         | 0.000 | 01:28 |                      |
| Diagnostics Check | OK    | 01:28 |                      |

Data Download Date/Time: 08/25/2024 10:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTON,GARY,W  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GELSOMINO,M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HILLIKER

DATE/TIME OF TEST: 08/10/2024 02:54:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:48 |
| Air Blank          | 0.000 | 02:48 |
| Control Test       | 0.079 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 1 | 0.159 | 02:49 |
| Breath Volume      | 1.210 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Air Blank          | 0.000 | 02:52 |
| Subject Sample # 2 | 0.161 | 02:53 |
| Breath Volume      | 2.378 | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Control Test       | 0.080 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Diagnostics Check  | OK    | 02:54 |

---

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTON,GARY,W  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BAUMGARTNER,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ESPINOZA LEMUS

DATE/TIME OF TEST: 08/10/2024 04:27:26

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:18 |                      |
| Air Blank         | 0.000 | 04:19 |                      |
| Control Test      | 0.079 | 04:19 |                      |
| Air Blank         | 0.000 | 04:19 |                      |
| Subject Sample    | # 1   | 04:22 | No Sample Provided   |
| Breath Volume     | 0.000 | 04:22 |                      |
| Air Blank         | 0.000 | 04:23 |                      |
| Air Blank         | 0.000 | 04:25 |                      |
| Subject Sample    | # 2   | 04:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:26 |                      |
| Air Blank         | 0.000 | 04:26 |                      |
| Control Test      | 0.080 | 04:26 |                      |
| Air Blank         | 0.000 | 04:27 |                      |
| Diagnostics Check | OK    | 04:27 |                      |

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTON,GARY,W  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GELSOMINO,M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: REVELO

DATE/TIME OF TEST: 08/11/2024 02:25:48

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:16 |                      |
| Air Blank         | 0.000 | 02:17 |                      |
| Control Test      | 0.079 | 02:17 |                      |
| Air Blank         | 0.000 | 02:18 |                      |
| Subject Sample    | # 1   | 02:21 | No Sample Provided   |
| Breath Volume     | 0.000 | 02:21 |                      |
| Air Blank         | 0.000 | 02:21 |                      |
| Air Blank         | 0.000 | 02:23 |                      |
| Subject Sample    | # 2   | 02:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:24 |                      |
| Air Blank         | 0.000 | 02:24 |                      |
| Control Test      | 0.080 | 02:25 |                      |
| Air Blank         | 0.000 | 02:25 |                      |
| Diagnostics Check | OK    | 02:25 |                      |

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTON,GARY,W  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GELSOMINO,M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: CUELLAR

DATE/TIME OF TEST: 08/11/2024 22:20:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:12 |
| Air Blank          | 0.000 | 22:13 |
| Control Test       | 0.079 | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Subject Sample # 1 | 0.199 | 22:16 |
| Breath Volume      | 1.195 | 22:16 |
| Air Blank          | 0.000 | 22:16 |
| Air Blank          | 0.000 | 22:18 |
| Subject Sample # 2 | 0.215 | 22:19 |
| Breath Volume      | 1.781 | 22:19 |
| Air Blank          | 0.000 | 22:19 |
| Control Test       | 0.078 | 22:20 |
| Air Blank          | 0.000 | 22:20 |
| Diagnostics Check  | OK    | 22:20 |

---

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A1  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: KAMISH

DATE/TIME OF TEST: 08/13/2024 01:14:07

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:05 |                      |
| Air Blank         | 0.000 | 01:06 |                      |
| Control Test      | 0.080 | 01:06 |                      |
| Air Blank         | 0.000 | 01:06 |                      |
| Subject Sample    | # 1   | 01:10 | No Sample Provided   |
| Breath Volume     | 0.000 | 01:10 |                      |
| Air Blank         | 0.000 | 01:10 |                      |
| Air Blank         | 0.000 | 01:12 |                      |
| Subject Sample    | # 2   | 01:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:12 |                      |
| Air Blank         | 0.000 | 01:13 |                      |
| Control Test      | 0.080 | 01:13 |                      |
| Air Blank         | 0.000 | 01:14 |                      |
| Diagnostics Check | OK    | 01:14 |                      |

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIDDERS,R,M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: MADRIGAL,K  
ARRESTING OFFICER AGENCY: Green Cove Springs  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A1  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 14:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 08/14/2024 14:59:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:52 |
| Air Blank          | 0.000 | 14:52 |
| Control Test       | 0.080 | 14:53 |
| Air Blank          | 0.000 | 14:53 |
| Subject Sample # 1 | 0.000 | 14:54 |
| Breath Volume      | 1.976 | 14:54 |
| Air Blank          | 0.000 | 14:54 |
| Air Blank          | 0.000 | 14:56 |
| Subject Sample # 2 | 0.000 | 14:57 |
| Breath Volume      | 2.453 | 14:57 |
| Air Blank          | 0.000 | 14:58 |
| Control Test       | 0.080 | 14:58 |
| Air Blank          | 0.000 | 14:58 |
| Diagnostics Check  | OK    | 14:59 |

---

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTON,GARY,W  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: JONES,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 08/15/2024 01:44:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Control Test       | 0.080 | 01:38 |
| Air Blank          | 0.000 | 01:39 |
| Subject Sample # 1 | 0.195 | 01:39 |
| Breath Volume      | 3.000 | 01:39 |
| Air Blank          | 0.000 | 01:40 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 2 | 0.190 | 01:43 |
| Breath Volume      | 3.453 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Control Test       | 0.079 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Diagnostics Check  | OK    | 01:44 |

---

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: JONES,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MCLAUGHIN

DATE/TIME OF TEST: 08/16/2024 02:25:04

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:16 |                      |
| Air Blank         | 0.000 | 02:16 |                      |
| Control Test      | 0.080 | 02:16 |                      |
| Air Blank         | 0.000 | 02:17 |                      |
| Subject Sample    | # 1   | 02:20 | No Sample Provided   |
| Breath Volume     | 0.000 | 02:20 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Air Blank         | 0.000 | 02:22 |                      |
| Subject Sample    | # 2   | 02:23 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:23 |                      |
| Air Blank         | 0.000 | 02:24 |                      |
| Control Test      | 0.079 | 02:24 |                      |
| Air Blank         | 0.000 | 02:24 |                      |
| Diagnostics Check | OK    | 02:25 |                      |

---

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A1  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 08/18/2024 04:28:40

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:20 |                      |
| Air Blank         | 0.000 | 04:20 |                      |
| Control Test      | 0.079 | 04:20 |                      |
| Air Blank         | 0.000 | 04:21 |                      |
| Subject Sample    | # 1   | 04:24 | No Sample Provided   |
| Breath Volume     | 0.000 | 04:24 |                      |
| Air Blank         | 0.000 | 04:25 |                      |
| Air Blank         | 0.000 | 04:27 |                      |
| Subject Sample    | # 2   | 04:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:27 |                      |
| Air Blank         | 0.000 | 04:27 |                      |
| Control Test      | 0.079 | 04:28 |                      |
| Air Blank         | 0.000 | 04:28 |                      |
| Diagnostics Check | OK    | 04:28 |                      |

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTON,GARY,W  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: JONES,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SKAFF

DATE/TIME OF TEST: 08/24/2024 01:23:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:16 |
| Air Blank          | 0.000 | 01:16 |
| Control Test       | 0.080 | 01:17 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 1 | 0.155 | 01:18 |
| Breath Volume      | 2.714 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Subject Sample # 2 | 0.162 | 01:21 |
| Breath Volume      | 3.222 | 01:21 |
| Air Blank          | 0.000 | 01:22 |
| Control Test       | 0.079 | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Diagnostics Check  | OK    | 01:23 |

---

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: JONES,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: WRAY

DATE/TIME OF TEST: 08/24/2024 01:34:24

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:26 |                      |
| Air Blank         | 0.000 | 01:26 |                      |
| Control Test      | 0.079 | 01:26 |                      |
| Air Blank         | 0.000 | 01:27 |                      |
| Subject Sample    | # 1   | 01:30 | No Sample Provided   |
| Breath Volume     | 0.000 | 01:30 |                      |
| Air Blank         | 0.000 | 01:30 |                      |
| Air Blank         | 0.000 | 01:32 |                      |
| Subject Sample    | # 2   | 01:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:33 |                      |
| Air Blank         | 0.000 | 01:33 |                      |
| Control Test      | 0.080 | 01:33 |                      |
| Air Blank         | 0.000 | 01:34 |                      |
| Diagnostics Check | OK    | 01:34 |                      |

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTON,GARY,W  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GELSOMINO,M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 08/24/2024 03:09:59

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:01 |                      |
| Air Blank         | 0.000 | 03:01 |                      |
| Control Test      | 0.080 | 03:02 |                      |
| Air Blank         | 0.000 | 03:02 |                      |
| Subject Sample    | # 1   | 03:05 | No Sample Provided   |
| Breath Volume     | 0.000 | 03:05 |                      |
| Air Blank         | 0.000 | 03:06 |                      |
| Air Blank         | 0.000 | 03:08 |                      |
| Subject Sample    | # 2   | 03:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:08 |                      |
| Air Blank         | 0.000 | 03:09 |                      |
| Control Test      | 0.080 | 03:09 |                      |
| Air Blank         | 0.000 | 03:09 |                      |
| Diagnostics Check | OK    | 03:09 |                      |

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: JONES,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 08/25/2024 03:24:30

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:13 |                    |
| Air Blank         | 0.000 | 03:13 |                    |
| Control Test      | 0.079 | 03:13 |                    |
| Air Blank         | 0.000 | 03:14 |                    |
| Subject Sample    | # 1   | 03:17 | No Sample Provided |
| Breath Volume     | 0.000 | 03:17 |                    |
| Air Blank         | 0.000 | 03:17 |                    |
| Air Blank         | 0.000 | 03:19 |                    |
| Subject Sample    | # 2   | 03:23 | Volume Not Met     |
| Breath Volume     | 0.156 | 03:23 |                    |
| Air Blank         | 0.000 | 03:23 |                    |
| Control Test      | 0.079 | 03:23 |                    |
| Air Blank         | 0.000 | 03:24 |                    |
| Diagnostics Check | OK    | 03:24 |                    |

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GELSOMINO,M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/25/2024 04:16:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.080 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 1 | 0.160 | 04:12 |
| Breath Volume      | 2.992 | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Subject Sample # 2 | 0.154 | 04:15 |
| Breath Volume      | 2.949 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Control Test       | 0.079 | 04:16 |
| Air Blank          | 0.000 | 04:16 |
| Diagnostics Check  | OK    | 04:16 |

---

Data Download Date/Time: 08/25/2024 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTON,GARY,W  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: JONES,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 08/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: EDENFIELD

DATE/TIME OF TEST: 08/26/2024 03:18:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:11 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.080 | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Subject Sample # 1 | 0.149 | 03:13 |
| Breath Volume      | 1.460 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Air Blank          | 0.000 | 03:15 |
| Subject Sample # 2 | 0.154 | 03:17 |
| Breath Volume      | 1.414 | 03:17 |
| Air Blank          | 0.000 | 03:17 |
| Control Test       | 0.080 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Diagnostics Check  | OK    | 03:18 |

---

Data Download Date/Time: 09/17/2024 11:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,M,T  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BUI,L  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 08/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 08/27/2024 20:52:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:44 |                      |
| Air Blank         | 0.000 | 20:44 |                      |
| Control Test      | 0.079 | 20:45 |                      |
| Air Blank         | 0.000 | 20:45 |                      |
| Subject Sample    | # 1   | 20:48 | No Sample Provided   |
| Breath Volume     | 0.000 | 20:48 |                      |
| Air Blank         | 0.000 | 20:49 |                      |
| Air Blank         | 0.000 | 20:51 |                      |
| Subject Sample    | # 2   | 20:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:51 |                      |
| Air Blank         | 0.000 | 20:52 |                      |
| Control Test      | 0.080 | 20:52 |                      |
| Air Blank         | 0.000 | 20:52 |                      |
| Diagnostics Check | OK    | 20:52 |                      |

---

Data Download Date/Time: 09/17/2024 11:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: JONES,N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 08/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07622080A2  
DRY GAS EXPIRY DATE: 05/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 08/30/2024 04:04:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.079 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 1 | 0.154 | 04:00 |
| Breath Volume      | 3.328 | 04:00 |
| Air Blank          | 0.000 | 04:00 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 2 | 0.152 | 04:03 |
| Breath Volume      | 3.984 | 04:03 |
| Air Blank          | 0.000 | 04:04 |
| Control Test       | 0.079 | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Diagnostics Check  | OK    | 04:04 |

---

Data Download Date/Time: 09/17/2024 11:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: KARPINSKI,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 08/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A1  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: FORRESTER

DATE/TIME OF TEST: 08/31/2024 03:02:53

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:54 |                      |
| Air Blank         | 0.000 | 02:54 |                      |
| Control Test      | 0.080 | 02:55 |                      |
| Air Blank         | 0.000 | 02:55 |                      |
| Subject Sample    | # 1   | 02:58 | No Sample Provided   |
| Breath Volume     | 0.000 | 02:58 |                      |
| Air Blank         | 0.000 | 02:59 |                      |
| Air Blank         | 0.000 | 03:01 |                      |
| Subject Sample    | # 2   | 03:01 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:01 |                      |
| Air Blank         | 0.000 | 03:01 |                      |
| Control Test      | 0.080 | 03:02 |                      |
| Air Blank         | 0.000 | 03:02 |                      |
| Diagnostics Check | OK    | 03:02 |                      |

Data Download Date/Time: 09/17/2024 11:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOWAY, GRANT, D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GALLOWAY, GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 08/04/2024 05:00:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:53 |
| Air Blank          | 0.000 | 04:53 |
| Control Test       | 0.079 | 04:53 |
| Air Blank          | 0.000 | 04:54 |
| Subject Sample # 1 | 0.150 | 04:55 |
| Breath Volume      | 1.906 | 04:55 |
| Air Blank          | 0.000 | 04:56 |
| Air Blank          | 0.000 | 04:58 |
| Subject Sample # 2 | 0.157 | 04:58 |
| Breath Volume      | 1.546 | 04:58 |
| Air Blank          | 0.000 | 04:59 |
| Control Test       | 0.079 | 04:59 |
| Air Blank          | 0.000 | 05:00 |
| Diagnostics Check  | OK    | 05:00 |

---

Data Download Date/Time: 09/14/2024 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZIZ, DAVID,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: AZIZ, DAVID  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 08/10/2024 00:32:39

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 00:29 |            |
| Air Blank         | 0.000 | 00:30 |            |
| Control Test      | 0.078 | 00:30 |            |
| Air Blank         | 0.000 | 00:31 |            |
| Subject Sample    | # 1   | 00:31 | RFI Detect |
| Breath Volume     | 0.000 | 00:31 |            |
| Air Blank         | 0.000 | 00:31 |            |
| Control Test      | 0.078 | 00:32 |            |
| Air Blank         | 0.000 | 00:32 |            |
| Diagnostics Check | OK    | 00:32 |            |

---

Data Download Date/Time: 09/14/2024 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZIZ,DAVID,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: AZIZ,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 08/10/2024 00:40:02

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 00:34 |            |
| Air Blank         | 0.000     | 00:34 |            |
| Control Test      | 0.078     | 00:34 |            |
| Air Blank         | 0.000     | 00:35 |            |
| Subject Sample    | # 1 0.130 | 00:35 |            |
| Breath Volume     | 2.531     | 00:35 |            |
| Air Blank         | 0.000     | 00:36 |            |
| Air Blank         | 0.000     | 00:38 |            |
| Subject Sample    | # 2       | 00:38 | RFI Detect |
| Breath Volume     | 0.691     | 00:38 |            |
| Air Blank         | 0.000     | 00:39 |            |
| Control Test      | 0.084     | 00:39 |            |
| Air Blank         | 0.000     | 00:39 |            |
| Diagnostics Check | OK        | 00:40 |            |

---

Data Download Date/Time: 09/14/2024 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZIZ,DAVID,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: AZIZ,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 08/10/2024 00:47:36

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 00:41 |                           |
| Air Blank         | 0.000     | 00:42 |                           |
| Control Test      | 0.078     | 00:42 |                           |
| Air Blank         | 0.000     | 00:42 |                           |
| Subject Sample    | # 1 0.124 | 00:43 |                           |
| Breath Volume     | 1.933     | 00:43 |                           |
| Air Blank         | 0.000     | 00:43 |                           |
| Air Blank         | 0.000     | 00:45 |                           |
| Subject Sample    | # 2 0.119 | 00:46 |                           |
| Breath Volume     | 3.285     | 00:46 |                           |
| Air Blank         | 0.000     | 00:46 |                           |
| Control Test      | 0.071     | 00:47 | Control Outside Tolerance |
| Air Blank         | 0.000     | 00:47 |                           |

---

Data Download Date/Time: 09/14/2024 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZIZ,DAVID,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: AZIZ,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 08/10/2024 00:57:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:51 |
| Air Blank          | 0.000 | 00:52 |
| Control Test       | 0.078 | 00:52 |
| Air Blank          | 0.000 | 00:53 |
| Subject Sample # 1 | 0.122 | 00:53 |
| Breath Volume      | 2.546 | 00:53 |
| Air Blank          | 0.000 | 00:54 |
| Air Blank          | 0.000 | 00:55 |
| Subject Sample # 2 | 0.132 | 00:56 |
| Breath Volume      | 2.652 | 00:56 |
| Air Blank          | 0.000 | 00:56 |
| Control Test       | 0.078 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Diagnostics Check  | OK    | 00:57 |

---

Data Download Date/Time: 09/14/2024 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAKELAND,KATLYN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WAKELAND,KATLYN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: HOSMER

DATE/TIME OF TEST: 08/10/2024 22:47:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Control Test       | 0.079 | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Subject Sample # 1 | 0.185 | 22:42 |
| Breath Volume      | 2.902 | 22:42 |
| Air Blank          | 0.000 | 22:43 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 2 | 0.183 | 22:45 |
| Breath Volume      | 2.128 | 22:45 |
| Air Blank          | 0.000 | 22:46 |
| Control Test       | 0.078 | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Diagnostics Check  | OK    | 22:47 |

---

Data Download Date/Time: 09/14/2024 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,JORDAN,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ROBERTS,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BRINSON

DATE/TIME OF TEST: 08/11/2024 02:02:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.080 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 1 | 0.132 | 01:58 |
| Breath Volume      | 2.605 | 01:58 |
| Air Blank          | 0.000 | 01:58 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 2 | 0.132 | 02:01 |
| Breath Volume      | 1.691 | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.079 | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Diagnostics Check  | OK    | 02:02 |

---

Data Download Date/Time: 09/14/2024 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERLYNE,R,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BERLYNE,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HAFNER

DATE/TIME OF TEST: 08/11/2024 03:12:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.079 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Subject Sample # 1 | 0.047 | 03:07 |
| Breath Volume      | 1.363 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Air Blank          | 0.000 | 03:10 |
| Subject Sample # 2 | 0.052 | 03:10 |
| Breath Volume      | 1.847 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.079 | 03:11 |
| Air Blank          | 0.000 | 03:11 |
| Diagnostics Check  | OK    | 03:12 |

---

Data Download Date/Time: 09/14/2024 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JASPER,ANDREW,M  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: VITO,C  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: PIPPINS

DATE/TIME OF TEST: 08/11/2024 05:28:26

---

|                    |       |       |                    |
|--------------------|-------|-------|--------------------|
| Diagnostics Check  | OK    | 05:19 |                    |
| Air Blank          | 0.000 | 05:20 |                    |
| Control Test       | 0.079 | 05:20 |                    |
| Air Blank          | 0.000 | 05:20 |                    |
| Subject Sample # 1 | 0.172 | 05:23 |                    |
| Breath Volume      | 1.765 | 05:23 |                    |
| Air Blank          | 0.000 | 05:24 |                    |
| Air Blank          | 0.000 | 05:26 |                    |
| Subject Sample # 2 | 0.169 | 05:26 |                    |
| Breath Volume      | 1.886 | 05:26 |                    |
| Air Blank          | 0.000 | 05:27 |                    |
| Control Test       | 0.088 | 05:27 | Interferent Detect |
| Air Blank          | 0.000 | 05:28 |                    |

---

Data Download Date/Time: 09/14/2024 12:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JASPER,ANDREW,M  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: VITO,C  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: PIPPINS

DATE/TIME OF TEST: 08/11/2024 05:56:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:47 |
| Air Blank          | 0.000 | 05:48 |
| Control Test       | 0.079 | 05:48 |
| Air Blank          | 0.000 | 05:48 |
| Subject Sample # 1 | 0.150 | 05:50 |
| Breath Volume      | 1.406 | 05:50 |
| Air Blank          | 0.000 | 05:50 |
| Air Blank          | 0.000 | 05:52 |
| Subject Sample # 2 | 0.148 | 05:55 |
| Breath Volume      | 2.664 | 05:55 |
| Air Blank          | 0.000 | 05:55 |
| Control Test       | 0.079 | 05:56 |
| Air Blank          | 0.000 | 05:56 |
| Diagnostics Check  | OK    | 05:56 |

---

Data Download Date/Time: 09/14/2024 12:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALDONADO, SOLOMON,  
BREATH TEST OPERATOR AGENCY: Zephyrhills PD  
ARRESTING OFFICER: ROMERO, LUIS  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: SYMES

DATE/TIME OF TEST: 08/12/2024 15:41:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:35 |
| Air Blank          | 0.000 | 15:35 |
| Control Test       | 0.080 | 15:35 |
| Air Blank          | 0.000 | 15:36 |
| Subject Sample # 1 | 0.000 | 15:36 |
| Breath Volume      | 3.207 | 15:36 |
| Air Blank          | 0.000 | 15:37 |
| Air Blank          | 0.000 | 15:39 |
| Subject Sample # 2 | 0.000 | 15:39 |
| Breath Volume      | 4.820 | 15:39 |
| Air Blank          | 0.000 | 15:40 |
| Control Test       | 0.080 | 15:40 |
| Air Blank          | 0.000 | 15:41 |
| Diagnostics Check  | OK    | 15:41 |

---

Data Download Date/Time: 09/14/2024 12:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,JORDAN,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ROBERTS,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SINKOVICH

DATE/TIME OF TEST: 08/19/2024 22:20:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Control Test       | 0.079 | 22:14 |
| Air Blank          | 0.000 | 22:15 |
| Subject Sample # 1 | 0.082 | 22:16 |
| Breath Volume      | 2.421 | 22:16 |
| Air Blank          | 0.000 | 22:16 |
| Air Blank          | 0.000 | 22:18 |
| Subject Sample # 2 | 0.086 | 22:19 |
| Breath Volume      | 2.734 | 22:19 |
| Air Blank          | 0.000 | 22:19 |
| Control Test       | 0.079 | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Diagnostics Check  | OK    | 22:20 |

---

Data Download Date/Time: 09/14/2024 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,WILLIAM,M  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: KISTLER,MARK  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: STACK

DATE/TIME OF TEST: 08/19/2024 23:42:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:36 |
| Air Blank          | 0.000 | 23:36 |
| Control Test       | 0.080 | 23:37 |
| Air Blank          | 0.000 | 23:37 |
| Subject Sample # 1 | 0.069 | 23:38 |
| Breath Volume      | 1.816 | 23:38 |
| Air Blank          | 0.000 | 23:38 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 2 | 0.068 | 23:40 |
| Breath Volume      | 1.875 | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Control Test       | 0.080 | 23:41 |
| Air Blank          | 0.000 | 23:42 |
| Diagnostics Check  | OK    | 23:42 |

---

Data Download Date/Time: 09/14/2024 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAKELAND,KATLYN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WAKELAND,KATLYN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: KNOWLES

DATE/TIME OF TEST: 08/20/2024 03:37:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Control Test       | 0.078 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 1 | 0.226 | 03:32 |
| Breath Volume      | 1.996 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Air Blank          | 0.000 | 03:34 |
| Subject Sample # 2 | 0.224 | 03:36 |
| Breath Volume      | 2.964 | 03:36 |
| Air Blank          | 0.000 | 03:36 |
| Control Test       | 0.078 | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Diagnostics Check  | OK    | 03:37 |

---

Data Download Date/Time: 09/14/2024 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JASPER,ANDREW,M  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: SYMONS,E  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: NUSBAUM

DATE/TIME OF TEST: 08/21/2024 03:49:37

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:36 |                |
| Air Blank         | 0.000     | 03:37 |                |
| Control Test      | 0.079     | 03:37 |                |
| Air Blank         | 0.000     | 03:37 |                |
| Subject Sample    | # 1 0.000 | 03:40 | Volume Not Met |
| Breath Volume     | 0.050     | 03:40 |                |
| Air Blank         | 0.000     | 03:41 |                |
| Air Blank         | 0.000     | 03:43 |                |
| Subject Sample    | # 2 0.000 | 03:44 |                |
| Breath Volume     | 1.292     | 03:44 |                |
| Air Blank         | 0.000     | 03:45 |                |
| Air Blank         | 0.000     | 03:47 |                |
| Subject Sample    | # 3 0.000 | 03:48 |                |
| Breath Volume     | 1.359     | 03:48 |                |
| Air Blank         | 0.000     | 03:48 |                |
| Control Test      | 0.079     | 03:49 |                |
| Air Blank         | 0.000     | 03:49 |                |
| Diagnostics Check | OK        | 03:49 |                |

---

Data Download Date/Time: 09/14/2024 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOWAY, GRANT, D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GALLOWAY, GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 08/23/2024 02:50:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:43 |
| Air Blank          | 0.000 | 02:44 |
| Control Test       | 0.079 | 02:44 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 1 | 0.022 | 02:45 |
| Breath Volume      | 2.089 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Air Blank          | 0.000 | 02:48 |
| Subject Sample # 2 | 0.022 | 02:49 |
| Breath Volume      | 2.519 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.079 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Diagnostics Check  | OK    | 02:50 |

---

Data Download Date/Time: 09/14/2024 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAKELAND,KATLYN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WAKELAND,KATLYN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CALAZANS MECHELLI

DATE/TIME OF TEST: 08/25/2024 03:55:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:48 |
| Air Blank          | 0.000 | 03:48 |
| Control Test       | 0.078 | 03:49 |
| Air Blank          | 0.000 | 03:49 |
| Subject Sample # 1 | 0.147 | 03:50 |
| Breath Volume      | 2.628 | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Subject Sample # 2 | 0.149 | 03:53 |
| Breath Volume      | 2.523 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Control Test       | 0.078 | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Diagnostics Check  | OK    | 03:55 |

---

Data Download Date/Time: 09/14/2024 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOWAY, GRANT, D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GALLOWAY, GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 08/28/2024 22:35:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:29 |
| Air Blank          | 0.000 | 22:29 |
| Control Test       | 0.080 | 22:29 |
| Air Blank          | 0.000 | 22:30 |
| Subject Sample # 1 | 0.149 | 22:30 |
| Breath Volume      | 2.085 | 22:30 |
| Air Blank          | 0.000 | 22:31 |
| Air Blank          | 0.000 | 22:33 |
| Subject Sample # 2 | 0.157 | 22:34 |
| Breath Volume      | 2.332 | 22:34 |
| Air Blank          | 0.000 | 22:34 |
| Control Test       | 0.080 | 22:34 |
| Air Blank          | 0.000 | 22:35 |
| Diagnostics Check  | OK    | 22:35 |

---

Data Download Date/Time: 09/14/2024 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOWAY, GRANT, D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GALLOWAY, GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 08/29/2024 01:52:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:45 |
| Air Blank          | 0.000 | 01:45 |
| Control Test       | 0.079 | 01:46 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 1 | 0.000 | 01:47 |
| Breath Volume      | 1.242 | 01:47 |
| Air Blank          | 0.000 | 01:48 |
| Air Blank          | 0.000 | 01:50 |
| Subject Sample # 2 | 0.000 | 01:50 |
| Breath Volume      | 1.628 | 01:50 |
| Air Blank          | 0.000 | 01:51 |
| Control Test       | 0.080 | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Diagnostics Check  | OK    | 01:52 |

---

Data Download Date/Time: 09/14/2024 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WIBERLY,RYAN,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WIBERLY,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LEMARIE

DATE/TIME OF TEST: 08/29/2024 20:45:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:39 |
| Air Blank          | 0.000 | 20:39 |
| Control Test       | 0.080 | 20:40 |
| Air Blank          | 0.000 | 20:40 |
| Subject Sample # 1 | 0.000 | 20:41 |
| Breath Volume      | 2.097 | 20:41 |
| Air Blank          | 0.000 | 20:41 |
| Air Blank          | 0.000 | 20:43 |
| Subject Sample # 2 | 0.000 | 20:44 |
| Breath Volume      | 1.750 | 20:44 |
| Air Blank          | 0.000 | 20:44 |
| Control Test       | 0.080 | 20:44 |
| Air Blank          | 0.000 | 20:45 |
| Diagnostics Check  | OK    | 20:45 |

---

Data Download Date/Time: 09/14/2024 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,B,T  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: THORNTON,B  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 08/02/2024 23:10:56

|                   |           |       |                 |
|-------------------|-----------|-------|-----------------|
| Diagnostics Check | OK        | 22:58 |                 |
| Air Blank         | 0.000     | 22:58 |                 |
| Control Test      | 0.080     | 22:59 |                 |
| Air Blank         | 0.000     | 22:59 |                 |
| Subject Sample    | # 1 0.073 | 23:02 | Slope Not Level |
| Breath Volume     | 1.210     | 23:02 |                 |
| Air Blank         | 0.000     | 23:03 |                 |
| Air Blank         | 0.000     | 23:05 |                 |
| Subject Sample    | # 2 0.075 | 23:05 |                 |
| Breath Volume     | 1.585     | 23:05 |                 |
| Air Blank         | 0.000     | 23:06 |                 |
| Air Blank         | 0.000     | 23:08 |                 |
| Subject Sample    | # 3 0.085 | 23:09 |                 |
| Breath Volume     | 1.523     | 23:09 |                 |
| Air Blank         | 0.000     | 23:10 |                 |
| Control Test      | 0.079     | 23:10 |                 |
| Air Blank         | 0.000     | 23:10 |                 |
| Diagnostics Check | OK        | 23:10 |                 |

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNGO, WILLIAM, F  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: FLIED, L  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 06:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: TOMBERLIN

DATE/TIME OF TEST: 08/04/2024 07:42:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:35 |
| Air Blank          | 0.000 | 07:36 |
| Control Test       | 0.080 | 07:36 |
| Air Blank          | 0.000 | 07:37 |
| Subject Sample # 1 | 0.000 | 07:38 |
| Breath Volume      | 2.441 | 07:38 |
| Air Blank          | 0.000 | 07:38 |
| Air Blank          | 0.000 | 07:40 |
| Subject Sample # 2 | 0.000 | 07:41 |
| Breath Volume      | 3.988 | 07:41 |
| Air Blank          | 0.000 | 07:41 |
| Control Test       | 0.079 | 07:42 |
| Air Blank          | 0.000 | 07:42 |
| Diagnostics Check  | OK    | 07:42 |

---

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTES,JOSEAS,I  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: VELETANLIC,S  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SPIVEY

DATE/TIME OF TEST: 08/12/2024 01:19:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:12 |
| Air Blank          | 0.000 | 01:12 |
| Control Test       | 0.079 | 01:13 |
| Air Blank          | 0.000 | 01:13 |
| Subject Sample # 1 | 0.163 | 01:14 |
| Breath Volume      | 1.339 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Subject Sample # 2 | 0.173 | 01:17 |
| Breath Volume      | 1.390 | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.080 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Diagnostics Check  | OK    | 01:19 |

---

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,B,A  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUTCHERSON,D  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 08/14/2024 03:26:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.080 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 1 | 0.181 | 03:22 |
| Breath Volume      | 2.222 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 2 | 0.176 | 03:25 |
| Breath Volume      | 1.789 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.080 | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Diagnostics Check  | OK    | 03:26 |

---

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTES,JOSEAS,I  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: EAST,Z  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MARRERO-POMALES

DATE/TIME OF TEST: 08/15/2024 01:18:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:15 |                      |
| Air Blank         | 0.000 | 01:15 |                      |
| Control Test      | 0.080 | 01:16 |                      |
| Air Blank         | 0.000 | 01:16 |                      |
| Subject Sample    | # 1   | 01:16 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:16 |                      |
| Air Blank         | 0.000 | 01:17 |                      |
| Control Test      | 0.080 | 01:17 |                      |
| Air Blank         | 0.000 | 01:18 |                      |
| Diagnostics Check | OK    | 01:18 |                      |

---

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,B,T  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HALL,J  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MILTON

DATE/TIME OF TEST: 08/18/2024 03:33:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.081 | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 1 | 0.070 | 03:26 |
| Breath Volume      | 1.105 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 2 | 0.075 | 03:31 |
| Breath Volume      | 1.296 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.080 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Diagnostics Check  | OK    | 03:33 |

---

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNEAD,C,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: PERALTA,C  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SKELTON

DATE/TIME OF TEST: 08/19/2024 05:08:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:59 |
| Air Blank          | 0.000 | 05:00 |
| Control Test       | 0.081 | 05:00 |
| Air Blank          | 0.000 | 05:01 |
| Subject Sample # 1 | 0.000 | 05:03 |
| Breath Volume      | 1.277 | 05:03 |
| Air Blank          | 0.000 | 05:03 |
| Air Blank          | 0.000 | 05:05 |
| Subject Sample # 2 | 0.000 | 05:07 |
| Breath Volume      | 1.550 | 05:07 |
| Air Blank          | 0.000 | 05:07 |
| Control Test       | 0.080 | 05:08 |
| Air Blank          | 0.000 | 05:08 |
| Diagnostics Check  | OK    | 05:08 |

---

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAL,JACOB,N  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: PENSA,MICHAEL  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 11:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: CARRIERO

DATE/TIME OF TEST: 08/20/2024 13:07:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:00 |
| Air Blank          | 0.000 | 13:01 |
| Control Test       | 0.081 | 13:01 |
| Air Blank          | 0.000 | 13:02 |
| Subject Sample # 1 | 0.000 | 13:03 |
| Breath Volume      | 2.761 | 13:03 |
| Air Blank          | 0.000 | 13:03 |
| Air Blank          | 0.000 | 13:05 |
| Subject Sample # 2 | 0.000 | 13:06 |
| Breath Volume      | 1.417 | 13:06 |
| Air Blank          | 0.000 | 13:06 |
| Control Test       | 0.080 | 13:06 |
| Air Blank          | 0.000 | 13:07 |
| Diagnostics Check  | OK    | 13:07 |

---

Data Download Date/Time: 08/28/2024 16:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,JAKOB,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: CREWS,GAVIN  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 08/24/2024 03:24:48

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:09 |                |
| Air Blank         | 0.000     | 03:09 |                |
| Control Test      | 0.079     | 03:10 |                |
| Air Blank         | 0.000     | 03:10 |                |
| Subject Sample    | # 1 0.233 | 03:12 |                |
| Breath Volume     | 1.183     | 03:12 |                |
| Air Blank         | 0.000     | 03:12 |                |
| Air Blank         | 0.000     | 03:14 |                |
| Subject Sample    | # 2 0.227 | 03:17 | Volume Not Met |
| Breath Volume     | 0.539     | 03:17 |                |
| Air Blank         | 0.000     | 03:18 |                |
| Air Blank         | 0.000     | 03:20 |                |
| Subject Sample    | # 3 0.238 | 03:23 | Volume Not Met |
| Breath Volume     | 0.441     | 03:23 |                |
| Air Blank         | 0.000     | 03:23 |                |
| Control Test      | 0.079     | 03:24 |                |
| Air Blank         | 0.000     | 03:24 |                |
| Diagnostics Check | OK        | 03:24 |                |

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,JAKOB,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: CUARTERO,L  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 12:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: WALDEN

DATE/TIME OF TEST: 08/25/2024 01:02:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:54 |
| Air Blank          | 0.000 | 00:54 |
| Control Test       | 0.080 | 00:54 |
| Air Blank          | 0.000 | 00:55 |
| Subject Sample # 1 | 0.099 | 00:56 |
| Breath Volume      | 2.968 | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Air Blank          | 0.000 | 00:59 |
| Subject Sample # 2 | 0.105 | 01:00 |
| Breath Volume      | 1.886 | 01:00 |
| Air Blank          | 0.000 | 01:01 |
| Control Test       | 0.079 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Diagnostics Check  | OK    | 01:02 |

---

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,JAKOB,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: BARNES,D  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 08/25/2024 03:21:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.080 | 03:13 |
| Air Blank          | 0.000 | 03:13 |
| Subject Sample # 1 | 0.107 | 03:16 |
| Breath Volume      | 1.480 | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Air Blank          | 0.000 | 03:18 |
| Subject Sample # 2 | 0.115 | 03:20 |
| Breath Volume      | 2.000 | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.079 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Diagnostics Check  | OK    | 03:21 |

---

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESTER, ARIELLE, J  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HILLIARD, NATHAN  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 07/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 08/26/2024 18:31:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:25 |
| Air Blank          | 0.000 | 18:25 |
| Control Test       | 0.080 | 18:25 |
| Air Blank          | 0.000 | 18:26 |
| Subject Sample # 1 | 0.243 | 18:26 |
| Breath Volume      | 1.937 | 18:26 |
| Air Blank          | 0.000 | 18:27 |
| Air Blank          | 0.000 | 18:29 |
| Subject Sample # 2 | 0.234 | 18:29 |
| Breath Volume      | 1.824 | 18:29 |
| Air Blank          | 0.000 | 18:30 |
| Control Test       | 0.078 | 18:30 |
| Air Blank          | 0.000 | 18:31 |
| Diagnostics Check  | OK    | 18:31 |

---

Data Download Date/Time: 08/28/2024 16:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,B,T  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUTCHERSON,D  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 08/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: ROWELL

DATE/TIME OF TEST: 08/30/2024 19:19:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:13 |
| Air Blank          | 0.000 | 19:13 |
| Control Test       | 0.079 | 19:14 |
| Air Blank          | 0.000 | 19:14 |
| Subject Sample # 1 | 0.172 | 19:15 |
| Breath Volume      | 2.207 | 19:15 |
| Air Blank          | 0.000 | 19:15 |
| Air Blank          | 0.000 | 19:17 |
| Subject Sample # 2 | 0.176 | 19:18 |
| Breath Volume      | 2.062 | 19:18 |
| Air Blank          | 0.000 | 19:18 |
| Control Test       | 0.080 | 19:19 |
| Air Blank          | 0.000 | 19:19 |
| Diagnostics Check  | OK    | 19:19 |

---

Data Download Date/Time: 09/25/2024 01:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,B,T  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS,W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 08/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: SWIFT

DATE/TIME OF TEST: 08/31/2024 04:28:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:19 |
| Air Blank          | 0.000 | 04:19 |
| Control Test       | 0.080 | 04:20 |
| Air Blank          | 0.000 | 04:20 |
| Subject Sample # 1 | 0.146 | 04:23 |
| Breath Volume      | 1.492 | 04:23 |
| Air Blank          | 0.000 | 04:24 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 2 | 0.141 | 04:27 |
| Breath Volume      | 1.574 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Control Test       | 0.079 | 04:28 |
| Air Blank          | 0.000 | 04:28 |
| Diagnostics Check  | OK    | 04:28 |

---

Data Download Date/Time: 09/25/2024 01:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ESPINDOLA AMBROSIO

DATE/TIME OF TEST: 08/17/2024 00:15:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:09 |
| Air Blank          | 0.000 | 00:09 |
| Control Test       | 0.080 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 1 | 0.140 | 00:10 |
| Breath Volume      | 2.375 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 2 | 0.141 | 00:13 |
| Breath Volume      | 1.925 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Control Test       | 0.080 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Diagnostics Check  | OK    | 00:15 |

---

Data Download Date/Time: 09/10/2024 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: IRVIS

DATE/TIME OF TEST: 08/17/2024 04:26:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:20 |
| Air Blank          | 0.000 | 04:20 |
| Control Test       | 0.081 | 04:21 |
| Air Blank          | 0.000 | 04:21 |
| Subject Sample # 1 | 0.159 | 04:22 |
| Breath Volume      | 1.492 | 04:22 |
| Air Blank          | 0.000 | 04:22 |
| Air Blank          | 0.000 | 04:24 |
| Subject Sample # 2 | 0.160 | 04:25 |
| Breath Volume      | 1.718 | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Control Test       | 0.081 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Diagnostics Check  | OK    | 04:26 |

---

Data Download Date/Time: 09/10/2024 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 08/18/2024 00:52:38

---

|                   |       |       |               |
|-------------------|-------|-------|---------------|
| Diagnostics Check | OK    | 00:49 |               |
| Air Blank         | 0.000 | 00:49 |               |
| Control Test      | 0.081 | 00:50 |               |
| Air Blank         | 0.000 | 00:50 |               |
| Subject Sample    | # 1   | 00:51 | Slope Not Met |
| Breath Volume     | 1.941 | 00:51 |               |
| Air Blank         |       | 00:51 | Purge Fail    |
| Air Blank         |       | 00:52 | Purge Fail    |

---

Data Download Date/Time: 09/10/2024 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 08/18/2024 01:08:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:02 |
| Air Blank          | 0.000 | 01:02 |
| Control Test       | 0.077 | 01:03 |
| Air Blank          | 0.000 | 01:03 |
| Subject Sample # 1 | 0.185 | 01:04 |
| Breath Volume      | 2.511 | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Air Blank          | 0.000 | 01:06 |
| Subject Sample # 2 | 0.189 | 01:06 |
| Breath Volume      | 2.660 | 01:06 |
| Air Blank          | 0.000 | 01:07 |
| Control Test       | 0.079 | 01:07 |
| Air Blank          | 0.000 | 01:08 |
| Diagnostics Check  | OK    | 01:08 |

---

Data Download Date/Time: 09/10/2024 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,4C07  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: PESINA

DATE/TIME OF TEST: 08/18/2024 01:28:43

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:17 |                    |
| Air Blank         | 0.000 | 01:17 |                    |
| Control Test      | 0.079 | 01:18 |                    |
| Air Blank         | 0.000 | 01:18 |                    |
| Subject Sample    | # 1   | 01:21 | No Sample Provided |
| Breath Volume     | 0.000 | 01:21 |                    |
| Air Blank         | 0.000 | 01:22 |                    |
| Air Blank         | 0.000 | 01:24 |                    |
| Subject Sample    | # 2   | 01:27 | No Sample Provided |
| Breath Volume     | 0.000 | 01:27 |                    |
| Air Blank         | 0.000 | 01:27 |                    |
| Control Test      | 0.078 | 01:28 |                    |
| Air Blank         | 0.000 | 01:28 |                    |
| Diagnostics Check | OK    | 01:28 |                    |

Data Download Date/Time: 09/10/2024 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: TURNER,L  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: SANTIAGO NEGRON

DATE/TIME OF TEST: 08/18/2024 02:26:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:19 |
| Air Blank          | 0.000 | 02:20 |
| Control Test       | 0.080 | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Subject Sample # 1 | 0.086 | 02:21 |
| Breath Volume      | 1.105 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Subject Sample # 2 | 0.086 | 02:24 |
| Breath Volume      | 1.441 | 02:24 |
| Air Blank          | 0.000 | 02:25 |
| Control Test       | 0.080 | 02:25 |
| Air Blank          | 0.000 | 02:25 |
| Diagnostics Check  | OK    | 02:26 |

---

Data Download Date/Time: 09/10/2024 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JANIAK,M  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 05:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 08/18/2024 05:54:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:48 |
| Air Blank          | 0.000 | 05:48 |
| Control Test       | 0.081 | 05:49 |
| Air Blank          | 0.000 | 05:49 |
| Subject Sample # 1 | 0.058 | 05:50 |
| Breath Volume      | 1.976 | 05:50 |
| Air Blank          | 0.000 | 05:50 |
| Air Blank          | 0.000 | 05:52 |
| Subject Sample # 2 | 0.055 | 05:53 |
| Breath Volume      | 1.738 | 05:53 |
| Air Blank          | 0.000 | 05:53 |
| Control Test       | 0.081 | 05:53 |
| Air Blank          | 0.000 | 05:54 |
| Diagnostics Check  | OK    | 05:54 |

---

Data Download Date/Time: 09/10/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 08/18/2024 19:23:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:16 |
| Air Blank          | 0.000 | 19:17 |
| Control Test       | 0.080 | 19:17 |
| Air Blank          | 0.000 | 19:18 |
| Subject Sample # 1 | 0.368 | 19:18 |
| Breath Volume      | 1.746 | 19:18 |
| Air Blank          | 0.000 | 19:19 |
| Air Blank          | 0.000 | 19:21 |
| Subject Sample # 2 | 0.365 | 19:21 |
| Breath Volume      | 1.433 | 19:21 |
| Air Blank          | 0.000 | 19:22 |
| Control Test       | 0.081 | 19:22 |
| Air Blank          | 0.000 | 19:22 |
| Diagnostics Check  | OK    | 19:23 |

---

Data Download Date/Time: 09/10/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,REGISTER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: VOGEL

DATE/TIME OF TEST: 08/27/2024 02:09:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Control Test       | 0.081 | 02:04 |
| Air Blank          | 0.000 | 02:04 |
| Subject Sample # 1 | 0.000 | 02:05 |
| Breath Volume      | 2.671 | 02:05 |
| Air Blank          | 0.000 | 02:05 |
| Air Blank          | 0.000 | 02:07 |
| Subject Sample # 2 | 0.000 | 02:08 |
| Breath Volume      | 3.921 | 02:08 |
| Air Blank          | 0.000 | 02:08 |
| Control Test       | 0.081 | 02:09 |
| Air Blank          | 0.000 | 02:09 |
| Diagnostics Check  | OK    | 02:09 |

---

Data Download Date/Time: 09/10/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SYKES

DATE/TIME OF TEST: 08/28/2024 02:23:52

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:12 |                    |
| Air Blank         | 0.000 | 02:12 |                    |
| Control Test      | 0.080 | 02:13 |                    |
| Air Blank         | 0.000 | 02:13 |                    |
| Subject Sample    | # 1   | 02:16 | No Sample Provided |
| Breath Volume     | 0.000 | 02:16 |                    |
| Air Blank         | 0.000 | 02:17 |                    |
| Air Blank         | 0.000 | 02:19 |                    |
| Subject Sample    | # 2   | 02:22 | No Sample Provided |
| Breath Volume     | 0.000 | 02:22 |                    |
| Air Blank         | 0.000 | 02:22 |                    |
| Control Test      | 0.079 | 02:23 |                    |
| Air Blank         | 0.000 | 02:23 |                    |
| Diagnostics Check | OK    | 02:23 |                    |

---

Data Download Date/Time: 09/10/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: ARMENDARIZ CASTILLO

DATE/TIME OF TEST: 08/30/2024 19:29:26

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 19:26 |            |
| Air Blank          | 0.000 | 19:26 |            |
| Control Test       | 0.081 | 19:27 |            |
| Air Blank          | 0.000 | 19:27 |            |
| Subject Sample # 1 | 0.000 | 19:27 |            |
| Breath Volume      | 3.796 | 19:27 |            |
| Air Blank          |       | 19:28 | RFI Detect |
| Air Blank          | 0.000 | 19:28 |            |
| Control Test       | 0.082 | 19:28 |            |
| Air Blank          | 0.000 | 19:29 |            |
| Diagnostics Check  | OK    | 19:29 |            |

---

Data Download Date/Time: 09/10/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: ARMENDARIZ CASTILLO

DATE/TIME OF TEST: 08/30/2024 19:37:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:31 |
| Air Blank          | 0.000 | 19:31 |
| Control Test       | 0.081 | 19:31 |
| Air Blank          | 0.000 | 19:32 |
| Subject Sample # 1 | 0.000 | 19:32 |
| Breath Volume      | 3.046 | 19:32 |
| Air Blank          | 0.000 | 19:33 |
| Air Blank          | 0.000 | 19:35 |
| Subject Sample # 2 | 0.000 | 19:35 |
| Breath Volume      | 3.421 | 19:35 |
| Air Blank          | 0.000 | 19:36 |
| Control Test       | 0.082 | 19:36 |
| Air Blank          | 0.000 | 19:36 |
| Diagnostics Check  | OK    | 19:37 |

---

Data Download Date/Time: 09/10/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CRUZ-RODRIGUEZ

DATE/TIME OF TEST: 08/31/2024 02:40:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:33 |
| Air Blank          | 0.000 | 02:33 |
| Control Test       | 0.081 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 1 | 0.095 | 02:35 |
| Breath Volume      | 1.804 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 2 | 0.095 | 02:38 |
| Breath Volume      | 1.843 | 02:38 |
| Air Blank          | 0.000 | 02:39 |
| Control Test       | 0.082 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Diagnostics Check  | OK    | 02:40 |

---

Data Download Date/Time: 09/10/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A2  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: GONZALEZ MELENDEZ

DATE/TIME OF TEST: 08/31/2024 22:37:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:30 |
| Air Blank          | 0.000 | 22:30 |
| Control Test       | 0.082 | 22:30 |
| Air Blank          | 0.000 | 22:31 |
| Subject Sample # 1 | 0.150 | 22:31 |
| Breath Volume      | 1.117 | 22:31 |
| Air Blank          | 0.000 | 22:32 |
| Air Blank          | 0.000 | 22:34 |
| Subject Sample # 2 | 0.158 | 22:35 |
| Breath Volume      | 1.558 | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Control Test       | 0.081 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Diagnostics Check  | OK    | 22:37 |

---

Data Download Date/Time: 09/10/2024 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,KALEB,R  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: WILSON,KALEB  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BOYETTE

DATE/TIME OF TEST: 08/03/2024 02:22:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:12 |
| Air Blank          | 0.000 | 02:13 |
| Control Test       | 0.079 | 02:13 |
| Air Blank          | 0.000 | 02:13 |
| Subject Sample # 1 | 0.168 | 02:15 |
| Breath Volume      | 1.277 | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Air Blank          | 0.000 | 02:18 |
| Subject Sample # 2 | 0.168 | 02:20 |
| Breath Volume      | 1.531 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Control Test       | 0.078 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Diagnostics Check  | OK    | 02:22 |

---

Data Download Date/Time: 08/26/2024 13:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,KALEB,R  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: WILSON,KALEB  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 08/11/2024 00:46:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Control Test       | 0.078 | 00:41 |
| Air Blank          | 0.000 | 00:41 |
| Subject Sample # 1 | 0.188 | 00:42 |
| Breath Volume      | 1.792 | 00:42 |
| Air Blank          | 0.000 | 00:42 |
| Air Blank          | 0.000 | 00:44 |
| Subject Sample # 2 | 0.185 | 00:45 |
| Breath Volume      | 1.332 | 00:45 |
| Air Blank          | 0.000 | 00:45 |
| Control Test       | 0.077 | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Diagnostics Check  | OK    | 00:46 |

---

Data Download Date/Time: 08/26/2024 13:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,KALEB,R  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: WILSON,KALEB  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SAUM

DATE/TIME OF TEST: 08/11/2024 03:52:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.079 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 1 | 0.184 | 03:47 |
| Breath Volume      | 1.144 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Air Blank          | 0.000 | 03:50 |
| Subject Sample # 2 | 0.180 | 03:50 |
| Breath Volume      | 1.507 | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.079 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Diagnostics Check  | OK    | 03:52 |

---

Data Download Date/Time: 08/26/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,KALEB,R  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: WILSON,KALEB  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 08/17/2024 02:27:19

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 02:26 |            |
| Air Blank         | 0.000 | 02:26 |            |
| Control Test      | 0.079 | 02:26 | RFI Detect |
| Air Blank         | 0.000 | 02:27 |            |

---

Data Download Date/Time: 08/26/2024 13:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,KALEB,R  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: WILSON,KALEB  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 08/17/2024 02:36:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.079 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Subject Sample # 1 | 0.198 | 02:32 |
| Breath Volume      | 2.781 | 02:32 |
| Air Blank          | 0.000 | 02:32 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 2 | 0.196 | 02:35 |
| Breath Volume      | 2.187 | 02:35 |
| Air Blank          | 0.000 | 02:35 |
| Control Test       | 0.078 | 02:36 |
| Air Blank          | 0.000 | 02:36 |
| Diagnostics Check  | OK    | 02:36 |

---

Data Download Date/Time: 08/26/2024 13:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,KALEB,R  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: WILSON,KALEB  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: KILMER

DATE/TIME OF TEST: 08/17/2024 05:11:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:05 |
| Air Blank          | 0.000 | 05:05 |
| Control Test       | 0.079 | 05:05 |
| Air Blank          | 0.000 | 05:06 |
| Subject Sample # 1 | 0.095 | 05:06 |
| Breath Volume      | 1.164 | 05:06 |
| Air Blank          | 0.000 | 05:07 |
| Air Blank          | 0.000 | 05:09 |
| Subject Sample # 2 | 0.095 | 05:10 |
| Breath Volume      | 1.644 | 05:10 |
| Air Blank          | 0.000 | 05:10 |
| Control Test       | 0.079 | 05:11 |
| Air Blank          | 0.000 | 05:11 |
| Diagnostics Check  | OK    | 05:11 |

---

Data Download Date/Time: 08/26/2024 13:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MINCIN,WESLEY,P  
BREATH TEST OPERATOR AGENCY: Gulfport PD  
ARRESTING OFFICER: MINCIN,WESLEY  
ARRESTING OFFICER AGENCY: Gulfport PD  
LAST AGENCY INSPECTION: 08/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: BUI  
ARREST TIME: 22:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: DUNHAM

DATE/TIME OF TEST: 08/30/2024 23:02:21

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:58 |                      |
| Air Blank         | 0.000 | 22:58 |                      |
| Control Test      | 0.080 | 22:59 |                      |
| Air Blank         | 0.000 | 22:59 |                      |
| Subject Sample    | # 1   | 23:00 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:00 |                      |
| Air Blank         | 0.000 | 23:01 |                      |
| Control Test      | 0.081 | 23:01 |                      |
| Air Blank         | 0.000 | 23:02 |                      |
| Diagnostics Check | OK    | 23:02 |                      |

Data Download Date/Time: 09/24/2024 02:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DERIET,RAFAEL,  
BREATH TEST OPERATOR AGENCY: Hollywood PD  
ARRESTING OFFICER: PALACIOS,JONATHAN  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402290677  
DRY GAS EXPIRY DATE: 12/07/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MONDOL

DATE/TIME OF TEST: 08/03/2024 02:58:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Control Test       | 0.077 | 02:52 |
| Air Blank          | 0.000 | 02:53 |
| Subject Sample # 1 | 0.189 | 02:53 |
| Breath Volume      | 2.753 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Air Blank          | 0.000 | 02:56 |
| Subject Sample # 2 | 0.188 | 02:57 |
| Breath Volume      | 2.441 | 02:57 |
| Air Blank          | 0.000 | 02:57 |
| Control Test       | 0.077 | 02:57 |
| Air Blank          | 0.000 | 02:58 |
| Diagnostics Check  | OK    | 02:58 |

---

Data Download Date/Time: 08/07/2024 11:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUGELHEIM,HENRY,S  
BREATH TEST OPERATOR AGENCY: Hollywood PD  
ARRESTING OFFICER: DIAZ,CARLOS  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 08/07/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402290677  
DRY GAS EXPIRY DATE: 12/07/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: REYES BATLLE

DATE/TIME OF TEST: 08/11/2024 03:52:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:45 |
| Air Blank          | 0.000 | 03:45 |
| Control Test       | 0.077 | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Subject Sample # 1 | 0.178 | 03:47 |
| Breath Volume      | 2.941 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Air Blank          | 0.000 | 03:50 |
| Subject Sample # 2 | 0.187 | 03:50 |
| Breath Volume      | 2.582 | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.077 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Diagnostics Check  | OK    | 03:52 |

---

Data Download Date/Time: 09/04/2024 11:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUGELHEIM,HENRY,S  
BREATH TEST OPERATOR AGENCY: Hollywood PD  
ARRESTING OFFICER: CARRASCO,ORLANDO  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 08/07/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402290677  
DRY GAS EXPIRY DATE: 12/07/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: PEREZ GODINEZ

DATE/TIME OF TEST: 08/24/2024 04:57:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:50 |
| Air Blank          | 0.000 | 04:51 |
| Control Test       | 0.077 | 04:51 |
| Air Blank          | 0.000 | 04:51 |
| Subject Sample # 1 | 0.190 | 04:52 |
| Breath Volume      | 3.484 | 04:52 |
| Air Blank          | 0.000 | 04:53 |
| Air Blank          | 0.000 | 04:55 |
| Subject Sample # 2 | 0.184 | 04:55 |
| Breath Volume      | 2.257 | 04:55 |
| Air Blank          | 0.000 | 04:56 |
| Control Test       | 0.077 | 04:56 |
| Air Blank          | 0.000 | 04:57 |
| Diagnostics Check  | OK    | 04:57 |

---

Data Download Date/Time: 09/04/2024 11:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,J,F  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: BROWN,J  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01923080A3  
DRY GAS EXPIRY DATE: 02/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 08/16/2024 21:50:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:44 |
| Air Blank          | 0.000 | 21:45 |
| Control Test       | 0.080 | 21:45 |
| Air Blank          | 0.000 | 21:45 |
| Subject Sample # 1 | 0.114 | 21:46 |
| Breath Volume      | 1.882 | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Air Blank          | 0.000 | 21:48 |
| Subject Sample # 2 | 0.107 | 21:49 |
| Breath Volume      | 2.007 | 21:49 |
| Air Blank          | 0.000 | 21:49 |
| Control Test       | 0.079 | 21:50 |
| Air Blank          | 0.000 | 21:50 |
| Diagnostics Check  | OK    | 21:50 |

---

Data Download Date/Time: 08/29/2024 13:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMER,LEEANN,N  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: INSALACO,M  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01923080A3  
DRY GAS EXPIRY DATE: 02/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: TISSUE

DATE/TIME OF TEST: 08/26/2024 21:40:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:33 |
| Air Blank          | 0.000 | 21:34 |
| Control Test       | 0.079 | 21:34 |
| Air Blank          | 0.000 | 21:34 |
| Subject Sample # 1 | 0.025 | 21:36 |
| Breath Volume      | 2.019 | 21:36 |
| Air Blank          | 0.000 | 21:36 |
| Air Blank          | 0.000 | 21:38 |
| Subject Sample # 2 | 0.023 | 21:39 |
| Breath Volume      | 1.375 | 21:39 |
| Air Blank          | 0.000 | 21:39 |
| Control Test       | 0.079 | 21:40 |
| Air Blank          | 0.000 | 21:40 |
| Diagnostics Check  | OK    | 21:40 |

---

Data Download Date/Time: 08/29/2024 13:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRCHGRABER,BRIAN,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: KIRCHGRABER,BRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 07/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG229803  
DRY GAS EXPIRY DATE: 10/25/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 08/08/2024 23:51:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:44 |
| Air Blank          | 0.000 | 23:45 |
| Control Test       | 0.078 | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Subject Sample # 1 | 0.250 | 23:47 |
| Breath Volume      | 4.042 | 23:47 |
| Air Blank          | 0.000 | 23:47 |
| Air Blank          | 0.000 | 23:49 |
| Subject Sample # 2 | 0.243 | 23:50 |
| Breath Volume      | 3.207 | 23:50 |
| Air Blank          | 0.000 | 23:51 |
| Control Test       | 0.077 | 23:51 |
| Air Blank          | 0.000 | 23:51 |
| Diagnostics Check  | OK    | 23:51 |

---

Data Download Date/Time: 08/22/2024 11:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ANDERSON,JASON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 08/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG229803  
DRY GAS EXPIRY DATE: 10/25/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: SLUSS

DATE/TIME OF TEST: 08/22/2024 21:56:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Control Test       | 0.078 | 21:51 |
| Air Blank          | 0.000 | 21:52 |
| Subject Sample # 1 | 0.000 | 21:52 |
| Breath Volume      | 1.382 | 21:52 |
| Air Blank          | 0.000 | 21:53 |
| Air Blank          | 0.000 | 21:55 |
| Subject Sample # 2 | 0.000 | 21:55 |
| Breath Volume      | 1.351 | 21:55 |
| Air Blank          | 0.000 | 21:55 |
| Control Test       | 0.080 | 21:56 |
| Air Blank          | 0.000 | 21:56 |
| Diagnostics Check  | OK    | 21:56 |

---

Data Download Date/Time: 09/19/2024 16:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLOWAY,DEVARIOUS,J  
BREATH TEST OPERATOR AGENCY: Miramar PD  
ARRESTING OFFICER: DORMILUS,CHRISTIAN  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477284  
DRY GAS EXPIRY DATE: 06/27/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MCCALLA

DATE/TIME OF TEST: 08/09/2024 22:51:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Control Test       | 0.077 | 22:44 |
| Air Blank          | 0.000 | 22:44 |
| Subject Sample # 1 | 0.244 | 22:46 |
| Breath Volume      | 1.906 | 22:46 |
| Air Blank          | 0.000 | 22:46 |
| Air Blank          | 0.000 | 22:48 |
| Subject Sample # 2 | 0.240 | 22:49 |
| Breath Volume      | 1.687 | 22:49 |
| Air Blank          | 0.000 | 22:50 |
| Control Test       | 0.077 | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Diagnostics Check  | OK    | 22:51 |

---

Data Download Date/Time: 08/07/2024 11:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,LAMON,A  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: NINA,JOEL  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 08/01/2024 02:12:25

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 02:10 |            |
| Air Blank         | 0.000 | 02:11 |            |
| Control Test      | 0.079 | 02:11 |            |
| Air Blank         |       | 02:11 | RFI Detect |
| Air Blank         | 0.000 | 02:12 |            |

---

Data Download Date/Time: 08/20/2024 18:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,LAMON,A  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: NINA,JOEL  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 08/01/2024 02:20:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:14 |
| Air Blank          | 0.000 | 02:14 |
| Control Test       | 0.079 | 02:15 |
| Air Blank          | 0.000 | 02:15 |
| Subject Sample # 1 | 0.132 | 02:16 |
| Breath Volume      | 2.339 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Air Blank          | 0.000 | 02:18 |
| Subject Sample # 2 | 0.131 | 02:19 |
| Breath Volume      | 3.667 | 02:19 |
| Air Blank          | 0.000 | 02:19 |
| Control Test       | 0.079 | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Diagnostics Check  | OK    | 02:20 |

---

Data Download Date/Time: 08/20/2024 18:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAY,OFFICER,C  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: RAY,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/02/2024 00:42:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:36 |
| Air Blank          | 0.000 | 00:36 |
| Control Test       | 0.079 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Subject Sample # 1 | 0.110 | 00:38 |
| Breath Volume      | 1.984 | 00:38 |
| Air Blank          | 0.000 | 00:38 |
| Air Blank          | 0.000 | 00:40 |
| Subject Sample # 2 | 0.102 | 00:40 |
| Breath Volume      | 1.484 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Control Test       | 0.079 | 00:41 |
| Air Blank          | 0.000 | 00:42 |
| Diagnostics Check  | OK    | 00:42 |

---

Data Download Date/Time: 08/20/2024 18:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: LIGHTFOOT,OFC  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CAPPAS

DATE/TIME OF TEST: 08/04/2024 03:17:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.079 | 03:11 |
| Air Blank          | 0.000 | 03:11 |
| Subject Sample # 1 | 0.094 | 03:12 |
| Breath Volume      | 1.460 | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Air Blank          | 0.000 | 03:14 |
| Subject Sample # 2 | 0.094 | 03:15 |
| Breath Volume      | 1.921 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.080 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Diagnostics Check  | OK    | 03:17 |

---

Data Download Date/Time: 08/20/2024 18:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERRY, CIMMONE,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: FROST, BRIAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: FERRY

DATE/TIME OF TEST: 08/04/2024 03:44:10

|                   |           |       |                   |
|-------------------|-----------|-------|-------------------|
| Diagnostics Check | OK        | 03:29 |                   |
| Air Blank         | 0.000     | 03:30 |                   |
| Control Test      | 0.078     | 03:30 |                   |
| Air Blank         | 0.000     | 03:31 |                   |
| Subject Sample    | # 1 0.213 | 03:32 | No .020 agreement |
| Breath Volume     | 1.203     | 03:32 |                   |
| Air Blank         | 0.000     | 03:33 |                   |
| Air Blank         | 0.000     | 03:34 |                   |
| Subject Sample    | # 2 0.172 | 03:38 | Volume Not Met    |
| Breath Volume     | 0.000     | 03:38 |                   |
| Air Blank         | 0.000     | 03:38 |                   |
| Air Blank         | 0.000     | 03:40 |                   |
| Subject Sample    | # 3 0.187 | 03:42 | No .020 agreement |
| Breath Volume     | 1.871     | 03:42 |                   |
| Air Blank         | 0.000     | 03:43 |                   |
| Control Test      | 0.077     | 03:43 |                   |
| Air Blank         | 0.000     | 03:44 |                   |
| Diagnostics Check | OK        | 03:44 |                   |

Data Download Date/Time: 08/20/2024 18:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERRY,CIMMONE,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: FROST,BRIAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: FERRY

DATE/TIME OF TEST: 08/04/2024 04:00:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Control Test       | 0.079 | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Subject Sample # 1 | 0.196 | 03:50 |
| Breath Volume      | 1.722 | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Air Blank          | 0.000 | 03:53 |
| Subject Sample # 2 | 0.175 | 03:54 |
| Breath Volume      | 1.183 | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Air Blank          | 0.000 | 03:57 |
| Subject Sample # 3 | 0.199 | 03:58 |
| Breath Volume      | 1.191 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.077 | 03:59 |
| Air Blank          | 0.000 | 03:59 |
| Diagnostics Check  | OK    | 04:00 |

---

Data Download Date/Time: 08/20/2024 18:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAY,OFFICER,C  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: CHANDLER,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 08/06/2024 22:04:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:56 |
| Air Blank          | 0.000 | 21:57 |
| Control Test       | 0.079 | 21:57 |
| Air Blank          | 0.000 | 21:57 |
| Subject Sample # 1 | 0.183 | 21:58 |
| Breath Volume      | 1.253 | 21:58 |
| Air Blank          | 0.000 | 21:59 |
| Air Blank          | 0.000 | 22:01 |
| Subject Sample # 2 | 0.179 | 22:02 |
| Breath Volume      | 1.605 | 22:02 |
| Air Blank          | 0.000 | 22:03 |
| Control Test       | 0.079 | 22:03 |
| Air Blank          | 0.000 | 22:03 |
| Diagnostics Check  | OK    | 22:04 |

---

Data Download Date/Time: 08/20/2024 18:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,LAMON,A  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: CHANDLER,RYAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: BAURES

DATE/TIME OF TEST: 08/11/2024 23:08:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:02 |
| Air Blank          | 0.000 | 23:03 |
| Control Test       | 0.079 | 23:03 |
| Air Blank          | 0.000 | 23:04 |
| Subject Sample # 1 | 0.000 | 23:04 |
| Breath Volume      | 3.921 | 23:04 |
| Air Blank          | 0.000 | 23:04 |
| Air Blank          | 0.000 | 23:06 |
| Subject Sample # 2 | 0.000 | 23:07 |
| Breath Volume      | 3.496 | 23:07 |
| Air Blank          | 0.000 | 23:07 |
| Control Test       | 0.079 | 23:08 |
| Air Blank          | 0.000 | 23:08 |
| Diagnostics Check  | OK    | 23:08 |

---

Data Download Date/Time: 08/20/2024 18:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILES,OFFICER,G  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: GILES,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 08/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 08/23/2024 02:54:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:48 |
| Air Blank          | 0.000 | 02:48 |
| Control Test       | 0.079 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 1 | 0.192 | 02:50 |
| Breath Volume      | 1.703 | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Air Blank          | 0.000 | 02:52 |
| Subject Sample # 2 | 0.204 | 02:53 |
| Breath Volume      | 1.367 | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Control Test       | 0.079 | 02:54 |
| Air Blank          | 0.000 | 02:54 |
| Diagnostics Check  | OK    | 02:54 |

---

Data Download Date/Time: 09/23/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YEATES,SCOTT,W  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: YEATES,SCOTT  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 08/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: ULLRICH

DATE/TIME OF TEST: 08/24/2024 20:21:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:14 |
| Air Blank          | 0.000 | 20:15 |
| Control Test       | 0.078 | 20:15 |
| Air Blank          | 0.000 | 20:16 |
| Subject Sample # 1 | 0.161 | 20:16 |
| Breath Volume      | 1.238 | 20:16 |
| Air Blank          | 0.000 | 20:17 |
| Air Blank          | 0.000 | 20:19 |
| Subject Sample # 2 | 0.153 | 20:20 |
| Breath Volume      | 1.300 | 20:20 |
| Air Blank          | 0.000 | 20:20 |
| Control Test       | 0.078 | 20:21 |
| Air Blank          | 0.000 | 20:21 |
| Diagnostics Check  | OK    | 20:21 |

---

Data Download Date/Time: 09/23/2024 16:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUIS,JUSTIN,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BUIS,JUSTIN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 08/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: SVECHENAYA

DATE/TIME OF TEST: 08/30/2024 21:21:17

---

|                   |           |       |                 |
|-------------------|-----------|-------|-----------------|
| Diagnostics Check | OK        | 21:07 |                 |
| Air Blank         | 0.000     | 21:07 |                 |
| Control Test      | 0.078     | 21:07 |                 |
| Air Blank         | 0.000     | 21:08 |                 |
| Subject Sample    | # 1 0.104 | 21:11 | Slope Not Level |
| Breath Volume     | 1.398     | 21:11 |                 |
| Air Blank         | 0.000     | 21:12 |                 |
| Air Blank         | 0.000     | 21:14 |                 |
| Subject Sample    | # 2 0.106 | 21:14 |                 |
| Breath Volume     | 1.636     | 21:14 |                 |
| Air Blank         | 0.000     | 21:15 |                 |
| Air Blank         | 0.000     | 21:17 |                 |
| Subject Sample    | # 3 0.113 | 21:19 |                 |
| Breath Volume     | 2.621     | 21:19 |                 |
| Air Blank         | 0.000     | 21:20 |                 |
| Control Test      | 0.079     | 21:20 |                 |
| Air Blank         | 0.000     | 21:21 |                 |
| Diagnostics Check | OK        | 21:21 |                 |

---

Data Download Date/Time: 09/23/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FROST,BRIAN,A  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: FROST,BRIAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 08/20/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: BERGER

DATE/TIME OF TEST: 08/31/2024 05:19:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:12 |
| Air Blank          | 0.000 | 05:13 |
| Control Test       | 0.078 | 05:13 |
| Air Blank          | 0.000 | 05:13 |
| Subject Sample # 1 | 0.210 | 05:14 |
| Breath Volume      | 2.328 | 05:14 |
| Air Blank          | 0.000 | 05:15 |
| Air Blank          | 0.000 | 05:17 |
| Subject Sample # 2 | 0.194 | 05:18 |
| Breath Volume      | 1.828 | 05:18 |
| Air Blank          | 0.000 | 05:18 |
| Control Test       | 0.079 | 05:18 |
| Air Blank          | 0.000 | 05:19 |
| Diagnostics Check  | OK    | 05:19 |

---

Data Download Date/Time: 09/23/2024 16:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: SORRELL,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477283  
DRY GAS EXPIRY DATE: 06/24/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 08/01/2024 00:59:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:56 |                      |
| Air Blank         | 0.000 | 00:57 |                      |
| Control Test      | 0.079 | 00:57 |                      |
| Air Blank         | 0.000 | 00:58 |                      |
| Subject Sample    | # 1   | 00:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:58 |                      |
| Air Blank         | 0.000 | 00:58 |                      |
| Control Test      | 0.080 | 00:59 |                      |
| Air Blank         | 0.000 | 00:59 |                      |
| Diagnostics Check | OK    | 00:59 |                      |

---

Data Download Date/Time: 08/30/2024 00:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELHOMME,MAXON,N  
BREATH TEST OPERATOR AGENCY: Wauchula PD  
ARRESTING OFFICER: SKINNER,OWEN  
ARRESTING OFFICER AGENCY: Wauchula PD  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477283  
DRY GAS EXPIRY DATE: 06/24/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: NOBLETT

DATE/TIME OF TEST: 08/04/2024 19:01:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:55 |
| Air Blank          | 0.000 | 18:55 |
| Control Test       | 0.079 | 18:55 |
| Air Blank          | 0.000 | 18:56 |
| Subject Sample # 1 | 0.341 | 18:57 |
| Breath Volume      | 2.222 | 18:57 |
| Air Blank          | 0.000 | 18:57 |
| Air Blank          | 0.000 | 18:59 |
| Subject Sample # 2 | 0.337 | 19:00 |
| Breath Volume      | 3.308 | 19:00 |
| Air Blank          | 0.000 | 19:00 |
| Control Test       | 0.079 | 19:01 |
| Air Blank          | 0.000 | 19:01 |
| Diagnostics Check  | OK    | 19:01 |

---

Data Download Date/Time: 08/30/2024 00:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELHOMME,MAXON,N  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: TYLER,ETHAN  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477283  
DRY GAS EXPIRY DATE: 06/24/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: FERIA

DATE/TIME OF TEST: 08/17/2024 23:38:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:31 |
| Air Blank          | 0.000 | 23:31 |
| Control Test       | 0.079 | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Subject Sample # 1 | 0.198 | 23:33 |
| Breath Volume      | 1.464 | 23:33 |
| Air Blank          | 0.000 | 23:34 |
| Air Blank          | 0.000 | 23:36 |
| Subject Sample # 2 | 0.204 | 23:36 |
| Breath Volume      | 1.796 | 23:36 |
| Air Blank          | 0.000 | 23:37 |
| Control Test       | 0.078 | 23:37 |
| Air Blank          | 0.000 | 23:38 |
| Diagnostics Check  | OK    | 23:38 |

---

Data Download Date/Time: 08/30/2024 00:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELHOMME,MAXON,N  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: WALKER,DUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477283  
DRY GAS EXPIRY DATE: 06/24/2025  
VIOLATION CODE: DUI  
ARREST TIME: 10:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: HIPOLITO-NICOLAS

DATE/TIME OF TEST: 08/18/2024 00:24:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:17 |
| Air Blank          | 0.000 | 00:18 |
| Control Test       | 0.079 | 00:18 |
| Air Blank          | 0.000 | 00:18 |
| Subject Sample # 1 | 0.186 | 00:19 |
| Breath Volume      | 2.210 | 00:19 |
| Air Blank          | 0.000 | 00:20 |
| Air Blank          | 0.000 | 00:22 |
| Subject Sample # 2 | 0.181 | 00:22 |
| Breath Volume      | 3.699 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Control Test       | 0.077 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Diagnostics Check  | OK    | 00:24 |

---

Data Download Date/Time: 08/30/2024 00:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELHOMME,MAXON,N  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: SKINNER,OWEN  
ARRESTING OFFICER AGENCY: Wauchula PD  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477283  
DRY GAS EXPIRY DATE: 06/24/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GOUGH

DATE/TIME OF TEST: 08/18/2024 01:23:59

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:21 |                      |
| Air Blank         | 0.000 | 01:21 |                      |
| Control Test      | 0.080 | 01:21 |                      |
| Air Blank         | 0.000 | 01:22 |                      |
| Subject Sample    | # 1   | 01:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:22 |                      |
| Air Blank         | 0.000 | 01:23 |                      |
| Control Test      | 0.078 | 01:23 |                      |
| Air Blank         | 0.000 | 01:23 |                      |
| Diagnostics Check | OK    | 01:23 |                      |

Data Download Date/Time: 08/30/2024 00:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLETCHER,MELISSA,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WAINWRIGHT,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: ROTTON

DATE/TIME OF TEST: 08/01/2024 19:48:36

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 19:31 |                    |
| Air Blank         | 0.000 | 19:31 |                    |
| Control Test      | 0.080 | 19:32 |                    |
| Air Blank         | 0.000 | 19:32 |                    |
| Subject Sample    | # 1   | 19:35 | No Sample Provided |
| Breath Volume     | 0.000 | 19:35 |                    |
| Air Blank         | 0.000 | 19:36 |                    |
| Air Blank         | 0.000 | 19:38 |                    |
| Subject Sample    | # 2   | 0.133 | 19:41              |
| Breath Volume     | 1.335 | 19:41 |                    |
| Air Blank         | 0.000 | 19:42 |                    |
| Air Blank         | 0.000 | 19:43 |                    |
| Subject Sample    | # 3   | 0.135 | 19:47              |
| Breath Volume     | 0.070 | 19:47 | Volume Not Met     |
| Air Blank         | 0.000 | 19:47 |                    |
| Control Test      | 0.081 | 19:48 |                    |
| Air Blank         | 0.000 | 19:48 |                    |
| Diagnostics Check | OK    | 19:48 |                    |

---

Data Download Date/Time: 08/13/2024 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLETCHER,MELISSA,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WAINWRIGHT,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 08/01/2024 22:07:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:01 |
| Air Blank          | 0.000 | 22:01 |
| Control Test       | 0.080 | 22:02 |
| Air Blank          | 0.000 | 22:02 |
| Subject Sample # 1 | 0.000 | 22:03 |
| Breath Volume      | 1.382 | 22:03 |
| Air Blank          | 0.000 | 22:03 |
| Air Blank          | 0.000 | 22:05 |
| Subject Sample # 2 | 0.000 | 22:06 |
| Breath Volume      | 1.386 | 22:06 |
| Air Blank          | 0.000 | 22:06 |
| Control Test       | 0.081 | 22:07 |
| Air Blank          | 0.000 | 22:07 |
| Diagnostics Check  | OK    | 22:07 |

---

Data Download Date/Time: 08/13/2024 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 08/02/2024 20:44:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:38 |
| Air Blank          | 0.000 | 20:39 |
| Control Test       | 0.079 | 20:39 |
| Air Blank          | 0.000 | 20:39 |
| Subject Sample # 1 | 0.000 | 20:40 |
| Breath Volume      | 2.605 | 20:40 |
| Air Blank          | 0.000 | 20:40 |
| Air Blank          | 0.000 | 20:42 |
| Subject Sample # 2 | 0.000 | 20:43 |
| Breath Volume      | 2.660 | 20:43 |
| Air Blank          | 0.000 | 20:43 |
| Control Test       | 0.080 | 20:43 |
| Air Blank          | 0.000 | 20:44 |
| Diagnostics Check  | OK    | 20:44 |

---

Data Download Date/Time: 08/13/2024 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CALISTE

DATE/TIME OF TEST: 08/03/2024 00:48:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.080 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Subject Sample # 1 | 0.161 | 00:44 |
| Breath Volume      | 1.683 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 2 | 0.161 | 00:47 |
| Breath Volume      | 1.601 | 00:47 |
| Air Blank          | 0.000 | 00:47 |
| Control Test       | 0.080 | 00:48 |
| Air Blank          | 0.000 | 00:48 |
| Diagnostics Check  | OK    | 00:48 |

---

Data Download Date/Time: 08/13/2024 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JANIAK,M  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MAGANA

DATE/TIME OF TEST: 08/03/2024 01:13:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:07 |
| Air Blank          | 0.000 | 01:07 |
| Control Test       | 0.080 | 01:08 |
| Air Blank          | 0.000 | 01:08 |
| Subject Sample # 1 | 0.165 | 01:09 |
| Breath Volume      | 1.589 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Air Blank          | 0.000 | 01:11 |
| Subject Sample # 2 | 0.160 | 01:12 |
| Breath Volume      | 1.683 | 01:12 |
| Air Blank          | 0.000 | 01:12 |
| Control Test       | 0.080 | 01:13 |
| Air Blank          | 0.000 | 01:13 |
| Diagnostics Check  | OK    | 01:13 |

---

Data Download Date/Time: 08/13/2024 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: SANJUAN FRANCISCO

DATE/TIME OF TEST: 08/04/2024 04:23:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:16 |
| Air Blank          | 0.000 | 04:16 |
| Control Test       | 0.079 | 04:17 |
| Air Blank          | 0.000 | 04:17 |
| Subject Sample # 1 | 0.185 | 04:18 |
| Breath Volume      | 1.867 | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Air Blank          | 0.000 | 04:21 |
| Subject Sample # 2 | 0.180 | 04:21 |
| Breath Volume      | 1.925 | 04:21 |
| Air Blank          | 0.000 | 04:22 |
| Control Test       | 0.078 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Diagnostics Check  | OK    | 04:23 |

---

Data Download Date/Time: 08/13/2024 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: DENMARK, HAROLD  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 09:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: STRIPLING

DATE/TIME OF TEST: 08/05/2024 22:47:05

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:35 |                    |
| Air Blank         | 0.000 | 22:36 |                    |
| Control Test      | 0.080 | 22:36 |                    |
| Air Blank         | 0.000 | 22:36 |                    |
| Subject Sample    | # 1   | 22:40 | No Sample Provided |
| Breath Volume     | 0.000 | 22:40 |                    |
| Air Blank         | 0.000 | 22:40 |                    |
| Air Blank         | 0.000 | 22:42 |                    |
| Subject Sample    | # 2   | 22:45 | No Sample Provided |
| Breath Volume     | 0.000 | 22:45 |                    |
| Air Blank         | 0.000 | 22:46 |                    |
| Control Test      | 0.081 | 22:46 |                    |
| Air Blank         | 0.000 | 22:47 |                    |
| Diagnostics Check | OK    | 22:47 |                    |

Data Download Date/Time: 08/13/2024 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WAINWRIGHT, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/06/2024 03:08:43

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:56 |                    |
| Air Blank         | 0.000 | 02:56 |                    |
| Control Test      | 0.081 | 02:56 |                    |
| Air Blank         | 0.000 | 02:57 |                    |
| Subject Sample    | # 1   | 03:00 | No Sample Provided |
| Breath Volume     | 0.000 | 03:00 |                    |
| Air Blank         | 0.000 | 03:01 |                    |
| Air Blank         | 0.000 | 03:03 |                    |
| Subject Sample    | # 2   | 0.209 | 03:03              |
| Breath Volume     | 1.777 | 03:03 |                    |
| Air Blank         | 0.000 | 03:04 |                    |
| Air Blank         | 0.000 | 03:06 |                    |
| Subject Sample    | # 3   | 0.195 | 03:07              |
| Breath Volume     | 1.296 | 03:07 |                    |
| Air Blank         | 0.000 | 03:07 |                    |
| Control Test      | 0.079 | 03:08 |                    |
| Air Blank         | 0.000 | 03:08 |                    |
| Diagnostics Check | OK    | 03:08 |                    |

Data Download Date/Time: 08/13/2024 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WAINWRIGHT, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: FORESYTH

DATE/TIME OF TEST: 08/06/2024 23:21:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:13 |
| Air Blank          | 0.000 | 23:13 |
| Control Test       | 0.080 | 23:14 |
| Air Blank          | 0.000 | 23:14 |
| Subject Sample # 1 | 0.193 | 23:16 |
| Breath Volume      | 1.726 | 23:16 |
| Air Blank          | 0.000 | 23:16 |
| Air Blank          | 0.000 | 23:18 |
| Subject Sample # 2 | 0.194 | 23:19 |
| Breath Volume      | 1.781 | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Control Test       | 0.078 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Diagnostics Check  | OK    | 23:21 |

---

Data Download Date/Time: 08/13/2024 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WAINWRIGHT, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 08/07/2024 00:52:06

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 00:40 |                    |
| Air Blank         | 0.000 | 00:41 |                    |
| Control Test      | 0.080 | 00:41 |                    |
| Air Blank         | 0.000 | 00:41 |                    |
| Subject Sample    | # 1   | 00:45 | No Sample Provided |
| Breath Volume     | 0.000 | 00:45 |                    |
| Air Blank         | 0.000 | 00:45 |                    |
| Air Blank         | 0.000 | 00:47 |                    |
| Subject Sample    | # 2   | 00:50 | No Sample Provided |
| Breath Volume     | 0.000 | 00:50 |                    |
| Air Blank         | 0.000 | 00:51 |                    |
| Control Test      | 0.082 | 00:51 |                    |
| Air Blank         | 0.000 | 00:52 |                    |
| Diagnostics Check | OK    | 00:52 |                    |

Data Download Date/Time: 08/13/2024 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: NESTER, LOGAN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 08/07/2024 03:39:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Control Test       | 0.080 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Subject Sample # 1 | 0.170 | 03:35 |
| Breath Volume      | 1.945 | 03:35 |
| Air Blank          | 0.000 | 03:35 |
| Air Blank          | 0.000 | 03:37 |
| Subject Sample # 2 | 0.164 | 03:38 |
| Breath Volume      | 1.835 | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Control Test       | 0.080 | 03:39 |
| Air Blank          | 0.000 | 03:39 |
| Diagnostics Check  | OK    | 03:39 |

---

Data Download Date/Time: 08/13/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ANGULO, DALLAS  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: TUTTLE

DATE/TIME OF TEST: 08/07/2024 03:56:15

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:44 |                    |
| Air Blank         | 0.000 | 03:45 |                    |
| Control Test      | 0.080 | 03:45 |                    |
| Air Blank         | 0.000 | 03:46 |                    |
| Subject Sample    | # 1   | 03:49 | No Sample Provided |
| Breath Volume     | 0.000 | 03:49 |                    |
| Air Blank         | 0.000 | 03:49 |                    |
| Air Blank         | 0.000 | 03:51 |                    |
| Subject Sample    | # 2   | 03:54 | No Sample Provided |
| Breath Volume     | 0.000 | 03:54 |                    |
| Air Blank         | 0.000 | 03:55 |                    |
| Control Test      | 0.080 | 03:55 |                    |
| Air Blank         | 0.000 | 03:56 |                    |
| Diagnostics Check | OK    | 03:56 |                    |

Data Download Date/Time: 08/13/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ALCALDE,ALEXANDER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 07:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:11

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 08/07/2024 08:41:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:34 |
| Air Blank          | 0.000 | 08:34 |
| Control Test       | 0.079 | 08:35 |
| Air Blank          | 0.000 | 08:35 |
| Subject Sample # 1 | 0.097 | 08:36 |
| Breath Volume      | 1.906 | 08:36 |
| Air Blank          | 0.000 | 08:36 |
| Air Blank          | 0.000 | 08:38 |
| Subject Sample # 2 | 0.093 | 08:39 |
| Breath Volume      | 2.636 | 08:39 |
| Air Blank          | 0.000 | 08:40 |
| Control Test       | 0.080 | 08:40 |
| Air Blank          | 0.000 | 08:40 |
| Diagnostics Check  | OK    | 08:41 |

---

Data Download Date/Time: 08/13/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DE LA CRUZ, LEONARDO, E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DE LA CRUZ, LEONARDO  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LEDUC RIVERA

DATE/TIME OF TEST: 08/08/2024 01:50:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Control Test       | 0.080 | 01:44 |
| Air Blank          | 0.000 | 01:44 |
| Subject Sample # 1 | 0.109 | 01:45 |
| Breath Volume      | 1.367 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Subject Sample # 2 | 0.108 | 01:48 |
| Breath Volume      | 1.621 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.078 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Diagnostics Check  | OK    | 01:50 |

---

Data Download Date/Time: 08/13/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLAY,JAMES,M  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ELLIOTT,JACOB  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/09/2024 07:05:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:59 |
| Air Blank          | 0.000 | 07:00 |
| Control Test       | 0.080 | 07:00 |
| Air Blank          | 0.000 | 07:01 |
| Subject Sample # 1 | 0.000 | 07:01 |
| Breath Volume      | 2.902 | 07:01 |
| Air Blank          | 0.000 | 07:02 |
| Air Blank          | 0.000 | 07:04 |
| Subject Sample # 2 | 0.000 | 07:04 |
| Breath Volume      | 2.843 | 07:04 |
| Air Blank          | 0.000 | 07:04 |
| Control Test       | 0.079 | 07:05 |
| Air Blank          | 0.000 | 07:05 |
| Diagnostics Check  | OK    | 07:05 |

---

Data Download Date/Time: 08/13/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOOTEN,NATHANIEL,K  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: JONES,ARIK  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: FALCON

DATE/TIME OF TEST: 08/10/2024 18:44:11

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 18:32 |                    |
| Air Blank         | 0.000 | 18:33 |                    |
| Control Test      | 0.079 | 18:33 |                    |
| Air Blank         | 0.000 | 18:34 |                    |
| Subject Sample    | # 1   | 18:37 | No Sample Provided |
| Breath Volume     | 0.000 | 18:37 |                    |
| Air Blank         | 0.000 | 18:37 |                    |
| Air Blank         | 0.000 | 18:39 |                    |
| Subject Sample    | # 2   | 18:42 | No Sample Provided |
| Breath Volume     | 0.000 | 18:42 |                    |
| Air Blank         | 0.000 | 18:43 |                    |
| Control Test      | 0.079 | 18:43 |                    |
| Air Blank         | 0.000 | 18:44 |                    |
| Diagnostics Check | OK    | 18:44 |                    |

---

Data Download Date/Time: 08/13/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARREN,BILLY,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: MARENCO,ALFREDO  
ARRESTING OFFICER AGENCY: Auburndale PD  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: GONZALEZ GARCIA

DATE/TIME OF TEST: 08/10/2024 22:41:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:32 |                      |
| Air Blank         | 0.000 | 22:32 |                      |
| Control Test      | 0.079 | 22:32 |                      |
| Air Blank         | 0.000 | 22:33 |                      |
| Subject Sample    | # 1   | 22:36 | No Sample Provided   |
| Breath Volume     | 0.000 | 22:36 |                      |
| Air Blank         | 0.000 | 22:37 |                      |
| Air Blank         | 0.000 | 22:39 |                      |
| Subject Sample    | # 2   | 22:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:39 |                      |
| Air Blank         | 0.000 | 22:40 |                      |
| Control Test      | 0.078 | 22:40 |                      |
| Air Blank         | 0.000 | 22:41 |                      |
| Diagnostics Check | OK    | 22:41 |                      |

---

Data Download Date/Time: 08/13/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ANGULO, DALLAS  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 08/11/2024 01:44:40

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 01:33 |                    |
| Air Blank         | 0.000     | 01:33 |                    |
| Control Test      | 0.080     | 01:34 |                    |
| Air Blank         | 0.000     | 01:34 |                    |
| Subject Sample    | # 1 0.113 | 01:37 | Volume Not Met     |
| Breath Volume     | 0.000     | 01:37 |                    |
| Air Blank         | 0.000     | 01:38 |                    |
| Air Blank         | 0.000     | 01:40 |                    |
| Subject Sample    | # 2       | 01:43 | No Sample Provided |
| Breath Volume     | 0.000     | 01:43 |                    |
| Air Blank         | 0.000     | 01:43 |                    |
| Control Test      | 0.079     | 01:44 |                    |
| Air Blank         | 0.000     | 01:44 |                    |
| Diagnostics Check | OK        | 01:44 |                    |

---

Data Download Date/Time: 08/13/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: PARSLEY

DATE/TIME OF TEST: 08/13/2024 01:07:31

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 00:56 |                    |
| Air Blank         | 0.000 | 00:56 |                    |
| Control Test      | 0.080 | 00:56 |                    |
| Air Blank         | 0.000 | 00:57 |                    |
| Subject Sample    | # 1   | 01:00 | No Sample Provided |
| Breath Volume     | 0.000 | 01:00 |                    |
| Air Blank         | 0.000 | 01:01 |                    |
| Air Blank         | 0.000 | 01:03 |                    |
| Subject Sample    | # 2   | 01:06 | No Sample Provided |
| Breath Volume     | 0.000 | 01:06 |                    |
| Air Blank         | 0.000 | 01:06 |                    |
| Control Test      | 0.080 | 01:06 |                    |
| Air Blank         | 0.000 | 01:07 |                    |
| Diagnostics Check | OK    | 01:07 |                    |

---

Data Download Date/Time: 08/13/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 08/13/2024 01:47:38

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:36 |                    |
| Air Blank         | 0.000 | 01:36 |                    |
| Control Test      | 0.080 | 01:37 |                    |
| Air Blank         | 0.000 | 01:37 |                    |
| Subject Sample    | # 1   | 01:40 | No Sample Provided |
| Breath Volume     | 0.000 | 01:40 |                    |
| Air Blank         | 0.000 | 01:41 |                    |
| Air Blank         | 0.000 | 01:43 |                    |
| Subject Sample    | # 2   | 01:46 | No Sample Provided |
| Breath Volume     | 0.000 | 01:46 |                    |
| Air Blank         | 0.000 | 01:46 |                    |
| Control Test      | 0.081 | 01:47 |                    |
| Air Blank         | 0.000 | 01:47 |                    |
| Diagnostics Check | OK    | 01:47 |                    |

Data Download Date/Time: 08/13/2024 10:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ALCALDE,ALEXANDER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 06:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: ASHLEY JR

DATE/TIME OF TEST: 08/13/2024 08:36:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:30 |
| Air Blank          | 0.000 | 08:31 |
| Control Test       | 0.080 | 08:31 |
| Air Blank          | 0.000 | 08:32 |
| Subject Sample # 1 | 0.000 | 08:32 |
| Breath Volume      | 2.121 | 08:32 |
| Air Blank          | 0.000 | 08:32 |
| Air Blank          | 0.000 | 08:34 |
| Subject Sample # 2 | 0.000 | 08:35 |
| Breath Volume      | 2.210 | 08:35 |
| Air Blank          | 0.000 | 08:35 |
| Control Test       | 0.080 | 08:36 |
| Air Blank          | 0.000 | 08:36 |
| Diagnostics Check  | OK    | 08:36 |

---

Data Download Date/Time: 08/13/2024 10:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ALCALDE,ALEXANDER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: PRECATADO

DATE/TIME OF TEST: 08/13/2024 17:38:53

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 17:27 |                    |
| Air Blank         | 0.000 | 17:27 |                    |
| Control Test      | 0.081 | 17:28 |                    |
| Air Blank         | 0.000 | 17:28 |                    |
| Subject Sample    | # 1   | 17:31 | No Sample Provided |
| Breath Volume     | 0.000 | 17:31 |                    |
| Air Blank         | 0.000 | 17:32 |                    |
| Air Blank         | 0.000 | 17:34 |                    |
| Subject Sample    | # 2   | 17:37 | No Sample Provided |
| Breath Volume     | 0.000 | 17:37 |                    |
| Air Blank         | 0.000 | 17:37 |                    |
| Control Test      | 0.081 | 17:38 |                    |
| Air Blank         | 0.000 | 17:38 |                    |
| Diagnostics Check | OK    | 17:38 |                    |

Data Download Date/Time: 09/10/2024 09:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 18:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: CARAMBOT OYOLA

DATE/TIME OF TEST: 08/13/2024 20:06:56

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 19:55 |                    |
| Air Blank         | 0.000 | 19:56 |                    |
| Control Test      | 0.081 | 19:56 |                    |
| Air Blank         | 0.000 | 19:56 |                    |
| Subject Sample    | # 1   | 19:59 | No Sample Provided |
| Breath Volume     | 0.000 | 19:59 |                    |
| Air Blank         | 0.000 | 20:00 |                    |
| Air Blank         | 0.000 | 20:02 |                    |
| Subject Sample    | # 2   | 20:05 | No Sample Provided |
| Breath Volume     | 0.000 | 20:05 |                    |
| Air Blank         | 0.000 | 20:06 |                    |
| Control Test      | 0.081 | 20:06 |                    |
| Air Blank         | 0.000 | 20:06 |                    |
| Diagnostics Check | OK    | 20:06 |                    |

---

Data Download Date/Time: 09/10/2024 09:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: MADISON

DATE/TIME OF TEST: 08/14/2024 00:32:23

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 00:20 |                    |
| Air Blank         | 0.000 | 00:21 |                    |
| Control Test      | 0.080 | 00:21 |                    |
| Air Blank         | 0.000 | 00:22 |                    |
| Subject Sample    | # 1   | 00:25 | No Sample Provided |
| Breath Volume     | 0.000 | 00:25 |                    |
| Air Blank         | 0.000 | 00:25 |                    |
| Air Blank         | 0.000 | 00:27 |                    |
| Subject Sample    | # 2   | 00:31 | No Sample Provided |
| Breath Volume     | 0.000 | 00:31 |                    |
| Air Blank         | 0.000 | 00:31 |                    |
| Control Test      | 0.080 | 00:31 |                    |
| Air Blank         | 0.000 | 00:32 |                    |
| Diagnostics Check | OK    | 00:32 |                    |

---

Data Download Date/Time: 09/10/2024 09:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: GRISSOM, TIMOTHY  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: TRINIDAD

DATE/TIME OF TEST: 08/14/2024 22:57:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:51 |
| Air Blank          | 0.000 | 22:51 |
| Control Test       | 0.080 | 22:51 |
| Air Blank          | 0.000 | 22:52 |
| Subject Sample # 1 | 0.191 | 22:52 |
| Breath Volume      | 2.027 | 22:52 |
| Air Blank          | 0.000 | 22:53 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 2 | 0.191 | 22:55 |
| Breath Volume      | 1.855 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Control Test       | 0.080 | 22:56 |
| Air Blank          | 0.000 | 22:57 |
| Diagnostics Check  | OK    | 22:57 |

---

Data Download Date/Time: 09/10/2024 09:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ, JAIRO, A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WOOTEN, NATHANIEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 08/15/2024 01:44:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:37 |
| Air Blank          | 0.000 | 01:38 |
| Control Test       | 0.080 | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Subject Sample # 1 | 0.070 | 01:40 |
| Breath Volume      | 3.234 | 01:40 |
| Air Blank          | 0.000 | 01:40 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 2 | 0.070 | 01:42 |
| Breath Volume      | 3.472 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Control Test       | 0.079 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Diagnostics Check  | OK    | 01:44 |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ALCALDE,ALEXANDER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 06:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:46

SUBJECT'S LAST NAME: PAILLANT

DATE/TIME OF TEST: 08/16/2024 08:22:41

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 08:11 |                    |
| Air Blank         | 0.000 | 08:11 |                    |
| Control Test      | 0.081 | 08:12 |                    |
| Air Blank         | 0.000 | 08:12 |                    |
| Subject Sample    | # 1   | 08:15 | No Sample Provided |
| Breath Volume     | 0.000 | 08:15 |                    |
| Air Blank         | 0.000 | 08:16 |                    |
| Air Blank         | 0.000 | 08:18 |                    |
| Subject Sample    | # 2   | 08:21 | No Sample Provided |
| Breath Volume     | 0.000 | 08:21 |                    |
| Air Blank         | 0.000 | 08:21 |                    |
| Control Test      | 0.081 | 08:22 |                    |
| Air Blank         | 0.000 | 08:22 |                    |
| Diagnostics Check | OK    | 08:22 |                    |

---

Data Download Date/Time: 09/10/2024 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: BERNARDEZ GUILTY

DATE/TIME OF TEST: 08/16/2024 23:05:11

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:53 |                    |
| Air Blank         | 0.000 | 22:54 |                    |
| Control Test      | 0.081 | 22:54 |                    |
| Air Blank         | 0.000 | 22:55 |                    |
| Subject Sample    | # 1   | 22:58 | No Sample Provided |
| Breath Volume     | 0.000 | 22:58 |                    |
| Air Blank         | 0.000 | 22:58 |                    |
| Air Blank         | 0.000 | 23:00 |                    |
| Subject Sample    | # 2   | 23:03 | No Sample Provided |
| Breath Volume     | 0.000 | 23:03 |                    |
| Air Blank         | 0.000 | 23:04 |                    |
| Control Test      | 0.080 | 23:04 |                    |
| Air Blank         | 0.000 | 23:05 |                    |
| Diagnostics Check | OK    | 23:05 |                    |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/17/2024 00:27:31

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 00:25 |            |
| Air Blank         | 0.000 | 00:26 |            |
| Control Test      | 0.080 | 00:26 |            |
| Air Blank         |       | 00:27 | RFI Detect |
| Air Blank         | 0.000 | 00:27 |            |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/17/2024 00:36:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:30 |
| Air Blank          | 0.000 | 00:30 |
| Control Test       | 0.079 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Subject Sample # 1 | 0.140 | 00:32 |
| Breath Volume      | 2.343 | 00:32 |
| Air Blank          | 0.000 | 00:32 |
| Air Blank          | 0.000 | 00:34 |
| Subject Sample # 2 | 0.136 | 00:35 |
| Breath Volume      | 2.410 | 00:35 |
| Air Blank          | 0.000 | 00:35 |
| Control Test       | 0.080 | 00:36 |
| Air Blank          | 0.000 | 00:36 |
| Diagnostics Check  | OK    | 00:36 |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 08/17/2024 01:36:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.080 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 1 | 0.000 | 01:32 |
| Breath Volume      | 1.753 | 01:32 |
| Air Blank          | 0.000 | 01:33 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.000 | 01:35 |
| Breath Volume      | 2.050 | 01:35 |
| Air Blank          | 0.000 | 01:35 |
| Control Test       | 0.082 | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Diagnostics Check  | OK    | 01:36 |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTO,MAVERICK,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SOTO,MAVERICK  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 08/17/2024 03:06:34

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:58 |                      |
| Air Blank         | 0.000 | 02:58 |                      |
| Control Test      | 0.080 | 02:58 |                      |
| Air Blank         | 0.000 | 02:59 |                      |
| Subject Sample    | # 1   | 03:02 | No Sample Provided   |
| Breath Volume     | 0.000 | 03:02 |                      |
| Air Blank         | 0.000 | 03:02 |                      |
| Air Blank         | 0.000 | 03:04 |                      |
| Subject Sample    | # 2   | 03:05 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:05 |                      |
| Air Blank         | 0.000 | 03:05 |                      |
| Control Test      | 0.081 | 03:05 |                      |
| Air Blank         | 0.000 | 03:06 |                      |
| Diagnostics Check | OK    | 03:06 |                      |

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: PUENTE VITERI

DATE/TIME OF TEST: 08/17/2024 22:17:43

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:06 |                |
| Air Blank         | 0.000     | 22:06 |                |
| Control Test      | 0.081     | 22:06 |                |
| Air Blank         | 0.000     | 22:07 |                |
| Subject Sample    | # 1 0.133 | 22:10 | Volume Not Met |
| Breath Volume     | 0.484     | 22:10 |                |
| Air Blank         | 0.000     | 22:11 |                |
| Air Blank         | 0.000     | 22:12 |                |
| Subject Sample    | # 2 0.118 | 22:16 | Volume Not Met |
| Breath Volume     | 0.359     | 22:16 |                |
| Air Blank         | 0.000     | 22:16 |                |
| Control Test      | 0.080     | 22:17 |                |
| Air Blank         | 0.000     | 22:17 |                |
| Diagnostics Check | OK        | 22:17 |                |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JANIAK,M  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BERNAL

DATE/TIME OF TEST: 08/17/2024 22:56:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:48 |
| Air Blank          | 0.000 | 22:48 |
| Control Test       | 0.081 | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Subject Sample # 1 | 0.169 | 22:50 |
| Breath Volume      | 1.464 | 22:50 |
| Air Blank          | 0.000 | 22:50 |
| Air Blank          | 0.000 | 22:52 |
| Subject Sample # 2 | 0.160 | 22:54 |
| Breath Volume      | 1.757 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Control Test       | 0.080 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Diagnostics Check  | OK    | 22:56 |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: METELLUS,SAM,S  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ELLIOTT,JACOB  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RIVERA TORRES

DATE/TIME OF TEST: 08/18/2024 00:37:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.080 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Subject Sample # 1 | 0.125 | 00:32 |
| Breath Volume      | 2.871 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Air Blank          | 0.000 | 00:35 |
| Subject Sample # 2 | 0.127 | 00:35 |
| Breath Volume      | 3.156 | 00:35 |
| Air Blank          | 0.000 | 00:36 |
| Control Test       | 0.078 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Diagnostics Check  | OK    | 00:37 |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAWSON JR,MICHAEL,E  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: DRAGAN,BRYAN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 08/18/2024 04:12:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:01 |
| Air Blank          | 0.000 | 04:01 |
| Control Test       | 0.080 | 04:02 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 1 | 0.219 | 04:04 |
| Breath Volume      | 1.808 | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Subject Sample # 2 | 0.197 | 04:07 |
| Breath Volume      | 2.015 | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Air Blank          | 0.000 | 04:10 |
| Subject Sample # 3 | 0.190 | 04:11 |
| Breath Volume      | 2.054 | 04:11 |
| Air Blank          | 0.000 | 04:11 |
| Control Test       | 0.077 | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Diagnostics Check  | OK    | 04:12 |

---

Data Download Date/Time: 09/10/2024 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ALCALDE,ALEXANDER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 06:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 08/18/2024 07:08:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:01 |
| Air Blank          | 0.000 | 07:02 |
| Control Test       | 0.080 | 07:02 |
| Air Blank          | 0.000 | 07:02 |
| Subject Sample # 1 | 0.143 | 07:03 |
| Breath Volume      | 3.460 | 07:03 |
| Air Blank          | 0.000 | 07:04 |
| Air Blank          | 0.000 | 07:06 |
| Subject Sample # 2 | 0.139 | 07:06 |
| Breath Volume      | 2.914 | 07:06 |
| Air Blank          | 0.000 | 07:07 |
| Control Test       | 0.080 | 07:07 |
| Air Blank          | 0.000 | 07:08 |
| Diagnostics Check  | OK    | 07:08 |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/18/2024 23:48:09

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 23:36 |                    |
| Air Blank         | 0.000 | 23:37 |                    |
| Control Test      | 0.081 | 23:37 |                    |
| Air Blank         | 0.000 | 23:38 |                    |
| Subject Sample    | # 1   | 23:41 | No Sample Provided |
| Breath Volume     | 0.000 | 23:41 |                    |
| Air Blank         | 0.000 | 23:41 |                    |
| Air Blank         | 0.000 | 23:43 |                    |
| Subject Sample    | # 2   | 23:46 | No Sample Provided |
| Breath Volume     | 0.000 | 23:46 |                    |
| Air Blank         | 0.000 | 23:47 |                    |
| Control Test      | 0.081 | 23:47 |                    |
| Air Blank         | 0.000 | 23:48 |                    |
| Diagnostics Check | OK    | 23:48 |                    |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: GRISSOM, TIMOTHY  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: CHMARA

DATE/TIME OF TEST: 08/20/2024 21:32:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:26 |
| Air Blank          | 0.000 | 21:26 |
| Control Test       | 0.079 | 21:26 |
| Air Blank          | 0.000 | 21:27 |
| Subject Sample # 1 | 0.337 | 21:28 |
| Breath Volume      | 1.843 | 21:28 |
| Air Blank          | 0.000 | 21:28 |
| Air Blank          | 0.000 | 21:30 |
| Subject Sample # 2 | 0.334 | 21:31 |
| Breath Volume      | 2.246 | 21:31 |
| Air Blank          | 0.000 | 21:31 |
| Control Test       | 0.076 | 21:32 |
| Air Blank          | 0.000 | 21:32 |
| Diagnostics Check  | OK    | 21:32 |

---

Data Download Date/Time: 09/10/2024 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: WARRIX

DATE/TIME OF TEST: 08/21/2024 21:06:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:00 |
| Air Blank          | 0.000 | 21:00 |
| Control Test       | 0.081 | 21:00 |
| Air Blank          | 0.000 | 21:01 |
| Subject Sample # 1 | 0.000 | 21:01 |
| Breath Volume      | 2.246 | 21:01 |
| Air Blank          | 0.000 | 21:02 |
| Air Blank          | 0.000 | 21:04 |
| Subject Sample # 2 | 0.000 | 21:04 |
| Breath Volume      | 1.777 | 21:04 |
| Air Blank          | 0.000 | 21:05 |
| Control Test       | 0.081 | 21:05 |
| Air Blank          | 0.000 | 21:06 |
| Diagnostics Check  | OK    | 21:06 |

---

Data Download Date/Time: 09/10/2024 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 08/22/2024 02:45:04

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:33 |                    |
| Air Blank         | 0.000 | 02:34 |                    |
| Control Test      | 0.081 | 02:34 |                    |
| Air Blank         | 0.000 | 02:34 |                    |
| Subject Sample    | # 1   | 02:38 | No Sample Provided |
| Breath Volume     | 0.000 | 02:38 |                    |
| Air Blank         | 0.000 | 02:38 |                    |
| Air Blank         | 0.000 | 02:40 |                    |
| Subject Sample    | # 2   | 02:43 | No Sample Provided |
| Breath Volume     | 0.000 | 02:43 |                    |
| Air Blank         | 0.000 | 02:44 |                    |
| Control Test      | 0.081 | 02:44 |                    |
| Air Blank         | 0.000 | 02:44 |                    |
| Diagnostics Check | OK    | 02:45 |                    |

---

Data Download Date/Time: 09/10/2024 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIROUX,ANDREW,R  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: GIROUX,ANDREW  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: TURPEAU

DATE/TIME OF TEST: 08/22/2024 04:59:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:51 |
| Air Blank          | 0.000 | 04:52 |
| Control Test       | 0.081 | 04:52 |
| Air Blank          | 0.000 | 04:52 |
| Subject Sample # 1 | 0.222 | 04:54 |
| Breath Volume      | 1.355 | 04:54 |
| Air Blank          | 0.000 | 04:54 |
| Air Blank          | 0.000 | 04:56 |
| Subject Sample # 2 | 0.224 | 04:57 |
| Breath Volume      | 1.445 | 04:57 |
| Air Blank          | 0.000 | 04:58 |
| Control Test       | 0.079 | 04:58 |
| Air Blank          | 0.000 | 04:59 |
| Diagnostics Check  | OK    | 04:59 |

---

Data Download Date/Time: 09/10/2024 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SALINAS,JOSE,  
BREATH TEST OPERATOR AGENCY: Winter Haven PD  
ARRESTING OFFICER: GRISSOM,TIMOTHY  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: HERNANDEZ RIVERA

DATE/TIME OF TEST: 08/24/2024 01:27:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:19 |
| Air Blank          | 0.000 | 01:19 |
| Control Test       | 0.080 | 01:20 |
| Air Blank          | 0.000 | 01:20 |
| Subject Sample # 1 | 0.122 | 01:22 |
| Breath Volume      | 1.101 | 01:22 |
| Air Blank          | 0.000 | 01:22 |
| Air Blank          | 0.000 | 01:24 |
| Subject Sample # 2 | 0.120 | 01:25 |
| Breath Volume      | 1.296 | 01:25 |
| Air Blank          | 0.000 | 01:26 |
| Control Test       | 0.079 | 01:26 |
| Air Blank          | 0.000 | 01:27 |
| Diagnostics Check  | OK    | 01:27 |

---

Data Download Date/Time: 09/10/2024 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALLISER,SHAUN,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: NASON,TAYLOR  
ARRESTING OFFICER AGENCY: Auburndale PD  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 15:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 08/24/2024 17:36:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:30 |
| Air Blank          | 0.000 | 17:30 |
| Control Test       | 0.081 | 17:30 |
| Air Blank          | 0.000 | 17:31 |
| Subject Sample # 1 | 0.000 | 17:31 |
| Breath Volume      | 2.667 | 17:31 |
| Air Blank          | 0.000 | 17:32 |
| Air Blank          | 0.000 | 17:34 |
| Subject Sample # 2 | 0.000 | 17:34 |
| Breath Volume      | 2.894 | 17:34 |
| Air Blank          | 0.000 | 17:35 |
| Control Test       | 0.080 | 17:35 |
| Air Blank          | 0.000 | 17:36 |
| Diagnostics Check  | OK    | 17:36 |

---

Data Download Date/Time: 09/10/2024 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLIPIN,A,  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: JONES,A  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 15:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/24/2024 21:29:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:22 |
| Air Blank          | 0.000 | 21:22 |
| Control Test       | 0.080 | 21:23 |
| Air Blank          | 0.000 | 21:23 |
| Subject Sample # 1 | 0.000 | 21:24 |
| Breath Volume      | 3.617 | 21:24 |
| Air Blank          | 0.000 | 21:24 |
| Air Blank          | 0.000 | 21:26 |
| Subject Sample # 2 | 0.000 | 21:28 |
| Breath Volume      | 3.023 | 21:28 |
| Air Blank          | 0.000 | 21:28 |
| Control Test       | 0.081 | 21:29 |
| Air Blank          | 0.000 | 21:29 |
| Diagnostics Check  | OK    | 21:29 |

---

Data Download Date/Time: 09/10/2024 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: GRISSOM, TIMOTHY  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: JACQUET

DATE/TIME OF TEST: 08/25/2024 01:14:20

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:02 |                    |
| Air Blank         | 0.000 | 01:03 |                    |
| Control Test      | 0.080 | 01:03 |                    |
| Air Blank         | 0.000 | 01:04 |                    |
| Subject Sample    | # 1   | 01:07 | No Sample Provided |
| Breath Volume     | 0.000 | 01:07 |                    |
| Air Blank         | 0.000 | 01:07 |                    |
| Air Blank         | 0.000 | 01:09 |                    |
| Subject Sample    | # 2   | 01:12 | No Sample Provided |
| Breath Volume     | 0.000 | 01:12 |                    |
| Air Blank         | 0.000 | 01:13 |                    |
| Control Test      | 0.079 | 01:13 |                    |
| Air Blank         | 0.000 | 01:14 |                    |
| Diagnostics Check | OK    | 01:14 |                    |

Data Download Date/Time: 09/10/2024 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLETCHER,MELISSA,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: CARTER,JACOB  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 08/25/2024 02:54:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:48 |
| Air Blank          | 0.000 | 02:48 |
| Control Test       | 0.080 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 1 | 0.052 | 02:50 |
| Breath Volume      | 2.695 | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Air Blank          | 0.000 | 02:52 |
| Subject Sample # 2 | 0.051 | 02:53 |
| Breath Volume      | 2.679 | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Control Test       | 0.079 | 02:54 |
| Air Blank          | 0.000 | 02:54 |
| Diagnostics Check  | OK    | 02:54 |

---

Data Download Date/Time: 09/10/2024 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDRON,DAVID,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WARREN,BILLY  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 08/25/2024 11:12:35

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 11:01 |                    |
| Air Blank         | 0.000 | 11:01 |                    |
| Control Test      | 0.080 | 11:01 |                    |
| Air Blank         | 0.000 | 11:02 |                    |
| Subject Sample    | # 1   | 11:05 | No Sample Provided |
| Breath Volume     | 0.000 | 11:05 |                    |
| Air Blank         | 0.000 | 11:06 |                    |
| Air Blank         | 0.000 | 11:08 |                    |
| Subject Sample    | # 2   | 11:11 | No Sample Provided |
| Breath Volume     | 0.000 | 11:11 |                    |
| Air Blank         | 0.000 | 11:11 |                    |
| Control Test      | 0.081 | 11:11 |                    |
| Air Blank         | 0.000 | 11:12 |                    |
| Diagnostics Check | OK    | 11:12 |                    |

---

Data Download Date/Time: 09/10/2024 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAIR,MARY,K  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ROQUE,ALEXANDER  
ARRESTING OFFICER AGENCY: Bartow PD  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: VELASQUEZ AGUILAR

DATE/TIME OF TEST: 08/26/2024 00:14:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Control Test       | 0.080 | 00:08 |
| Air Blank          | 0.000 | 00:08 |
| Subject Sample # 1 | 0.155 | 00:09 |
| Breath Volume      | 1.332 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Air Blank          | 0.000 | 00:12 |
| Subject Sample # 2 | 0.153 | 00:12 |
| Breath Volume      | 1.222 | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Control Test       | 0.079 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Diagnostics Check  | OK    | 00:14 |

---

Data Download Date/Time: 09/10/2024 09:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIROUX, ANDREW, R  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER, JUSTIN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: BROWN II

DATE/TIME OF TEST: 08/27/2024 23:25:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:19 |
| Air Blank          | 0.000 | 23:19 |
| Control Test       | 0.080 | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Subject Sample # 1 | 0.000 | 23:21 |
| Breath Volume      | 2.011 | 23:21 |
| Air Blank          | 0.000 | 23:21 |
| Air Blank          | 0.000 | 23:23 |
| Subject Sample # 2 | 0.000 | 23:24 |
| Breath Volume      | 1.234 | 23:24 |
| Air Blank          | 0.000 | 23:24 |
| Control Test       | 0.079 | 23:25 |
| Air Blank          | 0.000 | 23:25 |
| Diagnostics Check  | OK    | 23:25 |

---

Data Download Date/Time: 09/10/2024 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARREN,BILLY,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WARREN,BILLY  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: MACENA

DATE/TIME OF TEST: 08/29/2024 00:14:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:07 |
| Air Blank          | 0.000 | 00:07 |
| Control Test       | 0.079 | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Subject Sample # 1 | 0.285 | 00:09 |
| Breath Volume      | 2.378 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Subject Sample # 2 | 0.300 | 00:12 |
| Breath Volume      | 1.574 | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Control Test       | 0.078 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Diagnostics Check  | OK    | 00:14 |

---

Data Download Date/Time: 09/10/2024 09:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JANIAK,M  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: MATTA VASQUEZ

DATE/TIME OF TEST: 08/30/2024 22:04:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:57 |
| Air Blank          | 0.000 | 21:58 |
| Control Test       | 0.080 | 21:58 |
| Air Blank          | 0.000 | 21:59 |
| Subject Sample # 1 | 0.177 | 22:00 |
| Breath Volume      | 1.160 | 22:00 |
| Air Blank          | 0.000 | 22:00 |
| Air Blank          | 0.000 | 22:02 |
| Subject Sample # 2 | 0.175 | 22:02 |
| Breath Volume      | 1.230 | 22:02 |
| Air Blank          | 0.000 | 22:03 |
| Control Test       | 0.080 | 22:03 |
| Air Blank          | 0.000 | 22:04 |
| Diagnostics Check  | OK    | 22:04 |

---

Data Download Date/Time: 09/10/2024 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: KING,P  
ARRESTING OFFICER AGENCY: Lake Alfred PD  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 08/31/2024 03:19:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Control Test       | 0.080 | 03:13 |
| Air Blank          | 0.000 | 03:13 |
| Subject Sample # 1 | 0.206 | 03:14 |
| Breath Volume      | 1.722 | 03:14 |
| Air Blank          | 0.000 | 03:15 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 2 | 0.208 | 03:17 |
| Breath Volume      | 1.503 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Control Test       | 0.080 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Diagnostics Check  | OK    | 03:19 |

---

Data Download Date/Time: 09/10/2024 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: EVANS,KALEIGH  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 05:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: NOVOA-LOPEZ

DATE/TIME OF TEST: 08/31/2024 06:36:54

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 06:34 |            |
| Air Blank         | 0.000 | 06:35 |            |
| Control Test      | 0.081 | 06:35 |            |
| Air Blank         | 0.000 | 06:35 |            |
| Subject Sample    | # 1   | 06:36 | RFI Detect |
| Breath Volume     | 0.000 | 06:36 |            |
| Air Blank         |       | 06:36 | RFI Detect |
| Air Blank         | 0.000 | 06:36 |            |

---

Data Download Date/Time: 09/10/2024 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: EVANS,KALEIGH  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 05:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: NOVOA-LOPEZ

DATE/TIME OF TEST: 08/31/2024 06:50:19

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 06:38 |                    |
| Air Blank         | 0.000 | 06:39 |                    |
| Control Test      | 0.080 | 06:39 |                    |
| Air Blank         | 0.000 | 06:40 |                    |
| Subject Sample    | # 1   | 06:43 | No Sample Provided |
| Breath Volume     | 0.000 | 06:43 |                    |
| Air Blank         | 0.000 | 06:43 |                    |
| Air Blank         | 0.000 | 06:45 |                    |
| Subject Sample    | # 2   | 06:48 | No Sample Provided |
| Breath Volume     | 0.000 | 06:48 |                    |
| Air Blank         | 0.000 | 06:49 |                    |
| Control Test      | 0.081 | 06:49 |                    |
| Air Blank         | 0.000 | 06:50 |                    |
| Diagnostics Check | OK    | 06:50 |                    |

---

Data Download Date/Time: 09/10/2024 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAVEZ,EDSON,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: KING,PHILIP  
ARRESTING OFFICER AGENCY: Lake Alfred PD  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 05:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: LARIOS

DATE/TIME OF TEST: 08/31/2024 07:01:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 06:58 |                      |
| Air Blank         | 0.000 | 06:58 |                      |
| Control Test      | 0.080 | 06:59 |                      |
| Air Blank         | 0.000 | 06:59 |                      |
| Subject Sample    | # 1   | 07:00 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:00 |                      |
| Air Blank         | 0.000 | 07:00 |                      |
| Control Test      | 0.081 | 07:01 |                      |
| Air Blank         | 0.000 | 07:01 |                      |
| Diagnostics Check | OK    | 07:01 |                      |

---

Data Download Date/Time: 09/10/2024 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ELIOTT,JACOB  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 05:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: MANZANO-MELENDZ

DATE/TIME OF TEST: 08/31/2024 08:22:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:15 |
| Air Blank          | 0.000 | 08:16 |
| Control Test       | 0.081 | 08:16 |
| Air Blank          | 0.000 | 08:17 |
| Subject Sample # 1 | 0.133 | 08:18 |
| Breath Volume      | 3.523 | 08:18 |
| Air Blank          | 0.000 | 08:18 |
| Air Blank          | 0.000 | 08:20 |
| Subject Sample # 2 | 0.113 | 08:21 |
| Breath Volume      | 1.734 | 08:21 |
| Air Blank          | 0.000 | 08:21 |
| Control Test       | 0.080 | 08:21 |
| Air Blank          | 0.000 | 08:22 |
| Diagnostics Check  | OK    | 08:22 |

---

Data Download Date/Time: 09/10/2024 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAKELAND,KATLYN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WAKELAND,KATLYN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 08/01/2024 22:45:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Control Test       | 0.079 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Subject Sample # 1 | 0.110 | 22:40 |
| Breath Volume      | 2.371 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 2 | 0.103 | 22:43 |
| Breath Volume      | 2.136 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Control Test       | 0.079 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Diagnostics Check  | OK    | 22:45 |

---

Data Download Date/Time: 08/02/2024 12:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KISTLER, WILLIAM, E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: KISTLER, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 08/09/2024 05:12:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:05 |
| Air Blank          | 0.000 | 05:05 |
| Control Test       | 0.080 | 05:06 |
| Air Blank          | 0.000 | 05:06 |
| Subject Sample # 1 | 0.127 | 05:07 |
| Breath Volume      | 2.089 | 05:07 |
| Air Blank          | 0.000 | 05:08 |
| Air Blank          | 0.000 | 05:10 |
| Subject Sample # 2 | 0.119 | 05:10 |
| Breath Volume      | 1.562 | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.078 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Diagnostics Check  | OK    | 05:12 |

---

Data Download Date/Time: 09/14/2024 12:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAKELAND,KATLYN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WAKELAND,KATLYN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 08/10/2024 00:31:34

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:27 |                      |
| Air Blank         | 0.000 | 00:28 |                      |
| Control Test      | 0.079 | 00:28 |                      |
| Air Blank         | 0.000 | 00:29 |                      |
| Subject Sample    | # 1   | 00:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:30 |                      |
| Air Blank         | 0.000 | 00:30 |                      |
| Control Test      | 0.079 | 00:30 |                      |
| Air Blank         | 0.000 | 00:31 |                      |
| Diagnostics Check | OK    | 00:31 |                      |

---

Data Download Date/Time: 09/14/2024 12:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZIZ,DAVID,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: AZIZ,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GONZALEZ MAZARIEGOS

DATE/TIME OF TEST: 08/11/2024 01:57:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:49 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.079 | 01:50 |
| Air Blank          | 0.000 | 01:50 |
| Subject Sample # 1 | 0.141 | 01:52 |
| Breath Volume      | 1.351 | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Air Blank          | 0.000 | 01:55 |
| Subject Sample # 2 | 0.138 | 01:55 |
| Breath Volume      | 1.300 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.078 | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Diagnostics Check  | OK    | 01:57 |

---

Data Download Date/Time: 09/14/2024 12:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOWAY, GRANT, D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GALLOWAY, GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: REYES AMAYA

DATE/TIME OF TEST: 08/23/2024 02:47:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Control Test       | 0.079 | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Subject Sample # 1 | 0.134 | 02:43 |
| Breath Volume      | 1.875 | 02:43 |
| Air Blank          | 0.000 | 02:43 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 2 | 0.128 | 02:46 |
| Breath Volume      | 1.679 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Control Test       | 0.080 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Diagnostics Check  | OK    | 02:47 |

---

Data Download Date/Time: 09/14/2024 12:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLLINS,THOMAS,J  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: HELLER,JASON  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: NILA AMBROSIO

DATE/TIME OF TEST: 08/23/2024 23:37:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Control Test       | 0.080 | 23:31 |
| Air Blank          | 0.000 | 23:31 |
| Subject Sample # 1 | 0.000 | 23:32 |
| Breath Volume      | 1.808 | 23:32 |
| Air Blank          | 0.000 | 23:33 |
| Air Blank          | 0.000 | 23:35 |
| Subject Sample # 2 | 0.000 | 23:36 |
| Breath Volume      | 2.117 | 23:36 |
| Air Blank          | 0.000 | 23:36 |
| Control Test       | 0.080 | 23:36 |
| Air Blank          | 0.000 | 23:37 |
| Diagnostics Check  | OK    | 23:37 |

---

Data Download Date/Time: 09/14/2024 12:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,WILLIAM,M  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: VITO,COURTNEY  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PELAEZ

DATE/TIME OF TEST: 08/25/2024 04:01:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:58 |                      |
| Air Blank         | 0.000 | 03:58 |                      |
| Control Test      | 0.079 | 03:59 |                      |
| Air Blank         | 0.000 | 03:59 |                      |
| Subject Sample    | # 1   | 03:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:59 |                      |
| Air Blank         | 0.000 | 04:00 |                      |
| Control Test      | 0.079 | 04:00 |                      |
| Air Blank         | 0.000 | 04:01 |                      |
| Diagnostics Check | OK    | 04:01 |                      |

---

Data Download Date/Time: 09/14/2024 12:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZIZ,DAVID,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: AZIZ,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: DEMARCO

DATE/TIME OF TEST: 08/26/2024 02:05:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Control Test       | 0.080 | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 1 | 0.000 | 02:00 |
| Breath Volume      | 2.414 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Air Blank          | 0.000 | 02:03 |
| Subject Sample # 2 | 0.000 | 02:04 |
| Breath Volume      | 1.128 | 02:04 |
| Air Blank          | 0.000 | 02:04 |
| Control Test       | 0.080 | 02:04 |
| Air Blank          | 0.000 | 02:05 |
| Diagnostics Check  | OK    | 02:05 |

---

Data Download Date/Time: 09/14/2024 12:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STARLING,KELLEY,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: STARLING,KELLEY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 08/30/2024 20:10:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:04 |
| Air Blank          | 0.000 | 20:04 |
| Control Test       | 0.079 | 20:05 |
| Air Blank          | 0.000 | 20:05 |
| Subject Sample # 1 | 0.229 | 20:06 |
| Breath Volume      | 3.140 | 20:06 |
| Air Blank          | 0.000 | 20:06 |
| Air Blank          | 0.000 | 20:08 |
| Subject Sample # 2 | 0.226 | 20:09 |
| Breath Volume      | 1.218 | 20:09 |
| Air Blank          | 0.000 | 20:09 |
| Control Test       | 0.078 | 20:10 |
| Air Blank          | 0.000 | 20:10 |
| Diagnostics Check  | OK    | 20:10 |

---

Data Download Date/Time: 09/14/2024 12:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KISTLER, WILLIAM, E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: KISTLER, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A2  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: CARMELLA

DATE/TIME OF TEST: 08/31/2024 21:05:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:55 |
| Air Blank          | 0.000 | 20:55 |
| Control Test       | 0.079 | 20:56 |
| Air Blank          | 0.000 | 20:56 |
| Subject Sample # 1 | 0.279 | 20:57 |
| Breath Volume      | 1.386 | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Air Blank          | 0.000 | 20:59 |
| Subject Sample # 2 | 0.306 | 21:00 |
| Breath Volume      | 2.328 | 21:00 |
| Air Blank          | 0.000 | 21:01 |
| Air Blank          | 0.000 | 21:03 |
| Subject Sample # 3 | 0.289 | 21:03 |
| Breath Volume      | 1.917 | 21:03 |
| Air Blank          | 0.000 | 21:04 |
| Control Test       | 0.078 | 21:04 |
| Air Blank          | 0.000 | 21:05 |
| Diagnostics Check  | OK    | 21:05 |

---

Data Download Date/Time: 09/14/2024 12:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KARNs,BRANDON,  
BREATH TEST OPERATOR AGENCY: Plantation PD  
ARRESTING OFFICER: MAGANA,E  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 07/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231603  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 08/04/2024 22:57:52

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 22:51 |            |
| Air Blank         | 0.000     | 22:52 |            |
| Control Test      | 0.078     | 22:52 |            |
| Air Blank         | 0.000     | 22:53 |            |
| Subject Sample    | # 1 0.310 | 22:53 |            |
| Breath Volume     | 2.710     | 22:53 |            |
| Air Blank         | 0.000     | 22:54 |            |
| Air Blank         | 0.000     | 22:56 |            |
| Subject Sample    | # 2 0.332 | 22:56 |            |
| Breath Volume     | 2.648     | 22:56 |            |
| Air Blank         |           | 22:57 | Purge Fail |
| Air Blank         | 0.000     | 22:57 |            |

---

Data Download Date/Time: 08/30/2024 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KARNS,BRANDON,  
BREATH TEST OPERATOR AGENCY: Plantation PD  
ARRESTING OFFICER: MAGANA,E  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 07/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231603  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 08/04/2024 23:17:14

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 23:03 |                    |
| Air Blank         | 0.000     | 23:03 |                    |
| Control Test      | 0.076     | 23:04 |                    |
| Air Blank         | 0.000     | 23:04 |                    |
| Subject Sample    | # 1 0.325 | 23:05 |                    |
| Breath Volume     | 1.656     | 23:05 |                    |
| Air Blank         | 0.000     | 23:06 |                    |
| Air Blank         | 0.000     | 23:08 |                    |
| Subject Sample    | # 2       | 23:11 | No Sample Provided |
| Breath Volume     | 0.000     | 23:11 |                    |
| Air Blank         | 0.000     | 23:11 |                    |
| Air Blank         | 0.000     | 23:13 |                    |
| Subject Sample    | # 3 0.313 | 23:15 |                    |
| Breath Volume     | 2.492     | 23:15 |                    |
| Air Blank         |           | 23:16 | Purge Fail         |
| Air Blank         | 0.000     | 23:17 |                    |

Data Download Date/Time: 08/30/2024 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

|                                            |                                 |
|--------------------------------------------|---------------------------------|
| INSTRUMENT REGISTERED TO: PLANTATION PD    | Date or Time Changed: No        |
| SOFTWARE VERSION: 8100.27                  | DRY GAS STANDARD LOT #: 231603  |
| BREATH TEST OPERATOR: KARNS,BRANDON,       | DRY GAS EXPIRY DATE: 03/05/2026 |
| BREATH TEST OPERATOR AGENCY: Plantation PD | VIOLATION CODE: DUI             |
| ARRESTING OFFICER: MAGANA,E                | ARREST TIME: 22:04              |
| ARRESTING OFFICER AGENCY: Plantation PD    | GENDER: Male                    |
| LAST AGENCY INSPECTION: 07/30/2024         | OBSERVATION PERIOD BEGAN: 22:27 |

|                               |                               |          |
|-------------------------------|-------------------------------|----------|
| SUBJECT'S LAST NAME: THOMPSON | DATE/TIME OF TEST: 08/04/2024 | 23:53:10 |
|-------------------------------|-------------------------------|----------|

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Control Test       | 0.076 | 23:46 |
| Air Blank          | 0.000 | 23:47 |
| Subject Sample # 1 | 0.286 | 23:47 |
| Breath Volume      | 2.007 | 23:47 |
| Air Blank          | 0.000 | 23:48 |
| Air Blank          | 0.000 | 23:50 |
| Subject Sample # 2 | 0.299 | 23:51 |
| Breath Volume      | 2.464 | 23:51 |
| Air Blank          | 0.000 | 23:52 |
| Control Test       | 0.077 | 23:52 |
| Air Blank          | 0.000 | 23:53 |
| Diagnostics Check  | OK    | 23:53 |

---

Data Download Date/Time: 08/30/2024 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

|                                            |                                 |
|--------------------------------------------|---------------------------------|
| INSTRUMENT REGISTERED TO: PLANTATION PD    | Date or Time Changed: No        |
| SOFTWARE VERSION: 8100.27                  | DRY GAS STANDARD LOT #: 231603  |
| BREATH TEST OPERATOR: KARNS,BRANDON,       | DRY GAS EXPIRY DATE: 03/05/2026 |
| BREATH TEST OPERATOR AGENCY: Plantation PD | VIOLATION CODE: DUI             |
| ARRESTING OFFICER: BROSKIE,J               | ARREST TIME: 21:31              |
| ARRESTING OFFICER AGENCY: Plantation PD    | GENDER: Male                    |
| LAST AGENCY INSPECTION: 07/30/2024         | OBSERVATION PERIOD BEGAN: 22:12 |

|                              |                               |          |
|------------------------------|-------------------------------|----------|
| SUBJECT'S LAST NAME: ROBERTS | DATE/TIME OF TEST: 08/19/2024 | 22:43:02 |
|------------------------------|-------------------------------|----------|

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:36 |
| Air Blank          | 0.000 | 22:36 |
| Control Test       | 0.079 | 22:37 |
| Air Blank          | 0.000 | 22:37 |
| Subject Sample # 1 | 0.224 | 22:38 |
| Breath Volume      | 2.085 | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Subject Sample # 2 | 0.221 | 22:41 |
| Breath Volume      | 2.285 | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Control Test       | 0.078 | 22:42 |
| Air Blank          | 0.000 | 22:42 |
| Diagnostics Check  | OK    | 22:43 |

---

Data Download Date/Time: 08/30/2024 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ABRAHAM,ERIC  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 08/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BOSWELL

DATE/TIME OF TEST: 08/07/2024 23:47:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:41 |
| Air Blank          | 0.000 | 23:41 |
| Control Test       | 0.080 | 23:41 |
| Air Blank          | 0.000 | 23:42 |
| Subject Sample # 1 | 0.161 | 23:43 |
| Breath Volume      | 1.210 | 23:43 |
| Air Blank          | 0.000 | 23:43 |
| Air Blank          | 0.000 | 23:45 |
| Subject Sample # 2 | 0.155 | 23:45 |
| Breath Volume      | 1.820 | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Control Test       | 0.080 | 23:46 |
| Air Blank          | 0.000 | 23:47 |
| Diagnostics Check  | OK    | 23:47 |

---

Data Download Date/Time: 09/05/2024 13:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREEMAN,AUSTIN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: FREEMAN,AUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 08/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 08/08/2024 22:29:00

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:25 |                      |
| Air Blank         | 0.000 | 22:26 |                      |
| Control Test      | 0.079 | 22:26 |                      |
| Air Blank         | 0.000 | 22:27 |                      |
| Subject Sample    | # 1   | 22:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:27 |                      |
| Air Blank         | 0.000 | 22:28 |                      |
| Control Test      | 0.079 | 22:28 |                      |
| Air Blank         | 0.000 | 22:28 |                      |
| Diagnostics Check | OK    | 22:29 |                      |

---

Data Download Date/Time: 09/05/2024 13:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ABRAHAM,ERIC  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 08/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 08/16/2024 01:59:56

|                    |       |       |                      |
|--------------------|-------|-------|----------------------|
| Diagnostics Check  | OK    | 01:50 |                      |
| Air Blank          | 0.000 | 01:51 |                      |
| Control Test       | 0.080 | 01:51 |                      |
| Air Blank          | 0.000 | 01:51 |                      |
| Subject Sample # 1 | 0.174 | 01:55 |                      |
| Breath Volume      | 1.554 | 01:55 |                      |
| Air Blank          | 0.000 | 01:55 |                      |
| Air Blank          | 0.000 | 01:57 |                      |
| Subject Sample # 2 |       | 01:58 | Subject Test Refused |
| Breath Volume      | 0.000 | 01:58 |                      |
| Air Blank          | 0.000 | 01:59 |                      |
|                    |       | 01:59 | RFI Detect           |
| Air Blank          | 0.000 | 01:59 |                      |

Data Download Date/Time: 09/05/2024 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ABRAHAM,ERIC  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 08/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RUIZ JIMENEZ

DATE/TIME OF TEST: 08/16/2024 03:37:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.080 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 1 | 0.147 | 03:32 |
| Breath Volume      | 1.988 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Air Blank          | 0.000 | 03:35 |
| Subject Sample # 2 | 0.148 | 03:35 |
| Breath Volume      | 1.488 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Control Test       | 0.080 | 03:36 |
| Air Blank          | 0.000 | 03:36 |
| Diagnostics Check  | OK    | 03:37 |

---

Data Download Date/Time: 09/05/2024 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ABRAHAM,ERIC  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 08/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MONCADA FLORES

DATE/TIME OF TEST: 08/17/2024 05:21:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:11 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.081 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Subject Sample # 1 | 0.212 | 05:14 |
| Breath Volume      | 1.476 | 05:14 |
| Air Blank          | 0.000 | 05:15 |
| Air Blank          | 0.000 | 05:16 |
| Subject Sample # 2 | 0.249 | 05:17 |
| Breath Volume      | 3.441 | 05:17 |
| Air Blank          | 0.000 | 05:18 |
| Air Blank          | 0.000 | 05:20 |
| Subject Sample # 3 | 0.245 | 05:20 |
| Breath Volume      | 2.355 | 05:20 |
| Air Blank          | 0.000 | 05:21 |
| Control Test       | 0.080 | 05:21 |
| Air Blank          | 0.000 | 05:21 |
| Diagnostics Check  | OK    | 05:21 |

---

Data Download Date/Time: 09/05/2024 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RADCLIFF, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 08/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 08/18/2024 01:48:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.080 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 1 | 0.130 | 01:44 |
| Breath Volume      | 2.003 | 01:44 |
| Air Blank          | 0.000 | 01:44 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 2 | 0.130 | 01:47 |
| Breath Volume      | 3.023 | 01:47 |
| Air Blank          | 0.000 | 01:47 |
| Control Test       | 0.080 | 01:47 |
| Air Blank          | 0.000 | 01:48 |
| Diagnostics Check  | OK    | 01:48 |

---

Data Download Date/Time: 09/05/2024 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ABRAHAM,ERIC,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ELKINS,JUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 08/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 08/24/2024 03:02:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Control Test       | 0.080 | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Subject Sample # 1 | 0.192 | 02:58 |
| Breath Volume      | 1.414 | 02:58 |
| Air Blank          | 0.000 | 02:58 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 2 | 0.184 | 03:01 |
| Breath Volume      | 1.371 | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Control Test       | 0.080 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Diagnostics Check  | OK    | 03:02 |

---

Data Download Date/Time: 09/05/2024 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RADCLIFF, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 08/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2402328118  
DRY GAS EXPIRY DATE: 01/20/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 08/24/2024 04:43:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:36 |
| Air Blank          | 0.000 | 04:37 |
| Control Test       | 0.080 | 04:37 |
| Air Blank          | 0.000 | 04:38 |
| Subject Sample # 1 | 0.133 | 04:39 |
| Breath Volume      | 1.472 | 04:39 |
| Air Blank          | 0.000 | 04:39 |
| Air Blank          | 0.000 | 04:41 |
| Subject Sample # 2 | 0.124 | 04:42 |
| Breath Volume      | 1.273 | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Control Test       | 0.080 | 04:43 |
| Air Blank          | 0.000 | 04:43 |
| Diagnostics Check  | OK    | 04:43 |

---

Data Download Date/Time: 09/05/2024 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HITCHCOCK, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BRUNET

DATE/TIME OF TEST: 08/03/2024 00:54:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:48 |
| Air Blank          | 0.000 | 00:48 |
| Control Test       | 0.081 | 00:49 |
| Air Blank          | 0.000 | 00:49 |
| Subject Sample # 1 | 0.199 | 00:50 |
| Breath Volume      | 1.589 | 00:50 |
| Air Blank          | 0.000 | 00:50 |
| Air Blank          | 0.000 | 00:52 |
| Subject Sample # 2 | 0.198 | 00:53 |
| Breath Volume      | 1.527 | 00:53 |
| Air Blank          | 0.000 | 00:53 |
| Control Test       | 0.081 | 00:54 |
| Air Blank          | 0.000 | 00:54 |
| Diagnostics Check  | OK    | 00:54 |

---

Data Download Date/Time: 08/23/2024 01:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROVENCHER,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: PROVENCHER,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 08/07/2024 20:38:33

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 20:34 |                    |
| Air Blank         | 0.000 | 20:34 |                    |
| Control Test      | 0.081 | 20:35 |                    |
| Air Blank         | 0.000 | 20:35 |                    |
| Subject Sample    | # 1   | 20:37 | Interferent Detect |
| Breath Volume     | 1.492 | 20:37 |                    |
| Air Blank         | 0.000 | 20:37 |                    |
| Control Test      | 0.081 | 20:37 |                    |
| Air Blank         | 0.000 | 20:38 |                    |
| Diagnostics Check | OK    | 20:38 |                    |

---

Data Download Date/Time: 08/23/2024 01:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROVENCHER,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: PROVENCHER,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 08/07/2024 20:44:54

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 20:40 |                    |
| Air Blank         | 0.000 | 20:41 |                    |
| Control Test      | 0.081 | 20:41 |                    |
| Air Blank         | 0.000 | 20:42 |                    |
| Subject Sample    | # 1   | 20:43 | Interferent Detect |
| Breath Volume     | 1.277 | 20:43 |                    |
| Air Blank         | 0.000 | 20:44 |                    |
| Control Test      | 0.081 | 20:44 |                    |
| Air Blank         | 0.000 | 20:44 |                    |
| Diagnostics Check | OK    | 20:44 |                    |

---

Data Download Date/Time: 08/23/2024 01:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NETTLES,JONATHAN,K  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: NETTLES,JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402357111  
DRY GAS EXPIRY DATE: 02/10/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 08/22/2024 16:42:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:36 |
| Air Blank          | 0.000 | 16:36 |
| Control Test       | 0.082 | 16:36 |
| Air Blank          | 0.000 | 16:37 |
| Subject Sample # 1 | 0.082 | 16:37 |
| Breath Volume      | 2.574 | 16:37 |
| Air Blank          | 0.000 | 16:38 |
| Air Blank          | 0.000 | 16:40 |
| Subject Sample # 2 | 0.077 | 16:40 |
| Breath Volume      | 2.878 | 16:40 |
| Air Blank          | 0.000 | 16:41 |
| Control Test       | 0.081 | 16:41 |
| Air Blank          | 0.000 | 16:42 |
| Diagnostics Check  | OK    | 16:42 |

---

Data Download Date/Time: 08/23/2024 01:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILLIPS,LAMAR,D  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: HERRERA,BRAIN  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 08/03/2024 03:18:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:11 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.082 | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Subject Sample # 1 | 0.152 | 03:13 |
| Breath Volume      | 1.929 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 2 | 0.168 | 03:16 |
| Breath Volume      | 2.507 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Control Test       | 0.081 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Diagnostics Check  | OK    | 03:18 |

---

Data Download Date/Time: 08/31/2024 17:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARRIS,MICHAEL,W

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: GRAY,TREVOR

ARRESTING OFFICER AGENCY: OTHER Local

LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 04:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: MAENSIVU

DATE/TIME OF TEST: 08/24/2024 05:40:52

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 05:35 |                      |
| Air Blank         | 0.000     | 05:35 |                      |
| Control Test      | 0.082     | 05:35 |                      |
| Air Blank         | 0.000     | 05:36 |                      |
| Subject Sample    | # 1 0.187 | 05:36 |                      |
| Breath Volume     | 2.414     | 05:36 |                      |
| Air Blank         | 0.000     | 05:37 |                      |
| Air Blank         | 0.000     | 05:39 |                      |
| Subject Sample    | # 2       | 05:39 | Subject Test Refused |
| Breath Volume     | 0.000     | 05:39 |                      |
| Air Blank         | 0.000     | 05:39 |                      |
| Control Test      | 0.081     | 05:40 |                      |
| Air Blank         | 0.000     | 05:40 |                      |
| Diagnostics Check | OK        | 05:40 |                      |

---

Data Download Date/Time: 08/31/2024 17:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOPER, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: GRAHAM, AUSTIN  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: HAGGINS

DATE/TIME OF TEST: 08/25/2024 07:47:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:41 |
| Air Blank          | 0.000 | 07:42 |
| Control Test       | 0.082 | 07:42 |
| Air Blank          | 0.000 | 07:42 |
| Subject Sample # 1 | 0.097 | 07:43 |
| Breath Volume      | 2.246 | 07:43 |
| Air Blank          | 0.000 | 07:43 |
| Air Blank          | 0.000 | 07:45 |
| Subject Sample # 2 | 0.095 | 07:46 |
| Breath Volume      | 2.750 | 07:46 |
| Air Blank          | 0.000 | 07:46 |
| Control Test       | 0.083 | 07:47 |
| Air Blank          | 0.000 | 07:47 |
| Diagnostics Check  | OK    | 07:47 |

---

Data Download Date/Time: 08/31/2024 17:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COOPER, WILLIAM, J

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: DARCY, KAE LIN

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 14:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 08/25/2024 15:22:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:13 |
| Air Blank          | 0.000 | 15:14 |
| Control Test       | 0.082 | 15:14 |
| Air Blank          | 0.000 | 15:15 |
| Subject Sample # 1 | 0.185 | 15:17 |
| Breath Volume      | 2.519 | 15:17 |
| Air Blank          | 0.000 | 15:18 |
| Air Blank          | 0.000 | 15:20 |
| Subject Sample # 2 | 0.190 | 15:20 |
| Breath Volume      | 3.656 | 15:20 |
| Air Blank          | 0.000 | 15:21 |
| Control Test       | 0.082 | 15:21 |
| Air Blank          | 0.000 | 15:22 |
| Diagnostics Check  | OK    | 15:22 |

---

Data Download Date/Time: 08/31/2024 17:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARRIS,MICHAEL,W

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: LOVELESS,PAYTON

ARRESTING OFFICER AGENCY: Deland PD

LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 01:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 08/26/2024 01:45:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:39 |
| Air Blank          | 0.000 | 01:39 |
| Control Test       | 0.082 | 01:40 |
| Air Blank          | 0.000 | 01:40 |
| Subject Sample # 1 | 0.285 | 01:41 |
| Breath Volume      | 2.316 | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 2 | 0.280 | 01:43 |
| Breath Volume      | 1.714 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Control Test       | 0.081 | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Diagnostics Check  | OK    | 01:45 |

---

Data Download Date/Time: 08/31/2024 17:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,KRISTOPHER,B  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HALL,KRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402372586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: BELANDRIA SOSA

DATE/TIME OF TEST: 08/11/2024 05:26:17

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 05:11 |                |
| Air Blank         | 0.000     | 05:12 |                |
| Control Test      | 0.082     | 05:12 |                |
| Air Blank         | 0.000     | 05:12 |                |
| Subject Sample    | # 1 0.115 | 05:16 | Volume Not Met |
| Breath Volume     | 0.289     | 05:16 |                |
| Air Blank         | 0.000     | 05:16 |                |
| Air Blank         | 0.000     | 05:18 |                |
| Subject Sample    | # 2 0.109 | 05:20 |                |
| Breath Volume     | 1.277     | 05:20 |                |
| Air Blank         | 0.000     | 05:21 |                |
| Air Blank         | 0.000     | 05:23 |                |
| Subject Sample    | # 3 0.101 | 05:24 |                |
| Breath Volume     | 1.992     | 05:24 |                |
| Air Blank         | 0.000     | 05:25 |                |
| Control Test      | 0.081     | 05:25 |                |
| Air Blank         | 0.000     | 05:26 |                |
| Diagnostics Check | OK        | 05:26 |                |

---

Data Download Date/Time: 09/03/2024 09:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,ZECHARIAH,C  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: TYNDAL,D  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402372586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 10:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/14/2024 11:32:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:23 |
| Air Blank          | 0.000 | 11:24 |
| Control Test       | 0.082 | 11:24 |
| Air Blank          | 0.000 | 11:25 |
| Subject Sample # 1 | 0.255 | 11:26 |
| Breath Volume      | 1.316 | 11:26 |
| Air Blank          | 0.000 | 11:27 |
| Air Blank          | 0.000 | 11:29 |
| Subject Sample # 2 | 0.263 | 11:30 |
| Breath Volume      | 2.156 | 11:30 |
| Air Blank          | 0.000 | 11:31 |
| Control Test       | 0.081 | 11:31 |
| Air Blank          | 0.000 | 11:32 |
| Diagnostics Check  | OK    | 11:32 |

---

Data Download Date/Time: 09/03/2024 09:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,KRISTOPHER,B  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HALL,KRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402372586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:09

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 08/18/2024 06:43:08

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 06:34 |            |
| Air Blank         | 0.000     | 06:34 |            |
| Control Test      | 0.082     | 06:35 |            |
| Air Blank         | 0.000     | 06:35 |            |
| Subject Sample    | # 1 0.162 | 06:36 |            |
| Breath Volume     | 1.289     | 06:36 |            |
| Air Blank         | 0.000     | 06:37 |            |
| Air Blank         | 0.000     | 06:39 |            |
| Subject Sample    | # 2 0.160 | 06:42 |            |
| Breath Volume     | 1.402     | 06:42 |            |
| Air Blank         |           | 06:42 | RFI Detect |
| Air Blank         | 0.000     | 06:43 |            |

---

Data Download Date/Time: 09/03/2024 09:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,KRISTOPHER,B  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HALL,KRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402372586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: SLIKER

DATE/TIME OF TEST: 08/24/2024 01:49:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.081 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 1 | 0.188 | 01:44 |
| Breath Volume      | 2.578 | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Air Blank          | 0.000 | 01:47 |
| Subject Sample # 2 | 0.189 | 01:47 |
| Breath Volume      | 2.894 | 01:47 |
| Air Blank          | 0.000 | 01:48 |
| Control Test       | 0.081 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Diagnostics Check  | OK    | 01:49 |

---

Data Download Date/Time: 09/03/2024 09:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDREWS, TYREE, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: ELLIOTT, JACOB  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: CORTES-CUEVAS

DATE/TIME OF TEST: 08/02/2024 21:32:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:24 |
| Air Blank          | 0.000 | 21:25 |
| Control Test       | 0.081 | 21:25 |
| Air Blank          | 0.000 | 21:26 |
| Subject Sample # 1 | 0.045 | 21:27 |
| Breath Volume      | 1.230 | 21:27 |
| Air Blank          | 0.000 | 21:28 |
| Air Blank          | 0.000 | 21:30 |
| Subject Sample # 2 | 0.043 | 21:31 |
| Breath Volume      | 1.464 | 21:31 |
| Air Blank          | 0.000 | 21:31 |
| Control Test       | 0.079 | 21:32 |
| Air Blank          | 0.000 | 21:32 |
| Diagnostics Check  | OK    | 21:32 |

---

Data Download Date/Time: 08/19/2024 12:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIROUX,ANDREW,R  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: KING,PHILIP  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: BUZARD

DATE/TIME OF TEST: 08/02/2024 23:11:58

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:57 |                |
| Air Blank         | 0.000     | 22:58 |                |
| Control Test      | 0.081     | 22:58 |                |
| Air Blank         | 0.000     | 22:58 |                |
| Subject Sample    | # 1 0.165 | 23:02 | Volume Not Met |
| Breath Volume     | 0.531     | 23:02 |                |
| Air Blank         | 0.000     | 23:02 |                |
| Air Blank         | 0.000     | 23:04 |                |
| Subject Sample    | # 2 0.188 | 23:06 |                |
| Breath Volume     | 1.410     | 23:06 |                |
| Air Blank         | 0.000     | 23:06 |                |
| Air Blank         | 0.000     | 23:08 |                |
| Subject Sample    | # 3 0.199 | 23:10 |                |
| Breath Volume     | 1.660     | 23:10 |                |
| Air Blank         | 0.000     | 23:11 |                |
| Control Test      | 0.080     | 23:11 |                |
| Air Blank         | 0.000     | 23:11 |                |
| Diagnostics Check | OK        | 23:11 |                |

Data Download Date/Time: 08/19/2024 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOFFORD,CLINT,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JONES,ARIK  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 15:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: ZUST

DATE/TIME OF TEST: 08/04/2024 17:30:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:24 |
| Air Blank          | 0.000 | 17:24 |
| Control Test       | 0.079 | 17:25 |
| Air Blank          | 0.000 | 17:25 |
| Subject Sample # 1 | 0.295 | 17:26 |
| Breath Volume      | 2.578 | 17:26 |
| Air Blank          | 0.000 | 17:26 |
| Air Blank          | 0.000 | 17:28 |
| Subject Sample # 2 | 0.301 | 17:29 |
| Breath Volume      | 2.644 | 17:29 |
| Air Blank          | 0.000 | 17:30 |
| Control Test       | 0.078 | 17:30 |
| Air Blank          | 0.000 | 17:30 |
| Diagnostics Check  | OK    | 17:30 |

---

Data Download Date/Time: 08/19/2024 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDRON,DAVID,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WRIGHT,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/06/2024 06:17:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:09 |
| Air Blank          | 0.000 | 06:09 |
| Control Test       | 0.080 | 06:10 |
| Air Blank          | 0.000 | 06:10 |
| Subject Sample # 1 | 0.164 | 06:12 |
| Breath Volume      | 1.574 | 06:12 |
| Air Blank          | 0.000 | 06:13 |
| Air Blank          | 0.000 | 06:15 |
| Subject Sample # 2 | 0.166 | 06:16 |
| Breath Volume      | 3.781 | 06:16 |
| Air Blank          | 0.000 | 06:16 |
| Control Test       | 0.080 | 06:17 |
| Air Blank          | 0.000 | 06:17 |
| Diagnostics Check  | OK    | 06:17 |

---

Data Download Date/Time: 08/19/2024 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 07:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: HURD

DATE/TIME OF TEST: 08/07/2024 09:13:26

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 09:02 |                    |
| Air Blank         | 0.000 | 09:02 |                    |
| Control Test      | 0.080 | 09:02 |                    |
| Air Blank         | 0.000 | 09:03 |                    |
| Subject Sample    | # 1   | 09:06 | No Sample Provided |
| Breath Volume     | 0.000 | 09:06 |                    |
| Air Blank         | 0.000 | 09:06 |                    |
| Air Blank         | 0.000 | 09:08 |                    |
| Subject Sample    | # 2   | 09:12 | No Sample Provided |
| Breath Volume     | 0.000 | 09:12 |                    |
| Air Blank         | 0.000 | 09:12 |                    |
| Control Test      | 0.081 | 09:12 |                    |
| Air Blank         | 0.000 | 09:13 |                    |
| Diagnostics Check | OK    | 09:13 |                    |

---

Data Download Date/Time: 08/19/2024 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDREWS, TYREE, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JANIAK, MALIKHI  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: SANON

DATE/TIME OF TEST: 08/09/2024 00:51:25

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 00:40 |                    |
| Air Blank         | 0.000 | 00:40 |                    |
| Control Test      | 0.081 | 00:40 |                    |
| Air Blank         | 0.000 | 00:41 |                    |
| Subject Sample    | # 1   | 00:44 | No Sample Provided |
| Breath Volume     | 0.000 | 00:44 |                    |
| Air Blank         | 0.000 | 00:44 |                    |
| Air Blank         | 0.000 | 00:46 |                    |
| Subject Sample    | # 2   | 00:50 | No Sample Provided |
| Breath Volume     | 0.000 | 00:50 |                    |
| Air Blank         | 0.000 | 00:50 |                    |
| Control Test      | 0.080 | 00:50 |                    |
| Air Blank         | 0.000 | 00:51 |                    |
| Diagnostics Check | OK    | 00:51 |                    |

---

Data Download Date/Time: 08/19/2024 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLETCHER,MELISSA,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: DENMARK,HAROLD  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: ZIMMER

DATE/TIME OF TEST: 08/10/2024 20:40:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:33 |
| Air Blank          | 0.000 | 20:34 |
| Control Test       | 0.080 | 20:34 |
| Air Blank          | 0.000 | 20:34 |
| Subject Sample # 1 | 0.176 | 20:35 |
| Breath Volume      | 2.351 | 20:35 |
| Air Blank          | 0.000 | 20:36 |
| Air Blank          | 0.000 | 20:38 |
| Subject Sample # 2 | 0.171 | 20:38 |
| Breath Volume      | 2.312 | 20:38 |
| Air Blank          | 0.000 | 20:39 |
| Control Test       | 0.080 | 20:39 |
| Air Blank          | 0.000 | 20:39 |
| Diagnostics Check  | OK    | 20:40 |

---

Data Download Date/Time: 08/19/2024 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTO,MAVERICK,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ADIR,MARY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MATOS HANCE

DATE/TIME OF TEST: 08/11/2024 01:57:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:50 |
| Air Blank          | 0.000 | 01:51 |
| Control Test       | 0.080 | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Subject Sample # 1 | 0.202 | 01:52 |
| Breath Volume      | 1.449 | 01:52 |
| Air Blank          | 0.000 | 01:52 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 2 | 0.203 | 01:55 |
| Breath Volume      | 2.003 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.079 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Diagnostics Check  | OK    | 01:57 |

---

Data Download Date/Time: 08/19/2024 12:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTO,MAVERICK,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SOTO,MAVERICK  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 08/11/2024 04:56:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:48 |
| Air Blank          | 0.000 | 04:48 |
| Control Test       | 0.081 | 04:49 |
| Air Blank          | 0.000 | 04:49 |
| Subject Sample # 1 | 0.195 | 04:51 |
| Breath Volume      | 2.371 | 04:51 |
| Air Blank          | 0.000 | 04:52 |
| Air Blank          | 0.000 | 04:53 |
| Subject Sample # 2 | 0.191 | 04:54 |
| Breath Volume      | 2.339 | 04:54 |
| Air Blank          | 0.000 | 04:55 |
| Control Test       | 0.079 | 04:55 |
| Air Blank          | 0.000 | 04:56 |
| Diagnostics Check  | OK    | 04:56 |

---

Data Download Date/Time: 08/19/2024 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,MARCO,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: RODRIGUEZ,TRISTAN  
ARRESTING OFFICER AGENCY: Lake Wales PD  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 08:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: TOULOUTE

DATE/TIME OF TEST: 08/11/2024 09:41:01

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 09:29 |                    |
| Air Blank         | 0.000 | 09:30 |                    |
| Control Test      | 0.080 | 09:30 |                    |
| Air Blank         | 0.000 | 09:30 |                    |
| Subject Sample    | # 1   | 09:34 | No Sample Provided |
| Breath Volume     | 0.000 | 09:34 |                    |
| Air Blank         | 0.000 | 09:34 |                    |
| Air Blank         | 0.000 | 09:36 |                    |
| Subject Sample    | # 2   | 09:39 | No Sample Provided |
| Breath Volume     | 0.000 | 09:39 |                    |
| Air Blank         | 0.000 | 09:40 |                    |
| Control Test      | 0.080 | 09:40 |                    |
| Air Blank         | 0.000 | 09:40 |                    |
| Diagnostics Check | OK    | 09:41 |                    |

---

Data Download Date/Time: 08/19/2024 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALLISER,SHAUN,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WALLISER,SHAUN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/11/2024 17:23:46

---

|                   |           |       |              |
|-------------------|-----------|-------|--------------|
| Diagnostics Check | OK        | 17:18 |              |
| Air Blank         | 0.000     | 17:19 |              |
| Control Test      | 0.081     | 17:19 |              |
| Air Blank         | 0.000     | 17:20 |              |
| Subject Sample    | # 1 0.051 | 17:20 |              |
| Breath Volume     | 2.093     | 17:20 |              |
| Air Blank         | 0.000     | 17:21 |              |
| Air Blank         |           | 17:23 | Ambient Fail |
| Air Blank         | 0.000     | 17:23 |              |

---

Data Download Date/Time: 08/19/2024 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALLISER,SHAUN,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WALLISER,SHAUN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/11/2024 17:33:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:26 |
| Air Blank          | 0.000 | 17:27 |
| Control Test       | 0.080 | 17:27 |
| Air Blank          | 0.000 | 17:28 |
| Subject Sample # 1 | 0.047 | 17:28 |
| Breath Volume      | 4.070 | 17:28 |
| Air Blank          | 0.000 | 17:29 |
| Air Blank          | 0.000 | 17:31 |
| Subject Sample # 2 | 0.047 | 17:31 |
| Breath Volume      | 3.761 | 17:31 |
| Air Blank          | 0.000 | 17:32 |
| Control Test       | 0.080 | 17:32 |
| Air Blank          | 0.000 | 17:33 |
| Diagnostics Check  | OK    | 17:33 |

---

Data Download Date/Time: 08/19/2024 12:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Auburndale PD  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Auburndale PD  
LAST AGENCY INSPECTION: 07/09/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 18:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: DELAPAZ SANCHEZ

DATE/TIME OF TEST: 08/11/2024 19:47:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:44 |                      |
| Air Blank         | 0.000 | 19:45 |                      |
| Control Test      | 0.081 | 19:45 |                      |
| Air Blank         | 0.000 | 19:45 |                      |
| Subject Sample    | # 1   | 19:46 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:46 |                      |
| Air Blank         | 0.000 | 19:46 |                      |
| Control Test      | 0.081 | 19:47 |                      |
| Air Blank         | 0.000 | 19:47 |                      |
| Diagnostics Check | OK    | 19:47 |                      |

---

Data Download Date/Time: 08/19/2024 12:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAIR,MARY,K  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: JANIAK,MALILCHI  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: ISEME

DATE/TIME OF TEST: 08/17/2024 05:29:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:23 |
| Air Blank          | 0.000 | 05:23 |
| Control Test       | 0.081 | 05:24 |
| Air Blank          | 0.000 | 05:24 |
| Subject Sample # 1 | 0.000 | 05:25 |
| Breath Volume      | 1.851 | 05:25 |
| Air Blank          | 0.000 | 05:25 |
| Air Blank          | 0.000 | 05:27 |
| Subject Sample # 2 | 0.000 | 05:28 |
| Breath Volume      | 2.000 | 05:28 |
| Air Blank          | 0.000 | 05:28 |
| Control Test       | 0.081 | 05:29 |
| Air Blank          | 0.000 | 05:29 |
| Diagnostics Check  | OK    | 05:29 |

---

Data Download Date/Time: 08/19/2024 12:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAIR,MARY,K  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ADAIR,MARY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: SALES MARTIN

DATE/TIME OF TEST: 08/17/2024 05:49:51

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 05:37 |                |
| Air Blank         | 0.000     | 05:37 |                |
| Control Test      | 0.081     | 05:38 |                |
| Air Blank         | 0.000     | 05:38 |                |
| Subject Sample    | # 1 0.177 | 05:41 | Volume Not Met |
| Breath Volume     | 0.265     | 05:41 |                |
| Air Blank         | 0.000     | 05:42 |                |
| Air Blank         | 0.000     | 05:44 |                |
| Subject Sample    | # 2 0.201 | 05:44 |                |
| Breath Volume     | 1.460     | 05:44 |                |
| Air Blank         | 0.000     | 05:45 |                |
| Air Blank         | 0.000     | 05:47 |                |
| Subject Sample    | # 3 0.182 | 05:48 |                |
| Breath Volume     | 1.136     | 05:48 |                |
| Air Blank         | 0.000     | 05:48 |                |
| Control Test      | 0.080     | 05:49 |                |
| Air Blank         | 0.000     | 05:49 |                |
| Diagnostics Check | OK        | 05:49 |                |

---

Data Download Date/Time: 08/19/2024 12:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOFFORD,CLINT,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WOFFORD,CLINT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 11:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 08/17/2024 12:19:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:11 |
| Air Blank          | 0.000 | 12:12 |
| Control Test       | 0.081 | 12:12 |
| Air Blank          | 0.000 | 12:12 |
| Subject Sample # 1 | 0.268 | 12:13 |
| Breath Volume      | 2.175 | 12:13 |
| Air Blank          | 0.000 | 12:14 |
| Air Blank          | 0.000 | 12:16 |
| Subject Sample # 2 | 0.252 | 12:17 |
| Breath Volume      | 1.527 | 12:17 |
| Air Blank          | 0.000 | 12:18 |
| Control Test       | 0.080 | 12:18 |
| Air Blank          | 0.000 | 12:19 |
| Diagnostics Check  | OK    | 12:19 |

---

Data Download Date/Time: 08/19/2024 12:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 09:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:06

SUBJECT'S LAST NAME: ARANGO

DATE/TIME OF TEST: 08/22/2024 10:33:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:27 |
| Air Blank          | 0.000 | 10:27 |
| Control Test       | 0.080 | 10:28 |
| Air Blank          | 0.000 | 10:28 |
| Subject Sample # 1 | 0.000 | 10:29 |
| Breath Volume      | 1.359 | 10:29 |
| Air Blank          | 0.000 | 10:29 |
| Air Blank          | 0.000 | 10:31 |
| Subject Sample # 2 | 0.000 | 10:32 |
| Breath Volume      | 1.582 | 10:32 |
| Air Blank          | 0.000 | 10:32 |
| Control Test       | 0.080 | 10:33 |
| Air Blank          | 0.000 | 10:33 |
| Diagnostics Check  | OK    | 10:33 |

---

Data Download Date/Time: 09/10/2024 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUCAS,NATHANIEL,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: LUCAS,NATHANIEL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: LYNN

DATE/TIME OF TEST: 08/23/2024 20:32:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:26 |
| Air Blank          | 0.000 | 20:26 |
| Control Test       | 0.080 | 20:26 |
| Air Blank          | 0.000 | 20:27 |
| Subject Sample # 1 | 0.172 | 20:27 |
| Breath Volume      | 1.464 | 20:27 |
| Air Blank          | 0.000 | 20:28 |
| Air Blank          | 0.000 | 20:30 |
| Subject Sample # 2 | 0.175 | 20:30 |
| Breath Volume      | 1.546 | 20:30 |
| Air Blank          | 0.000 | 20:31 |
| Control Test       | 0.080 | 20:31 |
| Air Blank          | 0.000 | 20:32 |
| Diagnostics Check  | OK    | 20:32 |

---

Data Download Date/Time: 09/10/2024 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALLISER,SHAUN,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: NASON,TAYLOR  
ARRESTING OFFICER AGENCY: Auburndale PD  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 15:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 08/24/2024 17:22:30

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 17:11 |                    |
| Air Blank         | 0.000 | 17:11 |                    |
| Control Test      | 0.081 | 17:11 |                    |
| Air Blank         | 0.000 | 17:12 |                    |
| Subject Sample    | # 1   | 17:15 | No Sample Provided |
| Breath Volume     | 0.000 | 17:15 |                    |
| Air Blank         | 0.000 | 17:16 |                    |
| Air Blank         | 0.000 | 17:18 |                    |
| Subject Sample    | # 2   | 17:21 | Volume Not Met     |
| Breath Volume     | 0.015 | 17:21 |                    |
| Air Blank         | 0.000 | 17:21 |                    |
| Control Test      | 0.082 | 17:21 |                    |
| Air Blank         | 0.000 | 17:22 |                    |
| Diagnostics Check | OK    | 17:22 |                    |

Data Download Date/Time: 09/10/2024 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUCAS,NATHANIEL,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: LUCAS,NATHANIEL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 08/26/2024 17:17:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:11 |
| Air Blank          | 0.000 | 17:12 |
| Control Test       | 0.080 | 17:12 |
| Air Blank          | 0.000 | 17:12 |
| Subject Sample # 1 | 0.207 | 17:13 |
| Breath Volume      | 1.308 | 17:13 |
| Air Blank          | 0.000 | 17:14 |
| Air Blank          | 0.000 | 17:15 |
| Subject Sample # 2 | 0.222 | 17:16 |
| Breath Volume      | 2.261 | 17:16 |
| Air Blank          | 0.000 | 17:17 |
| Control Test       | 0.078 | 17:17 |
| Air Blank          | 0.000 | 17:17 |
| Diagnostics Check  | OK    | 17:17 |

---

Data Download Date/Time: 09/10/2024 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER,JUSTIN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: REGISTER,J  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: GARCIA LEIVA

DATE/TIME OF TEST: 08/26/2024 22:21:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:14 |
| Air Blank          | 0.000 | 22:14 |
| Control Test       | 0.081 | 22:14 |
| Air Blank          | 0.000 | 22:15 |
| Subject Sample # 1 | 0.000 | 22:17 |
| Breath Volume      | 1.292 | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Air Blank          | 0.000 | 22:19 |
| Subject Sample # 2 | 0.000 | 22:20 |
| Breath Volume      | 1.695 | 22:20 |
| Air Blank          | 0.000 | 22:20 |
| Control Test       | 0.081 | 22:21 |
| Air Blank          | 0.000 | 22:21 |
| Diagnostics Check  | OK    | 22:21 |

---

Data Download Date/Time: 09/10/2024 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALLISER,SHAUN,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: MARTIN,SAMUEL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 08/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 13:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:34

SUBJECT'S LAST NAME: REIS

DATE/TIME OF TEST: 08/28/2024 15:15:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:08 |
| Air Blank          | 0.000 | 15:09 |
| Control Test       | 0.080 | 15:09 |
| Air Blank          | 0.000 | 15:10 |
| Subject Sample # 1 | 0.269 | 15:10 |
| Breath Volume      | 3.058 | 15:10 |
| Air Blank          | 0.000 | 15:11 |
| Air Blank          | 0.000 | 15:13 |
| Subject Sample # 2 | 0.267 | 15:13 |
| Breath Volume      | 2.937 | 15:13 |
| Air Blank          | 0.000 | 15:14 |
| Control Test       | 0.079 | 15:14 |
| Air Blank          | 0.000 | 15:15 |
| Diagnostics Check  | OK    | 15:15 |

---

Data Download Date/Time: 09/10/2024 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRUMMOND,ALEX,V  
BREATH TEST OPERATOR AGENCY: Putnam County SO  
ARRESTING OFFICER: PLEIMAN,ERIC  
ARRESTING OFFICER AGENCY: Putnam County SO  
LAST AGENCY INSPECTION: 07/23/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26422080A1  
DRY GAS EXPIRY DATE: 11/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/16/2024 17:06:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:59 |
| Air Blank          | 0.000 | 17:00 |
| Control Test       | 0.079 | 17:00 |
| Air Blank          | 0.000 | 17:01 |
| Subject Sample # 1 | 0.008 | 17:02 |
| Breath Volume      | 3.152 | 17:02 |
| Air Blank          | 0.000 | 17:02 |
| Air Blank          | 0.000 | 17:04 |
| Subject Sample # 2 | 0.008 | 17:04 |
| Breath Volume      | 2.859 | 17:04 |
| Air Blank          | 0.000 | 17:05 |
| Control Test       | 0.081 | 17:05 |
| Air Blank          | 0.000 | 17:06 |
| Diagnostics Check  | OK    | 17:06 |

---

Data Download Date/Time: 08/30/2024 19:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,JOSE,L  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: RUSHWORTH,JOSEPH  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 4022872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 08/11/2024 22:38:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:31 |
| Air Blank          | 0.000 | 22:31 |
| Control Test       | 0.081 | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Subject Sample # 1 | 0.087 | 22:33 |
| Breath Volume      | 2.058 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 2 | 0.088 | 22:36 |
| Breath Volume      | 2.152 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Control Test       | 0.082 | 22:37 |
| Air Blank          | 0.000 | 22:38 |
| Diagnostics Check  | OK    | 22:38 |

---

Data Download Date/Time: 08/28/2024 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORGAN,MICHAEL,B  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: VARGAS,NATHAN  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 4022872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 07:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:04

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 08/12/2024 08:35:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:29 |
| Air Blank          | 0.000 | 08:30 |
| Control Test       | 0.081 | 08:30 |
| Air Blank          | 0.000 | 08:31 |
| Subject Sample # 1 | 0.143 | 08:31 |
| Breath Volume      | 1.492 | 08:31 |
| Air Blank          | 0.000 | 08:32 |
| Air Blank          | 0.000 | 08:34 |
| Subject Sample # 2 | 0.144 | 08:34 |
| Breath Volume      | 1.265 | 08:34 |
| Air Blank          | 0.000 | 08:35 |
| Control Test       | 0.081 | 08:35 |
| Air Blank          | 0.000 | 08:35 |
| Diagnostics Check  | OK    | 08:35 |

---

Data Download Date/Time: 08/28/2024 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTISTA,BRITTNEY,L  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: THOMAS,ETHAN  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 4022872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: WORLEY

DATE/TIME OF TEST: 08/13/2024 00:50:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Control Test       | 0.080 | 00:45 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 1 | 0.139 | 00:46 |
| Breath Volume      | 1.308 | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Air Blank          | 0.000 | 00:48 |
| Subject Sample # 2 | 0.139 | 00:48 |
| Breath Volume      | 1.324 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Control Test       | 0.080 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Diagnostics Check  | OK    | 00:50 |

---

Data Download Date/Time: 08/28/2024 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,JOSE,L  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: RIVERA,JOSE  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 4022872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 08/14/2024 23:36:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:30 |
| Air Blank          | 0.000 | 23:30 |
| Control Test       | 0.081 | 23:31 |
| Air Blank          | 0.000 | 23:31 |
| Subject Sample # 1 | 0.154 | 23:32 |
| Breath Volume      | 2.257 | 23:32 |
| Air Blank          | 0.000 | 23:32 |
| Air Blank          | 0.000 | 23:34 |
| Subject Sample # 2 | 0.154 | 23:35 |
| Breath Volume      | 2.523 | 23:35 |
| Air Blank          | 0.000 | 23:35 |
| Control Test       | 0.080 | 23:36 |
| Air Blank          | 0.000 | 23:36 |
| Diagnostics Check  | OK    | 23:36 |

---

Data Download Date/Time: 08/28/2024 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOODIN,STEPHEN,D  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: WOODIN,STEPHEN  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 07/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 4022872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 15:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/15/2024 15:55:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:48 |
| Air Blank          | 0.000 | 15:48 |
| Control Test       | 0.081 | 15:49 |
| Air Blank          | 0.000 | 15:49 |
| Subject Sample # 1 | 0.138 | 15:51 |
| Breath Volume      | 2.699 | 15:51 |
| Air Blank          | 0.000 | 15:51 |
| Air Blank          | 0.000 | 15:53 |
| Subject Sample # 2 | 0.145 | 15:54 |
| Breath Volume      | 2.906 | 15:54 |
| Air Blank          | 0.000 | 15:54 |
| Control Test       | 0.079 | 15:55 |
| Air Blank          | 0.000 | 15:55 |
| Diagnostics Check  | OK    | 15:55 |

---

Data Download Date/Time: 08/28/2024 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: MALETTO,NICHOLAS  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 08/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 4022872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: SANTIAGO GARCIA

DATE/TIME OF TEST: 08/30/2024 23:04:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:02 |                      |
| Air Blank         | 0.000 | 23:02 |                      |
| Control Test      | 0.082 | 23:02 |                      |
| Air Blank         | 0.000 | 23:03 |                      |
| Subject Sample    | # 1   | 23:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:03 |                      |
| Air Blank         | 0.000 | 23:04 |                      |
| Control Test      | 0.081 | 23:04 |                      |
| Air Blank         | 0.000 | 23:04 |                      |
| Diagnostics Check | OK    | 23:04 |                      |

---

Data Download Date/Time: 09/24/2024 16:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: MALETTO,NICHOLAS  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 08/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 4022872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: LOSTAUNAU

DATE/TIME OF TEST: 08/31/2024 03:48:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:42 |
| Air Blank          | 0.000 | 03:42 |
| Control Test       | 0.080 | 03:43 |
| Air Blank          | 0.000 | 03:43 |
| Subject Sample # 1 | 0.193 | 03:43 |
| Breath Volume      | 2.363 | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Air Blank          | 0.000 | 03:46 |
| Subject Sample # 2 | 0.191 | 03:46 |
| Breath Volume      | 1.425 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Control Test       | 0.081 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Diagnostics Check  | OK    | 03:48 |

---

Data Download Date/Time: 09/24/2024 16:13