

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MODEST

DATE/TIME OF TEST: 03/01/2025 04:26:53

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:15	
Control Test		0.079	04:16	
Air Blank		0.000	04:16	
Subject Sample	# 1		04:19	No Sample Provided
Breath Volume		0.000	04:19	
Air Blank		0.000	04:20	
Air Blank		0.000	04:22	
Subject Sample	# 2	0.104	04:22	
Breath Volume		1.695	04:22	
Air Blank		0.000	04:23	
Air Blank		0.000	04:25	
Subject Sample	# 3	0.103	04:25	
Breath Volume		1.757	04:25	
Air Blank		0.000	04:25	
Control Test		0.078	04:26	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 03/13/2025 01:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 03/02/2025 00:43:28

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.139	00:39
Breath Volume	1.714	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.144	00:41
Breath Volume	1.605	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 03/13/2025 01:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TAYLOR,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: PLONTNYY

DATE/TIME OF TEST: 03/03/2025 17:21:57

Diagnostics Check	OK	17:15
Air Blank	0.000	17:16
Control Test	0.079	17:16
Air Blank	0.000	17:17
Subject Sample # 1	0.261	17:17
Breath Volume	2.113	17:17
Air Blank	0.000	17:18
Air Blank	0.000	17:19
Subject Sample # 2	0.243	17:20
Breath Volume	3.394	17:20
Air Blank	0.000	17:21
Control Test	0.077	17:21
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

Data Download Date/Time: 03/13/2025 01:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BORIS GONZALEZ

DATE/TIME OF TEST: 03/04/2025 03:27:11

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.064	03:22
Breath Volume	2.578	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.064	03:25
Breath Volume	2.363	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 03/13/2025 01:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,M
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARIN SILVA

DATE/TIME OF TEST: 03/08/2025 03:47:35

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.048	03:42
Breath Volume	1.542	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.051	03:46
Breath Volume	2.117	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 03/13/2025 01:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPE,RYAN,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LAPE,RYAN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: NICOLETTI

DATE/TIME OF TEST: 03/08/2025 22:58:57

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:50	
Control Test	0.079	22:51	
Air Blank	0.000	22:51	
Subject Sample	# 1	22:54	No Sample Provided
Breath Volume	0.000	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample	# 2	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:58	
Control Test	0.079	22:58	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 03/13/2025 01:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: HANNA

DATE/TIME OF TEST: 03/09/2025 01:04:10

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.078	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 03/13/2025 01:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY, MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY, MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: GARCIA CARDENAS

DATE/TIME OF TEST: 03/09/2025 07:46:07

Diagnostics Check	OK	07:39
Air Blank	0.000	07:40
Control Test	0.079	07:40
Air Blank	0.000	07:41
Subject Sample # 1	0.242	07:41
Breath Volume	2.378	07:41
Air Blank	0.000	07:42
Air Blank	0.000	07:44
Subject Sample # 2	0.225	07:44
Breath Volume	1.351	07:44
Air Blank	0.000	07:45
Control Test	0.079	07:45
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

Data Download Date/Time: 03/13/2025 01:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DOUGLAS,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BARQUERO,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: CASTINEYRA

DATE/TIME OF TEST: 03/11/2025 22:14:09

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.079	22:06
Air Blank		0.000	22:07
Subject Sample	# 1	0.218	22:08
Breath Volume		2.085	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:10
Subject Sample	# 2	0.207	22:12
Breath Volume		1.460	22:12
Air Blank		0.000	22:13
Control Test		0.079	22:13
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 03/13/2025 01:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 03/14/2025 04:33:11

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.078	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.180	04:28
Breath Volume	1.351	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.173	04:31
Breath Volume	1.417	04:31
Air Blank	0.000	04:32
Control Test	0.078	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY, MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY, MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: MARTINEZ CARDENAS

DATE/TIME OF TEST: 03/15/2025 05:49:18

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.078	05:43
Air Blank	0.000	05:44
Subject Sample # 1	0.180	05:44
Breath Volume	2.226	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.176	05:47
Breath Volume	1.457	05:47
Air Blank	0.000	05:48
Control Test	0.078	05:48
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ADDERLEY

DATE/TIME OF TEST: 03/16/2025 00:14:45

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:12	
Control Test	0.078	00:12	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:13	Subject Test Refused
Breath Volume	0.000	00:13	
Air Blank	0.000	00:13	
Control Test	0.079	00:14	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALBAN,OMAR,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: GODINEZ,RANDY
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: TENORIO VALLADARES

DATE/TIME OF TEST: 03/16/2025 01:05:14

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.078	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.169	01:00
Breath Volume		1.460	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.164	01:03
Breath Volume		1.324	01:03
Air Blank		0.000	01:04
Control Test		0.076	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MENENCIER

DATE/TIME OF TEST: 03/17/2025 00:58:11

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.079	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 1	0.212	00:49	Volume Not Met
Breath Volume		0.542	00:49	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 2	0.227	00:53	
Breath Volume		1.441	00:53	
Air Blank		0.000	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 3	0.228	00:56	
Breath Volume		1.519	00:56	
Air Blank		0.000	00:57	
Control Test		0.077	00:57	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 03/18/2025 01:51:16

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:40	
Control Test		0.079	01:40	
Air Blank		0.000	01:40	
Subject Sample	# 1	0.142	01:44	Volume Not Met
Breath Volume		0.312	01:44	
Air Blank		0.000	01:44	
Air Blank		0.000	01:46	
Subject Sample	# 2	0.149	01:49	Volume Not Met
Breath Volume		0.035	01:49	
Air Blank		0.000	01:50	
Control Test		0.081	01:50	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,MOISES,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BLANCO,MOISES
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CUELI

DATE/TIME OF TEST: 03/19/2025 23:49:03

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.093	23:44
Breath Volume	1.160	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.087	23:47
Breath Volume	1.371	23:47
Air Blank	0.000	23:48
Control Test	0.077	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:49

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEBO,KEVIN,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: FEBO,KEVIN
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/22/2025 02:59:27

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.097	02:55
Breath Volume	2.449	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.095	02:57
Breath Volume	2.246	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPE,RYAN,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LAPE,RYAN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: BORREGO SALAZAR

DATE/TIME OF TEST: 03/22/2025 16:36:38

Diagnostics Check	OK	16:30
Air Blank	0.000	16:30
Control Test	0.078	16:30
Air Blank	0.000	16:31
Subject Sample # 1	0.157	16:32
Breath Volume	2.746	16:32
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.157	16:35
Breath Volume	2.656	16:35
Air Blank	0.000	16:35
Control Test	0.078	16:36
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORISMOND,WISNER,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: GODINEZ,RANDY
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BLANCO RAMIREZ

DATE/TIME OF TEST: 03/23/2025 00:54:02

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.079	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.144	00:49
Breath Volume	2.515	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.141	00:52
Breath Volume	2.578	00:52
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:54

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: PEREZ TRUJILLO

DATE/TIME OF TEST: 03/23/2025 04:54:52

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.139	04:49
Breath Volume	2.082	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.136	04:53
Breath Volume	2.246	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 03/24/2025 03:03:40

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.078	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURBELO MESA,CLAUDIA,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: CURBELO MESA,CLAUDIA
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MAZA

DATE/TIME OF TEST: 03/25/2025 02:50:52

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.206	02:46
Breath Volume		1.726	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:48
Subject Sample	# 2	0.213	02:49
Breath Volume		2.605	02:49
Air Blank		0.000	02:49
Control Test		0.076	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: GONZALEZ,A
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: CHAVARRIALOPEZ

DATE/TIME OF TEST: 03/29/2025 09:11:59

Diagnostics Check	OK	09:05
Air Blank	0.000	09:06
Control Test	0.079	09:06
Air Blank	0.000	09:07
Subject Sample # 1	0.193	09:07
Breath Volume	2.347	09:07
Air Blank	0.000	09:08
Air Blank	0.000	09:09
Subject Sample # 2	0.182	09:10
Breath Volume	1.523	09:10
Air Blank	0.000	09:11
Control Test	0.078	09:11
Air Blank	0.000	09:11
Diagnostics Check	OK	09:11

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLO,JHOJAN,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: GALLO,JHOJAN
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 03/29/2025 17:23:24

Diagnostics Check	OK	17:14	
Air Blank	0.000	17:15	
Control Test	0.079	17:15	
Air Blank	0.000	17:16	
Subject Sample	# 1 0.155	17:19	Volume Not Met
Breath Volume	1.042	17:19	
Air Blank	0.000	17:20	
Air Blank	0.000	17:21	
Subject Sample	# 2	17:22	RFI Detect
Breath Volume	0.000	17:22	
Air Blank	0.000	17:22	
Control Test	0.075	17:22	
Air Blank	0.000	17:23	
Diagnostics Check	OK	17:23	

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLO,JHOJAN,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: GALLO,JHOJAN
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 03/29/2025 17:34:16

Diagnostics Check		OK	17:28
Air Blank		0.000	17:28
Control Test		0.078	17:28
Air Blank		0.000	17:29
Subject Sample	# 1	0.180	17:29
Breath Volume		1.992	17:29
Air Blank		0.000	17:30
Air Blank		0.000	17:32
Subject Sample	# 2	0.190	17:32
Breath Volume		1.503	17:32
Air Blank		0.000	17:33
Control Test		0.077	17:33
Air Blank		0.000	17:34
Diagnostics Check		OK	17:34

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: HERNANDEZ FUNEZ

DATE/TIME OF TEST: 03/30/2025 23:34:34

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.297	23:29
Breath Volume	2.820	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.297	23:32
Breath Volume	2.046	23:32
Air Blank	0.000	23:33
Control Test	0.078	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 04/18/2025 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: CHAN VARGAS

DATE/TIME OF TEST: 03/13/2025 01:11:54

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.143	01:06
Breath Volume	1.171	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.156	01:10
Breath Volume	1.738	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 04/03/2025 03:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEBO,KEVIN,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: MCCLURE,TIMOTHY
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: BARRANTES

DATE/TIME OF TEST: 03/13/2025 03:35:01

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.165	03:29
Breath Volume	1.734	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.167	03:33
Breath Volume	2.171	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:35

Data Download Date/Time: 04/03/2025 03:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: EVELAND

DATE/TIME OF TEST: 03/08/2025 05:10:52

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.081	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.105	05:06
Breath Volume	2.175	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:08
Subject Sample # 2	0.109	05:09
Breath Volume	2.003	05:09
Air Blank	0.000	05:09
Control Test	0.081	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 03/13/2025 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/01/2025 00:49:38

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.241	00:45
Breath Volume		1.343	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.232	00:48
Breath Volume		1.804	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 03/14/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 03/02/2025 02:46:19

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.098	02:41
Breath Volume	1.546	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.102	02:44
Breath Volume	1.777	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 03/14/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: NETTLES

DATE/TIME OF TEST: 03/02/2025 04:22:05

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.037	04:17
Breath Volume	1.871	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.038	04:20
Breath Volume	1.625	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 03/14/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BONNER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 03/02/2025 09:19:15

Diagnostics Check		OK	09:12
Air Blank		0.000	09:13
Control Test		0.080	09:13
Air Blank		0.000	09:14
Subject Sample	# 1	0.184	09:14
Breath Volume		1.121	09:14
Air Blank		0.000	09:15
Air Blank		0.000	09:17
Subject Sample	# 2	0.177	09:17
Breath Volume		1.859	09:17
Air Blank		0.000	09:18
Control Test		0.079	09:18
Air Blank		0.000	09:19
Diagnostics Check		OK	09:19

Data Download Date/Time: 03/14/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: ALEMAN MENESES

DATE/TIME OF TEST: 03/02/2025 19:59:11

Diagnostics Check		OK	19:52
Air Blank		0.000	19:53
Control Test		0.080	19:53
Air Blank		0.000	19:54
Subject Sample	# 1	0.231	19:54
Breath Volume		1.613	19:54
Air Blank		0.000	19:55
Air Blank		0.000	19:57
Subject Sample	# 2	0.244	19:57
Breath Volume		1.660	19:57
Air Blank		0.000	19:58
Control Test		0.080	19:58
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 03/04/2025 18:23:07

Diagnostics Check	OK	18:20	
Air Blank	0.000	18:20	
Control Test	0.079	18:21	
Air Blank	0.000	18:21	
Subject Sample	# 1	18:21	Subject Test Refused
Breath Volume	0.000	18:21	
Air Blank	0.000	18:22	
Control Test	0.079	18:22	
Air Blank	0.000	18:23	
Diagnostics Check	OK	18:23	

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 03/04/2025 20:36:00

Diagnostics Check	OK	20:33	
Air Blank	0.000	20:33	
Control Test	0.080	20:33	
Air Blank	0.000	20:34	
Subject Sample	# 1	20:34	Subject Test Refused
Breath Volume	0.000	20:34	
Air Blank	0.000	20:35	
Control Test	0.080	20:35	
Air Blank	0.000	20:35	
Diagnostics Check	OK	20:36	

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KRAMER,S
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/05/2025 01:33:54

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Control Test	0.079	01:33	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHAMMED ALABBAS,A,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KELLY,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 03/05/2025 04:56:58

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.079	04:54	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:55	Subject Test Refused
Breath Volume	0.000	04:55	
Air Blank	0.000	04:56	
Control Test	0.080	04:56	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: RIVERO

DATE/TIME OF TEST: 03/05/2025 22:54:53

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.094	22:50
Breath Volume	1.308	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.091	22:53
Breath Volume	1.589	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,PETER,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RICH,T
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: THORE

DATE/TIME OF TEST: 03/06/2025 00:27:05

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.079	00:25	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:25	Subject Test Refused
Breath Volume	0.000	00:25	
Air Blank	0.000	00:26	
Control Test	0.079	00:26	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: ADMIN

DATE/TIME OF TEST: 03/06/2025 17:47:23

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.080	17:41
Air Blank	0.000	17:42
Subject Sample # 1	0.256	17:42
Breath Volume	3.644	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.252	17:45
Breath Volume	3.796	17:45
Air Blank	0.000	17:46
Control Test	0.079	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: ADMIN

DATE/TIME OF TEST: 03/06/2025 18:01:31

Diagnostics Check	OK	17:55
Air Blank	0.000	17:55
Control Test	0.079	17:56
Air Blank	0.000	17:56
Subject Sample # 1	0.250	17:57
Breath Volume	3.664	17:57
Air Blank	0.000	17:57
Air Blank	0.000	17:59
Subject Sample # 2	0.250	18:00
Breath Volume	3.656	18:00
Air Blank	0.000	18:00
Control Test	0.079	18:00
Air Blank	0.000	18:01
Diagnostics Check	OK	18:01

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: ADMIN

DATE/TIME OF TEST: 03/06/2025 18:06:20

Diagnostics Check	OK	18:03	
Air Blank	0.000	18:03	
Control Test	0.080	18:04	
Air Blank	0.000	18:04	
Subject Sample	# 1	18:04	Subject Test Refused
Breath Volume	0.000	18:04	
Air Blank	0.000	18:05	
Control Test	0.080	18:05	
Air Blank	0.000	18:06	
Diagnostics Check	OK	18:06	

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: TURBEVILLE

DATE/TIME OF TEST: 03/06/2025 18:12:43

Diagnostics Check	OK	18:09	
Air Blank	0.000	18:10	
Control Test	0.080	18:10	
Air Blank	0.000	18:11	
Subject Sample	# 1	18:11	Subject Test Refused
Breath Volume	0.000	18:11	
Air Blank	0.000	18:11	
Control Test	0.080	18:12	
Air Blank	0.000	18:12	
Diagnostics Check	OK	18:12	

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 03/08/2025 01:22:20

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.189	01:17
Breath Volume	2.394	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.176	01:20
Breath Volume	2.316	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: COLSON

DATE/TIME OF TEST: 03/08/2025 03:22:23

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.140	03:17
Breath Volume		2.070	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.140	03:20
Breath Volume		1.535	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMSON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 03/08/2025 05:12:23

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.142	05:07
Breath Volume	1.515	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.152	05:10
Breath Volume	1.941	05:10
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 03/11/2025 01:13:59

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.079	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:13	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 03/14/2025 04:08:42

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.080	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.080	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: POSADA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: TILLET

DATE/TIME OF TEST: 03/14/2025 07:43:00

Diagnostics Check	OK	07:35
Air Blank	0.000	07:35
Control Test	0.079	07:35
Air Blank	0.000	07:36
Subject Sample # 1	0.143	07:38
Breath Volume	1.750	07:38
Air Blank	0.000	07:38
Air Blank	0.000	07:40
Subject Sample # 2	0.143	07:41
Breath Volume	1.687	07:41
Air Blank	0.000	07:42
Control Test	0.079	07:42
Air Blank	0.000	07:42
Diagnostics Check	OK	07:43

Data Download Date/Time: 03/14/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMSON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: KISER

DATE/TIME OF TEST: 03/16/2025 02:01:07

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.193	01:56
Breath Volume	1.355	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.207	01:59
Breath Volume	2.289	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 04/22/2025 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 03/16/2025 04:45:58

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:34	
Control Test		0.080	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1	0.145	04:38	Volume Not Met
Breath Volume		0.160	04:38	
Air Blank		0.000	04:38	
Air Blank		0.000	04:40	
Subject Sample	# 2	0.153	04:41	
Breath Volume		1.757	04:41	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 3	0.148	04:44	
Breath Volume		1.371	04:44	
Air Blank		0.000	04:45	
Control Test		0.079	04:45	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 04/22/2025 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPENCER,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: PEVETO

DATE/TIME OF TEST: 03/18/2025 07:41:44

Diagnostics Check	OK	07:35
Air Blank	0.000	07:36
Control Test	0.081	07:36
Air Blank	0.000	07:36
Subject Sample # 1	0.066	07:37
Breath Volume	3.765	07:37
Air Blank	0.000	07:37
Air Blank	0.000	07:39
Subject Sample # 2	0.068	07:40
Breath Volume	4.738	07:40
Air Blank	0.000	07:40
Control Test	0.080	07:41
Air Blank	0.000	07:41
Diagnostics Check	OK	07:41

Data Download Date/Time: 04/22/2025 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MEJIA DIAZ

DATE/TIME OF TEST: 03/19/2025 01:40:27

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.106	01:36
Breath Volume	2.863	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.105	01:38
Breath Volume	2.023	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 04/22/2025 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTTER,JONAH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: ASHE

DATE/TIME OF TEST: 03/21/2025 23:24:44

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.131	23:19
Breath Volume	1.941	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.132	23:23
Breath Volume	2.660	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 03/22/2025 00:07:04

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.080	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.233	00:02
Breath Volume		2.695	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.234	00:05
Breath Volume		1.488	00:05
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:07

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTTER,JONAH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/22/2025 03:16:22

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.150	03:10
Breath Volume	2.503	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.154	03:14
Breath Volume	3.199	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTTER,JONAH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: LEMUZ

DATE/TIME OF TEST: 03/22/2025 04:08:37

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.051	04:04
Breath Volume	3.187	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.051	04:07
Breath Volume	2.933	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONAH,TROTTER,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: GAVOCI

DATE/TIME OF TEST: 03/22/2025 04:49:32

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.080	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.080	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTTER,JONAH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLING,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: COLSON

DATE/TIME OF TEST: 03/22/2025 05:43:00

Diagnostics Check	OK	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:38
Subject Sample # 1	0.000	05:38
Breath Volume	2.343	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.000	05:41
Breath Volume	1.808	05:41
Air Blank	0.000	05:42
Control Test	0.080	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:43

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,C,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/22/2025 16:38:46

Diagnostics Check	OK	16:35	
Air Blank	0.000	16:36	
Control Test	0.080	16:36	
Air Blank	0.000	16:37	
Subject Sample	# 1	16:37	Subject Test Refused
Breath Volume	0.000	16:37	
Air Blank	0.000	16:37	
Control Test	0.080	16:38	
Air Blank	0.000	16:38	
Diagnostics Check	OK	16:38	

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,M,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MANUKYAN,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 03/23/2025 02:52:49

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.080	02:49	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:52	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,M,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N.V.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: COKLEY

DATE/TIME OF TEST: 03/23/2025 05:47:30

Diagnostics Check		OK	05:40
Air Blank		0.000	05:41
Control Test		0.080	05:41
Air Blank		0.000	05:42
Subject Sample	# 1	0.113	05:42
Breath Volume		1.960	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.112	05:46
Breath Volume		2.312	05:46
Air Blank		0.000	05:46
Control Test		0.079	05:46
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GIVENS

DATE/TIME OF TEST: 03/24/2025 02:00:49

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.145	01:55
Breath Volume	1.324	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.143	01:59
Breath Volume	1.730	01:59
Air Blank	0.000	01:59
Control Test	0.078	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BEECHAM

DATE/TIME OF TEST: 03/25/2025 01:45:06

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.225	01:40
Breath Volume	1.382	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.227	01:43
Breath Volume	1.277	01:43
Air Blank	0.000	01:44
Control Test	0.077	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: GAILYARD

DATE/TIME OF TEST: 03/26/2025 22:24:49

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.149	22:20
Breath Volume	1.312	22:20
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.155	22:23
Breath Volume	1.792	22:23
Air Blank	0.000	22:23
Control Test	0.080	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ELLISON,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: GADSON

DATE/TIME OF TEST: 03/27/2025 06:25:31

Diagnostics Check	OK	06:19
Air Blank	0.000	06:19
Control Test	0.080	06:20
Air Blank	0.000	06:20
Subject Sample # 1	0.149	06:20
Breath Volume	1.750	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.131	06:24
Breath Volume	1.667	06:24
Air Blank	0.000	06:24
Control Test	0.080	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/28/2025 02:38:26

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:37	
Control Test	0.080	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 03/29/2025 01:37:24

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.044	01:32
Breath Volume	2.148	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.046	01:35
Breath Volume	2.261	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 03/29/2025 05:05:49

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.079	04:59
Air Blank		0.000	05:00
Subject Sample	# 1	0.113	05:01
Breath Volume		1.945	05:01
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.121	05:04
Breath Volume		2.730	05:04
Air Blank		0.000	05:04
Control Test		0.079	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,D
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: CORTES DE LA ROSA

DATE/TIME OF TEST: 03/29/2025 06:39:54

Diagnostics Check	OK	06:33
Air Blank	0.000	06:34
Control Test	0.079	06:34
Air Blank	0.000	06:34
Subject Sample # 1	0.092	06:35
Breath Volume	1.722	06:35
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample # 2	0.092	06:38
Breath Volume	1.500	06:38
Air Blank	0.000	06:38
Control Test	0.080	06:39
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 03/30/2025 00:01:10

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.079	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1	23:59	Subject Test Refused
Breath Volume	0.000	23:59	
Air Blank	0.000	00:00	
Control Test	0.080	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: CRUZ-CRUZ

DATE/TIME OF TEST: 03/30/2025 01:49:54

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.181	01:45
Breath Volume		1.636	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.180	01:48
Breath Volume		1.761	01:48
Air Blank		0.000	01:49
Control Test		0.078	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ECHEVERRIAVIDAL,X
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: RAMIREZ PEREZ

DATE/TIME OF TEST: 03/30/2025 04:40:57

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.271	04:36
Breath Volume	1.484	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.268	04:39
Breath Volume	2.101	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORGERT,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:24

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/30/2025 07:51:16

Diagnostics Check	OK	07:44
Air Blank	0.000	07:44
Control Test	0.079	07:45
Air Blank	0.000	07:45
Subject Sample # 1	0.077	07:46
Breath Volume	1.835	07:46
Air Blank	0.000	07:46
Air Blank	0.000	07:48
Subject Sample # 2	0.079	07:49
Breath Volume	1.390	07:49
Air Blank	0.000	07:50
Control Test	0.079	07:50
Air Blank	0.000	07:51
Diagnostics Check	OK	07:51

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/31/2025 03:31:54

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.136	03:27
Breath Volume	3.402	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.131	03:30
Breath Volume	3.535	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HAMBURGER

DATE/TIME OF TEST: 03/31/2025 19:57:31

Diagnostics Check	OK	19:50
Air Blank	0.000	19:50
Control Test	0.079	19:50
Air Blank	0.000	19:51
Subject Sample # 1	0.112	19:52
Breath Volume	2.468	19:52
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.118	19:56
Breath Volume	2.820	19:56
Air Blank	0.000	19:56
Control Test	0.079	19:56
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 04/22/2025 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,GLENN,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,GLENN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 03/04/2025 14:52:02

Diagnostics Check	OK	14:50	
Air Blank	0.000	14:50	
Control Test	0.082	14:50	
Air Blank	0.000	14:51	
Subject Sample	# 1	14:51	RFI Detect
Breath Volume	0.000	14:51	
Air Blank		14:51	RFI Detect
Air Blank	0.000	14:52	

Data Download Date/Time: 03/30/2025 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHILLER,JOSEPH,C
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: CALLOWAY,DANIEL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: SPURGEON

DATE/TIME OF TEST: 03/07/2025 12:26:12

Diagnostics Check	OK	12:20
Air Blank	0.000	12:20
Control Test	0.082	12:21
Air Blank	0.000	12:21
Subject Sample # 1	0.000	12:21
Breath Volume	1.980	12:21
Air Blank	0.000	12:22
Air Blank	0.000	12:24
Subject Sample # 2	0.000	12:24
Breath Volume	2.554	12:24
Air Blank	0.000	12:25
Control Test	0.080	12:25
Air Blank	0.000	12:26
Diagnostics Check	OK	12:26

Data Download Date/Time: 03/30/2025 17:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,ALAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: PALMER,ALAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 03/10/2025 01:47:17

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.181	01:42
Breath Volume	1.750	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.168	01:45
Breath Volume	1.992	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 03/30/2025 17:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,GLENN,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: CALLAWAY,DANIEL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 03/15/2025 00:02:00

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.082	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:57
Breath Volume	2.386	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:00
Breath Volume	2.023	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:02

Data Download Date/Time: 03/30/2025 17:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODKIN,JASON,M
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: BAILEY,ADAM
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024026353
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: SINNARD

DATE/TIME OF TEST: 03/09/2025 05:03:18

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.078	05:01	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:02	
Control Test	0.077	05:02	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 04/04/2025 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODKIN,JASON,M
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: HULSEN,JOHN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024026353
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: SPAETH

DATE/TIME OF TEST: 03/17/2025 19:48:35

Diagnostics Check	OK	19:45	
Air Blank	0.000	19:46	
Control Test	0.078	19:46	
Air Blank	0.000	19:46	
Subject Sample	# 1	19:47	Subject Test Refused
Breath Volume	0.000	19:47	
Air Blank	0.000	19:47	
Control Test	0.078	19:47	
Air Blank	0.000	19:48	
Diagnostics Check	OK	19:48	

Data Download Date/Time: 04/04/2025 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUZADER,I,L
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: WILLIAMSON,S
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024026353
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 03/29/2025 17:09:14

Diagnostics Check	OK	17:08	
Air Blank		17:08	RFI Detect
Air Blank	0.000	17:09	

Data Download Date/Time: 04/04/2025 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUZADER,I,L
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: WILLIAMSON,S
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024026353
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 03/29/2025 17:12:23

Diagnostics Check	OK	17:11	
Air Blank		17:11	RFI Detect
Air Blank	0.000	17:12	

Data Download Date/Time: 04/04/2025 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUZADER,I,L
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: WILLIAMSON,S
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024026353
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 03/29/2025 17:21:07

Diagnostics Check	OK	17:14
Air Blank	0.000	17:15
Control Test	0.077	17:15
Air Blank	0.000	17:15
Subject Sample # 1	0.179	17:16
Breath Volume	3.640	17:16
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample # 2	0.179	17:19
Breath Volume	4.460	17:19
Air Blank	0.000	17:20
Control Test	0.076	17:20
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

Data Download Date/Time: 04/04/2025 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,STEVEN,R
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: GAFFEY,TUCKER
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024026353
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: SANTILLI

DATE/TIME OF TEST: 03/29/2025 20:18:48

Diagnostics Check		OK	20:11
Air Blank		0.000	20:12
Control Test		0.077	20:12
Air Blank		0.000	20:12
Subject Sample	# 1	0.242	20:14
Breath Volume		1.445	20:14
Air Blank		0.000	20:15
Air Blank		0.000	20:16
Subject Sample	# 2	0.245	20:17
Breath Volume		1.265	20:17
Air Blank		0.000	20:17
Control Test		0.076	20:18
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 04/04/2025 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWLING, DAKOTA, A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DOWLING, DAKOTA
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024026353
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: DIAZ CASTRO

DATE/TIME OF TEST: 03/30/2025 16:49:59

Diagnostics Check	OK	16:48	
Air Blank	0.000	16:49	
Control Test	0.077	16:49	
Air Blank		16:49	RFI Detect
Air Blank		16:49	RFI Detect

Data Download Date/Time: 04/04/2025 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWLING,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DOWLING,DAKOTA
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024026353
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: DIAZ CASTRO

DATE/TIME OF TEST: 03/30/2025 16:58:39

Diagnostics Check		OK	16:52
Air Blank		0.000	16:52
Control Test		0.077	16:52
Air Blank		0.000	16:53
Subject Sample	# 1	0.219	16:53
Breath Volume		1.472	16:53
Air Blank		0.000	16:54
Air Blank		0.000	16:56
Subject Sample	# 2	0.235	16:56
Breath Volume		1.437	16:56
Air Blank		0.000	16:57
Control Test		0.076	16:58
Air Blank		0.000	16:58
Diagnostics Check		OK	16:58

Data Download Date/Time: 04/04/2025 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,PATRICK,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STACK,KEVIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: CAPUNFUERZA

DATE/TIME OF TEST: 03/06/2025 02:13:30

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.078	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:09	No Sample Provided
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Air Blank	0.000	02:11	
Subject Sample	# 2	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.077	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SLAUGHTER,JASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: ZABRISKIE

DATE/TIME OF TEST: 03/06/2025 19:39:07

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.079	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.194	19:34
Breath Volume	2.167	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.213	19:37
Breath Volume	2.730	19:37
Air Blank	0.000	19:38
Control Test	0.079	19:38
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GULLEDGE,JOHN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CORNMAN,KEVIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 03/08/2025 00:49:10

Diagnostics Check	OK	00:37		
Air Blank	0.000	00:37		
Control Test	0.078	00:38		
Air Blank	0.000	00:38		
Subject Sample	# 1	00:41	No Sample Provided	
Breath Volume	0.000	00:41		
Air Blank	0.000	00:42		
Air Blank	0.000	00:44		
Subject Sample	# 2	0.167	00:47	Volume Not Met
Breath Volume	0.574	00:47		
Air Blank	0.000	00:48		
Control Test	0.079	00:48		
Air Blank	0.000	00:49		
Diagnostics Check	OK	00:49		

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GULLEDGE,JOHN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CORNMAN,KEVIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 03/08/2025 01:09:39

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.078	01:00	
Air Blank		0.000	01:00	
Subject Sample	# 1	0.194	01:03	Volume Not Met
Breath Volume		0.136	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2		01:08	Subject Test Refused
Breath Volume		0.082	01:08	
Air Blank		0.000	01:08	
Control Test		0.078	01:09	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,STANLEY,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WOODS,KENNETH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 03/08/2025 01:24:19

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.078	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALBRECHT, SYDNEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/09/2025 01:38:33

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.078	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.078	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: SHAUGHNESSY

DATE/TIME OF TEST: 03/09/2025 16:56:58

Diagnostics Check	OK	16:50
Air Blank	0.000	16:50
Control Test	0.078	16:50
Air Blank	0.000	16:51
Subject Sample # 1	0.192	16:52
Breath Volume	2.687	16:52
Air Blank	0.000	16:52
Air Blank	0.000	16:54
Subject Sample # 2	0.194	16:55
Breath Volume	2.718	16:55
Air Blank	0.000	16:56
Control Test	0.078	16:56
Air Blank	0.000	16:56
Diagnostics Check	OK	16:56

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/12/2025 01:03:07

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.079	01:01	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.079	01:02	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMILLION,LANCE,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: GOMILLION,LANCE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/12/2025 21:45:22

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.000	21:41
Breath Volume	2.925	21:41
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.000	21:44
Breath Volume	2.093	21:44
Air Blank	0.000	21:44
Control Test	0.077	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH, BLAKE, T
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELCH, BLAKE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/12/2025 23:08:27

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.076	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.237	23:03
Breath Volume	2.972	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:05
Subject Sample # 2	0.236	23:06
Breath Volume	3.539	23:06
Air Blank	0.000	23:07
Control Test	0.077	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STACK,KEVIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LOCHRIDGE

DATE/TIME OF TEST: 03/15/2025 00:36:39

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.277	00:31
Breath Volume	2.667	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.281	00:35
Breath Volume	3.343	00:35
Air Blank	0.000	00:35
Control Test	0.078	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSLEY,KYLE,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOSLEY,KYLE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/16/2025 02:15:23

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.077	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1 0.241	02:10	
Breath Volume	2.863	02:10	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2 0.236	02:13	
Breath Volume	3.324	02:13	
Air Blank	0.000	02:14	
Control Test	0.074	02:14	Control Outside Tolerance
Air Blank	0.000	02:15	

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 03/19/2025 00:29:33

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.079	00:26	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.079	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: RAWLS

DATE/TIME OF TEST: 03/19/2025 23:20:56

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.093	23:16
Breath Volume	1.210	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.100	23:19
Breath Volume	1.734	23:19
Air Blank	0.000	23:20
Control Test	0.078	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSLEY,KYLE,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOSLEY,KYLE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 03/20/2025 02:33:36

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.127	02:29
Breath Volume	2.000	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.124	02:32
Breath Volume	1.835	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 04/06/2025 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKWOOD,PATRICK,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CASEY,SHELBY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: BUCY

DATE/TIME OF TEST: 03/20/2025 19:23:50

Diagnostics Check	OK	19:23	
Air Blank		19:23	RFI Detect
Air Blank	0.000	19:23	

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKWOOD,PATRICK,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CASEY,SHELBY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: BUCY

DATE/TIME OF TEST: 03/20/2025 19:39:39

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:36	
Control Test	0.078	19:36	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:38	Subject Test Refused
Breath Volume	0.000	19:38	
Air Blank	0.000	19:38	
Control Test	0.079	19:39	
Air Blank	0.000	19:39	
Diagnostics Check	OK	19:39	

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GULLEDGE,JOHN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROBIN,JERRY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: SELENT

DATE/TIME OF TEST: 03/23/2025 19:00:24

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:47	
Control Test	0.079	18:47	
Air Blank	0.000	18:48	
Subject Sample	# 1 0.164	18:51	Volume Not Met
Breath Volume	0.167	18:51	
Air Blank	0.000	18:52	
Air Blank	0.000	18:53	
Subject Sample	# 2 0.171	18:55	
Breath Volume	1.375	18:55	
Air Blank	0.000	18:56	
Air Blank	0.000	18:58	
Subject Sample	# 3 0.164	18:58	
Breath Volume	2.179	18:58	
Air Blank	0.000	18:59	
Control Test	0.077	18:59	
Air Blank	0.000	19:00	
Diagnostics Check	OK	19:00	

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ONEILL,BRAYDEN,P
BREATH TEST OPERATOR AGENCY: Milton PD
ARRESTING OFFICER: ONEILL,BRAYDEN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: KRASNOSKY

DATE/TIME OF TEST: 03/24/2025 01:44:55

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.078	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.156	01:40
Breath Volume	1.820	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.166	01:43
Breath Volume	1.304	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ONEILL,BRAYDEN,P
BREATH TEST OPERATOR AGENCY: Milton PD
ARRESTING OFFICER: ONEILL,BRAYDEN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/25/2025 04:17:25

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.000	04:12
Breath Volume	2.835	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.000	04:16
Breath Volume	2.382	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,STANLEY,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HANKINS,MICHAELA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/27/2025 21:01:05

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:58	
Control Test	0.079	20:58	
Air Blank	0.000	20:59	
Subject Sample	# 1	20:59	Subject Test Refused
Breath Volume	0.000	20:59	
Air Blank	0.000	21:00	
Control Test	0.079	21:00	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 13:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: GARMON

DATE/TIME OF TEST: 03/28/2025 14:58:35

Diagnostics Check	OK	14:51
Air Blank	0.000	14:52
Control Test	0.079	14:52
Air Blank	0.000	14:53
Subject Sample # 1	0.000	14:54
Breath Volume	1.628	14:54
Air Blank	0.000	14:54
Air Blank	0.000	14:56
Subject Sample # 2	0.000	14:57
Breath Volume	1.921	14:57
Air Blank	0.000	14:57
Control Test	0.078	14:57
Air Blank	0.000	14:58
Diagnostics Check	OK	14:58

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKWOOD,PATRICK,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CASEY,SHELBY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MACKS

DATE/TIME OF TEST: 03/29/2025 02:26:31

Diagnostics Check	OK	02:25	
Air Blank		02:26	RFI Detect
Air Blank	0.000	02:26	

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKWOOD,PATRICK,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CASEY,SHELBY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MACKS

DATE/TIME OF TEST: 03/29/2025 02:34:37

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.078	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.088	02:30
Breath Volume	3.433	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.080	02:33
Breath Volume	1.519	02:33
Air Blank	0.000	02:33
Control Test	0.078	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ONEILL,BRAYDEN,P
BREATH TEST OPERATOR AGENCY: Milton PD
ARRESTING OFFICER: ONEILL,BRAYDEN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 03/29/2025 03:02:28

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.059	02:58
Breath Volume	2.812	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.062	03:00
Breath Volume	2.960	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDD,RACHEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUDD,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: DOWNS-WHITE

DATE/TIME OF TEST: 03/29/2025 20:56:51

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.079	20:51
Air Blank	0.000	20:51
Subject Sample # 1	0.197	20:52
Breath Volume	2.511	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:54
Subject Sample # 2	0.198	20:55
Breath Volume	2.644	20:55
Air Blank	0.000	20:55
Control Test	0.078	20:56
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKWOOD,PATRICK,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MABIRE,TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: TOLENTINO

DATE/TIME OF TEST: 03/30/2025 01:56:19

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.190	01:51
Breath Volume	1.484	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.177	01:54
Breath Volume	1.347	01:54
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,PATRICK,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STACK,KEVIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BROWNING

DATE/TIME OF TEST: 03/30/2025 02:52:40

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.077	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.083	02:47
Breath Volume	1.667	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.088	02:51
Breath Volume	1.679	02:51
Air Blank	0.000	02:51
Control Test	0.078	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,PATRICK,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STEPHENSON,PATRICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: HARBIN JOHNSON

DATE/TIME OF TEST: 03/30/2025 04:33:23

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.078	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.320	04:28
Breath Volume	1.957	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.322	04:31
Breath Volume	1.281	04:31
Air Blank	0.000	04:32
Control Test	0.076	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKWOOD,PATRICK,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FELDHAUS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ALMENDAREZ

DATE/TIME OF TEST: 03/30/2025 22:05:35

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
Control Test	0.079	22:03	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Control Test	0.078	22:04	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPZ,BRIAN,T
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LOPEZ,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/30/2025 22:38:48

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.077	22:30	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:34	No Sample Provided
Breath Volume	0.000	22:34	
Air Blank	0.000	22:34	
Air Blank	0.000	22:36	
Subject Sample	# 2	22:37	Subject Test Refused
Breath Volume	0.000	22:37	
Air Blank	0.000	22:37	
Control Test	0.077	22:38	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:38	

Data Download Date/Time: 04/06/2025 20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ONEILL,BRAYDEN,P
BREATH TEST OPERATOR AGENCY: Milton PD
ARRESTING OFFICER: ONEILL,BRAYDEN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: KOHLMAN

DATE/TIME OF TEST: 03/01/2025 00:01:49

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.082	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.148	23:54	Volume Not Met
Breath Volume	0.281	23:54	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 1	00:00	No Sample Provided
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 03/05/2025 03:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ONEILL,BRAYDEN,P
BREATH TEST OPERATOR AGENCY: Milton PD
ARRESTING OFFICER: ONEILL,BRAYDEN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:03

SUBJECT'S LAST NAME: KOHLMAN

DATE/TIME OF TEST: 03/01/2025 00:13:31

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.157	00:08
Breath Volume	1.128	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.145	00:11
Breath Volume	1.105	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 03/05/2025 03:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MABIRE,TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MCKINNON

DATE/TIME OF TEST: 03/01/2025 02:21:57

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.082	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:21	
Control Test	0.083	02:21	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 03/05/2025 03:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BURKS

DATE/TIME OF TEST: 03/01/2025 23:11:55

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:00	
Control Test		0.082	23:01	
Air Blank		0.000	23:01	
Subject Sample	# 1	0.104	23:04	Volume Not Met
Breath Volume		0.050	23:04	
Air Blank		0.000	23:05	
Air Blank		0.000	23:07	
Subject Sample	# 2	0.102	23:10	Volume Not Met
Breath Volume		1.347	23:10	
Air Blank		0.000	23:11	
Control Test		0.082	23:11	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 03/05/2025 03:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,BRIAN,T
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACKENZIE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: LYON

DATE/TIME OF TEST: 03/02/2025 23:34:08

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.082	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.057	23:29
Breath Volume	2.140	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.059	23:32
Breath Volume	2.621	23:32
Air Blank	0.000	23:33
Control Test	0.083	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 03/05/2025 03:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSLEY,KYLE,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOSLEY,KYLE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/16/2025 02:27:20

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.239	02:23
Breath Volume	1.269	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.241	02:25
Breath Volume	1.742	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 04/06/2025 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/17/2025 02:53:50

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.083	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.083	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 04/06/2025 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSICK,ROBERT,C
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: BUSICK,ROBERT
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: WHITMIRE

DATE/TIME OF TEST: 03/01/2025 04:48:05

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.082	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.176	04:43
Breath Volume	2.269	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:45
Subject Sample # 2	0.187	04:46
Breath Volume	1.738	04:46
Air Blank	0.000	04:47
Control Test	0.082	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:48

Data Download Date/Time: 04/03/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYO,RICHARD,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MAYO,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 03/01/2025 20:29:04

Diagnostics Check		OK	20:22
Air Blank		0.000	20:23
Control Test		0.082	20:23
Air Blank		0.000	20:24
Subject Sample	# 1	0.225	20:24
Breath Volume		2.472	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:27
Subject Sample	# 2	0.235	20:27
Breath Volume		2.703	20:27
Air Blank		0.000	20:28
Control Test		0.081	20:28
Air Blank		0.000	20:28
Diagnostics Check		OK	20:29

Data Download Date/Time: 04/03/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRISP,JESSICA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LAMB,MICHAEL
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: PEEL

DATE/TIME OF TEST: 03/05/2025 15:43:19

Diagnostics Check	OK	15:36
Air Blank	0.000	15:37
Control Test	0.082	15:37
Air Blank	0.000	15:37
Subject Sample # 1	0.232	15:38
Breath Volume	1.988	15:38
Air Blank	0.000	15:39
Air Blank	0.000	15:40
Subject Sample # 2	0.230	15:41
Breath Volume	1.601	15:41
Air Blank	0.000	15:42
Control Test	0.080	15:42
Air Blank	0.000	15:43
Diagnostics Check	OK	15:43

Data Download Date/Time: 04/03/2025 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MALONEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: BASFORD

DATE/TIME OF TEST: 03/06/2025 22:33:47

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.083	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.367	22:29
Breath Volume	2.183	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.355	22:32
Breath Volume	2.066	22:32
Air Blank	0.000	22:32
Control Test	0.082	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 04/03/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MALONEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: BASFORD

DATE/TIME OF TEST: 03/06/2025 23:04:02

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.082	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.344	22:59
Breath Volume	2.406	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.359	23:02
Breath Volume	2.953	23:02
Air Blank	0.000	23:03
Control Test	0.082	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:04

Data Download Date/Time: 04/03/2025 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRISP,JESSICA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CRISP,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 03/29/2025 16:07:55

Diagnostics Check	OK	16:01
Air Blank	0.000	16:02
Control Test	0.083	16:02
Air Blank	0.000	16:03
Subject Sample # 1	0.000	16:03
Breath Volume	4.308	16:03
Air Blank	0.000	16:03
Air Blank	0.000	16:05
Subject Sample # 2	0.000	16:06
Breath Volume	4.210	16:06
Air Blank	0.000	16:07
Control Test	0.082	16:07
Air Blank	0.000	16:07
Diagnostics Check	OK	16:07

Data Download Date/Time: 04/03/2025 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,DONALD,R
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: GOFF,BETH
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GARZON

DATE/TIME OF TEST: 03/01/2025 00:05:57

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.075	23:55	
Air Blank	0.000	23:55	
Subject Sample	# 2 0.093	23:58	Volume Not Met
Breath Volume	0.285	23:58	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.098	00:04	Volume Not Met
Breath Volume	0.328	00:04	
Air Blank	0.000	00:05	
Control Test	0.077	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 04/27/2025 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,DONALD,R
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: GOFF,BETH
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GARZON

DATE/TIME OF TEST: 03/01/2025 00:25:01

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:10	
Control Test	0.075	00:10	
Air Blank	0.000	00:11	
Subject Sample # 1	0.133	00:12	
Breath Volume	1.394	00:12	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample # 2	0.073	00:17	Volume Not Met
Breath Volume	0.023	00:17	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample # 3	0.085	00:23	Volume Not Met
Breath Volume	0.093	00:23	
Air Blank	0.000	00:23	
Control Test	0.076	00:24	
Air Blank		00:24	RFI Detect
Air Blank	0.000	00:25	

Data Download Date/Time: 04/27/2025 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON, STACEY, C
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WOLFER

DATE/TIME OF TEST: 03/10/2025 21:36:14

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.079	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.119	21:31
Breath Volume	1.660	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:33
Subject Sample # 2	0.124	21:34
Breath Volume	3.359	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 03/18/2025 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSME, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: THOMPSON, JOSHUA
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/01/2025 03:59:19

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.082	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.164	03:54
Breath Volume	2.890	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.153	03:57
Breath Volume	1.585	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 03/19/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOOD,KRISTIN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 03/02/2025 19:38:05

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.082	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.071	19:33
Breath Volume	1.882	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.071	19:36
Breath Volume	1.808	19:36
Air Blank	0.000	19:37
Control Test	0.083	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 03/19/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: POLLARD,CAMERON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/08/2025 20:19:10

Diagnostics Check	OK	20:12
Air Blank	0.000	20:13
Control Test	0.081	20:13
Air Blank	0.000	20:14
Subject Sample # 1	0.193	20:14
Breath Volume	2.816	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.200	20:17
Breath Volume	2.777	20:17
Air Blank	0.000	20:18
Control Test	0.081	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 03/19/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDOVA,BRYON,A
BREATH TEST OPERATOR AGENCY: Springfield PD
ARRESTING OFFICER: CORDOVA,BRYON
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: HANKS

DATE/TIME OF TEST: 03/09/2025 06:22:17

Diagnostics Check	OK	06:16
Air Blank	0.000	06:16
Control Test	0.081	06:16
Air Blank	0.000	06:17
Subject Sample # 1	0.143	06:17
Breath Volume	2.621	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample # 2	0.146	06:20
Breath Volume	2.347	06:20
Air Blank	0.000	06:21
Control Test	0.082	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 03/19/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELTHROP, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BELTHROP, MATTHEW
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 03/12/2025 04:16:29

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.082	04:05	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:09	No Sample Provided
Breath Volume	0.000	04:09	
Air Blank	0.000	04:09	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Control Test	0.082	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 03/19/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISON,ALLISON,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: LOCKMAN,HEATHER
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: BOHNEN

DATE/TIME OF TEST: 03/14/2025 01:09:30

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.081	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:08	Slope Not Met
Breath Volume	0.628	01:08	
Air Blank	0.000	01:08	
Control Test	0.081	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 03/19/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISON,ALLISON,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: LOCKMAN,HEATHER
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: BOHNEN

DATE/TIME OF TEST: 03/14/2025 01:17:57

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.081	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.222	01:13
Breath Volume		1.156	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:15
Subject Sample	# 2	0.230	01:16
Breath Volume		1.453	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 03/19/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON CORDOVA,BRYON,A
BREATH TEST OPERATOR AGENCY: Springfield PD
ARRESTING OFFICER: DAWSON CORDOVA,BRYON
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 03/17/2025 07:36:47

Diagnostics Check	OK	07:30
Air Blank	0.000	07:30
Control Test	0.081	07:30
Air Blank	0.000	07:31
Subject Sample # 1	0.111	07:32
Breath Volume	1.707	07:32
Air Blank	0.000	07:32
Air Blank	0.000	07:34
Subject Sample # 2	0.110	07:35
Breath Volume	2.390	07:35
Air Blank	0.000	07:35
Control Test	0.081	07:36
Air Blank	0.000	07:36
Diagnostics Check	OK	07:36

Data Download Date/Time: 03/19/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELTHROP, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BELTHROP, MATTHEW
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 03/20/2025 04:12:51

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.082	04:02	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:05	No Sample Provided
Breath Volume	0.000	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2	04:11	No Sample Provided
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Control Test	0.081	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 04/23/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: BLANKS

DATE/TIME OF TEST: 03/22/2025 01:46:57

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.081	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Control Test	0.081	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 04/23/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,SEVEN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: ASTON

DATE/TIME OF TEST: 03/22/2025 05:41:25

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.082	05:36
Air Blank	0.000	05:36
Subject Sample # 1	0.108	05:37
Breath Volume	2.558	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.106	05:39
Breath Volume	2.578	05:39
Air Blank	0.000	05:40
Control Test	0.081	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 04/23/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,F,H
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TURNER,M
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 13:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/22/2025 14:35:07

Diagnostics Check	OK	14:28
Air Blank	0.000	14:28
Control Test	0.081	14:29
Air Blank	0.000	14:29
Subject Sample # 1	0.192	14:30
Breath Volume	1.464	14:30
Air Blank	0.000	14:30
Air Blank	0.000	14:32
Subject Sample # 2	0.182	14:33
Breath Volume	1.152	14:33
Air Blank	0.000	14:34
Control Test	0.081	14:34
Air Blank	0.000	14:35
Diagnostics Check	OK	14:35

Data Download Date/Time: 04/23/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/23/2025 04:17:34

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:14	
Control Test	0.081	04:14	
Air Blank	0.000	04:15	
Subject Sample # 1	0.102	04:15	
Breath Volume	1.292	04:15	
Air Blank		04:16	Purge Fail
Air Blank	0.000	04:17	

Data Download Date/Time: 04/23/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/23/2025 04:21:51

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.067	04:21	Control Outside Tolerance
Air Blank	0.000	04:21	

Data Download Date/Time: 04/23/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/23/2025 04:25:32

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.063	04:24	Control Outside Tolerance
Air Blank	0.000	04:25	

Data Download Date/Time: 04/23/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELTHROP, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BELTHROP, MATTHEW
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MOSELEY

DATE/TIME OF TEST: 03/25/2025 03:48:30

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.264	03:41
Breath Volume		1.847	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.260	03:47
Breath Volume		1.664	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 04/23/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHERMAN,DEVON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MUTLUTURK,FATIH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: HOLDER

DATE/TIME OF TEST: 03/26/2025 01:58:08

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.081	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.183	01:53
Breath Volume		2.910	01:53
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.188	01:56
Breath Volume		2.023	01:56
Air Blank		0.000	01:57
Control Test		0.081	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 04/23/2025 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHERMAN,DEVON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MUTLUTURK,FATIH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 03/28/2025 01:57:53

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.160	01:53
Breath Volume	2.253	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.178	01:56
Breath Volume	2.648	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 04/23/2025 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELTHROP, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BELTHROP, MATTHEW
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 03/29/2025 00:30:30

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.079	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:23	No Sample Provided
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Air Blank	0.000	00:26	
Subject Sample	# 2	00:29	No Sample Provided
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 04/23/2025 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELTHROP, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BELTHROP, MATTHEW
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: WIRGAU

DATE/TIME OF TEST: 03/30/2025 02:04:54

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.082	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.000	02:00
Breath Volume	1.636	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.000	02:03
Breath Volume	1.582	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 04/23/2025 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: MARTINEZ SOLORZANO

DATE/TIME OF TEST: 03/02/2025 00:13:01

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.080	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:06	No Sample Provided
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 2	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Control Test	0.080	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 03/26/2025 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: MARTINEZ SOLORZANO

DATE/TIME OF TEST: 03/02/2025 00:24:32

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:15	
Control Test	0.081	00:16	
Air Blank	0.000	00:16	
Subject Sample	# 1 0.204	00:18	
Breath Volume	1.167	00:18	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample	# 2	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.080	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 03/26/2025 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,MARC,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: TRUJILLO,MARC
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 16:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: FRANCE

DATE/TIME OF TEST: 03/08/2025 17:52:24

Diagnostics Check		OK	17:40	
Air Blank		0.000	17:41	
Control Test		0.079	17:41	
Air Blank		0.000	17:41	
Subject Sample	# 1	0.119	17:45	Volume Not Met
Breath Volume		0.113	17:45	
Air Blank		0.000	17:45	
Air Blank		0.000	17:47	
Subject Sample	# 2	0.127	17:50	Volume Not Met
Breath Volume		0.289	17:50	
Air Blank		0.000	17:51	
Control Test		0.077	17:51	
Air Blank		0.000	17:52	
Diagnostics Check		OK	17:52	

Data Download Date/Time: 03/26/2025 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,MARC,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: TRUJILLO,MARC
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 16:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: FRANCE

DATE/TIME OF TEST: 03/08/2025 18:05:42

Diagnostics Check		OK	17:55	
Air Blank		0.000	17:56	
Control Test		0.077	17:56	
Air Blank		0.000	17:57	
Subject Sample	# 1	0.117	18:00	Volume Not Met
Breath Volume		0.222	18:00	
Air Blank		0.000	18:01	
Air Blank		0.000	18:02	
Subject Sample	# 2		18:04	Subject Test Refused
Breath Volume		0.000	18:04	
Air Blank		0.000	18:04	
Control Test		0.075	18:05	
Air Blank		0.000	18:05	
Diagnostics Check		OK	18:05	

Data Download Date/Time: 03/26/2025 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: ALESHIRE,NATHANIEL
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MOELLER

DATE/TIME OF TEST: 03/20/2025 03:58:33

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.173	03:53
Breath Volume	2.160	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.186	03:56
Breath Volume	2.011	03:56
Air Blank	0.000	03:57
Control Test	0.077	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 03/26/2025 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BROSS

DATE/TIME OF TEST: 03/01/2025 01:22:28

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.081	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:15	No Sample Provided
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Air Blank	0.000	01:17	
Subject Sample	# 2	01:21	No Sample Provided
Breath Volume	0.000	01:21	
Air Blank	0.000	01:21	
Control Test	0.081	01:21	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BROSS

DATE/TIME OF TEST: 03/01/2025 01:30:39

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.166	01:26
Breath Volume	1.796	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.171	01:29
Breath Volume	2.453	01:29
Air Blank	0.000	01:29
Control Test	0.081	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DE LA CRUZ,JUAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: CORREA AGUIRRE

DATE/TIME OF TEST: 03/03/2025 01:15:03

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.081	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.146	01:10
Breath Volume	1.703	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.146	01:13
Breath Volume	1.289	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:15

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: OLIVER,SALVADOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: BRAND

DATE/TIME OF TEST: 03/05/2025 01:12:09

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.081	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.067	01:07
Breath Volume	2.183	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.067	01:10
Breath Volume	3.515	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: YOUNGBLOOD

DATE/TIME OF TEST: 03/05/2025 22:44:02

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.241	22:39
Breath Volume	2.960	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:41
Subject Sample # 2	0.242	22:42
Breath Volume	3.226	22:42
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:44

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MARTINEZ PALMA

DATE/TIME OF TEST: 03/06/2025 02:03:44

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.082	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.039	01:59
Breath Volume	1.335	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.041	02:02
Breath Volume	1.171	02:02
Air Blank	0.000	02:02
Control Test	0.082	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCALIA,ADAM,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ROJIK,CAMERON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 03/08/2025 00:01:55

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.151	23:57
Breath Volume	4.351	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.150	00:00
Breath Volume	4.191	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ROGERS,TRISTAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RUTLAND

DATE/TIME OF TEST: 03/08/2025 02:52:35

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.221	02:45
Breath Volume		1.777	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.196	02:47
Breath Volume		2.160	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 3	0.211	02:50
Breath Volume		3.269	02:50
Air Blank		0.000	02:51
Control Test		0.076	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TURNER,GLENN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WIARD

DATE/TIME OF TEST: 03/09/2025 03:05:01

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	03:00
Subject Sample # 1	0.116	03:00
Breath Volume	2.839	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.115	03:03
Breath Volume	2.421	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:05

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: OLIVER,SALVADOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 03/10/2025 01:34:24

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.209	01:29
Breath Volume	1.488	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.209	01:32
Breath Volume	2.121	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: CLANCY

DATE/TIME OF TEST: 03/12/2025 04:16:56

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.082	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:09	No Sample Provided
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.082	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 03/13/2025 19:08:19

Diagnostics Check	OK	19:01
Air Blank	0.000	19:02
Control Test	0.081	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.236	19:03
Breath Volume	3.140	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.239	19:06
Breath Volume	3.890	19:06
Air Blank	0.000	19:07
Control Test	0.081	19:07
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FULTON

DATE/TIME OF TEST: 03/14/2025 01:29:00

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.081	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.066	01:24
Breath Volume		2.398	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.063	01:27
Breath Volume		2.324	01:27
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:29

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: DEGNER

DATE/TIME OF TEST: 03/15/2025 04:05:41

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.211	04:01
Breath Volume	1.933	04:01
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.204	04:04
Breath Volume	1.511	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 03/31/2025 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAYA,ZACHARY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: SAEZ

DATE/TIME OF TEST: 03/15/2025 17:41:35

Diagnostics Check	OK	17:38	
Air Blank	0.000	17:39	
Control Test	0.080	17:39	
Air Blank	0.000	17:40	
Subject Sample	# 1	17:40	Subject Test Refused
Breath Volume	0.000	17:40	
Air Blank	0.000	17:40	
Control Test	0.081	17:40	
Air Blank	0.000	17:41	
Diagnostics Check	OK	17:41	

Data Download Date/Time: 03/31/2025 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: FERREIRA DA SILVA

DATE/TIME OF TEST: 03/16/2025 02:56:44

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.205	02:51
Breath Volume	1.816	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.220	02:55
Breath Volume	1.773	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 03/31/2025 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MIXON,MATTHEW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: DEBLANC

DATE/TIME OF TEST: 03/16/2025 23:33:34

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:31	
Control Test	0.081	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Control Test	0.081	23:32	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 03/31/2025 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: TOUPS

DATE/TIME OF TEST: 03/18/2025 00:08:41

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:04	
Control Test	0.082	00:04	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:08	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 03/31/2025 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDDY,DANIEL,D
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: EDDY,DANIEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 03/19/2025 09:07:15

Diagnostics Check	OK	09:00
Air Blank	0.000	09:01
Control Test	0.081	09:01
Air Blank	0.000	09:02
Subject Sample # 1	0.000	09:02
Breath Volume	3.902	09:02
Air Blank	0.000	09:03
Air Blank	0.000	09:05
Subject Sample # 2	0.000	09:05
Breath Volume	2.308	09:05
Air Blank	0.000	09:06
Control Test	0.081	09:06
Air Blank	0.000	09:07
Diagnostics Check	OK	09:07

Data Download Date/Time: 03/31/2025 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: COREA ELVIR

DATE/TIME OF TEST: 03/22/2025 23:38:21

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.082	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.000	23:34
Breath Volume	1.789	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.000	23:36
Breath Volume	1.582	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 03/31/2025 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ROJIK,CAMERON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: MCCULLOUGH

DATE/TIME OF TEST: 03/25/2025 20:21:23

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:12	
Control Test	0.080	20:12	
Air Blank	0.000	20:13	
Subject Sample	# 1	20:16	No Sample Provided
Breath Volume	0.000	20:16	
Air Blank	0.000	20:16	
Air Blank	0.000	20:18	
Subject Sample	# 2	20:20	Subject Test Refused
Breath Volume	0.000	20:20	
Air Blank	0.000	20:20	
Control Test	0.080	20:20	
Air Blank	0.000	20:21	
Diagnostics Check	OK	20:21	

Data Download Date/Time: 03/31/2025 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MURNER

DATE/TIME OF TEST: 03/27/2025 00:37:08

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.082	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.100	00:32
Breath Volume	1.582	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:34
Subject Sample # 2	0.098	00:35
Breath Volume	2.839	00:35
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 03/31/2025 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,ALEJANDRA,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,ALEJANDRA
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GOMEZ LOPEZ

DATE/TIME OF TEST: 03/15/2025 21:29:50

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.082	21:21
Air Blank	0.000	21:21
Subject Sample # 1	0.200	21:23
Breath Volume	2.476	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.197	21:28
Breath Volume	1.539	21:28
Air Blank	0.000	21:28
Control Test	0.079	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 04/23/2025 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,ALEJANDRA,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,ALEJANDRA
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MENDEZ CALDERON

DATE/TIME OF TEST: 03/15/2025 22:44:08

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:36	
Control Test	0.081	22:36	
Air Blank	0.000	22:37	
Subject Sample	# 1 0.183	22:38	
Breath Volume	1.421	22:38	
Air Blank	0.000	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 2	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.082	22:43	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 04/23/2025 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: MANSUR

DATE/TIME OF TEST: 03/19/2025 16:50:47

Diagnostics Check	OK	16:47	
Air Blank	0.000	16:48	
Control Test	0.082	16:48	
Air Blank	0.000	16:49	
Subject Sample	# 1	16:49	Subject Test Refused
Breath Volume	0.000	16:49	
Air Blank	0.000	16:49	
Control Test	0.081	16:50	
Air Blank	0.000	16:50	
Diagnostics Check	OK	16:50	

Data Download Date/Time: 04/23/2025 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,ALEJANDRA,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,ALEJANDRA
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 03/25/2025 21:58:09

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.083	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.000	21:53
Breath Volume	3.507	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.000	21:56
Breath Volume	2.937	21:56
Air Blank	0.000	21:57
Control Test	0.082	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 04/23/2025 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: ORTIZ RAMIREZ

DATE/TIME OF TEST: 03/07/2025 22:14:48

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:12	
Control Test	0.081	22:12	
Air Blank	0.000	22:13	
Subject Sample	# 1	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:13	
Control Test	0.081	22:14	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:14	

Data Download Date/Time: 03/13/2025 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,ALEJANDRA,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,ALEJANDRA
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 03/08/2025 01:39:26

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.127	01:34
Breath Volume	1.339	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.135	01:37
Breath Volume	1.195	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 03/13/2025 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,MARK,K
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CABRERA,VINCENT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:04

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/13/2025 08:31:19

Diagnostics Check	OK	08:25
Air Blank	0.000	08:25
Control Test	0.080	08:26
Air Blank	0.000	08:26
Subject Sample # 1	0.000	08:27
Breath Volume	1.367	08:27
Air Blank	0.000	08:27
Air Blank	0.000	08:29
Subject Sample # 2	0.000	08:29
Breath Volume	1.496	08:29
Air Blank	0.000	08:30
Control Test	0.081	08:30
Air Blank	0.000	08:31
Diagnostics Check	OK	08:31

Data Download Date/Time: 03/13/2025 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ESTRELLA,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: BARATTINI

DATE/TIME OF TEST: 03/01/2025 02:01:00

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 03/23/2025 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SIDES,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/01/2025 05:20:51

Diagnostics Check		OK	05:14
Air Blank		0.000	05:14
Control Test		0.080	05:14
Air Blank		0.000	05:15
Subject Sample	# 1	0.205	05:16
Breath Volume		1.234	05:16
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.202	05:19
Breath Volume		2.867	05:19
Air Blank		0.000	05:19
Control Test		0.079	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 03/23/2025 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KRAMER,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: DUFRESNE

DATE/TIME OF TEST: 03/02/2025 00:46:01

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.222	00:41
Breath Volume	3.156	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:43
Subject Sample # 2	0.219	00:44
Breath Volume	3.027	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:46

Data Download Date/Time: 03/23/2025 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOON,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DIAZ COLON

DATE/TIME OF TEST: 03/02/2025 03:22:04

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.179	03:17
Breath Volume	1.972	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.174	03:20
Breath Volume	1.851	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

Data Download Date/Time: 03/23/2025 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KRAMER,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: AYALA VICENTE

DATE/TIME OF TEST: 03/03/2025 02:16:35

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.080	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 03/23/2025 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUEVAS,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 03/04/2025 18:58:39

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.080	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.135	18:53
Breath Volume	1.636	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.145	18:57
Breath Volume	2.050	18:57
Air Blank	0.000	18:57
Control Test	0.080	18:58
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 03/23/2025 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,C
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ROSSI

DATE/TIME OF TEST: 03/07/2025 00:59:03

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.070	00:53
Breath Volume		1.667	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.070	00:57
Breath Volume		2.410	00:57
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:59

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,R,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARRIENTOS,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MCGILL

DATE/TIME OF TEST: 03/13/2025 02:46:17

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.096	02:41
Breath Volume	2.117	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.096	02:44
Breath Volume	1.792	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS, ALLYN, J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VENEGAS, SERGIO
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 08:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: DOWNES

DATE/TIME OF TEST: 03/15/2025 09:57:27

Diagnostics Check		OK	09:48	
Air Blank		0.000	09:48	
Control Test		0.080	09:49	
Air Blank		0.000	09:49	
Subject Sample	# 1	0.000	09:52	Volume Not Met
Breath Volume		0.082	09:52	
Air Blank		0.000	09:53	
Air Blank		0.000	09:55	
Subject Sample	# 2		09:56	Subject Test Refused
Breath Volume		0.000	09:56	
Air Blank		0.000	09:56	
Control Test		0.080	09:56	
Air Blank		0.000	09:57	
Diagnostics Check		OK	09:57	

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HUBBARD,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: HERNANDEZ-RAMIREZ

DATE/TIME OF TEST: 03/15/2025 23:28:15

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.081	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.194	23:23
Breath Volume	1.914	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.189	23:26
Breath Volume	1.257	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ESTRELLA,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HAUGLAND

DATE/TIME OF TEST: 03/16/2025 23:27:33

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.150	23:22
Breath Volume	1.394	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.159	23:26
Breath Volume	2.792	23:26
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILAR,RODRIGO,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: MATHEWS,MARY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: NOLAZCO GARCIA

DATE/TIME OF TEST: 03/16/2025 23:51:39

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:49	
Control Test	0.080	23:49	
Air Blank	0.000	23:50	
Subject Sample	# 1	23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:50	
Control Test	0.080	23:51	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,BRITTANY,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: OOLEY,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: QUESNEL

DATE/TIME OF TEST: 03/17/2025 13:22:19

Diagnostics Check	OK	13:15
Air Blank	0.000	13:16
Control Test	0.080	13:16
Air Blank	0.000	13:16
Subject Sample # 1	0.244	13:17
Breath Volume	1.960	13:17
Air Blank	0.000	13:18
Air Blank	0.000	13:20
Subject Sample # 2	0.226	13:20
Breath Volume	1.757	13:20
Air Blank	0.000	13:21
Control Test	0.080	13:21
Air Blank	0.000	13:22
Diagnostics Check	OK	13:22

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,BRITTANY,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CHOUINARD,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 03/17/2025 17:33:44

Diagnostics Check	OK	17:27
Air Blank	0.000	17:27
Control Test	0.081	17:28
Air Blank	0.000	17:28
Subject Sample # 1	0.191	17:29
Breath Volume	2.265	17:29
Air Blank	0.000	17:29
Air Blank	0.000	17:31
Subject Sample # 2	0.188	17:32
Breath Volume	1.710	17:32
Air Blank	0.000	17:32
Control Test	0.081	17:33
Air Blank	0.000	17:33
Diagnostics Check	OK	17:33

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SIDES,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 03/20/2025 21:54:49

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.081	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.000	21:50
Breath Volume	2.703	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.000	21:53
Breath Volume	1.898	21:53
Air Blank	0.000	21:53
Control Test	0.081	21:54
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CPL.,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KRAMER,DEP
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MANGEOT

DATE/TIME OF TEST: 03/20/2025 22:59:43

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:57	
Control Test	0.080	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:59	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LOWELL,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 03/22/2025 05:06:03

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.081	05:04	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:05	
Control Test	0.082	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:06	

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,SEBASTIEN,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GONZALEZ,SEBASTIEN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: WOODDELL

DATE/TIME OF TEST: 03/23/2025 01:51:11

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.131	01:46
Breath Volume		2.183	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.139	01:49
Breath Volume		2.667	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,R,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MAGILL,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/23/2025 05:29:17

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.080	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.156	05:24
Breath Volume	1.226	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:26
Subject Sample # 2	0.160	05:27
Breath Volume	1.925	05:27
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 03/23/2025 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUEVAS,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCGILL,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/24/2025 00:06:32

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.081	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.015	00:02
Breath Volume		2.613	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.015	00:05
Breath Volume		2.597	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 04/28/2025 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,BRITTANY,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/26/2025 11:33:07

Diagnostics Check	OK	11:26
Air Blank	0.000	11:27
Control Test	0.081	11:27
Air Blank	0.000	11:28
Subject Sample # 1	0.000	11:28
Breath Volume	2.886	11:28
Air Blank	0.000	11:29
Air Blank	0.000	11:31
Subject Sample # 2	0.000	11:31
Breath Volume	1.917	11:31
Air Blank	0.000	11:32
Control Test	0.082	11:32
Air Blank	0.000	11:33
Diagnostics Check	OK	11:33

Data Download Date/Time: 04/28/2025 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ESTRELLA,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/23/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: SCACCIO

DATE/TIME OF TEST: 03/29/2025 00:46:16

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.315	00:41
Breath Volume	2.582	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.315	00:44
Breath Volume	2.464	00:44
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 04/28/2025 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ORTEGA,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: SAFI

DATE/TIME OF TEST: 03/29/2025 02:05:17

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:04	
Control Test	0.081	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 04/28/2025 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MARTIN,Q
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/29/2025 02:54:35

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.081	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.007	02:50
Breath Volume	2.894	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.007	02:53
Breath Volume	2.761	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 04/28/2025 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,C
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MONTROSS

DATE/TIME OF TEST: 03/29/2025 03:27:36

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.060	03:23
Breath Volume	2.531	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.059	03:26
Breath Volume	1.722	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 04/28/2025 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DESCLOS

DATE/TIME OF TEST: 03/31/2025 01:31:14

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.080	01:29	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 04/28/2025 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALES,ANGEL,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 03/31/2025 23:15:29

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.081	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.190	23:10
Breath Volume	2.820	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.180	23:13
Breath Volume	2.097	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 04/28/2025 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: COLGATE

DATE/TIME OF TEST: 03/21/2025 18:53:44

Diagnostics Check	OK	18:47
Air Blank	0.000	18:47
Control Test	0.078	18:48
Air Blank	0.000	18:48
Subject Sample # 1	0.093	18:49
Breath Volume	1.750	18:49
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 2	0.086	18:52
Breath Volume	1.902	18:52
Air Blank	0.000	18:52
Control Test	0.078	18:53
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/21/2025 22:23:03

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.078	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.065	22:18
Breath Volume	2.027	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.063	22:21
Breath Volume	2.312	22:21
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:23

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE, RAYCHEL, E
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: PEREZ, ORLANDO
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: GARCIA BECERRA

DATE/TIME OF TEST: 03/22/2025 04:40:18

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.157	04:35
Breath Volume	3.699	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.149	04:38
Breath Volume	3.433	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES GONZALEZ,PEDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REYES GONZALEZ,PEDRO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 03/22/2025 05:16:34

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.079	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:15	
Control Test	0.079	05:15	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STROUD,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: KILLEA

DATE/TIME OF TEST: 03/23/2025 01:35:58

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.078	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.201	01:30
Breath Volume	3.476	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.202	01:34
Breath Volume	2.464	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES GONZALEZ,PEDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REYES GONZALEZ,PEDRO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: VARGAS OSORIA

DATE/TIME OF TEST: 03/24/2025 02:53:10

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.155	02:48
Breath Volume	1.636	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.165	02:51
Breath Volume	2.554	02:51
Air Blank	0.000	02:52
Control Test	0.075	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GOUDE

DATE/TIME OF TEST: 03/24/2025 03:03:08

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.077	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.154	02:58
Breath Volume	2.582	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.153	03:01
Breath Volume	1.867	03:01
Air Blank	0.000	03:02
Control Test	0.077	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BEATY,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: PLANK

DATE/TIME OF TEST: 03/26/2025 03:31:42

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.166	03:27
Breath Volume	2.644	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.164	03:30
Breath Volume	2.664	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACEY, BRANDON, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY, DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 03/27/2025 03:31:24

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,DAN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: ZARZECKI

DATE/TIME OF TEST: 03/27/2025 18:12:21

Diagnostics Check	OK	17:59	
Air Blank	0.000	17:59	
Control Test	0.078	18:00	
Air Blank	0.000	18:00	
Subject Sample	# 1 0.194	18:03	Volume Not Met
Breath Volume	0.179	18:03	
Air Blank	0.000	18:04	
Air Blank	0.000	18:06	
Subject Sample	# 2 0.194	18:07	
Breath Volume	1.246	18:07	
Air Blank	0.000	18:07	
Air Blank	0.000	18:09	
Subject Sample	# 3 0.194	18:10	
Breath Volume	1.289	18:10	
Air Blank	0.000	18:11	
Control Test	0.076	18:11	
Air Blank	0.000	18:12	
Diagnostics Check	OK	18:12	

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STROUD,JACKSON
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: KIMBLE

DATE/TIME OF TEST: 03/27/2025 21:43:22

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:33	
Control Test	0.078	21:33	
Air Blank	0.000	21:34	
Subject Sample	# 1 0.265	21:37	Volume Not Met
Breath Volume	0.160	21:37	
Air Blank	0.000	21:38	
Air Blank	0.000	21:39	
Subject Sample	# 2	21:41	Subject Test Refused
Breath Volume	0.000	21:41	
Air Blank	0.000	21:42	
Control Test	0.077	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 04/04/2025 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESNEY,KYLE,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: STOCKTON

DATE/TIME OF TEST: 03/28/2025 02:02:43

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.132	01:58
Breath Volume	1.914	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample # 2	0.130	02:01
Breath Volume	1.750	02:01
Air Blank	0.000	02:01
Control Test	0.078	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUGAN,KASEY,L
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: BANASIAK,MICHAEL
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SWEARINGEN

DATE/TIME OF TEST: 03/28/2025 03:32:01

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.138	03:27
Breath Volume	2.972	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.129	03:30
Breath Volume	2.808	03:30
Air Blank	0.000	03:31
Control Test	0.077	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:32

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERSTEN,DANIEL,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: NERSTEN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MLADENOVA

DATE/TIME OF TEST: 03/28/2025 23:58:30

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.078	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.197	23:54
Breath Volume	2.445	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.190	23:56
Breath Volume	2.144	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 03/29/2025 02:08:25

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.077	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:06	Slope Not Met
Breath Volume	1.343	02:06	
Air Blank	0.000	02:07	
Control Test	0.078	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 03/29/2025 02:18:22

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.077	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.238	02:13
Breath Volume		3.593	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.234	02:16
Breath Volume		3.332	02:16
Air Blank		0.000	02:17
Control Test		0.075	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,R,
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: MARTINEZ,O
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DE ALMEIDA CAMPOS JR

DATE/TIME OF TEST: 03/29/2025 03:39:10

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.078	03:30	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:33	No Sample Provided
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample	# 2	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.079	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: DISHER

DATE/TIME OF TEST: 03/30/2025 01:37:18

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.078	01:32
Breath Volume	1.503	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.079	01:35
Breath Volume	2.289	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JONES,NICHOLAS
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: TIPTON

DATE/TIME OF TEST: 03/30/2025 02:05:47

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.078	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.182	02:00
Breath Volume		1.828	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.188	02:04
Breath Volume		1.476	02:04
Air Blank		0.000	02:04
Control Test		0.078	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARSTON,KAITLYN
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: HEALY

DATE/TIME OF TEST: 03/30/2025 03:57:30

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.077	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.202	03:53
Breath Volume		1.589	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.206	03:55
Breath Volume		2.089	03:55
Air Blank		0.000	03:56
Control Test		0.077	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JONES,NICHOLAS
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: CORDOVA AGUILAR

DATE/TIME OF TEST: 03/30/2025 04:16:23

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.078	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.031	04:11
Breath Volume	1.648	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.031	04:14
Breath Volume	1.761	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: WANGEN

DATE/TIME OF TEST: 03/30/2025 16:30:30

Diagnostics Check	OK	16:24
Air Blank	0.000	16:24
Control Test	0.079	16:25
Air Blank	0.000	16:25
Subject Sample # 1	0.184	16:26
Breath Volume	3.710	16:26
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.190	16:28
Breath Volume	3.429	16:28
Air Blank	0.000	16:29
Control Test	0.078	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SATTERFIELD

DATE/TIME OF TEST: 03/31/2025 20:45:47

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.079	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.171	20:41
Breath Volume	3.371	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:43
Subject Sample # 2	0.182	20:44
Breath Volume	1.621	20:44
Air Blank	0.000	20:44
Control Test	0.079	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 04/04/2025 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPOTO, MATTHEW, R

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: SPOTO, MATTHEW

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 04125080A5

DRY GAS EXPIRY DATE: 03/05/2027

VIOLATION CODE: DUI

ARREST TIME: 01:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: AGUERO

DATE/TIME OF TEST: 03/29/2025 01:55:57

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.141	01:51
Breath Volume	2.695	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.143	01:54
Breath Volume	1.953	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 04/28/2025 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SPOTO, MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SANTIS HERNANDEZ

DATE/TIME OF TEST: 03/29/2025 18:31:06

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.081	18:24
Air Blank	0.000	18:25
Subject Sample # 1	0.304	18:26
Breath Volume	1.589	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.300	18:29
Breath Volume	1.921	18:29
Air Blank	0.000	18:30
Control Test	0.080	18:30
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 04/28/2025 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLLER,LOGAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: KOLLER,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GRANADOS

DATE/TIME OF TEST: 03/04/2025 00:29:22

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.218	00:24
Breath Volume		3.656	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:26
Subject Sample	# 2	0.202	00:27
Breath Volume		3.058	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 04/04/2025 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,HECTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,HECTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 03/08/2025 02:43:34

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.000	02:39
Breath Volume	1.902	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.000	02:42
Breath Volume	1.773	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 04/04/2025 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS, TYLER, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: HANOVER

DATE/TIME OF TEST: 03/26/2025 19:23:48

Diagnostics Check	OK	19:17
Air Blank	0.000	19:18
Control Test	0.080	19:18
Air Blank	0.000	19:18
Subject Sample # 1	0.000	19:19
Breath Volume	1.617	19:19
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.000	19:22
Breath Volume	1.492	19:22
Air Blank	0.000	19:22
Control Test	0.080	19:23
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 04/04/2025 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,BRADLEY,I
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JONES,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 03/31/2025 22:52:39

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.079	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.142	22:48
Breath Volume	3.054	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.133	22:51
Breath Volume	4.035	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 04/04/2025 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINMETZ,J,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MCCARTY,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/09/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: RAMIREZ-MENDEZ

DATE/TIME OF TEST: 03/09/2025 21:53:26

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.079	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.189	21:49
Breath Volume	1.394	21:49
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.180	21:51
Breath Volume	1.296	21:51
Air Blank	0.000	21:52
Control Test	0.078	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 04/24/2025 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELBOGIS, DAVID, R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: WILLIAMSON, STEVEN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012408082
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/02/2025 00:00:21

Diagnostics Check		OK	23:48	
Air Blank		0.000	23:48	
Control Test		0.081	23:49	
Air Blank		0.000	23:49	
Subject Sample	# 1	0.170	23:52	Volume Not Met
Breath Volume		0.339	23:52	
Air Blank		0.000	23:53	
Air Blank		0.000	23:55	
Subject Sample	# 2	0.147	23:58	Volume Not Met
Breath Volume		0.503	23:58	
Air Blank		0.000	23:59	
Control Test		0.078	23:59	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 03/30/2025 17:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELBOGIS, DAVID, R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DUNCAN, NICHOLAS
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012408082
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 03/03/2025 01:50:25

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.082	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.210	01:45
Breath Volume	1.738	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.201	01:48
Breath Volume	2.117	01:48
Air Blank	0.000	01:49
Control Test	0.082	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 03/30/2025 17:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELBOGIS, DAVID, R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: GOODSON, STACY
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012408082
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GUTERL

DATE/TIME OF TEST: 03/03/2025 02:07:10

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.082	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.264	02:02
Breath Volume	1.734	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.266	02:05
Breath Volume	1.527	02:05
Air Blank	0.000	02:06
Control Test	0.082	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 03/30/2025 17:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ROBERT,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012408082
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: KEITH

DATE/TIME OF TEST: 03/08/2025 00:21:07

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.081	00:19	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:20	
Control Test	0.081	00:20	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 03/30/2025 17:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JAIRUS,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCCONNELL,JAIRUS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012408082
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: DAVEAU

DATE/TIME OF TEST: 03/10/2025 22:22:28

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.082	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.000	22:18
Breath Volume	1.722	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.000	22:21
Breath Volume	2.050	22:21
Air Blank	0.000	22:21
Control Test	0.082	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 03/30/2025 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEYES,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KEYES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012408082
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CORBIN

DATE/TIME OF TEST: 03/13/2025 01:52:25

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.082	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.159	01:47
Breath Volume	3.699	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.159	01:50
Breath Volume	3.558	01:50
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 03/30/2025 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELLERT, DAVID, C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: CARTER, VICTORIA
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012408082
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: LAMAR

DATE/TIME OF TEST: 03/14/2025 21:21:48

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.081	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.190	21:16
Breath Volume	1.718	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.197	21:19
Breath Volume	1.621	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:21
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 03/30/2025 17:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELBOGIS, DAVID, R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: GAFFEY, TUCKER
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012408082
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/15/2025 17:52:45

Diagnostics Check		OK	17:38	
Air Blank		0.000	17:38	
Control Test		0.081	17:39	
Air Blank		0.000	17:39	
Subject Sample	# 1	0.196	17:42	Volume Not Met
Breath Volume		0.191	17:42	
Air Blank		0.000	17:43	
Air Blank		0.000	17:45	
Subject Sample	# 2	0.201	17:47	
Breath Volume		1.187	17:47	
Air Blank		0.000	17:48	
Air Blank		0.000	17:49	
Subject Sample	# 3	0.195	17:51	
Breath Volume		1.160	17:51	
Air Blank		0.000	17:51	
Control Test		0.078	17:52	
Air Blank		0.000	17:52	
Diagnostics Check		OK	17:52	

Data Download Date/Time: 03/30/2025 17:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEYES,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KEYES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012408082
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 03/21/2025 00:12:37

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.081	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.163	00:07
Breath Volume	1.292	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.158	00:11
Breath Volume	2.343	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 03/30/2025 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,KEVIN,R
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: HERNANDEZ,RAFAEL
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/15/2025 00:37:00

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.081	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:32	No Sample Provided
Breath Volume	0.000	00:32	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample	# 2	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:36	
Control Test	0.080	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 03/25/2025 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHMALL, NICHOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: DONOVAN II

DATE/TIME OF TEST: 03/01/2025 00:53:56

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.084	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.207	00:48
Breath Volume		1.449	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.190	00:52
Breath Volume		1.328	00:52
Air Blank		0.000	00:53
Control Test		0.082	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOUZA,MAYKOLL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MERASSAINT

DATE/TIME OF TEST: 03/01/2025 01:47:02

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.083	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.181	01:42
Breath Volume	1.265	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.184	01:45
Breath Volume	2.011	01:45
Air Blank	0.000	01:46
Control Test	0.083	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:47

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATIAS, ROLANDO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: VILLASENOR CURIEL

DATE/TIME OF TEST: 03/01/2025 02:00:13

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.083	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.179	01:55
Breath Volume	4.335	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.176	01:58
Breath Volume	3.957	01:58
Air Blank	0.000	01:59
Control Test	0.083	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MARQUEZARDILA

DATE/TIME OF TEST: 03/01/2025 02:25:26

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.083	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.145	02:20
Breath Volume	3.238	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.128	02:23
Breath Volume	1.253	02:23
Air Blank	0.000	02:24
Control Test	0.082	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LIBAG

DATE/TIME OF TEST: 03/01/2025 02:51:33

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.083	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.192	02:47
Breath Volume	1.164	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.194	02:50
Breath Volume	1.636	02:50
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OGLETREE,JERRELL
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MOTA ALVES DOS SANTO

DATE/TIME OF TEST: 03/01/2025 03:13:30

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.083	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.136	03:08
Breath Volume	1.171	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.135	03:12
Breath Volume	2.207	03:12
Air Blank	0.000	03:12
Control Test	0.083	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARDY,BRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: ROSAS

DATE/TIME OF TEST: 03/01/2025 03:59:02

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.083	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.145	03:54
Breath Volume	2.289	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.149	03:57
Breath Volume	3.273	03:57
Air Blank	0.000	03:58
Control Test	0.083	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:59

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOEL, KENSON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: ESTIME

DATE/TIME OF TEST: 03/01/2025 21:14:48

Diagnostics Check		OK	21:03	
Air Blank		0.000	21:03	
Control Test		0.083	21:04	
Air Blank		0.000	21:04	
Subject Sample	# 1	0.147	21:07	Volume Not Met
Breath Volume		0.347	21:07	
Air Blank		0.000	21:08	
Air Blank		0.000	21:10	
Subject Sample	# 2	0.146	21:13	Volume Not Met
Breath Volume		0.214	21:13	
Air Blank		0.000	21:13	
Control Test		0.080	21:14	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOEL, KENSON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: ESTIME

DATE/TIME OF TEST: 03/01/2025 21:23:38

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.081	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.152	21:19
Breath Volume	1.453	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.144	21:22
Breath Volume	1.144	21:22
Air Blank	0.000	21:22
Control Test	0.082	21:23
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE, JUNIOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: REYNOSO MORALES

DATE/TIME OF TEST: 03/01/2025 22:02:16

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.083	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.110	21:57
Breath Volume	1.820	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.110	22:00
Breath Volume	1.621	22:00
Air Blank	0.000	22:01
Control Test	0.081	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BISPO,STEPHANIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: HANKS

DATE/TIME OF TEST: 03/01/2025 23:20:42

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.082	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.164	23:16
Breath Volume	1.707	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.163	23:19
Breath Volume	1.714	23:19
Air Blank	0.000	23:19
Control Test	0.083	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALTINERLIELMAS,ADEM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/02/2025 01:50:56

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.083	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.129	01:46
Breath Volume	1.394	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.130	01:49
Breath Volume	2.464	01:49
Air Blank	0.000	01:50
Control Test	0.082	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: TICAS

DATE/TIME OF TEST: 03/02/2025 03:55:08

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.083	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.162	03:50
Breath Volume	2.464	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.169	03:53
Breath Volume	1.601	03:53
Air Blank	0.000	03:54
Control Test	0.082	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 03/20/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRAY,SPENCER
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: LAYNES JACOBO

DATE/TIME OF TEST: 03/02/2025 18:27:31

Diagnostics Check	OK	18:21
Air Blank	0.000	18:22
Control Test	0.084	18:22
Air Blank	0.000	18:22
Subject Sample # 1	0.279	18:23
Breath Volume	2.718	18:23
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.276	18:26
Breath Volume	2.703	18:26
Air Blank	0.000	18:26
Control Test	0.083	18:26
Air Blank	0.000	18:27
Diagnostics Check	OK	18:27

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES II,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/03/2025 02:39:46

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.083	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.162	02:35
Breath Volume	1.335	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.163	02:38
Breath Volume	1.382	02:38
Air Blank	0.000	02:38
Control Test	0.083	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA, PEDRO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 03/03/2025 04:17:55

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:15	
Control Test	0.083	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.083	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: APONTE,MIGUEL
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VASUDEV KAMATH

DATE/TIME OF TEST: 03/04/2025 03:22:01

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.084	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.200	03:17
Breath Volume		1.382	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.195	03:20
Breath Volume		1.144	03:20
Air Blank		0.000	03:21
Control Test		0.082	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:22

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: JORGE ARMANDO

DATE/TIME OF TEST: 03/04/2025 03:39:52

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.082	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1	0.200	03:32	Volume Not Met
Breath Volume		0.015	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 2	0.169	03:38	Volume Not Met
Breath Volume		0.121	03:38	
Air Blank		0.000	03:38	
Control Test		0.080	03:39	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOWLING,PETER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: HERRERA ANGARICA

DATE/TIME OF TEST: 03/04/2025 21:33:13

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.083	21:28
Air Blank	0.000	21:28
Subject Sample # 1	0.000	21:29
Breath Volume	1.898	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.000	21:31
Breath Volume	2.148	21:31
Air Blank	0.000	21:32
Control Test	0.082	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 03/04/2025 22:30:11

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:19	
Control Test	0.084	22:19	
Air Blank	0.000	22:19	
Subject Sample	# 1	22:23	No Sample Provided
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample	# 2	0.000	22:25
Breath Volume	2.199	22:25	
Air Blank	0.000	22:26	
Air Blank	0.000	22:28	
Subject Sample	# 3	0.000	22:28
Breath Volume	1.960	22:28	
Air Blank	0.000	22:29	
Control Test	0.083	22:29	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANCHEZ,JUAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: CULVER

DATE/TIME OF TEST: 03/07/2025 22:11:49

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.084	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.225	22:07
Breath Volume	2.753	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:09
Subject Sample # 2	0.230	22:10
Breath Volume	2.890	22:10
Air Blank	0.000	22:10
Control Test	0.083	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCMORROW, RYAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 03/07/2025 22:58:12

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.084	22:46	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:50	No Sample Provided
Breath Volume	0.000	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 2	0.218	22:53
Breath Volume	2.707	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 3	0.229	22:56
Breath Volume	2.449	22:56	
Air Blank	0.000	22:57	
Control Test	0.083	22:57	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALLEN,TALBOT
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 03/08/2025 00:45:41

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.084	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:44	Subject Test Refused
Breath Volume	0.000	00:44	
Air Blank	0.000	00:44	
Control Test	0.084	00:45	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALLEN,TALBOT
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 03/08/2025 00:57:22

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.085	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.168	00:51
Breath Volume	1.738	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.162	00:55
Breath Volume	1.437	00:55
Air Blank	0.000	00:56
Control Test	0.083	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO,SERGIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ALBELO CORDERO

DATE/TIME OF TEST: 03/08/2025 01:35:09

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.084	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:25	No Sample Provided
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 2	0.168	01:28
Breath Volume	1.593	01:28	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample	# 3	01:33	No Sample Provided
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.083	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO,SERGIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ALBELO CORDERO

DATE/TIME OF TEST: 03/08/2025 01:43:36

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.083	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.160	01:39
Breath Volume	1.457	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.155	01:42
Breath Volume	1.558	01:42
Air Blank	0.000	01:42
Control Test	0.083	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/08/2025 02:43:00

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.085	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.000	02:38
Breath Volume	2.710	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.000	02:41
Breath Volume	1.832	02:41
Air Blank	0.000	02:42
Control Test	0.084	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:43

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LARA,JORDAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BERNARDINO

DATE/TIME OF TEST: 03/08/2025 04:02:32

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.083	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.199	03:57
Breath Volume	2.175	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.209	04:01
Breath Volume	1.781	04:01
Air Blank	0.000	04:01
Control Test	0.083	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: RANGEL

DATE/TIME OF TEST: 03/08/2025 05:13:18

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.084	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.097	05:08
Breath Volume	1.601	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.098	05:11
Breath Volume	1.562	05:11
Air Blank	0.000	05:12
Control Test	0.084	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHEWS,ZACHARY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/08/2025 20:21:00

Diagnostics Check	OK	20:18	
Air Blank	0.000	20:18	
Control Test	0.085	20:19	
Air Blank	0.000	20:19	
Subject Sample	# 1	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:20	
Control Test	0.084	20:20	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:21	

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAMANNA,MATTHEW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BLAZICK

DATE/TIME OF TEST: 03/08/2025 22:48:31

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:35	
Control Test		0.084	22:36	
Air Blank		0.000	22:36	
Subject Sample	# 1	0.181	22:39	Volume Not Met
Breath Volume		1.058	22:39	
Air Blank		0.000	22:40	
Air Blank		0.000	22:42	
Subject Sample	# 2	0.227	22:43	
Breath Volume		1.226	22:43	
Air Blank		0.000	22:44	
Air Blank		0.000	22:45	
Subject Sample	# 3	0.207	22:47	
Breath Volume		1.523	22:47	
Air Blank		0.000	22:47	
Control Test		0.082	22:47	
Air Blank		0.000	22:48	
Diagnostics Check		OK	22:48	

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHIPLEY, DANIEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 03/09/2025 22:53:07

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.084	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.160	22:48
Breath Volume	1.722	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.161	22:51
Breath Volume	1.136	22:51
Air Blank	0.000	22:52
Control Test	0.083	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SMITH,JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: LAMASTUS

DATE/TIME OF TEST: 03/10/2025 02:06:01

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.083	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.215	02:01
Breath Volume		1.425	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.212	02:04
Breath Volume		1.597	02:04
Air Blank		0.000	02:05
Control Test		0.082	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:06

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAEZ,BAYRON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: HOSEIN

DATE/TIME OF TEST: 03/10/2025 02:27:50

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.083	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.102	02:22
Breath Volume		1.312	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.100	02:26
Breath Volume		1.960	02:26
Air Blank		0.000	02:26
Control Test		0.082	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA, PEDRO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: VILLEGAS VALLEJO

DATE/TIME OF TEST: 03/10/2025 04:35:38

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.082	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.209	04:31
Breath Volume	2.726	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.207	04:34
Breath Volume	3.253	04:34
Air Blank	0.000	04:34
Control Test	0.082	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 03/20/2025 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ESPINAL,KENDI
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 03/11/2025 00:27:32

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.083	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.242	00:22
Breath Volume	1.582	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.243	00:26
Breath Volume	2.640	00:26
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARBONE,ANDRES
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: HEIDMANN

DATE/TIME OF TEST: 03/11/2025 04:18:49

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.084	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.183	04:14
Breath Volume	4.531	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.180	04:17
Breath Volume	4.308	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOEL, KENSON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: MENDES OLIVEIRA

DATE/TIME OF TEST: 03/11/2025 20:11:38

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:05	
Control Test	0.083	20:05	
Air Blank	0.000	20:06	
Subject Sample # 1	0.349	20:07	
Breath Volume	1.718	20:07	
Air Blank	0.000	20:08	
Air Blank	0.000	20:09	
Subject Sample # 2		20:10	Slope Not Met
Breath Volume	1.351	20:10	
Air Blank	0.000	20:10	
Control Test	0.084	20:11	
Air Blank	0.000	20:11	
Diagnostics Check	OK	20:11	

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOEL, KENSON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: MENDES OLIVEIRA

DATE/TIME OF TEST: 03/11/2025 20:42:58

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.083	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.350	20:38
Breath Volume	1.648	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.337	20:41
Breath Volume	1.652	20:41
Air Blank	0.000	20:42
Control Test	0.082	20:42
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 03/12/2025 04:03:20

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.083	04:00	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.083	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ORTIZ USHPA

DATE/TIME OF TEST: 03/14/2025 03:51:30

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.084	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.190	03:47
Breath Volume	3.156	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.189	03:50
Breath Volume	3.050	03:50
Air Blank	0.000	03:50
Control Test	0.082	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARZA,MANUEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FLEURVILLE

DATE/TIME OF TEST: 03/14/2025 23:30:52

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.083	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.212	23:26
Breath Volume	1.496	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.212	23:29
Breath Volume	1.406	23:29
Air Blank	0.000	23:29
Control Test	0.083	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONTE,COREY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: PAGLIUCA

DATE/TIME OF TEST: 03/15/2025 01:16:44

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.083	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.201	01:11
Breath Volume		2.890	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.194	01:15
Breath Volume		3.007	01:15
Air Blank		0.000	01:15
Control Test		0.082	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES, BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ZAMARRIPA CANCHOLA

DATE/TIME OF TEST: 03/15/2025 02:43:26

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.084	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.246	02:39
Breath Volume		2.410	02:39
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.251	02:41
Breath Volume		2.996	02:41
Air Blank		0.000	02:42
Control Test		0.082	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRISINTE,ANGELO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: BOTTORFF

DATE/TIME OF TEST: 03/15/2025 19:36:24

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.083	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.000	19:32
Breath Volume	2.378	19:32
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.000	19:35
Breath Volume	2.058	19:35
Air Blank	0.000	19:35
Control Test	0.084	19:35
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRUITT, JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GUADALUPE

DATE/TIME OF TEST: 03/15/2025 21:24:20

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:13	
Control Test	0.083	21:13	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:17	No Sample Provided
Breath Volume	0.000	21:17	
Air Blank	0.000	21:17	
Air Blank	0.000	21:19	
Subject Sample	# 2	21:22	No Sample Provided
Breath Volume	0.000	21:22	
Air Blank	0.000	21:23	
Control Test	0.083	21:23	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEL TORO,MICHAEL
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 03/15/2025 22:53:43

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.083	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1 0.125	22:41	
Breath Volume	1.421	22:41	
Air Blank	0.000	22:42	
Air Blank	0.000	22:44	
Subject Sample	# 2	22:47	No Sample Provided
Breath Volume	0.000	22:47	
Air Blank	0.000	22:47	
Air Blank	0.000	22:49	
Subject Sample	# 3 0.123	22:52	
Breath Volume	1.843	22:52	
Air Blank	0.000	22:52	
Control Test	0.082	22:53	
Air Blank	0.000	22:53	
Diagnostics Check	OK	22:53	

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAAVEDRA, NIKOLAS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: FARLOW JR

DATE/TIME OF TEST: 03/16/2025 02:28:33

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.083	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.240	02:21
Breath Volume		2.445	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.261	02:24
Breath Volume		2.847	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 3	0.245	02:27
Breath Volume		2.492	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NEWMAN, PRESTON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: BAGGETT

DATE/TIME OF TEST: 03/16/2025 06:10:22

Diagnostics Check	OK	06:01	
Air Blank	0.000	06:02	
Control Test	0.082	06:02	
Air Blank	0.000	06:02	
Subject Sample	# 1	06:06	No Sample Provided
Breath Volume	0.000	06:06	
Air Blank	0.000	06:06	
Air Blank	0.000	06:08	
Subject Sample	# 2	06:09	Subject Test Refused
Breath Volume	0.000	06:09	
Air Blank	0.000	06:09	
Control Test	0.082	06:09	
Air Blank	0.000	06:10	
Diagnostics Check	OK	06:10	

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PUTMAN,FORREST
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: BERROSTORRES

DATE/TIME OF TEST: 03/16/2025 07:12:31

Diagnostics Check	OK	07:04
Air Blank	0.000	07:05
Control Test	0.084	07:05
Air Blank	0.000	07:06
Subject Sample # 1	0.168	07:06
Breath Volume	1.667	07:06
Air Blank	0.000	07:07
Air Blank	0.000	07:08
Subject Sample # 2	0.158	07:11
Breath Volume	2.230	07:11
Air Blank	0.000	07:11
Control Test	0.084	07:11
Air Blank	0.000	07:12
Diagnostics Check	OK	07:12

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RINCON,RENZO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: ROJASGONZALEZ

DATE/TIME OF TEST: 03/16/2025 07:56:22

Diagnostics Check	OK	07:49
Air Blank	0.000	07:50
Control Test	0.083	07:50
Air Blank	0.000	07:51
Subject Sample # 1	0.178	07:51
Breath Volume	1.675	07:51
Air Blank	0.000	07:52
Air Blank	0.000	07:53
Subject Sample # 2	0.170	07:54
Breath Volume	2.613	07:54
Air Blank	0.000	07:55
Control Test	0.083	07:55
Air Blank	0.000	07:56
Diagnostics Check	OK	07:56

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 03/17/2025 02:24:13

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.085	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.086	02:18
Breath Volume	1.191	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.082	02:22
Breath Volume	1.808	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIELS,DAMON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: BOTERO MUNOZ

DATE/TIME OF TEST: 03/17/2025 04:04:18

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.083	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.198	03:56
Breath Volume	1.753	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.220	03:59
Breath Volume	2.945	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 3	0.188	04:02
Breath Volume	1.246	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PUTMAN,FOREST
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: OSORIO ARAGOZA

DATE/TIME OF TEST: 03/17/2025 04:36:49

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.082	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.093	04:32
Breath Volume	1.773	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:34
Subject Sample # 2	0.088	04:35
Breath Volume	2.550	04:35
Air Blank	0.000	04:35
Control Test	0.082	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/17/2025 05:01:38

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.083	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.144	04:57
Breath Volume	1.933	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.161	05:00
Breath Volume	2.773	05:00
Air Blank	0.000	05:00
Control Test	0.082	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSA,BRANDON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: ALVITES NEIRA

DATE/TIME OF TEST: 03/18/2025 01:50:33

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.084	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.106	01:45
Breath Volume	1.687	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.111	01:49
Breath Volume	1.621	01:49
Air Blank	0.000	01:49
Control Test	0.082	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,LORENZO
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: BONDY

DATE/TIME OF TEST: 03/18/2025 23:00:36

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.084	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.176	22:55
Breath Volume	3.933	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.169	22:59
Breath Volume	3.566	22:59
Air Blank	0.000	22:59
Control Test	0.084	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 03/20/2025 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: FORTMANN

DATE/TIME OF TEST: 03/19/2025 03:19:40

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.083	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:12	No Sample Provided
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 2	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.083	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 03/20/2025 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI, ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: RAMOS RIVERA

DATE/TIME OF TEST: 03/19/2025 04:38:17

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.083	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.211	04:33
Breath Volume		2.011	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.228	04:36
Breath Volume		2.949	04:36
Air Blank		0.000	04:37
Control Test		0.083	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 03/20/2025 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ, MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: PEREZ BRAVO

DATE/TIME OF TEST: 03/19/2025 21:50:46

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.084	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.247	21:46
Breath Volume	1.207	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:48
Subject Sample # 2	0.248	21:49
Breath Volume	1.824	21:49
Air Blank	0.000	21:49
Control Test	0.083	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 03/20/2025 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: INMAN

DATE/TIME OF TEST: 03/19/2025 22:49:58

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.084	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.132	22:45
Breath Volume	1.375	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.131	22:48
Breath Volume	2.675	22:48
Air Blank	0.000	22:49
Control Test	0.084	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 03/20/2025 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 03/21/2025 21:41:27

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:40	
Control Test	0.088	21:40	Control Outside Tolerance
Air Blank	0.000	21:41	

Data Download Date/Time: 04/22/2025 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 03/21/2025 21:49:38

Diagnostics Check	OK	21:43
Air Blank	0.000	21:44
Control Test	0.085	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.190	21:45
Breath Volume	2.460	21:45
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.194	21:48
Breath Volume	2.367	21:48
Air Blank	0.000	21:48
Control Test	0.084	21:49
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 04/22/2025 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO,SERGIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: VALLEY GARCIA

DATE/TIME OF TEST: 03/23/2025 03:59:55

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.084	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.163	03:53
Breath Volume	1.871	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.153	03:58
Breath Volume	1.500	03:58
Air Blank	0.000	03:59
Control Test	0.084	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 04/22/2025 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCMORROW,RYAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: MARTINEZ CRUZ

DATE/TIME OF TEST: 03/23/2025 18:39:02

Diagnostics Check	OK	18:32
Air Blank	0.000	18:33
Control Test	0.084	18:33
Air Blank	0.000	18:33
Subject Sample # 1	0.213	18:34
Breath Volume	1.460	18:34
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample # 2	0.211	18:37
Breath Volume	1.242	18:37
Air Blank	0.000	18:38
Control Test	0.084	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:39

Data Download Date/Time: 04/22/2025 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/24/2025 01:15:53

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.084	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.000	01:11
Breath Volume	1.875	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.000	01:14
Breath Volume	2.359	01:14
Air Blank	0.000	01:15
Control Test	0.083	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 04/22/2025 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAYES,ARTHUR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/24/2025 21:43:58

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.216	21:38
Breath Volume	1.343	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.226	21:42
Breath Volume	1.738	21:42
Air Blank	0.000	21:43
Control Test	0.079	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 04/22/2025 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP, MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/02/2025
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: REMSBERG

DATE/TIME OF TEST: 03/12/2025 23:03:44

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.203	22:58
Breath Volume	3.078	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.210	23:01
Breath Volume	2.902	23:01
Air Blank	0.000	23:02
Control Test	0.082	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 04/22/2025 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ESQUIVEL,A
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 03/01/2025 15:19:58

Diagnostics Check	OK	15:12	
Air Blank	0.000	15:13	
Control Test	0.082	15:13	
Air Blank	0.000	15:13	
Subject Sample	# 1	15:17	No Sample Provided
Breath Volume	0.000	15:17	
Air Blank	0.000	15:17	
Air Blank		15:19	RFI Detect
Air Blank		15:19	RFI Detect
Air Blank	0.000	15:19	

Data Download Date/Time: 03/30/2025 02:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ESQUIVEL,A
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 03/01/2025 15:29:28

Diagnostics Check	OK	15:26	
Air Blank	0.000	15:27	
Control Test	0.083	15:27	
Air Blank	0.000	15:27	
Subject Sample	# 1	15:28	Subject Test Refused
Breath Volume	0.000	15:28	
Air Blank	0.000	15:28	
Control Test	0.083	15:28	
Air Blank	0.000	15:29	
Diagnostics Check	OK	15:29	

Data Download Date/Time: 03/30/2025 02:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HIDERHOFER,E
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: PATISHTAN MENDEZ

DATE/TIME OF TEST: 03/02/2025 05:19:58

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.082	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.178	05:15
Breath Volume		2.925	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:17
Subject Sample	# 2	0.179	05:18
Breath Volume		2.152	05:18
Air Blank		0.000	05:19
Control Test		0.081	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 03/30/2025 02:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FROELICH,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 03/03/2025 00:26:09

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.083	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.096	00:21
Breath Volume	1.953	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.090	00:24
Breath Volume	1.300	00:24
Air Blank	0.000	00:25
Control Test	0.082	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 03/30/2025 02:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,ABIGAIL,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: STOERMER,D
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LARA

DATE/TIME OF TEST: 03/08/2025 01:59:21

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.114	01:54
Breath Volume	1.214	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.119	01:57
Breath Volume	1.296	01:57
Air Blank	0.000	01:58
Control Test	0.083	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 03/30/2025 02:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON, ABIGAIL, M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOAGLAN, B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: KEIL

DATE/TIME OF TEST: 03/09/2025 05:36:59

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.080	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.092	05:32
Breath Volume	2.691	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.089	05:35
Breath Volume	2.480	05:35
Air Blank	0.000	05:36
Control Test	0.081	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 03/30/2025 02:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,ABIGAIL,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BRODERICK

DATE/TIME OF TEST: 03/09/2025 20:49:08

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.081	20:47	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:47	Subject Test Refused
Breath Volume	0.000	20:47	
Air Blank	0.000	20:48	
Control Test	0.079	20:48	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 03/30/2025 02:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENNICK,SARAH,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ORTIZ,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: YODER

DATE/TIME OF TEST: 03/23/2025 03:47:49

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.082	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.199	03:43
Breath Volume	3.414	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.192	03:46
Breath Volume	2.621	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 03/30/2025 02:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,KENNETH,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: VIEU,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:14

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 03/24/2025 07:41:42

Diagnostics Check	OK	07:34
Air Blank	0.000	07:35
Control Test	0.081	07:35
Air Blank	0.000	07:36
Subject Sample # 1	0.139	07:37
Breath Volume	1.445	07:37
Air Blank	0.000	07:37
Air Blank	0.000	07:39
Subject Sample # 2	0.138	07:40
Breath Volume	1.640	07:40
Air Blank	0.000	07:40
Control Test	0.081	07:41
Air Blank	0.000	07:41
Diagnostics Check	OK	07:41

Data Download Date/Time: 03/30/2025 02:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARRIS,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: DOVER

DATE/TIME OF TEST: 03/24/2025 23:27:58

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:16	
Control Test	0.082	23:16	
Air Blank	0.000	23:16	
Subject Sample	# 1	23:20	No Sample Provided
Breath Volume	0.000	23:20	
Air Blank	0.000	23:20	
Air Blank	0.000	23:22	
Subject Sample	# 2	0.176	23:23
Breath Volume	1.355	23:23	
Air Blank	0.000	23:23	
Air Blank	0.000	23:25	
Subject Sample	# 3	0.173	23:26
Breath Volume	1.218	23:26	
Air Blank	0.000	23:27	
Control Test	0.080	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 03/30/2025 02:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: REILY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: HEARD

DATE/TIME OF TEST: 03/25/2025 18:58:32

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.081	18:52
Air Blank	0.000	18:53
Subject Sample # 1	0.199	18:53
Breath Volume	1.343	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.197	18:57
Breath Volume	1.957	18:57
Air Blank	0.000	18:57
Control Test	0.081	18:57
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 03/30/2025 02:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GRIFFING

DATE/TIME OF TEST: 03/29/2025 23:28:04

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.159	23:23
Breath Volume	2.187	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.147	23:26
Breath Volume	1.554	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:28

Data Download Date/Time: 03/30/2025 02:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FROELICH,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CLICK

DATE/TIME OF TEST: 03/29/2025 23:57:31

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.000	23:52
Breath Volume	2.035	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.000	23:56
Breath Volume	2.078	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 03/30/2025 02:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOPEZ,FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: CUSTER

DATE/TIME OF TEST: 03/01/2025 21:00:49

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.083	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.008	20:56
Breath Volume	2.593	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.009	20:59
Breath Volume	2.726	20:59
Air Blank	0.000	20:59
Control Test	0.083	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 03/06/2025 01:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/02/2025 17:46:31

Diagnostics Check	OK	17:40
Air Blank	0.000	17:40
Control Test	0.083	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.169	17:42
Breath Volume	3.859	17:42
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 2	0.161	17:45
Breath Volume	1.910	17:45
Air Blank	0.000	17:45
Control Test	0.082	17:45
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 03/06/2025 01:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,TASHA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: DALE

DATE/TIME OF TEST: 03/04/2025 02:01:38

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.082	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.008	01:57
Breath Volume	2.113	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.007	02:00
Breath Volume	1.664	02:00
Air Blank	0.000	02:00
Control Test	0.083	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 03/06/2025 01:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LING,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SWENEY

DATE/TIME OF TEST: 03/02/2025 23:36:45

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.082	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:35	Interferent Detect
Breath Volume	2.605	23:35	
Air Blank	0.000	23:35	
Control Test	0.082	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 03/18/2025 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LING,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SWENEY

DATE/TIME OF TEST: 03/02/2025 23:42:14

Diagnostics Check	OK	23:41	
Air Blank		23:41	RFI Detect
Air Blank	0.000	23:42	

Data Download Date/Time: 03/18/2025 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LING,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SWENEY

DATE/TIME OF TEST: 03/02/2025 23:45:14

Diagnostics Check	OK	23:44	
Air Blank		23:44	RFI Detect
Air Blank	0.000	23:45	

Data Download Date/Time: 03/18/2025 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,JUAN,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MEDINA,JUAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MARIN SANTOS

DATE/TIME OF TEST: 03/22/2025 01:19:20

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.082	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.184	01:14
Breath Volume	2.542	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.178	01:17
Breath Volume	1.945	01:17
Air Blank	0.000	01:18
Control Test	0.082	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 04/09/2025 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BUONPANE,V
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:07

SUBJECT'S LAST NAME: MORALESJIMENEZ

DATE/TIME OF TEST: 03/26/2025 11:35:28

Diagnostics Check	OK	11:31	
Air Blank	0.000	11:32	
Control Test	0.081	11:32	
Air Blank	0.000	11:32	
Subject Sample	# 1 0.185	11:33	
Breath Volume	1.988	11:33	
Air Blank		11:34	RFI Detect
Air Blank	0.000	11:34	
Control Test	0.081	11:34	
Air Blank	0.000	11:35	
Diagnostics Check	OK	11:35	

Data Download Date/Time: 04/09/2025 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BUONPANE,V
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:07

SUBJECT'S LAST NAME: MORALESJIMENEZ

DATE/TIME OF TEST: 03/26/2025 11:43:06

Diagnostics Check	OK	11:36
Air Blank	0.000	11:37
Control Test	0.081	11:37
Air Blank	0.000	11:38
Subject Sample # 1	0.177	11:38
Breath Volume	1.597	11:38
Air Blank	0.000	11:39
Air Blank	0.000	11:41
Subject Sample # 2	0.172	11:41
Breath Volume	1.410	11:41
Air Blank	0.000	11:42
Control Test	0.081	11:42
Air Blank	0.000	11:43
Diagnostics Check	OK	11:43

Data Download Date/Time: 04/09/2025 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUZARDO,G,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: JOHNSON,N
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: URBINA MARTINEZ

DATE/TIME OF TEST: 03/02/2025 06:36:03

Diagnostics Check		OK	06:24
Air Blank		0.000	06:24
Control Test		0.081	06:25
Air Blank		0.000	06:25
Subject Sample	# 1	0.196	06:26
Breath Volume		3.335	06:26
Air Blank		0.000	06:27
Air Blank		0.000	06:28
Subject Sample	# 2	0.157	06:31
Breath Volume		1.289	06:31
Air Blank		0.000	06:32
Air Blank		0.000	06:33
Subject Sample	# 3	0.181	06:34
Breath Volume		3.460	06:34
Air Blank		0.000	06:35
Control Test		0.078	06:35
Air Blank		0.000	06:35
Diagnostics Check		OK	06:36

Data Download Date/Time: 03/26/2025 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALZUGARAY,ALEJANDRO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ALZUGARAY,ALEJANDRO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ALVARADO IXPATA

DATE/TIME OF TEST: 03/09/2025 00:48:13

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.168	00:43
Breath Volume	1.742	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.171	00:46
Breath Volume	1.906	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 03/26/2025 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,H,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCE,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: MASSO

DATE/TIME OF TEST: 03/29/2025 03:28:20

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.024	03:23
Breath Volume	1.371	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.024	03:26
Breath Volume	1.246	03:26
Air Blank	0.000	03:27
Control Test	0.082	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 04/30/2025 23:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,E,R
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ALMANZAR,S
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: URBANIAK

DATE/TIME OF TEST: 03/29/2025 08:55:07

Diagnostics Check	OK	08:44	
Air Blank	0.000	08:45	
Control Test	0.081	08:45	
Air Blank	0.000	08:46	
Subject Sample	# 1	08:49	No Sample Provided
Breath Volume	0.000	08:49	
Air Blank	0.000	08:49	
Air Blank	0.000	08:51	
Subject Sample	# 2	08:53	Subject Test Refused
Breath Volume	0.000	08:53	
Air Blank	0.000	08:54	
Control Test	0.081	08:54	
Air Blank	0.000	08:55	
Diagnostics Check	OK	08:55	

Data Download Date/Time: 04/30/2025 23:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUCH,DOMANIK,G
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: AUCH,DOMANIK
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/02/2025 02:11:16

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.082	02:00	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:03	No Sample Provided
Breath Volume	0.000	02:03	
Air Blank	0.000	02:04	
Air Blank	0.000	02:06	
Subject Sample	# 2	0.318	02:06
Breath Volume	2.980	02:06	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample	# 3	0.309	02:09
Breath Volume	2.304	02:09	
Air Blank	0.000	02:10	
Control Test	0.080	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 03/28/2025 21:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,LETOSHUA,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FLAHERTY,ZACHERY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 03/04/2025 13:05:23

Diagnostics Check	OK	12:59
Air Blank	0.000	12:59
Control Test	0.082	13:00
Air Blank	0.000	13:00
Subject Sample # 1	0.159	13:00
Breath Volume	2.515	13:00
Air Blank	0.000	13:01
Air Blank	0.000	13:03
Subject Sample # 2	0.152	13:03
Breath Volume	2.082	13:03
Air Blank	0.000	13:04
Control Test	0.082	13:04
Air Blank	0.000	13:05
Diagnostics Check	OK	13:05

Data Download Date/Time: 03/28/2025 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: LARSSON

DATE/TIME OF TEST: 03/05/2025 12:02:08

Diagnostics Check	OK	11:55
Air Blank	0.000	11:56
Control Test	0.082	11:56
Air Blank	0.000	11:57
Subject Sample # 1	0.092	11:57
Breath Volume	1.773	11:57
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.090	12:00
Breath Volume	1.699	12:00
Air Blank	0.000	12:01
Control Test	0.083	12:01
Air Blank	0.000	12:02
Diagnostics Check	OK	12:02

Data Download Date/Time: 03/28/2025 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,LETOSHUA,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREENE,LETOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BETTS

DATE/TIME OF TEST: 03/08/2025 14:20:03

Diagnostics Check	OK	14:15	
Air Blank	0.000	14:16	
Control Test	0.084	14:16	
Air Blank	0.000	14:17	
Subject Sample	# 1	14:19	Sequence Aborted
Breath Volume	0.148	14:19	
Air Blank	0.000	14:20	

Data Download Date/Time: 03/28/2025 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,LETOSHUA,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREENE,LETOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BETTS

DATE/TIME OF TEST: 03/08/2025 14:33:46

Diagnostics Check	OK	14:22	
Air Blank	0.000	14:22	
Control Test	0.083	14:23	
Air Blank	0.000	14:23	
Subject Sample	# 1 0.096	14:26	Volume Not Met
Breath Volume	0.125	14:26	
Air Blank	0.000	14:27	
Air Blank	0.000	14:29	
Subject Sample	# 2	14:32	No Sample Provided
Breath Volume	0.000	14:32	
Air Blank	0.000	14:32	
Control Test	0.082	14:33	
Air Blank	0.000	14:33	
Diagnostics Check	OK	14:33	

Data Download Date/Time: 03/28/2025 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 03/08/2025 21:51:07

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.081	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.205	21:46
Breath Volume	2.847	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:48
Subject Sample # 2	0.200	21:49
Breath Volume	1.933	21:49
Air Blank	0.000	21:50
Control Test	0.083	21:50
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 03/28/2025 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CUMMING,W
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/14/2025 07:27:39

Diagnostics Check	OK	07:21
Air Blank	0.000	07:21
Control Test	0.083	07:22
Air Blank	0.000	07:22
Subject Sample # 1	0.186	07:23
Breath Volume	2.312	07:23
Air Blank	0.000	07:23
Air Blank	0.000	07:25
Subject Sample # 2	0.187	07:26
Breath Volume	2.203	07:26
Air Blank	0.000	07:26
Control Test	0.081	07:27
Air Blank	0.000	07:27
Diagnostics Check	OK	07:27

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: WARD,GLENN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 03/14/2025 16:04:09

Diagnostics Check	OK	15:57
Air Blank	0.000	15:57
Control Test	0.083	15:57
Air Blank	0.000	15:58
Subject Sample # 1	0.280	15:58
Breath Volume	2.109	15:58
Air Blank	0.000	15:59
Air Blank	0.000	16:01
Subject Sample # 2	0.281	16:02
Breath Volume	1.328	16:02
Air Blank	0.000	16:03
Control Test	0.082	16:03
Air Blank	0.000	16:04
Diagnostics Check	OK	16:04

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 03/15/2025 21:01:20

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:50	
Control Test	0.082	20:51	
Air Blank	0.000	20:51	
Subject Sample	# 1 0.234	20:52	
Breath Volume	2.757	20:52	
Air Blank	0.000	20:53	
Air Blank	0.000	20:55	
Subject Sample	# 2 0.255	20:56	
Breath Volume	2.207	20:56	
Air Blank	0.000	20:57	
Air Blank	0.000	20:59	
Subject Sample	# 3 0.239	20:59	
Breath Volume	2.488	20:59	
Air Blank	0.000	21:00	
Control Test	0.081	21:00	
Air Blank		21:00	RFI Detect
Air Blank	0.000	21:01	

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UTLEY,BRETT,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: UTLEY,BRETT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 03/16/2025 22:04:24

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.082	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.201	21:59
Breath Volume	1.867	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:01
Subject Sample # 2	0.201	22:02
Breath Volume	2.445	22:02
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UTLEY,BRETT,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MANDILE,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 03/16/2025 22:13:28

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.081	22:08	
Air Blank	0.000	22:08	
Subject Sample	# 1 0.143	22:09	
Breath Volume	1.250	22:09	
Air Blank	0.000	22:09	
Air Blank	0.000	22:11	
Subject Sample	# 2 0.136	22:11	
Breath Volume	1.105	22:11	
Air Blank	0.000	22:12	
Control Test	0.081	22:12	
Air Blank		22:13	RFI Detect
Air Blank	0.000	22:13	

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UTLEY,BRETT,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MANDILE,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 03/16/2025 22:22:04

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.081	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.148	22:17
Breath Volume	1.941	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.139	22:20
Breath Volume	1.515	22:20
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:22

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DA SILVA,ANDRE,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: DA SILVA,ANDRE
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: MANDELBAUM

DATE/TIME OF TEST: 03/17/2025 19:05:54

Diagnostics Check	OK	18:54	
Air Blank	0.000	18:54	
Control Test	0.082	18:55	
Air Blank	0.000	18:55	
Subject Sample	# 1	18:58	No Sample Provided
Breath Volume	0.000	18:58	
Air Blank	0.000	18:59	
Air Blank	0.000	19:01	
Subject Sample	# 2	19:04	No Sample Provided
Breath Volume	0.000	19:04	
Air Blank	0.000	19:05	
Control Test	0.082	19:05	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: TEXIDO

DATE/TIME OF TEST: 03/20/2025 23:13:41

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.082	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.015	23:09
Breath Volume	1.542	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.016	23:12
Breath Volume	2.738	23:12
Air Blank	0.000	23:12
Control Test	0.083	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UTLEY,BRETT,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: UTLEY,BRETT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 03/23/2025 01:13:33

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.082	01:04	
Air Blank	0.000	01:05	
Subject Sample # 1	0.149	01:07	
Breath Volume	1.789	01:07	
Air Blank	0.000	01:08	
Air Blank	0.000	01:10	
Subject Sample # 2		01:11	Subject Test Refused
Breath Volume	0.195	01:11	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UTLEY,BRETT,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,M
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ANDRES

DATE/TIME OF TEST: 03/23/2025 02:03:57

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.082	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.145	01:59
Breath Volume	1.300	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.155	02:02
Breath Volume	1.359	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,J,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CUMMING,W
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:54

SUBJECT'S LAST NAME: GOLTRY

DATE/TIME OF TEST: 03/24/2025 16:22:59

Diagnostics Check	OK	16:14	
Air Blank	0.000	16:14	
Control Test	0.082	16:15	
Air Blank	0.000	16:15	
Subject Sample	# 1	16:18	No Sample Provided
Breath Volume	0.000	16:18	
Air Blank	0.000	16:19	
Air Blank	0.000	16:21	
Subject Sample	# 2	16:21	Subject Test Refused
Breath Volume	0.000	16:21	
Air Blank	0.000	16:22	
Control Test	0.081	16:22	
Air Blank	0.000	16:22	
Diagnostics Check	OK	16:22	

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/24/2025 23:23:11

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.082	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.126	23:17
Breath Volume	1.238	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample # 2	0.123	23:21
Breath Volume	1.289	23:21
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MARSH,JASON
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: LUND

DATE/TIME OF TEST: 03/25/2025 00:05:59

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.215	00:01
Breath Volume		2.609	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:03
Subject Sample	# 2	0.208	00:04
Breath Volume		2.738	00:04
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,M,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HINES,MIKO
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ANSPACH

DATE/TIME OF TEST: 03/26/2025 00:58:13

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.081	00:53
Breath Volume	2.390	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.085	00:56
Breath Volume	2.621	00:56
Air Blank	0.000	00:57
Control Test	0.082	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: RICKS

DATE/TIME OF TEST: 03/26/2025 16:05:41

Diagnostics Check	OK	15:57
Air Blank	0.000	15:58
Control Test	0.079	15:58
Air Blank	0.000	15:59
Subject Sample # 1	0.159	16:01
Breath Volume	2.015	16:01
Air Blank	0.000	16:01
Air Blank	0.000	16:03
Subject Sample # 2	0.143	16:04
Breath Volume	1.355	16:04
Air Blank	0.000	16:04
Control Test	0.077	16:05
Air Blank	0.000	16:05
Diagnostics Check	OK	16:05

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 03/27/2025 03:44:38

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.082	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.130	03:39
Breath Volume	1.343	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.125	03:43
Breath Volume	1.421	03:43
Air Blank	0.000	03:43
Control Test	0.082	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 03/28/2025 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,M,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: CORDOBA,ANTHONY
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LEVISMAN

DATE/TIME OF TEST: 03/01/2025 03:57:04

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.243	03:52
Breath Volume		2.519	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.242	03:55
Breath Volume		1.617	03:55
Air Blank		0.000	03:56
Control Test		0.076	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:57

Data Download Date/Time: 03/24/2025 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,M
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: CARRANZA

DATE/TIME OF TEST: 03/06/2025 02:00:20

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.079	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.079	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 03/24/2025 12:20