

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: GUTIERREZ MORALES

DATE/TIME OF TEST: 11/01/2024 18:32:59

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.079	18:26
Air Blank	0.000	18:27
Subject Sample # 1	0.223	18:27
Breath Volume	1.523	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.222	18:31
Breath Volume	2.066	18:31
Air Blank	0.000	18:32
Control Test	0.079	18:32
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 11/28/2024 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: AJU CHIYAL

DATE/TIME OF TEST: 11/08/2024 01:27:33

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.224	01:22
Breath Volume	1.363	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.225	01:25
Breath Volume	3.183	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 11/28/2024 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: HERNANDEZ,DANIEL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: AGUILAR PEREZ

DATE/TIME OF TEST: 11/26/2024 22:30:46

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.129	22:25
Breath Volume	2.355	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.131	22:29
Breath Volume	3.472	22:29
Air Blank	0.000	22:29
Control Test	0.079	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 11/28/2024 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERDT,ANDRU,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: GRAHAM,DARRER
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: THIGPEN

DATE/TIME OF TEST: 11/28/2024 04:01:14

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.078	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.177	03:56
Breath Volume	2.394	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.168	03:59
Breath Volume	2.425	03:59
Air Blank	0.000	04:00
Control Test	0.077	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 12/17/2024 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POST,HAYDEN,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: ABUSERYDZE,DAVID
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: DE LUNA GUZMAN

DATE/TIME OF TEST: 11/28/2024 23:21:47

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:18	
Control Test	0.077	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1 0.259	23:20	
Breath Volume	1.945	23:20	
Air Blank		23:21	Purge Fail
Air Blank	0.000	23:21	

Data Download Date/Time: 12/17/2024 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POST,HAYDEN,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: ABUSERYDZE,DAVID
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: DE LUNA GUZMAN

DATE/TIME OF TEST: 11/28/2024 23:31:03

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.075	23:24	
Air Blank	0.000	23:24	
Subject Sample	# 1 0.262	23:26	
Breath Volume	1.996	23:26	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2 0.248	23:29	
Breath Volume	2.261	23:29	
Air Blank	0.000	23:30	
Control Test	0.074	23:30	Control Outside Tolerance
Air Blank	0.000	23:31	

Data Download Date/Time: 12/17/2024 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERDT,ANDRU,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: ALLEN,BREANNA
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SAULISBURY

DATE/TIME OF TEST: 11/29/2024 04:44:19

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.078	04:42	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:43	
Control Test	0.076	04:43	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 12/17/2024 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERDT,ANDRU,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: SMITH,TERIE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: CLEMENT

DATE/TIME OF TEST: 11/29/2024 05:12:38

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:10	
Control Test	0.078	05:10	
Air Blank	0.000	05:11	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:11	
Control Test	0.077	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 12/17/2024 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 11/01/2024 02:38:10

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Control Test	0.080	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 11/23/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 11/19/2024 01:09:01

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:04	
Subject Sample # 1	0.000	01:05	
Breath Volume	1.406	01:05	
Air Blank	0.000	01:05	
Air Blank		01:07	RFI Detect
Air Blank	0.000	01:08	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 11/23/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 11/19/2024 01:17:54

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.000	01:13
Breath Volume	1.539	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.000	01:16
Breath Volume	1.269	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 11/23/2024 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY,MATTHEW,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WILLIAMS,ERIC
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SOKOL

DATE/TIME OF TEST: 11/25/2024 02:24:38

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.119	02:19
Breath Volume	1.722	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.120	02:22
Breath Volume	1.863	02:22
Air Blank	0.000	02:23
Control Test	0.078	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 12/26/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANEY, T
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: FAUNCE

DATE/TIME OF TEST: 11/29/2024 01:17:04

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.143	01:12
Breath Volume	4.343	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:14
Subject Sample # 2	0.129	01:15
Breath Volume	1.742	01:15
Air Blank	0.000	01:16
Control Test	0.078	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:17

Data Download Date/Time: 12/26/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: KEITH,JOHN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/05/2024 18:20:42

Diagnostics Check	OK	18:13
Air Blank	0.000	18:13
Control Test	0.081	18:14
Air Blank	0.000	18:14
Subject Sample # 1	0.000	18:16
Breath Volume	1.675	18:16
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.000	18:19
Breath Volume	1.648	18:19
Air Blank	0.000	18:19
Control Test	0.079	18:20
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 02/03/2025 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ GARCIA,HERNAN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: KEITH,JOHN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: ROJAS NUNEZ

DATE/TIME OF TEST: 11/15/2024 00:01:39

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.082	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.082	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 02/03/2025 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: KEITH,JOHN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 11/17/2024 00:35:05

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.083	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.222	00:29
Breath Volume	3.039	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.224	00:33
Breath Volume	2.433	00:33
Air Blank	0.000	00:34
Control Test	0.082	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 02/03/2025 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ GARCIA,HERNAN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DIBERNARDINO,BRANDON
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: FRANCISCO DIEGO

DATE/TIME OF TEST: 11/20/2024 01:22:34

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.082	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.111	01:17
Breath Volume	2.804	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.104	01:20
Breath Volume	3.425	01:20
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 02/03/2025 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DIBERNARDINO,BRANDON
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HOLMAN

DATE/TIME OF TEST: 11/26/2024 20:25:40

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.082	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.063	20:20
Breath Volume	2.843	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.066	20:23
Breath Volume	2.992	20:23
Air Blank	0.000	20:24
Control Test	0.082	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 02/03/2025 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ GARCIA,HERNAN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DIBERNARDINO,BRANDON
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: VILLALOBOS

DATE/TIME OF TEST: 11/30/2024 04:27:11

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.082	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.266	04:22
Breath Volume	1.644	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.254	04:25
Breath Volume	1.628	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 02/03/2025 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: VERA MUNOZ

DATE/TIME OF TEST: 11/01/2024 01:04:36

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.080	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.184	01:00
Breath Volume	2.000	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.185	01:02
Breath Volume	2.750	01:02
Air Blank	0.000	01:03
Control Test	0.078	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: OPORTA

DATE/TIME OF TEST: 11/05/2024 00:59:00

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:51	
Subject Sample # 1	0.116	00:54	
Breath Volume	1.605	00:54	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample # 2	0.122	00:57	
Breath Volume	1.933	00:57	
Air Blank	0.000	00:58	
Control Test	0.071	00:58	Control Outside Tolerance
Air Blank	0.000	00:59	

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: OPORTA

DATE/TIME OF TEST: 11/05/2024 01:07:43

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.077	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.122	01:02
Breath Volume	2.449	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:04
Subject Sample # 2	0.125	01:06
Breath Volume	1.695	01:06
Air Blank	0.000	01:06
Control Test	0.076	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: LOVERING

DATE/TIME OF TEST: 11/05/2024 23:45:02

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:39	
Control Test	0.077	23:39	
Air Blank	0.000	23:40	
Subject Sample # 1	0.000	23:40	
Breath Volume	3.914	23:40	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample # 2	0.000	23:43	
Breath Volume	3.355	23:43	
Air Blank	0.000	23:44	
Control Test	0.075	23:44	Interferent Detect
Air Blank	0.000	23:45	

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: ELL

DATE/TIME OF TEST: 11/14/2024 20:03:31

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.078	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.170	19:59
Breath Volume	3.417	19:59
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.162	20:01
Breath Volume	3.464	20:01
Air Blank	0.000	20:02
Control Test	0.077	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 11/15/2024 23:37:12

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.139	23:32
Breath Volume	2.656	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.136	23:35
Breath Volume	1.550	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GONZALEZ SANDREA

DATE/TIME OF TEST: 11/19/2024 03:37:28

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.158	03:32
Breath Volume	2.250	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.165	03:35
Breath Volume	1.898	03:35
Air Blank	0.000	03:36
Control Test	0.076	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MARTIN PEREZ

DATE/TIME OF TEST: 11/20/2024 01:03:46

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.173	00:59
Breath Volume	2.621	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.164	01:02
Breath Volume	2.355	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/17/2024 04:23:03

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.097	04:18
Breath Volume	2.972	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.097	04:21
Breath Volume	2.503	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:23

Data Download Date/Time: 11/29/2024 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ROUNTREE,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: BALLEW

DATE/TIME OF TEST: 11/22/2024 12:18:55

Diagnostics Check	OK	12:12
Air Blank	0.000	12:13
Control Test	0.082	12:13
Air Blank	0.000	12:13
Subject Sample # 1	0.000	12:14
Breath Volume	2.093	12:14
Air Blank	0.000	12:15
Air Blank	0.000	12:17
Subject Sample # 2	0.000	12:17
Breath Volume	2.097	12:17
Air Blank	0.000	12:17
Control Test	0.083	12:18
Air Blank	0.000	12:18
Diagnostics Check	OK	12:18

Data Download Date/Time: 11/29/2024 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HUGGINS,L
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BELLEFLEUR

DATE/TIME OF TEST: 11/28/2024 04:06:28

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1	04:01	No Sample Provided
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 2	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.081	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 11/29/2024 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: SHULER III

DATE/TIME OF TEST: 11/01/2024 00:33:17

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.063	00:28
Breath Volume	4.148	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.066	00:31
Breath Volume	4.667	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: DIEUMERCI

DATE/TIME OF TEST: 11/01/2024 05:44:21

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:38
Air Blank	0.000	05:39
Subject Sample # 1	0.111	05:39
Breath Volume	1.109	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.108	05:42
Breath Volume	1.554	05:42
Air Blank	0.000	05:43
Control Test	0.079	05:43
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ALVARADO VIVAS

DATE/TIME OF TEST: 11/02/2024 03:43:22

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.149	03:38
Breath Volume	1.238	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.147	03:41
Breath Volume	2.351	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SANTIAGO MALDONADO

DATE/TIME OF TEST: 11/02/2024 04:47:14

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.157	04:42
Breath Volume	2.539	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.151	04:45
Breath Volume	2.539	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RENTERIA,ELIZABETH
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 11/02/2024 05:35:15

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.079	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.219	05:30
Breath Volume	1.398	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:32
Subject Sample # 2	0.221	05:33
Breath Volume	1.375	05:33
Air Blank	0.000	05:34
Control Test	0.077	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: CHAJ-GONZALEZ

DATE/TIME OF TEST: 11/02/2024 21:18:31

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.080	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.246	21:13
Breath Volume	1.527	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.234	21:16
Breath Volume	1.906	21:16
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/03/2024 01:31:34

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.151	01:26
Breath Volume	2.066	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.155	01:30
Breath Volume	1.550	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 11/03/2024 02:19:16

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.135	02:14
Breath Volume	1.421	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.141	02:17
Breath Volume	2.621	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,RYAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: FERNANDO PASCUAL

DATE/TIME OF TEST: 11/03/2024 03:34:14

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.229	03:29
Breath Volume	2.660	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.225	03:32
Breath Volume	2.496	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: CLERVOYANT

DATE/TIME OF TEST: 11/03/2024 08:20:56

Diagnostics Check	OK	08:13
Air Blank	0.000	08:13
Control Test	0.080	08:14
Air Blank	0.000	08:14
Subject Sample # 1	0.162	08:16
Breath Volume	2.804	08:16
Air Blank	0.000	08:16
Air Blank	0.000	08:18
Subject Sample # 2	0.167	08:19
Breath Volume	2.628	08:19
Air Blank	0.000	08:20
Control Test	0.079	08:20
Air Blank	0.000	08:20
Diagnostics Check	OK	08:20

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: ROBLERO GONZALEZ

DATE/TIME OF TEST: 11/03/2024 22:59:48

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.266	22:55
Breath Volume	2.109	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.257	22:58
Breath Volume	2.566	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ALLIANCE

DATE/TIME OF TEST: 11/04/2024 01:49:53

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1 0.209	01:39	
Breath Volume	1.593	01:39	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2 0.145	01:45	Volume Not Met
Breath Volume	0.257	01:45	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample	# 3 0.205	01:48	
Breath Volume	1.574	01:48	
Air Blank	0.000	01:48	
Control Test	0.077	01:49	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRANADOS,AUIS
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/07/2024 03:04:32

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1 0.136	02:57	Volume Not Met
Breath Volume	0.066	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 2 0.123	03:02	Volume Not Met
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Control Test	0.078	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GELINA,PHILIP
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: ALLEN III

DATE/TIME OF TEST: 11/07/2024 11:14:25

Diagnostics Check	OK	11:09	
Air Blank	0.000	11:10	
Control Test	0.080	11:10	
Air Blank	0.000	11:11	
Subject Sample	# 1	11:13	Subject Test Refused
Breath Volume	0.000	11:13	
Air Blank	0.000	11:13	
Control Test	0.080	11:13	
Air Blank	0.000	11:14	
Diagnostics Check	OK	11:14	

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: MEJIA ACOSTA

DATE/TIME OF TEST: 11/07/2024 21:56:54

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.080	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.142	21:52
Breath Volume	2.542	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:54
Subject Sample # 2	0.144	21:55
Breath Volume	2.523	21:55
Air Blank	0.000	21:55
Control Test	0.079	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 11/09/2024 03:10:07

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.161	03:05
Breath Volume	1.300	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.169	03:08
Breath Volume	1.574	03:08
Air Blank	0.000	03:09
Control Test	0.077	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL, KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SANCHEZ VEGA

DATE/TIME OF TEST: 11/09/2024 03:34:15

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.181	03:28
Breath Volume	1.632	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.172	03:32
Breath Volume	1.589	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLACKMAN, VICTOR
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: JERONIMO SIMON

DATE/TIME OF TEST: 11/09/2024 06:08:51

Diagnostics Check	OK	06:02
Air Blank	0.000	06:02
Control Test	0.079	06:03
Air Blank	0.000	06:03
Subject Sample # 1	0.212	06:04
Breath Volume	2.378	06:04
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample # 2	0.202	06:07
Breath Volume	2.125	06:07
Air Blank	0.000	06:07
Control Test	0.078	06:08
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FRANQUI,JASON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: MARQUEZ SERRANO

DATE/TIME OF TEST: 11/09/2024 13:45:30

Diagnostics Check	OK	13:38
Air Blank	0.000	13:38
Control Test	0.080	13:39
Air Blank	0.000	13:39
Subject Sample # 1	0.234	13:40
Breath Volume	2.816	13:40
Air Blank	0.000	13:41
Air Blank	0.000	13:43
Subject Sample # 2	0.232	13:43
Breath Volume	1.382	13:43
Air Blank	0.000	13:44
Control Test	0.078	13:44
Air Blank	0.000	13:45
Diagnostics Check	OK	13:45

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,RYAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: AVALO ESTABAN

DATE/TIME OF TEST: 11/09/2024 22:25:45

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:13	
Control Test	0.080	22:13	
Air Blank	0.000	22:13	
Subject Sample	# 1 0.116	22:17	Volume Not Met
Breath Volume	0.472	22:17	
Air Blank	0.000	22:17	
Air Blank	0.000	22:19	
Subject Sample	# 2 0.144	22:21	
Breath Volume	1.687	22:21	
Air Blank	0.000	22:21	
Air Blank	0.000	22:23	
Subject Sample	# 3 0.148	22:24	
Breath Volume	1.699	22:24	
Air Blank	0.000	22:24	
Control Test	0.078	22:25	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN],
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: ROJO

DATE/TIME OF TEST: 11/10/2024 00:25:35

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.194	00:20
Breath Volume	1.460	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.192	00:24
Breath Volume	1.402	00:24
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXANDER,JACE
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MORALS-SANCHEZ

DATE/TIME OF TEST: 11/10/2024 00:50:45

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.078	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.258	00:46
Breath Volume	1.765	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.256	00:49
Breath Volume	2.042	00:49
Air Blank	0.000	00:49
Control Test	0.078	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BASTIEN

DATE/TIME OF TEST: 11/10/2024 02:23:35

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.132	02:15
Breath Volume	2.175	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.105	02:18
Breath Volume	1.117	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 3	0.125	02:22
Breath Volume	1.410	02:22
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HAGER

DATE/TIME OF TEST: 11/10/2024 03:00:36

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.078	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1 0.000	02:53	Volume Not Met
Breath Volume	0.128	02:53	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 2 0.000	02:59	Volume Not Met
Breath Volume	0.125	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBINSON,DAVIS
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 11/10/2024 04:47:22

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.148	04:42
Breath Volume	1.484	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.155	04:45
Breath Volume	1.605	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD,ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: MARCENAC

DATE/TIME OF TEST: 11/10/2024 05:42:04

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.079	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.135	05:37
Breath Volume	1.359	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.138	05:40
Breath Volume	2.125	05:40
Air Blank	0.000	05:41
Control Test	0.079	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:42

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JENNINGS,DAVON
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 11/12/2024 01:14:53

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.244	01:10
Breath Volume	1.425	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.239	01:13
Breath Volume	1.839	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: AMBROCIO TOJIL

DATE/TIME OF TEST: 11/12/2024 22:48:35

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:32	
Control Test	0.080	22:33	
Air Blank	0.000	22:33	
Subject Sample	# 1 0.194	22:36	Slope Not Level
Breath Volume	1.207	22:36	
Air Blank	0.000	22:37	
Air Blank	0.000	22:39	
Subject Sample	# 2 0.193	22:41	
Breath Volume	1.507	22:41	
Air Blank	0.000	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 3	22:46	Subject Test Refused
Breath Volume	0.300	22:46	
Air Blank	0.000	22:47	
Control Test	0.077	22:47	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 11/14/2024 02:50:08

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.210	02:44
Breath Volume	1.578	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.205	02:48
Breath Volume	2.019	02:48
Air Blank	0.000	02:49
Control Test	0.077	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: PASCUAL FRANCISCO

DATE/TIME OF TEST: 11/15/2024 01:04:13

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.165	00:59
Breath Volume	1.457	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.166	01:02
Breath Volume	3.222	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSTON,DEVIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: LONDONO GONZALEZ

DATE/TIME OF TEST: 11/16/2024 04:43:29

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.226	04:35
Breath Volume	1.257	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:37
Subject Sample # 2	0.254	04:38
Breath Volume	3.328	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 3	0.244	04:41
Breath Volume	2.246	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 12/06/2024 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARKER,JEFFREY
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: GAYLE

DATE/TIME OF TEST: 11/16/2024 05:37:30

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.080	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.160	05:32
Breath Volume	1.296	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.161	05:35
Breath Volume	1.628	05:35
Air Blank	0.000	05:36
Control Test	0.079	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 12/06/2024 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KLEIN, JEFFERY
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: TIEDEMANN

DATE/TIME OF TEST: 11/17/2024 03:07:45

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.128	03:02
Breath Volume	1.222	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.141	03:06
Breath Volume	1.496	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 12/06/2024 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: HIRSCHFIELD

DATE/TIME OF TEST: 11/17/2024 21:58:58

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.184	21:53
Breath Volume	2.109	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:55
Subject Sample # 2	0.174	21:57
Breath Volume	1.277	21:57
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 12/06/2024 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GIBSON, BRANDON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/18/2024 16:25:33

Diagnostics Check	OK	16:19
Air Blank	0.000	16:20
Control Test	0.080	16:20
Air Blank	0.000	16:21
Subject Sample # 1	0.000	16:21
Breath Volume	1.992	16:21
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample # 2	0.000	16:24
Breath Volume	2.925	16:24
Air Blank	0.000	16:24
Control Test	0.080	16:24
Air Blank	0.000	16:25
Diagnostics Check	OK	16:25

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: CUTEREZ LAREZ

DATE/TIME OF TEST: 11/19/2024 23:11:22

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.081	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.181	23:06
Breath Volume	2.816	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.182	23:09
Breath Volume	3.035	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: LALLA

DATE/TIME OF TEST: 11/22/2024 01:15:31

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.082	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.123	01:11
Breath Volume	1.875	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.124	01:13
Breath Volume	1.968	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO, AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: RENAZILE

DATE/TIME OF TEST: 11/22/2024 07:18:35

Diagnostics Check	OK	07:12
Air Blank	0.000	07:12
Control Test	0.081	07:12
Air Blank	0.000	07:13
Subject Sample # 1	0.112	07:14
Breath Volume	1.906	07:14
Air Blank	0.000	07:14
Air Blank	0.000	07:16
Subject Sample # 2	0.112	07:17
Breath Volume	2.949	07:17
Air Blank	0.000	07:17
Control Test	0.081	07:17
Air Blank	0.000	07:18
Diagnostics Check	OK	07:18

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BAGDZIUNAS, BRYAN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 11/23/2024 02:51:37

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.180	02:47
Breath Volume	1.906	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.180	02:50
Breath Volume	1.585	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: GUILCAMAIGUA CAISAGU

DATE/TIME OF TEST: 11/23/2024 23:06:06

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.081	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.144	23:01
Breath Volume	2.554	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.139	23:04
Breath Volume	1.812	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSTON,DEVIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: NIXON

DATE/TIME OF TEST: 11/24/2024 06:30:01

Diagnostics Check	OK	06:22
Air Blank	0.000	06:22
Control Test	0.081	06:22
Air Blank	0.000	06:23
Subject Sample # 1	0.076	06:25
Breath Volume	2.148	06:25
Air Blank	0.000	06:26
Air Blank	0.000	06:27
Subject Sample # 2	0.074	06:28
Breath Volume	2.355	06:28
Air Blank	0.000	06:29
Control Test	0.081	06:29
Air Blank	0.000	06:29
Diagnostics Check	OK	06:30

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOOHEY,DOUGLAS
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: TREJO MENDOZA

DATE/TIME OF TEST: 11/25/2024 03:53:51

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.145	03:49
Breath Volume	1.636	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.143	03:52
Breath Volume	2.039	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE, DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: ARAGONES

DATE/TIME OF TEST: 11/26/2024 02:02:36

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.191	01:57
Breath Volume	1.609	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.172	02:01
Breath Volume	1.496	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MORALES SALES

DATE/TIME OF TEST: 11/27/2024 00:16:34

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.081	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.176	00:12
Breath Volume	1.492	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.178	00:15
Breath Volume	2.562	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JENNINGS,DAVON
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CAMARENA

DATE/TIME OF TEST: 11/27/2024 01:54:36

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.056	01:50
Breath Volume	2.085	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.057	01:53
Breath Volume	1.816	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 11/27/2024 03:24:02

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.000	03:19
Breath Volume	1.808	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.000	03:22
Breath Volume	1.496	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:24

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEPEINE,KEN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 11/27/2024 06:57:26

Diagnostics Check	OK	06:50
Air Blank	0.000	06:51
Control Test	0.080	06:51
Air Blank	0.000	06:51
Subject Sample # 1	0.185	06:52
Breath Volume	1.488	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.175	06:55
Breath Volume	1.144	06:55
Air Blank	0.000	06:56
Control Test	0.079	06:56
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: BERKE

DATE/TIME OF TEST: 11/28/2024 03:01:42

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.174	02:57
Breath Volume	2.007	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.169	03:00
Breath Volume	1.476	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: RAMIREZ AGUILAR

DATE/TIME OF TEST: 11/28/2024 22:30:28

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.081	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.239	22:25
Breath Volume	2.714	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.232	22:28
Breath Volume	3.457	22:28
Air Blank	0.000	22:29
Control Test	0.079	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU, YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LOUISSAINT

DATE/TIME OF TEST: 11/28/2024 23:44:07

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.080	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.150	23:39
Breath Volume	1.777	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.154	23:42
Breath Volume	1.945	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: CLERVIN

DATE/TIME OF TEST: 11/29/2024 01:23:04

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.184	01:18
Breath Volume	1.996	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.180	01:21
Breath Volume	1.597	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:23

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ESCOBAR MALDONADO

DATE/TIME OF TEST: 11/29/2024 03:24:02

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.094	03:19
Breath Volume	2.371	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.091	03:22
Breath Volume	1.687	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:24

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE, RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 11/29/2024 20:43:21

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.081	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.218	20:38
Breath Volume	2.410	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.209	20:41
Breath Volume	1.804	20:41
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PROPHETE, STEVE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: ROSARIO PITZER

DATE/TIME OF TEST: 11/29/2024 21:57:06

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.227	21:52
Breath Volume	1.652	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.216	21:55
Breath Volume	1.531	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GUEVARA ZAMBRANO

DATE/TIME OF TEST: 11/30/2024 02:24:04

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.143	02:19
Breath Volume	1.875	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.138	02:22
Breath Volume	2.101	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:24

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BATES,IVAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/30/2024 03:32:33

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.069	03:28
Breath Volume	2.367	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.069	03:31
Breath Volume	1.453	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 12/06/2024 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL, KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/15/2024 22:02:11

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.132	21:57
Breath Volume	2.515	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.130	22:00
Breath Volume	1.921	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PRUDHOMME, DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BUSHEY JOYAS

DATE/TIME OF TEST: 11/16/2024 04:13:09

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.079	03:59	
Air Blank	0.000	03:59	
Subject Sample	# 1 0.105	04:02	Volume Not Met
Breath Volume	0.132	04:02	
Air Blank	0.000	04:03	
Air Blank	0.000	04:05	
Subject Sample	# 2 0.178	04:07	
Breath Volume	2.253	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 3 0.162	04:11	
Breath Volume	1.605	04:11	
Air Blank	0.000	04:12	
Control Test	0.079	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MANCILLA VELAZQUEZ

DATE/TIME OF TEST: 11/17/2024 00:46:55

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.143	00:42
Breath Volume	1.609	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.147	00:45
Breath Volume	2.804	00:45
Air Blank	0.000	00:46
Control Test	0.078	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/17/2024 03:57:19

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.099	03:56	Interferent Detect
Air Blank		03:57	Purge Fail

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/17/2024 04:30:41

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.080	04:28	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:29	Interferent Detect
Breath Volume	2.265	04:29	
Air Blank	0.000	04:29	
Control Test	0.061	04:30	Interferent Detect
Air Blank	0.000	04:30	

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: CHAJ CHAJ

DATE/TIME OF TEST: 11/18/2024 00:19:27

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.244	00:14
Breath Volume	1.378	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.250	00:17
Breath Volume	2.531	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 11/19/2024 03:47:07

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.077	03:42
Breath Volume	1.378	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.076	03:45
Breath Volume	1.953	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CADET

DATE/TIME OF TEST: 11/20/2024 04:59:57

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.080	04:46	
Air Blank	0.000	04:47	
Subject Sample	# 1 0.105	04:50	Volume Not Met
Breath Volume	0.367	04:50	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 2 0.117	04:55	
Breath Volume	1.906	04:55	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample	# 3 0.117	04:58	
Breath Volume	2.121	04:58	
Air Blank	0.000	04:59	
Control Test	0.078	04:59	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: SALES ORDONEZ

DATE/TIME OF TEST: 11/20/2024 21:10:16

Diagnostics Check	OK	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:03
Subject Sample # 1	0.268	21:04
Breath Volume	1.597	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.260	21:08
Breath Volume	1.312	21:08
Air Blank	0.000	21:09
Control Test	0.077	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 11/23/2024 00:32:23

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.144	00:27
Breath Volume	1.472	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.146	00:30
Breath Volume	1.785	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSTON,DEVIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: SANTOS NUNEZ

DATE/TIME OF TEST: 11/23/2024 01:25:50

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.080	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.169	01:21
Breath Volume	1.441	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.164	01:24
Breath Volume	1.656	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 12/06/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: BOVA

DATE/TIME OF TEST: 11/24/2024 00:46:07

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.080	00:40
Breath Volume	1.222	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.082	00:44
Breath Volume	1.589	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SANCHEZ SALES

DATE/TIME OF TEST: 11/24/2024 01:58:59

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.165	01:54
Breath Volume	3.101	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.164	01:57
Breath Volume	3.214	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCDONOUGH,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: GARCIA VASQUEZ

DATE/TIME OF TEST: 11/27/2024 00:59:10

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.204	00:53
Breath Volume	2.277	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.198	00:57
Breath Volume	2.652	00:57
Air Blank	0.000	00:58
Control Test	0.077	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: GOMEZ RAMOS

DATE/TIME OF TEST: 11/27/2024 02:56:41

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.129	02:52
Breath Volume	1.867	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.128	02:55
Breath Volume	1.941	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD,ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MOTLEY

DATE/TIME OF TEST: 11/28/2024 02:58:16

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.234	02:53
Breath Volume	1.113	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.232	02:56
Breath Volume	2.375	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: MILLARD

DATE/TIME OF TEST: 11/28/2024 05:54:51

Diagnostics Check	OK	05:47
Air Blank	0.000	05:48
Control Test	0.079	05:48
Air Blank	0.000	05:49
Subject Sample # 1	0.161	05:49
Breath Volume	1.632	05:49
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample # 2	0.157	05:53
Breath Volume	1.746	05:53
Air Blank	0.000	05:53
Control Test	0.079	05:54
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR, ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: BRUNSON

DATE/TIME OF TEST: 11/28/2024 21:39:55

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.080	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 1 0.140	21:28	
Breath Volume	1.390	21:28	
Air Blank	0.000	21:29	
Air Blank	0.000	21:31	
Subject Sample	# 2 0.134	21:34	Volume Not Met
Breath Volume	0.062	21:34	
Air Blank	0.000	21:35	
Air Blank	0.000	21:36	
Subject Sample	# 3 0.145	21:38	
Breath Volume	1.539	21:38	
Air Blank	0.000	21:38	
Control Test	0.076	21:39	
Air Blank	0.000	21:39	
Diagnostics Check	OK	21:39	

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: ANGEL VELAZQUEZ

DATE/TIME OF TEST: 11/28/2024 22:51:20

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.078	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.226	22:46
Breath Volume	2.402	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.208	22:49
Breath Volume	2.203	22:49
Air Blank	0.000	22:50
Control Test	0.076	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSTON, DEVON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: VELAZQUEZ MENEZ

DATE/TIME OF TEST: 11/29/2024 01:34:35

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.151	01:30
Breath Volume	2.031	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.148	01:33
Breath Volume	2.308	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HAYLES, ROSHAWN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BENNETT COLEMAN

DATE/TIME OF TEST: 11/29/2024 02:36:20

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.179	02:31
Breath Volume	1.265	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.182	02:34
Breath Volume	1.378	02:34
Air Blank	0.000	02:35
Control Test	0.078	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MERKLE

DATE/TIME OF TEST: 11/30/2024 01:54:53

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.051	01:50
Breath Volume	2.109	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:52
Subject Sample # 2	0.053	01:53
Breath Volume	1.707	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: FLEURIMON

DATE/TIME OF TEST: 11/30/2024 02:58:41

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSHAIER, ELIAS
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 11/30/2024 04:16:39

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.172	04:12
Breath Volume	1.605	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.168	04:15
Breath Volume	1.609	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 12/06/2024 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: LAUER

DATE/TIME OF TEST: 11/01/2024 00:42:19

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.081	00:28	
Air Blank	0.000	00:29	
Subject Sample # 1	0.185	00:30	
Breath Volume	1.609	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample # 2	0.143	00:36	Volume Not Met
Breath Volume	0.101	00:36	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample # 3		00:40	Subject Test Refused
Breath Volume	0.417	00:40	
Air Blank	0.000	00:41	
Control Test	0.079	00:41	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: DULA

DATE/TIME OF TEST: 11/01/2024 03:57:32

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.252	03:53
Breath Volume	1.968	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.257	03:55
Breath Volume	2.035	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HLUCH,MARIO
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: EMBT

DATE/TIME OF TEST: 11/01/2024 04:58:27

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.080	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.239	04:49
Breath Volume	2.851	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.199	04:53
Breath Volume	1.960	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:55
Subject Sample # 3	0.201	04:56
Breath Volume	1.136	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO, AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: SANTANA DE LEON

DATE/TIME OF TEST: 11/01/2024 21:43:46

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.081	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.271	21:39
Breath Volume	3.054	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.258	21:42
Breath Volume	2.148	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BULLARD

DATE/TIME OF TEST: 11/02/2024 04:32:57

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.115	04:28
Breath Volume	1.191	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.109	04:31
Breath Volume	1.773	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: JEAN BAPTISTE

DATE/TIME OF TEST: 11/02/2024 05:50:12

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.081	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.101	05:44
Breath Volume	1.273	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.096	05:48
Breath Volume	1.167	05:48
Air Blank	0.000	05:49
Control Test	0.080	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: GONGORA GALLARDO

DATE/TIME OF TEST: 11/03/2024 01:40:22

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.216	01:35
Breath Volume	1.539	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.217	01:38
Breath Volume	1.304	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 11/03/2024 01:56:56

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.177	01:52
Breath Volume	2.449	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample # 2	0.170	01:55
Breath Volume	2.425	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,JAIME
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: KOUNLAVONG

DATE/TIME OF TEST: 11/03/2024 02:31:23

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.147	02:26
Breath Volume	1.664	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.147	02:29
Breath Volume	2.085	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 11/03/2024 04:30:22

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.177	04:25
Breath Volume	2.570	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.184	04:28
Breath Volume	2.851	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: VELASQUEZ-CRUZ

DATE/TIME OF TEST: 11/03/2024 19:10:45

Diagnostics Check	OK	19:02
Air Blank	0.000	19:02
Control Test	0.080	19:03
Air Blank	0.000	19:03
Subject Sample # 1	0.226	19:04
Breath Volume	1.980	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.220	19:09
Breath Volume	2.460	19:09
Air Blank	0.000	19:09
Control Test	0.080	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BRITO DOBLE

DATE/TIME OF TEST: 11/04/2024 03:40:03

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.081	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:39	
Control Test	0.080	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: EKONOMOU

DATE/TIME OF TEST: 11/06/2024 00:31:38

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.177	00:27
Breath Volume	1.179	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.181	00:30
Breath Volume	1.976	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 11/08/2024 02:41:46

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.076	02:37
Breath Volume	2.906	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.075	02:40
Breath Volume	1.738	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CRISTOBAL LUCAS

DATE/TIME OF TEST: 11/08/2024 03:12:18

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.219	03:07
Breath Volume	1.796	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.220	03:10
Breath Volume	1.156	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STOUD, ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 11/08/2024 22:32:49

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.081	22:21	
Air Blank	0.000	22:22	
Subject Sample # 1	0.256	22:22	
Breath Volume	1.390	22:22	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample # 2	0.220	22:28	Slope Not Level
Breath Volume	1.707	22:28	
Air Blank	0.000	22:28	
Air Blank	0.000	22:30	
Subject Sample # 3	0.248	22:31	
Breath Volume	2.703	22:31	
Air Blank	0.000	22:31	
Control Test	0.079	22:32	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PRADO,MARJORIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 11/09/2024 05:36:57

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.174	05:32
Breath Volume	1.609	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.162	05:35
Breath Volume	1.792	05:35
Air Blank	0.000	05:36
Control Test	0.080	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PEREZ AGUILAR

DATE/TIME OF TEST: 11/09/2024 23:50:53

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.224	23:46
Breath Volume	1.789	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.224	23:49
Breath Volume	1.187	23:49
Air Blank	0.000	23:49
Control Test	0.080	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ZOLLICOFFER

DATE/TIME OF TEST: 11/10/2024 01:36:44

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.102	01:32
Breath Volume	2.132	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.099	01:35
Breath Volume	1.828	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: OWERS,PETER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: TOURLET

DATE/TIME OF TEST: 11/10/2024 02:45:36

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.158	02:39
Breath Volume	1.722	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.156	02:44
Breath Volume	1.800	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: VELASQUEZ ESCALANTE

DATE/TIME OF TEST: 11/10/2024 04:39:02

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.174	04:32
Breath Volume	2.308	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.172	04:37
Breath Volume	1.226	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:39

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/10/2024 05:43:19

Diagnostics Check	OK	05:32	
Air Blank	0.000	05:32	
Control Test	0.080	05:33	
Air Blank	0.000	05:33	
Subject Sample	# 1 0.165	05:36	Volume Not Met
Breath Volume	0.160	05:36	
Air Blank	0.000	05:37	
Air Blank	0.000	05:39	
Subject Sample	# 2	05:41	Subject Test Refused
Breath Volume	0.417	05:41	
Air Blank	0.000	05:42	
Control Test	0.079	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOPEZ, DINEEN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: SAMAREL

DATE/TIME OF TEST: 11/10/2024 23:31:02

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.081	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.133	23:26
Breath Volume	1.152	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:28
Subject Sample # 2	0.131	23:29
Breath Volume	1.671	23:29
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:31

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 11/11/2024 03:32:29

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.213	03:22
Breath Volume	1.308	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.188	03:27
Breath Volume	1.496	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 3	0.189	03:30
Breath Volume	1.457	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: HERNANDEZ GOMEZ

DATE/TIME OF TEST: 11/11/2024 04:28:06

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.221	04:23
Breath Volume	2.171	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.226	04:26
Breath Volume	3.035	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 11/15/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: ALEGRIE CASTELLANO

DATE/TIME OF TEST: 11/11/2024 20:22:33

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.080	20:17
Air Blank	0.000	20:17
Subject Sample # 1	0.107	20:18
Breath Volume	3.007	20:18
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.103	20:21
Breath Volume	3.066	20:21
Air Blank	0.000	20:21
Control Test	0.082	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 11/15/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: PUNCHES

DATE/TIME OF TEST: 11/12/2024 01:04:43

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.158	01:00
Breath Volume	2.570	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.152	01:03
Breath Volume	1.972	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 11/15/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: SHEEHAN

DATE/TIME OF TEST: 11/12/2024 02:43:25

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.162	02:38
Breath Volume	1.656	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.168	02:41
Breath Volume	2.957	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 11/15/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LARRABEE, JONATHAN
ARRESTING OFFICER AGENCY: FL Atlantic UPD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 11/12/2024 21:34:56

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.080	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.276	21:30
Breath Volume	2.667	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:32
Subject Sample # 2	0.267	21:33
Breath Volume	2.648	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 11/15/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NIEVES,JUAN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: AUTEN

DATE/TIME OF TEST: 11/13/2024 01:44:41

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.000	01:40
Breath Volume	2.558	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.000	01:43
Breath Volume	2.621	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 11/15/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KUZNETSOV, NIKITA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BECKERICH

DATE/TIME OF TEST: 11/13/2024 23:49:35

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.184	23:45
Breath Volume	1.562	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.190	23:48
Breath Volume	1.449	23:48
Air Blank	0.000	23:48
Control Test	0.078	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 11/15/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KUZNETSOV, NIKITA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 11/14/2024 22:11:42

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.081	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.000	22:07
Breath Volume	1.671	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.000	22:10
Breath Volume	1.488	22:10
Air Blank	0.000	22:10
Control Test	0.080	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 11/15/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NEILEY,STEVEN
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: FOLTZ

DATE/TIME OF TEST: 11/10/2024 02:48:36

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.000	02:44
Breath Volume	2.984	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.000	02:47
Breath Volume	3.375	02:47
Air Blank	0.000	02:47
Control Test	0.082	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 11/26/2024 19:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/11/2024 03:21:29

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.056	03:17
Breath Volume	3.191	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.057	03:19
Breath Volume	3.074	03:19
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 11/26/2024 19:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BROWN,JARMARUS
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CARO

DATE/TIME OF TEST: 11/19/2024 00:52:22

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.115	00:47
Breath Volume	1.585	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.117	00:50
Breath Volume	1.562	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 11/26/2024 19:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: METCALFE

DATE/TIME OF TEST: 11/24/2024 03:22:39

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.082	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.163	03:18
Breath Volume	2.550	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.167	03:20
Breath Volume	2.656	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 11/26/2024 19:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 11/06/2024 04:41:18

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.000	04:37
Breath Volume	2.109	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.000	04:39
Breath Volume	1.632	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 12/10/2024 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT, COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/17/2024 19:10:55

Diagnostics Check	OK	19:04
Air Blank	0.000	19:04
Control Test	0.082	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.000	19:06
Breath Volume	3.144	19:06
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.000	19:09
Breath Volume	3.339	19:09
Air Blank	0.000	19:09
Control Test	0.079	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 12/10/2024 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 13:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 11/21/2024 14:27:38

Diagnostics Check	OK	14:20
Air Blank	0.000	14:21
Control Test	0.082	14:21
Air Blank	0.000	14:22
Subject Sample # 1	0.000	14:22
Breath Volume	1.121	14:22
Air Blank	0.000	14:23
Air Blank	0.000	14:25
Subject Sample # 2	0.000	14:26
Breath Volume	1.687	14:26
Air Blank	0.000	14:26
Control Test	0.083	14:27
Air Blank	0.000	14:27
Diagnostics Check	OK	14:27

Data Download Date/Time: 12/10/2024 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: PEREZ GOMEZ

DATE/TIME OF TEST: 11/24/2024 02:42:09

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1 0.253	02:38	Volume Not Met
Breath Volume	0.097	02:38	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 2	02:41	Sequence Aborted
Breath Volume	0.000	02:41	
Air Blank	0.000	02:42	

Data Download Date/Time: 12/10/2024 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: RIGSBY

DATE/TIME OF TEST: 11/26/2024 20:25:49

Diagnostics Check	OK	20:19
Air Blank	0.000	20:20
Control Test	0.081	20:20
Air Blank	0.000	20:21
Subject Sample # 1	0.000	20:21
Breath Volume	2.398	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:24
Subject Sample # 2	0.000	20:24
Breath Volume	2.734	20:24
Air Blank	0.000	20:24
Control Test	0.079	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 12/10/2024 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,HARRY,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GITHER,ZACHARY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: MYLES

DATE/TIME OF TEST: 11/28/2024 20:15:39

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.082	20:09
Air Blank	0.000	20:09
Subject Sample # 1	0.093	20:10
Breath Volume	2.214	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.090	20:13
Breath Volume	2.464	20:13
Air Blank	0.000	20:14
Control Test	0.078	20:15
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 12/10/2024 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: STEVERSON

DATE/TIME OF TEST: 11/30/2024 21:44:34

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.082	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.145	21:39
Breath Volume	2.371	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.143	21:42
Breath Volume	1.894	21:42
Air Blank	0.000	21:43
Control Test	0.081	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 12/10/2024 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUZ,ASHLEY,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CANOVAS,ALEXANDER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 11/03/2024 02:31:42

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.079	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Control Test	0.079	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 11/27/2024 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITT,JESSE,T
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARVAJAL,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: BUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: CHIRINOS

DATE/TIME OF TEST: 11/11/2024 21:37:29

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:35	
Control Test	0.079	21:35	
Air Blank	0.000	21:35	
Subject Sample	# 1	21:36	Subject Test Refused
Breath Volume	0.000	21:36	
Air Blank	0.000	21:36	
Control Test	0.079	21:36	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 11/27/2024 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ECHEVERRI,DANIEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ECHEVERRI,DANIEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: DUI

ARREST TIME: 00:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 11/16/2024 02:26:12

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.074	02:25	Control Outside Tolerance
Air Blank	0.000	02:26	

Data Download Date/Time: 11/27/2024 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ECHEVERRI,DANIEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ECHEVERRI,DANIEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: DUI

ARREST TIME: 00:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 11/16/2024 02:33:54

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.073	02:33	Control Outside Tolerance
Air Blank	0.000	02:33	

Data Download Date/Time: 11/27/2024 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ECHEVERRI,DANIEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ECHEVERRI,DANIEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: DUI

ARREST TIME: 00:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 11/16/2024 02:37:07

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.073	02:36	Control Outside Tolerance
Air Blank	0.000	02:37	

Data Download Date/Time: 11/27/2024 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ECHEVERRI,DANIEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ECHEVERRI,DANIEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: DUI

ARREST TIME: 00:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: L234762350000

DATE/TIME OF TEST: 11/16/2024 02:51:14

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.073	02:50	Control Outside Tolerance
Air Blank	0.000	02:51	

Data Download Date/Time: 11/27/2024 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAZMURI,KEYSER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GAZMURI,KEYSER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: LORENZINO

DATE/TIME OF TEST: 11/21/2024 06:28:33

Diagnostics Check	OK	06:24	
Air Blank	0.000	06:24	
Control Test	0.079	06:24	
Air Blank	0.000	06:25	
Subject Sample	# 1	06:27	Subject Test Refused
Breath Volume	0.000	06:27	
Air Blank	0.000	06:27	
Control Test	0.079	06:27	
Air Blank	0.000	06:28	
Diagnostics Check	OK	06:28	

Data Download Date/Time: 11/27/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCNEIL,SHELDON,R

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MCNEIL,SHELDON

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402367487

DRY GAS EXPIRY DATE: 04/28/2025

VIOLATION CODE: DUI

ARREST TIME: 14:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: VERDECIA PEREZ

DATE/TIME OF TEST: 11/21/2024 16:16:46

Diagnostics Check	OK	16:13	
Air Blank	0.000	16:14	
Control Test	0.079	16:14	
Air Blank	0.000	16:15	
Subject Sample	# 1	16:15	Subject Test Refused
Breath Volume	0.000	16:15	
Air Blank	0.000	16:15	
Control Test	0.079	16:16	
Air Blank	0.000	16:16	
Diagnostics Check	OK	16:16	

Data Download Date/Time: 11/27/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KAYLA,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TOLIVER,BERNETTA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: CHAMORRO

DATE/TIME OF TEST: 11/22/2024 07:40:20

Diagnostics Check	OK	07:37	
Air Blank	0.000	07:37	
Control Test	0.079	07:38	
Air Blank	0.000	07:38	
Subject Sample	# 1	07:38	Subject Test Refused
Breath Volume	0.000	07:38	
Air Blank	0.000	07:39	
Control Test	0.078	07:39	
Air Blank	0.000	07:40	
Diagnostics Check	OK	07:40	

Data Download Date/Time: 11/27/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARIN,MARIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: PITTERS

DATE/TIME OF TEST: 11/02/2024 00:58:06

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.079	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.106	00:53
Breath Volume	1.593	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.112	00:56
Breath Volume	1.433	00:56
Air Blank	0.000	00:57
Control Test	0.078	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,OSCAR,D
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN,JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 11/03/2024 00:19:27

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.108	00:15
Breath Volume	1.898	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.107	00:17
Breath Volume	1.437	00:17
Air Blank	0.000	00:18
Control Test	0.077	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: PENNING

DATE/TIME OF TEST: 11/05/2024 20:44:56

Diagnostics Check	OK	20:33	
Air Blank	0.000	20:33	
Control Test	0.079	20:33	
Air Blank	0.000	20:34	
Subject Sample	# 1 0.201	20:37	Slope Not Level
Breath Volume	1.253	20:37	
Air Blank	0.000	20:38	
Air Blank	0.000	20:40	
Subject Sample	# 2 0.195	20:43	Volume Not Met
Breath Volume	1.042	20:43	
Air Blank	0.000	20:43	
Control Test	0.075	20:44	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: PENNING

DATE/TIME OF TEST: 11/05/2024 21:16:42

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:08	
Control Test	0.077	21:09	
Air Blank	0.000	21:09	
Subject Sample	# 1 0.182	21:12	Volume Not Met
Breath Volume	1.062	21:12	
Air Blank	0.000	21:13	
Air Blank	0.000	21:15	
Subject Sample	# 2	21:15	Subject Test Refused
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.075	21:16	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN, DAVID,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GUZMAN, DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 11/09/2024 21:00:33

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:57	
Control Test	0.080	20:57	
Air Blank	0.000	20:58	
Subject Sample	# 1	20:59	Subject Test Refused
Breath Volume	0.000	20:59	
Air Blank	0.000	20:59	
Control Test	0.078	20:59	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: UBIEDA BELISARIO

DATE/TIME OF TEST: 11/10/2024 01:55:10

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.095	01:50
Breath Volume	1.363	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.094	01:53
Breath Volume	1.687	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO-MARTINEZ, KEVEN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA, WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RODRIGUEZ ROJAS

DATE/TIME OF TEST: 11/11/2024 03:10:36

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 11/25/2024 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/11/2024 23:44:23

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.146	23:39
Breath Volume	2.656	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.147	23:42
Breath Volume	2.687	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 11/25/2024 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ, OSCAR, D
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: DOMINGUEZ, OSCAR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ROYERO BLANCO

DATE/TIME OF TEST: 11/16/2024 23:33:28

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:31	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Control Test	0.078	23:32	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 11/25/2024 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JAMIE,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MILLER,JAMIE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 11/18/2024 22:10:49

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.000	22:06
Breath Volume	1.992	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.000	22:09
Breath Volume	1.972	22:09
Air Blank	0.000	22:09
Control Test	0.080	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 11/25/2024 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN,JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: MC COLLUM

DATE/TIME OF TEST: 11/20/2024 21:45:36

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:42	
Control Test	0.079	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:44	Subject Test Refused
Breath Volume	0.000	21:44	
Air Blank	0.000	21:44	
Control Test	0.079	21:44	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 11/25/2024 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MILLER,JAIME
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 11/23/2024 02:24:59

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.077	02:20
Breath Volume	1.261	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.079	02:23
Breath Volume	1.644	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 11/25/2024 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDON DEL RIO, RODOLFO, E
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RONDON DEL RIO, RODOFLO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/24/2024 07:23:41

Diagnostics Check	OK	07:16
Air Blank	0.000	07:16
Control Test	0.079	07:17
Air Blank	0.000	07:17
Subject Sample # 1	0.069	07:18
Breath Volume	2.410	07:18
Air Blank	0.000	07:19
Air Blank	0.000	07:21
Subject Sample # 2	0.065	07:22
Breath Volume	2.167	07:22
Air Blank	0.000	07:22
Control Test	0.079	07:23
Air Blank	0.000	07:23
Diagnostics Check	OK	07:23

Data Download Date/Time: 11/25/2024 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSQUERA,SANTIAGO,D
BREATH TEST OPERATOR AGENCY: FFWCC (NW REGION)
ARRESTING OFFICER: MOSQUERA,SANTIAGO
ARRESTING OFFICER AGENCY: FFWCC (NW REGION)
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: BUI
ARREST TIME: 15:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 11/24/2024 16:43:50

Diagnostics Check	OK	16:36	
Air Blank	0.000	16:36	
Control Test	0.079	16:37	
Air Blank	0.000	16:37	
Subject Sample	# 1 0.236	16:38	
Breath Volume	2.250	16:38	
Air Blank	0.000	16:39	
Air Blank	0.000	16:41	
Subject Sample	# 2 0.235	16:41	
Breath Volume	2.500	16:41	
Air Blank	0.000	16:42	
Control Test	0.074	16:43	Control Outside Tolerance
Air Blank	0.000	16:43	

Data Download Date/Time: 11/25/2024 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 10/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 11/03/2024 06:53:38

Diagnostics Check	OK	06:47
Air Blank	0.000	06:47
Control Test	0.082	06:48
Air Blank	0.000	06:48
Subject Sample # 1	0.179	06:49
Breath Volume	1.105	06:49
Air Blank	0.000	06:49
Air Blank	0.000	06:51
Subject Sample # 2	0.175	06:52
Breath Volume	1.363	06:52
Air Blank	0.000	06:52
Control Test	0.081	06:53
Air Blank	0.000	06:53
Diagnostics Check	OK	06:53

Data Download Date/Time: 11/22/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 10/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 11/08/2024 02:53:33

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.082	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.226	02:48
Breath Volume	2.160	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.213	02:51
Breath Volume	2.078	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 11/22/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 10/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PEREZ DENIS

DATE/TIME OF TEST: 11/16/2024 03:49:47

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.082	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.165	03:44
Breath Volume	1.582	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.158	03:48
Breath Volume	1.539	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 11/22/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: TOWNE,SHAWN
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 10/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: GINES PINTO

DATE/TIME OF TEST: 11/16/2024 17:17:51

Diagnostics Check	OK	17:14	
Air Blank	0.000	17:15	
Control Test	0.082	17:15	
Air Blank	0.000	17:16	
Subject Sample	# 1	17:16	Subject Test Refused
Breath Volume	0.000	17:16	
Air Blank	0.000	17:16	
Control Test	0.081	17:17	
Air Blank	0.000	17:17	
Diagnostics Check	OK	17:17	

Data Download Date/Time: 11/22/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUEVARA,ANDRE
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 15:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: REYNOLDS JR

DATE/TIME OF TEST: 11/06/2024 17:28:23

Diagnostics Check	OK	17:21
Air Blank	0.000	17:21
Control Test	0.079	17:22
Air Blank	0.000	17:22
Subject Sample # 1	0.000	17:23
Breath Volume	1.910	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.000	17:27
Breath Volume	1.746	17:27
Air Blank	0.000	17:27
Control Test	0.078	17:27
Air Blank	0.000	17:28
Diagnostics Check	OK	17:28

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REDFEARN,JAMES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: PELTON

DATE/TIME OF TEST: 11/08/2024 02:30:24

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.078	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.105	02:24
Breath Volume	1.421	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.100	02:28
Breath Volume	1.585	02:28
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEONE,DANIEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: BERNAL

DATE/TIME OF TEST: 11/09/2024 07:58:37

Diagnostics Check	OK	07:52
Air Blank	0.000	07:52
Control Test	0.079	07:52
Air Blank	0.000	07:53
Subject Sample # 1	0.139	07:54
Breath Volume	1.859	07:54
Air Blank	0.000	07:54
Air Blank	0.000	07:56
Subject Sample # 2	0.140	07:57
Breath Volume	2.195	07:57
Air Blank	0.000	07:57
Control Test	0.079	07:58
Air Blank	0.000	07:58
Diagnostics Check	OK	07:58

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEONE,DANIEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BOUDRIE

DATE/TIME OF TEST: 11/14/2024 21:31:58

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.077	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.071	21:27
Breath Volume	1.250	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:29
Subject Sample # 2	0.067	21:30
Breath Volume	2.589	21:30
Air Blank	0.000	21:31
Control Test	0.077	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGUROS,BRIAN
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: JEAN-CHARLES

DATE/TIME OF TEST: 11/15/2024 19:00:45

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.077	18:55
Air Blank	0.000	18:55
Subject Sample # 1	0.292	18:56
Breath Volume	2.257	18:56
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.312	18:59
Breath Volume	2.992	18:59
Air Blank	0.000	18:59
Control Test	0.078	19:00
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOPEZ,JORDAN
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MARTINEZ RENDEROS

DATE/TIME OF TEST: 11/16/2024 04:14:15

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.246	04:08
Breath Volume	2.046	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.244	04:12
Breath Volume	2.839	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: GOLDBERG

DATE/TIME OF TEST: 11/17/2024 19:27:24

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.077	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.214	19:22
Breath Volume	2.632	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.216	19:25
Breath Volume	2.703	19:25
Air Blank	0.000	19:26
Control Test	0.078	19:26
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: DA SILVA NETO

DATE/TIME OF TEST: 11/17/2024 20:49:36

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.077	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.266	20:44
Breath Volume	1.617	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.268	20:47
Breath Volume	2.109	20:47
Air Blank	0.000	20:48
Control Test	0.077	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/18/2024 02:21:33

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.077	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.131	02:16
Breath Volume	1.589	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.132	02:19
Breath Volume	1.777	02:19
Air Blank	0.000	02:20
Control Test	0.077	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: FIGUEROA UMANZOR

DATE/TIME OF TEST: 11/19/2024 23:22:28

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.078	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.184	23:17
Breath Volume	2.335	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.186	23:20
Breath Volume	1.429	23:20
Air Blank	0.000	23:21
Control Test	0.078	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RYAN,JASON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: SANCHEZ RAMIREZ

DATE/TIME OF TEST: 11/22/2024 02:12:21

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.172	02:07
Breath Volume	1.542	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.158	02:10
Breath Volume	1.921	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PEREZ TAVERAS

DATE/TIME OF TEST: 11/24/2024 00:12:40

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.079	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.099	00:08
Breath Volume	2.652	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.098	00:11
Breath Volume	2.746	00:11
Air Blank	0.000	00:11
Control Test	0.078	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ANCIANO

DATE/TIME OF TEST: 11/24/2024 02:27:20

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.079	02:22
Breath Volume	2.449	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.069	02:25
Breath Volume	1.289	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 12/11/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: JAHOVIC

DATE/TIME OF TEST: 11/24/2024 22:19:29

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.078	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.263	22:14
Breath Volume	2.191	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.264	22:17
Breath Volume	2.671	22:17
Air Blank	0.000	22:18
Control Test	0.077	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 12/11/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PHILLIP,KYLE
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 11/24/2024 22:48:07

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.078	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1 0.206	22:40	Volume Not Met
Breath Volume	0.378	22:40	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample	# 2 0.205	22:46	Volume Not Met
Breath Volume	0.503	22:46	
Air Blank	0.000	22:47	
Control Test	0.076	22:47	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 12/11/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BLACKWOOD

DATE/TIME OF TEST: 11/24/2024 22:59:34

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.076	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.175	22:55
Breath Volume	1.312	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.182	22:57
Breath Volume	1.562	22:57
Air Blank	0.000	22:58
Control Test	0.076	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 12/11/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 11/25/2024 01:23:35

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.078	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.220	01:19
Breath Volume	1.585	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.230	01:22
Breath Volume	1.644	01:22
Air Blank	0.000	01:22
Control Test	0.077	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 12/11/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FAHQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MEDINA CHAVEZ

DATE/TIME OF TEST: 11/27/2024 01:28:44

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.078	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.154	01:24
Breath Volume	3.089	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.148	01:27
Breath Volume	3.093	01:27
Air Blank	0.000	01:27
Control Test	0.077	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 12/11/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: RICHMOND

DATE/TIME OF TEST: 11/29/2024 03:11:56

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.148	03:06
Breath Volume	2.457	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.144	03:10
Breath Volume	1.628	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 12/11/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOWE,RACHEL
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 11/29/2024 03:54:15

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.097	03:49
Breath Volume	1.707	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.096	03:52
Breath Volume	2.292	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 12/11/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KIRKLAND,MICHAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 11/29/2024 17:40:07

Diagnostics Check	OK	17:33
Air Blank	0.000	17:34
Control Test	0.078	17:34
Air Blank	0.000	17:35
Subject Sample # 1	0.279	17:35
Breath Volume	4.445	17:35
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample # 2	0.279	17:38
Breath Volume	3.242	17:38
Air Blank	0.000	17:39
Control Test	0.078	17:39
Air Blank	0.000	17:40
Diagnostics Check	OK	17:40

Data Download Date/Time: 12/11/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AYALA,RAUDEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 11/29/2024 20:50:36

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.078	20:45
Air Blank	0.000	20:45
Subject Sample # 1	0.203	20:46
Breath Volume	2.328	20:46
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.200	20:48
Breath Volume	2.628	20:48
Air Blank	0.000	20:49
Control Test	0.077	20:49
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 12/11/2024 09:47