

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001275

Instrument Registered To: HIALEAH GARDENS P.D.

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, ROLANDO,

Breath Test Operator Agency: Hialeah Gardens PD

Arresting Officer: RODRIGUEZ, ROLANDO

Arresting Officer Agency: Hialeah Gardens PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610001G

Dry Gas Expiry Date: 04/12/2008

Violation Code: DUI

Arrest Time: 21:31

Gender: Male

Observation Period Began: 21:45

Subject's Last Name: HERNANDEZ

Date/Time of Test: 12/01/2007 22:10

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:06	
Control Test	0.079	22:07	
Air Blank	0.000	22:07	
Subject Sample #1		22:09	Subject Test Refused
Breath Volume	0.000	22:09	
Air Blank	0.000	22:10	
Control Test	0.079	22:10	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

Data Download Date/Time: 12/11/2007 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 04:15

Gender: Male

Observation Period Began: 04:21

Subject's Last Name: SOTO

Date/Time of Test: 12/01/2007 04:58

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:55	
Control Test	0.078	04:55	
Air Blank	0.000	04:56	
Subject Sample #1		04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:57	
Control Test	0.079	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 12/11/2007 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: LEY, LAZARO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: ENRIQUEZ, JUAN

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 21:25

Gender: Male

Observation Period Began: 22:08

Subject's Last Name: COUTURE

Date/Time of Test: 12/07/2007 22:35

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.060	22:34	Control Outside Tolerance
Air Blank	0.000	22:35	

Data Download Date/Time: 12/11/2007 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: LEY, LAZARO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: ENRIQUEZ, JUAN

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 21:25

Gender: Male

Observation Period Began: 22:08

Subject's Last Name: COUTURE

Date/Time of Test: 12/07/2007 22:44

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.078	22:42	
Air Blank	0.000	22:43	
Subject Sample #1		22:43	Subject Test Refused
Breath Volume	0.000	22:43	
Air Blank	0.000	22:43	
Control Test	0.079	22:44	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 12/11/2007 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 00:09

Gender: Male

Observation Period Began: 00:29

Subject's Last Name: CHAVARRIA

Date/Time of Test: 12/08/2007 00:59

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Subject Sample #1	0.145	00:55
Breath Volume	2.531	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample #2	0.148	00:58
Breath Volume	1.367	00:58
Air Blank	0.000	00:58
Control Test	0.077	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 12/11/2007 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PARKE, JEANETTE,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: GUTIERREZ, MARIO

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 00:45

Gender: Male

Observation Period Began: 01:21

Subject's Last Name: SUMANO

Date/Time of Test: 12/08/2007 01:46

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.078	01:43	
Air Blank	0.000	01:44	
Subject Sample #1		01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.078	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 12/11/2007 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: CLOSIUS, RICHARD

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 01:45

Gender: Male

Observation Period Began: 02:06

Subject's Last Name: DIAZ

Date/Time of Test: 12/08/2007 02:34

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.059	02:29
Breath Volume	2.851	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample #2	0.055	02:32
Breath Volume	2.980	02:32
Air Blank	0.000	02:33
Control Test	0.076	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 12/11/2007 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: PACHECO, GEORGE

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 03:28

Gender: Male

Observation Period Began: 03:57

Subject's Last Name: VILLALOBOS

Date/Time of Test: 12/08/2007 04:31

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:29	
Control Test	0.078	04:29	
Air Blank	0.000	04:30	
Subject Sample #1		04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:30	
Control Test	0.078	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 12/11/2007 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: PACHECO, GEORGE

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 21:41

Gender: Male

Observation Period Began: 22:21

Subject's Last Name: OSORTO

Date/Time of Test: 12/21/2007 22:58

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.078	22:53
Air Blank	0.000	22:53
Subject Sample #1	0.158	22:54
Breath Volume	2.351	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample #2	0.172	22:57
Breath Volume	2.281	22:57
Air Blank	0.000	22:57
Control Test	0.078	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 01/18/2008 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: CLOSIUS, RICHARD

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 22:58

Gender: Male

Observation Period Began: 23:12

Subject's Last Name: MONIER SANCHEZ

Date/Time of Test: 12/21/2007 23:43

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:40	
Control Test	0.078	23:40	
Air Blank	0.000	23:41	
Subject Sample #1		23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:42	
Control Test	0.078	23:42	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 01/18/2008 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: LEY, LAZARO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: SIROIS, CHARLEY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 23:20

Gender: Male

Observation Period Began: 23:42

Subject's Last Name: ESTUPINAN

Date/Time of Test: 12/22/2007 00:12

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.078	00:07
Air Blank	0.000	00:08
Subject Sample #1	0.148	00:08
Breath Volume	1.953	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample #2	0.149	00:11
Breath Volume	1.734	00:11
Air Blank	0.000	00:12
Control Test	0.077	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 01/18/2008 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: RUTLEDGE, CHRISTOPHER

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 00:01

Gender: Male

Observation Period Began: 00:42

Subject's Last Name: HERNANDEZ

Date/Time of Test: 12/22/2007 01:18

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.078	01:13
Air Blank	0.000	01:13
Subject Sample #1	0.237	01:14
Breath Volume	2.394	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:16
Subject Sample #2	0.225	01:17
Breath Volume	2.382	01:17
Air Blank	0.000	01:18
Control Test	0.077	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 01/18/2008 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PARKE, JEANETTE,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DEROSA, TONY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 02:11

Subject's Last Name: CASTELLANOS

Date/Time of Test: 12/22/2007 02:37

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.078	02:35	
Air Blank	0.000	02:35	
Subject Sample #1		02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:36	
Control Test	0.078	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 01/18/2008 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: VAN GILS, PATRICK

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 01:20

Gender: Male

Observation Period Began: 02:09

Subject's Last Name: LOPEZ FERRER

Date/Time of Test: 12/22/2007 02:49

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:43
Air Blank	0.000	02:44
Subject Sample #1	0.098	02:45
Breath Volume	1.671	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample #2	0.091	02:48
Breath Volume	2.042	02:48
Air Blank	0.000	02:48
Control Test	0.077	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 01/18/2008 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: DAVENPORT, JERRY, J

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 02:02

Subject's Last Name: LORA

Date/Time of Test: 12/30/2007 02:29

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.078	02:26	
Air Blank	0.000	02:27	
Subject Sample #1		02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 01/18/2008 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001276

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 702401G

Dry Gas Expiry Date: 01/26/2009

Violation Code: DUI

Arrest Time: 02:15

Gender: Male

Observation Period Began: 02:33

Subject's Last Name: ALMENDARES

Date/Time of Test: 12/30/2007 03:10

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:04
Subject Sample #1	0.119	03:05
Breath Volume	1.503	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample #2	0.115	03:08
Breath Volume	2.109	03:08
Air Blank	0.000	03:09
Control Test	0.078	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 01/18/2008 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: FRANCO, GREGORY, R

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:55

Gender: Male

Observation Period Began: 02:09

Subject's Last Name: MORUARTE-JIMENEZ

Date/Time of Test: 12/05/2007 02:37

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample #1	0.136	02:33
Breath Volume	1.562	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample #2	0.124	02:36
Breath Volume	1.214	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 12/11/2007 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: FRANCO, GREGORY, R

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: PEREZ, CARLOS

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:42

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: GONZALEZ

Date/Time of Test: 12/06/2007 01:38

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.079	01:36	
Air Blank	0.000	01:36	
Subject Sample #1		01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 12/11/2007 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: FRANCO, GREGORY, R

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: CLOSIUS, RICK

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:32

Gender: Male

Observation Period Began: 02:06

Subject's Last Name: SIRVAS

Date/Time of Test: 12/06/2007 02:34

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:29
Subject Sample #1	0.220	02:30
Breath Volume	1.703	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample #2	0.222	02:33
Breath Volume	1.792	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 12/11/2007 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: FRANCO, GREGORY, R

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:47

Gender: Male

Observation Period Began: 02:02

Subject's Last Name: MALONE

Date/Time of Test: 12/07/2007 02:39

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Subject Sample #1	0.108	02:34
Breath Volume	1.398	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample #2	0.107	02:37
Breath Volume	1.167	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 12/11/2007 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: CLOSIUS, RICHARD

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:10

Gender: Female

Observation Period Began: 03:36

Subject's Last Name: VENTURI

Date/Time of Test: 12/07/2007 04:07

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:05	
Control Test	0.079	04:05	
Air Blank	0.000	04:06	
Subject Sample #1		04:06	Subject Test Refused
Breath Volume	0.000	04:06	
Air Blank	0.000	04:06	
Control Test	0.078	04:07	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 12/11/2007 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: DAVENPORT, JERRY, J

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:00

Gender: Female

Observation Period Began: 03:49

Subject's Last Name: ZANGRONIZ

Date/Time of Test: 12/07/2007 04:27

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Subject Sample #1	0.285	04:19
Breath Volume	1.863	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample #2	0.262	04:22
Breath Volume	1.371	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample #3	0.279	04:25
Breath Volume	2.203	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:27

Data Download Date/Time: 12/11/2007 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: PEREZ, CARLOS

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 02:45

Gender: Male

Observation Period Began: 03:02

Subject's Last Name: GARCIA

Date/Time of Test: 12/08/2007 03:28

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:27	
Subject Sample #1		03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Control Test	0.079	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 12/11/2007 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: DAVENPORT, JERRY, J

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 11/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 04:25

Gender: Male

Observation Period Began: 05:02

Subject's Last Name: BORJAS-GOMEZ

Date/Time of Test: 12/08/2007 05:39

Diagnostics Check	OK	05:30	
Air Blank	0.000	05:31	
Control Test	0.079	05:31	
Air Blank	0.000	05:32	
Subject Sample #1	0.236	05:32	
Breath Volume	1.894	05:32	
Air Blank	0.000	05:33	
Air Blank	0.000	05:35	
Subject Sample #2		05:38	Subject Test Refused
Breath Volume	0.000	05:38	
Air Blank	0.000	05:38	
Control Test	0.078	05:38	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:39	

Data Download Date/Time: 12/11/2007 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: FRANCO, GREGORY, R

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: SLIMAK, MARK

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 21:54

Gender: Male

Observation Period Began: 22:10

Subject's Last Name: ARIAS

Date/Time of Test: 12/13/2007 22:55

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:50
Subject Sample #1	0.000	22:51
Breath Volume	2.230	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample #2	0.000	22:54
Breath Volume	1.292	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:19

Gender: Male

Observation Period Began: 00:45

Subject's Last Name: MCCULLOUGH

Date/Time of Test: 12/14/2007 01:10

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:08	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Subject Sample #1		01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Control Test	0.080	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: PEREZ, CARLOS

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:00

Gender: Male

Observation Period Began: 01:13

Subject's Last Name: ARTILES MINOR

Date/Time of Test: 12/14/2007 01:41

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Subject Sample #1	0.163	01:37
Breath Volume	1.343	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample #2	0.183	01:39
Breath Volume	2.070	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: DAVENPORT, JERRY, J

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 01:52

Subject's Last Name: FERNANDEZ

Date/Time of Test: 12/14/2007 02:20

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Subject Sample #1	0.114	02:15
Breath Volume	2.140	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample #2	0.112	02:18
Breath Volume	3.402	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PACHECO, GEORGE,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: PACHECO, GEORGE

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:53

Gender: Male

Observation Period Began: 02:37

Subject's Last Name: SALINAS

Date/Time of Test: 12/14/2007 03:08

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Subject Sample #1	0.201	03:04
Breath Volume	1.890	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample #2	0.213	03:06
Breath Volume	1.300	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 04:05

Gender: Female

Observation Period Began: 04:20

Subject's Last Name: VALDEZ

Date/Time of Test: 12/14/2007 04:54

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:46
Subject Sample #1	0.082	04:47
Breath Volume	1.792	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample #2	0.081	04:52
Breath Volume	2.082	04:52
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PACHECO, GEORGE,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: PACHECO, GEORGE

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:24

Gender: Male

Observation Period Began: 01:05

Subject's Last Name: COOPER

Date/Time of Test: 12/15/2007 01:32

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:26
Subject Sample #1	0.057	01:27
Breath Volume	1.699	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample #2	0.058	01:30
Breath Volume	2.378	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PEREZ, CARLOS, E

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:28

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: CASTRO GAYOL

Date/Time of Test: 12/19/2007 02:10

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:02
Subject Sample #1	0.181	02:05
Breath Volume	1.339	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample #2	0.199	02:08
Breath Volume	1.167	02:08
Air Blank	0.000	02:09
Control Test	0.075	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:10

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PEREZ, CARLOS, E

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: CLOSIUS, RICHARD

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 02:20

Gender: Male

Observation Period Began: 02:34

Subject's Last Name: RODRIGUEZ-PERAZA

Date/Time of Test: 12/19/2007 03:02

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample #1	0.031	02:57
Breath Volume	2.636	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample #2	0.031	03:00
Breath Volume	2.527	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PEREZ, CARLOS, E

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 02:15

Gender: Male

Observation Period Began: 02:31

Subject's Last Name: BERMUDEZ

Date/Time of Test: 12/20/2007 02:56

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:54	
Subject Sample #1		02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PEREZ, CARLOS, E

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:35

Gender: Female

Observation Period Began: 04:06

Subject's Last Name: RIO

Date/Time of Test: 12/20/2007 04:30

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.080	04:28	
Air Blank	0.000	04:28	
Subject Sample #1		04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.079	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: CLOSIUS, RICHARD, L

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: SLIMAK, MARK

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:14

Gender: Male

Observation Period Began: 03:23

Subject's Last Name: PEREZ

Date/Time of Test: 12/21/2007 03:50

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.079	03:48	
Air Blank	0.000	03:48	
Subject Sample #1		03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.079	03:49	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:30

Gender: Male

Observation Period Began: 03:46

Subject's Last Name: GIL

Date/Time of Test: 12/21/2007 04:13

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample #1	0.196	04:09
Breath Volume	2.839	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:11
Subject Sample #2	0.198	04:12
Breath Volume	2.957	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 04:08

Gender: Female

Observation Period Began: 04:28

Subject's Last Name: OGRADY

Date/Time of Test: 12/21/2007 05:04

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:01	
Control Test	0.079	05:01	
Air Blank	0.000	05:01	
Subject Sample #1		05:02	Subject Test Refused
Breath Volume	0.000	05:02	
Air Blank	0.000	05:03	
Control Test	0.079	05:03	
Air Blank	0.000	05:04	
Diagnostics Check	OK	05:04	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: CLOSIUS, RICHARD, L

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: SORI, MARITZA

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 04:20

Gender: Male

Observation Period Began: 04:36

Subject's Last Name: RAMIREZ

Date/Time of Test: 12/21/2007 05:20

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:17	
Control Test	0.079	05:17	
Air Blank	0.000	05:18	
Subject Sample #1		05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:19	
Control Test	0.078	05:19	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PARKE, JEANETTE,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: MONTESINO, NESTOR

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 23:10

Gender: Male

Observation Period Began: 23:37

Subject's Last Name: YOUNG

Date/Time of Test: 12/22/2007 00:14

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.079	00:11	
Air Blank	0.000	00:12	
Subject Sample #1		00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:13	
Control Test	0.079	00:13	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: CLOSIUS, RICHARD, L

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: CLOSIUS, RICHARD

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:43

Gender: Male

Observation Period Began: 02:10

Subject's Last Name: DECCA

Date/Time of Test: 12/22/2007 02:33

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:32	
Subject Sample #1		02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.079	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: DAVENPORT, JERRY, J

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:35

Gender: Male

Observation Period Began: 04:06

Subject's Last Name: GUTIERREZ

Date/Time of Test: 12/22/2007 04:53

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.079	04:51	
Air Blank	0.000	04:52	
Subject Sample #1		04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:52	
Control Test	0.079	04:53	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PARKE, JEANETTE,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREG

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 22:19

Gender: Male

Observation Period Began: 22:39

Subject's Last Name: LOJANO-ELIZALDE

Date/Time of Test: 12/22/2007 23:08

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Subject Sample #1	0.228	23:03
Breath Volume	2.859	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample #2	0.219	23:06
Breath Volume	1.949	23:06
Air Blank	0.000	23:07
Control Test	0.078	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: FRANCO, GREGORY, R

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: CLOUSIS, RICHARD

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 23:00

Gender: Male

Observation Period Began: 23:35

Subject's Last Name: MONRROY

Date/Time of Test: 12/23/2007 00:05

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Subject Sample #1	0.127	23:59
Breath Volume	1.398	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:01
Subject Sample #2	0.118	00:04
Breath Volume	1.273	00:04
Air Blank	0.000	00:04
Control Test	0.078	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: FRANCO, GREGORY, R

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: SLIMAK, MARK

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:26

Gender: Male

Observation Period Began: 00:26

Subject's Last Name: MENDOZA-RUBIO

Date/Time of Test: 12/23/2007 01:30

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:25
Air Blank	0.000	01:25
Subject Sample #1	0.156	01:26
Breath Volume	1.636	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample #2	0.155	01:29
Breath Volume	1.625	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: DAVENPORT, JERRY, J

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:35

Gender: Male

Observation Period Began: 01:58

Subject's Last Name: BRADLEY

Date/Time of Test: 12/23/2007 02:41

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.079	02:39	
Air Blank	0.000	02:39	
Subject Sample #1		02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.079	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: FRANCO, GREGORY, R

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:00

Gender: Male

Observation Period Began: 03:38

Subject's Last Name: RICO

Date/Time of Test: 12/23/2007 04:19

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:13
Subject Sample #1	0.213	04:14
Breath Volume	1.433	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample #2	0.230	04:17
Breath Volume	1.562	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: DAVENPORT, JERRY, J

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 05:20

Gender: Male

Observation Period Began: 05:26

Subject's Last Name: RESTREPO

Date/Time of Test: 12/23/2007 05:54

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:51	
Control Test	0.079	05:52	
Air Blank	0.000	05:52	
Subject Sample #1		05:53	Subject Test Refused
Breath Volume	0.000	05:53	
Air Blank	0.000	05:53	
Control Test	0.079	05:53	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: FRANCO, GREGORY, R

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 21:23

Gender: Male

Observation Period Began: 21:43

Subject's Last Name: RUIZ

Date/Time of Test: 12/25/2007 22:15

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:10	
Control Test	0.079	22:10	
Air Blank	0.000	22:11	
Subject Sample #1		22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank	0.000	22:14	
Control Test	0.080	22:15	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: DAVENPORT, JERRY, J

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 02:02

Subject's Last Name: RAMIREZ VEGA

Date/Time of Test: 12/26/2007 02:34

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Subject Sample #1	0.000	02:30
Breath Volume	1.820	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample #2	0.000	02:33
Breath Volume	3.042	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: FRANCO, GREGORY, R

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:20

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: CARMENATE

Date/Time of Test: 12/27/2007 02:19

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample #1	0.161	02:15
Breath Volume	2.335	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample #2	0.164	02:18
Breath Volume	2.492	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: SLIMAK, MARK,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: SLIMAK, MARK

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 02:15

Gender: Female

Observation Period Began: 02:25

Subject's Last Name: ROIG

Date/Time of Test: 12/27/2007 02:51

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:46
Subject Sample #1	0.099	02:47
Breath Volume	1.789	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample #2	0.096	02:50
Breath Volume	2.464	02:50
Air Blank	0.000	02:50
Control Test	0.078	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PACHECO, GEORGE,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 01:30

Subject's Last Name: AGUIRRE

Date/Time of Test: 12/28/2007 02:16

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Subject Sample #1	0.031	02:11
Breath Volume	2.621	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample #2	0.035	02:14
Breath Volume	3.949	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: DAVENPORT, JERRY, J

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 01:47

Subject's Last Name: ALPIZAR

Date/Time of Test: 12/28/2007 02:53

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:52	
Subject Sample #1		02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.079	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: PACHECO, GEORGE,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: PACHECO, GEORGE

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 04:25

Gender: Male

Observation Period Began: 04:42

Subject's Last Name: CERRILLO

Date/Time of Test: 12/28/2007 05:32

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:29	
Control Test	0.078	05:30	
Air Blank	0.000	05:30	
Subject Sample #1		05:30	Subject Test Refused
Breath Volume	0.000	05:30	
Air Blank	0.000	05:31	
Control Test	0.078	05:31	
Air Blank	0.000	05:32	
Diagnostics Check	OK	05:32	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:32

Gender: Male

Observation Period Began: 01:23

Subject's Last Name: ALVAREZ

Date/Time of Test: 12/29/2007 01:52

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample #1	0.228	01:48
Breath Volume	1.937	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample #2	0.210	01:51
Breath Volume	1.617	01:51
Air Blank	0.000	01:52
Control Test	0.077	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:29

Gender: Female

Observation Period Began: 02:03

Subject's Last Name: GUIDO

Date/Time of Test: 12/29/2007 02:29

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.079	02:27	
Air Blank	0.000	02:28	
Subject Sample #1		02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.079	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: SLIMAK, MARK

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:18

Gender: Male

Observation Period Began: 03:34

Subject's Last Name: OLGUIN

Date/Time of Test: 12/29/2007 04:04

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:56	
Air Blank	0.000	03:56	
Subject Sample #1	0.159	03:59	Volume Not Met
Breath Volume	1.093	03:59	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample #2		04:02	Subject Test Refused
Breath Volume	0.820	04:02	
Air Blank	0.000	04:03	
Control Test	0.078	04:03	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:38

Gender: Male

Observation Period Began: 03:54

Subject's Last Name: CARVAJAL GUERRA

Date/Time of Test: 12/29/2007 04:32

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.079	04:27
Air Blank	0.000	04:27
Subject Sample #1	0.245	04:28
Breath Volume	1.683	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample #2	0.230	04:31
Breath Volume	2.617	04:31
Air Blank	0.000	04:31
Control Test	0.078	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: DAVENPORT, JERRY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 04:35

Gender: Female

Observation Period Began: 05:06

Subject's Last Name: ALVARENGA

Date/Time of Test: 12/29/2007 05:34

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:29
Subject Sample #1	0.052	05:29
Breath Volume	2.777	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample #2	0.051	05:32
Breath Volume	2.832	05:32
Air Blank	0.000	05:33
Control Test	0.079	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: PACHECO, GEORGE

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 04:54

Gender: Female

Observation Period Began: 05:13

Subject's Last Name: ZAMOT

Date/Time of Test: 12/29/2007 05:45

Diagnostics Check	OK	05:39
Air Blank	0.000	05:39
Control Test	0.079	05:40
Air Blank	0.000	05:40
Subject Sample #1	0.148	05:41
Breath Volume	1.191	05:41
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample #2	0.149	05:43
Breath Volume	1.273	05:43
Air Blank	0.000	05:44
Control Test	0.079	05:44
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: SLIMAK, MARK

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 06:03

Gender: Male

Observation Period Began: 06:19

Subject's Last Name: SOLORZANO

Date/Time of Test: 12/29/2007 06:51

Diagnostics Check	OK	06:44
Air Blank	0.000	06:44
Control Test	0.078	06:45
Air Blank	0.000	06:45
Subject Sample #1	0.128	06:46
Breath Volume	1.355	06:46
Air Blank	0.000	06:47
Air Blank	0.000	06:49
Subject Sample #2	0.123	06:49
Breath Volume	1.378	06:49
Air Blank	0.000	06:50
Control Test	0.078	06:50
Air Blank	0.000	06:51
Diagnostics Check	OK	06:51

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001277

Instrument Registered To: MIAMI-DADE PD

Software Version: 8100.27

Breath Test Operator: ESPINOZA, PABLO,

Breath Test Operator Agency: Miami-Dade PD

Arresting Officer: FRANCO, GREGORY

Arresting Officer Agency: Miami-Dade PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7149011

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 05:30

Gender: Male

Observation Period Began: 06:12

Subject's Last Name: GARCIA

Date/Time of Test: 12/29/2007 07:38

Diagnostics Check	OK	07:28
Air Blank	0.000	07:28
Control Test	0.078	07:28
Air Blank	0.000	07:29
Subject Sample #1	0.075	07:30
Breath Volume	2.152	07:30
Air Blank	0.000	07:30
Air Blank	0.000	07:32
Subject Sample #2	0.112	07:33
Breath Volume	2.023	07:33
Air Blank	0.000	07:34
Air Blank	0.000	07:36
Subject Sample #3	0.103	07:36
Breath Volume	3.890	07:36
Air Blank	0.000	07:37
Control Test	0.078	07:37
Air Blank	0.000	07:38
Diagnostics Check	OK	07:38

Data Download Date/Time: 01/18/2008 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BARRON, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 23:30

Gender: Male

Observation Period Began: 00:40

Subject's Last Name: PICKETT

Date/Time of Test: 12/01/2007 01:07

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample #1	0.281	01:02
Breath Volume	4.523	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample #2	0.286	01:05
Breath Volume	4.710	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: WOOLERY, W

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 23:40

Gender: Male

Observation Period Began: 01:55

Subject's Last Name: SHUMPERT

Date/Time of Test: 12/01/2007 02:21

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:16
Air Blank	0.000	02:16
Subject Sample #1	0.018	02:17
Breath Volume	3.390	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.019	02:19
Breath Volume	4.277	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BARRON, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 01:50

Gender: Male

Observation Period Began: 02:45

Subject's Last Name: RANDOLPH

Date/Time of Test: 12/01/2007 03:13

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample #1	0.170	03:08
Breath Volume	2.265	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample #2	0.169	03:11
Breath Volume	1.531	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: TAYLOR, J

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 03:00

Gender: Male

Observation Period Began: 03:23

Subject's Last Name: SNELLING

Date/Time of Test: 12/01/2007 03:51

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample #1	0.228	03:45
Breath Volume	1.425	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:47
Subject Sample #2	0.208	03:49
Breath Volume	1.367	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: MERRELL, J

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 03:00

Gender: Female

Observation Period Began: 04:53

Subject's Last Name: PEARE

Date/Time of Test: 12/01/2007 05:22

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Subject Sample #1	0.218	05:17
Breath Volume	2.363	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample #2	0.199	05:20
Breath Volume	1.847	05:20
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: LUEDERS, JUSTIN, G

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: FISSELL, M

Arresting Officer Agency: Atlantic Beach PD

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 02:40

Gender: Male

Observation Period Began: 06:42

Subject's Last Name: OTIS

Date/Time of Test: 12/02/2007 07:09

Diagnostics Check	OK	07:02
Air Blank	0.000	07:03
Control Test	0.080	07:03
Air Blank	0.000	07:03
Subject Sample #1	0.000	07:04
Breath Volume	3.894	07:04
Air Blank	0.000	07:05
Air Blank	0.000	07:07
Subject Sample #2	0.000	07:07
Breath Volume	3.941	07:07
Air Blank	0.000	07:08
Control Test	0.079	07:08
Air Blank	0.000	07:09
Diagnostics Check	OK	07:09

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BYRNE, J

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 00:45

Gender: Male

Observation Period Began: 01:35

Subject's Last Name: WALKER

Date/Time of Test: 12/03/2007 02:01

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample #1	0.141	01:57
Breath Volume	3.882	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample #2	0.121	02:00
Breath Volume	2.039	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: JACKSON, R

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: KITT

Date/Time of Test: 12/03/2007 03:06

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:01
Subject Sample #1	0.299	03:02
Breath Volume	1.769	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample #2	0.288	03:05
Breath Volume	1.667	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: SAOUD, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 01:55

Gender: Male

Observation Period Began: 02:50

Subject's Last Name: DUNBAR

Date/Time of Test: 12/04/2007 03:34

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample #1	0.231	03:30
Breath Volume	2.550	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample #2	0.247	03:33
Breath Volume	2.718	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: HOUSEND, B

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 01:00

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: DAVIS

Date/Time of Test: 12/06/2007 02:58

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:56	
Subject Sample #1		02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.080	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: EDMONDS, D

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 02:14

Gender: Male

Observation Period Began: 03:06

Subject's Last Name: JOHNSON

Date/Time of Test: 12/06/2007 03:27

Diagnostics Check	OK	03:26	
Air Blank		03:26	RFI Detect
Air Blank	0.000	03:27	

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: EDMONDS, D

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 02:14

Gender: Male

Observation Period Began: 03:06

Subject's Last Name: JOHNSON

Date/Time of Test: 12/06/2007 03:36

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample #1	0.135	03:31
Breath Volume	1.886	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample #2	0.142	03:35
Breath Volume	1.507	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: EDMONDS, D

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 01:37

Gender: Male

Observation Period Began: 02:14

Subject's Last Name: PALGUT

Date/Time of Test: 12/07/2007 02:49

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Subject Sample #1	0.023	02:44
Breath Volume	3.117	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample #2	0.022	02:47
Breath Volume	3.375	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001279

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BEESON, S, D

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: MCNAMEE, M

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 711301G

Dry Gas Expiry Date: 04/25/2009

Violation Code: DUI

Arrest Time: 03:00

Gender: Male

Observation Period Began: 03:58

Subject's Last Name: HEMENOVER

Date/Time of Test: 12/07/2007 04:30

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.080	04:20	
Air Blank	0.000	04:20	
Subject Sample #1	0.130	04:23	Volume Not Met
Breath Volume	0.035	04:23	
Air Blank	0.000	04:24	
Air Blank	0.000	04:26	
Subject Sample #2	0.092	04:29	Volume Not Met
Breath Volume	0.015	04:29	
Air Blank	0.000	04:30	
Control Test	0.078	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 12/13/2007 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001281

Instrument Registered To: NASSAU COUNTY SO

Software Version: 8100.27

Breath Test Operator: HUBBARD, BENITA, A

Breath Test Operator Agency: Nassau County SO

Arresting Officer: COOPER, TIM

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:02

Gender: Male

Observation Period Began: 21:20

Subject's Last Name: WELSH

Date/Time of Test: 12/02/2007 21:53

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.081	21:46
Air Blank	0.000	21:46
Subject Sample #1	0.132	21:47
Breath Volume	1.597	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample #2	0.122	21:51
Breath Volume	1.566	21:51
Air Blank	0.000	21:52
Control Test	0.079	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 12/03/2007 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001281

Instrument Registered To: NASSAU COUNTY SO

Software Version: 8100.27

Breath Test Operator: HUBBARD, BENITA, A

Breath Test Operator Agency: Nassau County SO

Arresting Officer: HEALY, G

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:32

Gender: Male

Observation Period Began: 22:45

Subject's Last Name: CHAMBERLAIN

Date/Time of Test: 12/02/2007 23:19

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:13
Air Blank	0.000	23:13
Subject Sample #1	0.169	23:15
Breath Volume	3.070	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample #2	0.168	23:18
Breath Volume	2.570	23:18
Air Blank	0.000	23:18
Control Test	0.079	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 12/03/2007 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001281

Instrument Registered To: NASSAU COUNTY SO

Software Version: 8100.27

Breath Test Operator: HUBBARD, BENITA, A

Breath Test Operator Agency: Nassau County SO

Arresting Officer: WITKOWSKI, J

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 12/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:09

Gender: Female

Observation Period Began: 00:25

Subject's Last Name: HAGINS

Date/Time of Test: 12/21/2007 01:04

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample #1	0.237	00:59
Breath Volume	2.070	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample #2	0.233	01:02
Breath Volume	1.492	01:02
Air Blank	0.000	01:03
Control Test	0.078	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 01/24/2008 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001281

Instrument Registered To: NASSAU COUNTY SO

Software Version: 8100.27

Breath Test Operator: CREWS, K, L

Breath Test Operator Agency: Nassau County SO

Arresting Officer: FRANK, E

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 12/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 17:20

Gender: Male

Observation Period Began: 17:45

Subject's Last Name: BRANNAN

Date/Time of Test: 12/21/2007 18:08

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:05	
Control Test	0.081	18:06	
Air Blank	0.000	18:06	
Subject Sample #1		18:06	Subject Test Refused
Breath Volume	0.000	18:06	
Air Blank	0.000	18:07	
Control Test	0.081	18:07	
Air Blank	0.000	18:08	
Diagnostics Check	OK	18:08	

Data Download Date/Time: 01/24/2008 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001281

Instrument Registered To: NASSAU COUNTY SO

Software Version: 8100.27

Breath Test Operator: CREWS, K, L

Breath Test Operator Agency: Nassau County SO

Arresting Officer: FAIR, B

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 12/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:13

Gender: Female

Observation Period Began: 04:00

Subject's Last Name: VINZANT

Date/Time of Test: 12/22/2007 04:27

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.082	04:21
Air Blank	0.000	04:22
Subject Sample #1	0.190	04:22
Breath Volume	2.140	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.192	04:25
Breath Volume	1.695	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 01/24/2008 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001281

Instrument Registered To: NASSAU COUNTY SO

Software Version: 8100.27

Breath Test Operator: ORR, I, J

Breath Test Operator Agency: Nassau County SO

Arresting Officer: LEWIS, D

Arresting Officer Agency: FDOT MCC

Last Agency Inspection: 12/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 14:34

Gender: Male

Observation Period Began: 15:12

Subject's Last Name: MARTIN

Date/Time of Test: 12/28/2007 15:39

Diagnostics Check	OK	15:33
Air Blank	0.000	15:33
Control Test	0.080	15:34
Air Blank	0.000	15:34
Subject Sample #1	0.086	15:35
Breath Volume	1.570	15:35
Air Blank	0.000	15:36
Air Blank	0.000	15:38
Subject Sample #2	0.087	15:38
Breath Volume	1.558	15:38
Air Blank	0.000	15:39
Control Test	0.079	15:39
Air Blank	0.000	15:39
Diagnostics Check	OK	15:39

Data Download Date/Time: 01/24/2008 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001281

Instrument Registered To: NASSAU COUNTY SO

Software Version: 8100.27

Breath Test Operator: NETTLES, J, W

Breath Test Operator Agency: Nassau County SO

Arresting Officer: HEPLER, J

Arresting Officer Agency: Fernandina Beach PD

Last Agency Inspection: 12/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:32

Gender: Male

Observation Period Began: 03:09

Subject's Last Name: O'GRADY

Date/Time of Test: 12/29/2007 03:36

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample #1	0.246	03:31
Breath Volume	2.230	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample #2	0.240	03:34
Breath Volume	2.335	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 01/24/2008 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001281

Instrument Registered To: NASSAU COUNTY SO

Software Version: 8100.27

Breath Test Operator: THIRSK, K, M

Breath Test Operator Agency: Nassau County SO

Arresting Officer: BALDWIN, R

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 12/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 15:48

Gender: Male

Observation Period Began: 15:58

Subject's Last Name: ORDAZ-VASQUEZ

Date/Time of Test: 12/30/2007 16:26

Diagnostics Check	OK	16:18
Air Blank	0.000	16:19
Control Test	0.081	16:19
Air Blank	0.000	16:20
Subject Sample #1	0.221	16:21
Breath Volume	1.535	16:21
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample #2	0.211	16:24
Breath Volume	2.820	16:24
Air Blank	0.000	16:25
Control Test	0.079	16:25
Air Blank	0.000	16:26
Diagnostics Check	OK	16:26

Data Download Date/Time: 01/24/2008 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: BRINSON, MARCUS, L

Breath Test Operator Agency: U of FI PD

Arresting Officer: BRINSON, MARCUS

Arresting Officer Agency: U of FI PD

Last Agency Inspection: 11/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 23:29

Gender: Male

Observation Period Began: 23:53

Subject's Last Name: MENDEZ

Date/Time of Test: 12/02/2007 00:31

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.077	00:25
Air Blank	0.000	00:26
Subject Sample #1	0.184	00:26
Breath Volume	1.875	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample #2	0.187	00:29
Breath Volume	1.636	00:29
Air Blank	0.000	00:30
Control Test	0.075	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 12/23/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: BELLEVILLE, HENRI, A

Breath Test Operator Agency: U of FI PD

Arresting Officer: BELLEVILLE, HENRI

Arresting Officer Agency: U of FI PD

Last Agency Inspection: 11/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 00:24

Gender: Male

Observation Period Began: 00:50

Subject's Last Name: TAXDAL

Date/Time of Test: 12/02/2007 01:26

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.078	01:23	
Air Blank	0.000	01:24	
Subject Sample #1		01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Control Test	0.078	01:25	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 12/23/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: PECK, TIMOTHY, L

Breath Test Operator Agency: U of FI PD

Arresting Officer: LEDGER, WILLIAM

Arresting Officer Agency: U of FI PD

Last Agency Inspection: 11/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 01:50

Gender: Male

Observation Period Began: 02:09

Subject's Last Name: EGER

Date/Time of Test: 12/02/2007 02:39

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.077	02:37	
Air Blank	0.000	02:38	
Subject Sample #1		02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:38	
Control Test	0.078	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 12/23/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: BELLEVILLE, HENRI, A

Breath Test Operator Agency: U of FI PD

Arresting Officer: BELLEVILLE, HENRI

Arresting Officer Agency: U of FI PD

Last Agency Inspection: 11/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 02:45

Gender: Female

Observation Period Began: 03:10

Subject's Last Name: MERRITT

Date/Time of Test: 12/06/2007 03:40

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.076	03:33
Air Blank	0.000	03:33
Subject Sample #1	0.167	03:34
Breath Volume	1.781	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample #2	0.180	03:38
Breath Volume	1.484	03:38
Air Blank	0.000	03:39
Control Test	0.076	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 12/23/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: BOYD, BRIAN, J

Breath Test Operator Agency: U of FI PD

Arresting Officer: BOYD, BRIAN

Arresting Officer Agency: U of FI PD

Last Agency Inspection: 11/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 00:16

Gender: Male

Observation Period Began: 00:35

Subject's Last Name: MCMULLEN

Date/Time of Test: 12/10/2007 01:05

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.078	01:03	
Air Blank	0.000	01:03	
Subject Sample #1		01:03	RFI Detect
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Control Test	0.079	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 12/23/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: BOYD, BRIAN, J

Breath Test Operator Agency: U of FI PD

Arresting Officer: BOYD, BRIAN

Arresting Officer Agency: U of FI PD

Last Agency Inspection: 11/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 00:16

Gender: Male

Observation Period Began: 00:35

Subject's Last Name: MCMULLEN

Date/Time of Test: 12/10/2007 01:13

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Subject Sample #1	0.149	01:09
Breath Volume	3.941	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:11
Subject Sample #2	0.143	01:12
Breath Volume	2.539	01:12
Air Blank	0.000	01:13
Control Test	0.077	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 12/23/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: SMITH, KEITH, J

Breath Test Operator Agency: U of FI PD

Arresting Officer: HAMBY, WALT

Arresting Officer Agency: U of FI PD

Last Agency Inspection: 11/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 02:11

Gender: Male

Observation Period Began: 02:27

Subject's Last Name: KENT

Date/Time of Test: 12/13/2007 03:09

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.078	03:03
Air Blank	0.000	03:04
Subject Sample #1	0.182	03:05
Breath Volume	3.367	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample #2	0.183	03:08
Breath Volume	2.738	03:08
Air Blank	0.000	03:08
Control Test	0.078	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 12/23/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: BOYD, BRIAN, J

Breath Test Operator Agency: U of FI PD

Arresting Officer: BOYD, BRIAN

Arresting Officer Agency: U of FI PD

Last Agency Inspection: 11/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 02:36

Gender: Male

Observation Period Began: 02:55

Subject's Last Name: COOK

Date/Time of Test: 12/13/2007 03:26

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.078	03:20
Air Blank	0.000	03:20
Subject Sample #1	0.112	03:21
Breath Volume	1.199	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample #2	0.112	03:24
Breath Volume	2.148	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 12/23/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: HUTCHINSON, ERIC, L

Breath Test Operator Agency: U of FI PD

Arresting Officer: WATSON, DAVID

Arresting Officer Agency: U of FI PD

Last Agency Inspection: 11/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 00:51

Gender: Female

Observation Period Began: 01:10

Subject's Last Name: KRUSZEWSKI

Date/Time of Test: 12/15/2007 01:43

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.077	01:37
Air Blank	0.000	01:37
Subject Sample #1	0.192	01:38
Breath Volume	1.671	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample #2	0.186	01:41
Breath Volume	1.558	01:41
Air Blank	0.000	01:42
Control Test	0.075	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 12/23/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: ADAMS, ANDY, B

Breath Test Operator Agency: U of FI PD

Arresting Officer: ADAMS, ANDY

Arresting Officer Agency: U of FI PD

Last Agency Inspection: 11/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 01:17

Gender: Male

Observation Period Began: 01:35

Subject's Last Name: GRIFFIN

Date/Time of Test: 12/15/2007 02:01

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.078	01:59	
Air Blank	0.000	01:59	
Subject Sample #1		02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.078	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 12/23/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001282

Instrument Registered To: U OF FLORIDA PD

Software Version: 8100.27

Breath Test Operator: PRIVETTE, MICHAEL, L

Breath Test Operator Agency: Alachua County SO

Arresting Officer: PRIVETTE, MICHAEL

Arresting Officer Agency: Alachua County SO

Last Agency Inspection: 12/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 18:55

Gender: Female

Observation Period Began: 19:30

Subject's Last Name: ANDERSON

Date/Time of Test: 12/30/2007 20:07

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.078	20:01
Air Blank	0.000	20:02
Subject Sample #1	0.294	20:03
Breath Volume	1.304	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:05
Subject Sample #2	0.286	20:06
Breath Volume	1.621	20:06
Air Blank	0.000	20:06
Control Test	0.077	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 01/30/2008 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001284

Instrument Registered To: GILCHRIST COUNTY SO

Software Version: 8100.27

Breath Test Operator: STALVEY, STEPHEN, W

Breath Test Operator Agency: Gilchrist County SO

Arresting Officer: STALVEY, STEPHEN

Arresting Officer Agency: Gilchrist County SO

Last Agency Inspection: 11/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 22:25

Gender: Male

Observation Period Began: 23:10

Subject's Last Name: HAMOR

Date/Time of Test: 12/01/2007 23:55

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.081	23:49
Air Blank	0.000	23:50
Subject Sample #1	0.123	23:50
Breath Volume	2.371	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:52
Subject Sample #2	0.132	23:53
Breath Volume	3.601	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 12/10/2007 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001284

Instrument Registered To: GILCHRIST COUNTY SO

Software Version: 8100.27

Breath Test Operator: STALVEY, STEPHEN, W

Breath Test Operator Agency: Gilchrist County SO

Arresting Officer: STALVEY, STEPHEN

Arresting Officer Agency: Gilchrist County SO

Last Agency Inspection: 11/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 22:03

Gender: Male

Observation Period Began: 23:10

Subject's Last Name: LANUM

Date/Time of Test: 12/02/2007 00:04

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Subject Sample #1	0.139	00:00
Breath Volume	1.710	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample #2	0.130	00:02
Breath Volume	1.488	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 12/10/2007 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001284

Instrument Registered To: GILCHRIST COUNTY SO

Software Version: 8100.27

Breath Test Operator: REEVES, DAVID, C

Breath Test Operator Agency: Gilchrist County SO

Arresting Officer: REEVES, DAVID

Arresting Officer Agency: Gilchrist County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 02:32

Gender: Male

Observation Period Began: 02:50

Subject's Last Name: OWENS

Date/Time of Test: 12/23/2007 03:26

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.082	03:24	
Air Blank	0.000	03:24	
Subject Sample #1		03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.081	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 01/26/2008 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001284

Instrument Registered To: GILCHRIST COUNTY SO

Software Version: 8100.27

Breath Test Operator: STALVEY, STEPHEN, W

Breath Test Operator Agency: Gilchrist County SO

Arresting Officer: DASHER, WAYNE

Arresting Officer Agency: Gilchrist County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 19:02

Gender: Male

Observation Period Began: 19:50

Subject's Last Name: HUGHES

Date/Time of Test: 12/24/2007 20:17

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.081	20:11
Air Blank	0.000	20:12
Subject Sample #1	0.184	20:12
Breath Volume	1.531	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample #2	0.179	20:15
Breath Volume	1.730	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 01/26/2008 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001286

Instrument Registered To: BAKER COUNTY SO

Software Version: 8100.27

Breath Test Operator: CONOVER, RICHARD, T

Breath Test Operator Agency: FHP Troop G

Arresting Officer: CONOVER, RICHARD

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 02:51

Gender: Male

Observation Period Began: 03:10

Subject's Last Name: KYER

Date/Time of Test: 12/12/2007 03:37

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.078	03:31
Air Blank	0.000	03:31
Subject Sample #1	0.195	03:32
Breath Volume	3.195	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.187	03:35
Breath Volume	3.792	03:35
Air Blank	0.000	03:36
Control Test	0.077	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time: 01/13/2008 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001286

Instrument Registered To: BAKER COUNTY SO

Software Version: 8100.27

Breath Test Operator: CONOVER, RICHARD, T

Breath Test Operator Agency: FHP Troop G

Arresting Officer: CONOVER, RICHARD

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 03:41

Gender: Male

Observation Period Began: 04:15

Subject's Last Name: BENNETT

Date/Time of Test: 12/13/2007 04:47

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:41
Subject Sample #1	0.184	04:42
Breath Volume	2.199	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample #2	0.183	04:45
Breath Volume	1.582	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 01/13/2008 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001286

Instrument Registered To: BAKER COUNTY SO

Software Version: 8100.27

Breath Test Operator: CICHOSKI, HENRY, J

Breath Test Operator Agency: FHP Troop G

Arresting Officer: SWEAT, GAVIN

Arresting Officer Agency: Baker County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 07:30

Gender: Male

Observation Period Began: 07:50

Subject's Last Name: BANNON II

Date/Time of Test: 12/15/2007 08:13

Diagnostics Check	OK	08:10	
Air Blank	0.000	08:10	
Control Test	0.079	08:11	
Air Blank	0.000	08:11	
Subject Sample #1		08:11	Subject Test Refused
Breath Volume	0.000	08:11	
Air Blank	0.000	08:12	
Control Test	0.078	08:12	
Air Blank	0.000	08:13	
Diagnostics Check	OK	08:13	

Data Download Date/Time: 01/13/2008 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001286

Instrument Registered To: BAKER COUNTY SO

Software Version: 8100.27

Breath Test Operator: THRIFT, CLINTON, D

Breath Test Operator Agency: Baker County SO

Arresting Officer: HATTLE, JEFFREY

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 23:32

Gender: Male

Observation Period Began: 23:45

Subject's Last Name: PREVATT

Date/Time of Test: 12/21/2007 00:12

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.079	00:07
Air Blank	0.000	00:07
Subject Sample #1	0.153	00:08
Breath Volume	2.191	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample #2	0.151	00:11
Breath Volume	1.582	00:11
Air Blank	0.000	00:11
Control Test	0.078	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 01/13/2008 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001286

Instrument Registered To: BAKER COUNTY SO

Software Version: 8100.27

Breath Test Operator: THRIFT, CLINTON, D

Breath Test Operator Agency: Baker County SO

Arresting Officer: LIMBAUGH, WAYNE

Arresting Officer Agency: Baker County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: COURT ORD

Arrest Time: 18:30

Gender: Male

Observation Period Began: 18:30

Subject's Last Name: LEWIS

Date/Time of Test: 12/28/2007 18:57

Diagnostics Check	OK	18:51
Air Blank	0.000	18:52
Control Test	0.078	18:52
Air Blank	0.000	18:52
Subject Sample #1	0.110	18:53
Breath Volume	1.234	18:53
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample #2	0.109	18:56
Breath Volume	1.261	18:56
Air Blank	0.000	18:57
Control Test	0.076	18:57
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 01/13/2008 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001286

Instrument Registered To: BAKER COUNTY SO

Software Version: 8100.27

Breath Test Operator: CICHOSKI, HENRY, J

Breath Test Operator Agency: FHP Troop G

Arresting Officer: RYAN, JEFF

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 23:34

Gender: Male

Observation Period Began: 00:16

Subject's Last Name: SPIVEY

Date/Time of Test: 12/29/2007 00:45

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.077	00:37
Air Blank	0.000	00:37
Subject Sample #1	0.204	00:38
Breath Volume	1.347	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample #2	0.226	00:41
Breath Volume	2.722	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample #3	0.205	00:44
Breath Volume	1.160	00:44
Air Blank	0.000	00:44
Control Test	0.077	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 01/13/2008 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001286

Instrument Registered To: BAKER COUNTY SO

Software Version: 8100.27

Breath Test Operator: CICHOSKI, HENRY, J

Breath Test Operator Agency: FHP Troop G

Arresting Officer: CICHOSKI, HENRY

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:31

Gender: Male

Observation Period Began: 02:07

Subject's Last Name: KIRK

Date/Time of Test: 12/29/2007 02:33

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.078	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.096	02:29
Breath Volume	2.796	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample #2	0.091	02:32
Breath Volume	3.785	02:32
Air Blank	0.000	02:32
Control Test	0.076	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 01/13/2008 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001288

Instrument Registered To: UNION COUNTY SO

Software Version: 8100.27

Breath Test Operator: CONOVER, RICHARD, T

Breath Test Operator Agency: FHP Troop G

Arresting Officer: CONOVER, RICHARD

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 12/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 5159011

Dry Gas Expiry Date: 06/13/2007

Violation Code: DUI

Arrest Time: 01:01

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: SINGLETARY

Date/Time of Test: 12/27/2007 02:09

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Subject Sample #1	0.194	02:04
Breath Volume	3.632	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample #2	0.204	02:07
Breath Volume	3.902	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 03/28/2008 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001290

Instrument Registered To: STARKE PD

Software Version: 8100.27

Breath Test Operator: KING, PAUL, A

Breath Test Operator Agency: Starke PD

Arresting Officer: SAPP, TOMMY

Arresting Officer Agency: Bradford County SO

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 14:12

Gender: Male

Observation Period Began: 15:30

Subject's Last Name: DONLEY

Date/Time of Test: 12/21/2007 16:08

Diagnostics Check	OK	16:01
Air Blank	0.000	16:01
Control Test	0.077	16:02
Air Blank	0.000	16:02
Subject Sample #1	0.033	16:03
Breath Volume	1.636	16:03
Air Blank	0.000	16:04
Air Blank	0.000	16:06
Subject Sample #2	0.029	16:06
Breath Volume	1.226	16:06
Air Blank	0.000	16:07
Control Test	0.077	16:07
Air Blank	0.000	16:08
Diagnostics Check	OK	16:08

Data Download Date/Time: 12/31/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001290

Instrument Registered To: STARKE PD

Software Version: 8100.27

Breath Test Operator: KING, PAUL, A

Breath Test Operator Agency: Starke PD

Arresting Officer: CREWS, R

Arresting Officer Agency: Starke PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 16:27

Gender: Male

Observation Period Began: 16:31

Subject's Last Name: MARR

Date/Time of Test: 12/27/2007 17:08

Diagnostics Check	OK	17:02
Air Blank	0.000	17:02
Control Test	0.079	17:03
Air Blank	0.000	17:03
Subject Sample #1	0.067	17:04
Breath Volume	3.136	17:04
Air Blank	0.000	17:04
Air Blank	0.000	17:06
Subject Sample #2	0.067	17:07
Breath Volume	3.265	17:07
Air Blank	0.000	17:07
Control Test	0.077	17:08
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 12/31/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001290

Instrument Registered To: STARKE PD

Software Version: 8100.27

Breath Test Operator: MURRAY, WILLIAM, A

Breath Test Operator Agency: Starke PD

Arresting Officer: MURRAY, WILLIAM

Arresting Officer Agency: Starke PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 21:35

Gender: Male

Observation Period Began: 22:15

Subject's Last Name: ONEAL

Date/Time of Test: 12/28/2007 22:58

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:50
Air Blank	0.000	22:50
Subject Sample #1	0.196	22:53
Breath Volume	3.121	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample #2	0.199	22:56
Breath Volume	3.691	22:56
Air Blank	0.000	22:57
Control Test	0.078	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:58

Data Download Date/Time: 12/31/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001290

Instrument Registered To: STARKE PD

Software Version: 8100.27

Breath Test Operator: MURRAY, WILLIAM, A

Breath Test Operator Agency: Starke PD

Arresting Officer: DAVIS, MICHELLE

Arresting Officer Agency: Starke PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 02:08

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: SMITH

Date/Time of Test: 12/30/2007 03:31

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.078	03:25
Air Blank	0.000	03:26
Subject Sample #1	0.023	03:26
Breath Volume	3.089	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample #2	0.023	03:29
Breath Volume	3.750	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 12/31/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001291

Instrument Registered To: BRADFORD COUNTY SO

Software Version: 8100.27

Breath Test Operator: BOATRIGHT JR, KENNETH, M

Breath Test Operator Agency: FHP Troop G

Arresting Officer: BOATRIGHT JR, KENNETH

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 11/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7061011

Dry Gas Expiry Date: 03/18/2009

Violation Code: DUI

Arrest Time: 02:05

Gender: Female

Observation Period Began: 02:35

Subject's Last Name: SAUNDERS

Date/Time of Test: 12/01/2007 03:02

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:56
Subject Sample #1	0.157	02:57
Breath Volume	3.074	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample #2	0.151	03:00
Breath Volume	1.640	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:02

Data Download Date/Time: 12/06/2007 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001291

Instrument Registered To: BRADFORD COUNTY SO

Software Version: 8100.27

Breath Test Operator: TAYLOR, DIANE, P

Breath Test Operator Agency: Bradford County SO

Arresting Officer: TAYLOR, DIANE

Arresting Officer Agency: Bradford County SO

Last Agency Inspection: 11/13/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 7061011

Dry Gas Expiry Date: 03/18/2009

Violation Code: ADMIN

Arrest Time: 18:30

Gender: Male

Observation Period Began: 18:30

Subject's Last Name: NETTLES

Date/Time of Test: 12/01/2007 19:00

Diagnostics Check	OK	18:54	
Air Blank	0.000	18:55	
Control Test	0.080	18:55	
Air Blank	0.000	18:55	
Subject Sample #1	0.000	18:56	
Breath Volume	2.988	18:56	
Air Blank	0.000	18:56	
Air Blank	0.000	18:58	
Subject Sample #2		18:59	Subject Test Refused
Breath Volume	0.000	18:59	
Air Blank	0.000	18:59	
Control Test	0.080	18:59	
Air Blank	0.000	19:00	
Diagnostics Check	OK	19:00	

Data Download Date/Time: 12/06/2007 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001291

Instrument Registered To: BRADFORD COUNTY SO

Software Version: 8100.27

Breath Test Operator: TAYLOR, DIANE, P

Breath Test Operator Agency: Bradford County SO

Arresting Officer: FORD, JASON

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 11/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7061011

Dry Gas Expiry Date: 03/18/2009

Violation Code: DUI

Arrest Time: 18:23

Gender: Female

Observation Period Began: 19:00

Subject's Last Name: DINKINS

Date/Time of Test: 12/01/2007 19:26

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.079	19:20
Air Blank	0.000	19:21
Subject Sample #1	0.144	19:21
Breath Volume	1.238	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample #2	0.128	19:24
Breath Volume	1.242	19:24
Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 12/06/2007 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001291

Instrument Registered To: BRADFORD COUNTY SO

Software Version: 8100.27

Breath Test Operator: BARRY, JASON, W

Breath Test Operator Agency: FHP Troop G

Arresting Officer: ETCHER, ERIC

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7061011

Dry Gas Expiry Date: 03/18/2009

Violation Code: DUI

Arrest Time: 17:17

Gender: Male

Observation Period Began: 17:42

Subject's Last Name: DRAKE

Date/Time of Test: 12/11/2007 18:13

Diagnostics Check	OK	18:03	
Air Blank	0.000	18:04	
Control Test	0.079	18:04	
Air Blank	0.000	18:05	
Subject Sample #1		18:08	No Sample Provided
Breath Volume	0.000	18:08	
Air Blank	0.000	18:08	
Air Blank	0.000	18:10	
Subject Sample #2		18:12	Sequence Aborted
Breath Volume	0.000	18:12	
Air Blank	0.000	18:13	

Data Download Date/Time: 01/22/2008 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001291

Instrument Registered To: BRADFORD COUNTY SO

Software Version: 8100.27

Breath Test Operator: BARRY, JASON, W

Breath Test Operator Agency: FHP Troop G

Arresting Officer: ERVIN, KYLE

Arresting Officer Agency: Hampton PD

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7061011

Dry Gas Expiry Date: 03/18/2009

Violation Code: DUI

Arrest Time: 18:22

Gender: Male

Observation Period Began: 19:24

Subject's Last Name: RODRIGUEZ

Date/Time of Test: 12/21/2007 19:51

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:46
Subject Sample #1	0.137	19:47
Breath Volume	1.984	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:49
Subject Sample #2	0.130	19:50
Breath Volume	1.695	19:50
Air Blank	0.000	19:50
Control Test	0.079	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 01/22/2008 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001291

Instrument Registered To: BRADFORD COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROWN, DAVID, T

Breath Test Operator Agency: Bradford County SO

Arresting Officer: LOCKHART, CLINT

Arresting Officer Agency: Starke PD

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7061011

Dry Gas Expiry Date: 03/18/2009

Violation Code: DUI

Arrest Time: 19:23

Gender: Male

Observation Period Began: 19:44

Subject's Last Name: BAMFORD

Date/Time of Test: 12/31/2007 20:15

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.079	20:09
Air Blank	0.000	20:09
Subject Sample #1	0.347	20:10
Breath Volume	4.156	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:12
Subject Sample #2	0.365	20:13
Breath Volume	4.433	20:13
Air Blank	0.000	20:14
Control Test	0.078	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 01/22/2008 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001302

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: HOLLEY, ROGER, D

Breath Test Operator Agency: Milton PD

Arresting Officer: SMITH, DWAYNE

Arresting Officer Agency: Milton PD

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 05:20

Gender: Male

Observation Period Began: 06:08

Subject's Last Name: PITTMAN

Date/Time of Test: 12/23/2007 06:47

Diagnostics Check	OK	06:41
Air Blank	0.000	06:42
Control Test	0.079	06:42
Air Blank	0.000	06:42
Subject Sample #1	0.107	06:43
Breath Volume	2.687	06:43
Air Blank	0.000	06:44
Air Blank	0.000	06:45
Subject Sample #2	0.101	06:46
Breath Volume	2.246	06:46
Air Blank	0.000	06:47
Control Test	0.079	06:47
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 01/13/2008 01:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: ROSENSWEIG, ERIC,

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: ROSENSWEIG, ERIC

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 11/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 00:13

Gender: Male

Observation Period Began: 00:39

Subject's Last Name: GALLOWAY

Date/Time of Test: 12/01/2007 01:12

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample #1	0.138	01:04
Breath Volume	1.175	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample #2	0.163	01:07
Breath Volume	3.441	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample #3	0.158	01:10
Breath Volume	2.960	01:10
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 12/08/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARRIS, JAMES, T

Breath Test Operator Agency: FHP Troop A

Arresting Officer: HARRIS, JAMES

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 05:23

Gender: Male

Observation Period Began: 06:15

Subject's Last Name: WILLIAMS

Date/Time of Test: 12/01/2007 06:54

Diagnostics Check	OK	06:48
Air Blank	0.000	06:48
Control Test	0.080	06:48
Air Blank	0.000	06:49
Subject Sample #1	0.140	06:50
Breath Volume	2.562	06:50
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample #2	0.124	06:53
Breath Volume	1.699	06:53
Air Blank	0.000	06:53
Control Test	0.080	06:53
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

Data Download Date/Time: 12/08/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: SWINDELL, SHAWN, T

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: SHEEHAN, TERI

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 11/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:53

Gender: Female

Observation Period Began: 01:05

Subject's Last Name: OLIVER

Date/Time of Test: 12/02/2007 01:28

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.079	01:26	
Air Blank	0.000	01:26	
Subject Sample #1		01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.079	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 12/08/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: STALCUP, TIMOTHY, C

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: STALCUP, TIMOTHY

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 11/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 20:53

Gender: Female

Observation Period Began: 21:19

Subject's Last Name: JOHNSON

Date/Time of Test: 12/03/2007 21:46

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.079	21:40
Air Blank	0.000	21:40
Subject Sample #1	0.252	21:41
Breath Volume	1.394	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:43
Subject Sample #2	0.254	21:44
Breath Volume	1.621	21:44
Air Blank	0.000	21:45
Control Test	0.079	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 12/08/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VINCENT, ED, S

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: VINCENT, ED

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 11/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 23:07

Gender: Male

Observation Period Began: 23:07

Subject's Last Name: ADAMS

Date/Time of Test: 12/04/2007 23:42

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample #1	0.170	23:37
Breath Volume	5.132	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample #2	0.152	23:40
Breath Volume	4.078	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:42

Data Download Date/Time: 12/08/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VINCENT, ED, S

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: VINCENT, ED

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 11/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 02:51

Gender: Male

Observation Period Began: 02:51

Subject's Last Name: HICKS

Date/Time of Test: 12/05/2007 03:27

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:26	
Subject Sample #1		03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.080	03:27	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 12/08/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, TIMOTHY, A

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: RODRIGUEZ, TIMOTHY

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 11/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 22:33

Gender: Male

Observation Period Began: 23:31

Subject's Last Name: CHIKOS

Date/Time of Test: 12/06/2007 00:07

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.080	23:52	
Air Blank	0.000	23:53	
Subject Sample #1	0.173	23:55	No .020 agreement
Breath Volume	1.292	23:55	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample #2	0.168	00:01	Volume Not Met
Breath Volume	0.152	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:03	
Subject Sample #3	0.151	00:05	No .020 agreement
Breath Volume	1.921	00:05	
Air Blank	0.000	00:06	
Control Test	0.076	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 12/08/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, TIMOTHY, A

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: RODRIGUEZ, TIMOTHY

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 11/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 22:33

Gender: Male

Observation Period Began: 00:35

Subject's Last Name: CHIKOS

Date/Time of Test: 12/06/2007 01:02

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:56	
Control Test	0.079	00:56	
Air Blank	0.000	00:57	
Subject Sample #1	0.137	00:58	
Breath Volume	1.925	00:58	
Air Blank	0.000	00:59	
Air Blank	0.000	01:00	
Subject Sample #2		01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.077	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 12/08/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: REAVES, WILLIAM, T

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: REAVES, WILLIAM

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 11/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 01:15

Gender: Female

Observation Period Began: 01:27

Subject's Last Name: KENWORTHY

Date/Time of Test: 12/06/2007 02:02

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:49	
Control Test	0.079	01:49	
Air Blank	0.000	01:49	
Subject Sample #1	0.115	01:52	Volume Not Met
Breath Volume	0.140	01:52	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample #2	0.117	01:57	
Breath Volume	2.335	01:57	
Air Blank	0.000	01:58	
Air Blank	0.000	02:00	
Subject Sample #3	0.115	02:01	
Breath Volume	1.421	02:01	
Air Blank	0.000	02:02	
Control Test	0.078	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 12/08/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VINCENT, ED, S

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: VINCENT, ED

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 11/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 04:55

Gender: Male

Observation Period Began: 05:03

Subject's Last Name: LIVESAY

Date/Time of Test: 12/07/2007 05:30

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Subject Sample #1	0.135	05:25
Breath Volume	3.171	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample #2	0.148	05:28
Breath Volume	3.441	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 12/08/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: ALLEN, BENJAMIN, M

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: VEYSEY, BRIAN

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 00:17

Gender: Male

Observation Period Began: 00:30

Subject's Last Name: WASHINGTON

Date/Time of Test: 12/09/2007 00:58

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:52
Subject Sample #1	0.203	00:53
Breath Volume	2.078	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample #2	0.207	00:56
Breath Volume	1.550	00:56
Air Blank	0.000	00:57
Control Test	0.078	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:58

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: RAY, MATTHEW, B

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: RAY, MATTHEW

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 00:46

Gender: Male

Observation Period Began: 01:37

Subject's Last Name: SEKUL

Date/Time of Test: 12/09/2007 02:06

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.079	01:58	
Air Blank	0.000	01:58	
Subject Sample #1		02:01	No Sample Provided
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample #2		02:06	RFI Detect
Breath Volume	0.000	02:06	
Air Blank		02:06	RFI Detect
Air Blank	0.000	02:06	

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: HENDRIX, TIMOTHY, D

Breath Test Operator Agency: FHP Troop A

Arresting Officer: HENDRIX, TIMOTHY

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 03:54

Gender: Male

Observation Period Began: 03:15

Subject's Last Name: SIMPSON

Date/Time of Test: 12/09/2007 04:44

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.080	04:39
Air Blank	0.000	04:39
Subject Sample #1	0.079	04:40
Breath Volume	2.703	04:40
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample #2	0.078	04:43
Breath Volume	3.660	04:43
Air Blank	0.000	04:43
Control Test	0.079	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: CAMPBELL, HENDERSOM, S

Breath Test Operator Agency: FHP Troop A

Arresting Officer: CAMPBELL, HENDERSON

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 17:46

Gender: Female

Observation Period Began: 18:45

Subject's Last Name: CRANE

Date/Time of Test: 12/11/2007 19:26

Diagnostics Check	OK	19:16	
Air Blank	0.000	19:16	
Control Test	0.078	19:16	
Air Blank	0.000	19:17	
Subject Sample #1	0.245	19:20	Volume Not Met
Breath Volume	0.214	19:20	
Air Blank	0.000	19:21	
Air Blank	0.000	19:22	
Subject Sample #2		19:25	Sequence Aborted
Breath Volume	1.085	19:25	
Air Blank	0.000	19:26	

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BURNS, KYLE, J

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: HARRISON, BRADLEY

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:02

Gender: Female

Observation Period Began: 00:38

Subject's Last Name: ADAMS

Date/Time of Test: 12/12/2007 01:12

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.078	00:59	
Air Blank	0.000	00:59	
Subject Sample #1	0.131	01:01	
Breath Volume	1.332	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample #2		01:07	No Sample Provided
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample #3	0.128	01:10	
Breath Volume	1.535	01:10	
Air Blank	0.000	01:11	
Control Test	0.078	01:11	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEWIS, ANNA, L

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: LEWIS, ANNA

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 20:25

Gender: Female

Observation Period Began: 20:55

Subject's Last Name: THELE

Date/Time of Test: 12/14/2007 21:25

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.079	21:18
Air Blank	0.000	21:19
Subject Sample #1	0.182	21:20
Breath Volume	2.890	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample #2	0.196	21:24
Breath Volume	1.570	21:24
Air Blank	0.000	21:25
Control Test	0.077	21:25
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: NELSON, DURON, A

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: ENTERKIN, WAYNE

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 22:40

Gender: Male

Observation Period Began: 23:18

Subject's Last Name: MCARTHUR

Date/Time of Test: 12/14/2007 23:45

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:39
Subject Sample #1	0.098	23:40
Breath Volume	2.285	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:42
Subject Sample #2	0.106	23:43
Breath Volume	2.945	23:43
Air Blank	0.000	23:44
Control Test	0.078	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VINCENT, ED, S

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: VEYSEY, BRIAN

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 00:39

Gender: Male

Observation Period Began: 00:39

Subject's Last Name: FISHER

Date/Time of Test: 12/16/2007 01:09

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Subject Sample #1		01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Control Test	0.079	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VINCENT, ED, S

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: VINCENT, ED

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 22:50

Gender: Male

Observation Period Began: 23:20

Subject's Last Name: BARR

Date/Time of Test: 12/16/2007 23:48

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.081	23:46	
Air Blank	0.000	23:46	
Subject Sample #1		23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.081	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: SWINDELL, SHAWN, T

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: SWINDELL, SHAWN

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:27

Gender: Male

Observation Period Began: 00:40

Subject's Last Name: PUGH

Date/Time of Test: 12/19/2007 01:07

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample #1	0.072	01:02
Breath Volume	1.308	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample #2	0.076	01:06
Breath Volume	1.507	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: STALCUP, TIMOTHY, C

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: STONE, ANDREW

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 15:34

Gender: Male

Observation Period Began: 16:20

Subject's Last Name: SMALLWOOD

Date/Time of Test: 12/19/2007 16:48

Diagnostics Check	OK	16:42
Air Blank	0.000	16:42
Control Test	0.080	16:43
Air Blank	0.000	16:43
Subject Sample #1	0.000	16:44
Breath Volume	2.804	16:44
Air Blank	0.000	16:44
Air Blank	0.000	16:46
Subject Sample #2	0.000	16:47
Breath Volume	2.117	16:47
Air Blank	0.000	16:47
Control Test	0.078	16:47
Air Blank	0.000	16:48
Diagnostics Check	OK	16:48

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: PHILLIPS, SAMUEL, S

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: PHILLIPS, SCOTT

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:44

Gender: Male

Observation Period Began: 00:04

Subject's Last Name: WHITE

Date/Time of Test: 12/21/2007 00:29

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.079	00:27	
Air Blank	0.000	00:27	
Subject Sample #1		00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: STALCUP, TIMOTHY, C

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: STALCUP, TIMOTHY

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 21:17

Gender: Male

Observation Period Began: 22:00

Subject's Last Name: WATERS

Date/Time of Test: 12/21/2007 22:35

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.079	22:21	
Air Blank	0.000	22:22	
Subject Sample #1	0.250	22:25	Volume Not Met
Breath Volume	0.808	22:25	
Air Blank	0.000	22:26	
Air Blank	0.000	22:27	
Subject Sample #2	0.243	22:30	
Breath Volume	1.578	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:33	
Subject Sample #3	0.246	22:33	
Breath Volume	1.937	22:33	
Air Blank	0.000	22:34	
Control Test	0.075	22:34	
Air Blank	0.000	22:35	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: MILLER, JAMES, L

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: MILLER, JAMES

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 22:09

Gender: Female

Observation Period Began: 22:57

Subject's Last Name: OWENS

Date/Time of Test: 12/21/2007 23:28

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Subject Sample #1	0.000	23:24
Breath Volume	2.167	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample #2	0.000	23:27
Breath Volume	2.812	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VEYSEY, BRIAN, M

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: VEYSEY, BRIAN

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: ADMIN

Arrest Time: 22:27

Gender: Male

Observation Period Began: 22:45

Subject's Last Name: JOHNSON

Date/Time of Test: 12/21/2007 23:52

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Subject Sample #1	0.118	23:48
Breath Volume	1.828	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample #2	0.125	23:51
Breath Volume	2.226	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, TIMOTHY, A

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: RODRIGUEZ, TIMOTHY

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 00:18

Gender: Male

Observation Period Began: 00:50

Subject's Last Name: KIRKLAND

Date/Time of Test: 12/22/2007 01:17

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.079	01:11
Air Blank	0.000	01:12
Subject Sample #1	0.139	01:12
Breath Volume	3.722	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample #2	0.137	01:16
Breath Volume	3.781	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEWIS, ANNA, L

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: LEWIS, ANNA

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 23:22

Gender: Male

Observation Period Began: 00:03

Subject's Last Name: WHITAKER

Date/Time of Test: 12/23/2007 00:30

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample #1	0.000	00:25
Breath Volume	2.562	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample #2	0.000	00:28
Breath Volume	2.109	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:30

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: DYKES, CLIFTON, R

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: DYKES, CLIFTON

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 08:09

Gender: Male

Observation Period Began: 08:15

Subject's Last Name: SANDERS

Date/Time of Test: 12/23/2007 08:54

Diagnostics Check	OK	08:47
Air Blank	0.000	08:47
Control Test	0.079	08:48
Air Blank	0.000	08:48
Subject Sample #1	0.205	08:49
Breath Volume	3.125	08:49
Air Blank	0.000	08:50
Air Blank	0.000	08:52
Subject Sample #2	0.199	08:52
Breath Volume	2.921	08:52
Air Blank	0.000	08:53
Control Test	0.077	08:53
Air Blank	0.000	08:54
Diagnostics Check	OK	08:54

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VINCENT, ED, S

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: VINCENT, ED

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: OTHER

Arrest Time: 00:20

Gender: Male

Observation Period Began: 00:20

Subject's Last Name: PERKINS

Date/Time of Test: 12/24/2007 00:48

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample #1	0.097	00:43
Breath Volume	4.296	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample #2	0.096	00:46
Breath Volume	4.648	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: STALCUP, TIMOTHY, C

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: YOUNGHANSE, JOHN

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 18:47

Gender: Female

Observation Period Began: 19:43

Subject's Last Name: YATES

Date/Time of Test: 12/25/2007 20:19

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:04	
Control Test	0.080	20:04	
Air Blank	0.000	20:04	
Subject Sample #1	0.141	20:08	Volume Not Met
Breath Volume	0.527	20:08	
Air Blank	0.000	20:08	
Air Blank	0.000	20:10	
Subject Sample #2	0.140	20:12	
Breath Volume	2.382	20:12	
Air Blank	0.000	20:13	
Air Blank	0.000	20:14	
Subject Sample #3	0.120	20:17	Volume Not Met
Breath Volume	0.687	20:17	
Air Blank	0.000	20:18	
Control Test	0.076	20:18	
Air Blank	0.000	20:19	
Diagnostics Check	OK	20:19	

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VINCENT, ED, S

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: VINCENT, ED

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 00:25

Gender: Female

Observation Period Began: 00:30

Subject's Last Name: HICKS

Date/Time of Test: 12/30/2007 01:06

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample #1	0.000	01:02
Breath Volume	2.312	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample #2	0.000	01:05
Breath Volume	2.449	01:05
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: DYKES, CLIFTON, R

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: DYKES, CLIFTON

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 12:23

Gender: Male

Observation Period Began: 14:23

Subject's Last Name: TORRES

Date/Time of Test: 12/30/2007 15:07

Diagnostics Check	OK	15:05	
Air Blank	0.000	15:06	
Control Test	0.000	15:06	RFI Detect
Air Blank	0.000	15:07	

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001303

Instrument Registered To: SANTA ROSA COUNTY SO

Software Version: 8100.27

Breath Test Operator: DYKES, CLIFTON, R

Breath Test Operator Agency: Santa Rosa County SO

Arresting Officer: DYKES, CLIFTON

Arresting Officer Agency: Santa Rosa County SO

Last Agency Inspection: 12/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 14:23

Gender: Male

Observation Period Began: 14:23

Subject's Last Name: TORRES

Date/Time of Test: 12/30/2007 15:19

Diagnostics Check	OK	15:13
Air Blank	0.000	15:13
Control Test	0.079	15:14
Air Blank	0.000	15:14
Subject Sample #1	0.183	15:15
Breath Volume	2.082	15:15
Air Blank	0.000	15:16
Air Blank	0.000	15:17
Subject Sample #2	0.184	15:18
Breath Volume	1.707	15:18
Air Blank	0.000	15:19
Control Test	0.078	15:19
Air Blank	0.000	15:19
Diagnostics Check	OK	15:19

Data Download Date/Time: 01/12/2008 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001305

Instrument Registered To: MARIANNA PD

Software Version: 8100.27

Breath Test Operator: DAVIS, FRANCIS, G

Breath Test Operator Agency: Marianna PD

Arresting Officer: BURKS, ROB

Arresting Officer Agency: Jackson County SO

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 03:15

Gender: Female

Observation Period Began: 04:15

Subject's Last Name: BROWN

Date/Time of Test: 12/12/2007 04:37

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.079	04:36	
Air Blank		04:37	RFI Detect
Air Blank	0.000	04:37	

Data Download Date/Time: 12/22/2007 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001305

Instrument Registered To: MARIANNA PD

Software Version: 8100.27

Breath Test Operator: DAVIS, FRANCIS, G

Breath Test Operator Agency: Marianna PD

Arresting Officer: BURKS, ROB

Arresting Officer Agency: Jackson County SO

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 03:15

Gender: Female

Observation Period Began: 04:15

Subject's Last Name: BROWN

Date/Time of Test: 12/12/2007 04:53

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.079	04:42	
Air Blank	0.000	04:42	
Subject Sample #1	0.105	04:45	Volume Not Met
Breath Volume	0.484	04:45	
Air Blank	0.000	04:46	
Air Blank	0.000	04:48	
Subject Sample #2	0.111	04:51	Volume Not Met
Breath Volume	0.980	04:51	
Air Blank	0.000	04:52	
Control Test	0.079	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 12/22/2007 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001305

Instrument Registered To: MARIANNA PD

Software Version: 8100.27

Breath Test Operator: JOHNSON, JAMES, D

Breath Test Operator Agency: FHP Troop A

Arresting Officer: JOHNSON, JAMES

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 20:50

Gender: Male

Observation Period Began: 21:15

Subject's Last Name: FAILS

Date/Time of Test: 12/12/2007 22:03

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.078	21:56
Air Blank	0.000	21:57
Subject Sample #1	0.055	21:58
Breath Volume	1.644	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:00
Subject Sample #2	0.059	22:01
Breath Volume	3.082	22:01
Air Blank	0.000	22:02
Control Test	0.076	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:03

Data Download Date/Time: 12/22/2007 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001305

Instrument Registered To: MARIANNA PD

Software Version: 8100.27

Breath Test Operator: FRENCH, J, K

Breath Test Operator Agency: FHP Troop A

Arresting Officer: FRENCH, J

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 11:03

Gender: Male

Observation Period Began: 11:03

Subject's Last Name: ZAMISKA

Date/Time of Test: 12/15/2007 11:41

Diagnostics Check	OK	11:35
Air Blank	0.000	11:35
Control Test	0.080	11:36
Air Blank	0.000	11:36
Subject Sample #1	0.069	11:37
Breath Volume	2.398	11:37
Air Blank	0.000	11:37
Air Blank	0.000	11:39
Subject Sample #2	0.070	11:40
Breath Volume	3.332	11:40
Air Blank	0.000	11:40
Control Test	0.079	11:40
Air Blank	0.000	11:41
Diagnostics Check	OK	11:41

Data Download Date/Time: 12/22/2007 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001305

Instrument Registered To: MARIANNA PD

Software Version: 8100.27

Breath Test Operator: DAVIS, FRANCIS, G

Breath Test Operator Agency: Marianna PD

Arresting Officer: SMITH, JOHN

Arresting Officer Agency: Marianna PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 00:01

Gender: Male

Observation Period Began: 00:25

Subject's Last Name: COX

Date/Time of Test: 12/16/2007 01:32

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:27
Subject Sample #1	0.160	01:27
Breath Volume	2.914	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample #2	0.167	01:30
Breath Volume	3.792	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 12/22/2007 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001305

Instrument Registered To: MARIANNA PD

Software Version: 8100.27

Breath Test Operator: JOHNSON, J., D

Breath Test Operator Agency: FHP Troop A

Arresting Officer: JOHNSON, J

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 12/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 11:45

Gender: Male

Observation Period Began: 11:45

Subject's Last Name: HENDRICKS

Date/Time of Test: 12/22/2007 12:12

Diagnostics Check	OK	12:06
Air Blank	0.000	12:06
Control Test	0.079	12:06
Air Blank	0.000	12:07
Subject Sample #1	0.011	12:08
Breath Volume	3.542	12:08
Air Blank	0.000	12:08
Air Blank	0.000	12:10
Subject Sample #2	0.009	12:11
Breath Volume	3.351	12:11
Air Blank	0.000	12:11
Control Test	0.078	12:11
Air Blank	0.000	12:12
Diagnostics Check	OK	12:12

Data Download Date/Time: 01/20/2008 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001305

Instrument Registered To: MARIANNA PD

Software Version: 8100.27

Breath Test Operator: PYLE JR, HARRY, R

Breath Test Operator Agency: FHP Troop A

Arresting Officer: VENNELL, JOHN

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 12/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 21:18

Gender: Male

Observation Period Began: 21:05

Subject's Last Name: MARTIN

Date/Time of Test: 12/24/2007 22:34

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Subject Sample #1	0.163	22:30
Breath Volume	3.031	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample #2	0.160	22:33
Breath Volume	1.628	22:33
Air Blank	0.000	22:33
Control Test	0.078	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 01/20/2008 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001305

Instrument Registered To: MARIANNA PD

Software Version: 8100.27

Breath Test Operator: BEASLEY, FRED, W

Breath Test Operator Agency: Marianna PD

Arresting Officer: DANIELS, GARY

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 12/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 16:52

Gender: Male

Observation Period Began: 17:00

Subject's Last Name: MARTIN

Date/Time of Test: 12/28/2007 17:40

Diagnostics Check	OK	17:33
Air Blank	0.000	17:34
Control Test	0.078	17:34
Air Blank	0.000	17:35
Subject Sample #1	0.252	17:35
Breath Volume	2.253	17:35
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample #2	0.252	17:38
Breath Volume	2.613	17:38
Air Blank	0.000	17:39
Control Test	0.077	17:39
Air Blank	0.000	17:39
Diagnostics Check	OK	17:40

Data Download Date/Time: 01/20/2008 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001305

Instrument Registered To: MARIANNA PD

Software Version: 8100.27

Breath Test Operator: JOHNSON, J, D

Breath Test Operator Agency: FHP Troop A

Arresting Officer: DANIELS, GARY

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 12/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 10:39

Gender: Male

Observation Period Began: 10:45

Subject's Last Name: PARKER

Date/Time of Test: 12/29/2007 11:24

Diagnostics Check	OK	11:18
Air Blank	0.000	11:18
Control Test	0.078	11:19
Air Blank	0.000	11:19
Subject Sample #1	0.280	11:20
Breath Volume	3.082	11:20
Air Blank	0.000	11:21
Air Blank	0.000	11:22
Subject Sample #2	0.266	11:23
Breath Volume	1.636	11:23
Air Blank	0.000	11:23
Control Test	0.078	11:24
Air Blank	0.000	11:24
Diagnostics Check	OK	11:24

Data Download Date/Time: 01/20/2008 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: SAILOR, TRAVIS, C

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: SAILOR, TRAVIS

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 21:57

Gender: Male

Observation Period Began: 22:18

Subject's Last Name: SPARKMAN

Date/Time of Test: 12/01/2007 22:49

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:47	
Control Test	0.079	22:47	
Air Blank	0.000	22:48	
Subject Sample #1		22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:48	
Control Test	0.078	22:48	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 12/19/2007 19:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: BOYD, ERIC, L

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: BOYD, ERIC

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 22:16

Gender: Male

Observation Period Began: 22:55

Subject's Last Name: PAYNE

Date/Time of Test: 12/12/2007 23:29

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.078	23:16	
Air Blank	0.000	23:17	
Subject Sample #1	0.211	23:18	
Breath Volume	1.972	23:18	
Air Blank	0.000	23:19	
Air Blank	0.000	23:20	
Subject Sample #2	0.158	23:23	Volume Not Met
Breath Volume	0.421	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:26	
Subject Sample #3		23:28	Subject Test Refused
Breath Volume	0.031	23:28	
Air Blank	0.000	23:28	
Control Test	0.073	23:29	Control Outside Tolerance
Air Blank	0.000	23:29	

Data Download Date/Time: 12/19/2007 19:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: HOUGH, ROBERT, F

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: HOUGH, ROBERT

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 00:11

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: COOK

Date/Time of Test: 12/16/2007 01:28

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample #1	0.204	01:23
Breath Volume	1.976	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample #2	0.204	01:26
Breath Volume	1.656	01:26
Air Blank	0.000	01:27
Control Test	0.077	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 12/19/2007 19:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: WOLL, MICHAEL,

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: WOLL, MICHAEL

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 03:24

Gender: Male

Observation Period Began: 03:55

Subject's Last Name: VELASQUEZ-CHAVEZ

Date/Time of Test: 12/16/2007 04:26

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Subject Sample #1	0.196	04:21
Breath Volume	2.535	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample #2	0.195	04:24
Breath Volume	1.859	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 12/19/2007 19:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: FINDLEY, BRANDEN, J

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: FINDLEY, BRANDEN

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 01:54

Gender: Male

Observation Period Began: 02:15

Subject's Last Name: PINTOR

Date/Time of Test: 12/17/2007 02:42

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample #1	0.138	02:37
Breath Volume	3.472	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.135	02:40
Breath Volume	2.449	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 12/19/2007 19:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: HOUGH, ROBERT, F

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: PEREIRA, JOSE

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 02:35

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: WALLACE

Date/Time of Test: 12/17/2007 03:30

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample #1	0.255	03:25
Breath Volume	1.117	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample #2	0.263	03:28
Breath Volume	2.703	03:28
Air Blank	0.000	03:29
Control Test	0.077	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 12/19/2007 19:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: HOUGH, ROBERT, F

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: PEREIRA, JOSE

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 02:56

Gender: Male

Observation Period Began: 03:20

Subject's Last Name: LUFRANO

Date/Time of Test: 12/18/2007 03:44

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.080	03:42	
Air Blank	0.000	03:43	
Subject Sample #1		03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.078	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 12/19/2007 19:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: REYNOLDS, EVAN, D

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: REYNOLDS, EVAN

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 12/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 01:26

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: STEPANOV

Date/Time of Test: 12/22/2007 02:53

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Subject Sample #1	0.196	02:45	Volume Not Met
Breath Volume	0.417	02:45	
Air Blank	0.000	02:46	
Air Blank	0.000	02:47	
Subject Sample #2	0.190	02:48	No .020 agreement
Breath Volume	2.218	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample #3	0.168	02:51	No .020 agreement
Breath Volume	2.105	02:51	
Air Blank	0.000	02:52	
Control Test	0.075	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 01/09/2008 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: WOLL, MICHAEL,

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: WOLL, MICHAEL

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 12/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 22:33

Gender: Male

Observation Period Began: 23:16

Subject's Last Name: CLIBURN

Date/Time of Test: 12/25/2007 23:39

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.078	23:37	
Air Blank	0.000	23:38	
Subject Sample #1		23:38	Subject Test Refused
Breath Volume	0.000	23:38	
Air Blank	0.000	23:39	
Control Test	0.077	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 01/09/2008 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: WOLL, MICHAEL,

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: JANKOWSKI, MICHAEL

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 12/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 23:03

Gender: Female

Observation Period Began: 23:40

Subject's Last Name: DAVIS

Date/Time of Test: 12/26/2007 00:13

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Subject Sample #1	0.171	00:09
Breath Volume	2.308	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample #2	0.171	00:12
Breath Volume	1.929	00:12
Air Blank	0.000	00:13
Control Test	0.078	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 01/09/2008 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: SAILOR, TRAVIS, C

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: HOGAN, MICHAEL

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 12/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 19:45

Gender: Male

Observation Period Began: 21:41

Subject's Last Name: CARTER

Date/Time of Test: 12/27/2007 22:08

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.079	22:03
Air Blank	0.000	22:04
Subject Sample #1	0.000	22:04
Breath Volume	2.949	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample #2	0.000	22:07
Breath Volume	3.425	22:07
Air Blank	0.000	22:07
Control Test	0.078	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 01/09/2008 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001306

Instrument Registered To: OKALOOSA SO

Software Version: 8100.27

Breath Test Operator: HOUGH, BRYAN, W

Breath Test Operator Agency: Ft. Walton Beach PD

Arresting Officer: HOUGH, BRYAN

Arresting Officer Agency: Ft. Walton Beach PD

Last Agency Inspection: 12/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 05/01/2008

Violation Code: DUI

Arrest Time: 22:23

Gender: Male

Observation Period Began: 22:38

Subject's Last Name: VICKERS

Date/Time of Test: 12/31/2007 23:07

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.083	23:01
Air Blank	0.000	23:02
Subject Sample #1	0.000	23:02
Breath Volume	2.347	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample #2	0.000	23:06
Breath Volume	2.164	23:06
Air Blank	0.000	23:06
Control Test	0.078	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 01/09/2008 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001307

Instrument Registered To: MADISON COUNTY SO

Software Version: 8100.27

Breath Test Operator: SMITH, JIMMY, L

Breath Test Operator Agency: Madison County SO

Arresting Officer: ALEXANDER, REGGIE

Arresting Officer Agency: Madison PD

Last Agency Inspection: 11/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 00:20

Gender: Male

Observation Period Began: 00:35

Subject's Last Name: BONNER

Date/Time of Test: 12/01/2007 01:16

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Subject Sample #1	0.342	01:12
Breath Volume	2.148	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample #2	0.349	01:14
Breath Volume	2.019	01:14
Air Blank	0.000	01:15
Control Test	0.077	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 12/10/2007 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001307

Instrument Registered To: MADISON COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCDANIEL, JESSE, O

Breath Test Operator Agency: Madison County SO

Arresting Officer: PITTS, DENNIS

Arresting Officer Agency: Madison County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 606002I

Dry Gas Expiry Date: 03/03/2008

Violation Code: DUI

Arrest Time: 00:55

Gender: Male

Observation Period Began: 00:59

Subject's Last Name: WYNN

Date/Time of Test: 12/23/2007 01:33

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.078	01:31	
Air Blank	0.000	01:31	
Subject Sample #1		01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 01/25/2008 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HERRINGTON, ROB, M

Breath Test Operator Agency: Panama City PD

Arresting Officer: HERRINGTON, ROB

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 00:53

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: TRAVASOS

Date/Time of Test: 12/01/2007 01:42

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:36
Subject Sample #1	0.197	01:37
Breath Volume	2.902	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample #2	0.195	01:40
Breath Volume	3.007	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:42

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 02:37

Gender: Male

Observation Period Began: 02:21

Subject's Last Name: JOUNG

Date/Time of Test: 12/01/2007 02:59

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.081	02:58	
Air Blank	0.000	02:58	
Subject Sample #1		02:59	Sequence Aborted
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 02:37

Gender: Male

Observation Period Began: 02:51

Subject's Last Name: JOUNG

Date/Time of Test: 12/01/2007 03:17

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.082	03:11
Air Blank	0.000	03:12
Subject Sample #1	0.112	03:13
Breath Volume	1.261	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample #2	0.114	03:16
Breath Volume	2.132	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 05:04

Gender: Female

Observation Period Began: 05:20

Subject's Last Name: PHILLIPS

Date/Time of Test: 12/01/2007 05:46

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.081	05:41
Air Blank	0.000	05:41
Subject Sample #1	0.000	05:42
Breath Volume	3.039	05:42
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample #2	0.000	05:45
Breath Volume	2.316	05:45
Air Blank	0.000	05:45
Control Test	0.081	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HERRINGTON, ROB, M

Breath Test Operator Agency: Panama City PD

Arresting Officer: HERRINGTON, ROB

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 21:14

Gender: Male

Observation Period Began: 21:35

Subject's Last Name: CORTEZ

Date/Time of Test: 12/01/2007 22:03

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.081	21:56
Air Blank	0.000	21:57
Subject Sample #1	0.149	21:58
Breath Volume	2.550	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:00
Subject Sample #2	0.147	22:01
Breath Volume	2.593	22:01
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 22:58

Gender: Male

Observation Period Began: 23:11

Subject's Last Name: BARA

Date/Time of Test: 12/01/2007 23:37

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.081	23:32
Air Blank	0.000	23:32
Subject Sample #1	0.023	23:32
Breath Volume	2.382	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample #2	0.020	23:35
Breath Volume	1.378	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HERRINGTON, ROB, M

Breath Test Operator Agency: Panama City PD

Arresting Officer: HERRINGTON, ROB

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 00:39

Gender: Male

Observation Period Began: 00:51

Subject's Last Name: SANDERS

Date/Time of Test: 12/02/2007 01:24

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Subject Sample #1		01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample #2		01:22	No Sample Provided
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 02:03

Subject's Last Name: GRYSZOWKA

Date/Time of Test: 12/02/2007 02:30

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Subject Sample #1	0.037	02:25
Breath Volume	2.976	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample #2	0.036	02:28
Breath Volume	2.632	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, STREICHERT

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 04:35

Gender: Male

Observation Period Began: 04:50

Subject's Last Name: MORGAN

Date/Time of Test: 12/02/2007 05:15

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:11	
Control Test	0.080	05:11	
Air Blank	0.000	05:12	
Subject Sample #1		05:13	Subject Test Refused
Breath Volume	0.000	05:13	
Air Blank	0.000	05:14	
Control Test	0.082	05:14	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WAGNER, BRETT, M

Breath Test Operator Agency: FHP Troop A

Arresting Officer: WAGNER, BRETT

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 17:35

Gender: Male

Observation Period Began: 18:30

Subject's Last Name: HILL

Date/Time of Test: 12/02/2007 19:07

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.080	19:01
Air Blank	0.000	19:02
Subject Sample #1	0.141	19:02
Breath Volume	4.335	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample #2	0.139	19:05
Breath Volume	4.582	19:05
Air Blank	0.000	19:06
Control Test	0.080	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WILBURN, BERNIE, J

Breath Test Operator Agency: Panama City PD

Arresting Officer: VANSTRANDER, NICK

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 08:31

Gender: Male

Observation Period Began: 08:45

Subject's Last Name: SPENCER

Date/Time of Test: 12/03/2007 09:12

Diagnostics Check	OK	09:06
Air Blank	0.000	09:06
Control Test	0.081	09:06
Air Blank	0.000	09:07
Subject Sample #1	0.000	09:07
Breath Volume	2.570	09:07
Air Blank	0.000	09:08
Air Blank	0.000	09:10
Subject Sample #2	0.000	09:10
Breath Volume	3.222	09:10
Air Blank	0.000	09:11
Control Test	0.080	09:11
Air Blank	0.000	09:11
Diagnostics Check	OK	09:12

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: BAILEY, RAYMOND, F

Breath Test Operator Agency: FHP Troop A

Arresting Officer: BAILEY, RAYMOND

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 15:30

Gender: Female

Observation Period Began: 16:30

Subject's Last Name: COLE

Date/Time of Test: 12/03/2007 17:04

Diagnostics Check	OK	16:50	
Air Blank	0.000	16:50	
Control Test	0.080	16:51	
Air Blank	0.000	16:51	
Subject Sample #1	0.098	16:52	
Breath Volume	1.570	16:52	
Air Blank	0.000	16:52	
Air Blank	0.000	16:54	
Subject Sample #2	0.085	16:57	Slope Not Level
Breath Volume	1.214	16:57	
Air Blank	0.000	16:58	
Air Blank	0.000	17:00	
Subject Sample #3	0.085	17:03	Slope Not Level
Breath Volume	1.375	17:03	
Air Blank	0.000	17:04	
Control Test	0.080	17:04	
Air Blank	0.000	17:04	
Diagnostics Check	OK	17:04	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: BAILEY, RAYMOND, F

Breath Test Operator Agency: FHP Troop A

Arresting Officer: BAILEY, RAYMOND

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 15:30

Gender: Female

Observation Period Began: 16:30

Subject's Last Name: COLE

Date/Time of Test: 12/03/2007 17:24

Diagnostics Check	OK	17:09	
Air Blank	0.000	17:10	
Control Test	0.081	17:10	
Air Blank	0.000	17:10	
Subject Sample #1	0.086	17:12	
Breath Volume	1.402	17:12	
Air Blank	0.000	17:13	
Air Blank	0.000	17:15	
Subject Sample #2		17:18	No Sample Provided
Breath Volume	0.000	17:18	
Air Blank	0.000	17:18	
Air Blank	0.000	17:20	
Subject Sample #3	0.080	17:23	
Breath Volume	1.367	17:23	
Air Blank	0.000	17:24	
Control Test	0.081	17:24	
Air Blank	0.000	17:24	
Diagnostics Check	OK	17:24	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: RIVERA, NICK,

Breath Test Operator Agency: FHP Troop A

Arresting Officer: RIVERA, NICK

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 22:03

Gender: Female

Observation Period Began: 22:35

Subject's Last Name: SAWICKI

Date/Time of Test: 12/03/2007 23:10

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.080	23:07	
Air Blank	0.000	23:08	
Subject Sample #1		23:09	Subject Test Refused
Breath Volume	0.000	23:09	
Air Blank	0.000	23:09	
Control Test	0.081	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARSEY, WESLEY, H

Breath Test Operator Agency: FHP Troop A

Arresting Officer: HARSEY, WESLEY

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 03:42

Gender: Female

Observation Period Began: 04:20

Subject's Last Name: SIROIS

Date/Time of Test: 12/04/2007 04:58

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:44	
Control Test	0.081	04:44	
Air Blank	0.000	04:45	
Subject Sample #1	0.170	04:48	Volume Not Met
Breath Volume	1.000	04:48	
Air Blank	0.000	04:48	
Air Blank	0.000	04:50	
Subject Sample #2	0.173	04:53	
Breath Volume	1.222	04:53	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample #3	0.185	04:57	
Breath Volume	1.789	04:57	
Air Blank	0.000	04:57	
Control Test	0.077	04:58	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HERRINGTON, ROB, M

Breath Test Operator Agency: Panama City PD

Arresting Officer: HERRINGTON, ROB

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 02:35

Gender: Male

Observation Period Began: 02:47

Subject's Last Name: LATHERY

Date/Time of Test: 12/05/2007 03:22

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.081	03:08	
Air Blank	0.000	03:08	
Subject Sample #1	0.111	03:12	Volume Not Met
Breath Volume	0.054	03:12	
Air Blank	0.000	03:12	
Air Blank	0.000	03:14	
Subject Sample #2	0.117	03:15	
Breath Volume	1.464	03:15	
Air Blank	0.000	03:16	
Air Blank	0.000	03:17	
Subject Sample #3	0.104	03:21	Volume Not Met
Breath Volume	0.015	03:21	
Air Blank	0.000	03:21	
Control Test	0.077	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: SWEET, CHRISTOPHER, W

Breath Test Operator Agency: Bay County SO

Arresting Officer: GRIMES, DENISE

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 17:53

Gender: Male

Observation Period Began: 18:10

Subject's Last Name: AULTMAN

Date/Time of Test: 12/05/2007 18:34

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:32	
Control Test	0.080	18:32	
Air Blank	0.000	18:33	
Subject Sample #1		18:33	Subject Test Refused
Breath Volume	0.000	18:33	
Air Blank	0.000	18:33	
Control Test	0.080	18:33	
Air Blank	0.000	18:34	
Diagnostics Check	OK	18:34	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HERRINGTON, ROB, M

Breath Test Operator Agency: Panama City PD

Arresting Officer: HERRINGTON, ROB

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 04:16

Gender: Male

Observation Period Began: 04:29

Subject's Last Name: ELLIOTT

Date/Time of Test: 12/06/2007 04:57

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.080	04:50
Air Blank	0.000	04:50
Subject Sample #1	0.200	04:52
Breath Volume	2.089	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample #2	0.193	04:55
Breath Volume	1.699	04:55
Air Blank	0.000	04:56
Control Test	0.078	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 00:14

Gender: Male

Observation Period Began: 00:50

Subject's Last Name: NEWELL

Date/Time of Test: 12/07/2007 01:18

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:12
Subject Sample #1	0.119	01:13
Breath Volume	3.312	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample #2	0.114	01:16
Breath Volume	3.742	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 02:45

Gender: Female

Observation Period Began: 03:00

Subject's Last Name: CHAMBERS

Date/Time of Test: 12/07/2007 03:29

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Subject Sample #1	0.061	03:24
Breath Volume	1.425	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:26
Subject Sample #2	0.062	03:28
Breath Volume	1.839	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HERRINGTON, ROB, M

Breath Test Operator Agency: Panama City PD

Arresting Officer: HERRINGTON, ROB

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 04:02

Gender: Male

Observation Period Began: 04:24

Subject's Last Name: SUNDY

Date/Time of Test: 12/07/2007 05:01

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:45	
Control Test	0.081	04:45	
Air Blank	0.000	04:46	
Subject Sample #1	0.160	04:49	Volume Not Met
Breath Volume	0.105	04:49	
Air Blank	0.000	04:50	
Air Blank	0.000	04:52	
Subject Sample #2	0.164	04:55	
Breath Volume	1.582	04:55	
Air Blank	0.000	04:56	
Air Blank	0.000	04:57	
Subject Sample #3	0.153	04:59	
Breath Volume	1.136	04:59	
Air Blank	0.000	05:00	
Control Test	0.072	05:00	Control Outside Tolerance
Air Blank	0.000	05:01	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 04:25

Gender: Female

Observation Period Began: 04:38

Subject's Last Name: YOUNGBLOOD

Date/Time of Test: 12/07/2007 05:09

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.076	05:04
Air Blank	0.000	05:04
Subject Sample #1	0.201	05:05
Breath Volume	1.277	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample #2	0.210	05:08
Breath Volume	1.535	05:08
Air Blank	0.000	05:08
Control Test	0.078	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: BAILEY, RAYMOND, F

Breath Test Operator Agency: FHP Troop A

Arresting Officer: BAILEY, RAYMOND

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 20:11

Gender: Female

Observation Period Began: 20:52

Subject's Last Name: GREEN

Date/Time of Test: 12/07/2007 21:16

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:13	
Control Test	0.080	21:13	
Air Blank	0.000	21:14	
Subject Sample #1		21:14	Subject Test Refused
Breath Volume	0.000	21:14	
Air Blank	0.000	21:15	
Control Test	0.080	21:15	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 01:13

Gender: Male

Observation Period Began: 01:38

Subject's Last Name: WARYWODA

Date/Time of Test: 12/08/2007 02:05

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	01:59
Subject Sample #1	0.150	02:01
Breath Volume	1.378	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample #2	0.153	02:04
Breath Volume	2.437	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARRIS, LISA, M

Breath Test Operator Agency: Bay County SO

Arresting Officer: SUMERALL, J

Arresting Officer Agency: Bay County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 01:48

Gender: Male

Observation Period Began: 02:22

Subject's Last Name: FERGUSON

Date/Time of Test: 12/08/2007 02:44

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Subject Sample #1		02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 03:13

Gender: Male

Observation Period Began: 03:34

Subject's Last Name: TALLEY

Date/Time of Test: 12/08/2007 04:01

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Subject Sample #1	0.166	03:56
Breath Volume	3.343	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample #2	0.161	03:59
Breath Volume	2.761	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTTLES, ANITA, A

Breath Test Operator Agency: Springfield PD

Arresting Officer: KOHLER, KYLE

Arresting Officer Agency: Springfield PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 22:37

Gender: Male

Observation Period Began: 22:50

Subject's Last Name: TRAVASOS

Date/Time of Test: 12/08/2007 23:19

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:15	
Control Test	0.081	23:15	
Air Blank	0.000	23:16	
Subject Sample #1	0.199	23:16	
Breath Volume	1.910	23:16	
Air Blank	0.000	23:17	
Air Blank		23:18	RFI Detect
Air Blank	0.000	23:19	
Control Test	0.081	23:19	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTTLES, ANITA, A

Breath Test Operator Agency: Springfield PD

Arresting Officer: KOHLER, KYLE

Arresting Officer Agency: Springfield PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 22:37

Gender: Male

Observation Period Began: 22:50

Subject's Last Name: TRAVASOS

Date/Time of Test: 12/08/2007 23:28

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Subject Sample #1	0.200	23:23
Breath Volume	2.203	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample #2	0.197	23:26
Breath Volume	1.714	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: TINDLE-HUSAR, TAMMY, S

Breath Test Operator Agency: FHP Troop A

Arresting Officer: TINDLE-HUSAR, TAMMY

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 23:38

Gender: Female

Observation Period Began: 00:32

Subject's Last Name: LASHURE

Date/Time of Test: 12/11/2007 00:59

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.081	00:54
Air Blank	0.000	00:54
Subject Sample #1	0.159	00:55
Breath Volume	2.968	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample #2	0.159	00:58
Breath Volume	3.210	00:58
Air Blank	0.000	00:58
Control Test	0.079	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WELCH, MARY, E

Breath Test Operator Agency: Bay County SO

Arresting Officer: VANSTRANDER, NICK

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 10:51

Gender: Female

Observation Period Began: 11:24

Subject's Last Name: HYDE

Date/Time of Test: 12/11/2007 11:56

Diagnostics Check	OK	11:44	
Air Blank	0.000	11:44	
Control Test	0.080	11:45	
Air Blank	0.000	11:45	
Subject Sample #1	0.272	11:46	
Breath Volume	1.574	11:46	
Air Blank	0.000	11:47	
Air Blank	0.000	11:49	
Subject Sample #2	0.259	11:52	Volume Not Met
Breath Volume	0.300	11:52	
Air Blank	0.000	11:52	
Air Blank	0.000	11:54	
Subject Sample #3	0.268	11:55	
Breath Volume	1.812	11:55	
Air Blank	0.000	11:55	
Control Test	0.078	11:56	
Air Blank	0.000	11:56	
Diagnostics Check	OK	11:56	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 03:03

Gender: Male

Observation Period Began: 03:21

Subject's Last Name: ANDERSON

Date/Time of Test: 12/12/2007 03:48

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43
Subject Sample #1	0.209	03:43
Breath Volume	3.425	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample #2	0.204	03:46
Breath Volume	3.351	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HERRINGTON, ROB, M

Breath Test Operator Agency: Panama City PD

Arresting Officer: HERRINGTON, ROB

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 23:37

Gender: Male

Observation Period Began: 23:58

Subject's Last Name: AGUADO

Date/Time of Test: 12/13/2007 00:30

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.080	00:19	
Air Blank	0.000	00:19	
Subject Sample #1		00:22	No Sample Provided
Breath Volume	0.000	00:22	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample #2	0.101	00:28	Volume Not Met
Breath Volume	0.824	00:28	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 01:39

Gender: Male

Observation Period Began: 01:58

Subject's Last Name: ARNOLD

Date/Time of Test: 12/13/2007 02:24

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample #1	0.179	02:20
Breath Volume	4.046	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample #2	0.176	02:23
Breath Volume	3.968	02:23
Air Blank	0.000	02:23
Control Test	0.079	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARRIS, LISA, M

Breath Test Operator Agency: Bay County SO

Arresting Officer: LANIER, D

Arresting Officer Agency: Bay County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 00:36

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: WARNER

Date/Time of Test: 12/14/2007 02:07

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample #1	0.152	02:02
Breath Volume	2.531	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample #2	0.148	02:05
Breath Volume	1.382	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HICKS, JARED, D

Breath Test Operator Agency: Panama City PD

Arresting Officer: HICKS, JARED

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 03:58

Gender: Male

Observation Period Began: 04:23

Subject's Last Name: TUTHILL

Date/Time of Test: 12/15/2007 04:54

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.081	04:45
Air Blank	0.000	04:46
Subject Sample #1	0.145	04:48
Breath Volume	2.144	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample #2	0.149	04:52
Breath Volume	1.933	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HICKS, JARED, D

Breath Test Operator Agency: Panama City PD

Arresting Officer: ATWELL, PAUL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 04:46

Gender: Male

Observation Period Began: 05:19

Subject's Last Name: MAJORS

Date/Time of Test: 12/15/2007 05:45

Diagnostics Check	OK	05:42	
Air Blank	0.000	05:42	
Control Test	0.081	05:43	
Air Blank	0.000	05:43	
Subject Sample #1		05:43	Subject Test Refused
Breath Volume	0.000	05:43	
Air Blank	0.000	05:44	
Control Test	0.080	05:44	
Air Blank	0.000	05:45	
Diagnostics Check	OK	05:45	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: GARRISON, ROBERT, L

Breath Test Operator Agency: Bay County SO

Arresting Officer: GARRISON, ROBERT

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 09:02

Gender: Male

Observation Period Began: 09:34

Subject's Last Name: NICHOLSON

Date/Time of Test: 12/16/2007 09:57

Diagnostics Check	OK	09:55	
Air Blank	0.000	09:55	
Control Test	0.081	09:56	
Air Blank	0.000	09:56	
Subject Sample #1		09:57	Sequence Aborted
Breath Volume	0.000	09:57	
Air Blank	0.000	09:57	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTTLES, ANITA, A

Breath Test Operator Agency: Springfield PD

Arresting Officer: LILLEY, K

Arresting Officer Agency: Bay County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 18:24

Gender: Male

Observation Period Began: 18:55

Subject's Last Name: MARTIN

Date/Time of Test: 12/16/2007 19:22

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.080	19:16
Air Blank	0.000	19:16
Subject Sample #1	0.290	19:17
Breath Volume	2.800	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:19
Subject Sample #2	0.273	19:20
Breath Volume	2.191	19:20
Air Blank	0.000	19:21
Control Test	0.079	19:21
Air Blank	0.000	19:21
Diagnostics Check	OK	19:22

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WILLIAMS, DAVID, R

Breath Test Operator Agency: Panama City PD

Arresting Officer: FRANCINGUES, LARRY

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 21:11

Gender: Female

Observation Period Began: 21:39

Subject's Last Name: JOHNSON

Date/Time of Test: 12/16/2007 22:06

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.082	22:00
Air Blank	0.000	22:00
Subject Sample #1	0.193	22:01
Breath Volume	1.339	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:03
Subject Sample #2	0.191	22:04
Breath Volume	1.804	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:06

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STEPHENS, SHIRLEY, G

Breath Test Operator Agency: Bay County SO

Arresting Officer: CROSBY, CLAYTON

Arresting Officer Agency: Bay County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 22:07

Gender: Male

Observation Period Began: 22:41

Subject's Last Name: TURNER

Date/Time of Test: 12/17/2007 23:04

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.081	23:02	
Air Blank	0.000	23:03	
Subject Sample #1		23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:04	
Control Test	0.081	23:04	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARRIS, LISA, M

Breath Test Operator Agency: Bay County SO

Arresting Officer: HILDEBRANDT, RYAN

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 18:40

Gender: Male

Observation Period Began: 18:53

Subject's Last Name: KILLGORE

Date/Time of Test: 12/25/2007 19:16

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:13	
Control Test	0.081	19:14	
Air Blank	0.000	19:14	
Subject Sample #1		19:14	Subject Test Refused
Breath Volume	0.000	19:14	
Air Blank	0.000	19:15	
Control Test	0.080	19:15	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:16	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STEPHENS, SHIRLEY, G

Breath Test Operator Agency: Bay County SO

Arresting Officer: NICHOL, CHRIS

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 02:45

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: WOLDEMICHAEL

Date/Time of Test: 12/30/2007 03:26

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.080	03:25	
Air Blank		03:25	RFI Detect
Air Blank	0.000	03:26	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STEPHENS, SHIRLEY, G

Breath Test Operator Agency: Bay County SO

Arresting Officer: NICHOL, CHRIS

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 02:45

Gender: Male

Observation Period Began: 03:26

Subject's Last Name: WOLDEMICHAEL

Date/Time of Test: 12/30/2007 03:55

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.081	03:47	
Air Blank	0.000	03:47	
Subject Sample #1	0.119	03:50	Volume Not Met
Breath Volume	0.066	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample #2		03:54	Subject Test Refused
Breath Volume	0.082	03:54	
Air Blank	0.000	03:54	
Control Test	0.079	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 12/30/2007 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ECKEL, BRIAN,

Breath Test Operator Agency: Panama City PD

Arresting Officer: ECKEL, BRIAN

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 12/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 01:18

Gender: Male

Observation Period Began: 02:08

Subject's Last Name: PRINCE

Date/Time of Test: 12/31/2007 02:35

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.081	02:30	
Air Blank	0.000	02:31	
Subject Sample #1		02:34	Subject Test Refused
Breath Volume	0.078	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 01/07/2008 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: SWEET, CHRISTOPHER, W

Breath Test Operator Agency: Bay County SO

Arresting Officer: LANIER, DAVID

Arresting Officer Agency: Bay County SO

Last Agency Inspection: 12/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 07:12

Gender: Male

Observation Period Began: 07:47

Subject's Last Name: FOX

Date/Time of Test: 12/31/2007 08:14

Diagnostics Check	OK	08:07
Air Blank	0.000	08:08
Control Test	0.081	08:08
Air Blank	0.000	08:09
Subject Sample #1	0.109	08:09
Breath Volume	3.906	08:09
Air Blank	0.000	08:10
Air Blank	0.000	08:12
Subject Sample #2	0.108	08:12
Breath Volume	3.359	08:12
Air Blank	0.000	08:13
Control Test	0.080	08:13
Air Blank	0.000	08:13
Diagnostics Check	OK	08:14

Data Download Date/Time: 01/07/2008 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: STREICHERT, MICHAEL, L

Breath Test Operator Agency: Panama City PD

Arresting Officer: STREICHERT, MICHAEL

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 12/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 12:47

Gender: Female

Observation Period Began: 13:10

Subject's Last Name: DIGIOVANNI

Date/Time of Test: 12/31/2007 13:37

Diagnostics Check	OK	13:31
Air Blank	0.000	13:31
Control Test	0.080	13:31
Air Blank	0.000	13:32
Subject Sample #1	0.000	13:32
Breath Volume	1.484	13:32
Air Blank	0.000	13:33
Air Blank	0.000	13:35
Subject Sample #2	0.000	13:36
Breath Volume	1.800	13:36
Air Blank	0.000	13:36
Control Test	0.080	13:36
Air Blank	0.000	13:37
Diagnostics Check	OK	13:37

Data Download Date/Time: 01/07/2008 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WELCH, MARY, E

Breath Test Operator Agency: Bay County SO

Arresting Officer: ROBERTS, BARRY

Arresting Officer Agency: Springfield PD

Last Agency Inspection: 12/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 20:53

Gender: Male

Observation Period Began: 21:14

Subject's Last Name: PAYNE

Date/Time of Test: 12/31/2007 21:37

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:35	
Control Test	0.080	21:35	
Air Blank	0.000	21:35	
Subject Sample #1		21:36	Subject Test Refused
Breath Volume	0.000	21:36	
Air Blank	0.000	21:36	
Control Test	0.079	21:36	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 01/07/2008 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001308

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARRIS, LISA, M

Breath Test Operator Agency: Bay County SO

Arresting Officer: MCCROAN, N

Arresting Officer Agency: Cedar Grove PD

Last Agency Inspection: 12/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 22:50

Gender: Male

Observation Period Began: 23:15

Subject's Last Name: SANDERS

Date/Time of Test: 12/31/2007 23:44

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample #1	0.147	23:38
Breath Volume	2.621	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:40
Subject Sample #2	0.144	23:42
Breath Volume	2.183	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 01/07/2008 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001309

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: CHAPMAN, C, D

Breath Test Operator Agency: FHP Troop A

Arresting Officer: CHAPMAN, C

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 21:40

Gender: Female

Observation Period Began: 22:08

Subject's Last Name: RICHARDS

Date/Time of Test: 12/04/2007 22:50

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.084	22:47	
Air Blank	0.000	22:48	
Subject Sample #1		22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:49	
Control Test	0.083	22:49	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 12/30/2007 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001309

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ELDRIDGE, BRANDON, A

Breath Test Operator Agency: FHP Troop A

Arresting Officer: ELDRIDGE, BRANDON

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 01:58

Gender: Male

Observation Period Began: 03:03

Subject's Last Name: MISSINNE

Date/Time of Test: 12/07/2007 03:39

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Subject Sample #1	0.090	03:28	Volume Not Met
Breath Volume	0.359	03:28	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample #2	0.104	03:32	
Breath Volume	1.703	03:32	
Air Blank	0.000	03:33	
Air Blank	0.000	03:35	
Subject Sample #3	0.105	03:38	Volume Not Met
Breath Volume	0.078	03:38	
Air Blank	0.000	03:38	
Control Test	0.083	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 12/30/2007 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001309

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ELDRIDGE, BRANDON, A

Breath Test Operator Agency: FHP Troop A

Arresting Officer: ELDRIDGE, BRANDON

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 01:58

Gender: Male

Observation Period Began: 03:39

Subject's Last Name: MISSINNE

Date/Time of Test: 12/07/2007 04:12

Diagnostics Check	OK	03:59	
Air Blank	0.000	04:00	
Control Test	0.082	04:00	
Air Blank	0.000	04:01	
Subject Sample #1	0.096	04:04	Volume Not Met
Breath Volume	0.050	04:04	
Air Blank	0.000	04:04	
Air Blank	0.000	04:06	
Subject Sample #2	0.094	04:08	
Breath Volume	1.878	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample #3	0.098	04:11	
Breath Volume	2.121	04:11	
Air Blank	0.000	04:12	
Control Test	0.082	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 12/30/2007 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001309

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ELDRIDGE, BRANDON, A

Breath Test Operator Agency: FHP Troop A

Arresting Officer: ELDRIDGE, BRANDON

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 02:02

Gender: Male

Observation Period Began: 02:45

Subject's Last Name: DONOHUE

Date/Time of Test: 12/27/2007 03:12

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.085	03:06
Air Blank	0.000	03:06
Subject Sample #1	0.202	03:07
Breath Volume	1.914	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample #2	0.195	03:10
Breath Volume	2.152	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 12/30/2007 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001309

Instrument Registered To: BAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: TAYLOR, CHRISTOPHER, W

Breath Test Operator Agency: Panama City PD

Arresting Officer: TAYLOR, CHRISTOPHER

Arresting Officer Agency: Panama City PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626903I

Dry Gas Expiry Date: 09/28/2008

Violation Code: DUI

Arrest Time: 23:50

Gender: Male

Observation Period Began: 00:09

Subject's Last Name: MCSPADDEN

Date/Time of Test: 12/29/2007 00:36

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.083	00:30
Air Blank	0.000	00:31
Subject Sample #1	0.225	00:31
Breath Volume	1.636	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample #2	0.215	00:34
Breath Volume	2.031	00:34
Air Blank	0.000	00:35
Control Test	0.082	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 12/30/2007 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001310

Instrument Registered To: NICEVILLE PD

Software Version: 8100.27

Breath Test Operator: CRUTHIRDS, JEREMY, D

Breath Test Operator Agency: Niceville PD

Arresting Officer: MURPHY, CHARLES

Arresting Officer Agency: Niceville PD

Last Agency Inspection: 12/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 00:34

Gender: Male

Observation Period Began: 01:05

Subject's Last Name: HAROLD

Date/Time of Test: 12/08/2007 01:31

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.078	01:26
Air Blank	0.000	01:26
Subject Sample #1	0.035	01:27
Breath Volume	1.574	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample #2	0.036	01:30
Breath Volume	1.398	01:30
Air Blank	0.000	01:31
Control Test	0.077	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 01/10/2008 05:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001310

Instrument Registered To: NICEVILLE PD

Software Version: 8100.27

Breath Test Operator: MERRYMAN, DANIEL, J

Breath Test Operator Agency: Okaloosa County SO

Arresting Officer: FORGIONE, JOSEPH

Arresting Officer Agency: Niceville PD

Last Agency Inspection: 12/07/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 03:10

Gender: Male

Observation Period Began: 04:16

Subject's Last Name: LEE

Date/Time of Test: 12/10/2007 04:52

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.078	04:45
Air Blank	0.000	04:46
Subject Sample #1	0.159	04:48
Breath Volume	1.597	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample #2	0.158	04:51
Breath Volume	1.757	04:51
Air Blank	0.000	04:51
Control Test	0.077	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 01/10/2008 05:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001310

Instrument Registered To: NICEVILLE PD

Software Version: 8100.27

Breath Test Operator: AUSTIN, RICHARD, A

Breath Test Operator Agency: Okaloosa County SO

Arresting Officer: MERRYMAN, DANIEL

Arresting Officer Agency: Niceville PD

Last Agency Inspection: 12/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 19:44

Gender: Female

Observation Period Began: 20:17

Subject's Last Name: GARCIA

Date/Time of Test: 12/13/2007 20:50

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:39	
Control Test	0.077	20:39	
Air Blank	0.000	20:40	
Subject Sample #1	0.000	20:43	Volume Not Met
Breath Volume	0.238	20:43	
Air Blank	0.000	20:43	
Air Blank	0.000	20:45	
Subject Sample #2	0.000	20:48	Volume Not Met
Breath Volume	0.015	20:48	
Air Blank	0.000	20:49	
Control Test	0.078	20:49	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time: 01/10/2008 05:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001310

Instrument Registered To: NICEVILLE PD

Software Version: 8100.27

Breath Test Operator: AUSTIN, RICHARD, A

Breath Test Operator Agency: Okaloosa County SO

Arresting Officer: MERRYMAN, DANIEL

Arresting Officer Agency: Niceville PD

Last Agency Inspection: 12/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 19:44

Gender: Female

Observation Period Began: 20:17

Subject's Last Name: GARCIA

Date/Time of Test: 12/13/2007 21:04

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.076	20:55
Air Blank	0.000	20:55
Subject Sample #1	0.000	20:58
Breath Volume	1.808	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample #2	0.000	21:02
Breath Volume	1.660	21:02
Air Blank	0.000	21:03
Control Test	0.077	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:04

Data Download Date/Time: 01/10/2008 05:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001310

Instrument Registered To: NICEVILLE PD

Software Version: 8100.27

Breath Test Operator: MURPHY, CHARLES, A

Breath Test Operator Agency: Niceville PD

Arresting Officer: COSTANTINI, ANTHONY

Arresting Officer Agency: Niceville PD

Last Agency Inspection: 12/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 03:50

Gender: Male

Observation Period Began: 04:22

Subject's Last Name: LOGAN

Date/Time of Test: 12/22/2007 04:56

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.077	04:48	
Air Blank	0.000	04:48	
Subject Sample #1		04:51	No Sample Provided
Breath Volume	0.000	04:51	
Air Blank	0.000	04:52	
Air Blank	0.000	04:54	
Subject Sample #2		04:55	Subject Test Refused
Breath Volume	0.000	04:55	
Air Blank	0.000	04:55	
Control Test	0.078	04:56	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 01/10/2008 05:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001310

Instrument Registered To: NICEVILLE PD

Software Version: 8100.27

Breath Test Operator: CRUTHIRDS, JEREMY, D

Breath Test Operator Agency: Niceville PD

Arresting Officer: CRUTHIRDS, JEREMY

Arresting Officer Agency: Niceville PD

Last Agency Inspection: 12/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 04:02

Gender: Male

Observation Period Began: 05:58

Subject's Last Name: ARMSTRONG

Date/Time of Test: 12/23/2007 06:24

Diagnostics Check	OK	06:18
Air Blank	0.000	06:18
Control Test	0.077	06:19
Air Blank	0.000	06:19
Subject Sample #1	0.139	06:20
Breath Volume	1.414	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:22
Subject Sample #2	0.155	06:23
Breath Volume	2.570	06:23
Air Blank	0.000	06:24
Control Test	0.077	06:24
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

Data Download Date/Time: 01/10/2008 05:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001310

Instrument Registered To: NICEVILLE PD

Software Version: 8100.27

Breath Test Operator: CRUTHIRDS, JEREMY, D

Breath Test Operator Agency: Niceville PD

Arresting Officer: COSTANTINI, TONY

Arresting Officer Agency: Niceville PD

Last Agency Inspection: 12/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 01:58

Gender: Male

Observation Period Began: 02:19

Subject's Last Name: RILEY

Date/Time of Test: 12/27/2007 02:42

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.078	02:40	
Air Blank	0.000	02:41	
Subject Sample #1		02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.078	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 01/10/2008 05:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: THOMPSON, SCOTT, R

Breath Test Operator Agency: Walton County SO

Arresting Officer: THOMPSON, SCOTT

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 18:00

Gender: Female

Observation Period Began: 18:00

Subject's Last Name: CURRINGTON

Date/Time of Test: 12/06/2007 19:22

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.079	19:16
Air Blank	0.000	19:17
Subject Sample #1	0.000	19:18
Breath Volume	1.828	19:18
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample #2	0.000	19:21
Breath Volume	1.882	19:21
Air Blank	0.000	19:22
Control Test	0.078	19:22
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 12/20/2007 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCDANIEL, KOREY, A

Breath Test Operator Agency: Walton County SO

Arresting Officer: MCDANIEL, KOREY

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 23:09

Gender: Male

Observation Period Began: 23:40

Subject's Last Name: BROWN

Date/Time of Test: 12/08/2007 00:18

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.078	00:13
Air Blank	0.000	00:13
Subject Sample #1	0.035	00:14
Breath Volume	3.937	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample #2	0.033	00:17
Breath Volume	4.632	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 12/20/2007 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCDANIEL, KOREY, A

Breath Test Operator Agency: Walton County SO

Arresting Officer: BURLISON, DUSTIN

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 23:41

Gender: Male

Observation Period Began: 00:17

Subject's Last Name: SIMMONS

Date/Time of Test: 12/08/2007 00:56

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Subject Sample #1	0.074	00:51
Breath Volume	2.714	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample #2	0.082	00:54
Breath Volume	5.414	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 12/20/2007 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCDANIEL, KOREY, A

Breath Test Operator Agency: Walton County SO

Arresting Officer: GIPSON, KELLIE

Arresting Officer Agency: OTHER State

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: PROB/PAROL

Arrest Time: 00:25

Gender: Male

Observation Period Began: 00:30

Subject's Last Name: MATTHEWS

Date/Time of Test: 12/08/2007 01:14

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Subject Sample #1	0.181	01:09
Breath Volume	1.730	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample #2	0.181	01:13
Breath Volume	2.210	01:13
Air Blank	0.000	01:13
Control Test	0.078	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 12/20/2007 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: AIMANOVICH, MATTHEW, J

Breath Test Operator Agency: Walton County SO

Arresting Officer: AIMANOVICH, MATTHEW

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 21:15

Gender: Male

Observation Period Began: 20:41

Subject's Last Name: HERRENDES

Date/Time of Test: 12/08/2007 21:29

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.079	21:23
Air Blank	0.000	21:24
Subject Sample #1	0.162	21:24
Breath Volume	1.625	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:26
Subject Sample #2	0.167	21:27
Breath Volume	1.832	21:27
Air Blank	0.000	21:28
Control Test	0.078	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:29

Data Download Date/Time: 12/20/2007 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOCK, JIM, J

Breath Test Operator Agency: Walton County SO

Arresting Officer: MOCK, JIM

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 23:59

Gender: Male

Observation Period Began: 04:55

Subject's Last Name: ADKINSON

Date/Time of Test: 12/10/2007 05:22

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Subject Sample #1	0.000	05:18
Breath Volume	4.695	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample #2	0.000	05:21
Breath Volume	5.074	05:21
Air Blank	0.000	05:21
Control Test	0.078	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 12/20/2007 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEDDON, JASON, N

Breath Test Operator Agency: Walton County SO

Arresting Officer: PETTIS, JOHN

Arresting Officer Agency: DeFuniak Springs PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 02:00

Gender: Female

Observation Period Began: 02:22

Subject's Last Name: ANDREWS

Date/Time of Test: 12/13/2007 02:59

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.078	02:54
Air Blank	0.000	02:54
Subject Sample #1	0.109	02:55
Breath Volume	2.078	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.106	02:58
Breath Volume	2.515	02:58
Air Blank	0.000	02:58
Control Test	0.078	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 12/20/2007 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: JOHNSON, DAVID, C

Breath Test Operator Agency: Walton County SO

Arresting Officer: JOHNSON, DAVID

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 20:52

Gender: Female

Observation Period Began: 21:30

Subject's Last Name: MAZZULLO

Date/Time of Test: 12/13/2007 22:04

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.077	21:59
Air Blank	0.000	22:00
Subject Sample #1	0.172	22:00
Breath Volume	1.492	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample #2	0.163	22:03
Breath Volume	1.578	22:03
Air Blank	0.000	22:04
Control Test	0.078	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 12/20/2007 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: AIMANOVICH, MATTHEW, J

Breath Test Operator Agency: Walton County SO

Arresting Officer: BRADWAY, KATHERINE

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 19:00

Gender: Male

Observation Period Began: 20:22

Subject's Last Name: BELL

Date/Time of Test: 12/14/2007 20:57

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.077	20:51
Air Blank	0.000	20:51
Subject Sample #1	0.040	20:52
Breath Volume	1.390	20:52
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample #2	0.037	20:55
Breath Volume	2.085	20:55
Air Blank	0.000	20:56
Control Test	0.076	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 12/20/2007 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOCK, JIM, J

Breath Test Operator Agency: Walton County SO

Arresting Officer: HARDAGE, TIM

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 22:25

Gender: Male

Observation Period Began: 01:10

Subject's Last Name: LOCKE

Date/Time of Test: 12/22/2007 01:53

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample #1	0.018	01:48
Breath Volume	2.316	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample #2	0.018	01:51
Breath Volume	2.691	01:51
Air Blank	0.000	01:52
Control Test	0.078	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 01/08/2008 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: BLACK, RICHARD, D

Breath Test Operator Agency: DeFuniak Springs PD

Arresting Officer: BLACK, RICHARD

Arresting Officer Agency: DeFuniak Springs PD

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 15:14

Gender: Male

Observation Period Began: 16:00

Subject's Last Name: HOWELL

Date/Time of Test: 12/23/2007 16:44

Diagnostics Check	OK	16:42	
Air Blank	0.000	16:42	
Control Test	0.078	16:43	
Air Blank	0.000	16:43	
Subject Sample #1		16:43	Sequence Aborted
Breath Volume	0.000	16:43	
Air Blank	0.000	16:44	

Data Download Date/Time: 01/08/2008 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: SLICK, LOREN, D

Breath Test Operator Agency: FHP Troop A

Arresting Officer: SLICK, LOREN

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 22:33

Gender: Female

Observation Period Began: 23:26

Subject's Last Name: BUTLER

Date/Time of Test: 12/23/2007 23:56

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.079	23:50
Air Blank	0.000	23:50
Subject Sample #1	0.208	23:52
Breath Volume	2.402	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample #2	0.223	23:55
Breath Volume	1.781	23:55
Air Blank	0.000	23:55
Control Test	0.078	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 01/08/2008 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEDDON, JASON, N

Breath Test Operator Agency: Walton County SO

Arresting Officer: PETTIS, JOHN

Arresting Officer Agency: DeFuniak Springs PD

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 03:00

Gender: Male

Observation Period Began: 03:43

Subject's Last Name: MONROY DAVILA

Date/Time of Test: 12/25/2007 04:14

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.079	04:12	
Air Blank	0.000	04:12	
Subject Sample #1		04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.079	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 01/08/2008 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCMILLIAN, RUSSELL, T

Breath Test Operator Agency: Walton County SO

Arresting Officer: MCMILLIAN, RUSSELL

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 04:21

Gender: Female

Observation Period Began: 05:13

Subject's Last Name: BATSON

Date/Time of Test: 12/28/2007 05:49

Diagnostics Check	OK	05:42	
Air Blank	0.000	05:43	
Control Test	0.077	05:43	
Air Blank	0.000	05:43	
Subject Sample #1	0.188	05:45	
Breath Volume	1.410	05:45	
Air Blank	0.000	05:45	
Air Blank	0.000	05:47	
Subject Sample #2	0.176	05:47	
Breath Volume	2.679	05:47	
Air Blank	0.000	05:48	
Control Test	0.074	05:48	Control Outside Tolerance
Air Blank	0.000	05:49	

Data Download Date/Time: 01/08/2008 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOCK, JIM, J

Breath Test Operator Agency: Walton County SO

Arresting Officer: MARTIN, NATHAN

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 22:17

Gender: Female

Observation Period Began: 22:54

Subject's Last Name: TANGUAY

Date/Time of Test: 12/28/2007 23:24

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.078	23:18
Air Blank	0.000	23:18
Subject Sample #1	0.175	23:19
Breath Volume	1.882	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample #2	0.191	23:22
Breath Volume	2.882	23:22
Air Blank	0.000	23:23
Control Test	0.078	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 01/08/2008 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: MARTIN, NATHAN, J

Breath Test Operator Agency: Walton County SO

Arresting Officer: PETTIS, JOHN

Arresting Officer Agency: DeFuniak Springs PD

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 02:07

Gender: Male

Observation Period Began: 02:49

Subject's Last Name: LUCIANO

Date/Time of Test: 12/29/2007 03:18

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.077	03:12	
Air Blank	0.000	03:13	
Subject Sample #1	0.165	03:13	
Breath Volume	2.437	03:13	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample #2		03:18	Sequence Aborted
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	

Data Download Date/Time: 01/08/2008 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: JOHNSON, DAVID, C

Breath Test Operator Agency: Walton County SO

Arresting Officer: JOHNSON, DAVID

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 19:06

Gender: Male

Observation Period Began: 20:30

Subject's Last Name: DUNCAN

Date/Time of Test: 12/29/2007 21:05

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:03	
Control Test	0.078	21:03	
Air Blank	0.000	21:03	
Subject Sample #1		21:04	Subject Test Refused
Breath Volume	0.000	21:04	
Air Blank	0.000	21:04	
Control Test	0.079	21:04	
Air Blank	0.000	21:05	
Diagnostics Check	OK	21:05	

Data Download Date/Time: 01/08/2008 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: JOHNSON, DAVID, C

Breath Test Operator Agency: Walton County SO

Arresting Officer: JORDAN, JOHN

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 20:02

Gender: Male

Observation Period Began: 20:30

Subject's Last Name: PHILLIPS

Date/Time of Test: 12/29/2007 21:14

Diagnostics Check	OK	21:07
Air Blank	0.000	21:08
Control Test	0.078	21:08
Air Blank	0.000	21:08
Subject Sample #1	0.233	21:09
Breath Volume	2.156	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample #2	0.236	21:12
Breath Volume	2.085	21:12
Air Blank	0.000	21:13
Control Test	0.077	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:14

Data Download Date/Time: 01/08/2008 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001311

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCDANIEL, KOREY, A

Breath Test Operator Agency: Walton County SO

Arresting Officer: HORNSBY, STEPHEN

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 22:04

Gender: Female

Observation Period Began: 22:40

Subject's Last Name: HECKINGER

Date/Time of Test: 12/31/2007 23:19

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.077	23:14
Air Blank	0.000	23:14
Subject Sample #1	0.055	23:14
Breath Volume	2.953	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample #2	0.056	23:17
Breath Volume	2.605	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 01/08/2008 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: HOGEBROOM, MICHAEL, S

Breath Test Operator Agency: Walton County SO

Arresting Officer: STANLEY, FRED

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 19:53

Gender: Male

Observation Period Began: 20:45

Subject's Last Name: GOMEZ-LOPEZ

Date/Time of Test: 12/01/2007 21:21

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.078	21:15
Air Blank	0.000	21:16
Subject Sample #1	0.202	21:16
Breath Volume	2.406	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample #2	0.202	21:19
Breath Volume	2.566	21:19
Air Blank	0.000	21:20
Control Test	0.077	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 12/20/2007 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: LIEB, JOHN, C

Breath Test Operator Agency: FHP Troop A

Arresting Officer: LIEB, JOHN

Arresting Officer Agency: FHP Troop A

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 00:43

Gender: Male

Observation Period Began: 01:02

Subject's Last Name: COE

Date/Time of Test: 12/02/2007 01:33

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.078	01:30	
Air Blank	0.000	01:31	
Subject Sample #1		01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 12/20/2007 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: GLENISTER, JOHN, D

Breath Test Operator Agency: Walton County SO

Arresting Officer: GLENISTER, JOHN

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 15:36

Gender: Male

Observation Period Began: 16:10

Subject's Last Name: FORBES

Date/Time of Test: 12/06/2007 16:49

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.079	16:41
Air Blank	0.000	16:41
Subject Sample #1	0.297	16:43
Breath Volume	1.156	16:43
Air Blank	0.000	16:43
Air Blank	0.000	16:45
Subject Sample #2	0.316	16:47
Breath Volume	2.402	16:47
Air Blank	0.000	16:48
Control Test	0.078	16:48
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 12/20/2007 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: SKIPPER, JEFFREY, L

Breath Test Operator Agency: Walton County SO

Arresting Officer: SKIPPER, JEFFREY

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 09:54

Gender: Male

Observation Period Began: 10:10

Subject's Last Name: MCCALVIN

Date/Time of Test: 12/07/2007 10:48

Diagnostics Check	OK	10:42
Air Blank	0.000	10:42
Control Test	0.078	10:43
Air Blank	0.000	10:43
Subject Sample #1	0.000	10:44
Breath Volume	3.296	10:44
Air Blank	0.000	10:44
Air Blank	0.000	10:46
Subject Sample #2	0.000	10:47
Breath Volume	3.609	10:47
Air Blank	0.000	10:47
Control Test	0.077	10:48
Air Blank	0.000	10:48
Diagnostics Check	OK	10:48

Data Download Date/Time: 12/20/2007 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: SKIPPER, JEFFREY, L

Breath Test Operator Agency: Walton County SO

Arresting Officer: SKIPPER, JEFFREY

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 13:20

Gender: Male

Observation Period Began: 13:42

Subject's Last Name: RALPH

Date/Time of Test: 12/08/2007 14:16

Diagnostics Check	OK	14:08
Air Blank	0.000	14:09
Control Test	0.079	14:09
Air Blank	0.000	14:10
Subject Sample #1	0.146	14:11
Breath Volume	1.445	14:11
Air Blank	0.000	14:12
Air Blank	0.000	14:14
Subject Sample #2	0.153	14:15
Breath Volume	1.730	14:15
Air Blank	0.000	14:16
Control Test	0.076	14:16
Air Blank	0.000	14:16
Diagnostics Check	OK	14:16

Data Download Date/Time: 12/20/2007 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: GRANTHAM, JOHNNY, R

Breath Test Operator Agency: Walton County SO

Arresting Officer: GRANTHAM, JOHNNY

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 06/13/2007

Violation Code: DUI

Arrest Time: 02:03

Gender: Female

Observation Period Began: 02:40

Subject's Last Name: BOODEN

Date/Time of Test: 12/09/2007 03:29

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Subject Sample #1	0.222	03:24
Breath Volume	2.906	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample #2	0.223	03:27
Breath Volume	2.699	03:27
Air Blank	0.000	03:28
Control Test	0.076	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 12/20/2007 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: TROMBLY, GAYLORD, L

Breath Test Operator Agency: Walton County SO

Arresting Officer: TROMBLY, GAYLORD

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 23:30

Gender: Male

Observation Period Began: 23:55

Subject's Last Name: ACKER

Date/Time of Test: 12/15/2007 00:33

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.078	00:26	
Air Blank	0.000	00:27	
Subject Sample #1	0.094	00:28	
Breath Volume	2.824	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:30	
Subject Sample #2	0.092	00:31	
Breath Volume	2.691	00:31	
Air Blank	0.000	00:32	
Control Test	0.078	00:32	
Air Blank		00:32	RFI Detect
Air Blank		00:33	RFI Detect

Data Download Date/Time: 12/20/2007 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAMBO, BRUCE, A

Breath Test Operator Agency: Walton County SO

Arresting Officer: SHAMBO, BRUCE

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 01:56

Gender: Male

Observation Period Began: 02:41

Subject's Last Name: STADELMAN

Date/Time of Test: 12/15/2007 03:08

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.078	03:06	
Air Blank	0.000	03:07	
Subject Sample #1		03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 12/20/2007 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEHMAN, CHRISTOPHER, J

Breath Test Operator Agency: Walton County SO

Arresting Officer: LEHMAN, CHRISTOPHER

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 03:14

Gender: Male

Observation Period Began: 04:37

Subject's Last Name: NERO

Date/Time of Test: 12/15/2007 05:04

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.078	04:59
Air Blank	0.000	04:59
Subject Sample #1	0.221	05:00
Breath Volume	3.257	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample #2	0.206	05:03
Breath Volume	2.414	05:03
Air Blank	0.000	05:03
Control Test	0.078	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 12/20/2007 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: COOK, LEONARD, C

Breath Test Operator Agency: Walton County SO

Arresting Officer: SHEPHERD, CHIRSTIAN

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 02:17

Subject's Last Name: ARSHAWSKY

Date/Time of Test: 12/20/2007 02:45

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.078	02:43	
Air Blank	0.000	02:43	
Subject Sample #1		02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 12/20/2007 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001312

Instrument Registered To: WALTON COUNTY SO

Software Version: 8100.27

Breath Test Operator: SAVAGE, DONALD, L

Breath Test Operator Agency: Walton County SO

Arresting Officer: SAVAGE, DONALD

Arresting Officer Agency: Walton County SO

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 15:20

Gender: Male

Observation Period Began: 15:58

Subject's Last Name: SMITH

Date/Time of Test: 12/20/2007 16:33

Diagnostics Check	OK	16:27
Air Blank	0.000	16:27
Control Test	0.078	16:27
Air Blank	0.000	16:28
Subject Sample #1	0.296	16:28
Breath Volume	2.105	16:28
Air Blank	0.000	16:29
Air Blank	0.000	16:31
Subject Sample #2	0.295	16:32
Breath Volume	4.417	16:32
Air Blank	0.000	16:32
Control Test	0.076	16:33
Air Blank	0.000	16:33
Diagnostics Check	OK	16:33

Data Download Date/Time: 01/08/2008 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: FERRELL, PAUL, E

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: FERRELL, PAUL

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: ADMIN

Arrest Time: 10:33

Gender: Male

Observation Period Began: 10:33

Subject's Last Name: KIBLER

Date/Time of Test: 12/04/2007 11:41

Diagnostics Check	OK	11:35
Air Blank	0.000	11:36
Control Test	0.079	11:36
Air Blank	0.000	11:37
Subject Sample #1	0.000	11:37
Breath Volume	4.511	11:37
Air Blank	0.000	11:37
Air Blank	0.000	11:39
Subject Sample #2	0.000	11:40
Breath Volume	3.417	11:40
Air Blank	0.000	11:40
Control Test	0.077	11:41
Air Blank	0.000	11:41
Diagnostics Check	OK	11:41

Data Download Date/Time: 12/27/2007 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: GARLAND, MARK, W

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: SHIREMAN, MARK

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 19:23

Gender: Male

Observation Period Began: 20:02

Subject's Last Name: YANEZ-PEREZ

Date/Time of Test: 12/08/2007 20:34

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.078	20:28
Air Blank	0.000	20:29
Subject Sample #1	0.199	20:30
Breath Volume	1.902	20:30
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample #2	0.201	20:33
Breath Volume	1.671	20:33
Air Blank	0.000	20:33
Control Test	0.077	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 12/27/2007 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: ASHBY, JOHN, R

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: ASHBY, JOHN

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: ADMIN

Arrest Time: 20:46

Gender: Male

Observation Period Began: 20:40

Subject's Last Name: MAERKI

Date/Time of Test: 12/08/2007 21:26

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.078	21:21
Air Blank	0.000	21:21
Subject Sample #1	0.000	21:22
Breath Volume	3.878	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample #2	0.000	21:24
Breath Volume	5.332	21:24
Air Blank	0.000	21:25
Control Test	0.077	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 12/27/2007 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: GARLAND, MARK, W

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: HANS, CHRIS

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 23:10

Gender: Male

Observation Period Began: 23:47

Subject's Last Name: CHALLANCIN

Date/Time of Test: 12/09/2007 00:10

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.077	00:08	
Air Blank	0.000	00:09	
Subject Sample #1		00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:10	
Control Test	0.077	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 12/27/2007 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: MAJERE, ARIC,

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: ASHBY, JOHN

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 20:07

Gender: Male

Observation Period Began: 20:30

Subject's Last Name: MARTIN

Date/Time of Test: 12/20/2007 20:57

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.077	20:51
Air Blank	0.000	20:52
Subject Sample #1	0.176	20:52
Breath Volume	3.449	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample #2	0.172	20:55
Breath Volume	3.414	20:55
Air Blank	0.000	20:56
Control Test	0.076	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 12/27/2007 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: FARLESS, ROSEMARY, U

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: FARLESS, ROSEMARY

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: ADMIN

Arrest Time: 15:45

Gender: Male

Observation Period Began: 16:09

Subject's Last Name: BORZILLO

Date/Time of Test: 12/21/2007 16:31

Diagnostics Check	OK	16:29	
Air Blank	0.000	16:30	
Control Test	0.079	16:30	
Air Blank		16:30	RFI Detect
Air Blank	0.000	16:31	

Data Download Date/Time: 12/27/2007 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: FARLESS, ROSEMARY, U

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: FARLESS, ROSEMARY

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: ADMIN

Arrest Time: 15:45

Gender: Male

Observation Period Began: 15:50

Subject's Last Name: BORZILLO

Date/Time of Test: 12/21/2007 16:39

Diagnostics Check	OK	16:33
Air Blank	0.000	16:34
Control Test	0.078	16:34
Air Blank	0.000	16:35
Subject Sample #1	0.000	16:35
Breath Volume	4.082	16:35
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample #2	0.000	16:38
Breath Volume	4.398	16:38
Air Blank	0.000	16:38
Control Test	0.076	16:39
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

Data Download Date/Time: 12/27/2007 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: GARLAND, MARK, W

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: SHIREMAN, MARK

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 18:56

Gender: Male

Observation Period Began: 19:16

Subject's Last Name: GUERRERO

Date/Time of Test: 12/21/2007 19:46

Diagnostics Check	OK	19:38	
Air Blank	0.000	19:38	
Control Test	0.078	19:38	
Air Blank	0.000	19:39	
Subject Sample #1	0.233	19:40	
Breath Volume	3.183	19:40	
Air Blank	0.000	19:41	
Air Blank	0.000	19:43	
Subject Sample #2	0.227	19:44	
Breath Volume	1.566	19:44	
Air Blank	0.000	19:45	
Control Test	0.074	19:45	Control Outside Tolerance
Air Blank	0.000	19:46	

Data Download Date/Time: 12/27/2007 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: GARLAND, MARK, W

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: HANS, CHRIS

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 22:19

Gender: Male

Observation Period Began: 22:38

Subject's Last Name: HERNANDEZ

Date/Time of Test: 12/22/2007 23:16

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.077	23:10
Air Blank	0.000	23:11
Subject Sample #1	0.208	23:11
Breath Volume	2.515	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample #2	0.220	23:14
Breath Volume	3.785	23:14
Air Blank	0.000	23:15
Control Test	0.077	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 12/27/2007 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: GARLAND, MARK, W

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: SHIREMAN, MARK

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 23:50

Gender: Male

Observation Period Began: 01:10

Subject's Last Name: VALENCIA-RUIZ

Date/Time of Test: 12/24/2007 01:33

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.078	01:31	
Air Blank	0.000	01:32	
Subject Sample #1		01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Control Test	0.079	01:33	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 12/27/2007 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001320

Instrument Registered To: OKEECHOBEE CO SO

Software Version: 8100.27

Breath Test Operator: MAJERE, ARIC,

Breath Test Operator Agency: Okeechobee County SO

Arresting Officer: ASHBY, JOHN

Arresting Officer Agency: Okeechobee County SO

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 00:27

Gender: Male

Observation Period Began: 01:01

Subject's Last Name: HOLLOWAY

Date/Time of Test: 12/30/2007 01:29

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.078	01:22
Air Blank	0.000	01:23
Subject Sample #1	0.200	01:24
Breath Volume	1.867	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample #2	0.203	01:27
Breath Volume	1.777	01:27
Air Blank	0.000	01:28
Control Test	0.075	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 02/04/2008 03:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001322

Instrument Registered To: OKEECHOBEE PD

Software Version: 8100.27

Breath Test Operator: BOON, JACK, D

Breath Test Operator Agency: Okeechobee PD

Arresting Officer: TROUTMAN, CHAD

Arresting Officer Agency: Okeechobee PD

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7061011

Dry Gas Expiry Date: 03/06/2009

Violation Code: DUI

Arrest Time: 01:28

Gender: Female

Observation Period Began: 01:49

Subject's Last Name: PATTERSON

Date/Time of Test: 12/06/2007 02:21

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:16	
Control Test	0.077	02:17	
Air Blank	0.000	02:17	
Subject Sample #1		02:20	Subject Test Refused
Breath Volume	0.371	02:20	
Air Blank	0.000	02:20	
Control Test	0.077	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 12/28/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001322

Instrument Registered To: OKEECHOBEE PD

Software Version: 8100.27

Breath Test Operator: BOON, JACK, D

Breath Test Operator Agency: Okeechobee PD

Arresting Officer: TROUTMAN, CHAD

Arresting Officer Agency: Okeechobee PD

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7061011

Dry Gas Expiry Date: 03/06/2009

Violation Code: DUI

Arrest Time: 01:28

Gender: Female

Observation Period Began: 01:49

Subject's Last Name: PATTERSON

Date/Time of Test: 12/06/2007 02:39

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.077	02:26	
Air Blank	0.000	02:26	
Subject Sample #1	0.177	02:28	
Breath Volume	1.292	02:28	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample #2	0.156	02:33	Slope Not Level
Breath Volume	1.273	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample #3		02:37	Subject Test Refused
Breath Volume	0.964	02:37	
Air Blank	0.000	02:38	
Control Test	0.076	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 12/28/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BRADY, KEVIN, E

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: OHARA, ROBERT

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 03:48

Gender: Male

Observation Period Began: 04:30

Subject's Last Name: ANDRADE

Date/Time of Test: 12/02/2007 05:05

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.077	05:00
Air Blank	0.000	05:00
Subject Sample #1	0.182	05:01
Breath Volume	2.945	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample #2	0.181	05:04
Breath Volume	1.523	05:04
Air Blank	0.000	05:04
Control Test	0.078	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: HARRIS, OFC, I
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: ZAMOYSKI, OFC
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 23:45
Gender: Male
Observation Period Began: 00:10

Subject's Last Name: SAPON-GARCIA

Date/Time of Test: 12/08/2007 00:38

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
Control Test	0.077	00:36	
Air Blank	0.000	00:36	
Subject Sample #1	0.133	00:37	
Breath Volume	1.320	00:37	
Air Blank		00:37	Improper Sample
Air Blank	0.000	00:37	
Control Test	0.077	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: ZAMOYSKI, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:45

Gender: Male

Observation Period Began: 00:10

Subject's Last Name: SAPON-GARCIA

Date/Time of Test: 12/08/2007 00:46

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.075	00:41
Air Blank	0.000	00:41
Subject Sample #1	0.139	00:42
Breath Volume	1.843	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample #2	0.145	00:45
Breath Volume	2.214	00:45
Air Blank	0.000	00:45
Control Test	0.077	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: LAWLER, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 00:44

Gender: Male

Observation Period Began: 01:12

Subject's Last Name: GAVIN

Date/Time of Test: 12/08/2007 01:41

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.077	01:35
Air Blank	0.000	01:35
Subject Sample #1	0.132	01:35
Breath Volume	1.636	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample #2	0.137	01:39
Breath Volume	1.191	01:39
Air Blank	0.000	01:40
Control Test	0.077	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: KERNS, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:38

Gender: Male

Observation Period Began: 01:50

Subject's Last Name: LANDRY

Date/Time of Test: 12/08/2007 02:17

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.077	02:11
Air Blank	0.000	02:12
Subject Sample #1	0.141	02:12
Breath Volume	3.148	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample #2	0.139	02:15
Breath Volume	2.894	02:15
Air Blank	0.000	02:16
Control Test	0.077	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: HARRIS, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:46

Gender: Male

Observation Period Began: 03:02

Subject's Last Name: IZETA LOPEZ

Date/Time of Test: 12/08/2007 03:42

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.077	03:37
Air Blank	0.000	03:37
Subject Sample #1	0.282	03:38
Breath Volume	2.511	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample #2	0.298	03:41
Breath Volume	1.394	03:41
Air Blank	0.000	03:41
Control Test	0.077	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: N/A, N/A

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: ADMIN

Arrest Time: 00:00

Gender: Male

Observation Period Began: 23:21

Subject's Last Name: BARNWELL

Date/Time of Test: 12/08/2007 23:54

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:49
Subject Sample #1	0.117	23:50
Breath Volume	1.566	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample #2	0.122	23:53
Breath Volume	2.390	23:53
Air Blank	0.000	23:53
Control Test	0.078	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: DE LA FUENTE, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:27

Gender: Male

Observation Period Began: 00:13

Subject's Last Name: AGUADO

Date/Time of Test: 12/09/2007 01:14

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.077	01:08
Air Blank	0.000	01:09
Subject Sample #1	0.168	01:09
Breath Volume	2.496	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample #2	0.161	01:12
Breath Volume	2.250	01:12
Air Blank	0.000	01:13
Control Test	0.077	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: CUTY, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:44

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: MULLETT

Date/Time of Test: 12/09/2007 03:36

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.078	03:31
Air Blank	0.000	03:31
Subject Sample #1	0.159	03:32
Breath Volume	2.445	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample #2	0.166	03:35
Breath Volume	2.257	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: LIDBETTER, DARREN, P
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: LOUPE, R
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 07:50
Gender: Male
Observation Period Began: 08:30

Subject's Last Name: FLORES

Date/Time of Test: 12/10/2007 09:08

Diagnostics Check	OK	09:02
Air Blank	0.000	09:03
Control Test	0.078	09:03
Air Blank	0.000	09:04
Subject Sample #1	0.000	09:04
Breath Volume	2.273	09:04
Air Blank	0.000	09:05
Air Blank	0.000	09:07
Subject Sample #2	0.000	09:07
Breath Volume	4.320	09:07
Air Blank	0.000	09:08
Control Test	0.078	09:08
Air Blank	0.000	09:08
Diagnostics Check	OK	09:08

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: O HARA, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 00:54

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: LESS

Date/Time of Test: 12/12/2007 01:54

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.076	01:49
Air Blank	0.000	01:49
Subject Sample #1	0.160	01:49
Breath Volume	2.968	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample #2	0.158	01:52
Breath Volume	2.875	01:52
Air Blank	0.000	01:53
Control Test	0.077	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: BRADY, KEVIN, E
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: DE LA FUENTE, KARL
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 23:49
Gender: Male
Observation Period Began: 00:01

Subject's Last Name: MOWL

Date/Time of Test: 12/14/2007 00:52

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.077	00:47
Air Blank	0.000	00:47
Subject Sample #1	0.238	00:48
Breath Volume	2.562	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample #2	0.225	00:51
Breath Volume	2.878	00:51
Air Blank	0.000	00:51
Control Test	0.076	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: CUTY, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:18

Gender: Male

Observation Period Began: 23:45

Subject's Last Name: ODEN

Date/Time of Test: 12/15/2007 00:13

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.077	00:07
Air Blank	0.000	00:08
Subject Sample #1	0.231	00:08
Breath Volume	2.636	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample #2	0.218	00:11
Breath Volume	2.640	00:11
Air Blank	0.000	00:12
Control Test	0.076	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: ZAMOYSKI, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 00:09

Gender: Female

Observation Period Began: 00:20

Subject's Last Name: CARLING

Date/Time of Test: 12/15/2007 00:50

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.077	00:43
Air Blank	0.000	00:43
Subject Sample #1	0.192	00:44
Breath Volume	2.128	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample #2	0.189	00:48
Breath Volume	2.570	00:48
Air Blank	0.000	00:49
Control Test	0.076	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: HARRIS, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:10

Gender: Female

Observation Period Began: 01:30

Subject's Last Name: SWARTZ

Date/Time of Test: 12/15/2007 02:00

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.077	01:54
Air Blank	0.000	01:55
Subject Sample #1	0.092	01:55
Breath Volume	3.140	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample #2	0.091	01:58
Breath Volume	3.546	01:58
Air Blank	0.000	01:59
Control Test	0.077	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: CUTY, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 03:06

Gender: Male

Observation Period Began: 03:35

Subject's Last Name: CHMURA

Date/Time of Test: 12/15/2007 03:58

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.073	03:58	Control Outside Tolerance
Air Blank	0.000	03:58	

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: CUTY, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 03:06

Gender: Male

Observation Period Began: 03:35

Subject's Last Name: CHMURA

Date/Time of Test: 12/15/2007 04:06

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.077	04:00
Air Blank	0.000	04:01
Subject Sample #1	0.235	04:01
Breath Volume	2.328	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample #2	0.239	04:05
Breath Volume	2.910	04:05
Air Blank	0.000	04:05
Control Test	0.077	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: ZAMOYSKI, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 03:14

Gender: Female

Observation Period Began: 03:40

Subject's Last Name: WIRTEL

Date/Time of Test: 12/15/2007 04:13

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.077	04:08
Air Blank	0.000	04:09
Subject Sample #1	0.102	04:09
Breath Volume	2.480	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample #2	0.101	04:12
Breath Volume	2.480	04:12
Air Blank	0.000	04:13
Control Test	0.076	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: KING, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 03:57

Gender: Male

Observation Period Began: 04:15

Subject's Last Name: DE LA CRUZ

Date/Time of Test: 12/15/2007 04:42

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.077	04:36
Air Blank	0.000	04:36
Subject Sample #1	0.114	04:37
Breath Volume	3.042	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:39
Subject Sample #2	0.115	04:40
Breath Volume	3.828	04:40
Air Blank	0.000	04:41
Control Test	0.077	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: HARRIS, OFC, I
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: LAWLER, OFC
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 03:57
Gender: Female
Observation Period Began: 05:16

Subject's Last Name: DE LA CRUZ

Date/Time of Test: 12/15/2007 05:43

Diagnostics Check	OK	05:37
Air Blank	0.000	05:37
Control Test	0.078	05:37
Air Blank	0.000	05:38
Subject Sample #1	0.169	05:38
Breath Volume	3.460	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample #2	0.168	05:41
Breath Volume	4.304	05:41
Air Blank	0.000	05:42
Control Test	0.076	05:42
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: HARRIS, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 20:34

Gender: Male

Observation Period Began: 21:00

Subject's Last Name: SOTO MARTINEZ

Date/Time of Test: 12/15/2007 21:34

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:25	
Control Test	0.078	21:25	
Air Blank	0.000	21:26	
Subject Sample #1	0.176	21:26	
Breath Volume	2.558	21:26	
Air Blank	0.000	21:27	
Air Blank	0.000	21:29	
Subject Sample #2	0.148	21:29	
Breath Volume	1.765	21:29	
Air Blank	0.000	21:30	
Air Blank	0.000	21:32	
Subject Sample #3		21:32	Slope Not Met
Breath Volume	0.746	21:32	
Air Blank	0.000	21:33	
Control Test	0.077	21:33	
Air Blank	0.000	21:34	
Diagnostics Check	OK	21:34	

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: HARRIS, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 20:34

Gender: Male

Observation Period Began: 21:00

Subject's Last Name: SOTO MARTINEZ

Date/Time of Test: 12/15/2007 21:39

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:36	
Control Test	0.077	21:36	
Air Blank	0.000	21:37	
Subject Sample #1		21:38	Slope Not Met
Breath Volume	0.914	21:38	
Air Blank	0.000	21:38	
Control Test	0.078	21:39	
Air Blank	0.000	21:39	
Diagnostics Check	OK	21:39	

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: HARRIS, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 20:34

Gender: Male

Observation Period Began: 21:00

Subject's Last Name: SOTO MARTINEZ

Date/Time of Test: 12/15/2007 21:48

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.077	21:42
Air Blank	0.000	21:42
Subject Sample #1	0.157	21:43
Breath Volume	1.656	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample #2	0.159	21:46
Breath Volume	2.667	21:46
Air Blank	0.000	21:47
Control Test	0.077	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: HARRIS, OFC, I
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: KERNS, OFC
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 00:17
Gender: Male
Observation Period Began: 00:45

Subject's Last Name: GRACIAK

Date/Time of Test: 12/16/2007 01:13

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.078	01:07
Air Blank	0.000	01:08
Subject Sample #1	0.173	01:09
Breath Volume	2.390	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:11
Subject Sample #2	0.189	01:12
Breath Volume	2.867	01:12
Air Blank	0.000	01:13
Control Test	0.077	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: HARRIS, OFC, I
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: HARRIS, OFC
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 01:36
Gender: Male
Observation Period Began: 01:50

Subject's Last Name: HERNANDEZ PEREZ

Date/Time of Test: 12/16/2007 02:16

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.078	02:11
Air Blank	0.000	02:11
Subject Sample #1	0.142	02:12
Breath Volume	3.656	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample #2	0.124	02:15
Breath Volume	3.082	02:15
Air Blank	0.000	02:16
Control Test	0.076	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: HARRIS, OFC, I
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: ZAMOYSKI, OFC
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 02:33
Gender: Male
Observation Period Began: 02:55

Subject's Last Name: KENNEDY

Date/Time of Test: 12/16/2007 03:22

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.078	03:16
Air Blank	0.000	03:16
Subject Sample #1	0.224	03:17
Breath Volume	3.722	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample #2	0.230	03:20
Breath Volume	4.175	03:20
Air Blank	0.000	03:21
Control Test	0.077	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: HARRIS, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 20:37

Gender: Female

Observation Period Began: 21:15

Subject's Last Name: SPANARKEL

Date/Time of Test: 12/21/2007 21:45

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:40
Subject Sample #1	0.226	21:40
Breath Volume	2.785	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample #2	0.207	21:43
Breath Volume	2.148	21:43
Air Blank	0.000	21:44
Control Test	0.077	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: BRADY, KEVIN, E
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: OHARA, BOB
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 01:10
Gender: Male
Observation Period Began: 01:30

Subject's Last Name: FORTNEY

Date/Time of Test: 12/23/2007 01:59

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.078	01:53	
Air Blank	0.000	01:54	
Subject Sample #1	0.158	01:54	
Breath Volume	4.683	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample #2		01:57	RFI Detect
Breath Volume	0.000	01:57	
Air Blank	0.000	01:58	
Control Test	0.077	01:58	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BRADY, KEVIN, E

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: OHARA, BOB

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:10

Gender: Male

Observation Period Began: 01:30

Subject's Last Name: FORTNEY

Date/Time of Test: 12/23/2007 02:06

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.078	02:01
Air Blank	0.000	02:01
Subject Sample #1	0.168	02:02
Breath Volume	3.851	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample #2	0.153	02:05
Breath Volume	2.351	02:05
Air Blank	0.000	02:05
Control Test	0.077	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: HARRIS, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: WALLACE

Date/Time of Test: 12/28/2007 02:29

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.078	02:24
Air Blank	0.000	02:24
Subject Sample #1	0.184	02:25
Breath Volume	3.664	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample #2	0.176	02:27
Breath Volume	3.300	02:27
Air Blank	0.000	02:28
Control Test	0.076	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: HARRIS, OFC, I
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: ZAMOYSKI, OFC
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 02:20
Gender: Male
Observation Period Began: 02:38

Subject's Last Name: SAMUEL

Date/Time of Test: 12/28/2007 03:07

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.077	03:00
Air Blank	0.000	03:01
Subject Sample #1	0.074	03:02
Breath Volume	1.542	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample #2	0.067	03:05
Breath Volume	1.562	03:05
Air Blank	0.000	03:06
Control Test	0.076	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: CUTY, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 11/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:07

Gender: Female

Observation Period Began: 01:30

Subject's Last Name: HOOZER

Date/Time of Test: 12/29/2007 01:58

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.078	01:52
Air Blank	0.000	01:53
Subject Sample #1	0.072	01:53
Breath Volume	2.722	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.072	01:56
Breath Volume	3.511	01:56
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: BRADY, KEVIN, E
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: LAWLER, ANDREW
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 02:53
Gender: Male
Observation Period Began: 03:12

Subject's Last Name: RHODES

Date/Time of Test: 12/29/2007 03:41

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.077	03:36
Air Blank	0.000	03:36
Subject Sample #1	0.052	03:36
Breath Volume	2.363	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample #2	0.053	03:40
Breath Volume	3.027	03:40
Air Blank	0.000	03:40
Control Test	0.078	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001323

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: HARRIS, OFC, I
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: ROBISON, OFC
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 11/23/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 701201G
Dry Gas Expiry Date: 01/15/2009
Violation Code: DUI
Arrest Time: 19:00
Gender: Male
Observation Period Began: 19:46

Subject's Last Name: SCOTESE

Date/Time of Test: 12/30/2007 20:12

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:07	
Control Test	0.077	20:08	
Air Blank	0.000	20:08	
Subject Sample #1		20:11	Slope Not Met
Breath Volume	0.941	20:11	
Air Blank	0.000	20:11	
Control Test	0.077	20:12	
Air Blank	0.000	20:12	
Diagnostics Check	OK	20:12	

Data Download Date/Time: 12/31/2007 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: DANIELS, RON

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 11/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:52

Gender: Female

Observation Period Began: 02:05

Subject's Last Name: ANDERSON

Date/Time of Test: 12/01/2007 02:42

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.079	02:40	
Air Blank	0.000	02:40	
Subject Sample #1		02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 12/06/2007 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PIGGOTT, MYRON, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: MORRELL, BRIAN

Arresting Officer Agency: Fellsmere PD

Last Agency Inspection: 11/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:35

Gender: Male

Observation Period Began: 01:50

Subject's Last Name: SEMENUCK

Date/Time of Test: 12/02/2007 02:25

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Subject Sample #1	0.096	02:20
Breath Volume	2.156	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample #2	0.086	02:24
Breath Volume	1.343	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 12/06/2007 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PIGGOTT, MYRON, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: MILLS, JASON

Arresting Officer Agency: Sebastian PD

Last Agency Inspection: 11/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 04:15

Gender: Male

Observation Period Began: 05:15

Subject's Last Name: SAVANT

Date/Time of Test: 12/02/2007 05:48

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:36	
Control Test	0.080	05:36	
Air Blank	0.000	05:36	
Subject Sample #1	0.000	05:39	Volume Not Met
Breath Volume	0.277	05:39	
Air Blank	0.000	05:40	
Air Blank	0.000	05:42	
Subject Sample #2	0.000	05:43	
Breath Volume	2.882	05:43	
Air Blank	0.000	05:44	
Air Blank	0.000	05:46	
Subject Sample #3	0.000	05:46	
Breath Volume	1.992	05:46	
Air Blank	0.000	05:47	
Control Test	0.078	05:47	
Air Blank	0.000	05:48	
Diagnostics Check	OK	05:48	

Data Download Date/Time: 12/06/2007 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: MINCEY, SHAWN, L

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: RAHMING, DERRICK

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 11/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 20:26

Gender: Male

Observation Period Began: 20:45

Subject's Last Name: OCONNOR

Date/Time of Test: 12/02/2007 21:31

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.079	21:25
Air Blank	0.000	21:26
Subject Sample #1	0.151	21:26
Breath Volume	1.835	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample #2	0.153	21:29
Breath Volume	1.261	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:31

Data Download Date/Time: 12/06/2007 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: CIRINO, ARTHUR, D

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: RICE, JAMES

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 11/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 19:19

Gender: Male

Observation Period Began: 19:35

Subject's Last Name: DEL AGUILA

Date/Time of Test: 12/05/2007 20:22

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Subject Sample #1	0.145	20:17
Breath Volume	1.738	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample #2	0.147	20:20
Breath Volume	1.468	20:20
Air Blank	0.000	20:21
Control Test	0.078	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 12/06/2007 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: CIRINO, ARTHUR, D

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: SNOWHILL, MIKE

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 16:44

Gender: Male

Observation Period Began: 17:30

Subject's Last Name: DEFOSSSES

Date/Time of Test: 12/06/2007 18:02

Diagnostics Check	OK	17:56
Air Blank	0.000	17:57
Control Test	0.079	17:57
Air Blank	0.000	17:58
Subject Sample #1	0.170	17:58
Breath Volume	2.773	17:58
Air Blank	0.000	17:59
Air Blank	0.000	18:00
Subject Sample #2	0.169	18:01
Breath Volume	2.207	18:01
Air Blank	0.000	18:02
Control Test	0.078	18:02
Air Blank	0.000	18:02
Diagnostics Check	OK	18:02

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: FLORES, DAVID

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:15

Gender: Female

Observation Period Began: 02:03

Subject's Last Name: REALE

Date/Time of Test: 12/07/2007 02:42

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Subject Sample #1	0.091	02:37
Breath Volume	1.566	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.088	02:40
Breath Volume	1.824	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: RUSSELL, DAVID

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:57

Gender: Male

Observation Period Began: 02:25

Subject's Last Name: FORKUSH

Date/Time of Test: 12/07/2007 03:18

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample #1	0.127	03:13
Breath Volume	1.132	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.140	03:16
Breath Volume	1.738	03:16
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: CLARK, JOHN

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:35

Gender: Male

Observation Period Began: 00:55

Subject's Last Name: JONES

Date/Time of Test: 12/08/2007 01:34

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Subject Sample #1		01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Control Test	0.079	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: SHARKEY, CHRISTOPHER, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: OOLEY, JAMES

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:57

Gender: Female

Observation Period Began: 02:10

Subject's Last Name: BOHATCHLIPTAK

Date/Time of Test: 12/08/2007 02:41

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:36
Air Blank	0.000	02:36
Subject Sample #1	0.194	02:37
Breath Volume	1.343	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample #2	0.183	02:40
Breath Volume	1.710	02:40
Air Blank	0.000	02:41
Control Test	0.077	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: OOLEY, JAMES

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 05:18

Gender: Male

Observation Period Began: 05:40

Subject's Last Name: MCCLOUD

Date/Time of Test: 12/08/2007 06:21

Diagnostics Check	OK	06:15
Air Blank	0.000	06:15
Control Test	0.079	06:16
Air Blank	0.000	06:16
Subject Sample #1	0.121	06:17
Breath Volume	3.265	06:17
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample #2	0.115	06:20
Breath Volume	2.082	06:20
Air Blank	0.000	06:20
Control Test	0.079	06:21
Air Blank	0.000	06:21
Diagnostics Check	OK	06:21

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PEEK, STEVE, K

Breath Test Operator Agency: FHP Troop L

Arresting Officer: PEEK, STEVE

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 21:55

Gender: Male

Observation Period Began: 21:55

Subject's Last Name: DIAZ

Date/Time of Test: 12/08/2007 23:00

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.079	22:54
Air Blank	0.000	22:55
Subject Sample #1	0.020	22:55
Breath Volume	2.160	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample #2	0.018	22:58
Breath Volume	4.335	22:58
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: MCCOMBS, SCOTT

Arresting Officer Agency: Fellsmere PD

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 21:50

Gender: Male

Observation Period Began: 00:38

Subject's Last Name: CORTEZ VEGA

Date/Time of Test: 12/09/2007 01:35

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:30
Subject Sample #1	0.181	01:31
Breath Volume	2.441	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample #2	0.170	01:34
Breath Volume	2.683	01:34
Air Blank	0.000	01:34
Control Test	0.077	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PAUL, GAUTIER,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: MCCOY, JASON

Arresting Officer Agency: Sebastian PD

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:23

Gender: Female

Observation Period Began: 01:45

Subject's Last Name: HUTCHISON

Date/Time of Test: 12/09/2007 02:38

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Subject Sample #1	0.205	02:33
Breath Volume	1.347	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample #2	0.205	02:36
Breath Volume	1.144	02:36
Air Blank	0.000	02:37
Control Test	0.077	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: FLORES, DAVID

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:41

Gender: Female

Observation Period Began: 02:00

Subject's Last Name: PARTIN

Date/Time of Test: 12/09/2007 03:15

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Subject Sample #1	0.217	03:06
Breath Volume	1.746	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample #2	0.196	03:09
Breath Volume	1.218	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample #3	0.194	03:14
Breath Volume	1.464	03:14
Air Blank	0.000	03:14
Control Test	0.077	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: MINCEY, SHAWN, L

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: GILLETTE, JASON

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 19:12

Gender: Male

Observation Period Began: 19:30

Subject's Last Name: REHBERGER

Date/Time of Test: 12/09/2007 20:26

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:23	
Control Test	0.079	20:23	
Air Blank	0.000	20:24	
Subject Sample #1		20:25	Subject Test Refused
Breath Volume	0.000	20:25	
Air Blank	0.000	20:25	
Control Test	0.080	20:25	
Air Blank	0.000	20:26	
Diagnostics Check	OK	20:26	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: MINCEY, SHAWN, L

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: CHASE, JASON

Arresting Officer Agency: Fellsmere PD

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 21:47

Gender: Male

Observation Period Began: 22:50

Subject's Last Name: LEON

Date/Time of Test: 12/09/2007 23:20

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.079	23:15	
Air Blank	0.000	23:15	
Subject Sample #1	0.154	23:15	
Breath Volume	2.839	23:15	
Air Blank	0.000	23:16	
Air Blank	0.000	23:18	
Subject Sample #2	0.154	23:18	
Breath Volume	4.023	23:18	
Air Blank	0.000	23:19	
Control Test	0.074	23:19	Control Outside Tolerance
Air Blank	0.000	23:20	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO
Software Version: 8100.27
Breath Test Operator: SHARKEY, CHRISTOPHER, J
Breath Test Operator Agency: Indian River Co SO
Arresting Officer: HARNAGE, SEAN
Arresting Officer Agency: Indian River Co SO
Last Agency Inspection: 12/06/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 714902I
Dry Gas Expiry Date: 05/31/2009
Violation Code: DUI
Arrest Time: 02:15
Gender: Female
Observation Period Began: 02:15

Subject's Last Name: SEILER

Date/Time of Test: 12/11/2007 02:55

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample #1	0.000	02:51
Breath Volume	1.714	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample #2	0.000	02:54
Breath Volume	1.648	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: CIRINO, ARTHUR, D

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: DELANEY, BEN

Arresting Officer Agency: Sebastian PD

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 20:56

Gender: Male

Observation Period Began: 21:20

Subject's Last Name: MCQUILLEN

Date/Time of Test: 12/13/2007 22:10

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:57	
Control Test	0.079	21:57	
Air Blank	0.000	21:58	
Subject Sample #1	0.179	21:58	
Breath Volume	1.976	21:58	
Air Blank	0.000	21:59	
Air Blank	0.000	22:01	
Subject Sample #2		22:04	No Sample Provided
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Air Blank	0.000	22:06	
Subject Sample #3	0.182	22:08	
Breath Volume	1.707	22:08	
Air Blank	0.000	22:09	
Control Test	0.078	22:09	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PIGGOTT, MYRON, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: GARDNER, SCOTT

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 23:58

Gender: Female

Observation Period Began: 00:14

Subject's Last Name: BROOKS

Date/Time of Test: 12/14/2007 00:46

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.078	00:40
Air Blank	0.000	00:40
Subject Sample #1	0.124	00:41
Breath Volume	1.609	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample #2	0.122	00:45
Breath Volume	1.542	00:45
Air Blank	0.000	00:45
Control Test	0.078	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: LIBERATORE, SCOTT, A

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: MEWBORN, PAUL

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 02:50

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: CURTIS

Date/Time of Test: 12/14/2007 03:48

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.078	03:41	
Air Blank	0.000	03:41	
Subject Sample #1		03:44	No Sample Provided
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Air Blank	0.000	03:47	
Subject Sample #2		03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.078	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PEEK, STEVE, K

Breath Test Operator Agency: FHP Troop L

Arresting Officer: PEEK, STEVE

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:08

Gender: Male

Observation Period Began: 01:08

Subject's Last Name: SANGALLI

Date/Time of Test: 12/15/2007 02:00

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.078	01:57	
Air Blank	0.000	01:58	
Subject Sample #1		01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.078	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PIGGOTT, MYRON, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: MEWBORN, PAUL

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:55

Gender: Male

Observation Period Began: 02:15

Subject's Last Name: HARRIS

Date/Time of Test: 12/16/2007 02:54

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.078	02:44	
Air Blank	0.000	02:44	
Subject Sample #1	0.253	02:47	Volume Not Met
Breath Volume	0.667	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample #2		02:52	Subject Test Refused
Breath Volume	0.265	02:52	
Air Blank	0.000	02:53	
Control Test	0.075	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: VALDERRAMA, MELVIN

Arresting Officer Agency: Fellsmere PD

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:55

Gender: Male

Observation Period Began: 02:45

Subject's Last Name: CERECERO

Date/Time of Test: 12/16/2007 03:34

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.078	03:28
Air Blank	0.000	03:29
Subject Sample #1	0.239	03:29
Breath Volume	2.390	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample #2	0.233	03:32
Breath Volume	2.500	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:34

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PEEK, STEVE, K

Breath Test Operator Agency: FHP Troop L

Arresting Officer: PEEK, STEVE

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 21:24

Gender: Male

Observation Period Began: 21:24

Subject's Last Name: CAMPA

Date/Time of Test: 12/16/2007 22:17

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Subject Sample #1	0.025	22:13
Breath Volume	2.640	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample #2	0.024	22:16
Breath Volume	1.878	22:16
Air Blank	0.000	22:16
Control Test	0.079	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: LIBERATORE, SCOTT, A

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: VINCENT, BERNEX

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 23:10

Gender: Female

Observation Period Began: 00:08

Subject's Last Name: CANEDY

Date/Time of Test: 12/17/2007 00:50

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:39	
Subject Sample #1	0.107	00:42	Volume Not Met
Breath Volume	0.269	00:42	
Air Blank	0.000	00:43	
Air Blank	0.000	00:45	
Subject Sample #2	0.136	00:46	
Breath Volume	2.039	00:46	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample #3	0.135	00:48	
Breath Volume	1.515	00:48	
Air Blank	0.000	00:49	
Control Test	0.078	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: SANTIGO, KIMBERLY, A

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: SANTIAGO, KIMBERLY

Arresting Officer Agency: OUT OF STATE

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:00

Gender: Male

Observation Period Began: 10:40

Subject's Last Name: LAURSEN

Date/Time of Test: 12/18/2007 11:06

Diagnostics Check	OK	11:00
Air Blank	0.000	11:01
Control Test	0.080	11:01
Air Blank	0.000	11:02
Subject Sample #1	0.000	11:02
Breath Volume	3.187	11:02
Air Blank	0.000	11:03
Air Blank	0.000	11:05
Subject Sample #2	0.000	11:05
Breath Volume	3.007	11:05
Air Blank	0.000	11:05
Control Test	0.080	11:06
Air Blank	0.000	11:06
Diagnostics Check	OK	11:06

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO
Software Version: 8100.27
Breath Test Operator: MINCEY, SHAWN, L
Breath Test Operator Agency: Indian River Co SO
Arresting Officer: SNOWHILL, MICHAEL
Arresting Officer Agency: Indian River Co SO
Last Agency Inspection: 12/06/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 714902I
Dry Gas Expiry Date: 05/31/2009
Violation Code: DUI
Arrest Time: 19:42
Gender: Male
Observation Period Began: 19:25

Subject's Last Name: HERSHBERGER

Date/Time of Test: 12/18/2007 20:59

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.080	20:53
Air Blank	0.000	20:54
Subject Sample #1	0.000	20:54
Breath Volume	1.707	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample #2	0.000	20:57
Breath Volume	2.152	20:57
Air Blank	0.000	20:58
Control Test	0.080	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: SHARKEY, CHRISTOPHER, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: OOLEY, JAMES

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:22

Gender: Female

Observation Period Began: 01:50

Subject's Last Name: MITCHELL

Date/Time of Test: 12/20/2007 02:17

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Subject Sample #1	0.104	02:13
Breath Volume	1.492	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample #2	0.103	02:16
Breath Volume	1.109	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: CIRINO, ARTHUR, D

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: BARR, STEVEN

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 20:40

Gender: Male

Observation Period Began: 21:35

Subject's Last Name: RIVERA

Date/Time of Test: 12/20/2007 22:12

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.080	22:07
Air Blank	0.000	22:07
Subject Sample #1	0.128	22:08
Breath Volume	1.613	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample #2	0.122	22:11
Breath Volume	1.386	22:11
Air Blank	0.000	22:11
Control Test	0.079	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO
Software Version: 8100.27
Breath Test Operator: GAUTIER, PAUL,
Breath Test Operator Agency: Indian River Co SO
Arresting Officer: GARDNER, SCOTT
Arresting Officer Agency: FHP Troop L
Last Agency Inspection: 12/06/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 714902I
Dry Gas Expiry Date: 05/31/2009
Violation Code: DUI
Arrest Time: 02:15
Gender: Male
Observation Period Began: 02:45

Subject's Last Name: YOUNGERMAN

Date/Time of Test: 12/21/2007 03:49

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample #1	0.272	03:45
Breath Volume	1.488	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample #2	0.265	03:47
Breath Volume	2.195	03:47
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PIGGOTT, MYRON, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: BYRD, TOMMY

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 05:12

Gender: Male

Observation Period Began: 05:30

Subject's Last Name: THORNTON

Date/Time of Test: 12/21/2007 05:56

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Subject Sample #1	0.139	05:51
Breath Volume	1.832	05:51
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample #2	0.137	05:54
Breath Volume	1.695	05:54
Air Blank	0.000	05:55
Control Test	0.079	05:55
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: BRAUER, JEFF, A

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: BRAUER, JEFF

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 10:40

Gender: Male

Observation Period Began: 10:40

Subject's Last Name: YOUNGERMAN

Date/Time of Test: 12/21/2007 11:11

Diagnostics Check	OK	11:04
Air Blank	0.000	11:05
Control Test	0.080	11:05
Air Blank	0.000	11:06
Subject Sample #1	0.103	11:06
Breath Volume	1.312	11:06
Air Blank	0.000	11:07
Air Blank	0.000	11:09
Subject Sample #2	0.107	11:09
Breath Volume	2.757	11:09
Air Blank	0.000	11:10
Control Test	0.078	11:10
Air Blank	0.000	11:11
Diagnostics Check	OK	11:11

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: BRAUER, JEFF, A

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: BRAUER, JEFF

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 12:30

Gender: Male

Observation Period Began: 12:30

Subject's Last Name: YOUNGERMAN

Date/Time of Test: 12/21/2007 12:58

Diagnostics Check	OK	12:52
Air Blank	0.000	12:53
Control Test	0.078	12:53
Air Blank	0.000	12:54
Subject Sample #1	0.075	12:54
Breath Volume	2.066	12:54
Air Blank	0.000	12:55
Air Blank	0.000	12:57
Subject Sample #2	0.067	12:57
Breath Volume	1.988	12:57
Air Blank	0.000	12:57
Control Test	0.079	12:58
Air Blank	0.000	12:58
Diagnostics Check	OK	12:58

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: YOUNGERMAN, GEORGE, W

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: LAVENDER, RANDALL

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 14:40

Subject's Last Name: YOUNGERMAN

Date/Time of Test: 12/21/2007 15:17

Diagnostics Check	OK	15:10
Air Blank	0.000	15:10
Control Test	0.079	15:11
Air Blank	0.000	15:11
Subject Sample #1	0.027	15:12
Breath Volume	1.949	15:12
Air Blank	0.000	15:12
Air Blank	0.000	15:14
Subject Sample #2	0.023	15:15
Breath Volume	2.210	15:15
Air Blank	0.000	15:16
Control Test	0.078	15:16
Air Blank	0.000	15:17
Diagnostics Check	OK	15:17

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: LAVENDER, RANDALL, W

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: LAVENDER, RANDALL

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 14:45

Subject's Last Name: YOUNGERMAN

Date/Time of Test: 12/21/2007 15:32

Diagnostics Check	OK	15:25
Air Blank	0.000	15:25
Control Test	0.079	15:26
Air Blank	0.000	15:26
Subject Sample #1	0.019	15:28
Breath Volume	1.894	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample #2	0.019	15:31
Breath Volume	2.578	15:31
Air Blank	0.000	15:31
Control Test	0.078	15:31
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO
Software Version: 8100.27
Breath Test Operator: MINCEY, SHAWN, L
Breath Test Operator Agency: Indian River Co SO
Arresting Officer: OOLEY, JAMES
Arresting Officer Agency: Indian River Co SO
Last Agency Inspection: 12/06/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 714902I
Dry Gas Expiry Date: 05/31/2009
Violation Code: DUI
Arrest Time: 17:52
Gender: Male
Observation Period Began: 18:40

Subject's Last Name: BAILEY

Date/Time of Test: 12/21/2007 19:12

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:09	
Control Test	0.080	19:10	
Air Blank	0.000	19:10	
Subject Sample #1		19:10	Subject Test Refused
Breath Volume	0.000	19:10	
Air Blank	0.000	19:11	
Control Test	0.080	19:11	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: SHARKEY, CHRISTOPHER, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: GARDNER, SCOTT

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:05

Gender: Female

Observation Period Began: 00:38

Subject's Last Name: MEYERS

Date/Time of Test: 12/22/2007 01:07

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample #1	0.089	01:03
Breath Volume	2.285	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample #2	0.092	01:06
Breath Volume	2.324	01:06
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO
Software Version: 8100.27
Breath Test Operator: PIGGOTT, MYRON, J
Breath Test Operator Agency: Indian River Co SO
Arresting Officer: OOLEY, JAMES
Arresting Officer Agency: Indian River Co SO
Last Agency Inspection: 12/06/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 714902I
Dry Gas Expiry Date: 05/31/2009
Violation Code: DUI
Arrest Time: 02:29
Gender: Female
Observation Period Began: 04:20

Subject's Last Name: RHODES

Date/Time of Test: 12/22/2007 04:58

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:50	
Control Test	0.080	04:51	
Air Blank	0.000	04:51	
Subject Sample #1		04:54	No Sample Provided
Breath Volume	0.000	04:54	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample #2		04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:57	
Control Test	0.080	04:58	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: SHARKEY, CHRISTOPHER, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: GARDNER, SCOTT

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:39

Gender: Male

Observation Period Began: 04:32

Subject's Last Name: SANDS

Date/Time of Test: 12/22/2007 05:24

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:14	
Control Test	0.079	05:15	
Air Blank	0.000	05:15	
Subject Sample #1	0.083	05:18	Volume Not Met
Breath Volume	0.050	05:18	
Air Blank	0.000	05:19	
Air Blank	0.000	05:21	
Subject Sample #2		05:22	Subject Test Refused
Breath Volume	0.554	05:22	
Air Blank	0.000	05:23	
Control Test	0.079	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PIGGOTT, MYRON, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: BONVIE, ROBERT

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 05:00

Gender: Male

Observation Period Began: 05:15

Subject's Last Name: MANZI

Date/Time of Test: 12/22/2007 06:07

Diagnostics Check	OK	06:01
Air Blank	0.000	06:02
Control Test	0.079	06:02
Air Blank	0.000	06:02
Subject Sample #1	0.156	06:03
Breath Volume	2.066	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:05
Subject Sample #2	0.156	06:06
Breath Volume	2.347	06:06
Air Blank	0.000	06:06
Control Test	0.079	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: CHASE, JASON

Arresting Officer Agency: Fellsmere PD

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 23:09

Gender: Male

Observation Period Began: 00:10

Subject's Last Name: CRUZ

Date/Time of Test: 12/23/2007 00:52

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.078	00:47
Air Blank	0.000	00:47
Subject Sample #1	0.135	00:48
Breath Volume	1.949	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample #2	0.135	00:51
Breath Volume	1.875	00:51
Air Blank	0.000	00:51
Control Test	0.078	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO
Software Version: 8100.27
Breath Test Operator: PIGGOTT, MYRON, J
Breath Test Operator Agency: Indian River Co SO
Arresting Officer: GARDNER, SCOTT
Arresting Officer Agency: FHP Troop L
Last Agency Inspection: 12/06/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 714902I
Dry Gas Expiry Date: 05/31/2009
Violation Code: DUI
Arrest Time: 02:25
Gender: Male
Observation Period Began: 03:00

Subject's Last Name: BAXTER

Date/Time of Test: 12/23/2007 03:27

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:22
Air Blank	0.000	03:22
Subject Sample #1	0.135	03:23
Breath Volume	1.589	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample #2	0.131	03:26
Breath Volume	1.652	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: SIMONS, NORA, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: SWEETING, TERRI

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 14:19

Gender: Male

Observation Period Began: 14:49

Subject's Last Name: CHAHIN

Date/Time of Test: 12/23/2007 15:18

Diagnostics Check	OK	15:10
Air Blank	0.000	15:10
Control Test	0.080	15:10
Air Blank	0.000	15:11
Subject Sample #1	0.211	15:12
Breath Volume	1.234	15:12
Air Blank	0.000	15:13
Air Blank	0.000	15:15
Subject Sample #2	0.202	15:16
Breath Volume	1.371	15:16
Air Blank	0.000	15:17
Control Test	0.078	15:17
Air Blank	0.000	15:17
Diagnostics Check	OK	15:18

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: LIBERATORE, SCOTT, A

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: LIBERATORE, SCOTT

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 07:10

Gender: Female

Observation Period Began: 07:10

Subject's Last Name: PIOTROWSKI

Date/Time of Test: 12/24/2007 07:40

Diagnostics Check	OK	07:34
Air Blank	0.000	07:34
Control Test	0.079	07:35
Air Blank	0.000	07:35
Subject Sample #1	0.000	07:35
Breath Volume	2.246	07:35
Air Blank	0.000	07:36
Air Blank	0.000	07:38
Subject Sample #2	0.000	07:38
Breath Volume	2.382	07:38
Air Blank	0.000	07:39
Control Test	0.078	07:39
Air Blank	0.000	07:40
Diagnostics Check	OK	07:40

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PEEK, STEVE, K

Breath Test Operator Agency: FHP Troop L

Arresting Officer: PEEK, STEVE

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 20:14

Gender: Male

Observation Period Began: 20:14

Subject's Last Name: LUCCA

Date/Time of Test: 12/25/2007 21:39

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:35
Subject Sample #1	0.000	21:35
Breath Volume	4.660	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample #2	0.000	21:38
Breath Volume	4.238	21:38
Air Blank	0.000	21:38
Control Test	0.079	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: MINCEY, SHAWN, L

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: VINCENT, BERNEX

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 22:52

Gender: Female

Observation Period Began: 23:05

Subject's Last Name: PRIESTER

Date/Time of Test: 12/25/2007 23:39

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:32
Subject Sample #1	0.127	23:34
Breath Volume	1.492	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample #2	0.125	23:37
Breath Volume	1.464	23:37
Air Blank	0.000	23:38
Control Test	0.078	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PIGGOTT, MYRON, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: GARDNER, SCOTT

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 00:44

Gender: Male

Observation Period Began: 01:20

Subject's Last Name: LONG-JONES

Date/Time of Test: 12/26/2007 01:51

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.079	01:49	
Air Blank	0.000	01:50	
Subject Sample #1		01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Control Test	0.080	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: MILLS, JASON

Arresting Officer Agency: Sebastian PD

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 04:10

Gender: Male

Observation Period Began: 04:10

Subject's Last Name: BROWNHILL

Date/Time of Test: 12/26/2007 05:43

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:41	
Control Test	0.080	05:41	
Air Blank	0.000	05:41	
Subject Sample #1		05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:42	
Control Test	0.079	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PIGGOTT, MYRON, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: GARDNER, SCOTT

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:31

Gender: Female

Observation Period Began: 01:31

Subject's Last Name: WALKER

Date/Time of Test: 12/27/2007 02:22

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Subject Sample #1	0.196	02:17
Breath Volume	1.582	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample #2	0.194	02:20
Breath Volume	1.406	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: CIRINO, ARTHUR, D

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: BYRD, TOMMY

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 19:21

Gender: Female

Observation Period Began: 20:20

Subject's Last Name: SHARPIN

Date/Time of Test: 12/27/2007 20:50

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:48	
Control Test	0.078	20:48	
Air Blank	0.000	20:49	
Subject Sample #1		20:49	Subject Test Refused
Breath Volume	0.000	20:49	
Air Blank	0.000	20:49	
Control Test	0.078	20:50	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: GARDNER, SCOTT

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 23:13

Gender: Male

Observation Period Began: 00:15

Subject's Last Name: DESPORTES

Date/Time of Test: 12/28/2007 00:43

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.078	00:36
Air Blank	0.000	00:37
Subject Sample #1	0.177	00:38
Breath Volume	1.300	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample #2	0.195	00:41
Breath Volume	1.234	00:41
Air Blank	0.000	00:42
Control Test	0.076	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:43

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GAUTIER, PAUL,

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: GARDNER, SCOTT

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 02:15

Gender: Male

Observation Period Began: 02:55

Subject's Last Name: STURGIS

Date/Time of Test: 12/28/2007 03:30

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:23
Subject Sample #1	0.131	03:25
Breath Volume	1.734	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample #2	0.129	03:28
Breath Volume	1.128	03:28
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: GOODMAN, ELLA, E

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: GOODMAN, ELLA

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 08:15

Gender: Female

Observation Period Began: 08:15

Subject's Last Name: ;6360323623026556=11

Date/Time of Test: 12/28/2007 08:51

Diagnostics Check	OK	08:45
Air Blank	0.000	08:45
Control Test	0.079	08:45
Air Blank	0.000	08:46
Subject Sample #1	0.000	08:46
Breath Volume	1.500	08:46
Air Blank	0.000	08:47
Air Blank	0.000	08:49
Subject Sample #2	0.000	08:50
Breath Volume	1.539	08:50
Air Blank	0.000	08:50
Control Test	0.079	08:50
Air Blank	0.000	08:51
Diagnostics Check	OK	08:51

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO
Software Version: 8100.27
Breath Test Operator: GAUTIER, PAUL,
Breath Test Operator Agency: Indian River Co SO
Arresting Officer: BONVIE, ROBERT
Arresting Officer Agency: Indian River Co SO
Last Agency Inspection: 12/06/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 714902I
Dry Gas Expiry Date: 05/31/2009
Violation Code: DUI
Arrest Time: 01:10
Gender: Male
Observation Period Began: 01:25

Subject's Last Name: JURCZYK

Date/Time of Test: 12/29/2007 02:11

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:05
Air Blank	0.000	02:05
Subject Sample #1	0.155	02:06
Breath Volume	1.468	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #2	0.147	02:09
Breath Volume	1.281	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: SHARKEY, CHRISTOPHER, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: BONVIE, ROBERT

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 04:23

Gender: Male

Observation Period Began: 04:33

Subject's Last Name: MCDONAGH-CROSS

Date/Time of Test: 12/29/2007 05:02

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:57
Subject Sample #1	0.195	04:57
Breath Volume	2.101	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample #2	0.195	05:00
Breath Volume	1.761	05:00
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: MINCEY, SHAWN, L

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: BONVIE, ROBERT

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 19:29

Gender: Male

Observation Period Began: 19:55

Subject's Last Name: MIDDLETON

Date/Time of Test: 12/29/2007 20:26

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:24	
Control Test	0.080	20:24	
Air Blank	0.000	20:25	
Subject Sample #1		20:25	Subject Test Refused
Breath Volume	0.000	20:25	
Air Blank	0.000	20:25	
Control Test	0.079	20:26	
Air Blank	0.000	20:26	
Diagnostics Check	OK	20:26	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PIGGOTT, MYRON, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: GARDNER, SCOTT

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 23:48

Gender: Male

Observation Period Began: 00:30

Subject's Last Name: WIRTH

Date/Time of Test: 12/30/2007 01:17

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.079	01:15	
Air Blank	0.000	01:16	
Subject Sample #1		01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.079	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001328

Instrument Registered To: INDIAN RIVER CO SO

Software Version: 8100.27

Breath Test Operator: PIGGOTT, MYRON, J

Breath Test Operator Agency: Indian River Co SO

Arresting Officer: BYRD, TOMMY

Arresting Officer Agency: Indian River Co SO

Last Agency Inspection: 12/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714902I

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:54

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: PRETORIUS

Date/Time of Test: 12/30/2007 04:43

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.079	04:41	
Air Blank	0.000	04:41	
Subject Sample #1		04:41	Subject Test Refused
Breath Volume	0.000	04:41	
Air Blank	0.000	04:42	
Control Test	0.079	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 01/07/2008 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001341

Instrument Registered To: DESOTO COUNTY SO

Software Version: 8100.27

Breath Test Operator: GLEOCKLER, LARRY, G

Breath Test Operator Agency: Arcadia PD

Arresting Officer: JONATHAN, CALEB

Arresting Officer Agency: Arcadia PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 23:37

Gender: Male

Observation Period Began: 23:45

Subject's Last Name: PRICE

Date/Time of Test: 12/08/2007 00:21

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.078	00:16
Air Blank	0.000	00:16
Subject Sample #1	0.179	00:17
Breath Volume	1.957	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:19
Subject Sample #2	0.178	00:20
Breath Volume	2.074	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 12/28/2007 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001341

Instrument Registered To: DESOTO COUNTY SO

Software Version: 8100.27

Breath Test Operator: RAIN, BENTON, C

Breath Test Operator Agency: DeSoto County SO

Arresting Officer: CLOUD, JAMES

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 18:51

Gender: Male

Observation Period Began: 19:03

Subject's Last Name: MARQUEZ

Date/Time of Test: 12/09/2007 19:36

Diagnostics Check	OK	19:29
Air Blank	0.000	19:30
Control Test	0.078	19:30
Air Blank	0.000	19:31
Subject Sample #1	0.205	19:31
Breath Volume	2.765	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample #2	0.202	19:34
Breath Volume	1.910	19:34
Air Blank	0.000	19:35
Control Test	0.078	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:36

Data Download Date/Time: 12/28/2007 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001341

Instrument Registered To: DESOTO COUNTY SO

Software Version: 8100.27

Breath Test Operator: RAIN, BENTON, C

Breath Test Operator Agency: DeSoto County SO

Arresting Officer: MARCUS, MARK

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 01:29

Gender: Female

Observation Period Began: 01:49

Subject's Last Name: WEEDEN

Date/Time of Test: 12/24/2007 02:33

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Subject Sample #1	0.111	02:26
Breath Volume	1.652	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample #2	0.109	02:31
Breath Volume	1.347	02:31
Air Blank	0.000	02:32
Control Test	0.077	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 12/28/2007 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001341

Instrument Registered To: DESOTO COUNTY SO

Software Version: 8100.27

Breath Test Operator: WOODS, ROBERT, L

Breath Test Operator Agency: Arcadia PD

Arresting Officer: WOODS, ROBERT

Arresting Officer Agency: Arcadia PD

Last Agency Inspection: 12/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 22:10

Gender: Male

Observation Period Began: 22:15

Subject's Last Name: CASTRO

Date/Time of Test: 12/29/2007 22:50

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.079	22:47	
Air Blank	0.000	22:47	
Subject Sample #1		22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:49	
Control Test	0.079	22:49	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 01/25/2008 07:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001345

Instrument Registered To: EUSTIS P.D.

Software Version: 8100.27

Breath Test Operator: CHAPPELL, MELISSA, A

Breath Test Operator Agency: Eustis PD

Arresting Officer: HUGHES, HAROLD

Arresting Officer Agency: Eustis PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 00:55

Gender: Female

Observation Period Began: 01:10

Subject's Last Name: STEVERSON

Date/Time of Test: 12/08/2007 01:46

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample #1	0.219	01:41
Breath Volume	1.144	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample #2	0.216	01:44
Breath Volume	1.484	01:44
Air Blank	0.000	01:45
Control Test	0.077	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 12/30/2007 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001345

Instrument Registered To: EUSTIS P.D.

Software Version: 8100.27

Breath Test Operator: BANNER, JEAN, A

Breath Test Operator Agency: Eustis PD

Arresting Officer: JOHNSON, CHRIS

Arresting Officer Agency: Eustis PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 11:08

Gender: Male

Observation Period Began: 11:25

Subject's Last Name: SPENCER

Date/Time of Test: 12/08/2007 12:03

Diagnostics Check	OK	11:57
Air Blank	0.000	11:57
Control Test	0.080	11:57
Air Blank	0.000	11:58
Subject Sample #1	0.167	11:58
Breath Volume	1.496	11:58
Air Blank	0.000	11:59
Air Blank	0.000	12:01
Subject Sample #2	0.168	12:01
Breath Volume	1.628	12:01
Air Blank	0.000	12:02
Control Test	0.079	12:02
Air Blank	0.000	12:03
Diagnostics Check	OK	12:03

Data Download Date/Time: 12/30/2007 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001345

Instrument Registered To: EUSTIS P.D.

Software Version: 8100.27

Breath Test Operator: CHAPPELL, MELISSA, A

Breath Test Operator Agency: Eustis PD

Arresting Officer: KOLLER, J

Arresting Officer Agency: Eustis PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 19:29

Gender: Male

Observation Period Began: 19:45

Subject's Last Name: MADRID

Date/Time of Test: 12/09/2007 20:46

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.077	20:40
Air Blank	0.000	20:41
Subject Sample #1	0.151	20:42
Breath Volume	2.218	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample #2	0.155	20:45
Breath Volume	2.292	20:45
Air Blank	0.000	20:45
Control Test	0.078	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 12/30/2007 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001345

Instrument Registered To: EUSTIS P.D.

Software Version: 8100.27

Breath Test Operator: GRAHAM, THERESA, S

Breath Test Operator Agency: Eustis PD

Arresting Officer: HORST, CHRIS

Arresting Officer Agency: Eustis PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7141011

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 19:36

Gender: Male

Observation Period Began: 20:25

Subject's Last Name: RICCELLI

Date/Time of Test: 12/19/2007 21:30

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:28	
Control Test	0.080	21:28	
Air Blank	0.000	21:29	
Subject Sample #1		21:29	Subject Test Refused
Breath Volume	0.000	21:29	
Air Blank	0.000	21:29	
Control Test	0.079	21:30	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 12/30/2007 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VENTURI, GEORGE, L

Breath Test Operator Agency: FDLE

Arresting Officer: TEST, GEORGE

Arresting Officer Agency: FDLE

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: SYS CHECK

Arrest Time: 13:00

Gender: Male

Observation Period Began: 13:00

Subject's Last Name: TEST

Date/Time of Test: 12/04/2007 14:04

Diagnostics Check	OK	13:57
Air Blank	0.000	13:58
Control Test	0.079	13:58
Air Blank	0.000	13:58
Subject Sample #1	0.085	13:59
Breath Volume	2.230	13:59
Air Blank	0.000	14:00
Air Blank	0.000	14:02
Subject Sample #2	0.082	14:02
Breath Volume	1.464	14:02
Air Blank	0.000	14:03
Control Test	0.079	14:03
Air Blank	0.000	14:03
Diagnostics Check	OK	14:04

Data Download Date/Time: 12/04/2007 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VENTURI, GEORGE, L

Breath Test Operator Agency: FDLE

Arresting Officer: TEST, GEORGE

Arresting Officer Agency: FDLE

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: SYS CHECK

Arrest Time: 13:30

Gender: Male

Observation Period Began: 13:30

Subject's Last Name: TEST

Date/Time of Test: 12/04/2007 14:18

Diagnostics Check	OK	14:10
Air Blank	0.000	14:11
Control Test	0.080	14:11
Air Blank	0.000	14:12
Subject Sample #1	0.083	14:12
Breath Volume	2.886	14:12
Air Blank	0.000	14:13
Air Blank	0.000	14:15
Subject Sample #2	0.082	14:16
Breath Volume	2.828	14:16
Air Blank	0.000	14:17
Control Test	0.079	14:17
Air Blank	0.000	14:18
Diagnostics Check	OK	14:18

Data Download Date/Time: 12/04/2007 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: FLINT, CHARLES

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 00:05

Gender: Female

Observation Period Began: 00:05

Subject's Last Name: OLMSTEAD

Date/Time of Test: 12/22/2007 00:33

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Subject Sample #1	0.061	00:28
Breath Volume	1.371	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample #2	0.058	00:32
Breath Volume	1.449	00:32
Air Blank	0.000	00:32
Control Test	0.078	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: PINTI, JENNIFER

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 00:18

Gender: Male

Observation Period Began: 00:18

Subject's Last Name: FRESSILLI

Date/Time of Test: 12/22/2007 00:49

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.078	00:42
Air Blank	0.000	00:43
Subject Sample #1	0.098	00:44
Breath Volume	2.023	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample #2	0.100	00:47
Breath Volume	2.531	00:47
Air Blank	0.000	00:48
Control Test	0.078	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:49

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: BRANDAU, STACEY

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 00:23

Gender: Female

Observation Period Began: 00:30

Subject's Last Name: TABER

Date/Time of Test: 12/22/2007 01:08

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:02	
Control Test	0.079	01:02	
Air Blank	0.000	01:03	
Subject Sample #1	0.199	01:04	
Breath Volume	2.742	01:04	
Air Blank	0.000	01:04	
Air Blank	0.000	01:06	
Subject Sample #2		01:06	RFI Detect
Breath Volume	0.000	01:06	
Air Blank	0.000	01:07	
Control Test	0.078	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: BRANDAU, STACEY

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 00:23

Gender: Female

Observation Period Began: 00:30

Subject's Last Name: TABER

Date/Time of Test: 12/22/2007 01:15

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Subject Sample #1	0.199	01:11
Breath Volume	2.734	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample #2	0.194	01:14
Breath Volume	2.671	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: BALES, TIM

Arresting Officer Agency: Sarasota PD

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 01:10

Gender: Male

Observation Period Began: 01:12

Subject's Last Name: CHAVES

Date/Time of Test: 12/22/2007 01:38

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.079	01:36	
Air Blank	0.000	01:36	
Subject Sample #1		01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: KING, OFFICER

Arresting Officer Agency: Sarasota PD

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 01:14

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: BLOOM

Date/Time of Test: 12/22/2007 01:54

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Subject Sample #1	0.199	01:50
Breath Volume	4.984	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample #2	0.213	01:53
Breath Volume	4.675	01:53
Air Blank	0.000	01:53
Control Test	0.078	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: BRANDAU, STACEY

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 01:37

Gender: Male

Observation Period Began: 01:37

Subject's Last Name: BENNETT

Date/Time of Test: 12/22/2007 02:08

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Subject Sample #1	0.114	02:04
Breath Volume	3.136	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample #2	0.127	02:07
Breath Volume	3.585	02:07
Air Blank	0.000	02:08
Control Test	0.078	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: PINTI, JENNIFER

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 01:54

Gender: Male

Observation Period Began: 01:55

Subject's Last Name: AGBAGA

Date/Time of Test: 12/22/2007 02:22

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Subject Sample #1	0.120	02:17
Breath Volume	1.371	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample #2	0.126	02:20
Breath Volume	2.386	02:20
Air Blank	0.000	02:21
Control Test	0.077	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: JOHNSON, RICHARD

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 00:50

Gender: Male

Observation Period Began: 01:53

Subject's Last Name: EASTERLIN

Date/Time of Test: 12/22/2007 02:34

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Subject Sample #1	0.090	02:30
Breath Volume	3.085	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample #2	0.092	02:33
Breath Volume	2.425	02:33
Air Blank	0.000	02:34
Control Test	0.078	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: FLINT, CHARLES

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 02:01

Gender: Male

Observation Period Began: 02:10

Subject's Last Name: DAIGLE

Date/Time of Test: 12/22/2007 02:50

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Subject Sample #1	0.145	02:45
Breath Volume	2.093	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample #2	0.140	02:48
Breath Volume	2.664	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: NICK, AARON

Arresting Officer Agency: North Port PD

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 02:31

Gender: Male

Observation Period Began: 02:32

Subject's Last Name: NEWSTEAD

Date/Time of Test: 12/22/2007 03:00

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:57	
Subject Sample #1		02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: IVINGS, DEPUTY

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 22:28

Gender: Male

Observation Period Began: 22:47

Subject's Last Name: HACKNEY

Date/Time of Test: 12/28/2007 23:14

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Subject Sample #1	0.223	23:09
Breath Volume	2.351	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample #2	0.227	23:12
Breath Volume	2.429	23:12
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:14

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: BRANDAU, STACEY

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 23:53

Gender: Male

Observation Period Began: 23:56

Subject's Last Name: PHILLIPS

Date/Time of Test: 12/29/2007 00:38

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.078	00:23	
Air Blank	0.000	00:23	
Subject Sample #1	0.193	00:26	Volume Not Met
Breath Volume	0.062	00:26	
Air Blank	0.000	00:27	
Air Blank	0.000	00:29	
Subject Sample #2	0.234	00:31	
Breath Volume	1.335	00:31	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample #3	0.215	00:36	Volume Not Met
Breath Volume	0.378	00:36	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BUTLER, CHRIS, T

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: BRANDAU, STACEY

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 23:53

Gender: Male

Observation Period Began: 23:56

Subject's Last Name: PHILLIPS

Date/Time of Test: 12/29/2007 00:52

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.077	00:40	
Air Blank	0.000	00:41	
Subject Sample #1	0.235	00:42	
Breath Volume	1.179	00:42	
Air Blank	0.000	00:43	
Air Blank	0.000	00:45	
Subject Sample #2	0.228	00:48	Volume Not Met
Breath Volume	0.019	00:48	
Air Blank	0.000	00:48	
Air Blank	0.000	00:50	
Subject Sample #3		00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Control Test	0.076	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: FRANKS, SIMON, A

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: ROMAN, CHRIS

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 21:01

Gender: Male

Observation Period Began: 21:02

Subject's Last Name: GREGSTON

Date/Time of Test: 12/29/2007 21:30

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.078	21:24
Air Blank	0.000	21:24
Subject Sample #1	0.306	21:25
Breath Volume	2.390	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample #2	0.319	21:28
Breath Volume	3.742	21:28
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: FRANKS, SIMON, A

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: BRANDAU, STACEY

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 22:51

Gender: Female

Observation Period Began: 22:52

Subject's Last Name: GAYLORD

Date/Time of Test: 12/29/2007 23:19

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.077	23:13
Air Blank	0.000	23:14
Subject Sample #1	0.240	23:15
Breath Volume	2.585	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample #2	0.243	23:18
Breath Volume	1.738	23:18
Air Blank	0.000	23:18
Control Test	0.078	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: FRANKS, SIMON, A

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: ROMAN, CHRIS

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 00:18

Gender: Male

Observation Period Began: 00:23

Subject's Last Name: CARRILLO

Date/Time of Test: 12/30/2007 00:50

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.078	00:44	
Air Blank	0.000	00:45	
Subject Sample #1	0.190	00:46	
Breath Volume	2.375	00:46	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample #2	0.185	00:49	
Breath Volume	2.273	00:49	
Air Blank		00:50	RFI Detect
Air Blank	0.000	00:50	

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: FRANKS, SIMON, A

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: ROMAN, CHRIS

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 00:18

Gender: Male

Observation Period Began: 00:50

Subject's Last Name: CARRILLO

Date/Time of Test: 12/30/2007 01:17

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:11
Subject Sample #1	0.193	01:12
Breath Volume	2.242	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample #2	0.189	01:16
Breath Volume	3.335	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: FRANKS, SIMON, A

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: BRANDAU, STACEY

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 02:13

Gender: Male

Observation Period Began: 02:14

Subject's Last Name: MAGZINO

Date/Time of Test: 12/30/2007 02:41

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.078	02:38	
Air Blank	0.000	02:38	
Subject Sample #1		02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Control Test	0.077	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001346

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BRAYLEY, HARRY, B

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: BRANDAU, STACEY

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 02:38

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: SERRANO QUAN TIP

Date/Time of Test: 12/31/2007 03:07

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:01
Subject Sample #1	0.103	03:02
Breath Volume	3.160	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample #2	0.099	03:05
Breath Volume	3.605	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:07

Data Download Date/Time: 01/07/2008 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001347

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VENTURI, GEORGE, L

Breath Test Operator Agency: FDLE

Arresting Officer: TEST, GEORGE

Arresting Officer Agency: FDLE

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: SYS CHECK

Arrest Time: 13:00

Gender: Male

Observation Period Began: 13:00

Subject's Last Name: TEST

Date/Time of Test: 12/04/2007 14:01

Diagnostics Check	OK	13:53
Air Blank	0.000	13:53
Control Test	0.078	13:53
Air Blank	0.000	13:54
Subject Sample #1	0.083	13:55
Breath Volume	1.730	13:55
Air Blank	0.000	13:55
Air Blank	0.000	13:57
Subject Sample #2	0.081	13:59
Breath Volume	1.953	13:59
Air Blank	0.000	14:00
Control Test	0.078	14:00
Air Blank	0.000	14:01
Diagnostics Check	OK	14:01

Data Download Date/Time: 12/04/2007 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001347

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VENTURI, GEORGE, L

Breath Test Operator Agency: FDLE

Arresting Officer: TEST, GEORGE

Arresting Officer Agency: FDLE

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: SYS CHECK

Arrest Time: 13:30

Gender: Male

Observation Period Began: 13:30

Subject's Last Name: TEST

Date/Time of Test: 12/04/2007 14:16

Diagnostics Check	OK	14:09
Air Blank	0.000	14:10
Control Test	0.078	14:10
Air Blank	0.000	14:11
Subject Sample #1	0.081	14:11
Breath Volume	1.535	14:11
Air Blank	0.000	14:12
Air Blank	0.000	14:14
Subject Sample #2	0.080	14:15
Breath Volume	1.500	14:15
Air Blank	0.000	14:15
Control Test	0.078	14:16
Air Blank	0.000	14:16
Diagnostics Check	OK	14:16

Data Download Date/Time: 12/04/2007 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001348

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: PINTI, JENNIFER, K

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: PINTI, JENNIFER

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 11/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 01:23

Gender: Female

Observation Period Began: 01:38

Subject's Last Name: KIRMO

Date/Time of Test: 12/01/2007 02:05

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Subject Sample #1	0.237	02:00
Breath Volume	2.312	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample #2	0.217	02:03
Breath Volume	1.171	02:03
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 12/04/2007 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001348

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: BRANDAU, STACY, L

Breath Test Operator Agency: Sarasota County SO

Arresting Officer: BRANDAU, STACY

Arresting Officer Agency: Sarasota County SO

Last Agency Inspection: 11/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: DUI

Arrest Time: 01:18

Gender: Male

Observation Period Began: 01:25

Subject's Last Name: SOTOMAYOR-ESTARALLAS

Date/Time of Test: 12/02/2007 01:52

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample #1	0.155	01:48
Breath Volume	3.203	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample #2	0.153	01:51
Breath Volume	2.765	01:51
Air Blank	0.000	01:51
Control Test	0.078	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 12/04/2007 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001348

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VENTURI, GEORGE, L

Breath Test Operator Agency: FDLE

Arresting Officer: TEST, GEORGE

Arresting Officer Agency: FDLE

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: SYS CHECK

Arrest Time: 13:00

Gender: Male

Observation Period Began: 13:00

Subject's Last Name: TEST

Date/Time of Test: 12/04/2007 13:59

Diagnostics Check	OK	13:51
Air Blank	0.000	13:51
Control Test	0.081	13:51
Air Blank	0.000	13:52
Subject Sample #1	0.080	13:53
Breath Volume	1.515	13:53
Air Blank	0.000	13:54
Air Blank	0.000	13:56
Subject Sample #2	0.080	13:58
Breath Volume	1.792	13:58
Air Blank	0.000	13:59
Control Test	0.082	13:59
Air Blank	0.000	13:59
Diagnostics Check	OK	13:59

Data Download Date/Time: 12/04/2007 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001348

Instrument Registered To: SARASOTA COUNTY SO

Software Version: 8100.27

Breath Test Operator: VENTURI, GEORGE, L

Breath Test Operator Agency: FDLE

Arresting Officer: TEST, TEST

Arresting Officer Agency: FDLE

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 608301G

Dry Gas Expiry Date: 03/27/2008

Violation Code: SYS CHECK

Arrest Time: 13:30

Gender: Male

Observation Period Began: 13:30

Subject's Last Name: TEST

Date/Time of Test: 12/04/2007 14:15

Diagnostics Check	OK	14:07
Air Blank	0.000	14:08
Control Test	0.081	14:08
Air Blank	0.000	14:09
Subject Sample #1	0.079	14:10
Breath Volume	1.773	14:10
Air Blank	0.000	14:11
Air Blank	0.000	14:12
Subject Sample #2	0.078	14:13
Breath Volume	1.652	14:13
Air Blank	0.000	14:14
Control Test	0.081	14:14
Air Blank	0.000	14:14
Diagnostics Check	OK	14:15

Data Download Date/Time: 12/04/2007 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001350

Instrument Registered To: FRUITLAND PARK P.D.

Software Version: 8100.27

Breath Test Operator: MATEY, JOHN, P

Breath Test Operator Agency: Fruitland Park PD

Arresting Officer: MATEY, JOHN

Arresting Officer Agency: Fruitland Park PD

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:52

Gender: Male

Observation Period Began: 01:25

Subject's Last Name: CORTEZ

Date/Time of Test: 12/08/2007 02:43

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.082	02:36
Air Blank	0.000	02:36
Subject Sample #1	0.138	02:38
Breath Volume	1.558	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample #2	0.139	02:41
Breath Volume	1.265	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:43

Data Download Date/Time: 01/11/2008 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001350

Instrument Registered To: FRUITLAND PARK P.D.

Software Version: 8100.27

Breath Test Operator: MATEY, JOHN, P

Breath Test Operator Agency: Fruitland Park PD

Arresting Officer: DESANTIS, JEFF

Arresting Officer Agency: Lake County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: JAMES

Date/Time of Test: 12/13/2007 02:12

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample #1	0.118	02:07
Breath Volume	1.527	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample #2	0.104	02:10
Breath Volume	1.652	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 01/11/2008 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001350

Instrument Registered To: FRUITLAND PARK P.D.

Software Version: 8100.27

Breath Test Operator: MATEY, JOHN, P

Breath Test Operator Agency: Fruitland Park PD

Arresting Officer: MATEY, JOHN

Arresting Officer Agency: Fruitland Park PD

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 22:23

Gender: Male

Observation Period Began: 22:15

Subject's Last Name: FLORES VENTURA

Date/Time of Test: 12/13/2007 23:15

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.082	23:08
Air Blank	0.000	23:09
Subject Sample #1	0.136	23:10
Breath Volume	2.753	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample #2	0.141	23:13
Breath Volume	2.847	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 01/11/2008 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001350

Instrument Registered To: FRUITLAND PARK P.D.

Software Version: 8100.27

Breath Test Operator: MATEY, JOHN, P

Breath Test Operator Agency: Fruitland Park PD

Arresting Officer: MATEY, JOHN

Arresting Officer Agency: Fruitland Park PD

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 00:20

Gender: Female

Observation Period Began: 00:20

Subject's Last Name: RITTIRUCKSA

Date/Time of Test: 12/19/2007 00:45

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.082	00:42	
Air Blank	0.000	00:43	
Subject Sample #1		00:44	Subject Test Refused
Breath Volume	0.000	00:44	
Air Blank	0.000	00:44	
Control Test	0.082	00:45	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 01/11/2008 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001350

Instrument Registered To: FRUITLAND PARK P.D.

Software Version: 8100.27

Breath Test Operator: FOSTER, STEVEN, M

Breath Test Operator Agency: Fruitland Park PD

Arresting Officer: ISOM, JAMES

Arresting Officer Agency: Fruitland Park PD

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:52

Gender: Male

Observation Period Began: 01:55

Subject's Last Name: RICKER

Date/Time of Test: 12/23/2007 02:36

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.082	02:31
Air Blank	0.000	02:31
Subject Sample #1	0.174	02:32
Breath Volume	2.937	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample #2	0.170	02:34
Breath Volume	2.519	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 01/11/2008 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001350

Instrument Registered To: FRUITLAND PARK P.D.

Software Version: 8100.27

Breath Test Operator: PARISH, R, A

Breath Test Operator Agency: Fruitland Park PD

Arresting Officer: WHERRY, V

Arresting Officer Agency: Lady Lake PD

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 23:32

Gender: Female

Observation Period Began: 23:45

Subject's Last Name: JONES

Date/Time of Test: 12/30/2007 00:16

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.082	00:10
Air Blank	0.000	00:11
Subject Sample #1	0.124	00:11
Breath Volume	1.832	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample #2	0.121	00:14
Breath Volume	2.031	00:14
Air Blank	0.000	00:15
Control Test	0.082	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 01/11/2008 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001350

Instrument Registered To: FRUITLAND PARK P.D.

Software Version: 8100.27

Breath Test Operator: PARISH, R', A

Breath Test Operator Agency: Fruitland Park PD

Arresting Officer: HOLLOWAY, T

Arresting Officer Agency: Lake County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:25

Gender: Male

Observation Period Began: 01:25

Subject's Last Name: WILLIAMS

Date/Time of Test: 12/30/2007 01:53

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.082	01:47
Air Blank	0.000	01:47
Subject Sample #1	0.180	01:48
Breath Volume	1.679	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample #2	0.184	01:51
Breath Volume	1.839	01:51
Air Blank	0.000	01:52
Control Test	0.082	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:53

Data Download Date/Time: 01/11/2008 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001350

Instrument Registered To: FRUITLAND PARK P.D.

Software Version: 8100.27

Breath Test Operator: CUTTER, DENNIS, L

Breath Test Operator Agency: Fruitland Park PD

Arresting Officer: VANCE, SHAWN

Arresting Officer Agency: Lake County SO

Last Agency Inspection: 12/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 22:38

Gender: Male

Observation Period Began: 22:15

Subject's Last Name: COCHRAN

Date/Time of Test: 12/31/2007 22:49

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.083	22:44
Air Blank	0.000	22:44
Subject Sample #1	0.107	22:44
Breath Volume	1.632	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample #2	0.105	22:47
Breath Volume	2.093	22:47
Air Blank	0.000	22:48
Control Test	0.081	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 01/11/2008 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001361

Instrument Registered To: FL HIGHWAY PATROL

Software Version: 8100.27

Breath Test Operator: ASBILL, R, L

Breath Test Operator Agency: FHP Troop D

Arresting Officer: ASBILL, R

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 12/02/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 23:53

Gender: Male

Observation Period Began: 00:20

Subject's Last Name: HALL

Date/Time of Test: 12/29/2007 01:08

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Subject Sample #1		01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 01/05/2008 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001363

Instrument Registered To: CHARLOTTE COUNTY SO

Software Version: 8100.27

Breath Test Operator: GRAVEL, CHRISTINA, D

Breath Test Operator Agency: Charlotte County SO

Arresting Officer: JONES, M

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611501G

Dry Gas Expiry Date: 04/27/2008

Violation Code: DUI

Arrest Time: 02:22

Gender: Male

Observation Period Began: 02:24

Subject's Last Name: ELLIS

Date/Time of Test: 12/29/2007 03:02

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.075	02:53	
Air Blank	0.000	02:54	
Subject Sample #1	0.242	02:54	
Breath Volume	3.164	02:54	
Air Blank	0.000	02:55	
Air Blank	0.000	02:57	
Subject Sample #2	0.213	02:57	
Breath Volume	1.414	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample #3	0.238	03:00	
Breath Volume	1.812	03:00	
Air Blank	0.000	03:01	
Control Test	0.073	03:01	Control Outside Tolerance
Air Blank	0.000	03:02	

Data Download Date/Time: 01/17/2008 04:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001363

Instrument Registered To: CHARLOTTE COUNTY SO

Software Version: 8100.27

Breath Test Operator: EGAN, MELISSA,

Breath Test Operator Agency: Charlotte County SO

Arresting Officer: ORR, MICHAEL

Arresting Officer Agency: Charlotte County SO

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611501G

Dry Gas Expiry Date: 04/27/2008

Violation Code: DUI

Arrest Time: 02:26

Gender: Male

Observation Period Began: 03:20

Subject's Last Name: DEUBEL

Date/Time of Test: 12/30/2007 03:47

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample #1	0.123	03:42
Breath Volume	2.058	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample #2	0.111	03:46
Breath Volume	3.042	03:46
Air Blank	0.000	03:46
Control Test	0.076	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 01/17/2008 04:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001363

Instrument Registered To: CHARLOTTE COUNTY SO

Software Version: 8100.27

Breath Test Operator: EGAN, MELISSA,

Breath Test Operator Agency: Punta Gorda PD

Arresting Officer: AYERS, HARVEY

Arresting Officer Agency: Punta Gorda PD

Last Agency Inspection: 12/20/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 611501G

Dry Gas Expiry Date: 04/27/2008

Violation Code: DUI

Arrest Time: 03:59

Gender: Male

Observation Period Began: 04:15

Subject's Last Name: BOYD

Date/Time of Test: 12/30/2007 04:53

Diagnostics Check	OK	04:48	
Air Blank	0.000	04:49	
Control Test	0.079	04:49	
Air Blank	0.000	04:50	
Subject Sample #1		04:51	Subject Test Refused
Breath Volume	0.457	04:51	
Air Blank	0.000	04:52	
Control Test	0.080	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 01/17/2008 04:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001363

Instrument Registered To: CHARLOTTE COUNTY SO

Software Version: 8100.27

Breath Test Operator: EGAN, MELISSA,

Breath Test Operator Agency: Charlotte County SO

Arresting Officer: SCOTT, JOSEPH

Arresting Officer Agency: Charlotte County SO

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611501G

Dry Gas Expiry Date: 04/27/2008

Violation Code: DUI

Arrest Time: 00:36

Gender: Male

Observation Period Began: 00:55

Subject's Last Name: OLESON

Date/Time of Test: 12/31/2007 01:25

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.079	01:20
Air Blank	0.000	01:20
Subject Sample #1	0.201	01:21
Breath Volume	1.636	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample #2	0.214	01:24
Breath Volume	1.765	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 01/17/2008 04:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001363

Instrument Registered To: CHARLOTTE COUNTY SO

Software Version: 8100.27

Breath Test Operator: VAN ARSDALE, JEFFREY, A

Breath Test Operator Agency: FHP Troop F

Arresting Officer: VAN ARSDALE, JEFFREY

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611501G

Dry Gas Expiry Date: 04/27/2008

Violation Code: DUI

Arrest Time: 15:37

Gender: Male

Observation Period Began: 16:15

Subject's Last Name: KEPPELER

Date/Time of Test: 12/31/2007 16:55

Diagnostics Check	OK	16:48
Air Blank	0.000	16:49
Control Test	0.080	16:49
Air Blank	0.000	16:50
Subject Sample #1	0.000	16:50
Breath Volume	2.062	16:50
Air Blank	0.000	16:51
Air Blank	0.000	16:53
Subject Sample #2	0.000	16:53
Breath Volume	1.957	16:53
Air Blank	0.000	16:54
Control Test	0.080	16:54
Air Blank	0.000	16:54
Diagnostics Check	OK	16:55

Data Download Date/Time: 01/17/2008 04:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001363

Instrument Registered To: CHARLOTTE COUNTY SO

Software Version: 8100.27

Breath Test Operator: KIRKLAND, GREGGORY, W

Breath Test Operator Agency: Charlotte County SO

Arresting Officer: RIVERA, L

Arresting Officer Agency: Charlotte County SO

Last Agency Inspection: 12/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611501G

Dry Gas Expiry Date: 04/27/2008

Violation Code: DUI

Arrest Time: 21:58

Gender: Male

Observation Period Began: 22:25

Subject's Last Name: SIMARD

Date/Time of Test: 12/31/2007 23:03

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Subject Sample #1	0.042	22:59
Breath Volume	3.804	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample #2	0.041	23:02
Breath Volume	4.128	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 01/17/2008 04:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 23:12

Gender: Female

Observation Period Began: 23:40

Subject's Last Name: MORGAN

Date/Time of Test: 12/01/2007 00:12

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Subject Sample #1	0.149	00:03	
Breath Volume	1.105	00:03	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample #2	0.120	00:08	Volume Not Met
Breath Volume	0.250	00:08	
Air Blank	0.000	00:09	
Air Blank	0.000	00:11	
Subject Sample #3		00:11	RFI Detect
Breath Volume	0.000	00:11	
Air Blank		00:11	RFI Detect
Air Blank	0.000	00:12	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 23:12

Gender: Female

Observation Period Began: 00:25

Subject's Last Name: MORGAN

Date/Time of Test: 12/01/2007 01:02

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.079	00:47	
Air Blank	0.000	00:47	
Subject Sample #1	0.111	00:50	Volume Not Met
Breath Volume	1.050	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample #2	0.115	00:55	
Breath Volume	1.148	00:55	
Air Blank	0.000	00:56	
Air Blank	0.000	00:57	
Subject Sample #3	0.111	01:01	Volume Not Met
Breath Volume	0.894	01:01	
Air Blank	0.000	01:01	
Control Test	0.075	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 03:45

Gender: Female

Observation Period Began: 04:05

Subject's Last Name: BLECKER

Date/Time of Test: 12/01/2007 04:48

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:34	
Control Test	0.080	04:35	
Air Blank	0.000	04:35	
Subject Sample #1	0.222	04:38	Volume Not Met
Breath Volume	0.628	04:38	
Air Blank	0.000	04:39	
Air Blank	0.000	04:41	
Subject Sample #2	0.204	04:43	
Breath Volume	1.847	04:43	
Air Blank	0.000	04:44	
Air Blank	0.000	04:45	
Subject Sample #3	0.216	04:46	
Breath Volume	2.300	04:46	
Air Blank	0.000	04:47	
Control Test	0.078	04:48	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: TRUMP, JEFFREY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: CAVETT, MICHAEL

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 00:35

Gender: Male

Observation Period Began: 00:51

Subject's Last Name: FRANK

Date/Time of Test: 12/02/2007 01:26

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Subject Sample #1		01:19	No Sample Provided
Breath Volume	0.000	01:19	
Air Blank	0.000	01:20	
Air Blank	0.000	01:22	
Subject Sample #2		01:25	No Sample Provided
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Control Test	0.079	01:26	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DONNELLY, PATRICK,

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DUELZ, JOSEPH

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 04:00

Gender: Male

Observation Period Began: 04:25

Subject's Last Name: BURROUGHS

Date/Time of Test: 12/02/2007 04:56

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:45	
Control Test	0.080	04:45	
Air Blank	0.000	04:46	
Subject Sample #1		04:49	No Sample Provided
Breath Volume	0.000	04:49	
Air Blank	0.000	04:50	
Air Blank	0.000	04:52	
Subject Sample #2		04:55	No Sample Provided
Breath Volume	0.000	04:55	
Air Blank	0.000	04:55	
Control Test	0.080	04:55	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: PLANK, TRACY, C

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DELUCCA, SAM

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 11:18

Gender: Male

Observation Period Began: 11:25

Subject's Last Name: FORD

Date/Time of Test: 12/02/2007 12:05

Diagnostics Check	OK	11:58	
Air Blank	0.000	11:58	
Control Test	0.080	11:59	
Air Blank	0.000	11:59	
Subject Sample #1		12:02	No Sample Provided
Breath Volume	0.000	12:02	
Air Blank	0.000	12:03	
Air Blank		12:04	Sequence Aborted
Air Blank	0.000	12:05	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: LEPIANKA, DAVID, V

Breath Test Operator Agency: Gainesville PD

Arresting Officer: LEPIANKA, DAVID

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 01:01

Gender: Male

Observation Period Began: 01:22

Subject's Last Name: SNELGROVE

Date/Time of Test: 12/04/2007 01:51

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample #1	0.146	01:47
Breath Volume	2.601	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:49
Subject Sample #2	0.152	01:50
Breath Volume	2.085	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: KENNEDY, ROBERT, J

Breath Test Operator Agency: Gainesville PD

Arresting Officer: KENNEDY, ROBERT

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 20:00

Gender: Male

Observation Period Began: 20:10

Subject's Last Name: FLARITY

Date/Time of Test: 12/04/2007 20:35

Diagnostics Check	OK	20:35	
Air Blank		20:35	RFI Detect
Air Blank	0.000	20:35	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: KENNEDY, ROBERT, J

Breath Test Operator Agency: Gainesville PD

Arresting Officer: KENNEDY, ROBERT

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 20:00

Gender: Male

Observation Period Began: 20:10

Subject's Last Name: FLARITY

Date/Time of Test: 12/04/2007 20:43

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:40	
Control Test	0.081	20:40	
Air Blank	0.000	20:41	
Subject Sample #1		20:43	Subject Test Refused
Breath Volume	0.000	20:43	
Air Blank		20:43	RFI Detect
Air Blank	0.000	20:43	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: LEPIANKA, DAVID, V

Breath Test Operator Agency: Gainesville PD

Arresting Officer: LEPIANKA, DAVID

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 23:27

Gender: Female

Observation Period Began: 00:01

Subject's Last Name: SURLES-WEBB

Date/Time of Test: 12/06/2007 00:29

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.080	00:22	
Air Blank	0.000	00:22	
Subject Sample #1		00:25	No Sample Provided
Breath Volume	0.000	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:28	
Subject Sample #2		00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.080	00:29	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 03:10

Gender: Male

Observation Period Began: 03:17

Subject's Last Name: DIXON

Date/Time of Test: 12/06/2007 03:44

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Subject Sample #1	0.237	03:39
Breath Volume	2.921	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample #2	0.220	03:42
Breath Volume	2.546	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: TALTON, C, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: TALTON, C

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 03:26

Gender: Male

Observation Period Began: 04:00

Subject's Last Name: ESPIN

Date/Time of Test: 12/07/2007 04:30

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample #1	0.113	04:25
Breath Volume	1.597	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample #2	0.111	04:28
Breath Volume	1.476	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 00:50

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: RAY

Date/Time of Test: 12/08/2007 01:27

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:21
Subject Sample #1	0.200	01:22
Breath Volume	1.656	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample #2	0.196	01:26
Breath Volume	2.414	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: GUYAN, GINA, L

Breath Test Operator Agency: Gainesville PD

Arresting Officer: GUYAN, GINA

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 01:46

Subject's Last Name: EDGAR

Date/Time of Test: 12/08/2007 02:15

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample #1	0.144	02:10
Breath Volume	3.730	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample #2	0.125	02:13
Breath Volume	3.328	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: SURRENCY, D, M

Breath Test Operator Agency: Gainesville PD

Arresting Officer: SURRENCY, D

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 02:35

Gender: Male

Observation Period Began: 01:52

Subject's Last Name: EVANS III

Date/Time of Test: 12/09/2007 02:23

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.078	02:13	
Air Blank	0.000	02:13	
Subject Sample #1		02:16	No Sample Provided
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample #2		02:22	No Sample Provided
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 02:05

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: CHANCY

Date/Time of Test: 12/09/2007 03:10

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.081	03:00	
Air Blank	0.000	03:00	
Subject Sample #1		03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample #2		03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.081	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: TALTON, C, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: TALTON, C

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 02:45

Gender: Male

Observation Period Began: 03:15

Subject's Last Name: GASKINS

Date/Time of Test: 12/09/2007 03:46

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample #1	0.173	03:41
Breath Volume	2.496	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample #2	0.178	03:44
Breath Volume	1.808	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: HENAGAN, MICHAEL, G

Breath Test Operator Agency: Gainesville PD

Arresting Officer: HENAGAN, MICHAEL

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 18:30

Gender: Male

Observation Period Began: 18:35

Subject's Last Name: GUTIERREZ

Date/Time of Test: 12/09/2007 19:08

Diagnostics Check	OK	19:07	
Air Blank		19:08	RFI Detect
Air Blank	0.000	19:08	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: HENAGAN, MICHAEL, G

Breath Test Operator Agency: Gainesville PD

Arresting Officer: HENAGAN, MICHAEL

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 18:30

Gender: Male

Observation Period Began: 18:35

Subject's Last Name: GUTIERREZ

Date/Time of Test: 12/09/2007 19:17

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.080	19:10
Air Blank	0.000	19:11
Subject Sample #1	0.217	19:12
Breath Volume	2.890	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample #2	0.206	19:15
Breath Volume	2.496	19:15
Air Blank	0.000	19:16
Control Test	0.079	19:16
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: TALTON, C, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: BRISTOW, S

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 00:48

Gender: Male

Observation Period Began: 01:10

Subject's Last Name: REYES

Date/Time of Test: 12/10/2007 01:42

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.080	01:31	
Air Blank	0.000	01:32	
Subject Sample #1		01:35	No Sample Provided
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample #2		01:41	No Sample Provided
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: KENNEDY, ROBERT, J

Breath Test Operator Agency: Gainesville PD

Arresting Officer: KENNEDY, ROBERT

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 03:15

Gender: Male

Observation Period Began: 03:30

Subject's Last Name: GUTHRIE

Date/Time of Test: 12/12/2007 03:54

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Subject Sample #1		03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.079	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 02:40

Gender: Female

Observation Period Began: 02:50

Subject's Last Name: CHIRICO

Date/Time of Test: 12/13/2007 03:25

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:18
Subject Sample #1	0.148	03:19
Breath Volume	2.792	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:21
Subject Sample #2	0.138	03:23
Breath Volume	2.445	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: SURRENCY, D, M

Breath Test Operator Agency: Gainesville PD

Arresting Officer: SURRENCY, D

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 01:35

Gender: Male

Observation Period Began: 01:50

Subject's Last Name: MILLER

Date/Time of Test: 12/14/2007 02:17

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample #1	0.168	02:12
Breath Volume	1.964	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample #2	0.159	02:15
Breath Volume	2.183	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 02:50

Gender: Female

Observation Period Began: 03:10

Subject's Last Name: TAYLOR

Date/Time of Test: 12/15/2007 03:47

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Subject Sample #1	0.203	03:39
Breath Volume	1.699	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample #2	0.178	03:42
Breath Volume	1.250	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample #3	0.173	03:45
Breath Volume	2.289	03:45
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 00:40

Gender: Male

Observation Period Began: 00:50

Subject's Last Name: BULSON

Date/Time of Test: 12/16/2007 01:25

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:18
Subject Sample #1	0.230	01:20
Breath Volume	2.589	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample #2	0.229	01:23
Breath Volume	3.140	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: LEPIANKA, DAVID, V

Breath Test Operator Agency: Gainesville PD

Arresting Officer: LEPIANKA, DAVID

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 04:15

Gender: Male

Observation Period Began: 04:40

Subject's Last Name: WILSON

Date/Time of Test: 12/18/2007 05:07

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:04	
Control Test	0.081	05:04	
Air Blank	0.000	05:05	
Subject Sample #1		05:05	Subject Test Refused
Breath Volume	0.000	05:05	
Air Blank	0.000	05:06	
Control Test	0.080	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: PRESTON, MICHEAL, D

Breath Test Operator Agency: Gainesville PD

Arresting Officer: PRESTON, MICHEAL

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 15:20

Gender: Female

Observation Period Began: 15:38

Subject's Last Name: MACHADO

Date/Time of Test: 12/19/2007 16:13

Diagnostics Check	OK	16:07
Air Blank	0.000	16:07
Control Test	0.081	16:08
Air Blank	0.000	16:08
Subject Sample #1	0.044	16:09
Breath Volume	1.824	16:09
Air Blank	0.000	16:10
Air Blank	0.000	16:11
Subject Sample #2	0.040	16:12
Breath Volume	2.960	16:12
Air Blank	0.000	16:13
Control Test	0.080	16:13
Air Blank	0.000	16:13
Diagnostics Check	OK	16:13

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: KENNEDY, ROBERT, J

Breath Test Operator Agency: Gainesville PD

Arresting Officer: KENNEDY, ROBERT

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 01:02

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: HASLAM

Date/Time of Test: 12/21/2007 01:49

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:48	
Subject Sample #1		01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: SURRENCY, D, M

Breath Test Operator Agency: Gainesville PD

Arresting Officer: SURENCY, D

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 02:25

Gender: Female

Observation Period Began: 02:35

Subject's Last Name: STEARNS

Date/Time of Test: 12/21/2007 03:02

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Subject Sample #1	0.187	02:58
Breath Volume	1.281	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample #2	0.189	03:01
Breath Volume	2.234	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 23:45

Gender: Female

Observation Period Began: 00:20

Subject's Last Name: HAYES

Date/Time of Test: 12/22/2007 00:56

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.079	00:44	
Air Blank	0.000	00:45	
Subject Sample #1	0.164	00:48	Volume Not Met
Breath Volume	0.015	00:48	
Air Blank	0.000	00:49	
Air Blank	0.000	00:50	
Subject Sample #2	0.170	00:51	
Breath Volume	1.437	00:51	
Air Blank	0.000	00:52	
Air Blank	0.000	00:54	
Subject Sample #3	0.156	00:55	
Breath Volume	1.500	00:55	
Air Blank	0.000	00:55	
Control Test	0.078	00:56	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 02:15

Gender: Female

Observation Period Began: 02:38

Subject's Last Name: CLINE

Date/Time of Test: 12/22/2007 03:10

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	02:59	
Subject Sample #1		03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample #2		03:08	No Sample Provided
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 00:50

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: HUTTON

Date/Time of Test: 12/23/2007 01:49

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.081	01:39	
Air Blank	0.000	01:39	
Subject Sample #1		01:42	No Sample Provided
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Air Blank	0.000	01:45	
Subject Sample #2		01:48	No Sample Provided
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.080	01:49	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: DEVORE, JOEY, A

Breath Test Operator Agency: Gainesville PD

Arresting Officer: DEVORE, JOEY

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 03:43

Gender: Male

Observation Period Began: 04:04

Subject's Last Name: MONTIEL

Date/Time of Test: 12/23/2007 04:38

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:32
Subject Sample #1	0.179	04:33
Breath Volume	2.011	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample #2	0.178	04:36
Breath Volume	1.312	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:38

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: KENNEDY, ROBERT, J

Breath Test Operator Agency: Gainesville PD

Arresting Officer: KENNEDY, ROBERT

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 00:24

Gender: Male

Observation Period Began: 00:40

Subject's Last Name: UPSHAW

Date/Time of Test: 12/27/2007 01:08

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.080	01:04	
Air Blank	0.000	01:04	
Subject Sample #1		01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 12/27/2007 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: KENNEDY, ROBERT, J

Breath Test Operator Agency: Gainesville PD

Arresting Officer: KENNEDY, ROBERT

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 00:34

Gender: Male

Observation Period Began: 00:59

Subject's Last Name: GRIGGS

Date/Time of Test: 12/28/2007 01:24

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:22	
Control Test	0.080	01:22	
Air Blank	0.000	01:22	
Subject Sample #1		01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 01/31/2008 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001364

Instrument Registered To: GAINESVILLE PD

Software Version: 8100.27

Breath Test Operator: JONES, STEVEN, E

Breath Test Operator Agency: Gainesville PD

Arresting Officer: JONES, STEVEN

Arresting Officer Agency: Gainesville PD

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 724801G

Dry Gas Expiry Date: 09/06/2009

Violation Code: DUI

Arrest Time: 21:45

Gender: Male

Observation Period Began: 21:51

Subject's Last Name: SIMPSON

Date/Time of Test: 12/30/2007 22:22

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.079	22:20	
Air Blank	0.000	22:20	
Subject Sample #1		22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.081	22:21	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 01/31/2008 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001365

Instrument Registered To: LEVY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ANDERS, J, P

Breath Test Operator Agency: Levy County SO

Arresting Officer: MERRITT, J

Arresting Officer Agency: Cedar Keys PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 01:25

Gender: Female

Observation Period Began: 02:06

Subject's Last Name: MILLER

Date/Time of Test: 12/08/2007 02:44

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample #1	0.189	02:39
Breath Volume	1.746	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample #2	0.178	02:43
Breath Volume	2.027	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 12/31/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001365

Instrument Registered To: LEVY COUNTY SO

Software Version: 8100.27

Breath Test Operator: MARTIN, SCOTT, G

Breath Test Operator Agency: FHP Troop B

Arresting Officer: MARTIN, SCOTT

Arresting Officer Agency: FHP Troop B

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 15:26

Gender: Male

Observation Period Began: 15:55

Subject's Last Name: MEEKS

Date/Time of Test: 12/09/2007 16:27

Diagnostics Check	OK	16:15	
Air Blank	0.000	16:16	
Control Test	0.078	16:16	
Air Blank	0.000	16:17	
Subject Sample #1		16:20	No Sample Provided
Breath Volume	0.000	16:20	
Air Blank	0.000	16:20	
Air Blank	0.000	16:22	
Subject Sample #2		16:25	No Sample Provided
Breath Volume	0.000	16:25	
Air Blank	0.000	16:26	
Control Test	0.079	16:26	
Air Blank	0.000	16:27	
Diagnostics Check	OK	16:27	

Data Download Date/Time: 12/31/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001365

Instrument Registered To: LEVY COUNTY SO

Software Version: 8100.27

Breath Test Operator: GORE, JOSEPH, L

Breath Test Operator Agency: Levy County SO

Arresting Officer: BOLTON, JAY

Arresting Officer Agency: Chiefland PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 11:24

Gender: Male

Observation Period Began: 12:27

Subject's Last Name: DEMASTERS

Date/Time of Test: 12/13/2007 13:11

Diagnostics Check	OK	13:03
Air Blank	0.000	13:04
Control Test	0.075	13:04
Air Blank	0.000	13:05
Subject Sample #1	0.218	13:06
Breath Volume	1.242	13:06
Air Blank	0.000	13:07
Air Blank	0.000	13:08
Subject Sample #2	0.226	13:09
Breath Volume	1.167	13:09
Air Blank	0.000	13:10
Control Test	0.075	13:11
Air Blank	0.000	13:11
Diagnostics Check	OK	13:11

Data Download Date/Time: 12/31/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001365

Instrument Registered To: LEVY COUNTY SO

Software Version: 8100.27

Breath Test Operator: KING, MATTHEW, L

Breath Test Operator Agency: Levy County SO

Arresting Officer: BREEDING, A

Arresting Officer Agency: Levy County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 01:20

Gender: Male

Observation Period Began: 01:55

Subject's Last Name: RIDDLE

Date/Time of Test: 12/16/2007 04:58

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:56	
Control Test	0.080	04:56	
Air Blank	0.000	04:56	
Subject Sample #1		04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:57	
Control Test	0.079	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 12/31/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001365

Instrument Registered To: LEVY COUNTY SO

Software Version: 8100.27

Breath Test Operator: GRIFFETH, D, L

Breath Test Operator Agency: Levy County SO

Arresting Officer: TEST, TEST

Arresting Officer Agency: Levy County SO

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: ADMIN

Arrest Time: 21:00

Gender: Male

Observation Period Began: 21:00

Subject's Last Name: HOLDER

Date/Time of Test: 12/17/2007 21:32

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27
Subject Sample #1	0.000	21:27
Breath Volume	3.667	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample #2	0.000	21:30
Breath Volume	3.648	21:30
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 12/31/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001365

Instrument Registered To: LEVY COUNTY SO

Software Version: 8100.27

Breath Test Operator: MERRITT, JAMES, E

Breath Test Operator Agency: Cedar Keys PD

Arresting Officer: MERRITT, JAMES

Arresting Officer Agency: Cedar Keys PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 22:49

Gender: Male

Observation Period Began: 00:04

Subject's Last Name: ANDERSON

Date/Time of Test: 12/20/2007 00:42

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:36
Air Blank	0.000	00:36
Subject Sample #1	0.176	00:37
Breath Volume	1.800	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample #2	0.178	00:40
Breath Volume	2.089	00:40
Air Blank	0.000	00:41
Control Test	0.076	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 12/31/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001365

Instrument Registered To: LEVY COUNTY SO

Software Version: 8100.27

Breath Test Operator: GRIFFETH, D, L

Breath Test Operator Agency: Levy County SO

Arresting Officer: MERRITT, E

Arresting Officer Agency: Cedar Keys PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: COMM VEHIC

Arrest Time: 00:04

Gender: Male

Observation Period Began: 00:49

Subject's Last Name: BEAUCHAMP

Date/Time of Test: 12/22/2007 01:26

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Subject Sample #1	0.135	01:21
Breath Volume	2.519	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample #2	0.131	01:24
Breath Volume	1.925	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 12/31/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001365

Instrument Registered To: LEVY COUNTY SO

Software Version: 8100.27

Breath Test Operator: GRIFFETH, D, L

Breath Test Operator Agency: Levy County SO

Arresting Officer: MERRITT, E

Arresting Officer Agency: Cedar Keys PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 23:01

Gender: Male

Observation Period Began: 23:58

Subject's Last Name: BAKER

Date/Time of Test: 12/23/2007 00:42

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:40	
Subject Sample #1		00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:41	
Control Test	0.079	00:41	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 12/31/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001365

Instrument Registered To: LEVY COUNTY SO

Software Version: 8100.27

Breath Test Operator: KING, MATTHEW, L

Breath Test Operator Agency: Levy County SO

Arresting Officer: COSSEBOOM, MICHAEL

Arresting Officer Agency: Cedar Keys PD

Last Agency Inspection: 11/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 20:48

Gender: Male

Observation Period Began: 21:30

Subject's Last Name: BAILEY

Date/Time of Test: 12/25/2007 22:04

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.080	21:58
Air Blank	0.000	21:58
Subject Sample #1	0.105	21:59
Breath Volume	1.804	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:01
Subject Sample #2	0.108	22:02
Breath Volume	2.378	22:02
Air Blank	0.000	22:03
Control Test	0.077	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 12/31/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001365

Instrument Registered To: LEVY COUNTY SO
Software Version: 8100.27
Breath Test Operator: WILKINSON, MICHAEL, B
Breath Test Operator Agency: Levy County SO
Arresting Officer: PAOLI, JASON
Arresting Officer Agency: Levy County SO
Last Agency Inspection: 11/27/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 706102I
Dry Gas Expiry Date: 03/05/2009
Violation Code: DUI
Arrest Time: 09:28
Gender: Female
Observation Period Began: 10:09

Subject's Last Name: HARLOW

Date/Time of Test: 12/29/2007 10:36

Diagnostics Check	OK	10:29
Air Blank	0.000	10:30
Control Test	0.078	10:30
Air Blank	0.000	10:31
Subject Sample #1	0.196	10:32
Breath Volume	2.457	10:32
Air Blank	0.000	10:32
Air Blank	0.000	10:34
Subject Sample #2	0.188	10:35
Breath Volume	1.796	10:35
Air Blank	0.000	10:35
Control Test	0.077	10:36
Air Blank	0.000	10:36
Diagnostics Check	OK	10:36

Data Download Date/Time: 12/31/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: PARKOS, KATHRYN, M

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: SCHROEDER, E

Arresting Officer Agency: Pinellas Park PD

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 02:41

Subject's Last Name: RICHARD

Date/Time of Test: 12/01/2007 03:10

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Subject Sample #1	0.150	03:06
Breath Volume	1.585	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample #2	0.163	03:09
Breath Volume	1.652	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MARTENS, J, T

Breath Test Operator Agency: Largo PD

Arresting Officer: MARTENS, J

Arresting Officer Agency: Largo PD

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 21:17

Gender: Male

Observation Period Began: 21:31

Subject's Last Name: GARDNER

Date/Time of Test: 12/01/2007 22:01

Diagnostics Check	OK	21:54
Air Blank	0.000	21:55
Control Test	0.080	21:55
Air Blank	0.000	21:56
Subject Sample #1	0.000	21:56
Breath Volume	2.253	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample #2	0.000	21:59
Breath Volume	1.929	21:59
Air Blank	0.000	22:00
Control Test	0.080	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MARTENS, J, T

Breath Test Operator Agency: Largo PD

Arresting Officer: ARLINGTON, B

Arresting Officer Agency: Largo PD

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 16:23

Gender: Female

Observation Period Began: 16:55

Subject's Last Name: KELLY

Date/Time of Test: 12/03/2007 17:22

Diagnostics Check	OK	17:15
Air Blank	0.000	17:16
Control Test	0.080	17:16
Air Blank	0.000	17:16
Subject Sample #1	0.000	17:17
Breath Volume	2.570	17:17
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample #2	0.000	17:21
Breath Volume	2.777	17:21
Air Blank	0.000	17:21
Control Test	0.079	17:21
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: BRUNEAU, DAVID, N

Breath Test Operator Agency: Clearwater PD

Arresting Officer: BRUNEAU, DAVID

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:03

Gender: Female

Observation Period Began: 23:44

Subject's Last Name: NOREN

Date/Time of Test: 12/06/2007 00:15

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Subject Sample #1	0.183	00:10
Breath Volume	1.488	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample #2	0.181	00:14
Breath Volume	1.351	00:14
Air Blank	0.000	00:14
Control Test	0.080	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: CITRANO, ANTHONY,

Breath Test Operator Agency: Largo PD

Arresting Officer: CITRANO, ANTHONY

Arresting Officer Agency: Largo PD

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 22:17

Gender: Female

Observation Period Began: 21:42

Subject's Last Name: WEATHERBEE

Date/Time of Test: 12/06/2007 23:06

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Subject Sample #1	0.032	23:00
Breath Volume	1.410	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample #2	0.030	23:04
Breath Volume	1.171	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: CITRANO, ANTHONY,

Breath Test Operator Agency: Largo PD

Arresting Officer: CITRANO, ANTHONY

Arresting Officer Agency: Largo PD

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 22:56

Gender: Male

Observation Period Began: 23:22

Subject's Last Name: BAILEY

Date/Time of Test: 12/07/2007 23:48

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample #1	0.000	23:44
Breath Volume	1.660	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample #2	0.000	23:47
Breath Volume	1.328	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: LUDLOW, CHRIS, W

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: LUDLOW, CHRIS

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:21

Gender: Male

Observation Period Began: 01:48

Subject's Last Name: VARGO

Date/Time of Test: 12/08/2007 02:15

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample #1	0.083	02:10
Breath Volume	1.527	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample #2	0.071	02:13
Breath Volume	1.847	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: PARKOS, KATHRYN, M

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: MASON, G

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:21

Gender: Female

Observation Period Began: 02:56

Subject's Last Name: TERSTEEGEN

Date/Time of Test: 12/09/2007 03:19

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Subject Sample #1		03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Control Test	0.080	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEANE, M,

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: WISEMAN, J

Arresting Officer Agency: Indian Shores PD

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 00:53

Gender: Female

Observation Period Began: 02:09

Subject's Last Name: BAKER

Date/Time of Test: 12/12/2007 02:41

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:33
Subject Sample #1	0.068	02:36
Breath Volume	1.347	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample #2	0.069	02:39
Breath Volume	1.140	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: PARKOS, KATHRYN, M

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: TRAYNOR, K

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 19:15

Gender: Male

Observation Period Began: 20:52

Subject's Last Name: WARREN

Date/Time of Test: 12/12/2007 21:26

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Subject Sample #1	0.181	21:21
Breath Volume	1.312	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample #2	0.172	21:24
Breath Volume	1.675	21:24
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: CITRANO, ANTHONY,

Breath Test Operator Agency: Largo PD

Arresting Officer: CITRANO, ANTHONY

Arresting Officer Agency: Largo PD

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 22:58

Gender: Male

Observation Period Began: 23:14

Subject's Last Name: MCMANUS

Date/Time of Test: 12/13/2007 23:38

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.079	23:36	
Air Blank	0.000	23:36	
Subject Sample #1		23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.080	23:37	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: LANGLAIS, P, A

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: LANGLAIS, P

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:43

Gender: Female

Observation Period Began: 03:04

Subject's Last Name: MOORE

Date/Time of Test: 12/15/2007 03:30

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample #1	0.123	03:26
Breath Volume	1.761	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample #2	0.123	03:29
Breath Volume	1.613	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MORRIS, BRIDGETT, A

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: MORRIS, BRIDGETT

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 03:57

Gender: Male

Observation Period Began: 04:40

Subject's Last Name: GREENE

Date/Time of Test: 12/15/2007 05:06

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:01
Subject Sample #1	0.141	05:02
Breath Volume	2.246	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample #2	0.145	05:05
Breath Volume	1.375	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MORRIS, BRIDGETT, A

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: WINCELOWICZ, CHARLES

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 05:30

Gender: Male

Observation Period Began: 05:49

Subject's Last Name: MCNALLY

Date/Time of Test: 12/15/2007 06:16

Diagnostics Check	OK	06:09
Air Blank	0.000	06:10
Control Test	0.080	06:10
Air Blank	0.000	06:10
Subject Sample #1	0.141	06:11
Breath Volume	1.480	06:11
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample #2	0.129	06:14
Breath Volume	1.253	06:14
Air Blank	0.000	06:15
Control Test	0.080	06:15
Air Blank	0.000	06:15
Diagnostics Check	OK	06:16

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: CITRANO, ANTHONY,

Breath Test Operator Agency: Largo PD

Arresting Officer: CITRANO, ANTHONY

Arresting Officer Agency: Largo PD

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:30

Gender: Female

Observation Period Began: 23:46

Subject's Last Name: COLEMAN

Date/Time of Test: 12/16/2007 00:12

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample #1	0.000	00:08
Breath Volume	1.273	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample #2	0.000	00:11
Breath Volume	1.312	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEANE, M,

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: FIEDLER, P

Arresting Officer Agency: Pinellas Park PD

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:00

Gender: Female

Observation Period Began: 00:20

Subject's Last Name: STRAUBEL

Date/Time of Test: 12/16/2007 00:55

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:49
Subject Sample #1	0.233	00:50
Breath Volume	2.031	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:52
Subject Sample #2	0.227	00:53
Breath Volume	1.386	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:55

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: LACHANCE, DEPUTY, D

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: LACHANCE, DEPUTY

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:26

Gender: Female

Observation Period Began: 01:48

Subject's Last Name: HOLMAN

Date/Time of Test: 12/16/2007 02:17

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:09	
Control Test	0.079	02:09	
Air Blank	0.000	02:09	
Subject Sample #1	0.201	02:13	Volume Not Met
Breath Volume	0.351	02:13	
Air Blank	0.000	02:13	
Air Blank	0.000	02:15	
Subject Sample #2		02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Control Test	0.078	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 12/16/2007 03:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: CITRANO, ANTHONY,

Breath Test Operator Agency: Largo PD

Arresting Officer: CITRANO, ANTHONY

Arresting Officer Agency: Largo PD

Last Agency Inspection: 12/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 19:40

Gender: Male

Observation Period Began: 19:56

Subject's Last Name: MIRANDA

Date/Time of Test: 12/20/2007 20:22

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Subject Sample #1	0.000	20:17
Breath Volume	2.914	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample #2	0.000	20:20
Breath Volume	2.734	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 01/16/2008 00:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: CITRANO, ANTHONY,

Breath Test Operator Agency: Largo PD

Arresting Officer: CITRANO, ANTHONY

Arresting Officer Agency: Largo PD

Last Agency Inspection: 12/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:42

Gender: Female

Observation Period Began: 00:12

Subject's Last Name: SHEA

Date/Time of Test: 12/21/2007 00:37

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.080	00:35	
Air Blank	0.000	00:35	
Subject Sample #1		00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:36	
Control Test	0.080	00:36	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 01/16/2008 00:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: CITRANO, ANTHONY,

Breath Test Operator Agency: Largo PD

Arresting Officer: CITRANO, ANTHONY

Arresting Officer Agency: Largo PD

Last Agency Inspection: 12/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 20:21

Gender: Female

Observation Period Began: 21:07

Subject's Last Name: BROWN

Date/Time of Test: 12/21/2007 21:33

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Subject Sample #1	0.000	21:29
Breath Volume	1.566	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample #2	0.000	21:32
Breath Volume	1.617	21:32
Air Blank	0.000	21:32
Control Test	0.080	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 01/16/2008 00:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCNEIL, J, R

Breath Test Operator Agency: Pinellas Park PD

Arresting Officer: MCNEIL, J

Arresting Officer Agency: Pinellas Park PD

Last Agency Inspection: 12/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:19

Gender: Female

Observation Period Began: 03:34

Subject's Last Name: FOSS

Date/Time of Test: 12/23/2007 04:08

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample #1	0.136	04:04
Breath Volume	2.781	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample #2	0.134	04:07
Breath Volume	2.816	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 01/16/2008 00:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCNEIL, J, R

Breath Test Operator Agency: Pinellas Park PD

Arresting Officer: MCNEIL, J

Arresting Officer Agency: Pinellas Park PD

Last Agency Inspection: 12/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:36

Gender: Male

Observation Period Began: 02:26

Subject's Last Name: BRADBERRY

Date/Time of Test: 12/27/2007 03:00

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Subject Sample #1	0.154	02:55
Breath Volume	1.230	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample #2	0.154	02:58
Breath Volume	3.050	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 01/16/2008 00:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCNEIL, J, R

Breath Test Operator Agency: Pinellas Park PD

Arresting Officer: MCNEIL, J

Arresting Officer Agency: Pinellas Park PD

Last Agency Inspection: 12/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 18:55

Gender: Male

Observation Period Began: 19:35

Subject's Last Name: RICHMOND

Date/Time of Test: 12/27/2007 20:02

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.080	19:59	
Air Blank	0.000	20:00	
Subject Sample #1		20:00	Subject Test Refused
Breath Volume	0.000	20:00	
Air Blank	0.000	20:01	
Control Test	0.080	20:01	
Air Blank	0.000	20:01	
Diagnostics Check	OK	20:02	

Data Download Date/Time: 01/16/2008 00:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001367

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MARTENS, J, T

Breath Test Operator Agency: Largo PD

Arresting Officer: VEGENSKI, M

Arresting Officer Agency: Largo PD

Last Agency Inspection: 12/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 20:11

Gender: Male

Observation Period Began: 21:20

Subject's Last Name: RUBIO

Date/Time of Test: 12/30/2007 21:43

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.081	21:41	
Air Blank	0.000	21:41	
Subject Sample #1		21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Control Test	0.081	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 01/16/2008 00:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001368

Instrument Registered To: MAYPORT NAVAL BASE

Software Version: 8100.27

Breath Test Operator: SCHLEGEL, ALLAN, L

Breath Test Operator Agency: Mayport Naval Base

Arresting Officer: PETERSON, ROBERT

Arresting Officer Agency: Mayport Naval Base

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7061011

Dry Gas Expiry Date: 03/06/2009

Violation Code: DUI

Arrest Time: 21:46

Gender: Male

Observation Period Began: 22:25

Subject's Last Name: ARROYOS

Date/Time of Test: 12/20/2007 23:22

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.078	23:11	
Air Blank	0.000	23:12	
Subject Sample #1		23:15	No Sample Provided
Breath Volume	0.000	23:15	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample #2		23:20	No Sample Provided
Breath Volume	0.000	23:20	
Air Blank	0.000	23:21	
Control Test	0.076	23:21	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 12/27/2007 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOSLEY, JENNIFER,

Breath Test Operator Agency: Orange County SO

Arresting Officer: MCKENZIE, W

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 23:48

Gender: Male

Observation Period Began: 00:57

Subject's Last Name: EATON

Date/Time of Test: 12/01/2007 01:29

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.081	01:24
Air Blank	0.000	01:24
Subject Sample #1	0.134	01:25
Breath Volume	1.914	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample #2	0.142	01:27
Breath Volume	2.183	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOSLEY, JENNIFER,

Breath Test Operator Agency: Orange County SO

Arresting Officer: HAWKINS, D

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:53

Gender: Male

Observation Period Began: 04:40

Subject's Last Name: MARRAKCHI

Date/Time of Test: 12/01/2007 05:08

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Subject Sample #1	0.056	05:04
Breath Volume	2.003	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample #2	0.054	05:07
Breath Volume	2.292	05:07
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: JOHNSON, CASPER

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 06:07

Gender: Male

Observation Period Began: 06:48

Subject's Last Name: BROOME

Date/Time of Test: 12/01/2007 07:29

Diagnostics Check	OK	07:15	
Air Blank	0.000	07:16	
Control Test	0.081	07:16	
Air Blank	0.000	07:17	
Subject Sample #1	0.214	07:20	Volume Not Met
Breath Volume	0.117	07:20	
Air Blank	0.000	07:20	
Air Blank	0.000	07:22	
Subject Sample #2	0.245	07:23	No .020 agreement
Breath Volume	1.574	07:23	
Air Blank	0.000	07:24	
Air Blank	0.000	07:26	
Subject Sample #3	0.266	07:28	No .020 agreement
Breath Volume	1.113	07:28	
Air Blank	0.000	07:28	
Control Test	0.079	07:29	
Air Blank	0.000	07:29	
Diagnostics Check	OK	07:29	

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: CASPER, JOHNSON

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 06:07

Gender: Male

Observation Period Began: 06:48

Subject's Last Name: BROOME

Date/Time of Test: 12/01/2007 07:40

Diagnostics Check	OK	07:32
Air Blank	0.000	07:33
Control Test	0.079	07:33
Air Blank	0.000	07:33
Subject Sample #1	0.242	07:35
Breath Volume	2.527	07:35
Air Blank	0.000	07:36
Air Blank	0.000	07:38
Subject Sample #2	0.252	07:39
Breath Volume	2.074	07:39
Air Blank	0.000	07:39
Control Test	0.079	07:40
Air Blank	0.000	07:40
Diagnostics Check	OK	07:40

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LANDROVE, ROBERT,

Breath Test Operator Agency: Orange County SO

Arresting Officer: HALL, F

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 07:09

Gender: Male

Observation Period Began: 07:38

Subject's Last Name: LEVINE

Date/Time of Test: 12/01/2007 08:30

Diagnostics Check	OK	08:15	
Air Blank	0.000	08:15	
Control Test	0.080	08:16	
Air Blank	0.000	08:16	
Subject Sample #1	0.135	08:19	Slope Not Level
Breath Volume	1.164	08:19	
Air Blank	0.000	08:20	
Air Blank	0.000	08:22	
Subject Sample #2	0.143	08:23	
Breath Volume	1.261	08:23	
Air Blank	0.000	08:23	
Air Blank	0.000	08:25	
Subject Sample #3	0.133	08:28	Volume Not Met
Breath Volume	0.960	08:28	
Air Blank	0.000	08:29	
Control Test	0.080	08:29	
Air Blank	0.000	08:30	
Diagnostics Check	OK	08:30	

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LANDROVE, ROBERT,

Breath Test Operator Agency: Orange County SO

Arresting Officer: HALL, F

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 07:09

Gender: Male

Observation Period Began: 07:38

Subject's Last Name: LEVINE

Date/Time of Test: 12/01/2007 08:39

Diagnostics Check	OK	08:32	
Air Blank	0.000	08:33	
Control Test	0.080	08:33	
Air Blank	0.000	08:33	
Subject Sample #1	0.146	08:35	
Breath Volume	1.988	08:35	
Air Blank	0.000	08:35	
Air Blank	0.000	08:37	
Subject Sample #2		08:37	Subject Test Refused
Breath Volume	0.000	08:37	
Air Blank	0.000	08:38	
Control Test	0.079	08:38	
Air Blank	0.000	08:39	
Diagnostics Check	OK	08:39	

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOSLEY, JENNIFER,

Breath Test Operator Agency: Orange County SO

Arresting Officer: GONZALEZ, S

Arresting Officer Agency: Maitland PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 00:41

Gender: Female

Observation Period Began: 01:00

Subject's Last Name: SELLERS

Date/Time of Test: 12/02/2007 01:28

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample #1	0.237	01:24
Breath Volume	1.312	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample #2	0.242	01:27
Breath Volume	1.554	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LANDROVE, ROBERT,

Breath Test Operator Agency: Orange County SO

Arresting Officer: RENARD, A

Arresting Officer Agency: FHP Troop K

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 07:54

Gender: Male

Observation Period Began: 08:31

Subject's Last Name: ROJAS ACOSTA

Date/Time of Test: 12/02/2007 09:01

Diagnostics Check	OK	08:54
Air Blank	0.000	08:54
Control Test	0.081	08:55
Air Blank	0.000	08:55
Subject Sample #1	0.178	08:57
Breath Volume	3.246	08:57
Air Blank	0.000	08:57
Air Blank	0.000	08:59
Subject Sample #2	0.177	09:00
Breath Volume	2.765	09:00
Air Blank	0.000	09:01
Control Test	0.079	09:01
Air Blank	0.000	09:01
Diagnostics Check	OK	09:01

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MILLS, A, G

Breath Test Operator Agency: Orange County SO

Arresting Officer: RAMIREZ, FABIAN

Arresting Officer Agency: Central FL Univ PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:21

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: CLARKE

Date/Time of Test: 12/04/2007 03:31

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.082	03:24
Air Blank	0.000	03:24
Subject Sample #1	0.000	03:27
Breath Volume	2.082	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample #2	0.000	03:30
Breath Volume	1.355	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: GONZALEZ, SABASTIAN

Arresting Officer Agency: Maitland PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:14

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: GRIFFIN

Date/Time of Test: 12/05/2007 02:14

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Subject Sample #1	0.203	02:10
Breath Volume	1.988	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample #2	0.200	02:13
Breath Volume	1.945	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LANDROVE, ROBERT,

Breath Test Operator Agency: Orange County SO

Arresting Officer: VAUGHN, J

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:28

Gender: Male

Observation Period Began: 02:22

Subject's Last Name: PUNDAY

Date/Time of Test: 12/05/2007 02:53

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:46
Subject Sample #1	0.111	02:48
Breath Volume	2.378	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample #2	0.106	02:52
Breath Volume	1.625	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: ADAMS, STEVE

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:20

Gender: Female

Observation Period Began: 02:54

Subject's Last Name: POWELL

Date/Time of Test: 12/05/2007 03:23

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Subject Sample #1	0.126	03:19
Breath Volume	1.351	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:21
Subject Sample #2	0.126	03:22
Breath Volume	1.125	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MORALES, MARIA, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: SCHELLHORN, R

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:15

Gender: Male

Observation Period Began: 03:30

Subject's Last Name: LIMA

Date/Time of Test: 12/05/2007 04:09

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:59	
Subject Sample #1	0.169	04:02	Volume Not Met
Breath Volume	0.089	04:02	
Air Blank	0.000	04:03	
Air Blank	0.000	04:04	
Subject Sample #2	0.144	04:07	Volume Not Met
Breath Volume	0.089	04:07	
Air Blank	0.000	04:08	
Control Test	0.078	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOSLEY, JENNIFER,

Breath Test Operator Agency: Orange County SO

Arresting Officer: SHAHEED, M

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 00:52

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: RAZKI

Date/Time of Test: 12/06/2007 01:32

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.081	01:25
Air Blank	0.000	01:26
Subject Sample #1	0.167	01:28
Breath Volume	3.140	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample #2	0.165	01:31
Breath Volume	1.785	01:31
Air Blank	0.000	01:31
Control Test	0.078	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOSLEY, JENNIFER,

Breath Test Operator Agency: Orange County SO

Arresting Officer: LOTTER, J

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:18

Gender: Male

Observation Period Began: 04:30

Subject's Last Name: PANGLE

Date/Time of Test: 12/06/2007 05:06

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.081	05:01
Air Blank	0.000	05:01
Subject Sample #1	0.133	05:02
Breath Volume	2.222	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample #2	0.134	05:05
Breath Volume	2.761	05:05
Air Blank	0.000	05:05
Control Test	0.080	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 12/06/2007 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOSLEY, JENNIFER,

Breath Test Operator Agency: Orange County SO

Arresting Officer: LOTTER, J

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 23:16

Gender: Female

Observation Period Began: 00:16

Subject's Last Name: ABSALON

Date/Time of Test: 12/07/2007 00:59

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:48	
Control Test	0.081	00:48	
Air Blank	0.000	00:49	
Subject Sample #1	0.197	00:52	Volume Not Met
Breath Volume	0.125	00:52	
Air Blank	0.000	00:52	
Air Blank	0.000	00:54	
Subject Sample #2	0.168	00:57	Volume Not Met
Breath Volume	0.683	00:57	
Air Blank	0.000	00:58	
Control Test	0.078	00:58	
Air Blank	0.000	00:59	
Diagnostics Check	OK	00:59	

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOSLEY, JENNIFER,

Breath Test Operator Agency: Orange County SO

Arresting Officer: CROSBY, C

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:12

Gender: Male

Observation Period Began: 03:33

Subject's Last Name: MERCADO

Date/Time of Test: 12/07/2007 04:08

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample #1	0.160	04:03
Breath Volume	1.238	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample #2	0.163	04:07
Breath Volume	2.261	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: ADAMS, STEVE

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 21:15

Gender: Male

Observation Period Began: 21:37

Subject's Last Name: SIMPKINS

Date/Time of Test: 12/07/2007 22:06

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Subject Sample #1	0.109	22:02
Breath Volume	1.882	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:04
Subject Sample #2	0.102	22:05
Breath Volume	2.289	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: BUTLER, KIM

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 23:36

Gender: Female

Observation Period Began: 00:25

Subject's Last Name: GALE

Date/Time of Test: 12/08/2007 00:56

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:50
Subject Sample #1	0.194	00:51
Breath Volume	1.371	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample #2	0.181	00:55
Breath Volume	2.089	00:55
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: ROSS, DAVID

Arresting Officer Agency: Maitland PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 00:48

Gender: Male

Observation Period Began: 01:28

Subject's Last Name: CARPENTER

Date/Time of Test: 12/08/2007 01:58

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:53
Air Blank	0.000	01:53
Subject Sample #1	0.160	01:54
Breath Volume	1.851	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.161	01:56
Breath Volume	1.582	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: LOTTER, J

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:29

Gender: Male

Observation Period Began: 03:03

Subject's Last Name: RICCIARDI

Date/Time of Test: 12/08/2007 03:30

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.081	03:28	
Air Blank	0.000	03:28	
Subject Sample #1		03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Control Test	0.080	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MILLS, A, G

Breath Test Operator Agency: Orange County SO

Arresting Officer: DENOCOLA, R

Arresting Officer Agency: Maitland PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:42

Gender: Male

Observation Period Began: 04:34

Subject's Last Name: GIZELT

Date/Time of Test: 12/08/2007 05:05

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Subject Sample #1	0.274	05:00
Breath Volume	1.859	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample #2	0.267	05:04
Breath Volume	2.066	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: HARMAN, MICHAEL

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 05:26

Gender: Male

Observation Period Began: 06:35

Subject's Last Name: GAYLE

Date/Time of Test: 12/08/2007 07:18

Diagnostics Check	OK	07:05	
Air Blank	0.000	07:06	
Control Test	0.081	07:06	
Air Blank	0.000	07:06	
Subject Sample #1	0.123	07:10	Volume Not Met
Breath Volume	0.632	07:10	
Air Blank	0.000	07:10	
Air Blank	0.000	07:12	
Subject Sample #2	0.171	07:14	
Breath Volume	2.710	07:14	
Air Blank	0.000	07:14	
Air Blank	0.000	07:16	
Subject Sample #3	0.153	07:17	
Breath Volume	1.355	07:17	
Air Blank	0.000	07:17	
Control Test	0.078	07:18	
Air Blank	0.000	07:18	
Diagnostics Check	OK	07:18	

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: STRIKLAND, STEVE

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 07:26

Gender: Male

Observation Period Began: 08:20

Subject's Last Name: KALAM

Date/Time of Test: 12/08/2007 08:50

Diagnostics Check	OK	08:43
Air Blank	0.000	08:43
Control Test	0.081	08:43
Air Blank	0.000	08:44
Subject Sample #1	0.116	08:45
Breath Volume	1.742	08:45
Air Blank	0.000	08:46
Air Blank	0.000	08:47
Subject Sample #2	0.117	08:48
Breath Volume	1.996	08:48
Air Blank	0.000	08:49
Control Test	0.079	08:49
Air Blank	0.000	08:50
Diagnostics Check	OK	08:50

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOSLEY, JENNIFER,

Breath Test Operator Agency: Orange County SO

Arresting Officer: ADAMS, S

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 20:10

Gender: Male

Observation Period Began: 20:22

Subject's Last Name: MILLER

Date/Time of Test: 12/08/2007 20:52

Diagnostics Check	OK	20:44
Air Blank	0.000	20:45
Control Test	0.081	20:45
Air Blank	0.000	20:46
Subject Sample #1	0.147	20:47
Breath Volume	1.289	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample #2	0.155	20:50
Breath Volume	1.984	20:50
Air Blank	0.000	20:51
Control Test	0.079	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: ADAMS, STEVE

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:45

Gender: Male

Observation Period Began: 02:01

Subject's Last Name: LOPEZ-PEREIRA

Date/Time of Test: 12/09/2007 02:35

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample #1	0.156	02:30
Breath Volume	1.328	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample #2	0.158	02:33
Breath Volume	1.855	02:33
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOSLEY, JENNIFER,

Breath Test Operator Agency: Orange County SO

Arresting Officer: FAST, W

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 02:56

Subject's Last Name: MERCADO

Date/Time of Test: 12/09/2007 03:23

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Subject Sample #1	0.086	03:19
Breath Volume	1.789	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample #2	0.084	03:22
Breath Volume	1.910	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: STONE, WILLIAM

Arresting Officer Agency: Central FL Univ PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:31

Gender: Male

Observation Period Began: 04:02

Subject's Last Name: REYES MARTINEZ

Date/Time of Test: 12/09/2007 04:35

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Subject Sample #1	0.176	04:28
Breath Volume	1.664	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample #2	0.167	04:34
Breath Volume	1.484	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOSLEY, JENNIFER,

Breath Test Operator Agency: Orange County SO

Arresting Officer: WILLIAMS, R

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:30

Gender: Male

Observation Period Began: 04:24

Subject's Last Name: OCONNOR

Date/Time of Test: 12/09/2007 04:52

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:46
Subject Sample #1	0.195	04:47
Breath Volume	1.339	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:49
Subject Sample #2	0.209	04:50
Breath Volume	2.078	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: WILLIAMS, STEPHEN

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 05:15

Gender: Male

Observation Period Began: 06:23

Subject's Last Name: ARBELO

Date/Time of Test: 12/09/2007 06:53

Diagnostics Check	OK	06:46
Air Blank	0.000	06:47
Control Test	0.080	06:47
Air Blank	0.000	06:47
Subject Sample #1	0.210	06:49
Breath Volume	2.656	06:49
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample #2	0.216	06:52
Breath Volume	1.644	06:52
Air Blank	0.000	06:52
Control Test	0.079	06:53
Air Blank	0.000	06:53
Diagnostics Check	OK	06:53

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MILLS, A, G

Breath Test Operator Agency: Orange County SO

Arresting Officer: WALWYN, CG

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:27

Gender: Male

Observation Period Began: 04:03

Subject's Last Name: FOURNIER GINES

Date/Time of Test: 12/10/2007 04:34

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.081	04:27
Air Blank	0.000	04:27
Subject Sample #1	0.132	04:29
Breath Volume	2.050	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:31
Subject Sample #2	0.131	04:33
Breath Volume	2.160	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: SCHELLHORN, ROBERT

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:55

Gender: Male

Observation Period Began: 02:08

Subject's Last Name: DELGADO

Date/Time of Test: 12/12/2007 02:38

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:33
Air Blank	0.000	02:33
Subject Sample #1	0.079	02:33
Breath Volume	2.402	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample #2	0.081	02:36
Breath Volume	2.148	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 12/12/2007 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MORALES, MARIA, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: DEWITT, T

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 22:36

Gender: Male

Observation Period Began: 23:22

Subject's Last Name: WILSON

Date/Time of Test: 12/12/2007 23:59

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:53
Subject Sample #1	0.185	23:54
Breath Volume	3.363	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:56
Subject Sample #2	0.185	23:57
Breath Volume	3.222	23:57
Air Blank	0.000	23:58
Control Test	0.080	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: BARLEY, JUSTIN

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:38

Gender: Male

Observation Period Began: 02:20

Subject's Last Name: GHOLSTON

Date/Time of Test: 12/13/2007 02:56

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.079	02:45	
Air Blank	0.000	02:46	
Subject Sample #1	0.181	02:49	Volume Not Met
Breath Volume	0.410	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample #2	0.194	02:55	Volume Not Met
Breath Volume	0.250	02:55	
Air Blank	0.000	02:55	
Control Test	0.080	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: BARLEY, JUSTIN

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:38

Gender: Male

Observation Period Began: 02:20

Subject's Last Name: GHOLSTON

Date/Time of Test: 12/13/2007 03:05

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Subject Sample #1		03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.080	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MORALES, MARIA, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: RECCA, M

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:56

Gender: Male

Observation Period Began: 03:47

Subject's Last Name: SILVERMAN

Date/Time of Test: 12/15/2007 04:16

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:11
Subject Sample #1	0.123	04:12
Breath Volume	2.812	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample #2	0.127	04:15
Breath Volume	2.554	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: STANLEY, M

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 00:45

Gender: Female

Observation Period Began: 01:23

Subject's Last Name: COOLEY

Date/Time of Test: 12/18/2007 02:00

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.081	01:49	
Air Blank	0.000	01:50	
Subject Sample #1	0.218	01:53	Volume Not Met
Breath Volume	0.421	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample #2	0.235	01:59	Volume Not Met
Breath Volume	0.207	01:59	
Air Blank	0.000	02:00	
Control Test	0.077	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: STANLEY, M

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 00:45

Gender: Female

Observation Period Began: 01:23

Subject's Last Name: COOLEY

Date/Time of Test: 12/18/2007 02:18

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:04	
Control Test	0.078	02:04	
Air Blank	0.000	02:04	
Subject Sample #1	0.225	02:06	
Breath Volume	1.390	02:06	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample #2	0.196	02:12	Volume Not Met
Breath Volume	0.679	02:12	
Air Blank	0.000	02:13	
Air Blank	0.000	02:14	
Subject Sample #3		02:16	Subject Test Refused
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Control Test	0.076	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: MURCH, N

Arresting Officer Agency: Winter Garden PD

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:18

Gender: Male

Observation Period Began: 02:56

Subject's Last Name: STEWART

Date/Time of Test: 12/18/2007 03:41

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Subject Sample #1	0.132	03:37
Breath Volume	3.234	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample #2	0.134	03:40
Breath Volume	3.308	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: YUKNAVAGE, B

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:35

Gender: Male

Observation Period Began: 03:50

Subject's Last Name: WEIGLE

Date/Time of Test: 12/18/2007 04:40

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.081	04:28	
Air Blank	0.000	04:28	
Subject Sample #1	0.138	04:31	Volume Not Met
Breath Volume	0.171	04:31	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample #2	0.135	04:35	
Breath Volume	2.054	04:35	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample #3	0.131	04:39	
Breath Volume	1.437	04:39	
Air Blank	0.000	04:39	
Control Test	0.080	04:40	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: MC CARTHY, ROBERT

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 00:22

Gender: Male

Observation Period Began: 00:38

Subject's Last Name: ANGUIANOSANCHEZ

Date/Time of Test: 12/19/2007 01:10

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.081	01:04
Air Blank	0.000	01:04
Subject Sample #1	0.137	01:06
Breath Volume	1.691	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample #2	0.147	01:09
Breath Volume	3.609	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: RECCA, MARC

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:40

Gender: Female

Observation Period Began: 02:18

Subject's Last Name: MUSA

Date/Time of Test: 12/19/2007 02:50

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample #1	0.176	02:45
Breath Volume	1.757	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample #2	0.174	02:48
Breath Volume	1.515	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: GARDINER, TODD

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:46

Gender: Male

Observation Period Began: 02:16

Subject's Last Name: MACKENZIE

Date/Time of Test: 12/20/2007 02:48

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Subject Sample #1	0.141	02:44
Breath Volume	1.277	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample #2	0.148	02:47
Breath Volume	2.222	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 12/20/2007 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: YUKNAVAGE, LEONARD

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 23:43

Gender: Male

Observation Period Began: 00:33

Subject's Last Name: PRICHNICK

Date/Time of Test: 12/22/2007 01:05

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	01:00
Subject Sample #1	0.144	01:00
Breath Volume	2.621	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample #2	0.140	01:04
Breath Volume	1.824	01:04
Air Blank	0.000	01:04
Control Test	0.081	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 12/27/2007 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: SCHELLHORN, R

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:55

Gender: Male

Observation Period Began: 02:15

Subject's Last Name: VALDIVIESO

Date/Time of Test: 12/22/2007 02:49

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:41
Air Blank	0.000	02:41
Subject Sample #1	0.189	02:42
Breath Volume	3.027	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample #2	0.213	02:45
Breath Volume	1.562	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample #3	0.204	02:48
Breath Volume	2.761	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 12/27/2007 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: ADAMS, STEVE

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:25

Gender: Male

Observation Period Began: 03:24

Subject's Last Name: LACE

Date/Time of Test: 12/22/2007 03:57

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:55	
Subject Sample #1		03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:56	
Control Test	0.081	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 12/27/2007 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: SANCHEZ, J

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 04:21

Gender: Male

Observation Period Began: 05:25

Subject's Last Name: FILLMAN

Date/Time of Test: 12/22/2007 05:53

Diagnostics Check	OK	05:47
Air Blank	0.000	05:48
Control Test	0.081	05:48
Air Blank	0.000	05:49
Subject Sample #1	0.166	05:49
Breath Volume	1.281	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample #2	0.167	05:52
Breath Volume	1.281	05:52
Air Blank	0.000	05:53
Control Test	0.080	05:53
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 12/27/2007 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: RIGGS, JEFF

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:38

Gender: Male

Observation Period Began: 03:23

Subject's Last Name: HANIF

Date/Time of Test: 12/23/2007 04:03

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.081	03:48	
Air Blank	0.000	03:49	
Subject Sample #1	0.212	03:51	
Breath Volume	1.738	03:51	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample #2	0.190	03:56	Volume Not Met
Breath Volume	0.601	03:56	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample #3	0.213	04:02	
Breath Volume	1.218	04:02	
Air Blank	0.000	04:02	
Control Test	0.077	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 12/27/2007 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: WILSON, A

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 04:21

Gender: Male

Observation Period Began: 05:06

Subject's Last Name: RUIZMARTUNO

Date/Time of Test: 12/29/2007 05:42

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.081	05:36
Air Blank	0.000	05:36
Subject Sample #1	0.193	05:37
Breath Volume	2.378	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample #2	0.190	05:41
Breath Volume	2.519	05:41
Air Blank	0.000	05:41
Control Test	0.080	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 01/03/2008 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LANDROVE, ROBERT,

Breath Test Operator Agency: Orange County SO

Arresting Officer: RECCA, M

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:08

Gender: Male

Observation Period Began: 02:45

Subject's Last Name: DINGMAN

Date/Time of Test: 12/30/2007 03:19

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Subject Sample #1	0.169	03:13
Breath Volume	1.183	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample #2	0.166	03:17
Breath Volume	2.011	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 01/03/2008 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: WILSON, AARON

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:02

Gender: Male

Observation Period Began: 03:40

Subject's Last Name: HAMMAN

Date/Time of Test: 12/31/2007 04:23

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.081	04:09	
Air Blank	0.000	04:09	
Subject Sample #1	0.154	04:12	Volume Not Met
Breath Volume	0.179	04:12	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample #2	0.219	04:15	
Breath Volume	1.832	04:15	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample #3		04:21	Slope Not Met
Breath Volume	0.820	04:21	
Air Blank	0.000	04:22	
Control Test	0.080	04:22	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 01/03/2008 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001416

Instrument Registered To: ORANGE COUNTY SO

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: WILSON, AARON

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 12/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:02

Gender: Male

Observation Period Began: 03:40

Subject's Last Name: HAMMAN

Date/Time of Test: 12/31/2007 04:33

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:26	
Subject Sample #1	0.187	04:29	Volume Not Met
Breath Volume	0.984	04:29	
Air Blank	0.000	04:30	
Air Blank	0.000	04:31	
Subject Sample #2		04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.078	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 01/03/2008 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: GRANT, CHARLES

Arresting Officer Agency: FHP Troop K

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 00:31

Gender: Male

Observation Period Began: 01:04

Subject's Last Name: BUSCAGLIA

Date/Time of Test: 12/01/2007 01:34

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample #1	0.080	01:28
Breath Volume	1.238	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample #2	0.084	01:32
Breath Volume	1.164	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 12/06/2007 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: SMITH, B

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 23:27

Gender: Female

Observation Period Began: 23:57

Subject's Last Name: TAYLOR

Date/Time of Test: 12/02/2007 00:28

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:23	
Control Test	0.079	00:24	
Air Blank	0.000	00:24	
Subject Sample #1		00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Control Test	0.080	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 12/06/2007 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: REMIREZ, J

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 04:57

Gender: Male

Observation Period Began: 05:20

Subject's Last Name: IVESTER

Date/Time of Test: 12/02/2007 06:00

Diagnostics Check	OK	05:48	
Air Blank	0.000	05:49	
Control Test	0.080	05:49	
Air Blank	0.000	05:49	
Subject Sample #1	0.124	05:50	
Breath Volume	1.906	05:50	
Air Blank	0.000	05:51	
Air Blank	0.000	05:53	
Subject Sample #2	0.120	05:56	Volume Not Met
Breath Volume	0.550	05:56	
Air Blank	0.000	05:57	
Air Blank	0.000	05:58	
Subject Sample #3	0.117	05:59	
Breath Volume	1.917	05:59	
Air Blank	0.000	05:59	
Control Test	0.077	06:00	
Air Blank	0.000	06:00	
Diagnostics Check	OK	06:00	

Data Download Date/Time: 12/06/2007 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: REGAN, W

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 22:54

Gender: Male

Observation Period Began: 23:18

Subject's Last Name: SIERZEGA

Date/Time of Test: 12/03/2007 23:47

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:42
Air Blank	0.000	23:42
Subject Sample #1	0.271	23:43
Breath Volume	1.867	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample #2	0.269	23:46
Breath Volume	2.691	23:46
Air Blank	0.000	23:46
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 12/06/2007 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: REGAN, W

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:58

Gender: Male

Observation Period Began: 03:19

Subject's Last Name: FLETCHER

Date/Time of Test: 12/04/2007 03:45

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:40
Air Blank	0.000	03:40
Subject Sample #1	0.049	03:41
Breath Volume	2.789	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample #2	0.046	03:44
Breath Volume	1.722	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 12/06/2007 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: VAUGHN, J

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:21

Gender: Male

Observation Period Began: 04:09

Subject's Last Name: SHELTON JR

Date/Time of Test: 12/04/2007 04:49

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:36	
Control Test	0.081	04:36	
Air Blank	0.000	04:37	
Subject Sample #1	0.108	04:40	Volume Not Met
Breath Volume	0.160	04:40	
Air Blank	0.000	04:41	
Air Blank	0.000	04:42	
Subject Sample #2	0.105	04:44	
Breath Volume	1.183	04:44	
Air Blank	0.000	04:45	
Air Blank	0.000	04:47	
Subject Sample #3	0.099	04:48	
Breath Volume	1.171	04:48	
Air Blank	0.000	04:48	
Control Test	0.076	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 12/06/2007 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: HAWKINS, D

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:14

Gender: Female

Observation Period Began: 01:53

Subject's Last Name: GROSECLOSE

Date/Time of Test: 12/05/2007 02:21

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample #1	0.073	02:17
Breath Volume	3.136	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:19
Subject Sample #2	0.066	02:20
Breath Volume	1.253	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 12/06/2007 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: LOTTER, J

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:08

Gender: Female

Observation Period Began: 03:01

Subject's Last Name: MEZABOLTON

Date/Time of Test: 12/05/2007 03:28

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample #1	0.117	03:24
Breath Volume	2.253	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample #2	0.107	03:26
Breath Volume	1.492	03:26
Air Blank	0.000	03:27
Control Test	0.077	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 12/06/2007 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: WILSON, AARON

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:34

Gender: Male

Observation Period Began: 03:12

Subject's Last Name: FISCHER

Date/Time of Test: 12/07/2007 03:46

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Subject Sample #1	0.023	03:41
Breath Volume	3.984	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample #2	0.020	03:44
Breath Volume	3.546	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 12/12/2007 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: CANER, OSVALDO,

Breath Test Operator Agency: Orange County SO

Arresting Officer: CLAYTON, LOURDES

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:38

Gender: Male

Observation Period Began: 05:30

Subject's Last Name: LANCE

Date/Time of Test: 12/07/2007 06:06

Diagnostics Check	OK	06:00
Air Blank	0.000	06:00
Control Test	0.080	06:00
Air Blank	0.000	06:01
Subject Sample #1	0.062	06:01
Breath Volume	3.734	06:01
Air Blank	0.000	06:02
Air Blank	0.000	06:04
Subject Sample #2	0.064	06:04
Breath Volume	3.613	06:04
Air Blank	0.000	06:05
Control Test	0.080	06:05
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 12/12/2007 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: RODRIGUEZ, MICHAEL, J

Breath Test Operator Agency: Orange County SO

Arresting Officer: KRALICH, B

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 01:26

Gender: Male

Observation Period Began: 01:38

Subject's Last Name: KEO KOSAL

Date/Time of Test: 12/08/2007 02:09

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Subject Sample #1	0.037	02:04
Breath Volume	2.054	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample #2	0.039	02:07
Breath Volume	2.597	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:09

Data Download Date/Time: 12/12/2007 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: GORBERS, J

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:28

Gender: Male

Observation Period Began: 04:39

Subject's Last Name: YOUNG

Date/Time of Test: 12/09/2007 05:21

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:10	
Control Test	0.080	05:10	
Air Blank	0.000	05:11	
Subject Sample #1	0.201	05:14	Volume Not Met
Breath Volume	0.015	05:14	
Air Blank	0.000	05:15	
Air Blank	0.000	05:16	
Subject Sample #2	0.215	05:19	Volume Not Met
Breath Volume	0.085	05:19	
Air Blank	0.000	05:20	
Control Test	0.077	05:20	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 12/12/2007 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: REGAN, W

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 22:59

Gender: Male

Observation Period Began: 23:29

Subject's Last Name: MEDINA

Date/Time of Test: 12/09/2007 23:58

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:53
Subject Sample #1	0.165	23:54
Breath Volume	2.980	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:56
Subject Sample #2	0.164	23:57
Breath Volume	3.546	23:57
Air Blank	0.000	23:57
Control Test	0.079	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 12/12/2007 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: SCHELLHORN, R

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 03:30

Gender: Male

Observation Period Began: 03:42

Subject's Last Name: SANTIAGO

Date/Time of Test: 12/10/2007 04:15

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample #1	0.021	04:11
Breath Volume	2.394	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample #2	0.021	04:14
Breath Volume	2.300	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 12/12/2007 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: GATZEY, J

Arresting Officer Agency: Orange County SO

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 21:25

Gender: Male

Observation Period Began: 22:30

Subject's Last Name: GRIFFIN JR

Date/Time of Test: 12/10/2007 23:00

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Subject Sample #1	0.062	22:56
Breath Volume	1.707	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample #2	0.061	22:59
Breath Volume	1.597	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 12/12/2007 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001418

Instrument Registered To: ORANGE COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: ARMAND, ELVIE, D

Breath Test Operator Agency: Orange County SO

Arresting Officer: CLARKE, A

Arresting Officer Agency: Orlando PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 714102I

Dry Gas Expiry Date: 05/22/2009

Violation Code: DUI

Arrest Time: 02:16

Gender: Female

Observation Period Began: 02:39

Subject's Last Name: OBRIEN

Date/Time of Test: 12/11/2007 03:08

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.081	03:06	
Air Blank	0.000	03:06	
Subject Sample #1		03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 12/12/2007 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001439

Instrument Registered To: MARTIN COUNTY SO

Software Version: 8100.27

Breath Test Operator: FOLBRECHT, JON, H

Breath Test Operator Agency: Martin County SO

Arresting Officer: FOLBRECHT, JON

Arresting Officer Agency: Martin County SO

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 61910I

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 19:09

Gender: Male

Observation Period Began: 19:20

Subject's Last Name: TEPAZ-SUY

Date/Time of Test: 12/29/2007 19:50

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Subject Sample #1	0.186	19:44
Breath Volume	1.957	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample #2	0.199	19:48
Breath Volume	2.011	19:48
Air Blank	0.000	19:49
Control Test	0.079	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 01/28/2008 21:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: GATTIS, GARY, M

Breath Test Operator Agency: Clay County SO

Arresting Officer: FROST, K

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 02:18

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: PALMER

Date/Time of Test: 12/01/2007 03:41

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Subject Sample #1	0.152	03:36
Breath Volume	2.054	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample #2	0.155	03:40
Breath Volume	1.503	03:40
Air Blank	0.000	03:40
Control Test	0.077	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 12/10/2007 13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: FRISCO, ROXANNE,

Breath Test Operator Agency: Clay County SO

Arresting Officer: FRISCO, ROXANNE

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 11/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: ADMIN

Arrest Time: 09:00

Gender: Female

Observation Period Began: 09:50

Subject's Last Name: CRISSMAN

Date/Time of Test: 12/05/2007 10:20

Diagnostics Check	OK	10:13
Air Blank	0.000	10:13
Control Test	0.082	10:13
Air Blank	0.000	10:14
Subject Sample #1	0.052	10:15
Breath Volume	1.250	10:15
Air Blank	0.000	10:15
Air Blank	0.000	10:17
Subject Sample #2	0.051	10:19
Breath Volume	1.800	10:19
Air Blank	0.000	10:20
Control Test	0.081	10:20
Air Blank	0.000	10:20
Diagnostics Check	OK	10:20

Data Download Date/Time: 12/10/2007 13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: TOMLINSON, JAMES, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: BURKE, ANDY

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 02:03

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: LAHAY

Date/Time of Test: 12/13/2007 03:02

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Subject Sample #1	0.095	02:55	Volume Not Met
Breath Volume	0.367	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample #2		03:01	No Sample Provided
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.079	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARWOOD, C, M

Breath Test Operator Agency: Clay County SO

Arresting Officer: BURKE, A

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 20:01

Gender: Male

Observation Period Began: 21:17

Subject's Last Name: GARCIA

Date/Time of Test: 12/13/2007 21:49

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.081	21:41	
Air Blank	0.000	21:42	
Subject Sample #1	0.160	21:45	Volume Not Met
Breath Volume	0.167	21:45	
Air Blank	0.000	21:46	
Air Blank	0.000	21:48	
Subject Sample #2		21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:48	
Control Test	0.078	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: BURCH, CHASON, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: TENNANT, R

Arresting Officer Agency: Orange Park PD

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 01:11

Gender: Female

Observation Period Began: 02:51

Subject's Last Name: THOMAS

Date/Time of Test: 12/19/2007 03:20

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:13	
Subject Sample #1		03:16	No Sample Provided
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:19	
Subject Sample #2		03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: TOMLINSON, JAMES, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: CONNORS, S

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 23:01

Gender: Male

Observation Period Began: 23:59

Subject's Last Name: O BOYLE

Date/Time of Test: 12/22/2007 00:31

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.082	00:21	
Air Blank	0.000	00:21	
Subject Sample #1		00:24	No Sample Provided
Breath Volume	0.000	00:24	
Air Blank	0.000	00:25	
Air Blank	0.000	00:27	
Subject Sample #2		00:30	No Sample Provided
Breath Volume	0.000	00:30	
Air Blank	0.000	00:30	
Control Test	0.081	00:31	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WILKERSON, M, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: CLARK, J

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 03:49

Gender: Male

Observation Period Began: 04:05

Subject's Last Name: JONES

Date/Time of Test: 12/22/2007 04:32

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.081	04:26
Air Blank	0.000	04:26
Subject Sample #1	0.146	04:27
Breath Volume	1.453	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample #2	0.142	04:31
Breath Volume	1.527	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: FRISCO, ROXANNE,

Breath Test Operator Agency: Clay County SO

Arresting Officer: BRIGGS, G

Arresting Officer Agency: Orange Park PD

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 03:18

Gender: Male

Observation Period Began: 06:07

Subject's Last Name: DRURY

Date/Time of Test: 12/22/2007 06:34

Diagnostics Check	OK	06:27
Air Blank	0.000	06:28
Control Test	0.081	06:28
Air Blank	0.000	06:28
Subject Sample #1	0.045	06:29
Breath Volume	1.722	06:29
Air Blank	0.000	06:30
Air Blank	0.000	06:31
Subject Sample #2	0.042	06:32
Breath Volume	1.367	06:32
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: DAY, KENNETH, J

Breath Test Operator Agency: Clay County SO

Arresting Officer: RYAN, J

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 18:48

Gender: Male

Observation Period Began: 19:21

Subject's Last Name: HERNANDEZ

Date/Time of Test: 12/23/2007 19:48

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.081	19:42
Air Blank	0.000	19:43
Subject Sample #1	0.132	19:43
Breath Volume	2.359	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample #2	0.130	19:46
Breath Volume	2.300	19:46
Air Blank	0.000	19:47
Control Test	0.078	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: TOMLINSON, JAMES, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: BROWN, TJ

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 01:04

Gender: Female

Observation Period Began: 01:41

Subject's Last Name: WHITE

Date/Time of Test: 12/27/2007 02:12

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.082	02:05
Air Blank	0.000	02:06
Subject Sample #1	0.152	02:07
Breath Volume	1.675	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample #2	0.151	02:10
Breath Volume	1.140	02:10
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:12

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WILKERSON, M, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: BROWN, T

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 01:13

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: FIALLOS

Date/Time of Test: 12/28/2007 02:28

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample #1	0.150	02:24
Breath Volume	1.550	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample #2	0.149	02:27
Breath Volume	1.757	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: BURCH, CHASON, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: KLIDIES, G

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 01:51

Gender: Male

Observation Period Began: 02:44

Subject's Last Name: GIBSON

Date/Time of Test: 12/28/2007 03:12

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Subject Sample #1	0.092	03:07
Breath Volume	1.312	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample #2	0.088	03:10
Breath Volume	2.027	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: FRISCO, ROXANNE,

Breath Test Operator Agency: Clay County SO

Arresting Officer: BROWN, T

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 02:36

Gender: Male

Observation Period Began: 03:10

Subject's Last Name: EUSTICE

Date/Time of Test: 12/29/2007 03:43

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:32	
Subject Sample #1	0.128	03:35	Volume Not Met
Breath Volume	0.816	03:35	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample #2	0.155	03:38	
Breath Volume	2.210	03:38	
Air Blank	0.000	03:39	
Air Blank	0.000	03:40	
Subject Sample #3	0.163	03:41	
Breath Volume	2.128	03:41	
Air Blank	0.000	03:42	
Control Test	0.079	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: DUFFY, DALE, V

Breath Test Operator Agency: Clay County SO

Arresting Officer: BROWN, THOMAS

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: DUI

Arrest Time: 23:22

Gender: Male

Observation Period Began: 00:12

Subject's Last Name: WHITNEY

Date/Time of Test: 12/30/2007 00:45

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Subject Sample #1	0.213	00:40
Breath Volume	1.562	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample #2	0.213	00:43
Breath Volume	2.039	00:43
Air Blank	0.000	00:44
Control Test	0.077	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001443

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: FRISCO, RICKY, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: FRISCO, RICKY

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 12/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7107011

Dry Gas Expiry Date: 04/18/2009

Violation Code: ADMIN

Arrest Time: 14:00

Gender: Female

Observation Period Began: 14:00

Subject's Last Name: OSGOOD

Date/Time of Test: 12/30/2007 14:33

Diagnostics Check	OK	14:27	
Air Blank	0.000	14:28	
Control Test	0.081	14:28	
Air Blank	0.000	14:28	
Subject Sample #1		14:31	Subject Test Refused
Breath Volume	0.050	14:31	
Air Blank	0.000	14:32	
Control Test	0.082	14:32	
Air Blank	0.000	14:33	
Diagnostics Check	OK	14:33	

Data Download Date/Time: 01/18/2008 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001469

Instrument Registered To: HURLBURT FIELD PD

Software Version: 8100.27

Breath Test Operator: PADILLA, ALBERTO,

Breath Test Operator Agency: Bay County SO

Arresting Officer: LLOYD, TIMOTHY

Arresting Officer Agency: Bay County SO

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7053011

Dry Gas Expiry Date: 02/27/2009

Violation Code: DUI

Arrest Time: 01:25

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: HOHENSTEIN

Date/Time of Test: 12/15/2007 02:13

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Subject Sample #1	0.000	02:06	Volume Not Met
Breath Volume	0.457	02:06	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample #2	0.000	02:12	Volume Not Met
Breath Volume	0.605	02:12	
Air Blank	0.000	02:12	
Control Test	0.080	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 12/15/2007 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001469

Instrument Registered To: HURLBURT FIELD PD

Software Version: 8100.27

Breath Test Operator: PADILLA, ALBERTO,

Breath Test Operator Agency: Bay County SO

Arresting Officer: LLOYD, TIMOTHY

Arresting Officer Agency: Bay County SO

Last Agency Inspection: 11/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7053011

Dry Gas Expiry Date: 02/27/2009

Violation Code: DUI

Arrest Time: 01:25

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: HOHENSTEIN

Date/Time of Test: 12/15/2007 02:32

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.079	02:17	
Air Blank	0.000	02:17	
Subject Sample #1	0.000	02:21	Volume Not Met
Breath Volume	0.523	02:21	
Air Blank	0.000	02:21	
Air Blank	0.000	02:23	
Subject Sample #2	0.000	02:26	
Breath Volume	1.925	02:26	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample #3	0.000	02:31	
Breath Volume	1.777	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 12/15/2007 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001469

Instrument Registered To: HURLBURT FIELD PD

Software Version: 8100.27

Breath Test Operator: SPEAKMAN, ROBERT, J

Breath Test Operator Agency: Hurlburt Field PD

Arresting Officer: CAIN, MAURICE

Arresting Officer Agency: Hurlburt Field PD

Last Agency Inspection: 12/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7053011

Dry Gas Expiry Date: 02/27/2009

Violation Code: DUI

Arrest Time: 10:17

Gender: Male

Observation Period Began: 10:50

Subject's Last Name: LOPEZ

Date/Time of Test: 12/23/2007 11:34

Diagnostics Check	OK	11:27
Air Blank	0.000	11:27
Control Test	0.080	11:27
Air Blank	0.000	11:28
Subject Sample #1	0.173	11:29
Breath Volume	2.625	11:29
Air Blank	0.000	11:30
Air Blank	0.000	11:31
Subject Sample #2	0.167	11:32
Breath Volume	2.531	11:32
Air Blank	0.000	11:33
Control Test	0.079	11:33
Air Blank	0.000	11:34
Diagnostics Check	OK	11:34

Data Download Date/Time: 01/12/2008 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001649

Instrument Registered To: KENNETH CITY PD

Software Version: 8100.27

Breath Test Operator: SUERETH, DONALD, P

Breath Test Operator Agency: FDLE

Arresting Officer: OFFICER, ARRESTING

Arresting Officer Agency: FDLE

Last Agency Inspection: 12/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 7096011

Dry Gas Expiry Date: 04/08/2009

Violation Code: SYS CHECK

Arrest Time: 10:30

Gender: Male

Observation Period Began: 10:45

Subject's Last Name: SUBJECT

Date/Time of Test: 12/18/2007 11:17

Diagnostics Check	OK	11:11
Air Blank	0.000	11:11
Control Test	0.080	11:11
Air Blank	0.000	11:12
Subject Sample #1	0.000	11:13
Breath Volume	2.488	11:13
Air Blank	0.000	11:13
Air Blank	0.000	11:15
Subject Sample #2	0.000	11:16
Breath Volume	2.796	11:16
Air Blank	0.000	11:16
Control Test	0.078	11:16
Air Blank	0.000	11:17
Diagnostics Check	OK	11:17

Data Download Date/Time: 12/18/2007 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: SALLS, DANIEL, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: SALLS, DANIEL

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 11/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 02:38

Gender: Male

Observation Period Began: 03:36

Subject's Last Name: CHAVEZ

Date/Time of Test: 12/01/2007 04:08

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:01
Subject Sample #1	0.099	04:03
Breath Volume	1.816	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample #2	0.103	04:07
Breath Volume	1.324	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 12/04/2007 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: GREENE, DOUGLAS, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: BLACKWELL, STEPHEN

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 11/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 20:54

Gender: Male

Observation Period Began: 21:30

Subject's Last Name: ACEVEDO

Date/Time of Test: 12/01/2007 22:00

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.080	21:55
Air Blank	0.000	21:55
Subject Sample #1	0.187	21:56
Breath Volume	4.152	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:58
Subject Sample #2	0.191	21:59
Breath Volume	4.210	21:59
Air Blank	0.000	21:59
Control Test	0.079	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 12/04/2007 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: GREENE, DOUGLAS, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: SMITH, RASHAAD

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 11/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 00:09

Gender: Male

Observation Period Began: 00:40

Subject's Last Name: LANDEROS-LARA

Date/Time of Test: 12/02/2007 01:10

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.079	01:05
Air Blank	0.000	01:05
Subject Sample #1	0.173	01:05
Breath Volume	1.875	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample #2	0.172	01:08
Breath Volume	2.929	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 12/04/2007 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: SALLS, DANIEL, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: SALLS, DANIEL

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:34

Gender: Male

Observation Period Began: 04:02

Subject's Last Name: DELGADO

Date/Time of Test: 12/08/2007 04:30

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:25
Air Blank	0.000	04:25
Subject Sample #1	0.204	04:26
Breath Volume	1.414	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample #2	0.196	04:28
Breath Volume	1.589	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: INGRAM, GEORGE, D

Breath Test Operator Agency: Collier County SO

Arresting Officer: TESS, DANIEL

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 14:37

Gender: Male

Observation Period Began: 14:50

Subject's Last Name: VASQUEZ

Date/Time of Test: 12/08/2007 15:32

Diagnostics Check	OK	15:26
Air Blank	0.000	15:27
Control Test	0.079	15:27
Air Blank	0.000	15:27
Subject Sample #1	0.288	15:28
Breath Volume	1.804	15:28
Air Blank	0.000	15:29
Air Blank	0.000	15:30
Subject Sample #2	0.297	15:31
Breath Volume	1.742	15:31
Air Blank	0.000	15:31
Control Test	0.079	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: SALLS, DANIEL, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: SALLS, DANIEL

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 00:29

Gender: Male

Observation Period Began: 00:50

Subject's Last Name: SANTANA

Date/Time of Test: 12/15/2007 01:19

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Subject Sample #1	0.288	01:15
Breath Volume	2.277	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample #2	0.278	01:18
Breath Volume	1.191	01:18
Air Blank	0.000	01:18
Control Test	0.077	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: SALLS, DANIEL, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: SALLS, DANIEL

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 00:11

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: JUAREZ GARCIA

Date/Time of Test: 12/15/2007 01:45

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.077	01:39
Air Blank	0.000	01:39
Subject Sample #1	0.173	01:39
Breath Volume	1.597	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample #2	0.167	01:43
Breath Volume	1.722	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: SALLS, DANIEL, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: SALLS, DANIEL

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 23:58

Gender: Male

Observation Period Began: 01:17

Subject's Last Name: GARCIA

Date/Time of Test: 12/15/2007 01:55

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Subject Sample #1	0.130	01:49
Breath Volume	1.332	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample #2	0.136	01:53
Breath Volume	2.015	01:53
Air Blank	0.000	01:54
Control Test	0.078	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: LOPEZ, JACOB,

Breath Test Operator Agency: Collier County SO

Arresting Officer: MCCLENITHAN, ERNEST

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 13:29

Gender: Male

Observation Period Began: 14:05

Subject's Last Name: MENDOZALOPEZ

Date/Time of Test: 12/15/2007 14:40

Diagnostics Check	OK	14:33
Air Blank	0.000	14:34
Control Test	0.079	14:34
Air Blank	0.000	14:35
Subject Sample #1	0.190	14:35
Breath Volume	2.343	14:35
Air Blank	0.000	14:36
Air Blank	0.000	14:38
Subject Sample #2	0.172	14:38
Breath Volume	1.500	14:38
Air Blank	0.000	14:39
Control Test	0.079	14:39
Air Blank	0.000	14:40
Diagnostics Check	OK	14:40

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: GREENE, DOUGLAS, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: ALONSO, JESUS

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 20:28

Gender: Male

Observation Period Began: 21:00

Subject's Last Name: GONZALEZ

Date/Time of Test: 12/16/2007 21:30

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:25
Subject Sample #1	0.196	21:25
Breath Volume	1.136	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:27
Subject Sample #2	0.180	21:29
Breath Volume	2.183	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: SALLS, DANIEL, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: SALLS, DANIEL

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 00:45

Gender: Male

Observation Period Began: 00:56

Subject's Last Name: SEMEIN

Date/Time of Test: 12/22/2007 01:30

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample #1	0.070	01:25
Breath Volume	1.300	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample #2	0.073	01:28
Breath Volume	1.496	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: SALLS, DANIEL, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: SALLS, DANIEL

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 00:27

Gender: Male

Observation Period Began: 00:56

Subject's Last Name: HERNANDEZ

Date/Time of Test: 12/22/2007 01:38

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample #1	0.167	01:34
Breath Volume	1.582	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample #2	0.168	01:37
Breath Volume	1.632	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: SALLS, DANIEL, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: SALLS, DANIEL

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 01:14

Gender: Male

Observation Period Began: 01:46

Subject's Last Name: GASCA

Date/Time of Test: 12/22/2007 02:14

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample #1	0.201	02:09
Breath Volume	2.824	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample #2	0.196	02:12
Breath Volume	2.410	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: FLOWERS, KRISTY, G

Breath Test Operator Agency: Collier County SO

Arresting Officer: PEDRAZA, JOHN

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 20:53

Gender: Male

Observation Period Began: 21:15

Subject's Last Name: SANCHEZ

Date/Time of Test: 12/23/2007 21:44

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.078	21:38
Air Blank	0.000	21:38
Subject Sample #1	0.252	21:39
Breath Volume	1.648	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample #2	0.239	21:42
Breath Volume	2.218	21:42
Air Blank	0.000	21:43
Control Test	0.078	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001651

Instrument Registered To: COLLIER COUNTY SO

Software Version: 8100.27

Breath Test Operator: SALLS, DANIEL, S

Breath Test Operator Agency: Collier County SO

Arresting Officer: SALLS, DANIEL

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 12/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 02:39

Gender: Male

Observation Period Began: 03:02

Subject's Last Name: HEREDIA-DIAZ

Date/Time of Test: 12/29/2007 03:37

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Subject Sample #1	0.196	03:29	Volume Not Met
Breath Volume	0.101	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample #2	0.208	03:35	Volume Not Met
Breath Volume	0.054	03:35	
Air Blank	0.000	03:36	
Control Test	0.076	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 01/01/2008 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:29

Gender: Male

Observation Period Began: 02:12

Subject's Last Name: TYREE

Date/Time of Test: 12/06/2007 02:41

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:39	
Control Test	0.078	02:39	
Air Blank	0.000	02:40	
Subject Sample #1		02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Control Test	0.078	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:49

Gender: Male

Observation Period Began: 23:34

Subject's Last Name: MCLEAN

Date/Time of Test: 12/13/2007 00:02

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.077	23:56
Air Blank	0.000	23:56
Subject Sample #1	0.245	23:57
Breath Volume	1.207	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample #2	0.237	00:00
Breath Volume	2.066	00:00
Air Blank	0.000	00:01
Control Test	0.077	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: CARTER, MATTHEW

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:34

Gender: Female

Observation Period Began: 00:47

Subject's Last Name: COOKLOCK

Date/Time of Test: 12/13/2007 01:19

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.077	01:08	
Air Blank	0.000	01:08	
Subject Sample #1	0.151	01:11	Volume Not Met
Breath Volume	0.675	01:11	
Air Blank	0.000	01:12	
Air Blank	0.000	01:14	
Subject Sample #2	0.157	01:17	Volume Not Met
Breath Volume	0.218	01:17	
Air Blank	0.000	01:18	
Control Test	0.075	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: CARTER, MATTHEW

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:34

Gender: Female

Observation Period Began: 00:47

Subject's Last Name: COOKLOCK

Date/Time of Test: 12/13/2007 01:34

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.076	01:21	
Air Blank	0.000	01:21	
Subject Sample #1	0.168	01:24	Volume Not Met
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample #2	0.167	01:28	
Breath Volume	1.535	01:28	
Air Blank	0.000	01:29	
Air Blank	0.000	01:31	
Subject Sample #3	0.156	01:32	
Breath Volume	1.671	01:32	
Air Blank	0.000	01:33	
Control Test	0.075	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:16

Gender: Male

Observation Period Began: 01:53

Subject's Last Name: BENNETT

Date/Time of Test: 12/13/2007 02:40

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.077	02:33	
Air Blank	0.000	02:33	
Subject Sample #1	0.146	02:35	
Breath Volume	1.183	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:38	
Subject Sample #2	0.136	02:38	
Breath Volume	1.898	02:38	
Air Blank	0.000	02:39	
Control Test	0.073	02:39	Control Outside Tolerance
Air Blank	0.000	02:40	

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:16

Gender: Male

Observation Period Began: 01:53

Subject's Last Name: BENNETT

Date/Time of Test: 12/13/2007 02:47

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.077	02:43	
Air Blank	0.000	02:43	
Subject Sample #1		02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.077	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:16

Gender: Male

Observation Period Began: 01:53

Subject's Last Name: BENNETT

Date/Time of Test: 12/13/2007 02:56

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.077	02:50
Air Blank	0.000	02:51
Subject Sample #1	0.140	02:52
Breath Volume	1.820	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample #2	0.139	02:55
Breath Volume	1.828	02:55
Air Blank	0.000	02:56
Control Test	0.077	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: CARTER, MATTHEW

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 03:14

Gender: Female

Observation Period Began: 03:07

Subject's Last Name: JOHNSON

Date/Time of Test: 12/13/2007 03:33

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.077	03:28	
Air Blank	0.000	03:29	
Subject Sample #1		03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.077	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: CARTER, MATTHEW

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 03:14

Gender: Female

Observation Period Began: 03:07

Subject's Last Name: JOHNSON

Date/Time of Test: 12/13/2007 03:47

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.077	03:37	
Air Blank	0.000	03:37	
Subject Sample #1	0.159	03:41	Volume Not Met
Breath Volume	0.136	03:41	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample #2		03:45	Subject Test Refused
Breath Volume	0.050	03:45	
Air Blank	0.000	03:46	
Control Test	0.075	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: HEER, JOSEPH, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:55

Gender: Male

Observation Period Began: 00:55

Subject's Last Name: MCCLENDON

Date/Time of Test: 12/15/2007 01:26

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.076	01:16	
Air Blank	0.000	01:17	
Subject Sample #1		01:20	No Sample Provided
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample #2		01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Control Test	0.077	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 03:16

Gender: Female

Observation Period Began: 02:48

Subject's Last Name: FRAZIER

Date/Time of Test: 12/15/2007 03:33

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.078	03:31	
Air Blank	0.000	03:31	
Subject Sample #1		03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.077	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:16

Gender: Male

Observation Period Began: 22:55

Subject's Last Name: TYRE

Date/Time of Test: 12/19/2007 23:27

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:21
Air Blank	0.000	23:22
Subject Sample #1	0.091	23:22
Breath Volume	2.808	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample #2	0.091	23:25
Breath Volume	2.609	23:25
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001653

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: CARTER, MATTHEW

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 11/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:37

Gender: Male

Observation Period Began: 23:40

Subject's Last Name: BAUER

Date/Time of Test: 12/20/2007 00:13

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.078	00:11	
Air Blank	0.000	00:11	
Subject Sample #1		00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:12	
Control Test	0.078	00:12	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 12/20/2007 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001654

Instrument Registered To: MIAMI P.D.

Software Version: 8100.27

Breath Test Operator: BELTRAN, PEDRO,

Breath Test Operator Agency: Miami PD

Arresting Officer: KUNCAS, ANDREW

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 11/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 03:10

Gender: Male

Observation Period Began: 03:35

Subject's Last Name: CALDERON

Date/Time of Test: 12/01/2007 04:01

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.079	03:58	
Air Blank	0.000	03:59	
Subject Sample #1		04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 12/11/2007 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001654

Instrument Registered To: MIAMI P.D.

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M,

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 23:50

Gender: Female

Observation Period Began: 23:50

Subject's Last Name: BETANCOURT

Date/Time of Test: 12/21/2007 00:19

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.079	00:11
Air Blank	0.000	00:11
Subject Sample #1	0.129	00:14
Breath Volume	2.160	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample #2	0.145	00:17
Breath Volume	1.589	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 01/22/2008 18:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001654

Instrument Registered To: MIAMI P.D.

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M,

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 20:30

Gender: Male

Observation Period Began: 20:30

Subject's Last Name: PONCE

Date/Time of Test: 12/21/2007 20:59

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.079	20:53
Air Blank	0.000	20:54
Subject Sample #1	0.187	20:55
Breath Volume	2.652	20:55
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample #2	0.177	20:58
Breath Volume	1.984	20:58
Air Blank	0.000	20:58
Control Test	0.079	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 01/22/2008 18:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001654

Instrument Registered To: MIAMI P.D.

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M,

Breath Test Operator Agency: South Miami PD

Arresting Officer: DEJESUS, R

Arresting Officer Agency: North Miami PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 22:40

Gender: Male

Observation Period Began: 22:41

Subject's Last Name: SUASTEGUI

Date/Time of Test: 12/21/2007 23:08

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:03
Subject Sample #1	0.148	23:03
Breath Volume	1.351	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample #2	0.147	23:06
Breath Volume	1.789	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 01/22/2008 18:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001654

Instrument Registered To: MIAMI P.D.

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M,

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 23:30

Gender: Male

Observation Period Began: 23:30

Subject's Last Name: WILLIS

Date/Time of Test: 12/21/2007 23:57

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Subject Sample #1	0.017	23:53
Breath Volume	1.839	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample #2	0.016	23:56
Breath Volume	2.656	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 01/22/2008 18:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001654

Instrument Registered To: MIAMI P.D.

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M,

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 12/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 00:01

Gender: Male

Observation Period Began: 00:01

Subject's Last Name: LANATTA

Date/Time of Test: 12/22/2007 00:27

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Subject Sample #1	0.000	00:23
Breath Volume	1.820	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample #2	0.000	00:25
Breath Volume	2.480	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 01/22/2008 18:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: JIMENEZ, JOSE, A

Breath Test Operator Agency: St. John's County SO

Arresting Officer: JIMENEZ, JOSE

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610201G

Dry Gas Expiry Date: 04/17/2008

Violation Code: DUI

Arrest Time: 11:19

Gender: Male

Observation Period Began: 00:14

Subject's Last Name: BUELL

Date/Time of Test: 12/01/2007 00:40

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.079	00:35
Air Blank	0.000	00:35
Subject Sample #1	0.151	00:36
Breath Volume	1.992	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample #2	0.145	00:39
Breath Volume	4.199	00:39
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610201G

Dry Gas Expiry Date: 04/17/2008

Violation Code: DUI

Arrest Time: 01:48

Gender: Female

Observation Period Began: 02:10

Subject's Last Name: BROWN

Date/Time of Test: 12/01/2007 02:46

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.079	02:33	
Air Blank	0.000	02:33	
Subject Sample #1	0.140	02:36	Volume Not Met
Breath Volume	0.425	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample #2	0.166	02:41	
Breath Volume	1.238	02:41	
Air Blank	0.000	02:41	
Air Blank	0.000	02:43	
Subject Sample #3	0.171	02:44	
Breath Volume	1.425	02:44	
Air Blank	0.000	02:45	
Control Test	0.077	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: ROSADO, ANGELO, D

Breath Test Operator Agency: St. John's County SO

Arresting Officer: GONZALEZ, MANNY

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610201G

Dry Gas Expiry Date: 04/17/2008

Violation Code: DUI

Arrest Time: 08:17

Gender: Female

Observation Period Began: 09:30

Subject's Last Name: ROGERS

Date/Time of Test: 12/03/2007 09:57

Diagnostics Check	OK	09:51
Air Blank	0.000	09:51
Control Test	0.080	09:52
Air Blank	0.000	09:52
Subject Sample #1	0.000	09:52
Breath Volume	1.460	09:52
Air Blank	0.000	09:53
Air Blank	0.000	09:55
Subject Sample #2	0.000	09:55
Breath Volume	1.929	09:55
Air Blank	0.000	09:56
Control Test	0.080	09:56
Air Blank	0.000	09:57
Diagnostics Check	OK	09:57

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: RIOS, RICARDO, M

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BELLAMY, MATTHEW

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610201G

Dry Gas Expiry Date: 04/17/2008

Violation Code: DUI

Arrest Time: 01:04

Gender: Female

Observation Period Began: 01:50

Subject's Last Name: KIRBY

Date/Time of Test: 12/04/2007 02:17

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample #1	0.181	02:12
Breath Volume	1.269	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample #2	0.170	02:16
Breath Volume	2.023	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: SCHNEIDER, LOREN, M

Breath Test Operator Agency: St. John's County SO

Arresting Officer: EDENFIELD, EARL

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610201G

Dry Gas Expiry Date: 04/17/2008

Violation Code: DUI

Arrest Time: 01:33

Gender: Female

Observation Period Began: 02:27

Subject's Last Name: KONEN

Date/Time of Test: 12/05/2007 02:56

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:48
Subject Sample #1	0.280	02:49
Breath Volume	1.371	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample #2	0.269	02:55
Breath Volume	1.468	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610201G

Dry Gas Expiry Date: 04/17/2008

Violation Code: DUI

Arrest Time: 21:27

Gender: Male

Observation Period Began: 22:05

Subject's Last Name: UBERMUTH

Date/Time of Test: 12/06/2007 22:34

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Subject Sample #1	0.078	22:30
Breath Volume	1.832	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample #2	0.073	22:33
Breath Volume	3.355	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610201G

Dry Gas Expiry Date: 04/17/2008

Violation Code: DUI

Arrest Time: 00:22

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: HENSLEY

Date/Time of Test: 12/07/2007 01:30

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample #1	0.047	01:25
Breath Volume	3.417	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample #2	0.044	01:28
Breath Volume	2.796	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:30

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610201G

Dry Gas Expiry Date: 04/17/2008

Violation Code: DUI

Arrest Time: 03:20

Gender: Male

Observation Period Began: 03:40

Subject's Last Name: COLON

Date/Time of Test: 12/07/2007 04:19

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample #1	0.171	04:15
Breath Volume	2.468	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.172	04:18
Breath Volume	2.523	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610201G

Dry Gas Expiry Date: 04/17/2008

Violation Code: DUI

Arrest Time: 21:27

Gender: Male

Observation Period Began: 22:00

Subject's Last Name: LLORCA

Date/Time of Test: 12/10/2007 22:31

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.081	22:25
Air Blank	0.000	22:26
Subject Sample #1	0.156	22:26
Breath Volume	3.964	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample #2	0.149	22:29
Breath Volume	1.367	22:29
Air Blank	0.000	22:30
Control Test	0.080	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 03:15

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: CRAIG

Date/Time of Test: 12/15/2007 04:54

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.077	04:48
Air Blank	0.000	04:48
Subject Sample #1	0.135	04:49
Breath Volume	5.183	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample #2	0.132	04:52
Breath Volume	2.425	04:52
Air Blank	0.000	04:53
Control Test	0.076	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:54

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 21:04

Gender: Male

Observation Period Began: 21:20

Subject's Last Name: HUMMELL

Date/Time of Test: 12/16/2007 21:50

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.078	21:44
Air Blank	0.000	21:45
Subject Sample #1	0.127	21:46
Breath Volume	3.343	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample #2	0.136	21:49
Breath Volume	3.863	21:49
Air Blank	0.000	21:49
Control Test	0.077	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 01:55

Gender: Male

Observation Period Began: 02:10

Subject's Last Name: YOUNG

Date/Time of Test: 12/17/2007 02:47

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Subject Sample #1	0.239	02:42
Breath Volume	1.824	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample #2	0.237	02:45
Breath Volume	2.386	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 01:46

Gender: Female

Observation Period Began: 02:20

Subject's Last Name: FANKHAUSER

Date/Time of Test: 12/20/2007 02:50

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Subject Sample #1	0.105	02:46
Breath Volume	1.480	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample #2	0.120	02:49
Breath Volume	2.339	02:49
Air Blank	0.000	02:49
Control Test	0.078	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: RIOS, RICARDO, M

Breath Test Operator Agency: St. John's County SO

Arresting Officer: HAMMONDS, FRANKIE

Arresting Officer Agency: St. Augustine Beach

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 00:47

Gender: Male

Observation Period Began: 01:30

Subject's Last Name: BRUNER

Date/Time of Test: 12/23/2007 01:57

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.078	01:51
Air Blank	0.000	01:51
Subject Sample #1	0.204	01:52
Breath Volume	2.019	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample #2	0.219	01:55
Breath Volume	1.617	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:57

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: SEAY, BRYAN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 15:54

Gender: Female

Observation Period Began: 16:45

Subject's Last Name: HOLEVA

Date/Time of Test: 12/25/2007 17:14

Diagnostics Check	OK	17:07
Air Blank	0.000	17:07
Control Test	0.078	17:08
Air Blank	0.000	17:08
Subject Sample #1	0.211	17:09
Breath Volume	1.523	17:09
Air Blank	0.000	17:10
Air Blank	0.000	17:12
Subject Sample #2	0.216	17:13
Breath Volume	1.832	17:13
Air Blank	0.000	17:13
Control Test	0.076	17:14
Air Blank	0.000	17:14
Diagnostics Check	OK	17:14

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 00:05

Gender: Male

Observation Period Began: 00:30

Subject's Last Name: COLLEY

Date/Time of Test: 12/26/2007 00:59

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Subject Sample #1	0.178	00:54
Breath Volume	2.398	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample #2	0.166	00:58
Breath Volume	3.484	00:58
Air Blank	0.000	00:58
Control Test	0.077	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 12/27/2007 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 23:28

Gender: Male

Observation Period Began: 23:55

Subject's Last Name: HOGAN

Date/Time of Test: 12/29/2007 00:25

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:20
Subject Sample #1	0.218	00:21
Breath Volume	3.398	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample #2	0.220	00:24
Breath Volume	3.101	00:24
Air Blank	0.000	00:24
Control Test	0.078	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 01/31/2008 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: RIOS, RICARDO, M

Breath Test Operator Agency: St. John's County SO

Arresting Officer: MATUSE, TONY

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 00:36

Gender: Male

Observation Period Began: 01:45

Subject's Last Name: ANDERSON

Date/Time of Test: 12/29/2007 02:11

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:06
Air Blank	0.000	02:06
Subject Sample #1	0.138	02:06
Breath Volume	2.238	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #2	0.139	02:09
Breath Volume	1.367	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 01/31/2008 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 02:01

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: ASBURY

Date/Time of Test: 12/29/2007 03:18

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.077	03:11
Air Blank	0.000	03:11
Subject Sample #1	0.091	03:13
Breath Volume	1.320	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample #2	0.085	03:17
Breath Volume	2.128	03:17
Air Blank	0.000	03:17
Control Test	0.077	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 01/31/2008 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: PORTEUS, ROBERT

Arresting Officer Agency: St. Augustine PD

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 02:06

Gender: Male

Observation Period Began: 03:40

Subject's Last Name: WILSON

Date/Time of Test: 12/29/2007 04:09

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.077	04:04
Air Blank	0.000	04:04
Subject Sample #1	0.045	04:05
Breath Volume	1.492	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample #2	0.047	04:08
Breath Volume	1.789	04:08
Air Blank	0.000	04:09
Control Test	0.076	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 01/31/2008 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 23:47

Gender: Male

Observation Period Began: 00:50

Subject's Last Name: TARLETON

Date/Time of Test: 12/30/2007 01:18

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:12
Subject Sample #1	0.101	01:13
Breath Volume	4.265	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample #2	0.102	01:16
Breath Volume	4.414	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 01/31/2008 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: BECKTON, QUENTIN, Q

Breath Test Operator Agency: St. John's County SO

Arresting Officer: BECKTON, QUENTIN

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 22:30

Gender: Male

Observation Period Began: 23:05

Subject's Last Name: SUMNER

Date/Time of Test: 12/31/2007 23:32

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:27
Subject Sample #1	0.222	23:28
Breath Volume	3.285	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample #2	0.218	23:31
Breath Volume	2.781	23:31
Air Blank	0.000	23:31
Control Test	0.078	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 01/31/2008 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001655

Instrument Registered To: ST JOHN'S COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCGINNIS, J.D.,

Breath Test Operator Agency: St. John's County SO

Arresting Officer: MCGINNIS, J.D.

Arresting Officer Agency: St. John's County SO

Last Agency Inspection: 12/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 719102I

Dry Gas Expiry Date: 07/15/2009

Violation Code: DUI

Arrest Time: 23:00

Gender: Female

Observation Period Began: 23:25

Subject's Last Name: CATES

Date/Time of Test: 12/31/2007 23:53

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.078	23:46
Air Blank	0.000	23:46
Subject Sample #1	0.207	23:49
Breath Volume	1.324	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample #2	0.209	23:52
Breath Volume	1.894	23:52
Air Blank	0.000	23:52
Control Test	0.076	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 01/31/2008 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001656

Instrument Registered To: NORTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: JEANITE, DUHAMEL,

Breath Test Operator Agency: North Miami PD

Arresting Officer: JEANITE, DUHAMEL

Arresting Officer Agency: North Miami PD

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 71490211

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 06:40

Gender: Male

Observation Period Began: 06:50

Subject's Last Name: CHAR

Date/Time of Test: 12/08/2007 07:13

Diagnostics Check	OK	07:13	
Air Blank	0.000	07:13	
		07:13	RFI Detect
Air Blank		07:13	RFI Detect

Data Download Date/Time: 12/24/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001656

Instrument Registered To: NORTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: JEANITE, DUHAMEL,

Breath Test Operator Agency: North Miami PD

Arresting Officer: JEANITE, DUHAMEL

Arresting Officer Agency: North Miami PD

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 71490211

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 06:40

Gender: Male

Observation Period Began: 06:50

Subject's Last Name: CHAR

Date/Time of Test: 12/08/2007 07:25

Diagnostics Check	OK	07:16
Air Blank	0.000	07:17
Control Test	0.078	07:17
Air Blank	0.000	07:18
Subject Sample #1	0.097	07:20
Breath Volume	2.003	07:20
Air Blank	0.000	07:21
Air Blank	0.000	07:22
Subject Sample #2	0.088	07:23
Breath Volume	1.738	07:23
Air Blank	0.000	07:24
Control Test	0.077	07:24
Air Blank	0.000	07:25
Diagnostics Check	OK	07:25

Data Download Date/Time: 12/24/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001656

Instrument Registered To: NORTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: ZUNIGA, ABRAHAM, R

Breath Test Operator Agency: North Miami PD

Arresting Officer: ZUNIGA, ABRAHAM

Arresting Officer Agency: North Miami PD

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 71490211

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 01:10

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: MOMPREMIER

Date/Time of Test: 12/09/2007 02:10

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:04
Air Blank	0.000	02:04
Subject Sample #1	0.107	02:05
Breath Volume	1.617	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample #2	0.102	02:09
Breath Volume	1.390	02:09
Air Blank	0.000	02:09
Control Test	0.077	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 12/24/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001656

Instrument Registered To: NORTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: MALDONADO, GALO, M

Breath Test Operator Agency: North Miami PD

Arresting Officer: REYES, E

Arresting Officer Agency: North Miami PD

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 71490211

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 20:15

Gender: Male

Observation Period Began: 20:15

Subject's Last Name: MALDONADO

Date/Time of Test: 12/11/2007 20:54

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.078	20:47
Air Blank	0.000	20:48
Subject Sample #1	0.000	20:48
Breath Volume	1.562	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample #2	0.000	20:52
Breath Volume	1.906	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 12/24/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001656

Instrument Registered To: NORTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: JEANITE, DUHAMEL,

Breath Test Operator Agency: North Miami PD

Arresting Officer: JEANITE, DUHAMEL

Arresting Officer Agency: North Miami PD

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 71490211

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:20

Gender: Male

Observation Period Began: 03:20

Subject's Last Name: CUTINO

Date/Time of Test: 12/17/2007 04:00

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.077	03:52	
Air Blank	0.000	03:52	
Subject Sample #1	0.204	03:53	
Breath Volume	1.406	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample #2	0.174	03:57	
Breath Volume	2.136	03:57	
Air Blank	0.000	03:57	
Air Blank		03:59	Improper Sample
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 12/24/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001656

Instrument Registered To: NORTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: JEANITE, DUHAMEL,

Breath Test Operator Agency: North Miami PD

Arresting Officer: JEANITE, DUHAMEL

Arresting Officer Agency: North Miami PD

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 71490211

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:20

Gender: Male

Observation Period Began: 03:20

Subject's Last Name: CUTINO

Date/Time of Test: 12/17/2007 04:09

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.077	04:04
Air Blank	0.000	04:04
Subject Sample #1	0.190	04:05
Breath Volume	2.054	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.197	04:08
Breath Volume	2.292	04:08
Air Blank	0.000	04:08
Control Test	0.077	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 12/24/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001656

Instrument Registered To: NORTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: REYES, ERNESTO,

Breath Test Operator Agency: North Miami PD

Arresting Officer: REYES, ERNESTO

Arresting Officer Agency: North Miami PD

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 71490211

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 19:15

Gender: Male

Observation Period Began: 19:00

Subject's Last Name: BLAIR

Date/Time of Test: 12/19/2007 20:07

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:03	
Control Test	0.077	20:03	
Air Blank	0.000	20:04	
Subject Sample #1		20:05	Subject Test Refused
Breath Volume	0.000	20:05	
Air Blank	0.000	20:06	
Control Test	0.078	20:06	
Air Blank	0.000	20:06	
Diagnostics Check	OK	20:07	

Data Download Date/Time: 12/24/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001656

Instrument Registered To: NORTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: PITTMAN, RODERICK, S

Breath Test Operator Agency: North Miami PD

Arresting Officer: HOLLANT, EMILE

Arresting Officer Agency: North Miami PD

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 71490211

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 05:58

Gender: Male

Observation Period Began: 06:47

Subject's Last Name: JOHNSON

Date/Time of Test: 12/21/2007 07:38

Diagnostics Check	OK	07:32
Air Blank	0.000	07:32
Control Test	0.078	07:32
Air Blank	0.000	07:33
Subject Sample #1	0.189	07:33
Breath Volume	1.351	07:33
Air Blank	0.000	07:34
Air Blank	0.000	07:36
Subject Sample #2	0.196	07:36
Breath Volume	3.152	07:36
Air Blank	0.000	07:37
Control Test	0.076	07:37
Air Blank	0.000	07:38
Diagnostics Check	OK	07:38

Data Download Date/Time: 12/24/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001656

Instrument Registered To: NORTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: BARNES, KENNETH, R

Breath Test Operator Agency: OTHER Local

Arresting Officer: BARNES, KENNETH

Arresting Officer Agency: OTHER Local

Last Agency Inspection: 11/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 71490211

Dry Gas Expiry Date: 05/31/2009

Violation Code: DUI

Arrest Time: 03:30

Gender: Female

Observation Period Began: 05:15

Subject's Last Name: VIVOLO

Date/Time of Test: 12/22/2007 06:36

Diagnostics Check	OK	06:30
Air Blank	0.000	06:31
Control Test	0.078	06:31
Air Blank	0.000	06:31
Subject Sample #1	0.103	06:32
Breath Volume	2.628	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:34
Subject Sample #2	0.100	06:35
Breath Volume	2.375	06:35
Air Blank	0.000	06:36
Control Test	0.077	06:36
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time: 12/24/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001720

Instrument Registered To: ALACHUA PD

Software Version: 8100.27

Breath Test Operator: PARKS, DONALD, M

Breath Test Operator Agency: Alachua PD

Arresting Officer: NAHRWOLD, CHARLIE

Arresting Officer Agency: Alachua PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632002I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 23:56

Gender: Female

Observation Period Began: 00:05

Subject's Last Name: KELLY

Date/Time of Test: 12/02/2007 00:39

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.079	00:29	
Air Blank	0.000	00:29	
Subject Sample #1		00:32	No Sample Provided
Breath Volume	0.000	00:32	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample #2		00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Control Test	0.080	00:39	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 12/28/2007 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001720

Instrument Registered To: ALACHUA PD

Software Version: 8100.27

Breath Test Operator: PARKS, DONALD, M

Breath Test Operator Agency: Alachua PD

Arresting Officer: PARKS, DONALD

Arresting Officer Agency: Alachua PD

Last Agency Inspection: 11/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632002I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:43

Gender: Female

Observation Period Began: 02:00

Subject's Last Name: ANERINE

Date/Time of Test: 12/15/2007 02:31

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.079	02:21	
Air Blank	0.000	02:21	
Subject Sample #1		02:24	No Sample Provided
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample #2		02:30	No Sample Provided
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 12/28/2007 20:45