

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 01/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 23:48:52	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C	<b>0.08 Gas Lot #</b> 5202011
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007	<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 2	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

Time-Date changed. 20: Control Outside Tolerance.

**Data Download Date/Time**      01/01/2007      23:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 02/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 00:08:17	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C	<b>0.08 Gas Lot #</b> 5202011
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007	<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      02/03/2007      00:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 03/05/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 23:40:27	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C	<b>0.08 Gas Lot #</b> 5202011
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007	<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time**      03/05/2007      23:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 04/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 23:22:26	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 5202011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 04/03/2007 23:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 05/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 00:51:13	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C	<b>0.08 Gas Lot #</b> 5202011
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007	<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time**      05/01/2007      00:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 06/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 23:44:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 5202011
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

00: RFI Detect.

**Data Download Date/Time**      06/01/2007      23:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 07/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 23:39:45	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 5202011
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

Time-Date changed. 05: RFI Detect.

**Data Download Date/Time**      07/02/2007      23:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 08/07/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 23:16:49	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 7171011
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time**      08/07/2007      23:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 09/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 03:02:53	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 7171011
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time**      09/01/2007      03:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 09/27/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SKIPPER,ROGER,G	<b>Time</b> 10:22:15	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 719402I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 07/15/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b> 0.0007	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4 (Retry)	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.077
		Alcohol Free Test #5 (Retry)	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.077
		Alcohol Free Test #6 (Retry)	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.077
		Alcohol Free Test #7 (Retry)	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.077
		Alcohol Free Test #8 (Retry)	0.000
		0.05 g/210L Test #8	0.051

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

0.08 g/210L Test #8	0.081
0.20 g/210L Test #8	0.200
0.08 g/210L Dry Gas Std Test #8	0.077
Alcohol Free Test #9 (Retry)	0.000
0.05 g/210L Test #9	0.051
0.08 g/210L Test #9	0.081
0.20 g/210L Test #9	0.200
0.08 g/210L Dry Gas Std Test #9	0.076
Alcohol Free Test #10 (Retry)	0.000
0.05 g/210L Test #10	0.052
0.08 g/210L Test #10	0.082
0.20 g/210L Test #10	0.200
0.08 g/210L Dry Gas Std Test #10	0.077
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

#### Remarks:

00: Ambient Fail SAMPLE CHAMBER DID NOT PURGE M/A TEST. . 0.00 ALCOHOL FREE SIM TESTED REPEATED.

Data Download Date/Time 09/27/2007 10:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 10/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 23:38:36	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 717011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

Int Det:LOOSE HOSE CONNECTION.

**Data Download Date/Time** 10/01/2007 23:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 11/07/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 00:30:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 7171011
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

TIME AND DATE CHANGE

**Data Download Date/Time**      11/07/2007      00:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001441

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 12/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGHTOWER,AARON,R	<b>Time</b> 01:02:49	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 7171011
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

A F / M A : .

**Data Download Date/Time**      12/01/2007      01:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 01/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 11:27:53	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 5202011
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 04/21/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time**      01/31/2007      11:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 02/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 12:43:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 5202011
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 04/21/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time**      02/02/2007      12:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 03/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 12:42:13	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 04/21/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time**      03/26/2007      12:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 04/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 14:35:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time**      04/11/2007      14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 05/16/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 13:13:42	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time**      05/16/2007      13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 06/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 11:28:37	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time**      06/19/2007      12:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 07/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 11:55:27	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.082

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time**      07/18/2007      12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 08/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VELAZQUEZ,KIMBERLY,	<b>Time</b> 11:27:48	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      08/22/2007      11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 09/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 13:39:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      09/11/2007      13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 10/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VELAZQUEZ,KIMBERLY,	<b>Time</b> 10:23:36	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

PSA ESPINOZA PRESENT FOR AI

**Data Download Date/Time**      10/10/2007      17:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 10/10/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 17:20:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0007	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.083
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.190
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.083
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.083
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.192
		0.08 g/210L Dry Gas Std Test #4	0.082
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.192
		0.08 g/210L Dry Gas Std Test #5	0.083
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.192
		0.08 g/210L Dry Gas Std Test #6	0.082
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.193
		0.08 g/210L Dry Gas Std Test #7	0.082
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.193
0.08 g/210L Dry Gas Std Test #8	0.082
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.049
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.194
0.08 g/210L Dry Gas Std Test #9	0.082
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.049
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.194
0.08 g/210L Dry Gas Std Test #10	0.082
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0014
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

OFFICER K. VELAZQUEZ & PSA ESPINOZA WERE PRESENT FOR INSPEC

**Data Download Date/Time**      10/10/2007      17:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 11/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VELAZQUEZ,KIMBERLY,	<b>Time</b> 13:04:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

00: Ambient Fail/CLEARED AREA/PSA ESPINOZA PRESENT

**Data Download Date/Time**      11/13/2007      12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001442

<b>Agency</b> BISCAYNE NAT. PARK	<b>Date</b> 12/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VELAZQUEZ,KIMBERLY,	<b>Time</b> 12:02:29	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

PSA ESPINOZA PRESENT FOR AI

**Data Download Date/Time**      12/11/2007      12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 01/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MEADOR,RICKY,A	<b>Time</b> 12:40:40	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 5202011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 01/28/2007 12:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 02/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MEADOR,RICKY,A	<b>Time</b> 01:23:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 5202011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/26/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.082

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 02/20/2007 01:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 03/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MEADOR,RICKY,A	<b>Time</b> 09:19:41	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 03/25/2007 09:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 04/19/2007	<b>Type</b> Department Inspection
<b>Inspector</b> MALHIOT, MATTHEW, E	<b>Time</b> 10:21:56	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0003	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.194
		0.08 g/210L Dry Gas Std Test #4	0.081
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.076
		0.20 g/210L Test #5	0.194
		0.08 g/210L Dry Gas Std Test #5	0.081
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.195
		0.08 g/210L Dry Gas Std Test #6	0.081
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.047
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.194
		0.08 g/210L Dry Gas Std Test #7	0.081
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.047

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

0.08 g/210L Test #8	0.077
0.20 g/210L Test #8	0.194
0.08 g/210L Dry Gas Std Test #8	0.080
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.046
0.08 g/210L Test #9	0.077
0.20 g/210L Test #9	0.195
0.08 g/210L Dry Gas Std Test #9	0.081
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.047
0.08 g/210L Test #10	0.077
0.20 g/210L Test #10	0.194
0.08 g/210L Dry Gas Std Test #10	0.081
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 04/22/2007 00:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 04/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MEADOR,RICKY,A	<b>Time</b> 00:28:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time**    04/22/2007    00:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 05/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MEADOR,RICKY,A	<b>Time</b> 05:17:46	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 05/20/2007 05:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 06/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MEADOR,RICKY,A	<b>Time</b> 23:13:54	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time**    06/15/2007    23:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 07/05/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MEADOR,RICKY,A	<b>Time</b> 21:12:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082

**Remarks:**

A F / M A : .

**Data Download Date/Time**      07/05/2007      21:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 08/06/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MEADOR,RICKY,A	<b>Time</b> 22:41:48	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.082

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time**      08/06/2007      22:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 09/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MEADOR,RICKY,A	<b>Time</b> 13:22:31	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7107011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 04/18/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 09/26/2007 07:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 10/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MEADOR,RICKY,A	<b>Time</b> 21:56:18	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7107011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 04/18/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 10/15/2007 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 11/07/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARTER,CARL,L	<b>Time</b> 12:00:35	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7107011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 04/18/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 11/07/2007 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 12/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARTER,CARL,L	<b>Time</b> 13:40:09	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7107011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 04/18/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 12/10/2007 13:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 01/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 11:02:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5201011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/22/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.082

**Remarks:**

**Data Download Date/Time** 01/27/2007 11:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 02/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 12:10:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5201011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/22/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 02/25/2007 13:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 03/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 10:03:49	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5201011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/22/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

Time-Date changed. Int Det: .

**Data Download Date/Time**      03/25/2007      11:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 04/21/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 10:09:32	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5201011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/22/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time**      04/21/2007      11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 05/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 13:18:45	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5201011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/22/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 05/23/2007 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 06/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 13:16:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7053011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 02/27/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

20: Control Outside Tolerance. RECHECKED CONNECTIONS

**Data Download Date/Time**      06/30/2007      14:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 07/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 14:55:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7053011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 02/27/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

20: Control Outside Tolerance.-RECHECKED CONNECTIONS/SEALS

**Data Download Date/Time**      07/20/2007      15:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 08/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 14:17:02	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Int Det:TIGHTENED SIM JAR AFTER 1ST ATTEMPT (LOOSE), . Non-compliance:INT FAILURE\_POSSIBLE POOR ACETONE SOLUTION.

**Data Download Date/Time** 09/03/2007 16:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 08/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 13:14:59	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7053011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 02/27/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.182
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

00: Ambient Fail-DID NOT DISCONNECT CONTAMINATED MOUTHPEICE. 20: Control Outside Tolerance:RECHECKED CONNECTIONS.

**Data Download Date/Time**      09/03/2007      16:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 09/06/2007	<b>Type</b> Department Inspection
<b>Inspector</b> GEDDINGS,MARGARET,M	<b>Time</b> 13:12:37	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 719402I
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 07/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.080
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.081
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.080
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.080
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.198
0.08 g/210L Dry Gas Std Test #8	0.080
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.048
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.198
0.08 g/210L Dry Gas Std Test #9	0.081
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.048
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.198
0.08 g/210L Dry Gas Std Test #10	0.080
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time      10/20/2007      11:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 10/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 10:50:03	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7053011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 02/27/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.187
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.197
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.195
		0.20 g/210L Test #2 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

08: Control Outside Tolerance: RE-SEATED SIM ADAPTORS. 20: Control Outside Tolerance: RECHECKED SIM FOR AIR LEAK.

**Data Download Date/Time** 10/20/2007 11:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 11/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 10:29:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 7053011
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 02/27/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1 (Retry)	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3 (Retry)	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

05: Ambient Fail. OPEN ARS NEAR BREATH TUBE

**Data Download Date/Time**      11/18/2007      11:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001469

<b>Agency</b> HURLBURT FIELD PD	<b>Date</b> 12/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPE,EDWARD,J	<b>Time</b> 08:52:01	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200708C	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 7053011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 02/27/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 12/15/2007 09:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK A.F.B.	<b>Date</b> 01/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GARRITY,THOMAS,A	<b>Time</b> 11:34:41	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Non-compliance: .

**Data Download Date/Time** 02/02/2007 10:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK A.F.B.	<b>Date</b> 01/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GARRITY,THOMAS,A	<b>Time</b> 12:13:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200504C	<b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 518802I		
<b>Expiration</b> 04/21/2007	<b>Expiration</b> 04/21/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 07/12/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.047
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test #2 (Retry)	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Alcohol Free Test #3	0.000
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

00: Control Outside Tolerance.

**Data Download Date/Time**      02/02/2007      10:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK A.F.B.	<b>Date</b> 02/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GARRITY,THOMAS,A	<b>Time</b> 10:54:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200504C	<b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 518802I		
<b>Expiration</b> 04/21/2007	<b>Expiration</b> 04/21/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 07/12/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      02/02/2007      10:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK A.F.B.	<b>Date</b> 03/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GARRITY,THOMAS,A	<b>Time</b> 12:08:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200504C	<b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 518802I		
<b>Expiration</b> 04/21/2007	<b>Expiration</b> 04/21/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 07/12/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.075

**Remarks:**

**Data Download Date/Time**      03/30/2007      12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK A.F.B.	<b>Date</b> 04/17/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GARRITY,THOMAS,A	<b>Time</b> 09:19:08	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200504C	<b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 518802I		
<b>Expiration</b> 04/21/2007	<b>Expiration</b> 04/21/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 07/12/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.075

**Remarks:**

**Data Download Date/Time**      04/17/2007      09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK A.F.B.	<b>Date</b> 06/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGLEY,CHARLES,J	<b>Time</b> 02:50:03	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 518802I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 07/12/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.074
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.074

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time**      06/29/2007      03:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

**Agency** PATRICK A.F.B. **Date** 06/29/2007 **Type** Agency Inspection  
**Inspector** LANGLEY,CHARLES,J **Time** 03:11:19 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Number of Simulators** 0 **Average Standard Deviation** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

**Data Download Date/Time** 06/29/2007 03:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

**Agency** PATRICK A.F.B. **Date** 06/29/2007 **Type** Agency Inspection  
**Inspector** LANGLEY,CHARLES,J **Time** 03:13:37 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Number of Simulators** 0 **Average Standard Deviation** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

**Data Download Date/Time** 06/29/2007 03:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK A.F.B.	<b>Date</b> 06/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGLEY,CHARLES,J	<b>Time</b> 03:16:04	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 0	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

**Data Download Date/Time**      06/29/2007      03:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK A.F.B.	<b>Date</b> 06/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGLEY,CHARLES,J	<b>Time</b> 03:31:05	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 0	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

**Data Download Date/Time** 07/30/2007 17:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK A.F.B.	<b>Date</b> 07/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WAGNER,ROBERT,F	<b>Time</b> 18:45:55	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 518802I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 07/12/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.022
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.073
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.073
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.026
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.047
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	No	0.20 g/210L Test #2	0.199
		0.08 g/210L Dry Gas Std Test #2	0.073
		0.08 g/210L Dry Gas Std Test #2 (Retry)	0.073
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.029
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.073
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.072

**Remarks:**

Time-Date changed. 05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. GAS CYLINDER EXPIRED.

**Data Download Date/Time** 09/03/2007 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK AFB SFS	<b>Date</b> 10/02/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SKIPPER,ROGER,G	<b>Time</b> 13:57:34	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7194021		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 07/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.082
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.202
		0.08 g/210L Dry Gas Std Test #5	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.080
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.080
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.050

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

0.08 g/210L Test #8	0.082
0.20 g/210L Test #8	0.202
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.082
0.20 g/210L Test #9	0.202
0.08 g/210L Dry Gas Std Test #9	0.080
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.050
0.08 g/210L Test #10	0.082
0.20 g/210L Test #10	0.203
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

#### Remarks:

INSPECTION AFTER REPAIR

Data Download Date/Time 10/02/2007 13:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK AFB SFS	<b>Date</b> 10/09/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WAGNER,ROBERT,F	<b>Time</b> 15:19:34	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 706102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 03/05/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.082
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.083
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.084
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.084
		0.05 g/210L Test #3 (Retry)	0.052
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.205
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 10/09/2007 16:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK AFB SFS	<b>Date</b> 11/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WAGNER,ROBERT,F	<b>Time</b> 13:01:06	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 723402I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 08/23/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.206
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time**      11/26/2007      13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK AFB SFS	<b>Date</b> 12/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WAGNER,ROBERT,F	<b>Time</b> 08:04:00	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

SIMULATOR OFF/COLD. INT DET COLD SIM.

**Data Download Date/Time** 12/18/2007 09:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

**Agency** PATRICK AFB SFS **Date** 12/18/2007 **Type** Agency Inspection  
**Inspector** WAGNER,ROBERT,F **Time** 08:31:53 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Number of Simulators** 5 **Average Standard Deviation** **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

ACETONE DEPLETED. ACETONE DEPLETED.

**Data Download Date/Time** 12/18/2007 09:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001471

<b>Agency</b> PATRICK AFB SFS	<b>Date</b> 12/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WAGNER,ROBERT,F	<b>Time</b> 09:09:59	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 723402I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 08/23/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.082
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

Int Det: Purge Fail.

**Data Download Date/Time** 12/18/2007 09:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 01/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 16:33:52	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 01/26/2007 17:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 02/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 21:06:04	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      02/25/2007      22:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 03/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 19:13:20	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      03/28/2007      20:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 04/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 19:34:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 04/25/2007 20:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 05/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 18:02:16	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 05/27/2007 19:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 06/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 19:51:53	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 06/28/2007 20:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 07/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 19:17:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 07/28/2007 20:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 08/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 12:49:45	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 08/25/2007 13:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 09/10/2007	<b>Type</b> Department Inspection
<b>Inspector</b> GEDDINGS,MARGARET,M	<b>Time</b> 13:37:58	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612D <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 719402I
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 07/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0003	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.080
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.080
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.080
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
		Alcohol Free Test #7	0.000

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.078
0.20 g/210L Test #7	0.196
0.08 g/210L Dry Gas Std Test #7	0.080
0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047
0.08 g/210L Test #8	0.078
0.20 g/210L Test #8	0.196
0.08 g/210L Dry Gas Std Test #8	0.080
0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.047
0.08 g/210L Test #9	0.078
0.20 g/210L Test #9	0.196
0.08 g/210L Dry Gas Std Test #9	0.080
0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.047
0.08 g/210L Test #10	0.078
0.20 g/210L Test #10	0.197
0.08 g/210L Dry Gas Std Test #10	0.080
0.08 g/210L Dry Gas Std Test #10 (Retry)	0.080
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0253
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

#### Remarks:

08: Control Outside Tolerance HOOKED UP HOSE.

Data Download Date/Time 09/12/2007 19:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 09/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 12:59:46	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time**      09/24/2007      19:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 10/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 12:56:34	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 10/31/2007 22:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 11/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 00:22:04	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 11/28/2007 01:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001646

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 12/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD,WILLIAM,E	<b>Time</b> 15:18:34	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 12/30/2007 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 01/24/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 15:36:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 01/24/2007 15:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 02/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 17:42:16	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 02/02/2007 17:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 03/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 17:00:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 03/02/2007 16:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 04/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 16:54:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 04/03/2007 17:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 05/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 10:36:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 05/03/2007 10:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 06/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 16:36:03	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 06/15/2007 16:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 07/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 15:58:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      07/13/2007      16:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 08/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 09:07:26	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 634801G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      08/30/2007      09:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 09/05/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VENTURI,GEORGE,L	<b>Time</b> 12:53:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6261011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.077
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.078
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.077
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.077
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.048

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

0.08 g/210L Test #8	0.078
0.20 g/210L Test #8	0.197
0.08 g/210L Dry Gas Std Test #8	0.078
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.049
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.078
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.049
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.197
0.08 g/210L Dry Gas Std Test #10	0.078
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 09/05/2007 12:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 09/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 17:27:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 09/12/2007 17:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 10/05/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 14:16:31	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7235011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 10/05/2007 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 11/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 16:11:24	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7235011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 11/20/2007 16:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001647

<b>Agency</b> CAPE CORAL PD	<b>Date</b> 12/04/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 17:38:54	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7235011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 12/05/2007 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 01/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VIENO,MICHAEL,J	<b>Time</b> 21:39:40	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 2	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 01/28/2007 23:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 02/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VIENO,MICHAEL,J	<b>Time</b> 00:22:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 2	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      02/28/2007      00:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 03/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VIENO,MICHAEL,J	<b>Time</b> 22:06:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 2	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

20: MOIST. IN HOSE BETWEEN SIM. INSTRUMENT.

**Data Download Date/Time**      03/27/2007      07:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 04/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VIENO,MICHAEL,J	<b>Time</b> 23:12:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 2	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 04/29/2007 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 05/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VIENO,MICHAEL,J	<b>Time</b> 09:42:22	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605C
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 2	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

RADIO INTERF..

**Data Download Date/Time** 05/29/2007 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 06/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GOLDBERG,THOMAS,R	<b>Time</b> 10:16:46	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.191
		0.20 g/210L Test #3 (Retry)	0.192
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

20: Control Outside Tolerance- MOISTUREIN TUBE

**Data Download Date/Time**      06/26/2007      10:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 07/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VIENO,MICHAEL,J	<b>Time</b> 10:23:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008

**Number of Simulators** 2      **Average Standard Deviation**      **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1 (Retry)	0.153
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.079
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.161
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3 (Retry)	0.156

**Remarks:**

20: RFI Detect/CELL PHONE, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time**      07/27/2007      14:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 07/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VIENO,MICHAEL,J	<b>Time</b> 14:22:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 2	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.034
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.203
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.05 g/210L Test #3 (Retry)	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time**      07/27/2007      14:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 08/12/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SUERETH,DONALD,P	<b>Time</b> 16:26:45	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7096011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b> 0.0005	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.078
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.077
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.077
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.077
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.051

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

0.08 g/210L Test #8	0.082
0.20 g/210L Test #8	0.198
0.08 g/210L Dry Gas Std Test #8	0.077
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.051
0.08 g/210L Test #9	0.082
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.077
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.051
0.08 g/210L Test #10	0.081
0.20 g/210L Test #10	0.198
0.08 g/210L Dry Gas Std Test #10	0.077
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

OFFICER VIENO PRESENT

**Data Download Date/Time**      08/12/2007      17:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 08/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VIENO,MICHAEL,J	<b>Time</b> 17:18:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7096011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      08/12/2007      17:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 09/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VIENO,MICHAEL,J	<b>Time</b> 06:37:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 09/21/2007 06:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 10/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VIENO,MICHAEL,J	<b>Time</b> 06:19:00	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 10/16/2007 07:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 12/18/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SUERETH,DONALD,P	<b>Time</b> 11:06:35	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7096011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0009	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.052
		0.08 g/210L Test #4	0.083
		0.20 g/210L Test #4	0.203
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.052
		0.08 g/210L Test #5	0.083
		0.20 g/210L Test #5	0.203
		0.08 g/210L Dry Gas Std Test #5	0.078
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.082
		0.20 g/210L Test #6	0.203
		0.08 g/210L Dry Gas Std Test #6	0.078
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.203
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.053

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

0.08 g/210L Test #8	0.082
0.20 g/210L Test #8	0.203
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.054
0.08 g/210L Test #9	0.082
0.20 g/210L Test #9	0.202
0.08 g/210L Dry Gas Std Test #9	0.078
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.053
0.08 g/210L Test #10	0.082
0.20 g/210L Test #10	0.202
0.08 g/210L Dry Gas Std Test #10	0.078
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0018
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time      12/18/2007      11:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 12/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SUERETH,DONALD,P	<b>Time</b> 11:08:45	<b>Software</b> 8100.27
<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>
<b>0.08 Gas Lot #</b>		
<b>Number of Simulators</b> 0	<b>Average Standard Deviation</b>	<b>Compliance</b> No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

DIAGNOSTIC INTENTIONALLY FAILED TO CONDUCT SUBJ. TEST.Non-compliance: .

**Data Download Date/Time** 12/18/2007 11:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001649

<b>Agency</b> KENNETH CITY PD	<b>Date</b> 12/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GOLDBERG,THOMAS,R	<b>Time</b> 06:25:09	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 12/19/2007 06:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 01/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 08:38:26	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

TESTED AT IJC MOVED FROM P TO B

**Data Download Date/Time** 01/01/2007 08:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 02/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 18:14:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

TESTED AT IJC AS B

**Data Download Date/Time** 02/01/2007 18:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 03/14/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 09:30:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

TESTED AT IJC B

**Data Download Date/Time** 03/14/2007 09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 04/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 16:55:06	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

TESTED AT IJC B

**Data Download Date/Time** 04/02/2007 17:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 05/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PEABODY,MICHAEL,F	<b>Time</b> 08:57:36	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

IJC OBSERVED BY CPL. NEIDERHISER

**Data Download Date/Time**      05/01/2007      08:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 06/04/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 17:19:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

IJC P

**Data Download Date/Time** 06/04/2007 17:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 07/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 16:42:54	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

TESTED AT IJC OBSERVED BY CPL PEABODY 2279

**Data Download Date/Time** 07/02/2007 16:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 08/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PEABODY,MICHAEL,F	<b>Time</b> 08:50:22	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

TESTED AT IJC CPL. NEIDERHISER OBSERVED TESTING

**Data Download Date/Time** 08/01/2007 08:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 09/05/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PEABODY,MICHAEL,F	<b>Time</b> 10:02:22	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

TESTED AT IJC.

**Data Download Date/Time** 09/05/2007 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 09/13/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VENTURI,GEORGE,L	<b>Time</b> 13:15:28	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6261011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0006	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.078
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.078
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.078
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.077
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.051

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.200
0.08 g/210L Dry Gas Std Test #8	0.078
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.201
0.08 g/210L Dry Gas Std Test #9	0.077
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.050
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.201
0.08 g/210L Dry Gas Std Test #10	0.077
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0012
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time    09/13/2007    16:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 10/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 15:01:23	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

TESTED AT IJC OBSERVED BY CPL PEABODY

**Data Download Date/Time** 10/01/2007 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 11/05/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 15:01:07	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

TESTED AT IJC OBSERVED BY CPL PEABODY

**Data Download Date/Time** 11/05/2007 15:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001651

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 12/04/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PEABODY,MICHAEL,F	<b>Time</b> 09:07:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 622802I		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

TESTED AT IJC.

**Data Download Date/Time** 12/04/2007 09:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 01/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 16:02:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5126011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 05/10/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.054
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 01/31/2007 16:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 02/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 21:32:53	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5126011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 05/10/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 02/28/2007 21:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 03/21/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 17:00:38	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5126011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 05/10/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 03/21/2007 18:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 04/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CHANG,DAVID,	<b>Time</b> 16:11:46	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5126011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**    04/10/2007    17:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 05/09/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 17:56:13	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5126011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
<b>Expiration</b> 05/10/2007		
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 05/09/2007 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 06/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 19:17:24	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 06/22/2007 20:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 07/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 18:18:40	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 07/19/2007 12:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 08/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 19:13:09	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 08/23/2007 19:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 09/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PITA,JORGE,A	<b>Time</b> 16:48:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 09/24/2007 10:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 10/09/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 15:59:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time** 10/15/2007 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 10/15/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 14:37:30	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7096011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0002	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.083
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.078
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.208
		0.08 g/210L Dry Gas Std Test #3	0.078
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.083
		0.20 g/210L Test #4	0.207
		0.08 g/210L Dry Gas Std Test #4	0.078
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.083
		0.20 g/210L Test #5	0.207
		0.08 g/210L Dry Gas Std Test #5	0.078
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.078
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.083
		0.20 g/210L Test #6	0.207
		0.08 g/210L Dry Gas Std Test #6	0.078
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.078
		Alcohol Free Test #7	0.000

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

0.05 g/210L Test #7	0.051
0.08 g/210L Test #7	0.083
0.20 g/210L Test #7	0.208
0.08 g/210L Dry Gas Std Test #7	0.078
0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.051
0.08 g/210L Test #8	0.083
0.20 g/210L Test #8	0.207
0.08 g/210L Dry Gas Std Test #8	0.077
0.08 g/210L Dry Gas Std Test #8 (Retry)	0.078
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.051
0.08 g/210L Test #9	0.083
0.20 g/210L Test #9	0.207
0.08 g/210L Dry Gas Std Test #9	0.077
0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.051
0.08 g/210L Test #10	0.083
0.20 g/210L Test #10	0.207
0.08 g/210L Dry Gas Std Test #10	0.078
0.08 g/210L Dry Gas Std Test #10 (Retry)	0.078
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0245
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

#### Remarks:

08:DRY/GAS HOSE WAS NOT CONNECTED

Data Download Date/Time 10/15/2007 15:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 11/21/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 22:31:27	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D	<b>0.08 Gas Lot #</b> 626102I
<b>Expiration</b> 08/01/2009 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.085
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      11/21/2007      22:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001652

<b>Agency</b> WEST MIAMI PD	<b>Date</b> 12/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 21:16:58	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 12/11/2007 23:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 01/17/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALBERTSON,JASON,W	<b>Time</b> 22:09:18	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

SIMS DRY AND CLEAN

**Data Download Date/Time**      01/17/2007      22:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 02/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALBERTSON,JASON,W	<b>Time</b> 20:43:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

SIMS DRY AND CLEAN

**Data Download Date/Time** 02/15/2007 20:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 03/21/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALBERTSON,JASON,W	<b>Time</b> 23:17:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

SIMS DRY AND CLEAN

**Data Download Date/Time** 03/21/2007 23:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 04/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALBERTSON,JASON,W	<b>Time</b> 23:16:13	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

SIMS DRY AND CLEAN

**Data Download Date/Time** 04/18/2007 23:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 05/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ARKOVICH,ROBBIE,S	<b>Time</b> 20:42:31	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 05/03/2007 20:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 06/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WEISKOPF,MICHAEL,D	<b>Time</b> 21:44:48	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 635301G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 12/21/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      06/13/2007      21:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 07/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WEISKOPF,MICHAEL,D	<b>Time</b> 18:27:01	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 635301G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 12/21/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 07/11/2007 18:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 08/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WEISKOPF,MICHAEL,D	<b>Time</b> 01:13:54	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 635301G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 12/21/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 08/11/2007 01:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 09/11/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SUERETH,DONALD,P	<b>Time</b> 15:46:11	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7096011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0008	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4 (Retry)	0.000
		0.05 g/210L Test #4	0.052
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.078
		Alcohol Free Test #5 (Retry)	0.000
		0.05 g/210L Test #5	0.052
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.203
		0.08 g/210L Dry Gas Std Test #5	0.078
		Alcohol Free Test #6 (Retry)	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.078
		Alcohol Free Test #7 (Retry)	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.203
		0.08 g/210L Dry Gas Std Test #7	0.078
		Alcohol Free Test #8 (Retry)	0.000
		0.05 g/210L Test #8	0.052

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.204
0.08 g/210L Dry Gas Std Test #8	0.078
Alcohol Free Test #9 (Retry)	0.000
0.05 g/210L Test #9	0.052
0.08 g/210L Test #9	0.081
0.20 g/210L Test #9	0.203
0.08 g/210L Dry Gas Std Test #9	0.078
Alcohol Free Test #10 (Retry)	0.000
0.05 g/210L Test #10	0.052
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.203
0.08 g/210L Dry Gas Std Test #10	0.078
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0014
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

#### Remarks:

00: Ambient Fail. ROOM FANNED..

Data Download Date/Time 09/11/2007 15:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 09/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WEISKOPF,MICHAEL,D	<b>Time</b> 21:26:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 635301G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
<b>Expiration</b> 12/21/2008		
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time**      09/26/2007      21:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 10/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WEISKOPF,MICHAEL,D	<b>Time</b> 12:18:03	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 635301G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
<b>Expiration</b> 12/21/2008		
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 10/11/2007 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 11/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WEISKOPF,MICHAEL,D	<b>Time</b> 13:03:06	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 635301G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
<b>Expiration</b> 12/21/2008		
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 11/19/2007 13:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001653

<b>Agency</b> ST PETERSBURG PD	<b>Date</b> 12/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WEISKOPF,MICHAEL,D	<b>Time</b> 21:20:20	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 635301G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
<b>Expiration</b> 12/21/2008		
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 12/20/2007 21:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI PD	<b>Date</b> 01/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 16:05:38	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 01/31/2007 16:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI PD	<b>Date</b> 02/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 21:47:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 02/13/2007 21:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI PD	<b>Date</b> 03/21/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 16:54:36	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time**      03/21/2007      18:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI PD	<b>Date</b> 04/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CHANG,DAVID,	<b>Time</b> 16:38:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 04/10/2007 17:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI PD	<b>Date</b> 05/09/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 18:14:29	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 05/09/2007 18:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI PD	<b>Date</b> 06/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 18:53:19	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 06/22/2007 19:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI PD	<b>Date</b> 07/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 18:03:36	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 07/20/2007 17:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI PD	<b>Date</b> 08/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 19:19:41	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 08/23/2007 19:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI P.D.	<b>Date</b> 09/12/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 15:15:59	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0005	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.078
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.078
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.077
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.077
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.200
0.08 g/210L Dry Gas Std Test #8	0.077
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.201
0.08 g/210L Dry Gas Std Test #9	0.077
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.049
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.201
0.08 g/210L Dry Gas Std Test #10	0.076
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

#### Remarks:

OFFICER PEDRO BELTRAN WAS PRESENT DURING THE INSPECTION.

Data Download Date/Time 09/12/2007 18:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI P.D.	<b>Date</b> 09/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 18:20:33	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 09/12/2007 18:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI P.D.	<b>Date</b> 10/09/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 15:13:16	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b>	No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Int Det:SIMULATOR CONNECTOR TO INLET WAS LOOSE . Non-compliance:A.I. ERROR FOR NO PROPERLY CONNECTING SIM. .

**Data Download Date/Time** 10/17/2007 08:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI P.D.	<b>Date</b> 10/09/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 16:02:15	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 10/17/2007 08:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI P.D.	<b>Date</b> 11/14/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 21:45:00	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      11/14/2007      22:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001654

<b>Agency</b> MIAMI P.D.	<b>Date</b> 12/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 20:54:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

Int Det:FORGOT TO ADD ACETONE.

**Data Download Date/Time** 12/11/2007 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 01/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROBERTSON,TIMTOHY,B	<b>Time</b> 14:35:24	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

00: Ambient Fail.

**Data Download Date/Time**      01/31/2007      14:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 02/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROBERTSON,TIMOTHY,B	<b>Time</b> 15:24:22	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.183
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.185
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.184
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time**      02/27/2007      15:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 03/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROBERTSON,TIMTOHY,B	<b>Time</b> 07:04:57	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 03/31/2007 07:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 04/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROBERTSON,TIMOTHY,B	<b>Time</b> 18:22:26	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 04/30/2007 18:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 05/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROBERTSON,TIMOTHY,B	<b>Time</b> 20:33:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 05/31/2007 20:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 06/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROBERTSON,TIMOTHY,B	<b>Time</b> 13:44:34	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

A F / M A: RFI Detect.

**Data Download Date/Time** 06/30/2007 13:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 07/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROBERTSON,TIMOTHY,B	<b>Time</b> 07:21:09	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 08/03/2007 05:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 08/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROBERTSON,TIMOTHY,B	<b>Time</b> 15:17:38	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 09/02/2007 23:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

**Agency** ST. JOHNS CO SO **Date** 09/29/2007 **Type** Agency Inspection  
**Inspector** ROSADO,ANGELO,D **Time** 18:20:34 **Software** 8100.27

**0.05 Lot #** 200612C **0.08 Lot #** 200612D **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** 12/14/2008 **Expiration** 12/14/2008

**Number of Simulators** 5 **Average Standard Deviation** **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.047
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.078
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #3	0.078
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Non-compliance: .

**Data Download Date/Time** 09/29/2007 18:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 09/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROSADO,ANGELO,D	<b>Time</b> 18:50:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      09/29/2007      18:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

**Agency** ST. JOHNS CO SO **Date** 10/28/2007 **Type** Agency Inspection  
**Inspector** ROSADO,ANGELO,D **Time** 18:09:03 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Number of Simulators** 5 **Average Standard Deviation** **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: . Non-compliance: .

**Data Download Date/Time** 10/30/2007 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 10/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROSADO,ANGELO,D	<b>Time</b> 18:16:19	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

N. Non-compliance: .

**Data Download Date/Time** 10/30/2007 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST. JOHNS CO SO	<b>Date</b> 10/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROSADO,ANGELO,D	<b>Time</b> 19:02:24	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 10/30/2007 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST JOHN'S COUNTY SO	<b>Date</b> 10/30/2007	<b>Type</b> Department Inspection
<b>Inspector</b> MALHIOT, MATTHEW, E	<b>Time</b> 11:13:18	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 6271011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 10/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0005	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.194
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.193
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.194
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.193
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.048

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.192
0.08 g/210L Dry Gas Std Test #8	0.078
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.048
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.193
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.048
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.193
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time      10/30/2007      11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST JOHN'S COUNTY SO	<b>Date</b> 11/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROSADO,ANGELO,D	<b>Time</b> 08:43:33	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 610201G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 04/17/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

FAILED TO USE MA SOLUTION.

**Data Download Date/Time** 11/29/2007 08:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001655

<b>Agency</b> ST JOHN'S COUNTY SO	<b>Date</b> 12/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROSADO,ANGELO,D	<b>Time</b> 14:33:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 719102I
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 07/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time**      12/27/2007      14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 03/06/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VENTURI,GEORGE,L	<b>Time</b> 14:59:45	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 525202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0005	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.077
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.077
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.077
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.078
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.052

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.197
0.08 g/210L Dry Gas Std Test #8	0.078
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.052
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.078
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.052
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.196
0.08 g/210L Dry Gas Std Test #10	0.077
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time    04/10/2007    17:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 04/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 17:13:19	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 525202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time**    04/10/2007    17:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 05/09/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 18:13:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 525202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.076
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.077
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.077

**Remarks:**

08: Control Outside Tolerance, HOSE WAS NOT CONNECTED . FORGOT TO CONNECT DRY GAS HOSE

**Data Download Date/Time** 05/09/2007 18:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 06/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 14:47:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 525202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 06/27/2007 16:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 07/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 18:12:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 525202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time** 07/20/2007 16:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 08/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 19:08:42	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 525202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 08/23/2007 19:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 09/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> DEJESUS, RAMON,	<b>Time</b> 21:07:32	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 525202I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 09/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 09/13/2007 14:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 10/25/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 13:47:31	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 714902I		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 05/31/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0005	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.193
		0.08 g/210L Dry Gas Std Test #4	0.078
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.193
		0.08 g/210L Dry Gas Std Test #5	0.077
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.193
		0.08 g/210L Dry Gas Std Test #6	0.078
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.193
		0.08 g/210L Dry Gas Std Test #7	0.078
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.194
0.08 g/210L Dry Gas Std Test #8	0.078
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.048
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.193
0.08 g/210L Dry Gas Std Test #9	0.078
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.049
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.193
0.08 g/210L Dry Gas Std Test #10	0.078
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

#### Remarks:

OFFICER R. DEJESUS WAS PRESENT DURING THE INSPECTION.

Data Download Date/Time      10/25/2007      14:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 10/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> DEJESUS, RAMON,	<b>Time</b> 14:29:46	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 7149021		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 05/31/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 10/25/2007 14:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 11/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> DE JESUS, RAMON,	<b>Time</b> 20:30:03	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 714902I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 05/31/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      11/29/2007      20:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 12/24/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> DE JESUS, RAMON,	<b>Time</b> 13:25:06	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b>	No

<b>Check or Test</b>	<b>Result</b>	<b>Test</b>	<b>Result</b>
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

00: Ambient Fail. Non-compliance: .

**Data Download Date/Time** 12/24/2007 14:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001656

<b>Agency</b> NORTH MIAMI PD	<b>Date</b> 12/24/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> DE JESUS, RAMON,	<b>Time</b> 14:01:34	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 714902I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 05/31/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

00: Ambient Fail.

**Data Download Date/Time** 12/24/2007 14:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 01/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 19:35:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 531401G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time** 01/19/2007 19:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 02/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 12:32:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 531401G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 04/21/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time**      02/02/2007      12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 03/08/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 13:49:49	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200504C <b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200505C	<b>0.08 Gas Lot #</b> 531401G
<b>Expiration</b> 04/21/2007 <b>Expiration</b> 04/21/2007	<b>Expiration</b> 05/05/2007	<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time**      03/08/2007      13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 04/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VELAZQUEZ,KIMBERLY,	<b>Time</b> 14:05:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 531401G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

PSA ESPINOZA PRESENT FOR AI

**Data Download Date/Time**      04/11/2007      14:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 05/16/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VELAZQUEZ,KIMBERLY,	<b>Time</b> 13:15:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 531401G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 11/14/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

PSA ESPINOZA PRESENT FOR AI

**Data Download Date/Time**      05/16/2007      13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 06/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 11:23:08	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 702401G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 01/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

OFC VELAZQUEZ

**Data Download Date/Time**      06/19/2007      11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 07/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VELAZQUEZ,KIMBERLY,	<b>Time</b> 11:44:23	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 702401G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 01/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

PSA ESPINOZA PRESENT FOR AI

**Data Download Date/Time**      07/18/2007      12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 08/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 11:48:29	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 702401G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 01/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time** 08/15/2007 12:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 09/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 13:45:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 702401G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 01/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 09/03/2007 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 10/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VELAZQUEZ,KIMBERLY,	<b>Time</b> 10:16:16	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 702401G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 01/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

PSA ESPINOZA PRESENT FOR AI

**Data Download Date/Time**      10/10/2007      17:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 10/10/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 17:04:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612C	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 702401G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 01/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0006	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.077
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.077
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.077
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.078
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.048

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.197
0.08 g/210L Dry Gas Std Test #8	0.077
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.048
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.077
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.048
0.08 g/210L Test #10	0.081
0.20 g/210L Test #10	0.197
0.08 g/210L Dry Gas Std Test #10	0.078
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0007
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

#### Remarks:

OFFICER K. VELAZQUEZ & PSA ESPINOZA WERE PRESENT FOR INSPEC

Data Download Date/Time 10/10/2007 17:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 11/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 22:41:40	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 702401G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 01/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 11/30/2007 21:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

**Agency** MICCOSUKEE PD **Date** 11/29/2007 **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO, **Time** 22:45:01 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Number of Simulators** 0 **Average Standard Deviation** **Compliance** Yes

<b>Check or Test</b>	<b>Result</b>	<b>Test</b>	<b>Result</b>
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

**Data Download Date/Time** 11/30/2007 21:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001657

<b>Agency</b> MICCOSUKEE PD	<b>Date</b> 12/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 11:29:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 702401G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 01/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

OFC VELAZQUEZ PRESENT FOR AI

**Data Download Date/Time** 12/11/2007 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

<b>Agency</b> UNIV OF FLORIDA PD	<b>Date</b> 04/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MALLO,NICOLE,L	<b>Time</b> 14:48:53	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 629302I
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 12/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      04/27/2007      14:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

<b>Agency</b> UNIV OF FLORIDA PD	<b>Date</b> 05/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SASSER,WILLIAM,C	<b>Time</b> 16:52:42	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A]
<b>0.08 Gas Lot #</b> 629302I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 10/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      05/10/2007      17:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

<b>Agency</b> UNIV OF FLORIDA PD	<b>Date</b> 06/13/2007	<b>Type</b> Department Inspection
<b>Inspector</b> MALHIOT, MATTHEW, E	<b>Time</b> 11:44:07	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5187011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/08/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0003	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.081
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.081
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.195
		0.08 g/210L Dry Gas Std Test #6	0.081
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.080
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.050

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.196
0.08 g/210L Dry Gas Std Test #8	0.081
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.196
0.08 g/210L Dry Gas Std Test #9	0.080
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.051
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.196
0.08 g/210L Dry Gas Std Test #10	0.080
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time      06/23/2007      18:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

<b>Agency</b> UNIV OF FLORIDA PD	<b>Date</b> 06/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MALLO,NICOLE,L	<b>Time</b> 17:56:31	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 629302I
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 10/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      06/23/2007      18:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

<b>Agency</b> UNIV OF FLORIDA PD	<b>Date</b> 07/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MALLO,NICOLE,L	<b>Time</b> 12:21:41	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 629302I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 10/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

A F / M A: Range Exceeded.

**Data Download Date/Time** 07/26/2007 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

<b>Agency</b> UNIV OF FLORIDA PD	<b>Date</b> 08/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MALLO,NICOLE,L	<b>Time</b> 22:01:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 629302I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 10/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

Time-Date changed.

**Data Download Date/Time**      08/18/2007      22:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

<b>Agency</b> UNIV OF FLORIDA PD	<b>Date</b> 09/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SASSER,WILLIAM,C	<b>Time</b> 17:32:04	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 629302I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 10/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 10/09/2007 09:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

<b>Agency</b> UNIV OF FLORIDA PD	<b>Date</b> 10/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SASSER,WILLIAM,C	<b>Time</b> 17:13:08	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 724801G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 09/06/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 10/30/2007 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

<b>Agency</b> UNIV OF FLORIDA PD	<b>Date</b> 11/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MALLO,NICOLE,L	<b>Time</b> 02:36:13	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 724801G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 09/06/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time**      11/30/2007      11:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001714

<b>Agency</b> UNIV OF FLORIDA PD	<b>Date</b> 12/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MALLO,NICOLE,L	<b>Time</b> 10:16:14	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 724801G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 09/06/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

A F / M A:TOO LITTLE MOUTH ALCOHOL/STILL HAD M/A FROM PRIOR

**Data Download Date/Time**      12/23/2007      10:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 01/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 07:33:06	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 5159011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 06/13/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 01/17/2007 18:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 02/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 06:59:54	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
<b>Number of Simulators</b> 4		<b>Average Standard Deviation</b>
		<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 02/12/2007 07:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 03/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 07:35:45	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 5159011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 06/13/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**    03/25/2007    08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 04/24/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 17:38:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 5159011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 06/13/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 04/24/2007 18:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 05/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 07:12:13	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 5159011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 06/13/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 05/19/2007 07:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 06/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 07:31:19	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7050011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 02/22/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      06/28/2007      07:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 07/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 02:38:32	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7050011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 02/22/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

A F / M A: Sequence Aborted.

**Data Download Date/Time**      07/25/2007      04:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 07/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 03:32:22	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7050011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 02/22/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      07/25/2007      04:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 08/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 07:31:18	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7050011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 02/22/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      08/27/2007      08:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 09/05/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 09:01:23	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7050011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 02/22/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 09/05/2007 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 09/05/2007	<b>Type</b> Department Inspection
<b>Inspector</b> HACKNEY,DWITE,N	<b>Time</b> 15:06:37	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 6272011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 10/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0003	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.081
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.081
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.081
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.081
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.051

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.197
0.08 g/210L Dry Gas Std Test #8	0.081
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.051
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.081
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.050
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.198
0.08 g/210L Dry Gas Std Test #10	0.081
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time    10/26/2007    07:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 10/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 07:31:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7050011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 02/22/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 10/26/2007 07:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 11/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 07:01:35	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7050011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 02/22/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 11/28/2007 07:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001715

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 12/17/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CAVINEE,KEVIN,A	<b>Time</b> 06:50:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7050011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 02/22/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

A F / M A: Improper Sample.

**Data Download Date/Time** 12/17/2007 07:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

<b>Agency</b> ALACHUA PD	<b>Date</b> 04/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> FULWOOD,GORDON,H	<b>Time</b> 16:39:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 5187011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/08/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 04/10/2007 16:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

<b>Agency</b> ALACHUA PD	<b>Date</b> 05/16/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> FULWOOD,GORDON,H	<b>Time</b> 15:23:30	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 5187011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 07/08/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 05/16/2007 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

<b>Agency</b> ALACHUA PD	<b>Date</b> 06/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> FULWOOD,GORDON,H	<b>Time</b> 15:03:31	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 632002I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      06/13/2007      15:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

<b>Agency</b> ALACHUA PD	<b>Date</b> 07/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> FULWOOD,GORDON,H	<b>Time</b> 10:41:04	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 632002I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 07/20/2007 10:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

<b>Agency</b> ALACHUA PD	<b>Date</b> 08/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> FULWOOD,GORDON,H	<b>Time</b> 17:33:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 632002I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      08/31/2007      17:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

<b>Agency</b> ALACHUA PD	<b>Date</b> 09/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PARKS,DONALD,M	<b>Time</b> 20:54:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 632002I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

Time-Date changed.

**Data Download Date/Time**      09/27/2007      20:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

<b>Agency</b> ALACHUA PD	<b>Date</b> 10/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PARKS,DONALD,M	<b>Time</b> 20:45:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 632002I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 10/29/2007 20:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

<b>Agency</b> ALACHUA PD	<b>Date</b> 10/31/2007	<b>Type</b> Department Inspection
<b>Inspector</b> MALHIOT, MATTHEW, E	<b>Time</b> 11:08:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 6271011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 10/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0003	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.080
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.080
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.080
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.198
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.049
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.198
0.08 g/210L Dry Gas Std Test #9	0.080
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.048
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.198
0.08 g/210L Dry Gas Std Test #10	0.080
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time      11/28/2007      01:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

<b>Agency</b> ALACHUA PD	<b>Date</b> 11/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PARKS,DONALD,M	<b>Time</b> 01:08:42	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200708C	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 632002I		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 11/28/2007 01:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001720

<b>Agency</b> ALACHUA PD	<b>Date</b> 12/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PARKS,DONALD,M	<b>Time</b> 20:40:26	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200708C	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 632002I		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 12/28/2007 20:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 01/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MORGAN,DARIN,P	<b>Time</b> 23:22:33	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200504C	<b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 502701G		
<b>Expiration</b> 04/21/2007	<b>Expiration</b> 04/21/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 02/02/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 01/25/2007 23:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 02/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MORGAN,DARIN,P	<b>Time</b> 23:42:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200504C	<b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 510902I		
<b>Expiration</b> 04/21/2007	<b>Expiration</b> 04/21/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 04/21/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

TOO MUCH MOUTH ALCOHOL.

**Data Download Date/Time**      02/22/2007      23:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 03/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MORGAN,DARIN,P	<b>Time</b> 12:16:31	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200504C	<b>0.08 Lot #</b> 200504D	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 510902I		
<b>Expiration</b> 04/21/2007	<b>Expiration</b> 04/21/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 04/21/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      03/23/2007      12:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 04/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MORGAN,DARIN,P	<b>Time</b> 20:27:34	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time**      04/26/2007      20:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 05/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MORGAN,DARIN,P	<b>Time</b> 19:17:54	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      05/30/2007      19:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 06/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MORGAN,DARIN,P	<b>Time</b> 16:55:16	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 06/12/2007 16:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 07/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MARCINIK,DAVID,E	<b>Time</b> 18:54:55	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

Time-Date changed. A F / M A:BAD MOUTH PIECE.

**Data Download Date/Time**      07/26/2007      18:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 08/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MARCINIK,DAVID,E	<b>Time</b> 18:35:41	<b>Software</b> 8100.27
<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>
<b>0.08 Gas Lot #</b>		
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Time-Date changed. A F / M A:OPER ERROR. Non-compliance: .

**Data Download Date/Time** 08/30/2007 19:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 08/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MARCINIK,DAVID,E	<b>Time</b> 19:24:40	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

OPER ERR.

**Data Download Date/Time** 08/30/2007 19:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 09/20/2007	<b>Type</b> Department Inspection
<b>Inspector</b> HACKNEY,DWITE,N	<b>Time</b> 10:21:48	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 6272011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 10/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0005	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.081
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.195
		0.08 g/210L Dry Gas Std Test #6	0.081
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.196
		0.08 g/210L Dry Gas Std Test #7	0.080
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.196
0.08 g/210L Dry Gas Std Test #8	0.080
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.049
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.196
0.08 g/210L Dry Gas Std Test #9	0.081
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.049
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.197
0.08 g/210L Dry Gas Std Test #10	0.080
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time    09/20/2007    13:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 09/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MARCINIK,DAVID,E	<b>Time</b> 12:58:18	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 706102I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 03/05/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      09/20/2007      13:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 10/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MARCINIK,DAVID,E	<b>Time</b> 13:22:49	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 714102I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 05/22/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

Time-Date changed.

**Data Download Date/Time**      10/29/2007      13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 11/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MARCINIK,DAVID,E	<b>Time</b> 20:47:26	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 714102I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 05/22/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.045
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

08: Control Outside Tolerance/LOOSE CONNECTION.

**Data Download Date/Time** 11/24/2007 04:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001721

<b>Agency</b> PALM BAY PD	<b>Date</b> 12/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MARCINIK,DAVID,E	<b>Time</b> 19:44:53	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 714102I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 05/22/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 12/27/2007 19:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 01/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 11:01:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 504502I		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 02/17/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 01/26/2007 10:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 02/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 12:12:58	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 504502I		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 02/17/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      02/02/2007      12:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 03/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 13:01:40	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time**      03/02/2007      12:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 04/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 13:34:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 04/03/2007 13:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 05/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 16:26:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200505B	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 05/02/2007 16:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 06/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 10:34:18	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 06/15/2007 10:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 07/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 13:15:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 612201G
<b>0.08 Gas Lot #</b> 701201G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2008
		<b>Expiration</b> 01/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 07/13/2007 13:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 08/16/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 12:27:15	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 612202G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 05/05/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.080

**Remarks:**

08: Control Outside Tolerance GAS TUBE NOT ATTACHED . OPERATOR ERROR

**Data Download Date/Time** 08/22/2007 11:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 08/22/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VENTURI,GEORGE,L	<b>Time</b> 15:52:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6261011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.199
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.049
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.199
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.049
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.199
0.08 g/210L Dry Gas Std Test #10	0.080
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0012
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time    08/22/2007    16:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 09/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 12:01:42	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 612201G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 05/05/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 09/12/2007 12:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 10/05/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 11:10:52	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 612202G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 05/05/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079

**Remarks:**

08: Control Outside Tolerance TANK NOT ATTACHED. OP ERROR

**Data Download Date/Time** 10/18/2007 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 11/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 10:47:53	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 702401G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 01/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 11/20/2007 10:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001722

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 12/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN,DAVID,S	<b>Time</b> 15:29:58	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 702401G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 01/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 12/11/2007 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001723

<b>Agency</b> VENICE PD	<b>Date</b> 01/24/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VENTURI,GEORGE,L	<b>Time</b> 17:47:54	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6261011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0002	<b>Compliance</b> No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.050

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001723

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.198
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.050
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.197
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

#### Remarks:

COMPLIES WITH 11D-8 F.A.C..

Data Download Date/Time 01/24/2007 18:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001723

<b>Agency</b> VENICE PD	<b>Date</b> 10/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BUTLER,CHRIS,T	<b>Time</b> 11:58:29	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

Time-Date changed.PUT INTO SERVICE AT S JAIL.

**Data Download Date/Time** 10/15/2007 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001723

<b>Agency</b> VENICE PD	<b>Date</b> 11/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BUTLER,CHRIS,T	<b>Time</b> 12:34:15	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 608301G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 03/27/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

INSPECTION DONE AT S JAIL.

**Data Download Date/Time** 11/13/2007 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001723

<b>Agency</b> VENICE PD	<b>Date</b> 12/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BUTLER,CHRIS,T	<b>Time</b> 09:26:16	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7235011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 08/26/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

INSPECTION DONE AT SOUTH JAIL

**Data Download Date/Time** 12/03/2007 09:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 01/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 10:36:58	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 01/18/2007 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 02/21/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 14:02:46	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      02/21/2007      13:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 03/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 12:18:38	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.20 g/210L Test #3 (Retry)	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

SIMULATOR TURNED OFF TEMP TOO LOW.

**Data Download Date/Time**      03/20/2007      12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 04/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 11:57:52	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 04/11/2007 11:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 05/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 11:33:44	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 518602I		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/07/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.206
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.051
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.207
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.044
		0.05 g/210L Test #3 (Retry)	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.206
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

LEAK AT SIMULATOR PORT

**Data Download Date/Time**      05/25/2007      11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 06/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 11:16:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 518602I		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/07/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      06/27/2007      11:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 07/18/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SUERETH,DONALD,P	<b>Time</b> 14:02:09	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6324011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/22/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.201
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.051

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.202
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.202
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.051
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.202
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time    07/19/2007    11:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 07/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 11:41:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6324011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 11/22/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      07/19/2007      11:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 08/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 11:25:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6324011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 11/22/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 08/28/2007 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 09/21/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 10:42:20	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6324011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/22/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.072
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

08: TUBING LOOSE AT SIMULATOR PORT

**Data Download Date/Time**      09/21/2007      10:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 10/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 10:01:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6324011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/22/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 10/15/2007 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 11/21/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 09:30:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 3624011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/22/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 11/21/2007 09:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001731

<b>Agency</b> HILLSBOROUGH SO	<b>Date</b> 12/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GRAY,BEVERLY,A	<b>Time</b> 10:10:09	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6324011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/22/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 12/20/2007 10:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001737

<b>Agency</b> SOUTH MIAMI P.D.	<b>Date</b> 09/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 12:52:06	<b>Software</b> 8100.26

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 0	<b>Average Standard Deviation</b>	<b>Compliance</b> No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

NOT IN SERVICE Non-compliance: .

**Data Download Date/Time**      09/12/2007      13:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001737

<b>Agency</b> SOUTH MIAMI P.D.	<b>Date</b> 09/12/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 16:13:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7096011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/05/2008
<b>Expiration</b> 04/08/2009		
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0006	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.080
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.077
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.080
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.080
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001737

0.08 g/210L Test #8	0.078
0.20 g/210L Test #8	0.197
0.08 g/210L Dry Gas Std Test #8	0.080
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.048
0.08 g/210L Test #9	0.078
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.080
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.049
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.197
0.08 g/210L Dry Gas Std Test #10	0.080
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

OFFICER PEDRO BELTRAN WAS PRESENT DURING THE INSPECTION.

**Data Download Date/Time**      09/12/2007      16:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001737

<b>Agency</b> SOUTH MIAMI P.D.	<b>Date</b> 09/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BELTRAN,PEDRO,	<b>Time</b> 18:30:59	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 09/12/2007 18:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001737

<b>Agency</b> SOUTH MIAMI P.D.	<b>Date</b> 10/09/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WEISSBERG,M,	<b>Time</b> 14:56:16	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 531505G
<b>Expiration</b> 05/04/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time**      10/11/2007      22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001737

<b>Agency</b> SOUTH MIAMI P.D.	<b>Date</b> 11/14/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WEISSBERG,M,	<b>Time</b> 17:42:57	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D	<b>0.08 Gas Lot #</b> 531505G
<b>Expiration</b> 08/01/2009 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.191
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

Time-Date changed.

**Data Download Date/Time**      11/14/2007      17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001737

<b>Agency</b> SOUTH MIAMI P.D.	<b>Date</b> 12/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WEISSBERG,MICHAEL,W	<b>Time</b> 21:58:27	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 626102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

A F / M A: DID NOT USE MOUTH ALCOHOL SOLN.

**Data Download Date/Time** 12/11/2007 22:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 01/09/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEICHER,DALE,A	<b>Time</b> 21:04:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

JANUARY INSPECTION

**Data Download Date/Time** 01/09/2007 20:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 02/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEICHER,DALE,A	<b>Time</b> 11:51:06	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 02/11/2007 18:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 02/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEICHER,DALE,A	<b>Time</b> 18:48:20	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

RETEST OF FEBRUARY AGENCY INSPECTION

**Data Download Date/Time** 02/11/2007 18:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 03/09/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEICHER,DALE,A	<b>Time</b> 08:10:49	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200505C	<b>0.08 Gas Lot #</b> 531505G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/05/2007	<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

MARCH 2007 AGENCY INSPECTION

**Data Download Date/Time**      03/09/2007      09:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 04/21/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PUTMAN,BRIAN,A	<b>Time</b> 19:41:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200505C	<b>0.08 Gas Lot #</b> 531505G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/05/2007	<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

APRIL INSPECTION

**Data Download Date/Time**      04/21/2007      19:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 05/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEICHER,DALE,A	<b>Time</b> 12:32:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 531505G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

MAY 2007 AGENCY INSPECTION

**Data Download Date/Time**      05/23/2007      12:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 06/17/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEICHER,DALE,A	<b>Time</b> 10:07:58	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

JUNE AGENCY INSP

**Data Download Date/Time** 06/17/2007 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 07/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEICHER,DALE,A	<b>Time</b> 13:37:26	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

FAIL PURGE TEST.

**Data Download Date/Time**      07/28/2007      15:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 07/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEICHER,DALE,A	<b>Time</b> 15:22:35	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 531505G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.043
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.044
		0.05 g/210L Test #3 (Retry)	0.046
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time**      07/28/2007      15:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 08/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEICHER,DALE,A	<b>Time</b> 14:14:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 531505G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 3	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

AUGUST AGENCY INSPECTION

**Data Download Date/Time**      08/11/2007      14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 08/30/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VENTURI,GEORGE,L	<b>Time</b> 14:23:16	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6261011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.195
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.195
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.194
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.047
		0.08 g/210L Test #7	0.078
		0.20 g/210L Test #7	0.195
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.047

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

0.08 g/210L Test #8	0.078
0.20 g/210L Test #8	0.195
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.047
0.08 g/210L Test #9	0.078
0.20 g/210L Test #9	0.195
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.046
0.08 g/210L Test #10	0.078
0.20 g/210L Test #10	0.195
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 08/30/2007 14:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 09/08/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SPEICHER,DALE,A	<b>Time</b> 14:36:14	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 531505G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/15/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

OPERATOR ERROR, REMOVED FROM SERVICE FOR PM

**Data Download Date/Time** 09/08/2007 14:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 11/14/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VENTURI,GEORGE,L	<b>Time</b> 16:04:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605C
<b>0.08 Gas Lot #</b> 6261011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.078
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.078
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.077
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.050

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.198
0.08 g/210L Dry Gas Std Test #8	0.078
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.198
0.08 g/210L Dry Gas Std Test #9	0.078
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.050
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.198
0.08 g/210L Dry Gas Std Test #10	0.078
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

#### Remarks:

INSPECTION FOLLOWING RETURN FROM AUTHORIZED REPAIR FACILITY

Data Download Date/Time 11/14/2007 16:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

**Agency** CHARLOTTE COUNTY SO **Date** 11/16/2007 **Type** Agency Inspection  
**Inspector** WRIGHT,TAMMY,M **Time** 21:45:34 **Software** 8100.27

**0.05 Lot #** 610301G **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** 04/18/2008

**Number of Simulators** 5 **Average Standard Deviation** **Compliance** No

<u>Check or Test</u>	<u>Result</u>	<u>Test</u>	<u>Result</u>
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.048
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.049
0.05 g/210L Test: OK	Yes		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

USED WRONG LOT NUMBERS AND EXPIRATION DATES. REDO AG INSP USED WRONG LOT / EXP INFORMATION.

**Data Download Date/Time** 11/16/2007 22:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 11/16/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WRIGHT,TAMMY,M	<b>Time</b> 22:26:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 610301G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 04/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

RETESTED TEST PASS

**Data Download Date/Time**      11/16/2007      22:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001739

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 12/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WRIGHT,TAMMY,M	<b>Time</b> 17:31:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 610301G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 04/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

AGENCY INSPECTION MONTH OF DECEMBER

**Data Download Date/Time** 12/20/2007 17:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 01/24/2007	<b>Type</b> Agency Inspection	
<b>Inspector</b> MILLER,THEODORE,J	<b>Time</b> 16:59:37	<b>Software</b> 8100.27	
<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 0	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes	

<b>Check or Test</b>	<b>Result</b>	<b>Test</b>	<b>Result</b>
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

**Data Download Date/Time** 01/26/2007 07:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 01/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILNE,THOMAS,P	<b>Time</b> 06:51:53	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6153011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 06/05/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time**      01/26/2007      07:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 02/16/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLER,THEODORE,J	<b>Time</b> 12:52:36	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6153011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/05/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

A F / M A: No Sample Provided.

**Data Download Date/Time**      02/16/2007      12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 02/20/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SUERETH,DONALD,P	<b>Time</b> 13:50:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 518602I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/07/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b> 0.0006	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.052
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.080
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.080
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.052

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

0.08 g/210L Test #8	0.081
0.20 g/210L Test #8	0.200
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.052
0.08 g/210L Test #9	0.081
0.20 g/210L Test #9	0.199
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.052
0.08 g/210L Test #10	0.081
0.20 g/210L Test #10	0.200
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time      02/20/2007      13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 03/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILNE,THOMAS,P	<b>Time</b> 04:50:46	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6153011		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 06/05/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time**      03/23/2007      04:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 04/10/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLER,THEODORE,J	<b>Time</b> 18:59:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6153011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/05/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 04/10/2007 19:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 05/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILNE,THOMAS,P	<b>Time</b> 04:04:24	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6153011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/05/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.179
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.180
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.182
		0.20 g/210L Test #3 (Retry)	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time**      05/25/2007      04:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 06/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLER,THEODORE,J	<b>Time</b> 09:59:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6153011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/05/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

MOUCH ALCOHOL LOT 2006-A ACETONE LOT

**Data Download Date/Time** 06/22/2007 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 07/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILNE,THOMAS,P	<b>Time</b> 05:06:29	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6153011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/05/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time**      07/27/2007      05:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 08/16/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLER,THEODORE,J	<b>Time</b> 17:05:07	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 615802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/09/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

MOUTH ALCOHOL LOT 2006-A ACETONE LOT 2006-D

**Data Download Date/Time** 08/16/2007 17:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 09/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILNE,THOMAS,P	<b>Time</b> 15:32:44	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 615802I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/09/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

INST TAKEN O/S FOR CALIBRATION CHECKS

**Data Download Date/Time**      09/11/2007      16:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 11/08/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SUERETH,DONALD,P	<b>Time</b> 14:08:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7096011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b> 0.0006	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.080
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.080
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.080
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.048

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.198
0.08 g/210L Dry Gas Std Test #8	0.080
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.048
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.198
0.08 g/210L Dry Gas Std Test #9	0.080
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.047
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.198
0.08 g/210L Dry Gas Std Test #10	0.081
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0015
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time      11/08/2007      14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

<b>Agency</b> CLEARWATER PD	<b>Date</b> 11/28/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SUERETH,DONALD,P	<b>Time</b> 17:39:53	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605C
<b>0.08 Gas Lot #</b> 7096011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0008	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.000
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.047
		0.08 g/210L Test #4	0.076
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.000
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.080
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.047
		0.08 g/210L Test #5	0.076
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.000
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.047
		0.08 g/210L Test #6	0.077
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.000
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.080
		Alcohol Free Test #7	0.000

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001740

0.05 g/210L Test #7	0.047
0.08 g/210L Test #7	0.077
0.20 g/210L Test #7	0.197
0.08 g/210L Dry Gas Std Test #7	0.000
0.08 g/210L Dry Gas Std Test #7 (Retry)	0.080
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.047
0.08 g/210L Test #8	0.076
0.20 g/210L Test #8	0.197
0.08 g/210L Dry Gas Std Test #8	0.000
0.08 g/210L Dry Gas Std Test #8 (Retry)	0.080
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.048
0.08 g/210L Test #9	0.077
0.20 g/210L Test #9	0.198
0.08 g/210L Dry Gas Std Test #9	0.000
0.08 g/210L Dry Gas Std Test #9 (Retry)	0.080
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.048
0.08 g/210L Test #10	0.077
0.20 g/210L Test #10	0.197
0.08 g/210L Dry Gas Std Test #10	0.000
0.08 g/210L Dry Gas Std Test #10 (Retry)	0.081
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0007
0.20 g/210L Test (Std Dev)	0.0016
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

#### Remarks:

08: Control Outside Tolerance. GAS HOSE DISCONNECTED.

Data Download Date/Time 11/28/2007 17:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 01/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 12:18:22	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 610001G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 01/12/2007 12:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 02/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 08:49:32	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 632102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      02/23/2007      08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 03/16/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 10:46:41	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 632102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      03/16/2007      10:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 04/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 11:44:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 632102I
<b>Expiration</b> 05/04/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      04/13/2007      14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 05/04/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 09:22:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7061011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 03/05/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 05/04/2007 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 06/08/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 12:11:57	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 7061011
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 03/05/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      06/08/2007      12:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 07/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 11:02:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      07/13/2007      11:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 07/26/2007	<b>Type</b> Department Inspection
<b>Inspector</b> HACKNEY,DWITE,N	<b>Time</b> 14:53:09	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 6272011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2006	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 10/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.080
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.080
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.048
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.080
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.197
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.048
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.080
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.048
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.197
0.08 g/210L Dry Gas Std Test #10	0.080
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time      08/17/2007      11:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 08/17/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 10:59:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 632102I
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      08/17/2007      11:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 09/14/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 12:13:28	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      09/14/2007      12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 10/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 10:19:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      10/19/2007      10:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 11/21/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 11:23:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D	<b>0.08 Gas Lot #</b> 7149011
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 05/31/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      11/21/2007      11:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001741

<b>Agency</b> PALM BEACH CO SO	<b>Date</b> 12/14/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CROUCHER,GREGORY,P	<b>Time</b> 12:29:09	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D	<b>0.08 Gas Lot #</b> 610001G
<b>Expiration</b> 08/01/2009 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 04/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      12/14/2007      12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 01/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 09:17:36	<b>Software</b> 8100.27

<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008		

**Number of Simulators** 5      **Average Standard Deviation**      **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test #2	0.046
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test #2	0.075
Alcohol Free Test: OK	Yes	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.046
0.08 g/210L Test: OK	No	0.08 g/210L Test #3	0.076
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

08: Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time**      01/23/2007      09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 03/21/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VENTURI,GEORGE,L	<b>Time</b> 14:09:27	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6261011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 09/20/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.078
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.077
		0.08 g/210L Dry Gas Std Test #4 (Retry)	0.078
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.077
		0.08 g/210L Dry Gas Std Test #5 (Retry)	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.079
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		0.08 g/210L Dry Gas Std Test #6 (Retry)	0.078
		Alcohol Free Test #7	0.000

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

0.05 g/210L Test #7	0.049
0.08 g/210L Test #7	0.079
0.20 g/210L Test #7	0.198
0.08 g/210L Dry Gas Std Test #7	0.078
0.08 g/210L Dry Gas Std Test #7 (Retry)	0.078
Alcohol Free Test #8	0.000
0.05 g/210L Test #8	0.049
0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.198
0.08 g/210L Dry Gas Std Test #8	0.078
0.08 g/210L Dry Gas Std Test #8 (Retry)	0.078
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.079
0.20 g/210L Test #9	0.198
0.08 g/210L Dry Gas Std Test #9	0.078
0.08 g/210L Dry Gas Std Test #9 (Retry)	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.049
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.198
0.08 g/210L Dry Gas Std Test #10	0.078
0.08 g/210L Dry Gas Std Test #10 (Retry)	0.078
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0245
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

#### Remarks:

08: Control Outside Tolerance: DGS HOSE NOT CONNECTED.

Data Download Date/Time 03/21/2007 14:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 04/17/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 11:51:11	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 606002I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 03/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 04/17/2007 11:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 05/08/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 14:23:06	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 606002I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 03/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 05/08/2007 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 06/05/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 11:42:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 606002I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 03/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

Time-Date changed.

**Data Download Date/Time**      06/05/2007      11:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 07/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 12:58:28	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 606002I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 03/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 07/05/2007 09:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 08/01/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 12:36:00	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 606002I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 03/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      08/01/2007      12:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 09/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 09:53:34	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 606002I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 03/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      09/12/2007      09:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 10/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 16:06:13	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 606002I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 03/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 10/02/2007 16:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 11/06/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 16:24:38	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 717102I		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 11/06/2007 16:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 12/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 13:02:51	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Time-Date changed. A F / M A: .

**Data Download Date/Time**      12/13/2007      14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 12/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 13:33:28	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>0.08 Gas Lot #</b>
	<b>Expiration</b> 12/05/2008	

**Number of Simulators** 5      **Average Standard Deviation**      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.077
Alcohol Free Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test #3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197

**Remarks:**

**Data Download Date/Time**      12/13/2007      14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001742

<b>Agency</b> BRADENTON PD	<b>Date</b> 12/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ANDREWS,ROBERT,P	<b>Time</b> 14:09:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 717102I		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 12/13/2007 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 01/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 13:37:06	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 01/19/2007 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 02/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 12:41:41	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 02/13/2007 12:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 03/05/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SUERETH,DONALD,P	<b>Time</b> 15:40:42	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b> 0.0005	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.082
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.082
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.052
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.082
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.052
		0.08 g/210L Test #5	0.082
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.082
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.053
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.199
		0.08 g/210L Dry Gas Std Test #6	0.082
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.053
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.199
		0.08 g/210L Dry Gas Std Test #7	0.082
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.053

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

0.08 g/210L Test #8	0.081
0.20 g/210L Test #8	0.200
0.08 g/210L Dry Gas Std Test #8	0.082
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.052
0.08 g/210L Test #9	0.082
0.20 g/210L Test #9	0.200
0.08 g/210L Dry Gas Std Test #9	0.082
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.052
0.08 g/210L Test #10	0.082
0.20 g/210L Test #10	0.200
0.08 g/210L Dry Gas Std Test #10	0.082
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time      03/05/2007      15:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 03/14/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 12:10:18	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.082
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 03/14/2007 12:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 04/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 11:38:36	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/24/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 04/27/2007 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 05/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 10:44:16	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 05/18/2007 10:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 06/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 12:29:52	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 06/26/2007 12:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 07/24/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 12:05:34	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time**      07/24/2007      12:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 08/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 12:42:15	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.076
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      08/27/2007      12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 09/24/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 12:23:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 09/24/2007 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 10/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 11:44:15	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 10/24/2007 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 11/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 12:13:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 11/29/2007 08:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001865

<b>Agency</b> LAKELAND PD	<b>Date</b> 12/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> KNOBLOCH,WILLIAM,T	<b>Time</b> 12:47:08	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6191011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 07/12/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 12/26/2007 12:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 01/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SMITH,ELMER,H	<b>Time</b> 22:25:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.000
		0.20 g/210L Test #3 (Retry)	0.193
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

20: CONTROL OUTSIDE TOLERANCE LOOSE HOOK UP

**Data Download Date/Time** 02/12/2007 16:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 02/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SMITH,ELMER,H	<b>Time</b> 20:39:16	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 535402I
<b>Expiration</b> 05/04/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079

**Remarks:**

TANK NOT CONNECTED ON FIRST TEST.

**Data Download Date/Time**      03/08/2007      22:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 03/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> SMITH,ELMER,H	<b>Time</b> 19:15:56	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 535402I
<b>Expiration</b> 05/04/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      04/03/2007      14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 04/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> RICE,NYLES,L	<b>Time</b> 05:48:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.188
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.190
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.190
		0.05 g/210L Test #3 (Retry)	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 05/02/2007 08:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 06/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> RICE,NYLES,L	<b>Time</b> 07:04:32	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test #3	0.000
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Int Det: .

**Data Download Date/Time** 06/03/2007 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 06/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> RICE,NYLES,L	<b>Time</b> 09:08:24	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.078
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.079

**Remarks:**

08: RFI Detect.

**Data Download Date/Time**      06/03/2007      09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 07/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> RICE,NYLES,L	<b>Time</b> 15:33:22	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

A F / M A: Range Exceeded.

**Data Download Date/Time**      07/31/2007      15:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 08/01/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SUERETH,DONALD,P	<b>Time</b> 09:58:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 7096011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.197
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.196
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.051

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

0.08 g/210L Test #8	0.081
0.20 g/210L Test #8	0.197
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.052
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.052
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.197
0.08 g/210L Dry Gas Std Test #10	0.078
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time      08/01/2007      10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 09/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> RICE,NYLES,L	<b>Time</b> 15:33:42	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.075
		0.08 g/210L Test #3 (Retry)	0.077
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 10/10/2007 14:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 10/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> RICE,NYLES,L	<b>Time</b> 08:52:37	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 10/19/2007 09:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001866

<b>Agency</b> US AIR FORCE MACDILL	<b>Date</b> 12/04/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> RICE,NYLES,L	<b>Time</b> 00:13:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 535402I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 12/22/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.185
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.187
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.189
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.193
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 12/07/2007 07:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 01/02/2007	<b>Type</b> Department Inspection
<b>Inspector</b> HACKNEY,DWITE,N	<b>Time</b> 15:40:03	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 518702I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 07/08/2007
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.077
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.078
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.078
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.077
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.077
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.051

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

0.08 g/210L Test #8	0.077
0.20 g/210L Test #8	0.198
0.08 g/210L Dry Gas Std Test #8	0.078
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.077
0.20 g/210L Test #9	0.197
0.08 g/210L Dry Gas Std Test #9	0.078
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.050
0.08 g/210L Test #10	0.077
0.20 g/210L Test #10	0.198
0.08 g/210L Dry Gas Std Test #10	0.078
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time    01/03/2007    11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 01/03/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS, IAN,	<b>Time</b> 11:38:30	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 515902I
<b>Expiration</b> 05/04/2008 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 06/13/2007
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      01/03/2007      11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 02/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 17:58:45	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 02/28/2007 18:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 03/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 13:16:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time**      03/26/2007      13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 04/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 11:42:07	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time** 04/11/2007 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 05/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 12:32:52	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time** 05/31/2007 12:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 06/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 22:23:12	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time**      06/27/2007      22:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 07/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 16:40:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time**      07/29/2007      16:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 08/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 16:38:25	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.075

**Remarks:**

DID NOT HAVE SIMULATOR HOOKED UP FOR 1ST PART OF 1ST SAMPLE.

**Data Download Date/Time** 08/28/2007 16:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 09/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 15:28:07	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.070
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.071
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.071
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

INLET TUBE NOT SEATED SNUG.

**Data Download Date/Time** 10/04/2007 15:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 10/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 04:54:32	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time** 10/31/2007 05:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 11/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 23:00:03	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.076
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time**      11/23/2007      23:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-001962

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 12/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HARRIS,OFC,I	<b>Time</b> 16:42:41	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.075
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.076

**Remarks:**

**Data Download Date/Time** 12/31/2007 17:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 01/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 09:08:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 612501G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	No	0.20 g/210L Test #1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.000
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

A F / M A: Range Exceeded. 08: Control Outside DUE TO AIR SEAL NOT MAINTAINED .

**Data Download Date/Time** 01/19/2007 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 02/16/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 10:25:42	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A	<b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 612501G		
<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.084
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

A F / M A: Range Exceeded.

**Data Download Date/Time**    02/16/2007    10:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 03/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 11:18:18	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 612501G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

A F / M A: Improper Sample.

**Data Download Date/Time**      03/15/2007      11:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 04/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 09:55:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 612501G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      04/25/2007      11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 05/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 08:34:26	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 612501G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 05/25/2007 08:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 06/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 09:38:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 612501G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 06/26/2007 10:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 07/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 08:02:20	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 612501G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      07/27/2007      08:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 08/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 08:54:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 612501G		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 08/28/2007 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 09/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 08:58:14	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 612501G
<b>Expiration</b> 05/04/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

A F / M A: Range ExceededRange Exceeded.

**Data Download Date/Time**      09/27/2007      09:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 10/25/2007	<b>Type</b> Department Inspection
<b>Inspector</b> MALHIOT, MATTHEW, E	<b>Time</b> 09:55:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 6271011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 10/03/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0005	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.083
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.053
		0.08 g/210L Test #4	0.083
		0.20 g/210L Test #4	0.202
		0.08 g/210L Dry Gas Std Test #4	0.078
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.052
		0.08 g/210L Test #5	0.084
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.084
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.083
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.052

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

0.08 g/210L Test #8	0.084
0.20 g/210L Test #8	0.202
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.052
0.08 g/210L Test #9	0.083
0.20 g/210L Test #9	0.202
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.053
0.08 g/210L Test #10	0.083
0.20 g/210L Test #10	0.202
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 10/29/2007 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 10/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 08:58:03	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 612501G
<b>Expiration</b> 05/04/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      10/29/2007      09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 11/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 09:46:57	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 612501G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 11/20/2007 09:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002097

<b>Agency</b> NAS JACKSONVILLE	<b>Date</b> 12/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> HEMPHILL,CLOTE,	<b>Time</b> 08:39:23	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 612501G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 05/08/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      12/28/2007      08:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002170

<b>Agency</b> NAS KEY WEST	<b>Date</b> 03/08/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BLAIR,PETER,B	<b>Time</b> 14:01:25	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 0	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

**Data Download Date/Time** 03/08/2007 15:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002170

<b>Agency</b> NAS KEY WEST	<b>Date</b> 03/08/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BLAIR,PETER,B	<b>Time</b> 14:03:41	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 2	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

**Data Download Date/Time**      03/08/2007      15:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002170

<b>Agency</b> NAS KEY WEST	<b>Date</b> 03/08/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BLAIR,PETER,B	<b>Time</b> 14:11:26	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 2	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: .

**Data Download Date/Time** 03/08/2007 15:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002170

**Agency** NAS KEY WEST **Date** 03/08/2007 **Type** Agency Inspection  
**Inspector** BLAIR,PETER,B **Time** 14:33:47 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Number of Simulators** 1 **Average Standard Deviation** **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test #3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Int Det:Y, . Non-compliance: .

**Data Download Date/Time** 03/08/2007 15:35



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002170

<b>Agency</b> NAS KEY WEST	<b>Date</b> 06/15/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> BLAIR,PETER,B	<b>Time</b> 12:25:29	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

Int Det: RFI Detect DUE TO CELL PHONE.

**Data Download Date/Time** 06/19/2007 14:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 01/24/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 09:08:58	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 03/14/2007 08:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 02/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 08:04:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.083
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 03/14/2007 08:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 03/14/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 08:29:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      03/14/2007      08:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 04/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 07:53:00	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

A F / M A: Subject Test Refused.

**Data Download Date/Time**      04/19/2007      07:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 05/23/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 07:44:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      05/23/2007      07:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 06/14/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 14:17:43	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

00: Ambient Fail.

**Data Download Date/Time** 06/14/2007 14:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 07/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 08:05:45	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time**      07/12/2007      08:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 08/17/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 10:52:13	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 08/17/2007 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 09/05/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 09:56:57	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.046
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

A F / M A: Range Exceeded.

**Data Download Date/Time**      09/05/2007      10:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 10/18/2007	<b>Type</b> Department Inspection
<b>Inspector</b> GEDDINGS,MARGARET,M	<b>Time</b> 16:41:39	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 719402I		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 07/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2 (Retry)	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4 (Retry)	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.197
		0.08 g/210L Dry Gas Std Test #4	0.078
		Alcohol Free Test #5 (Retry)	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.079
		0.20 g/210L Test #5	0.198
		0.08 g/210L Dry Gas Std Test #5	0.078
		Alcohol Free Test #6 (Retry)	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.198
		0.08 g/210L Dry Gas Std Test #6	0.078
		Alcohol Free Test #7 (Retry)	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.198
		0.08 g/210L Dry Gas Std Test #7	0.078
		Alcohol Free Test #8 (Retry)	0.000
		0.05 g/210L Test #8	0.050

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.198
0.08 g/210L Dry Gas Std Test #8	0.078
Alcohol Free Test #9 (Retry)	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.198
0.08 g/210L Dry Gas Std Test #9	0.078
Alcohol Free Test #10 (Retry)	0.000
0.05 g/210L Test #10	0.050
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.198
0.08 g/210L Dry Gas Std Test #10	0.078
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

#### Remarks:

00: Ambient FailNO VENTILATION IN ROOM OPEN DOORS.

Data Download Date/Time 10/29/2007 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 10/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 10:16:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

**Data Download Date/Time** 10/29/2007 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 11/02/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 07:57:19	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1 (Retry)	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.20 g/210L Test #3 (Retry)	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

A F / M A: Sequence Aborted. 00: Ambient Fail. 20: Ambient Fail.

**Data Download Date/Time** 11/02/2007 08:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002171

<b>Agency</b> FSU POLICE	<b>Date</b> 12/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> GIANGROSSO,JOE,P	<b>Time</b> 08:12:55	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 617202I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 12/13/2007 08:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 01/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROOS,CHRISTOPHER,R	<b>Time</b> 13:07:33	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6209011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/02/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 01/22/2007 13:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 02/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROOS,CHRISTOPHER,R	<b>Time</b> 10:35:07	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6209011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/02/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

Time-Date changed.

**Data Download Date/Time**      02/28/2007      10:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 03/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROOS,CHRISTOPHER,R	<b>Time</b> 10:49:59	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6209011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/02/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 03/28/2007 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 04/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROOS,CHRISTOPHER,R	<b>Time</b> 12:28:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6209011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/02/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 04/26/2007 12:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 05/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROOS,CHRISTOPHER,R	<b>Time</b> 18:22:01	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6209011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/02/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 05/30/2007 18:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 06/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WILLIAMS,DAVID,B	<b>Time</b> 16:55:50	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 6209011		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 08/02/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2 (Retry)	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.196
		0.20 g/210L Test #3 (Retry)	0.196
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

20: RFI Detect/CELL PHONE

**Data Download Date/Time**      06/30/2007      16:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 07/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WILLIAMS,DAVID,B	<b>Time</b> 21:50:38	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 717102I		
<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.05 g/210L Test #3 (Retry)	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

05: RFI Detect/CELLULAR PHONE.

**Data Download Date/Time**      07/31/2007      21:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 08/28/2007	<b>Type</b> Department Inspection
<b>Inspector</b> SKIPPER,ROGER,G	<b>Time</b> 16:39:44	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7194021		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 07/15/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.052
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.052
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.204
		0.08 g/210L Dry Gas Std Test #3	0.080
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.204
		0.08 g/210L Dry Gas Std Test #4	0.080
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.052
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.203
		0.08 g/210L Dry Gas Std Test #5	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.052
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.204
		0.08 g/210L Dry Gas Std Test #6	0.081
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.052
		0.08 g/210L Test #7	0.082
		0.20 g/210L Test #7	0.204
		0.08 g/210L Dry Gas Std Test #7	0.081
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.052

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

0.08 g/210L Test #8	0.082
0.20 g/210L Test #8	0.204
0.08 g/210L Dry Gas Std Test #8	0.080
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.052
0.08 g/210L Test #9	0.082
0.20 g/210L Test #9	0.204
0.08 g/210L Dry Gas Std Test #9	0.080
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.053
0.08 g/210L Test #10	0.081
0.20 g/210L Test #10	0.204
0.08 g/210L Dry Gas Std Test #10	0.081
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time      08/29/2007      01:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 08/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> WILLIAMS,DAVID,B	<b>Time</b> 01:41:06	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 717102I		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1 (Retry)	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2 (Retry)	0.081
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080
		0.08 g/210L Dry Gas Std Test #3 (Retry)	0.081

**Remarks:**

08: DRY GAS NOT CONNECTED.

**Data Download Date/Time**      08/29/2007      01:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 09/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROOS,CHRISTOPHER,R	<b>Time</b> 21:42:35	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 717102I		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 09/24/2007 20:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 10/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROOS,CHRISTOPHER,R	<b>Time</b> 17:26:03	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 717102I		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 10/25/2007 17:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002231

<b>Agency</b> NEW SYMRNA BEACH PD	<b>Date</b> 11/13/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ROOS,CHRISTOPHER,R	<b>Time</b> 11:14:14	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 717102I
<b>Expiration</b> 08/01/2009 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 06/21/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.048
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.05 g/210L Test #3 (Retry)	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

05: SIMULATOR WAS NOT HOOKED UP DURING FIRST .05 TEST.

**Data Download Date/Time**      11/13/2007      11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 01/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 23:59:55	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 622802I
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 08/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time**      01/19/2007      23:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 02/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 07:38:21	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 634801G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      02/27/2007      07:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 03/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 09:15:56	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b> 09/22/2007		

**Number of Simulators**      5      **Average Standard Deviation**      **Compliance**      No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Barometric Pressure Sensor Check: OK		Alcohol Free Test #2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test #2	0.050
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #3	0.050
0.05 g/210L Test: OK	Yes		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

Non-compliance: .

**Data Download Date/Time**      03/27/2007      10:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 03/27/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 09:56:44	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 634801G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

PREVIOUS TEST MISTAKENLY TERMINATED DUE WRONG KEY STROKE

**Data Download Date/Time**      03/27/2007      10:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 04/24/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 09:29:51	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200509A <b>0.08 Lot #</b> 200509B	<b>0.20 Lot #</b> 200509C	<b>0.08 Gas Lot #</b> 634801G
<b>Expiration</b> 09/22/2007 <b>Expiration</b> 09/22/2007	<b>Expiration</b> 09/22/2007	<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      04/24/2007      10:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 05/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 07:15:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      05/30/2007      07:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 06/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 07:42:35	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      06/20/2007      08:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 07/31/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 00:10:29	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      07/31/2007      00:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 08/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 07:28:05	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.077
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

Int Det:SIMULATOR LID TIGHTENED FOR BETTER SEAL.

**Data Download Date/Time**      08/28/2007      07:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 09/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 07:53:00	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 634801G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 12/18/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.034
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.038
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.045
		0.08 g/210L Test #3 (Retry)	0.078
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

08: SIMULATOR TURNED OFF BEFORE TEST, REHEATED AND TEST RUN.

**Data Download Date/Time**      09/26/2007      08:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 10/25/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 13:29:01	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 7141011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 05/22/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0006	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.192
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.193
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.194
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.048
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.194
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.194
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.048

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.195
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.048
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.195
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.048
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.195
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0017
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

**Remarks:**

OFFICER KEVIN MILLAN WAS PRESENT FOR INSPECTION.

**Data Download Date/Time**      10/26/2007      00:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 10/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 00:05:46	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 7141011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 05/22/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.202
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time** 10/26/2007 00:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 11/29/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 08:00:36	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D	<b>0.08 Gas Lot #</b> 7141011
<b>Expiration</b> 08/01/2009 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 05/22/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

AI DIONNE PRESENT

**Data Download Date/Time**      11/29/2007      08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002232

<b>Agency</b> MIAMI BEACH PD	<b>Date</b> 12/20/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 00:59:45	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D	<b>0.08 Gas Lot #</b> 724701G
<b>Expiration</b> 08/01/2009 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 09/06/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

AI M. DIONNE PRESENT FOR INSPECTION.

**Data Download Date/Time**      12/20/2007      01:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002462

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 06/14/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 13:54:15	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b>
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>0.08 Gas Lot #</b>

**Number of Simulators** 5      **Average Standard Deviation**      **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test #1 (Retry)	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.074
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.074
0.08 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test #3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test #3	0.080
		0.08 g/210L Test #3 (Retry)	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.08 g/210L Test #4 (Retry)	0.080
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.048
		0.08 g/210L Test #5	0.079
		0.08 g/210L Test #5 (Retry)	0.075
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.073
		0.08 g/210L Test #6 (Retry)	0.080
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.073
		0.08 g/210L Test #7 (Retry)	0.075
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049
		0.08 g/210L Test #8	0.073
		0.08 g/210L Test #8 (Retry)	0.074
		Alcohol Free Test #9	0.000
		0.05 g/210L Test #9	0.049
		0.08 g/210L Test #9	0.080
		0.08 g/210L Test #9 (Retry)	0.075
		Alcohol Free Test #10	0.000

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002462

0.05 g/210L Test #10	0.049
0.08 g/210L Test #10	0.080
0.08 g/210L Test #10 (Retry)	0.081
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0033
0.08 g/210L Test (Std Dev)	0.0027

#### Remarks:

08: Control Outside Tolerance. LOOSE CONNECTION., Control O. RE RETURN TO FACTORYNon-compliance: .

**Data Download Date/Time**      10/15/2007      14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002462

<b>Agency</b> HIALEAH P. D.	<b>Date</b> 10/08/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 14:58:15	<b>Software</b> 8100.27
<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>
<b>0.08 Gas Lot #</b>		
<b>Number of Simulators</b> 0	<b>Average Standard Deviation</b>	<b>Compliance</b> No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

RETURNED FROM CMI, NO AI INSPECTIONNon-compliance: .

**Data Download Date/Time** 10/15/2007 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002462

<b>Agency</b> HIALEAH P. D.	<b>Date</b> 10/08/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 17:23:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 7096011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0007	<b>Compliance</b> No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.000
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.000
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.195
		0.08 g/210L Dry Gas Std Test #3	0.000
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.196
		0.08 g/210L Dry Gas Std Test #4	0.000
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.196
		0.08 g/210L Dry Gas Std Test #5	0.000
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.197
		0.08 g/210L Dry Gas Std Test #6	0.000
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.197
		0.08 g/210L Dry Gas Std Test #7	0.000
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.050

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002462

0.08 g/210L Test #8	0.081
0.20 g/210L Test #8	0.197
0.08 g/210L Dry Gas Std Test #8	0.000
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.051
0.08 g/210L Test #9	0.081
0.20 g/210L Test #9	0.198
0.08 g/210L Dry Gas Std Test #9	0.000
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.050
0.08 g/210L Test #10	0.081
0.20 g/210L Test #10	0.199
0.08 g/210L Dry Gas Std Test #10	0.000
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0020
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

#### Remarks:

08: DRY GAS NOT PROPERLY CONNECTED	Non-compliance:DRY GAS.
------------------------------------	-------------------------

Data Download Date/Time      10/15/2007      14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002462

<b>Agency</b> HIALEAH P. D.	<b>Date</b> 10/15/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 14:35:30	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 7096011
<b>Expiration</b> 08/01/2009 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.203
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.051
		0.08 g/210L Test #4	0.081
		0.20 g/210L Test #4	0.203
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.051
		0.08 g/210L Test #5	0.081
		0.20 g/210L Test #5	0.203
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.051
		0.08 g/210L Test #6	0.081
		0.20 g/210L Test #6	0.202
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.202
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.051

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002462

0.08 g/210L Test #8	0.082
0.20 g/210L Test #8	0.202
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.051
0.08 g/210L Test #9	0.081
0.20 g/210L Test #9	0.203
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.051
0.08 g/210L Test #10	0.082
0.20 g/210L Test #10	0.203
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

**Remarks:**

M/A NEED MORE SOLUTION.OFF A. ALVAREZ WAS PRESENT DURING IN

**Data Download Date/Time**      10/15/2007      14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002462

<b>Agency</b> HIALEAH P. D.	<b>Date</b> 12/11/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MONTALVO,M,	<b>Time</b> 10:47:17	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200708C	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 7050011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 02/22/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.077

**Remarks:**

**Data Download Date/Time** 12/11/2007 10:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002463

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 06/14/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 14:37:52	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6321011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.079
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.078
		0.20 g/210L Test #5	0.201
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.078
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.079
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.050

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002463

0.08 g/210L Test #8	0.079
0.20 g/210L Test #8	0.201
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.078
0.20 g/210L Test #9	0.201
0.08 g/210L Dry Gas Std Test #9	0.078
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.051
0.08 g/210L Test #10	0.079
0.20 g/210L Test #10	0.201
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time      06/14/2007      14:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002463

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 07/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALVAREZ,AMADO,L	<b>Time</b> 10:47:29	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 6322011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.065
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.055
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.173
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.196
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.056
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.054
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.193
		0.20 g/210L Test #2 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.054
		0.05 g/210L Test #3 (Retry)	0.053
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.20 g/210L Test #3 (Retry)	0.197
		0.08 g/210L Dry Gas Std Test #3	0.078

**Remarks:**

A F / M A : . 05: Control Outside Tolerance. 20: Control Outside Tolerance.

**Data Download Date/Time**      07/25/2007      10:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002463

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 08/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALVAREZ,AMADO,L	<b>Time</b> 13:44:36	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 6320011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

N/A

**Data Download Date/Time**      08/30/2007      13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002463

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 09/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALVAREZ,AMADO,L	<b>Time</b> 08:23:58	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b>
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>0.08 Gas Lot #</b>

**Number of Simulators** 5      **Average Standard Deviation**      **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test #2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test #2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test #2	0.066
0.05 g/210L Test: OK	Yes	Alcohol Free Test #3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test #3	0.049
0.20 g/210L Test: OK	No	0.08 g/210L Test #3	0.068

**Remarks:**

08: Control Outside Tolerance, RFI Detect. RFI FROM POLICE RADIO.

**Data Download Date/Time**      09/25/2007      10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002463

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 09/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MONTALVO,M,	<b>Time</b> 10:40:53	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 6320011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/22/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      09/25/2007      10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002463

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 10/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MONTALVO,M,	<b>Time</b> 01:03:35	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200708C	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 6320011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.082
		0.20 g/210L Test #3	0.206
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time** 10/19/2007 02:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002463

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 11/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALVAREZ,AMADO,L	<b>Time</b> 08:52:26	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200708C	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 6320011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      11/19/2007      08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002463

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 12/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALVAREZ,AMADO,L	<b>Time</b> 08:33:33	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: Range Exceeded TOO MUCH MOUTH ALCOHOL. Non-compliance:TOO MUCH MOUTH ALCOHOL, OPERATOR ERROR..

**Data Download Date/Time** 12/12/2007 08:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002463

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 12/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALVAREZ,AMADO,L	<b>Time</b> 09:15:02	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200708C	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 607401G		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 03/23/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 12/12/2007 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 06/14/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 13:39:34	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200509C
<b>0.08 Gas Lot #</b> 6321011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 09/22/2007
		<b>Expiration</b> 11/20/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0003	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.079
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.049
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.199
		0.08 g/210L Dry Gas Std Test #4	0.079
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.079
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.079
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.050
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.079
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.050

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.200
0.08 g/210L Dry Gas Std Test #8	0.079
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.050
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.200
0.08 g/210L Dry Gas Std Test #9	0.079
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.050
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.200
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time    06/14/2007    13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 07/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALVAREZ,AMADO,L	<b>Time</b> 09:39:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 6311011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/09/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.034
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.053
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.077
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.055
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.053
		0.05 g/210L Test #3 (Retry)	0.052
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

A F / M A: Range Exceeded. 05: Control Outside Tolerance.

**Data Download Date/Time**      07/25/2007      09:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 08/28/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MONTALVO,MANUEL,	<b>Time</b> 13:45:32	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 6311011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/09/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.047
		0.08 g/210L Test #3	0.077
		0.08 g/210L Test #3 (Retry)	0.079
		0.20 g/210L Test #3	0.196
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time**      08/28/2007      14:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 09/04/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 13:21:14	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A	<b>0.08 Gas Lot #</b> 7096011
<b>Expiration</b> 08/01/2009 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008	<b>Expiration</b> 04/08/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b> 0.0006	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.078
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.048
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.078
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.048
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.200
		0.08 g/210L Dry Gas Std Test #4	0.078
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.049
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.200
		0.08 g/210L Dry Gas Std Test #5	0.078
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.049
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.201
		0.08 g/210L Dry Gas Std Test #6	0.078
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.049
		0.08 g/210L Test #7	0.080
		0.20 g/210L Test #7	0.201
		0.08 g/210L Dry Gas Std Test #7	0.078
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

0.08 g/210L Test #8	0.080
0.20 g/210L Test #8	0.201
0.08 g/210L Dry Gas Std Test #8	0.078
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.049
0.08 g/210L Test #9	0.080
0.20 g/210L Test #9	0.201
0.08 g/210L Dry Gas Std Test #9	0.078
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.050
0.08 g/210L Test #10	0.080
0.20 g/210L Test #10	0.201
0.08 g/210L Dry Gas Std Test #10	0.079
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

#### Remarks:

OFFICER A. ALVAREZ WAS PRESENT DURING THE INSPECTION.

Data Download Date/Time 09/04/2007 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 09/24/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALVAREZ,AMADO,L	<b>Time</b> 15:33:20	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
<b>0.08 Gas Lot #</b>		

**Number of Simulators** 5      **Average Standard Deviation**      **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1 (Retry)	
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test #2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.192
0.20 g/210L Test: OK	No	Alcohol Free Test #3	0.000
0.08 g/210L Dry Gas Std Test: OK	No	0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.194

**Remarks:**

20: Control Outside Tolerance, RFI Detect. Non-compliance: .

**Data Download Date/Time**      09/24/2007      15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 09/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> ALVAREZ,AMADO,L	<b>Time</b> 09:19:46	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200612A
<b>0.08 Gas Lot #</b> 6311011		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/14/2008	<b>Expiration</b> 12/05/2008
		<b>Expiration</b> 11/09/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1 (Retry)	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.079
0.08 g/210L Test: OK	Yes	0.08 g/210L Test #2 (Retry)	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3 (Retry)	0.080
		0.20 g/210L Test #3	0.201
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

08: RFI Detect.

**Data Download Date/Time**      09/25/2007      09:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 10/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MONTALVO,M,	<b>Time</b> 02:40:25	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> No	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: RFI Detect. Non-compliance:OFFICER CAME IN WITH A CELL-PHONE.

**Data Download Date/Time** 10/19/2007 02:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 10/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> MONTALVO,M,	<b>Time</b> 03:12:17	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A	<b>0.08 Lot #</b> 200708C	<b>0.20 Lot #</b> 200708D
<b>0.08 Gas Lot #</b> 6311011		
<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009
		<b>Expiration</b> 11/09/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time** 10/19/2007 02:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002464

<b>Agency</b> HIALEAH P.D.	<b>Date</b> 11/19/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> RODRIGUEZ,MANUEL,	<b>Time</b> 16:14:23	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 2007708A <b>0.08 Lot #</b> 200708C	<b>0.20 Lot #</b> 200708D	<b>0.08 Gas Lot #</b> 6311011
<b>Expiration</b> 08/01/2009 <b>Expiration</b> 08/01/2009	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 11/09/2008
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.079

**Remarks:**

**Data Download Date/Time**      11/19/2007      16:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002566

<b>Agency</b> SURFSIDE P.D.	<b>Date</b> 07/18/2007	<b>Type</b> Department Inspection
<b>Inspector</b> VEIGA,SANDRA,F	<b>Time</b> 16:07:56	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 711301G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/25/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0006	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test #1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.080
		Alcohol Free Test #4	0.000
		0.05 g/210L Test #4	0.050
		0.08 g/210L Test #4	0.080
		0.20 g/210L Test #4	0.198
		0.08 g/210L Dry Gas Std Test #4	0.080
		Alcohol Free Test #5	0.000
		0.05 g/210L Test #5	0.050
		0.08 g/210L Test #5	0.080
		0.20 g/210L Test #5	0.199
		0.08 g/210L Dry Gas Std Test #5	0.080
		Alcohol Free Test #6	0.000
		0.05 g/210L Test #6	0.050
		0.08 g/210L Test #6	0.080
		0.20 g/210L Test #6	0.200
		0.08 g/210L Dry Gas Std Test #6	0.080
		Alcohol Free Test #7	0.000
		0.05 g/210L Test #7	0.051
		0.08 g/210L Test #7	0.081
		0.20 g/210L Test #7	0.200
		0.08 g/210L Dry Gas Std Test #7	0.080
		Alcohol Free Test #8	0.000
		0.05 g/210L Test #8	0.051

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002566

0.08 g/210L Test #8	0.081
0.20 g/210L Test #8	0.201
0.08 g/210L Dry Gas Std Test #8	0.080
Alcohol Free Test #9	0.000
0.05 g/210L Test #9	0.051
0.08 g/210L Test #9	0.081
0.20 g/210L Test #9	0.201
0.08 g/210L Dry Gas Std Test #9	0.080
Alcohol Free Test #10	0.000
0.05 g/210L Test #10	0.051
0.08 g/210L Test #10	0.081
0.20 g/210L Test #10	0.199
0.08 g/210L Dry Gas Std Test #10	0.080
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0012
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

**Remarks:**

OFFICER E. PINEDA WAS PRESENT DURING INSPECTION.

**Data Download Date/Time**      07/18/2007      17:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002566

<b>Agency</b> SURFSIDE P.D.	<b>Date</b> 07/18/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PINEDA, EDWIN,	<b>Time</b> 17:46:17	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 711301G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/25/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.081
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.199
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

OFC. K. VELAZQUEZ PRESENT FOR A.I.

**Data Download Date/Time**      07/18/2007      17:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002566

<b>Agency</b> SURFSIDE P.D.	<b>Date</b> 08/22/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PINEDA, EDWIN,	<b>Time</b> 11:35:16	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 711301G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/25/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.079
		0.20 g/210L Test #3	0.198
		0.08 g/210L Dry Gas Std Test #3	0.081

**Remarks:**

**Data Download Date/Time**      08/22/2007      11:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002566

<b>Agency</b> SURFSIDE P.D.	<b>Date</b> 09/26/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PINEDA, EDWIN,	<b>Time</b> 13:40:26	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200708A <b>0.08 Lot #</b> 200612D	<b>0.20 Lot #</b> 200708D	<b>0.08 Gas Lot #</b> 711301G
<b>Expiration</b> 08/01/2009 <b>Expiration</b> 12/14/2008	<b>Expiration</b> 08/01/2009	<b>Expiration</b> 04/25/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	
Date and/or Time Adjusted	No	Alcohol Free Test #1 (Retry)	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test #1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test #2 (Retry)	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test #2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3 (Retry)	0.000
		0.05 g/210L Test #3	0.049
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.197
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

00: Interferent Detect HOOKUP WRONG SIMULATOR W/ACETONE.

**Data Download Date/Time**      09/28/2007      10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002566

<b>Agency</b> SURFSIDE P.D.	<b>Date</b> 10/30/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PINEDA, EDWIN,	<b>Time</b> 13:40:06	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C <b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605C	<b>0.08 Gas Lot #</b> 711301G
<b>Expiration</b> 12/14/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/25/2009
<b>Number of Simulators</b> 4	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.040
Minimum Sample Volume Check: OK		0.05 g/210L Test #1 (Retry)	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test #1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test #1	0.188
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test #1 (Retry)	0.193
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test #1	0.080
Alcohol Free Test: OK	Yes	Alcohol Free Test #2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test #2	0.041
0.08 g/210L Test: OK	Yes	0.05 g/210L Test #2 (Retry)	0.048
0.20 g/210L Test: OK	Yes	0.08 g/210L Test #2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test #2	0.191
		0.20 g/210L Test #2 (Retry)	0.195
		0.08 g/210L Dry Gas Std Test #2	0.080
		Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.033
		0.05 g/210L Test #3 (Retry)	0.047
		0.08 g/210L Test #3	0.078
		0.20 g/210L Test #3	0.192
		0.20 g/210L Test #3 (Retry)	0.194
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

A F / M A: BLEW ALCOHOL FREE. 05: Control Outside/NEW SIM. IMPROPER ATTACHMENT. 20: Control Outside/TIGHTENED SIM. TOP.

**Data Download Date/Time**      10/30/2007      13:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002566

<b>Agency</b> SURFSIDE P.D.	<b>Date</b> 11/12/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> PINEDA, EDWIN,	<b>Time</b> 12:46:36	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200612C	<b>0.08 Lot #</b> 200605C	<b>0.20 Lot #</b> 200605F
<b>0.08 Gas Lot #</b> 711301G		
<b>Expiration</b> 12/14/2008	<b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008
		<b>Expiration</b> 04/25/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.050
		0.08 g/210L Test #3	0.080
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      11/13/2007      22:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number:80-002566

<b>Agency</b> SURFSIDE P.D.	<b>Date</b> 12/25/2007	<b>Type</b> Agency Inspection
<b>Inspector</b> CHANG,DAVID,	<b>Time</b> 15:48:47	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 200605B <b>0.08 Lot #</b> N	<b>0.20 Lot #</b> 200605F	<b>0.08 Gas Lot #</b> 711301G
<b>Expiration</b> 05/04/2008 <b>Expiration</b> 05/04/2008	<b>Expiration</b> 05/10/2008	<b>Expiration</b> 04/25/2009
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b>	<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test #1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test #1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test #1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test #1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test #1	0.079
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test #2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test #2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test #2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test #2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test #2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test #3	0.000
		0.05 g/210L Test #3	0.051
		0.08 g/210L Test #3	0.081
		0.20 g/210L Test #3	0.200
		0.08 g/210L Dry Gas Std Test #3	0.080

**Remarks:**

**Data Download Date/Time**      12/25/2007      16:25