

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

Agency: FDLE

Date/Time: 03/25/2009 08:22:32

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/06/2009 13:59

Air Blank	0.000	08:21
Control Test	0.081	08:22
Air Blank	0.000	08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

Agency: FDLE FMROC

Date/Time: 03/19/2009 15:55:54

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/11/2009 16:05

Air Blank	0.000	15:55
Control Test	0.078	15:55
Air Blank	0.000	15:55

Agency: FDLE FMROC

Date/Time: 03/26/2009 12:09:17

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/11/2009 16:05

Air Blank	0.000	12:08
Control Test	0.000	12:08
Air Blank	0.000	12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

Agency: FDLE

Date/Time: 12/20/2009 10:39:11

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/20/2009 11:36

Air Blank	0.000	10:38
Control Test	0.081	10:38
Air Blank	0.000	10:39

Agency: FDLE

Date/Time: 12/20/2009 10:41:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/20/2009 11:36

Air Blank	0.000	10:40
Control Test	0.049	10:41
Air Blank	0.000	10:41

Agency: FDLE

Date/Time: 12/20/2009 10:43:35

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/20/2009 11:36

Air Blank	0.000	10:42
Control Test	0.081	10:43
Air Blank	0.000	10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

Agency: FDLE

Date/Time: 12/20/2009 10:45:59

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/20/2009 11:36

Air Blank	0.000	10:44
Control Test	0.202	10:45
Air Blank	0.000	10:45

Florida Department of Law Enforcement Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

Agency: FT WALTON BEACH PD

Date/Time: 11/30/2009 21:57:30

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/06/2009 04:41

Air Blank	0.000	21:56
Control Test	0.071	21:56
Air Blank	0.000	21:57

Agency: FT WALTON BEACH PD

Date/Time: 11/30/2009 21:59:29

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/06/2009 04:41

Air Blank	0.000	21:58
Control Test	0.074	21:58
Air Blank	0.000	21:59

Agency: FT WALTON BEACH PD

Date/Time: 11/30/2009 22:01:49

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/06/2009 04:41

Air Blank	0.000	22:00
Control Test	0.195	22:01
Air Blank	0.000	22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

Agency: OKALOOSA COUNTY SO

Date/Time: 08/26/2009 18:31:18

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/31/2009 07:56

Air Blank	0.000	18:30
Control Test	0.081	18:30
Air Blank	0.000	18:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 14:56:52

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	14:56	
		14:56	RFI Detect
Air Blank	0.000	14:56	

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 14:58:50

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	14:57	
Control Test	0.080	14:58	
Air Blank		14:58	RFI Detect
Air Blank	0.000	14:58	

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:00:41

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	14:59	
Control Test	0.080	15:00	
Air Blank	0.000	15:00	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:03:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:02
Control Test	0.077	15:02
Air Blank	0.000	15:03

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:04:59

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:03
Control Test	0.078	15:04
Air Blank	0.000	15:04

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:06:44

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:05
Control Test	0.077	15:06
Air Blank	0.000	15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:08:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:07
Control Test	0.194	15:08
Air Blank	0.000	15:08

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:10:36

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:09
Control Test	0.197	15:10
Air Blank	0.000	15:10

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:12:29

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:11
Control Test	0.198	15:11
Air Blank	0.000	15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:14:05

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:13
Control Test	0.080	15:13
Air Blank	0.000	15:14

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:15:24

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:14
Control Test	0.079	15:14
Air Blank	0.000	15:15

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:16:44

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:15
Control Test	0.080	15:16
Air Blank	0.000	15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:18:53

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:17
Control Test	0.046	15:18
Air Blank	0.000	15:18

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:20:35

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:19
Control Test	0.048	15:20
Air Blank	0.000	15:20

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/25/2009 15:22:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/25/2009 18:41

Air Blank	0.000	15:21
Control Test	0.047	15:21
Air Blank	0.000	15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

Agency: ESCAMBIA COUNTY SO

Date/Time: 07/28/2009 13:19:27

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/28/2009 16:03

Air Blank	0.000	13:18	
Control Test	0.012	13:18	Interferent Detect
Air Blank	0.000	13:19	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

Agency: ESCAMBIA COUNTY SO

Date/Time: 06/03/2009 17:20:07

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/04/2009 09:35

Air Blank		17:19	Sequence Aborted
-----------	--	-------	------------------

Air Blank	0.000	17:20	
-----------	-------	-------	--

Agency: ESCAMBIA COUNTY SO

Date/Time: 06/18/2009 14:52:24

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/18/2009 19:33

Air Blank	0.000	14:51	
-----------	-------	-------	--

Control Test	0.033	14:51	
--------------	-------	-------	--

Air Blank	0.000	14:52	
-----------	-------	-------	--

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

Agency: JACKSON COUNTY SO

Date/Time: 10/20/2009 10:34:17

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/30/2009 14:54

Air Blank	0.000	10:33	
Control Test	0.080	10:33	Interferent Detect
Air Blank		10:34	Purge Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

Agency: SANTA ROSA COUNTY SO

Date/Time: 09/18/2009 16:26:03

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 09/29/2009 10:55

Air Blank	0.000	16:25
Control Test	0.075	16:25
Air Blank	0.000	16:26

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

Agency: HAMILTON COUNTY SO

Date/Time: 01/29/2009 02:31:09

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/29/2009 03:27

Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

Agency: NAS JACKSONVILLE PD

Date/Time: 08/04/2009 13:47:31

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/04/2009 14:35

Air Blank	0.000	13:46
Control Test	0.000	13:47
Air Blank	0.000	13:47

Agency: NAS JACKSONVILLE PD

Date/Time: 08/04/2009 13:49:49

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/04/2009 14:35

Air Blank	0.000	13:48
Control Test	0.000	13:49
Air Blank	0.000	13:49

Agency: NAS JACKSONVILLE PD

Date/Time: 08/04/2009 13:52:17

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/04/2009 14:35

Air Blank	0.000	13:51
Control Test	0.194	13:51
Air Blank	0.000	13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

Agency: NAS JACKSONVILLE PD

Date/Time: 08/04/2009 13:54:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/04/2009 14:35

Air Blank	0.000	13:53
Control Test	0.000	13:53
Air Blank	0.000	13:54

Agency: NAS JACKSONVILLE PD

Date/Time: 08/04/2009 13:57:29

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/04/2009 14:35

Air Blank	0.000	13:56	
Control Test	0.000	13:56	Interferent Detect
Air Blank	0.000	13:57	

Agency: NAS JACKSONVILLE PD

Date/Time: 09/16/2009 01:11:24

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/16/2009 01:49

Air Blank	0.000	01:10	
Control Test	0.000	01:10	Interferent Detect
Air Blank	0.000	01:11	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

Agency: NAS JACKSONVILLE PD

Date/Time: 09/16/2009 01:13:39

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/16/2009 01:49

Air Blank	0.000	01:12
Control Test	0.192	01:13
Air Blank	0.000	01:13

Agency: NAS JACKSONVILLE PD

Date/Time: 10/19/2009 01:37:02

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/19/2009 03:03

Air Blank	0.000	01:35
Control Test	0.000	01:36
Air Blank	0.000	01:37

Agency: NAS JACKSONVILLE PD

Date/Time: 10/19/2009 01:39:36

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/19/2009 03:03

Air Blank	0.000	01:38
Control Test	0.195	01:39
Air Blank	0.000	01:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

Agency: NAS JACKSONVILLE PD

Date/Time: 10/19/2009 01:45:00

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/19/2009 03:03

Air Blank	0.000	01:43
Control Test	0.000	01:44
Air Blank	0.000	01:45

Agency: NAS JACKSONVILLE PD

Date/Time: 10/19/2009 01:57:43

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/19/2009 03:03

Air Blank	0.000	01:56
Control Test	0.000	01:57
Air Blank	0.000	01:57

Agency: NAS JACKSONVILLE PD

Date/Time: 10/19/2009 02:00:03

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/19/2009 03:03

Air Blank	0.000	01:58
Control Test	0.000	01:59
Air Blank	0.000	02:00

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

Agency: NAS JACKSONVILLE PD

Date/Time: 10/19/2009 02:04:57

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/19/2009 03:03

Air Blank 0.000 02:03

Control Test 0.000 02:04

Air Blank 0.000 02:04

Agency: NAS JACKSONVILLE PD

Date/Time: 10/19/2009 02:08:56

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/19/2009 03:03

Air Blank 0.000 02:07

Control Test 0.000 02:08

Air Blank 0.000 02:08

Agency: NAS JACKSONVILLE PD

Date/Time: 10/19/2009 02:23:05

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/19/2009 03:03

Air Blank 0.000 02:21

Control Test 0.000 02:22 Interferent Detect

Air Blank 0.000 02:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

Agency: NAS JACKSONVILLE PD

Date/Time: 11/24/2009 13:47:56

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/24/2009 14:33

Air Blank	0.000	13:46	
Control Test	0.000	13:47	Interferent Detect
Air Blank	0.000	13:47	

Agency: NAS JACKSONVILLE PD

Date/Time: 11/24/2009 13:50:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/24/2009 14:33

Air Blank	0.000	13:49	
Control Test	0.190	13:49	
Air Blank	0.000	13:50	

Agency: NAS JACKSONVILLE PD

Date/Time: 11/24/2009 13:54:47

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/24/2009 14:33

Air Blank	0.000	13:53	
Control Test	0.193	13:54	
Air Blank	0.000	13:54	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

Agency: NAS JACKSONVILLE PD

Date/Time: 12/22/2009 09:07:42

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/22/2009 09:49

Air Blank	0.000	09:06	
Control Test	0.000	09:07	Interferent Detect
Air Blank	0.000	09:07	

Agency: NAS JACKSONVILLE PD

Date/Time: 12/22/2009 09:09:32

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/22/2009 09:49

Air Blank	0.000	09:08	
Control Test	0.048	09:09	
Air Blank	0.000	09:09	

Agency: NAS JACKSONVILLE PD

Date/Time: 12/22/2009 09:11:24

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/22/2009 09:49

Air Blank	0.000	09:10	
Control Test	0.193	09:10	
Air Blank	0.000	09:11	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

Agency: MADISON COUNTY SO

Date/Time: 03/17/2009 12:12:39

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/22/2009 14:58

Air Blank	0.000	12:11
Control Test	0.075	12:12
Air Blank	0.000	12:12

Agency: MADISON COUNTY SO

Date/Time: 05/27/2009 13:42:00

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/06/2009 18:10

Air Blank	0.000	13:41
Control Test	0.075	13:41
Air Blank	0.000	13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

Agency: WASHINGTON COUNTY SO

Date/Time: 07/21/2009 13:05:55

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/21/2009 18:23

Air Blank	0.000	13:05	
Control Test	0.078	13:05	Interferent Detect
Air Blank	0.000	13:05	

Agency: WASHINGTON COUNTY SO

Date/Time: 07/21/2009 13:13:45

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/21/2009 18:23

Air Blank	0.000	13:12	
Control Test	0.079	13:13	Interferent Detect
Air Blank	0.000	13:13	

Agency: WASHINGTON COUNTY SO

Date/Time: 07/21/2009 13:15:06

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/21/2009 18:23

Air Blank	0.000	13:14	
Control Test	0.078	13:14	Interferent Detect
Air Blank	0.000	13:15	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

Agency: WASHINGTON COUNTY SO

Date/Time: 07/21/2009 13:25:54

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/21/2009 18:23

Air Blank	0.000	13:25	
Control Test	0.081	13:25	Interferent Detect
Air Blank	0.000	13:25	

Agency: WASHINGTON COUNTY SO

Date/Time: 07/21/2009 13:34:36

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/21/2009 18:23

Air Blank	0.000	13:33	
Control Test	0.079	13:34	Interferent Detect
Air Blank	0.000	13:34	

Agency: WASHINGTON COUNTY SO

Date/Time: 07/21/2009 14:14:11

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/21/2009 18:23

Air Blank	0.000	14:13	
Control Test	0.080	14:13	
Air Blank	0.000	14:14	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

Agency: HOLMES COUNTY SO

Date/Time: 07/28/2009 09:08:18

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/28/2009 12:13

Air Blank	0.000	09:07	
Control Test	0.009	09:07	Interferent Detect
Air Blank	0.000	09:08	

Agency: HOLMES COUNTY SO

Date/Time: 07/28/2009 10:18:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/28/2009 12:13

Air Blank	0.000	10:17	
Control Test	0.008	10:17	Interferent Detect
Air Blank	0.000	10:18	

Agency: HOLMES COUNTY SO

Date/Time: 07/28/2009 10:21:01

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/28/2009 12:13

Air Blank	0.000	10:19	
Control Test	0.045	10:20	
Air Blank	0.000	10:21	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

Agency: HOLMES COUNTY SO

Date/Time: 07/28/2009 10:22:43

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/28/2009 12:13

Air Blank	0.000	10:21
Control Test	0.045	10:22
Air Blank	0.000	10:22

Agency: HOLMES COUNTY SO

Date/Time: 07/28/2009 10:29:01

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/28/2009 12:13

Air Blank	0.000	10:27
Control Test	0.046	10:28
Air Blank	0.000	10:29

Agency: HOLMES COUNTY SO

Date/Time: 07/28/2009 10:31:04

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/28/2009 12:13

Air Blank	0.000	10:29
Control Test	0.077	10:30
Air Blank	0.000	10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

Agency: HOLMES COUNTY SO

Date/Time: 07/28/2009 09:11:08

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/28/2009 12:10

Air Blank	0.000	09:10	
Control Test	0.009	09:10	RFI Detect
Air Blank	0.000	09:11	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

Agency: ST LUCIE COUNTY SO

Date/Time: 06/27/2009 23:58:43

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/31/2009 22:46

Air Blank	0.000	23:57
Control Test	0.069	23:58
Air Blank	0.000	23:58

Agency: ST LUCIE COUNTY SO

Date/Time: 06/28/2009 00:06:13

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/31/2009 22:46

Air Blank	0.000	00:05	
Control Test	0.074	00:05	Interferent Detect
Air Blank	0.000	00:06	

Agency: ST LUCIE COUNTY SO

Date/Time: 06/28/2009 00:07:34

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/31/2009 22:46

Air Blank	0.000	00:06
Control Test	0.075	00:07
Air Blank	0.000	00:07

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

Agency: ST LUCIE COUNTY SO

Date/Time: 06/28/2009 00:09:37

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/31/2009 22:46

Air Blank		00:09	RFI Detect
Air Blank	0.000	00:09	

Agency: ST LUCIE COUNTY SO

Date/Time: 06/28/2009 00:10:57

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/31/2009 22:46

Air Blank	0.000	00:10	
Control Test	0.075	00:10	
Air Blank	0.000	00:10	

Agency: ST LUCIE COUNTY SO

Date/Time: 06/28/2009 00:22:34

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/31/2009 22:46

Air Blank	0.000	00:21	
Control Test	0.079	00:22	
Air Blank	0.000	00:22	

Agency: ST LUCIE COUNTY SO

Date/Time: 12/31/2009 19:45:44

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/30/2010 21:01

Air Blank	0.000	19:44	
Control Test	0.079	19:45	
Air Blank	0.000	19:45	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 01/30/2009 14:24:48

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/30/2009 15:16

Air Blank	0.000	14:23
Control Test	0.077	14:24
Air Blank	0.000	14:24

Agency: VERO BEACH PD

Date/Time: 01/30/2009 14:26:49

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/30/2009 15:16

Air Blank	0.000	14:25
Control Test	0.000	14:26
Air Blank	0.000	14:26

Agency: VERO BEACH PD

Date/Time: 01/30/2009 14:28:30

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/30/2009 15:16

Air Blank	0.000	14:27
Control Test	0.189	14:28
Air Blank	0.000	14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 01/30/2009 14:29:28

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/30/2009 15:16

Air Blank		14:28	Ambient Fail
Air Blank	0.000	14:29	

Agency: VERO BEACH PD

Date/Time: 01/30/2009 14:30:26

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/30/2009 15:16

Air Blank		14:29	Ambient Fail
Air Blank	0.000	14:30	

Agency: VERO BEACH PD

Date/Time: 01/30/2009 14:32:16

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/30/2009 15:16

Air Blank	0.000	14:31	
Control Test	0.199	14:31	
Air Blank	0.000	14:32	

Agency: VERO BEACH PD

Date/Time: 01/30/2009 14:34:06

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/30/2009 15:16

Air Blank	0.000	14:32	
Control Test	0.194	14:33	
Air Blank	0.000	14:34	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 02/20/2009 14:52:52

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:51
ITP Check	0.079	14:52
Air Blank	0.000	14:52

Agency: VERO BEACH PD

Date/Time: 02/20/2009 14:54:18

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:53
Control Test	0.080	14:53
Air Blank	0.000	14:54

Agency: VERO BEACH PD

Date/Time: 02/28/2009 10:39:57

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	10:38
Control Test	0.079	10:39
Air Blank	0.000	10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 02/28/2009 10:41:59

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	10:40
Control Test	0.199	10:41
Air Blank	0.000	10:41

Agency: VERO BEACH PD

Date/Time: 03/05/2009 14:10:39

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:09
Control Test	0.159	14:10
Air Blank	0.000	14:10

Agency: VERO BEACH PD

Date/Time: 03/05/2009 14:12:48

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:11
Control Test	0.173	14:12
Air Blank	0.000	14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 03/05/2009 14:15:14

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:13	
Control Test	0.177	14:14	
Air Blank		14:14	RFI Detect
Air Blank	0.000	14:15	

Agency: VERO BEACH PD

Date/Time: 03/05/2009 14:18:05

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:16	
Control Test	0.183	14:17	
Air Blank	0.000	14:18	

Agency: VERO BEACH PD

Date/Time: 03/05/2009 14:21:54

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:20	
Control Test	0.189	14:21	
Air Blank	0.000	14:21	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 03/05/2009 14:26:01

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:24
Control Test	0.191	14:25
Air Blank	0.000	14:26

Agency: VERO BEACH PD

Date/Time: 03/05/2009 14:27:52

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:26
Control Test	0.078	14:27
Air Blank	0.000	14:27

Agency: VERO BEACH PD

Date/Time: 03/05/2009 14:30:04

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:28
Control Test	0.048	14:29
Air Blank	0.000	14:30

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 03/05/2009 14:32:00

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/05/2009 15:15

Air Blank	0.000	14:30	
Control Test	0.007	14:31	Interferent Detect
Air Blank	0.000	14:32	

Agency: VERO BEACH PD

Date/Time: 04/30/2009 16:31:29

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/30/2009 17:05

Air Blank	0.000	16:30	
Control Test	0.193	16:30	
Air Blank	0.000	16:31	

Agency: VERO BEACH PD

Date/Time: 04/30/2009 16:33:16

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/30/2009 17:05

Air Blank	0.000	16:32	
Control Test	0.076	16:32	
Air Blank	0.000	16:33	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 05/28/2009 11:30:11

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/28/2009 12:09

Air Blank	0.000	11:28	
Control Test	0.000	11:29	Interferent Detect
Air Blank	0.000	11:30	

Agency: VERO BEACH PD

Date/Time: 05/28/2009 11:32:28

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/28/2009 12:09

Air Blank	0.000	11:31	
Control Test	0.049	11:31	
Air Blank	0.000	11:32	

Agency: VERO BEACH PD

Date/Time: 05/28/2009 11:34:55

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/28/2009 12:09

Air Blank	0.000	11:33	
Control Test	0.077	11:34	
Air Blank	0.000	11:34	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 05/28/2009 11:36:36

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/28/2009 12:09

Air Blank	0.000	11:35
Control Test	0.197	11:36
Air Blank	0.000	11:36

Agency: VERO BEACH PD

Date/Time: 06/07/2009 12:15:14

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/07/2009 13:04

Air Blank	0.000	12:14
Control Test	0.079	12:14
Air Blank	0.000	12:15

Agency: VERO BEACH PD

Date/Time: 06/07/2009 12:16:58

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/07/2009 13:04

Air Blank	0.000	12:16
Control Test	0.079	12:16
Air Blank	0.000	12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 06/07/2009 12:18:49

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/07/2009 13:04

Air Blank	0.000	12:17
Control Test	0.000	12:18
Air Blank	0.000	12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

Agency: HILLSBOROUGH CO SO

Date/Time: 04/27/2009 13:13:57

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/01/2009 09:49

Air Blank	0.000	13:13
Control Test	0.078	13:13
Air Blank	0.000	13:13

Agency: HILLSBOROUGH CO SO

Date/Time: 04/27/2009 13:16:03

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/01/2009 09:49

Air Blank	0.000	13:15
Control Test	0.079	13:15
Air Blank	0.000	13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

Agency: MARION COUNTY SO

Date/Time: 07/13/2009 01:36:53

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/19/2009 08:07

Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36

Agency: MARION COUNTY SO

Date/Time: 12/13/2009 05:25:22

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/17/2009 09:07

Air Blank	0.000	05:24
Control Test	0.078	05:24
Air Blank	0.000	05:25

Agency: MARION COUNTY SO

Date/Time: 12/13/2009 05:26:48

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/17/2009 09:07

Air Blank	0.000	05:25
Control Test	0.078	05:26
Air Blank	0.000	05:26

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 01/31/2009 12:21:48

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 01/31/2009 14:07

Air Blank	0.000	12:20
ITP Check	0.079	12:21
Air Blank	0.000	12:21

Agency: CITRUS COUNTY SO

Date/Time: 01/31/2009 12:25:52

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/31/2009 14:07

Air Blank	0.000	12:25
Control Test	0.076	12:25
Air Blank	0.000	12:25

Agency: CITRUS COUNTY SO

Date/Time: 01/31/2009 12:30:26

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/31/2009 14:07

Air Blank	0.000	12:29
Control Test	0.079	12:29
Air Blank	0.000	12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 01/31/2009 12:32:30

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/31/2009 14:07

Air Blank	0.000	12:31
Control Test	0.198	12:31
Air Blank	0.000	12:32

Agency: CITRUS COUNTY SO

Date/Time: 01/31/2009 12:37:53

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/31/2009 14:07

Air Blank	0.000	12:36	
Control Test	0.007	12:37	Interferent Detect
Air Blank	0.000	12:37	

Agency: CITRUS COUNTY SO

Date/Time: 01/31/2009 12:39:46

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/31/2009 14:07

Air Blank	0.000	12:38
Control Test	0.050	12:39
Air Blank	0.000	12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 01/31/2009 12:45:44

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/31/2009 14:07

Air Blank 0.000 12:44

Control Test 0.000 12:45

Air Blank 12:45 Sequence Aborted

Air Blank 0.000 12:45

Agency: CITRUS COUNTY SO

Date/Time: 01/31/2009 12:47:24

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/31/2009 14:07

Air Blank 0.000 12:46

Control Test 0.000 12:46

Air Blank 0.000 12:47

Agency: CITRUS COUNTY SO

Date/Time: 02/18/2009 15:06:43

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/18/2009 15:45

Air Blank 15:06 Sequence Aborted

Air Blank 0.000 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 03/16/2009 18:17:19

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/09/2009 08:48

Air Blank		18:16	Ambient Fail
Air Blank	0.000	18:17	

Agency: CITRUS COUNTY SO

Date/Time: 03/16/2009 18:20:37

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/09/2009 08:48

Air Blank	0.000	18:19	
Control Test	0.021	18:20	
Air Blank	0.000	18:20	

Agency: CITRUS COUNTY SO

Date/Time: 03/16/2009 18:22:46

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/09/2009 08:48

Air Blank	0.000	18:21	
Control Test	0.000	18:22	
Air Blank	0.000	18:22	

Agency: CITRUS COUNTY SO

Date/Time: 04/20/2009 10:35:56

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	10:35	
Control Test	0.076	10:35	
Air Blank	0.000	10:35	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 04/20/2009 10:37:38

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	10:36
Control Test	0.076	10:37
Air Blank	0.000	10:37

Agency: CITRUS COUNTY SO

Date/Time: 04/20/2009 10:42:07

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	10:41
ITP Check	0.078	10:41
Air Blank	0.000	10:42

Agency: CITRUS COUNTY SO

Date/Time: 04/20/2009 11:05:49

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	11:04
Control Test	0.195	11:05
Air Blank	0.000	11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 04/20/2009 11:10:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	11:08
Control Test	0.198	11:09
Air Blank	0.000	11:10

Agency: CITRUS COUNTY SO

Date/Time: 04/22/2009 15:01:16

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	15:00
ITP Check	0.078	15:00
Air Blank	0.000	15:01

Agency: CITRUS COUNTY SO

Date/Time: 04/22/2009 15:04:20

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	15:03
Control Test	0.049	15:03
Air Blank	0.000	15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 04/23/2009 14:46:21

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	14:45
ITP Check	0.079	14:45
Air Blank	0.000	14:46

Agency: CITRUS COUNTY SO

Date/Time: 04/23/2009 14:47:53

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	14:47
Control Test	0.076	14:47
Air Blank	0.000	14:47

Agency: CITRUS COUNTY SO

Date/Time: 04/23/2009 14:50:10

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	14:48
Control Test	0.074	14:49
Air Blank	0.000	14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 04/23/2009 14:51:54

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	14:50
Control Test	0.075	14:51
Air Blank	0.000	14:51

Agency: CITRUS COUNTY SO

Date/Time: 04/23/2009 14:55:41

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	14:54
Control Test	0.000	14:55
Air Blank	0.000	14:55

Agency: CITRUS COUNTY SO

Date/Time: 04/23/2009 14:57:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	14:56
Control Test	0.048	14:57
Air Blank	0.000	14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 04/23/2009 14:59:58

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	14:58
Control Test	0.200	14:59
Air Blank	0.000	14:59

Agency: CITRUS COUNTY SO

Date/Time: 04/23/2009 16:19:44

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 17:02

Air Blank	0.000	16:18
Control Test	0.082	16:19
Air Blank	0.000	16:19

Agency: CITRUS COUNTY SO

Date/Time: 06/25/2009 15:11:09

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/25/2009 16:13

Air Blank	0.000	15:10
ITP Check	0.077	15:10
Air Blank	0.000	15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 06/25/2009 15:12:42

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/25/2009 16:13

Air Blank	0.000	15:11
ITP Check	0.078	15:12
Air Blank	0.000	15:12

Agency: CITRUS COUNTY SO

Date/Time: 06/25/2009 15:16:36

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/25/2009 16:13

Air Blank	0.000	15:15
Control Test	0.083	15:16
Air Blank	0.000	15:16

Agency: CITRUS COUNTY SO

Date/Time: 06/25/2009 15:19:03

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/25/2009 16:13

Air Blank	0.000	15:17
Control Test	0.082	15:18
Air Blank	0.000	15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 06/25/2009 15:21:17

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/25/2009 16:13

Air Blank	0.000	15:20
Control Test	0.049	15:20
Air Blank	0.000	15:21

Agency: CITRUS COUNTY SO

Date/Time: 06/25/2009 15:23:33

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/25/2009 16:13

Air Blank	0.000	15:22
Control Test	0.198	15:22
Air Blank	0.000	15:23

Agency: CITRUS COUNTY SO

Date/Time: 10/23/2009 16:05:40

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/23/2009 17:38

Air Blank	0.000	16:04
Control Test	0.082	16:04
Air Blank	0.000	16:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 10/23/2009 16:07:16

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 10/23/2009 17:38

Air Blank	0.000	16:06
ITP Check	0.077	16:06
Air Blank	0.000	16:07

Agency: CITRUS COUNTY SO

Date/Time: 10/23/2009 16:20:02

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/23/2009 17:38

Air Blank	0.000	16:18
Control Test	0.079	16:19
Air Blank	0.000	16:20

Agency: CITRUS COUNTY SO

Date/Time: 11/30/2009 11:00:17

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/30/2009 11:50

Air Blank	0.000	10:59
Control Test	0.081	10:59
Air Blank	0.000	11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 11/30/2009 11:03:39

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/30/2009 11:50

Air Blank	0.000	11:02
Control Test	0.079	11:02
Air Blank	0.000	11:03

Agency: CITRUS COUNTY SO

Date/Time: 11/30/2009 11:06:40

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/30/2009 11:50

Air Blank	0.000	11:05
Control Test	0.199	11:05
Air Blank	0.000	11:06

Agency: CITRUS COUNTY SO

Date/Time: 11/30/2009 11:09:23

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/30/2009 11:50

Air Blank	0.000	11:07
Control Test	0.050	11:08
Air Blank	0.000	11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 11/30/2009 11:13:34

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 11/30/2009 11:50

Air Blank	0.000	11:12
ITP Check	0.077	11:13
Air Blank	0.000	11:13

Agency: CITRUS COUNTY SO

Date/Time: 12/29/2009 17:22:32

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 12/29/2009 18:07

Air Blank	0.000	17:21
ITP Check	0.077	17:22
Air Blank	0.000	17:22

Agency: CITRUS COUNTY SO

Date/Time: 12/29/2009 17:26:58

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/29/2009 18:07

Air Blank	0.000	17:25
Control Test	0.082	17:26
Air Blank	0.000	17:26

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 12/29/2009 17:29:03

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/29/2009 18:07

Air Blank	0.000	17:27
Control Test	0.079	17:28
Air Blank	0.000	17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency: MARION COUNTY SO

Date/Time: 07/13/2009 01:38:43

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/19/2009 08:10

Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000829

Agency: HILLSBOROUGH CO S0

Date/Time: 10/22/2009 10:33:17

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/22/2009 11:26

Air Blank	0.000	10:32
Control Test	0.191	10:32
Air Blank	0.000	10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency: HILLSBOROUGH CO SO

Date/Time: 11/09/2009 08:51:07

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/24/2009 10:22

Air Blank	0.000	08:50
Control Test	0.082	08:50
Air Blank	0.000	08:51

Agency: HILLSBOROUGH CO SO

Date/Time: 11/24/2009 09:28:08

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/24/2009 10:22

Air Blank	0.000	09:26
Control Test	0.000	09:27
Air Blank	0.000	09:28

Agency: HILLSBOROUGH CO SO

Date/Time: 11/24/2009 09:30:32

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/24/2009 10:22

Air Blank	0.000	09:29
Control Test	0.202	09:30
Air Blank	0.000	09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency: LYNN HAVEN PD

Date/Time: 04/11/2009 07:34:54

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	07:34
Control Test	0.074	07:34
Air Blank	0.000	07:34

Agency: LYNN HAVEN PD

Date/Time: 04/11/2009 07:36:19

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	07:35
Control Test	0.074	07:35
Air Blank	0.000	07:36

Agency: LYNN HAVEN PD

Date/Time: 04/11/2009 07:39:09

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	07:38
Control Test	0.075	07:38
Air Blank	0.000	07:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency: LYNN HAVEN PD

Date/Time: 04/11/2009 07:40:28

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	07:39
Control Test	0.074	07:39
Air Blank	0.000	07:40

Agency: LYNN HAVEN PD

Date/Time: 04/11/2009 08:20:51

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	08:20
Control Test	0.076	08:20
Air Blank	0.000	08:20

Agency: LYNN HAVEN PD

Date/Time: 04/11/2009 08:22:11

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	08:21
Control Test	0.076	08:21
Air Blank	0.000	08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency: LYNN HAVEN PD

Date/Time: 04/11/2009 08:23:43

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	08:22
Control Test	0.076	08:23
Air Blank	0.000	08:23

Agency: LYNN HAVEN PD

Date/Time: 04/11/2009 08:25:42

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	08:24
Control Test	0.077	08:25
Air Blank	0.000	08:25

Agency: LYNN HAVEN PD

Date/Time: 04/20/2009 06:51:43

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	06:50
Control Test	0.077	06:51
Air Blank	0.000	06:51

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency: LYNN HAVEN PD

Date/Time: 04/20/2009 06:53:01

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	06:52
Control Test	0.077	06:52
Air Blank	0.000	06:53

Agency: LYNN HAVEN PD

Date/Time: 04/20/2009 06:55:21

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/29/2009 11:22

Air Blank	0.000	06:54
Control Test	0.078	06:54
Air Blank	0.000	06:55

Agency: LYNN HAVEN PD

Date/Time: 06/24/2009 23:57:01

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/25/2009 02:05

Air Blank	0.000	23:55
Control Test	0.192	23:56
Air Blank	0.000	23:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency: LYNN HAVEN PD

Date/Time: 06/24/2009 23:58:44

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/25/2009 02:05

Air Blank	0.000	23:57
-----------	-------	-------

Control Test	0.194	23:58
--------------	-------	-------

Air Blank	0.000	23:58
-----------	-------	-------

Agency: LYNN HAVEN PD

Date/Time: 06/25/2009 00:06:07

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/25/2009 02:05

Air Blank	0.000	00:04
-----------	-------	-------

Control Test	0.073	00:05
--------------	-------	-------

Air Blank	0.000	00:06
-----------	-------	-------

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Agency: CAPE CORAL PD

Date/Time: 05/07/2009 14:37:23

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 05/12/2009 09:46

Air Blank 0.000 14:36

ITP Check 0.080 14:36

Air Blank 0.000 14:37

Agency: CAPE CORAL PD

Date/Time: 06/23/2009 07:23:52

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 07/03/2009 13:02

Air Blank 0.000 07:22

ITP Check 0.080 07:23

Air Blank 0.000 07:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency: CORAL GABLES PD

Date/Time: 11/27/2009 10:37:46

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/27/2009 11:21

Air Blank	0.000	10:36
Control Test	0.081	10:37
Air Blank	0.000	10:37

Agency: CORAL GABLES PD

Date/Time: 11/27/2009 10:39:32

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/27/2009 11:21

Air Blank	0.000	10:38
Control Test	0.081	10:39
Air Blank	0.000	10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

Agency: PINELLAS COUNTY SO

Date/Time: 09/24/2009 01:52:02

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 09/24/2009 22:49

Air Blank	0.000	01:51
ITP Check	0.078	01:51
Air Blank	0.000	01:52

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency: FFWCC TAMPA

Date/Time: 02/06/2009 10:30:22

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/06/2009 18:42

Air Blank	0.000	10:29
Control Test	0.080	10:29
Air Blank	0.000	10:30

Agency: FFWCC TAMPA

Date/Time: 02/06/2009 10:32:43

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/06/2009 18:42

Air Blank	0.000	10:31
Control Test	0.081	10:32
Air Blank	0.000	10:32

Agency: FFWCC TAMPA

Date/Time: 02/06/2009 10:35:19

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/06/2009 18:42

Air Blank	0.000	10:34
Control Test	0.081	10:34
Air Blank	0.000	10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency: FFWCC TAMPA

Date/Time: 02/07/2009 08:54:02

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/06/2009 18:42

Air Blank	0.000	08:53
Control Test	0.079	08:53
Air Blank	0.000	08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

Agency: PINECREST PD

Date/Time: 10/22/2009 15:01:27

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/22/2009 16:01

Air Blank		15:00	Sequence Aborted
-----------	--	-------	------------------

Air Blank	0.000	15:01	
-----------	-------	-------	--

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000913

Agency: MANATEE COUNTY SO

Date/Time: 06/27/2009 03:29:52

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/08/2009 11:00

Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:29

Agency: MANATEE COUNTY SO

Date/Time: 08/26/2009 10:32:44

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/26/2009 11:45

Air Blank	0.000	10:31
Control Test	0.000	10:32
Air Blank	0.000	10:32

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

Agency: FORT MYERS PD

Date/Time: 06/19/2009 22:56:21

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 07/03/2009 10:38

Air Blank	0.000	22:55
-----------	-------	-------

ITP Check	0.079	22:55
-----------	-------	-------

Air Blank	0.000	22:56
-----------	-------	-------

Agency: FORT MYERS PD

Date/Time: 06/19/2009 22:58:31

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 07/03/2009 10:38

Air Blank	0.000	22:57
-----------	-------	-------

ITP Check	0.078	22:58
-----------	-------	-------

Air Blank	0.000	22:58
-----------	-------	-------

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 20:22:19

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/08/2009 00:03

Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 21:39:36

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/08/2009 00:03

Air Blank	0.000	21:38
Control Test	0.000	21:39
Air Blank	0.000	21:39

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 21:41:43

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/08/2009 00:03

Air Blank	0.000	21:40
ITP Check	0.079	21:41
Air Blank	0.000	21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 21:43:46

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/08/2009 00:03

Air Blank	0.000	21:42
Control Test	0.202	21:43
Air Blank	0.000	21:43

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 21:48:27

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/08/2009 00:03

Air Blank	0.000	21:47
Control Test	0.000	21:47
Air Blank	0.000	21:48

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 21:51:10

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/08/2009 00:03

Air Blank	0.000	21:50
ITP Check	0.078	21:50
Air Blank	0.000	21:51

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 22:34:08

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/08/2009 00:03

Air Blank	0.000	22:32
Control Test	0.000	22:33
Air Blank	0.000	22:34

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 22:36:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/08/2009 00:03

Air Blank	0.000	22:35
Control Test	0.000	22:36
Air Blank	0.000	22:36

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 22:46:05

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/08/2009 00:03

Air Blank	0.000	22:44	
Control Test	0.000	22:45	Interferent Detect
Air Blank	0.000	22:46	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 22:47:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/08/2009 00:03

Air Blank	0.000	22:46
Control Test	0.000	22:47
Air Blank	0.000	22:47

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

Agency: COLLIER COUNTY SO

Date/Time: 04/01/2009 08:39:06

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/02/2009 12:40

Air Blank	0.000	08:37
Control Test	0.000	08:38
Air Blank	0.000	08:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

Agency: COLLIER COUNTY SO

Date/Time: 07/01/2009 08:26:14

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/01/2009 09:18

Air Blank	0.000	08:25
Control Test	0.203	08:25
Air Blank	0.000	08:26

Agency: COLLIER COUNTY SO

Date/Time: 07/01/2009 08:30:04

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/01/2009 09:18

Air Blank	0.000	08:28	
Control Test	0.000	08:29	Interferent Detect
Air Blank	0.000	08:30	

Agency: COLLIER COUNTY SO

Date/Time: 08/05/2009 07:42:03

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/05/2009 17:20

Air Blank	0.000	07:40
Control Test	0.000	07:41
Air Blank	0.000	07:42

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

Agency: COLLIER COUNTY SO

Date/Time: 08/05/2009 07:44:48

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/05/2009 17:20

Air Blank	0.000	07:43
Control Test	0.046	07:44
Air Blank	0.000	07:44

Agency: COLLIER COUNTY SO

Date/Time: 08/05/2009 08:06:34

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/05/2009 17:20

Air Blank	0.000	08:05
Control Test	0.000	08:06
Air Blank	0.000	08:06

Agency: COLLIER COUNTY SO

Date/Time: 08/05/2009 08:21:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/05/2009 17:20

Air Blank	0.000	08:20	
Control Test	0.000	08:21	Interferent Detect
Air Blank	0.000	08:21	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

Agency: COLLIER COUNTY SO

Date/Time: 10/01/2009 07:38:09

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/01/2009 08:28

Air Blank	0.000	07:36	
Control Test	0.000	07:37	Interferent Detect
Air Blank	0.000	07:38	

Agency: COLLIER COUNTY SO

Date/Time: 10/01/2009 07:41:23

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/01/2009 08:28

Air Blank	0.000	07:40	
Control Test	0.047	07:40	
Air Blank	0.000	07:41	

Agency: COLLIER COUNTY SO

Date/Time: 12/01/2009 07:58:33

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/01/2009 08:41

Air Blank	0.000	07:57	
Control Test	0.000	07:58	Interferent Detect
Air Blank	0.000	07:58	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

Agency: CHARLOTTE COUNTY SO

Date/Time: 01/14/2009 07:20:07

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/14/2009 08:33

Air Blank	0.000	07:18
Control Test	0.047	07:19
Air Blank	0.000	07:20

Agency: CHARLOTTE COUNTY SO

Date/Time: 02/08/2009 08:43:42

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/08/2009 09:50

Air Blank	0.000	08:42	
Control Test	0.000	08:43	Interferent Detect
Air Blank	0.000	08:43	

Agency: CHARLOTTE COUNTY SO

Date/Time: 02/08/2009 08:46:46

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/08/2009 09:50

Air Blank	0.000	08:45
Control Test	0.050	08:46
Air Blank	0.000	08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

Agency: CHARLOTTE COUNTY SO

Date/Time: 02/08/2009 08:49:04

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/08/2009 09:50

Air Blank	0.000	08:47
Control Test	0.078	08:48
Air Blank	0.000	08:49

Agency: CHARLOTTE COUNTY SO

Date/Time: 02/08/2009 08:51:18

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/08/2009 09:50

Air Blank	0.000	08:50
Control Test	0.195	08:50
Air Blank	0.000	08:51

Agency: CHARLOTTE COUNTY SO

Date/Time: 02/18/2009 11:22:37

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/11/2009 10:49

Air Blank	0.000	11:21
Control Test	0.081	11:22
Air Blank	0.000	11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

Agency: CHARLOTTE COUNTY SO

Date/Time: 09/13/2009 15:13:22

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/13/2009 16:24

Air Blank	0.000	15:12
Control Test	0.000	15:12
Air Blank	0.000	15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

Agency: CHARLOTTE COUNTY SO

Date/Time: 12/16/2009 11:31:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/16/2009 13:53

Air Blank	0.000	11:30
Control Test	0.203	11:30
Air Blank	0.000	11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

Agency: GLADES COUNTY S.O.

Date/Time: 01/26/2009 08:34:50

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 17:38

Air Blank	0.000	08:34
Control Test	0.086	08:34
Air Blank	0.000	08:34

Agency: GLADES COUNTY S.O.

Date/Time: 01/26/2009 08:37:13

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 17:38

Air Blank	0.000	08:36
Control Test	0.086	08:36
Air Blank	0.000	08:37

Agency: GLADES COUNTY S.O.

Date/Time: 01/26/2009 08:38:37

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 17:38

Air Blank	0.000	08:37
Control Test	0.086	08:38
Air Blank	0.000	08:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

Agency: GLADES COUNTY S.O.

Date/Time: 01/26/2009 08:42:41

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 17:38

Air Blank	0.000	08:41
Control Test	0.086	08:42
Air Blank	0.000	08:42

Agency: GLADES COUNTY S.O.

Date/Time: 01/26/2009 08:44:51

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 17:38

Air Blank	0.000	08:43
Control Test	0.086	08:44
Air Blank	0.000	08:44

Agency: GLADES COUNTY S.O.

Date/Time: 01/26/2009 08:47:17

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 17:38

Air Blank	0.000	08:46
Control Test	0.085	08:46
Air Blank	0.000	08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

Agency: GLADES COUNTY S.O.

Date/Time: 01/26/2009 08:33:56

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 16:52

Air Blank	0.000	08:33
Control Test	0.082	08:33
Air Blank	0.000	08:33

Agency: GLADES COUNTY S.O.

Date/Time: 01/26/2009 08:35:35

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 16:52

Air Blank	0.000	08:34
Control Test	0.080	08:35
Air Blank	0.000	08:35

Agency: GLADES COUNTY S.O.

Date/Time: 01/26/2009 08:41:41

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 16:52

Air Blank	0.000	08:40
Control Test	0.074	08:41
Air Blank	0.000	08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

Agency: FRANKLIN COUNTY SO

Date/Time: 05/28/2009 13:54:04

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/29/2009 09:32

Air Blank	0.000	13:53
Control Test	0.085	13:53
Air Blank	0.000	13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

Agency: LIBERTY COUNTY SO

Date/Time: 09/09/2009 13:20:53

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/09/2009 14:05

Air Blank	0.000	13:19
Control Test	0.047	13:20
Air Blank	0.000	13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

Agency: WAKULLA COUNTY SO

Date/Time: 06/22/2009 11:20:35

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/24/2009 07:27

Air Blank	0.000	11:19
Control Test	0.067	11:20
Air Blank	0.000	11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

Agency: WAKULLA COUNTY SO

Date/Time: 12/09/2009 10:42:53

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/01/2010 13:14

Air Blank	0.000	10:42
Control Test	0.077	10:42
Air Blank	0.000	10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

Agency: QUINCY PD

Date/Time: 04/08/2009 10:36:29

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/07/2009 11:43

Air Blank	0.000	10:35
Control Test	0.076	10:35
Air Blank	0.000	10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

Agency: WINTER PARK P.D.

Date/Time: 01/15/2009 16:49:43

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/28/2009 16:05

Air Blank	0.000	16:48
Control Test	0.079	16:49
Air Blank	0.000	16:49

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 07/31/2009 14:10:40

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/31/2009 14:59

Air Blank	0.000	14:09
Control Test	0.000	14:10
Air Blank	0.000	14:10

Agency: TITUSVILLE P.D.

Date/Time: 08/31/2009 19:43:56

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 08/31/2009 20:39

Air Blank	0.000	19:43
ITP Check	0.079	19:43
Air Blank	0.000	19:43

Agency: TITUSVILLE P.D.

Date/Time: 09/30/2009 16:37:22

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/30/2009 17:37

Air Blank	0.000	16:36	
Control Test	0.000	16:36	Interferent Detect
Air Blank	0.000	16:37	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 10/31/2009 12:42:34

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/31/2009 12:36

Air Blank	0.000	12:41
Control Test	0.000	12:42
Air Blank	0.000	12:42

Agency: TITUSVILLE P.D.

Date/Time: 10/31/2009 12:44:31

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/31/2009 12:36

Air Blank	0.000	12:43	
Control Test	0.000	12:44	Interferent Detect
Air Blank	0.000	12:44	

Agency: TITUSVILLE P.D.

Date/Time: 11/30/2009 13:47:56

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/30/2009 14:34

Air Blank	0.000	13:46	
Control Test	0.000	13:47	Interferent Detect
Air Blank	0.000	13:47	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 12/31/2009 16:26:17

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 17:42

Air Blank	0.000	16:25
-----------	-------	-------

Control Test	0.000	16:25
--------------	-------	-------

Air Blank	0.000	16:26
-----------	-------	-------

Agency: TITUSVILLE P.D.

Date/Time: 12/31/2009 16:28:33

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 17:42

Air Blank	0.000	16:27
-----------	-------	-------

Control Test	0.000	16:28	Interferent Detect
--------------	-------	-------	--------------------

Air Blank	0.000	16:28
-----------	-------	-------

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

Agency: SEMINOLE COUNTY SO

Date/Time: 07/04/2009 03:40:54

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/29/2009 23:24

Air Blank	0.000	03:40
Control Test	0.077	03:40
Air Blank	0.000	03:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

Agency: COCOA P.D.

Date/Time: 02/26/2009 19:44:45

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/30/2009 17:16

Air Blank	0.000	19:43
Control Test	0.080	19:44
Air Blank	0.000	19:44

Agency: COCOA P.D.

Date/Time: 03/30/2009 17:12:30

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/30/2009 17:16

Air Blank	0.000	17:11
Control Test	0.079	17:12
Air Blank	0.000	17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 02/19/2009 16:04:53

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/20/2009 16:02

Air Blank	0.000	16:03
ITP Check	0.081	16:04
Air Blank	0.000	16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 08/17/2009 15:28:13

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/17/2009 16:21

Air Blank	0.000	15:27
Control Test	0.077	15:27
Air Blank	0.000	15:28

Agency: HAINES CITY PD

Date/Time: 08/17/2009 15:30:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/17/2009 16:21

Air Blank	0.000	15:29
Control Test	0.195	15:30
Air Blank	0.000	15:30

Agency: HAINES CITY PD

Date/Time: 09/09/2009 15:49:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/09/2009 17:17

Air Blank	0.000	15:48
Control Test	0.047	15:49
Air Blank	0.000	15:49

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 09/09/2009 15:54:08

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/09/2009 17:17

Air Blank	0.000	15:52
Control Test	0.000	15:53
Air Blank	0.000	15:54

Agency: HAINES CITY PD

Date/Time: 09/09/2009 15:57:02

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/09/2009 17:17

Air Blank	0.000	15:55	
Control Test	0.000	15:56	Interferent Detect
Air Blank	0.000	15:57	

Agency: HAINES CITY PD

Date/Time: 10/07/2009 18:28:59

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/07/2009 19:16

Air Blank	0.000	18:27
Control Test	0.194	18:28
Air Blank	0.000	18:28

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 10/07/2009 18:31:09

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/07/2009 19:16

Air Blank	0.000	18:29
Control Test	0.076	18:30
Air Blank	0.000	18:31

Agency: HAINES CITY PD

Date/Time: 10/07/2009 18:33:30

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/07/2009 19:16

Air Blank	0.000	18:32
Control Test	0.048	18:32
Air Blank	0.000	18:33

Agency: HAINES CITY PD

Date/Time: 11/06/2009 04:01:31

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/09/2009 17:24

Air Blank	0.000	04:00
Control Test	0.104	04:01
Air Blank	0.000	04:01

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 11/09/2009 16:42:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/09/2009 17:24

Air Blank	0.000	16:41
Control Test	0.193	16:41
Air Blank	0.000	16:42

Agency: HAINES CITY PD

Date/Time: 12/11/2009 18:17:53

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/11/2009 19:31

Air Blank	0.000	18:16
Control Test	0.196	18:17
Air Blank	0.000	18:17

Agency: HAINES CITY PD

Date/Time: 12/11/2009 18:19:56

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/11/2009 19:31

Air Blank	0.000	18:18
Control Test	0.078	18:19
Air Blank	0.000	18:19

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

Agency: JACKSONVILLE SO

Date/Time: 02/20/2009 06:51:10

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/26/2009 13:50

Air Blank	0.000	06:50
ITP Check	0.080	06:50
Air Blank	0.000	06:51

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 01/30/2009 19:24:23

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/03/2009 15:03

Air Blank	0.000	19:23	
Control Test	0.000	19:23	Interferent Detect
Air Blank	0.000	19:24	

Agency: PINELLAS PARK PD

Date/Time: 01/30/2009 19:26:59

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/03/2009 15:03

Air Blank	0.000	19:25	
Control Test	0.049	19:26	
Air Blank	0.000	19:26	

Agency: PINELLAS PARK PD

Date/Time: 01/30/2009 19:29:48

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/03/2009 15:03

Air Blank	0.000	19:29	
Control Test	0.084	19:29	
Air Blank	0.000	19:29	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 02/12/2009 12:09:44

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/12/2009 14:07

Air Blank	0.000	12:08
Control Test	0.084	12:09
Air Blank	0.000	12:09

Agency: PINELLAS PARK PD

Date/Time: 02/12/2009 12:57:10

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/12/2009 14:07

Air Blank	0.000	12:55
Control Test	0.050	12:56
Air Blank	0.000	12:57

Agency: PINELLAS PARK PD

Date/Time: 02/12/2009 12:59:31

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/12/2009 14:07

Air Blank	0.000	12:58
Control Test	0.080	12:58
Air Blank	0.000	12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 03/17/2009 08:12:55

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/17/2009 09:52

Air Blank	0.000	08:12
Control Test	0.083	08:12
Air Blank	0.000	08:12

Agency: PINELLAS PARK PD

Date/Time: 03/17/2009 08:15:57

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/17/2009 09:52

Air Blank	0.000	08:14
Control Test	0.050	08:15
Air Blank	0.000	08:15

Agency: PINELLAS PARK PD

Date/Time: 03/17/2009 08:18:20

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/17/2009 09:52

Air Blank	0.000	08:17
Control Test	0.081	08:17
Air Blank	0.000	08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 04/09/2009 11:05:36

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/09/2009 12:12

Air Blank	0.000	11:04
Control Test	0.083	11:05
Air Blank	0.000	11:05

Agency: PINELLAS PARK PD

Date/Time: 04/09/2009 11:07:33

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/09/2009 12:12

Air Blank	0.000	11:06
Control Test	0.050	11:07
Air Blank	0.000	11:07

Agency: PINELLAS PARK PD

Date/Time: 04/09/2009 11:09:16

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/09/2009 12:12

Air Blank	0.000	11:08
Control Test	0.081	11:08
Air Blank	0.000	11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 05/19/2009 09:23:25

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/19/2009 12:07

Air Blank	0.000	09:22
Control Test	0.083	09:22
Air Blank	0.000	09:23

Agency: PINELLAS PARK PD

Date/Time: 05/19/2009 09:25:41

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/19/2009 12:07

Air Blank	0.000	09:24
Control Test	0.080	09:25
Air Blank	0.000	09:25

Agency: PINELLAS PARK PD

Date/Time: 05/19/2009 09:27:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/19/2009 12:07

Air Blank	0.000	09:26
Control Test	0.050	09:27
Air Blank	0.000	09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 06/10/2009 11:43:53

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/16/2009 11:59

Air Blank	0.000	11:43
Control Test	0.084	11:43
Air Blank	0.000	11:43

Agency: PINELLAS PARK PD

Date/Time: 06/16/2009 10:59:49

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/16/2009 11:59

Air Blank	0.000	10:59
Control Test	0.084	10:59
Air Blank	0.000	10:59

Agency: PINELLAS PARK PD

Date/Time: 06/16/2009 11:01:52

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/16/2009 11:59

Air Blank	0.000	11:00
Control Test	0.050	11:01
Air Blank	0.000	11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 06/16/2009 11:03:48

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/16/2009 11:59

Air Blank	0.000	11:02
Control Test	0.079	11:03
Air Blank	0.000	11:03

Agency: PINELLAS PARK PD

Date/Time: 07/08/2009 10:18:11

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/08/2009 11:36

Air Blank	0.000	10:17
Control Test	0.084	10:17
Air Blank	0.000	10:18

Agency: PINELLAS PARK PD

Date/Time: 07/08/2009 10:20:17

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/08/2009 11:36

Air Blank	0.000	10:19
Control Test	0.049	10:19
Air Blank	0.000	10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 07/08/2009 10:21:34

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/08/2009 11:36

Air Blank	10:20	Ambient Fail
Air Blank	10:21	Purge Fail

Agency: PINELLAS PARK PD

Date/Time: 07/08/2009 10:27:17

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/08/2009 11:36

Air Blank	0.000	10:26
Control Test	0.080	10:26
Air Blank	0.000	10:27

Agency: PINELLAS PARK PD

Date/Time: 08/12/2009 08:37:21

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/12/2009 09:40

Air Blank	0.000	08:36
Control Test	0.083	08:36
Air Blank	0.000	08:37

Agency: PINELLAS PARK PD

Date/Time: 08/12/2009 08:39:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/12/2009 09:40

Air Blank	0.000	08:38
Control Test	0.049	08:38
Air Blank	0.000	08:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 08/12/2009 08:41:10

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/12/2009 09:40

Air Blank	0.000	08:39
Control Test	0.080	08:40
Air Blank	0.000	08:41

Agency: PINELLAS PARK PD

Date/Time: 09/10/2009 11:36:21

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 09/10/2009 13:04

Air Blank	0.000	11:35
Control Test	0.084	11:35
Air Blank	0.000	11:36

Agency: PINELLAS PARK PD

Date/Time: 09/10/2009 11:38:26

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/10/2009 13:04

Air Blank	0.000	11:37
Control Test	0.050	11:37
Air Blank	0.000	11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 09/10/2009 11:40:18

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/10/2009 13:04

Air Blank	0.000	11:39
Control Test	0.079	11:39
Air Blank	0.000	11:40

Agency: PINELLAS PARK PD

Date/Time: 10/14/2009 10:24:03

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/14/2009 11:38

Air Blank	0.000	10:23
Control Test	0.083	10:23
Air Blank	0.000	10:24

Agency: PINELLAS PARK PD

Date/Time: 10/14/2009 10:27:35

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/14/2009 11:38

Air Blank	0.000	10:26
Control Test	0.077	10:27
Air Blank	0.000	10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 10/14/2009 10:29:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/14/2009 11:38

Air Blank	0.000	10:28
Control Test	0.048	10:29
Air Blank	0.000	10:29

Agency: PINELLAS PARK PD

Date/Time: 11/12/2009 10:55:34

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/12/2009 12:10

Air Blank	0.000	10:54
Control Test	0.084	10:55
Air Blank	0.000	10:55

Agency: PINELLAS PARK PD

Date/Time: 11/12/2009 10:57:33

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/12/2009 12:10

Air Blank	0.000	10:56
Control Test	0.048	10:57
Air Blank	0.000	10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 11/12/2009 10:59:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/12/2009 12:10

Air Blank	0.000	10:58
Control Test	0.078	10:59
Air Blank	0.000	10:59

Agency: PINELLAS PARK PD

Date/Time: 12/09/2009 10:34:29

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/09/2009 15:47

Air Blank	0.000	10:33
Control Test	0.082	10:33
Air Blank	0.000	10:34

Agency: PINELLAS PARK PD

Date/Time: 12/09/2009 10:36:53

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/09/2009 15:47

Air Blank	0.000	10:35
Control Test	0.048	10:36
Air Blank	0.000	10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 12/09/2009 10:37:51

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/09/2009 15:47

Air Blank		10:37	Ambient Fail
-----------	--	-------	--------------

Air Blank	0.000	10:37	
-----------	-------	-------	--

Agency: PINELLAS PARK PD

Date/Time: 12/09/2009 10:39:34

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/09/2009 15:47

Air Blank	0.000	10:38	
-----------	-------	-------	--

Control Test	0.077	10:39	
--------------	-------	-------	--

Air Blank	0.000	10:39	
-----------	-------	-------	--

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

Agency: COCONUT CREEK PD

Date/Time: 02/02/2009 15:11:30

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/02/2009 15:43

Air Blank	0.000	15:10
Control Test	0.078	15:11
Air Blank	0.000	15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

Agency: JACKSONVILLE SO

Date/Time: 07/24/2009 05:14:46

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/30/2009 14:09

Air Blank	0.000	05:13
Control Test	0.080	05:14
Air Blank	0.000	05:14

Florida Department of Law Enforcement Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

Agency: LAKE ALFRED PD

Date/Time: 08/06/2009 12:31:50

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 08/06/2009 15:48

Air Blank		12:31	Sequence Aborted
Air Blank	0.000	12:31	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 06/18/2009 17:01:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/19/2009 10:22

Air Blank		17:01	Sequence Aborted
Air Blank	0.000	17:01	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency: CORAL SPRINGS PD

Date/Time: 09/15/2009 09:29:19

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 09/25/2009 09:52

Air Blank	0.000	09:28
Control Test	0.079	09:28
Air Blank	0.000	09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency: LAKELAND PD

Date/Time: 07/13/2009 08:58:46

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	08:57
Control Test	0.083	08:58
Air Blank	0.000	08:58

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:00:23

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	08:59
Control Test	0.084	08:59
Air Blank	0.000	09:00

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:01:43

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:00
Control Test	0.083	09:01
Air Blank	0.000	09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:05:17

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:04
Control Test	0.083	09:04
Air Blank	0.000	09:05

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:06:46

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:05
Control Test	0.083	09:06
Air Blank	0.000	09:06

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:08:25

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:07
Control Test	0.084	09:07
Air Blank	0.000	09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:31:31

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:30
Control Test	0.083	09:31
Air Blank	0.000	09:31

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:32:55

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:32
Control Test	0.082	09:32
Air Blank	0.000	09:32

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:34:18

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:33
Control Test	0.083	09:33
Air Blank	0.000	09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:35:36

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:34
Control Test	0.082	09:35
Air Blank	0.000	09:35

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:37:47

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:36
Control Test	0.082	09:37
Air Blank	0.000	09:37

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:40:18

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:39
Control Test	0.082	09:39
Air Blank	0.000	09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:43:53

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/15/2009 10:19

Air Blank	0.000	09:43
Control Test	0.082	09:43
Air Blank	0.000	09:43

Agency: LAKELAND PD

Date/Time: 10/04/2009 08:29:35

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/04/2009 09:37

Air Blank	0.000	08:28
Control Test	0.049	08:29
Air Blank	0.000	08:29

Agency: LAKELAND PD

Date/Time: 10/04/2009 08:31:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/04/2009 09:37

Air Blank	0.000	08:30
Control Test	0.194	08:31
Air Blank	0.000	08:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency: LAKELAND PD

Date/Time: 12/08/2009 10:00:14

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/08/2009 14:33

Air Blank 0.000 09:59

09:59

Sequence Aborted

Air Blank 0.000 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency: WINTER HAVEN PD

Date/Time: 04/19/2009 09:19:28

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/19/2009 10:41

Air Blank	0.000	09:18
Control Test	0.048	09:18
Air Blank	0.000	09:19

Agency: WINTER HAVEN PD

Date/Time: 04/19/2009 09:22:31

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/19/2009 10:41

Air Blank	0.000	09:21
Control Test	0.078	09:21
Air Blank	0.000	09:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency: GULFPORT PD

Date/Time: 03/13/2009 23:53:35

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/30/2009 06:52

Air Blank	0.000	23:52
Control Test	0.080	23:53
Air Blank	0.000	23:53

Agency: GULFPORT PD

Date/Time: 06/25/2009 15:57:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/25/2009 16:57

Air Blank	0.000	15:56
Control Test	0.081	15:57
Air Blank	0.000	15:57

Agency: GULFPORT PD

Date/Time: 08/27/2009 16:05:31

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/27/2009 17:07

Air Blank	0.000	16:04
Control Test	0.049	16:04
Air Blank	0.000	16:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency: GULFPORT PD

Date/Time: 08/27/2009 16:07:53

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/27/2009 17:07

Air Blank	0.000	16:06
Control Test	0.078	16:07
Air Blank	0.000	16:07

Agency: GULFPORT PD

Date/Time: 12/27/2009 22:13:30

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/27/2009 23:33

Air Blank	0.000	22:12
Control Test	0.041	22:13
Air Blank	0.000	22:13

Agency: GULFPORT PD

Date/Time: 12/27/2009 22:16:01

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/27/2009 23:33

Air Blank	0.000	22:14
Control Test	0.046	22:15
Air Blank	0.000	22:16

Florida Department of Law Enforcement Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency: GULFPORT PD

Date/Time: 12/27/2009 22:18:37

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/27/2009 23:33

Air Blank	0.000	22:17
Control Test	0.079	22:18
Air Blank	0.000	22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

Agency: TREASURE ISLAND PD

Date/Time: 06/11/2009 10:06:41

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/12/2009 10:45

Air Blank	0.000	10:05
ITP Check	0.081	10:06
Air Blank	0.000	10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency: NEW PORT RICHEY PD

Date/Time: 03/25/2009 13:03:01

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/25/2009 14:59

Air Blank	0.000	13:02
ITP Check	0.079	13:02
Air Blank	0.000	13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency: BROWARD COUNTY SO

Date/Time: 08/02/2009 16:41:00

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/21/2009 10:22

Air Blank	0.000	16:40
Control Test	0.079	16:40
Air Blank	0.000	16:41

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 03/22/2009 05:24:15

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/23/2009 14:44

Air Blank	0.000	05:23
Control Test	0.075	05:23
Air Blank	0.000	05:24

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency: JACKSONVILLE SO

Date/Time: 10/20/2009 09:02:39

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/27/2009 14:33

Air Blank		09:02	Sequence Aborted
Air Blank	0.000	09:02	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

Agency: CEDAR KEY PD

Date/Time: 06/11/2009 16:36:39

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/29/2009 21:42

Air Blank	0.000	16:35
Control Test	0.000	16:36
Air Blank	0.000	16:36

Agency: CEDAR KEY PD

Date/Time: 06/11/2009 16:40:13

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/29/2009 21:42

Air Blank	0.000	16:39
Control Test	0.051	16:39
Air Blank	0.000	16:40

Agency: CEDAR KEY PD

Date/Time: 11/29/2009 19:57:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/29/2009 21:21

Air Blank	0.000	19:56
Control Test	0.052	19:56
Air Blank	0.000	19:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency: DADE CITY PD

Date/Time: 08/03/2009 10:58:22

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/19/2009 12:39

Air Blank	0.000	10:57
Control Test	0.075	10:57
Air Blank	0.000	10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 02/20/2009 15:32:26

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/20/2009 18:14

Air Blank	0.000	15:31
Control Test	0.083	15:31
Air Blank	0.000	15:32

Agency: LARGO PD

Date/Time: 05/26/2009 09:40:21

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/26/2009 12:02

Air Blank	0.000	09:39
Control Test	0.078	09:39
Air Blank	0.000	09:40

Agency: LARGO PD

Date/Time: 05/26/2009 09:42:25

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 05/26/2009 12:02

Air Blank	0.000	09:41
ITP Check	0.078	09:41
Air Blank	0.000	09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 01/27/2009 18:38:39

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/27/2009 19:26

Air Blank	0.000	18:37
Control Test	0.047	18:38
Air Blank	0.000	18:38

Agency: PLANTATION PD

Date/Time: 01/27/2009 18:40:37

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/27/2009 19:26

Air Blank	0.000	18:39
Control Test	0.082	18:40
Air Blank	0.000	18:40

Agency: PLANTATION PD

Date/Time: 01/27/2009 18:42:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/27/2009 19:26

Air Blank	0.000	18:41
Control Test	0.201	18:41
Air Blank	0.000	18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 01/27/2009 18:43:45

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/27/2009 19:26

Air Blank	0.000	18:42
Control Test	0.079	18:43
Air Blank	0.000	18:43

Agency: PLANTATION PD

Date/Time: 02/17/2009 19:21:18

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/17/2009 20:04

Air Blank	0.000	19:20
Control Test	0.080	19:20
Air Blank	0.000	19:21

Agency: PLANTATION PD

Date/Time: 02/17/2009 19:23:27

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/17/2009 20:04

Air Blank	0.000	19:22
Control Test	0.048	19:22
Air Blank	0.000	19:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 02/17/2009 19:25:27

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/17/2009 20:04

Air Blank	0.000	19:24
Control Test	0.082	19:24
Air Blank	0.000	19:25

Agency: PLANTATION PD

Date/Time: 02/17/2009 19:27:28

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/17/2009 20:04

Air Blank	0.000	19:26
Control Test	0.200	19:26
Air Blank	0.000	19:27

Agency: PLANTATION PD

Date/Time: 03/26/2009 19:44:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/26/2009 20:27

Air Blank	0.000	19:43
Control Test	0.049	19:44
Air Blank	0.000	19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 03/26/2009 19:46:32

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/26/2009 20:27

Air Blank	0.000	19:45
Control Test	0.082	19:45
Air Blank	0.000	19:46

Agency: PLANTATION PD

Date/Time: 03/26/2009 19:48:37

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/26/2009 20:27

Air Blank	0.000	19:47
Control Test	0.201	19:48
Air Blank	0.000	19:48

Agency: PLANTATION PD

Date/Time: 03/26/2009 19:49:57

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/26/2009 20:27

Air Blank	0.000	19:49
Control Test	0.080	19:49
Air Blank	0.000	19:49

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 04/22/2009 19:29:47

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/22/2009 20:12

Air Blank	0.000	19:28
Control Test	0.049	19:29
Air Blank	0.000	19:29

Agency: PLANTATION PD

Date/Time: 04/22/2009 19:31:53

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/22/2009 20:12

Air Blank	0.000	19:30
Control Test	0.081	19:31
Air Blank	0.000	19:31

Agency: PLANTATION PD

Date/Time: 04/22/2009 19:34:01

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/22/2009 20:12

Air Blank	0.000	19:32
Control Test	0.199	19:33
Air Blank	0.000	19:34

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 04/22/2009 19:35:26

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/22/2009 20:12

Air Blank	0.000	19:34
Control Test	0.079	19:34
Air Blank	0.000	19:35

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:22:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:21
Control Test	0.000	18:21
Air Blank	0.000	18:22

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:24:48

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:23
Control Test	0.000	18:24
Air Blank	0.000	18:24

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:26:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:25
Control Test	0.000	18:26
Air Blank	0.000	18:26

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:29:33

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:28
Control Test	0.081	18:29
Air Blank	0.000	18:29

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:31:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:30
Control Test	0.080	18:31
Air Blank	0.000	18:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:34:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:33
Control Test	0.049	18:33
Air Blank	0.000	18:34

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:35:48

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:35
Control Test	0.055	18:35
Air Blank	0.000	18:35

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:37:05

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	18:36	Ambient Fail
Air Blank	18:37	Purge Fail

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:38:39

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:37
Control Test	0.080	18:38
Air Blank	0.000	18:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:42:09

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:40
Control Test	0.050	18:41
Air Blank	0.000	18:42

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:52:41

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:51
Control Test	0.081	18:52
Air Blank	0.000	18:52

Agency: PLANTATION PD

Date/Time: 05/06/2009 18:55:04

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/06/2009 19:32

Air Blank	0.000	18:53
Control Test	0.201	18:54
Air Blank	0.000	18:55

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 06/27/2009 18:46:01

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/27/2009 19:34

Air Blank		18:45	RFI Detect
Air Blank	0.000	18:46	

Agency: PLANTATION PD

Date/Time: 06/27/2009 18:47:26

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/27/2009 19:34

Air Blank	0.000	18:46	
Control Test	0.079	18:46	
Air Blank	0.000	18:47	

Agency: PLANTATION PD

Date/Time: 06/27/2009 18:49:39

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/27/2009 19:34

Air Blank	0.000	18:48	
Control Test	0.050	18:49	
Air Blank	0.000	18:49	

Agency: PLANTATION PD

Date/Time: 06/27/2009 18:51:50

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/27/2009 19:34

Air Blank	0.000	18:50	
Control Test	0.081	18:51	
Air Blank	0.000	18:51	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 06/27/2009 18:53:56

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/27/2009 19:34

Air Blank	0.000	18:52
Control Test	0.201	18:53
Air Blank	0.000	18:53

Agency: PLANTATION PD

Date/Time: 06/27/2009 18:55:52

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/27/2009 19:34

Air Blank	0.000	18:54
Control Test	0.000	18:55
Air Blank	0.000	18:55

Agency: PLANTATION PD

Date/Time: 07/14/2009 19:57:57

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/14/2009 20:41

Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 07/14/2009 20:00:11

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/14/2009 20:41

Air Blank	0.000	19:58
Control Test	0.049	19:59
Air Blank	0.000	20:00

Agency: PLANTATION PD

Date/Time: 07/14/2009 20:02:09

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/14/2009 20:41

Air Blank	0.000	20:00
Control Test	0.082	20:01
Air Blank	0.000	20:02

Agency: PLANTATION PD

Date/Time: 07/14/2009 20:04:16

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/14/2009 20:41

Air Blank	0.000	20:03
Control Test	0.202	20:03
Air Blank	0.000	20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 08/19/2009 21:37:35

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/20/2009 19:13

Air Blank	0.000	21:36
Control Test	0.078	21:37
Air Blank	0.000	21:37

Agency: PLANTATION PD

Date/Time: 08/20/2009 18:29:35

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/20/2009 19:13

Air Blank	0.000	18:28
Control Test	0.078	18:29
Air Blank	0.000	18:29

Agency: PLANTATION PD

Date/Time: 08/20/2009 18:31:43

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/20/2009 19:13

Air Blank	0.000	18:30
Control Test	0.049	18:31
Air Blank	0.000	18:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 08/20/2009 18:33:49

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/20/2009 19:13

Air Blank	0.000	18:32
Control Test	0.082	18:33
Air Blank	0.000	18:33

Agency: PLANTATION PD

Date/Time: 08/20/2009 18:35:52

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/20/2009 19:13

Air Blank	0.000	18:34
Control Test	0.202	18:35
Air Blank	0.000	18:35

Agency: PLANTATION PD

Date/Time: 09/16/2009 14:37:27

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 09/16/2009 19:56

Air Blank	0.000	14:36
Control Test	0.078	14:36
Air Blank	0.000	14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 09/16/2009 19:10:58

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 09/16/2009 19:56

Air Blank	0.000	19:10
Control Test	0.078	19:10
Air Blank	0.000	19:10

Agency: PLANTATION PD

Date/Time: 09/16/2009 19:13:20

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/16/2009 19:56

Air Blank	0.000	19:12
Control Test	0.050	19:12
Air Blank	0.000	19:13

Agency: PLANTATION PD

Date/Time: 09/16/2009 19:15:31

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/16/2009 19:56

Air Blank	0.000	19:14
Control Test	0.080	19:14
Air Blank	0.000	19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 09/16/2009 19:17:35

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/16/2009 19:56

Air Blank	0.000	19:16
Control Test	0.203	19:17
Air Blank	0.000	19:17

Agency: PLANTATION PD

Date/Time: 10/21/2009 19:15:16

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/21/2009 20:01

Air Blank	0.000	19:14
Control Test	0.079	19:14
Air Blank	0.000	19:15

Agency: PLANTATION PD

Date/Time: 10/21/2009 19:17:24

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/21/2009 20:01

Air Blank	0.000	19:16
Control Test	0.050	19:16
Air Blank	0.000	19:17

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 10/21/2009 19:19:26

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/21/2009 20:01

Air Blank	0.000	19:18
Control Test	0.080	19:18
Air Blank	0.000	19:19

Agency: PLANTATION PD

Date/Time: 10/21/2009 19:21:22

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/21/2009 20:01

Air Blank	0.000	19:20
Control Test	0.200	19:20
Air Blank	0.000	19:21

Agency: PLANTATION PD

Date/Time: 11/21/2009 11:25:17

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/24/2009 08:46

Air Blank	0.000	11:24
Control Test	0.078	11:24
Air Blank	0.000	11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 11/21/2009 11:28:06

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/24/2009 08:46

Air Blank	0.000	11:26
Control Test	0.050	11:27
Air Blank	0.000	11:28

Agency: PLANTATION PD

Date/Time: 11/21/2009 11:30:09

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/24/2009 08:46

Air Blank	0.000	11:28
Control Test	0.081	11:29
Air Blank	0.000	11:30

Agency: PLANTATION PD

Date/Time: 11/21/2009 11:32:27

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/24/2009 08:46

Air Blank	0.000	11:31
Control Test	0.201	11:31
Air Blank	0.000	11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 11/21/2009 12:23:41

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/24/2009 08:46

Air Blank	0.000	12:22
Control Test	0.081	12:23
Air Blank	0.000	12:23

Agency: PLANTATION PD

Date/Time: 11/21/2009 12:25:11

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/24/2009 08:46

Air Blank	0.000	12:24
Control Test	0.081	12:24
Air Blank	0.000	12:25

Agency: PLANTATION PD

Date/Time: 11/21/2009 12:26:09

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/24/2009 08:46

Air Blank		12:25	Ambient Fail
Air Blank	0.000	12:26	

Agency: PLANTATION PD

Date/Time: 11/21/2009 12:27:42

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/24/2009 08:46

Air Blank	0.000	12:26
Control Test	0.081	12:27
Air Blank	0.000	12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 12/26/2009 18:34:03

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/26/2009 19:18

Air Blank	0.000	18:33
Control Test	0.082	18:33
Air Blank	0.000	18:34

Agency: PLANTATION PD

Date/Time: 12/26/2009 18:36:19

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/26/2009 19:18

Air Blank	0.000	18:35
Control Test	0.050	18:35
Air Blank	0.000	18:36

Agency: PLANTATION PD

Date/Time: 12/26/2009 18:38:27

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/26/2009 19:18

Air Blank	0.000	18:37
Control Test	0.081	18:37
Air Blank	0.000	18:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 12/26/2009 18:40:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/26/2009 19:18

Air Blank	0.000	18:39
Control Test	0.201	18:40
Air Blank	0.000	18:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency: JACKSONVILLE SO

Date/Time: 04/16/2009 15:09:02

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/30/2009 07:25

Air Blank	0.000	15:08
Control Test	0.078	15:08
Air Blank	0.000	15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency: FHP TROOP A

Date/Time: 01/27/2009 13:07:40

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/27/2009 15:08

Air Blank	0.000	13:06
Control Test	0.190	13:07
Air Blank	0.000	13:07

Agency: FHP TROOP A

Date/Time: 01/27/2009 13:09:37

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/27/2009 15:08

Air Blank	0.000	13:08
Control Test	0.192	13:09
Air Blank	0.000	13:09

Agency: FHP TROOP A

Date/Time: 01/27/2009 13:11:46

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/27/2009 15:08

Air Blank	0.000	13:10
Control Test	0.077	13:11
Air Blank	0.000	13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency: FHP TROOP A

Date/Time: 01/27/2009 13:13:50

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/27/2009 15:08

Air Blank	0.000	13:12
Control Test	0.046	13:13
Air Blank	0.000	13:13

Agency: FHP TROOP A

Date/Time: 04/30/2009 00:38:59

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/30/2009 02:33

Air Blank	0.000	00:37
Control Test	0.193	00:38
Air Blank	0.000	00:38

Agency: FHP TROOP A

Date/Time: 08/07/2009 09:05:27

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/07/2009 11:28

Air Blank	0.000	09:04
Control Test	0.192	09:04
Air Blank	0.000	09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency: FHP TROOP A

Date/Time: 12/30/2009 19:26:43

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	0.000	19:25
Control Test	0.184	19:26
Air Blank	0.000	19:26

Agency: FHP TROOP A

Date/Time: 12/30/2009 19:28:28

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	0.000	19:27
Control Test	0.178	19:27
Air Blank	0.000	19:28

Agency: FHP TROOP A

Date/Time: 12/30/2009 19:37:57

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	0.000	19:36
Control Test	0.193	19:37
Air Blank	0.000	19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency: FHP TROOP A

Date/Time: 12/30/2009 19:40:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	0.000	19:39
Control Test	0.189	19:40
Air Blank	0.000	19:40

Agency: FHP TROOP A

Date/Time: 12/31/2009 14:52:34

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	0.000	14:51
Control Test	0.000	14:52
Air Blank	0.000	14:52

Agency: FHP TROOP A

Date/Time: 12/31/2009 14:53:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank		14:52	Improper Sample
Air Blank	0.000	14:53	

Agency: FHP TROOP A

Date/Time: 12/31/2009 14:54:32

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank		14:53	Ambient Fail
Air Blank		14:54	Purge Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency: FHP TROOP A

Date/Time: 12/31/2009 14:56:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	0.000	14:55
Control Test	0.048	14:55
Air Blank	0.000	14:56

Agency: FHP TROOP A

Date/Time: 12/31/2009 14:57:58

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	0.000	14:56
Control Test	0.076	14:57
Air Blank	0.000	14:57

Agency: FHP TROOP A

Date/Time: 12/31/2009 14:59:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	14:58	Ambient Fail
Air Blank	14:59	Purge Fail

Agency: FHP TROOP A

Date/Time: 12/31/2009 15:00:13

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	14:59	Ambient Fail
Air Blank	0.000	15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency: FHP TROOP A

Date/Time: 12/31/2009 15:01:29

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	15:00	Ambient Fail
Air Blank	15:01	Purge Fail

Agency: FHP TROOP A

Date/Time: 12/31/2009 15:03:31

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/31/2009 16:44

Air Blank	0.000	15:02
Control Test	0.197	15:03
Air Blank	0.000	15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency: FHP

Date/Time: 04/04/2009 02:12:11

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/26/2009 20:05

Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency: PUTNAM COUNTY SO

Date/Time: 01/27/2009 10:44:08

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 11:18

Air Blank	0.000	10:43
Control Test	0.079	10:43
Air Blank	0.000	10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

Agency: BROWARD COUNTY SO

Date/Time: 01/12/2009 18:40:41

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/22/2009 18:42

Air Blank	0.000	18:39
Control Test	0.080	18:40
Air Blank	0.000	18:40

Agency: BROWARD COUNTY SO

Date/Time: 03/26/2009 14:24:31

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/09/2009 15:24

Air Blank	0.000	14:23
Control Test	0.080	14:24
Air Blank	0.000	14:24

Agency: BROWARD COUNTY SO

Date/Time: 05/07/2009 19:21:29

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/14/2009 17:09

Air Blank	0.000	19:20
Control Test	0.078	19:20
Air Blank	0.000	19:21

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

Agency: BROWARD COUNTY SO

Date/Time: 07/16/2009 17:02:07

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/13/2009 16:00

Air Blank	0.000	17:00
Control Test	0.076	17:01
Air Blank	0.000	17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency: HAINES CITY PD

Date/Time: 06/02/2009 14:13:28

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/06/2009 15:31

Air Blank	0.000	14:12	
Control Test	0.000	14:12	Interferent Detect
Air Blank	0.000	14:13	

Agency: HAINES CITY PD

Date/Time: 06/02/2009 14:15:39

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/06/2009 15:31

Air Blank	0.000	14:14	
Control Test	0.049	14:15	
Air Blank	0.000	14:15	

Agency: HAINES CITY PD

Date/Time: 06/02/2009 14:18:16

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/06/2009 15:31

Air Blank	0.000	14:17	
Control Test	0.198	14:17	
Air Blank	0.000	14:18	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency: HAINES CITY PD

Date/Time: 06/02/2009 14:20:37

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/06/2009 15:31

Air Blank	0.000	14:19
Control Test	0.076	14:20
Air Blank	0.000	14:20

Agency: HAINES CITY PD

Date/Time: 08/17/2009 15:29:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/17/2009 16:19

Air Blank	0.000	15:28
Control Test	0.000	15:28
Air Blank	0.000	15:29

Agency: HAINES CITY PD

Date/Time: 08/17/2009 15:32:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/17/2009 16:19

Air Blank	0.000	15:31	
Control Test	0.000	15:31	Interferent Detect
Air Blank	0.000	15:32	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency: HAINES CITY PD

Date/Time: 08/17/2009 15:34:50

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/17/2009 16:19

Air Blank	0.000	15:33
Control Test	0.050	15:34
Air Blank	0.000	15:34

Agency: HAINES CITY PD

Date/Time: 09/09/2009 15:50:50

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/09/2009 17:23

Air Blank	0.000	15:49
Control Test	0.078	15:50
Air Blank	0.000	15:50

Agency: HAINES CITY PD

Date/Time: 09/09/2009 15:53:08

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/09/2009 17:23

Air Blank	0.000	15:51
Control Test	0.199	15:52
Air Blank	0.000	15:53

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency: HAINES CITY PD

Date/Time: 10/07/2009 18:30:10

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/07/2009 19:12

Air Blank	0.000	18:29
Control Test	0.000	18:29
Air Blank	0.000	18:30

Agency: HAINES CITY PD

Date/Time: 10/07/2009 18:32:25

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/07/2009 19:12

Air Blank	0.000	18:31	
Control Test	0.000	18:31	Interferent Detect
Air Blank	0.000	18:32	

Agency: HAINES CITY PD

Date/Time: 11/09/2009 16:43:36

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/09/2009 17:27

Air Blank	0.000	16:42	
Control Test	0.009	16:43	Interferent Detect
Air Blank	0.000	16:43	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency: HAINES CITY PD

Date/Time: 11/09/2009 16:45:50

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/09/2009 17:27

Air Blank	0.000	16:44
Control Test	0.050	16:45
Air Blank	0.000	16:45

Agency: HAINES CITY PD

Date/Time: 11/09/2009 16:47:52

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/09/2009 17:27

Air Blank	0.000	16:46
Control Test	0.079	16:47
Air Blank	0.000	16:47

Agency: HAINES CITY PD

Date/Time: 11/29/2009 19:38:38

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/11/2009 19:27

Air Blank	0.000	19:37
Control Test	0.082	19:38
Air Blank	0.000	19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency: HAINES CITY PD

Date/Time: 12/11/2009 18:18:57

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/11/2009 19:27

Air Blank	0.000	18:17	
Control Test	0.008	18:18	Interferent Detect
Air Blank	0.000	18:18	

Agency: HAINES CITY PD

Date/Time: 12/11/2009 18:21:04

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/11/2009 19:27

Air Blank	0.000	18:19	
Control Test	0.050	18:20	
Air Blank	0.000	18:21	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

Agency: BREVARD COUNTY S.O.

Date/Time: 04/15/2009 13:11:01

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/28/2009 15:43

Air Blank	0.000	13:10
Control Test	0.000	13:10
Air Blank	0.000	13:11

Agency: BREVARD COUNTY S.O.

Date/Time: 04/15/2009 13:12:50

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/28/2009 15:43

Air Blank	0.000	13:11
Control Test	0.079	13:12
Air Blank	0.000	13:12

Agency: BREVARD COUNTY S.O.

Date/Time: 04/15/2009 13:14:56

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/28/2009 15:43

Air Blank	0.000	13:14
Control Test	0.080	13:14
Air Blank	0.000	13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

Agency: CLERMONT P.D.

Date/Time: 01/20/2009 04:21:39

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/20/2009 05:08

Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21

Agency: CLERMONT P.D.

Date/Time: 12/29/2009 04:03:09

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/29/2009 05:07

Air Blank	0.000	04:01
Control Test	0.043	04:02
Air Blank	0.000	04:03

Agency: CLERMONT P.D.

Date/Time: 12/29/2009 04:05:02

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/29/2009 05:07

Air Blank	0.000	04:03
Control Test	0.045	04:04
Air Blank	0.000	04:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

Agency: CLERMONT P.D.

Date/Time: 12/29/2009 04:07:18

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/29/2009 05:07

Air Blank	0.000	04:06
Control Test	0.068	04:06
Air Blank	0.000	04:07

Agency: CLERMONT P.D.

Date/Time: 12/29/2009 04:09:04

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/29/2009 05:07

Air Blank	0.000	04:07
Control Test	0.071	04:08
Air Blank	0.000	04:09

Agency: CLERMONT P.D.

Date/Time: 12/29/2009 04:11:03

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/29/2009 05:07

Air Blank	0.000	04:09
Control Test	0.189	04:10
Air Blank	0.000	04:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

Agency: CLERMONT P.D.

Date/Time: 12/29/2009 04:13:37

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/29/2009 05:07

Air Blank	0.000	04:12
Control Test	0.074	04:13
Air Blank	0.000	04:13

Agency: CLERMONT P.D.

Date/Time: 12/29/2009 05:02:44

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/22/2010 03:22

Air Blank	0.000	05:01
Control Test	0.068	05:02
Air Blank	0.000	05:02

Agency: CLERMONT P.D.

Date/Time: 12/29/2009 05:04:36

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/22/2010 03:22

Air Blank	0.000	05:03
Control Test	0.075	05:04
Air Blank	0.000	05:04

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: LADY LAKE P.D.

Date/Time: 08/12/2009 03:25:36

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/25/2009 11:39

Air Blank	0.000	03:24
Control Test	0.000	03:25
Air Blank	0.000	03:25

Agency: LADY LAKE P.D.

Date/Time: 08/12/2009 03:36:58

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 09/25/2009 11:39

Air Blank	0.000	03:36
ITP Check	0.079	03:36
Air Blank	0.000	03:36

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

Agency: BOCA RATON PD

Date/Time: 04/10/2009 12:17:24

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/15/2009 14:14

Air Blank	0.000	12:16
Control Test	0.079	12:16
Air Blank	0.000	12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 10/21/2009 15:13:57

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/21/2009 16:11

Air Blank	0.000	15:13
Control Test	0.079	15:13
Air Blank	0.000	15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

Agency: BOYNTON BEACH PD

Date/Time: 11/17/2009 15:59:24

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/17/2009 17:27

Air Blank		15:58	Sequence Aborted
Air Blank	0.000	15:59	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 05/09/2009 12:17:23

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/07/2009 23:52

Air Blank	0.000	12:16
ITP Check	0.081	12:16
Air Blank	0.000	12:17

Agency: COLLIER COUNTY SO

Date/Time: 05/20/2009 10:14:34

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/07/2009 23:52

Air Blank	0.000	10:13
Control Test	0.078	10:14
Air Blank	0.000	10:14

Agency: COLLIER COUNTY SO

Date/Time: 05/20/2009 10:16:43

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/07/2009 23:52

Air Blank	0.000	10:15
Control Test	0.079	10:16
Air Blank	0.000	10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 05/20/2009 10:18:22

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/07/2009 23:52

Air Blank	0.000	10:17
Control Test	0.079	10:17
Air Blank	0.000	10:18

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 20:25:07

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/07/2009 23:52

Air Blank	0.000	20:23
Control Test	0.047	20:24
Air Blank	0.000	20:25

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 21:16:47

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/07/2009 23:52

Air Blank	0.000	21:15
Control Test	0.000	21:16
Air Blank	0.000	21:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 06/07/2009 21:20:47

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/07/2009 23:52

Air Blank	0.000	21:19
Control Test	0.000	21:20
Air Blank	0.000	21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency: COLLIER COUNTY SO

Date/Time: 07/01/2009 08:11:35

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/01/2009 13:12

Air Blank	0.000	08:10
Control Test	0.000	08:11
Air Blank	0.000	08:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency: PANAMA CITY PD

Date/Time: 06/24/2009 12:09:43

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/24/2009 14:19

Air Blank	0.000	12:08
Control Test	0.243	12:09
Air Blank	0.000	12:09

Agency: PANAMA CITY PD

Date/Time: 06/24/2009 12:11:56

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/24/2009 14:19

Air Blank	0.000	12:10
Control Test	0.196	12:11
Air Blank	0.000	12:11

Agency: PANAMA CITY PD

Date/Time: 06/24/2009 12:13:40

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/24/2009 14:19

Air Blank	0.000	12:12
Control Test	0.048	12:13
Air Blank	0.000	12:13

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency: PANAMA CITY PD

Date/Time: 06/24/2009 13:22:53

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/30/2009 10:30

Air Blank	0.000	13:21
Control Test	0.085	13:22
Air Blank	0.000	13:22

Agency: PANAMA CITY PD

Date/Time: 08/24/2009 06:52:39

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/24/2009 08:40

Air Blank	0.000	06:51
Control Test	0.077	06:52
Air Blank	0.000	06:52

Agency: PANAMA CITY PD

Date/Time: 09/30/2009 06:54:59

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/30/2009 09:23

Air Blank	0.000	06:53
Control Test	0.053	06:54
Air Blank	0.000	06:54

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency: PANAMA CITY PD

Date/Time: 09/30/2009 06:56:44

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/30/2009 09:23

Air Blank	0.000	06:55
Control Test	0.078	06:56
Air Blank	0.000	06:56

Agency: PANAMA CITY PD

Date/Time: 10/29/2009 14:18:32

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/29/2009 16:58

Air Blank	0.000	14:17
Control Test	0.077	14:18
Air Blank	0.000	14:18

Agency: PANAMA CITY PD

Date/Time: 12/05/2009 15:03:23

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/16/2009 12:16

Air Blank	0.000	15:02
Control Test	0.080	15:02
Air Blank	0.000	15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency: TYNDALL AFB

Date/Time: 04/23/2009 10:15:48

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/05/2009 06:34

Air Blank	0.000	10:15
Control Test	0.077	10:15
Air Blank	0.000	10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 01/31/2009 08:13:31

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/31/2009 09:08

Air Blank	0.000	08:12
Control Test	0.048	08:13
Air Blank	0.000	08:13

Agency: AUBURNDALE PD

Date/Time: 01/31/2009 08:16:00

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/31/2009 09:08

Air Blank	0.000	08:14
Control Test	0.081	08:15
Air Blank	0.000	08:16

Agency: AUBURNDALE PD

Date/Time: 01/31/2009 08:18:03

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/31/2009 09:08

Air Blank	0.000	08:16
Control Test	0.198	08:17
Air Blank	0.000	08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 02/27/2009 18:20:19

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/27/2009 19:10

Air Blank	0.000	18:19
Control Test	0.048	18:19
Air Blank	0.000	18:20

Agency: AUBURNDALE PD

Date/Time: 02/27/2009 18:22:08

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/27/2009 19:10

Air Blank	0.000	18:20
Control Test	0.079	18:21
Air Blank	0.000	18:22

Agency: AUBURNDALE PD

Date/Time: 02/27/2009 18:23:50

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/27/2009 19:10

Air Blank	0.000	18:22
Control Test	0.199	18:23
Air Blank	0.000	18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 03/15/2009 10:00:41

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/15/2009 10:46

Air Blank	0.000	09:59
Control Test	0.048	10:00
Air Blank	0.000	10:00

Agency: AUBURNDALE PD

Date/Time: 03/15/2009 10:02:26

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/15/2009 10:46

Air Blank	0.000	10:01
Control Test	0.079	10:01
Air Blank	0.000	10:02

Agency: AUBURNDALE PD

Date/Time: 03/15/2009 10:04:20

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/15/2009 10:46

Air Blank	0.000	10:03
Control Test	0.198	10:03
Air Blank	0.000	10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 04/26/2009 08:44:37

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/26/2009 09:35

Air Blank	0.000	08:43
Control Test	0.048	08:44
Air Blank	0.000	08:44

Agency: AUBURNDALE PD

Date/Time: 04/26/2009 08:46:47

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/26/2009 09:35

Air Blank	0.000	08:45
Control Test	0.078	08:46
Air Blank	0.000	08:46

Agency: AUBURNDALE PD

Date/Time: 04/26/2009 08:48:29

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/26/2009 09:35

Air Blank	0.000	08:47
Control Test	0.205	08:47
Air Blank	0.000	08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 05/24/2009 09:09:14

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/24/2009 09:56

Air Blank	0.000	09:08
Control Test	0.048	09:08
Air Blank	0.000	09:09

Agency: AUBURNDALE PD

Date/Time: 05/24/2009 09:11:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/24/2009 09:56

Air Blank	0.000	09:10
Control Test	0.078	09:10
Air Blank	0.000	09:11

Agency: AUBURNDALE PD

Date/Time: 05/24/2009 09:13:08

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/24/2009 09:56

Air Blank	0.000	09:11
Control Test	0.197	09:12
Air Blank	0.000	09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 06/29/2009 17:06:35

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/29/2009 17:51

Air Blank	0.000	17:05
Control Test	0.048	17:06
Air Blank	0.000	17:06

Agency: AUBURNDALE PD

Date/Time: 06/29/2009 17:08:22

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/29/2009 17:51

Air Blank	0.000	17:07
Control Test	0.080	17:07
Air Blank	0.000	17:08

Agency: AUBURNDALE PD

Date/Time: 06/29/2009 17:10:05

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/29/2009 17:51

Air Blank	0.000	17:08
Control Test	0.198	17:09
Air Blank	0.000	17:10

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 07/19/2009 09:54:12

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/19/2009 10:48

Air Blank	0.000	09:53
Control Test	0.048	09:53
Air Blank	0.000	09:54

Agency: AUBURNDALE PD

Date/Time: 07/19/2009 09:56:12

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/19/2009 10:48

Air Blank	0.000	09:55
Control Test	0.079	09:55
Air Blank	0.000	09:56

Agency: AUBURNDALE PD

Date/Time: 07/19/2009 09:57:59

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/19/2009 10:48

Air Blank	0.000	09:56
Control Test	0.198	09:57
Air Blank	0.000	09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 08/29/2009 14:27:57

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/29/2009 19:15

Air Blank	0.000	14:26
Control Test	0.048	14:27
Air Blank	0.000	14:27

Agency: AUBURNDALE PD

Date/Time: 08/29/2009 14:30:35

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/29/2009 19:15

Air Blank	0.000	14:29
Control Test	0.078	14:30
Air Blank	0.000	14:30

Agency: AUBURNDALE PD

Date/Time: 08/29/2009 14:32:35

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/29/2009 19:15

Air Blank	0.000	14:31
Control Test	0.197	14:32
Air Blank	0.000	14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 09/13/2009 16:52:41

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/13/2009 17:50

Air Blank	0.000	16:51
Control Test	0.079	16:52
Air Blank	0.000	16:52

Agency: AUBURNDALE PD

Date/Time: 09/13/2009 16:54:40

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/13/2009 17:50

Air Blank	0.000	16:53
Control Test	0.049	16:54
Air Blank	0.000	16:54

Agency: AUBURNDALE PD

Date/Time: 09/13/2009 16:57:42

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/13/2009 17:50

Air Blank	0.000	16:56
Control Test	0.197	16:57
Air Blank	0.000	16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 10/24/2009 15:48:59

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/24/2009 16:33

Air Blank	0.000	15:47
Control Test	0.049	15:48
Air Blank	0.000	15:48

Agency: AUBURNDALE PD

Date/Time: 10/24/2009 15:50:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/24/2009 16:33

Air Blank	0.000	15:49
Control Test	0.078	15:50
Air Blank	0.000	15:50

Agency: AUBURNDALE PD

Date/Time: 10/24/2009 15:52:27

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/24/2009 16:33

Air Blank	0.000	15:51
Control Test	0.198	15:51
Air Blank	0.000	15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 11/22/2009 06:48:14

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/22/2009 15:06

Air Blank	0.000	06:47
Control Test	0.047	06:47
Air Blank	0.000	06:48

Agency: AUBURNDALE PD

Date/Time: 11/22/2009 06:50:43

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/22/2009 15:06

Air Blank	0.000	06:49
Control Test	0.077	06:50
Air Blank	0.000	06:50

Agency: AUBURNDALE PD

Date/Time: 11/22/2009 06:52:27

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/22/2009 15:06

Air Blank	0.000	06:51
Control Test	0.199	06:51
Air Blank	0.000	06:52

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency: DAYTONA BEACH SHORES

Date/Time: 10/01/2009 07:38:01

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/26/2009 15:27

Air Blank		07:37	RFI Detect
-----------	--	-------	------------

Air Blank	0.000	07:38	
-----------	-------	-------	--

Agency: DAYTONA BEACH SHORES

Date/Time: 10/01/2009 07:40:07

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/26/2009 15:27

Air Blank	0.000	07:39	
-----------	-------	-------	--

Control Test	0.080	07:39	
--------------	-------	-------	--

Air Blank	0.000	07:40	
-----------	-------	-------	--

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:16:57

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank	0.000	18:16
Control Test	0.078	18:16
Air Blank	0.000	18:16

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:18:40

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank	0.000	18:17
Control Test	0.078	18:18
Air Blank	0.000	18:18

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:20:01

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank	0.000	18:19
Control Test	0.079	18:19
Air Blank	0.000	18:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:21:20

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank	0.000	18:20
Control Test	0.079	18:20
Air Blank	0.000	18:21

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:22:39

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank	0.000	18:21
Control Test	0.078	18:22
Air Blank	0.000	18:22

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:23:58

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank	0.000	18:23
Control Test	0.078	18:23
Air Blank	0.000	18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:25:18

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank	0.000	18:24
Control Test	0.078	18:24
Air Blank	0.000	18:25

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:26:36

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank	0.000	18:25
Control Test	0.079	18:26
Air Blank	0.000	18:26

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:27:34

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank		18:27	Ambient Fail
Air Blank	0.000	18:27	

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:29:09

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank	0.000	18:28
Control Test	0.079	18:28
Air Blank	0.000	18:29

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency: PORT ORANGE P.D.

Date/Time: 07/03/2009 18:30:35

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/03/2009 21:02

Air Blank	0.000	18:29
Control Test	0.078	18:30
Air Blank	0.000	18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency: BROWARD COUNTY SO

Date/Time: 07/30/2009 12:08:51

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/13/2009 14:46

Air Blank	0.000	12:08
Control Test	0.080	12:08
Air Blank	0.000	12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

Agency: BROWARD COUNTY SO

Date/Time: 05/14/2009 17:17:46

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/17/2009 16:43

Air Blank	0.000	17:16
Control Test	0.079	17:17
Air Blank	0.000	17:17

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency: BROWARD COUNTY SO

Date/Time: 07/30/2009 12:29:28

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/13/2009 14:52

Air Blank	0.000	12:28
Control Test	0.078	12:28
Air Blank	0.000	12:29

Agency: BROWARD COUNTY SO

Date/Time: 08/20/2009 14:42:46

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/27/2009 15:41

Air Blank	0.000	14:41
Control Test	0.079	14:42
Air Blank	0.000	14:42

Agency: BROWARD COUNTY SO

Date/Time: 08/20/2009 14:44:17

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/27/2009 15:41

Air Blank	0.000	14:43
Control Test	0.079	14:43
Air Blank	0.000	14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency: BREVARD COUNTY S.O.

Date/Time: 11/18/2009 17:11:11

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/18/2009 18:49

Air Blank	0.000	17:09	
Control Test	0.000	17:10	Interferent Detect
Air Blank	0.000	17:11	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency: COCOA P.D.

Date/Time: 02/26/2009 19:39:45

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/30/2009 17:20

Air Blank	0.000	19:38
Control Test	0.081	19:39
Air Blank	0.000	19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency: ROCKLEDGE P.D.

Date/Time: 11/24/2009 09:37:14

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/24/2009 12:42

Air Blank		09:36	Sequence Aborted
-----------	--	-------	------------------

Air Blank	0.000	09:37	
-----------	-------	-------	--

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency: SEMINOLE COUNTY S.O.

Date/Time: 04/25/2009 00:42:34

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/26/2009 21:49

Air Blank	0.000	00:41
-----------	-------	-------

Control Test	0.079	00:42
--------------	-------	-------

Air Blank	0.000	00:42
-----------	-------	-------

Agency: SEMINOLE COUNTY S.O.

Date/Time: 09/19/2009 04:14:54

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 09/21/2009 22:16

Air Blank	0.000	04:14
-----------	-------	-------

Control Test	0.079	04:14
--------------	-------	-------

Air Blank	0.000	04:14
-----------	-------	-------

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

Agency: MIAMI-DADE PD

Date/Time: 12/19/2009 00:38:16

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/07/2010 11:44

Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 06/20/2009 02:20:50

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/09/2009 13:53

Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:20

Agency: MIAMI-DADE PD

Date/Time: 11/20/2009 23:17:58

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/10/2009 13:36

Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:17

Agency: MIAMI-DADE PD

Date/Time: 11/20/2009 23:20:57

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/10/2009 13:36

Air Blank	0.000	23:20
Control Test	0.082	23:20
Air Blank	0.000	23:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 11/21/2009 00:48:30

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/10/2009 13:36

Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:48

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency: JACKSONVILLE SO

Date/Time: 07/08/2009 14:56:20

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 07/17/2009 14:08

Air Blank	0.000	14:55
ITP Check	0.080	14:55
Air Blank	0.000	14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency: JACKSONVILLE SO

Date/Time: 04/16/2009 08:37:36

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/30/2009 07:15

Air Blank	0.000	08:36
Control Test	0.079	08:37
Air Blank	0.000	08:37

Agency: JACKSONVILLE SO

Date/Time: 04/23/2009 06:17:25

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/30/2009 07:15

Air Blank	0.000	06:16
Control Test	0.080	06:16
Air Blank	0.000	06:17

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

Agency: NASSAU COUNTY SO

Date/Time: 06/08/2009 05:28:16

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/12/2009 15:19

Air Blank		05:27	Improper Sample
Air Blank	0.000	05:28	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

Agency: UNIVERSITY OF FL PD

Date/Time: 09/21/2009 14:46:56

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/30/2009 14:40

Air Blank	0.000	14:45
Control Test	0.058	14:46
Air Blank	0.000	14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

Agency: HIGH SPRINGS PD

Date/Time: 10/01/2009 09:35:00

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/01/2009 10:01

Air Blank	0.000	09:34
-----------	-------	-------

Control Test	0.069	09:34
--------------	-------	-------

Air Blank	0.000	09:35
-----------	-------	-------

Agency: HIGH SPRINGS PD

Date/Time: 10/01/2009 09:49:03

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/01/2009 10:01

Air Blank	0.000	09:47
-----------	-------	-------

Control Test	0.000	09:48
--------------	-------	-------

Air Blank	0.000	09:49
-----------	-------	-------

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

Agency: GILCHRIST COUNTY SO

Date/Time: 03/25/2009 08:52:49

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/20/2009 15:23

Air Blank	0.000	08:52
Control Test	0.079	08:52
Air Blank	0.000	08:52

Agency: GILCHRIST COUNTY SO

Date/Time: 04/20/2009 14:31:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/20/2009 15:23

Air Blank	0.000	14:30
Control Test	0.189	14:30
Air Blank	0.000	14:31

Agency: GILCHRIST COUNTY SO

Date/Time: 04/20/2009 14:33:06

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/20/2009 15:23

Air Blank	0.000	14:31
Control Test	0.191	14:32
Air Blank	0.000	14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

Agency: GILCHRIST COUNTY SO

Date/Time: 04/20/2009 14:34:25

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/20/2009 15:23

Air Blank	0.000	14:33
Control Test	0.079	14:33
Air Blank	0.000	14:34

Agency: GILCHRIST COUNTY SO

Date/Time: 04/20/2009 14:36:20

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/20/2009 15:23

Air Blank	0.000	14:35
Control Test	0.075	14:35
Air Blank	0.000	14:36

Agency: GILCHRIST COUNTY SO

Date/Time: 04/20/2009 14:38:48

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/20/2009 15:23

Air Blank	0.000	14:37
Control Test	0.046	14:38
Air Blank	0.000	14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

Agency: GILCHRIST COUNTY SO

Date/Time: 07/31/2009 21:35:25

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/01/2009 00:26

Air Blank	0.000	21:34
Control Test	0.000	21:34
Air Blank	0.000	21:35

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

Agency: GILCHRIST COUNTY SO

Date/Time: 05/15/2009 07:38:32

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/15/2009 08:20

Air Blank	0.000	07:37
Control Test	0.047	07:38
Air Blank	0.000	07:38

Agency: GILCHRIST COUNTY SO

Date/Time: 05/15/2009 07:40:05

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/15/2009 08:20

Air Blank	0.000	07:39
Control Test	0.078	07:39
Air Blank	0.000	07:40

Agency: GILCHRIST COUNTY SO

Date/Time: 06/17/2009 09:15:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/17/2009 09:57

Air Blank	0.000	09:14
Control Test	0.129	09:15
Air Blank	0.000	09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

Agency: GILCHRIST COUNTY SO

Date/Time: 07/31/2009 21:40:43

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/01/2009 00:28

Air Blank	0.000	21:39
Control Test	0.000	21:40
Air Blank	0.000	21:40

Agency: GILCHRIST COUNTY SO

Date/Time: 07/31/2009 22:03:01

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/01/2009 00:28

Air Blank	0.000	22:01
Control Test	0.000	22:02
Air Blank	0.000	22:03

Agency: GILCHRIST COUNTY SO

Date/Time: 07/31/2009 22:06:42

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/01/2009 00:28

Air Blank	0.000	22:05
Control Test	0.199	22:06
Air Blank	0.000	22:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

Agency: GILCHRIST COUNTY SO

Date/Time: 07/31/2009 22:08:37

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/01/2009 00:28

Air Blank	0.000	22:07
Control Test	0.000	22:08
Air Blank	0.000	22:08

Agency: GILCHRIST COUNTY SO

Date/Time: 07/31/2009 22:18:44

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/01/2009 00:28

Air Blank	0.000	22:17
Control Test	0.000	22:18
Air Blank	0.000	22:18

Agency: GILCHRIST COUNTY SO

Date/Time: 07/31/2009 22:23:41

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/01/2009 00:28

Air Blank	0.000	22:22
Control Test	0.000	22:23
Air Blank	0.000	22:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

Agency: GILCHRIST COUNTY SO

Date/Time: 07/31/2009 22:42:58

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/01/2009 00:28

Air Blank	0.000	22:41
Control Test	0.000	22:42
Air Blank	0.000	22:42

Agency: GILCHRIST COUNTY SO

Date/Time: 07/31/2009 23:05:44

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/01/2009 00:28

Air Blank	0.000	23:04
Control Test	0.200	23:05
Air Blank	0.000	23:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001286

Agency: BAKER COUNTY SO

Date/Time: 05/29/2009 15:31:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/16/2009 13:04

Air Blank	0.000	15:30
Control Test	0.019	15:30
Air Blank	0.000	15:31

Agency: BAKER COUNTY SO

Date/Time: 05/29/2009 15:44:13

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/16/2009 13:04

Air Blank	0.000	15:43
Control Test	0.018	15:43
Air Blank	0.000	15:44

Agency: BAKER COUNTY SO

Date/Time: 05/29/2009 15:48:16

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/16/2009 13:04

Air Blank	0.000	15:47
Control Test	0.196	15:47
Air Blank	0.000	15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

Agency: UNION COUNTY SO

Date/Time: 08/26/2009 09:53:41

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/26/2009 15:19

Air Blank	0.000	09:52	
Control Test	0.017	09:53	Interferent Detect
Air Blank	0.000	09:53	

Agency: UNION COUNTY SO

Date/Time: 08/26/2009 09:59:10

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/26/2009 15:19

Air Blank	0.000	09:57	
Control Test	0.077	09:58	
Air Blank	0.000	09:59	

Agency: UNION COUNTY SO

Date/Time: 08/26/2009 10:01:53

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/26/2009 15:19

Air Blank	0.000	10:00	
Control Test	0.048	10:01	
Air Blank	0.000	10:01	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

Agency: UNION COUNTY SO

Date/Time: 09/25/2009 09:56:52

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/25/2009 14:36

Air Blank	0.000	09:55
Control Test	0.071	09:56
Air Blank	0.000	09:56

Agency: UNION COUNTY SO

Date/Time: 09/25/2009 09:59:38

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/25/2009 14:36

Air Blank	0.000	09:58	
Control Test	0.015	09:59	Interferent Detect
Air Blank	0.000	09:59	

Agency: UNION COUNTY SO

Date/Time: 09/25/2009 10:03:13

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/25/2009 14:36

Air Blank	0.000	10:02
Control Test	0.047	10:02
Air Blank	0.000	10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

Agency: UNION COUNTY SO

Date/Time: 09/25/2009 10:05:42

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/25/2009 14:36

Air Blank	0.000	10:04
Control Test	0.192	10:05
Air Blank	0.000	10:05

Agency: UNION COUNTY SO

Date/Time: 09/25/2009 10:09:00

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/25/2009 14:36

Air Blank	0.000	10:07
Control Test	0.076	10:08
Air Blank	0.000	10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 09/18/2009 16:28:34

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 09/25/2009 10:46

Air Blank	0.000	16:27
Control Test	0.078	16:28
Air Blank	0.000	16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

Agency: BAY COUNTY SO

Date/Time: 12/10/2009 11:53:17

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 12/11/2009 10:43

Air Blank	0.000	11:52
-----------	-------	-------

Control Test	0.081	11:52
--------------	-------	-------

Air Blank	0.000	11:53
-----------	-------	-------

Agency: BAY COUNTY SO

Date/Time: 12/11/2009 08:58:11

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/11/2009 10:43

Air Blank	0.000	08:56
-----------	-------	-------

Control Test	0.207	08:57
--------------	-------	-------

Air Blank	0.000	08:58
-----------	-------	-------

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

Agency: DESOTO COUNTY SO

Date/Time: 05/20/2009 23:34:46

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/22/2009 23:42

Air Blank	0.000	23:33	
Control Test	0.000	23:34	Interferent Detect
Air Blank	0.000	23:34	

Agency: DESOTO COUNTY SO

Date/Time: 12/29/2009 10:40:36

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/02/2010 19:12

Air Blank	0.000	10:39	
Control Test	0.078	10:40	
Air Blank	0.000	10:40	

Agency: DESOTO COUNTY SO

Date/Time: 12/29/2009 10:42:26

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/02/2010 19:12

Air Blank	0.000	10:41	
Control Test	0.078	10:41	
Air Blank	0.000	10:42	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

Agency: DESOTO COUNTY SO

Date/Time: 02/24/2009 19:18:49

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/28/2009 17:07

Air Blank	0.000	19:18
Control Test	0.080	19:18
Air Blank	0.000	19:18

Agency: DESOTO COUNTY SO

Date/Time: 05/21/2009 23:18:42

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/22/2009 23:48

Air Blank	0.000	23:17	
Control Test	0.000	23:18	Interferent Detect
Air Blank	0.000	23:18	

Agency: DESOTO COUNTY SO

Date/Time: 05/21/2009 23:20:26

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/22/2009 23:48

Air Blank	0.000	23:19
Control Test	0.051	23:19
Air Blank	0.000	23:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

Agency: DESOTO COUNTY SO

Date/Time: 05/21/2009 23:22:08

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/22/2009 23:48

Air Blank	0.000	23:20
-----------	-------	-------

Control Test	0.079	23:21
--------------	-------	-------

Air Blank	0.000	23:22
-----------	-------	-------

Agency: DESOTO COUNTY SO

Date/Time: 05/21/2009 23:23:52

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/22/2009 23:48

Air Blank	0.000	23:22
-----------	-------	-------

Control Test	0.201	23:23
--------------	-------	-------

Air Blank	0.000	23:23
-----------	-------	-------

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

Agency: SARASOTA COUNTY SO

Date/Time: 03/04/2009 13:53:24

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/21/2009 00:03

Air Blank	0.000	13:52
Control Test	0.078	13:52
Air Blank	0.000	13:53

Agency: SARASOTA COUNTY SO

Date/Time: 06/03/2009 15:52:06

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/15/2009 10:13

Air Blank	0.000	15:51
Control Test	0.078	15:51
Air Blank	0.000	15:52

Agency: SARASOTA COUNTY SO

Date/Time: 07/16/2009 09:40:56

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/16/2009 10:47

Air Blank	0.000	09:39
Control Test	0.078	09:40
Air Blank	0.000	09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

Agency: SARASOTA COUNTY SO

Date/Time: 09/01/2009 10:47:21

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 09/03/2009 11:22

Air Blank	0.000	10:46
ITP Check	0.081	10:46
Air Blank	0.000	10:47

Agency: SARASOTA COUNTY SO

Date/Time: 09/01/2009 10:49:01

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 09/03/2009 11:22

Air Blank	0.000	10:48
Control Test	0.078	10:48
Air Blank	0.000	10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency: SARASOTA COUNTY SO

Date/Time: 01/09/2009 21:06:34

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/23/2009 13:16

Air Blank	0.000	21:05
Control Test	0.080	21:06
Air Blank	0.000	21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY SO

Date/Time: 02/08/2009 08:45:40

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/08/2009 09:53

Air Blank	0.000	08:44	
Control Test	0.000	08:45	Interferent Detect
Air Blank	0.000	08:45	

Agency: CHARLOTTE COUNTY SO

Date/Time: 02/08/2009 08:47:51

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/08/2009 09:53

Air Blank	0.000	08:46	
Control Test	0.048	08:47	
Air Blank	0.000	08:47	

Agency: CHARLOTTE COUNTY SO

Date/Time: 02/08/2009 08:50:13

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/08/2009 09:53

Air Blank	0.000	08:49	
Control Test	0.079	08:49	
Air Blank	0.000	08:50	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY SO

Date/Time: 02/08/2009 08:52:24

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/08/2009 09:53

Air Blank	0.000	08:51
Control Test	0.196	08:51
Air Blank	0.000	08:52

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/11/2009 07:19:31

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/11/2009 10:52

Air Blank	0.000	07:18
Control Test	0.000	07:19
Air Blank	0.000	07:19

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/11/2009 07:22:44

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/11/2009 10:52

Air Blank	0.000	07:21
Control Test	0.046	07:22
Air Blank	0.000	07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/11/2009 07:31:24

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/11/2009 10:52

Air Blank	0.000	07:30
Control Test	0.058	07:30
Air Blank	0.000	07:31

Agency: CHARLOTTE COUNTY SO

Date/Time: 08/09/2009 08:45:04

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/09/2009 10:37

Air Blank	0.000	08:43	
Control Test	0.000	08:44	Interferent Detect
Air Blank	0.000	08:45	

Agency: CHARLOTTE COUNTY SO

Date/Time: 10/21/2009 10:03:49

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/21/2009 11:41

Air Blank	0.000	10:02	
Control Test	0.000	10:03	Interferent Detect
Air Blank	0.000	10:03	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency: LEVY COUNTY SO

Date/Time: 01/11/2009 18:57:26

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/31/2009 04:41

Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:57

Agency: LEVY COUNTY SO

Date/Time: 06/24/2009 13:59:16

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/30/2009 00:08

Air Blank	0.000	13:58
Control Test	0.000	13:58
Air Blank	0.000	13:59

Agency: LEVY COUNTY SO

Date/Time: 06/24/2009 14:01:07

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/30/2009 00:08

Air Blank	0.000	13:59
Control Test	0.000	14:00
Air Blank	0.000	14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency: LEVY COUNTY SO

Date/Time: 06/24/2009 15:29:40

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/30/2009 00:08

Air Blank	0.000	15:28
Control Test	0.000	15:29
Air Blank	0.000	15:29

Agency: LEVY COUNTY SO

Date/Time: 06/24/2009 15:31:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/30/2009 00:08

Air Blank	0.000	15:30
Control Test	0.000	15:30
Air Blank	0.000	15:31

Agency: LEVY COUNTY SO

Date/Time: 06/24/2009 15:58:02

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/30/2009 00:08

Air Blank	0.000	15:56	
Control Test	0.000	15:57	Interferent Detect
Air Blank	0.000	15:58	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency: LEVY COUNTY SO

Date/Time: 06/24/2009 15:59:47

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/30/2009 00:08

Air Blank	0.000	15:58	
Control Test	0.000	15:59	Interferent Detect
Air Blank	0.000	15:59	

Agency: LEVY COUNTY SO

Date/Time: 07/31/2009 22:48:20

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/30/2009 20:57

Air Blank	0.000	22:47	
Control Test	0.000	22:47	
Air Blank	0.000	22:48	

Agency: LEVY COUNTY SO

Date/Time: 07/31/2009 23:06:11

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/30/2009 20:57

Air Blank	0.000	23:04	
Control Test	0.000	23:05	Interferent Detect
Air Blank	0.000	23:06	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency: ORANGE COUNTY S.O.

Date/Time: 09/28/2009 08:45:09

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 10/01/2009 09:22

Air Blank	0.000	08:44
Control Test	0.000	08:44
Air Blank	0.000	08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001441

Agency: KENNEDY SPACE CENTER

Date/Time: 06/18/2009 17:10:41

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 06/19/2009 10:50

Air Blank		17:10	Sequence Aborted
Air Blank	0.000	17:10	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001469

Agency: HURLBURT FIELD PD

Date/Time: 01/12/2009 23:01:15

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/30/2009 14:11

Air Blank	0.000	23:00
Control Test	0.076	23:00
Air Blank	0.000	23:01

Agency: HURLBURT FIELD PD

Date/Time: 01/18/2009 06:49:57

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/30/2009 14:11

Air Blank	0.000	06:49
Control Test	0.076	06:49
Air Blank	0.000	06:49

Agency: HURLBURT FIELD PD

Date/Time: 01/30/2009 11:59:47

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/30/2009 14:11

Air Blank	0.000	11:58
Control Test	0.077	11:59
Air Blank	0.000	11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001469

Agency: HURLBURT FIELD PD

Date/Time: 03/03/2009 16:14:08

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/03/2009 18:25

Air Blank	0.000	16:13
Control Test	0.076	16:13
Air Blank	0.000	16:14

Agency: HURLBURT FIELD PD

Date/Time: 03/05/2009 14:11:38

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/02/2009 16:19

Air Blank	0.000	14:10
Control Test	0.075	14:11
Air Blank	0.000	14:11

Agency: HURLBURT FIELD PD

Date/Time: 03/16/2009 10:33:31

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/02/2009 16:19

Air Blank	0.000	10:32
Control Test	0.076	10:33
Air Blank	0.000	10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency: BAY COUNTY SO

Date/Time: 02/15/2009 22:19:47

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/23/2009 15:59

Air Blank	0.000	22:18
Control Test	0.075	22:19
Air Blank	0.000	22:19

Agency: BAY COUNTY SO

Date/Time: 02/15/2009 22:21:57

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/23/2009 15:59

Air Blank	0.000	22:21
Control Test	0.076	22:21
Air Blank	0.000	22:21

Agency: BAY COUNTY SO

Date/Time: 02/15/2009 22:23:53

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/23/2009 15:59

Air Blank	0.000	22:23
Control Test	0.077	22:23
Air Blank	0.000	22:23

Florida Department of Law Enforcement Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency: COLLIER COUNTY SO

Date/Time: 04/01/2009 08:43:04

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/07/2009 08:51

Air Blank	0.000	08:41
Control Test	0.000	08:42
Air Blank	0.000	08:43

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency: OSCEOLA CORRECTIONS

Date/Time: 07/20/2009 12:18:11

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 07/20/2009 15:04

Air Blank	0.000	12:16	
ITP Check	0.081	12:17	
Air Blank		12:17	Sequence Aborted
Air Blank	0.000	12:18	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency: PALM BAY P.D.

Date/Time: 01/07/2009 07:22:55

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/08/2009 10:45

Air Blank	0.000	07:22
Control Test	0.081	07:22
Air Blank	0.000	07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY SO

Date/Time: 04/08/2009 09:10:23

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/08/2009 10:51

Air Blank	0.000	09:09	
Control Test	0.007	09:09	Interferent Detect
Air Blank	0.000	09:10	

Agency: CHARLOTTE COUNTY SO

Date/Time: 04/23/2009 17:09:17

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 05/21/2009 11:08

Air Blank	0.000	17:08	
ITP Check	0.081	17:08	
Air Blank	0.000	17:09	

Agency: CHARLOTTE COUNTY SO

Date/Time: 07/02/2009 08:09:08

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 07/02/2009 09:56

Air Blank	0.000	08:07	
Control Test	0.000	08:08	Interferent Detect
Air Blank	0.000	08:09	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 04/16/2009 16:18:01

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 04/16/2009 17:52

Air Blank	0.000	16:17
Control Test	0.082	16:17
Air Blank	0.000	16:18

Agency: CLEARWATER PD

Date/Time: 04/16/2009 16:20:21

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/16/2009 17:52

Air Blank	0.000	16:19
Control Test	0.050	16:19
Air Blank	0.000	16:20

Agency: CLEARWATER PD

Date/Time: 04/16/2009 16:22:11

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/16/2009 17:52

Air Blank	0.000	16:20
Control Test	0.077	16:21
Air Blank	0.000	16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 04/16/2009 16:23:56

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/16/2009 17:52

Air Blank	0.000	16:22
Control Test	0.203	16:23
Air Blank	0.000	16:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency: PALM BEACH CO SO

Date/Time: 11/14/2009 15:09:21

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/20/2009 11:25

Air Blank	0.000	15:08
Control Test	0.078	15:08
Air Blank	0.000	15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency: BRADENTON PD

Date/Time: 06/17/2009 11:19:34

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/17/2009 11:25

Air Blank	0.000	11:18
Control Test	0.078	11:19
Air Blank	0.000	11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:07:39

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank	0.000	09:06
Control Test	0.080	09:07
Air Blank	0.000	09:07

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:10:55

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank	0.000	09:10
Control Test	0.081	09:10
Air Blank	0.000	09:10

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:13:49

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank	0.000	09:13
Control Test	0.081	09:13
Air Blank	0.000	09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:18:16

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank	0.000	09:17
Control Test	0.081	09:17
Air Blank	0.000	09:18

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:32:42

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank	0.000	09:31
Control Test	0.081	09:32
Air Blank	0.000	09:32

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:34:20

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank	0.000	09:33
Control Test	0.082	09:33
Air Blank	0.000	09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:35:47

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank	0.000	09:34
Control Test	0.082	09:35
Air Blank	0.000	09:35

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:37:03

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank		09:36	Ambient Fail
Air Blank		09:37	Purge Fail

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:38:27

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank	0.000	09:37
Control Test	0.082	09:37
Air Blank	0.000	09:38

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:41:12

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank	0.000	09:40
Control Test	0.082	09:40
Air Blank	0.000	09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

Agency: LAKELAND PD

Date/Time: 07/13/2009 09:42:53

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 08/09/2009 08:32

Air Blank	0.000	09:42
Control Test	0.082	09:42
Air Blank	0.000	09:42

Agency: LAKELAND PD

Date/Time: 10/04/2009 07:15:44

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/04/2009 08:03

Air Blank	0.000	07:14	
Control Test	0.000	07:15	Interferent Detect
Air Blank	0.000	07:15	

Agency: LAKELAND PD

Date/Time: 10/04/2009 07:18:01

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/04/2009 08:03

Air Blank	0.000	07:16
Control Test	0.049	07:17
Air Blank	0.000	07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 06/18/2009 10:04:20

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/18/2009 11:25

Air Blank		10:03	Sequence Aborted
-----------	--	-------	------------------

Air Blank	0.000	10:04	
-----------	-------	-------	--

Agency: US AIR FORCE MACDILL

Date/Time: 07/29/2009 10:34:34

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 08/12/2009 10:58

Air Blank	0.000	10:33	
-----------	-------	-------	--

Control Test	0.186	10:34	
--------------	-------	-------	--

Air Blank	0.000	10:34	
-----------	-------	-------	--

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency: NAS JACKSONVILLE PD

Date/Time: 01/20/2009 12:27:37

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 01/24/2009 00:10

Air Blank	0.000	12:26
Control Test	0.000	12:27
Air Blank	0.000	12:27

Agency: NAS JACKSONVILLE PD

Date/Time: 01/20/2009 12:29:45

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/24/2009 00:10

Air Blank	0.000	12:28	
Control Test	0.000	12:29	Interferent Detect
Air Blank	0.000	12:29	

Agency: NAS JACKSONVILLE PD

Date/Time: 01/20/2009 12:31:43

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 01/24/2009 00:10

Air Blank	0.000	12:30
Control Test	0.196	12:31
Air Blank	0.000	12:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency: NAS JACKSONVILLE PD

Date/Time: 02/17/2009 13:31:14

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/17/2009 14:14

Air Blank	0.000	13:30	
Control Test	0.008	13:30	Interferent Detect
Air Blank	0.000	13:31	

Agency: NAS JACKSONVILLE PD

Date/Time: 02/17/2009 13:34:07

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 02/17/2009 14:14

Air Blank	0.000	13:32	
Control Test	0.194	13:33	
Air Blank	0.000	13:34	

Agency: NAS JACKSONVILLE PD

Date/Time: 03/17/2009 12:53:11

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/17/2009 13:37

Air Blank	0.000	12:51	
Control Test	0.000	12:52	Interferent Detect
Air Blank	0.000	12:53	

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency: NAS JACKSONVILLE PD

Date/Time: 03/17/2009 12:54:56

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/17/2009 13:37

Air Blank	0.000	12:53
Control Test	0.190	12:54
Air Blank	0.000	12:54

Agency: NAS JACKSONVILLE PD

Date/Time: 03/17/2009 12:57:06

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 03/17/2009 13:37

Air Blank	0.000	12:55
Control Test	0.192	12:56
Air Blank	0.000	12:57

Agency: NAS JACKSONVILLE PD

Date/Time: 04/23/2009 08:45:01

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 09:29

Air Blank	0.000	08:43	
Control Test	0.000	08:44	Interferent Detect
Air Blank	0.000	08:45	

Florida Department of Law Enforcement Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency: NAS JACKSONVILLE PD

Date/Time: 04/23/2009 08:46:52

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 04/23/2009 09:29

Air Blank	0.000	08:45
Control Test	0.191	08:46
Air Blank	0.000	08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Agency: NAS KEY WEST

Date/Time: 11/12/2009 11:29:28

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/12/2009 12:14

Air Blank	0.000	11:28
Control Test	0.084	11:28
Air Blank	0.000	11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency: MIAMI BEACH PD

Date/Time: 02/26/2009 09:25:15

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/19/2009 16:10

Air Blank	0.000	09:24
-----------	-------	-------

Control Test	0.078	09:24
--------------	-------	-------

Air Blank	0.000	09:25
-----------	-------	-------

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Agency: FFWCC CARRABELLE

Date/Time: 10/23/2009 11:28:57

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/23/2009 11:45

Air Blank	0.000	11:27
Control Test	0.193	11:28
Air Blank	0.000	11:28

Agency: FFWCC CARRABELLE

Date/Time: 10/23/2009 11:31:08

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/23/2009 11:45

Air Blank	0.000	11:29
Control Test	0.195	11:30
Air Blank	0.000	11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency: MIAMI BEACH PD

Date/Time: 02/12/2009 10:26:39

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 02/26/2009 08:44

Air Blank	0.000	10:25
Control Test	0.079	10:26
Air Blank	0.000	10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency: PASCO COUNTY SO

Date/Time: 04/11/2009 03:49:22

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/11/2009 08:39

Air Blank	0.000	03:48
Control Test	0.070	03:48
Air Blank	0.000	03:49

Agency: PASCO COUNTY SO

Date/Time: 04/11/2009 03:50:42

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 05/11/2009 08:39

Air Blank	0.000	03:49
Control Test	0.072	03:50
Air Blank	0.000	03:50

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency: HILLSBOROUGH CO SO

Date/Time: 06/04/2009 07:04:01

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/04/2009 13:59

Air Blank	0.000	07:03
Control Test	0.077	07:03
Air Blank	0.000	07:04

Agency: HILLSBOROUGH CO SO

Date/Time: 06/04/2009 07:07:22

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/04/2009 13:59

Air Blank	0.000	07:06
Control Test	0.077	07:06
Air Blank	0.000	07:07

Agency: HILLSBOROUGH CO SO

Date/Time: 06/04/2009 07:08:41

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/04/2009 13:59

Air Blank	0.000	07:07
Control Test	0.077	07:08
Air Blank	0.000	07:08

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency: HILLSBOROUGH CO SO

Date/Time: 06/04/2009 07:10:12

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/04/2009 13:59

Air Blank	0.000	07:09
Control Test	0.077	07:09
Air Blank	0.000	07:10

Agency: HILLSBOROUGH CO SO

Date/Time: 09/24/2009 13:17:01

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/25/2009 11:38

Air Blank	0.000	13:15
Control Test	0.195	13:16
Air Blank	0.000	13:17

Agency: HILLSBOROUGH CO SO

Date/Time: 09/24/2009 13:21:59

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 09/25/2009 11:38

Air Blank	0.000	13:20
Control Test	0.199	13:21
Air Blank	0.000	13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency: HILLSBOROUGH CO SO

Date/Time: 11/24/2009 11:32:40

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 11/24/2009 12:41

Air Blank	0.000	11:31
Control Test	0.196	11:32
Air Blank	0.000	11:32

Agency: HILLSBOROUGH CO SO

Date/Time: 12/15/2009 11:23:58

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 12/15/2009 12:07

Air Blank	0.000	11:22
Control Test	0.199	11:23
Air Blank	0.000	11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency: HILLSBOROUGH CO SO

Date/Time: 06/04/2009 07:16:35

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/04/2009 14:02

Air Blank	0.000	07:15
Control Test	0.076	07:16
Air Blank	0.000	07:16

Agency: HILLSBOROUGH CO SO

Date/Time: 06/04/2009 07:21:52

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/04/2009 14:02

Air Blank	0.000	07:21
Control Test	0.076	07:21
Air Blank	0.000	07:21

Agency: HILLSBOROUGH CO SO

Date/Time: 06/04/2009 07:23:58

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/04/2009 14:02

Air Blank	0.000	07:23
Control Test	0.077	07:23
Air Blank	0.000	07:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency: HILLSBOROUGH CO SO

Date/Time: 06/04/2009 07:25:18

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 06/04/2009 14:02

Air Blank	0.000	07:24
Control Test	0.076	07:24
Air Blank	0.000	07:25

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency: FL HIGHWAY PATROL

Date/Time: 07/10/2009 11:53:39

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 07/14/2009 12:56

Air Blank	0.000	11:52
Control Test	0.074	11:53
Air Blank	0.000	11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency: FL HIGHWAY PATROL

Date/Time: 05/27/2009 09:44:15

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/27/2009 11:41

Air Blank	0.000	09:43
Control Test	0.050	09:43
Air Blank	0.000	09:44

Agency: FL HIGHWAY PATROL

Date/Time: 05/27/2009 09:46:25

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/27/2009 11:41

Air Blank	0.000	09:45
Control Test	0.050	09:45
Air Blank	0.000	09:46

Agency: FL HIGHWAY PATROL

Date/Time: 05/27/2009 09:48:36

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/27/2009 11:41

Air Blank	0.000	09:47
Control Test	0.051	09:48
Air Blank	0.000	09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency: FL HIGHWAY PATROL

Date/Time: 05/27/2009 09:52:03

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/27/2009 11:41

Air Blank	0.000	09:50
Control Test	0.071	09:51
Air Blank	0.000	09:52

Agency: FL HIGHWAY PATROL

Date/Time: 05/27/2009 09:54:27

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/27/2009 11:41

Air Blank	0.000	09:53
Control Test	0.075	09:53
Air Blank	0.000	09:54

Agency: FL HIGHWAY PATROL

Date/Time: 05/27/2009 09:57:22

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/27/2009 11:41

Air Blank	0.000	09:56
Control Test	0.076	09:56
Air Blank	0.000	09:57

Florida Department of Law Enforcement Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency: FL HIGHWAY PATROL

Date/Time: 05/27/2009 10:01:16

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 05/27/2009 11:41

Air Blank	0.000	10:00
Control Test	0.077	10:00
Air Blank	0.000	10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency: LAKELAND PD

Date/Time: 08/17/2009 11:01:43

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 09/20/2009 07:59

Air Blank	0.000	11:00
Control Test	0.078	11:01
Air Blank	0.000	11:01

Agency: LAKELAND PD

Date/Time: 10/04/2009 08:20:12

Standard Type: Wet bath

Event Source: Local

Data Download Date/Time: 10/04/2009 09:03

Air Blank	0.000	08:18
Control Test	0.049	08:19
Air Blank	0.000	08:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency: HURLBURT FIELD PD

Date/Time: 10/30/2009 13:27:47

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 11/02/2009 09:52

Air Blank	0.000	13:26
Control Test	0.084	13:27
Air Blank	0.000	13:27

Agency: HURLBURT FIELD PD

Date/Time: 12/30/2009 08:56:18

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/22/2010 18:32

Air Blank	0.000	08:55
Control Test	0.085	08:55
Air Blank	0.000	08:56

Agency: HURLBURT FIELD PD

Date/Time: 12/30/2009 09:09:37

Standard Type: Dry gas

Event Source: Local

Data Download Date/Time: 03/22/2010 18:32

Air Blank	0.000	09:08
Control Test	0.085	09:09
Air Blank	0.000	09:09
