

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: HUGHES, P

Arresting Officer Agency: Sanford PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:18

Gender: Male

Observation Period Began: 01:46

Subject's Last Name: GASPAR

Date/Time of Test: 04/01/2007 02:17

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample #1	0.232	02:11
Breath Volume	2.269	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample #2	0.216	02:15
Breath Volume	2.945	02:15
Air Blank	0.000	02:16
Control Test	0.076	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: FUENTES, VICTOR

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 02:07

Gender: Female

Observation Period Began: 02:47

Subject's Last Name: TILLIS

Date/Time of Test: 04/01/2007 03:16

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Subject Sample #1	0.321	03:09
Breath Volume	2.769	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample #2	0.295	03:11
Breath Volume	1.191	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample #3	0.297	03:14
Breath Volume	1.582	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: ALBRECHT, C

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 02:43

Gender: Male

Observation Period Began: 03:21

Subject's Last Name: MERRITT

Date/Time of Test: 04/01/2007 03:47

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample #1	0.115	03:42
Breath Volume	2.378	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample #2	0.112	03:45
Breath Volume	3.488	03:45
Air Blank	0.000	03:46
Control Test	0.077	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: PICKLESIMER, V

Arresting Officer Agency: Lake Mary PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 03:46

Gender: Male

Observation Period Began: 04:14

Subject's Last Name: FUENTES

Date/Time of Test: 04/01/2007 04:40

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:35
Subject Sample #1	0.138	04:36
Breath Volume	3.203	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:38
Subject Sample #2	0.142	04:39
Breath Volume	2.480	04:39
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: CHRISTIANSON, M

Arresting Officer Agency: Winter Springs PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 19:53

Gender: Male

Observation Period Began: 20:07

Subject's Last Name: SCHANBACHER

Date/Time of Test: 04/01/2007 20:42

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:37
Subject Sample #1	0.101	20:38
Breath Volume	2.625	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample #2	0.102	20:41
Breath Volume	2.250	20:41
Air Blank	0.000	20:41
Control Test	0.080	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: VAN CLEAVE, A

Arresting Officer Agency: Altamonte Springs PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 20:10

Gender: Male

Observation Period Began: 21:25

Subject's Last Name: SANDERSON

Date/Time of Test: 04/01/2007 22:01

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:55
Air Blank	0.000	21:56
Subject Sample #1	0.061	21:57
Breath Volume	3.246	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample #2	0.059	22:00
Breath Volume	3.546	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MARCARTONI, W

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 21:10

Gender: Male

Observation Period Began: 21:20

Subject's Last Name: GIOVANELLI

Date/Time of Test: 04/04/2007 21:59

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.079	21:53
Air Blank	0.000	21:54
Subject Sample #1	0.067	21:55
Breath Volume	2.070	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample #2	0.067	21:57
Breath Volume	2.464	21:57
Air Blank	0.000	21:58
Control Test	0.079	21:58
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: FRY, GENE

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 23:11

Gender: Male

Observation Period Began: 23:50

Subject's Last Name: DAVIS

Date/Time of Test: 04/05/2007 00:42

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample #1	0.140	00:37
Breath Volume	2.015	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample #2	0.159	00:40
Breath Volume	2.191	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: PIORRO, C

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 20:50

Gender: Male

Observation Period Began: 21:16

Subject's Last Name: BRACE

Date/Time of Test: 04/05/2007 21:45

Diagnostics Check	OK	21:39
Air Blank	0.000	21:40
Control Test	0.080	21:40
Air Blank	0.000	21:40
Subject Sample #1	0.151	21:41
Breath Volume	2.105	21:41
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample #2	0.141	21:44
Breath Volume	1.820	21:44
Air Blank	0.000	21:44
Control Test	0.079	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: BURNETT, TRP

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 21:16

Gender: Male

Observation Period Began: 22:24

Subject's Last Name: GUSTUS

Date/Time of Test: 04/05/2007 22:51

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Subject Sample #1	0.069	22:46
Breath Volume	2.652	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample #2	0.063	22:50
Breath Volume	1.664	22:50
Air Blank	0.000	22:50
Control Test	0.077	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: WALKER, MATTHEW

Arresting Officer Agency: Sanford PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:26

Gender: Male

Observation Period Began: 02:06

Subject's Last Name: GODWIN

Date/Time of Test: 04/06/2007 02:32

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample #1	0.164	02:28
Breath Volume	1.191	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample #2	0.183	02:31
Breath Volume	2.667	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: PICKLESIMER, VIRGIL

Arresting Officer Agency: Lake Mary PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:59

Gender: Male

Observation Period Began: 02:49

Subject's Last Name: MCGRADY

Date/Time of Test: 04/06/2007 03:15

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample #1	0.033	03:11
Breath Volume	3.777	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample #2	0.031	03:14
Breath Volume	3.992	03:14
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: SMITH, ROBERT

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 02:44

Gender: Male

Observation Period Began: 03:12

Subject's Last Name: DOXTATER

Date/Time of Test: 04/06/2007 03:40

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample #1	0.099	03:35
Breath Volume	1.167	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample #2	0.098	03:38
Breath Volume	1.359	03:38
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: GENTRY, A

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 22:27

Gender: Female

Observation Period Began: 23:23

Subject's Last Name: DIGIOVANNI

Date/Time of Test: 04/06/2007 23:48

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample #1		23:47	Subject Test Refused
Breath Volume	0.117	23:47	
Air Blank	0.000	23:47	
Control Test	0.082	23:48	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: PICKLESIMER, VIRGIL

Arresting Officer Agency: Lake Mary PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 23:47

Gender: Male

Observation Period Began: 00:49

Subject's Last Name: MANUEL

Date/Time of Test: 04/07/2007 01:19

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample #1	0.160	01:14
Breath Volume	1.457	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample #2	0.152	01:17
Breath Volume	2.101	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MCNEIL, OFC

Arresting Officer Agency: Oviedo PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 02:07

Gender: Male

Observation Period Began: 02:51

Subject's Last Name: HARMON

Date/Time of Test: 04/07/2007 03:26

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:13	
Subject Sample #1	0.223	03:16	Volume Not Met
Breath Volume	0.050	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:18	
Subject Sample #2	0.212	03:21	
Breath Volume	1.308	03:21	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample #3	0.225	03:24	
Breath Volume	1.164	03:24	
Air Blank	0.000	03:25	
Control Test	0.075	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: NAS, JOSEPH

Arresting Officer Agency: Casselberry PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 02:23

Gender: Male

Observation Period Began: 03:04

Subject's Last Name: SHAW

Date/Time of Test: 04/07/2007 03:41

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Subject Sample #1	0.258	03:36
Breath Volume	2.351	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample #2	0.257	03:39
Breath Volume	1.527	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: RHODES, JAMES

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 04:19

Gender: Male

Observation Period Began: 04:40

Subject's Last Name: HENYECZ

Date/Time of Test: 04/07/2007 05:23

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.079	05:18
Air Blank	0.000	05:18
Subject Sample #1	0.173	05:19
Breath Volume	3.253	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:21
Subject Sample #2	0.171	05:22
Breath Volume	3.417	05:22
Air Blank	0.000	05:22
Control Test	0.078	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MCINTOSH, BRIAN

Arresting Officer Agency: Sanford PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 23:05

Gender: Male

Observation Period Began: 23:44

Subject's Last Name: PEREZ

Date/Time of Test: 04/08/2007 00:11

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:06
Subject Sample #1	0.234	00:06
Breath Volume	2.253	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample #2	0.234	00:09
Breath Volume	2.250	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: SERAFIN, L

Arresting Officer Agency: Oviedo PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:35

Gender: Male

Observation Period Began: 02:02

Subject's Last Name: COLON

Date/Time of Test: 04/08/2007 02:28

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.082	02:23
Air Blank	0.000	02:23
Subject Sample #1	0.254	02:24
Breath Volume	1.953	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample #2	0.263	02:26
Breath Volume	1.539	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: BOLTON, JOHN

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 18:30

Gender: Male

Observation Period Began: 19:10

Subject's Last Name: DEITCH

Date/Time of Test: 04/09/2007 19:44

Diagnostics Check	OK	19:37
Air Blank	0.000	19:38
Control Test	0.077	19:38
Air Blank	0.000	19:39
Subject Sample #1	0.000	19:39
Breath Volume	2.886	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample #2	0.000	19:42
Breath Volume	2.773	19:42
Air Blank	0.000	19:43
Control Test	0.079	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: ERWIN, WILLIAM

Arresting Officer Agency: Sanford PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 18:55

Gender: Male

Observation Period Began: 19:19

Subject's Last Name: BARANSKI

Date/Time of Test: 04/09/2007 20:41

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.080	20:34
Air Blank	0.000	20:35
Subject Sample #1	0.203	20:36
Breath Volume	2.671	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample #2	0.194	20:39
Breath Volume	2.113	20:39
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: NAS, J

Arresting Officer Agency: Casselberry PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 18:10

Gender: Male

Observation Period Began: 19:00

Subject's Last Name: MARCANTEL

Date/Time of Test: 04/11/2007 19:51

Diagnostics Check	OK	19:44
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:45
Subject Sample #1	0.221	19:46
Breath Volume	1.882	19:46
Air Blank	0.000	19:47
Air Blank	0.000	19:48
Subject Sample #2	0.219	19:49
Breath Volume	1.289	19:49
Air Blank	0.000	19:50
Control Test	0.080	19:50
Air Blank	0.000	19:50
Diagnostics Check	OK	19:51

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: JACOBS, D

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 18:51

Gender: Female

Observation Period Began: 19:20

Subject's Last Name: KUMM

Date/Time of Test: 04/11/2007 20:12

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:11	
Control Test	0.080	20:11	
Air Blank	0.000	20:11	
Subject Sample #1		20:12	Sequence Aborted
Breath Volume	0.000	20:12	
Air Blank	0.000	20:12	

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: JACOBS, D

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 18:51

Gender: Female

Observation Period Began: 19:20

Subject's Last Name: KUMM

Date/Time of Test: 04/11/2007 20:46

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.080	20:40
Air Blank	0.000	20:40
Subject Sample #1	0.111	20:41
Breath Volume	1.460	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample #2	0.111	20:45
Breath Volume	1.390	20:45
Air Blank	0.000	20:45
Control Test	0.080	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: GLENN, Z

Arresting Officer Agency: Sanford PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 20:27

Gender: Male

Observation Period Began: 21:08

Subject's Last Name: COHEN

Date/Time of Test: 04/12/2007 21:34

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.080	21:29
Air Blank	0.000	21:30
Subject Sample #1	0.204	21:30
Breath Volume	3.734	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample #2	0.196	21:33
Breath Volume	3.902	21:33
Air Blank	0.000	21:33
Control Test	0.079	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MARCANTONI, WALLACE

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 19:19

Gender: Male

Observation Period Began: 20:16

Subject's Last Name: BARTON

Date/Time of Test: 04/13/2007 20:45

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:38
Subject Sample #1	0.054	20:40
Breath Volume	1.332	20:40
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample #2	0.049	20:43
Breath Volume	1.269	20:43
Air Blank	0.000	20:44
Control Test	0.078	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: HOOVER, DANIEL, C

Breath Test Operator Agency: Altamonte Springs PD

Arresting Officer: HOOVER, DANIEL

Arresting Officer Agency: Altamonte Springs PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:13

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: HAYES

Date/Time of Test: 04/14/2007 02:08

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Subject Sample #1	0.151	02:04
Breath Volume	2.992	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample #2	0.147	02:07
Breath Volume	2.855	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: FRY, G

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 03:55

Gender: Male

Observation Period Began: 04:24

Subject's Last Name: LEON

Date/Time of Test: 04/14/2007 04:53

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:44	
Control Test	0.080	04:45	
Air Blank	0.000	04:45	
Subject Sample #1	0.134	04:48	Volume Not Met
Breath Volume	0.265	04:48	
Air Blank	0.000	04:49	
Air Blank	0.000	04:51	
Subject Sample #2		04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:52	
Control Test	0.077	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: PRICE, F

Arresting Officer Agency: Longwood PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 03:16

Gender: Male

Observation Period Began: 03:54

Subject's Last Name: MARINEZ

Date/Time of Test: 04/15/2007 04:22

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Subject Sample #1	0.101	04:17
Breath Volume	1.882	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample #2	0.096	04:20
Breath Volume	1.144	04:20
Air Blank	0.000	04:21
Control Test	0.077	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:22

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: FREEMAN, R

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 03:30

Gender: Female

Observation Period Began: 04:15

Subject's Last Name: IDROVO GALVEZ

Date/Time of Test: 04/15/2007 04:46

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:41
Subject Sample #1	0.200	04:41
Breath Volume	1.132	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample #2	0.203	04:44
Breath Volume	1.445	04:44
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: HEATH, B, W

Breath Test Operator Agency: Winter Springs PD

Arresting Officer: IKEGUCHT, D

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 14:55

Gender: Female

Observation Period Began: 16:11

Subject's Last Name: BRAYMAN

Date/Time of Test: 04/16/2007 16:45

Diagnostics Check	OK	16:38
Air Blank	0.000	16:38
Control Test	0.080	16:38
Air Blank	0.000	16:39
Subject Sample #1	0.108	16:40
Breath Volume	2.617	16:40
Air Blank	0.000	16:41
Air Blank	0.000	16:43
Subject Sample #2	0.100	16:44
Breath Volume	2.320	16:44
Air Blank	0.000	16:44
Control Test	0.079	16:45
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: SCHARZMAN, D

Arresting Officer Agency: Lake Mary PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:44

Gender: Male

Observation Period Began: 02:25

Subject's Last Name: MCKENNA

Date/Time of Test: 04/18/2007 03:17

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:10
Air Blank	0.000	03:10
Subject Sample #1	0.135	03:12
Breath Volume	1.320	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample #2	0.128	03:15
Breath Volume	1.117	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MCNEIL, B

Arresting Officer Agency: Oviedo PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 03:40

Gender: Male

Observation Period Began: 04:25

Subject's Last Name: AZNARAN

Date/Time of Test: 04/18/2007 05:10

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Subject Sample #1	0.139	05:05
Breath Volume	1.507	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample #2	0.146	05:09
Breath Volume	3.343	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MCNEIL, BRETT

Arresting Officer Agency: Oviedo PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 03:23

Gender: Male

Observation Period Began: 03:55

Subject's Last Name: AGUILAR

Date/Time of Test: 04/21/2007 04:22

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample #1	0.175	04:18
Breath Volume	2.515	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample #2	0.177	04:21
Breath Volume	2.937	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: HEATH, B, W

Breath Test Operator Agency: Winter Springs PD

Arresting Officer: MOJICA, K

Arresting Officer Agency: Oviedo PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 06:10

Gender: Female

Observation Period Began: 06:50

Subject's Last Name: FELIX

Date/Time of Test: 04/21/2007 07:19

Diagnostics Check	OK	07:13
Air Blank	0.000	07:13
Control Test	0.080	07:13
Air Blank	0.000	07:14
Subject Sample #1	0.098	07:14
Breath Volume	2.003	07:14
Air Blank	0.000	07:15
Air Blank	0.000	07:17
Subject Sample #2	0.089	07:17
Breath Volume	2.605	07:17
Air Blank	0.000	07:18
Control Test	0.079	07:19
Air Blank	0.000	07:19
Diagnostics Check	OK	07:19

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: HEATH, B, W

Breath Test Operator Agency: Winter Springs PD

Arresting Officer: NAGOWSKI, D

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 06:16

Gender: Male

Observation Period Began: 07:02

Subject's Last Name: VILLMOW

Date/Time of Test: 04/21/2007 07:47

Diagnostics Check	OK	07:40
Air Blank	0.000	07:41
Control Test	0.080	07:41
Air Blank	0.000	07:42
Subject Sample #1	0.145	07:42
Breath Volume	2.835	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:45
Subject Sample #2	0.126	07:45
Breath Volume	1.187	07:45
Air Blank	0.000	07:46
Control Test	0.078	07:46
Air Blank	0.000	07:47
Diagnostics Check	OK	07:47

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: SHEA, JOHN

Arresting Officer Agency: Winter Springs PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 22:04

Gender: Female

Observation Period Began: 22:40

Subject's Last Name: MEADOWS

Date/Time of Test: 04/21/2007 23:27

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:13	
Control Test	0.079	23:13	
Air Blank	0.000	23:14	
Subject Sample #1	0.148	23:16	
Breath Volume	1.433	23:16	
Air Blank	0.000	23:16	
Air Blank	0.000	23:18	
Subject Sample #2	0.150	23:21	Volume Not Met
Breath Volume	0.507	23:21	
Air Blank	0.000	23:22	
Air Blank	0.000	23:24	
Subject Sample #3	0.161	23:25	
Breath Volume	2.433	23:25	
Air Blank	0.000	23:26	
Control Test	0.080	23:26	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MATTINGLY, D

Arresting Officer Agency: Oviedo PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 01:55

Subject's Last Name: LAMETTRY

Date/Time of Test: 04/22/2007 02:34

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Subject Sample #1	0.118	02:30
Breath Volume	4.476	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample #2	0.110	02:33
Breath Volume	4.304	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: STANLEY, H

Arresting Officer Agency: Lake Mary PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 02:38

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: MOSES

Date/Time of Test: 04/22/2007 03:31

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Subject Sample #1	0.128	03:27
Breath Volume	3.925	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample #2	0.128	03:30
Breath Volume	4.062	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MOATES, A

Arresting Officer Agency: Casselberry PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 22:45

Gender: Male

Observation Period Began: 23:14

Subject's Last Name: NEWMAN

Date/Time of Test: 04/22/2007 23:42

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample #1	0.106	23:37
Breath Volume	4.460	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample #2	0.112	23:41
Breath Volume	4.691	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: COOK, JAMES

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: SWIGER

Date/Time of Test: 04/23/2007 03:04

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Subject Sample #1	0.088	02:59
Breath Volume	2.917	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample #2	0.082	03:03
Breath Volume	2.992	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: JASPER, D

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 22:52

Gender: Male

Observation Period Began: 23:20

Subject's Last Name: PILGRIM

Date/Time of Test: 04/23/2007 23:51

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:45
Subject Sample #1	0.139	23:46
Breath Volume	2.781	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample #2	0.131	23:49
Breath Volume	2.636	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MCNEIL, BRETT

Arresting Officer Agency: Oviedo PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 04:20

Gender: Male

Observation Period Began: 04:57

Subject's Last Name: GREMILLON

Date/Time of Test: 04/25/2007 05:24

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:21	
Control Test	0.081	05:22	
Air Blank	0.000	05:22	
Subject Sample #1		05:23	Subject Test Refused
Breath Volume	0.000	05:23	
Air Blank	0.000	05:23	
Control Test	0.080	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MCNEIL, B

Arresting Officer Agency: Oviedo PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 20:56

Gender: Female

Observation Period Began: 21:50

Subject's Last Name: SURETTE

Date/Time of Test: 04/25/2007 22:27

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:12	
Control Test	0.081	22:12	
Air Blank	0.000	22:12	
Subject Sample #1	0.135	22:16	Volume Not Met
Breath Volume	0.031	22:16	
Air Blank	0.000	22:16	
Air Blank	0.000	22:18	
Subject Sample #2	0.140	22:20	
Breath Volume	1.410	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:23	
Subject Sample #3	0.146	22:26	
Breath Volume	1.265	22:26	
Air Blank	0.000	22:26	
Control Test	0.079	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: NAS, J

Arresting Officer Agency: Casselberry PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 19:57

Gender: Female

Observation Period Began: 20:36

Subject's Last Name: ROQUE-CRUZ

Date/Time of Test: 04/26/2007 21:19

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:14
Subject Sample #1	0.154	21:15
Breath Volume	1.207	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:17
Subject Sample #2	0.157	21:18
Breath Volume	1.859	21:18
Air Blank	0.000	21:18
Control Test	0.078	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MUNN, SCOTT

Arresting Officer Agency: Casselberry PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 21:44

Gender: Male

Observation Period Began: 22:30

Subject's Last Name: PERRY

Date/Time of Test: 04/26/2007 23:03

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:53
Air Blank	0.000	22:53
Subject Sample #1	0.150	22:56
Breath Volume	1.125	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample #2	0.138	23:01
Breath Volume	1.562	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:03

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: COOK, J

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 00:10

Gender: Male

Observation Period Began: 01:03

Subject's Last Name: THURMOND

Date/Time of Test: 04/27/2007 01:29

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample #1	0.186	01:25
Breath Volume	3.378	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample #2	0.168	01:27
Breath Volume	1.679	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: PEDERSON, K

Arresting Officer Agency: Casselberry PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:25

Gender: Male

Observation Period Began: 01:51

Subject's Last Name: BEALS

Date/Time of Test: 04/27/2007 02:17

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample #1	0.160	02:13
Breath Volume	4.164	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample #2	0.156	02:16
Breath Volume	4.179	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MUNN, S

Arresting Officer Agency: Casselberry PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 02:34

Subject's Last Name: HERNANDEZ

Date/Time of Test: 04/27/2007 03:00

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Subject Sample #1	0.146	02:55
Breath Volume	3.574	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample #2	0.141	02:58
Breath Volume	2.464	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: PICKLESIMER, V, M

Breath Test Operator Agency: Lake Mary PD

Arresting Officer: STEIN, PAUL

Arresting Officer Agency: Winter Springs PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 04:26

Gender: Male

Observation Period Began: 04:55

Subject's Last Name: RIVERA

Date/Time of Test: 04/27/2007 05:33

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.079	05:28
Air Blank	0.000	05:28
Subject Sample #1	0.208	05:28
Breath Volume	3.195	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample #2	0.206	05:32
Breath Volume	3.585	05:32
Air Blank	0.000	05:32
Control Test	0.079	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: RIGGINS, PAUL, J

Breath Test Operator Agency: Seminole County SO

Arresting Officer: JENSEN, T

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 12:00

Gender: Male

Observation Period Began: 12:50

Subject's Last Name: SHORT

Date/Time of Test: 04/27/2007 13:24

Diagnostics Check	OK	13:18
Air Blank	0.000	13:19
Control Test	0.080	13:19
Air Blank	0.000	13:20
Subject Sample #1	0.106	13:20
Breath Volume	2.125	13:20
Air Blank	0.000	13:21
Air Blank	0.000	13:22
Subject Sample #2	0.104	13:23
Breath Volume	2.609	13:23
Air Blank	0.000	13:23
Control Test	0.080	13:24
Air Blank	0.000	13:24
Diagnostics Check	OK	13:24

Data Download Date/Time: 04/27/2007 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BETHAM, KEITH, D

Breath Test Operator Agency: Seminole County SO

Arresting Officer: SMITH, ROBERT

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 00:48

Gender: Female

Observation Period Began: 01:29

Subject's Last Name: VINCENT

Date/Time of Test: 04/28/2007 02:05

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.080	01:51	
Air Blank	0.000	01:51	
Subject Sample #1	0.162	01:54	Volume Not Met
Breath Volume	0.054	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample #2	0.142	02:00	
Breath Volume	1.796	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:02	
Subject Sample #3	0.141	02:04	
Breath Volume	1.585	02:04	
Air Blank	0.000	02:04	
Control Test	0.077	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 05/23/2007 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: SMITH, R

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:06

Gender: Female

Observation Period Began: 01:31

Subject's Last Name: RICE

Date/Time of Test: 04/29/2007 01:57

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Subject Sample #1	0.119	01:53
Breath Volume	2.539	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample #2	0.115	01:55
Breath Volume	1.394	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 05/23/2007 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: OKEEFE, M

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 02:45

Gender: Female

Observation Period Began: 03:42

Subject's Last Name: PITTER

Date/Time of Test: 04/29/2007 04:14

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.081	04:03	
Air Blank	0.000	04:03	
Subject Sample #1	0.089	04:06	Volume Not Met
Breath Volume	0.074	04:06	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample #2	0.097	04:12	Volume Not Met
Breath Volume	0.218	04:12	
Air Blank	0.000	04:13	
Control Test	0.078	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 05/23/2007 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: BERRIOS, STEPHANIE, A

Breath Test Operator Agency: Seminole County SO

Arresting Officer: OKEEFE, M

Arresting Officer Agency: Seminole County SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 02:45

Gender: Female

Observation Period Began: 03:42

Subject's Last Name: PITTER

Date/Time of Test: 04/29/2007 04:30

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:16	
Control Test	0.079	04:17	
Air Blank	0.000	04:17	
Subject Sample #1	0.086	04:20	Volume Not Met
Breath Volume	0.496	04:20	
Air Blank	0.000	04:21	
Air Blank	0.000	04:23	
Subject Sample #2	0.095	04:24	
Breath Volume	1.144	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample #3	0.100	04:29	
Breath Volume	1.523	04:29	
Air Blank	0.000	04:29	
Control Test	0.078	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 05/23/2007 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: RIGGINS, PAUL, J

Breath Test Operator Agency: Seminole County SO

Arresting Officer: POTTER, K

Arresting Officer Agency: Sanford PD

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 05:42

Gender: Male

Observation Period Began: 05:50

Subject's Last Name: HERNENDEZ-SANTIAGO

Date/Time of Test: 04/29/2007 06:26

Diagnostics Check	OK	06:20
Air Blank	0.000	06:20
Control Test	0.080	06:20
Air Blank	0.000	06:21
Subject Sample #1	0.141	06:22
Breath Volume	2.652	06:22
Air Blank	0.000	06:23
Air Blank	0.000	06:24
Subject Sample #2	0.144	06:25
Breath Volume	2.441	06:25
Air Blank	0.000	06:25
Control Test	0.079	06:26
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 05/23/2007 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: CREMEANS, PATTI, M

Breath Test Operator Agency: Altamonte Springs PD

Arresting Officer: CREMEANS, PATTI

Arresting Officer Agency: Altamonte Springs PD

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 16:45

Gender: Male

Observation Period Began: 18:20

Subject's Last Name: MALANOWSKI

Date/Time of Test: 04/29/2007 18:50

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:45
Subject Sample #1	0.169	18:45
Breath Volume	2.652	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample #2	0.162	18:48
Breath Volume	3.058	18:48
Air Blank	0.000	18:49
Control Test	0.080	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 05/23/2007 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-000998

Instrument Registered To: SEMINOLE S.O.

Software Version: 8100.27

Breath Test Operator: GARCIA, RAY,

Breath Test Operator Agency: Seminole County SO

Arresting Officer: MARCANTONI, WALLACE

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 23:03

Gender: Male

Observation Period Began: 23:50

Subject's Last Name: VAZQUEZ

Date/Time of Test: 04/30/2007 00:42

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Subject Sample #1	0.191	00:38
Breath Volume	1.527	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample #2	0.197	00:40
Breath Volume	2.738	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 05/23/2007 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001000

Instrument Registered To: COCOA PD

Software Version: 8100.27

Breath Test Operator: BRADSHAW, C,

Breath Test Operator Agency: Cocoa PD

Arresting Officer: ROSENFELD, M

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 03/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:20

Gender: Male

Observation Period Began: 02:55

Subject's Last Name: STEELE

Date/Time of Test: 04/01/2007 03:28

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.078	03:18
Air Blank	0.000	03:18
Subject Sample #1	0.176	03:20
Breath Volume	1.144	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample #2	0.200	03:23
Breath Volume	1.191	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample #3	0.196	03:27
Breath Volume	1.667	03:27
Air Blank	0.000	03:27
Control Test	0.077	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 04/17/2007 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001000

Instrument Registered To: COCOA PD

Software Version: 8100.27

Breath Test Operator: BRADSHAW, C,

Breath Test Operator Agency: Cocoa PD

Arresting Officer: BRADSHAW, C

Arresting Officer Agency: Cocoa PD

Last Agency Inspection: 03/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:27

Gender: Male

Observation Period Began: 01:55

Subject's Last Name: BUCKNER

Date/Time of Test: 04/14/2007 02:21

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:16
Subject Sample #1	0.160	02:17
Breath Volume	1.722	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.158	02:20
Breath Volume	1.421	02:20
Air Blank	0.000	02:20
Control Test	0.077	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 04/17/2007 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001000

Instrument Registered To: COCOA PD

Software Version: 8100.27

Breath Test Operator: BRADSHAW, C,

Breath Test Operator Agency: Cocoa PD

Arresting Officer: BRADSHAW, C

Arresting Officer Agency: Cocoa PD

Last Agency Inspection: 03/07/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 00:45

Gender: Male

Observation Period Began: 00:55

Subject's Last Name: ANDERSON

Date/Time of Test: 04/15/2007 01:23

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:16
Air Blank	0.000	01:17
Subject Sample #1	0.000	01:19
Breath Volume	1.437	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample #2	0.000	01:22
Breath Volume	1.492	01:22
Air Blank	0.000	01:22
Control Test	0.077	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 04/17/2007 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001000

Instrument Registered To: COCOA PD

Software Version: 8100.27

Breath Test Operator: BRADSHAW, C,

Breath Test Operator Agency: Cocoa PD

Arresting Officer: BRADSHAW, C

Arresting Officer Agency: Cocoa PD

Last Agency Inspection: 04/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 00:10

Gender: Male

Observation Period Began: 00:10

Subject's Last Name: SAGE

Date/Time of Test: 04/19/2007 00:52

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:50	
Control Test	0.078	00:50	
Air Blank	0.000	00:50	
Subject Sample #1		00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 05/15/2007 09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: COOK, JEFFREY, A

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: COOK, J

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 03/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 02:40

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: SHEPHERD

Date/Time of Test: 04/01/2007 03:23

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.079	03:13	
Air Blank	0.000	03:13	
Subject Sample #1		03:16	No Sample Provided
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:19	
Subject Sample #2		03:22	No Sample Provided
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: COOK, JEFFREY, A

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: CAPORIZZO, DEPUTY

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 03/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 02:00

Gender: Female

Observation Period Began: 02:00

Subject's Last Name: FEHRENBACH

Date/Time of Test: 04/02/2007 02:57

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample #1	0.114	02:52
Breath Volume	1.234	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample #2	0.119	02:55
Breath Volume	2.316	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: SOUSA, PAUL, M

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: SOUSA, PAUL

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 03/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 02:34

Gender: Female

Observation Period Began: 02:55

Subject's Last Name: CHANG

Date/Time of Test: 04/03/2007 03:26

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Subject Sample #1	0.069	03:21
Breath Volume	1.449	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample #2	0.066	03:25
Breath Volume	1.242	03:25
Air Blank	0.000	03:25
Control Test	0.081	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: COOK, JEFFREY, A

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: MASSOLINI, DEPUTY

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 03/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 00:49

Gender: Male

Observation Period Began: 00:49

Subject's Last Name: GHALI

Date/Time of Test: 04/05/2007 01:21

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:15
Air Blank	0.000	01:15
Subject Sample #1	0.141	01:16
Breath Volume	1.542	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample #2	0.139	01:19
Breath Volume	1.550	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: ANDERSON, ERIC, G

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: ANDERSON, ERIC

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 03/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 23:18

Gender: Female

Observation Period Began: 23:30

Subject's Last Name: PEARSON

Date/Time of Test: 04/07/2007 00:06

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Subject Sample #1	0.192	00:02
Breath Volume	2.195	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample #2	0.192	00:05
Breath Volume	1.722	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: MEYERS, BRIAN, K

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: MEYERS, BRIAN

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 03/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 00:19

Gender: Male

Observation Period Began: 00:01

Subject's Last Name: MUZZEY

Date/Time of Test: 04/07/2007 01:00

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Subject Sample #1	0.169	00:55
Breath Volume	2.429	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample #2	0.158	00:58
Breath Volume	1.847	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: COOK, JEFFREY, A

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: COOK, JEFFREY

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 03/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 00:10

Gender: Male

Observation Period Began: 01:18

Subject's Last Name: PANASIS

Date/Time of Test: 04/07/2007 02:07

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:57	
Subject Sample #1		02:00	No Sample Provided
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample #2		02:06	No Sample Provided
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: ANDERSON, ERIC, G

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: ANDERSON, ERIC

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 03/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 03:01

Subject's Last Name: LIENAU

Date/Time of Test: 04/08/2007 03:31

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
		03:30	RFI Detect
Air Blank	0.000	03:31	

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: ANDERSON, ERIC, G

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: ANDERSON, ERIC

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 03/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 03:01

Subject's Last Name: LIENAU

Date/Time of Test: 04/08/2007 03:38

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample #1	0.160	03:34
Breath Volume	2.656	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample #2	0.165	03:37
Breath Volume	2.839	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: MCHUGH, CHRIS, J

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: FROST, DARREN

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 03/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 21:21

Gender: Male

Observation Period Began: 21:30

Subject's Last Name: MCNAIR

Date/Time of Test: 04/08/2007 22:10

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.080	22:05
Air Blank	0.000	22:05
Subject Sample #1	0.166	22:06
Breath Volume	2.777	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:08
Subject Sample #2	0.162	22:09
Breath Volume	3.093	22:09
Air Blank	0.000	22:09
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: ANDERSON, ERIC, G

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: ANDERSON, ERIC

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:03

Gender: Male

Observation Period Began: 01:03

Subject's Last Name: MCPHERSON

Date/Time of Test: 04/14/2007 02:10

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:04
Air Blank	0.000	02:04
Subject Sample #1	0.035	02:06
Breath Volume	2.417	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample #2	0.032	02:09
Breath Volume	2.246	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: FERNEZ, STEPHEN, J

Breath Test Operator Agency: Brevard County SO

Arresting Officer: FERNEZ, STEPHEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:04

Gender: Male

Observation Period Began: 01:41

Subject's Last Name: BRAY

Date/Time of Test: 04/14/2007 02:21

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample #1	0.224	02:16
Breath Volume	1.988	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.214	02:20
Breath Volume	1.957	02:20
Air Blank	0.000	02:21
Control Test	0.076	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 04/17/2007 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: COOK, JEFFERY, A

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: FREEMAN, DEPUTY

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:21

Gender: Male

Observation Period Began: 01:21

Subject's Last Name: ERCOLANI

Date/Time of Test: 04/19/2007 02:06

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample #1	0.156	02:02
Breath Volume	1.796	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:04
Subject Sample #2	0.152	02:05
Breath Volume	1.527	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 05/15/2007 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: COOK, JEFFREY, A

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: MEYERS, BRIAN

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:52

Gender: Female

Observation Period Began: 01:52

Subject's Last Name: REDWING

Date/Time of Test: 04/21/2007 02:32

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.080	02:21	
Air Blank	0.000	02:22	
Subject Sample #1	0.200	02:25	Volume Not Met
Breath Volume	0.066	02:25	
Air Blank	0.000	02:26	
Air Blank	0.000	02:27	
Subject Sample #2		02:31	No Sample Provided
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.078	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 05/15/2007 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: COOK, JEFFREY, A

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: MEYERS, BRIAN

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:52

Gender: Female

Observation Period Began: 01:52

Subject's Last Name: REDWING

Date/Time of Test: 04/21/2007 02:46

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.080	02:34	
Air Blank	0.000	02:35	
Subject Sample #1	0.187	02:38	Volume Not Met
Breath Volume	0.597	02:38	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample #2	0.199	02:41	
Breath Volume	1.996	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:43	
Subject Sample #3	0.204	02:44	
Breath Volume	1.777	02:44	
Air Blank	0.000	02:45	
Control Test	0.076	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 05/15/2007 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: FERNEZ, STEPHEN, J

Breath Test Operator Agency: Brevard County SO

Arresting Officer: FERNEZ, STEPHEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 02:02

Gender: Male

Observation Period Began: 01:56

Subject's Last Name: DONALDSON

Date/Time of Test: 04/22/2007 02:25

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:20
Subject Sample #1	0.165	02:21
Breath Volume	3.027	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample #2	0.170	02:24
Breath Volume	2.898	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 05/15/2007 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: FERNEZ, STEPHEN, J

Breath Test Operator Agency: Brevard County SO

Arresting Officer: FERNEZ, STEPHEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:40

Gender: Female

Observation Period Began: 01:30

Subject's Last Name: WICKENS

Date/Time of Test: 04/23/2007 02:11

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Subject Sample #1	0.197	02:06
Breath Volume	1.394	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #2	0.201	02:09
Breath Volume	1.382	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 05/15/2007 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: COOK, JEFFREY, A

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: MILLER, J

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 02:15

Gender: Male

Observation Period Began: 03:24

Subject's Last Name: SMITH

Date/Time of Test: 04/25/2007 03:57

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample #1	0.167	03:53
Breath Volume	1.500	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample #2	0.184	03:56
Breath Volume	1.812	03:56
Air Blank	0.000	03:57
Control Test	0.077	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 05/15/2007 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: MCHUGH, CHRIS, J

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: MCHUGH, CHRIS

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 23:07

Gender: Male

Observation Period Began: 23:35

Subject's Last Name: MERICOLA

Date/Time of Test: 04/26/2007 00:07

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Subject Sample #1	0.091	00:02
Breath Volume	3.679	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample #2	0.088	00:05
Breath Volume	2.660	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 05/15/2007 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001001

Instrument Registered To: SATELLITE BEACH PD

Software Version: 8100.27

Breath Test Operator: MCHUGH, CHRIS, J

Breath Test Operator Agency: Satellite Beach PD

Arresting Officer: MCHUGH, CHRIS

Arresting Officer Agency: Satellite Beach PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:19

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: BISSETT

Date/Time of Test: 04/28/2007 02:29

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Subject Sample #1	0.196	02:25
Breath Volume	3.472	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample #2	0.188	02:28
Breath Volume	2.988	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 05/15/2007 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: THOMAS, GLEN, P

Breath Test Operator Agency: Brevard County SO

Arresting Officer: THOMAS, GLEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 17:23

Gender: Male

Observation Period Began: 17:23

Subject's Last Name: BAKER

Date/Time of Test: 04/04/2007 18:42

Diagnostics Check	OK	18:36	
Air Blank	0.000	18:37	
Control Test	0.081	18:37	
Air Blank	0.000	18:38	
Subject Sample #1	0.128	18:39	
Breath Volume	2.277	18:39	
Air Blank	0.000	18:39	
Air Blank		18:41	RFI Detect
Air Blank	0.000	18:41	
Control Test	0.082	18:42	
Air Blank	0.000	18:42	
Diagnostics Check	OK	18:42	

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: THOMAS, GLEN, P

Breath Test Operator Agency: Brevard County SO

Arresting Officer: THOMAS, GLEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 17:23

Gender: Male

Observation Period Began: 17:23

Subject's Last Name: BAKER

Date/Time of Test: 04/04/2007 18:53

Diagnostics Check	OK	18:47
Air Blank	0.000	18:48
Control Test	0.081	18:48
Air Blank	0.000	18:48
Subject Sample #1	0.126	18:49
Breath Volume	1.687	18:49
Air Blank	0.000	18:50
Air Blank	0.000	18:51
Subject Sample #2	0.134	18:52
Breath Volume	2.972	18:52
Air Blank	0.000	18:53
Control Test	0.081	18:53
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: MANLEY, ROBERT, W

Breath Test Operator Agency: Brevard County SO

Arresting Officer: LEARY, EDWARD

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 21:19

Gender: Male

Observation Period Began: 21:19

Subject's Last Name: PETRUSKI

Date/Time of Test: 04/10/2007 21:52

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:47
Subject Sample #1	0.176	21:48
Breath Volume	3.285	21:48
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample #2	0.171	21:50
Breath Volume	3.890	21:50
Air Blank	0.000	21:51
Control Test	0.076	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: MANLEY, ROBERT, W

Breath Test Operator Agency: Brevard County SO

Arresting Officer: MANLEY, ROBERT

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 01:21

Gender: Male

Observation Period Began: 01:21

Subject's Last Name: SCOTT

Date/Time of Test: 04/12/2007 02:13

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample #1	0.222	02:08
Breath Volume	2.214	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample #2	0.224	02:11
Breath Volume	1.593	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: MANLEY, ROBERT, W

Breath Test Operator Agency: Brevard County SO

Arresting Officer: LOCKHART, BRETT

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 00:54

Gender: Male

Observation Period Began: 01:35

Subject's Last Name: FRANKS

Date/Time of Test: 04/13/2007 02:03

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Subject Sample #1	0.155	01:58
Breath Volume	3.753	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample #2	0.152	02:01
Breath Volume	3.890	02:01
Air Blank	0.000	02:02
Control Test	0.081	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: NEGRON, ALEXANDER,

Breath Test Operator Agency: Brevard County SO

Arresting Officer: WEST, JASON

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 03:03

Gender: Male

Observation Period Began: 03:03

Subject's Last Name: KIRKEY

Date/Time of Test: 04/16/2007 04:04

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.082	03:59
Air Blank	0.000	04:00
Subject Sample #1	0.170	04:00
Breath Volume	1.859	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample #2	0.167	04:03
Breath Volume	2.160	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: MANLEY, ROBERT, W

Breath Test Operator Agency: Brevard County SO

Arresting Officer: ANDERSON, SAM

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 20:22

Gender: Male

Observation Period Began: 20:52

Subject's Last Name: RAAB

Date/Time of Test: 04/19/2007 21:23

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.080	21:18
Air Blank	0.000	21:18
Subject Sample #1	0.088	21:19
Breath Volume	1.308	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample #2	0.102	21:22
Breath Volume	3.105	21:22
Air Blank	0.000	21:22
Control Test	0.083	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: MANLEY, ROBERT, W

Breath Test Operator Agency: Brevard County SO

Arresting Officer: BENNETT, ROBERT

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 21:14

Gender: Male

Observation Period Began: 21:14

Subject's Last Name: POLLARD

Date/Time of Test: 04/20/2007 22:14

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.082	22:09
Air Blank	0.000	22:09
Subject Sample #1	0.219	22:10
Breath Volume	2.773	22:10
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample #2	0.220	22:12
Breath Volume	2.410	22:12
Air Blank	0.000	22:13
Control Test	0.081	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: LEWIS, DAVID, A

Breath Test Operator Agency: Brevard County SO

Arresting Officer: ANDERSON, SAMUEL

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 22:00

Gender: Male

Observation Period Began: 22:30

Subject's Last Name: BARDING

Date/Time of Test: 04/21/2007 23:10

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Subject Sample #1	0.258	23:05
Breath Volume	1.238	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample #2	0.265	23:08
Breath Volume	2.320	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: GARRITY, THOMAS, A

Breath Test Operator Agency: Brevard County SO

Arresting Officer: JOINER, TELLY

Arresting Officer Agency: Titusville PD

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 00:18

Gender: Male

Observation Period Began: 02:47

Subject's Last Name: THOMPSON

Date/Time of Test: 04/22/2007 03:17

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Subject Sample #1	0.210	03:12
Breath Volume	2.007	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample #2	0.220	03:15
Breath Volume	1.226	03:15
Air Blank	0.000	03:16
Control Test	0.083	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: GARRITY, THOMAS, A

Breath Test Operator Agency: Brevard County SO

Arresting Officer: DIAZ, ORLANDO

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 02:52

Gender: Female

Observation Period Began: 02:52

Subject's Last Name: BUTLER

Date/Time of Test: 04/22/2007 04:23

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.082	04:18
Air Blank	0.000	04:18
Subject Sample #1	0.114	04:19
Breath Volume	3.222	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample #2	0.112	04:22
Breath Volume	2.679	04:22
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: DOWDY, BRUCE,

Breath Test Operator Agency: Brevard County SO

Arresting Officer: KERSEY, ETHAN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 11:47

Gender: Female

Observation Period Began: 10:27

Subject's Last Name: LANG

Date/Time of Test: 04/25/2007 13:09

Diagnostics Check	OK	13:03
Air Blank	0.000	13:03
Control Test	0.081	13:03
Air Blank	0.000	13:04
Subject Sample #1	0.000	13:04
Breath Volume	2.332	13:04
Air Blank	0.000	13:05
Air Blank	0.000	13:07
Subject Sample #2	0.000	13:07
Breath Volume	2.867	13:07
Air Blank	0.000	13:08
Control Test	0.081	13:08
Air Blank	0.000	13:09
Diagnostics Check	OK	13:09

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001004

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: DELAUGHTER, MARC, C

Breath Test Operator Agency: Brevard County SO

Arresting Officer: DELAUGHTER, MARC

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 21:32

Gender: Male

Observation Period Began: 21:32

Subject's Last Name: DEUTSCH

Date/Time of Test: 04/27/2007 23:08

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.082	23:02
Air Blank	0.000	23:03
Subject Sample #1	0.117	23:03
Breath Volume	2.343	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample #2	0.116	23:07
Breath Volume	1.628	23:07
Air Blank	0.000	23:08
Control Test	0.081	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 05/14/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: REAM, RANDY, J

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: REAM, RANDY

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 23:30

Gender: Male

Observation Period Began: 00:49

Subject's Last Name: BRUCE

Date/Time of Test: 04/01/2007 01:15

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample #1	0.000	01:11
Breath Volume	1.777	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample #2	0.000	01:14
Breath Volume	1.738	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEANE, M,

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: CLARK, A

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 23:22

Gender: Male

Observation Period Began: 00:20

Subject's Last Name: HERSTEL

Date/Time of Test: 04/02/2007 00:48

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.081	00:41
Air Blank	0.000	00:42
Subject Sample #1	0.204	00:43
Breath Volume	1.636	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample #2	0.184	00:46
Breath Volume	1.417	00:46
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: TAYLOR, G, S

Breath Test Operator Agency: FHP Troop C

Arresting Officer: TAYLOR, G

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 03:12

Gender: Male

Observation Period Began: 04:07

Subject's Last Name: HARRISON

Date/Time of Test: 04/02/2007 04:30

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:28	
Control Test	0.080	04:28	
Air Blank	0.000	04:29	
Subject Sample #1		04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.080	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: RECLA, T, A

Breath Test Operator Agency: Largo PD

Arresting Officer: RECLA, T

Arresting Officer Agency: Largo PD

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 19:45

Gender: Male

Observation Period Began: 20:08

Subject's Last Name: ALMAND

Date/Time of Test: 04/02/2007 20:35

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.080	20:30
Air Blank	0.000	20:30
Subject Sample #1	0.231	20:31
Breath Volume	3.257	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample #2	0.226	20:34
Breath Volume	3.109	20:34
Air Blank	0.000	20:34
Control Test	0.080	20:34
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: REAM, RANDY, J

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: REAM, RANDY

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 02:03

Gender: Male

Observation Period Began: 02:54

Subject's Last Name: SHAFFER

Date/Time of Test: 04/05/2007 03:20

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:15
Air Blank	0.000	03:15
Subject Sample #1	0.129	03:16
Breath Volume	2.214	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample #2	0.117	03:19
Breath Volume	1.117	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: LACHANCE, DEPUTY, D

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: LACHANCE, DEPUTY

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 02:28

Gender: Male

Observation Period Began: 02:53

Subject's Last Name: PITCHER

Date/Time of Test: 04/06/2007 03:20

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample #1	0.131	03:15
Breath Volume	1.214	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample #2	0.141	03:18
Breath Volume	1.714	03:18
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: PANIAGUA, MICHAEL, A

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: PANIAGUA, MICHAEL

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 02:02

Gender: Female

Observation Period Began: 02:32

Subject's Last Name: LYNDS

Date/Time of Test: 04/07/2007 02:59

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample #1	0.271	02:55
Breath Volume	2.300	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.271	02:58
Breath Volume	2.246	02:58
Air Blank	0.000	02:58
Control Test	0.081	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: EASTTY, MARK, L

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: EASTTY, MARK

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 21:16

Gender: Male

Observation Period Began: 21:31

Subject's Last Name: HUMPHREYS

Date/Time of Test: 04/07/2007 22:04

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.080	21:59
Air Blank	0.000	21:59
Subject Sample #1	0.154	22:00
Breath Volume	2.027	22:00
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample #2	0.159	22:03
Breath Volume	1.601	22:03
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: LACHANCE, DEPUTY, D

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: LACHANCE, DEPUTY

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 02:54

Gender: Male

Observation Period Began: 03:14

Subject's Last Name: INTHAVONG

Date/Time of Test: 04/08/2007 03:40

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:35
Subject Sample #1	0.115	03:36
Breath Volume	1.578	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample #2	0.115	03:39
Breath Volume	1.574	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: WILHELM, JAMES, D

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: WILHELM, JAMES

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 02:36

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: DOUGHERTY

Date/Time of Test: 04/10/2007 03:36

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample #1	0.197	03:31
Breath Volume	3.550	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample #2	0.183	03:34
Breath Volume	2.906	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: SKIPPER, CHARLES, H

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: GROSS, SPENCER

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 15:05

Gender: Female

Observation Period Began: 15:48

Subject's Last Name: TYLER

Date/Time of Test: 04/11/2007 16:19

Diagnostics Check	OK	16:18	
Air Blank	0.000	16:18	
Control Test	0.080	16:19	
Air Blank		16:19	RFI Detect
Air Blank	0.000	16:19	

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: SKIPER, CHARLES, H

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: GROSS, SPENCER

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 15:05

Gender: Female

Observation Period Began: 15:48

Subject's Last Name: TYLER

Date/Time of Test: 04/11/2007 16:24

Diagnostics Check	OK	16:22	
Air Blank	0.000	16:22	
Control Test	0.080	16:22	
Air Blank	0.000	16:23	
Subject Sample #1		16:24	Sequence Aborted
Breath Volume	0.000	16:24	
Air Blank	0.000	16:24	

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: PEACOCK, CHERYL, A

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: KOLBICKA, L

Arresting Officer Agency: Pinellas Park PD

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 02:38

Subject's Last Name: PETERS

Date/Time of Test: 04/12/2007 03:05

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Subject Sample #1	0.183	03:00
Breath Volume	2.285	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample #2	0.186	03:03
Breath Volume	2.308	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: PREISING, J, S

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: SHOULTA, J

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 16:15

Gender: Male

Observation Period Began: 17:02

Subject's Last Name: O'BRIEN

Date/Time of Test: 04/12/2007 17:31

Diagnostics Check	OK	17:22
Air Blank	0.000	17:23
Control Test	0.080	17:23
Air Blank	0.000	17:24
Subject Sample #1	0.259	17:26
Breath Volume	2.335	17:26
Air Blank	0.000	17:27
Air Blank	0.000	17:29
Subject Sample #2	0.244	17:29
Breath Volume	1.562	17:29
Air Blank	0.000	17:30
Control Test	0.084	17:30
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: LACHANCE, DEPUTY, D

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: LACHANCE, DEPUTY

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 03:02

Gender: Male

Observation Period Began: 03:27

Subject's Last Name: RUBI

Date/Time of Test: 04/14/2007 03:54

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:49
Subject Sample #1	0.065	03:50
Breath Volume	2.476	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample #2	0.075	03:53
Breath Volume	2.136	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MCKNIGHT, CLINTON, H

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: MCKNIGHT, CLINTON

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 03/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 08:45

Gender: Male

Observation Period Began: 09:18

Subject's Last Name: JENNINGS

Date/Time of Test: 04/15/2007 09:47

Diagnostics Check	OK	09:41
Air Blank	0.000	09:41
Control Test	0.081	09:42
Air Blank	0.000	09:42
Subject Sample #1	0.000	09:43
Breath Volume	3.246	09:43
Air Blank	0.000	09:43
Air Blank	0.000	09:45
Subject Sample #2	0.000	09:45
Breath Volume	1.941	09:45
Air Blank	0.000	09:46
Control Test	0.081	09:46
Air Blank	0.000	09:47
Diagnostics Check	OK	09:47

Data Download Date/Time: 04/15/2007 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEANE, M,

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: WOODMAN, T

Arresting Officer Agency: Gulfport PD

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 23:00

Gender: Female

Observation Period Began: 23:37

Subject's Last Name: POWELL

Date/Time of Test: 04/16/2007 00:12

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.081	00:10	
Air Blank	0.000	00:11	
Subject Sample #1		00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Control Test	0.081	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 04/16/2007 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEANE, M,

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: MCKENZIE, J

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 00:12

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: TIPTON

Date/Time of Test: 04/17/2007 01:27

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:25	
Control Test	0.082	01:25	
Air Blank	0.000	01:26	
Subject Sample #1		01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.081	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEANE, M,

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: DIX, N

Arresting Officer Agency: Largo PD

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 03:43

Gender: Male

Observation Period Began: 04:30

Subject's Last Name: MULLEN

Date/Time of Test: 04/17/2007 05:03

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.082	04:57
Air Blank	0.000	04:58
Subject Sample #1	0.167	04:58
Breath Volume	3.257	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample #2	0.162	05:01
Breath Volume	2.859	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEANE, M,

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: DEIBERN, A

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 01:05

Gender: Female

Observation Period Began: 01:55

Subject's Last Name: FATKIN

Date/Time of Test: 04/18/2007 02:19

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.081	02:17	
Air Blank	0.000	02:18	
Subject Sample #1		02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Control Test	0.080	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: PEACOCK, CHERYL, A

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: IRELAND, B

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 01:40

Gender: Female

Observation Period Began: 02:04

Subject's Last Name: LASALLA

Date/Time of Test: 04/19/2007 02:30

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.082	02:25
Air Blank	0.000	02:25
Subject Sample #1	0.147	02:26
Breath Volume	1.871	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample #2	0.139	02:29
Breath Volume	2.398	02:29
Air Blank	0.000	02:29
Control Test	0.081	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: LACHANCE, DEPUTY, D

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: LACHANCE, DEPUTY

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 02:25

Gender: Female

Observation Period Began: 02:50

Subject's Last Name: XUEREB

Date/Time of Test: 04/19/2007 03:13

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.081	03:11	
Air Blank	0.000	03:11	
Subject Sample #1		03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.081	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: EASTTY, MARK, L

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: EASTTY, MARK

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 20:38

Gender: Female

Observation Period Began: 20:48

Subject's Last Name: LEMUS

Date/Time of Test: 04/19/2007 21:20

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.081	21:15
Air Blank	0.000	21:15
Subject Sample #1	0.000	21:16
Breath Volume	1.683	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample #2	0.000	21:19
Breath Volume	1.433	21:19
Air Blank	0.000	21:20
Control Test	0.081	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: CITRANO, ANTHONY,

Breath Test Operator Agency: Largo PD

Arresting Officer: CITRANO, ANTHONY

Arresting Officer Agency: Largo PD

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 04:01

Gender: Male

Observation Period Began: 04:10

Subject's Last Name: ADAMS

Date/Time of Test: 04/20/2007 04:35

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:33	
Subject Sample #1		04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:34	
Control Test	0.081	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: SKAGGS, H, E

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: SKAGGS, H

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 00:57

Gender: Male

Observation Period Began: 01:34

Subject's Last Name: ARHONDIS

Date/Time of Test: 04/21/2007 02:02

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.082	01:56
Air Blank	0.000	01:56
Subject Sample #1	0.177	01:57
Breath Volume	1.425	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample #2	0.169	02:00
Breath Volume	1.710	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:02

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: VIENO, MICHAEL, J

Breath Test Operator Agency: Kenneth City PD

Arresting Officer: VIENO, MICHAEL

Arresting Officer Agency: Kenneth City PD

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 02:40

Gender: Male

Observation Period Began: 03:32

Subject's Last Name: BROWN

Date/Time of Test: 04/21/2007 03:59

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Subject Sample #1	0.000	03:54
Breath Volume	1.210	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample #2	0.000	03:57
Breath Volume	2.023	03:57
Air Blank	0.000	03:58
Control Test	0.082	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: REAM, RANDY, J

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: N/A, N/A

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: PROB/PAROL

Arrest Time: 19:30

Gender: Male

Observation Period Began: 19:30

Subject's Last Name: MURRAY

Date/Time of Test: 04/21/2007 20:04

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.080	19:59
Air Blank	0.000	19:59
Subject Sample #1	0.000	20:00
Breath Volume	3.242	20:00
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample #2	0.000	20:03
Breath Volume	2.906	20:03
Air Blank	0.000	20:03
Control Test	0.081	20:04
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MARSHALL, DAVID, O

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MARSHALL, DAVID

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/15/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 19:05

Gender: Female

Observation Period Began: 20:31

Subject's Last Name: SIMRELL

Date/Time of Test: 04/22/2007 21:06

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:01
Subject Sample #1	0.243	21:01
Breath Volume	2.355	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample #2	0.245	21:05
Breath Volume	2.566	21:05
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MAUL, CHRISTOPHER, J

Breath Test Operator Agency: FHP Troop C

Arresting Officer: MAUL, CHRISTOPHER

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6188011

Dry Gas Expiry Date: 07/10/2008

Violation Code: DUI

Arrest Time: 18:50

Gender: Male

Observation Period Began: 19:10

Subject's Last Name: MARSH

Date/Time of Test: 04/23/2007 19:53

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.081	19:46
Air Blank	0.000	19:47
Subject Sample #1	0.145	19:47
Breath Volume	1.371	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample #2	0.149	19:52
Breath Volume	2.230	19:52
Air Blank	0.000	19:52
Control Test	0.082	19:52
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: PEACOCK, CHERYL, A

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: ANDERSON, M

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:09

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: TEJAMANIL PENA

Date/Time of Test: 04/28/2007 02:19

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.081	02:05	
Air Blank	0.000	02:05	
Subject Sample #1	0.180	02:08	No .020 agreement
Breath Volume	1.742	02:08	
Air Blank	0.000	02:09	
Air Blank	0.000	02:10	
Subject Sample #2	0.175	02:14	Volume Not Met
Breath Volume	0.773	02:14	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample #3	0.155	02:17	No .020 agreement
Breath Volume	1.523	02:17	
Air Blank	0.000	02:18	
Control Test	0.080	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: PEACOCK, CHERYL, A

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: ANDERSON, M

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:09

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: TEJAMANIL PENA

Date/Time of Test: 04/28/2007 02:27

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Subject Sample #1	0.169	02:22
Breath Volume	1.539	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample #2	0.158	02:26
Breath Volume	1.597	02:26
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: REAM, RANDY, J

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: GRIFFIN, S

Arresting Officer Agency: Indian Shores PD

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 01:48

Gender: Male

Observation Period Began: 02:49

Subject's Last Name: MORGAN

Date/Time of Test: 04/28/2007 03:21

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:15
Subject Sample #1	0.197	03:16
Breath Volume	1.656	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample #2	0.216	03:20
Breath Volume	1.949	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: PEACOCK, CHERYL, A

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: NEVILLE, C

Arresting Officer Agency: Pinellas County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 21:00

Gender: Male

Observation Period Began: 21:55

Subject's Last Name: MARCINIAK

Date/Time of Test: 04/28/2007 22:26

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample #1	0.253	22:21
Breath Volume	1.925	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:23
Subject Sample #2	0.255	22:24
Breath Volume	2.171	22:24
Air Blank	0.000	22:25
Control Test	0.081	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:26

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEANE, M,

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: NICHOLS, T

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 18:44

Gender: Male

Observation Period Began: 19:08

Subject's Last Name: HURST

Date/Time of Test: 04/29/2007 19:37

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.081	19:32
Air Blank	0.000	19:32
Subject Sample #1	0.190	19:33
Breath Volume	3.300	19:33
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample #2	0.195	19:36
Breath Volume	2.828	19:36
Air Blank	0.000	19:36
Control Test	0.081	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001005

Instrument Registered To: PINELLAS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEANE, M,

Breath Test Operator Agency: Pinellas County SO

Arresting Officer: BUTINSKI, M

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 19:00

Gender: Male

Observation Period Began: 20:00

Subject's Last Name: WILLIAMS

Date/Time of Test: 04/30/2007 20:30

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.081	20:22
Air Blank	0.000	20:23
Subject Sample #1	0.200	20:26
Breath Volume	1.285	20:26
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample #2	0.183	20:29
Breath Volume	2.042	20:29
Air Blank	0.000	20:30
Control Test	0.081	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 05/13/2007 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001006

Instrument Registered To: INDIALANTIC PD

Software Version: 8100.27

Breath Test Operator: FERNEZ, STEPHEN, J

Breath Test Operator Agency: Brevard County SO

Arresting Officer: FERNEZ, STEPHEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 2

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 00:15

Gender: Male

Observation Period Began: 00:15

Subject's Last Name: DEAN

Date/Time of Test: 04/16/2007 01:48

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.080	01:37	
Air Blank	0.000	01:37	
Subject Sample #1	0.189	01:40	Volume Not Met
Breath Volume	0.359	01:40	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample #2	0.178	01:46	Volume Not Met
Breath Volume	0.242	01:46	
Air Blank	0.000	01:47	
Control Test	0.073	01:47	Control Outside Tolerance
Air Blank	0.000	01:48	

Data Download Date/Time: 04/27/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001006

Instrument Registered To: INDIALANTIC PD

Software Version: 8100.27

Breath Test Operator: GRISWOLD, MELANIE, S

Breath Test Operator Agency: Indialantic PD

Arresting Officer: GRISWOLD, MELANIE

Arresting Officer Agency: Melbourne Beach PD

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 02:48

Gender: Female

Observation Period Began: 03:15

Subject's Last Name: CHILBERG

Date/Time of Test: 04/19/2007 04:28

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Subject Sample #1	0.076	04:23
Breath Volume	1.535	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample #2	0.076	04:26
Breath Volume	1.164	04:26
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 04/27/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001007

Instrument Registered To: HAINES CITY PD

Software Version: 8100.27

Breath Test Operator: TEAGUE, MICHAEL, E

Breath Test Operator Agency: Haines City PD

Arresting Officer: TEAGUE, MICHAEL

Arresting Officer Agency: Haines City PD

Last Agency Inspection: 03/09/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 23:56

Gender: Male

Observation Period Began: 00:00

Subject's Last Name: ADAMS

Date/Time of Test: 04/02/2007 00:50

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Subject Sample #1	0.082	00:45
Breath Volume	2.386	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample #2	0.083	00:48
Breath Volume	2.867	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 04/17/2007 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001007

Instrument Registered To: HAINES CITY PD
Software Version: 8100.27
Breath Test Operator: PIRTLE, REGINA, L
Breath Test Operator Agency: Haines City PD
Arresting Officer: SMITH, CHAD
Arresting Officer Agency: Haines City PD
Last Agency Inspection: 03/09/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 629302I
Dry Gas Expiry Date: 10/23/2008
Violation Code: DUI
Arrest Time: 00:18
Gender: Male
Observation Period Began: 00:40

Subject's Last Name: MARTINEZ

Date/Time of Test: 04/09/2007 01:32

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Subject Sample #1	0.192	01:19	
Breath Volume	1.332	01:19	
Air Blank	0.000	01:20	
Air Blank	0.000	01:22	
Subject Sample #2	0.196	01:25	Volume Not Met
Breath Volume	0.070	01:25	
Air Blank	0.000	01:26	
Air Blank	0.000	01:28	
Subject Sample #3	0.192	01:31	Volume Not Met
Breath Volume	0.082	01:31	
Air Blank	0.000	01:31	
Control Test	0.078	01:32	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 04/17/2007 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001007

Instrument Registered To: HAINES CITY PD

Software Version: 8100.27

Breath Test Operator: ADAMS, RONALD, G

Breath Test Operator Agency: Haines City PD

Arresting Officer: ADAMS, RONALD

Arresting Officer Agency: Haines City PD

Last Agency Inspection: 03/09/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 03:00

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: DAVID

Date/Time of Test: 04/15/2007 03:34

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample #1	0.000	03:30
Breath Volume	2.425	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample #2	0.000	03:33
Breath Volume	2.089	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 04/17/2007 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001007

Instrument Registered To: HAINES CITY PD

Software Version: 8100.27

Breath Test Operator: PIRTLE, REGINA, L

Breath Test Operator Agency: Haines City PD

Arresting Officer: DAY, JUSTIN

Arresting Officer Agency: Haines City PD

Last Agency Inspection: 04/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 23:33

Gender: Male

Observation Period Began: 00:15

Subject's Last Name: GUTIERREZ-ANDRADE

Date/Time of Test: 04/20/2007 00:52

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.081	00:39	
Air Blank	0.000	00:40	
Subject Sample #1	0.000	00:43	Volume Not Met
Breath Volume	0.566	00:43	
Air Blank	0.000	00:43	
Air Blank	0.000	00:45	
Subject Sample #2	0.000	00:48	
Breath Volume	1.226	00:48	
Air Blank	0.000	00:48	
Air Blank	0.000	00:50	
Subject Sample #3	0.000	00:51	
Breath Volume	1.222	00:51	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 05/30/2007 15:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001007

Instrument Registered To: HAINES CITY PD

Software Version: 8100.27

Breath Test Operator: MONTALBANO, JONATHAN, D

Breath Test Operator Agency: Polk County SO

Arresting Officer: MONTALBANO, JONATHAN

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 04/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 22:57

Gender: Male

Observation Period Began: 23:30

Subject's Last Name: KUCHERA

Date/Time of Test: 04/27/2007 23:57

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample #1	0.119	23:52
Breath Volume	2.937	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample #2	0.121	23:55
Breath Volume	3.250	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 05/30/2007 15:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001007

Instrument Registered To: HAINES CITY PD

Software Version: 8100.27

Breath Test Operator: TEAGUE, MICHAEL, E

Breath Test Operator Agency: Haines City PD

Arresting Officer: TEAGUE, MICHAEL

Arresting Officer Agency: Haines City PD

Last Agency Inspection: 04/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 02:22

Gender: Male

Observation Period Began: 03:20

Subject's Last Name: MARTINEZ-ROJAS

Date/Time of Test: 04/28/2007 03:56

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Subject Sample #1	0.194	03:51
Breath Volume	1.667	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample #2	0.199	03:54
Breath Volume	2.222	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 05/30/2007 15:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001007

Instrument Registered To: HAINES CITY PD

Software Version: 8100.27

Breath Test Operator: TEAGUE, MICHAEL, E

Breath Test Operator Agency: Haines City PD

Arresting Officer: CAAMANO, JUAN

Arresting Officer Agency: Haines City PD

Last Agency Inspection: 04/17/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 03:40

Gender: Male

Observation Period Began: 04:30

Subject's Last Name: SANDOVAL

Date/Time of Test: 04/28/2007 05:02

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.079	04:59	
Air Blank	0.000	05:00	
Subject Sample #1		05:02	Sequence Aborted
Breath Volume	0.000	05:02	
Air Blank	0.000	05:02	

Data Download Date/Time: 05/30/2007 15:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001008

Instrument Registered To: WEST MELBOURNE PD

Software Version: 8100.27

Breath Test Operator: CANTEES, JUSTIN, A

Breath Test Operator Agency: West Melbourne PD

Arresting Officer: CANTEES, JUSTIN

Arresting Officer Agency: West Melbourne PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 02:27

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: ALLMAN

Date/Time of Test: 04/20/2007 03:15

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Subject Sample #1	0.206	03:10
Breath Volume	4.015	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample #2	0.199	03:13
Breath Volume	3.753	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 04/29/2007 08:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001010

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: THOMAS, GLEN, P

Breath Test Operator Agency: Brevard County SO

Arresting Officer: THOMAS, GLEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 13:27

Gender: Female

Observation Period Began: 13:27

Subject's Last Name: DUVALL

Date/Time of Test: 04/13/2007 15:16

Diagnostics Check	OK	15:09
Air Blank	0.000	15:10
Control Test	0.081	15:10
Air Blank	0.000	15:10
Subject Sample #1	0.000	15:11
Breath Volume	2.390	15:11
Air Blank	0.000	15:12
Air Blank	0.000	15:14
Subject Sample #2	0.000	15:14
Breath Volume	1.960	15:14
Air Blank	0.000	15:15
Control Test	0.080	15:15
Air Blank	0.000	15:15
Diagnostics Check	OK	15:16

Data Download Date/Time: 05/09/2007 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001010

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: THOMAS, GLEN, P

Breath Test Operator Agency: Brevard County SO

Arresting Officer: THOMAS, GLEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 21:41

Gender: Male

Observation Period Began: 21:41

Subject's Last Name: ARCHILA

Date/Time of Test: 04/14/2007 22:14

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.079	22:09
Air Blank	0.000	22:09
Subject Sample #1	0.150	22:10
Breath Volume	2.636	22:10
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample #2	0.141	22:13
Breath Volume	2.343	22:13
Air Blank	0.000	22:13
Control Test	0.078	22:14
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 05/09/2007 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001010

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: WHEELER, WENDY, M

Breath Test Operator Agency: Brevard County SO

Arresting Officer: WHEELER, WENDY

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 22:29

Gender: Male

Observation Period Began: 22:29

Subject's Last Name: INMAN

Date/Time of Test: 04/14/2007 23:16

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.079	23:05	
Air Blank	0.000	23:06	
Subject Sample #1	0.202	23:09	Volume Not Met
Breath Volume	0.390	23:09	
Air Blank	0.000	23:09	
Air Blank	0.000	23:11	
Subject Sample #2	0.168	23:14	Volume Not Met
Breath Volume	0.351	23:14	
Air Blank	0.000	23:15	
Control Test	0.077	23:15	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 05/09/2007 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001010

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: HARRISON, JAMES, C

Breath Test Operator Agency: Brevard County SO

Arresting Officer: HARRISON, JAMES

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:23

Gender: Male

Observation Period Began: 01:23

Subject's Last Name: RODRIGUEZ

Date/Time of Test: 04/15/2007 02:01

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.078	01:52
Air Blank	0.000	01:53
Subject Sample #1	0.147	01:53
Breath Volume	2.023	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.117	01:57
Breath Volume	1.101	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample #3	0.139	02:00
Breath Volume	3.093	02:00
Air Blank	0.000	02:00
Control Test	0.078	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 05/09/2007 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001010

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: HARRISON, JAMES, C

Breath Test Operator Agency: Brevard County SO

Arresting Officer: THOMAS, GLEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:36

Gender: Male

Observation Period Began: 01:36

Subject's Last Name: SMITH

Date/Time of Test: 04/15/2007 02:19

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.079	02:07	
Air Blank	0.000	02:08	
Subject Sample #1	0.188	02:11	Volume Not Met
Breath Volume	0.562	02:11	
Air Blank	0.000	02:12	
Air Blank	0.000	02:13	
Subject Sample #2	0.215	02:14	
Breath Volume	1.632	02:14	
Air Blank	0.000	02:15	
Air Blank	0.000	02:16	
Subject Sample #3	0.205	02:17	
Breath Volume	1.792	02:17	
Air Blank	0.000	02:18	
Control Test	0.075	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 05/09/2007 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001010

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: THOMAS, GLEN, P

Breath Test Operator Agency: Brevard County SO

Arresting Officer: THOMAS, GLEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 22:48

Gender: Female

Observation Period Began: 22:48

Subject's Last Name: FRANCIS

Date/Time of Test: 04/19/2007 23:23

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:18
Subject Sample #1	0.264	23:18
Breath Volume	1.769	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample #2	0.276	23:21
Breath Volume	1.429	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 05/09/2007 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001010

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: THOMAS, GLEN, P

Breath Test Operator Agency: Brevard County SO

Arresting Officer: THOMAS, GLEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 23:46

Gender: Male

Observation Period Began: 23:46

Subject's Last Name: JAMES

Date/Time of Test: 04/21/2007 00:24

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.081	00:18	
Air Blank	0.000	00:18	
Subject Sample #1		00:21	No Sample Provided
Breath Volume	0.000	00:21	
Air Blank	0.000	00:22	
Air Blank	0.000	00:24	
Subject Sample #2		00:24	Sequence Aborted
Breath Volume	0.000	00:24	
Air Blank	0.000	00:24	

Data Download Date/Time: 05/09/2007 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001010

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: THOMAS, GLEN, P

Breath Test Operator Agency: Brevard County SO

Arresting Officer: THOMAS, GLEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 01:30

Subject's Last Name: COLLINS

Date/Time of Test: 04/26/2007 02:03

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Subject Sample #1	0.090	01:58
Breath Volume	3.125	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample #2	0.089	02:01
Breath Volume	2.628	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 05/09/2007 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001010

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: THOMAS, GLEN, P

Breath Test Operator Agency: Brevard County SO

Arresting Officer: WHEELER, WENDY

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 21:24

Gender: Male

Observation Period Began: 21:24

Subject's Last Name: COWELL

Date/Time of Test: 04/26/2007 22:19

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:14
Air Blank	0.000	22:14
Subject Sample #1	0.024	22:15
Breath Volume	2.734	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample #2	0.022	22:17
Breath Volume	1.488	22:17
Air Blank	0.000	22:18
Control Test	0.079	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 05/09/2007 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001010

Instrument Registered To: BREVARD S.O.

Software Version: 8100.27

Breath Test Operator: THOMAS, GLEN, P

Breath Test Operator Agency: Brevard County SO

Arresting Officer: THOMAS, GLEN

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/04/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 23:57

Gender: Male

Observation Period Began: 23:57

Subject's Last Name: SANTIAGO

Date/Time of Test: 04/27/2007 00:36

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.080	00:32	
Air Blank	0.000	00:32	
Subject Sample #1	0.000	00:33	
Breath Volume	3.000	00:33	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample #2	0.000	00:35	
Breath Volume	1.777	00:35	
Air Blank		00:36	RFI Detect
Air Blank	0.000	00:36	

Data Download Date/Time: 05/09/2007 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: CHERRY, TYNA, M

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: STANLEY, T. L.

Arresting Officer Agency: Atlantic Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:12

Gender: Male

Observation Period Began: 23:58

Subject's Last Name: GETTS

Date/Time of Test: 04/01/2007 00:26

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:20	
Control Test	0.080	00:20	
Air Blank	0.000	00:20	
Subject Sample #1	0.088	00:21	
Breath Volume	2.707	00:21	
Air Blank	0.000	00:22	
Air Blank	0.000	00:24	
Subject Sample #2	0.088	00:24	
Breath Volume	2.878	00:24	
Air Blank	0.000	00:25	
Control Test	0.080	00:25	
Air Blank	0.000	00:26	
Diagnostics Check	Fail	00:26	Temperature Regulation Test
Air Blank	0.000	00:26	

Data Download Date/Time: 04/16/2007 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BAKER, DAVID, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BYRNE, J

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 20:40

Gender: Female

Observation Period Began: 22:25

Subject's Last Name: WALTON

Date/Time of Test: 04/04/2007 22:52

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Subject Sample #1	0.057	22:47
Breath Volume	1.214	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample #2	0.062	22:50
Breath Volume	1.121	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 04/16/2007 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO
Software Version: 8100.27
Breath Test Operator: CLAYMAN, EDWARD, B
Breath Test Operator Agency: Jacksonville SO
Arresting Officer: GUTHRIE, K
Arresting Officer Agency: Jacksonville SO
Last Agency Inspection: 03/27/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 630702I
Dry Gas Expiry Date: 11/06/2008
Violation Code: DUI
Arrest Time: 18:15
Gender: Female
Observation Period Began: 19:39

Subject's Last Name: WALTERS

Date/Time of Test: 04/08/2007 20:08

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.082	20:02
Air Blank	0.000	20:03
Subject Sample #1	0.000	20:04
Breath Volume	1.921	20:04
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample #2	0.000	20:07
Breath Volume	1.960	20:07
Air Blank	0.000	20:07
Control Test	0.081	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 04/16/2007 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BAKER, DAVID, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BORGERT, M.

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 17:20

Gender: Female

Observation Period Began: 18:45

Subject's Last Name: CANALES

Date/Time of Test: 04/13/2007 19:23

Diagnostics Check	OK	19:16
Air Blank	0.000	19:17
Control Test	0.081	19:17
Air Blank	0.000	19:17
Subject Sample #1	0.211	19:19
Breath Volume	1.578	19:19
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample #2	0.213	19:22
Breath Volume	1.675	19:22
Air Blank	0.000	19:22
Control Test	0.080	19:23
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 04/16/2007 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BAKER, DAVID, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: LOGAN, D

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 16:20

Gender: Male

Observation Period Began: 17:50

Subject's Last Name: GAZZAM

Date/Time of Test: 04/16/2007 18:16

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.080	18:10
Air Blank	0.000	18:11
Subject Sample #1	0.193	18:12
Breath Volume	2.796	18:12
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample #2	0.192	18:15
Breath Volume	2.617	18:15
Air Blank	0.000	18:15
Control Test	0.081	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 04/27/2007 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: HAMLIN, DAN

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:05

Gender: Female

Observation Period Began: 02:42

Subject's Last Name: STONEBURNER

Date/Time of Test: 04/18/2007 03:11

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:05
Subject Sample #1	0.156	03:06
Breath Volume	2.410	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample #2	0.160	03:09
Breath Volume	2.566	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 04/27/2007 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BAKER, DAVID, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BLACKSTONE, P.

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 15:30

Gender: Male

Observation Period Began: 17:00

Subject's Last Name: ROBERTS

Date/Time of Test: 04/18/2007 17:38

Diagnostics Check	OK	17:31
Air Blank	0.000	17:32
Control Test	0.080	17:32
Air Blank	0.000	17:32
Subject Sample #1	0.154	17:33
Breath Volume	1.972	17:33
Air Blank	0.000	17:34
Air Blank	0.000	17:35
Subject Sample #2	0.144	17:37
Breath Volume	1.449	17:37
Air Blank	0.000	17:37
Control Test	0.081	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time: 04/27/2007 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CHRISTMAS, F

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:15

Gender: Female

Observation Period Began: 03:10

Subject's Last Name: ELGIE

Date/Time of Test: 04/20/2007 03:37

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:32
Air Blank	0.000	03:32
Subject Sample #1	0.170	03:33
Breath Volume	3.019	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.159	03:36
Breath Volume	1.875	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/27/2007 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: MORRIS, DONNIE

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:10

Gender: Male

Observation Period Began: 03:55

Subject's Last Name: REDSHAW

Date/Time of Test: 04/20/2007 04:25

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:22	
Control Test	0.079	04:23	
Air Blank	0.000	04:23	
Subject Sample #1		04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.080	04:24	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 04/27/2007 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: WOOLERY, W

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:52

Gender: Female

Observation Period Began: 05:10

Subject's Last Name: VAZQUEZ

Date/Time of Test: 04/20/2007 05:37

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.079	05:31
Air Blank	0.000	05:31
Subject Sample #1	0.104	05:32
Breath Volume	1.925	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample #2	0.105	05:35
Breath Volume	1.839	05:35
Air Blank	0.000	05:36
Control Test	0.079	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 04/27/2007 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: FROST, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:20

Gender: Female

Observation Period Began: 00:30

Subject's Last Name: LOYD

Date/Time of Test: 04/22/2007 01:02

Diagnostics Check	OK	00:59	
Air Blank	0.000	01:00	
Control Test	0.079	01:00	
Air Blank	0.000	01:01	
Subject Sample #1		01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.079	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 04/27/2007 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: JACKSON, R

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:15

Gender: Female

Observation Period Began: 01:00

Subject's Last Name: BEEBE

Date/Time of Test: 04/22/2007 01:32

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Subject Sample #1	0.150	01:27
Breath Volume	2.742	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample #2	0.158	01:30
Breath Volume	2.632	01:30
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 04/27/2007 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: HICKS, B

Arresting Officer Agency: Atlantic Beach PD

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:07

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: HANNIS

Date/Time of Test: 04/22/2007 02:24

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.079	02:21	
Air Blank	0.000	02:22	
Subject Sample #1		02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 04/27/2007 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO
Software Version: 8100.27
Breath Test Operator: PRUETT, RICHARD, L
Breath Test Operator Agency: Jacksonville SO
Arresting Officer: REGESTER, S
Arresting Officer Agency: Jacksonville SO
Last Agency Inspection: 04/16/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 630702I
Dry Gas Expiry Date: 11/06/2008
Violation Code: DUI
Arrest Time: 03:00
Gender: Male
Observation Period Began: 03:50

Subject's Last Name: GOBER

Date/Time of Test: 04/22/2007 04:16

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Subject Sample #1	0.211	04:12
Breath Volume	1.148	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample #2	0.219	04:15
Breath Volume	2.675	04:15
Air Blank	0.000	04:15
Control Test	0.081	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 04/27/2007 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001018

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BAKER, DAVID, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: RYAN, J.

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 16:00

Gender: Male

Observation Period Began: 17:38

Subject's Last Name: FULFORD

Date/Time of Test: 04/27/2007 18:15

Diagnostics Check	OK	18:08
Air Blank	0.000	18:09
Control Test	0.080	18:09
Air Blank	0.000	18:10
Subject Sample #1	0.012	18:10
Breath Volume	2.937	18:10
Air Blank	0.000	18:11
Air Blank	0.000	18:13
Subject Sample #2	0.010	18:13
Breath Volume	1.601	18:13
Air Blank	0.000	18:14
Control Test	0.080	18:14
Air Blank	0.000	18:14
Diagnostics Check	OK	18:15

Data Download Date/Time: 05/01/2007 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MURPHY, JAMES,

Breath Test Operator Agency: Highlands County SO

Arresting Officer: BARLAUG, G

Arresting Officer Agency: Sebring PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 02:20

Gender: Female

Observation Period Began: 02:35

Subject's Last Name: BELL

Date/Time of Test: 04/01/2007 03:01

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Subject Sample #1	0.161	02:56
Breath Volume	1.632	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.156	02:59
Breath Volume	1.363	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:01

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MURPHY, JAMES,

Breath Test Operator Agency: Highlands County SO

Arresting Officer: LOW, B

Arresting Officer Agency: Sebring PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 23:28

Gender: Male

Observation Period Began: 23:40

Subject's Last Name: OLIVER

Date/Time of Test: 04/05/2007 00:06

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.077	00:01
Air Blank	0.000	00:01
Subject Sample #1	0.000	00:02
Breath Volume	2.332	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample #2	0.000	00:05
Breath Volume	2.156	00:05
Air Blank	0.000	00:05
Control Test	0.078	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MURPHY, JAMES,

Breath Test Operator Agency: Highlands County SO

Arresting Officer: THOMAS, T

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 00:56

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: MOOREHEAD

Date/Time of Test: 04/06/2007 01:38

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.077	01:36	
Air Blank	0.000	01:36	
Subject Sample #1		01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.078	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: SWEET, DAVID, N

Breath Test Operator Agency: Highlands County SO

Arresting Officer: SINGLES, ROBINN

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 14:16

Gender: Male

Observation Period Began: 15:09

Subject's Last Name: LOCKLEAR

Date/Time of Test: 04/06/2007 15:37

Diagnostics Check	OK	15:30
Air Blank	0.000	15:30
Control Test	0.079	15:31
Air Blank	0.000	15:31
Subject Sample #1	0.222	15:32
Breath Volume	1.578	15:32
Air Blank	0.000	15:33
Air Blank	0.000	15:35
Subject Sample #2	0.225	15:35
Breath Volume	1.179	15:35
Air Blank	0.000	15:36
Control Test	0.077	15:36
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: RITENOUR, CHRISTOPHER, D

Breath Test Operator Agency: Highlands County SO

Arresting Officer: CASTILLO, DAVID

Arresting Officer Agency: Sebring PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:15

Gender: Male

Observation Period Began: 22:47

Subject's Last Name: RIVERA

Date/Time of Test: 04/07/2007 23:15

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.079	23:09
Air Blank	0.000	23:09
Subject Sample #1	0.170	23:11
Breath Volume	2.734	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample #2	0.166	23:14
Breath Volume	2.234	23:14
Air Blank	0.000	23:15
Control Test	0.078	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: RITENOUR, CHRISTOPHER, D

Breath Test Operator Agency: Highlands County SO

Arresting Officer: BARLAUG, GREG

Arresting Officer Agency: Sebring PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:45

Gender: Male

Observation Period Began: 01:14

Subject's Last Name: MARTINEZ

Date/Time of Test: 04/08/2007 01:46

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.078	01:35
Air Blank	0.000	01:36
Subject Sample #1	0.232	01:37
Breath Volume	1.496	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample #2	0.206	01:40
Breath Volume	2.222	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample #3	0.204	01:44
Breath Volume	2.500	01:44
Air Blank	0.000	01:45
Control Test	0.078	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: RITENOUR, CHRISTOPHER, D

Breath Test Operator Agency: Highlands County SO

Arresting Officer: BUEFORD, SEAN

Arresting Officer Agency: Sebring PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:47

Gender: Male

Observation Period Began: 03:14

Subject's Last Name: DEAN

Date/Time of Test: 04/08/2007 03:46

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:36	
Air Blank	0.000	03:36	
Subject Sample #1		03:39	No Sample Provided
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample #2		03:45	No Sample Provided
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Control Test	0.079	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: RITENOUR, CHRISTOPHER, D

Breath Test Operator Agency: Highlands County SO

Arresting Officer: WILLIS, ADAM

Arresting Officer Agency: Avon Park PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:32

Gender: Male

Observation Period Began: 01:55

Subject's Last Name: DORSEY

Date/Time of Test: 04/09/2007 02:22

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:16
Air Blank	0.000	02:16
Subject Sample #1	0.109	02:17
Breath Volume	2.890	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.102	02:20
Breath Volume	2.625	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: DAVIDSON, L

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 08:30

Gender: Male

Observation Period Began: 08:30

Subject's Last Name: OTERO

Date/Time of Test: 04/09/2007 09:17

Diagnostics Check	OK	09:10
Air Blank	0.000	09:11
Control Test	0.079	09:11
Air Blank	0.000	09:12
Subject Sample #1	0.095	09:12
Breath Volume	3.144	09:12
Air Blank	0.000	09:13
Air Blank	0.000	09:15
Subject Sample #2	0.097	09:15
Breath Volume	3.089	09:15
Air Blank	0.000	09:16
Control Test	0.078	09:16
Air Blank	0.000	09:17
Diagnostics Check	OK	09:17

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: DAVIDSON, L

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 12:00

Gender: Male

Observation Period Began: 12:00

Subject's Last Name: OTERO

Date/Time of Test: 04/09/2007 12:31

Diagnostics Check	OK	12:25
Air Blank	0.000	12:25
Control Test	0.078	12:25
Air Blank	0.000	12:26
Subject Sample #1	0.034	12:26
Breath Volume	2.953	12:26
Air Blank	0.000	12:27
Air Blank	0.000	12:29
Subject Sample #2	0.034	12:29
Breath Volume	2.734	12:29
Air Blank	0.000	12:30
Control Test	0.079	12:30
Air Blank	0.000	12:31
Diagnostics Check	OK	12:31

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MURPHY, JAMES,

Breath Test Operator Agency: Highlands County SO

Arresting Officer: THOMAS, T

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:27

Gender: Male

Observation Period Began: 22:45

Subject's Last Name: BLACK

Date/Time of Test: 04/10/2007 23:15

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.077	23:09
Air Blank	0.000	23:10
Subject Sample #1	0.000	23:10
Breath Volume	2.332	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample #2	0.000	23:13
Breath Volume	2.597	23:13
Air Blank	0.000	23:14
Control Test	0.077	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:15

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: SWEET, DAVID, N

Breath Test Operator Agency: Highlands County SO

Arresting Officer: BARLAUG, GREG

Arresting Officer Agency: Sebring PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:24

Gender: Male

Observation Period Began: 23:43

Subject's Last Name: SLAVENS

Date/Time of Test: 04/12/2007 00:06

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.077	00:05	
Air Blank		00:06	RFI Detect
Air Blank	0.000	00:06	

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: SWEET, DAVID, N

Breath Test Operator Agency: Highlands County SO

Arresting Officer: BARLAUG, GREG

Arresting Officer Agency: Sebring PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:24

Gender: Male

Observation Period Began: 23:43

Subject's Last Name: SLAVENS

Date/Time of Test: 04/12/2007 00:14

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.078	00:08
Air Blank	0.000	00:09
Subject Sample #1	0.140	00:09
Breath Volume	4.105	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample #2	0.142	00:13
Breath Volume	3.824	00:13
Air Blank	0.000	00:13
Control Test	0.077	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: SWEET, DAVID, N

Breath Test Operator Agency: Highlands County SO

Arresting Officer: MCCLELLAN, LARRY

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:50

Gender: Female

Observation Period Began: 21:30

Subject's Last Name: STEVENSON

Date/Time of Test: 04/12/2007 21:57

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.077	21:52
Air Blank	0.000	21:52
Subject Sample #1	0.215	21:53
Breath Volume	2.000	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample #2	0.210	21:56
Breath Volume	1.812	21:56
Air Blank	0.000	21:56
Control Test	0.077	21:57
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: SWEET, DAVID, N

Breath Test Operator Agency: Highlands County SO

Arresting Officer: BARLAUG, GREG

Arresting Officer Agency: Sebring PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:12

Gender: Male

Observation Period Began: 22:06

Subject's Last Name: TAYLOR

Date/Time of Test: 04/12/2007 22:35

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.077	22:29
Air Blank	0.000	22:29
Subject Sample #1	0.117	22:30
Breath Volume	3.652	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample #2	0.120	22:33
Breath Volume	3.144	22:33
Air Blank	0.000	22:34
Control Test	0.078	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: CUTOLO, M

Arresting Officer Agency: Sebring PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:16

Gender: Male

Observation Period Began: 23:45

Subject's Last Name: GRANTHAN

Date/Time of Test: 04/14/2007 00:13

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.078	00:07
Air Blank	0.000	00:08
Subject Sample #1	0.140	00:08
Breath Volume	2.082	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample #2	0.146	00:11
Breath Volume	2.679	00:11
Air Blank	0.000	00:12
Control Test	0.078	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: THOMAS, T

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:12

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: SMITH

Date/Time of Test: 04/14/2007 03:05

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.078	03:02	
Air Blank	0.000	03:03	
Subject Sample #1		03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.078	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: HARBAUGH, C

Arresting Officer Agency: Avon Park PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:24

Gender: Male

Observation Period Began: 03:10

Subject's Last Name: VILLATORO

Date/Time of Test: 04/14/2007 03:46

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.077	03:31	
Air Blank	0.000	03:32	
Subject Sample #1	0.161	03:34	
Breath Volume	2.160	03:34	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample #2	0.132	03:39	Volume Not Met
Breath Volume	0.261	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample #3	0.154	03:45	Volume Not Met
Breath Volume	0.121	03:45	
Air Blank	0.000	03:46	
Control Test	0.077	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MURPHY, JAMES,

Breath Test Operator Agency: Highlands County SO

Arresting Officer: HARBAUGH, C

Arresting Officer Agency: Avon Park PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 04:21

Gender: Male

Observation Period Began: 05:38

Subject's Last Name: SIMMONS

Date/Time of Test: 04/14/2007 06:04

Diagnostics Check	OK	05:58
Air Blank	0.000	05:59
Control Test	0.077	05:59
Air Blank	0.000	06:00
Subject Sample #1	0.152	06:00
Breath Volume	2.167	06:00
Air Blank	0.000	06:01
Air Blank	0.000	06:02
Subject Sample #2	0.145	06:03
Breath Volume	3.128	06:03
Air Blank	0.000	06:03
Control Test	0.077	06:04
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MURPHY, JAMES,

Breath Test Operator Agency: Highlands County SO

Arresting Officer: NEW, S

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 06:17

Gender: Male

Observation Period Began: 06:55

Subject's Last Name: CARIDI

Date/Time of Test: 04/14/2007 07:18

Diagnostics Check	OK	07:16	
Air Blank	0.000	07:16	
Control Test	0.076	07:16	
Air Blank	0.000	07:17	
Subject Sample #1		07:17	Subject Test Refused
Breath Volume	0.000	07:17	
Air Blank	0.000	07:17	
Control Test	0.077	07:18	
Air Blank	0.000	07:18	
Diagnostics Check	OK	07:18	

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: MARCUS, M

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:13

Gender: Male

Observation Period Began: 23:42

Subject's Last Name: WAY

Date/Time of Test: 04/15/2007 00:11

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.077	00:05
Air Blank	0.000	00:06
Subject Sample #1	0.081	00:06
Breath Volume	1.511	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample #2	0.075	00:10
Breath Volume	1.371	00:10
Air Blank	0.000	00:10
Control Test	0.078	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: MIDDELTON, A

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:30

Gender: Female

Observation Period Began: 00:40

Subject's Last Name: OBERDIER

Date/Time of Test: 04/15/2007 01:14

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.078	01:01	
Air Blank	0.000	01:02	
Subject Sample #1	0.071	01:03	
Breath Volume	1.179	01:03	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample #2	0.074	01:08	Volume Not Met
Breath Volume	0.667	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample #3	0.074	01:12	
Breath Volume	1.695	01:12	
Air Blank	0.000	01:13	
Control Test	0.078	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: THOMAS, T

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:01

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: MENDOZA

Date/Time of Test: 04/15/2007 01:39

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.078	01:36	
Air Blank	0.000	01:37	
Subject Sample #1		01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	
Control Test	0.078	01:38	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: ZAHLER, J

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:24

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: BRONSON

Date/Time of Test: 04/15/2007 03:26

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample #1	0.057	03:22
Breath Volume	2.937	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample #2	0.056	03:25
Breath Volume	3.050	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: CLOUD, J

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:30

Gender: Male

Observation Period Began: 04:25

Subject's Last Name: FLORES

Date/Time of Test: 04/16/2007 04:58

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.078	04:46	
Air Blank	0.000	04:47	
Subject Sample #1	0.165	04:50	Volume Not Met
Breath Volume	0.273	04:50	
Air Blank	0.000	04:51	
Air Blank	0.000	04:52	
Subject Sample #2	0.174	04:53	
Breath Volume	2.171	04:53	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample #3	0.172	04:56	
Breath Volume	2.078	04:56	
Air Blank	0.000	04:57	
Control Test	0.078	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MURPHY, JAMES,

Breath Test Operator Agency: Highlands County SO

Arresting Officer: MILLER, P

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 07:44

Gender: Female

Observation Period Began: 08:49

Subject's Last Name: MADRAY

Date/Time of Test: 04/18/2007 09:13

Diagnostics Check	OK	09:09	
Air Blank	0.000	09:09	
Control Test	0.077	09:10	
Air Blank	0.000	09:10	
Subject Sample #1	0.000	09:10	
Breath Volume	1.449	09:10	
Air Blank	0.000	09:11	
Air Blank		09:13	RFI Detect
Air Blank		09:13	RFI Detect
Air Blank		09:13	RFI Detect

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MURPHY, JAMES,

Breath Test Operator Agency: Highlands County SO

Arresting Officer: MILLER, P

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 07:44

Gender: Female

Observation Period Began: 08:49

Subject's Last Name: MADRAY

Date/Time of Test: 04/18/2007 09:22

Diagnostics Check	OK	09:16
Air Blank	0.000	09:16
Control Test	0.077	09:17
Air Blank	0.000	09:17
Subject Sample #1	0.000	09:18
Breath Volume	1.398	09:18
Air Blank	0.000	09:18
Air Blank	0.000	09:20
Subject Sample #2	0.000	09:21
Breath Volume	1.554	09:21
Air Blank	0.000	09:21
Control Test	0.077	09:21
Air Blank	0.000	09:22
Diagnostics Check	OK	09:22

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: THOMAS, T

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: GOODALL

Date/Time of Test: 04/19/2007 03:07

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.077	03:01
Air Blank	0.000	03:02
Subject Sample #1	0.107	03:03
Breath Volume	2.218	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample #2	0.101	03:06
Breath Volume	1.621	03:06
Air Blank	0.000	03:06
Control Test	0.077	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: BRACKETT, MARK, O

Breath Test Operator Agency: Highlands County SO

Arresting Officer: MARCUS, MARK

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 16:55

Gender: Male

Observation Period Began: 17:25

Subject's Last Name: DOMINGUEZ

Date/Time of Test: 04/20/2007 17:52

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.077	17:46
Air Blank	0.000	17:47
Subject Sample #1	0.192	17:47
Breath Volume	2.875	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample #2	0.187	17:50
Breath Volume	1.144	17:50
Air Blank	0.000	17:51
Control Test	0.076	17:51
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: MURPHY, JAMES,

Breath Test Operator Agency: Highlands County SO

Arresting Officer: THOMAS, AMANDA

Arresting Officer Agency: OTHER State

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: ZERO TOL

Arrest Time: 11:15

Gender: Male

Observation Period Began: 11:26

Subject's Last Name: BENNETT

Date/Time of Test: 04/24/2007 12:06

Diagnostics Check	OK	12:00
Air Blank	0.000	12:01
Control Test	0.077	12:01
Air Blank	0.000	12:02
Subject Sample #1	0.059	12:02
Breath Volume	1.921	12:02
Air Blank	0.000	12:03
Air Blank	0.000	12:04
Subject Sample #2	0.058	12:05
Breath Volume	1.722	12:05
Air Blank	0.000	12:05
Control Test	0.078	12:06
Air Blank	0.000	12:06
Diagnostics Check	OK	12:06

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: DANIELSON, LORAN, L

Breath Test Operator Agency: Highlands County SO

Arresting Officer: GENTRY, WILLIAM

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:43

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: CARIAS

Date/Time of Test: 04/25/2007 03:34

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.078	03:30	
Air Blank	0.000	03:30	
Subject Sample #1		03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.078	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: RITENOUR, CHRISTOPHER, D

Breath Test Operator Agency: Highlands County SO

Arresting Officer: BUEFORD, SEAN

Arresting Officer Agency: Sebring PD

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:41

Gender: Male

Observation Period Began: 03:14

Subject's Last Name: HERBERT

Date/Time of Test: 04/29/2007 03:41

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Subject Sample #1	0.119	03:36
Breath Volume	2.300	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample #2	0.111	03:40
Breath Volume	1.503	03:40
Air Blank	0.000	03:40
Control Test	0.077	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001043

Instrument Registered To: HIGHLANDS COUNTY SO

Software Version: 8100.27

Breath Test Operator: RITENOUR, CHRISTOPHER, D

Breath Test Operator Agency: Highlands County SO

Arresting Officer: YOUNG, KENNETH

Arresting Officer Agency: Highlands County SO

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:25

Gender: Male

Observation Period Began: 03:50

Subject's Last Name: EASON

Date/Time of Test: 04/29/2007 04:17

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.078	04:11
Air Blank	0.000	04:12
Subject Sample #1	0.105	04:12
Breath Volume	1.394	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample #2	0.105	04:15
Breath Volume	1.605	04:15
Air Blank	0.000	04:16
Control Test	0.077	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 04/30/2007 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001044

Instrument Registered To: COCONUT CREEK PD

Software Version: 8100.27

Breath Test Operator: KARP, DARREN, M

Breath Test Operator Agency: Coconut Creek Pub Sf

Arresting Officer: ESCOBAR, A

Arresting Officer Agency: Coconut Creek Pub Sf

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 609301G

Dry Gas Expiry Date: 04/05/2008

Violation Code: DUI

Arrest Time: 02:56

Gender: Female

Observation Period Began: 03:12

Subject's Last Name: VOLK

Date/Time of Test: 04/02/2007 03:46

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.077	03:40
Air Blank	0.000	03:41
Subject Sample #1	0.181	03:41
Breath Volume	1.550	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample #2	0.186	03:44
Breath Volume	1.664	03:44
Air Blank	0.000	03:45
Control Test	0.077	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 04/03/2007 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001044

Instrument Registered To: COCONUT CREEK PD

Software Version: 8100.27

Breath Test Operator: EBERLY, DANIEL, R

Breath Test Operator Agency: Coconut Creek Pub Sf

Arresting Officer: WHITFORD, RYAN

Arresting Officer Agency: Coconut Creek Pub Sf

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 609301G

Dry Gas Expiry Date: 04/05/2008

Violation Code: DUI

Arrest Time: 00:22

Gender: Male

Observation Period Began: 00:38

Subject's Last Name: SAMPLES

Date/Time of Test: 04/03/2007 01:03

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.077	01:01	
Air Blank	0.000	01:01	
Subject Sample #1		01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:02	
Control Test	0.078	01:03	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 04/03/2007 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001044

Instrument Registered To: COCONUT CREEK PD

Software Version: 8100.27

Breath Test Operator: KARP, DARREN, M

Breath Test Operator Agency: Coconut Creek Pub Sf

Arresting Officer: ESCOBAR, A

Arresting Officer Agency: Coconut Creek Pub Sf

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 609301G

Dry Gas Expiry Date: 04/05/2008

Violation Code: DUI

Arrest Time: 00:44

Gender: Female

Observation Period Began: 01:01

Subject's Last Name: MONTI

Date/Time of Test: 04/10/2007 01:59

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:57	
Control Test	0.077	01:57	
Air Blank	0.000	01:58	
Subject Sample #1		01:58	Sequence Aborted
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	

Data Download Date/Time: 05/11/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001044

Instrument Registered To: COCONUT CREEK PD

Software Version: 8100.27

Breath Test Operator: KARP, DARREN, M

Breath Test Operator Agency: Coconut Creek Pub Sf

Arresting Officer: ESCOBAR, A

Arresting Officer Agency: Coconut Creek Pub Sf

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 609301G

Dry Gas Expiry Date: 04/05/2008

Violation Code: DUI

Arrest Time: 00:44

Gender: Female

Observation Period Began: 01:01

Subject's Last Name: MONTI

Date/Time of Test: 04/10/2007 02:03

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.078	02:02	
Air Blank	0.000	02:03	
Subject Sample #1		02:03	Sequence Aborted
Breath Volume	0.000	02:03	
Air Blank	0.000	02:03	

Data Download Date/Time: 05/11/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001044

Instrument Registered To: COCONUT CREEK PD

Software Version: 8100.27

Breath Test Operator: DANIELLO, STEVEN, J

Breath Test Operator Agency: Coconut Creek Pub Sf

Arresting Officer: MANCUSO, ANTHONY

Arresting Officer Agency: Coconut Creek Pub Sf

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 609301G

Dry Gas Expiry Date: 04/05/2008

Violation Code: DUI

Arrest Time: 02:35

Gender: Female

Observation Period Began: 02:55

Subject's Last Name: PAUL

Date/Time of Test: 04/15/2007 03:26

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.077	03:24	
Air Blank	0.000	03:24	
Subject Sample #1		03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.077	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 05/11/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001044

Instrument Registered To: COCONUT CREEK PD

Software Version: 8100.27

Breath Test Operator: DANIELLO, STEVEN, J

Breath Test Operator Agency: Coconut Creek Pub Sf

Arresting Officer: KARP, DARREN

Arresting Officer Agency: Coconut Creek Pub Sf

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 609301G

Dry Gas Expiry Date: 04/05/2008

Violation Code: DUI

Arrest Time: 21:52

Gender: Male

Observation Period Began: 22:15

Subject's Last Name: QUIRINDONGO

Date/Time of Test: 04/15/2007 22:44

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.077	22:36
Air Blank	0.000	22:36
Subject Sample #1	0.290	22:38
Breath Volume	1.410	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample #2	0.282	22:42
Breath Volume	1.796	22:42
Air Blank	0.000	22:43
Control Test	0.077	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 05/11/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001044

Instrument Registered To: COCONUT CREEK PD

Software Version: 8100.27

Breath Test Operator: DANIELLO, STEVEN, J

Breath Test Operator Agency: Coconut Creek Pub Sf

Arresting Officer: KARP, DARREN

Arresting Officer Agency: Coconut Creek Pub Sf

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 609301G

Dry Gas Expiry Date: 04/05/2008

Violation Code: DUI

Arrest Time: 03:31

Gender: Female

Observation Period Began: 03:50

Subject's Last Name: BALBIER

Date/Time of Test: 04/19/2007 04:30

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.077	04:18	
Air Blank	0.000	04:19	
Subject Sample #1		04:22	No Sample Provided
Breath Volume	0.000	04:22	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample #2	0.169	04:25	
Breath Volume	1.472	04:25	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample #3	0.162	04:28	
Breath Volume	2.101	04:28	
Air Blank	0.000	04:29	
Control Test	0.078	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 05/11/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001044

Instrument Registered To: COCONUT CREEK PD

Software Version: 8100.27

Breath Test Operator: EBERLY, DANIEL, R

Breath Test Operator Agency: Coconut Creek Pub Sf

Arresting Officer: ROBERTSON, MICHAEL

Arresting Officer Agency: Coconut Creek Pub Sf

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 609301G

Dry Gas Expiry Date: 04/05/2008

Violation Code: DUI

Arrest Time: 21:52

Gender: Male

Observation Period Began: 22:07

Subject's Last Name: HONORAT

Date/Time of Test: 04/21/2007 22:41

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.077	22:37	
Air Blank	0.000	22:38	
Subject Sample #1		22:40	Subject Test Refused
Breath Volume	0.226	22:40	
Air Blank	0.000	22:40	
Control Test	0.078	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 05/11/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001044

Instrument Registered To: COCONUT CREEK PD

Software Version: 8100.27

Breath Test Operator: EBERLY, DANIEL, R

Breath Test Operator Agency: Coconut Creek Pub Sf

Arresting Officer: EBERLY, DANIEL

Arresting Officer Agency: Coconut Creek Pub Sf

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 609301G

Dry Gas Expiry Date: 04/05/2008

Violation Code: DUI

Arrest Time: 00:25

Gender: Male

Observation Period Began: 00:55

Subject's Last Name: BRAYTON

Date/Time of Test: 04/22/2007 01:51

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.077	01:49	
Air Blank	0.000	01:49	
Subject Sample #1		01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Control Test	0.077	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 05/11/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001044

Instrument Registered To: COCONUT CREEK PD

Software Version: 8100.27

Breath Test Operator: DANIELLO, STEVEN, J

Breath Test Operator Agency: Coconut Creek Pub Sf

Arresting Officer: ESCOBAR, ALEX

Arresting Officer Agency: Coconut Creek Pub Sf

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 609301G

Dry Gas Expiry Date: 04/05/2008

Violation Code: DUI

Arrest Time: 01:14

Gender: Male

Observation Period Began: 01:35

Subject's Last Name: CARDOSO

Date/Time of Test: 04/29/2007 02:08

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.077	02:03
Air Blank	0.000	02:03
Subject Sample #1	0.186	02:03
Breath Volume	2.011	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample #2	0.193	02:06
Breath Volume	2.410	02:06
Air Blank	0.000	02:07
Control Test	0.077	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 05/11/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARWOOD, C, M

Breath Test Operator Agency: Clay County SO

Arresting Officer: FRAZHO, T

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 01:49

Gender: Male

Observation Period Began: 02:32

Subject's Last Name: DAMERON

Date/Time of Test: 04/01/2007 03:05

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	03:00
Subject Sample #1	0.222	03:00
Breath Volume	2.753	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample #2	0.227	03:04
Breath Volume	2.636	03:04
Air Blank	0.000	03:04
Control Test	0.077	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: PADGETT, MELINDA, S

Breath Test Operator Agency: Clay County SO

Arresting Officer: RICKS, LESTER

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 11:55

Gender: Female

Observation Period Began: 12:35

Subject's Last Name: HENRY

Date/Time of Test: 04/04/2007 13:07

Diagnostics Check	OK	12:58	
Air Blank	0.000	12:59	
Control Test	0.078	12:59	
Air Blank	0.000	12:59	
Subject Sample #1		13:03	No Sample Provided
Breath Volume	0.000	13:03	
Air Blank	0.000	13:03	
Air Blank	0.000	13:05	
Subject Sample #2		13:05	Subject Test Refused
Breath Volume	0.000	13:05	
Air Blank	0.000	13:06	
Control Test	0.078	13:06	
Air Blank	0.000	13:07	
Diagnostics Check	OK	13:07	

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WILKERSON, M, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: MARITATO, D

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 03:31

Gender: Female

Observation Period Began: 04:00

Subject's Last Name: TETZEL

Date/Time of Test: 04/06/2007 04:28

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.078	04:23	
Air Blank	0.000	04:23	
Subject Sample #1	0.192	04:24	
Breath Volume	2.597	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample #2	0.190	04:27	
Breath Volume	2.285	04:27	
Air Blank		04:27	Improper Sample
Air Blank	0.000	04:28	

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: DUFFY, DALE, V

Breath Test Operator Agency: Clay County SO

Arresting Officer: WARD, GLENN

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 04:00

Gender: Male

Observation Period Began: 04:32

Subject's Last Name: PREZAZ

Date/Time of Test: 04/12/2007 05:05

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.078	05:00
Air Blank	0.000	05:00
Subject Sample #1	0.150	05:01
Breath Volume	3.019	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample #2	0.130	05:04
Breath Volume	1.750	05:04
Air Blank	0.000	05:04
Control Test	0.076	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: MAIJALA, CRAIG, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: HYSLER, DAMON

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 01:45

Gender: Male

Observation Period Began: 02:15

Subject's Last Name: FITZHUGH

Date/Time of Test: 04/13/2007 03:20

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:14
Subject Sample #1	0.089	03:16
Breath Volume	2.515	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample #2	0.085	03:18
Breath Volume	1.792	03:18
Air Blank	0.000	03:19
Control Test	0.077	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: SMITH, S, E

Breath Test Operator Agency: Clay County SO

Arresting Officer: SIMMONS, B

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 02:20

Gender: Male

Observation Period Began: 03:30

Subject's Last Name: JENKINS

Date/Time of Test: 04/13/2007 04:03

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:58
Air Blank	0.000	03:58
Subject Sample #1	0.041	03:59
Breath Volume	2.964	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample #2	0.034	04:02
Breath Volume	1.390	04:02
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARWOOD, C, M

Breath Test Operator Agency: Clay County SO

Arresting Officer: SPENCE, G

Arresting Officer Agency: Orange Park PD

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 20:04

Gender: Female

Observation Period Began: 21:54

Subject's Last Name: GRIMSLEY

Date/Time of Test: 04/13/2007 22:30

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.077	22:24
Air Blank	0.000	22:24
Subject Sample #1	0.176	22:25
Breath Volume	2.742	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample #2	0.189	22:28
Breath Volume	2.820	22:28
Air Blank	0.000	22:29
Control Test	0.077	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WILKERSON, M, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: BROWN, T

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 22:03

Gender: Female

Observation Period Began: 22:40

Subject's Last Name: QUIGLEY

Date/Time of Test: 04/13/2007 23:09

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.078	23:01	
Air Blank	0.000	23:02	
Subject Sample #1		23:05	No Sample Provided
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Air Blank	0.000	23:07	
Subject Sample #2		23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:08	
Control Test	0.078	23:08	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARWOOD, C, M

Breath Test Operator Agency: Clay County SO

Arresting Officer: BROWN, T.J.

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 02:11

Gender: Male

Observation Period Began: 02:53

Subject's Last Name: HEY

Date/Time of Test: 04/14/2007 03:21

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:16
Subject Sample #1	0.205	03:16
Breath Volume	5.386	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample #2	0.195	03:19
Breath Volume	4.937	03:19
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HAMLETT, JOSHUA, D

Breath Test Operator Agency: Clay County SO

Arresting Officer: KEPUNGER, T

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: ADMIN

Arrest Time: 01:23

Gender: Male

Observation Period Began: 13:00

Subject's Last Name: CAMPBELL

Date/Time of Test: 04/14/2007 13:58

Diagnostics Check	OK	13:51
Air Blank	0.000	13:52
Control Test	0.077	13:52
Air Blank	0.000	13:53
Subject Sample #1	0.087	13:53
Breath Volume	2.128	13:53
Air Blank	0.000	13:54
Air Blank	0.000	13:56
Subject Sample #2	0.096	13:56
Breath Volume	3.421	13:56
Air Blank	0.000	13:57
Control Test	0.078	13:57
Air Blank	0.000	13:58
Diagnostics Check	OK	13:58

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: TOMLINSON, JAMES, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: DOPP, T

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 00:05

Gender: Male

Observation Period Began: 00:56

Subject's Last Name: WOHLHART

Date/Time of Test: 04/15/2007 01:24

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:18
Subject Sample #1	0.228	01:19
Breath Volume	3.296	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample #2	0.221	01:22
Breath Volume	2.785	01:22
Air Blank	0.000	01:23
Control Test	0.077	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: FRISCO, RICKY, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: BONSER, C

Arresting Officer Agency: Orange Park PD

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 01:57

Gender: Male

Observation Period Began: 03:30

Subject's Last Name: POTTER

Date/Time of Test: 04/18/2007 04:08

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.078	03:56	
Air Blank	0.000	03:56	
Subject Sample #1	0.134	03:59	Volume Not Met
Breath Volume	0.234	03:59	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample #2	0.141	04:03	
Breath Volume	1.785	04:03	
Air Blank	0.000	04:03	
Air Blank	0.000	04:05	
Subject Sample #3	0.146	04:06	
Breath Volume	2.695	04:06	
Air Blank	0.000	04:07	
Control Test	0.077	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HAMLETT, JOSHUA, D

Breath Test Operator Agency: Clay County SO

Arresting Officer: BONSER, C

Arresting Officer Agency: Orange Park PD

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: ADMIN

Arrest Time: 01:57

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: POTTER

Date/Time of Test: 04/18/2007 10:40

Diagnostics Check	OK	10:34
Air Blank	0.000	10:34
Control Test	0.078	10:35
Air Blank	0.000	10:35
Subject Sample #1	0.040	10:35
Breath Volume	2.324	10:35
Air Blank	0.000	10:36
Air Blank	0.000	10:38
Subject Sample #2	0.039	10:38
Breath Volume	3.000	10:38
Air Blank	0.000	10:39
Control Test	0.078	10:39
Air Blank	0.000	10:40
Diagnostics Check	OK	10:40

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ADAMSON, HEATHER, L

Breath Test Operator Agency: Clay County SO

Arresting Officer: ADAMSON, HEATHER

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: ADMIN

Arrest Time: 10:05

Gender: Female

Observation Period Began: 10:05

Subject's Last Name: NICHOLAS

Date/Time of Test: 04/19/2007 10:33

Diagnostics Check	OK	10:27
Air Blank	0.000	10:27
Control Test	0.078	10:28
Air Blank	0.000	10:28
Subject Sample #1	0.000	10:29
Breath Volume	2.214	10:29
Air Blank	0.000	10:29
Air Blank	0.000	10:31
Subject Sample #2	0.000	10:32
Breath Volume	2.910	10:32
Air Blank	0.000	10:32
Control Test	0.077	10:33
Air Blank	0.000	10:33
Diagnostics Check	OK	10:33

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARWOOD, C, M

Breath Test Operator Agency: Clay County SO

Arresting Officer: EHRENFELD, G

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 17:17

Gender: Male

Observation Period Began: 18:25

Subject's Last Name: MERCHANT

Date/Time of Test: 04/19/2007 19:03

Diagnostics Check	OK	18:56
Air Blank	0.000	18:57
Control Test	0.077	18:57
Air Blank	0.000	18:58
Subject Sample #1	0.193	18:58
Breath Volume	2.617	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample #2	0.190	19:02
Breath Volume	2.882	19:02
Air Blank	0.000	19:02
Control Test	0.077	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 04/22/2007 00:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARNOLD, WILLIAM, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: HYSLER, D

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 04/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 01:52

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: FOSTER

Date/Time of Test: 04/23/2007 03:04

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.078	02:58
Air Blank	0.000	02:59
Subject Sample #1	0.099	03:00
Breath Volume	2.515	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample #2	0.093	03:03
Breath Volume	1.316	03:03
Air Blank	0.000	03:03
Control Test	0.077	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 05/20/2007 05:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARNOLD, WILLIAM, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: ISARAJ, K

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 04/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 01:43

Gender: Male

Observation Period Began: 04:05

Subject's Last Name: WELLS

Date/Time of Test: 04/26/2007 04:31

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.077	04:25
Air Blank	0.000	04:26
Subject Sample #1	0.144	04:26
Breath Volume	3.757	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample #2	0.138	04:29
Breath Volume	3.058	04:29
Air Blank	0.000	04:30
Control Test	0.077	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 05/20/2007 05:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HOWARD, MICHAEL, J

Breath Test Operator Agency: Clay County SO

Arresting Officer: HOWARD, M

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 04/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: ADMIN

Arrest Time: 10:00

Gender: Male

Observation Period Began: 10:00

Subject's Last Name: THORNTON

Date/Time of Test: 04/26/2007 10:35

Diagnostics Check	OK	10:29
Air Blank	0.000	10:29
Control Test	0.077	10:29
Air Blank	0.000	10:30
Subject Sample #1	0.065	10:31
Breath Volume	2.207	10:31
Air Blank	0.000	10:31
Air Blank	0.000	10:33
Subject Sample #2	0.064	10:34
Breath Volume	2.078	10:34
Air Blank	0.000	10:34
Control Test	0.077	10:35
Air Blank	0.000	10:35
Diagnostics Check	OK	10:35

Data Download Date/Time: 05/20/2007 05:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARNOLD, WILLIAM, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: MAYLON, B

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 04/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 18:54

Gender: Male

Observation Period Began: 22:05

Subject's Last Name: VAUGHN

Date/Time of Test: 04/26/2007 22:32

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.078	22:30	
Air Blank	0.000	22:30	
Subject Sample #1		22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:31	
Control Test	0.078	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 05/20/2007 05:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEESE, J, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: SMITH, D

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 04/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 02:19

Gender: Female

Observation Period Began: 03:05

Subject's Last Name: WRIGHT

Date/Time of Test: 04/27/2007 03:33

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Subject Sample #1	0.132	03:28
Breath Volume	3.667	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample #2	0.134	03:32
Breath Volume	3.253	03:32
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 05/20/2007 05:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: DEESE, J, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: SMITH, D

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 04/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 23:45

Gender: Male

Observation Period Began: 23:49

Subject's Last Name: ROJAS DORGAN

Date/Time of Test: 04/28/2007 00:17

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:11
Subject Sample #1	0.215	00:12
Breath Volume	3.718	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample #2	0.204	00:15
Breath Volume	2.828	00:15
Air Blank	0.000	00:16
Control Test	0.078	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:17

Data Download Date/Time: 05/20/2007 05:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WENDORFF, DANIEL, E

Breath Test Operator Agency: Clay County SO

Arresting Officer: GRANT, J

Arresting Officer Agency: Orange Park PD

Last Agency Inspection: 04/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 06:49

Gender: Male

Observation Period Began: 09:20

Subject's Last Name: BOWMAN

Date/Time of Test: 04/28/2007 09:48

Diagnostics Check	OK	09:41
Air Blank	0.000	09:41
Control Test	0.077	09:42
Air Blank	0.000	09:42
Subject Sample #1	0.094	09:43
Breath Volume	1.804	09:43
Air Blank	0.000	09:44
Air Blank	0.000	09:46
Subject Sample #2	0.096	09:47
Breath Volume	2.074	09:47
Air Blank	0.000	09:47
Control Test	0.077	09:48
Air Blank	0.000	09:48
Diagnostics Check	OK	09:48

Data Download Date/Time: 05/20/2007 05:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: WILKERSON, M, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: C, BONSER

Arresting Officer Agency: Orange Park PD

Last Agency Inspection: 04/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 02:11

Gender: Male

Observation Period Began: 03:39

Subject's Last Name: SAUER

Date/Time of Test: 04/29/2007 04:02

Diagnostics Check	OK	03:59	
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:00	
Subject Sample #1		04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Control Test	0.078	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 05/20/2007 05:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HAMLETT, JOSHUA, D

Breath Test Operator Agency: Clay County SO

Arresting Officer: BONSER, C

Arresting Officer Agency: Orange Park PD

Last Agency Inspection: 04/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: ADMIN

Arrest Time: 09:00

Gender: Male

Observation Period Began: 10:00

Subject's Last Name: SAUER

Date/Time of Test: 04/29/2007 10:50

Diagnostics Check	OK	10:49	
Air Blank	0.000	10:50	
Control Test	0.078	10:50	RFI Detect
Air Blank	0.000	10:50	

Data Download Date/Time: 05/20/2007 05:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001046

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARWOOD, C, M

Breath Test Operator Agency: Clay County SO

Arresting Officer: SUTHERLAND, D

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 04/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 16:44

Gender: Male

Observation Period Began: 18:02

Subject's Last Name: MICKEY

Date/Time of Test: 04/29/2007 18:32

Diagnostics Check	OK	18:25
Air Blank	0.000	18:25
Control Test	0.078	18:25
Air Blank	0.000	18:26
Subject Sample #1	0.345	18:27
Breath Volume	4.398	18:27
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample #2	0.348	18:30
Breath Volume	4.164	18:30
Air Blank	0.000	18:31
Control Test	0.077	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:32

Data Download Date/Time: 05/20/2007 05:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001048

Instrument Registered To: LAKE ALFRED PD

Software Version: 8100.27

Breath Test Operator: BERNDT, RICHARD, W

Breath Test Operator Agency: Haines City PD

Arresting Officer: BLOSE, JEFFREY

Arresting Officer Agency: Lake Alfred PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 5187011

Dry Gas Expiry Date: 07/08/2007

Violation Code: DUI

Arrest Time: 15:32

Gender: Male

Observation Period Began: 16:20

Subject's Last Name: CRAMTON

Date/Time of Test: 04/07/2007 17:08

Diagnostics Check	OK	16:57	
Air Blank	0.000	16:57	
Control Test	0.079	16:58	
Air Blank	0.000	16:58	
Subject Sample #1		17:01	No Sample Provided
Breath Volume	0.000	17:01	
Air Blank	0.000	17:02	
Air Blank	0.000	17:04	
Subject Sample #2		17:07	No Sample Provided
Breath Volume	0.000	17:07	
Air Blank	0.000	17:07	
Control Test	0.080	17:08	
Air Blank	0.000	17:08	
Diagnostics Check	OK	17:08	

Data Download Date/Time: 04/30/2007 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001048

Instrument Registered To: LAKE ALFRED PD

Software Version: 8100.27

Breath Test Operator: GETTLE, ANTHONY, W

Breath Test Operator Agency: Lake Alfred PD

Arresting Officer: GETTLE, ANTHONY

Arresting Officer Agency: Lake Alfred PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 5187011

Dry Gas Expiry Date: 07/08/2007

Violation Code: DUI

Arrest Time: 23:46

Gender: Male

Observation Period Began: 00:38

Subject's Last Name: WEST JR

Date/Time of Test: 04/11/2007 01:21

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Subject Sample #1	0.025	01:17
Breath Volume	1.808	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample #2	0.024	01:19
Breath Volume	2.800	01:19
Air Blank	0.000	01:20
Control Test	0.077	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 04/30/2007 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001048

Instrument Registered To: LAKE ALFRED PD

Software Version: 8100.27

Breath Test Operator: RICHARDS, MICHAEL, B

Breath Test Operator Agency: Lake Alfred PD

Arresting Officer: RICHARDS, MICHAEL

Arresting Officer Agency: Lake Alfred PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 5187011

Dry Gas Expiry Date: 07/08/2007

Violation Code: DUI

Arrest Time: 23:47

Gender: Male

Observation Period Began: 00:00

Subject's Last Name: HORACE

Date/Time of Test: 04/21/2007 00:34

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:28
Subject Sample #1	0.186	00:29
Breath Volume	1.882	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample #2	0.188	00:32
Breath Volume	3.566	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 04/30/2007 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001048

Instrument Registered To: LAKE ALFRED PD

Software Version: 8100.27

Breath Test Operator: GETTLE, ANTHONY, W

Breath Test Operator Agency: Lake Alfred PD

Arresting Officer: GETTLE, ANTHONY

Arresting Officer Agency: Lake Alfred PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 5187011

Dry Gas Expiry Date: 07/08/2007

Violation Code: DUI

Arrest Time: 23:46

Gender: Male

Observation Period Began: 00:28

Subject's Last Name: JONES

Date/Time of Test: 04/28/2007 00:56

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample #1	0.139	00:52
Breath Volume	3.339	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:54
Subject Sample #2	0.132	00:55
Breath Volume	3.992	00:55
Air Blank	0.000	00:55
Control Test	0.077	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 04/30/2007 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO
Software Version: 8100.27
Breath Test Operator: TOCZYLOWSKI, MICHAEL, A
Breath Test Operator Agency: Pasco County SO
Arresting Officer: TOCZYLOWSKI, MICHAEL
Arresting Officer Agency: Pasco County SO
Last Agency Inspection: 03/13/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 617202I
Dry Gas Expiry Date: 06/23/2008
Violation Code: DUI
Arrest Time: 00:37
Gender: Male
Observation Period Began: 01:41

Subject's Last Name: ELDRIDGE

Date/Time of Test: 04/04/2007 02:08

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.082	02:03
Air Blank	0.000	02:03
Subject Sample #1	0.000	02:04
Breath Volume	2.269	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample #2	0.000	02:07
Breath Volume	2.515	02:07
Air Blank	0.000	02:07
Control Test	0.083	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: CARMEN, G, F

Breath Test Operator Agency: Pasco County SO

Arresting Officer: CARMEN, G

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 01:30

Subject's Last Name: LOVELL JR

Date/Time of Test: 04/04/2007 02:19

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.082	02:13
Air Blank	0.000	02:14
Subject Sample #1	0.136	02:14
Breath Volume	1.277	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample #2	0.150	02:17
Breath Volume	2.957	02:17
Air Blank	0.000	02:18
Control Test	0.083	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BELL, C,

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BELL, C

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 17:11

Gender: Female

Observation Period Began: 17:45

Subject's Last Name: KOZLOWSKI

Date/Time of Test: 04/05/2007 18:12

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:06	
Control Test	0.083	18:06	
Air Blank	0.000	18:07	
Subject Sample #1	0.000	18:07	
Breath Volume	1.578	18:07	
Air Blank	0.000	18:08	
Air Blank	0.000	18:10	
Subject Sample #2	0.000	18:11	
Breath Volume	1.523	18:11	
Air Blank	0.000	18:12	
Control Test	0.000	18:12	RFI Detect
Air Blank	0.000	18:12	

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BELL, C,

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BELL, C

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 17:11

Gender: Female

Observation Period Began: 17:45

Subject's Last Name: KOZLOWSKI

Date/Time of Test: 04/05/2007 18:33

Diagnostics Check	OK	18:26
Air Blank	0.000	18:26
Control Test	0.082	18:27
Air Blank	0.000	18:27
Subject Sample #1	0.000	18:28
Breath Volume	1.480	18:28
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample #2	0.000	18:31
Breath Volume	1.671	18:31
Air Blank	0.000	18:32
Control Test	0.083	18:32
Air Blank	0.000	18:33
Diagnostics Check	OK	18:33

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BALDWIN, ROBERT, K

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BALDWIN, ROBERT

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 00:23

Gender: Male

Observation Period Began: 00:50

Subject's Last Name: CHRISTY

Date/Time of Test: 04/06/2007 01:17

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.082	01:12
Air Blank	0.000	01:13
Subject Sample #1	0.079	01:13
Breath Volume	2.238	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample #2	0.076	01:16
Breath Volume	2.265	01:16
Air Blank	0.000	01:16
Control Test	0.082	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: TOCZYLOWSKI, MICHAEL, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: TOCZYLOWSKI, MICHAEL

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 03:35

Gender: Male

Observation Period Began: 03:47

Subject's Last Name: PFEIFFER

Date/Time of Test: 04/06/2007 04:16

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.082	04:11
Air Blank	0.000	04:11
Subject Sample #1	0.127	04:12
Breath Volume	2.039	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample #2	0.122	04:15
Breath Volume	2.328	04:15
Air Blank	0.000	04:16
Control Test	0.082	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BELL, C,

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BELL, C

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 00:17

Gender: Male

Observation Period Began: 00:45

Subject's Last Name: MARTINEZ

Date/Time of Test: 04/07/2007 01:12

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.083	01:06
Air Blank	0.000	01:07
Subject Sample #1	0.260	01:07
Breath Volume	1.453	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample #2	0.240	01:10
Breath Volume	1.679	01:10
Air Blank	0.000	01:11
Control Test	0.082	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: WINTERS, LANE, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: WINTERS, LANE

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:11

Gender: Male

Observation Period Began: 01:19

Subject's Last Name: FISCHER

Date/Time of Test: 04/12/2007 01:45

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Subject Sample #1	0.141	01:40
Breath Volume	2.363	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample #2	0.142	01:44
Breath Volume	2.324	01:44
Air Blank	0.000	01:44
Control Test	0.083	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: TOCZYLOWSKI, MICHAEL, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: TOCZYLOWSKI, MICHAEL

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:37

Gender: Male

Observation Period Began: 01:48

Subject's Last Name: PREVITI IV

Date/Time of Test: 04/13/2007 02:12

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.082	02:10	
Air Blank	0.000	02:10	
Subject Sample #1		02:11	Interferent Detect
Breath Volume	2.589	02:11	
Air Blank	0.000	02:11	
Control Test	0.083	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: TOCZYLOWSKI, MICHAEL, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: TOCZYLOWSKI, MICHAEL

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:37

Gender: Male

Observation Period Began: 02:12

Subject's Last Name: PREVITI IV

Date/Time of Test: 04/13/2007 02:35

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.082	02:33	
Air Blank	0.000	02:33	
Subject Sample #1		02:34	Interferent Detect
Breath Volume	1.863	02:34	
Air Blank	0.000	02:34	
Control Test	0.082	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: WINTERS, LANE, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: WINTERS, LANE

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: SYS CHECK

Arrest Time: 01:30

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: DOE

Date/Time of Test: 04/13/2007 02:43

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.082	02:38
Air Blank	0.000	02:39
Subject Sample #1	0.000	02:39
Breath Volume	2.355	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample #2	0.000	02:42
Breath Volume	2.320	02:42
Air Blank	0.000	02:42
Control Test	0.082	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: WINTERS, LANE, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: WINTERS, LANE

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 02:07

Gender: Female

Observation Period Began: 02:20

Subject's Last Name: BRODEUR

Date/Time of Test: 04/13/2007 02:53

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.082	02:46
Air Blank	0.000	02:47
Subject Sample #1	0.092	02:48
Breath Volume	1.527	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample #2	0.091	02:51
Breath Volume	1.519	02:51
Air Blank	0.000	02:52
Control Test	0.083	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: TOCZYLOWSKI, MICHAEL, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: TOCZYLOWSKI, MICHAEL

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:37

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: PREVITI IV

Date/Time of Test: 04/13/2007 03:08

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.082	03:02
Air Blank	0.000	03:02
Subject Sample #1	0.124	03:03
Breath Volume	1.246	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample #2	0.104	03:07
Breath Volume	1.820	03:07
Air Blank	0.000	03:07
Control Test	0.084	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 04/14/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BELL, C,

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BELL, C

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 00:02

Gender: Male

Observation Period Began: 00:15

Subject's Last Name: LOWERY

Date/Time of Test: 04/15/2007 00:44

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:37
Air Blank	0.000	00:38
Subject Sample #1	0.158	00:39
Breath Volume	2.144	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample #2	0.154	00:42
Breath Volume	1.972	00:42
Air Blank	0.000	00:43
Control Test	0.081	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:44

Data Download Date/Time: 05/16/2007 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BALDWIN, ROBERT, K

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BALDWIN, ROBERT

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 23:43

Gender: Male

Observation Period Began: 00:10

Subject's Last Name: HAIGLER

Date/Time of Test: 04/19/2007 00:38

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.082	00:33
Air Blank	0.000	00:33
Subject Sample #1	0.185	00:34
Breath Volume	2.792	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample #2	0.181	00:37
Breath Volume	2.859	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 05/16/2007 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: RIFFE, DAVID, J

Breath Test Operator Agency: Pasco County SO

Arresting Officer: STANJESKI, JOHN

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 03:39

Gender: Male

Observation Period Began: 04:03

Subject's Last Name: CORTES

Date/Time of Test: 04/19/2007 04:33

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.083	04:27
Air Blank	0.000	04:28
Subject Sample #1	0.141	04:28
Breath Volume	3.039	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:30
Subject Sample #2	0.146	04:31
Breath Volume	3.839	04:31
Air Blank	0.000	04:32
Control Test	0.083	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:33

Data Download Date/Time: 05/16/2007 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: TOCZYLOWSKI, MICHAEL, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: TOCZYLOWSKI, MICHAEL

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 22:20

Gender: Male

Observation Period Began: 23:40

Subject's Last Name: EVANS

Date/Time of Test: 04/21/2007 00:17

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.083	00:12
Air Blank	0.000	00:12
Subject Sample #1	0.000	00:13
Breath Volume	2.105	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample #2	0.000	00:16
Breath Volume	1.761	00:16
Air Blank	0.000	00:17
Control Test	0.083	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 05/16/2007 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BALDWIN, ROBERT, K

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BALDWIN, ROBERT

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:41

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: SCHROEDER

Date/Time of Test: 04/22/2007 03:00

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.082	02:54
Air Blank	0.000	02:54
Subject Sample #1	0.167	02:55
Breath Volume	1.652	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample #2	0.180	02:59
Breath Volume	1.753	02:59
Air Blank	0.000	02:59
Control Test	0.083	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 05/16/2007 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001049

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: ORNDORFF, WILLIAM, J

Breath Test Operator Agency: Pasco County SO

Arresting Officer: ORNDORFF, WILLIAM

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:24

Gender: Male

Observation Period Began: 01:24

Subject's Last Name: GOLDSMITH

Date/Time of Test: 04/27/2007 02:20

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.082	02:12
Air Blank	0.000	02:12
Subject Sample #1	0.170	02:14
Breath Volume	1.519	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample #2	0.155	02:19
Breath Volume	1.742	02:19
Air Blank	0.000	02:20
Control Test	0.083	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 05/16/2007 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001050

Instrument Registered To: CORAL SPRINGS PD

Software Version: 8100.27

Breath Test Operator: SAXON, LAWRENCE, D

Breath Test Operator Agency: Coral Springs PD

Arresting Officer: WILLIAMSON, CRISTINA

Arresting Officer Agency: Coral Springs PD

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 22:09

Gender: Male

Observation Period Began: 23:15

Subject's Last Name: FLINK

Date/Time of Test: 04/01/2007 23:53

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Subject Sample #1	0.000	23:48
Breath Volume	3.500	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample #2	0.000	23:51
Breath Volume	4.253	23:51
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 04/24/2007 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001050

Instrument Registered To: CORAL SPRINGS PD

Software Version: 8100.27

Breath Test Operator: SAXON, LAWRENCE, D

Breath Test Operator Agency: Coral Springs PD

Arresting Officer: BURTON, TIM

Arresting Officer Agency: Coral Springs PD

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 01:42

Gender: Male

Observation Period Began: 02:45

Subject's Last Name: MILLICAN

Date/Time of Test: 04/08/2007 03:35

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Subject Sample #1		03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:34	
Control Test	0.080	03:35	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 04/24/2007 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001050

Instrument Registered To: CORAL SPRINGS PD

Software Version: 8100.27

Breath Test Operator: MATONAK, GLENN, A

Breath Test Operator Agency: Coral Springs PD

Arresting Officer: WILLIAMSON, CHRISTINA

Arresting Officer Agency: Coral Springs PD

Last Agency Inspection: 03/30/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 02:23

Gender: Male

Observation Period Began: 02:37

Subject's Last Name: RILEY

Date/Time of Test: 04/14/2007 03:05

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.080	03:03	
Air Blank	0.000	03:03	
Subject Sample #1		03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.080	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	Fail	03:05	Temperature Regulation Test
Air Blank	0.000	03:05	

Data Download Date/Time: 04/24/2007 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001050

Instrument Registered To: CORAL SPRINGS PD

Software Version: 8100.27

Breath Test Operator: MATONAK, GLENN, A

Breath Test Operator Agency: Coral Springs PD

Arresting Officer: DEGERDON, STEPHEN

Arresting Officer Agency: Coral Springs PD

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 23:41

Gender: Male

Observation Period Began: 23:54

Subject's Last Name: MALDONADO

Date/Time of Test: 04/15/2007 00:28

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Subject Sample #1	0.198	00:23
Breath Volume	5.484	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample #2	0.193	00:26
Breath Volume	5.152	00:26
Air Blank	0.000	00:27
Control Test	0.077	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 04/24/2007 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:35

Gender: Female

Observation Period Began: 23:53

Subject's Last Name: WARD

Date/Time of Test: 04/01/2007 00:23

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample #1	0.175	00:17
Breath Volume	1.671	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample #2	0.159	00:21
Breath Volume	1.691	00:21
Air Blank	0.000	00:22
Control Test	0.081	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ALBERTSON, JASON

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:56

Gender: Male

Observation Period Began: 01:15

Subject's Last Name: FERRARO

Date/Time of Test: 04/01/2007 01:42

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample #1	0.082	01:38
Breath Volume	3.386	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample #2	0.076	01:41
Breath Volume	2.332	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: KING, SCOTT, E

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: CARNES, WILLIAM

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 16:03

Gender: Male

Observation Period Began: 16:25

Subject's Last Name: KOTLESZKA

Date/Time of Test: 04/01/2007 16:58

Diagnostics Check	OK	16:49
Air Blank	0.000	16:49
Control Test	0.082	16:50
Air Blank	0.000	16:50
Subject Sample #1	0.414	16:51
Breath Volume	2.769	16:51
Air Blank	0.000	16:52
Air Blank	0.000	16:53
Subject Sample #2	0.389	16:54
Breath Volume	3.480	16:54
Air Blank	0.000	16:55
Air Blank	0.000	16:56
Subject Sample #3	0.411	16:57
Breath Volume	3.300	16:57
Air Blank	0.000	16:58
Control Test	0.080	16:58
Air Blank	0.000	16:58
Diagnostics Check	OK	16:58

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MIKE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:16

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: NIX

Date/Time of Test: 04/05/2007 02:57

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.079	02:57	RFI Detect
Air Blank	0.000	02:57	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MIKE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:16

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: NIX

Date/Time of Test: 04/05/2007 03:03

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.082	03:01	
Air Blank	0.000	03:01	
Subject Sample #1		03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.082	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MIKE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 20:43

Gender: Male

Observation Period Began: 21:01

Subject's Last Name: LANE

Date/Time of Test: 04/05/2007 21:24

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:21	
Control Test	0.081	21:22	
Air Blank	0.000	21:22	
Subject Sample #1		21:22	Subject Test Refused
Breath Volume	0.000	21:22	
Air Blank	0.000	21:23	
Control Test	0.081	21:23	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 21:33

Gender: Female

Observation Period Began: 21:54

Subject's Last Name: SMITH

Date/Time of Test: 04/05/2007 22:18

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:16	
Control Test	0.081	22:16	
Air Blank	0.000	22:17	
Subject Sample #1		22:17	Sequence Aborted
Breath Volume	0.000	22:17	
Air Blank	0.000	22:18	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 21:33

Gender: Female

Observation Period Began: 22:18

Subject's Last Name: SMITH

Date/Time of Test: 04/05/2007 22:44

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Subject Sample #1	0.118	22:39
Breath Volume	1.617	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample #2	0.117	22:42
Breath Volume	1.496	22:42
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MIKE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:21

Gender: Male

Observation Period Began: 02:32

Subject's Last Name: CULLITON

Date/Time of Test: 04/06/2007 02:59

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample #1	0.138	02:54
Breath Volume	2.980	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.139	02:57
Breath Volume	3.628	02:57
Air Blank	0.000	02:58
Control Test	0.081	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:30

Gender: Male

Observation Period Began: 02:50

Subject's Last Name: LOVARIO

Date/Time of Test: 04/06/2007 03:24

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample #1	0.144	03:19
Breath Volume	1.945	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample #2	0.153	03:22
Breath Volume	3.183	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: CARNES, VINCENT

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 19:19

Gender: Male

Observation Period Began: 19:46

Subject's Last Name: DREW

Date/Time of Test: 04/06/2007 20:14

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.082	20:07
Air Blank	0.000	20:08
Subject Sample #1	0.047	20:09
Breath Volume	1.425	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:11
Subject Sample #2	0.047	20:13
Breath Volume	1.429	20:13
Air Blank	0.000	20:13
Control Test	0.082	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: NAGLE, TERRI

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 22:03

Gender: Male

Observation Period Began: 21:40

Subject's Last Name: DINNOCENZO

Date/Time of Test: 04/06/2007 22:19

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.080	22:11
Air Blank	0.000	22:11
Subject Sample #1	0.097	22:14
Breath Volume	1.574	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample #2	0.088	22:18
Breath Volume	1.449	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ALBERTSON, JASON

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:02

Gender: Male

Observation Period Began: 23:17

Subject's Last Name: BUTLER

Date/Time of Test: 04/06/2007 23:45

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:39
Subject Sample #1	0.229	23:40
Breath Volume	1.359	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample #2	0.240	23:43
Breath Volume	2.328	23:43
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: NAGLE, TERRI

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:20

Gender: Male

Observation Period Began: 02:32

Subject's Last Name: DIMITROV

Date/Time of Test: 04/07/2007 02:57

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.077	02:55	
Air Blank	0.000	02:56	
Subject Sample #1		02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.077	02:57	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: NAGLE, TERRI

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 18:38

Gender: Male

Observation Period Began: 18:26

Subject's Last Name: SMITH

Date/Time of Test: 04/07/2007 18:50

Diagnostics Check	OK	18:47	
Air Blank	0.000	18:47	
Control Test	0.080	18:47	
Air Blank	0.000	18:48	
Subject Sample #1		18:48	Subject Test Refused
Breath Volume	0.000	18:48	
Air Blank	0.000	18:49	
Control Test	0.081	18:49	
Air Blank	0.000	18:49	
Diagnostics Check	OK	18:50	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: NAGLE, TERRI

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:10

Gender: Female

Observation Period Began: 23:03

Subject's Last Name: PUZEL

Date/Time of Test: 04/07/2007 23:31

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.082	23:26
Air Blank	0.000	23:26
Subject Sample #1	0.000	23:27
Breath Volume	1.414	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample #2	0.000	23:30
Breath Volume	2.242	23:30
Air Blank	0.000	23:30
Control Test	0.082	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MIKE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:36

Gender: Male

Observation Period Began: 02:41

Subject's Last Name: SMITH JR

Date/Time of Test: 04/08/2007 03:08

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:03
Air Blank	0.000	03:03
Subject Sample #1	0.157	03:04
Breath Volume	3.351	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample #2	0.154	03:07
Breath Volume	2.515	03:07
Air Blank	0.000	03:07
Control Test	0.082	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: NAGLE, TERRI

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:58

Gender: Female

Observation Period Began: 03:22

Subject's Last Name: BROWN

Date/Time of Test: 04/08/2007 03:45

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.082	03:42	
Air Blank	0.000	03:43	
Subject Sample #1		03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 22:35

Gender: Female

Observation Period Began: 23:04

Subject's Last Name: SULLIVAN

Date/Time of Test: 04/11/2007 23:35

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample #1	0.117	23:28
Breath Volume	1.617	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample #2	0.117	23:33
Breath Volume	1.500	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: SESSIONS, CHRISTOPHER

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:45

Gender: Female

Observation Period Began: 00:10

Subject's Last Name: MCCOY

Date/Time of Test: 04/12/2007 00:40

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Subject Sample #1	0.196	00:35
Breath Volume	1.953	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample #2	0.194	00:38
Breath Volume	1.582	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: JOCKERS, MICHAEL, F

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: JOCKERS, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 13:38

Gender: Male

Observation Period Began: 14:03

Subject's Last Name: ROBINSON

Date/Time of Test: 04/12/2007 14:33

Diagnostics Check	OK	14:24	
Air Blank	0.000	14:24	
Control Test	0.079	14:24	
Air Blank	0.000	14:25	
Subject Sample #1	0.277	14:28	Volume Not Met
Breath Volume	0.421	14:28	
Air Blank	0.000	14:29	
Air Blank	0.000	14:31	
Subject Sample #2		14:31	Subject Test Refused
Breath Volume	0.667	14:31	
Air Blank	0.000	14:32	
Control Test	0.078	14:32	
Air Blank	0.000	14:33	
Diagnostics Check	OK	14:33	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: NAGLE, TERRI, E

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 22:20

Gender: Female

Observation Period Began: 22:40

Subject's Last Name: LESNIAK

Date/Time of Test: 04/12/2007 23:07

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:02
Subject Sample #1	0.000	23:03
Breath Volume	1.515	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample #2	0.000	23:06
Breath Volume	1.695	23:06
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ALBERTSON, JASON

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:51

Gender: Female

Observation Period Began: 00:36

Subject's Last Name: LUCAS

Date/Time of Test: 04/13/2007 01:08

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Subject Sample #1	0.128	01:04
Breath Volume	2.183	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample #2	0.129	01:07
Breath Volume	2.148	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: NAGLE, TERRI

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:00

Gender: Male

Observation Period Began: 22:51

Subject's Last Name: UHLMANN

Date/Time of Test: 04/13/2007 23:17

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.081	23:12
Air Blank	0.000	23:12
Subject Sample #1	0.276	23:13
Breath Volume	2.554	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample #2	0.283	23:16
Breath Volume	2.953	23:16
Air Blank	0.000	23:16
Control Test	0.081	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ALBERTSON, J

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:00

Gender: Female

Observation Period Began: 00:14

Subject's Last Name: BALLOTTA

Date/Time of Test: 04/14/2007 00:40

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample #1	0.312	00:36
Breath Volume	1.835	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample #2	0.306	00:39
Breath Volume	2.417	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: NAGLE, TERRI, E

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:06

Gender: Male

Observation Period Began: 02:24

Subject's Last Name: HICKSON

Date/Time of Test: 04/14/2007 02:47

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.082	02:44	
Air Blank	0.000	02:45	
Subject Sample #1		02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.081	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ALBERTSON, JASON

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:09

Gender: Female

Observation Period Began: 02:25

Subject's Last Name: FOX

Date/Time of Test: 04/15/2007 02:53

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:47
Subject Sample #1	0.153	02:48
Breath Volume	2.351	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample #2	0.156	02:51
Breath Volume	2.472	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: NAGLE, TERRI, E

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: NAGLE, TERRI

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:32

Gender: Male

Observation Period Began: 02:57

Subject's Last Name: MAUMONT

Date/Time of Test: 04/15/2007 03:24

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:18
Subject Sample #1	0.210	03:19
Breath Volume	2.171	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample #2	0.207	03:22
Breath Volume	2.664	03:22
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: NAGLE, TERRI, E

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ALBERTSON, JASON

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 04:41

Gender: Male

Observation Period Began: 04:27

Subject's Last Name: STYKA

Date/Time of Test: 04/15/2007 04:54

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:49
Subject Sample #1	0.214	04:49
Breath Volume	2.515	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample #2	0.208	04:52
Breath Volume	2.546	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: FORSHIER, NOLAN, J

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: DOUGLAS, JOHN

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 10:33

Gender: Male

Observation Period Began: 10:58

Subject's Last Name: FORSHIER

Date/Time of Test: 04/15/2007 11:19

Diagnostics Check	OK	11:19	
Air Blank		11:19	RFI Detect
Air Blank	0.000	11:19	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: FORSHIER, NOLAN, J

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: DOUGLAS, JOHN

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 10:33

Gender: Male

Observation Period Began: 10:58

Subject's Last Name: F626630890610

Date/Time of Test: 04/15/2007 11:24

Diagnostics Check	OK	11:23	
Air Blank	0.000	11:24	
Control Test	0.064	11:24	Control Outside Tolerance
Air Blank	0.000	11:24	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: FORSHIER, NOLAN, J

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: DOUGLAS, JOHN

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 10:33

Gender: Male

Observation Period Began: 10:58

Subject's Last Name: FORSHIER

Date/Time of Test: 04/15/2007 11:29

Diagnostics Check	OK	11:28	
Air Blank		11:29	Ambient Fail
Air Blank	0.000	11:29	

Data Download Date/Time: 04/18/2007 22:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 01:57

Gender: Male

Observation Period Began: 02:16

Subject's Last Name: KAELBER

Date/Time of Test: 04/19/2007 02:44

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:39
Subject Sample #1	0.000	02:40
Breath Volume	2.121	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample #2	0.000	02:43
Breath Volume	2.632	02:43
Air Blank	0.000	02:43
Control Test	0.082	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: NAGLE, TERRI

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 20:30

Gender: Female

Observation Period Began: 20:46

Subject's Last Name: BARRICK

Date/Time of Test: 04/19/2007 21:15

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:11	
Control Test	0.080	21:11	
Air Blank	0.000	21:12	
Subject Sample #1		21:13	Subject Test Refused
Breath Volume	0.000	21:13	
Air Blank	0.000	21:14	
Control Test	0.081	21:14	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:14

Gender: Male

Observation Period Began: 00:24

Subject's Last Name: FLORES-PACHECO

Date/Time of Test: 04/20/2007 00:53

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:47
Subject Sample #1	0.140	00:48
Breath Volume	3.296	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample #2	0.144	00:51
Breath Volume	2.964	00:51
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: NAGLE, TERRI

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:35

Gender: Male

Observation Period Began: 23:53

Subject's Last Name: MURILLO

Date/Time of Test: 04/21/2007 00:19

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:14
Subject Sample #1	0.212	00:15
Breath Volume	1.890	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample #2	0.228	00:18
Breath Volume	1.105	00:18
Air Blank	0.000	00:18
Control Test	0.081	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 01:19

Gender: Male

Observation Period Began: 01:32

Subject's Last Name: JONES JR

Date/Time of Test: 04/21/2007 01:54

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:53	
Control Test	0.081	01:53	
Air Blank		01:54	RFI Detect
Air Blank		01:54	RFI Detect

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 01:19

Gender: Male

Observation Period Began: 01:32

Subject's Last Name: JONES JR

Date/Time of Test: 04/21/2007 02:00

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.082	01:57	
Air Blank	0.000	01:58	
Subject Sample #1		01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.082	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ALBERTSON, JASON

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:02

Gender: Male

Observation Period Began: 02:18

Subject's Last Name: JANUSKEVICIUS

Date/Time of Test: 04/21/2007 02:49

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:39
Air Blank	0.000	02:39
Subject Sample #1	0.208	02:40
Breath Volume	1.960	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample #2	0.179	02:44
Breath Volume	1.218	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample #3	0.217	02:47
Breath Volume	2.046	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: KING, SCOTT, E

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: JACWIN, THOMAS

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 16:36

Gender: Male

Observation Period Began: 17:15

Subject's Last Name: FREE

Date/Time of Test: 04/21/2007 17:52

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.081	17:47
Air Blank	0.000	17:47
Subject Sample #1	0.283	17:48
Breath Volume	1.605	17:48
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample #2	0.291	17:51
Breath Volume	1.617	17:51
Air Blank	0.000	17:51
Control Test	0.082	17:52
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: GADDIS, SARAH, E

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: CARTER, MATTHEW

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 20:22

Gender: Male

Observation Period Began: 20:59

Subject's Last Name: KIRSTENSEN

Date/Time of Test: 04/22/2007 21:29

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.080	21:23
Air Blank	0.000	21:24
Subject Sample #1	0.220	21:25
Breath Volume	1.410	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample #2	0.219	21:27
Breath Volume	1.257	21:27
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: KING, SCOTT, E

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: COX, LENARD

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 23:33

Gender: Male

Observation Period Began: 00:05

Subject's Last Name: PAUL

Date/Time of Test: 04/24/2007 00:35

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Subject Sample #1	0.212	00:30
Breath Volume	1.214	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample #2	0.232	00:33
Breath Volume	1.562	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: JOCKERS, MICHAEL, F

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: JOCKERS, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 19:47

Gender: Male

Observation Period Began: 19:47

Subject's Last Name: CHENEY

Date/Time of Test: 04/24/2007 20:15

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.081	20:09
Air Blank	0.000	20:09
Subject Sample #1	0.084	20:10
Breath Volume	2.664	20:10
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample #2	0.084	20:13
Breath Volume	2.375	20:13
Air Blank	0.000	20:14
Control Test	0.082	20:14
Air Blank	0.000	20:14
Diagnostics Check	OK	20:15

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ARKOVICH, ROBBIE, S

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MICHAEL

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 19:03

Gender: Male

Observation Period Began: 19:21

Subject's Last Name: HILA

Date/Time of Test: 04/26/2007 19:53

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Subject Sample #1	0.354	19:48
Breath Volume	1.742	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample #2	0.347	19:52
Breath Volume	2.324	19:52
Air Blank	0.000	19:52
Control Test	0.081	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: NAGLE, TERRI

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 22:00

Gender: Male

Observation Period Began: 22:18

Subject's Last Name: HORNER

Date/Time of Test: 04/26/2007 22:41

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:39	
Control Test	0.081	22:39	
Air Blank	0.000	22:39	
Subject Sample #1		22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Control Test	0.080	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:22

Gender: Male

Observation Period Began: 00:32

Subject's Last Name: BRONKEMA

Date/Time of Test: 04/27/2007 00:59

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.081	00:53
Air Blank	0.000	00:54
Subject Sample #1	0.212	00:55
Breath Volume	2.968	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample #2	0.220	00:58
Breath Volume	2.320	00:58
Air Blank	0.000	00:58
Control Test	0.080	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 01:43

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: RUIZ

Date/Time of Test: 04/27/2007 02:27

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Subject Sample #1	0.194	02:22
Breath Volume	3.003	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample #2	0.198	02:25
Breath Volume	3.281	02:25
Air Blank	0.000	02:26
Control Test	0.082	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ALBERTSON, JASON

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 03:05

Gender: Male

Observation Period Began: 03:21

Subject's Last Name: PERSONS

Date/Time of Test: 04/27/2007 03:43

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.081	03:41	
Air Blank	0.000	03:42	
Subject Sample #1		03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: ALBERTSON, JASON, W

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: WEISKOPF, MIKE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 01:34

Gender: Male

Observation Period Began: 01:38

Subject's Last Name: LINER

Date/Time of Test: 04/28/2007 02:03

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.081	02:00	
Air Blank	0.000	02:01	
Subject Sample #1		02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: NAGLE, TERRI, E

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 02:07

Gender: Male

Observation Period Began: 02:24

Subject's Last Name: LOZADA

Date/Time of Test: 04/28/2007 02:50

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.081	02:46	
Air Blank	0.000	02:47	
Subject Sample #1		02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.082	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: WEISKOPF, MICHAEL, D

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ALBERTSON, JASON

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 00:39

Gender: Male

Observation Period Began: 00:57

Subject's Last Name: CROMARTIE

Date/Time of Test: 04/29/2007 01:26

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.082	01:18	
Air Blank	0.000	01:19	
Subject Sample #1	0.228	01:20	
Breath Volume	1.855	01:20	
Air Blank	0.000	01:20	
Air Blank	0.000	01:22	
Subject Sample #2		01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Control Test	0.082	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: NAGLE, TERRI, E

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: ARKOVICH, ROBBIE

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 01:38

Gender: Female

Observation Period Began: 01:47

Subject's Last Name: HOUSE

Date/Time of Test: 04/29/2007 02:21

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.082	02:15
Air Blank	0.000	02:16
Subject Sample #1	0.000	02:17
Breath Volume	2.039	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.000	02:19
Breath Volume	1.261	02:19
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001051

Instrument Registered To: ST PETERSBURG PD

Software Version: 8100.27

Breath Test Operator: SESSIONS, C,

Breath Test Operator Agency: St. Petersburg PD

Arresting Officer: SESSIONS, C

Arresting Officer Agency: St. Petersburg PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 635301G

Dry Gas Expiry Date: 12/21/2008

Violation Code: DUI

Arrest Time: 22:01

Gender: Male

Observation Period Began: 22:45

Subject's Last Name: LEMERY

Date/Time of Test: 04/30/2007 23:09

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.082	23:06	
Air Blank	0.000	23:06	
Subject Sample #1		23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:08	
Control Test	0.081	23:08	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 05/03/2007 20:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SNODGRASS, SHANNON, C

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CHRISTMAS, F

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:05

Gender: Male

Observation Period Began: 00:24

Subject's Last Name: BROOKS

Date/Time of Test: 04/01/2007 00:47

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:45	
Control Test	0.079	00:45	
Air Blank	0.000	00:46	
Subject Sample #1		00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Control Test	0.079	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SNODGRASS, SHANNON, C

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: WOOLERY, W

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:10

Gender: Male

Observation Period Began: 02:58

Subject's Last Name: THOMPSON

Date/Time of Test: 04/01/2007 03:33

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Subject Sample #1		03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.078	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SNODGRASS, SHANNON, C

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: VALERIO, E

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:35

Gender: Male

Observation Period Began: 03:35

Subject's Last Name: MCIVER

Date/Time of Test: 04/01/2007 04:20

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.078	04:14
Air Blank	0.000	04:15
Subject Sample #1	0.130	04:15
Breath Volume	2.031	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample #2	0.132	04:19
Breath Volume	2.902	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: PEREZ, S

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:59

Gender: Male

Observation Period Began: 00:30

Subject's Last Name: MIMS

Date/Time of Test: 04/02/2007 01:05

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.079	01:04	
Air Blank		01:04	RFI Detect
Air Blank		01:05	RFI Detect

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: PEREZ, S

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:59

Gender: Male

Observation Period Began: 00:30

Subject's Last Name: MIMS

Date/Time of Test: 04/02/2007 01:13

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.078	01:07
Air Blank	0.000	01:07
Subject Sample #1	0.236	01:08
Breath Volume	2.742	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample #2	0.245	01:11
Breath Volume	1.816	01:11
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:13

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: GUTHRIE, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:15

Gender: Female

Observation Period Began: 01:38

Subject's Last Name: CHANDLER

Date/Time of Test: 04/02/2007 02:03

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.079	02:00	
Air Blank	0.000	02:00	
Subject Sample #1		02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	
Control Test	0.079	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: VALRIO, E

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:20

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: JOHNSON

Date/Time of Test: 04/02/2007 03:03

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample #1	0.105	02:58
Breath Volume	1.953	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample #2	0.100	03:01
Breath Volume	1.402	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: JACKSON, R

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:45

Gender: Male

Observation Period Began: 04:30

Subject's Last Name: STANLEY

Date/Time of Test: 04/02/2007 04:58

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Subject Sample #1	0.162	04:51
Breath Volume	1.812	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample #2	0.171	04:56
Breath Volume	2.019	04:56
Air Blank	0.000	04:57
Control Test	0.082	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CULLEN, J

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:13

Gender: Male

Observation Period Began: 01:55

Subject's Last Name: SWEEZY

Date/Time of Test: 04/03/2007 02:19

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:16	
Control Test	0.079	02:17	
Air Blank	0.000	02:17	
Subject Sample #1		02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.079	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CHRISTMAS, F

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 03:18

Subject's Last Name: ALLEN

Date/Time of Test: 04/03/2007 03:45

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample #1	0.129	03:41
Breath Volume	1.664	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample #2	0.122	03:44
Breath Volume	1.222	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CULLEN, DOUG

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:15

Gender: Male

Observation Period Began: 23:58

Subject's Last Name: DENHAM

Date/Time of Test: 04/05/2007 00:26

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:21
Subject Sample #1	0.222	00:21
Breath Volume	2.046	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample #2	0.228	00:24
Breath Volume	1.265	00:24
Air Blank	0.000	00:25
Control Test	0.082	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:26

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CHRISTMAS, FOLIN

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:45

Gender: Male

Observation Period Began: 02:56

Subject's Last Name: MATTER

Date/Time of Test: 04/05/2007 03:19

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Subject Sample #1		03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.078	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PERRY, S, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: GUTHRIE, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:00

Gender: Male

Observation Period Began: 22:08

Subject's Last Name: SETZLER

Date/Time of Test: 04/06/2007 22:31

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:29	
Control Test	0.079	22:29	
Air Blank	0.000	22:30	
Subject Sample #1		22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Control Test	0.080	22:31	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: HARTING, MATT

Arresting Officer Agency: Jacksonville Beach

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 20:32

Gender: Male

Observation Period Began: 22:56

Subject's Last Name: ADAMS

Date/Time of Test: 04/06/2007 23:23

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Subject Sample #1	0.195	23:18
Breath Volume	4.503	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample #2	0.201	23:21
Breath Volume	4.285	23:21
Air Blank	0.000	23:22
Control Test	0.082	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CULLEN, DOUG

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:35

Gender: Male

Observation Period Began: 01:14

Subject's Last Name: ISBELL

Date/Time of Test: 04/07/2007 01:40

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample #1	0.197	01:36
Breath Volume	3.175	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample #2	0.192	01:39
Breath Volume	3.410	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CULLEN, JOHN

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:05

Gender: Male

Observation Period Began: 02:21

Subject's Last Name: SANCHEZ FLORES

Date/Time of Test: 04/07/2007 02:54

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample #1	0.222	02:44
Breath Volume	1.179	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample #2	0.189	02:48
Breath Volume	1.890	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample #3	0.216	02:53
Breath Volume	1.304	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: HOOVER, JOHN

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 03:23

Subject's Last Name: PARKE

Date/Time of Test: 04/07/2007 03:49

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample #1	0.133	03:45
Breath Volume	3.687	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample #2	0.137	03:48
Breath Volume	3.777	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: JAMISON, CHARLES

Arresting Officer Agency: Atlantic Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:32

Gender: Male

Observation Period Began: 04:04

Subject's Last Name: HARRIS

Date/Time of Test: 04/07/2007 04:31

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Subject Sample #1	0.161	04:26
Breath Volume	2.015	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample #2	0.159	04:29
Breath Volume	1.855	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:31

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: MCNAMEE, MORGAN

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:15

Gender: Male

Observation Period Began: 04:42

Subject's Last Name: JONES

Date/Time of Test: 04/07/2007 05:08

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.080	05:03
Air Blank	0.000	05:03
Subject Sample #1	0.132	05:04
Breath Volume	2.734	05:04
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample #2	0.135	05:07
Breath Volume	3.253	05:07
Air Blank	0.000	05:07
Control Test	0.081	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BAKER, DAVID, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: EDMONDS, D

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 17:44

Gender: Male

Observation Period Began: 18:55

Subject's Last Name: MELCHING

Date/Time of Test: 04/07/2007 19:22

Diagnostics Check	OK	19:15
Air Blank	0.000	19:15
Control Test	0.080	19:16
Air Blank	0.000	19:16
Subject Sample #1	0.259	19:16
Breath Volume	1.382	19:16
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample #2	0.262	19:20
Breath Volume	1.945	19:20
Air Blank	0.000	19:21
Control Test	0.083	19:21
Air Blank	0.000	19:21
Diagnostics Check	OK	19:22

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: BAKER, DAVID, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: MCNAMEE, M

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 18:40

Gender: Male

Observation Period Began: 20:20

Subject's Last Name: NAVA

Date/Time of Test: 04/07/2007 20:46

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:44	
Control Test	0.080	20:44	
Air Blank	0.000	20:45	
Subject Sample #1		20:45	Subject Test Refused
Breath Volume	0.000	20:45	
Air Blank	0.000	20:46	
Control Test	0.080	20:46	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: VALERIO, ERNESTO

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:55

Gender: Male

Observation Period Began: 23:34

Subject's Last Name: MURDZA

Date/Time of Test: 04/10/2007 00:01

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.079	23:56
Air Blank	0.000	23:56
Subject Sample #1	0.056	23:57
Breath Volume	3.460	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample #2	0.054	23:59
Breath Volume	3.390	23:59
Air Blank	0.000	00:00
Control Test	0.081	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BORGERT, MATT

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:25

Gender: Male

Observation Period Began: 01:34

Subject's Last Name: SCOTT

Date/Time of Test: 04/10/2007 02:01

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Subject Sample #1	0.100	01:56
Breath Volume	2.570	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample #2	0.099	01:59
Breath Volume	2.507	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BORGERT, M

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:25

Gender: Male

Observation Period Began: 23:38

Subject's Last Name: HAMMAN

Date/Time of Test: 04/11/2007 00:07

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.079	00:02	
Air Blank	0.000	00:02	
Subject Sample #1	0.127	00:03	
Breath Volume	4.503	00:03	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample #2	0.120	00:06	
Breath Volume	3.351	00:06	
Air Blank	0.000	00:06	
Control Test	0.078	00:07	
Air Blank		00:07	RFI Detect
Air Blank	0.000	00:07	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BORGERT, M

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:25

Gender: Male

Observation Period Began: 23:38

Subject's Last Name: HAMMAN

Date/Time of Test: 04/11/2007 00:16

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Subject Sample #1	0.118	00:11
Breath Volume	2.457	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample #2	0.116	00:14
Breath Volume	4.507	00:14
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: GRAHAM, T

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:55

Gender: Female

Observation Period Began: 01:00

Subject's Last Name: ALLEN

Date/Time of Test: 04/11/2007 01:28

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample #1	0.181	01:23
Breath Volume	2.000	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:25
Subject Sample #2	0.183	01:26
Breath Volume	1.246	01:26
Air Blank	0.000	01:27
Control Test	0.078	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: VALERIO, E

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:56

Gender: Male

Observation Period Began: 02:50

Subject's Last Name: CROSS

Date/Time of Test: 04/11/2007 03:17

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Subject Sample #1	0.145	03:12
Breath Volume	2.636	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.143	03:15
Breath Volume	1.304	03:15
Air Blank	0.000	03:16
Control Test	0.078	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: HOOVER, J

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:00

Gender: Female

Observation Period Began: 00:30

Subject's Last Name: BROOKS

Date/Time of Test: 04/12/2007 00:58

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.079	00:54	
Air Blank	0.000	00:55	
Subject Sample #1		00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.078	00:58	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BARRON, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:05

Gender: Female

Observation Period Began: 01:15

Subject's Last Name: PATE

Date/Time of Test: 04/12/2007 01:41

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.078	01:39	
Air Blank	0.000	01:40	
Subject Sample #1		01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.079	01:41	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BYRNE, JAMES

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:30

Gender: Female

Observation Period Began: 22:54

Subject's Last Name: FROST

Date/Time of Test: 04/12/2007 23:21

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.080	23:15
Air Blank	0.000	23:15
Subject Sample #1	0.106	23:16
Breath Volume	1.671	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:18
Subject Sample #2	0.102	23:19
Breath Volume	1.320	23:19
Air Blank	0.000	23:20
Control Test	0.081	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BORGERT, MATT

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:30

Gender: Male

Observation Period Began: 00:34

Subject's Last Name: WOLFE

Date/Time of Test: 04/13/2007 00:57

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.081	00:55	
Air Blank	0.000	00:56	
Subject Sample #1		00:56	Subject Test Refused
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Control Test	0.080	00:57	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: PEREZ, SERAFIN

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 02:34

Subject's Last Name: SCHUMACHER

Date/Time of Test: 04/13/2007 02:57

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.081	02:55	
Air Blank	0.000	02:55	
Subject Sample #1		02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.081	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: MOSCHBERGER, SHAWN

Arresting Officer Agency: Atlantic Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:06

Gender: Female

Observation Period Began: 04:47

Subject's Last Name: GARCIA

Date/Time of Test: 04/13/2007 05:10

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.080	05:08	
Air Blank	0.000	05:08	
Subject Sample #1		05:09	Subject Test Refused
Breath Volume	0.000	05:09	
Air Blank	0.000	05:09	
Control Test	0.080	05:09	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRICE, T, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BRYAN, D

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 05:02

Gender: Male

Observation Period Began: 06:40

Subject's Last Name: REYNOSO

Date/Time of Test: 04/13/2007 07:06

Diagnostics Check	OK	07:00
Air Blank	0.000	07:01
Control Test	0.080	07:01
Air Blank	0.000	07:01
Subject Sample #1	0.119	07:02
Breath Volume	3.210	07:02
Air Blank	0.000	07:03
Air Blank	0.000	07:05
Subject Sample #2	0.116	07:05
Breath Volume	3.113	07:05
Air Blank	0.000	07:06
Control Test	0.080	07:06
Air Blank	0.000	07:06
Diagnostics Check	OK	07:06

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: EDMONDS, DARRELL

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:27

Gender: Female

Observation Period Began: 03:02

Subject's Last Name: MARKS

Date/Time of Test: 04/15/2007 03:29

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:23
Subject Sample #1	0.165	03:24
Breath Volume	3.480	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample #2	0.159	03:27
Breath Volume	3.125	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: GRAHAM, TIM

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:00

Gender: Male

Observation Period Began: 04:06

Subject's Last Name: LOPEZ

Date/Time of Test: 04/15/2007 04:39

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample #1	0.196	04:35
Breath Volume	1.863	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample #2	0.201	04:38
Breath Volume	1.941	04:38
Air Blank	0.000	04:38
Control Test	0.079	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: DYAL, CARROLL

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 04:00

Gender: Male

Observation Period Began: 05:14

Subject's Last Name: KENNEDY

Date/Time of Test: 04/15/2007 05:52

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.078	05:47
Air Blank	0.000	05:47
Subject Sample #1	0.000	05:48
Breath Volume	2.808	05:48
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample #2	0.000	05:51
Breath Volume	3.035	05:51
Air Blank	0.000	05:51
Control Test	0.079	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRICE, T, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: ROLISON, M

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 10:20

Gender: Male

Observation Period Began: 11:17

Subject's Last Name: SHOOK

Date/Time of Test: 04/15/2007 11:44

Diagnostics Check	OK	11:37
Air Blank	0.000	11:38
Control Test	0.079	11:38
Air Blank	0.000	11:39
Subject Sample #1	0.229	11:39
Breath Volume	1.652	11:39
Air Blank	0.000	11:40
Air Blank	0.000	11:42
Subject Sample #2	0.224	11:42
Breath Volume	1.449	11:42
Air Blank	0.000	11:43
Control Test	0.078	11:43
Air Blank	0.000	11:44
Diagnostics Check	OK	11:44

Data Download Date/Time: 04/16/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: JACKSON, R

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:05

Gender: Male

Observation Period Began: 00:26

Subject's Last Name: SNELL

Date/Time of Test: 04/21/2007 00:49

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:47	
Subject Sample #1		00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.079	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BORGERT, M

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:45

Gender: Female

Observation Period Began: 01:50

Subject's Last Name: HOWELL

Date/Time of Test: 04/21/2007 02:16

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Subject Sample #1	0.155	02:12
Breath Volume	1.765	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample #2	0.148	02:15
Breath Volume	2.136	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BORGERT, M

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:30

Gender: Female

Observation Period Began: 02:30

Subject's Last Name: WILLIAMS

Date/Time of Test: 04/21/2007 02:53

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.079	02:51	
Air Blank	0.000	02:51	
Subject Sample #1		02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.079	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: VALERIO, E

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 03:10

Subject's Last Name: DURRANT

Date/Time of Test: 04/21/2007 03:38

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Subject Sample #1	0.137	03:33
Breath Volume	1.148	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.150	03:37
Breath Volume	1.722	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: WELLS, S

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:10

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: LOCKWOOD

Date/Time of Test: 04/21/2007 04:48

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:44	
Control Test	0.078	04:45	
Air Blank	0.000	04:45	
Subject Sample #1		04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:47	
Control Test	0.079	04:47	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CULLEN, D

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:45

Gender: Male

Observation Period Began: 00:34

Subject's Last Name: SUBECK

Date/Time of Test: 04/23/2007 01:01

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Subject Sample #1	0.217	00:57
Breath Volume	2.007	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample #2	0.214	01:00
Breath Volume	1.378	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CULLEN, D

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:05

Gender: Male

Observation Period Began: 00:55

Subject's Last Name: EDELEN

Date/Time of Test: 04/24/2007 01:22

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Subject Sample #1	0.103	01:18
Breath Volume	2.203	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample #2	0.099	01:21
Breath Volume	1.718	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRUETT, RICHARD, L

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BARRON, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 02:15

Subject's Last Name: BRIGHT

Date/Time of Test: 04/24/2007 02:41

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:36
Subject Sample #1	0.052	02:37
Breath Volume	3.375	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample #2	0.044	02:40
Breath Volume	1.570	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CHRISTMAS, FOLIN

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:05

Gender: Male

Observation Period Began: 00:19

Subject's Last Name: SUAREZ

Date/Time of Test: 04/25/2007 01:05

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.079	01:00
Air Blank	0.000	01:00
Subject Sample #1	0.183	01:01
Breath Volume	1.691	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample #2	0.195	01:04
Breath Volume	1.375	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BRIDGES, ERNEST

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:06

Gender: Male

Observation Period Began: 01:58

Subject's Last Name: MILLS

Date/Time of Test: 04/25/2007 02:25

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample #1	0.125	02:20
Breath Volume	1.628	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample #2	0.118	02:23
Breath Volume	1.222	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: HARTING, MATT

Arresting Officer Agency: Jacksonville Beach

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:43

Gender: Female

Observation Period Began: 06:18

Subject's Last Name: RYAN

Date/Time of Test: 04/25/2007 06:41

Diagnostics Check	OK	06:38	
Air Blank	0.000	06:39	
Control Test	0.080	06:39	
Air Blank	0.000	06:39	
Subject Sample #1		06:40	Subject Test Refused
Breath Volume	0.000	06:40	
Air Blank	0.000	06:40	
Control Test	0.078	06:40	
Air Blank	0.000	06:41	
Diagnostics Check	OK	06:41	

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BARTHOLOMEW, JASON

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:51

Gender: Male

Observation Period Began: 23:54

Subject's Last Name: SHARPE

Date/Time of Test: 04/26/2007 00:21

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:15
Subject Sample #1	0.254	00:16
Breath Volume	2.523	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:18
Subject Sample #2	0.243	00:19
Breath Volume	2.003	00:19
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: CHRISTMAS, FOLIN

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: CRAPPS

Date/Time of Test: 04/26/2007 03:33

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Subject Sample #1	0.091	03:28
Breath Volume	1.593	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample #2	0.084	03:31
Breath Volume	1.636	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRICE, T, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: HOOVER, J

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 20:50

Gender: Male

Observation Period Began: 20:10

Subject's Last Name: WIMMER

Date/Time of Test: 04/26/2007 22:35

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:29
Air Blank	0.000	22:29
Subject Sample #1	0.122	22:30
Breath Volume	1.726	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:32
Subject Sample #2	0.116	22:34
Breath Volume	2.621	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: GUTHRIE, KEITH

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:40

Gender: Male

Observation Period Began: 01:49

Subject's Last Name: BROOKS

Date/Time of Test: 04/27/2007 02:16

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Subject Sample #1	0.151	02:11
Breath Volume	1.718	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample #2	0.165	02:15
Breath Volume	1.722	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: PRESCOTT, NELSON

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:00

Gender: Male

Observation Period Began: 04:48

Subject's Last Name: JOHNSON

Date/Time of Test: 04/27/2007 05:14

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.077	05:09
Air Blank	0.000	05:09
Subject Sample #1	0.152	05:10
Breath Volume	2.035	05:10
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample #2	0.152	05:13
Breath Volume	2.210	05:13
Air Blank	0.000	05:13
Control Test	0.079	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRICE, T, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: MCMILLIAN, V

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 09:01

Gender: Male

Observation Period Began: 10:56

Subject's Last Name: JONES

Date/Time of Test: 04/27/2007 11:19

Diagnostics Check	OK	11:16	
Air Blank	0.000	11:17	
Control Test	0.079	11:17	
Air Blank	0.000	11:18	
Subject Sample #1		11:18	Subject Test Refused
Breath Volume	0.000	11:18	
Air Blank	0.000	11:18	
Control Test	0.078	11:19	
Air Blank	0.000	11:19	
Diagnostics Check	OK	11:19	

Data Download Date/Time: 04/27/2007 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BLACKSTONE, PHIL

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:00

Gender: Male

Observation Period Began: 22:58

Subject's Last Name: BAKER

Date/Time of Test: 04/27/2007 23:25

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample #1	0.215	23:20
Breath Volume	2.550	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:22
Subject Sample #2	0.209	23:23
Breath Volume	1.730	23:23
Air Blank	0.000	23:24
Control Test	0.078	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BARRON, KEITH

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:50

Gender: Male

Observation Period Began: 00:57

Subject's Last Name: CRAWFORD

Date/Time of Test: 04/28/2007 01:25

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.080	01:20
Air Blank	0.000	01:20
Subject Sample #1	0.077	01:20
Breath Volume	2.363	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample #2	0.073	01:23
Breath Volume	2.609	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: POTTER, CHRIS

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:55

Gender: Female

Observation Period Began: 02:55

Subject's Last Name: WELSH

Date/Time of Test: 04/28/2007 03:18

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.078	03:15	
Air Blank	0.000	03:16	
Subject Sample #1		03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: EISERT, WILLIAM

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:25

Gender: Male

Observation Period Began: 05:27

Subject's Last Name: MANZANO

Date/Time of Test: 04/28/2007 06:09

Diagnostics Check	OK	06:03
Air Blank	0.000	06:03
Control Test	0.079	06:03
Air Blank	0.000	06:04
Subject Sample #1	0.183	06:04
Breath Volume	2.929	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample #2	0.176	06:07
Breath Volume	2.824	06:07
Air Blank	0.000	06:08
Control Test	0.079	06:08
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BRIDGES, ERNEST

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:45

Gender: Male

Observation Period Began: 00:54

Subject's Last Name: GARCIA-BONILLA

Date/Time of Test: 04/29/2007 01:26

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample #1	0.136	01:22
Breath Volume	3.015	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample #2	0.130	01:25
Breath Volume	2.507	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: VALERIO, ERNESTO

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:50

Gender: Male

Observation Period Began: 04:02

Subject's Last Name: IZQUIERDO LEYVA

Date/Time of Test: 04/29/2007 04:29

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:23
Air Blank	0.000	04:23
Subject Sample #1	0.184	04:24
Breath Volume	3.625	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample #2	0.184	04:27
Breath Volume	3.050	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRESCOTT, EARL, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: FISSEL, MICHEAL

Arresting Officer Agency: Atlantic Beach PD

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 04:08

Gender: Male

Observation Period Began: 05:40

Subject's Last Name: STROMBERG

Date/Time of Test: 04/29/2007 06:07

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:02
Subject Sample #1	0.173	06:02
Breath Volume	2.476	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample #2	0.169	06:05
Breath Volume	3.070	06:05
Air Blank	0.000	06:06
Control Test	0.079	06:06
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRICE, T, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: GOLINO, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 08:00

Gender: Male

Observation Period Began: 09:17

Subject's Last Name: MCCARTY

Date/Time of Test: 04/29/2007 09:41

Diagnostics Check	OK	09:38	
Air Blank	0.000	09:39	
Control Test	0.079	09:39	
Air Blank	0.000	09:40	
Subject Sample #1		09:40	Subject Test Refused
Breath Volume	0.000	09:40	
Air Blank	0.000	09:40	
Control Test	0.079	09:41	
Air Blank	0.000	09:41	
Diagnostics Check	OK	09:41	

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: CLAYMAN, EDWARD, B

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BARRON, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:45

Gender: Male

Observation Period Began: 20:49

Subject's Last Name: HERNANDEZ

Date/Time of Test: 04/29/2007 21:17

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Subject Sample #1	0.185	21:12
Breath Volume	1.773	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample #2	0.191	21:15
Breath Volume	2.730	21:15
Air Blank	0.000	21:16
Control Test	0.078	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BYRNE, J

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:00

Gender: Male

Observation Period Began: 00:04

Subject's Last Name: WILLIAMS

Date/Time of Test: 04/30/2007 00:28

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.078	00:25	
Air Blank	0.000	00:26	
Subject Sample #1		00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:27	
Control Test	0.078	00:27	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: VALERIO, E

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:50

Gender: Male

Observation Period Began: 02:10

Subject's Last Name: HERRERA

Date/Time of Test: 04/30/2007 02:37

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Subject Sample #1	0.192	02:32
Breath Volume	3.125	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample #2	0.193	02:36
Breath Volume	3.035	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: GRAHAM, T

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:45

Gender: Male

Observation Period Began: 02:58

Subject's Last Name: CHORLTON

Date/Time of Test: 04/30/2007 03:21

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Subject Sample #1		03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.079	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, W

Breath Test Operator Agency: Atlantic Beach PD

Arresting Officer: FISSEL, M

Arresting Officer Agency: Atlantic Beach PD

Last Agency Inspection: 04/27/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:58

Gender: Male

Observation Period Began: 03:35

Subject's Last Name: JENKINS

Date/Time of Test: 04/30/2007 03:58

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:56	
Subject Sample #1		03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: SMITH, KEVIN, M

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: BARRON, K

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:40

Gender: Male

Observation Period Began: 04:09

Subject's Last Name: ALONSO

Date/Time of Test: 04/30/2007 04:36

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:30
Subject Sample #1	0.000	04:31
Breath Volume	2.613	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample #2	0.000	04:34
Breath Volume	2.242	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:36

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001052

Instrument Registered To: JACKSONVILLE SO

Software Version: 8100.27

Breath Test Operator: PRICE, T, A

Breath Test Operator Agency: Jacksonville SO

Arresting Officer: SHOUSE, J

Arresting Officer Agency: Jacksonville SO

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:50

Gender: Male

Observation Period Began: 21:17

Subject's Last Name: ANDREAU

Date/Time of Test: 04/30/2007 21:42

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:39	
Control Test	0.079	21:39	
Air Blank	0.000	21:40	
Subject Sample #1		21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:41	
Control Test	0.079	21:41	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 05/16/2007 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: CORBITT, DANIEL, S

Breath Test Operator Agency: Nassau County SO

Arresting Officer: CARTER, EREMY

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:20

Gender: Male

Observation Period Began: 22:30

Subject's Last Name: NAYLOR

Date/Time of Test: 04/01/2007 23:00

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.078	22:54
Air Blank	0.000	22:55
Subject Sample #1	0.191	22:55
Breath Volume	3.113	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample #2	0.188	22:58
Breath Volume	2.128	22:58
Air Blank	0.000	22:59
Control Test	0.076	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	23:00

Data Download Date/Time: 04/04/2007 16:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LEE, ROBERT, T

Breath Test Operator Agency: Nassau County SO

Arresting Officer: N/A, N/A

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: ADMIN

Arrest Time: 19:00

Gender: Male

Observation Period Began: 19:30

Subject's Last Name: PITCHIAYA

Date/Time of Test: 04/06/2007 19:59

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.078	19:51
Air Blank	0.000	19:52
Subject Sample #1	0.266	19:52
Breath Volume	1.386	19:52
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample #2	0.270	19:57
Breath Volume	1.867	19:57
Air Blank	0.000	19:58
Control Test	0.079	19:58
Air Blank	0.000	19:59
Diagnostics Check	OK	19:59

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LEE, ROBERT, T

Breath Test Operator Agency: Nassau County SO

Arresting Officer: WITKOWSKI, JARED

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:08

Gender: Male

Observation Period Began: 00:35

Subject's Last Name: KOSTIN

Date/Time of Test: 04/07/2007 01:07

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Subject Sample #1		01:07	Sequence Aborted
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LEE, ROBERT, T

Breath Test Operator Agency: Nassau County SO

Arresting Officer: PATTERSON, ROBIN

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:21

Gender: Male

Observation Period Began: 19:45

Subject's Last Name: BENNETT

Date/Time of Test: 04/07/2007 20:08

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.080	20:07	
Air Blank	0.000	20:07	
Subject Sample #1		20:07	Sequence Aborted
Breath Volume	0.000	20:07	
Air Blank	0.000	20:08	

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LEE', ROBERT, T

Breath Test Operator Agency: Nassau County SO

Arresting Officer: ENNIS, JAMES

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:39

Gender: Female

Observation Period Began: 21:55

Subject's Last Name: KUTI

Date/Time of Test: 04/07/2007 22:18

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:17	
Control Test	0.080	22:17	
Air Blank	0.000	22:17	
Subject Sample #1		22:18	Sequence Aborted
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LEE, ROBERT, T

Breath Test Operator Agency: Nassau County SO

Arresting Officer: JACKSON, S

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:00

Gender: Male

Observation Period Began: 22:22

Subject's Last Name: COLLINS

Date/Time of Test: 04/08/2007 22:53

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.078	22:47
Air Blank	0.000	22:48
Subject Sample #1	0.172	22:48
Breath Volume	3.304	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample #2	0.175	22:52
Breath Volume	3.566	22:52
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: CORBITT, DANIEL, S

Breath Test Operator Agency: FHP Troop G

Arresting Officer: HEALY, G

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:50

Gender: Male

Observation Period Began: 23:10

Subject's Last Name: HADDOCK

Date/Time of Test: 04/14/2007 23:43

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.078	23:42	
Air Blank	0.000	23:42	
Subject Sample #1		23:43	Sequence Aborted
Breath Volume	0.000	23:43	
Air Blank	0.000	23:43	

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: ORR, I, J

Breath Test Operator Agency: Nassau County SO

Arresting Officer: BALDWIN, R

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 10:51

Gender: Male

Observation Period Began: 11:04

Subject's Last Name: WALDRIP

Date/Time of Test: 04/22/2007 11:32

Diagnostics Check	OK	11:24
Air Blank	0.000	11:25
Control Test	0.078	11:25
Air Blank	0.000	11:25
Subject Sample #1	0.198	11:28
Breath Volume	1.750	11:28
Air Blank	0.000	11:28
Air Blank	0.000	11:30
Subject Sample #2	0.201	11:31
Breath Volume	2.433	11:31
Air Blank	0.000	11:31
Control Test	0.077	11:32
Air Blank	0.000	11:32
Diagnostics Check	OK	11:32

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LLANIER, LORIE, L

Breath Test Operator Agency: Nassau County SO

Arresting Officer: COLE, E

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:45

Gender: Male

Observation Period Began: 03:32

Subject's Last Name: TYSON

Date/Time of Test: 04/24/2007 04:03

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Subject Sample #1	0.223	03:58
Breath Volume	3.054	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample #2	0.227	04:01
Breath Volume	2.351	04:01
Air Blank	0.000	04:02
Control Test	0.077	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: CORBITT, DANIEL, S

Breath Test Operator Agency: Nassau County SO

Arresting Officer: COLES, ERNIE

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:24

Gender: Male

Observation Period Began: 19:55

Subject's Last Name: SMITH

Date/Time of Test: 04/24/2007 20:22

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.078	20:16
Air Blank	0.000	20:17
Subject Sample #1	0.000	20:17
Breath Volume	3.328	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample #2	0.000	20:20
Breath Volume	1.867	20:20
Air Blank	0.000	20:21
Control Test	0.077	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:22

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LANIER, LORIE, L

Breath Test Operator Agency: Nassau County SO

Arresting Officer: CLARK, E

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:55

Gender: Female

Observation Period Began: 02:15

Subject's Last Name: WYCKOFF

Date/Time of Test: 04/25/2007 02:47

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:39	
Control Test	0.078	02:39	
Air Blank	0.000	02:39	
Subject Sample #1	0.183	02:40	
Breath Volume	1.683	02:40	
Air Blank	0.000	02:41	
Air Blank	0.000	02:42	
Subject Sample #2	0.156	02:43	
Breath Volume	1.402	02:43	
Air Blank	0.000	02:44	
Air Blank	0.000	02:45	
Subject Sample #3		02:47	Sequence Aborted
Breath Volume	1.089	02:47	
Air Blank	0.000	02:47	

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LEE, ROBERT, T

Breath Test Operator Agency: Nassau County SO

Arresting Officer: PATTERSON, ROBIN

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:19

Gender: Male

Observation Period Began: 03:43

Subject's Last Name: LOUBE

Date/Time of Test: 04/26/2007 04:21

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Subject Sample #1	0.170	04:15
Breath Volume	1.980	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:17
Subject Sample #2	0.155	04:19
Breath Volume	1.785	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: CORBITT, DANIEL, S

Breath Test Operator Agency: Nassau County SO

Arresting Officer: FOSTER, GREG

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 17:38

Gender: Male

Observation Period Began: 18:28

Subject's Last Name: ANDERSON

Date/Time of Test: 04/28/2007 18:56

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.079	18:50
Air Blank	0.000	18:51
Subject Sample #1	0.180	18:51
Breath Volume	1.765	18:51
Air Blank	0.000	18:52
Air Blank	0.000	18:54
Subject Sample #2	0.193	18:55
Breath Volume	1.535	18:55
Air Blank	0.000	18:55
Control Test	0.078	18:56
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LLANIER, LORIE, L

Breath Test Operator Agency: Nassau County SO

Arresting Officer: WHITLEY, M

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 18:37

Gender: Female

Observation Period Began: 19:00

Subject's Last Name: VAZQUEZ

Date/Time of Test: 04/28/2007 19:31

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.079	19:21
Air Blank	0.000	19:22
Subject Sample #1	0.142	19:24
Breath Volume	1.324	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample #2	0.140	19:29
Breath Volume	1.261	19:29
Air Blank	0.000	19:30
Control Test	0.080	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LANIER, LORIE, L

Breath Test Operator Agency: Nassau County SO

Arresting Officer: TURNER, B

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 04/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:23

Gender: Male

Observation Period Began: 03:04

Subject's Last Name: HESTER

Date/Time of Test: 04/29/2007 03:28

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.078	03:26	
Air Blank	0.000	03:26	
Subject Sample #1		03:27	Sequence Aborted
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: CORBITT, DANIEL, S

Breath Test Operator Agency: Nassau County SO

Arresting Officer: CLARK, RUSS

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 04/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:15

Gender: Female

Observation Period Began: 03:57

Subject's Last Name: HARRIS

Date/Time of Test: 04/29/2007 04:29

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.078	04:24	
Air Blank	0.000	04:24	
Subject Sample #1	0.096	04:27	Volume Not Met
Breath Volume	0.402	04:27	
Air Blank		04:28	Purge Fail
Air Blank	0.000	04:29	

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001053

Instrument Registered To: NASSAU CO SO

Software Version: 8100.27

Breath Test Operator: LLANIER, LORIE, L

Breath Test Operator Agency: Nassau County SO

Arresting Officer: LANIER, LORIE

Arresting Officer Agency: Nassau County SO

Last Agency Inspection: 04/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: COURT ORD

Arrest Time: 21:20

Gender: Female

Observation Period Began: 21:20

Subject's Last Name: NICKLAS

Date/Time of Test: 04/29/2007 21:48

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.079	21:43
Air Blank	0.000	21:44
Subject Sample #1	0.000	21:44
Breath Volume	2.675	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample #2	0.000	21:47
Breath Volume	2.703	21:47
Air Blank	0.000	21:48
Control Test	0.078	21:48
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 05/02/2007 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD

Software Version: 8100.27

Breath Test Operator: SANCHEZ, MAIKER,

Breath Test Operator Agency: Davie PD

Arresting Officer: BART, JEFF

Arresting Officer Agency: Davie PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 06:00

Gender: Male

Observation Period Began: 06:10

Subject's Last Name: STAUBLE

Date/Time of Test: 04/01/2007 06:45

Diagnostics Check	OK	06:39
Air Blank	0.000	06:39
Control Test	0.077	06:40
Air Blank	0.000	06:40
Subject Sample #1	0.164	06:41
Breath Volume	1.410	06:41
Air Blank	0.000	06:42
Air Blank	0.000	06:43
Subject Sample #2	0.146	06:44
Breath Volume	1.496	06:44
Air Blank	0.000	06:44
Control Test	0.077	06:45
Air Blank	0.000	06:45
Diagnostics Check	OK	06:45

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD
Software Version: 8100.27
Breath Test Operator: CAIN, SHELIA, R
Breath Test Operator Agency: Davie PD
Arresting Officer: BART, S
Arresting Officer Agency: Davie PD
Last Agency Inspection: 03/31/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 630702I
Dry Gas Expiry Date: 11/06/2008
Violation Code: DUI
Arrest Time: 06:40
Gender: Male
Observation Period Began: 07:00

Subject's Last Name: HAX

Date/Time of Test: 04/02/2007 07:37

Diagnostics Check	OK	07:31
Air Blank	0.000	07:31
Control Test	0.077	07:32
Air Blank	0.000	07:32
Subject Sample #1	0.230	07:33
Breath Volume	3.171	07:33
Air Blank	0.000	07:34
Air Blank	0.000	07:35
Subject Sample #2	0.225	07:36
Breath Volume	2.640	07:36
Air Blank	0.000	07:36
Control Test	0.077	07:37
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD

Software Version: 8100.27

Breath Test Operator: SANCHEZ, MAIKER,

Breath Test Operator Agency: Davie PD

Arresting Officer: LARGHI, MARK

Arresting Officer Agency: Davie PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:05

Gender: Female

Observation Period Began: 02:20

Subject's Last Name: HUSHNER

Date/Time of Test: 04/04/2007 02:52

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.078	02:44
Air Blank	0.000	02:45
Subject Sample #1	0.240	02:47
Breath Volume	1.816	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample #2	0.237	02:50
Breath Volume	1.937	02:50
Air Blank	0.000	02:51
Control Test	0.075	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD

Software Version: 8100.27

Breath Test Operator: ARNDT, JEFFREY, R

Breath Test Operator Agency: Davie PD

Arresting Officer: LEONE, MARK

Arresting Officer Agency: Davie PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:00

Gender: Female

Observation Period Began: 22:20

Subject's Last Name: SINKU

Date/Time of Test: 04/04/2007 22:58

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.079	22:51
Air Blank	0.000	22:52
Subject Sample #1	0.017	22:53
Breath Volume	3.058	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample #2	0.014	22:56
Breath Volume	3.054	22:56
Air Blank	0.000	22:57
Control Test	0.077	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD

Software Version: 8100.27

Breath Test Operator: DONOHUE, TIMOTHY, S

Breath Test Operator Agency: Davie PD

Arresting Officer: DONOHUE, TIMOTHY

Arresting Officer Agency: Davie PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 20:54

Gender: Female

Observation Period Began: 21:20

Subject's Last Name: HIEB

Date/Time of Test: 04/12/2007 22:00

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.077	21:54
Air Blank	0.000	21:55
Subject Sample #1	0.000	21:55
Breath Volume	2.093	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample #2	0.000	21:59
Breath Volume	2.214	21:59
Air Blank	0.000	21:59
Control Test	0.077	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD

Software Version: 8100.27

Breath Test Operator: DONOHUE, TIMOTHY, S

Breath Test Operator Agency: Davie PD

Arresting Officer: BYRNE, BRIAN

Arresting Officer Agency: Davie PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:00

Gender: Male

Observation Period Began: 22:10

Subject's Last Name: WHITE

Date/Time of Test: 04/13/2007 22:56

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.078	22:50
Air Blank	0.000	22:51
Subject Sample #1	0.207	22:51
Breath Volume	2.847	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:53
Subject Sample #2	0.193	22:54
Breath Volume	3.132	22:54
Air Blank	0.000	22:55
Control Test	0.076	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:56

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD

Software Version: 8100.27

Breath Test Operator: ARNDT, JEFFREY, R

Breath Test Operator Agency: Davie PD

Arresting Officer: LARGHI, MARK

Arresting Officer Agency: Davie PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:30

Gender: Female

Observation Period Began: 01:40

Subject's Last Name: LAGUNA

Date/Time of Test: 04/19/2007 02:15

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.079	02:06	
Air Blank	0.000	02:06	
Subject Sample #1		02:09	No Sample Provided
Breath Volume	0.000	02:09	
Air Blank	0.000	02:10	
Air Blank	0.000	02:12	
Subject Sample #2		02:14	Sequence Aborted
Breath Volume	0.015	02:14	
Air Blank	0.000	02:15	

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD
Software Version: 8100.27
Breath Test Operator: PEREZ, DANIEL,
Breath Test Operator Agency: Davie PD
Arresting Officer: BERARDO, TONY
Arresting Officer Agency: Davie PD
Last Agency Inspection: 03/31/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 630702I
Dry Gas Expiry Date: 11/06/2008
Violation Code: DUI
Arrest Time: 01:00
Gender: Male
Observation Period Began: 01:20

Subject's Last Name: NOTTBOHM

Date/Time of Test: 04/21/2007 01:55

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.078	01:46
Air Blank	0.000	01:47
Subject Sample #1	0.188	01:48
Breath Volume	3.773	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample #2	0.167	01:51
Breath Volume	1.699	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample #3	0.174	01:54
Breath Volume	1.593	01:54
Air Blank	0.000	01:54
Control Test	0.076	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD

Software Version: 8100.27

Breath Test Operator: SANCHEZ, MAIKER,

Breath Test Operator Agency: Davie PD

Arresting Officer: BERARDO, TONY

Arresting Officer Agency: Davie PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 04:55

Gender: Male

Observation Period Began: 05:05

Subject's Last Name: ALVAREZ

Date/Time of Test: 04/21/2007 05:36

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.078	05:30
Air Blank	0.000	05:30
Subject Sample #1	0.184	05:31
Breath Volume	3.257	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:33
Subject Sample #2	0.187	05:34
Breath Volume	2.863	05:34
Air Blank	0.000	05:35
Control Test	0.077	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD

Software Version: 8100.27

Breath Test Operator: DONOHUE, TIMOTHY, S

Breath Test Operator Agency: Davie PD

Arresting Officer: ROSEN, CHAD

Arresting Officer Agency: Davie PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 17:40

Gender: Male

Observation Period Began: 18:10

Subject's Last Name: SCANLAN

Date/Time of Test: 04/25/2007 18:49

Diagnostics Check	OK	18:43
Air Blank	0.000	18:43
Control Test	0.079	18:43
Air Blank	0.000	18:44
Subject Sample #1	0.000	18:45
Breath Volume	1.937	18:45
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample #2	0.000	18:48
Breath Volume	2.304	18:48
Air Blank	0.000	18:48
Control Test	0.076	18:49
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001056

Instrument Registered To: DAVIE PD

Software Version: 8100.27

Breath Test Operator: PEREZ, DANIEL,

Breath Test Operator Agency: Davie PD

Arresting Officer: FRAMPTON, ANDREW

Arresting Officer Agency: Davie PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:05

Gender: Male

Observation Period Began: 00:20

Subject's Last Name: PALMER

Date/Time of Test: 04/30/2007 00:44

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
		00:44	Sequence Aborted
Air Blank	0.000	00:44	

Data Download Date/Time: 04/30/2007 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:58

Gender: Male

Observation Period Began: 02:45

Subject's Last Name: ZUNIGA-CANTARERO

Date/Time of Test: 04/01/2007 03:12

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.082	03:10	
Air Blank	0.000	03:10	
Subject Sample #1		03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.081	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: BAKER, DONALD, P

Breath Test Operator Agency: Lakeland PD

Arresting Officer: BAKER, DONALD

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 20:04

Gender: Male

Observation Period Began: 20:59

Subject's Last Name: SOMERS

Date/Time of Test: 04/03/2007 21:35

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.080	21:27
Air Blank	0.000	21:28
Subject Sample #1	0.195	21:30
Breath Volume	2.136	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:32
Subject Sample #2	0.176	21:33
Breath Volume	1.480	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FITZGERALD, F, F

Breath Test Operator Agency: Lakeland PD

Arresting Officer: FITZGERALD, F

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:12

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: GODINEZ

Date/Time of Test: 04/04/2007 03:08

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:03
Subject Sample #1	0.152	03:03
Breath Volume	1.250	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample #2	0.153	03:06
Breath Volume	1.183	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FINNEY, S, E

Breath Test Operator Agency: Lakeland PD

Arresting Officer: FINNEY, S

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 02:24

Subject's Last Name: TERRY

Date/Time of Test: 04/06/2007 02:52

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.082	02:45
Air Blank	0.000	02:46
Subject Sample #1	0.131	02:47
Breath Volume	1.718	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample #2	0.135	02:50
Breath Volume	1.386	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FINNEY, S, E

Breath Test Operator Agency: Lakeland PD

Arresting Officer: FINNEY, S

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 18:25

Gender: Male

Observation Period Began: 18:15

Subject's Last Name: BOYER

Date/Time of Test: 04/06/2007 18:41

Diagnostics Check	OK	18:38	
Air Blank	0.000	18:39	
Control Test	0.082	18:39	
Air Blank	0.000	18:40	
Subject Sample #1		18:41	Sequence Aborted
Breath Volume	0.000	18:41	
Air Blank	0.000	18:41	

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FINNEY, S, E

Breath Test Operator Agency: Lakeland PD

Arresting Officer: FINNEY, S

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 18:25

Gender: Male

Observation Period Began: 18:30

Subject's Last Name: BOYER

Date/Time of Test: 04/06/2007 18:57

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.081	18:51
Air Blank	0.000	18:51
Subject Sample #1	0.180	18:52
Breath Volume	1.578	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample #2	0.195	18:55
Breath Volume	1.574	18:55
Air Blank	0.000	18:56
Control Test	0.081	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 18:26

Gender: Female

Observation Period Began: 19:00

Subject's Last Name: WALTON

Date/Time of Test: 04/06/2007 19:26

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.082	19:21
Air Blank	0.000	19:22
Subject Sample #1	0.000	19:22
Breath Volume	2.605	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample #2	0.000	19:25
Breath Volume	2.609	19:25
Air Blank	0.000	19:25
Control Test	0.082	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FITZGERALD, F, F

Breath Test Operator Agency: Lakeland PD

Arresting Officer: FITZGERALD, F

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:24

Gender: Female

Observation Period Began: 00:15

Subject's Last Name: CURTIS

Date/Time of Test: 04/07/2007 01:43

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.082	01:38
Air Blank	0.000	01:39
Subject Sample #1	0.182	01:39
Breath Volume	1.125	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample #2	0.163	01:42
Breath Volume	1.332	01:42
Air Blank	0.000	01:43
Control Test	0.082	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FITZGERALD, F, F

Breath Test Operator Agency: Lakeland PD

Arresting Officer: FITZGERALD, F

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:21

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: CICARELLI

Date/Time of Test: 04/07/2007 03:06

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.082	03:01
Air Blank	0.000	03:01
Subject Sample #1	0.185	03:02
Breath Volume	4.753	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample #2	0.177	03:04
Breath Volume	4.414	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:33

Gender: Male

Observation Period Began: 03:14

Subject's Last Name: LARSON

Date/Time of Test: 04/07/2007 03:42

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.082	03:39	
Air Blank	0.000	03:40	
Subject Sample #1		03:40	Subject Test Refused
Breath Volume	0.000	03:40	
Air Blank	0.000	03:41	
Control Test	0.082	03:41	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: LLOYD, RAYMOND, W

Breath Test Operator Agency: Lakeland PD

Arresting Officer: LLOYD, RAYMOND

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:09

Gender: Female

Observation Period Began: 00:35

Subject's Last Name: ZIELINSKI

Date/Time of Test: 04/08/2007 01:05

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.082	00:56
Air Blank	0.000	00:57
Subject Sample #1	0.200	00:57
Breath Volume	1.992	00:57
Air Blank	0.000	00:58
Air Blank	0.000	00:59
Subject Sample #2	0.176	01:01
Breath Volume	2.488	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample #3	0.186	01:04
Breath Volume	1.855	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FITZGERALD, F, F

Breath Test Operator Agency: Lakeland PD

Arresting Officer: FITZGERALD, F

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:07

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: SCHORNAGEL

Date/Time of Test: 04/08/2007 03:29

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:23
Air Blank	0.000	03:23
Subject Sample #1	0.257	03:24
Breath Volume	1.757	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample #2	0.272	03:27
Breath Volume	1.144	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:29

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: BREWIN, THOMAS, H

Breath Test Operator Agency: Polk County SO

Arresting Officer: BREWIN, THOMAS

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:25

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: HAWKINS

Date/Time of Test: 04/08/2007 03:37

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.081	03:32
Air Blank	0.000	03:32
Subject Sample #1	0.000	03:33
Breath Volume	4.269	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.000	03:35
Breath Volume	3.269	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FITZGERALD, F, F

Breath Test Operator Agency: Lakeland PD

Arresting Officer: HENNIS, G

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:36

Gender: Male

Observation Period Began: 02:05

Subject's Last Name: HELMLY

Date/Time of Test: 04/09/2007 02:32

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.082	02:26	
Air Blank	0.000	02:27	
Subject Sample #1	0.141	02:27	
Breath Volume	2.039	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample #2	0.132	02:30	
Breath Volume	1.882	02:30	
Air Blank	0.000	02:31	
Control Test	0.086	02:31	Control Outside Tolerance
Air Blank	0.000	02:32	

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FINNEY, S, E

Breath Test Operator Agency: Lakeland PD

Arresting Officer: FINNEY, S

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:59

Gender: Male

Observation Period Began: 00:14

Subject's Last Name: SAMS

Date/Time of Test: 04/10/2007 00:49

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.083	00:43
Air Blank	0.000	00:44
Subject Sample #1	0.061	00:45
Breath Volume	1.812	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample #2	0.059	00:48
Breath Volume	1.496	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FINNEY, S, E

Breath Test Operator Agency: Lakeland PD

Arresting Officer: FINNEY, S

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:38

Gender: Male

Observation Period Began: 22:40

Subject's Last Name: MAUERMAN

Date/Time of Test: 04/10/2007 23:13

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.082	23:04
Air Blank	0.000	23:04
Subject Sample #1	0.156	23:05
Breath Volume	1.496	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample #2	0.190	23:08
Breath Volume	1.546	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:10
Subject Sample #3	0.176	23:11
Breath Volume	3.218	23:11
Air Blank	0.000	23:12
Control Test	0.078	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:13

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: LLOYD, RAYMOND, W

Breath Test Operator Agency: Lakeland PD

Arresting Officer: LLOYD, RAYMOND

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:09

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: MILLIGAN

Date/Time of Test: 04/12/2007 02:54

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Subject Sample #1		02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:55

Gender: Female

Observation Period Began: 04:15

Subject's Last Name: HOLLAND

Date/Time of Test: 04/14/2007 04:42

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:36
Subject Sample #1	0.174	04:37
Breath Volume	2.140	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample #2	0.180	04:40
Breath Volume	2.660	04:40
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:01

Gender: Male

Observation Period Began: 04:45

Subject's Last Name: MENTADO-ANGELES

Date/Time of Test: 04/14/2007 05:15

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.081	05:08
Air Blank	0.000	05:09
Subject Sample #1	0.118	05:10
Breath Volume	1.632	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample #2	0.103	05:14
Breath Volume	1.937	05:14
Air Blank	0.000	05:15
Control Test	0.081	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:49

Gender: Male

Observation Period Began: 05:45

Subject's Last Name: KEBBERLY

Date/Time of Test: 04/14/2007 06:09

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:05	
Control Test	0.081	06:06	
Air Blank	0.000	06:06	
Subject Sample #1		06:07	Subject Test Refused
Breath Volume	0.000	06:07	
Air Blank	0.000	06:08	
Control Test	0.081	06:08	
Air Blank	0.000	06:09	
Diagnostics Check	OK	06:09	

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:28

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: BERESFORD

Date/Time of Test: 04/15/2007 03:06

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample #1	0.174	03:02
Breath Volume	3.472	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample #2	0.177	03:05
Breath Volume	3.144	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:10

Gender: Male

Observation Period Began: 22:40

Subject's Last Name: WYATT

Date/Time of Test: 04/15/2007 23:07

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.083	23:05	
Air Blank	0.000	23:05	
Subject Sample #1		23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:06	
Control Test	0.082	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: LLOYD, RAYMOND, W

Breath Test Operator Agency: Lakeland PD

Arresting Officer: LLOYD, RAYMOND

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:40

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: LANHAM

Date/Time of Test: 04/17/2007 01:31

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.083	01:26
Air Blank	0.000	01:27
Subject Sample #1	0.126	01:27
Breath Volume	1.492	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample #2	0.123	01:30
Breath Volume	1.343	01:30
Air Blank	0.000	01:30
Control Test	0.082	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:25

Gender: Male

Observation Period Began: 20:30

Subject's Last Name: STEWART

Date/Time of Test: 04/20/2007 20:59

Diagnostics Check	OK	20:52
Air Blank	0.000	20:52
Control Test	0.081	20:53
Air Blank	0.000	20:53
Subject Sample #1	0.217	20:54
Breath Volume	1.714	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample #2	0.214	20:58
Breath Volume	2.210	20:58
Air Blank	0.000	20:58
Control Test	0.080	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: LLOYD, RAYMOND, W

Breath Test Operator Agency: Lakeland PD

Arresting Officer: LLOYD, RAYMOND

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:19

Gender: Male

Observation Period Began: 22:38

Subject's Last Name: MCDONALD

Date/Time of Test: 04/21/2007 23:06

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:00
Subject Sample #1	0.085	23:01
Breath Volume	2.074	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample #2	0.090	23:04
Breath Volume	2.628	23:04
Air Blank	0.000	23:05
Control Test	0.081	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: QUACKENBUSH, BRET, C

Breath Test Operator Agency: Lakeland PD

Arresting Officer: QUACKENBUSH, BRET

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:29

Gender: Male

Observation Period Began: 23:00

Subject's Last Name: REININGER

Date/Time of Test: 04/22/2007 23:28

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:25	
Control Test	0.081	23:25	
Air Blank	0.000	23:26	
Subject Sample #1		23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:27	
Control Test	0.081	23:27	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: QUACKENBUSH, BRET, C

Breath Test Operator Agency: Lakeland PD

Arresting Officer: WATERMAN, DAVID

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:40

Gender: Male

Observation Period Began: 22:59

Subject's Last Name: GONZALEZ

Date/Time of Test: 04/22/2007 23:47

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Subject Sample #1	0.207	23:43
Breath Volume	3.628	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample #2	0.213	23:46
Breath Volume	2.238	23:46
Air Blank	0.000	23:46
Control Test	0.081	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: LLOYD, RAYMOND, W

Breath Test Operator Agency: Lakeland PD

Arresting Officer: LLOYD, RAYMOND

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:31

Gender: Male

Observation Period Began: 23:50

Subject's Last Name: NEGRON

Date/Time of Test: 04/24/2007 00:16

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.082	00:10
Air Blank	0.000	00:11
Subject Sample #1	0.082	00:11
Breath Volume	3.246	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample #2	0.075	00:15
Breath Volume	2.617	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FINNEY, S, E

Breath Test Operator Agency: Lakeland PD

Arresting Officer: FINNEY, S

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:55

Gender: Male

Observation Period Began: 04:07

Subject's Last Name: STANTON

Date/Time of Test: 04/25/2007 04:40

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:34
Subject Sample #1	0.181	04:35
Breath Volume	1.992	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample #2	0.180	04:39
Breath Volume	1.476	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:10

Gender: Male

Observation Period Began: 21:10

Subject's Last Name: MACIAS

Date/Time of Test: 04/25/2007 21:40

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:32	
Control Test	0.081	21:32	
Air Blank	0.000	21:33	
Subject Sample #1	0.180	21:36	Volume Not Met
Breath Volume	0.136	21:36	
Air Blank	0.000	21:36	
Air Blank	0.000	21:38	
Subject Sample #2		21:39	Subject Test Refused
Breath Volume	0.085	21:39	
Air Blank	0.000	21:40	
Control Test	0.079	21:40	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:13

Gender: Male

Observation Period Began: 21:20

Subject's Last Name: FUTRELL

Date/Time of Test: 04/26/2007 21:47

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.081	21:41
Air Blank	0.000	21:42
Subject Sample #1	0.000	21:43
Breath Volume	1.804	21:43
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample #2	0.000	21:46
Breath Volume	1.566	21:46
Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:18

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: SCHAFF

Date/Time of Test: 04/27/2007 02:07

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.081	02:02
Air Blank	0.000	02:02
Subject Sample #1	0.241	02:03
Breath Volume	2.152	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample #2	0.248	02:06
Breath Volume	2.128	02:06
Air Blank	0.000	02:06
Control Test	0.081	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 04/27/2007 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 04:16

Gender: Male

Observation Period Began: 04:45

Subject's Last Name: VAN METER

Date/Time of Test: 04/28/2007 05:14

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.081	05:09
Air Blank	0.000	05:09
Subject Sample #1	0.204	05:10
Breath Volume	1.859	05:10
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample #2	0.192	05:13
Breath Volume	1.878	05:13
Air Blank	0.000	05:13
Control Test	0.081	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 05/18/2007 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: MOSSER, D, M

Breath Test Operator Agency: Lakeland PD

Arresting Officer: MOSSER, D

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:35

Gender: Male

Observation Period Began: 00:53

Subject's Last Name: BELLOTTO

Date/Time of Test: 04/29/2007 01:16

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.081	01:14	
Air Blank	0.000	01:15	
Subject Sample #1		01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:16	
Control Test	0.081	01:16	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 05/18/2007 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001057

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: JETT, EDWIN, A

Breath Test Operator Agency: Lakeland PD

Arresting Officer: KEAPOCK, KEVIN

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:39

Gender: Male

Observation Period Began: 00:45

Subject's Last Name: BOURQUIN

Date/Time of Test: 04/30/2007 01:16

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:10
Subject Sample #1	0.199	01:11
Breath Volume	1.308	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample #2	0.219	01:15
Breath Volume	1.863	01:15
Air Blank	0.000	01:16
Control Test	0.077	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 05/18/2007 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001059

Instrument Registered To: FERNANDINA BEACH PD

Software Version: 8100.27

Breath Test Operator: SANDERSON, RHONDA, S

Breath Test Operator Agency: Fernandina Beach PD

Arresting Officer: HEPLER, JASON

Arresting Officer Agency: Fernandina Beach PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 03:48

Gender: Male

Observation Period Began: 04:41

Subject's Last Name: HOLLOWAY

Date/Time of Test: 04/08/2007 05:17

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Subject Sample #1	0.000	05:12
Breath Volume	1.164	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample #2	0.000	05:16
Breath Volume	2.335	05:16
Air Blank	0.000	05:17
Control Test	0.080	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 04/26/2007 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001059

Instrument Registered To: FERNANDINA BEACH PD

Software Version: 8100.27

Breath Test Operator: MCCOY, SEAN, P

Breath Test Operator Agency: Fernandina Beach PD

Arresting Officer: WILLIAMS, CORY

Arresting Officer Agency: Fernandina Beach PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 14:19

Gender: Male

Observation Period Began: 14:50

Subject's Last Name: WILDER

Date/Time of Test: 04/11/2007 15:23

Diagnostics Check	OK	15:15
Air Blank	0.000	15:16
Control Test	0.079	15:16
Air Blank	0.000	15:16
Subject Sample #1	0.094	15:18
Breath Volume	1.808	15:18
Air Blank	0.000	15:19
Air Blank	0.000	15:20
Subject Sample #2	0.092	15:22
Breath Volume	2.027	15:22
Air Blank	0.000	15:22
Control Test	0.079	15:23
Air Blank	0.000	15:23
Diagnostics Check	OK	15:23

Data Download Date/Time: 04/26/2007 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001059

Instrument Registered To: FERNANDINA BEACH PD

Software Version: 8100.27

Breath Test Operator: MINAHAN, STEPHEN, J

Breath Test Operator Agency: Fernandina Beach PD

Arresting Officer: PAETSCH, JOSH

Arresting Officer Agency: Fernandina Beach PD

Last Agency Inspection: 03/31/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 22:58

Gender: Female

Observation Period Began: 23:10

Subject's Last Name: HOLBROOK

Date/Time of Test: 04/17/2007 23:38

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:35	
Control Test	0.079	23:35	
Air Blank	0.000	23:36	
Subject Sample #1		23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 04/26/2007 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001059

Instrument Registered To: FERNANDINA BEACH PD

Software Version: 8100.27

Breath Test Operator: MINAHAN, STEPHEN, J

Breath Test Operator Agency: Fernandina Beach PD

Arresting Officer: SHOEMAKER, RICHARD

Arresting Officer Agency: Fernandina Beach PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 22:10

Gender: Male

Observation Period Began: 22:25

Subject's Last Name: ROBLES

Date/Time of Test: 04/30/2007 22:51

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.079	22:46	
Air Blank	0.000	22:47	
Subject Sample #1	0.200	22:50	Volume Not Met
Breath Volume	0.699	22:50	
Air Blank		22:50	Purge Fail
Air Blank		22:51	Purge Fail

Data Download Date/Time: 06/04/2007 19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: DAVIS, JEREMY, L

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: DAVIS, JEREMY

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 01:56

Gender: Male

Observation Period Began: 02:08

Subject's Last Name: RAMKISHUN

Date/Time of Test: 04/07/2007 02:39

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Subject Sample #1	0.190	02:35
Breath Volume	1.218	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample #2	0.170	02:38
Breath Volume	1.589	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: SANTOS, WILTON, L

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: SANTOS, WILTON

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 02:12

Gender: Male

Observation Period Began: 02:20

Subject's Last Name: WOOD

Date/Time of Test: 04/07/2007 02:50

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:45	
Subject Sample #1	0.148	02:46	
Breath Volume	4.542	02:46	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample #2	0.146	02:49	
Breath Volume	4.140	02:49	
Air Blank	0.000	02:49	
Control Test	0.093	02:50	Control Outside Tolerance
Air Blank	0.000	02:50	

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: MEADOWS, MITCHELL, L

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: MERCER, JOSHUA

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 02:17

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: DOWNING

Date/Time of Test: 04/15/2007 03:04

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.081	03:02	
Air Blank	0.000	03:03	
Subject Sample #1		03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.069	03:04	Control Outside Tolerance
Air Blank	0.000	03:04	

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: ELLIOTT, JENNIFER, M

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: SANTOS, WILTON

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 00:26

Gender: Male

Observation Period Began: 00:51

Subject's Last Name: STOWELL

Date/Time of Test: 04/17/2007 01:22

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Subject Sample #1	0.181	01:17
Breath Volume	2.621	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample #2	0.171	01:20
Breath Volume	2.531	01:20
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: ELLIOTT, JENNIFER, M

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: SANTOS, WILTON

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 02:05

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: SMITH

Date/Time of Test: 04/18/2007 02:58

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:52	
Subject Sample #1	0.163	02:53	
Breath Volume	1.410	02:53	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample #2		02:57	Sequence Aborted
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: LOVELACE, JAMES, R

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: FLORES, RICARDO

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 21:35

Gender: Male

Observation Period Began: 21:50

Subject's Last Name: BENTON

Date/Time of Test: 04/18/2007 22:22

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:18	
Control Test	0.079	22:19	
Air Blank	0.000	22:19	
Subject Sample #1		22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.079	22:21	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: GASKIN, DANIEL, E

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: MERCER, JOSHUA

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 02:41

Gender: Male

Observation Period Began: 02:41

Subject's Last Name: BAXTER

Date/Time of Test: 04/19/2007 03:19

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Subject Sample #1		03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: CASTLE, DAVID, L

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: SANTOS, WILTON

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 01:59

Gender: Male

Observation Period Began: 02:20

Subject's Last Name: JOHNSON

Date/Time of Test: 04/21/2007 03:03

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:52	
Control Test	0.078	02:52	
Air Blank	0.000	02:52	
Subject Sample #1		02:55	No Sample Provided
Breath Volume	0.000	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample #2	0.122	02:58	
Breath Volume	4.714	02:58	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample #3	0.116	03:02	
Breath Volume	2.996	03:02	
Air Blank	0.000	03:02	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: SANTOS, WILTON, L

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: DAVIS, JEREMY

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 02:29

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: ADERHOLT

Date/Time of Test: 04/21/2007 03:37

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Subject Sample #1	0.237	03:32
Breath Volume	1.191	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.239	03:36
Breath Volume	2.542	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: DAVIS, JEREMY, L

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: SANTOS, WILTON

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 02:04

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: FELTON

Date/Time of Test: 04/22/2007 03:25

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Subject Sample #1		03:24	Sequence Aborted
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: DAVIS, JEREMY, L

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: SANTOS, WILTON

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 02:23

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: FELTON

Date/Time of Test: 04/22/2007 03:30

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank	0.000	03:29	
Subject Sample #1		03:29	Sequence Aborted
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	

Data Download Date/Time: 04/26/2007 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD
Software Version: 8100.27
Breath Test Operator: MEADOWS, MITCHELL, L
Breath Test Operator Agency: Winter Haven PD
Arresting Officer: MERCER, JOSHUA
Arresting Officer Agency: Winter Haven PD
Last Agency Inspection: 04/26/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 6261011
Dry Gas Expiry Date: 09/20/2008
Violation Code: DUI
Arrest Time: 03:19
Gender: Male
Observation Period Began: 03:25

Subject's Last Name: OWENS

Date/Time of Test: 04/28/2007 03:58

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:56	
Air Blank	0.000	03:56	
Subject Sample #1		03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:57	
Control Test	0.079	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 05/28/2007 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD
Software Version: 8100.27
Breath Test Operator: MEADOWS, MITCHELL, L
Breath Test Operator Agency: Winter Haven PD
Arresting Officer: MERCER, JOSHUA
Arresting Officer Agency: Winter Haven PD
Last Agency Inspection: 04/26/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 6261011
Dry Gas Expiry Date: 09/20/2008
Violation Code: DUI
Arrest Time: 02:29
Gender: Male
Observation Period Began: 02:50

Subject's Last Name: LYNCH

Date/Time of Test: 04/29/2007 03:19

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample #1	0.134	03:15
Breath Volume	4.144	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample #2	0.130	03:18
Breath Volume	3.640	03:18
Air Blank	0.000	03:18
Control Test	0.077	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 05/28/2007 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001060

Instrument Registered To: WINTER HAVEN PD

Software Version: 8100.27

Breath Test Operator: BONSALL, ROBERT, E

Breath Test Operator Agency: Winter Haven PD

Arresting Officer: MERCER, JOSH

Arresting Officer Agency: Winter Haven PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6261011

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 05:24

Gender: Male

Observation Period Began: 05:25

Subject's Last Name: GLENDENNING

Date/Time of Test: 04/29/2007 06:13

Diagnostics Check	OK	06:10	
Air Blank	0.000	06:10	
Control Test	0.079	06:10	
Air Blank	0.000	06:11	
Subject Sample #1		06:11	Subject Test Refused
Breath Volume	0.000	06:11	
Air Blank	0.000	06:12	
Control Test	0.079	06:12	
Air Blank	0.000	06:13	
Diagnostics Check	OK	06:13	

Data Download Date/Time: 05/28/2007 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001061

Instrument Registered To: GULFPORT PD

Software Version: 8100.27

Breath Test Operator: VANDENBERG, MICHAEL, J

Breath Test Operator Agency: Gulfport PD

Arresting Officer: VANDENBERG, MICHAEL

Arresting Officer Agency: Gulfport PD

Last Agency Inspection: 03/30/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 22:40

Gender: Male

Observation Period Began: 22:55

Subject's Last Name: SIMMONDS

Date/Time of Test: 04/07/2007 23:29

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Subject Sample #1	0.150	23:25
Breath Volume	3.058	23:25
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample #2	0.145	23:28
Breath Volume	2.628	23:28
Air Blank	0.000	23:28
Control Test	0.079	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 04/24/2007 00:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: BOCK, PHILIP, L

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: BOCK, PHILIP

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 03/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:54

Gender: Female

Observation Period Began: 02:00

Subject's Last Name: WARMACK

Date/Time of Test: 04/01/2007 03:07

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Subject Sample #1	0.177	03:01
Breath Volume	1.500	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample #2	0.176	03:06
Breath Volume	1.496	03:06
Air Blank	0.000	03:06
Control Test	0.078	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 04/06/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: BOCK, PHILIP, L

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: HILLAND, KENNETH

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 03/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:07

Gender: Female

Observation Period Began: 19:08

Subject's Last Name: BRANCHE

Date/Time of Test: 04/01/2007 19:43

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:38
Subject Sample #1	0.169	19:39
Breath Volume	2.375	19:39
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample #2	0.161	19:42
Breath Volume	1.757	19:42
Air Blank	0.000	19:42
Control Test	0.080	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 04/06/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: BOCK, PHILIP, L

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: LIGHTFIELD, ADAM

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 03/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:05

Gender: Male

Observation Period Began: 00:35

Subject's Last Name: WILLIAMSON

Date/Time of Test: 04/06/2007 01:05

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.079	01:01	
Air Blank	0.000	01:02	
Subject Sample #1		01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:04	
Control Test	0.080	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 04/06/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: DIAZ, HECTOR,

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: HILLAND, KENNY

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 04/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 13:50

Gender: Male

Observation Period Began: 13:55

Subject's Last Name: TONETTI

Date/Time of Test: 04/06/2007 14:35

Diagnostics Check	OK	14:27	
Air Blank	0.000	14:27	
Control Test	0.079	14:28	
Air Blank	0.000	14:28	
Subject Sample #1		14:31	No Sample Provided
Breath Volume	0.000	14:31	
Air Blank	0.000	14:32	
Air Blank	0.000	14:34	
Subject Sample #2		14:35	Sequence Aborted
Breath Volume	0.000	14:35	
Air Blank	0.000	14:35	

Data Download Date/Time: 05/06/2007 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: DIAZ, HECTOR,

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: HILLAND, KENNETH

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 04/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 13:50

Gender: Male

Observation Period Began: 13:55

Subject's Last Name: TONETTI

Date/Time of Test: 04/06/2007 14:52

Diagnostics Check	OK	14:51	
Air Blank		14:51	Sequence Aborted
Air Blank	0.000	14:52	

Data Download Date/Time: 05/06/2007 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: DIAZ, HECTOR,

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: HILLAND, KENNETH

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 04/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 13:50

Gender: Male

Observation Period Began: 14:40

Subject's Last Name: TONETTI

Date/Time of Test: 04/06/2007 15:04

Diagnostics Check	OK	15:01	
Air Blank	0.000	15:02	
Control Test	0.079	15:02	
Air Blank	0.000	15:03	
Subject Sample #1		15:03	Subject Test Refused
Breath Volume	0.000	15:03	
Air Blank	0.000	15:03	
Control Test	0.080	15:04	
Air Blank	0.000	15:04	
Diagnostics Check	OK	15:04	

Data Download Date/Time: 05/06/2007 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: CHIAPUTTI, DAREN, A

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: LIGHTFIELD, ADAM

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 04/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:04

Gender: Male

Observation Period Began: 01:37

Subject's Last Name: JOHNSON

Date/Time of Test: 04/13/2007 02:04

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample #1	0.000	02:00
Breath Volume	3.007	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample #2	0.000	02:03
Breath Volume	2.863	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 05/06/2007 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: BOCK, PHILIP, L

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: LIGHTFIELD, ADAM

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 04/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:16

Gender: Female

Observation Period Began: 02:23

Subject's Last Name: MALONE

Date/Time of Test: 04/21/2007 02:53

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample #1		02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 05/06/2007 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: CHIAPUTTI, DAREN, A

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: LIGHTFIELD, ADAM

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 04/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:36

Gender: Male

Observation Period Began: 00:49

Subject's Last Name: DECKER

Date/Time of Test: 04/26/2007 01:21

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Subject Sample #1	0.161	01:16
Breath Volume	1.968	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample #2	0.161	01:19
Breath Volume	2.121	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 05/06/2007 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: CHIAPUTTI, DAREN, A

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: LIGHTFIELD, ADAM

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 04/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:25

Gender: Female

Observation Period Began: 01:37

Subject's Last Name: BARTOLI-MUSSIO

Date/Time of Test: 04/26/2007 02:17

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.079	02:06	
Air Blank	0.000	02:07	
Subject Sample #1	0.133	02:10	Volume Not Met
Breath Volume	0.062	02:10	
Air Blank	0.000	02:10	
Air Blank	0.000	02:12	
Subject Sample #2	0.000	02:15	Volume Not Met
Breath Volume	0.019	02:15	
Air Blank	0.000	02:16	
Control Test	0.078	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 05/06/2007 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: CHIAPUTTI, DAREN, A

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: LIGHTFIELD, ADAM

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 04/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:25

Gender: Female

Observation Period Began: 02:55

Subject's Last Name: BARTOLI-MUSSIO

Date/Time of Test: 04/26/2007 03:29

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:17	
Subject Sample #1	0.166	03:20	Volume Not Met
Breath Volume	0.421	03:20	
Air Blank	0.000	03:21	
Air Blank	0.000	03:23	
Subject Sample #2	0.170	03:24	
Breath Volume	1.339	03:24	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample #3	0.160	03:27	
Breath Volume	1.195	03:27	
Air Blank	0.000	03:28	
Control Test	0.077	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 05/06/2007 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: BOCK, PHILIP, L

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: CHIAPUTTI, D.

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 04/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:05

Gender: Male

Observation Period Began: 21:12

Subject's Last Name: KAISNER

Date/Time of Test: 04/27/2007 21:54

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.079	21:49
Air Blank	0.000	21:49
Subject Sample #1	0.103	21:50
Breath Volume	2.621	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample #2	0.104	21:53
Breath Volume	2.835	21:53
Air Blank	0.000	21:53
Control Test	0.081	21:54
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 05/06/2007 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001062

Instrument Registered To: TREASURE ISLAND PD

Software Version: 8100.27

Breath Test Operator: BOCK, PHILIP, L

Breath Test Operator Agency: Treasure Island PD

Arresting Officer: SAVARESE, D.

Arresting Officer Agency: Treasure Island PD

Last Agency Inspection: 04/06/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:01

Gender: Male

Observation Period Began: 03:13

Subject's Last Name: LEVERING

Date/Time of Test: 04/29/2007 03:58

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:50	
Subject Sample #1		03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample #2		03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 05/06/2007 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001063

Instrument Registered To: HOLLYWOOD PD

Software Version: 8100.27

Breath Test Operator: RILLO, R,

Breath Test Operator Agency: Hollywood PD

Arresting Officer: RILLO, R

Arresting Officer Agency: Hollywood PD

Last Agency Inspection: 03/24/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 20:35

Gender: Male

Observation Period Began: 21:05

Subject's Last Name: VOGEL

Date/Time of Test: 04/02/2007 21:34

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:30
Subject Sample #1	0.138	21:30
Breath Volume	2.851	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample #2	0.137	21:33
Breath Volume	1.554	21:33
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 04/27/2007 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001063

Instrument Registered To: HOLLYWOOD PD

Software Version: 8100.27

Breath Test Operator: RILLO, R,

Breath Test Operator Agency: Hollywood PD

Arresting Officer: RILLO, R

Arresting Officer Agency: Hollywood PD

Last Agency Inspection: 03/24/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 19:35

Gender: Male

Observation Period Began: 19:55

Subject's Last Name: CRAIG

Date/Time of Test: 04/10/2007 20:19

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:18	
Control Test	0.078	20:18	
Air Blank	0.000	20:19	
Subject Sample #1		20:19	Sequence Aborted
Breath Volume	0.000	20:19	
Air Blank	0.000	20:19	

Data Download Date/Time: 04/27/2007 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001063

Instrument Registered To: HOLLYWOOD PD

Software Version: 8100.27

Breath Test Operator: BURGESS, LAWRENCE, B

Breath Test Operator Agency: Hollywood PD

Arresting Officer: BURGESS, LAWRENCE

Arresting Officer Agency: Hollywood PD

Last Agency Inspection: 03/24/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 03:35

Gender: Male

Observation Period Began: 04:30

Subject's Last Name: MACKLIFF

Date/Time of Test: 04/11/2007 05:05

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.078	04:58
Air Blank	0.000	04:58
Subject Sample #1	0.255	05:00
Breath Volume	1.429	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample #2	0.252	05:04
Breath Volume	2.437	05:04
Air Blank	0.000	05:04
Control Test	0.078	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 04/27/2007 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001063

Instrument Registered To: HOLLYWOOD PD

Software Version: 8100.27

Breath Test Operator: BURGESS, LAWRENCE, B

Breath Test Operator Agency: Hollywood PD

Arresting Officer: BURGESS, LAWRENCE

Arresting Officer Agency: Hollywood PD

Last Agency Inspection: 03/24/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 01:22

Gender: Male

Observation Period Began: 02:05

Subject's Last Name: PEREZ

Date/Time of Test: 04/12/2007 02:39

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.078	02:34
Air Blank	0.000	02:34
Subject Sample #1	0.093	02:35
Breath Volume	1.441	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample #2	0.084	02:38
Breath Volume	1.644	02:38
Air Blank	0.000	02:38
Control Test	0.078	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 04/27/2007 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001063

Instrument Registered To: HOLLYWOOD PD

Software Version: 8100.27

Breath Test Operator: RILLO, R,

Breath Test Operator Agency: Hollywood PD

Arresting Officer: RILLO, R

Arresting Officer Agency: Hollywood PD

Last Agency Inspection: 03/24/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 00:13

Gender: Female

Observation Period Began: 00:30

Subject's Last Name: MINERLEY-WIESENTHAL

Date/Time of Test: 04/16/2007 01:31

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.078	01:19	
Air Blank	0.000	01:19	
Subject Sample #1	0.133	01:22	Volume Not Met
Breath Volume	0.261	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample #2	0.132	01:26	
Breath Volume	1.554	01:26	
Air Blank	0.000	01:26	
Air Blank	0.000	01:28	
Subject Sample #3	0.133	01:29	
Breath Volume	1.843	01:29	
Air Blank	0.000	01:30	
Control Test	0.078	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 04/27/2007 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001063

Instrument Registered To: HOLLYWOOD PD

Software Version: 8100.27

Breath Test Operator: BURGESS, LARRY, B

Breath Test Operator Agency: Hollywood PD

Arresting Officer: PEREZ, OSVALDO

Arresting Officer Agency: Hollywood PD

Last Agency Inspection: 03/24/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 03:05

Gender: Male

Observation Period Began: 03:20

Subject's Last Name: THOMPKINS

Date/Time of Test: 04/19/2007 03:52

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.077	03:46
Air Blank	0.000	03:47
Subject Sample #1	0.209	03:47
Breath Volume	2.410	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample #2	0.213	03:50
Breath Volume	2.566	03:50
Air Blank	0.000	03:51
Control Test	0.077	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 04/27/2007 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: JAY, ERIK, J

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: JAY, ERIK

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:23

Gender: Male

Observation Period Began: 22:30

Subject's Last Name: MILLER

Date/Time of Test: 04/05/2007 22:57

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:51
Subject Sample #1	0.177	22:52
Breath Volume	1.566	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample #2	0.182	22:56
Breath Volume	2.292	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 04/19/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: KIRKPATRICK, J,

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: KIRKPATRICK, J

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:48

Gender: Male

Observation Period Began: 01:58

Subject's Last Name: GILBERT

Date/Time of Test: 04/06/2007 02:25

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample #1	0.168	02:20
Breath Volume	2.000	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample #2	0.180	02:23
Breath Volume	1.730	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:25

Data Download Date/Time: 04/19/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: KIRKPATRICK, J,

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: KIRKPATRICK, J

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 20:30

Gender: Female

Observation Period Began: 20:40

Subject's Last Name: ADAMS

Date/Time of Test: 04/06/2007 21:08

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.080	21:02
Air Blank	0.000	21:02
Subject Sample #1	0.206	21:03
Breath Volume	3.300	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample #2	0.198	21:06
Breath Volume	2.617	21:06
Air Blank	0.000	21:07
Control Test	0.079	21:07
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 04/19/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: KIRKPATRICK, J,

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: KIRKPATRICK, J

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:21

Gender: Male

Observation Period Began: 22:35

Subject's Last Name: WEATHERS

Date/Time of Test: 04/07/2007 23:02

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:57
Subject Sample #1	0.099	22:58
Breath Volume	2.007	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample #2	0.095	23:01
Breath Volume	1.433	23:01
Air Blank	0.000	23:01
Control Test	0.081	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 04/19/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: WOODWARD, PATRICK, E

Breath Test Operator Agency: Port Richey PD

Arresting Officer: WOODWARD, PATRICK

Arresting Officer Agency: Port Richey PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:18

Gender: Female

Observation Period Began: 19:18

Subject's Last Name: PHILLIPS

Date/Time of Test: 04/18/2007 20:04

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.079	19:56
Air Blank	0.000	19:56
Subject Sample #1	0.254	19:58
Breath Volume	1.917	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample #2	0.270	20:03
Breath Volume	1.464	20:03
Air Blank	0.000	20:03
Control Test	0.078	20:03
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 04/19/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: KIRKPATRICK, J,

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: KIRKPATRICK, J

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:40

Gender: Male

Observation Period Began: 21:45

Subject's Last Name: SHANLEY

Date/Time of Test: 04/18/2007 22:13

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:06
Subject Sample #1	0.252	22:08
Breath Volume	2.175	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample #2	0.256	22:11
Breath Volume	2.050	22:11
Air Blank	0.000	22:12
Control Test	0.078	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 04/19/2007 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: JAY, ERIK, J

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: JAY, E

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 18:36

Gender: Male

Observation Period Began: 18:50

Subject's Last Name: ZIEGEL

Date/Time of Test: 04/25/2007 19:17

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.079	19:11
Air Blank	0.000	19:12
Subject Sample #1	0.175	19:12
Breath Volume	1.246	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample #2	0.178	19:16
Breath Volume	2.082	19:16
Air Blank	0.000	19:16
Control Test	0.079	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 05/25/2007 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: KIRKPATRICK, J,

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: KIRKPATRICK, J

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:20

Gender: Male

Observation Period Began: 22:30

Subject's Last Name: MANGRU

Date/Time of Test: 04/25/2007 23:02

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:56
Subject Sample #1	0.204	22:57
Breath Volume	1.730	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample #2	0.189	23:01
Breath Volume	2.703	23:01
Air Blank	0.000	23:01
Control Test	0.078	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 05/25/2007 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: THIVENER, DERRICK, J

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: JAY, ERIC

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:35

Gender: Female

Observation Period Began: 00:45

Subject's Last Name: MOORE

Date/Time of Test: 04/26/2007 01:12

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample #1	0.173	01:08
Breath Volume	1.164	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample #2	0.166	01:11
Breath Volume	2.554	01:11
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 05/25/2007 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: KIRKPATRICK, J,

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: KIRKPATRICK, J

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:30

Gender: Male

Observation Period Began: 21:40

Subject's Last Name: WOREL

Date/Time of Test: 04/26/2007 22:07

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Subject Sample #1	0.143	22:02
Breath Volume	2.886	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample #2	0.145	22:06
Breath Volume	2.449	22:06
Air Blank	0.000	22:06
Control Test	0.079	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 05/25/2007 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: THIVENER, DERRICK, J

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: JAY, ERIC

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:33

Gender: Male

Observation Period Began: 02:37

Subject's Last Name: KING

Date/Time of Test: 04/27/2007 03:04

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.078	02:58
Air Blank	0.000	02:59
Subject Sample #1	0.228	02:59
Breath Volume	3.257	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample #2	0.210	03:03
Breath Volume	2.195	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 05/25/2007 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: JAY, ERIK, J

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: FRYE, D

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:11

Gender: Male

Observation Period Began: 02:37

Subject's Last Name: LOWE

Date/Time of Test: 04/27/2007 03:16

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Subject Sample #1	0.142	03:11
Breath Volume	1.449	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample #2	0.131	03:14
Breath Volume	1.695	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 05/25/2007 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: KIRKPATRICK, J,

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: KIRKPATRICK, J

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 20:45

Gender: Male

Observation Period Began: 20:50

Subject's Last Name: PLAZEWSKI

Date/Time of Test: 04/28/2007 21:17

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.081	21:11
Air Blank	0.000	21:12
Subject Sample #1	0.319	21:12
Breath Volume	3.808	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample #2	0.314	21:15
Breath Volume	3.257	21:15
Air Blank	0.000	21:16
Control Test	0.079	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 05/25/2007 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: JAY, ERIK, J

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: JAY, ERIK

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:52

Gender: Female

Observation Period Began: 01:06

Subject's Last Name: STEUART

Date/Time of Test: 04/29/2007 01:33

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Subject Sample #1	0.189	01:28
Breath Volume	1.699	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample #2	0.190	01:31
Breath Volume	2.082	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 05/25/2007 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001064

Instrument Registered To: NEW PORT RICHEY PD

Software Version: 8100.27

Breath Test Operator: KIRKPATRICK, J,

Breath Test Operator Agency: New Port Richey PD

Arresting Officer: KIRKPATRICK, J

Arresting Officer Agency: New Port Richey PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:20

Gender: Female

Observation Period Began: 01:36

Subject's Last Name: MITCHELL

Date/Time of Test: 04/29/2007 02:05

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample #1	0.084	02:00
Breath Volume	2.378	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample #2	0.083	02:04
Breath Volume	1.894	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 05/25/2007 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001065

Instrument Registered To: TARPON SPRINGS PD

Software Version: 8100.27

Breath Test Operator: OSARCZUK, SCOTT, R

Breath Test Operator Agency: Tarpon Springs PD

Arresting Officer: LORD, JOHN

Arresting Officer Agency: Tarpon Springs PD

Last Agency Inspection: 03/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 18:14

Gender: Male

Observation Period Began: 19:54

Subject's Last Name: BUTLER

Date/Time of Test: 04/09/2007 20:29

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.082	20:23
Air Blank	0.000	20:23
Subject Sample #1	0.187	20:25
Breath Volume	1.867	20:25
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample #2	0.186	20:27
Breath Volume	2.000	20:27
Air Blank	0.000	20:28
Control Test	0.082	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 04/16/2007 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001065

Instrument Registered To: TARPON SPRINGS PD

Software Version: 8100.27

Breath Test Operator: OSARCZUK, SCOTT, R

Breath Test Operator Agency: Tarpon Springs PD

Arresting Officer: OSARCZUK, SCOTT

Arresting Officer Agency: Tarpon Springs PD

Last Agency Inspection: 03/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 19:38

Gender: Male

Observation Period Began: 20:00

Subject's Last Name: VEGA

Date/Time of Test: 04/10/2007 20:35

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.082	20:29
Air Blank	0.000	20:30
Subject Sample #1	0.044	20:30
Breath Volume	1.691	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample #2	0.041	20:34
Breath Volume	1.804	20:34
Air Blank	0.000	20:34
Control Test	0.082	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 04/16/2007 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001065

Instrument Registered To: TARPON SPRINGS PD

Software Version: 8100.27

Breath Test Operator: GASSEN, S,

Breath Test Operator Agency: Tarpon Springs PD

Arresting Officer: GASSEN, S

Arresting Officer Agency: Tarpon Springs PD

Last Agency Inspection: 04/16/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:00

Gender: Male

Observation Period Began: 00:31

Subject's Last Name: BELLES

Date/Time of Test: 04/17/2007 00:57

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.082	00:52
Air Blank	0.000	00:52
Subject Sample #1	0.196	00:53
Breath Volume	3.601	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample #2	0.196	00:56
Breath Volume	3.027	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 05/26/2007 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001065

Instrument Registered To: TARPON SPRINGS PD

Software Version: 8100.27

Breath Test Operator: GEER, MATT, T

Breath Test Operator Agency: Tarpon Springs PD

Arresting Officer: GEER, MATT

Arresting Officer Agency: Tarpon Springs PD

Last Agency Inspection: 04/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6307011

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 11:36

Gender: Male

Observation Period Began: 11:58

Subject's Last Name: HULL

Date/Time of Test: 04/23/2007 13:32

Diagnostics Check	OK	13:30	
Air Blank	0.000	13:30	
Control Test	0.083	13:30	
Air Blank	0.000	13:31	
Subject Sample #1		13:31	Subject Test Refused
Breath Volume	0.000	13:31	
Air Blank	0.000	13:32	
Control Test	0.082	13:32	
Air Blank	0.000	13:32	
Diagnostics Check	OK	13:32	

Data Download Date/Time: 05/26/2007 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001066

Instrument Registered To: ST PETE BEACH PD

Software Version: 8100.27

Breath Test Operator: WEIL, ROBERT, A

Breath Test Operator Agency: St. Pete Beach PD

Arresting Officer: WEIL, ROBERT

Arresting Officer Agency: St. Pete Beach PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:17

Gender: Male

Observation Period Began: 02:40

Subject's Last Name: DOBRIJEVIC

Date/Time of Test: 04/08/2007 03:09

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:03
Subject Sample #1	0.113	03:04
Breath Volume	2.695	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample #2	0.122	03:07
Breath Volume	3.519	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 04/19/2007 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001066

Instrument Registered To: ST PETE BEACH PD
Software Version: 8100.27
Breath Test Operator: WEIL, ROBERT, A
Breath Test Operator Agency: St. Pete Beach PD
Arresting Officer: WEIL, ROBERT
Arresting Officer Agency: St. Pete Beach PD
Last Agency Inspection: 03/22/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 630602I
Dry Gas Expiry Date: 11/06/2008
Violation Code: DUI
Arrest Time: 01:42
Gender: Male
Observation Period Began: 01:52

Subject's Last Name: CAMPBELL

Date/Time of Test: 04/09/2007 02:25

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.080	02:14	
Air Blank	0.000	02:14	
Subject Sample #1	0.194	02:17	Volume Not Met
Breath Volume	0.042	02:17	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample #2	0.220	02:23	Volume Not Met
Breath Volume	0.058	02:23	
Air Blank	0.000	02:24	
Control Test	0.078	02:24	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 04/19/2007 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001066

Instrument Registered To: ST PETE BEACH PD

Software Version: 8100.27

Breath Test Operator: WEIL, ROBERT, A

Breath Test Operator Agency: St. Pete Beach PD

Arresting Officer: WEIL, ROBERT

Arresting Officer Agency: St. Pete Beach PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:40

Gender: Female

Observation Period Began: 02:57

Subject's Last Name: BENNETT

Date/Time of Test: 04/10/2007 03:31

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.079	03:18	
Air Blank	0.000	03:18	
Subject Sample #1	0.177	03:20	
Breath Volume	1.691	03:20	
Air Blank	0.000	03:21	
Air Blank	0.000	03:23	
Subject Sample #2	0.164	03:26	Volume Not Met
Breath Volume	0.390	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:28	
Subject Sample #3	0.159	03:30	
Breath Volume	2.050	03:30	
Air Blank	0.000	03:30	
Control Test	0.075	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 04/19/2007 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001066

Instrument Registered To: ST PETE BEACH PD

Software Version: 8100.27

Breath Test Operator: WEIL, ROBERT, A

Breath Test Operator Agency: St. Pete Beach PD

Arresting Officer: CURTIN, DENNIS

Arresting Officer Agency: St. Pete Beach PD

Last Agency Inspection: 03/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 12:55

Gender: Female

Observation Period Began: 01:14

Subject's Last Name: FORSTER

Date/Time of Test: 04/19/2007 01:37

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:35	
Air Blank	0.000	01:35	
Subject Sample #1		01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Control Test	0.081	01:36	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 04/19/2007 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001066

Instrument Registered To: ST PETE BEACH PD
Software Version: 8100.27
Breath Test Operator: NOVAK, R, A
Breath Test Operator Agency: St. Pete Beach PD
Arresting Officer: RALEIGH, A
Arresting Officer Agency: St. Pete Beach PD
Last Agency Inspection: 04/19/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 630602I
Dry Gas Expiry Date: 11/06/2008
Violation Code: DUI
Arrest Time: 01:13
Gender: Male
Observation Period Began: 01:23

Subject's Last Name: CLOUSER

Date/Time of Test: 04/21/2007 01:49

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample #1	0.164	01:45
Breath Volume	3.445	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:47
Subject Sample #2	0.169	01:48
Breath Volume	2.421	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 05/22/2007 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001066

Instrument Registered To: ST PETE BEACH PD

Software Version: 8100.27

Breath Test Operator: WEIL, ROBERT, A

Breath Test Operator Agency: St. Pete Beach PD

Arresting Officer: WEIL, ROBERT

Arresting Officer Agency: St. Pete Beach PD

Last Agency Inspection: 04/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:42

Gender: Male

Observation Period Began: 23:52

Subject's Last Name: RIORDAN

Date/Time of Test: 04/22/2007 00:19

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Subject Sample #1	0.221	00:15
Breath Volume	3.617	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample #2	0.223	00:18
Breath Volume	3.910	00:18
Air Blank	0.000	00:18
Control Test	0.078	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 05/22/2007 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001066

Instrument Registered To: ST PETE BEACH PD

Software Version: 8100.27

Breath Test Operator: NOVAK, R, A

Breath Test Operator Agency: St. Pete Beach PD

Arresting Officer: RALEIGH, A

Arresting Officer Agency: St. Pete Beach PD

Last Agency Inspection: 04/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:36

Gender: Male

Observation Period Began: 02:45

Subject's Last Name: ENGLETON

Date/Time of Test: 04/27/2007 03:08

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.081	03:06	
Air Blank	0.000	03:07	
Subject Sample #1		03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.081	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 05/22/2007 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001066

Instrument Registered To: ST PETE BEACH PD

Software Version: 8100.27

Breath Test Operator: BAAR, JOSHUA, M

Breath Test Operator Agency: St. Pete Beach PD

Arresting Officer: LEGENDRE, KYLE

Arresting Officer Agency: St. Pete Beach PD

Last Agency Inspection: 04/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630602I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 18:56

Gender: Male

Observation Period Began: 19:05

Subject's Last Name: MASTERS

Date/Time of Test: 04/29/2007 19:31

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.080	19:26
Air Blank	0.000	19:26
Subject Sample #1	0.188	19:26
Breath Volume	2.359	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample #2	0.172	19:29
Breath Volume	1.867	19:29
Air Blank	0.000	19:30
Control Test	0.080	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 05/22/2007 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001066

Instrument Registered To: ST PETE BEACH PD
Software Version: 8100.27
Breath Test Operator: JOHNSON, BRUCE, M
Breath Test Operator Agency: St. Pete Beach PD
Arresting Officer: RALEIGH, ADDIE
Arresting Officer Agency: St. Pete Beach PD
Last Agency Inspection: 04/19/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 630602I
Dry Gas Expiry Date: 11/06/2008
Violation Code: DUI
Arrest Time: 08:26
Gender: Male
Observation Period Began: 09:09

Subject's Last Name: WIDES

Date/Time of Test: 04/30/2007 09:35

Diagnostics Check	OK	09:32	
Air Blank	0.000	09:32	
Control Test	0.080	09:32	
Air Blank	0.000	09:33	
Subject Sample #1		09:33	Subject Test Refused
Breath Volume	0.000	09:33	
Air Blank	0.000	09:34	
Control Test	0.080	09:34	
Air Blank	0.000	09:35	
Diagnostics Check	OK	09:35	

Data Download Date/Time: 05/22/2007 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001068

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: BODLE, R, S

Breath Test Operator Agency: Clearwater PD

Arresting Officer: LULEK, T

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 615802I

Dry Gas Expiry Date: 06/09/2008

Violation Code: DUI

Arrest Time: 21:20

Gender: Female

Observation Period Began: 21:30

Subject's Last Name: HATKINS

Date/Time of Test: 04/28/2007 22:04

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:51	
Control Test	0.081	21:51	
Air Blank	0.000	21:52	
Subject Sample #1	0.000	21:55	Volume Not Met
Breath Volume	0.539	21:55	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample #2	0.000	21:59	
Breath Volume	2.050	21:59	
Air Blank	0.000	22:00	
Air Blank	0.000	22:02	
Subject Sample #3	0.000	22:02	
Breath Volume	1.953	22:02	
Air Blank	0.000	22:03	
Control Test	0.081	22:03	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 05/25/2007 06:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001070

Instrument Registered To: FT MEADE PD

Software Version: 8100.27

Breath Test Operator: VALLLE, CARLOS,

Breath Test Operator Agency: Fort Meade PD

Arresting Officer: VALLE, CARLOS

Arresting Officer Agency: Fort Meade PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 03:15

Gender: Male

Observation Period Began: 03:50

Subject's Last Name: SOUKSAVATH

Date/Time of Test: 04/15/2007 04:19

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.077	04:13
Air Blank	0.000	04:14
Subject Sample #1	0.112	04:15
Breath Volume	2.558	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.114	04:18
Breath Volume	2.671	04:18
Air Blank	0.000	04:18
Control Test	0.077	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 05/07/2007 19:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001071

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: DURRETTE, TRAVIS, K

Breath Test Operator Agency: Clay County SO

Arresting Officer: DURRETTE, T

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: DUI

Arrest Time: 13:00

Gender: Male

Observation Period Began: 14:10

Subject's Last Name: KOWALCZK

Date/Time of Test: 04/20/2007 14:40

Diagnostics Check	OK	14:34
Air Blank	0.000	14:35
Control Test	0.079	14:35
Air Blank	0.000	14:36
Subject Sample #1	0.000	14:36
Breath Volume	3.078	14:36
Air Blank	0.000	14:37
Air Blank	0.000	14:39
Subject Sample #2	0.000	14:39
Breath Volume	3.410	14:39
Air Blank	0.000	14:39
Control Test	0.079	14:40
Air Blank	0.000	14:40
Diagnostics Check	OK	14:40

Data Download Date/Time: 04/22/2007 00:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001071

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARNOLD, WILLIAM, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: CRAVEN, SGT

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: ADMIN

Arrest Time: 19:30

Gender: Male

Observation Period Began: 19:34

Subject's Last Name: EILAND

Date/Time of Test: 04/20/2007 20:00

Diagnostics Check	OK	19:54
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:55
Subject Sample #1	0.000	19:56
Breath Volume	2.164	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample #2	0.000	19:59
Breath Volume	3.234	19:59
Air Blank	0.000	19:59
Control Test	0.079	20:00
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 04/22/2007 00:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001071

Instrument Registered To: CLAY COUNTY SO

Software Version: 8100.27

Breath Test Operator: ARNOLD, WILLIAM, A

Breath Test Operator Agency: Clay County SO

Arresting Officer: CRAVEN, SGT

Arresting Officer Agency: Clay County SO

Last Agency Inspection: 03/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 520102I

Dry Gas Expiry Date: 07/22/2007

Violation Code: ADMIN

Arrest Time: 19:30

Gender: Male

Observation Period Began: 19:40

Subject's Last Name: DAWSON

Date/Time of Test: 04/20/2007 20:10

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.079	20:04
Air Blank	0.000	20:05
Subject Sample #1	0.168	20:05
Breath Volume	2.300	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample #2	0.168	20:08
Breath Volume	2.835	20:08
Air Blank	0.000	20:09
Control Test	0.080	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 04/22/2007 00:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: ALONSO, HIGINIO, L

Breath Test Operator Agency: FHP Troop E

Arresting Officer: HARDY, JOE

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 10:51

Gender: Male

Observation Period Began: 00:30

Subject's Last Name: ALONSO

Date/Time of Test: 04/03/2007 01:05

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	01:00
Subject Sample #1	0.151	01:01
Breath Volume	2.480	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample #2	0.144	01:04
Breath Volume	2.605	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: GUTIERREZ, ENRIQUE, F

Breath Test Operator Agency: FHP Troop E

Arresting Officer: GUTIERREZ, ENRIQUE

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 17:44

Gender: Female

Observation Period Began: 19:05

Subject's Last Name: BERNALES

Date/Time of Test: 04/03/2007 19:30

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:28	
Control Test	0.081	19:28	
Air Blank	0.000	19:29	
Subject Sample #1		19:29	Subject Test Refused
Breath Volume	0.000	19:29	
Air Blank	0.000	19:29	
Control Test	0.080	19:30	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: MARDIS, JOHN, D

Breath Test Operator Agency: FHP Troop E

Arresting Officer: ROGERS, WALTER

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 02:30

Gender: Male

Observation Period Began: 02:50

Subject's Last Name: MARONI

Date/Time of Test: 04/04/2007 04:55

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:53	
Control Test	0.081	04:53	
Air Blank	0.000	04:54	
Subject Sample #1		04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:54	
Control Test	0.081	04:55	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: GUTIERREZ, ENRIQUE, F

Breath Test Operator Agency: FHP Troop E

Arresting Officer: GUTIERREZ, ENRIQUE

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 19:38

Gender: Female

Observation Period Began: 20:25

Subject's Last Name: ROTGER

Date/Time of Test: 04/06/2007 20:59

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:47	
Control Test	0.081	20:48	
Air Blank	0.000	20:48	
Subject Sample #1	0.183	20:51	Volume Not Met
Breath Volume	0.058	20:51	
Air Blank	0.000	20:52	
Air Blank	0.000	20:54	
Subject Sample #2	0.251	20:57	Volume Not Met
Breath Volume	0.015	20:57	
Air Blank	0.000	20:58	
Control Test	0.078	20:58	
Air Blank	0.000	20:59	
Diagnostics Check	OK	20:59	

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: GALLARDO, JORGE, L

Breath Test Operator Agency: FHP Troop E

Arresting Officer: GALLARDO, JORGE

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 21:33

Gender: Male

Observation Period Began: 22:00

Subject's Last Name: CEDENO

Date/Time of Test: 04/06/2007 22:55

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:53	
Control Test	0.080	22:53	
Air Blank	0.000	22:54	
Subject Sample #1		22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:54	
Control Test	0.081	22:55	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: ARIAS, ELOY,

Breath Test Operator Agency: FHP Troop E

Arresting Officer: ARIAS, ELOY

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 15:08

Gender: Male

Observation Period Began: 16:03

Subject's Last Name: HENRIQUEZ

Date/Time of Test: 04/07/2007 16:29

Diagnostics Check	OK	16:23
Air Blank	0.000	16:23
Control Test	0.082	16:24
Air Blank	0.000	16:24
Subject Sample #1	0.105	16:25
Breath Volume	2.585	16:25
Air Blank	0.000	16:25
Air Blank	0.000	16:27
Subject Sample #2	0.103	16:28
Breath Volume	2.453	16:28
Air Blank	0.000	16:28
Control Test	0.080	16:29
Air Blank	0.000	16:29
Diagnostics Check	OK	16:29

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: MARDIS, JOHN, D

Breath Test Operator Agency: FHP Troop E

Arresting Officer: MARDIS, JOHN

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 04:40

Gender: Male

Observation Period Began: 06:00

Subject's Last Name: ESCOBAR

Date/Time of Test: 04/08/2007 06:28

Diagnostics Check	OK	06:21
Air Blank	0.000	06:21
Control Test	0.081	06:22
Air Blank	0.000	06:22
Subject Sample #1	0.080	06:23
Breath Volume	1.339	06:23
Air Blank	0.000	06:24
Air Blank	0.000	06:25
Subject Sample #2	0.081	06:26
Breath Volume	2.839	06:26
Air Blank	0.000	06:27
Control Test	0.080	06:27
Air Blank	0.000	06:27
Diagnostics Check	OK	06:28

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: DEPESTRE, GREGORY,

Breath Test Operator Agency: FHP Troop E

Arresting Officer: DEPESTRE, GREGORY

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 23:20

Gender: Male

Observation Period Began: 01:30

Subject's Last Name: APOLINARIO

Date/Time of Test: 04/09/2007 02:54

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.081	02:50	
Air Blank	0.000	02:51	
Subject Sample #1		02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank	0.000	02:53	
Control Test	0.080	02:54	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: LOPEZ, WILLIAM, N

Breath Test Operator Agency: FHP Troop E

Arresting Officer: LOPEZ, WILLIAM

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 03:21

Gender: Female

Observation Period Began: 03:21

Subject's Last Name: DOMBKOWSKI

Date/Time of Test: 04/09/2007 04:31

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Subject Sample #1	0.175	04:26
Breath Volume	2.113	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample #2	0.163	04:29
Breath Volume	1.750	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: OLDS, LEE, V

Breath Test Operator Agency: FHP Troop E

Arresting Officer: OLDS, LEE

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:15

Gender: Female

Observation Period Began: 02:25

Subject's Last Name: CLAPPERTON

Date/Time of Test: 04/13/2007 03:12

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:02	
Subject Sample #1		03:05	No Sample Provided
Breath Volume	0.000	03:05	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample #2		03:10	No Sample Provided
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.081	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: MORISSEAU, JACKSON,

Breath Test Operator Agency: FHP Troop E

Arresting Officer: MORISSEAU, JACKSON

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 18:03

Gender: Male

Observation Period Began: 18:15

Subject's Last Name: HIERRO

Date/Time of Test: 04/21/2007 19:11

Diagnostics Check	OK	19:04
Air Blank	0.000	19:05
Control Test	0.081	19:05
Air Blank	0.000	19:05
Subject Sample #1	0.235	19:06
Breath Volume	1.960	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample #2	0.242	19:09
Breath Volume	1.750	19:09
Air Blank	0.000	19:10
Control Test	0.079	19:10
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: ADDISON, RAYMOND, E

Breath Test Operator Agency: FHP Troop E

Arresting Officer: WILFONG, MICHAEL

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 17:56

Gender: Male

Observation Period Began: 18:45

Subject's Last Name: HERNANDEZ

Date/Time of Test: 04/22/2007 19:22

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.081	19:16
Air Blank	0.000	19:16
Subject Sample #1	0.125	19:17
Breath Volume	1.753	19:17
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample #2	0.131	19:21
Breath Volume	2.222	19:21
Air Blank	0.000	19:21
Control Test	0.081	19:22
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: FRAIOLI, JAY, A

Breath Test Operator Agency: FHP Troop E

Arresting Officer: FRAIOLI, JAY

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 22:33

Gender: Male

Observation Period Began: 23:15

Subject's Last Name: INCLAN

Date/Time of Test: 04/25/2007 23:41

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample #1	0.189	23:37
Breath Volume	4.261	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample #2	0.192	23:40
Breath Volume	4.871	23:40
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: JONES, EUGENE, V

Breath Test Operator Agency: FHP Troop E

Arresting Officer: JONES, EUGENE

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 02:18

Subject's Last Name: MARTINEZ

Date/Time of Test: 04/29/2007 02:53

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.081	02:50	
Air Blank	0.000	02:50	
Subject Sample #1		02:53	Sequence Aborted
Breath Volume	0.000	02:53	
Air Blank	0.000	02:53	

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: JONES, EUGENE, V

Breath Test Operator Agency: FHP Troop E

Arresting Officer: JONES, EUGENE

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 02:18

Subject's Last Name: MARTINEZ

Date/Time of Test: 04/29/2007 03:02

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.082	03:00	
Air Blank	0.000	03:01	
Subject Sample #1		03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.080	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001075

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: JOHNSON, SAMUEL, O

Breath Test Operator Agency: FHP Troop E

Arresting Officer: JOHNSON, SAMUEL

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 02:57

Gender: Male

Observation Period Began: 04:52

Subject's Last Name: LOPEZ

Date/Time of Test: 04/30/2007 05:26

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:24	
Control Test	0.081	05:24	
Air Blank	0.000	05:24	
Subject Sample #1		05:25	Subject Test Refused
Breath Volume	0.000	05:25	
Air Blank	0.000	05:25	
Control Test	0.081	05:25	
Air Blank	0.000	05:26	
Diagnostics Check	OK	05:26	

Data Download Date/Time: 04/30/2007 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: STONE, HERSHEL, A

Breath Test Operator Agency: Hardee County SO

Arresting Officer: ARCE, JOE

Arresting Officer Agency: Hardee County SO

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 00:30

Gender: Male

Observation Period Began: 00:30

Subject's Last Name: HERNANDEZ

Date/Time of Test: 04/01/2007 01:15

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:10
Subject Sample #1	0.214	01:11
Breath Volume	3.156	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample #2	0.220	01:14
Breath Volume	3.410	01:14
Air Blank	0.000	01:15
Control Test	0.078	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 04/12/2007 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: ALBRITTON, BRIAN, S

Breath Test Operator Agency: Hardee County SO

Arresting Officer: WOODY, NATHAN

Arresting Officer Agency: Hardee County SO

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 00:15

Gender: Male

Observation Period Began: 00:35

Subject's Last Name: PEREZ

Date/Time of Test: 04/02/2007 01:06

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample #1	0.243	01:02
Breath Volume	1.437	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample #2	0.247	01:05
Breath Volume	1.117	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 04/12/2007 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: OBRYAN, DANNY, L

Breath Test Operator Agency: Hardee County SO

Arresting Officer: SCHEILD, JOHN

Arresting Officer Agency: Bowling Green PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 12:49

Gender: Male

Observation Period Began: 13:55

Subject's Last Name: PRICE

Date/Time of Test: 04/06/2007 14:37

Diagnostics Check	OK	14:25	
Air Blank	0.000	14:26	
Control Test	0.080	14:26	
Air Blank	0.000	14:26	
Subject Sample #1		14:30	No Sample Provided
Breath Volume	0.000	14:30	
Air Blank	0.000	14:30	
Air Blank	0.000	14:32	
Subject Sample #2		14:35	No Sample Provided
Breath Volume	0.000	14:35	
Air Blank	0.000	14:36	
Control Test	0.081	14:36	
Air Blank	0.000	14:36	
Diagnostics Check	OK	14:37	

Data Download Date/Time: 04/12/2007 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: ALBRITTON, BRIAN, S

Breath Test Operator Agency: Hardee County SO

Arresting Officer: GRACE, SCOTT

Arresting Officer Agency: Bowling Green PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 23:56

Gender: Female

Observation Period Began: 01:03

Subject's Last Name: THOMAS

Date/Time of Test: 04/10/2007 01:32

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:27
Air Blank	0.000	01:27
Subject Sample #1	0.095	01:27
Breath Volume	2.476	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample #2	0.093	01:31
Breath Volume	2.390	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 04/12/2007 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LINDER, JAYNE, S

Breath Test Operator Agency: Hardee County SO

Arresting Officer: WHATLEY, MATTHEW

Arresting Officer Agency: Wauchula PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 02:26

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: FLORES

Date/Time of Test: 04/13/2007 03:37

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample #1	0.199	03:30
Breath Volume	2.562	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample #2	0.189	03:36
Breath Volume	2.839	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/27/2007 01:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: ALBRITTON, BRIAN, S

Breath Test Operator Agency: Hardee County SO

Arresting Officer: BENAVIDEZ, KIM

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 15:13

Gender: Male

Observation Period Began: 15:30

Subject's Last Name: HARRISON

Date/Time of Test: 04/18/2007 15:57

Diagnostics Check	OK	15:51
Air Blank	0.000	15:51
Control Test	0.080	15:51
Air Blank	0.000	15:52
Subject Sample #1	0.009	15:52
Breath Volume	2.453	15:52
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample #2	0.008	15:56
Breath Volume	1.425	15:56
Air Blank	0.000	15:56
Control Test	0.079	15:56
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

Data Download Date/Time: 04/27/2007 01:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: OBRYAN, DANNY, L

Breath Test Operator Agency: Hardee County SO

Arresting Officer: BRIDGES, JEREMY

Arresting Officer Agency: Bowling Green PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:25

Gender: Male

Observation Period Began: 01:50

Subject's Last Name: ROMO

Date/Time of Test: 04/22/2007 02:23

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.078	02:13	
Air Blank	0.000	02:13	
Subject Sample #1		02:16	No Sample Provided
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample #2		02:22	No Sample Provided
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.078	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 04/27/2007 01:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: OBRYAN, DANNY, L

Breath Test Operator Agency: Hardee County SO

Arresting Officer: WHATLEY, MATTHEW

Arresting Officer Agency: Wauchula PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 21:38

Gender: Male

Observation Period Began: 21:55

Subject's Last Name: MARTINEZ

Date/Time of Test: 04/22/2007 22:24

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:17
Subject Sample #1	0.195	22:18
Breath Volume	1.390	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample #2	0.190	22:22
Breath Volume	1.234	22:22
Air Blank	0.000	22:23
Control Test	0.077	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:24

Data Download Date/Time: 04/27/2007 01:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: POWELL, BRIAN, K

Breath Test Operator Agency: Hardee County SO

Arresting Officer: GARZA, GABE

Arresting Officer Agency: Wauchula PD

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: OTHER

Arrest Time: 10:30

Gender: Male

Observation Period Began: 22:35

Subject's Last Name: MEDINA

Date/Time of Test: 04/27/2007 23:04

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:58
Subject Sample #1	0.264	22:59
Breath Volume	1.625	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample #2	0.278	23:02
Breath Volume	2.964	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 05/30/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CARTWRIGHT, GARY, W

Breath Test Operator Agency: Hardee County SO

Arresting Officer: WHATLEY, MATTHEW

Arresting Officer Agency: Wauchula PD

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: MARTINEZ

Date/Time of Test: 04/29/2007 02:23

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:20	
Control Test	0.081	02:20	
Air Blank	0.000	02:21	
Subject Sample #1		02:22	RFI Detect
Breath Volume	0.183	02:22	
Air Blank		02:22	RFI Detect
Air Blank	0.000	02:23	

Data Download Date/Time: 05/30/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001077

Instrument Registered To: HARDEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CARTWRIGHT, GARY, W

Breath Test Operator Agency: Hardee County SO

Arresting Officer: WHATLEY, MATTHEW

Arresting Officer Agency: Wauchula PD

Last Agency Inspection: 04/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: MARTINEZ

Date/Time of Test: 04/29/2007 02:32

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Subject Sample #1	0.088	02:28
Breath Volume	2.089	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample #2	0.076	02:31
Breath Volume	1.195	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 05/30/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001078

Instrument Registered To: DUNDEE PD

Software Version: 8100.27

Breath Test Operator: YODONIS, MICHAEL, C

Breath Test Operator Agency: Lake Wales PD

Arresting Officer: YODONIS, MICHAEL

Arresting Officer Agency: Lake Wales PD

Last Agency Inspection: 03/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 23:10

Gender: Male

Observation Period Began: 23:40

Subject's Last Name: MAKOWSKI

Date/Time of Test: 04/07/2007 00:11

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:06
Subject Sample #1	0.202	00:06
Breath Volume	1.382	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample #2	0.196	00:09
Breath Volume	2.144	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 04/09/2007 03:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001078

Instrument Registered To: DUNDEE PD

Software Version: 8100.27

Breath Test Operator: GARNER, ALEXANDER, W

Breath Test Operator Agency: Dundee PD

Arresting Officer: HULVERSON, REGINA

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 04/09/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 02:52

Gender: Male

Observation Period Began: 03:23

Subject's Last Name: WHITELOW

Date/Time of Test: 04/27/2007 03:55

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:44	
Subject Sample #1		03:48	No Sample Provided
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample #2		03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Control Test	0.081	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 05/13/2007 20:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: LACEY, TIMOTHY, W

Breath Test Operator Agency: Polk County SO

Arresting Officer: BAIN, MICHAEL

Arresting Officer Agency: Bartow PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:20

Gender: Male

Observation Period Began: 01:31

Subject's Last Name: MCCAIN

Date/Time of Test: 04/04/2007 01:59

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Subject Sample #1	0.161	01:55
Breath Volume	2.800	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:57
Subject Sample #2	0.164	01:58
Breath Volume	2.511	01:58
Air Blank	0.000	01:58
Control Test	0.079	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BREWIN, THOMAS, H

Breath Test Operator Agency: Polk County SO

Arresting Officer: BREWIN, THOMAS

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:32

Gender: Male

Observation Period Began: 00:53

Subject's Last Name: SALINAS-MARTINEZ

Date/Time of Test: 04/05/2007 01:48

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:44	
Subject Sample #1		01:46	Subject Test Refused
Breath Volume	0.093	01:46	
Air Blank	0.000	01:47	
Control Test	0.080	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROWN, DAVE, D

Breath Test Operator Agency: Polk County SO

Arresting Officer: HULVERSON, REGINA

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:08

Gender: Male

Observation Period Began: 03:52

Subject's Last Name: CHARRAN SINGH

Date/Time of Test: 04/05/2007 04:19

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample #1	0.228	04:14
Breath Volume	1.230	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.231	04:17
Breath Volume	2.019	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROWN, DAVE, D

Breath Test Operator Agency: Polk County SO

Arresting Officer: MONTALBANO, JONATHAN

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:34

Gender: Male

Observation Period Began: 04:00

Subject's Last Name: STAPLETON

Date/Time of Test: 04/05/2007 04:30

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.079	04:27	
Air Blank	0.000	04:28	
Subject Sample #1		04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.080	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BREWIN, THOMAS, H

Breath Test Operator Agency: Polk County SO

Arresting Officer: BREWIN, THOMAS

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:45

Gender: Male

Observation Period Began: 01:05

Subject's Last Name: MCCULLOUGH

Date/Time of Test: 04/06/2007 01:44

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample #1		01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOYER, G,

Breath Test Operator Agency: Polk County SO

Arresting Officer: HEATH, C

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 16:45

Gender: Male

Observation Period Began: 18:23

Subject's Last Name: ALBOS

Date/Time of Test: 04/06/2007 18:56

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.080	18:51
Air Blank	0.000	18:52
Subject Sample #1	0.008	18:52
Breath Volume	2.687	18:52
Air Blank	0.000	18:52
Air Blank	0.000	18:54
Subject Sample #2	0.008	18:55
Breath Volume	1.875	18:55
Air Blank	0.000	18:55
Control Test	0.080	18:56
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BREWIN, THOMAS, H

Breath Test Operator Agency: Polk County SO

Arresting Officer: BREWIN, THOMAS

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:05

Gender: Male

Observation Period Began: 22:30

Subject's Last Name: KING

Date/Time of Test: 04/06/2007 23:43

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:38
Subject Sample #1	0.245	23:38
Breath Volume	3.324	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample #2	0.237	23:41
Breath Volume	2.488	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: YODONIS, MICHAEL, C

Breath Test Operator Agency: Lake Wales PD

Arresting Officer: YODONIS, MICHAEL

Arresting Officer Agency: Lake Wales PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:59

Gender: Male

Observation Period Began: 00:00

Subject's Last Name: SACHS

Date/Time of Test: 04/08/2007 00:32

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Subject Sample #1	0.281	00:25
Breath Volume	2.617	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample #2	0.259	00:28
Breath Volume	4.378	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample #3	0.259	00:31
Breath Volume	4.027	00:31
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: NEIL, STEVEN, E

Breath Test Operator Agency: Polk County SO

Arresting Officer: ANDERSON, ROBERT

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:08

Gender: Male

Observation Period Began: 03:18

Subject's Last Name: RODRIGUEZ

Date/Time of Test: 04/08/2007 04:05

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Subject Sample #1	0.100	04:00
Breath Volume	2.421	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample #2	0.100	04:03
Breath Volume	1.992	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: YODONIS, MICHAEL, C

Breath Test Operator Agency: Lake Wales PD

Arresting Officer: YODONIS, MICHAEL

Arresting Officer Agency: Lake Wales PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:30

Gender: Male

Observation Period Began: 22:50

Subject's Last Name: SMITH

Date/Time of Test: 04/08/2007 23:27

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:21
Subject Sample #1	0.269	23:22
Breath Volume	2.707	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample #2	0.260	23:25
Breath Volume	2.000	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:27

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROWN, DAVE, D

Breath Test Operator Agency: Polk County SO

Arresting Officer: MONTALBANO, JONATHAN

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 17:35

Gender: Male

Observation Period Began: 18:05

Subject's Last Name: MITCHELL

Date/Time of Test: 04/09/2007 18:50

Diagnostics Check	OK	18:37	
Air Blank	0.000	18:38	
Control Test	0.079	18:38	
Air Blank	0.000	18:39	
Subject Sample #1	0.302	18:42	Volume Not Met
Breath Volume	0.078	18:42	
Air Blank	0.000	18:43	
Air Blank	0.000	18:44	
Subject Sample #2	0.291	18:45	
Breath Volume	1.117	18:45	
Air Blank	0.000	18:46	
Air Blank	0.000	18:47	
Subject Sample #3	0.293	18:48	
Breath Volume	2.015	18:48	
Air Blank	0.000	18:49	
Control Test	0.077	18:49	
Air Blank	0.000	18:49	
Diagnostics Check	OK	18:50	

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROWN, DAVE, D

Breath Test Operator Agency: Polk County SO

Arresting Officer: MONTALBANO, JONATHAN

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:41

Gender: Male

Observation Period Began: 03:08

Subject's Last Name: TAYLOR

Date/Time of Test: 04/11/2007 03:41

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample #1	0.000	03:37
Breath Volume	2.453	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample #2	0.000	03:40
Breath Volume	2.945	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOYER, G,

Breath Test Operator Agency: Polk County SO

Arresting Officer: BREWIN, T

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 00:26

Gender: Male

Observation Period Began: 01:25

Subject's Last Name: ORTIZ

Date/Time of Test: 04/13/2007 01:58

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:46	
Subject Sample #1		01:50	No Sample Provided
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Air Blank	0.000	01:52	
Subject Sample #2	0.098	01:54	
Breath Volume	2.484	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:56	
Subject Sample #3	0.100	01:57	
Breath Volume	2.261	01:57	
Air Blank	0.000	01:57	
Control Test	0.079	01:58	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: TARTE, PHILLIP, A

Breath Test Operator Agency: Polk County SO

Arresting Officer: MILLS, QUENTIN

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 08:40

Gender: Male

Observation Period Began: 09:25

Subject's Last Name: LEON MARTINEZ

Date/Time of Test: 04/14/2007 10:04

Diagnostics Check	OK	09:58
Air Blank	0.000	09:58
Control Test	0.079	09:58
Air Blank	0.000	09:59
Subject Sample #1	0.052	09:59
Breath Volume	1.410	09:59
Air Blank	0.000	10:00
Air Blank	0.000	10:02
Subject Sample #2	0.053	10:02
Breath Volume	1.246	10:02
Air Blank	0.000	10:03
Control Test	0.079	10:03
Air Blank	0.000	10:04
Diagnostics Check	OK	10:04

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROWN, DAVE, D

Breath Test Operator Agency: Polk County SO

Arresting Officer: HAMPTON, JAMES

Arresting Officer Agency: Lake Wales PD

Last Agency Inspection: 03/21/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 22:40

Gender: Male

Observation Period Began: 23:30

Subject's Last Name: PATRICK

Date/Time of Test: 04/15/2007 00:15

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.079	00:13	
Air Blank	0.000	00:14	
Subject Sample #1		00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 04/15/2007 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROWN, DAVE, D

Breath Test Operator Agency: Polk County SO

Arresting Officer: MYERS, LINDSEY

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 03:25

Gender: Female

Observation Period Began: 04:19

Subject's Last Name: MCCULLUM

Date/Time of Test: 04/19/2007 04:54

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:51	
Control Test	0.079	04:52	
Air Blank	0.000	04:52	
Subject Sample #1		04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:53	
Control Test	0.080	04:53	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 05/16/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOYER, G,

Breath Test Operator Agency: Polk County SO

Arresting Officer: COOK, C

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 12:52

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: SAVORY

Date/Time of Test: 04/21/2007 02:13

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Subject Sample #1	0.121	02:07
Breath Volume	1.437	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #2	0.123	02:11
Breath Volume	2.089	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:13

Data Download Date/Time: 05/16/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOYER, G,

Breath Test Operator Agency: Polk County SO

Arresting Officer: CHURCH, J

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:33

Gender: Male

Observation Period Began: 04:00

Subject's Last Name: BALTAZAR

Date/Time of Test: 04/22/2007 04:30

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Subject Sample #1	0.209	04:23
Breath Volume	1.214	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.196	04:28
Breath Volume	1.132	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 05/16/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: MOYER, G, 4

Breath Test Operator Agency: Polk County SO

Arresting Officer: OSBORNE, B

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:21

Gender: Male

Observation Period Began: 00:08

Subject's Last Name: ORTEGA

Date/Time of Test: 04/23/2007 00:52

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.078	00:48	
Air Blank	0.000	00:48	
Subject Sample #1		00:51	Subject Test Refused
Breath Volume	0.316	00:51	
Air Blank	0.000	00:51	
Control Test	0.079	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 05/16/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: ROBERTS, DANIEL, J

Breath Test Operator Agency: Polk County SO

Arresting Officer: KEEL, RONALD

Arresting Officer Agency: Lake Wales PD

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:58

Gender: Male

Observation Period Began: 00:43

Subject's Last Name: GARCIA

Date/Time of Test: 04/25/2007 01:19

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:14
Subject Sample #1	0.200	01:14
Breath Volume	1.683	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample #2	0.187	01:17
Breath Volume	2.402	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 05/16/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: ROBERTS, DANIEL, J

Breath Test Operator Agency: Polk County SO

Arresting Officer: DEWEESE, KEN

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 23:47

Gender: Female

Observation Period Began: 01:12

Subject's Last Name: OLSON

Date/Time of Test: 04/25/2007 01:45

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.077	01:35
Air Blank	0.000	01:36
Subject Sample #1	0.258	01:39
Breath Volume	1.800	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample #2	0.260	01:43
Breath Volume	2.234	01:43
Air Blank	0.000	01:44
Control Test	0.077	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 05/16/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROWN, DAVE, D

Breath Test Operator Agency: Polk County SO

Arresting Officer: MOVAHEDI, SARAH

Arresting Officer Agency: Mulberry PD

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:20

Gender: Male

Observation Period Began: 03:50

Subject's Last Name: VARNES

Date/Time of Test: 04/25/2007 04:29

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Subject Sample #1	0.080	04:25
Breath Volume	3.792	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample #2	0.076	04:28
Breath Volume	2.917	04:28
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 05/16/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROWN, DAVE, D

Breath Test Operator Agency: Polk County SO

Arresting Officer: RIVERA, HUMBERTO

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 21:08

Gender: Male

Observation Period Began: 21:45

Subject's Last Name: ROLON

Date/Time of Test: 04/27/2007 22:27

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.079	22:17
Air Blank	0.000	22:17
Subject Sample #1	0.175	22:19
Breath Volume	1.632	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample #2	0.207	22:22
Breath Volume	2.273	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample #3	0.197	22:25
Breath Volume	2.019	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 05/16/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: MAGYAROSI, TODD, A

Breath Test Operator Agency: Polk County SO

Arresting Officer: MAGYAROSI, TODD

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 02:24

Gender: Male

Observation Period Began: 02:57

Subject's Last Name: HATCHER

Date/Time of Test: 04/29/2007 03:24

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample #1	0.172	03:20
Breath Volume	2.316	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample #2	0.176	03:23
Breath Volume	3.648	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 05/16/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001080

Instrument Registered To: POLK COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROWN, DAVE, D

Breath Test Operator Agency: Polk County SO

Arresting Officer: RUSHING, WILLIAM

Arresting Officer Agency: Polk County SO

Last Agency Inspection: 04/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 01:34

Gender: Female

Observation Period Began: 01:55

Subject's Last Name: MILLER

Date/Time of Test: 04/30/2007 02:22

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Subject Sample #1	0.241	02:18
Breath Volume	2.398	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample #2	0.242	02:21
Breath Volume	2.492	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 05/16/2007 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001082

Instrument Registered To: PLANTATION PD

Software Version: 8100.27

Breath Test Operator: FOX, OFC T,

Breath Test Operator Agency: Plantation PD

Arresting Officer: MASON, OFC K

Arresting Officer Agency: Plantation PD

Last Agency Inspection: 03/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 23:29

Gender: Male

Observation Period Began: 23:45

Subject's Last Name: ABREGO

Date/Time of Test: 04/01/2007 00:38

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:32
Air Blank	0.000	00:32
Subject Sample #1	0.134	00:34
Breath Volume	1.929	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample #2	0.115	00:37
Breath Volume	1.308	00:37
Air Blank	0.000	00:37
Control Test	0.078	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 04/18/2007 19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001082

Instrument Registered To: PLANTATION PD

Software Version: 8100.27

Breath Test Operator: DORFMAN, A,

Breath Test Operator Agency: Plantation PD

Arresting Officer: DIAZ, J

Arresting Officer Agency: Plantation PD

Last Agency Inspection: 03/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 03:51

Gender: Male

Observation Period Began: 04:15

Subject's Last Name: HUNTER

Date/Time of Test: 04/01/2007 04:49

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:41
Subject Sample #1	0.113	04:43
Breath Volume	1.777	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample #2	0.132	04:47
Breath Volume	1.347	04:47
Air Blank	0.000	04:48
Control Test	0.078	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 04/18/2007 19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001082

Instrument Registered To: PLANTATION PD

Software Version: 8100.27

Breath Test Operator: FOX, OFC T,

Breath Test Operator Agency: Plantation PD

Arresting Officer: HIGHFIELD, OFC S

Arresting Officer Agency: Plantation PD

Last Agency Inspection: 03/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 05:04

Gender: Male

Observation Period Began: 05:45

Subject's Last Name: BAKER

Date/Time of Test: 04/09/2007 06:27

Diagnostics Check	OK	06:19
Air Blank	0.000	06:20
Control Test	0.079	06:20
Air Blank	0.000	06:20
Subject Sample #1	0.222	06:22
Breath Volume	2.699	06:22
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample #2	0.232	06:25
Breath Volume	3.667	06:25
Air Blank	0.000	06:26
Control Test	0.077	06:26
Air Blank	0.000	06:26
Diagnostics Check	OK	06:27

Data Download Date/Time: 04/18/2007 19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001082

Instrument Registered To: PLANTATION PD

Software Version: 8100.27

Breath Test Operator: DORFMAN, A,

Breath Test Operator Agency: Plantation PD

Arresting Officer: MASON, K

Arresting Officer Agency: Plantation PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 02:42

Gender: Male

Observation Period Began: 02:50

Subject's Last Name: LEON

Date/Time of Test: 04/27/2007 03:18

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:12
Air Blank	0.000	03:12
Subject Sample #1	0.024	03:13
Breath Volume	1.367	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.025	03:16
Breath Volume	2.527	03:16
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:18

Data Download Date/Time: 05/17/2007 20:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001082

Instrument Registered To: PLANTATION PD

Software Version: 8100.27

Breath Test Operator: FOX, OFC T,

Breath Test Operator Agency: Plantation PD

Arresting Officer: MASON, OFC K

Arresting Officer Agency: Plantation PD

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6272011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 00:04

Gender: Male

Observation Period Began: 00:30

Subject's Last Name: JACOBSON

Date/Time of Test: 04/30/2007 01:10

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.078	01:04
Air Blank	0.000	01:05
Subject Sample #1	0.000	01:06
Breath Volume	2.656	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample #2	0.000	01:09
Breath Volume	3.285	01:09
Air Blank	0.000	01:09
Control Test	0.078	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 05/17/2007 20:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001109

Instrument Registered To: FL HIGHWAY PATROL

Software Version: 8100.27

Breath Test Operator: JOHNSON, J,

Breath Test Operator Agency: FHP Troop C

Arresting Officer: HOLLIS, J

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 4314011

Dry Gas Expiry Date: 11/11/2006

Violation Code: DUI

Arrest Time: 21:25

Gender: Male

Observation Period Began: 22:15

Subject's Last Name: KARALIS

Date/Time of Test: 04/16/2007 22:43

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Subject Sample #1	0.144	22:38
Breath Volume	2.050	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample #2	0.151	22:42
Breath Volume	2.191	22:42
Air Blank	0.000	22:42
Control Test	0.080	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 05/11/2007 16:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001109

Instrument Registered To: FL HIGHWAY PATROL

Software Version: 8100.27

Breath Test Operator: JOHNSON, JEFFREY, L

Breath Test Operator Agency: FHP Troop C

Arresting Officer: HOLLIS, JACQUELYN

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 630702I

Dry Gas Expiry Date: 11/06/2008

Violation Code: DUI

Arrest Time: 18:08

Gender: Male

Observation Period Began: 19:15

Subject's Last Name: KOVALESKI

Date/Time of Test: 04/22/2007 19:47

Diagnostics Check	OK	19:37
Air Blank	0.000	19:38
Control Test	0.079	19:38
Air Blank	0.000	19:39
Subject Sample #1	0.232	19:39
Breath Volume	3.519	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample #2	0.199	19:43
Breath Volume	2.144	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample #3	0.194	19:46
Breath Volume	1.347	19:46
Air Blank	0.000	19:46
Control Test	0.078	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 05/11/2007 16:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001112

Instrument Registered To: FL HIGHWAY PATROL

Software Version: 8100.27

Breath Test Operator: LEROY, STODDARD, B

Breath Test Operator Agency: FHP Troop D

Arresting Officer: HOWES, RICHARD

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 04/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 515902I

Dry Gas Expiry Date: 06/13/2007

Violation Code: DUI

Arrest Time: 00:12

Gender: Male

Observation Period Began: 00:12

Subject's Last Name: BURD

Date/Time of Test: 04/10/2007 00:59

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:54
Air Blank	0.000	00:54
Subject Sample #1	0.183	00:55
Breath Volume	2.144	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample #2	0.182	00:58
Breath Volume	2.464	00:58
Air Blank	0.000	00:58
Control Test	0.079	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 05/14/2007 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001112

Instrument Registered To: FL HIGHWAY PATROL

Software Version: 8100.27

Breath Test Operator: CARPENTER, DARWIN, R

Breath Test Operator Agency: FHP Troop D

Arresting Officer: HOWES, RICK

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 04/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 515902I

Dry Gas Expiry Date: 06/13/2007

Violation Code: DUI

Arrest Time: 11:15

Gender: Male

Observation Period Began: 12:15

Subject's Last Name: CARPENTER

Date/Time of Test: 04/20/2007 12:40

Diagnostics Check	OK	12:36	
Air Blank	0.000	12:36	
Control Test	0.079	12:37	
Air Blank	0.000	12:37	
Subject Sample #1		12:40	Sequence Aborted
Breath Volume	0.066	12:40	
Air Blank	0.000	12:40	

Data Download Date/Time: 05/14/2007 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001112

Instrument Registered To: FL HIGHWAY PATROL

Software Version: 8100.27

Breath Test Operator: DEWITT, THOMAS, R

Breath Test Operator Agency: FHP Troop D

Arresting Officer: BARROW, ERIC

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 04/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 515902I

Dry Gas Expiry Date: 06/13/2007

Violation Code: DUI

Arrest Time: 03:45

Gender: Male

Observation Period Began: 04:09

Subject's Last Name: MEYERS JR

Date/Time of Test: 04/21/2007 04:35

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:30
Subject Sample #1	0.124	04:31
Breath Volume	2.179	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample #2	0.119	04:34
Breath Volume	3.203	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 05/14/2007 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001112

Instrument Registered To: FL HIGHWAY PATROL

Software Version: 8100.27

Breath Test Operator: LEROY, STODDARD, B

Breath Test Operator Agency: FHP Troop D

Arresting Officer: BARROW, ERIC

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 04/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 515902I

Dry Gas Expiry Date: 06/13/2007

Violation Code: DUI

Arrest Time: 23:05

Gender: Male

Observation Period Began: 23:05

Subject's Last Name: PLAISANCE

Date/Time of Test: 04/22/2007 23:36

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:31
Subject Sample #1	0.139	23:31
Breath Volume	1.425	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample #2	0.136	23:34
Breath Volume	1.183	23:34
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:36

Data Download Date/Time: 05/14/2007 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001112

Instrument Registered To: FL HIGHWAY PATROL

Software Version: 8100.27

Breath Test Operator: LEROY, STODDARD, B

Breath Test Operator Agency: FHP Troop D

Arresting Officer: LOPEZ, SANTIAGO

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 04/05/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 515902I

Dry Gas Expiry Date: 06/13/2007

Violation Code: DUI

Arrest Time: 23:12

Gender: Male

Observation Period Began: 23:15

Subject's Last Name: HUDSON

Date/Time of Test: 04/22/2007 23:52

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:47
Air Blank	0.000	23:47
Subject Sample #1	0.210	23:48
Breath Volume	2.929	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample #2	0.213	23:51
Breath Volume	2.289	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 05/14/2007 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001116

Instrument Registered To: EGLIN AFB POLICE

Software Version: 8100.27

Breath Test Operator: WATSON, HEATHER, L

Breath Test Operator Agency: Eglin AFB PD

Arresting Officer: BURKEY, MATTHEW

Arresting Officer Agency: Eglin AFB PD

Last Agency Inspection: 04/09/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 21:00

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: BALLARD

Date/Time of Test: 04/19/2007 02:34

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample #1	0.030	02:30
Breath Volume	1.808	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample #2	0.030	02:33
Breath Volume	1.667	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 05/31/2007 04:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001116

Instrument Registered To: EGLIN AFB POLICE

Software Version: 8100.27

Breath Test Operator: WATSON, HEATHER, L

Breath Test Operator Agency: Eglin AFB PD

Arresting Officer: QUALLE, JOSHUA

Arresting Officer Agency: Eglin AFB PD

Last Agency Inspection: 04/09/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 23:00

Gender: Female

Observation Period Began: 23:20

Subject's Last Name: CAVE

Date/Time of Test: 04/25/2007 23:47

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Subject Sample #1	0.169	23:43
Breath Volume	1.886	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample #2	0.173	23:46
Breath Volume	2.812	23:46
Air Blank	0.000	23:46
Control Test	0.077	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 05/31/2007 04:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001116

Instrument Registered To: EGLIN AFB POLICE

Software Version: 8100.27

Breath Test Operator: WATSON, HEATHER, L

Breath Test Operator Agency: Eglin AFB PD

Arresting Officer: LAMB, BRIAN

Arresting Officer Agency: Eglin AFB PD

Last Agency Inspection: 04/09/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 04:15

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: VENSEL

Date/Time of Test: 04/28/2007 04:48

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.078	04:43
Air Blank	0.000	04:43
Subject Sample #1	0.142	04:44
Breath Volume	2.527	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample #2	0.135	04:47
Breath Volume	1.812	04:47
Air Blank	0.000	04:47
Control Test	0.078	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 05/31/2007 04:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, DONALD, E

Breath Test Operator Agency: Pasco County SO

Arresting Officer: SHAW, DONALD

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 02:34

Gender: Male

Observation Period Began: 03:42

Subject's Last Name: SEQUEN

Date/Time of Test: 04/01/2007 04:09

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Subject Sample #1	0.141	04:04
Breath Volume	1.402	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.137	04:07
Breath Volume	1.746	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, DONALD, E

Breath Test Operator Agency: Pasco County SO

Arresting Officer: SHAW, DONALD

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 03:30

Gender: Female

Observation Period Began: 03:50

Subject's Last Name: MASLYN

Date/Time of Test: 04/01/2007 04:18

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Subject Sample #1	0.139	04:14
Breath Volume	2.488	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample #2	0.137	04:17
Breath Volume	2.429	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: GARDNER, ALLAN, W

Breath Test Operator Agency: Pasco County SO

Arresting Officer: DARNAL, MILES

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 06:06

Gender: Male

Observation Period Began: 07:05

Subject's Last Name: MURILLO

Date/Time of Test: 04/01/2007 07:37

Diagnostics Check	OK	07:31
Air Blank	0.000	07:31
Control Test	0.080	07:31
Air Blank	0.000	07:32
Subject Sample #1	0.121	07:32
Breath Volume	2.527	07:32
Air Blank	0.000	07:33
Air Blank	0.000	07:35
Subject Sample #2	0.120	07:35
Breath Volume	2.300	07:35
Air Blank	0.000	07:36
Control Test	0.079	07:36
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: JOHNSON, J,

Breath Test Operator Agency: FHP Troop C

Arresting Officer: JOHNSON, J

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 14:50

Gender: Male

Observation Period Began: 17:10

Subject's Last Name: WILEY

Date/Time of Test: 04/01/2007 17:33

Diagnostics Check	OK	17:30	
Air Blank	0.000	17:31	
Control Test	0.081	17:31	
Air Blank	0.000	17:32	
Subject Sample #1		17:32	Sequence Aborted
Breath Volume	0.000	17:32	
Air Blank	0.000	17:33	

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: JOHNSON, J,

Breath Test Operator Agency: FHP Troop C

Arresting Officer: HOLLIS, J

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 18:10

Gender: Male

Observation Period Began: 19:25

Subject's Last Name: WHITTLER

Date/Time of Test: 04/01/2007 19:48

Diagnostics Check	OK	19:46	
Air Blank	0.000	19:47	
Control Test	0.080	19:47	
Air Blank	0.000	19:48	
Subject Sample #1		19:48	Sequence Aborted
Breath Volume	0.000	19:48	
Air Blank	0.000	19:48	

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: JONES, MICHAEL, S

Breath Test Operator Agency: Pasco County SO

Arresting Officer: MALLOY, J

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 02:57

Gender: Female

Observation Period Began: 02:57

Subject's Last Name: HERNANDEZ

Date/Time of Test: 04/02/2007 03:52

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample #1	0.000	03:47
Breath Volume	2.531	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample #2	0.000	03:50
Breath Volume	2.089	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: RAYMER, SCOTT, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: MALLOY, JOSHUA

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/13/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 00:18

Gender: Male

Observation Period Began: 00:18

Subject's Last Name: THERIAULT

Date/Time of Test: 04/03/2007 00:59

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:53
Subject Sample #1	0.104	00:54
Breath Volume	3.910	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample #2	0.105	00:57
Breath Volume	4.128	00:57
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO
Software Version: 8100.27
Breath Test Operator: KIRKPATRICK, J,
Breath Test Operator Agency: New Port Richey PD
Arresting Officer: JAY, E
Arresting Officer Agency: New Port Richey PD
Last Agency Inspection: 03/13/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 6172021
Dry Gas Expiry Date: 06/23/2008
Violation Code: DUI
Arrest Time: 02:56
Gender: Male
Observation Period Began: 03:15

Subject's Last Name: BOHART

Date/Time of Test: 04/06/2007 03:44

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Subject Sample #1	0.197	03:39
Breath Volume	1.640	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample #2	0.193	03:43
Breath Volume	1.613	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: WINTERS, LANE, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: SCHMUTZ, DAVID

Arresting Officer Agency: FHP Troop K

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 02:10

Gender: Male

Observation Period Began: 03:10

Subject's Last Name: STANLEY

Date/Time of Test: 04/07/2007 03:45

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Subject Sample #1	0.102	03:40
Breath Volume	1.445	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample #2	0.100	03:44
Breath Volume	1.980	03:44
Air Blank	0.000	03:44
Control Test	0.083	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BALDWIN, ROBERT, K

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BALDWIN, ROBERT

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 23:24

Gender: Male

Observation Period Began: 00:15

Subject's Last Name: REITEMEIER

Date/Time of Test: 04/08/2007 00:43

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:38
Air Blank	0.000	00:38
Subject Sample #1	0.146	00:39
Breath Volume	2.613	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample #2	0.140	00:42
Breath Volume	2.652	00:42
Air Blank	0.000	00:43
Control Test	0.081	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, DONALD, E

Breath Test Operator Agency: Pasco County SO

Arresting Officer: SHAW, DONALD

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:27

Gender: Male

Observation Period Began: 01:31

Subject's Last Name: WISE

Date/Time of Test: 04/08/2007 02:00

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Subject Sample #1	0.183	01:56
Breath Volume	2.109	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:58
Subject Sample #2	0.173	01:59
Breath Volume	1.738	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BALDWIN, ROBERT, K

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BALDWIN, ROBERT

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:49

Gender: Female

Observation Period Began: 02:15

Subject's Last Name: NORLIE

Date/Time of Test: 04/08/2007 02:42

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Subject Sample #1	0.141	02:37
Breath Volume	2.226	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.132	02:40
Breath Volume	1.738	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BELL, C,

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BELL, C

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 02:44

Gender: Male

Observation Period Began: 03:10

Subject's Last Name: TOBER

Date/Time of Test: 04/08/2007 03:37

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample #1	0.179	03:33
Breath Volume	1.910	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.191	03:35
Breath Volume	3.312	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: JOHNSON, J,

Breath Test Operator Agency: FHP Troop C

Arresting Officer: JOHNSON, J

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 16:00

Gender: Female

Observation Period Began: 17:10

Subject's Last Name: FORGEY

Date/Time of Test: 04/10/2007 17:37

Diagnostics Check	OK	17:31
Air Blank	0.000	17:31
Control Test	0.079	17:32
Air Blank	0.000	17:32
Subject Sample #1	0.181	17:32
Breath Volume	1.367	17:32
Air Blank	0.000	17:33
Air Blank	0.000	17:35
Subject Sample #2	0.177	17:35
Breath Volume	2.015	17:35
Air Blank	0.000	17:36
Control Test	0.080	17:36
Air Blank	0.000	17:37
Diagnostics Check	OK	17:37

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: ROQUE, DAVID, S

Breath Test Operator Agency: Pasco County SO

Arresting Officer: METZLER, SHANE

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:07

Gender: Male

Observation Period Began: 01:29

Subject's Last Name: FISHER

Date/Time of Test: 04/11/2007 01:58

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:53
Air Blank	0.000	01:53
Subject Sample #1	0.143	01:54
Breath Volume	3.000	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.133	01:57
Breath Volume	1.515	01:57
Air Blank	0.000	01:57
Control Test	0.081	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: CARMEN, G, F

Breath Test Operator Agency: Pasco County SO

Arresting Officer: CARMEN, G

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 19:38

Gender: Male

Observation Period Began: 19:50

Subject's Last Name: PERKINS

Date/Time of Test: 04/11/2007 20:20

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.080	20:14
Air Blank	0.000	20:15
Subject Sample #1	0.186	20:16
Breath Volume	3.175	20:16
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample #2	0.189	20:19
Breath Volume	2.945	20:19
Air Blank	0.000	20:19
Control Test	0.080	20:20
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, DONALD, E

Breath Test Operator Agency: Pasco County SO

Arresting Officer: SHAW, DONALD

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 21:36

Gender: Male

Observation Period Began: 21:53

Subject's Last Name: MONTES

Date/Time of Test: 04/12/2007 22:21

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.080	22:16
Air Blank	0.000	22:16
Subject Sample #1	0.171	22:17
Breath Volume	1.535	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample #2	0.170	22:19
Breath Volume	2.027	22:19
Air Blank	0.000	22:20
Control Test	0.080	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, DONALD, E

Breath Test Operator Agency: Pasco County SO

Arresting Officer: SHAW, DONALD

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:01

Gender: Male

Observation Period Began: 01:33

Subject's Last Name: JACOBSON

Date/Time of Test: 04/13/2007 01:59

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.081	01:55	
Air Blank	0.000	01:55	
Subject Sample #1	0.159	01:57	
Breath Volume	1.316	01:57	
Air Blank		01:58	Purge Fail
Air Blank	0.000	01:59	

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, DONALD, E

Breath Test Operator Agency: Pasco County SO

Arresting Officer: SHAW, DONALD

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:01

Gender: Male

Observation Period Began: 01:33

Subject's Last Name: JACOBSON

Date/Time of Test: 04/13/2007 02:09

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample #1	0.162	02:03
Breath Volume	2.183	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample #2	0.162	02:08
Breath Volume	1.765	02:08
Air Blank	0.000	02:09
Control Test	0.084	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: MORRIS, ADAM, M

Breath Test Operator Agency: FHP Troop C

Arresting Officer: MORRIS, ADAM

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 02:57

Gender: Male

Observation Period Began: 02:57

Subject's Last Name: STRAUBEL

Date/Time of Test: 04/13/2007 03:55

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:50
Subject Sample #1	0.164	03:51
Breath Volume	1.628	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample #2	0.177	03:54
Breath Volume	2.007	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, DONALD, E

Breath Test Operator Agency: Pasco County SO

Arresting Officer: SHAW, DONALD

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 21:22

Gender: Male

Observation Period Began: 21:37

Subject's Last Name: ORTIZ

Date/Time of Test: 04/13/2007 22:06

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
Control Test	0.081	22:02	
Air Blank	0.000	22:03	
Subject Sample #1		22:05	Subject Test Refused
Breath Volume	0.000	22:05	
Air Blank	0.000	22:05	
Control Test	0.080	22:05	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: MORRIS, ADAM, M

Breath Test Operator Agency: FHP Troop C

Arresting Officer: KRAUS, D

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 00:11

Gender: Male

Observation Period Began: 00:11

Subject's Last Name: WILSON

Date/Time of Test: 04/14/2007 01:11

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.080	01:05	
Air Blank	0.000	01:06	
Subject Sample #1	0.197	01:07	
Breath Volume	2.144	01:07	
Air Blank	0.000	01:08	
Air Blank	0.000	01:09	
Subject Sample #2		01:10	RFI Detect
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Control Test	0.079	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: MORRIS, ADAM, M

Breath Test Operator Agency: FHP Troop C

Arresting Officer: KRAUS, D

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 00:11

Gender: Male

Observation Period Began: 00:11

Subject's Last Name: WILSON

Date/Time of Test: 04/14/2007 01:19

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:13
Subject Sample #1	0.196	01:14
Breath Volume	2.511	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:16
Subject Sample #2	0.207	01:17
Breath Volume	2.007	01:17
Air Blank	0.000	01:18
Control Test	0.081	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: MORRIS, ADAM, M

Breath Test Operator Agency: FHP Troop C

Arresting Officer: MORRIS, ADAM

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 02:17

Gender: Female

Observation Period Began: 02:17

Subject's Last Name: DIDOMENICO

Date/Time of Test: 04/14/2007 02:59

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample #1	0.139	02:55
Breath Volume	2.667	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.139	02:58
Breath Volume	2.074	02:58
Air Blank	0.000	02:58
Control Test	0.081	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: TOCZYLOWSKI, MICHAEL, A

Breath Test Operator Agency: Pasco County SO

Arresting Officer: TOCZYLOWSKI, MICHAEL

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 02:30

Gender: Female

Observation Period Began: 02:45

Subject's Last Name: MEDLOCK

Date/Time of Test: 04/14/2007 03:13

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample #1	0.124	03:08
Breath Volume	1.675	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample #2	0.130	03:11
Breath Volume	2.070	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:13

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BELL, C,

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BELL, C

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 03/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 03:29

Gender: Male

Observation Period Began: 03:50

Subject's Last Name: BROOKS

Date/Time of Test: 04/14/2007 04:25

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:20
Subject Sample #1	0.177	04:21
Breath Volume	2.097	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:23
Subject Sample #2	0.176	04:24
Breath Volume	1.988	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 04/14/2007 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: CARMEN, G, F

Breath Test Operator Agency: Pasco County SO

Arresting Officer: CARMEN, G

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 04:46

Gender: Male

Observation Period Began: 04:48

Subject's Last Name: COTHRON

Date/Time of Test: 04/15/2007 05:21

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.079	05:12
Air Blank	0.000	05:12
Subject Sample #1	0.145	05:13
Breath Volume	3.351	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample #2	0.120	05:16
Breath Volume	1.578	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample #3	0.138	05:19
Breath Volume	4.199	05:19
Air Blank	0.000	05:20
Control Test	0.081	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: ORNDORFF, WILLIAM, J

Breath Test Operator Agency: Pasco County SO

Arresting Officer: ORNDORFF, WILLIAM

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 22:27

Gender: Male

Observation Period Began: 22:27

Subject's Last Name: HENDERSON

Date/Time of Test: 04/17/2007 23:16

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:11
Subject Sample #1	0.269	23:12
Breath Volume	3.281	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:14
Subject Sample #2	0.279	23:15
Breath Volume	3.050	23:15
Air Blank	0.000	23:15
Control Test	0.082	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BOOKER, ERIK, V

Breath Test Operator Agency: Pasco County SO

Arresting Officer: CARMEN, G

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 03:59

Gender: Male

Observation Period Began: 04:14

Subject's Last Name: BOOKER

Date/Time of Test: 04/20/2007 04:41

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample #1	0.193	04:37
Breath Volume	2.324	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample #2	0.202	04:40
Breath Volume	2.593	04:40
Air Blank	0.000	04:40
Control Test	0.079	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BELL, C,

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BELL, C

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 03:20

Gender: Male

Observation Period Began: 04:00

Subject's Last Name: CHIARULLI

Date/Time of Test: 04/21/2007 04:27

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Subject Sample #1	0.000	04:22
Breath Volume	2.562	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.000	04:25
Breath Volume	2.117	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BELL, C,

Breath Test Operator Agency: Pasco County SO

Arresting Officer: GOSWICK, T

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 02:48

Gender: Male

Observation Period Began: 03:30

Subject's Last Name: PELEROSI

Date/Time of Test: 04/21/2007 05:25

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:19
Subject Sample #1	0.102	05:19
Breath Volume	1.617	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample #2	0.112	05:23
Breath Volume	2.046	05:23
Air Blank	0.000	05:24
Control Test	0.082	05:24
Air Blank	0.000	05:24
Diagnostics Check	OK	05:25

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BELL, C,

Breath Test Operator Agency: Pasco County SO

Arresting Officer: GOSWICK, T

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 03:31

Gender: Male

Observation Period Began: 04:10

Subject's Last Name: BILECZ

Date/Time of Test: 04/22/2007 04:40

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample #1	0.117	04:36
Breath Volume	1.625	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample #2	0.119	04:39
Breath Volume	1.582	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BROOKS, WILLIAM, F

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BORIA, JASON

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 15:14

Gender: Male

Observation Period Began: 15:50

Subject's Last Name: VELEZ

Date/Time of Test: 04/23/2007 16:25

Diagnostics Check	OK	16:17	
Air Blank	0.000	16:17	
Control Test	0.080	16:18	
Air Blank	0.000	16:18	
Subject Sample #1	0.243	16:21	Volume Not Met
Breath Volume	0.191	16:21	
Air Blank	0.000	16:22	
Air Blank	0.000	16:24	
Subject Sample #2		16:24	Sequence Aborted
Breath Volume	0.152	16:24	
Air Blank	0.000	16:25	

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: MORRIS, ADAM, M

Breath Test Operator Agency: FHP Troop C

Arresting Officer: MORRIS, ADAM

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:13

Gender: Male

Observation Period Began: 01:13

Subject's Last Name: BEALS

Date/Time of Test: 04/25/2007 02:14

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample #1	0.164	02:10
Breath Volume	4.589	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample #2	0.160	02:13
Breath Volume	4.097	02:13
Air Blank	0.000	02:14
Control Test	0.082	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BELL, C,

Breath Test Operator Agency: Pasco County SO

Arresting Officer: BELL, C

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 19:27

Gender: Male

Observation Period Began: 19:45

Subject's Last Name: ROBERTSON

Date/Time of Test: 04/26/2007 20:10

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:08	
Control Test	0.079	20:08	
Air Blank	0.000	20:08	
Subject Sample #1		20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:09	
Control Test	0.080	20:09	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: FODEN, WILLIAM, R

Breath Test Operator Agency: FHP Troop C

Arresting Officer: FODEN, WILLIAM

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 23:04

Gender: Male

Observation Period Began: 00:00

Subject's Last Name: BERRY

Date/Time of Test: 04/27/2007 00:30

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:25
Air Blank	0.000	00:25
Subject Sample #1	0.166	00:26
Breath Volume	1.808	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample #2	0.169	00:29
Breath Volume	1.867	00:29
Air Blank	0.000	00:29
Control Test	0.082	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: CARMEN, G, F

Breath Test Operator Agency: Pasco County SO

Arresting Officer: CARMEN, G

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 01:08

Gender: Male

Observation Period Began: 01:35

Subject's Last Name: DUBOSE

Date/Time of Test: 04/27/2007 02:03

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:58
Air Blank	0.000	01:58
Subject Sample #1	0.193	01:58
Breath Volume	2.695	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample #2	0.195	02:01
Breath Volume	2.730	02:01
Air Blank	0.000	02:02
Control Test	0.081	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: FODEN, WILLIAM, R

Breath Test Operator Agency: FHP Troop C

Arresting Officer: FODEN, WILLIAM

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 02:23

Gender: Female

Observation Period Began: 03:40

Subject's Last Name: MARSEE

Date/Time of Test: 04/28/2007 04:09

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:06	
Control Test	0.080	04:07	
Air Blank	0.000	04:07	
Subject Sample #1		04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Control Test	0.080	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: RIFFE, DAVID, J

Breath Test Operator Agency: Pasco County SO

Arresting Officer: RIFFE, DAVID

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 617202I

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 04:11

Gender: Male

Observation Period Began: 04:50

Subject's Last Name: PATTERSON

Date/Time of Test: 04/28/2007 05:25

Diagnostics Check	OK	05:18
Air Blank	0.000	05:19
Control Test	0.080	05:19
Air Blank	0.000	05:20
Subject Sample #1	0.132	05:20
Breath Volume	1.445	05:20
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample #2	0.127	05:24
Breath Volume	1.574	05:24
Air Blank	0.000	05:24
Control Test	0.081	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: BALDWIN, ROBERT, K

Breath Test Operator Agency: Pasco County SO

Arresting Officer: METZLER, SHANE

Arresting Officer Agency: Pasco County SO

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 22:20

Gender: Female

Observation Period Began: 23:35

Subject's Last Name: FLECTHER

Date/Time of Test: 04/29/2007 00:02

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:56
Subject Sample #1	0.000	23:57
Breath Volume	1.324	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample #2	0.000	00:00
Breath Volume	1.632	00:00
Air Blank	0.000	00:01
Control Test	0.082	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:02

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001117

Instrument Registered To: PASCO COUNTY SO

Software Version: 8100.27

Breath Test Operator: FODEN, WILLIAM, R

Breath Test Operator Agency: FHP Troop C

Arresting Officer: FODEN, WILLIAM

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 04/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6172021

Dry Gas Expiry Date: 06/23/2008

Violation Code: DUI

Arrest Time: 23:44

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: SPERRY

Date/Time of Test: 04/30/2007 02:09

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.080	02:07	
Air Blank	0.000	02:08	
Subject Sample #1		02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:09	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 05/16/2007 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001119

Instrument Registered To: FHP TROOP D - DELAND

Software Version: 8100.27

Breath Test Operator: DYCKMAN, JAMES, F

Breath Test Operator Agency: FHP Troop D

Arresting Officer: DELEON, LUIS

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 19:00

Gender: Male

Observation Period Began: 20:45

Subject's Last Name: DYCKMAN

Date/Time of Test: 04/19/2007 21:32

Diagnostics Check	OK	21:23	
Air Blank	0.000	21:24	
Control Test	0.081	21:24	
Air Blank	0.000	21:24	
Subject Sample #1	0.174	21:27	
Breath Volume	1.562	21:27	
Air Blank	0.000	21:28	
Air Blank	0.000	21:29	
Subject Sample #2	0.168	21:30	
Breath Volume	1.644	21:30	
Air Blank		21:31	RFI Detect
Air Blank	0.000	21:32	

Data Download Date/Time: 05/25/2007 18:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001119

Instrument Registered To: FHP TROOP D - DELAND

Software Version: 8100.27

Breath Test Operator: TORRES, RAFAEL,

Breath Test Operator Agency: FHP Troop D

Arresting Officer: DELEON, LUIS

Arresting Officer Agency: FHP Troop D

Last Agency Inspection: 04/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610301G

Dry Gas Expiry Date: 04/18/2008

Violation Code: DUI

Arrest Time: 19:00

Gender: Male

Observation Period Began: 20:45

Subject's Last Name: DYCKMAN

Date/Time of Test: 04/19/2007 21:43

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample #1	0.170	21:39
Breath Volume	1.890	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:41
Subject Sample #2	0.166	21:42
Breath Volume	1.894	21:42
Air Blank	0.000	21:42
Control Test	0.079	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 05/25/2007 18:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001122

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: CAREY, VINCENT, E

Breath Test Operator Agency: FHP Troop K

Arresting Officer: CAREY, VINCENT

Arresting Officer Agency: FHP Troop K

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 22:16

Gender: Male

Observation Period Began: 23:00

Subject's Last Name: RIOS MARTINEZ

Date/Time of Test: 04/12/2007 23:23

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:22	
Subject Sample #1		23:22	Sequence Aborted
Breath Volume	0.000	23:22	
Air Blank	0.000	23:23	

Data Download Date/Time: 04/30/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001122

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: CAREY, VINCENT, E

Breath Test Operator Agency: FHP Troop K

Arresting Officer: CAREY, VINCENT

Arresting Officer Agency: FHP Troop K

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 22:16

Gender: Male

Observation Period Began: 23:00

Subject's Last Name: RIOS MARTINEZ

Date/Time of Test: 04/12/2007 23:36

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
		23:36	RFI Detect
Air Blank	0.000	23:36	

Data Download Date/Time: 04/30/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001122

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: CAREY, VINCENT, E

Breath Test Operator Agency: FHP Troop K

Arresting Officer: CAREY, VINCENT

Arresting Officer Agency: FHP Troop K

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 22:16

Gender: Male

Observation Period Began: 23:00

Subject's Last Name: RIOS MARTINEZ

Date/Time of Test: 04/12/2007 23:42

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.080	23:40	
Air Blank	0.000	23:41	
Subject Sample #1		23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Control Test	0.080	23:42	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 04/30/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001122

Instrument Registered To: FHP TROOP E MIAMI

Software Version: 8100.27

Breath Test Operator: BAZILE, AVENS, E

Breath Test Operator Agency: FHP Troop K

Arresting Officer: BAZILE, AVENS

Arresting Officer Agency: FHP Troop K

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6271011

Dry Gas Expiry Date: 10/03/2008

Violation Code: DUI

Arrest Time: 11:30

Gender: Male

Observation Period Began: 11:25

Subject's Last Name: CHAVEZ

Date/Time of Test: 04/30/2007 12:17

Diagnostics Check	OK	12:08
Air Blank	0.000	12:08
Control Test	0.081	12:08
Air Blank	0.000	12:09
Subject Sample #1	0.277	12:12
Breath Volume	2.753	12:12
Air Blank	0.000	12:12
Air Blank	0.000	12:14
Subject Sample #2	0.279	12:15
Breath Volume	2.484	12:15
Air Blank	0.000	12:16
Control Test	0.080	12:16
Air Blank	0.000	12:17
Diagnostics Check	OK	12:17

Data Download Date/Time: 04/30/2007 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001123

Instrument Registered To: FHP TROOP F FT MYERS

Software Version: 8100.27

Breath Test Operator: FERRELL, BOOKER, T

Breath Test Operator Agency: FHP Troop F

Arresting Officer: FERRELL, BOOKER

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 17:23

Gender: Female

Observation Period Began: 18:31

Subject's Last Name: ZANDRI

Date/Time of Test: 04/15/2007 19:00

Diagnostics Check	OK	18:55	
Air Blank	0.000	18:55	
Control Test	0.079	18:56	
Air Blank	0.000	18:56	
Subject Sample #1	0.027	18:57	
Breath Volume	1.753	18:57	
Air Blank	0.000	18:57	
Air Blank		18:59	Ambient Fail
Air Blank	0.000	19:00	

Data Download Date/Time: 04/23/2007 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001123

Instrument Registered To: FHP TROOP F FT MYERS

Software Version: 8100.27

Breath Test Operator: FERRELL, BOOKER, T

Breath Test Operator Agency: FHP Troop F

Arresting Officer: FERRELL, BOOKER

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 17:23

Gender: Female

Observation Period Began: 18:31

Subject's Last Name: ZANDRI

Date/Time of Test: 04/15/2007 19:09

Diagnostics Check	OK	19:02
Air Blank	0.000	19:03
Control Test	0.079	19:03
Air Blank	0.000	19:03
Subject Sample #1	0.025	19:04
Breath Volume	1.843	19:04
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample #2	0.022	19:07
Breath Volume	1.617	19:07
Air Blank	0.000	19:08
Control Test	0.078	19:08
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 04/23/2007 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001123

Instrument Registered To: FHP TROOP F FT MYERS

Software Version: 8100.27

Breath Test Operator: FERRELL, BOOKER, T

Breath Test Operator Agency: FHP Troop F

Arresting Officer: EDISON, GREGORY

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 03/29/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 20:25

Gender: Female

Observation Period Began: 21:50

Subject's Last Name: JANSSEN

Date/Time of Test: 04/15/2007 22:23

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:18
Air Blank	0.000	22:18
Subject Sample #1	0.162	22:18
Breath Volume	1.570	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample #2	0.158	22:21
Breath Volume	1.937	22:21
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 04/23/2007 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001123

Instrument Registered To: FHP TROOP F FT MYERS

Software Version: 8100.27

Breath Test Operator: FERRELL, BOOKER, T

Breath Test Operator Agency: FHP Troop F

Arresting Officer: MERRITT, MICHAEL

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 04/22/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 629302I

Dry Gas Expiry Date: 10/23/2008

Violation Code: DUI

Arrest Time: 20:03

Gender: Male

Observation Period Began: 21:10

Subject's Last Name: ZOELLNER

Date/Time of Test: 04/24/2007 21:56

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:51
Subject Sample #1	0.000	21:52
Breath Volume	3.628	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample #2	0.000	21:55
Breath Volume	1.968	21:55
Air Blank	0.000	21:55
Control Test	0.078	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 06/03/2007 07:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001128

Instrument Registered To: PUTNAM COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: VEGA, KELVIN,

Breath Test Operator Agency: FHP Troop G

Arresting Officer: VEGA, KELVIN

Arresting Officer Agency: FHP Troop G

Last Agency Inspection: 03/08/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 604401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 23:15

Gender: Male

Observation Period Began: 23:30

Subject's Last Name: VELASQUEZ

Date/Time of Test: 04/01/2007 00:16

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Subject Sample #1	0.200	00:11
Breath Volume	1.929	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample #2	0.187	00:15
Breath Volume	1.531	00:15
Air Blank	0.000	00:15
Control Test	0.077	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 04/10/2007 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001128

Instrument Registered To: PUTNAM COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: TRABER, KIMBERLY, L

Breath Test Operator Agency: Putnam County SO

Arresting Officer: FRILL, TIM

Arresting Officer Agency: Putnam County SO

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 604401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 18:56

Gender: Male

Observation Period Began: 20:20

Subject's Last Name: SMITH

Date/Time of Test: 04/13/2007 20:59

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:48	
Control Test	0.079	20:48	
Air Blank	0.000	20:49	
Subject Sample #1		20:52	No Sample Provided
Breath Volume	0.000	20:52	
Air Blank	0.000	20:52	
Air Blank	0.000	20:54	
Subject Sample #2		20:57	No Sample Provided
Breath Volume	0.000	20:57	
Air Blank	0.000	20:58	
Control Test	0.080	20:58	
Air Blank	0.000	20:59	
Diagnostics Check	OK	20:59	

Data Download Date/Time: 05/22/2007 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001128

Instrument Registered To: PUTNAM COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: OWENS, TOMMY, J

Breath Test Operator Agency: Putnam County SO

Arresting Officer: RUIZ, PETE

Arresting Officer Agency: Palatka PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 604401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 21:53

Gender: Male

Observation Period Began: 22:20

Subject's Last Name: BACHMAN

Date/Time of Test: 04/16/2007 23:03

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:48	
Control Test	0.079	22:48	
Air Blank	0.000	22:49	
Subject Sample #1	0.223	22:50	
Breath Volume	1.492	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:52	
Subject Sample #2	0.184	22:55	Volume Not Met
Breath Volume	0.058	22:55	
Air Blank	0.000	22:56	
Air Blank	0.000	22:58	
Subject Sample #3	0.169	23:01	Volume Not Met
Breath Volume	0.050	23:01	
Air Blank	0.000	23:02	
Control Test	0.077	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 05/22/2007 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001128

Instrument Registered To: PUTNAM COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: LECLAIR, LAWRENCE, W

Breath Test Operator Agency: Putnam County SO

Arresting Officer: WELBANKE, BELINDA

Arresting Officer Agency: Palatka PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 604401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 23:03

Gender: Female

Observation Period Began: 23:30

Subject's Last Name: CULJAK

Date/Time of Test: 04/18/2007 23:59

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:53
Subject Sample #1	0.123	23:54
Breath Volume	2.164	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample #2	0.124	23:57
Breath Volume	2.574	23:57
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:59

Data Download Date/Time: 05/22/2007 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001128

Instrument Registered To: PUTNAM COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: BYERS, TERRY, S

Breath Test Operator Agency: Putnam County SO

Arresting Officer: BYERS, TERRY

Arresting Officer Agency: Putnam County SO

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 604401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 18:08

Gender: Male

Observation Period Began: 18:08

Subject's Last Name: POLLARD

Date/Time of Test: 04/22/2007 19:18

Diagnostics Check	OK	19:12
Air Blank	0.000	19:13
Control Test	0.080	19:13
Air Blank	0.000	19:13
Subject Sample #1	0.186	19:14
Breath Volume	3.117	19:14
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample #2	0.180	19:17
Breath Volume	2.695	19:17
Air Blank	0.000	19:17
Control Test	0.079	19:18
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 05/22/2007 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001128

Instrument Registered To: PUTNAM COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: JORDAN, DONNY, R

Breath Test Operator Agency: Putnam County SO

Arresting Officer: TRABER, KIM

Arresting Officer Agency: Putnam County SO

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 604401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 22:30

Gender: Male

Observation Period Began: 23:05

Subject's Last Name: NOONAN

Date/Time of Test: 04/25/2007 23:37

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:33	
Subject Sample #1	0.188	23:33	
Breath Volume	1.816	23:33	
Air Blank	0.000	23:34	
Air Blank	0.000	23:36	
Subject Sample #2		23:36	RFI Detect
Breath Volume	0.000	23:36	
Air Blank		23:36	RFI Detect
Air Blank	0.000	23:37	

Data Download Date/Time: 05/22/2007 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001128

Instrument Registered To: PUTNAM COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: JORDAN, DONNY, R

Breath Test Operator Agency: Putnam County SO

Arresting Officer: TRABER, KIM

Arresting Officer Agency: Putnam County SO

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 604401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 22:30

Gender: Male

Observation Period Began: 23:05

Subject's Last Name: NOONAN

Date/Time of Test: 04/25/2007 23:48

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:42
Subject Sample #1	0.181	23:43
Breath Volume	1.304	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample #2	0.179	23:46
Breath Volume	1.476	23:46
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:48

Data Download Date/Time: 05/22/2007 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001128

Instrument Registered To: PUTNAM COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: EASON, BENJAMIN, E

Breath Test Operator Agency: FFWCC (Ocala)

Arresting Officer: MARTIN, JONATHON

Arresting Officer Agency: FFWCC (Ocala)

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 604401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 20:32

Gender: Male

Observation Period Began: 21:00

Subject's Last Name: BENNETT

Date/Time of Test: 04/27/2007 22:09

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:03
Subject Sample #1	0.080	22:04
Breath Volume	1.550	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample #2	0.077	22:08
Breath Volume	1.863	22:08
Air Blank	0.000	22:08
Control Test	0.077	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 05/22/2007 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001128

Instrument Registered To: PUTNAM COUNTY S.O.

Software Version: 8100.27

Breath Test Operator: BYERS, TERRY, S

Breath Test Operator Agency: Putnam County SO

Arresting Officer: BYERS, TERRY

Arresting Officer Agency: Putnam County SO

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 604401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 02:50

Gender: Male

Observation Period Began: 03:10

Subject's Last Name: ADKINS

Date/Time of Test: 04/28/2007 03:55

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Subject Sample #1	0.078	03:50
Breath Volume	1.253	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample #2	0.075	03:53
Breath Volume	2.691	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 05/22/2007 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001130

Instrument Registered To: VOLUSIA S.O.

Software Version: 8100.27

Breath Test Operator: MCGUIRE, SHON, D

Breath Test Operator Agency: Volusia County SO

Arresting Officer: MAYS, LAURALYNN

Arresting Officer Agency: Volusia County SO

Last Agency Inspection: 04/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 5188011

Dry Gas Expiry Date: 07/12/2007

Violation Code: DUI

Arrest Time: 02:08

Gender: Male

Observation Period Began: 02:08

Subject's Last Name: SHOPOVICK

Date/Time of Test: 04/30/2007 02:37

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.079	02:35	
Air Blank	0.000	02:35	
Subject Sample #1		02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 05/15/2007 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001131

Instrument Registered To: VOLUSIA S.O.

Software Version: 8100.27

Breath Test Operator: IHNKEN, MATTHEW, D

Breath Test Operator Agency: Volusia County SO

Arresting Officer: IHNKEN, MATTHEW

Arresting Officer Agency: Volusia County SO

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 08/01/2008

Violation Code: DUI

Arrest Time: 21:50

Gender: Male

Observation Period Began: 22:15

Subject's Last Name: ROSADO

Date/Time of Test: 04/06/2007 22:46

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:40
Subject Sample #1	0.212	22:42
Breath Volume	1.949	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample #2	0.212	22:44
Breath Volume	1.804	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 04/28/2007 19:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001131

Instrument Registered To: VOLUSIA S.O.

Software Version: 8100.27

Breath Test Operator: MCGUIRE, SHON, D

Breath Test Operator Agency: Volusia County SO

Arresting Officer: MCGUIRE, SHON

Arresting Officer Agency: Volusia County SO

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 08/01/2008

Violation Code: DUI

Arrest Time: 03:42

Gender: Male

Observation Period Began: 03:35

Subject's Last Name: MAJTENYI

Date/Time of Test: 04/10/2007 04:12

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:11	
Subject Sample #1		04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Control Test	0.081	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 04/28/2007 19:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001131

Instrument Registered To: VOLUSIA S.O.

Software Version: 8100.27

Breath Test Operator: MCGUIRE, SHON, D

Breath Test Operator Agency: Volusia County SO

Arresting Officer: MCGUIRE, SHON

Arresting Officer Agency: Volusia County SO

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 611801G

Dry Gas Expiry Date: 08/01/2008

Violation Code: DUI

Arrest Time: 21:05

Gender: Male

Observation Period Began: 21:30

Subject's Last Name: ROSADO

Date/Time of Test: 04/19/2007 22:19

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.078	22:11	
Air Blank	0.000	22:12	
Subject Sample #1	0.189	22:15	Volume Not Met
Breath Volume	0.652	22:15	
Air Blank	0.000	22:16	
Air Blank	0.000	22:18	
Subject Sample #2		22:18	Subject Test Refused
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	
Control Test	0.078	22:18	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 04/28/2007 19:32