

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 595761
BREATH TEST OPERATOR: LAMBERT, BRIAN, K	DRY GAS EXPIRY DATE: 04/13/2011
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: ADMIN
ARRESTING OFFICER: LAMBERT, BRIAN	ARREST TIME: 09:00
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/02/2010	OBSERVATION PERIOD BEGAN: 10:00
SUBJECT'S LAST NAME: DOE	DATE/TIME OF TEST: 07/08/2010 11:02:29

---

Diagnostics Check	OK	11:02	
Air Blank		11:02	RFI Detect
Air Blank		11:02	RFI Detect

---

Data Download Date/Time	07/28/2010	16:23
-------------------------	------------	-------

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAMBERT, BRIAN, K

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: RIDAUGHT, GARY

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 595761

DRY GAS EXPIRY DATE: 04/13/2011

VIOLATION CODE: BUI

ARREST TIME: 17:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 07/09/2010 18:09:32

---

Diagnostics Check	OK	18:03
Air Blank	0.000	18:03
Control Test	0.078	18:04
Air Blank	0.000	18:04
Subject Sample #1	0.104	18:05
Breath Volume	2.089	18:05
Air Blank	0.000	18:05
Air Blank	0.000	18:07
Subject Sample #2	0.103	18:08
Breath Volume	1.417	18:08
Air Blank	0.000	18:08
Control Test	0.078	18:08
Air Blank	0.000	18:09
Diagnostics Check	OK	18:09

---

Data Download Date/Time 07/28/2010 16:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAMBERT, BRIAN, K

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: DEARBORN, RICHARD

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 06/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 595761

DRY GAS EXPIRY DATE: 04/13/2011

VIOLATION CODE: BUI

ARREST TIME: 16:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 07/10/2010 17:31:14

---

Diagnostics Check	OK	17:22
Air Blank	0.000	17:23
Control Test	0.079	17:23
Air Blank	0.000	17:24
Subject Sample #1	0.143	17:25
Breath Volume	3.972	17:25
Air Blank	0.000	17:26
Air Blank	0.000	17:28
Subject Sample #2	0.150	17:29
Breath Volume	3.128	17:29
Air Blank	0.000	17:30
Control Test	0.079	17:30
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

---

Data Download Date/Time 07/28/2010 16:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VICKERS, NICK, B  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: FERGUSON, J  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 07/01/2010 20:34:22

---

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.078	20:28
Air Blank	0.000	20:29
Subject Sample #1	0.310	20:29
Breath Volume	1.656	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample #2	0.290	20:32
Breath Volume	1.359	20:32
Air Blank	0.000	20:33
Control Test	0.076	20:33
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

---

Data Download Date/Time 08/04/2010 13:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VICKES, N, B  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: ERVIN, J  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 07/01/2010 22:44:01

---

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:35	
Control Test	0.078	22:36	
Air Blank	0.000	22:36	
Subject Sample #1		22:39	No Sample Provided
Breath Volume	0.000	22:39	
Air Blank	0.000	22:40	
Air Blank	0.000	22:42	
Subject Sample #2		22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.078	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:44	

---

Data Download Date/Time 08/04/2010 13:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VICKERS, NICK, B  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: JOHNSTON, J  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/01/2010 23:03:33

---

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:49	
Control Test	0.078	22:49	
Air Blank	0.000	22:50	
Subject Sample #1	0.233	22:53	Volume Not Met
Breath Volume	1.062	22:53	
Air Blank	0.000	22:53	
Air Blank	0.000	22:55	
Subject Sample #2	0.238	22:56	
Breath Volume	1.457	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample #3	0.232	23:01	
Breath Volume	1.847	23:01	
Air Blank	0.000	23:02	
Control Test	0.076	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

---

Data Download Date/Time 08/04/2010 13:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD, ERIC, L  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: BOYD, ERIC  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 07/03/2010 01:15:09

---

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:13	
Subject Sample #1		01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:14	
Control Test	0.079	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: MARTIN, JAMES  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: RIGG

DATE/TIME OF TEST: 07/05/2010 03:27:06

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample #1	0.127	03:22
Breath Volume	3.660	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample #2	0.126	03:25
Breath Volume	3.167	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHLAGER, JONATHAN, D  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: SCHLAGER, JONATHAN  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 07/05/2010 20:22:18

---

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:11	
Control Test	0.080	20:11	
Air Blank	0.000	20:12	
Subject Sample #1		20:15	No Sample Provided
Breath Volume	0.000	20:15	
Air Blank	0.000	20:15	
Air Blank	0.000	20:17	
Subject Sample #2		20:20	No Sample Provided
Breath Volume	0.000	20:20	
Air Blank	0.000	20:21	
Control Test	0.080	20:21	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD, ERIC, L  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: BOYD, ERIC  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 07/10/2010 03:52:43

---

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.078	03:47
Air Blank	0.000	03:47
Subject Sample #1	0.180	03:48
Breath Volume	2.480	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample #2	0.177	03:51
Breath Volume	2.097	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHEFFIELD, SHELDON, L  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: SHEFFIELD, SHELDON  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: FRAWLEY

DATE/TIME OF TEST: 07/11/2010 00:16:57

---

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.079	00:07	
Air Blank	0.000	00:07	
Subject Sample #1		00:10	No Sample Provided
Breath Volume	0.000	00:10	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample #2		00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

---

Data Download Date/Time 08/31/2010 12:08

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREIRA, JOSE, L  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: PEREIRA, JOSE  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SANTIZO PEREZ

DATE/TIME OF TEST: 07/11/2010 01:43:34

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Subject Sample #1	0.086	01:38
Breath Volume	1.890	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample #2	0.092	01:42
Breath Volume	2.792	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREIRA, JOSE, L  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: PEREIRA, JOSE  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 17:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/11/2010 18:37:05

---

Diagnostics Check	OK	18:34	
Air Blank	0.000	18:34	
Control Test	0.079	18:34	
Air Blank	0.000	18:35	
Subject Sample #1		18:35	Subject Test Refused
Breath Volume	0.000	18:35	
Air Blank	0.000	18:36	
Control Test	0.078	18:36	
Air Blank	0.000	18:37	
Diagnostics Check	OK	18:37	

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUGH, BRYAN, W  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: HOUGH, BRYAN  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: DELAROSA

DATE/TIME OF TEST: 07/14/2010 00:50:18

---

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.080	00:47	
Air Blank	0.000	00:47	
Subject Sample #1		00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUNGER, JOSEPH, W  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: POLSON, BRIAN  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 07/14/2010 02:52:02

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Subject Sample #1	0.210	02:47
Breath Volume	1.937	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample #2	0.218	02:50
Breath Volume	3.675	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:52

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: POLSON, JAMES  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/17/2010 02:29:38

---

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.081	02:26	
Air Blank	0.000	02:26	
Subject Sample #1		02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.079	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHLAGER, JONATHAN, D  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: POLSON, BRIAN  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/18/2010 06:56:49

---

Diagnostics Check	OK	06:51	
Air Blank	0.000	06:51	
Control Test	0.080	06:52	
Air Blank	0.000	06:52	
Subject Sample #1		06:55	Subject Test Refused
Breath Volume	0.000	06:55	
Air Blank	0.000	06:55	
Control Test	0.080	06:56	
Air Blank	0.000	06:56	
Diagnostics Check	OK	06:56	

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD, ERIC, L  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: BOYD, ERIC  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 07/20/2010 22:06:21

---

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:03	
Control Test	0.080	22:04	
Air Blank	0.000	22:04	
Subject Sample #1		22:05	Subject Test Refused
Breath Volume	0.000	22:05	
Air Blank	0.000	22:05	
Control Test	0.080	22:05	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:06	

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD, ERIC, L  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: BOYD, ERIC  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: URBAN

DATE/TIME OF TEST: 07/21/2010 05:28:31

---

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.080	05:25	
Air Blank	0.000	05:26	
Subject Sample #1		05:27	Subject Test Refused
Breath Volume	0.000	05:27	
Air Blank	0.000	05:27	
Control Test	0.079	05:27	
Air Blank	0.000	05:28	
Diagnostics Check	OK	05:28	

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUGH, ROBERT, F  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: POLSON, BRIAN  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 07/30/2010 01:11:06

---

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Subject Sample #1	0.033	01:06
Breath Volume	2.183	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample #2	0.031	01:09
Breath Volume	2.039	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD, ERIC, L  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: BOYD, ERIC  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 07/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/31/2010 03:07:33

---

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.080	02:54	
Air Blank	0.000	02:54	
Subject Sample #1	0.199	02:57	Volume Not Met
Breath Volume	0.625	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample #2	0.246	03:00	
Breath Volume	2.160	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample #3		03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Control Test	0.078	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

---

Data Download Date/Time 08/31/2010 12:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERS, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: PETERS, MICHAEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 16:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: STUBBLEFIELD

DATE/TIME OF TEST: 07/03/2010 18:48:50

---

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:46	
Control Test	0.079	18:46	
Air Blank	0.000	18:47	
Subject Sample #1		18:47	Subject Test Refused
Breath Volume	0.000	18:47	
Air Blank	0.000	18:47	
Control Test	0.079	18:48	
Air Blank	0.000	18:48	
Diagnostics Check	OK	18:48	

---

Data Download Date/Time 07/29/2010 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JACKSON, C A, A

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: JACKSON, C

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 16:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 07/06/2010 18:30:51

---

Diagnostics Check	OK	18:24
Air Blank	0.000	18:25
Control Test	0.078	18:25
Air Blank	0.000	18:25
Subject Sample #1	0.153	18:26
Breath Volume	2.160	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:28
Subject Sample #2	0.153	18:29
Breath Volume	1.890	18:29
Air Blank	0.000	18:29
Control Test	0.078	18:30
Air Blank	0.000	18:30
Diagnostics Check	OK	18:30

---

Data Download Date/Time 07/29/2010 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JENKS, JOHN, R  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: JENKS, JOHN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 07/06/2010 22:40:37

---

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.077	22:32
Air Blank	0.000	22:33
Subject Sample #1	0.282	22:35
Breath Volume	4.058	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample #2	0.280	22:39
Breath Volume	3.355	22:39
Air Blank	0.000	22:39
Control Test	0.078	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

---

Data Download Date/Time 07/29/2010 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUHLBACH, BRIAN, C  
BREATH TEST OPERATOR AGENCY: Crestview PD  
ARRESTING OFFICER: MUHLBACH, BRIAN  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ROSINE

DATE/TIME OF TEST: 07/07/2010 02:29:44

---

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.078	02:27	
Air Blank	0.000	02:28	
Subject Sample #1		02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.078	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

---

Data Download Date/Time 07/29/2010 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON, HAROLD, E  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MASON, HAROLD  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SPURLOCK

DATE/TIME OF TEST: 07/13/2010 03:08:27

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Subject Sample #1	0.105	03:04
Breath Volume	2.398	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample #2	0.102	03:06
Breath Volume	1.949	03:06
Air Blank	0.000	03:07
Control Test	0.077	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

---

Data Download Date/Time 07/29/2010 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MUHLBACH, BRIAN, C

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: MUHLBACH, BRIAN

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 03:58

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CROOKS

DATE/TIME OF TEST: 07/15/2010 04:36:52

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:31
Subject Sample #1	0.192	04:32
Breath Volume	1.628	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample #2	0.191	04:35
Breath Volume	1.187	04:35
Air Blank	0.000	04:35
Control Test	0.078	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time 07/29/2010 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON, HAROLD, E  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MASON, HAROLD  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: VONTROTT

DATE/TIME OF TEST: 07/16/2010 03:36:56

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Subject Sample #1	0.108	03:32
Breath Volume	3.683	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample #2	0.107	03:35
Breath Volume	1.980	03:35
Air Blank	0.000	03:36
Control Test	0.077	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time 07/29/2010 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERS, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: PETERS, MICHAEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: WERLITZ

DATE/TIME OF TEST: 07/17/2010 06:57:52

---

Diagnostics Check	OK	06:50
Air Blank	0.000	06:51
Control Test	0.077	06:51
Air Blank	0.000	06:52
Subject Sample #1	0.178	06:52
Breath Volume	2.421	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample #2	0.174	06:56
Breath Volume	1.804	06:56
Air Blank	0.000	06:56
Control Test	0.076	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

---

Data Download Date/Time 07/29/2010 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON, HAROLD, E  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MASON, HAROLD  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/24/2010 01:08:46

---

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.078	01:03
Air Blank	0.000	01:03
Subject Sample #1	0.031	01:04
Breath Volume	2.734	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample #2	0.030	01:07
Breath Volume	3.164	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

---

Data Download Date/Time 07/29/2010 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON, HAROLD, E  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MASON, HAROLD  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 07/25/2010 01:40:42

---

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.079	01:38	
Air Blank	0.000	01:39	
Subject Sample #1		01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.078	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

---

Data Download Date/Time 07/29/2010 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTMAS, MATTHEW, S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: CHRISTMAS, MATTHEW  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 07/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 15:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: WOOTEN

DATE/TIME OF TEST: 07/30/2010 20:54:21

---

Diagnostics Check	OK	20:47
Air Blank	0.000	20:48
Control Test	0.077	20:48
Air Blank	0.000	20:49
Subject Sample #1	0.134	20:49
Breath Volume	1.156	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample #2	0.142	20:52
Breath Volume	1.425	20:52
Air Blank	0.000	20:53
Control Test	0.077	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

---

Data Download Date/Time 08/31/2010 15:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILEY, WILLIAM, J

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: WILEY, WILLIAM

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 01:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/01/2010 02:00:39

---

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.079	01:55
Air Blank	0.000	01:55
Subject Sample #1	0.191	01:56
Breath Volume	2.203	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample #2	0.206	01:59
Breath Volume	1.152	01:59
Air Blank	0.000	01:59
Control Test	0.079	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEETERS, OZZIE, H  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: TEETERS, OZZIE  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CORREDERA

DATE/TIME OF TEST: 07/01/2010 02:51:31

---

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Subject Sample #1	0.126	02:46
Breath Volume	1.605	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample #2	0.116	02:49
Breath Volume	1.316	02:49
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREEMAN, JOHNNY, A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: FREEMAN, JOHNNY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DIONNE

DATE/TIME OF TEST: 07/03/2010 00:35:20

---

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.078	00:29
Air Blank	0.000	00:30
Subject Sample #1	0.203	00:30
Breath Volume	1.613	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample #2	0.196	00:33
Breath Volume	1.250	00:33
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEYES, GABRIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KEYES, GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 00:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/03/2010 01:11:42

---

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Subject Sample #1	0.190	01:07
Breath Volume	2.855	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample #2	0.186	01:10
Breath Volume	2.148	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL, CHARLES, L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MITCHELL, CHARLES  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GALLOWAY

DATE/TIME OF TEST: 07/03/2010 02:17:15

---

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Subject Sample #1	0.210	02:12
Breath Volume	2.769	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample #2	0.221	02:15
Breath Volume	2.027	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JERNIGAN, DAVID, L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BERRYMAN, DOUGLAS  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: BUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: CHADBOURNE

DATE/TIME OF TEST: 07/03/2010 21:04:26

---

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.078	20:58
Air Blank	0.000	20:59
Subject Sample #1	0.178	20:59
Breath Volume	3.054	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample #2	0.177	21:02
Breath Volume	2.324	21:02
Air Blank	0.000	21:03
Control Test	0.079	21:03
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEELY, SCOTT, J  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: MOSES, JEFFERY  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/04/2010 01:47:01

---

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:42
Subject Sample #1	0.000	01:42
Breath Volume	1.695	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample #2	0.000	01:45
Breath Volume	4.023	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:47

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEELY, SCOTT, J  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: MOSES, JEFFERY  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/04/2010 01:58:57

---

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample #1	0.000	01:54
Breath Volume	4.394	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.000	01:57
Breath Volume	4.507	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEYES, GABRIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KEYES, GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 03:14

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 07/04/2010 04:12:43

---

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample #1	0.252	04:08
Breath Volume	1.136	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.248	04:11
Breath Volume	1.164	04:11
Air Blank	0.000	04:11
Control Test	0.078	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK, KEITH, A  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BURNSD, DAVID  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 16:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: VANDERBURG

DATE/TIME OF TEST: 07/04/2010 18:15:06

---

Diagnostics Check	OK	18:12	
Air Blank	0.000	18:12	
Control Test	0.080	18:12	
Air Blank	0.000	18:13	
Subject Sample #1		18:13	Subject Test Refused
Breath Volume	0.000	18:13	
Air Blank	0.000	18:14	
Control Test	0.079	18:14	
Air Blank	0.000	18:15	
Diagnostics Check	OK	18:15	

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEELY, SCOTT, J  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SEELY, SCOTT  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 07/05/2010 01:21:02

---

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:15
Subject Sample #1	0.250	01:16
Breath Volume	3.265	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample #2	0.259	01:19
Breath Volume	3.207	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:21

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEYES, GABRIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KEYES, GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 01:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/05/2010 02:06:36

---

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.081	02:01
Air Blank	0.000	02:01
Subject Sample #1	0.172	02:02
Breath Volume	1.781	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample #2	0.175	02:05
Breath Volume	2.046	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEYES, GABRIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KEYES, GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 04:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/05/2010 04:51:12

---

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Subject Sample #1	0.037	04:46
Breath Volume	2.644	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample #2	0.037	04:49
Breath Volume	2.714	04:49
Air Blank	0.000	04:50
Control Test	0.078	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREEMAN, JOHNNY, A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PANICI, PAUL  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 07/05/2010 07:39:15

---

Diagnostics Check	OK	07:31
Air Blank	0.000	07:31
Control Test	0.080	07:32
Air Blank	0.000	07:32
Subject Sample #1	0.212	07:33
Breath Volume	3.453	07:33
Air Blank	0.000	07:34
Air Blank	0.000	07:35
Subject Sample #2	0.218	07:37
Breath Volume	2.187	07:37
Air Blank	0.000	07:38
Control Test	0.079	07:38
Air Blank	0.000	07:39
Diagnostics Check	OK	07:39

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL, CHARLES, L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MITCHELL, CHARLES  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 07/05/2010 21:50:52

---

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.076	21:45
Air Blank	0.000	21:46
Subject Sample #1	0.299	21:46
Breath Volume	2.402	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:48
Subject Sample #2	0.297	21:49
Breath Volume	2.277	21:49
Air Blank	0.000	21:49
Control Test	0.078	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILEY, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WILEY, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: DOLAMORE

DATE/TIME OF TEST: 07/06/2010 02:03:11

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Subject Sample #1	0.189	01:58
Breath Volume	2.339	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample #2	0.207	02:01
Breath Volume	2.507	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TEETERS, OZZIE, H

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: TEETERS, OZZIE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 04:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 07/06/2010 04:28:05

---

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.079	04:26	
Air Blank	0.000	04:26	
Subject Sample #1		04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWANSON, JEFF, E  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: VELTKAMP, PAM  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CORVO

DATE/TIME OF TEST: 07/07/2010 04:52:13

---

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.078	04:46
Air Blank	0.000	04:46
Subject Sample #1	0.209	04:47
Breath Volume	2.636	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample #2	0.208	04:50
Breath Volume	2.019	04:50
Air Blank	0.000	04:51
Control Test	0.078	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILEY, WILLIAM, J

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: WILEY, WILLIAM

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 05:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/08/2010 07:28:39

---

Diagnostics Check	OK	07:20
Air Blank	0.000	07:21
Control Test	0.080	07:21
Air Blank	0.000	07:22
Subject Sample #1	0.187	07:24
Breath Volume	2.570	07:24
Air Blank	0.000	07:24
Air Blank	0.000	07:26
Subject Sample #2	0.192	07:27
Breath Volume	2.183	07:27
Air Blank	0.000	07:27
Control Test	0.078	07:28
Air Blank	0.000	07:28
Diagnostics Check	OK	07:28

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEYES, GABRIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KEYES, GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 04:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 07/09/2010 06:14:58

---

Diagnostics Check	OK	06:08
Air Blank	0.000	06:08
Control Test	0.078	06:09
Air Blank	0.000	06:09
Subject Sample #1	0.055	06:10
Breath Volume	1.773	06:10
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample #2	0.054	06:13
Breath Volume	1.332	06:13
Air Blank	0.000	06:14
Control Test	0.079	06:14
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEYES, GABRIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KEYES, GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 22:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CUSHING

DATE/TIME OF TEST: 07/09/2010 23:37:40

---

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Subject Sample #1	0.204	23:32
Breath Volume	2.664	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample #2	0.205	23:36
Breath Volume	2.433	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL, CHARLES, L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MITCHELL, CHARLES  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 12:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: TOUSSANT

DATE/TIME OF TEST: 07/10/2010 01:45:43

---

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:43	
Subject Sample #1		01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Control Test	0.079	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEYES, GABRIEL, S  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: KEYES, GABRIEL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 07/10/2010 04:18:50

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:13
Subject Sample #1	0.104	04:14
Breath Volume	2.632	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample #2	0.105	04:17
Breath Volume	2.960	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEETERS, OZZIE, H  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: TEETERS, OZZIE  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: KIEK

DATE/TIME OF TEST: 07/10/2010 20:25:08

---

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.079	20:19
Air Blank	0.000	20:20
Subject Sample #1	0.227	20:20
Breath Volume	1.722	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample #2	0.221	20:23
Breath Volume	1.667	20:23
Air Blank	0.000	20:24
Control Test	0.080	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULT, AMY, L  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: DOCKERY, SHAWN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 11:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 07/11/2010 11:43:25

---

Diagnostics Check	OK	11:37
Air Blank	0.000	11:37
Control Test	0.080	11:38
Air Blank	0.000	11:38
Subject Sample #1	0.000	11:39
Breath Volume	2.230	11:39
Air Blank	0.000	11:39
Air Blank	0.000	11:41
Subject Sample #2	0.000	11:42
Breath Volume	2.324	11:42
Air Blank	0.000	11:42
Control Test	0.079	11:42
Air Blank	0.000	11:43
Diagnostics Check	OK	11:43

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNOTTS, CHRISTOPHER, W  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: DUNLAP, BENJAMIN  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 17:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 07/11/2010 20:14:33

---

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:12	
Control Test	0.079	20:12	
Air Blank	0.000	20:12	
Subject Sample #1		20:13	Subject Test Refused
Breath Volume	0.000	20:13	
Air Blank	0.000	20:13	
Control Test	0.079	20:13	
Air Blank	0.000	20:14	
Diagnostics Check	OK	20:14	

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUNT, JESSE, F

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: HUNT, JESSE

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 19:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/11/2010 20:39:22

---

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.080	20:31	
Air Blank	0.000	20:31	
Subject Sample #1		20:35	No Sample Provided
Breath Volume	0.000	20:35	
Air Blank	0.000	20:35	
Air Blank	0.000	20:37	
Subject Sample #2		20:38	Subject Test Refused
Breath Volume	0.000	20:38	
Air Blank	0.000	20:38	
Control Test	0.080	20:38	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEYES, GABRIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KEYES, GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 01:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 07/12/2010 02:29:54

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Subject Sample #1	0.176	02:25
Breath Volume	2.292	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample #2	0.166	02:28
Breath Volume	1.300	02:28
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILEY, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WILEY, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DUNSFORD

DATE/TIME OF TEST: 07/13/2010 03:30:49

---

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Subject Sample #1	0.203	03:26
Breath Volume	2.683	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample #2	0.193	03:29
Breath Volume	1.886	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREEMAN, JOHNNY, A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: FREEMAN, 1S03  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: PUTWEN

DATE/TIME OF TEST: 07/13/2010 07:46:25

---

Diagnostics Check	OK	07:39
Air Blank	0.000	07:40
Control Test	0.080	07:40
Air Blank	0.000	07:40
Subject Sample #1	0.179	07:41
Breath Volume	1.417	07:41
Air Blank	0.000	07:42
Air Blank	0.000	07:43
Subject Sample #2	0.182	07:44
Breath Volume	1.234	07:44
Air Blank	0.000	07:45
Control Test	0.079	07:45
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TEETERS, OZZIE, H

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: TEETERS, OZZIE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 01:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HENCYE

DATE/TIME OF TEST: 07/16/2010 02:06:19

---

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Subject Sample #1	0.124	02:01
Breath Volume	2.062	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample #2	0.124	02:04
Breath Volume	1.589	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TREMMEL, MELISSA, A  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: HOLLOWAY, JACOB  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/16/2010 23:15:10

---

Diagnostics Check	OK	23:14	
Air Blank		23:14	Ambient Fail
Air Blank		23:15	Improper Sample

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TREMMEL, MELISSA, A  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: HOLLOWAY, JACOB  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/16/2010 23:32:46

---

Diagnostics Check	OK	23:31	
Air Blank		23:32	Ambient Fail
Air Blank	0.000	23:32	

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TREMMEL, MELISSA, A  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: HOLLOWAY, JACOB  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ROGER

DATE/TIME OF TEST: 07/16/2010 23:39:02

---

Diagnostics Check	OK	23:38	
Air Blank		23:38	Ambient Fail
Air Blank	0.000	23:39	

---

Data Download Date/Time	07/23/2010	12:46
-------------------------	------------	-------

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAJORS, CHRIS, L  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: HOLLOWAY, JACOB  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/16/2010 23:44:18

---

Diagnostics Check	OK	23:43	
Air Blank		23:44	Improper Sample
Air Blank		23:44	Improper Sample

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TUCKER, JOSHUA, M

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: TUCKER, JOSHUA

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 627099

DRY GAS EXPIRY DATE: 06/03/2011

VIOLATION CODE: DUI

ARREST TIME: 06:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:52

SUBJECT'S LAST NAME: KILLAM

DATE/TIME OF TEST: 07/18/2010 08:16:30

---

Diagnostics Check	OK	08:12	
Air Blank	0.000	08:13	
Control Test	0.080	08:13	
Air Blank	0.000	08:14	
Subject Sample #1		08:15	Subject Test Refused
Breath Volume	0.000	08:15	
Air Blank	0.000	08:15	
Control Test	0.079	08:15	
Air Blank	0.000	08:16	
Diagnostics Check	OK	08:16	

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TATUM, JIMMIE, H  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: TATUM, JIMMIE  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 17:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: BURCH

DATE/TIME OF TEST: 07/18/2010 18:21:59

---

Diagnostics Check	OK	18:18	
Air Blank	0.000	18:18	
Control Test	0.080	18:19	
Air Blank	0.000	18:19	
Subject Sample #1		18:20	Subject Test Refused
Breath Volume	0.000	18:20	
Air Blank	0.000	18:21	
Control Test	0.079	18:21	
Air Blank	0.000	18:21	
Diagnostics Check	OK	18:21	

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TATUM, JIMMIE, H  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: TATUM, JIMMIE  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SADOCK

DATE/TIME OF TEST: 07/18/2010 21:35:06

---

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.078	21:27
Air Blank	0.000	21:28
Subject Sample #1	0.179	21:30
Breath Volume	1.199	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample #2	0.176	21:33
Breath Volume	1.914	21:33
Air Blank	0.000	21:34
Control Test	0.077	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WRIGHT, L,  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WRIGHT, L  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 627099  
DRY GAS EXPIRY DATE: 06/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 13:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: STUBBLEFIELD

DATE/TIME OF TEST: 07/20/2010 14:35:23

---

Diagnostics Check	OK	14:29
Air Blank	0.000	14:29
Control Test	0.080	14:30
Air Blank	0.000	14:30
Subject Sample #1	0.000	14:31
Breath Volume	2.703	14:31
Air Blank	0.000	14:31
Air Blank	0.000	14:33
Subject Sample #2	0.000	14:34
Breath Volume	1.539	14:34
Air Blank	0.000	14:34
Control Test	0.079	14:34
Air Blank	0.000	14:35
Diagnostics Check	OK	14:35

---

Data Download Date/Time 07/23/2010 12:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAMER, JOHNATHAN, M  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: CRAMER, JOHNATHAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 07/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: FRICK

DATE/TIME OF TEST: 07/23/2010 19:54:42

---

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.079	19:48
Air Blank	0.000	19:49
Subject Sample #1	0.226	19:49
Breath Volume	1.195	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample #2	0.217	19:53
Breath Volume	1.628	19:53
Air Blank	0.000	19:53
Control Test	0.081	19:54
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

---

Data Download Date/Time 08/19/2010 15:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAMER, JOHNATHAN, M  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: CRAMER, JOHNATHAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 07/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 07/24/2010 01:36:01

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample #1	0.236	01:31
Breath Volume	1.441	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample #2	0.226	01:34
Breath Volume	1.519	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:36

---

Data Download Date/Time 08/19/2010 15:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUNT, JESSE, F

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: HUNT, JESSE

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 07/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 622245

DRY GAS EXPIRY DATE: 05/28/2011

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: STROMBERG

DATE/TIME OF TEST: 07/24/2010 02:28:11

---

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample #1	0.198	02:23
Breath Volume	2.167	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample #2	0.194	02:26
Breath Volume	2.125	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

---

Data Download Date/Time 08/19/2010 15:08

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROPER, DAVID, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ROPER, DAVID  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: A69072  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 07/26/2010 03:15:13

---

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Subject Sample #1	0.200	03:10
Breath Volume	1.488	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample #2	0.208	03:13
Breath Volume	1.171	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time 08/19/2010 15:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTIAGO, BRIAN,  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: SANTIAGO, BRIAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 07/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG907202  
DRY GAS EXPIRY DATE: 03/13/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MONEY

DATE/TIME OF TEST: 07/28/2010 04:27:43

---

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample #1	0.209	04:23
Breath Volume	3.070	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.206	04:26
Breath Volume	3.593	04:26
Air Blank	0.000	04:26
Control Test	0.081	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time 08/19/2010 15:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KHUNE, RONALD, I  
BREATH TEST OPERATOR AGENCY: FDOT MCC  
ARRESTING OFFICER: KHUNE, RONALD  
ARRESTING OFFICER AGENCY: FDOT MCC  
LAST AGENCY INSPECTION: 07/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG907202  
DRY GAS EXPIRY DATE: 03/13/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HATCHER JR

DATE/TIME OF TEST: 07/29/2010 00:03:09

---

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:53
Subject Sample #1	0.275	23:55
Breath Volume	1.347	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:57
Subject Sample #2	0.246	23:58
Breath Volume	3.230	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample #3	0.238	00:01
Breath Volume	3.890	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time 08/19/2010 15:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GORMLEY, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: GORMLEY, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: STICKLE

DATE/TIME OF TEST: 07/30/2010 22:25:51

---

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:18	
Control Test	0.079	22:18	
Air Blank	0.000	22:19	
Subject Sample #1	0.210	22:20	
Breath Volume	2.378	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:22	
Subject Sample #2	0.214	22:24	
Breath Volume	1.691	22:24	
Air Blank	0.000	22:24	
Control Test	0.095	22:25	Interferent Detect
Air Blank		22:25	Purge Fail

---

Data Download Date/Time 08/19/2010 15:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GORMLEY, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: GORMLEY, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: STICKLE

DATE/TIME OF TEST: 07/30/2010 22:33:23

---

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.078	22:29	
Air Blank	0.000	22:30	
Subject Sample #1		22:31	Interferent Detect
Breath Volume	1.515	22:31	
Air Blank	0.000	22:32	
Control Test	0.056	22:32	Interferent Detect
Air Blank	0.000	22:33	

---

Data Download Date/Time 08/19/2010 15:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL, JAMES, R  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: ZABELLE, JOHN  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 07/03/2010 01:53:41

---

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.077	01:51	
Air Blank	0.000	01:51	
Subject Sample #1		01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.075	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

---

Data Download Date/Time 07/23/2010 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILEY, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WILEY, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 07/05/2010 01:53:34

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.077	01:47
Air Blank	0.000	01:48
Subject Sample #1	0.117	01:49
Breath Volume	5.527	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample #2	0.114	01:52
Breath Volume	3.464	01:52
Air Blank	0.000	01:52
Control Test	0.078	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

---

Data Download Date/Time 07/23/2010 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TREMMEL, MELISSA, A  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: HOLLOWAY, JACOB  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/17/2010 00:01:41

---

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.077	23:56
Air Blank	0.000	23:56
Subject Sample #1	0.222	23:57
Breath Volume	4.031	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample #2	0.240	00:00
Breath Volume	6.050	00:00
Air Blank	0.000	00:00
Control Test	0.076	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

---

Data Download Date/Time 07/23/2010 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROEDEL, JUSTIN, M

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: ROEDEL, JUSTIN

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 622245

DRY GAS EXPIRY DATE: 05/28/2011

VIOLATION CODE: DUI

ARREST TIME: 23:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/17/2010 00:28:29

---

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.076	00:22
Air Blank	0.000	00:22
Subject Sample #1	0.281	00:24
Breath Volume	2.203	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample #2	0.264	00:27
Breath Volume	1.398	00:27
Air Blank	0.000	00:27
Control Test	0.076	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

---

Data Download Date/Time 07/23/2010 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TREMMEL, MELISSA, A  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: RIVKIN, SCOTT  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 07/17/2010 01:37:08

---

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.077	01:31
Air Blank	0.000	01:31
Subject Sample #1	0.130	01:32
Breath Volume	4.464	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample #2	0.126	01:35
Breath Volume	3.964	01:35
Air Blank	0.000	01:36
Control Test	0.077	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

---

Data Download Date/Time 07/23/2010 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RITTENHOUSE, WILLIE, P  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RITTENHOUSE, WILLIE  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/17/2010 02:37:33

---

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.077	02:31
Air Blank	0.000	02:32
Subject Sample #1	0.148	02:32
Breath Volume	2.941	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample #2	0.148	02:36
Breath Volume	3.136	02:36
Air Blank	0.000	02:36
Control Test	0.077	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time 07/23/2010 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARRIS, MELISSA, G

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: HARRIS, MELISSA

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 07/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 622245

DRY GAS EXPIRY DATE: 05/28/2011

VIOLATION CODE: DUI

ARREST TIME: 02:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: STALLWORTH

DATE/TIME OF TEST: 07/26/2010 03:06:57

---

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.077	03:01
Air Blank	0.000	03:01
Subject Sample #1	0.225	03:02
Breath Volume	3.722	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample #2	0.229	03:05
Breath Volume	4.609	03:05
Air Blank	0.000	03:06
Control Test	0.077	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time 08/19/2010 15:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECK, WILLIAM, M  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: GALLOP, CARY  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 07/30/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: DISTEFANO

DATE/TIME OF TEST: 07/30/2010 22:35:25

---

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Subject Sample #1	0.138	22:30
Breath Volume	2.402	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample #2	0.129	22:33
Breath Volume	1.933	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

---

Data Download Date/Time 08/17/2010 10:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATTEN, WILLIAM, R  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PATTEN, WILLIAM  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 07/30/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 07/30/2010 23:27:57

---

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Subject Sample #1	0.074	23:23
Breath Volume	2.121	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample #2	0.074	23:26
Breath Volume	2.078	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

---

Data Download Date/Time 08/17/2010 10:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUTNEL, THOMAS, K  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PUTNEL, THOMAS  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 07/30/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/31/2010 20:44:04

---

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.079	20:38
Air Blank	0.000	20:39
Subject Sample #1	0.262	20:39
Breath Volume	2.632	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:41
Subject Sample #2	0.248	20:42
Breath Volume	2.011	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:44

---

Data Download Date/Time 08/17/2010 10:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER, C, L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ALEXANDER, C  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 07/01/2010 04:18:34

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:13
Subject Sample #1	0.209	04:13
Breath Volume	1.691	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample #2	0.204	04:17
Breath Volume	1.765	04:17
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: SHIDEMANTLE

DATE/TIME OF TEST: 07/01/2010 04:45:00

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Subject Sample #1	0.145	04:40
Breath Volume	1.503	04:40
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample #2	0.144	04:43
Breath Volume	2.074	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:45

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD, ERIC, L  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: BOYD, ERIC  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 07/01/2010 05:26:26

---

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:23	
Control Test	0.078	05:24	
Air Blank	0.000	05:24	
Subject Sample #1		05:25	Subject Test Refused
Breath Volume	0.000	05:25	
Air Blank	0.000	05:25	
Control Test	0.077	05:25	
Air Blank	0.000	05:26	
Diagnostics Check	OK	05:26	

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: BORN

DATE/TIME OF TEST: 07/02/2010 04:37:58

---

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.078	04:31
Air Blank	0.000	04:32
Subject Sample #1	0.183	04:33
Breath Volume	1.675	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:35
Subject Sample #2	0.187	04:36
Breath Volume	2.019	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD, ERIC, L  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: BOYD, ERIC  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 07/02/2010 05:01:36

---

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.078	04:55
Air Blank	0.000	04:56
Subject Sample #1	0.218	04:56
Breath Volume	4.781	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample #2	0.216	04:59
Breath Volume	4.691	04:59
Air Blank	0.000	05:00
Control Test	0.078	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: FUSCALDO

DATE/TIME OF TEST: 07/03/2010 03:08:07

---

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.079	03:06	
Air Blank	0.000	03:06	
Subject Sample #1		03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.078	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: HINRICHS

DATE/TIME OF TEST: 07/04/2010 03:33:00

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample #1	0.151	03:26
Breath Volume	1.875	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample #2	0.150	03:31
Breath Volume	2.136	03:31
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:33

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALEXANDER, C, L

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: ALEXANDER, C

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 829401G

DRY GAS EXPIRY DATE: 11/07/2010

VIOLATION CODE: DUI

ARREST TIME: 02:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/04/2010 04:07:16

---

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Subject Sample #1	0.109	04:02
Breath Volume	1.910	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample #2	0.107	04:05
Breath Volume	1.734	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REWIS, CHAD, L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: REWIS, CHAD  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: LANDERS

DATE/TIME OF TEST: 07/04/2010 23:19:55

---

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:14
Subject Sample #1	0.195	23:15
Breath Volume	1.789	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:17
Subject Sample #2	0.183	23:18
Breath Volume	2.570	23:18
Air Blank	0.000	23:19
Control Test	0.078	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REWIS, CHAD, L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: REWIS, CHAD  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 07/05/2010 04:42:37

---

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Subject Sample #1	0.188	04:37
Breath Volume	1.765	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample #2	0.188	04:41
Breath Volume	1.351	04:41
Air Blank	0.000	04:41
Control Test	0.079	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAICHE, MARK, J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: DRISCOLL JR, KENNETH  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 13:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/05/2010 15:37:53

---

Diagnostics Check	OK	15:29
Air Blank	0.000	15:30
Control Test	0.078	15:30
Air Blank	0.000	15:31
Subject Sample #1	0.123	15:33
Breath Volume	1.226	15:33
Air Blank	0.000	15:34
Air Blank	0.000	15:36
Subject Sample #2	0.130	15:36
Breath Volume	1.324	15:36
Air Blank	0.000	15:36
Control Test	0.078	15:37
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 07/09/2010 01:01:35

---

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.078	00:59	
Air Blank	0.000	00:59	
Subject Sample #1		01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Control Test	0.078	01:00	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILBERT, JEREMY, A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: GILBERT, JEREMY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: TEER

DATE/TIME OF TEST: 07/09/2010 01:48:32

---

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:46	
Subject Sample #1		01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON, ROBERT, S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WEEKS, NOLAN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 21:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 07/10/2010 23:08:51

---

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.079	23:06	
Air Blank	0.000	23:07	
Subject Sample #1		23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:07	
Control Test	0.078	23:08	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: CAVAZOS

DATE/TIME OF TEST: 07/11/2010 05:03:24

---

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.078	04:56
Air Blank	0.000	04:57
Subject Sample #1	0.201	04:58
Breath Volume	2.199	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample #2	0.208	05:01
Breath Volume	1.917	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SCHILL, SAMANTHA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: BALDERRAMA

DATE/TIME OF TEST: 07/13/2010 23:02:14

---

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Subject Sample #1	0.126	22:57
Breath Volume	1.722	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample #2	0.131	23:00
Breath Volume	2.285	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON, ROBERT, S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MILEY, CARL  
ARRESTING OFFICER AGENCY: Shalimar PD  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ROHL

DATE/TIME OF TEST: 07/15/2010 00:25:58

---

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:19
Subject Sample #1	0.215	00:20
Breath Volume	2.375	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample #2	0.204	00:24
Breath Volume	2.179	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEMENT, LANCE, C  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BEMENT, LANCE  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/16/2010 23:39:42

---

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Subject Sample #1	0.188	23:34
Breath Volume	1.777	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample #2	0.204	23:38
Breath Volume	1.160	23:38
Air Blank	0.000	23:38
Control Test	0.078	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 07/17/2010 03:48:19

---

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample #1	0.136	03:43
Breath Volume	2.855	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample #2	0.131	03:46
Breath Volume	2.363	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER, C, L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ALEXANDER, C  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 05:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/17/2010 06:29:33

---

Diagnostics Check	OK	06:26	
Air Blank	0.000	06:27	
Control Test	0.078	06:27	
Air Blank	0.000	06:27	
Subject Sample #1		06:28	Subject Test Refused
Breath Volume	0.000	06:28	
Air Blank	0.000	06:28	
Control Test	0.077	06:28	
Air Blank	0.000	06:29	
Diagnostics Check	OK	06:29	

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOLTERS JR, LESTER, R  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WOLTERS JR., LESTER  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 07/18/2010 01:45:01

---

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample #1	0.155	01:40
Breath Volume	2.597	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample #2	0.155	01:43
Breath Volume	3.273	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:45

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/18/2010 02:04:31

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.078	01:58
Air Blank	0.000	01:58
Subject Sample #1	0.147	02:00
Breath Volume	2.554	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample #2	0.145	02:03
Breath Volume	2.234	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALEXANDER, C, L

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: ALEXANDER, C

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 829401G

DRY GAS EXPIRY DATE: 11/07/2010

VIOLATION CODE: DUI

ARREST TIME: 03:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: JOUN

DATE/TIME OF TEST: 07/20/2010 04:41:29

---

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample #1	0.156	04:36
Breath Volume	4.132	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample #2	0.149	04:39
Breath Volume	3.398	04:39
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MOLINARI

DATE/TIME OF TEST: 07/22/2010 00:37:41

---

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.079	00:26	
Air Blank	0.000	00:27	
Subject Sample #1	0.201	00:30	Volume Not Met
Breath Volume	0.121	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:32	
Subject Sample #2	0.195	00:36	Volume Not Met
Breath Volume	0.769	00:36	
Air Blank	0.000	00:36	
Control Test	0.077	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ENGLAND

DATE/TIME OF TEST: 07/24/2010 03:45:23

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Subject Sample #1	0.233	03:40
Breath Volume	1.207	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample #2	0.219	03:43
Breath Volume	2.121	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUNTZ, MARK, A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: KUNTZ, MARK  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 07/24/2010 05:59:30

---

Diagnostics Check	OK	05:52
Air Blank	0.000	05:53
Control Test	0.078	05:53
Air Blank	0.000	05:54
Subject Sample #1	0.053	05:54
Breath Volume	3.378	05:54
Air Blank	0.000	05:55
Air Blank	0.000	05:57
Subject Sample #2	0.051	05:57
Breath Volume	3.074	05:57
Air Blank	0.000	05:58
Control Test	0.076	05:58
Air Blank	0.000	05:59
Diagnostics Check	OK	05:59

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CLOUATRE

DATE/TIME OF TEST: 07/25/2010 05:12:31

---

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.079	05:06
Air Blank	0.000	05:06
Subject Sample #1	0.135	05:08
Breath Volume	1.152	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample #2	0.135	05:11
Breath Volume	2.101	05:11
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

---

Data Download Date/Time 07/28/2010 10:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER, C, L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ALEXANDER, C  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/31/2010 04:15:14

---

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:02	
Subject Sample #1	0.118	04:05	
Breath Volume	1.203	04:05	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample #2	0.119	04:10	Volume Not Met
Breath Volume	1.011	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample #3	0.123	04:13	
Breath Volume	1.457	04:13	
Air Blank	0.000	04:14	
Control Test	0.076	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

---

Data Download Date/Time 09/05/2010 07:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM, J, D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM, J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 07/31/2010 04:38:25

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.077	04:31
Air Blank	0.000	04:31
Subject Sample #1	0.135	04:33
Breath Volume	2.003	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample #2	0.153	04:36
Breath Volume	1.695	04:36
Air Blank	0.000	04:37
Control Test	0.077	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

---

Data Download Date/Time 09/05/2010 07:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANTZ, WILLIAM, R  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: LANTZ, WILLIAM  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: KRASNOV

DATE/TIME OF TEST: 07/31/2010 05:10:59

---

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:08	
Control Test	0.078	05:08	
Air Blank	0.000	05:09	
Subject Sample #1		05:09	Subject Test Refused
Breath Volume	0.000	05:09	
Air Blank	0.000	05:10	
Control Test	0.077	05:10	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

---

Data Download Date/Time 09/05/2010 07:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINKLER, RICHARD, J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: OSULLIVAN, MARCUS  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 04:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: SIMPLICIANO

DATE/TIME OF TEST: 07/31/2010 05:35:29

---

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.078	05:28
Air Blank	0.000	05:29
Subject Sample #1	0.126	05:30
Breath Volume	1.480	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample #2	0.133	05:33
Breath Volume	1.574	05:33
Air Blank	0.000	05:34
Control Test	0.077	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

---

Data Download Date/Time 09/05/2010 07:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEMENT, LANCE, C  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BEMENT, LANCE  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 829401G  
DRY GAS EXPIRY DATE: 11/07/2010  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: ES

DATE/TIME OF TEST: 07/31/2010 06:10:05

---

Diagnostics Check	OK	06:07	
Air Blank	0.000	06:07	
Control Test	0.077	06:07	
Air Blank	0.000	06:08	
Subject Sample #1		06:08	Subject Test Refused
Breath Volume	0.000	06:08	
Air Blank	0.000	06:09	
Control Test	0.078	06:09	
Air Blank	0.000	06:10	
Diagnostics Check	OK	06:10	

---

Data Download Date/Time 09/05/2010 07:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, WILLIAM, C

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: SMITH, WILLIAM

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 710643

DRY GAS EXPIRY DATE: 12/28/2011

VIOLATION CODE: DUI

ARREST TIME: 01:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: SCOLARI

DATE/TIME OF TEST: 07/04/2010 05:23:38

---

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.079	05:17
Air Blank	0.000	05:18
Subject Sample #1	0.279	05:19
Breath Volume	3.304	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample #2	0.278	05:22
Breath Volume	2.332	05:22
Air Blank	0.000	05:22
Control Test	0.078	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

---

Data Download Date/Time 07/12/2010 12:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SPIRAKIS, JOE  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: SARGENT

DATE/TIME OF TEST: 07/04/2010 05:38:05

---

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.077	05:32
Air Blank	0.000	05:33
Subject Sample #1	0.179	05:33
Breath Volume	2.503	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample #2	0.176	05:36
Breath Volume	3.019	05:36
Air Blank	0.000	05:37
Control Test	0.078	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:38

---

Data Download Date/Time 07/12/2010 12:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOODWIN, LORI, A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: GOODWIN, LORI  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 09:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: HARRIGAN

DATE/TIME OF TEST: 07/07/2010 10:16:46

---

Diagnostics Check	OK	10:09
Air Blank	0.000	10:10
Control Test	0.079	10:10
Air Blank	0.000	10:11
Subject Sample #1	0.039	10:12
Breath Volume	1.562	10:12
Air Blank	0.000	10:12
Air Blank	0.000	10:14
Subject Sample #2	0.038	10:15
Breath Volume	1.902	10:15
Air Blank	0.000	10:15
Control Test	0.078	10:16
Air Blank	0.000	10:16
Diagnostics Check	OK	10:16

---

Data Download Date/Time 07/12/2010 12:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, BRIAN, A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MAGDALANEY, ANDREW  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: NUCKOLS

DATE/TIME OF TEST: 07/10/2010 20:23:40

---

Diagnostics Check	OK	20:16
Air Blank	0.000	20:17
Control Test	0.078	20:17
Air Blank	0.000	20:18
Subject Sample #1	0.201	20:19
Breath Volume	1.359	20:19
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample #2	0.197	20:22
Breath Volume	1.222	20:22
Air Blank	0.000	20:22
Control Test	0.076	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

---

Data Download Date/Time 07/12/2010 12:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIRD, LANCE, R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HORN, KEVIN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: HAMBRICK

DATE/TIME OF TEST: 07/13/2010 19:21:37

---

Diagnostics Check	OK	19:10	
Air Blank	0.000	19:10	
Control Test	0.078	19:10	
Air Blank	0.000	19:11	
Subject Sample #1	0.171	19:14	Volume Not Met
Breath Volume	0.675	19:14	
Air Blank	0.000	19:15	
Air Blank	0.000	19:17	
Subject Sample #2	0.193	19:20	Volume Not Met
Breath Volume	0.136	19:20	
Air Blank	0.000	19:20	
Control Test	0.077	19:21	
Air Blank	0.000	19:21	
Diagnostics Check	OK	19:21	

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEICHNER, ADAM, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TEICHNER, ADAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: RING

DATE/TIME OF TEST: 07/16/2010 23:05:57

---

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:51	
Control Test	0.078	22:51	
Air Blank	0.000	22:51	
Subject Sample #1	0.214	22:53	
Breath Volume	1.906	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:55	
Subject Sample #2	0.222	22:59	Volume Not Met
Breath Volume	0.085	22:59	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample #3	0.200	23:04	
Breath Volume	2.324	23:04	
Air Blank	0.000	23:05	
Control Test	0.077	23:05	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONDON, BRYAN, W  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: LONDON, BRYAN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: KOCON

DATE/TIME OF TEST: 07/17/2010 04:00:54

---

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.078	03:55
Air Blank	0.000	03:55
Subject Sample #1	0.189	03:56
Breath Volume	3.746	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample #2	0.189	03:59
Breath Volume	2.507	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SMITH, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: SCARBROUGH

DATE/TIME OF TEST: 07/17/2010 18:10:16

---

Diagnostics Check	OK	17:58	
Air Blank	0.000	17:59	
Control Test	0.078	17:59	
Air Blank	0.000	18:00	
Subject Sample #1	0.047	18:03	Volume Not Met
Breath Volume	0.246	18:03	
Air Blank	0.000	18:03	
Air Blank	0.000	18:05	
Subject Sample #2	0.030	18:08	Volume Not Met
Breath Volume	0.367	18:08	
Air Blank	0.000	18:09	
Control Test	0.077	18:09	
Air Blank	0.000	18:10	
Diagnostics Check	OK	18:10	

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KAHALLEY, JAMEY, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: KAHALLEY, JAMEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 13:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: LIPP

DATE/TIME OF TEST: 07/18/2010 13:52:41

---

Diagnostics Check	Fail	13:52	Voltage/Current Test
Air Blank	0.000	13:52	

---

Data Download Date/Time	08/18/2010	14:23
-------------------------	------------	-------

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KAHALLEY, JAMEY, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: KAHALLEY, JAMEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 13:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: LIPP

DATE/TIME OF TEST: 07/18/2010 14:00:27

---

Diagnostics Check	OK	13:54
Air Blank	0.000	13:54
Control Test	0.079	13:55
Air Blank	0.000	13:55
Subject Sample #1	0.000	13:56
Breath Volume	2.769	13:56
Air Blank	0.000	13:56
Air Blank	0.000	13:58
Subject Sample #2	0.000	13:59
Breath Volume	2.996	13:59
Air Blank	0.000	13:59
Control Test	0.077	13:59
Air Blank	0.000	14:00
Diagnostics Check	OK	14:00

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEICHNER, ADAM, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TEICHNER, ADAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: FOLSOM

DATE/TIME OF TEST: 07/20/2010 02:23:07

---

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.078	02:20	
Air Blank	0.000	02:20	
Subject Sample #1		02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Control Test	0.079	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON, DURON, A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: NELSON, DURON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 07/23/2010 03:07:24

---

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.078	02:56	
Air Blank	0.000	02:57	
Subject Sample #1		03:00	No Sample Provided
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:02	
Subject Sample #2		03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Control Test	0.077	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEICHNER, ADAM, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TEICHNER, ADAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 07/24/2010 01:57:30

---

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:47	
Subject Sample #1		01:50	No Sample Provided
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample #2		01:56	No Sample Provided
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.078	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, CLAYTON, M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SMITH, CLAYTON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GALLOWAY

DATE/TIME OF TEST: 07/24/2010 03:38:18

---

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.078	03:35	
Air Blank	0.000	03:36	
Subject Sample #1		03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.078	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIS, BRIAN, E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WILLIS, BRIAN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 07/24/2010 23:30:51

---

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Subject Sample #1	0.151	23:26
Breath Volume	1.312	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample #2	0.160	23:29
Breath Volume	1.636	23:29
Air Blank	0.000	23:29
Control Test	0.078	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROPER, DAVID, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ROPER, DAVID  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 07/25/2010 06:59:19

---

Diagnostics Check	OK	06:52
Air Blank	0.000	06:52
Control Test	0.078	06:53
Air Blank	0.000	06:53
Subject Sample #1	0.000	06:54
Breath Volume	1.320	06:54
Air Blank	0.000	06:55
Air Blank	0.000	06:57
Subject Sample #2	0.000	06:57
Breath Volume	2.484	06:57
Air Blank	0.000	06:58
Control Test	0.078	06:58
Air Blank	0.000	06:59
Diagnostics Check	OK	06:59

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON, BRADLEY, D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: JORDAN, LUCAS  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: RAK

DATE/TIME OF TEST: 07/26/2010 02:06:41

---

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.079	01:59	
Air Blank	0.000	01:59	
Subject Sample #1		02:02	No Sample Provided
Breath Volume	0.000	02:02	
Air Blank	0.000	02:03	
Air Blank	0.000	02:05	
Subject Sample #2		02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:05	
Control Test	0.078	02:06	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIS, BRIAN, E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WILLIS, BRIAN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 07/27/2010 05:27:46

---

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:23	
Control Test	0.079	05:23	
Air Blank	0.000	05:23	
Subject Sample #1		05:26	Subject Test Refused
Breath Volume	0.000	05:26	
Air Blank	0.000	05:26	
Control Test	0.079	05:27	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

---

Data Download Date/Time 08/18/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FALGOUT, DAVID, G  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: FALGOUT, DAVID  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581913  
DRY GAS EXPIRY DATE: 02/23/2011  
VIOLATION CODE: ZERO TOL  
ARREST TIME: 17:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 07/26/2010 17:54:03

---

Diagnostics Check	OK	17:47
Air Blank	0.000	17:48
Control Test	0.081	17:48
Air Blank	0.000	17:49
Subject Sample #1	0.000	17:49
Breath Volume	2.449	17:49
Air Blank	0.000	17:50
Air Blank	0.000	17:52
Subject Sample #2	0.000	17:52
Breath Volume	2.152	17:52
Air Blank	0.000	17:53
Control Test	0.080	17:53
Air Blank	0.000	17:53
Diagnostics Check	OK	17:54

---

Data Download Date/Time 08/06/2010 09:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILL, RICHARD, W  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: 07262010, RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581913  
DRY GAS EXPIRY DATE: 02/23/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: TOMPKINS

DATE/TIME OF TEST: 07/26/2010 20:54:53

---

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.081	20:47
Air Blank	0.000	20:48
Subject Sample #1	0.172	20:50
Breath Volume	2.023	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample #2	0.168	20:53
Breath Volume	1.640	20:53
Air Blank	0.000	20:53
Control Test	0.079	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

---

Data Download Date/Time 08/06/2010 09:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCINTYRE, CODY, A  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: MCINTYRE, CODY  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581913  
DRY GAS EXPIRY DATE: 02/23/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LENARD

DATE/TIME OF TEST: 07/27/2010 02:23:12

---

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:17
Air Blank	0.000	02:17
Subject Sample #1	0.191	02:18
Breath Volume	2.394	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample #2	0.194	02:21
Breath Volume	2.437	02:21
Air Blank	0.000	02:22
Control Test	0.082	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

---

Data Download Date/Time 08/06/2010 09:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEE, CHARLES, L

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: LEE, CHARLES

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 21:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SOTO-VASQUEZ

DATE/TIME OF TEST: 07/31/2010 22:42:01

---

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.082	22:35
Air Blank	0.000	22:36
Subject Sample #1	0.190	22:37
Breath Volume	1.906	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample #2	0.191	22:40
Breath Volume	1.562	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:42

---

Data Download Date/Time 08/06/2010 09:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, TAMMY, L  
BREATH TEST OPERATOR AGENCY: Jennings PD  
ARRESTING OFFICER: GLOVER, GEORGE  
ARRESTING OFFICER AGENCY: Jennings PD  
LAST AGENCY INSPECTION: 07/06/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 16:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: CHINN

DATE/TIME OF TEST: 07/17/2010 17:02:20

---

Diagnostics Check	OK	16:55
Air Blank	0.000	16:56
Control Test	0.079	16:56
Air Blank	0.000	16:57
Subject Sample #1	0.302	16:57
Breath Volume	2.074	16:57
Air Blank	0.000	16:58
Air Blank	0.000	17:00
Subject Sample #2	0.307	17:00
Breath Volume	1.507	17:00
Air Blank	0.000	17:01
Control Test	0.079	17:01
Air Blank	0.000	17:02
Diagnostics Check	OK	17:02

---

Data Download Date/Time 07/29/2010 15:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEICHNER, ADAM, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TEICHNER, ADAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: CREMEANS

DATE/TIME OF TEST: 07/01/2010 07:18:04

---

Diagnostics Check	OK	07:10
Air Blank	0.000	07:11
Control Test	0.080	07:11
Air Blank	0.000	07:12
Subject Sample #1	0.000	07:13
Breath Volume	1.394	07:13
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample #2	0.000	07:16
Breath Volume	1.339	07:16
Air Blank	0.000	07:17
Control Test	0.079	07:17
Air Blank	0.000	07:17
Diagnostics Check	OK	07:18

---

Data Download Date/Time 07/12/2010 13:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAVES, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WARD, JOSEPH  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 16:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: COFFEY

DATE/TIME OF TEST: 07/01/2010 17:20:24

---

Diagnostics Check	OK	17:17	
Air Blank	0.000	17:17	
Control Test	0.080	17:18	
Air Blank	0.000	17:18	
Subject Sample #1		17:19	Subject Test Refused
Breath Volume	0.000	17:19	
Air Blank	0.000	17:19	
Control Test	0.080	17:19	
Air Blank	0.000	17:20	
Diagnostics Check	OK	17:20	

---

Data Download Date/Time 07/12/2010 13:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENFINGER, THOMAS, D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: GOODWIN, JOSHUA  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: WYERS

DATE/TIME OF TEST: 07/03/2010 05:02:18

---

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.081	04:55
Air Blank	0.000	04:55
Subject Sample #1	0.257	04:57
Breath Volume	2.445	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample #2	0.264	05:00
Breath Volume	2.464	05:00
Air Blank	0.000	05:01
Control Test	0.078	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

---

Data Download Date/Time 07/12/2010 13:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS, ANNA, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: OAKES, GEOFFREY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GETCHEL

DATE/TIME OF TEST: 07/04/2010 01:25:12

---

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Subject Sample #1	0.195	01:20
Breath Volume	1.562	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample #2	0.180	01:23
Breath Volume	1.597	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

---

Data Download Date/Time 07/12/2010 13:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILEY, WILLIAM, J

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: WILEY, WILLIAM

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 710643

DRY GAS EXPIRY DATE: 12/28/2011

VIOLATION CODE: DUI

ARREST TIME: 02:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: BANN

DATE/TIME OF TEST: 07/08/2010 03:15:13

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:10
Subject Sample #1	0.238	03:10
Breath Volume	2.605	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample #2	0.224	03:13
Breath Volume	1.585	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time 07/12/2010 13:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEICHNER, ADAM, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TEICHNER, ADAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 07/08/2010 06:09:52

---

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:58	
Control Test	0.082	05:59	
Air Blank	0.000	05:59	
Subject Sample #1	0.145	06:02	Volume Not Met
Breath Volume	0.089	06:02	
Air Blank	0.000	06:03	
Air Blank	0.000	06:05	
Subject Sample #2	0.136	06:08	Volume Not Met
Breath Volume	0.019	06:08	
Air Blank	0.000	06:08	
Control Test	0.080	06:09	
Air Blank	0.000	06:09	
Diagnostics Check	OK	06:09	

---

Data Download Date/Time 07/12/2010 13:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEICHNER, ADAM, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TEICHNER, ADAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: HEBERT

DATE/TIME OF TEST: 07/10/2010 01:43:28

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample #1	0.176	01:39
Breath Volume	2.140	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample #2	0.187	01:41
Breath Volume	1.535	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time 07/12/2010 13:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENFINGER, THOMAS, D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ENFINGER, THOMAS  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: DEUBLER

DATE/TIME OF TEST: 07/10/2010 04:59:42

---

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:54
Air Blank	0.000	04:54
Subject Sample #1	0.219	04:55
Breath Volume	3.570	04:55
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample #2	0.224	04:58
Breath Volume	3.582	04:58
Air Blank	0.000	04:58
Control Test	0.080	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

---

Data Download Date/Time 07/12/2010 13:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENFINGER, THOMAS, D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ENFINGER, THOMAS  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: HUYNH

DATE/TIME OF TEST: 07/11/2010 04:05:43

---

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample #1	0.020	04:01
Breath Volume	4.414	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample #2	0.015	04:04
Breath Volume	3.500	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

---

Data Download Date/Time 07/12/2010 13:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENFINGER, THOMAS, D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ENFINGER, THOMAS  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: HUYNH

DATE/TIME OF TEST: 07/11/2010 04:19:56

---

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:15
Subject Sample #1	0.015	04:15
Breath Volume	2.421	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample #2	0.014	04:18
Breath Volume	2.457	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

---

Data Download Date/Time 07/12/2010 13:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CYPRET, STEPHEN, D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CYPRET, STEPHEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 07/12/2010 20:19:11

---

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:14	
Control Test	0.081	20:15	
Air Blank	0.000	20:15	
Subject Sample #1		20:17	Subject Test Refused
Breath Volume	0.000	20:17	
Air Blank	0.000	20:18	
Control Test	0.080	20:18	
Air Blank	0.000	20:19	
Diagnostics Check	OK	20:19	

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VINCENT, ED, S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: VINCENT, ED  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 17:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: CHATTMAN

DATE/TIME OF TEST: 07/15/2010 18:29:30

---

Diagnostics Check	OK	18:22
Air Blank	0.000	18:23
Control Test	0.081	18:23
Air Blank	0.000	18:24
Subject Sample #1	0.245	18:24
Breath Volume	1.835	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample #2	0.262	18:27
Breath Volume	3.394	18:27
Air Blank	0.000	18:28
Control Test	0.079	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEICHNER, ADAM, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TEICHNER, ADAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: RING

DATE/TIME OF TEST: 07/16/2010 22:49:04

---

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.081	22:38	
Air Blank	0.000	22:38	
Subject Sample #1	0.253	22:41	Volume Not Met
Breath Volume	0.125	22:41	
Air Blank	0.000	22:42	
Air Blank	0.000	22:44	
Subject Sample #2	0.237	22:47	Volume Not Met
Breath Volume	0.332	22:47	
Air Blank	0.000	22:48	
Control Test	0.078	22:48	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:49	

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, JAMES, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: GOODWIN, JOSH  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 07/17/2010 04:41:27

---

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.081	04:39	
Air Blank	0.000	04:39	
Subject Sample #1		04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.081	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RITTENHOUSE, WILLIE, P  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RITTENHOUSE, WILLIE  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: LENTINI

DATE/TIME OF TEST: 07/18/2010 23:59:06

---

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.082	23:57	
Air Blank	0.000	23:57	
Subject Sample #1		23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.081	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON, DURON, A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: NELSON, DURON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 07/23/2010 03:11:59

---

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:05
Subject Sample #1	0.208	03:06
Breath Volume	1.187	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample #2	0.196	03:10
Breath Volume	2.792	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, JAMES, T  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: HARRIS, JAMES  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 07/24/2010 01:02:03

---

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Subject Sample #1	0.120	00:49	
Breath Volume	1.757	00:49	
Air Blank	0.000	00:49	
Air Blank	0.000	00:51	
Subject Sample #2	0.120	00:54	Volume Not Met
Breath Volume	0.882	00:54	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample #3	0.105	01:00	Volume Not Met
Breath Volume	0.402	01:00	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:02	

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON, DURON, A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BANEY, BRAD  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/29/2010 02:37:45

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.081	02:31
Air Blank	0.000	02:31
Subject Sample #1	0.279	02:33
Breath Volume	2.125	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample #2	0.291	02:36
Breath Volume	1.152	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROPER, DAVID, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ROPER, DAVID  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HOLIFIELD

DATE/TIME OF TEST: 07/30/2010 00:39:23

---

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.081	00:25	
Air Blank	0.000	00:26	
Subject Sample #1	0.211	00:26	
Breath Volume	2.304	00:26	
Air Blank	0.000	00:27	
Air Blank	0.000	00:29	
Subject Sample #2	0.203	00:32	Volume Not Met
Breath Volume	0.023	00:32	
Air Blank	0.000	00:33	
Air Blank	0.000	00:34	
Subject Sample #3	0.153	00:37	Volume Not Met
Breath Volume	0.000	00:37	
Air Blank	0.000	00:38	
Control Test	0.078	00:38	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SMITH, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 07/30/2010 17:06:44

---

Diagnostics Check	OK	16:54	
Air Blank	0.000	16:54	
Control Test	0.081	16:54	
Air Blank	0.000	16:55	
Subject Sample #1	0.000	16:58	Volume Not Met
Breath Volume	0.761	16:58	
Air Blank	0.000	16:59	
Air Blank	0.000	17:01	
Subject Sample #2	0.000	17:02	
Breath Volume	2.308	17:02	
Air Blank	0.000	17:02	
Air Blank	0.000	17:04	
Subject Sample #3	0.000	17:05	
Breath Volume	2.046	17:05	
Air Blank	0.000	17:05	
Control Test	0.078	17:06	
Air Blank	0.000	17:06	
Diagnostics Check	OK	17:06	

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLSTER, ROBERT, D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HOLSTER, ROBERT  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 10:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/31/2010 11:07:25

---

Diagnostics Check	OK	11:00
Air Blank	0.000	11:01
Control Test	0.081	11:01
Air Blank	0.000	11:02
Subject Sample #1	0.000	11:02
Breath Volume	3.484	11:02
Air Blank	0.000	11:03
Air Blank	0.000	11:05
Subject Sample #2	0.000	11:06
Breath Volume	2.503	11:06
Air Blank	0.000	11:06
Control Test	0.079	11:06
Air Blank	0.000	11:07
Diagnostics Check	OK	11:07

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIRD, LANCE, R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: LEACH, JOSHUA  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 07/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/31/2010 21:17:38

---

Diagnostics Check	OK	21:06	
Air Blank	0.000	21:06	
Control Test	0.080	21:07	
Air Blank	0.000	21:07	
Subject Sample #1		21:10	No Sample Provided
Breath Volume	0.000	21:10	
Air Blank	0.000	21:11	
Air Blank	0.000	21:13	
Subject Sample #2		21:16	No Sample Provided
Breath Volume	0.000	21:16	
Air Blank	0.000	21:16	
Control Test	0.081	21:17	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:17	

---

Data Download Date/Time 08/18/2010 14:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PACE, KEVIN, M

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: CIESLIK, JESSE

ARRESTING OFFICER AGENCY: Columbia County SO

LAST AGENCY INSPECTION: 06/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 643837

DRY GAS EXPIRY DATE: 07/16/2011

VIOLATION CODE: DUI

ARREST TIME: 03:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/01/2010 04:03:19

---

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.078	04:01	
Air Blank	0.000	04:01	
Subject Sample #1		04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

---

Data Download Date/Time 07/28/2010 13:11

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON, JIMMY, R  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: WATSON, JIMMY  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 06/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 643837  
DRY GAS EXPIRY DATE: 07/16/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 07/03/2010 03:26:56

---

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.077	03:20
Air Blank	0.000	03:21
Subject Sample #1	0.140	03:21
Breath Volume	1.792	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample #2	0.140	03:25
Breath Volume	1.480	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

---

Data Download Date/Time 07/28/2010 13:11

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARNELL, JONATHAN, M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: FARNELL, JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 643837  
DRY GAS EXPIRY DATE: 07/16/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/26/2010 00:04:10

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Subject Sample #1	0.167	23:59
Breath Volume	1.433	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample #2	0.166	00:02
Breath Volume	2.167	00:02
Air Blank	0.000	00:03
Control Test	0.076	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

---

Data Download Date/Time 07/28/2010 13:11

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTIZ, HENRY,  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CRUZ, J  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 643837  
DRY GAS EXPIRY DATE: 07/16/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HUTCH

DATE/TIME OF TEST: 07/31/2010 01:39:39

---

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:25	
Control Test	0.078	01:25	
Air Blank	0.000	01:25	
Subject Sample #1	0.254	01:29	Volume Not Met
Breath Volume	0.046	01:29	
Air Blank	0.000	01:29	
Air Blank	0.000	01:31	
Subject Sample #2	0.260	01:34	
Breath Volume	1.445	01:34	
Air Blank	0.000	01:35	
Air Blank	0.000	01:36	
Subject Sample #3	0.254	01:38	
Breath Volume	1.800	01:38	
Air Blank	0.000	01:38	
Control Test	0.074	01:39	Control Outside Tolerance
Air Blank	0.000	01:39	

---

Data Download Date/Time 08/27/2010 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BEAUFORD, JIM, M

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: KIRKLAND, KEITH

ARRESTING OFFICER AGENCY: Madison County SO

LAST AGENCY INSPECTION: 07/08/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 643837

DRY GAS EXPIRY DATE: 07/17/2011

VIOLATION CODE: DUI

ARREST TIME: 01:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 07/09/2010 03:20:20

---

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:18	
Subject Sample #1	0.000	03:18	
Breath Volume	3.722	03:18	
Air Blank		03:18	Improper Sample
Air Blank	0.000	03:19	
Control Test	0.079	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

---

Data Download Date/Time 08/17/2010 09:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BEAUFORD, JIM, M

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: BEAUFORD, JIM

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 07/08/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 643837

DRY GAS EXPIRY DATE: 07/17/2011

VIOLATION CODE: DUI

ARREST TIME: 01:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: AJANEL SICA

DATE/TIME OF TEST: 07/10/2010 02:54:26

---

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample #1	0.185	02:49
Breath Volume	4.148	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample #2	0.186	02:52
Breath Volume	3.492	02:52
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

---

Data Download Date/Time 08/17/2010 09:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, STEWART, G  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: PARKER, JAMES  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/08/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 643837  
DRY GAS EXPIRY DATE: 07/17/2011  
VIOLATION CODE: DUI  
ARREST TIME: 07:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:06

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 07/14/2010 08:46:09

---

Diagnostics Check	OK	08:39
Air Blank	0.000	08:39
Control Test	0.078	08:40
Air Blank	0.000	08:40
Subject Sample #1	0.151	08:41
Breath Volume	1.757	08:41
Air Blank	0.000	08:42
Air Blank	0.000	08:43
Subject Sample #2	0.155	08:44
Breath Volume	2.976	08:44
Air Blank	0.000	08:45
Control Test	0.075	08:45
Air Blank	0.000	08:46
Diagnostics Check	OK	08:46

---

Data Download Date/Time 08/17/2010 09:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: WASHINGTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITHD, DOUGLAS,  
BREATH TEST OPERATOR AGENCY: Washington County SO  
ARRESTING OFFICER: SMITH, D  
ARRESTING OFFICER AGENCY: Washington County SO  
LAST AGENCY INSPECTION: 06/30/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/23/2010 21:10:24

---

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.078	21:04
Air Blank	0.000	21:04
Subject Sample #1	0.055	21:05
Breath Volume	1.203	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample #2	0.054	21:08
Breath Volume	1.437	21:08
Air Blank	0.000	21:09
Control Test	0.077	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

---

Data Download Date/Time 07/29/2010 17:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHERROUSE, CRAIG, F  
BREATH TEST OPERATOR AGENCY: Holmes County SO  
ARRESTING OFFICER: BERBERT, HEATHER  
ARRESTING OFFICER AGENCY: Holmes County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832201I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 18:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 07/16/2010 19:08:53

---

Diagnostics Check	OK	19:01
Air Blank	0.000	19:02
Control Test	0.079	19:02
Air Blank	0.000	19:03
Subject Sample #1	0.000	19:04
Breath Volume	3.054	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample #2	0.000	19:07
Breath Volume	2.031	19:07
Air Blank	0.000	19:07
Control Test	0.080	19:08
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

---

Data Download Date/Time 08/29/2010 19:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHERROUSE, CRAIG, F  
BREATH TEST OPERATOR AGENCY: Holmes County SO  
ARRESTING OFFICER: BERBERT, HEATHER  
ARRESTING OFFICER AGENCY: Holmes County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832201I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 18:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 07/16/2010 19:22:34

---

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.080	19:16
Air Blank	0.000	19:17
Subject Sample #1	0.000	19:18
Breath Volume	2.394	19:18
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample #2	0.000	19:21
Breath Volume	2.679	19:21
Air Blank	0.000	19:21
Control Test	0.081	19:21
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

---

Data Download Date/Time 08/29/2010 19:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHERROUSE, CRAIG, F  
BREATH TEST OPERATOR AGENCY: Holmes County SO  
ARRESTING OFFICER: MAHAN, WILLIAM  
ARRESTING OFFICER AGENCY: Holmes County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832201I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: HEDLUND

DATE/TIME OF TEST: 07/18/2010 22:08:11

---

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:02
Air Blank	0.000	22:02
Subject Sample #1	0.167	22:03
Breath Volume	1.867	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample #2	0.173	22:06
Breath Volume	2.203	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

---

Data Download Date/Time 08/29/2010 19:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE, CHARLES, G  
BREATH TEST OPERATOR AGENCY: Holmes County SO  
ARRESTING OFFICER: SMITH, D  
ARRESTING OFFICER AGENCY: Washington County SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832201I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 15:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/21/2010 16:06:05

---

Diagnostics Check	OK	15:58
Air Blank	0.000	15:59
Control Test	0.079	15:59
Air Blank	0.000	16:00
Subject Sample #1	0.049	16:01
Breath Volume	2.996	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:03
Subject Sample #2	0.047	16:04
Breath Volume	2.644	16:04
Air Blank	0.000	16:05
Control Test	0.079	16:05
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

---

Data Download Date/Time 08/29/2010 19:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON, DALYN, J  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WILSON, DALYN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: BRAZWELL

DATE/TIME OF TEST: 07/03/2010 22:09:32

---

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:07	
Control Test	0.079	22:07	
Air Blank	0.000	22:07	
Subject Sample #1		22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:08	
Control Test	0.080	22:08	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

---

Data Download Date/Time 07/18/2010 22:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGILL, ROBERT, W  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: HARRIGILL, ROBERT  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: FIFE

DATE/TIME OF TEST: 07/04/2010 23:55:56

---

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:49
Subject Sample #1	0.121	23:51
Breath Volume	1.375	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample #2	0.128	23:54
Breath Volume	1.890	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

---

Data Download Date/Time 07/18/2010 22:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NGUYEN, TAI, T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: NGUYEN, TAI  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: D'ANGELO

DATE/TIME OF TEST: 07/07/2010 01:16:05

---

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Subject Sample #1	0.154	01:11
Breath Volume	1.125	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample #2	0.160	01:14
Breath Volume	1.144	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

---

Data Download Date/Time 07/18/2010 22:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, GREG, P  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BAKER, GREG  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: GODCHARLES

DATE/TIME OF TEST: 07/08/2010 02:57:16

---

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.078	02:54	
Air Blank	0.000	02:55	
Subject Sample #1		02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

---

Data Download Date/Time 07/18/2010 22:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENNEY, JASON, L  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: DENNEY, JASON  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 12:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:34

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/10/2010 00:38:25

---

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.080	00:34	
Air Blank	0.000	00:35	
Subject Sample #1		00:37	Subject Test Refused
Breath Volume	0.000	00:37	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

---

Data Download Date/Time 07/18/2010 22:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NGUYEN, TAI, T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: DENNEY, JASON  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BLEWER

DATE/TIME OF TEST: 07/11/2010 03:08:34

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Subject Sample #1	0.182	03:03
Breath Volume	2.968	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample #2	0.171	03:06
Breath Volume	2.843	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

---

Data Download Date/Time 07/18/2010 22:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK, KEITH, A  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: ROWDEN, SHEILA  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 07/13/2010 16:35:38

---

Diagnostics Check	OK	16:28
Air Blank	0.000	16:28
Control Test	0.079	16:29
Air Blank	0.000	16:29
Subject Sample #1	0.188	16:30
Breath Volume	1.691	16:30
Air Blank	0.000	16:31
Air Blank	0.000	16:33
Subject Sample #2	0.187	16:34
Breath Volume	2.113	16:34
Air Blank	0.000	16:34
Control Test	0.078	16:35
Air Blank	0.000	16:35
Diagnostics Check	OK	16:35

---

Data Download Date/Time 07/18/2010 22:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEFF, STEPHEN,  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WILSON, DALYN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: CRABTREE

DATE/TIME OF TEST: 07/17/2010 02:05:38

---

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample #1	0.233	02:00
Breath Volume	3.851	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample #2	0.233	02:04
Breath Volume	2.027	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

---

Data Download Date/Time 07/18/2010 22:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, GREG, P  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SKELTON, KEVIN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: SACCO

DATE/TIME OF TEST: 07/18/2010 02:42:32

---

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Subject Sample #1		02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

---

Data Download Date/Time 07/18/2010 22:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEFF, STEPHEN,  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BAKER, GREG  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: ARZBERGER

DATE/TIME OF TEST: 07/18/2010 22:42:56

---

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Subject Sample #1	0.264	22:38
Breath Volume	1.539	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample #2	0.261	22:41
Breath Volume	1.468	22:41
Air Blank	0.000	22:42
Control Test	0.078	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

---

Data Download Date/Time 08/06/2010 02:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POOL, HAROLD, M  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: POOL, HAROLD  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 07/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: HURLEY

DATE/TIME OF TEST: 07/22/2010 23:58:58

---

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.080	23:56	
Air Blank	0.000	23:57	
Subject Sample #1		23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.078	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

---

Data Download Date/Time 08/06/2010 02:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENNEY, JASON, L  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WHITE, TRAVIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: DODGION

DATE/TIME OF TEST: 07/24/2010 23:12:24

---

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Subject Sample #1	0.233	23:07
Breath Volume	1.812	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:09
Subject Sample #2	0.222	23:10
Breath Volume	1.718	23:10
Air Blank	0.000	23:11
Control Test	0.078	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

---

Data Download Date/Time 08/06/2010 02:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, GREG, P  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: NEFF, STEFF  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 07/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG930301  
DRY GAS EXPIRY DATE: 10/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 07/30/2010 22:22:02

---

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:19	
Control Test	0.078	22:19	
Air Blank	0.000	22:20	
Subject Sample #1		22:20	Subject Test Refused
Breath Volume	0.000	22:20	
Air Blank	0.000	22:21	
Control Test	0.077	22:21	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:22	

---

Data Download Date/Time 08/06/2010 02:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOMINGUEZ, OSCAR,  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: SKOGLUND, ROBERT  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG004701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SINFUENTES-ARRESOLA

DATE/TIME OF TEST: 07/10/2010 02:24:09

---

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.078	02:15	
Air Blank	0.000	02:15	
Subject Sample #1	0.271	02:16	
Breath Volume	1.671	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:18	
Subject Sample #2	0.245	02:19	
Breath Volume	2.210	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:21	
Subject Sample #3	0.267	02:22	
Breath Volume	3.500	02:22	
Air Blank	0.000	02:23	
Control Test	0.074	02:23	Control Outside Tolerance
Air Blank	0.000	02:24	

---

Data Download Date/Time 07/26/2010 01:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORAN, CHARLES, S  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: DOYLE, JUSTIN  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG004701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: BECOTTE

DATE/TIME OF TEST: 07/13/2010 20:12:31

---

Diagnostics Check	OK	20:03
Air Blank	0.000	20:04
Control Test	0.079	20:04
Air Blank	0.000	20:04
Subject Sample #1	0.184	20:06
Breath Volume	1.363	20:06
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample #2	0.165	20:10
Breath Volume	1.468	20:10
Air Blank	0.000	20:11
Control Test	0.078	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

---

Data Download Date/Time 07/26/2010 01:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLELLAN, FLETCHER, C  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: DEACETIS, DENNIS  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG004701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 07/24/2010 23:39:53

---

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:26	
Control Test	0.077	23:26	
Air Blank	0.000	23:27	
Subject Sample #1	0.196	23:29	
Breath Volume	1.468	23:29	
Air Blank	0.000	23:29	
Air Blank	0.000	23:31	
Subject Sample #2	0.177	23:34	Volume Not Met
Breath Volume	1.046	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample #3	0.201	23:38	
Breath Volume	1.707	23:38	
Air Blank	0.000	23:38	
Control Test	0.077	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

---

Data Download Date/Time 07/26/2010 01:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLELLAN, FLETCHER, C  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: DEACETIS, DENNIS  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 07/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG004701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PINKNEY

DATE/TIME OF TEST: 07/29/2010 02:04:44

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:58
Air Blank	0.000	01:58
Subject Sample #1	0.107	02:00
Breath Volume	1.507	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample #2	0.114	02:03
Breath Volume	1.613	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

---

Data Download Date/Time 08/31/2010 09:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEDERSEN, JOHN, P  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: TOOLE, SEAN  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 07/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG004701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 17:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: WILLAMAN

DATE/TIME OF TEST: 07/29/2010 18:14:16

---

Diagnostics Check	OK	18:11	
Air Blank	0.000	18:11	
Control Test	0.077	18:11	
Air Blank	0.000	18:12	
Subject Sample #1		18:12	Subject Test Refused
Breath Volume	0.000	18:12	
Air Blank	0.000	18:13	
Control Test	0.077	18:13	
Air Blank	0.000	18:14	
Diagnostics Check	OK	18:14	

---

Data Download Date/Time 08/31/2010 09:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDDY, PHILIP, J  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: BRANDES, MICHAEL  
ARRESTING OFFICER AGENCY: Sebastian PD  
LAST AGENCY INSPECTION: 07/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG004701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 10:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: FILAKOSKY

DATE/TIME OF TEST: 07/31/2010 12:46:36

---

Diagnostics Check	OK	12:40
Air Blank	0.000	12:40
Control Test	0.078	12:41
Air Blank	0.000	12:41
Subject Sample #1	0.000	12:42
Breath Volume	3.984	12:42
Air Blank	0.000	12:42
Air Blank	0.000	12:44
Subject Sample #2	0.000	12:45
Breath Volume	4.320	12:45
Air Blank	0.000	12:45
Control Test	0.078	12:46
Air Blank	0.000	12:46
Diagnostics Check	OK	12:46

---

Data Download Date/Time 08/31/2010 09:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEE, COREY, C

BREATH TEST OPERATOR AGENCY: FHP Troop L

ARRESTING OFFICER: LEE, COREY

ARRESTING OFFICER AGENCY: FHP Troop L

LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 685214

DRY GAS EXPIRY DATE: 10/28/2011

VIOLATION CODE: DUI

ARREST TIME: 23:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 07/29/2010 00:45:01

---

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.077	00:38
Air Blank	0.000	00:39
Subject Sample #1	0.273	00:39
Breath Volume	1.207	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample #2	0.264	00:43
Breath Volume	1.476	00:43
Air Blank	0.000	00:44
Control Test	0.076	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:45

---

Data Download Date/Time 07/30/2010 20:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TIMMONS, ALLAN, S

BREATH TEST OPERATOR AGENCY: FHP Troop K

ARRESTING OFFICER: TIMMONS, ALLAN

ARRESTING OFFICER AGENCY: FHP Troop K

LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 685214

DRY GAS EXPIRY DATE: 10/28/2011

VIOLATION CODE: DUI

ARREST TIME: 19:08

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: MORRELL

DATE/TIME OF TEST: 07/29/2010 20:17:05

---

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:14	
Control Test	0.077	20:14	
Air Blank	0.000	20:15	
Subject Sample #1		20:15	Subject Test Refused
Breath Volume	0.000	20:15	
Air Blank	0.000	20:16	
Control Test	0.077	20:16	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:17	

---

Data Download Date/Time 07/30/2010 20:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, COREY, C  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: SWANSON, M  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/30/2010 00:15:36

---

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:12	
Control Test	0.078	00:12	
Air Blank	0.000	00:13	
Subject Sample #1		00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:14	
Control Test	0.078	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

---

Data Download Date/Time 07/30/2010 20:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMBORGHINI, PETER, J  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LAMBORGHINI, PETER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WETHERILL

DATE/TIME OF TEST: 07/30/2010 01:12:41

---

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.078	01:06
Air Blank	0.000	01:07
Subject Sample #1	0.150	01:07
Breath Volume	2.015	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample #2	0.152	01:11
Breath Volume	2.109	01:11
Air Blank	0.000	01:11
Control Test	0.078	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

---

Data Download Date/Time 07/30/2010 20:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, COREY, C  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: LEE, COREY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: TIPTON

DATE/TIME OF TEST: 07/30/2010 03:52:38

---

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.077	03:46
Air Blank	0.000	03:47
Subject Sample #1	0.116	03:47
Breath Volume	1.652	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample #2	0.115	03:51
Breath Volume	1.597	03:51
Air Blank	0.000	03:51
Control Test	0.077	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time 07/30/2010 20:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SCHEETZ, MELLOW, D

BREATH TEST OPERATOR AGENCY: FHP Troop L

ARRESTING OFFICER: SCHEETZ, MELLOW

ARRESTING OFFICER AGENCY: FHP Troop L

LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 685214

DRY GAS EXPIRY DATE: 10/28/2011

VIOLATION CODE: DUI

ARREST TIME: 08:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:02

SUBJECT'S LAST NAME: LOWELL

DATE/TIME OF TEST: 07/31/2010 10:36:04

---

Diagnostics Check	OK	10:27
Air Blank	0.000	10:27
Control Test	0.077	10:28
Air Blank	0.000	10:28
Subject Sample #1	0.000	10:31
Breath Volume	1.253	10:31
Air Blank	0.000	10:31
Air Blank	0.000	10:33
Subject Sample #2	0.000	10:34
Breath Volume	2.312	10:34
Air Blank	0.000	10:35
Control Test	0.078	10:35
Air Blank	0.000	10:35
Diagnostics Check	OK	10:36

---

Data Download Date/Time 08/28/2010 12:57

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GROTHE

DATE/TIME OF TEST: 07/03/2010 03:09:13

---

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Subject Sample #1	0.181	03:04
Breath Volume	2.445	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample #2	0.162	03:07
Breath Volume	2.808	03:07
Air Blank	0.000	03:08
Control Test	0.078	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMBORGHINI, PETER, J  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LAMBORGHINI, PETER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 07/03/2010 22:41:22

---

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Subject Sample #1	0.272	22:33
Breath Volume	1.667	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample #2	0.240	22:36
Breath Volume	2.234	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample #3	0.263	22:39
Breath Volume	2.261	22:39
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMBORGHINI, PETER, J  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LAMBORGHINI, PETER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: VARGULISH

DATE/TIME OF TEST: 07/04/2010 03:24:43

---

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.078	03:17
Air Blank	0.000	03:17
Subject Sample #1	0.197	03:19
Breath Volume	1.113	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:21
Subject Sample #2	0.190	03:23
Breath Volume	1.148	03:23
Air Blank	0.000	03:23
Control Test	0.078	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUTY, J, K  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: CUTY, J  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 07/04/2010 03:36:40

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.077	03:29
Air Blank	0.000	03:30
Subject Sample #1	0.263	03:32
Breath Volume	2.320	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample #2	0.263	03:35
Breath Volume	2.328	03:35
Air Blank	0.000	03:35
Control Test	0.075	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: STEINKE

DATE/TIME OF TEST: 07/04/2010 21:05:34

---

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.079	21:00
Air Blank	0.000	21:00
Subject Sample #1	0.166	21:01
Breath Volume	3.933	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample #2	0.164	21:04
Breath Volume	2.882	21:04
Air Blank	0.000	21:04
Control Test	0.078	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CECKANOWICZ, SCOTT,  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: PETRY, DAVID  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 07/06/2010 16:37:06

---

Diagnostics Check	OK	16:29
Air Blank	0.000	16:30
Control Test	0.078	16:30
Air Blank	0.000	16:30
Subject Sample #1	0.171	16:32
Breath Volume	1.515	16:32
Air Blank	0.000	16:33
Air Blank	0.000	16:35
Subject Sample #2	0.172	16:35
Breath Volume	1.582	16:35
Air Blank	0.000	16:36
Control Test	0.077	16:36
Air Blank	0.000	16:37
Diagnostics Check	OK	16:37

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CECKANOWICZ, SCOTT,  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: SMITH, CHARLES  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: TOMMIE

DATE/TIME OF TEST: 07/07/2010 01:38:39

---

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.078	01:24	
Air Blank	0.000	01:24	
Subject Sample #1	0.105	01:27	Volume Not Met
Breath Volume	0.320	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample #2	0.122	01:32	
Breath Volume	1.535	01:32	
Air Blank	0.000	01:33	
Air Blank	0.000	01:35	
Subject Sample #3	0.128	01:37	
Breath Volume	1.156	01:37	
Air Blank	0.000	01:37	
Control Test	0.076	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, COREY, C  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: LEE, COREY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SMALLWOOD

DATE/TIME OF TEST: 07/08/2010 02:00:30

---

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.079	01:47	
Air Blank	0.000	01:47	
Subject Sample #1	0.230	01:48	
Breath Volume	1.234	01:48	
Air Blank	0.000	01:48	
Air Blank	0.000	01:50	
Subject Sample #2	0.191	01:53	Volume Not Met
Breath Volume	0.164	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample #3	0.224	01:58	
Breath Volume	1.394	01:58	
Air Blank	0.000	01:59	
Control Test	0.077	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GIGUERE

DATE/TIME OF TEST: 07/10/2010 01:32:10

---

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:27
Subject Sample #1	0.144	01:27
Breath Volume	1.218	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample #2	0.146	01:30
Breath Volume	1.382	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 12:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/10/2010 01:44:56

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:37
Subject Sample #1	0.097	01:39
Breath Volume	1.597	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample #2	0.093	01:43
Breath Volume	1.496	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FULSTON, TIMOTHY, J  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: FULSTON, TIMOTHY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 14:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: WENTZ

DATE/TIME OF TEST: 07/10/2010 15:32:27

---

Diagnostics Check	OK	15:18	
Air Blank	0.000	15:19	
Control Test	0.078	15:19	
Air Blank	0.000	15:20	
Subject Sample #1	0.150	15:23	Volume Not Met
Breath Volume	0.101	15:23	
Air Blank	0.000	15:24	
Air Blank	0.000	15:25	
Subject Sample #2	0.154	15:27	
Breath Volume	1.902	15:27	
Air Blank	0.000	15:27	
Air Blank	0.000	15:29	
Subject Sample #3	0.155	15:30	
Breath Volume	1.484	15:30	
Air Blank	0.000	15:31	
Control Test	0.075	15:31	
Air Blank	0.000	15:32	
Diagnostics Check	OK	15:32	

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: COLBURN

DATE/TIME OF TEST: 07/10/2010 23:38:09

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Subject Sample #1	0.180	23:32
Breath Volume	1.480	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample #2	0.172	23:36
Breath Volume	1.652	23:36
Air Blank	0.000	23:37
Control Test	0.078	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 07/11/2010 22:39:46

---

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.078	22:32
Air Blank	0.000	22:32
Subject Sample #1	0.107	22:35
Breath Volume	1.585	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample #2	0.095	22:38
Breath Volume	1.699	22:38
Air Blank	0.000	22:38
Control Test	0.077	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CECKANOWICZ, SCOTT,  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: CECKANOWICZ, SCOTT  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 07/11/2010 23:51:41

---

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.078	23:44
Air Blank	0.000	23:45
Subject Sample #1	0.187	23:46
Breath Volume	1.238	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample #2	0.171	23:50
Breath Volume	1.585	23:50
Air Blank	0.000	23:50
Control Test	0.076	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: BONVILLE

DATE/TIME OF TEST: 07/12/2010 23:27:40

---

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:15	
Control Test	0.078	23:16	
Air Blank	0.000	23:16	
Subject Sample #1	0.104	23:19	Volume Not Met
Breath Volume	0.214	23:19	
Air Blank	0.000	23:20	
Air Blank	0.000	23:22	
Subject Sample #2	0.115	23:23	
Breath Volume	1.957	23:23	
Air Blank	0.000	23:23	
Air Blank	0.000	23:25	
Subject Sample #3	0.118	23:26	
Breath Volume	1.488	23:26	
Air Blank	0.000	23:26	
Control Test	0.077	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CECKANOWICZ, SCOTT,  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: CECKANOWICZ, SCOTT  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: DOVER

DATE/TIME OF TEST: 07/12/2010 23:39:15

---

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.077	23:32
Air Blank	0.000	23:32
Subject Sample #1	0.165	23:34
Breath Volume	1.156	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample #2	0.165	23:37
Breath Volume	1.273	23:37
Air Blank	0.000	23:38
Control Test	0.077	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CECKANOWICZ, SCOTT,  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: SLATER, VERNON  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: ZAKLIN

DATE/TIME OF TEST: 07/12/2010 23:57:59

---

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.077	23:47	
Air Blank	0.000	23:47	
Subject Sample #1	0.108	23:50	Volume Not Met
Breath Volume	0.152	23:50	
Air Blank	0.000	23:51	
Air Blank	0.000	23:53	
Subject Sample #2	0.106	23:56	Volume Not Met
Breath Volume	0.957	23:56	
Air Blank	0.000	23:57	
Control Test	0.077	23:57	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILROY, DANIEL, J  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: GILROY, DANIEL  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: KOBYLARSKI

DATE/TIME OF TEST: 07/14/2010 04:26:57

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Subject Sample #1	0.248	04:22
Breath Volume	2.453	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample #2	0.246	04:25
Breath Volume	2.125	04:25
Air Blank	0.000	04:26
Control Test	0.076	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YUNAS, RYAN, S

BREATH TEST OPERATOR AGENCY: FHP Troop L

ARRESTING OFFICER: YUNAS, RYAN

ARRESTING OFFICER AGENCY: FHP Troop L

LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 685214

DRY GAS EXPIRY DATE: 10/28/2011

VIOLATION CODE: DUI

ARREST TIME: 02:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SODERQUIST

DATE/TIME OF TEST: 07/17/2010 03:18:00

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.078	03:12
Air Blank	0.000	03:12
Subject Sample #1	0.163	03:13
Breath Volume	2.957	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample #2	0.166	03:16
Breath Volume	2.132	03:16
Air Blank	0.000	03:17
Control Test	0.077	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:18

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: CASIERI

DATE/TIME OF TEST: 07/18/2010 02:45:59

---

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.079	02:42	
Air Blank	0.000	02:43	
Subject Sample #1		02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Control Test	0.078	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: EL

DATE/TIME OF TEST: 07/18/2010 23:23:33

---

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.078	23:18
Air Blank	0.000	23:18
Subject Sample #1	0.160	23:19
Breath Volume	3.667	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample #2	0.162	23:22
Breath Volume	3.511	23:22
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, COREY, C  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: LEE, COREY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WOJCIESZAK

DATE/TIME OF TEST: 07/22/2010 03:49:39

---

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.079	03:35	
Air Blank	0.000	03:35	
Subject Sample #1	0.176	03:38	Volume Not Met
Breath Volume	0.097	03:38	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample #2	0.171	03:43	
Breath Volume	1.460	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample #3	0.173	03:48	
Breath Volume	1.738	03:48	
Air Blank	0.000	03:48	
Control Test	0.077	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEE, COREY, C

BREATH TEST OPERATOR AGENCY: FHP Troop L

ARRESTING OFFICER: LEE, COREY

ARRESTING OFFICER AGENCY: FHP Troop L

LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 685214

DRY GAS EXPIRY DATE: 10/28/2011

VIOLATION CODE: DUI

ARREST TIME: 01:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: PALMA

DATE/TIME OF TEST: 07/24/2010 02:33:55

---

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.144	02:29
Breath Volume	2.445	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample #2	0.144	02:32
Breath Volume	2.734	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/24/2010 03:24:00

---

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.078	03:21	
Air Blank	0.000	03:22	
Subject Sample #1		03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Control Test	0.078	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:24	

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LAMBORGHINI, PETE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: VOLPE

DATE/TIME OF TEST: 07/24/2010 05:12:12

---

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.078	05:06
Air Blank	0.000	05:07
Subject Sample #1	0.109	05:07
Breath Volume	1.984	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample #2	0.108	05:10
Breath Volume	1.312	05:10
Air Blank	0.000	05:11
Control Test	0.078	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAYFORD, GREGG, R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HAYFORD, GREGG  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 07/25/2010 03:02:50

---

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:55
Subject Sample #1	0.241	02:56
Breath Volume	1.234	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.235	03:01
Breath Volume	1.558	03:01
Air Blank	0.000	03:01
Control Test	0.077	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARNER, TODD, A  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: WARNER, TODD  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 07:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 07/25/2010 07:53:35

---

Diagnostics Check	OK	07:45
Air Blank	0.000	07:46
Control Test	0.079	07:46
Air Blank	0.000	07:46
Subject Sample #1	0.219	07:48
Breath Volume	1.671	07:48
Air Blank	0.000	07:49
Air Blank	0.000	07:50
Subject Sample #2	0.220	07:51
Breath Volume	1.191	07:51
Air Blank	0.000	07:52
Control Test	0.076	07:52
Air Blank	0.000	07:53
Diagnostics Check	OK	07:53

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI, DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: BERTLEY

DATE/TIME OF TEST: 07/25/2010 21:58:08

---

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.078	21:52
Air Blank	0.000	21:52
Subject Sample #1	0.192	21:53
Breath Volume	4.085	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:55
Subject Sample #2	0.188	21:56
Breath Volume	3.074	21:56
Air Blank	0.000	21:57
Control Test	0.078	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GORDINEER, CHRISTOPHER, C

BREATH TEST OPERATOR AGENCY: Fort Pierce PD

ARRESTING OFFICER: DAVIS, CD

ARRESTING OFFICER AGENCY: Fort Pierce PD

LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 685214

DRY GAS EXPIRY DATE: 10/28/2011

VIOLATION CODE: DUI

ARREST TIME: 18:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: DASILVA

DATE/TIME OF TEST: 07/27/2010 19:28:18

---

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:12	
Control Test	0.079	19:12	
Air Blank	0.000	19:12	
Subject Sample #1	0.270	19:16	Volume Not Met
Breath Volume	0.179	19:16	
Air Blank	0.000	19:16	
Air Blank	0.000	19:18	
Subject Sample #2	0.346	19:20	
Breath Volume	1.269	19:20	
Air Blank	0.000	19:21	
Air Blank	0.000	19:23	
Subject Sample #3	0.360	19:26	
Breath Volume	1.488	19:26	
Air Blank	0.000	19:27	
Control Test	0.076	19:27	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, COREY, C  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: LEE, COREY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: FULWOOD

DATE/TIME OF TEST: 07/28/2010 04:16:52

---

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.079	04:06	
Air Blank	0.000	04:06	
Subject Sample #1	0.127	04:09	Volume Not Met
Breath Volume	0.359	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample #2	0.124	04:15	Volume Not Met
Breath Volume	0.035	04:15	
Air Blank	0.000	04:15	
Control Test	0.078	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

---

Data Download Date/Time 07/29/2010 13:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STARNES, KENNETH, E  
BREATH TEST OPERATOR AGENCY: Gulf County SO  
ARRESTING OFFICER: STARNES, KENNETH  
ARRESTING OFFICER AGENCY: Gulf County SO  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: CHISM

DATE/TIME OF TEST: 07/04/2010 01:30:33

---

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample #1	0.104	01:25
Breath Volume	1.722	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample #2	0.102	01:29
Breath Volume	2.011	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

---

Data Download Date/Time 07/27/2010 16:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, JEROME, C  
BREATH TEST OPERATOR AGENCY: Gulf County SO  
ARRESTING OFFICER: WILLIAMS, JEROME  
ARRESTING OFFICER AGENCY: Port St. Joe PD  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 14:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/24/2010 14:56:53

---

Diagnostics Check	OK	14:50
Air Blank	0.000	14:51
Control Test	0.078	14:51
Air Blank	0.000	14:51
Subject Sample #1	0.192	14:52
Breath Volume	3.082	14:52
Air Blank	0.000	14:53
Air Blank	0.000	14:55
Subject Sample #2	0.197	14:55
Breath Volume	2.160	14:55
Air Blank	0.000	14:56
Control Test	0.079	14:56
Air Blank	0.000	14:56
Diagnostics Check	OK	14:56

---

Data Download Date/Time 07/27/2010 16:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWSOME, JAMES, A  
BREATH TEST OPERATOR AGENCY: Gulf County SO  
ARRESTING OFFICER: OQUENDO, JOEL  
ARRESTING OFFICER AGENCY: Gulf County SO  
LAST AGENCY INSPECTION: 07/27/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/31/2010 06:15:07

---

Diagnostics Check	OK	06:07
Air Blank	0.000	06:08
Control Test	0.079	06:08
Air Blank	0.000	06:09
Subject Sample #1	0.171	06:10
Breath Volume	2.789	06:10
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample #2	0.165	06:13
Breath Volume	1.589	06:13
Air Blank	0.000	06:14
Control Test	0.079	06:14
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

---

Data Download Date/Time 08/31/2010 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: LONGBOAT KEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN, JOHN, P  
BREATH TEST OPERATOR AGENCY: Longboat Key PD  
ARRESTING OFFICER: FISHER, AMMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 635301G  
DRY GAS EXPIRY DATE: 12/21/2008  
VIOLATION CODE: DUI  
ARREST TIME: 20:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 07/17/2010 23:21:02

---

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.081	23:15
Air Blank	0.000	23:15
Subject Sample #1	0.128	23:16
Breath Volume	2.503	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample #2	0.138	23:19
Breath Volume	2.628	23:19
Air Blank	0.000	23:20
Control Test	0.081	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:21

---

Data Download Date/Time 07/21/2010 11:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: LONGBOAT KEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MISLYAN, STEPHEN, A  
BREATH TEST OPERATOR AGENCY: Longboat Key PD  
ARRESTING OFFICER: MISLYAN, STEPHEN  
ARRESTING OFFICER AGENCY: Longboat Key PD  
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: CFCSRA  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BECKHAM

DATE/TIME OF TEST: 07/21/2010 01:14:10

---

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:08
Air Blank	0.000	01:08
Subject Sample #1	0.319	01:09
Breath Volume	3.394	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:11
Subject Sample #2	0.336	01:12
Breath Volume	3.218	01:12
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

---

Data Download Date/Time 07/21/2010 11:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKEE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: BEARD

DATE/TIME OF TEST: 07/01/2010 03:58:31

---

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Subject Sample #1	0.183	03:54
Breath Volume	2.839	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample #2	0.178	03:56
Breath Volume	3.015	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

---

Data Download Date/Time 07/02/2010 11:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LEWIS, I  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: TOTH

DATE/TIME OF TEST: 07/02/2010 02:48:35

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.082	02:42
Air Blank	0.000	02:43
Subject Sample #1	0.217	02:43
Breath Volume	1.234	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample #2	0.201	02:46
Breath Volume	1.156	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

---

Data Download Date/Time 07/02/2010 11:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORRELL, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FIDDELKE

DATE/TIME OF TEST: 07/02/2010 04:47:08

---

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Subject Sample #1	0.111	04:42
Breath Volume	2.187	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample #2	0.115	04:45
Breath Volume	1.976	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

---

Data Download Date/Time 07/02/2010 11:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBSON, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/03/2010 01:34:35

---

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Subject Sample #1		01:33	Subject Test Refused
Breath Volume	0.152	01:33	
Air Blank	0.000	01:33	
Control Test	0.081	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILLAN-CINTRON, E  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 07/03/2010 02:20:04

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Subject Sample #1	0.153	02:15
Breath Volume	1.644	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample #2	0.164	02:18
Breath Volume	2.136	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:20

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RAMER, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 07/04/2010 04:04:49

---

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample #1	0.186	03:59
Breath Volume	3.257	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample #2	0.190	04:03
Breath Volume	2.343	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PARKER, S  
ARRESTING OFFICER AGENCY: U of So FL Campus PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/04/2010 04:29:08

---

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:23
Subject Sample #1	0.160	04:24
Breath Volume	1.441	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:26
Subject Sample #2	0.170	04:27
Breath Volume	1.152	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ACZMAREK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MOUNCE

DATE/TIME OF TEST: 07/04/2010 21:48:40

---

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.081	21:43
Air Blank	0.000	21:43
Subject Sample #1	0.000	21:43
Breath Volume	1.136	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample #2	0.000	21:47
Breath Volume	2.152	21:47
Air Blank	0.000	21:47
Control Test	0.082	21:48
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: FAURA

DATE/TIME OF TEST: 07/04/2010 23:13:05

---

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:07
Subject Sample #1	0.208	23:08
Breath Volume	1.222	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:10
Subject Sample #2	0.212	23:11
Breath Volume	2.109	23:11
Air Blank	0.000	23:12
Control Test	0.081	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:13

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCCALLA, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: SAYROO

DATE/TIME OF TEST: 07/05/2010 05:24:35

---

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.080	05:13	
Air Blank	0.000	05:14	
Subject Sample #1	0.265	05:14	
Breath Volume	1.566	05:14	
Air Blank	0.000	05:15	
Air Blank	0.000	05:17	
Subject Sample #2		05:20	No Sample Provided
Breath Volume	0.000	05:20	
Air Blank	0.000	05:21	
Air Blank	0.000	05:23	
Subject Sample #3		05:23	Subject Test Refused
Breath Volume	0.000	05:23	
Air Blank	0.000	05:23	
Control Test	0.081	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: BROWDER

DATE/TIME OF TEST: 07/06/2010 20:39:06

---

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.081	20:33
Air Blank	0.000	20:34
Subject Sample #1	0.000	20:34
Breath Volume	2.742	20:34
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample #2	0.000	20:37
Breath Volume	2.031	20:37
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORRIS, T  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 07/07/2010 01:00:05

---

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.082	00:54
Air Blank	0.000	00:55
Subject Sample #1	0.161	00:55
Breath Volume	2.378	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample #2	0.156	00:58
Breath Volume	2.988	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BORDEN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: EMMANUEL JR

DATE/TIME OF TEST: 07/07/2010 04:04:36

---

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample #1	0.125	03:59
Breath Volume	1.140	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample #2	0.122	04:03
Breath Volume	1.269	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KACZMAREK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: LANCASTER

DATE/TIME OF TEST: 07/07/2010 22:55:34

---

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.081	22:50
Air Blank	0.000	22:50
Subject Sample #1	0.210	22:51
Breath Volume	1.894	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample #2	0.213	22:54
Breath Volume	1.757	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORET, F  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: FREAHAH

DATE/TIME OF TEST: 07/08/2010 00:31:21

---

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Subject Sample #1	0.000	00:27
Breath Volume	3.257	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample #2	0.000	00:29
Breath Volume	2.089	00:29
Air Blank	0.000	00:30
Control Test	0.081	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAVIS, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LACKEY II

DATE/TIME OF TEST: 07/08/2010 00:41:40

---

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:36
Subject Sample #1	0.000	00:37
Breath Volume	1.289	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample #2	0.000	00:40
Breath Volume	1.953	00:40
Air Blank	0.000	00:40
Control Test	0.082	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TESTER, P  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CASEBOLT

DATE/TIME OF TEST: 07/08/2010 04:25:04

---

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.081	04:19
Air Blank	0.000	04:19
Subject Sample #1	0.145	04:20
Breath Volume	2.082	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample #2	0.141	04:23
Breath Volume	1.679	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:25

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORET, F  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: CANALES

DATE/TIME OF TEST: 07/08/2010 04:34:01

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:28
Air Blank	0.000	04:28
Subject Sample #1	0.059	04:29
Breath Volume	1.433	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:31
Subject Sample #2	0.057	04:32
Breath Volume	1.242	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:34

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORRELL, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/08/2010 04:53:20

---

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:39	
Subject Sample #1	0.137	04:42	Volume Not Met
Breath Volume	0.148	04:42	
Air Blank	0.000	04:42	
Air Blank	0.000	04:44	
Subject Sample #2	0.166	04:46	
Breath Volume	1.152	04:46	
Air Blank	0.000	04:46	
Air Blank	0.000	04:48	
Subject Sample #3	0.149	04:51	Volume Not Met
Breath Volume	0.132	04:51	
Air Blank	0.000	04:52	
Control Test	0.077	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORRELL, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/08/2010 05:11:54

---

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:56	
Control Test	0.077	04:56	
Air Blank	0.000	04:57	
Subject Sample #1	0.134	05:00	Volume Not Met
Breath Volume	0.972	05:00	
Air Blank	0.000	05:00	
Air Blank	0.000	05:02	
Subject Sample #2	0.154	05:04	
Breath Volume	1.660	05:04	
Air Blank	0.000	05:05	
Air Blank	0.000	05:07	
Subject Sample #3	0.156	05:10	Volume Not Met
Breath Volume	0.136	05:10	
Air Blank	0.000	05:10	
Control Test	0.078	05:11	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:11	

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORRELL, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LEDESMA

DATE/TIME OF TEST: 07/09/2010 03:22:02

---

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:16
Subject Sample #1	0.189	03:17
Breath Volume	2.796	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample #2	0.185	03:20
Breath Volume	2.523	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KLEEMAN, R  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 07/10/2010 02:28:55

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Subject Sample #1	0.000	02:24
Breath Volume	3.140	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample #2	0.000	02:27
Breath Volume	2.906	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SAVO

DATE/TIME OF TEST: 07/10/2010 03:45:33

---

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample #1	0.086	03:41
Breath Volume	2.601	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample #2	0.080	03:44
Breath Volume	2.289	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCLAIN, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 07/10/2010 04:38:22

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Subject Sample #1	0.165	04:29
Breath Volume	1.285	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample #2	0.144	04:33
Breath Volume	1.476	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample #3	0.150	04:36
Breath Volume	2.882	04:36
Air Blank	0.000	04:37
Control Test	0.078	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SHERERTZ, R  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: BUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: GEIL

DATE/TIME OF TEST: 07/10/2010 20:00:10

---

Diagnostics Check	OK	19:53
Air Blank	0.000	19:53
Control Test	0.081	19:53
Air Blank	0.000	19:54
Subject Sample #1	0.053	19:55
Breath Volume	1.476	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample #2	0.052	19:58
Breath Volume	1.664	19:58
Air Blank	0.000	19:59
Control Test	0.081	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: INSIGNARES, P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: OCAMPO-VAZQUEZ

DATE/TIME OF TEST: 07/10/2010 23:20:44

---

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.082	23:15
Air Blank	0.000	23:15
Subject Sample #1	0.196	23:16
Breath Volume	1.269	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample #2	0.202	23:19
Breath Volume	1.746	23:19
Air Blank	0.000	23:19
Control Test	0.081	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KENNEDY, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: YOUNGMAN

DATE/TIME OF TEST: 07/11/2010 01:25:10

---

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:20
Subject Sample #1	0.000	01:20
Breath Volume	1.972	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample #2	0.000	01:23
Breath Volume	2.011	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: JUDGE

DATE/TIME OF TEST: 07/11/2010 02:43:55

---

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample #1	0.146	02:38
Breath Volume	1.949	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample #2	0.141	02:42
Breath Volume	1.929	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MATAS, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 07/11/2010 03:27:45

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample #1	0.093	03:23
Breath Volume	1.375	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample #2	0.090	03:26
Breath Volume	1.492	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/11/2010 04:18:01

---

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:13
Subject Sample #1	0.206	04:13
Breath Volume	2.808	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample #2	0.214	04:16
Breath Volume	3.523	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:18

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVARES, N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: TREVITHICK

DATE/TIME OF TEST: 07/11/2010 05:05:27

---

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Subject Sample #1	0.214	05:00
Breath Volume	1.500	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample #2	0.225	05:03
Breath Volume	1.503	05:03
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: INSIGNARES, P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: BALDASAR

DATE/TIME OF TEST: 07/11/2010 21:16:05

---

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:11
Subject Sample #1	0.213	21:11
Breath Volume	2.574	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample #2	0.203	21:14
Breath Volume	1.664	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:16

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCLAIN, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 07/11/2010 23:23:46

---

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Subject Sample #1	0.211	23:19
Breath Volume	2.363	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample #2	0.216	23:22
Breath Volume	1.453	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 07/12/2010 04:11:16

---

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample #1	0.160	04:05
Breath Volume	1.214	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.158	04:09
Breath Volume	1.117	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCLAIN, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: CAJIGAS

DATE/TIME OF TEST: 07/12/2010 04:18:47

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Subject Sample #1	0.178	04:14
Breath Volume	1.160	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample #2	0.192	04:17
Breath Volume	1.382	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KENNEDY, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: BEAZLEY

DATE/TIME OF TEST: 07/12/2010 05:01:55

---

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:56	
Control Test	0.080	04:56	
Air Blank	0.000	04:57	
Subject Sample #1	0.209	04:57	
Breath Volume	1.246	04:57	
Air Blank	0.000	04:58	
Air Blank	0.000	05:00	
Subject Sample #2		05:00	Subject Test Refused
Breath Volume	0.000	05:00	
Air Blank	0.000	05:01	
Control Test	0.080	05:01	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:01	

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FOSTER, E  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: TRIPP

DATE/TIME OF TEST: 07/12/2010 23:51:41

---

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.081	23:48	
Air Blank	0.000	23:48	
Subject Sample #1		23:50	Subject Test Refused
Breath Volume	0.285	23:50	
Air Blank	0.000	23:50	
Control Test	0.081	23:51	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KACZMAREK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: BROWNE

DATE/TIME OF TEST: 07/13/2010 01:05:55

---

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.081	01:00
Air Blank	0.000	01:00
Subject Sample #1	0.132	01:01
Breath Volume	1.406	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:03
Subject Sample #2	0.144	01:04
Breath Volume	1.453	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WHITEHEAD, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 07/13/2010 02:55:06

---

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Subject Sample #1	0.079	02:50
Breath Volume	1.847	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample #2	0.073	02:53
Breath Volume	1.488	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

---

Data Download Date/Time 07/13/2010 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CALDERON, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 11:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:02

SUBJECT'S LAST NAME: MINTON

DATE/TIME OF TEST: 07/13/2010 12:29:09

---

Diagnostics Check	OK	12:22
Air Blank	0.000	12:22
Control Test	0.083	12:22
Air Blank	0.000	12:23
Subject Sample #1	0.000	12:24
Breath Volume	1.378	12:24
Air Blank	0.000	12:25
Air Blank	0.000	12:27
Subject Sample #2	0.000	12:27
Breath Volume	1.464	12:27
Air Blank	0.000	12:28
Control Test	0.081	12:28
Air Blank	0.000	12:29
Diagnostics Check	OK	12:29

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SANCHEZ, R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/14/2010 03:36:47

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.082	03:30
Air Blank	0.000	03:31
Subject Sample #1	0.020	03:32
Breath Volume	1.523	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample #2	0.016	03:35
Breath Volume	1.832	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBSON, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/16/2010 00:54:28

---

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.082	00:49
Air Blank	0.000	00:49
Subject Sample #1	0.012	00:50
Breath Volume	2.207	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample #2	0.011	00:53
Breath Volume	2.398	00:53
Air Blank	0.000	00:53
Control Test	0.082	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHANAN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: GILSTRAP

DATE/TIME OF TEST: 07/16/2010 02:39:28

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.082	02:33
Air Blank	0.000	02:34
Subject Sample #1	0.218	02:34
Breath Volume	3.925	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample #2	0.220	02:37
Breath Volume	3.019	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BORDEN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/16/2010 04:58:21

---

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:50
Subject Sample #1	0.203	04:53
Breath Volume	1.265	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample #2	0.206	04:56
Breath Volume	1.183	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORREY, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: BENOIT

DATE/TIME OF TEST: 07/16/2010 19:50:22

---

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.081	19:40
Air Blank	0.000	19:41
Subject Sample #1	0.229	19:42
Breath Volume	1.460	19:42
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample #2	0.203	19:45
Breath Volume	1.191	19:45
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample #3	0.222	19:48
Breath Volume	2.156	19:48
Air Blank	0.000	19:49
Control Test	0.077	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORET, F  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: HICKS JR

DATE/TIME OF TEST: 07/16/2010 23:53:26

---

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:48
Subject Sample #1	0.193	23:48
Breath Volume	1.832	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample #2	0.195	23:51
Breath Volume	2.359	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILLAN-CINTRON, E  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: AGUADO-MANZO

DATE/TIME OF TEST: 07/17/2010 01:02:57

---

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.082	00:57
Air Blank	0.000	00:57
Subject Sample #1	0.213	00:58
Breath Volume	2.519	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:00
Subject Sample #2	0.220	01:01
Breath Volume	2.390	01:01
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMS, G  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: RIDDICK

DATE/TIME OF TEST: 07/17/2010 02:58:42

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Subject Sample #1	0.135	02:54
Breath Volume	3.195	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.141	02:57
Breath Volume	2.679	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORET, F  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: ECHEVERRI

DATE/TIME OF TEST: 07/17/2010 04:27:52

---

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.081	04:22
Air Blank	0.000	04:22
Subject Sample #1	0.080	04:23
Breath Volume	2.804	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.080	04:26
Breath Volume	2.820	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHANAN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/17/2010 04:56:28

---

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Subject Sample #1	0.238	04:51
Breath Volume	1.527	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample #2	0.246	04:54
Breath Volume	1.832	04:54
Air Blank	0.000	04:55
Control Test	0.082	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUSWELL, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: OJEDA

DATE/TIME OF TEST: 07/18/2010 02:12:37

---

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.082	02:07
Air Blank	0.000	02:07
Subject Sample #1	0.117	02:08
Breath Volume	1.675	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample #2	0.117	02:11
Breath Volume	1.585	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORREY, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: COPPENS

DATE/TIME OF TEST: 07/18/2010 03:50:30

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample #1	0.131	03:44
Breath Volume	2.011	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample #2	0.137	03:48
Breath Volume	1.343	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORET, F  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: FOISSET

DATE/TIME OF TEST: 07/18/2010 04:02:13

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:57
Subject Sample #1	0.096	03:57
Breath Volume	2.566	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample #2	0.098	04:00
Breath Volume	2.460	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMS, G  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/18/2010 04:27:24

---

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.081	04:22
Air Blank	0.000	04:22
Subject Sample #1	0.079	04:23
Breath Volume	1.863	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.080	04:25
Breath Volume	1.753	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WHITEHEAD, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/18/2010 04:38:21

---

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Subject Sample #1	0.244	04:33
Breath Volume	1.710	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample #2	0.246	04:36
Breath Volume	1.218	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FOSTER, E  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: SANDERSON

DATE/TIME OF TEST: 07/18/2010 19:49:13

---

Diagnostics Check	OK	19:46	
Air Blank	0.000	19:46	
Control Test	0.082	19:47	
Air Blank	0.000	19:47	
Subject Sample #1		19:47	Subject Test Refused
Breath Volume	0.000	19:47	
Air Blank	0.000	19:48	
Control Test	0.081	19:48	
Air Blank	0.000	19:49	
Diagnostics Check	OK	19:49	

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RAMER, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 07/19/2010 02:02:50

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.082	01:56
Air Blank	0.000	01:57
Subject Sample #1	0.000	01:58
Breath Volume	1.312	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample #2	0.000	02:01
Breath Volume	3.277	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCLAIN, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 07/20/2010 00:39:29

---

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.082	00:34
Air Blank	0.000	00:34
Subject Sample #1	0.255	00:35
Breath Volume	1.761	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample #2	0.254	00:37
Breath Volume	2.234	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

---

Data Download Date/Time 07/20/2010 11:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCLAIN, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 18:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: LOVELL-MONTGOMERY

DATE/TIME OF TEST: 07/20/2010 19:49:09

---

Diagnostics Check	OK	19:42
Air Blank	0.000	19:43
Control Test	0.081	19:43
Air Blank	0.000	19:44
Subject Sample #1	0.000	19:44
Breath Volume	2.976	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample #2	0.000	19:47
Breath Volume	3.156	19:47
Air Blank	0.000	19:48
Control Test	0.081	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIVERA, E  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/21/2010 00:39:43

---

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:34
Subject Sample #1	0.000	00:35
Breath Volume	1.921	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample #2	0.000	00:38
Breath Volume	1.734	00:38
Air Blank	0.000	00:38
Control Test	0.080	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ELSEDOODI, A  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: TORRES-SIERRA

DATE/TIME OF TEST: 07/21/2010 01:06:41

---

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample #1	0.185	01:02
Breath Volume	2.984	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample #2	0.183	01:05
Breath Volume	3.027	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: TILLMAN

DATE/TIME OF TEST: 07/21/2010 02:52:29

---

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample #1	0.000	02:48
Breath Volume	1.519	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample #2	0.000	02:51
Breath Volume	1.351	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARBOSA, F  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 07/21/2010 04:22:49

---

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:16
Subject Sample #1	0.089	04:18
Breath Volume	1.488	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample #2	0.091	04:21
Breath Volume	1.531	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAVIS, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: NOVELLI

DATE/TIME OF TEST: 07/22/2010 23:04:00

---

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.081	22:58
Air Blank	0.000	22:58
Subject Sample #1	0.000	22:59
Breath Volume	1.796	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample #2	0.000	23:02
Breath Volume	2.234	23:02
Air Blank	0.000	23:03
Control Test	0.081	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:04

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KACZMAREK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/23/2010 03:43:24

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:37
Subject Sample #1	0.196	03:38
Breath Volume	2.695	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample #2	0.195	03:41
Breath Volume	2.695	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRUZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: HERNANDEZ-ASTELLO

DATE/TIME OF TEST: 07/24/2010 00:29:38

---

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.081	00:27	
Air Blank	0.000	00:27	
Subject Sample #1		00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.081	00:29	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCMURPHY, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LARRY

DATE/TIME OF TEST: 07/24/2010 01:31:10

---

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:25
Air Blank	0.000	01:25
Subject Sample #1	0.227	01:26
Breath Volume	1.156	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample #2	0.237	01:29
Breath Volume	2.250	01:29
Air Blank	0.000	01:30
Control Test	0.081	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBSON, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: REDDEN

DATE/TIME OF TEST: 07/24/2010 03:24:47

---

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Subject Sample #1	0.249	03:17	Volume Not Met
Breath Volume	0.339	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:19	
Subject Sample #2	0.254	03:23	Volume Not Met
Breath Volume	0.191	03:23	
Air Blank	0.000	03:23	
Control Test	0.077	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KLEEMAN, R  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: CRANDON

DATE/TIME OF TEST: 07/24/2010 05:02:31

---

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.078	04:56
Air Blank	0.000	04:57
Subject Sample #1	0.219	04:57
Breath Volume	1.449	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample #2	0.236	05:00
Breath Volume	2.101	05:00
Air Blank	0.000	05:01
Control Test	0.078	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, REBECCA, J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BORDEN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 06:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: BREA

DATE/TIME OF TEST: 07/24/2010 07:02:32

---

Diagnostics Check	OK	06:55
Air Blank	0.000	06:55
Control Test	0.080	06:56
Air Blank	0.000	06:56
Subject Sample #1	0.190	06:58
Breath Volume	2.339	06:58
Air Blank	0.000	06:58
Air Blank	0.000	07:00
Subject Sample #2	0.192	07:00
Breath Volume	2.289	07:00
Air Blank	0.000	07:01
Control Test	0.079	07:01
Air Blank	0.000	07:02
Diagnostics Check	OK	07:02

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCMURPHY, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/24/2010 21:38:08

---

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Subject Sample #1	0.173	21:33
Breath Volume	3.292	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample #2	0.175	21:36
Breath Volume	3.085	21:36
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 07/25/2010 02:06:58

---

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:01
Air Blank	0.000	02:01
Subject Sample #1	0.192	02:02
Breath Volume	2.203	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample #2	0.192	02:05
Breath Volume	2.253	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VALLEJO, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: CIPRIANO

DATE/TIME OF TEST: 07/25/2010 04:06:54

---

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample #1	0.172	04:02
Breath Volume	2.949	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample #2	0.169	04:05
Breath Volume	2.496	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ROBERT, B  
ARRESTING OFFICER AGENCY: Macdill Air Force Ba  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: SHIPLEY

DATE/TIME OF TEST: 07/25/2010 04:39:28

---

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Subject Sample #1	0.081	04:34
Breath Volume	2.660	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample #2	0.080	04:37
Breath Volume	2.628	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 07/25/2010 04:55:57

---

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:48
Subject Sample #1	0.217	04:50
Breath Volume	2.210	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:52
Subject Sample #2	0.215	04:54
Breath Volume	1.816	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: UTRERA-VIERA

DATE/TIME OF TEST: 07/25/2010 05:00:27

---

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.078	04:57	
Air Blank	0.000	04:58	
Subject Sample #1		04:59	Subject Test Refused
Breath Volume	0.000	04:59	
Air Blank	0.000	04:59	
Control Test	0.079	04:59	
Air Blank	0.000	05:00	
Diagnostics Check	OK	05:00	

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: SEGURA-HERNANDEZ

DATE/TIME OF TEST: 07/25/2010 23:02:15

---

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:52
Air Blank	0.000	22:52
Subject Sample #1	0.190	22:55
Breath Volume	1.816	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:57
Subject Sample #2	0.189	23:00
Breath Volume	1.765	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: STENSON

DATE/TIME OF TEST: 07/26/2010 02:18:52

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Subject Sample #1	0.163	02:14
Breath Volume	1.101	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample #2	0.170	02:17
Breath Volume	2.335	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CUSTALOW

DATE/TIME OF TEST: 07/26/2010 04:18:36

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample #1	0.177	04:14
Breath Volume	2.609	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample #2	0.184	04:17
Breath Volume	2.558	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: HOUGH

DATE/TIME OF TEST: 07/26/2010 04:36:21

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:31
Air Blank	0.000	04:31
Subject Sample #1	0.103	04:31
Breath Volume	2.933	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample #2	0.103	04:34
Breath Volume	3.566	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCLAIN, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: ARTHUR

DATE/TIME OF TEST: 07/26/2010 05:21:12

---

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:16
Subject Sample #1	0.000	05:16
Breath Volume	2.664	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample #2	0.000	05:19
Breath Volume	3.101	05:19
Air Blank	0.000	05:20
Control Test	0.081	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVERAS, N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: SNIPES JR

DATE/TIME OF TEST: 07/26/2010 20:55:49

---

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.081	20:49
Air Blank	0.000	20:49
Subject Sample #1	0.212	20:50
Breath Volume	1.812	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample #2	0.212	20:54
Breath Volume	1.894	20:54
Air Blank	0.000	20:54
Control Test	0.080	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVERAS, N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/27/2010 04:09:11

---

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Subject Sample #1	0.167	04:04
Breath Volume	3.718	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.173	04:07
Breath Volume	2.496	04:07
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RENTAS

DATE/TIME OF TEST: 07/28/2010 04:31:10

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.082	04:26
Air Blank	0.000	04:26
Subject Sample #1	0.000	04:27
Breath Volume	2.527	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample #2	0.000	04:29
Breath Volume	2.488	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORRELL, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: BLACKBURN

DATE/TIME OF TEST: 07/29/2010 01:46:48

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:37
Air Blank	0.000	01:37
Subject Sample #1	0.180	01:39
Breath Volume	1.261	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample #2	0.202	01:42
Breath Volume	2.203	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample #3	0.203	01:45
Breath Volume	1.683	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MATAS, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 07/29/2010 03:35:15

---

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:29
Subject Sample #1	0.164	03:30
Breath Volume	1.566	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample #2	0.164	03:33
Breath Volume	1.519	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCLAIN, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SCHADE

DATE/TIME OF TEST: 07/29/2010 03:57:01

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Subject Sample #1	0.207	03:52
Breath Volume	2.292	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample #2	0.211	03:55
Breath Volume	1.804	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, REBECCA, J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SHEPPARD, R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 06:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 07/29/2010 07:47:23

---

Diagnostics Check	OK	07:41
Air Blank	0.000	07:41
Control Test	0.080	07:42
Air Blank	0.000	07:42
Subject Sample #1	0.196	07:42
Breath Volume	1.753	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:45
Subject Sample #2	0.197	07:45
Breath Volume	1.828	07:45
Air Blank	0.000	07:46
Control Test	0.080	07:46
Air Blank	0.000	07:47
Diagnostics Check	OK	07:47

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: LOCK

DATE/TIME OF TEST: 07/29/2010 23:36:29

---

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:31
Subject Sample #1	0.116	23:32
Breath Volume	1.269	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample #2	0.123	23:34
Breath Volume	1.234	23:34
Air Blank	0.000	23:35
Control Test	0.079	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARBOSA, F  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 07/30/2010 03:09:57

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:04
Subject Sample #1	0.077	03:05
Breath Volume	3.214	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:07
Subject Sample #2	0.076	03:08
Breath Volume	3.187	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KLEEMAN, R  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: GINTHER

DATE/TIME OF TEST: 07/30/2010 03:36:07

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:30
Subject Sample #1	0.215	03:31
Breath Volume	1.402	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample #2	0.222	03:34
Breath Volume	1.375	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FERNANDEZ, R  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SANKHI

DATE/TIME OF TEST: 07/30/2010 04:12:59

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Subject Sample #1	0.156	04:07
Breath Volume	1.191	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample #2	0.152	04:11
Breath Volume	2.218	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CAMPBELL, J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: VON DEM HAGEN

DATE/TIME OF TEST: 07/30/2010 04:50:27

---

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:45
Air Blank	0.000	04:45
Subject Sample #1	0.000	04:46
Breath Volume	1.695	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample #2	0.000	04:49
Breath Volume	1.742	04:49
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 07/30/2010 05:48:12

---

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:38	
Control Test	0.081	05:39	
Air Blank	0.000	05:39	
Subject Sample #1	0.147	05:42	Volume Not Met
Breath Volume	0.148	05:42	
Air Blank	0.000	05:43	
Air Blank	0.000	05:45	
Subject Sample #2		05:46	Subject Test Refused
Breath Volume	0.019	05:46	
Air Blank	0.000	05:47	
Control Test	0.078	05:47	
Air Blank	0.000	05:48	
Diagnostics Check	OK	05:48	

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 07/30/2010 05:53:51

---

Diagnostics Check	OK	05:50	
Air Blank	0.000	05:51	
Control Test	0.079	05:51	
Air Blank	0.000	05:52	
Subject Sample #1		05:52	Subject Test Refused
Breath Volume	0.000	05:52	
Air Blank	0.000	05:52	
Control Test	0.079	05:53	
Air Blank	0.000	05:53	
Diagnostics Check	OK	05:53	

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUSWELL, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 07/31/2010 00:03:19

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Subject Sample #1	0.000	23:59
Breath Volume	3.191	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample #2	0.000	00:01
Breath Volume	2.382	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: STEAD, K  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 747256  
DRY GAS EXPIRY DATE: 03/25/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MGUIRE

DATE/TIME OF TEST: 07/31/2010 04:00:53

---

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample #1		03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.081	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

---

Data Download Date/Time 08/06/2010 13:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANUS, GLEN, A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: GANUS, GLEN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/14/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 16:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/04/2010 17:57:58

---

Diagnostics Check	OK	17:51
Air Blank	0.000	17:52
Control Test	0.078	17:52
Air Blank	0.000	17:53
Subject Sample #1	0.215	17:53
Breath Volume	1.605	17:53
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample #2	0.214	17:56
Breath Volume	2.460	17:56
Air Blank	0.000	17:57
Control Test	0.079	17:57
Air Blank	0.000	17:57
Diagnostics Check	OK	17:57

---

Data Download Date/Time 07/20/2010 10:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANUS, GLEN, A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: GANUS, GLEN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/14/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/12/2010 20:52:13

---

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.080	20:46
Air Blank	0.000	20:46
Subject Sample #1	0.041	20:47
Breath Volume	2.011	20:47
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample #2	0.041	20:50
Breath Volume	1.246	20:50
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

---

Data Download Date/Time 07/20/2010 10:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUGHES, ALFRED, L  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DOWNING, RICHARD  
ARRESTING OFFICER AGENCY: Dixie County SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 17:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: SECREST

DATE/TIME OF TEST: 07/26/2010 18:42:05

---

Diagnostics Check	OK	18:37	
Air Blank	0.000	18:37	
Control Test	0.079	18:38	
Air Blank	0.000	18:38	
Subject Sample #1	0.226	18:39	
Breath Volume	2.332	18:39	
Air Blank		18:40	Improper Sample
Air Blank	0.000	18:41	
Control Test	0.080	18:41	
Air Blank	0.000	18:42	
Diagnostics Check	OK	18:42	

---

Data Download Date/Time 08/26/2010 10:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUGHES, ALFRED, L  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DOWNING, RICHARD  
ARRESTING OFFICER AGENCY: Dixie County SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 17:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: SECREST

DATE/TIME OF TEST: 07/26/2010 18:50:15

---

Diagnostics Check	OK	18:44
Air Blank	0.000	18:44
Control Test	0.079	18:44
Air Blank	0.000	18:45
Subject Sample #1	0.234	18:45
Breath Volume	2.375	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample #2	0.233	18:48
Breath Volume	1.812	18:48
Air Blank	0.000	18:49
Control Test	0.079	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

---

Data Download Date/Time 08/26/2010 10:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GANUS, GLEN, A

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: GANUS, GLEN

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 828901G

DRY GAS EXPIRY DATE: 10/21/2010

VIOLATION CODE: DUI

ARREST TIME: 00:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 07/30/2010 01:06:34

---

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:05	
Subject Sample #1		01:06	Sequence Aborted
Breath Volume	0.000	01:06	
Air Blank	0.000	01:06	

---

Data Download Date/Time 08/26/2010 10:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURPHY, TIMOTHY, L  
BREATH TEST OPERATOR AGENCY: Perry PD  
ARRESTING OFFICER: MURPHY, TIMOTHY  
ARRESTING OFFICER AGENCY: Perry PD  
LAST AGENCY INSPECTION: 06/27/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 07/09/2010 01:15:59

---

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.078	01:09
Air Blank	0.000	01:09
Subject Sample #1	0.173	01:11
Breath Volume	1.570	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample #2	0.178	01:14
Breath Volume	1.500	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

---

Data Download Date/Time 07/26/2010 02:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUNDY, WILLIAM, B  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: LUNDY, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/27/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: DORMEYER

DATE/TIME OF TEST: 07/10/2010 23:15:16

---

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Subject Sample #1	0.000	23:09
Breath Volume	1.410	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample #2	0.000	23:13
Breath Volume	1.375	23:13
Air Blank	0.000	23:14
Control Test	0.078	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

---

Data Download Date/Time 07/26/2010 02:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS, MANUEL, J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SMYRNIOS, MANUEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/27/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 07/16/2010 23:17:56

---

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:12
Subject Sample #1	0.226	23:13
Breath Volume	2.480	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample #2	0.228	23:16
Breath Volume	2.031	23:16
Air Blank	0.000	23:17
Control Test	0.076	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

---

Data Download Date/Time 07/26/2010 02:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEAUFORD, JIM, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: BEAUFORD, JIM  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/27/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/17/2010 00:43:34

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:37
Subject Sample #1	0.260	00:38
Breath Volume	2.246	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample #2	0.252	00:42
Breath Volume	2.484	00:42
Air Blank	0.000	00:42
Control Test	0.078	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

---

Data Download Date/Time 07/26/2010 02:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS, MANUEL, J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SMYRNIOS, MANUEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/27/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: LARA

DATE/TIME OF TEST: 07/17/2010 23:14:06

---

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Subject Sample #1	0.123	23:09
Breath Volume	1.843	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample #2	0.120	23:12
Breath Volume	1.824	23:12
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

---

Data Download Date/Time 07/26/2010 02:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS, MANUEL, L  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SMYRNIOS, MANUEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/27/2010

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/24/2010 19:49:19

---

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:47	
Control Test	0.077	19:48	
Air Blank	0.000	19:48	
		19:48	Improper Sample
Air Blank	0.000	19:49	

---

Data Download Date/Time 07/26/2010 02:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS, MANUEL, J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SMYRNIOS, MANUEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/27/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 828901G  
DRY GAS EXPIRY DATE: 10/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/24/2010 19:58:38

---

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.078	19:52
Air Blank	0.000	19:53
Subject Sample #1	0.171	19:53
Breath Volume	2.019	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample #2	0.166	19:57
Breath Volume	2.308	19:57
Air Blank	0.000	19:57
Control Test	0.078	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

---

Data Download Date/Time 07/26/2010 02:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS, MANUEL, J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SMYRNIOS, MANUEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG005701  
DRY GAS EXPIRY DATE: 02/26/2012  
VIOLATION CODE: DUI  
ARREST TIME: 15:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 07/30/2010 16:22:05

---

Diagnostics Check	OK	16:15
Air Blank	0.000	16:16
Control Test	0.078	16:16
Air Blank	0.000	16:17
Subject Sample #1	0.208	16:17
Breath Volume	1.968	16:17
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample #2	0.215	16:20
Breath Volume	2.632	16:20
Air Blank	0.000	16:21
Control Test	0.078	16:21
Air Blank	0.000	16:21
Diagnostics Check	OK	16:22

---

Data Download Date/Time 08/30/2010 16:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASH, JAMES, C  
BREATH TEST OPERATOR AGENCY: Perry PD  
ARRESTING OFFICER: BURFORD, LANCE  
ARRESTING OFFICER AGENCY: Taylor County SO  
LAST AGENCY INSPECTION: 07/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG005701  
DRY GAS EXPIRY DATE: 02/26/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SPEES

DATE/TIME OF TEST: 07/31/2010 05:42:14

---

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.079	05:35
Air Blank	0.000	05:35
Subject Sample #1	0.134	05:36
Breath Volume	3.777	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample #2	0.141	05:40
Breath Volume	4.273	05:40
Air Blank	0.000	05:41
Control Test	0.079	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

---

Data Download Date/Time 08/30/2010 16:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: JOHNSON, S  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 17:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: MCSORLEY

DATE/TIME OF TEST: 07/28/2010 18:36:23

---

Diagnostics Check	OK	18:29
Air Blank	0.000	18:30
Control Test	0.080	18:30
Air Blank	0.000	18:31
Subject Sample #1	0.273	18:31
Breath Volume	1.734	18:31
Air Blank	0.000	18:32
Air Blank	0.000	18:34
Subject Sample #2	0.260	18:34
Breath Volume	1.890	18:34
Air Blank	0.000	18:35
Control Test	0.078	18:35
Air Blank	0.000	18:36
Diagnostics Check	OK	18:36

---

Data Download Date/Time 07/28/2010 19:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, A,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: WRIGHT, D  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DICKMAN

DATE/TIME OF TEST: 07/11/2010 04:13:31

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Subject Sample #1	0.226	04:05
Breath Volume	2.058	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.202	04:08
Breath Volume	1.281	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample #3	0.202	04:11
Breath Volume	1.292	04:11
Air Blank	0.000	04:12
Control Test	0.076	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, GARY, A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER, GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SOLER MENENDEZ

DATE/TIME OF TEST: 07/14/2010 00:08:51

---

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Subject Sample #1	0.075	00:04
Breath Volume	1.839	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:06
Subject Sample #2	0.069	00:07
Breath Volume	1.289	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: WRIGHT, M  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: CUBERO

DATE/TIME OF TEST: 07/14/2010 04:04:50

---

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:57
Subject Sample #1	0.077	03:59
Breath Volume	1.113	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample #2	0.075	04:03
Breath Volume	1.507	04:03
Air Blank	0.000	04:03
Control Test	0.077	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, A,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: GALLUP, C  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: DECESARE

DATE/TIME OF TEST: 07/15/2010 22:42:40

---

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Subject Sample #1	0.000	22:37
Breath Volume	1.585	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample #2	0.000	22:41
Breath Volume	2.164	22:41
Air Blank	0.000	22:41
Control Test	0.081	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROADERICK, CPL, C  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: HOLLAND, R  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 07/16/2010 01:59:29

---

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:54	
Control Test	0.079	01:54	
Air Blank	0.000	01:54	
Subject Sample #1	0.162	01:58	Volume Not Met
Breath Volume	0.300	01:58	
Air Blank		01:58	Purge Fail
Air Blank	0.000	01:59	

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROADERICK, CPL, C  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: HOLLAND, R  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 07/16/2010 02:05:22

---

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.078	02:01	
Air Blank	0.000	02:02	
Subject Sample #1		02:03	Subject Test Refused
Breath Volume	0.089	02:03	
Air Blank	0.000	02:04	
Control Test	0.078	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSHER, KENNETH, R  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BATTS, C  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: KUEGLER

DATE/TIME OF TEST: 07/16/2010 23:01:23

---

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Subject Sample #1	0.141	22:56
Breath Volume	1.738	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample #2	0.137	22:59
Breath Volume	2.281	22:59
Air Blank	0.000	23:00
Control Test	0.078	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRITCHETT, A, M  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: SIMPSON, S  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: HANNAR

DATE/TIME OF TEST: 07/17/2010 00:30:49

---

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Subject Sample #1	0.121	00:25
Breath Volume	1.609	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample #2	0.117	00:29
Breath Volume	2.316	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSHER, KENNETH, R  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER, D  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 07/17/2010 02:04:04

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample #1	0.014	01:59
Breath Volume	2.152	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample #2	0.012	02:02
Breath Volume	2.527	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:04

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, GARY, A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER, GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: COTTRELL

DATE/TIME OF TEST: 07/17/2010 02:21:18

---

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Subject Sample #1	0.112	02:16
Breath Volume	2.550	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.109	02:19
Breath Volume	1.906	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSHER, KENNETH, R  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PINDER, D  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 07/17/2010 20:25:32

---

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.078	20:19
Air Blank	0.000	20:20
Subject Sample #1	0.169	20:20
Breath Volume	2.417	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample #2	0.179	20:23
Breath Volume	4.425	20:23
Air Blank	0.000	20:24
Control Test	0.078	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, GARY, A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER, GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: DEHOYOS

DATE/TIME OF TEST: 07/17/2010 21:48:43

---

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.079	21:43
Air Blank	0.000	21:43
Subject Sample #1	0.099	21:44
Breath Volume	1.726	21:44
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample #2	0.101	21:47
Breath Volume	1.222	21:47
Air Blank	0.000	21:47
Control Test	0.079	21:48
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, GARY, A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER, GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: FALL

DATE/TIME OF TEST: 07/18/2010 00:59:14

---

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:53
Subject Sample #1	0.129	00:54
Breath Volume	1.996	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:56
Subject Sample #2	0.129	00:57
Breath Volume	1.597	00:57
Air Blank	0.000	00:58
Control Test	0.078	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROADERICK, CPL, C  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: KWOCZKA, G  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: GENTILE

DATE/TIME OF TEST: 07/20/2010 02:10:43

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:05
Subject Sample #1	0.148	02:06
Breath Volume	1.488	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample #2	0.149	02:09
Breath Volume	1.488	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: THOMPSON, L  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 07/21/2010 22:59:55

---

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.078	22:56	
Air Blank	0.000	22:57	
Subject Sample #1	0.190	22:58	
Breath Volume	1.273	22:58	
Air Blank		22:59	Purge Fail
Air Blank		22:59	RFI Detect

---

Data Download Date/Time 07/25/2010 18:49

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: THOMPSON, L  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 07/21/2010 23:13:28

---

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.077	23:03
Air Blank	0.000	23:04
Subject Sample #1	0.186	23:05
Breath Volume	1.441	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample #2	0.211	23:08
Breath Volume	1.843	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample #3	0.208	23:11
Breath Volume	2.921	23:11
Air Blank	0.000	23:12
Control Test	0.078	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PINDER, D  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/22/2010 01:48:16

---

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:45	
Control Test	0.079	01:45	
Air Blank	0.000	01:46	
Subject Sample #1		01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: DOODY, P  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: VANSANT

DATE/TIME OF TEST: 07/22/2010 04:03:20

---

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:57
Subject Sample #1	0.182	03:58
Breath Volume	2.289	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample #2	0.184	04:01
Breath Volume	2.562	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DEWEESE, D  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 07/22/2010 23:15:48

---

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Subject Sample #1	0.164	23:11
Breath Volume	2.085	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample #2	0.173	23:14
Breath Volume	1.765	23:14
Air Blank	0.000	23:14
Control Test	0.080	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: STROW, D.J.  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BULLOCK

DATE/TIME OF TEST: 07/22/2010 23:30:56

---

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.078	23:28	
Air Blank	0.000	23:28	
Subject Sample #1	0.193	23:29	
Breath Volume	2.863	23:29	
Air Blank		23:30	Purge Fail
Air Blank	0.000	23:30	

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: STROW, D.J.  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BULLOCK

DATE/TIME OF TEST: 07/22/2010 23:41:07

---

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
Control Test	0.078	23:33	
Air Blank	0.000	23:33	
Subject Sample #1		23:36	No Sample Provided
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample #2		23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:40	
Control Test	0.078	23:40	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ, J,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MCNEILL, C  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: EARLY

DATE/TIME OF TEST: 07/23/2010 15:45:03

---

Diagnostics Check	OK	15:44	
Air Blank		15:44	RFI Detect
Air Blank	0.000	15:45	

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ, J,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MCNEILL, C  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: EARLY

DATE/TIME OF TEST: 07/23/2010 15:49:55

---

Diagnostics Check	OK	15:47	
Air Blank	0.000	15:47	
Control Test	0.079	15:47	
Air Blank	0.000	15:48	
Subject Sample #1		15:48	Subject Test Refused
Breath Volume	0.000	15:48	
Air Blank	0.000	15:49	
Control Test	0.079	15:49	
Air Blank	0.000	15:49	
Diagnostics Check	OK	15:49	

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, A,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: JOEDICKE, W  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: FLAGG

DATE/TIME OF TEST: 07/24/2010 00:52:46

---

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:43	
Control Test	0.079	00:43	
Air Blank	0.000	00:43	
Subject Sample #1	0.129	00:47	Volume Not Met
Breath Volume	0.296	00:47	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample #2		00:51	Subject Test Refused
Breath Volume	0.062	00:51	
Air Blank	0.000	00:51	
Control Test	0.078	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, A,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: LEE, A  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FAVORS

DATE/TIME OF TEST: 07/24/2010 03:44:42

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample #1	0.114	03:39
Breath Volume	2.175	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample #2	0.110	03:43
Breath Volume	1.828	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOWERS, D, W  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PORTER, S  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: LAIZ

DATE/TIME OF TEST: 07/25/2010 00:29:43

---

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample #1	0.124	00:25
Breath Volume	1.964	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample #2	0.123	00:28
Breath Volume	1.347	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROADERICK, CPL, C  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: SOLLAZZO, C  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/25/2010 01:11:25

---

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Subject Sample #1	0.277	01:06
Breath Volume	1.863	01:06
Air Blank	0.000	01:08
Air Blank	0.000	01:09
Subject Sample #2	0.284	01:09
Breath Volume	1.292	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOWERS, D, W  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BLACKBURN, L  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BRINKMANN

DATE/TIME OF TEST: 07/25/2010 02:13:09

---

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Subject Sample #1	0.000	02:08
Breath Volume	1.367	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample #2	0.000	02:11
Breath Volume	1.453	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

---

Data Download Date/Time 07/25/2010 18:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, GARY, A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER, GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 07/28/2010 01:16:52

---

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample #1	0.078	01:12
Breath Volume	1.917	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:14
Subject Sample #2	0.076	01:15
Breath Volume	1.882	01:15
Air Blank	0.000	01:15
Control Test	0.078	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: DOODY, P  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ECKERDT

DATE/TIME OF TEST: 07/28/2010 02:59:23

---

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.079	02:57	
Air Blank	0.000	02:57	
Subject Sample #1		02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Control Test	0.079	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIX, BRYAN, D  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: THORESEN, E  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 16:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/28/2010 17:35:29

---

Diagnostics Check	OK	17:32	
Air Blank	0.000	17:32	
Control Test	0.079	17:33	
Air Blank	0.000	17:33	
Subject Sample #1		17:34	Subject Test Refused
Breath Volume	0.000	17:34	
Air Blank	0.000	17:34	
Control Test	0.079	17:34	
Air Blank	0.000	17:35	
Diagnostics Check	OK	17:35	

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DOBOSIEWICZ, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GLISSON

DATE/TIME OF TEST: 07/29/2010 03:58:58

---

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Subject Sample #1	0.116	03:53
Breath Volume	1.886	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample #2	0.115	03:57
Breath Volume	2.468	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUNTER, ROBERT, M

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: MCNEILL, C

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581914

DRY GAS EXPIRY DATE: 02/26/2011

VIOLATION CODE: DUI

ARREST TIME: 22:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/30/2010 00:28:43

---

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Subject Sample #1	0.201	00:24
Breath Volume	4.238	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample #2	0.205	00:27
Breath Volume	4.175	00:27
Air Blank	0.000	00:27
Control Test	0.079	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: VALDEZ, Y  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: REINOSO-CASTENADA

DATE/TIME OF TEST: 07/31/2010 02:44:12

---

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:37
Subject Sample #1	0.168	02:39
Breath Volume	1.414	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample #2	0.177	02:42
Breath Volume	1.367	02:42
Air Blank	0.000	02:43
Control Test	0.075	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: JOEDICKE, W  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 07/31/2010 03:19:33

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:13
Subject Sample #1	0.151	03:14
Breath Volume	1.296	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample #2	0.153	03:17
Breath Volume	1.804	03:17
Air Blank	0.000	03:18
Control Test	0.077	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHALEY, DAVID, G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: FROST, C  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: JARMAN

DATE/TIME OF TEST: 07/31/2010 04:00:47

---

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.078	03:56	
Air Blank	0.000	03:57	
Subject Sample #1		03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.079	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSHER, KENNETH, R  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BAXLEY, D  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: DINO

DATE/TIME OF TEST: 07/31/2010 04:35:43

---

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:33	
Control Test	0.079	04:33	
Air Blank	0.000	04:34	
Subject Sample #1		04:34	Subject Test Refused
Breath Volume	0.000	04:34	
Air Blank	0.000	04:34	
Control Test	0.078	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, KRISTIN, K  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: COHEN, JESSICA  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 581914  
DRY GAS EXPIRY DATE: 02/26/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: CUBBAGE

DATE/TIME OF TEST: 07/31/2010 06:07:59

---

Diagnostics Check	OK	05:58
Air Blank	0.000	05:59
Control Test	0.078	05:59
Air Blank	0.000	05:59
Subject Sample #1	0.153	06:01
Breath Volume	2.625	06:01
Air Blank	0.000	06:02
Air Blank	0.000	06:04
Subject Sample #2	0.133	06:06
Breath Volume	1.949	06:06
Air Blank	0.000	06:07
Control Test	0.077	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEALE, BRADLEY, L  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: SWARBRICK, M  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/31/2010 16:14:41

---

Diagnostics Check	OK	16:08
Air Blank	0.000	16:08
Control Test	0.079	16:08
Air Blank	0.000	16:09
Subject Sample #1	0.283	16:10
Breath Volume	3.148	16:10
Air Blank	0.000	16:10
Air Blank	0.000	16:12
Subject Sample #2	0.284	16:13
Breath Volume	3.281	16:13
Air Blank	0.000	16:13
Control Test	0.077	16:14
Air Blank	0.000	16:14
Diagnostics Check	OK	16:14

---

Data Download Date/Time 08/21/2010 18:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: O'DONNELL, T  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SUGGS

DATE/TIME OF TEST: 07/03/2010 23:51:51

---

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:49	
Control Test	0.083	23:49	
Air Blank	0.000	23:50	
Subject Sample #1		23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:50	
Control Test	0.084	23:51	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: O'DONNELL, T  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 21:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 07/04/2010 22:47:51

---

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.084	22:45	
Air Blank	0.000	22:46	
Subject Sample #1		22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Control Test	0.084	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: MULVERHILL, G  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/06/2010 23:40:27

---

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.082	23:38	
Air Blank	0.000	23:38	
Subject Sample #1		23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:39	
Control Test	0.082	23:39	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: O'DONNELL, T  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/07/2010 20:01:59

---

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.083	19:56
Air Blank	0.000	19:56
Subject Sample #1	0.025	19:57
Breath Volume	1.699	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample #2	0.026	20:00
Breath Volume	3.042	20:00
Air Blank	0.000	20:01
Control Test	0.083	20:01
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CRIMA, G  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: DEHART

DATE/TIME OF TEST: 07/07/2010 21:42:19

---

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.084	21:39	
Air Blank	0.000	21:40	
Subject Sample #1		21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:41	
Control Test	0.083	21:41	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: O'DONNELL, T  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: FISCHER

DATE/TIME OF TEST: 07/09/2010 03:07:27

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.082	03:01
Air Blank	0.000	03:02
Subject Sample #1	0.105	03:03
Breath Volume	2.484	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample #2	0.104	03:05
Breath Volume	3.160	03:05
Air Blank	0.000	03:06
Control Test	0.082	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: COOPER, W  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 16:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: MESIC

DATE/TIME OF TEST: 07/09/2010 17:51:15

---

Diagnostics Check	OK	17:48	
Air Blank	0.000	17:48	
Control Test	0.083	17:49	
Air Blank	0.000	17:49	
Subject Sample #1		17:49	Subject Test Refused
Breath Volume	0.000	17:49	
Air Blank	0.000	17:50	
Control Test	0.083	17:50	
Air Blank	0.000	17:51	
Diagnostics Check	OK	17:51	

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: HARTZIG, R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 07/10/2010 00:13:10

---

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.083	00:06
Air Blank	0.000	00:06
Subject Sample #1	0.148	00:08
Breath Volume	2.144	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample #2	0.149	00:11
Breath Volume	1.695	00:11
Air Blank	0.000	00:12
Control Test	0.083	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: HARTZIG, R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TICE

DATE/TIME OF TEST: 07/10/2010 03:42:28

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.083	03:36
Air Blank	0.000	03:37
Subject Sample #1	0.167	03:38
Breath Volume	2.867	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample #2	0.155	03:40
Breath Volume	2.136	03:40
Air Blank	0.000	03:41
Control Test	0.082	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: HARTZIG, R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BENGGE

DATE/TIME OF TEST: 07/10/2010 03:58:13

---

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.083	03:52
Air Blank	0.000	03:52
Subject Sample #1	0.269	03:53
Breath Volume	2.609	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample #2	0.270	03:56
Breath Volume	3.152	03:56
Air Blank	0.000	03:57
Control Test	0.082	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: MULVERHILL, G  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SWANSTON

DATE/TIME OF TEST: 07/12/2010 01:32:48

---

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.083	01:27
Air Blank	0.000	01:27
Subject Sample #1	0.000	01:28
Breath Volume	3.605	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample #2	0.000	01:31
Breath Volume	3.488	01:31
Air Blank	0.000	01:31
Control Test	0.083	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: PETROVICH, D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 15:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: NESBITT

DATE/TIME OF TEST: 07/12/2010 16:59:10

---

Diagnostics Check	OK	16:55	
Air Blank	0.000	16:55	
Control Test	0.083	16:56	
Air Blank	0.000	16:56	
Subject Sample #1		16:57	Subject Test Refused
Breath Volume	0.000	16:57	
Air Blank	0.000	16:58	
Control Test	0.083	16:58	
Air Blank	0.000	16:59	
Diagnostics Check	OK	16:59	

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: PALESE, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: HERING

DATE/TIME OF TEST: 07/17/2010 00:32:39

---

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.083	00:27
Air Blank	0.000	00:27
Subject Sample #1	0.103	00:28
Breath Volume	1.800	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample #2	0.097	00:31
Breath Volume	1.535	00:31
Air Blank	0.000	00:31
Control Test	0.083	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: HARTZIG, BOBBY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MISHOE

DATE/TIME OF TEST: 07/17/2010 01:05:26

---

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.083	01:00
Air Blank	0.000	01:00
Subject Sample #1	0.201	01:01
Breath Volume	1.589	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample #2	0.203	01:03
Breath Volume	1.757	01:03
Air Blank	0.000	01:04
Control Test	0.083	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: HARTZIG, BOBBY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/17/2010 02:38:29

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.084	02:32
Air Blank	0.000	02:33
Subject Sample #1	0.123	02:34
Breath Volume	1.632	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample #2	0.127	02:36
Breath Volume	2.003	02:36
Air Blank	0.000	02:37
Control Test	0.083	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M

BREATH TEST OPERATOR AGENCY: Hernando County SO

ARRESTING OFFICER: RODRIGUEZ, JOEL

ARRESTING OFFICER AGENCY: Brooksville PD

LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 11:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 11:58

SUBJECT'S LAST NAME: HOUSE

DATE/TIME OF TEST: 07/17/2010 12:34:31

---

Diagnostics Check	OK	12:28
Air Blank	0.000	12:29
Control Test	0.083	12:29
Air Blank	0.000	12:29
Subject Sample #1	0.000	12:30
Breath Volume	1.718	12:30
Air Blank	0.000	12:30
Air Blank	0.000	12:32
Subject Sample #2	0.000	12:33
Breath Volume	1.816	12:33
Air Blank	0.000	12:33
Control Test	0.084	12:33
Air Blank	0.000	12:34
Diagnostics Check	OK	12:34

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: O'DONNELL, T  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LYSAKOWSKI

DATE/TIME OF TEST: 07/18/2010 01:25:48

---

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.085	01:19
Air Blank	0.000	01:20
Subject Sample #1	0.220	01:21
Breath Volume	2.917	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample #2	0.225	01:24
Breath Volume	2.917	01:24
Air Blank	0.000	01:24
Control Test	0.083	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: O'DONNELL, T  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/18/2010 03:56:48

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.083	03:50
Air Blank	0.000	03:51
Subject Sample #1	0.276	03:52
Breath Volume	2.648	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample #2	0.263	03:55
Breath Volume	2.273	03:55
Air Blank	0.000	03:55
Control Test	0.083	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: O'DONNELL, T  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: BEHRENS

DATE/TIME OF TEST: 07/18/2010 20:14:22

---

Diagnostics Check	OK	20:07
Air Blank	0.000	20:08
Control Test	0.083	20:08
Air Blank	0.000	20:09
Subject Sample #1	0.239	20:10
Breath Volume	2.058	20:10
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample #2	0.226	20:12
Breath Volume	1.738	20:12
Air Blank	0.000	20:13
Control Test	0.083	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: HOAGLAN, BRANDON  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/22/2010 05:44:48

---

Diagnostics Check	OK	05:38
Air Blank	0.000	05:39
Control Test	0.084	05:39
Air Blank	0.000	05:39
Subject Sample #1	0.000	05:40
Breath Volume	2.371	05:40
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample #2	0.000	05:43
Breath Volume	2.535	05:43
Air Blank	0.000	05:43
Control Test	0.084	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: HOAGLAN, BRANDON  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: COMBS

DATE/TIME OF TEST: 07/23/2010 04:19:12

---

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.084	04:13
Air Blank	0.000	04:14
Subject Sample #1	0.168	04:14
Breath Volume	1.769	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.150	04:17
Breath Volume	1.601	04:17
Air Blank	0.000	04:18
Control Test	0.084	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: MULVERHILL, G  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GREY

DATE/TIME OF TEST: 07/24/2010 03:59:19

---

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.083	03:50
Air Blank	0.000	03:51
Subject Sample #1	0.291	03:51
Breath Volume	3.210	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample #2	0.253	03:54
Breath Volume	2.144	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample #3	0.288	03:57
Breath Volume	2.687	03:57
Air Blank	0.000	03:58
Control Test	0.083	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: MULVERHILL, G  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/24/2010 04:16:49

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.084	04:10
Air Blank	0.000	04:10
Subject Sample #1	0.000	04:12
Breath Volume	1.566	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample #2	0.000	04:15
Breath Volume	1.480	04:15
Air Blank	0.000	04:15
Control Test	0.083	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: SERRANO, E  
ARRESTING OFFICER AGENCY: Brooksville PD  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 07/25/2010 00:44:46

---

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.084	00:38
Air Blank	0.000	00:39
Subject Sample #1	0.143	00:40
Breath Volume	2.542	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample #2	0.131	00:43
Breath Volume	2.515	00:43
Air Blank	0.000	00:43
Control Test	0.084	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: JOHNSON, S  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MAHATCEK

DATE/TIME OF TEST: 07/25/2010 02:20:39

---

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.083	02:14
Air Blank	0.000	02:15
Subject Sample #1	0.181	02:16
Breath Volume	3.250	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample #2	0.191	02:19
Breath Volume	2.605	02:19
Air Blank	0.000	02:19
Control Test	0.083	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: HARTZIG, R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: BARBOUR

DATE/TIME OF TEST: 07/26/2010 00:18:59

---

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.083	00:12
Air Blank	0.000	00:13
Subject Sample #1	0.191	00:14
Breath Volume	2.445	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample #2	0.195	00:17
Breath Volume	1.933	00:17
Air Blank	0.000	00:18
Control Test	0.082	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: O'DONNELL, TYLER  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 07/27/2010 03:04:07

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.083	02:58
Air Blank	0.000	02:59
Subject Sample #1	0.182	02:59
Breath Volume	2.304	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample #2	0.187	03:02
Breath Volume	1.925	03:02
Air Blank	0.000	03:03
Control Test	0.083	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: FREDRICKSEN, S  
ARRESTING OFFICER AGENCY: Brooksville PD  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: PATRIZI

DATE/TIME OF TEST: 07/27/2010 03:26:29

---

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.082	03:23	
Air Blank	0.000	03:24	
Subject Sample #1		03:24	Slope Not Met
Breath Volume	0.835	03:24	
Air Blank	0.000	03:25	
Control Test	0.082	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: FREDRICKSEN, S  
ARRESTING OFFICER AGENCY: Brooksville PD  
LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: PATRIZI

DATE/TIME OF TEST: 07/27/2010 03:33:51

---

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.082	03:28
Air Blank	0.000	03:29
Subject Sample #1	0.138	03:29
Breath Volume	1.687	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample #2	0.131	03:32
Breath Volume	1.863	03:32
Air Blank	0.000	03:32
Control Test	0.082	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M

BREATH TEST OPERATOR AGENCY: Hernando County SO

ARRESTING OFFICER: LAMIA, S

ARRESTING OFFICER AGENCY: Hernando County SO

LAST AGENCY INSPECTION: 06/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 23:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 07/28/2010 00:18:20

---

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:15	
Control Test	0.082	00:16	
Air Blank	0.000	00:16	
Subject Sample #1		00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:17	
Control Test	0.082	00:17	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

---

Data Download Date/Time 07/28/2010 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREIFENBERGER, FRANCES, M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: JOHNSON, S  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: FANTER

DATE/TIME OF TEST: 07/28/2010 23:24:32

---

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.083	23:22	
Air Blank	0.000	23:22	
Subject Sample #1		23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:23	
Control Test	0.083	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

---

Data Download Date/Time 08/27/2010 23:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: JOHNSON, S  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: LOISELLE

DATE/TIME OF TEST: 07/30/2010 01:57:28

---

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Subject Sample #1	0.146	01:52
Breath Volume	3.042	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample #2	0.145	01:55
Breath Volume	2.730	01:55
Air Blank	0.000	01:56
Control Test	0.082	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

---

Data Download Date/Time 08/27/2010 23:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: ORIGON, L  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 16:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: BAKLE

DATE/TIME OF TEST: 07/30/2010 17:02:29

---

Diagnostics Check	OK	16:56
Air Blank	0.000	16:56
Control Test	0.082	16:56
Air Blank	0.000	16:57
Subject Sample #1	0.000	16:58
Breath Volume	2.468	16:58
Air Blank	0.000	16:58
Air Blank	0.000	17:00
Subject Sample #2	0.000	17:01
Breath Volume	1.996	17:01
Air Blank	0.000	17:01
Control Test	0.082	17:01
Air Blank	0.000	17:02
Diagnostics Check	OK	17:02

---

Data Download Date/Time 08/27/2010 23:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: O'DONNELL, T  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 07/31/2010 03:14:21

---

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.083	03:08
Air Blank	0.000	03:09
Subject Sample #1	0.145	03:09
Breath Volume	2.171	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample #2	0.147	03:12
Breath Volume	2.308	03:12
Air Blank	0.000	03:13
Control Test	0.083	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

---

Data Download Date/Time 08/27/2010 23:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWREY, J, D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: O'DONNELL, T  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 825601I  
DRY GAS EXPIRY DATE: 10/28/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: BON

DATE/TIME OF TEST: 07/31/2010 20:16:55

---

Diagnostics Check	OK	20:10
Air Blank	0.000	20:10
Control Test	0.083	20:11
Air Blank	0.000	20:11
Subject Sample #1	0.114	20:12
Breath Volume	1.941	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:14
Subject Sample #2	0.116	20:15
Breath Volume	2.546	20:15
Air Blank	0.000	20:16
Control Test	0.084	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

---

Data Download Date/Time 08/27/2010 23:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, BRENDA, R  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: RAPPLEYEA, BILL  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 704221  
DRY GAS EXPIRY DATE: 12/11/2011  
VIOLATION CODE: DUI  
ARREST TIME: 12:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: SCHUSTER

DATE/TIME OF TEST: 07/01/2010 01:47:56

---

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Subject Sample #1	0.213	01:43
Breath Volume	2.343	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample #2	0.209	01:46
Breath Volume	1.679	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

---

Data Download Date/Time 07/25/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WIGGLESWORTH, JOE, B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: PELTON, D  
ARRESTING OFFICER AGENCY: Wildwood PD  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 704221  
DRY GAS EXPIRY DATE: 12/11/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 07/03/2010 00:46:25

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:41	
Subject Sample #1	0.229	00:44	Volume Not Met
Breath Volume	1.125	00:44	
Air Blank		00:45	Purge Fail
Air Blank	0.000	00:46	

---

Data Download Date/Time 07/25/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WIGGLESWORTH, JOE, B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: PELTON, D  
ARRESTING OFFICER AGENCY: Wildwood PD  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 704221  
DRY GAS EXPIRY DATE: 12/11/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 07/03/2010 00:55:54

---

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Subject Sample #1	0.244	00:50
Breath Volume	1.441	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample #2	0.236	00:54
Breath Volume	1.398	00:54
Air Blank	0.000	00:54
Control Test	0.081	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

---

Data Download Date/Time 07/25/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTLER, E, Y  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WILLIAMS, J  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 704221  
DRY GAS EXPIRY DATE: 12/11/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/07/2010 19:54:24

---

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.079	19:48
Air Blank	0.000	19:48
Subject Sample #1	0.000	19:49
Breath Volume	1.953	19:49
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample #2	0.000	19:53
Breath Volume	1.714	19:53
Air Blank	0.000	19:53
Control Test	0.081	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

---

Data Download Date/Time 07/25/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLANKER, KENNETH, B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: FERNANDEZ, LEWIS  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 704221  
DRY GAS EXPIRY DATE: 12/11/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 07/14/2010 20:35:41

---

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.080	20:29
Air Blank	0.000	20:30
Subject Sample #1	0.250	20:31
Breath Volume	1.746	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample #2	0.244	20:33
Breath Volume	1.738	20:33
Air Blank	0.000	20:34
Control Test	0.080	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

---

Data Download Date/Time 07/25/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLANKER, KENNETH, B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: KATICH, JOESPH  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 704221  
DRY GAS EXPIRY DATE: 12/11/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/14/2010 23:32:53

---

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:30	
Control Test	0.080	23:30	
Air Blank	0.000	23:31	
Subject Sample #1		23:31	Subject Test Refused
Breath Volume	0.000	23:31	
Air Blank	0.000	23:31	
Control Test	0.080	23:32	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

---

Data Download Date/Time 07/25/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REID, JAMES, D  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: HARTZIG, B  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 704221  
DRY GAS EXPIRY DATE: 12/11/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: MULDOON

DATE/TIME OF TEST: 07/15/2010 23:31:23

---

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample #1	0.146	23:26
Breath Volume	3.917	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample #2	0.146	23:29
Breath Volume	4.238	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

---

Data Download Date/Time 07/25/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAREFIELD, MOLLY, I  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: FUJIMOTO, RICHARD  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 704221  
DRY GAS EXPIRY DATE: 12/11/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 07/17/2010 02:19:17

---

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Subject Sample #1	0.151	02:14
Breath Volume	2.746	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample #2	0.153	02:17
Breath Volume	2.234	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time 07/25/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTLER, E, Y  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: FRANKLIN, S  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/29/2010

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 704221  
DRY GAS EXPIRY DATE: 12/11/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: WALSTON

DATE/TIME OF TEST: 07/18/2010 22:17:10

---

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.080	22:15	
Air Blank	0.000	22:15	
Subject Sample #1		22:15	Subject Test Refused
Breath Volume	0.000	22:15	
Air Blank	0.000	22:16	
Control Test	0.081	22:16	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

---

Data Download Date/Time 07/25/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITNEY, REBECCA, S  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOLTZHOUSE, J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: SMELKO

DATE/TIME OF TEST: 07/01/2010 02:55:15

---

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:49
Subject Sample #1	0.250	02:50
Breath Volume	1.933	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample #2	0.245	02:53
Breath Volume	2.207	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWETOKOS, JAMES, P  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: LAUGHLIN, J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 07/02/2010 05:48:16

---

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:43	
Control Test	0.079	05:44	
Air Blank	0.000	05:44	
Subject Sample #1		05:46	Subject Test Refused
Breath Volume	0.000	05:46	
Air Blank	0.000	05:47	
Control Test	0.079	05:47	
Air Blank	0.000	05:48	
Diagnostics Check	OK	05:48	

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIBILIA, CHRISTOPHER, R  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: DOWLING, T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MASHBURN

DATE/TIME OF TEST: 07/02/2010 21:56:40

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:48
Subject Sample #1	0.211	21:51
Breath Volume	2.328	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample #2	0.205	21:55
Breath Volume	2.378	21:55
Air Blank	0.000	21:55
Control Test	0.078	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARIS, CECIL, D  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: JONES, J  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: BUI  
ARREST TIME: 19:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 07/03/2010 21:19:37

---

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:13
Subject Sample #1	0.173	21:15
Breath Volume	2.421	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample #2	0.178	21:18
Breath Volume	1.925	21:18
Air Blank	0.000	21:18
Control Test	0.079	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIBILIA, CHRISTOPHER, R  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: STEELFOX, M  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: BUI  
ARREST TIME: 18:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 07/04/2010 19:51:16

---

Diagnostics Check	OK	19:48	
Air Blank	0.000	19:48	
Control Test	0.080	19:49	
Air Blank	0.000	19:49	
Subject Sample #1		19:49	Subject Test Refused
Breath Volume	0.000	19:49	
Air Blank	0.000	19:50	
Control Test	0.079	19:50	
Air Blank	0.000	19:51	
Diagnostics Check	OK	19:51	

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIBILIA, CHRISTOPHER, R  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: CRANE, R  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: BUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 07/04/2010 21:39:16

---

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.080	21:32
Air Blank	0.000	21:33
Subject Sample #1	0.255	21:34
Breath Volume	2.265	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample #2	0.251	21:37
Breath Volume	2.011	21:37
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWETOKOS, JAMES, P  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: RICCI, T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MINER

DATE/TIME OF TEST: 07/06/2010 04:01:29

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:54
Subject Sample #1	0.215	03:56
Breath Volume	2.902	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample #2	0.217	03:59
Breath Volume	2.582	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWETOKOS, JAMES, P  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOLTZHOUSE, J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: TRUMPOLD

DATE/TIME OF TEST: 07/07/2010 01:15:35

---

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Subject Sample #1	0.219	01:10
Breath Volume	1.160	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample #2	0.230	01:13
Breath Volume	1.335	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIBILIA, CHRISTOPHER, R  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: CANFIELD, L  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/08/2010 00:15:35

---

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample #1	0.215	00:11
Breath Volume	1.984	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample #2	0.219	00:14
Breath Volume	1.917	00:14
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWETOKOS, JAMES, P  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: PATTERSON, D  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VERTREES

DATE/TIME OF TEST: 07/11/2010 03:05:19

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Subject Sample #1	0.209	03:00
Breath Volume	2.410	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample #2	0.216	03:03
Breath Volume	2.011	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWETOKOS, JAMES, P  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOLTZHOUSE, J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 07/11/2010 04:00:28

---

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Subject Sample #1	0.117	03:55
Breath Volume	1.996	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample #2	0.116	03:58
Breath Volume	1.562	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTA, SUSAN, L  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOLTZHOUSE, J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: RUSTEK

DATE/TIME OF TEST: 07/12/2010 04:27:03

---

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:20
Air Blank	0.000	04:20
Subject Sample #1	0.222	04:22
Breath Volume	1.707	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample #2	0.213	04:25
Breath Volume	1.734	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:27

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWETOKOS, JAMES, P  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOTLZHOUSE, J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SPEICHER

DATE/TIME OF TEST: 07/16/2010 02:24:52

---

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample #1	0.173	02:20
Breath Volume	2.097	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample #2	0.163	02:23
Breath Volume	2.117	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIBILIA, CHRISTOPHER, R  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: CANFIELD, L  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 07/17/2010 20:31:55

---

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.080	20:26
Air Blank	0.000	20:26
Subject Sample #1	0.141	20:27
Breath Volume	1.472	20:27
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample #2	0.137	20:30
Breath Volume	1.238	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARIS, CECIL, D  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: F, N/A  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: ZERO TOL  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: HEIN

DATE/TIME OF TEST: 07/22/2010 18:17:24

---

Diagnostics Check	OK	18:10
Air Blank	0.000	18:11
Control Test	0.080	18:11
Air Blank	0.000	18:12
Subject Sample #1	0.068	18:12
Breath Volume	2.121	18:12
Air Blank	0.000	18:13
Air Blank	0.000	18:14
Subject Sample #2	0.065	18:15
Breath Volume	1.664	18:15
Air Blank	0.000	18:16
Control Test	0.080	18:16
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARIS, CECIL, D  
BREATH TEST OPERATOR AGENCY: Citrus County SO  
ARRESTING OFFICER: CORBIN, T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: ARABIA

DATE/TIME OF TEST: 07/22/2010 19:42:17

---

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Subject Sample #1	0.061	19:36
Breath Volume	1.316	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample #2	0.055	19:40
Breath Volume	1.246	19:40
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAUER, PHYLLIS, J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOLTZHOUSE, J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 12:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 07/24/2010 13:21:48

---

Diagnostics Check	OK	13:17	
Air Blank	0.000	13:17	
Control Test	0.080	13:17	
Air Blank	0.000	13:18	
Subject Sample #1		13:20	Subject Test Refused
Breath Volume	0.113	13:20	
Air Blank	0.000	13:20	
Control Test	0.081	13:21	
Air Blank	0.000	13:21	
Diagnostics Check	OK	13:21	

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWETOKOS, JAMES, P  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: RICCI, T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 17:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: WILBER

DATE/TIME OF TEST: 07/24/2010 18:29:29

---

Diagnostics Check	OK	18:22
Air Blank	0.000	18:23
Control Test	0.080	18:23
Air Blank	0.000	18:24
Subject Sample #1	0.074	18:24
Breath Volume	2.316	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample #2	0.072	18:27
Breath Volume	2.562	18:27
Air Blank	0.000	18:28
Control Test	0.080	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTA, SUSAN, L  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOWARD, R  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MANFRATES

DATE/TIME OF TEST: 07/25/2010 00:26:11

---

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.080	00:20
Air Blank	0.000	00:20
Subject Sample #1	0.000	00:21
Breath Volume	2.691	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample #2	0.000	00:24
Breath Volume	2.437	00:24
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWETOKOS, JAMES, P  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOLTZHOUSE, J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 17:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 07/25/2010 18:24:59

---

Diagnostics Check	OK	18:18
Air Blank	0.000	18:19
Control Test	0.079	18:19
Air Blank	0.000	18:19
Subject Sample #1	0.208	18:20
Breath Volume	1.609	18:20
Air Blank	0.000	18:21
Air Blank	0.000	18:22
Subject Sample #2	0.210	18:23
Breath Volume	1.925	18:23
Air Blank	0.000	18:24
Control Test	0.080	18:24
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

---

Data Download Date/Time 07/27/2010 13:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTA, SUSAN, L  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: DOYLE, C  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: SEISE

DATE/TIME OF TEST: 07/29/2010 00:20:11

---

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Subject Sample #1	0.226	00:15
Breath Volume	1.789	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample #2	0.217	00:18
Breath Volume	1.210	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

---

Data Download Date/Time 08/31/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIBILIA, CHRISTOPHER, R  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HORTON, T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 07/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: IVANYI

DATE/TIME OF TEST: 07/31/2010 22:08:16

---

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.079	22:03
Air Blank	0.000	22:03
Subject Sample #1	0.000	22:04
Breath Volume	1.167	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample #2	0.000	22:06
Breath Volume	1.109	22:06
Air Blank	0.000	22:07
Control Test	0.078	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

---

Data Download Date/Time 08/31/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUERETH, DONALD, P  
BREATH TEST OPERATOR AGENCY: FDLE  
ARRESTING OFFICER: OFFICER, ARRESTING  
ARRESTING OFFICER AGENCY: FDLE  
LAST AGENCY INSPECTION: 07/27/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 14:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 07/27/2010 14:35:38

---

Diagnostics Check	OK	14:29
Air Blank	0.000	14:30
Control Test	0.081	14:30
Air Blank	0.000	14:30
Subject Sample #1	0.000	14:31
Breath Volume	1.679	14:31
Air Blank	0.000	14:31
Air Blank	0.000	14:33
Subject Sample #2	0.000	14:34
Breath Volume	1.933	14:34
Air Blank	0.000	14:34
Control Test	0.081	14:35
Air Blank	0.000	14:35
Diagnostics Check	OK	14:35

---

Data Download Date/Time 07/27/2010 14:39

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DILLON, LARRY, G

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: POWERS, CHRIS

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 18:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 07/07/2010 20:40:39

---

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.081	20:35
Air Blank	0.000	20:35
Subject Sample #1	0.204	20:36
Breath Volume	4.152	20:36
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample #2	0.200	20:39
Breath Volume	3.996	20:39
Air Blank	0.000	20:39
Control Test	0.082	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

---

Data Download Date/Time 07/24/2010 08:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLOWAY, JAMES, T  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: HOLLOWAY, JAMES  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: KITCHEN

DATE/TIME OF TEST: 07/04/2010 22:43:33

---

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Subject Sample #1	0.086	22:38
Breath Volume	1.750	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample #2	0.084	22:41
Breath Volume	2.464	22:41
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROW, P, L  
BREATH TEST OPERATOR AGENCY: Tavares PD  
ARRESTING OFFICER: DOZIER, J  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 07/07/2010 05:46:59

---

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.080	05:41
Air Blank	0.000	05:41
Subject Sample #1	0.133	05:42
Breath Volume	2.593	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:44
Subject Sample #2	0.121	05:45
Breath Volume	2.093	05:45
Air Blank	0.000	05:46
Control Test	0.080	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BLESSING, JARED, A

BREATH TEST OPERATOR AGENCY: Tavares PD

ARRESTING OFFICER: BLESSING, JARED

ARRESTING OFFICER AGENCY: Tavares PD

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 00:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/08/2010 00:33:12

---

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Subject Sample #1	0.202	00:28
Breath Volume	2.992	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample #2	0.194	00:31
Breath Volume	3.250	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARGE, SUSAN, M  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: MILLER, GARY  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 11:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: STAHLMAN

DATE/TIME OF TEST: 07/08/2010 12:16:38

---

Diagnostics Check	OK	12:10
Air Blank	0.000	12:10
Control Test	0.081	12:11
Air Blank	0.000	12:11
Subject Sample #1	0.000	12:12
Breath Volume	1.339	12:12
Air Blank	0.000	12:12
Air Blank	0.000	12:14
Subject Sample #2	0.000	12:15
Breath Volume	2.472	12:15
Air Blank	0.000	12:15
Control Test	0.080	12:16
Air Blank	0.000	12:16
Diagnostics Check	OK	12:16

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, JONATHAN, B  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: CANDERSON, D  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 07/09/2010 00:27:47

---

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.081	00:22
Air Blank	0.000	00:22
Subject Sample #1	0.008	00:23
Breath Volume	4.078	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample #2	0.007	00:26
Breath Volume	4.335	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PRINCE, ERIC, J

BREATH TEST OPERATOR AGENCY: Tavares PD

ARRESTING OFFICER: REED, W

ARRESTING OFFICER AGENCY: Lake County SO

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 12:40

GENDER: Female

OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 07/10/2010 14:05:40

---

Diagnostics Check	OK	13:59
Air Blank	0.000	13:59
Control Test	0.081	14:00
Air Blank	0.000	14:00
Subject Sample #1	0.176	14:01
Breath Volume	1.628	14:01
Air Blank	0.000	14:01
Air Blank	0.000	14:03
Subject Sample #2	0.168	14:04
Breath Volume	1.347	14:04
Air Blank	0.000	14:04
Control Test	0.080	14:05
Air Blank	0.000	14:05
Diagnostics Check	OK	14:05

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RATLIFF, JOSEPH, A

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: POWERS, CHRISTOPHER

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 00:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: LEWARE

DATE/TIME OF TEST: 07/11/2010 01:46:45

---

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:45	
Subject Sample #1		01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROW, P, L  
BREATH TEST OPERATOR AGENCY: Tavares PD  
ARRESTING OFFICER: DOZIER, J  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: MESTER

DATE/TIME OF TEST: 07/11/2010 04:12:18

---

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.080	04:10	
Air Blank	0.000	04:10	
Subject Sample #1		04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARPER, BILLY, J  
BREATH TEST OPERATOR AGENCY: Tavares PD  
ARRESTING OFFICER: HARPER, BILLY  
ARRESTING OFFICER AGENCY: Tavares PD  
LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 10:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 07/14/2010 10:53:25

---

Diagnostics Check	OK	10:46
Air Blank	0.000	10:47
Control Test	0.080	10:47
Air Blank	0.000	10:48
Subject Sample #1	0.234	10:48
Breath Volume	2.476	10:48
Air Blank	0.000	10:49
Air Blank	0.000	10:51
Subject Sample #2	0.225	10:51
Breath Volume	2.464	10:51
Air Blank	0.000	10:52
Control Test	0.080	10:52
Air Blank	0.000	10:53
Diagnostics Check	OK	10:53

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HALL, JON, M

BREATH TEST OPERATOR AGENCY: Tavares PD

ARRESTING OFFICER: GRAHAM, JOHN

ARRESTING OFFICER AGENCY: Lake County SO

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 10:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:23

SUBJECT'S LAST NAME: SPERRY

DATE/TIME OF TEST: 07/17/2010 11:32:18

---

Diagnostics Check	OK	11:25
Air Blank	0.000	11:26
Control Test	0.080	11:26
Air Blank	0.000	11:27
Subject Sample #1	0.083	11:27
Breath Volume	2.152	11:27
Air Blank	0.000	11:28
Air Blank	0.000	11:30
Subject Sample #2	0.081	11:30
Breath Volume	2.824	11:30
Air Blank	0.000	11:31
Control Test	0.080	11:31
Air Blank	0.000	11:32
Diagnostics Check	OK	11:32

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL, JON, M  
BREATH TEST OPERATOR AGENCY: Tavares PD  
ARRESTING OFFICER: GRAHAM, JOHN  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 10:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:23

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 07/17/2010 12:01:08

---

Diagnostics Check	OK	11:48	
Air Blank	0.000	11:49	
Control Test	0.081	11:49	
Air Blank	0.000	11:49	
Subject Sample #1	0.083	11:53	Slope Not Level
Breath Volume	1.597	11:53	
Air Blank	0.000	11:53	
Air Blank	0.000	11:55	
Subject Sample #2	0.092	11:56	
Breath Volume	1.765	11:56	
Air Blank	0.000	11:56	
Air Blank	0.000	11:58	
Subject Sample #3	0.081	11:59	
Breath Volume	1.542	11:59	
Air Blank	0.000	12:00	
Control Test	0.079	12:00	
Air Blank		12:00	RFI Detect
Air Blank	0.000	12:01	

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HALL, JON, M

BREATH TEST OPERATOR AGENCY: Tavares PD

ARRESTING OFFICER: GRAHAM, JOHN

ARRESTING OFFICER AGENCY: Lake County SO

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 10:42

GENDER: Female

OBSERVATION PERIOD BEGAN: 10:23

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 07/17/2010 12:10:40

---

Diagnostics Check	OK	12:04
Air Blank	0.000	12:05
Control Test	0.080	12:05
Air Blank	0.000	12:05
Subject Sample #1	0.085	12:06
Breath Volume	1.410	12:06
Air Blank	0.000	12:06
Air Blank	0.000	12:08
Subject Sample #2	0.085	12:09
Breath Volume	1.203	12:09
Air Blank	0.000	12:09
Control Test	0.080	12:10
Air Blank	0.000	12:10
Diagnostics Check	OK	12:10

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PALINSKI, T,

BREATH TEST OPERATOR AGENCY: Astatula PD

ARRESTING OFFICER: ROBBINS, C

ARRESTING OFFICER AGENCY: Lake County SO

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 19:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: GAGLIOTI

DATE/TIME OF TEST: 07/19/2010 20:42:53

---

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.081	20:37
Air Blank	0.000	20:38
Subject Sample #1	0.200	20:38
Breath Volume	1.304	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:40
Subject Sample #2	0.195	20:41
Breath Volume	1.421	20:41
Air Blank	0.000	20:41
Control Test	0.080	20:42
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

---

Data Download Date/Time 07/24/2010 08:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CROW, P, L

BREATH TEST OPERATOR AGENCY: Tavares PD

ARRESTING OFFICER: WILLIX, D

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 07/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 16:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 07/24/2010 18:16:26

---

Diagnostics Check	OK	18:08
Air Blank	0.000	18:09
Control Test	0.080	18:09
Air Blank	0.000	18:09
Subject Sample #1	0.266	18:11
Breath Volume	3.203	18:11
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample #2	0.246	18:14
Breath Volume	1.187	18:14
Air Blank	0.000	18:15
Control Test	0.080	18:15
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

---

Data Download Date/Time 08/23/2010 09:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROW, P, L  
BREATH TEST OPERATOR AGENCY: Tavares PD  
ARRESTING OFFICER: DOZIER, G  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CISNEROS

DATE/TIME OF TEST: 07/25/2010 00:52:49

---

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:47
Air Blank	0.000	00:47
Subject Sample #1	0.242	00:48
Breath Volume	1.351	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample #2	0.248	00:51
Breath Volume	1.539	00:51
Air Blank	0.000	00:51
Control Test	0.081	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

---

Data Download Date/Time 08/23/2010 09:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CROW, P, L

BREATH TEST OPERATOR AGENCY: Tavares PD

ARRESTING OFFICER: DOZIER, G

ARRESTING OFFICER AGENCY: Lake County SO

LAST AGENCY INSPECTION: 07/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 02:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: STANIEWICZ

DATE/TIME OF TEST: 07/25/2010 03:14:42

---

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.079	03:12	
Air Blank	0.000	03:12	
Subject Sample #1		03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

---

Data Download Date/Time 08/23/2010 09:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CROW, P, L

BREATH TEST OPERATOR AGENCY: Tavares PD

ARRESTING OFFICER: COSDEN, B

ARRESTING OFFICER AGENCY: Lake County SO

LAST AGENCY INSPECTION: 07/24/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 02:27

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 07/25/2010 03:31:41

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:26
Subject Sample #1	0.000	03:27
Breath Volume	1.566	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample #2	0.000	03:30
Breath Volume	1.378	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

---

Data Download Date/Time 08/23/2010 09:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER, SANDRA, W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PADGETT, D  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 17:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 07/25/2010 19:20:52

---

Diagnostics Check	OK	19:14
Air Blank	0.000	19:14
Control Test	0.081	19:14
Air Blank	0.000	19:15
Subject Sample #1	0.000	19:16
Breath Volume	3.464	19:16
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample #2	0.000	19:19
Breath Volume	2.800	19:19
Air Blank	0.000	19:19
Control Test	0.077	19:20
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

---

Data Download Date/Time 08/23/2010 09:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORRIER, LISA, M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: OBRIEN, TIMOTHY  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 07/27/2010 02:42:43

---

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample #1	0.181	02:37
Breath Volume	3.050	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.182	02:41
Breath Volume	2.574	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

---

Data Download Date/Time 08/23/2010 09:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER, SANDRA, W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER, SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DENMARK

DATE/TIME OF TEST: 07/27/2010 20:25:59

---

Diagnostics Check	OK	20:22	
Air Blank	0.000	20:23	
Control Test	0.081	20:23	
Air Blank	0.000	20:24	
Subject Sample #1		20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:25	
Control Test	0.080	20:25	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

---

Data Download Date/Time 08/23/2010 09:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUSCOTT, DANA, E  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: STRAUB, DAVE  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 07/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: BUI  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 07/30/2010 21:00:44

---

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.081	20:55
Air Blank	0.000	20:55
Subject Sample #1	0.152	20:56
Breath Volume	3.796	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample #2	0.141	20:59
Breath Volume	3.410	20:59
Air Blank	0.000	20:59
Control Test	0.077	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

---

Data Download Date/Time 08/23/2010 09:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BENINATO, LORI, S  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: OBRIEN, TIMOTHY  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 07/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 07/31/2010 22:02:04

---

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:53
Subject Sample #1	0.264	21:56
Breath Volume	1.917	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:58
Subject Sample #2	0.270	22:00
Breath Volume	2.816	22:00
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:02

---

Data Download Date/Time 08/23/2010 09:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEVENS, JEREMY, L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: STEVENS, JEREMY

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 07/02/2010 03:31:21

---

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample #1	0.030	03:26
Breath Volume	1.851	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample #2	0.029	03:29
Breath Volume	1.914	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUSEY, JOHN, T  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: JAMES, SARAH  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 07/03/2010 21:57:04

---

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:52
Subject Sample #1	0.170	21:52
Breath Volume	1.371	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:54
Subject Sample #2	0.172	21:55
Breath Volume	1.308	21:55
Air Blank	0.000	21:56
Control Test	0.078	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:57

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, JONATHAN, B  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: WILLIAMS, JONATHAN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/04/2010 23:11:57

---

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.078	23:04
Air Blank	0.000	23:05
Subject Sample #1	0.181	23:07
Breath Volume	2.171	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample #2	0.172	23:10
Breath Volume	3.160	23:10
Air Blank	0.000	23:11
Control Test	0.077	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARSONS, CHRISTOPHER, L  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: CICERO, CHRISTINE  
ARRESTING OFFICER AGENCY: Lady Lake PD  
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MADDEN

DATE/TIME OF TEST: 07/05/2010 00:36:42

---

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:31
Subject Sample #1	0.120	00:32
Breath Volume	1.453	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample #2	0.117	00:35
Breath Volume	1.332	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KETCHUM, CHARLES, A  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: LISTON, JOHN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 08:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: MCLENDON

DATE/TIME OF TEST: 07/10/2010 09:14:17

---

Diagnostics Check	OK	09:10	
Air Blank	0.000	09:10	
Control Test	0.080	09:11	
Air Blank	0.000	09:11	
Subject Sample #1		09:12	Subject Test Refused
Breath Volume	0.000	09:12	
Air Blank	0.000	09:13	
Control Test	0.080	09:13	
Air Blank	0.000	09:14	
Diagnostics Check	OK	09:14	

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, JONATHAN, B  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: WILLIAMS, JONATHAN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PONDER

DATE/TIME OF TEST: 07/13/2010 03:14:53

---

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:09
Air Blank	0.000	03:09
Subject Sample #1	0.100	03:10
Breath Volume	3.921	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample #2	0.089	03:13
Breath Volume	1.718	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CUNNINGHAM, ALISA, B

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: VANCE, S

ARRESTING OFFICER AGENCY: Lake County SO

LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 01:51

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 07/14/2010 02:13:20

---

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:11	
Subject Sample #1		02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.079	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEVENS, JEREMY, L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: STEVENS, JEREMY

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 00:23

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: STODDARD

DATE/TIME OF TEST: 07/15/2010 01:32:14

---

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Subject Sample #1	0.000	01:27
Breath Volume	2.292	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample #2	0.000	01:30
Breath Volume	1.636	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAUSEY, JOHN, T

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: PARSONS, CHRISTOPHER

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 02:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CARMICHAEL

DATE/TIME OF TEST: 07/17/2010 03:28:56

---

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample #1	0.136	03:24
Breath Volume	2.605	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample #2	0.136	03:27
Breath Volume	1.488	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAUSEY, JOHN, T

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WILLIAMS, JONATHAN

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 22:29

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: PENTURFF

DATE/TIME OF TEST: 07/18/2010 23:42:17

---

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Subject Sample #1	0.223	23:37
Breath Volume	1.960	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample #2	0.230	23:40
Breath Volume	1.917	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEVENS, JEREMY, L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: LORENTZEN, EDWARD

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 22:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: KRANICK

DATE/TIME OF TEST: 07/20/2010 23:46:04

---

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.079	23:43	
Air Blank	0.000	23:44	
Subject Sample #1		23:44	RFI Detect
Breath Volume	0.000	23:44	
Air Blank	0.000	23:45	
Control Test	0.078	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:46	

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEVENS, JEREMY, L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: LORENTZEN, EDWARD

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 22:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: KRANICK

DATE/TIME OF TEST: 07/21/2010 00:23:02

---

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.078	00:12	
Air Blank	0.000	00:13	
Subject Sample #1	0.153	00:16	Volume Not Met
Breath Volume	0.406	00:16	
Air Blank	0.000	00:17	
Air Blank	0.000	00:18	
Subject Sample #2	0.178	00:22	Volume Not Met
Breath Volume	0.152	00:22	
Air Blank		00:22	RFI Detect
Air Blank	0.000	00:23	

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEVENS, JEREMY, L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: LORENTZEN, EDWARD

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 22:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: KRANICK

DATE/TIME OF TEST: 07/21/2010 00:40:11

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Subject Sample #1	0.159	00:34
Breath Volume	1.230	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:36
Subject Sample #2	0.161	00:38
Breath Volume	1.441	00:38
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRENSHAW, JARED, M  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: CRENSHAW, JARED  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: RANGEL JR

DATE/TIME OF TEST: 07/21/2010 01:50:36

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:44
Subject Sample #1	0.168	01:46
Breath Volume	2.117	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample #2	0.162	01:49
Breath Volume	2.335	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, JONATHAN, B  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: WILLIAMS, JONATHAN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VANVLACK

DATE/TIME OF TEST: 07/22/2010 01:20:48

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:15
Subject Sample #1	0.171	01:15
Breath Volume	2.066	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample #2	0.177	01:19
Breath Volume	2.753	01:19
Air Blank	0.000	01:19
Control Test	0.079	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEVENS, JEREMY, L  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: STEVENS, JEREMY  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/23/2010 23:17:47

---

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:15	
Control Test	0.080	23:15	
Air Blank	0.000	23:16	
Subject Sample #1		23:16	Subject Test Refused
Breath Volume	0.000	23:16	
Air Blank	0.000	23:16	
Control Test	0.079	23:17	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEVENS, JEREMY, L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WALSH, SHANNON

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 02:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: COATS

DATE/TIME OF TEST: 07/24/2010 03:26:48

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:21
Air Blank	0.000	03:21
Subject Sample #1	0.000	03:22
Breath Volume	2.308	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample #2	0.000	03:25
Breath Volume	2.253	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALSH, SHANNON, L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: STEVENS, JEREMY

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 01:46

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 07/29/2010 02:39:06

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:33
Subject Sample #1	0.151	02:34
Breath Volume	2.046	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample #2	0.145	02:37
Breath Volume	2.261	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

---

Data Download Date/Time 07/29/2010 05:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEVENS, JEREMY, L  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: STEVENS, JEREMY  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 07/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 07/30/2010 01:52:34

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample #1	0.114	01:48
Breath Volume	2.507	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample #2	0.107	01:51
Breath Volume	2.015	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time 08/08/2010 21:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEVENS, JEREMY, L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: BARRETT, GARY

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 07/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 02:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 07/30/2010 03:17:12

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample #1	0.160	03:12
Breath Volume	2.585	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.148	03:15
Breath Volume	2.207	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time 08/08/2010 21:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, JONATHAN, B  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: WILLIAMS, JONATHAN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 07/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GOWENS

DATE/TIME OF TEST: 07/31/2010 01:28:02

---

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:22
Air Blank	0.000	01:22
Subject Sample #1	0.076	01:23
Breath Volume	3.464	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample #2	0.070	01:26
Breath Volume	1.898	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:28

---

Data Download Date/Time 08/08/2010 21:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, JONATHAN, B  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: PARSONS, CHRIS  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 07/29/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/31/2010 02:55:52

---

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample #1	0.205	02:51
Breath Volume	2.054	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample #2	0.204	02:54
Breath Volume	1.675	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

---

Data Download Date/Time 08/08/2010 21:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DANCEL, KAREN, J

BREATH TEST OPERATOR AGENCY: Mount Dora PD

ARRESTING OFFICER: DANCEL, KAREN

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 595761

DRY GAS EXPIRY DATE: 04/13/2011

VIOLATION CODE: DUI

ARREST TIME: 00:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BELLEROSE

DATE/TIME OF TEST: 07/04/2010 01:32:24

---

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:26
Subject Sample #1	0.168	01:27
Breath Volume	2.394	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample #2	0.178	01:30
Breath Volume	2.074	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time 07/26/2010 19:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DANCEL, KAREN, J

BREATH TEST OPERATOR AGENCY: Mount Dora PD

ARRESTING OFFICER: GREGORY, CHARLES

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 595761

DRY GAS EXPIRY DATE: 04/13/2011

VIOLATION CODE: DUI

ARREST TIME: 03:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BORTHWICK

DATE/TIME OF TEST: 07/05/2010 03:33:54

---

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.078	03:31	
Air Blank	0.000	03:32	
Subject Sample #1		03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.078	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

---

Data Download Date/Time 07/26/2010 19:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DANCEL, KAREN, J

BREATH TEST OPERATOR AGENCY: Mount Dora PD

ARRESTING OFFICER: GREGORY, CHARLES

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 595761

DRY GAS EXPIRY DATE: 04/13/2011

VIOLATION CODE: DUI

ARREST TIME: 02:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: WIEGAND

DATE/TIME OF TEST: 07/09/2010 02:57:36

---

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Subject Sample #1	0.157	02:53
Breath Volume	4.500	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample #2	0.164	02:56
Breath Volume	4.035	02:56
Air Blank	0.000	02:56
Control Test	0.076	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time 07/26/2010 19:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HELFANT, ADAM, M

BREATH TEST OPERATOR AGENCY: Mount Dora PD

ARRESTING OFFICER: COSDEN, BROCC

ARRESTING OFFICER AGENCY: Lake County SO

LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 595761

DRY GAS EXPIRY DATE: 04/13/2011

VIOLATION CODE: DUI

ARREST TIME: 22:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/11/2010 00:07:45

---

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.078	00:01
Air Blank	0.000	00:02
Subject Sample #1	0.284	00:03
Breath Volume	3.425	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample #2	0.293	00:06
Breath Volume	3.691	00:06
Air Blank	0.000	00:06
Control Test	0.078	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

---

Data Download Date/Time 07/26/2010 19:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HELFANT, ADAM, M

BREATH TEST OPERATOR AGENCY: Mount Dora PD

ARRESTING OFFICER: HELFANT, ADAM

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 595761

DRY GAS EXPIRY DATE: 04/13/2011

VIOLATION CODE: DUI

ARREST TIME: 23:23

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 07/15/2010 00:06:49

---

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.078	00:01
Air Blank	0.000	00:01
Subject Sample #1	0.011	00:02
Breath Volume	3.589	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample #2	0.012	00:05
Breath Volume	3.621	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

---

Data Download Date/Time 07/26/2010 19:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREGORY, CHARLES, L  
BREATH TEST OPERATOR AGENCY: Mount Dora PD  
ARRESTING OFFICER: GREGORY, CHARLES  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 595761  
DRY GAS EXPIRY DATE: 04/13/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 07/17/2010 02:48:31

---

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Subject Sample #1	0.115	02:43
Breath Volume	1.125	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample #2	0.117	02:47
Breath Volume	1.496	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

---

Data Download Date/Time 07/26/2010 19:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREGORY, CHARLES, L  
BREATH TEST OPERATOR AGENCY: Mount Dora PD  
ARRESTING OFFICER: GREGORY, CHARLES  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 595761  
DRY GAS EXPIRY DATE: 04/13/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: OBERDICK

DATE/TIME OF TEST: 07/18/2010 01:36:57

---

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.078	01:31
Air Blank	0.000	01:31
Subject Sample #1	0.082	01:32
Breath Volume	2.351	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample #2	0.084	01:35
Breath Volume	4.035	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time 07/26/2010 19:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREGORY, CHARLES, L  
BREATH TEST OPERATOR AGENCY: Mount Dora PD  
ARRESTING OFFICER: GREGORY, CHARLES  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 595761  
DRY GAS EXPIRY DATE: 04/13/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DURKIN

DATE/TIME OF TEST: 07/19/2010 01:20:07

---

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:15
Subject Sample #1	0.241	01:15
Breath Volume	4.527	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample #2	0.241	01:18
Breath Volume	4.339	01:18
Air Blank	0.000	01:19
Control Test	0.077	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

---

Data Download Date/Time 07/26/2010 19:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREGORY, CHARLES, L

BREATH TEST OPERATOR AGENCY: Mount Dora PD

ARRESTING OFFICER: GREGORY, CHARLES

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 595761

DRY GAS EXPIRY DATE: 04/13/2011

VIOLATION CODE: DUI

ARREST TIME: 22:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/21/2010 23:15:16

---

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:09
Subject Sample #1	0.159	23:10
Breath Volume	2.789	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample #2	0.170	23:13
Breath Volume	3.050	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

---

Data Download Date/Time 07/26/2010 19:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ, J,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: THOMPSON, L  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 07/01/2010 01:52:22

---

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.082	01:46
Air Blank	0.000	01:47
Subject Sample #1	0.164	01:47
Breath Volume	3.445	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample #2	0.165	01:50
Breath Volume	3.250	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ, J,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PORCELLI, F  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BONANNO

DATE/TIME OF TEST: 07/01/2010 02:16:58

---

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample #1	0.147	02:12
Breath Volume	3.062	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample #2	0.146	02:15
Breath Volume	3.347	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIX, BRYAN, D  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: GALLUP, C.  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/01/2010 05:10:38

---

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.080	05:00	
Air Blank	0.000	05:00	
Subject Sample #1		05:03	No Sample Provided
Breath Volume	0.000	05:03	
Air Blank	0.000	05:04	
Air Blank	0.000	05:06	
Subject Sample #2		05:09	No Sample Provided
Breath Volume	0.000	05:09	
Air Blank	0.000	05:09	
Control Test	0.081	05:10	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, GARY, A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER, GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 07/02/2010 20:09:28

---

Diagnostics Check	OK	20:03
Air Blank	0.000	20:03
Control Test	0.081	20:04
Air Blank	0.000	20:04
Subject Sample #1	0.000	20:05
Breath Volume	1.117	20:05
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample #2	0.000	20:08
Breath Volume	1.234	20:08
Air Blank	0.000	20:08
Control Test	0.080	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRITCHETT, A, M  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: JOEDICKE, W  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: GANTZ

DATE/TIME OF TEST: 07/03/2010 01:54:56

---

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.081	01:49
Air Blank	0.000	01:49
Subject Sample #1	0.113	01:50
Breath Volume	1.851	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample #2	0.120	01:53
Breath Volume	2.019	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, GARY, A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER, GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DEMPSEY

DATE/TIME OF TEST: 07/03/2010 23:42:07

---

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Subject Sample #1	0.185	23:37
Breath Volume	2.320	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample #2	0.175	23:40
Breath Volume	1.488	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIOS, CARLOS, F  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: SAMS, M  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/04/2010 03:35:03

---

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:33	
Air Blank	0.000	03:33	
Subject Sample #1		03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:35	

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRITCHETT, A, M  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: JONES, A  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 07/05/2010 01:44:36

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Subject Sample #1	0.130	01:39
Breath Volume	1.808	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample #2	0.135	01:43
Breath Volume	1.972	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATERS, CYNTHIA, M  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BRUNER, D  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 11:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:29

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/06/2010 12:06:22

---

Diagnostics Check	OK	12:00
Air Blank	0.000	12:00
Control Test	0.081	12:01
Air Blank	0.000	12:01
Subject Sample #1	0.181	12:01
Breath Volume	1.378	12:01
Air Blank	0.000	12:02
Air Blank	0.000	12:04
Subject Sample #2	0.180	12:04
Breath Volume	1.140	12:04
Air Blank	0.000	12:05
Control Test	0.081	12:05
Air Blank	0.000	12:06
Diagnostics Check	OK	12:06

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNTER, ROBERT, M  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: KWOCZKA, G  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BUETTI

DATE/TIME OF TEST: 07/07/2010 00:37:46

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.081	00:32
Air Blank	0.000	00:32
Subject Sample #1	0.000	00:33
Breath Volume	1.687	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample #2	0.000	00:36
Breath Volume	2.550	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, GARY, A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER, GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 07/08/2010 23:30:48

---

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.080	23:25
Air Blank	0.000	23:25
Subject Sample #1	0.106	23:26
Breath Volume	1.570	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample #2	0.104	23:29
Breath Volume	1.453	23:29
Air Blank	0.000	23:29
Control Test	0.081	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE, M,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MOORE, M  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 07/09/2010 02:35:17

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:26
Subject Sample #1	0.174	02:28
Breath Volume	2.089	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample #2	0.174	02:33
Breath Volume	1.960	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZIEGLER, BRANDON, R  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: WRIGHT, D  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 06/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 07/09/2010 21:30:42

---

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.081	21:24
Air Blank	0.000	21:25
Subject Sample #1	0.333	21:25
Breath Volume	1.789	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample #2	0.352	21:29
Breath Volume	1.593	21:29
Air Blank	0.000	21:29
Control Test	0.080	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

---

Data Download Date/Time 07/25/2010 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L.D.  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 07/24/2010 03:18:42

---

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:16	
Subject Sample #1		03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.079	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

---

Data Download Date/Time 08/06/2010 13:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L.D.  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 07/20/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 07/24/2010 03:31:25

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample #1	0.125	03:27
Breath Volume	2.265	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample #2	0.123	03:29
Breath Volume	1.707	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

---

Data Download Date/Time 08/06/2010 13:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, BRADFORD, B  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BAKER, BRADFORD  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CALI

DATE/TIME OF TEST: 07/01/2010 03:03:30

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:57
Subject Sample #1	0.112	02:58
Breath Volume	1.332	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample #2	0.109	03:01
Breath Volume	1.414	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BULLOCK, JOSEPH, J  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: BULLOCK, JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: TANIS

DATE/TIME OF TEST: 07/01/2010 17:37:05

---

Diagnostics Check	OK	17:30
Air Blank	0.000	17:30
Control Test	0.079	17:31
Air Blank	0.000	17:31
Subject Sample #1	0.000	17:32
Breath Volume	4.746	17:32
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample #2	0.000	17:35
Breath Volume	6.148	17:35
Air Blank	0.000	17:36
Control Test	0.079	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:37

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: LANSIDEL

DATE/TIME OF TEST: 07/02/2010 23:37:27

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Subject Sample #1	0.218	23:32
Breath Volume	2.613	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample #2	0.218	23:35
Breath Volume	1.320	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, BRADFORD, B  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BAKER, BRADFORD  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LAUDEN

DATE/TIME OF TEST: 07/06/2010 00:15:32

---

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.079	00:10
Air Blank	0.000	00:10
Subject Sample #1	0.235	00:10
Breath Volume	3.019	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample #2	0.221	00:13
Breath Volume	2.597	00:13
Air Blank	0.000	00:14
Control Test	0.077	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: DELANEY

DATE/TIME OF TEST: 07/06/2010 23:06:11

---

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Subject Sample #1	0.186	23:01
Breath Volume	2.578	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample #2	0.184	23:04
Breath Volume	2.632	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 07/09/2010 04:01:39

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Subject Sample #1	0.146	03:57
Breath Volume	3.011	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample #2	0.147	04:00
Breath Volume	2.933	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARLOW, ROBERT, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BARLOW, ROBERT  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/10/2010 03:42:25

---

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample #1	0.085	03:37
Breath Volume	1.289	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample #2	0.071	03:40
Breath Volume	1.511	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARLOW, ROBERT, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BARLOW, ROBERT  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: BUFORD

DATE/TIME OF TEST: 07/11/2010 03:37:03

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:31
Subject Sample #1	0.075	03:32
Breath Volume	1.878	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample #2	0.083	03:35
Breath Volume	2.976	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PANARO, RONNIE, M  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: PANARO, RONNIE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: GENTRY

DATE/TIME OF TEST: 07/11/2010 16:57:08

---

Diagnostics Check	OK	16:50
Air Blank	0.000	16:51
Control Test	0.079	16:51
Air Blank	0.000	16:52
Subject Sample #1	0.137	16:52
Breath Volume	3.226	16:52
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample #2	0.131	16:55
Breath Volume	3.351	16:55
Air Blank	0.000	16:56
Control Test	0.079	16:56
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARLOW, ROBERT, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BARLOW, ROBERT  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 07/12/2010 02:34:25

---

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.151	02:29
Breath Volume	4.546	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample #2	0.151	02:32
Breath Volume	5.621	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, BRADFORD, B  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BAKER, BRADFORD  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/12/2010 02:45:43

---

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.078	02:43	
Air Blank	0.000	02:44	
Subject Sample #1		02:44	RFI Detect
Breath Volume	0.000	02:44	
Air Blank	0.000	02:44	
Control Test	0.079	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, BRADFORD, B  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BAKER, BRADFORD  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/12/2010 02:50:47

---

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.079	02:48	
Air Blank	0.000	02:48	
Subject Sample #1		02:49	Slope Not Met
Breath Volume	1.527	02:49	
Air Blank	0.000	02:49	
Control Test	0.079	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, BRADFORD, B  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BAKER, BRADFORD  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/12/2010 03:15:55

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.078	03:10
Air Blank	0.000	03:11
Subject Sample #1	0.220	03:11
Breath Volume	1.253	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:13
Subject Sample #2	0.220	03:14
Breath Volume	1.386	03:14
Air Blank	0.000	03:14
Control Test	0.078	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FRERICHS

DATE/TIME OF TEST: 07/15/2010 23:46:04

---

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:40	
Control Test	0.080	23:40	
Air Blank	0.000	23:41	
Subject Sample #1	0.147	23:41	
Breath Volume	3.421	23:41	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample #2	0.136	23:44	
Breath Volume	2.648	23:44	
Air Blank	0.000	23:45	
Control Test	0.073	23:45	Control Outside Tolerance
Air Blank	0.000	23:46	

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FRERICHS

DATE/TIME OF TEST: 07/15/2010 23:54:12

---

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.079	23:48
Air Blank	0.000	23:48
Subject Sample #1	0.139	23:49
Breath Volume	2.609	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:51
Subject Sample #2	0.139	23:52
Breath Volume	2.441	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: MULQUEEN

DATE/TIME OF TEST: 07/16/2010 20:03:33

---

Diagnostics Check	OK	19:56
Air Blank	0.000	19:57
Control Test	0.078	19:57
Air Blank	0.000	19:57
Subject Sample #1	0.220	19:58
Breath Volume	2.695	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample #2	0.218	20:01
Breath Volume	2.125	20:01
Air Blank	0.000	20:02
Control Test	0.078	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: STEEN

DATE/TIME OF TEST: 07/17/2010 02:46:07

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:39
Subject Sample #1	0.220	02:40
Breath Volume	1.949	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample #2	0.200	02:44
Breath Volume	1.542	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BULLOCK, JOSEPH, J  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: BULLOCK, JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GUADALUPE

DATE/TIME OF TEST: 07/17/2010 03:51:51

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample #1	0.174	03:47
Breath Volume	3.308	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample #2	0.175	03:50
Breath Volume	4.132	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS, CHRISTOPHER, L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 07/18/2010 03:18:30

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.078	03:12
Air Blank	0.000	03:12
Subject Sample #1	0.220	03:13
Breath Volume	2.660	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.214	03:16
Breath Volume	2.000	03:16
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARLOW, ROBERT, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: KORKOWSKI, RAYMOND  
ARRESTING OFFICER AGENCY: Tequesta PD  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 07/19/2010 02:15:26

---

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:13	
Subject Sample #1	0.206	02:13	
Breath Volume	2.078	02:13	
Air Blank		02:13	Improper Sample
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARLOW, ROBERT, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: KORKOWSKI, RAYMOND  
ARRESTING OFFICER AGENCY: Tequesta PD  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 07/19/2010 02:27:58

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Subject Sample #1	0.185	02:23
Breath Volume	2.625	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample #2	0.190	02:26
Breath Volume	2.046	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, BRADFORD, B  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BAKER, BRADFORD  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/20/2010 00:58:00

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample #1	0.208	00:53
Breath Volume	3.921	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample #2	0.207	00:56
Breath Volume	2.875	00:56
Air Blank	0.000	00:57
Control Test	0.077	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:58

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: THRASHER

DATE/TIME OF TEST: 07/21/2010 00:50:52

---

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample #1	0.197	00:46
Breath Volume	2.878	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample #2	0.195	00:49
Breath Volume	2.960	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CARRIZO

DATE/TIME OF TEST: 07/23/2010 03:39:36

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Subject Sample #1	0.176	03:35
Breath Volume	1.363	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample #2	0.173	03:38
Breath Volume	1.578	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: ROOD

DATE/TIME OF TEST: 07/24/2010 01:38:11

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample #1	0.178	01:33
Breath Volume	3.308	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample #2	0.168	01:36
Breath Volume	3.023	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREGORY, AMANDA, M  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: GREGORY, AMANDA  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 06/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 16:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: MANUEL

DATE/TIME OF TEST: 07/24/2010 19:54:35

---

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.080	19:48
Air Blank	0.000	19:48
Subject Sample #1	0.143	19:50
Breath Volume	2.527	19:50
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample #2	0.125	19:52
Breath Volume	1.457	19:52
Air Blank	0.000	19:53
Control Test	0.078	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

---

Data Download Date/Time 07/25/2010 10:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWIEZY, STAN, L  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: BUCK, AMY  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: BUI  
ARREST TIME: 14:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: SWIEZY

DATE/TIME OF TEST: 07/25/2010 15:59:57

---

Diagnostics Check	OK	15:48	
Air Blank	0.000	15:48	
Control Test	0.080	15:49	
Air Blank	0.000	15:49	
Subject Sample #1		15:52	No Sample Provided
Breath Volume	0.000	15:52	
Air Blank	0.000	15:53	
Air Blank	0.000	15:55	
Subject Sample #2		15:58	No Sample Provided
Breath Volume	0.000	15:58	
Air Blank	0.000	15:59	
Control Test	0.080	15:59	
Air Blank	0.000	15:59	
Diagnostics Check	OK	15:59	

---

Data Download Date/Time 08/30/2010 11:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARLOW, ROBERT, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BARLOW, ROBERT  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: KUNST

DATE/TIME OF TEST: 07/26/2010 23:26:28

---

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:20
Subject Sample #1	0.228	23:21
Breath Volume	4.882	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample #2	0.231	23:24
Breath Volume	5.375	23:24
Air Blank	0.000	23:25
Control Test	0.078	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

---

Data Download Date/Time 08/30/2010 11:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIMMONS, BRIAN, K  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: TIMMONS, BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WONGPROM

DATE/TIME OF TEST: 07/28/2010 23:44:07

---

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Subject Sample #1	0.186	23:38	
Breath Volume	2.578	23:38	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample #2		23:42	Slope Not Met
Breath Volume	0.812	23:42	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

---

Data Download Date/Time 08/30/2010 11:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIMMONS, BRIAN, K  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: TIMMONS, BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WONGPROM

DATE/TIME OF TEST: 07/28/2010 23:52:51

---

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:46
Subject Sample #1	0.178	23:47
Breath Volume	2.523	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample #2	0.188	23:51
Breath Volume	2.308	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

---

Data Download Date/Time 08/30/2010 11:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, BRADFORD, B  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: MUSE, GLEN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 07/30/2010 02:36:45

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:31
Air Blank	0.000	02:31
Subject Sample #1	0.186	02:32
Breath Volume	2.476	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample #2	0.191	02:35
Breath Volume	3.019	02:35
Air Blank	0.000	02:35
Control Test	0.077	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time 08/30/2010 11:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIMM, GEORGE, W  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: PRIMM, GEORGE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 07/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TOWNSHEND

DATE/TIME OF TEST: 07/31/2010 00:10:03

---

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:04
Air Blank	0.000	00:04
Subject Sample #1	0.147	00:05
Breath Volume	5.035	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:07
Subject Sample #2	0.150	00:08
Breath Volume	5.664	00:08
Air Blank	0.000	00:09
Control Test	0.077	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:10

---

Data Download Date/Time 08/30/2010 11:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POOLE, LIAS, M  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: POOLE, LIAS  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 593673  
DRY GAS EXPIRY DATE: 03/16/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 07/04/2010 23:55:21

---

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.075	23:53	
Air Blank	0.000	23:53	
Subject Sample #1		23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:54	
Control Test	0.075	23:54	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

---

Data Download Date/Time 08/04/2010 11:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOTZONG, DUSTIN, R  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: BOTZONG, DUSTIN  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 593673  
DRY GAS EXPIRY DATE: 03/16/2011  
VIOLATION CODE: DUI  
ARREST TIME: 16:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: SEXTON JR

DATE/TIME OF TEST: 07/05/2010 17:23:36

---

Diagnostics Check	OK	17:22	
Air Blank	0.000	17:22	
Control Test	0.074	17:23	Control Outside Tolerance
Air Blank	0.000	17:23	

---

Data Download Date/Time 08/04/2010 11:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOTZONG, DUSTIN, R  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: BOTZONG, DUSTIN  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 593673  
DRY GAS EXPIRY DATE: 03/16/2011  
VIOLATION CODE: DUI  
ARREST TIME: 16:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: SEXTON JR

DATE/TIME OF TEST: 07/05/2010 17:28:32

---

Diagnostics Check	OK	17:27	
Air Blank	0.000	17:27	
Control Test	0.074	17:27	Control Outside Tolerance
Air Blank	0.000	17:28	

---

Data Download Date/Time 08/04/2010 11:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON, PATRICK, T  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: HAMILTON, PATRICK  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 593673  
DRY GAS EXPIRY DATE: 03/16/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: EDWINS

DATE/TIME OF TEST: 07/07/2010 23:22:01

---

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.074	23:21	Control Outside Tolerance
Air Blank	0.000	23:22	

---

Data Download Date/Time 08/04/2010 11:16

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON, PATRICK, T  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: HAMILTON, PATRICK  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 593673  
DRY GAS EXPIRY DATE: 03/16/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: EDWINS

DATE/TIME OF TEST: 07/07/2010 23:25:13

---

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.074	23:24	Control Outside Tolerance
Air Blank	0.000	23:25	

---

Data Download Date/Time 08/04/2010 11:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON, PATRICK, T  
BREATH TEST OPERATOR AGENCY: Lynn Haven PD  
ARRESTING OFFICER: HAMILTON, PATRICK  
ARRESTING OFFICER AGENCY: Lynn Haven PD  
LAST AGENCY INSPECTION: 06/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 593673  
DRY GAS EXPIRY DATE: 03/16/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: EDWINS

DATE/TIME OF TEST: 07/07/2010 23:28:19

---

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.074	23:27	Control Outside Tolerance
Air Blank	0.000	23:28	

---

Data Download Date/Time 08/04/2010 11:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'GRADY, PATRICK,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MORGAN, JOSHUA  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: IRICK

DATE/TIME OF TEST: 07/02/2010 00:13:08

---

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Subject Sample #1	0.000	00:08
Breath Volume	1.199	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample #2	0.000	00:11
Breath Volume	2.062	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

---

Data Download Date/Time 07/09/2010 11:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'GRADY, PATRICK,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MORGAN, JOSHUA  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 07/03/2010 00:16:57

---

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.078	00:11
Air Blank	0.000	00:11
Subject Sample #1	0.066	00:12
Breath Volume	2.589	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample #2	0.067	00:15
Breath Volume	3.148	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

---

Data Download Date/Time 07/09/2010 11:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'GRADY, PATRICK,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CIESLINSKI, JOSEPH  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/30/2010 03:43:10

---

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.077	03:40	
Air Blank	0.000	03:40	
Subject Sample #1		03:41	Interferent Detect
Breath Volume	1.273	03:41	
Air Blank	0.000	03:42	
Control Test	0.078	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

---

Data Download Date/Time 08/13/2010 11:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'GRADY, PATRICK,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CIESLINSKI, JOSEPH  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 07/09/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 08/10/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/30/2010 04:08:37

---

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.077	04:02
Air Blank	0.000	04:03
Subject Sample #1	0.079	04:04
Breath Volume	1.953	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample #2	0.083	04:07
Breath Volume	2.351	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time 08/13/2010 11:19