

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY, ROBERT, P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BURKE, FLOYD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 807901I

DRY GAS EXPIRY DATE: 03/21/2010

VIOLATION CODE: DUI

ARREST TIME: 18:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: KAPUSHY

DATE/TIME OF TEST: 03/03/2010 19:27:14

---

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.082	19:21
Air Blank	0.000	19:22
Subject Sample #1	0.160	19:22
Breath Volume	1.691	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample #2	0.161	19:25
Breath Volume	1.714	19:25
Air Blank	0.000	19:26
Control Test	0.081	19:26
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

---

Data Download Date/Time 03/24/2010 17:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY, ROBERT, P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BURKE, FLOYD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 807901I

DRY GAS EXPIRY DATE: 03/21/2010

VIOLATION CODE: DUI

ARREST TIME: 02:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 03/10/2010 04:03:32

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Subject Sample #1	0.159	03:58
Breath Volume	1.390	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample #2	0.147	04:01
Breath Volume	1.488	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time 03/24/2010 17:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY, ROBERT, P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BURKE, FLOYD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 807901I

DRY GAS EXPIRY DATE: 03/21/2010

VIOLATION CODE: DUI

ARREST TIME: 20:33

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: GRIFFITHS

DATE/TIME OF TEST: 03/11/2010 21:05:51

---

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.081	21:00
Air Blank	0.000	21:00
Subject Sample #1	0.102	21:01
Breath Volume	1.527	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample #2	0.099	21:04
Breath Volume	1.687	21:04
Air Blank	0.000	21:04
Control Test	0.081	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

---

Data Download Date/Time 03/24/2010 17:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TOBAR, FRANK, J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BURKE, FLOYDE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 807901I

DRY GAS EXPIRY DATE: 03/21/2010

VIOLATION CODE: DUI

ARREST TIME: 21:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 03/15/2010 21:39:25

---

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.082	21:33
Air Blank	0.000	21:34
Subject Sample #1	0.000	21:35
Breath Volume	3.394	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample #2	0.000	21:38
Breath Volume	2.511	21:38
Air Blank	0.000	21:38
Control Test	0.081	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

---

Data Download Date/Time 03/24/2010 17:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TOBAR, FRANK, J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BENEDITH, HILDA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 807901I

DRY GAS EXPIRY DATE: 03/21/2010

VIOLATION CODE: DUI

ARREST TIME: 23:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VILLANE

DATE/TIME OF TEST: 03/17/2010 00:20:13

---

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:14
Subject Sample #1	0.091	00:15
Breath Volume	1.835	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample #2	0.086	00:18
Breath Volume	1.703	00:18
Air Blank	0.000	00:19
Control Test	0.081	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

---

Data Download Date/Time 03/24/2010 17:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MORGAN, DARIN, P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BIRMER, S

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 807901I

DRY GAS EXPIRY DATE: 03/21/2010

VIOLATION CODE: DUI

ARREST TIME: 16:57

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/20/2010 18:32:41

---

Diagnostics Check	OK	18:26
Air Blank	0.000	18:27
Control Test	0.082	18:27
Air Blank	0.000	18:27
Subject Sample #1	0.000	18:28
Breath Volume	2.847	18:28
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample #2	0.000	18:31
Breath Volume	1.464	18:31
Air Blank	0.000	18:31
Control Test	0.081	18:32
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

---

Data Download Date/Time 03/24/2010 17:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MORGAN, DARIN, P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: NIEVES, J

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 07:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: WELDON

DATE/TIME OF TEST: 03/30/2010 09:17:36

---

Diagnostics Check	OK	09:09
Air Blank	0.000	09:10
Control Test	0.080	09:10
Air Blank	0.000	09:11
Subject Sample #1	0.000	09:12
Breath Volume	1.734	09:12
Air Blank	0.000	09:13
Air Blank	0.000	09:15
Subject Sample #2	0.000	09:16
Breath Volume	2.015	09:16
Air Blank	0.000	09:16
Control Test	0.080	09:17
Air Blank	0.000	09:17
Diagnostics Check	OK	09:17

---

Data Download Date/Time 05/01/2010 03:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MORGAN, DARIN, P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MORGAN, DARIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 14:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/30/2010 15:03:20

---

Diagnostics Check	OK	14:57
Air Blank	0.000	14:57
Control Test	0.080	14:58
Air Blank	0.000	14:58
Subject Sample #1	0.000	14:59
Breath Volume	3.187	14:59
Air Blank	0.000	14:59
Air Blank	0.000	15:01
Subject Sample #2	0.000	15:01
Breath Volume	2.234	15:01
Air Blank	0.000	15:02
Control Test	0.080	15:02
Air Blank	0.000	15:03
Diagnostics Check	OK	15:03

---

Data Download Date/Time 05/01/2010 03:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, LEIDY, Y  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: WALKOWIAK, BRIAN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 03/01/2010 03:32:39

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Subject Sample #1	0.194	03:27
Breath Volume	2.894	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample #2	0.184	03:30
Breath Volume	1.472	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIEL, PATRICK, J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MARLOW, BRANDON  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: DE PINA

DATE/TIME OF TEST: 03/01/2010 04:49:32

---

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Subject Sample #1	0.191	04:44
Breath Volume	2.292	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample #2	0.182	04:47
Breath Volume	1.554	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KONIECZKI, DOMINIC, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KONIECZKI, DOMINIC  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: BUNTING

DATE/TIME OF TEST: 03/01/2010 20:57:51

---

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Subject Sample #1	0.000	20:53
Breath Volume	1.308	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample #2	0.000	20:56
Breath Volume	1.371	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:57
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES, GLENN, H  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: JONES, GLENN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: BAYZIE

DATE/TIME OF TEST: 03/01/2010 22:21:18

---

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.079	22:15
Air Blank	0.000	22:16
Subject Sample #1	0.058	22:16
Breath Volume	1.390	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample #2	0.061	22:19
Breath Volume	2.578	22:19
Air Blank	0.000	22:20
Control Test	0.080	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES, GLENN, H  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: JONES, GLENN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MCCULLOUGH

DATE/TIME OF TEST: 03/02/2010 02:33:54

---

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.079	02:17	
Air Blank	0.000	02:18	
Subject Sample #1	0.174	02:21	
Breath Volume	1.457	02:21	
Air Blank	0.000	02:22	
Air Blank	0.000	02:23	
Subject Sample #2	0.145	02:27	Volume Not Met
Breath Volume	0.496	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:29	
Subject Sample #3	0.175	02:31	
Breath Volume	2.128	02:31	
Air Blank	0.000	02:32	
Control Test	0.074	02:33	Control Outside Tolerance
Air Blank	0.000	02:33	

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ, SUNELLY, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ESTERLE, HUGH  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: DE MOULIN

DATE/TIME OF TEST: 03/02/2010 03:06:16

---

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.076	02:59
Air Blank	0.000	03:00
Subject Sample #1	0.179	03:01
Breath Volume	2.757	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample #2	0.176	03:04
Breath Volume	2.656	03:04
Air Blank	0.000	03:05
Control Test	0.077	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOOTH, MATTHEW, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCHULZE, MATTHEW  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 03/03/2010 03:59:31

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample #1	0.087	03:54
Breath Volume	1.972	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample #2	0.084	03:57
Breath Volume	2.292	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS, MATTHEW, K  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: ABBEY, MORGAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 03/04/2010 01:39:34

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:32
Subject Sample #1	0.143	01:33
Breath Volume	2.449	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample #2	0.142	01:37
Breath Volume	2.410	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELSAID, MARK, K  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PIGNATARO, PAUL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 03/04/2010 02:43:17

---

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample #1	0.203	02:37
Breath Volume	2.562	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.205	02:41
Breath Volume	2.601	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS, GEORGE, T  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMYRNIOS, GEORGE  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BOUDRIE

DATE/TIME OF TEST: 03/04/2010 03:45:28

---

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Subject Sample #1	0.169	03:40
Breath Volume	3.734	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample #2	0.161	03:43
Breath Volume	2.484	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REISINGER, ANDREAS,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BEATTY, CHARLES  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 20:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: JELENIC

DATE/TIME OF TEST: 03/04/2010 22:03:55

---

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:58	
Control Test	0.080	21:59	
Air Blank	0.000	21:59	
Subject Sample #1	0.198	22:02	
Breath Volume	1.558	22:02	
Air Blank		22:03	Purge Fail
Air Blank	0.000	22:03	

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REISINGER, ANDREAS,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BEATTY, CHARLES  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 20:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: JELENIC

DATE/TIME OF TEST: 03/04/2010 22:22:06

---

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.075	22:13
Air Blank	0.000	22:13
Subject Sample #1	0.188	22:16
Breath Volume	2.003	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:18
Subject Sample #2	0.199	22:20
Breath Volume	2.109	22:20
Air Blank	0.000	22:21
Control Test	0.075	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REISINGER, ANDREAS,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PATTON, ROB  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 03/04/2010 22:34:03

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.077	22:28
Air Blank	0.000	22:29
Subject Sample #1	0.125	22:29
Breath Volume	2.566	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample #2	0.113	22:32
Breath Volume	1.519	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:34

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELSAID, MARK, K  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ALLEN, BRIAN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/05/2010 00:09:35

---

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Subject Sample #1	0.263	00:04
Breath Volume	3.500	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample #2	0.273	00:07
Breath Volume	3.601	00:07
Air Blank	0.000	00:08
Control Test	0.079	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEYSE, JAIME, M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEMOS, M  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 04:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/05/2010 06:07:36

---

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.080	06:01
Air Blank	0.000	06:02
Subject Sample #1	0.189	06:02
Breath Volume	1.488	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample #2	0.201	06:05
Breath Volume	1.773	06:05
Air Blank	0.000	06:06
Control Test	0.079	06:06
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIEROT, JEFFREY, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PIEROT, JEFFREY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: SANCHEZ-LEDEZMA

DATE/TIME OF TEST: 03/05/2010 23:56:53

---

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:50
Subject Sample #1	0.187	23:52
Breath Volume	2.246	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample #2	0.195	23:55
Breath Volume	1.355	23:55
Air Blank	0.000	23:55
Control Test	0.077	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ, SUNELLY, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GAGNON, ALAIN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 16:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 03/06/2010 17:27:48

---

Diagnostics Check	OK	17:19
Air Blank	0.000	17:20
Control Test	0.080	17:20
Air Blank	0.000	17:21
Subject Sample #1	0.206	17:22
Breath Volume	2.554	17:22
Air Blank	0.000	17:23
Air Blank	0.000	17:24
Subject Sample #2	0.214	17:25
Breath Volume	3.097	17:25
Air Blank	0.000	17:26
Control Test	0.076	17:27
Air Blank	0.000	17:27
Diagnostics Check	OK	17:27

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOOTH, MATTHEW, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: FACCILONGA, LESLIE  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 16:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: GREENWOOD

DATE/TIME OF TEST: 03/06/2010 17:36:53

---

Diagnostics Check	OK	17:30
Air Blank	0.000	17:30
Control Test	0.077	17:31
Air Blank	0.000	17:31
Subject Sample #1	0.280	17:32
Breath Volume	1.242	17:32
Air Blank	0.000	17:33
Air Blank	0.000	17:34
Subject Sample #2	0.281	17:35
Breath Volume	1.335	17:35
Air Blank	0.000	17:35
Control Test	0.079	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIEROT, JEFFREY, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PIEROT, JEFFREY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GARRICK

DATE/TIME OF TEST: 03/07/2010 01:17:36

---

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample #1	0.000	01:13
Breath Volume	2.031	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample #2	0.000	01:16
Breath Volume	2.039	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIEROT, JEFFREY, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: NILAND, JEREMY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/07/2010 02:26:13

---

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.080	02:22	
Air Blank	0.000	02:22	
Subject Sample #1		02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.080	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BELL, BRAD  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: CATLIN

DATE/TIME OF TEST: 03/07/2010 21:12:18

---

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.080	21:04
Air Blank	0.000	21:04
Subject Sample #1	0.114	21:07
Breath Volume	1.621	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample #2	0.122	21:10
Breath Volume	2.140	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, SELENA, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEE, SELENA  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/07/2010 22:45:23

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Subject Sample #1	0.083	22:40
Breath Volume	1.398	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample #2	0.076	22:43
Breath Volume	1.257	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HILL, SCOTT  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HOENIE

DATE/TIME OF TEST: 03/08/2010 04:19:43

---

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:14
Air Blank	0.000	04:14
Subject Sample #1	0.124	04:15
Breath Volume	2.320	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.123	04:18
Breath Volume	2.917	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, LEIDY, Y  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: METZGER, MICHAEL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 16:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 03/09/2010 18:24:09

---

Diagnostics Check	OK	18:17
Air Blank	0.000	18:18
Control Test	0.080	18:18
Air Blank	0.000	18:19
Subject Sample #1	0.007	18:19
Breath Volume	3.839	18:19
Air Blank	0.000	18:20
Air Blank	0.000	18:22
Subject Sample #2	0.000	18:22
Breath Volume	3.503	18:22
Air Blank	0.000	18:23
Control Test	0.080	18:23
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIEL, PATRICK, J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: JERZ, DANIEL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 18:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SMERALDO

DATE/TIME OF TEST: 03/09/2010 20:35:17

---

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.081	20:29
Air Blank	0.000	20:30
Subject Sample #1	0.000	20:30
Breath Volume	2.308	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample #2	0.000	20:33
Breath Volume	2.285	20:33
Air Blank	0.000	20:34
Control Test	0.081	20:34
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, LEIDY, Y  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MORENCY, DANIEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 20:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: COWEN

DATE/TIME OF TEST: 03/09/2010 21:03:53

---

Diagnostics Check	OK	20:57
Air Blank	0.000	20:57
Control Test	0.081	20:58
Air Blank	0.000	20:58
Subject Sample #1	0.000	20:59
Breath Volume	1.996	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample #2	0.000	21:02
Breath Volume	1.578	21:02
Air Blank	0.000	21:02
Control Test	0.080	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, LEIDY, Y  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CRAVEN, LISA  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: DEMARIO

DATE/TIME OF TEST: 03/09/2010 21:16:51

---

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.080	21:10
Air Blank	0.000	21:11
Subject Sample #1	0.187	21:12
Breath Volume	2.433	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample #2	0.170	21:15
Breath Volume	1.980	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD, LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD, LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: DA PRATO

DATE/TIME OF TEST: 03/10/2010 00:36:43

---

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.080	00:27	
Air Blank	0.000	00:27	
Subject Sample #1	0.218	00:29	
Breath Volume	1.765	00:29	
Air Blank	0.000	00:30	
Air Blank	0.000	00:31	
Subject Sample #2	0.170	00:35	Volume Not Met
Breath Volume	0.882	00:35	
Air Blank	0.000	00:36	
Air Blank	0.000	00:36	

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD, LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD, LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: DA PRATO

DATE/TIME OF TEST: 03/10/2010 00:48:28

---

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:40
Subject Sample #1	0.227	00:40
Breath Volume	2.191	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample #2	0.190	00:43
Breath Volume	1.359	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample #3	0.229	00:46
Breath Volume	1.652	00:46
Air Blank	0.000	00:47
Control Test	0.078	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KONIECZKI, DOMINIC, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KONIECZKI, DOMINIC  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HENEHAN

DATE/TIME OF TEST: 03/11/2010 20:27:13

---

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.081	20:21
Air Blank	0.000	20:22
Subject Sample #1	0.120	20:22
Breath Volume	1.964	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample #2	0.114	20:25
Breath Volume	1.703	20:25
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOOTH, MATTHEW, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GORDON, JOHN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: OLINGER

DATE/TIME OF TEST: 03/11/2010 21:50:30

---

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.080	21:45
Air Blank	0.000	21:45
Subject Sample #1	0.000	21:46
Breath Volume	3.246	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample #2	0.000	21:49
Breath Volume	3.386	21:49
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

---

Data Download Date/Time 03/12/2010 10:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEINBERG, MAX, H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: WEINBERG, MAX  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 09:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 03/12/2010 10:52:41

---

Diagnostics Check	OK	10:46
Air Blank	0.000	10:47
Control Test	0.081	10:47
Air Blank	0.000	10:47
Subject Sample #1	0.000	10:48
Breath Volume	3.890	10:48
Air Blank	0.000	10:48
Air Blank	0.000	10:50
Subject Sample #2	0.000	10:51
Breath Volume	3.601	10:51
Air Blank	0.000	10:51
Control Test	0.080	10:52
Air Blank	0.000	10:52
Diagnostics Check	OK	10:52

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELSAID, MARK, K  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KAMAUFF, DONAVON  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 03/12/2010 23:35:11

---

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.081	23:29
Air Blank	0.000	23:30
Subject Sample #1	0.122	23:30
Breath Volume	1.953	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:32
Subject Sample #2	0.123	23:33
Breath Volume	1.859	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD, LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD, LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: WACHA

DATE/TIME OF TEST: 03/13/2010 00:30:54

---

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:23
Subject Sample #1	0.114	00:25
Breath Volume	1.210	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample #2	0.113	00:29
Breath Volume	1.429	00:29
Air Blank	0.000	00:29
Control Test	0.078	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES, GLENN, H  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: JONES, GLENN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ZEPF

DATE/TIME OF TEST: 03/13/2010 00:43:06

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:38
Subject Sample #1	0.242	00:38
Breath Volume	2.253	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample #2	0.240	00:41
Breath Volume	2.089	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, LEIDY, Y  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEMOS, MATTHEW  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/13/2010 02:10:10

---

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:06	
Subject Sample #1		02:08	Subject Test Refused
Breath Volume	1.429	02:08	
Air Blank	0.000	02:09	
Control Test	0.078	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, LEIDY, Y  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ROSARIO, WILLIAM  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: LUBY

DATE/TIME OF TEST: 03/13/2010 02:23:02

---

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:15	
Control Test	0.078	02:15	
Air Blank	0.000	02:16	
Subject Sample #1	0.092	02:17	
Breath Volume	1.386	02:17	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample #2		02:21	Subject Test Refused
Breath Volume	0.726	02:21	
Air Blank	0.000	02:22	
Control Test	0.079	02:22	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:23	

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, SELENA, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEE, SELENA  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/13/2010 04:22:24

---

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample #1	0.000	04:18
Breath Volume	2.808	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample #2	0.000	04:21
Breath Volume	2.464	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRYOR, MANDI, L  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMART, D  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 13:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: WELTY

DATE/TIME OF TEST: 03/13/2010 14:22:46

---

Diagnostics Check	OK	14:14
Air Blank	0.000	14:15
Control Test	0.080	14:15
Air Blank	0.000	14:16
Subject Sample #1	0.273	14:17
Breath Volume	1.593	14:17
Air Blank	0.000	14:18
Air Blank	0.000	14:20
Subject Sample #2	0.257	14:20
Breath Volume	1.796	14:20
Air Blank	0.000	14:21
Control Test	0.075	14:22
Air Blank	0.000	14:22
Diagnostics Check	OK	14:22

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRYOR, MANDI, L  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CARROLL, E  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 16:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 03/13/2010 17:00:26

---

Diagnostics Check	OK	16:53
Air Blank	0.000	16:53
Control Test	0.079	16:54
Air Blank	0.000	16:54
Subject Sample #1	0.200	16:55
Breath Volume	2.050	16:55
Air Blank	0.000	16:56
Air Blank	0.000	16:58
Subject Sample #2	0.198	16:58
Breath Volume	1.660	16:58
Air Blank	0.000	16:59
Control Test	0.076	16:59
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, SELENA, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEE, SELENA  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 03/14/2010 03:33:41

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample #1	0.000	03:28
Breath Volume	2.398	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample #2	0.000	03:32
Breath Volume	3.515	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ALTMAN, JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: SPEAKER

DATE/TIME OF TEST: 03/14/2010 03:45:02

---

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:39
Air Blank	0.000	03:39
Subject Sample #1	0.138	03:40
Breath Volume	2.148	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:42
Subject Sample #2	0.136	03:43
Breath Volume	2.218	03:43
Air Blank	0.000	03:44
Control Test	0.077	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:45

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIDIO, JAMES, F  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DIDIO, JAMES  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/14/2010 04:39:27

---

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Subject Sample #1	0.207	04:34
Breath Volume	1.476	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample #2	0.204	04:37
Breath Volume	1.441	04:37
Air Blank	0.000	04:38
Control Test	0.076	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOMMERS, EDWARD, J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CANTRELL, JAROD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 11:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:34

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 03/15/2010 13:00:17

---

Diagnostics Check	OK	12:56	
Air Blank	0.000	12:56	
Control Test	0.081	12:57	
Air Blank	0.000	12:57	
Subject Sample #1		12:58	Subject Test Refused
Breath Volume	0.000	12:58	
Air Blank	0.000	12:59	
Control Test	0.081	12:59	
Air Blank	0.000	13:00	
Diagnostics Check	OK	13:00	

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOOTH, MATTHEW, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HURD, ERIC  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 21:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: TALLARDY

DATE/TIME OF TEST: 03/15/2010 22:21:51

---

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.081	22:16
Air Blank	0.000	22:16
Subject Sample #1	0.199	22:17
Breath Volume	3.707	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample #2	0.199	22:20
Breath Volume	4.046	22:20
Air Blank	0.000	22:20
Control Test	0.079	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIEROT, JEFFREY, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PIEROT, JEFFREY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: KUNST

DATE/TIME OF TEST: 03/16/2010 02:51:56

---

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:44
Air Blank	0.000	02:44
Subject Sample #1	0.134	02:47
Breath Volume	1.914	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample #2	0.128	02:50
Breath Volume	1.687	02:50
Air Blank	0.000	02:51
Control Test	0.077	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIEROT, JEFFREY, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BABOR, TIM  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 03/16/2010 04:13:05

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample #1	0.224	04:06
Breath Volume	1.441	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample #2	0.206	04:11
Breath Volume	1.527	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:13

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KONIECZKI, DOMINIC, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KONIECZKI, DOMINIC  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 03/17/2010 05:06:40

---

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Subject Sample #1	0.109	05:01
Breath Volume	1.632	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample #2	0.105	05:04
Breath Volume	1.554	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SARGIS, ANDREW, J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ANDUJAR, JOSEPH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 08:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 03/17/2010 09:44:24

---

Diagnostics Check	OK	09:37
Air Blank	0.000	09:38
Control Test	0.080	09:38
Air Blank	0.000	09:39
Subject Sample #1	0.121	09:39
Breath Volume	2.269	09:39
Air Blank	0.000	09:40
Air Blank	0.000	09:42
Subject Sample #2	0.125	09:42
Breath Volume	2.992	09:42
Air Blank	0.000	09:43
Control Test	0.080	09:43
Air Blank	0.000	09:44
Diagnostics Check	OK	09:44

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELSAID, MARK, K  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CHRISTENSON, ADAM  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 03/18/2010 03:28:09

---

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Subject Sample #1	0.186	03:23	Volume Not Met
Breath Volume	0.234	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:25	
Subject Sample #2		03:26	Subject Test Refused
Breath Volume	0.085	03:26	
Air Blank	0.000	03:27	
Control Test	0.077	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, LEIDY, Y  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MCHENRY, JASON  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: SWINDLE

DATE/TIME OF TEST: 03/18/2010 03:53:13

---

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:49	
Subject Sample #1		03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Air Blank	0.000	03:53	

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARGRANDER, ROBERT, C  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MARGRANDER, ROBERT  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/18/2010 04:04:52

---

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.081	03:58	
Air Blank	0.000	03:58	
Subject Sample #1	0.143	03:59	
Breath Volume	2.570	03:59	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample #2	0.139	04:03	
Breath Volume	2.039	04:03	
Air Blank		04:04	Purge Fail
Air Blank		04:04	Purge Fail

---

Data Download Date/Time 03/30/2010 13:14

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLAFLIN, DAVID, S  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: TESTER, TESTR  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 832401I  
DRY GAS EXPIRY DATE: 11/21/2010  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 12:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: TESTEE

DATE/TIME OF TEST: 03/18/2010 12:44:02

---

Diagnostics Check	OK	12:42	
Air Blank	0.000	12:43	
Control Test	0.062	12:43	Control Outside Tolerance
Air Blank	0.000	12:44	

---

Data Download Date/Time 03/30/2010 13:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AGUIAR, JESUS, R  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: JOHNSON, CHRIS  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 822501I  
DRY GAS EXPIRY DATE: 08/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: RHENALS

DATE/TIME OF TEST: 03/01/2010 00:45:55

---

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.076	00:39	
Air Blank	0.000	00:40	
Subject Sample #1	0.155	00:41	
Breath Volume	3.175	00:41	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample #2	0.155	00:44	
Breath Volume	4.105	00:44	
Air Blank	0.000	00:45	
Control Test	0.074	00:45	Control Outside Tolerance
Air Blank	0.000	00:45	

---

Data Download Date/Time 04/01/2010 11:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AGUIAR, JESUS, R  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: JOHNSON, CHRIS  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 822501I  
DRY GAS EXPIRY DATE: 08/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: RHENALS

DATE/TIME OF TEST: 03/01/2010 01:18:02

---

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.076	01:07	
Air Blank	0.000	01:07	
Subject Sample #1	0.099	01:10	Volume Not Met
Breath Volume	0.179	01:10	
Air Blank	0.000	01:11	
Air Blank	0.000	01:13	
Subject Sample #2	0.096	01:16	Volume Not Met
Breath Volume	1.082	01:16	
Air Blank	0.000	01:17	
Control Test	0.075	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:18	

---

Data Download Date/Time 04/01/2010 11:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ, JOSE, A  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: VIJIL, J  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 822501I  
DRY GAS EXPIRY DATE: 08/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FELTON

DATE/TIME OF TEST: 03/03/2010 04:08:30

---

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.076	04:06	
Air Blank	0.000	04:06	
Subject Sample #1		04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.075	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

---

Data Download Date/Time 04/01/2010 11:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ, JOSE, A  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: BUKENS, R  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 822501I  
DRY GAS EXPIRY DATE: 08/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/04/2010 04:12:00

---

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.076	04:08	
Air Blank	0.000	04:09	
Subject Sample #1		04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Control Test	0.076	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:12	

---

Data Download Date/Time 04/01/2010 11:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ, JOSE, A  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: VIJIL, J  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 822501I  
DRY GAS EXPIRY DATE: 08/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 03/04/2010 04:45:03

---

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.076	04:39	
Air Blank	0.000	04:39	
Subject Sample #1	0.171	04:40	
Breath Volume	1.875	04:40	
Air Blank	0.000	04:41	
Air Blank	0.000	04:42	
Subject Sample #2	0.168	04:43	
Breath Volume	1.679	04:43	
Air Blank	0.000	04:44	
Control Test	0.073	04:44	Control Outside Tolerance
Air Blank	0.000	04:45	

---

Data Download Date/Time 04/01/2010 11:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ, JOSE, A  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: JOHNSON, C  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 822501I  
DRY GAS EXPIRY DATE: 08/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ESCASENA

DATE/TIME OF TEST: 03/06/2010 01:02:15

---

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.076	01:00	
Air Blank	0.000	01:00	
Subject Sample #1		01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.076	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

---

Data Download Date/Time 04/01/2010 11:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ, JOSE, A  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: BARTRA, E  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 822501I  
DRY GAS EXPIRY DATE: 08/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 03/11/2010 04:40:39

---

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:38	
Control Test	0.075	04:38	
Air Blank	0.000	04:39	
Subject Sample #1		04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:39	
Control Test	0.076	04:40	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

---

Data Download Date/Time 04/01/2010 11:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, PAUL, Q  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: RODRIGUEZ, PAUL  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 822501I  
DRY GAS EXPIRY DATE: 08/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/14/2010 21:51:55

---

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:49	
Control Test	0.076	21:49	
Air Blank	0.000	21:50	
Subject Sample #1		21:50	Subject Test Refused
Breath Volume	0.000	21:50	
Air Blank	0.000	21:50	
Control Test	0.076	21:51	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

---

Data Download Date/Time 04/01/2010 11:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ, JOSE, A  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: JOHNSON, C  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 822501I  
DRY GAS EXPIRY DATE: 08/21/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/27/2010 02:33:28

---

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.076	02:22	
Air Blank	0.000	02:23	
Subject Sample #1	0.165	02:26	Volume Not Met
Breath Volume	0.023	02:26	
Air Blank	0.000	02:27	
Air Blank	0.000	02:28	
Subject Sample #2	0.172	02:31	Volume Not Met
Breath Volume	0.277	02:31	
Air Blank	0.000	02:32	
Control Test	0.072	02:32	Control Outside Tolerance
Air Blank	0.000	02:33	

---

Data Download Date/Time 04/01/2010 11:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BODLE, RANDY, S  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BODLE, RANDY  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ILLIANO

DATE/TIME OF TEST: 03/01/2010 02:49:17

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.083	02:42
Air Blank	0.000	02:43
Subject Sample #1	0.145	02:44
Breath Volume	1.324	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample #2	0.153	02:47
Breath Volume	1.800	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOLE, S, A

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: WIERZBA, R

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 02:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: NOWLIN

DATE/TIME OF TEST: 03/01/2010 03:01:59

---

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Subject Sample #1	0.235	02:55	
Breath Volume	2.132	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:57	
Subject Sample #2		03:00	Subject Test Refused
Breath Volume	0.347	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOLE, S, A

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: AIRADO, R

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 22:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/02/2010 00:00:16

---

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.083	23:57	
Air Blank	0.000	23:58	
Subject Sample #1		23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Control Test	0.084	23:59	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARSHALL, DAVID, O  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: MARSHALL, DAVID  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: KRIETEMEYER

DATE/TIME OF TEST: 03/05/2010 04:33:14

---

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.084	04:26
Air Blank	0.000	04:27
Subject Sample #1	0.059	04:28
Breath Volume	2.425	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample #2	0.051	04:31
Breath Volume	2.140	04:31
Air Blank	0.000	04:32
Control Test	0.084	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARRIS, OFFICER, M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: HARRIS, OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 03:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: VONKAENEL

DATE/TIME OF TEST: 03/06/2010 04:15:09

---

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.084	04:08
Air Blank	0.000	04:09
Subject Sample #1	0.193	04:10
Breath Volume	1.636	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample #2	0.190	04:13
Breath Volume	1.773	04:13
Air Blank	0.000	04:14
Control Test	0.082	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARRIS, OFFICER, M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: HARRIS, OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LARSEN

DATE/TIME OF TEST: 03/07/2010 03:06:45

---

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.083	03:00
Air Blank	0.000	03:01
Subject Sample #1	0.151	03:01
Breath Volume	1.500	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample #2	0.148	03:05
Breath Volume	1.621	03:05
Air Blank	0.000	03:05
Control Test	0.083	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOLE, S, A

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: HOLE, S

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 01:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/08/2010 02:16:51

---

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.084	02:10
Air Blank	0.000	02:11
Subject Sample #1	0.164	02:12
Breath Volume	2.210	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample #2	0.171	02:15
Breath Volume	1.707	02:15
Air Blank	0.000	02:15
Control Test	0.082	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOLE, S, A

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: HOLE, S

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 03:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: BAIN

DATE/TIME OF TEST: 03/08/2010 04:38:57

---

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.083	04:32
Air Blank	0.000	04:33
Subject Sample #1	0.182	04:34
Breath Volume	3.101	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample #2	0.177	04:37
Breath Volume	3.613	04:37
Air Blank	0.000	04:38
Control Test	0.082	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BODLE, RANDY, S  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BODLE, RANDY  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BOBRAN

DATE/TIME OF TEST: 03/08/2010 05:05:41

---

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.083	04:56
Air Blank	0.000	04:56
Subject Sample #1	0.187	04:57
Breath Volume	4.648	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample #2	0.153	05:00
Breath Volume	1.460	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:02
Subject Sample #3	0.192	05:04
Breath Volume	4.109	05:04
Air Blank	0.000	05:04
Control Test	0.082	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KARNES, BRYAN, J  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: EATON, N  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 14:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: KIBLER

DATE/TIME OF TEST: 03/08/2010 15:47:56

---

Diagnostics Check	OK	15:41
Air Blank	0.000	15:41
Control Test	0.084	15:41
Air Blank	0.000	15:42
Subject Sample #1	0.170	15:42
Breath Volume	1.363	15:42
Air Blank	0.000	15:43
Air Blank	0.000	15:45
Subject Sample #2	0.157	15:46
Breath Volume	1.414	15:46
Air Blank	0.000	15:47
Control Test	0.083	15:47
Air Blank	0.000	15:47
Diagnostics Check	OK	15:47

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BODLE, RANDY, S

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: AIRADO, R

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 22:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: DOBBIN

DATE/TIME OF TEST: 03/09/2010 23:37:43

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.084	23:31
Air Blank	0.000	23:32
Subject Sample #1	0.000	23:32
Breath Volume	5.746	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample #2	0.000	23:36
Breath Volume	7.210	23:36
Air Blank	0.000	23:36
Control Test	0.083	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RESLER, OFC.K., G  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: RESLER, OFC.K.  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: PANAGOULIAS

DATE/TIME OF TEST: 03/11/2010 02:46:10

---

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.083	02:42	
Air Blank	0.000	02:43	
Subject Sample #1		02:45	Sequence Aborted
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RESLER, OFC.K., G  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: RESLER, OFC.K.  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: PANAGOULIAS

DATE/TIME OF TEST: 03/11/2010 02:59:00

---

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.084	02:57	
Air Blank	0.000	02:58	
Subject Sample #1		02:58	Sequence Aborted
Breath Volume	0.000	02:58	
Air Blank	0.000	02:59	

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RESLER, OFC.K., G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: RESLER, OFC.K.

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 01:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: PANAGOULIAS

DATE/TIME OF TEST: 03/11/2010 03:46:02

---

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.083	03:40	
Air Blank	0.000	03:40	
Subject Sample #1	0.121	03:44	Volume Not Met
Breath Volume	0.757	03:44	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUEN, JONATHON, K  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: CAPPA, JONATHAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: FARNUM

DATE/TIME OF TEST: 03/11/2010 20:22:25

---

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:11	
Control Test	0.083	20:11	
Air Blank	0.000	20:12	
Subject Sample #1	0.251	20:13	No .020 agreement
Breath Volume	1.851	20:13	
Air Blank	0.000	20:13	
Air Blank	0.000	20:15	
Subject Sample #2	0.228	20:16	No .020 agreement
Breath Volume	1.757	20:16	
Air Blank	0.000	20:17	
Air Blank	0.000	20:19	
Subject Sample #3	0.274	20:20	No .020 agreement
Breath Volume	3.722	20:20	
Air Blank	0.000	20:21	
Control Test	0.081	20:21	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUEN, JONATHON, K  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: CAPPA, JONATHAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: FARNUM

DATE/TIME OF TEST: 03/11/2010 20:38:13

---

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:27	
Control Test	0.081	20:28	
Air Blank	0.000	20:28	
Subject Sample #1	0.266	20:30	No .020 agreement
Breath Volume	4.667	20:30	
Air Blank	0.000	20:30	
Air Blank	0.000	20:32	
Subject Sample #2	0.200	20:33	No .020 agreement
Breath Volume	1.289	20:33	
Air Blank	0.000	20:33	
Air Blank	0.000	20:35	
Subject Sample #3	0.244	20:36	No .020 agreement
Breath Volume	3.175	20:36	
Air Blank	0.000	20:37	
Control Test	0.081	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YUEN, JONATHON, K

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: CAPPA, JONATHAN

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 19:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: FARNUM

DATE/TIME OF TEST: 03/11/2010 21:04:51

---

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.081	20:58
Air Blank	0.000	20:59
Subject Sample #1	0.243	21:00
Breath Volume	3.328	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample #2	0.232	21:03
Breath Volume	2.339	21:03
Air Blank	0.000	21:03
Control Test	0.081	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILNE, THOMAS, P

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: HASTY, MICHAEL

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 01:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/12/2010 02:21:35

---

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.083	02:15
Air Blank	0.000	02:16
Subject Sample #1	0.191	02:16
Breath Volume	1.875	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.181	02:19
Breath Volume	1.562	02:19
Air Blank	0.000	02:20
Control Test	0.082	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BAILEY, DOUGLASS, J

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: BAILEY, DOUGLASS

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 07:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/12/2010 08:30:00

---

Diagnostics Check	OK	08:23
Air Blank	0.000	08:24
Control Test	0.083	08:24
Air Blank	0.000	08:25
Subject Sample #1	0.000	08:25
Breath Volume	2.515	08:25
Air Blank	0.000	08:26
Air Blank	0.000	08:28
Subject Sample #2	0.000	08:28
Breath Volume	2.332	08:28
Air Blank	0.000	08:29
Control Test	0.083	08:29
Air Blank	0.000	08:29
Diagnostics Check	OK	08:30

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WIERZBA, ROBERT, J

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: WIERZBA, ROBERT

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 01:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/15/2010 02:40:54

---

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.084	02:33
Air Blank	0.000	02:34
Subject Sample #1	0.237	02:35
Breath Volume	5.453	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample #2	0.255	02:39
Breath Volume	6.648	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RESLER, OFC.K., G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: RESLER, OFC.K.

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 03:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DALEY

DATE/TIME OF TEST: 03/16/2010 04:06:40

---

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.083	04:01
Air Blank	0.000	04:01
Subject Sample #1	0.176	04:02
Breath Volume	2.351	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample #2	0.172	04:05
Breath Volume	1.734	04:05
Air Blank	0.000	04:05
Control Test	0.082	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOLE, S, A

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: HOLE, S

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 02:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 03/17/2010 03:53:14

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.083	03:47
Air Blank	0.000	03:48
Subject Sample #1	0.205	03:48
Breath Volume	3.968	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample #2	0.186	03:51
Breath Volume	2.335	03:51
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TURPACK, T,

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: TURPACK, T

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 23:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: DONER

DATE/TIME OF TEST: 03/18/2010 00:41:08

---

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.084	00:35
Air Blank	0.000	00:35
Subject Sample #1	0.143	00:36
Breath Volume	5.039	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample #2	0.142	00:39
Breath Volume	5.742	00:39
Air Blank	0.000	00:40
Control Test	0.082	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILNE, THOMAS, P  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: MILNE, THOMAS  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: CRUZ ANGELES

DATE/TIME OF TEST: 03/19/2010 02:48:44

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.083	02:43
Air Blank	0.000	02:43
Subject Sample #1	0.112	02:44
Breath Volume	2.125	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample #2	0.109	02:47
Breath Volume	1.785	02:47
Air Blank	0.000	02:47
Control Test	0.082	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YUEN, JONATHON, K

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: SCHMIDT, JOHN

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 18:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 03/19/2010 19:09:01

---

Diagnostics Check	OK	19:01
Air Blank	0.000	19:02
Control Test	0.084	19:02
Air Blank	0.000	19:03
Subject Sample #1	0.212	19:04
Breath Volume	3.679	19:04
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample #2	0.215	19:07
Breath Volume	2.582	19:07
Air Blank	0.000	19:08
Control Test	0.082	19:08
Air Blank	0.000	19:08
Diagnostics Check	OK	19:09

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TURPACK, T,

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: TURPACK, T

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 17:02

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: OLIVERA

DATE/TIME OF TEST: 03/20/2010 17:53:38

---

Diagnostics Check	OK	17:47
Air Blank	0.000	17:47
Control Test	0.083	17:47
Air Blank	0.000	17:48
Subject Sample #1	0.144	17:48
Breath Volume	3.335	17:48
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample #2	0.147	17:52
Breath Volume	3.230	17:52
Air Blank	0.000	17:52
Control Test	0.083	17:53
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILNE, THOMAS, P

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: MILNE, THOMAS

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 01:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MIKULSKI

DATE/TIME OF TEST: 03/21/2010 02:41:09

---

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.083	02:34	
Air Blank	0.000	02:35	
Subject Sample #1	0.107	02:35	
Breath Volume	1.570	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:38	
Subject Sample #2		02:39	Slope Not Met
Breath Volume	0.578	02:39	
Air Blank	0.000	02:40	
Control Test	0.082	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILNE, THOMAS, P  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: MILNE, THOMAS  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MIKULSKI

DATE/TIME OF TEST: 03/21/2010 03:12:09

---

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.083	03:06
Air Blank	0.000	03:06
Subject Sample #1	0.092	03:07
Breath Volume	1.503	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample #2	0.090	03:10
Breath Volume	1.503	03:10
Air Blank	0.000	03:11
Control Test	0.083	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELLUCCI, ANDREW, A  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: JENKINS, RHOBBY  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: MULLIS

DATE/TIME OF TEST: 03/22/2010 16:23:48

---

Diagnostics Check	OK	16:20	
Air Blank	0.000	16:21	
Control Test	0.083	16:21	
Air Blank	0.000	16:22	
Subject Sample #1		16:22	Subject Test Refused
Breath Volume	0.000	16:22	
Air Blank	0.000	16:22	
Control Test	0.083	16:23	
Air Blank	0.000	16:23	
Diagnostics Check	OK	16:23	

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BODLE, RANDY, S

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: BODLE, RANDY

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 02:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: KLEINICK

DATE/TIME OF TEST: 03/23/2010 02:58:24

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.083	02:52
Air Blank	0.000	02:53
Subject Sample #1	0.143	02:53
Breath Volume	1.335	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.143	02:56
Breath Volume	1.449	02:56
Air Blank	0.000	02:57
Control Test	0.082	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLE, S, A  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: HOLE, S  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: NORRYKE

DATE/TIME OF TEST: 03/24/2010 05:14:51

---

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.083	05:09
Air Blank	0.000	05:09
Subject Sample #1	0.147	05:10
Breath Volume	2.949	05:10
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample #2	0.143	05:13
Breath Volume	3.000	05:13
Air Blank	0.000	05:13
Control Test	0.081	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RESLER, OFC.K., G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: RESLER, OFC.K.

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 05:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 03/24/2010 06:13:44

---

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.083	06:08
Air Blank	0.000	06:08
Subject Sample #1	0.220	06:09
Breath Volume	2.781	06:09
Air Blank	0.000	06:09
Air Blank	0.000	06:11
Subject Sample #2	0.208	06:12
Breath Volume	2.800	06:12
Air Blank	0.000	06:12
Control Test	0.081	06:13
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

---

Data Download Date/Time 03/26/2010 05:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILNE, THOMAS, P

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: MILNE, THOMAS

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 00:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 03/27/2010 01:42:52

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.083	01:36
Air Blank	0.000	01:37
Subject Sample #1	0.070	01:38
Breath Volume	2.207	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample #2	0.061	01:41
Breath Volume	1.929	01:41
Air Blank	0.000	01:41
Control Test	0.082	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

---

Data Download Date/Time 03/28/2010 03:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RENFROE, CLYDE, W

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: RENFROE, CLYDE

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 02:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WHITLEY

DATE/TIME OF TEST: 03/27/2010 03:30:49

---

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.083	03:28	
Air Blank	0.000	03:28	
Subject Sample #1		03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Control Test	0.084	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

---

Data Download Date/Time 03/28/2010 03:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAILEY, DOUGLASS, J  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: LAKE, K  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: STONEBROOK

DATE/TIME OF TEST: 03/28/2010 02:33:43

---

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.083	02:31	
Air Blank	0.000	02:31	
Subject Sample #1		02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.083	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

---

Data Download Date/Time 03/28/2010 03:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAILEY, DOUGLASS, J  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: LAKE, K  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: STONEBROOK

DATE/TIME OF TEST: 03/28/2010 02:43:02

---

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.083	02:37
Air Blank	0.000	02:37
Subject Sample #1	0.147	02:38
Breath Volume	2.148	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample #2	0.152	02:41
Breath Volume	1.859	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:43

---

Data Download Date/Time 03/28/2010 03:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BODLE, RANDY, S

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: BODLE, RANDY

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 22:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MALEK

DATE/TIME OF TEST: 03/28/2010 23:27:33

---

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.083	23:21
Air Blank	0.000	23:22
Subject Sample #1	0.251	23:23
Breath Volume	3.210	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample #2	0.249	23:25
Breath Volume	2.558	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

---

Data Download Date/Time 04/30/2010 04:38

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BODLE, R, S  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BODLE, R  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 668180  
DRY GAS EXPIRY DATE: 09/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: VIGEANT

DATE/TIME OF TEST: 03/30/2010 02:36:14

---

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.084	02:30
Air Blank	0.000	02:31
Subject Sample #1	0.153	02:31
Breath Volume	3.015	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample #2	0.146	02:34
Breath Volume	1.789	02:34
Air Blank	0.000	02:35
Control Test	0.082	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time 04/30/2010 04:38

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BODLE, RANDY, S

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: BODLE, RANDY

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 02:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/31/2010 03:32:02

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.083	03:26
Air Blank	0.000	03:26
Subject Sample #1	0.037	03:27
Breath Volume	3.375	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample #2	0.033	03:30
Breath Volume	3.417	03:30
Air Blank	0.000	03:31
Control Test	0.082	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:32

---

Data Download Date/Time 04/30/2010 04:38

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLINE, GLENN, A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: CLINE, GLENN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BEAL

DATE/TIME OF TEST: 03/01/2010 02:17:23

---

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Subject Sample #1	0.210	02:12
Breath Volume	2.378	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample #2	0.201	02:15
Breath Volume	3.253	02:15
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DE JESUS, JULIA,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: DE JESUS, JULIA  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 18:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: STURM

DATE/TIME OF TEST: 03/05/2010 19:50:28

---

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.078	19:44
Air Blank	0.000	19:45
Subject Sample #1	0.170	19:45
Breath Volume	1.558	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample #2	0.166	19:48
Breath Volume	1.347	19:48
Air Blank	0.000	19:49
Control Test	0.078	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: COLEMAN, WILLIAM  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ROMANENKO

DATE/TIME OF TEST: 03/06/2010 01:28:03

---

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.078	01:12	
Air Blank	0.000	01:12	
Subject Sample #1	0.227	01:15	Volume Not Met
Breath Volume	1.078	01:15	
Air Blank	0.000	01:16	
Air Blank	0.000	01:18	
Subject Sample #2	0.227	01:20	
Breath Volume	2.316	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample #3	0.203	01:26	Volume Not Met
Breath Volume	0.734	01:26	
Air Blank	0.000	01:27	
Control Test	0.076	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:28	

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: COLEMAN, WILLIAM  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ROMANENKO

DATE/TIME OF TEST: 03/06/2010 01:38:52

---

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.076	01:36	
Air Blank	0.000	01:37	
Subject Sample #1		01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.077	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPURLOCK, ROBERT, V  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SPURLOCK, ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GORDIANO-SANTANA

DATE/TIME OF TEST: 03/06/2010 02:42:51

---

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.078	02:39	
Air Blank	0.000	02:40	
Subject Sample #1	0.218	02:41	
Breath Volume	2.667	02:41	
Air Blank		02:42	Purge Fail
Air Blank	0.000	02:42	

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPURLOCK, ROBERT, V  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SPURLOCK, ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GORDIANO-SANTANA

DATE/TIME OF TEST: 03/06/2010 02:45:41

---

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.067	02:45	Control Outside Tolerance
Air Blank	0.000	02:45	

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPURLOCK, ROBERT, V  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SPURLOCK, ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GORDIANO-SANTANA

DATE/TIME OF TEST: 03/06/2010 02:58:50

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:53
Air Blank	0.000	02:53
Subject Sample #1	0.223	02:54
Breath Volume	2.109	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.218	02:57
Breath Volume	1.550	02:57
Air Blank	0.000	02:57
Control Test	0.078	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTHERLAND, MICHAEL, P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SUTHERLAND, MICHAEL J  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DEBLASIS LOSEY

DATE/TIME OF TEST: 03/06/2010 04:10:07

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Subject Sample #1	0.147	04:05
Breath Volume	1.839	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:07
Subject Sample #2	0.152	04:08
Breath Volume	1.441	04:08
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: COLEMAN, WILLIAM  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BALTAZAR-VALERIANO

DATE/TIME OF TEST: 03/06/2010 23:11:04

---

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:03
Subject Sample #1	0.167	23:05
Breath Volume	1.664	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample #2	0.166	23:09
Breath Volume	1.125	23:09
Air Blank	0.000	23:10
Control Test	0.078	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:11

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPURLOCK, ROBERT, V  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SPURLOCK, ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: GUZMAN-RAMIREZ

DATE/TIME OF TEST: 03/06/2010 23:28:52

---

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.075	23:23
Air Blank	0.000	23:24
Subject Sample #1	0.249	23:24
Breath Volume	1.949	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:26
Subject Sample #2	0.262	23:27
Breath Volume	2.019	23:27
Air Blank	0.000	23:27
Control Test	0.076	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: COLEMAN, WILLIAM  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 03/07/2010 02:17:34

---

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:15	
Control Test	0.079	02:15	
Air Blank	0.000	02:16	
Subject Sample #1		02:16	Subject Test Refused
Breath Volume	0.000	02:16	
Air Blank	0.000	02:16	
Control Test	0.079	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: COLEMAN, WILLIAM  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 03/07/2010 23:36:46

---

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:34	
Control Test	0.079	23:34	
Air Blank	0.000	23:35	
Subject Sample #1		23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Control Test	0.080	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: COLEMAN, WILLIAM  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/08/2010 02:50:12

---

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.079	02:39	
Air Blank	0.000	02:39	
Subject Sample #1	0.169	02:42	Volume Not Met
Breath Volume	0.171	02:42	
Air Blank	0.000	02:43	
Air Blank	0.000	02:45	
Subject Sample #2	0.186	02:45	
Breath Volume	1.156	02:45	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample #3	0.188	02:48	
Breath Volume	1.312	02:48	
Air Blank	0.000	02:49	
Control Test	0.075	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLINE, GLENN, A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: CLINE, GLENN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 20:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 03/08/2010 21:41:15

---

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.078	21:35
Air Blank	0.000	21:36
Subject Sample #1	0.080	21:36
Breath Volume	1.683	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample #2	0.082	21:39
Breath Volume	2.207	21:39
Air Blank	0.000	21:40
Control Test	0.078	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRONCZAK, CHERYLE, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: FRONCZAK, CHERYLE  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HEINICKE

DATE/TIME OF TEST: 03/09/2010 01:02:10

---

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Subject Sample #1	0.177	00:57
Breath Volume	1.890	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample #2	0.175	01:00
Breath Volume	1.921	01:00
Air Blank	0.000	01:01
Control Test	0.078	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLINE, GLENN, A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: CLINE, GLENN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SUMMERALL

DATE/TIME OF TEST: 03/09/2010 03:59:50

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.078	03:54
Air Blank	0.000	03:54
Subject Sample #1	0.170	03:55
Breath Volume	2.984	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:57
Subject Sample #2	0.166	03:58
Breath Volume	2.718	03:58
Air Blank	0.000	03:58
Control Test	0.078	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: METZGER, CHRISTOPHER, A  
BREATH TEST OPERATOR AGENCY: Palmetto PD  
ARRESTING OFFICER: METZGER, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palmetto PD  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 11:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 03/10/2010 12:07:35

---

Diagnostics Check	OK	12:00
Air Blank	0.000	12:01
Control Test	0.078	12:01
Air Blank	0.000	12:02
Subject Sample #1	0.227	12:02
Breath Volume	2.453	12:02
Air Blank	0.000	12:03
Air Blank	0.000	12:04
Subject Sample #2	0.212	12:05
Breath Volume	2.093	12:05
Air Blank	0.000	12:06
Control Test	0.077	12:06
Air Blank	0.000	12:07
Diagnostics Check	OK	12:07

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLINE, GLENN, A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: CLINE, GLENN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DEJESUS-SALAZAR

DATE/TIME OF TEST: 03/13/2010 00:34:42

---

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.078	00:28
Air Blank	0.000	00:28
Subject Sample #1	0.108	00:29
Breath Volume	1.851	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample #2	0.106	00:33
Breath Volume	2.289	00:33
Air Blank	0.000	00:33
Control Test	0.078	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TREADWAY, MARK, R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TREADWAY, MARK  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 831802I  
DRY GAS EXPIRY DATE: 11/22/2010  
VIOLATION CODE: DUI  
ARREST TIME: 12:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: IACONI

DATE/TIME OF TEST: 03/13/2010 02:01:17

---

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.078	01:50	
Air Blank	0.000	01:51	
Subject Sample #1	0.211	01:54	Volume Not Met
Breath Volume	0.906	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:56	
Subject Sample #2		01:59	RFI Detect
Breath Volume	0.000	01:59	
Air Blank	0.000	02:00	
Control Test	0.076	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: BRADENTON PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARTINEZ, LUIS, A

BREATH TEST OPERATOR AGENCY: Palmetto PD

ARRESTING OFFICER: MARTINEZ, LUIS

ARRESTING OFFICER AGENCY: Palmetto PD

LAST AGENCY INSPECTION: 02/10/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831802I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 01:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 03/13/2010 03:01:29

---

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:56
Air Blank	0.000	02:56
Subject Sample #1	0.180	02:56
Breath Volume	4.410	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.181	02:59
Breath Volume	4.683	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

---

Data Download Date/Time 03/17/2010 11:36

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KNOBLOCH, WILLIAM, T

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: STRICKLAND, STEVE

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 02/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 07:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 03/06/2010 08:25:08

---

Diagnostics Check	OK	08:18
Air Blank	0.000	08:19
Control Test	0.082	08:19
Air Blank	0.000	08:20
Subject Sample #1	0.184	08:20
Breath Volume	3.808	08:20
Air Blank	0.000	08:21
Air Blank	0.000	08:23
Subject Sample #2	0.181	08:23
Breath Volume	4.351	08:23
Air Blank	0.000	08:24
Control Test	0.084	08:24
Air Blank	0.000	08:25
Diagnostics Check	OK	08:25

---

Data Download Date/Time 03/21/2010 08:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: QUACKENBUSH, BRET, C

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: QUACKENBUSH, BRET

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 02/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 02:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 03/07/2010 03:01:14

---

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.084	02:55
Air Blank	0.000	02:56
Subject Sample #1	0.108	02:56
Breath Volume	3.144	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.099	02:59
Breath Volume	2.617	02:59
Air Blank	0.000	03:00
Control Test	0.082	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

---

Data Download Date/Time 03/21/2010 08:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 02/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 00:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/14/2010 01:28:42

---

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.082	01:18
Air Blank	0.000	01:19
Subject Sample #1	0.197	01:20
Breath Volume	2.019	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample #2	0.168	01:23
Breath Volume	1.363	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample #3	0.185	01:27
Breath Volume	2.121	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

---

Data Download Date/Time 03/21/2010 08:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 02/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 21:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 03/17/2010 22:03:57

---

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:01	
Control Test	0.082	22:01	
Air Blank	0.000	22:02	
Subject Sample #1		22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank	0.000	22:03	
Control Test	0.082	22:03	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:03	

---

Data Download Date/Time 03/21/2010 08:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 02/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 00:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ANTONIO

DATE/TIME OF TEST: 03/18/2010 01:26:34

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.082	01:21
Air Blank	0.000	01:21
Subject Sample #1	0.090	01:22
Breath Volume	1.710	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample #2	0.092	01:25
Breath Volume	1.472	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time 03/21/2010 08:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 02/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HARGROVE

DATE/TIME OF TEST: 03/18/2010 02:44:08

---

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:37
Subject Sample #1	0.238	02:39
Breath Volume	2.496	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample #2	0.236	02:42
Breath Volume	1.652	02:42
Air Blank	0.000	02:43
Control Test	0.082	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time 03/21/2010 08:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BAKER, DONALD, P

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: BAKER, DONALD

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 21:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 03/29/2010 21:57:25

---

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.081	21:52
Air Blank	0.000	21:52
Subject Sample #1	0.178	21:53
Breath Volume	2.742	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample #2	0.188	21:55
Breath Volume	2.906	21:55
Air Blank	0.000	21:56
Control Test	0.082	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

---

Data Download Date/Time 04/18/2010 08:44

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: QUACKENBUSH, BRET, C

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: QUACKENBUSH, BRET

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 00:43

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 03/31/2010 01:26:49

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.082	01:21
Air Blank	0.000	01:22
Subject Sample #1	0.000	01:22
Breath Volume	3.109	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample #2	0.000	01:25
Breath Volume	2.800	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time 04/18/2010 08:44

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAHLER, JOSHUA,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: MAHLER, JOSHUA  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/27/2010 01:24:54

---

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:19
Subject Sample #1	0.165	01:20
Breath Volume	4.011	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample #2	0.163	01:23
Breath Volume	3.347	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time 04/29/2010 02:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAHLER, JOSHUA,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: AUGUSTE, MIA  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: OWCZAREK

DATE/TIME OF TEST: 03/27/2010 01:44:54

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample #1	0.183	01:40
Breath Volume	2.167	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample #2	0.190	01:43
Breath Volume	1.210	01:43
Air Blank	0.000	01:44
Control Test	0.075	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

---

Data Download Date/Time 04/29/2010 02:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAHLER, JOSHUA,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: MAHLER, JOSHUA  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SORGENFREY

DATE/TIME OF TEST: 03/28/2010 02:02:24

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:57
Air Blank	0.000	01:57
Subject Sample #1	0.204	01:58
Breath Volume	2.562	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample #2	0.187	02:00
Breath Volume	2.195	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time 04/29/2010 02:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS, OLIVE, M  
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD  
ARRESTING OFFICER: WATSON, MALCOLM  
ARRESTING OFFICER AGENCY: NAS Jacksonville PD  
LAST AGENCY INSPECTION: 02/25/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/25/2010 02:41:18

---

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.082	02:39	
Air Blank	0.000	02:39	
Subject Sample #1		02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.083	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

---

Data Download Date/Time 03/27/2010 01:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DATKA, JOSHUA, L  
BREATH TEST OPERATOR AGENCY: NAS Key West  
ARRESTING OFFICER: GOULDING, MARK  
ARRESTING OFFICER AGENCY: NAS Key West  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: OTHER  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PIERREPHILIPPE

DATE/TIME OF TEST: 03/11/2010 00:06:01

---

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:01
Subject Sample #1	0.000	00:01
Breath Volume	1.660	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample #2	0.000	00:04
Breath Volume	1.640	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:06

---

Data Download Date/Time 04/22/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DATKA, JOSHUA, L  
BREATH TEST OPERATOR AGENCY: NAS Key West  
ARRESTING OFFICER: MONTGOMERY, ROBERT  
ARRESTING OFFICER AGENCY: NAS Key West  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: OTHER  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: FIGUEROAGUZMAN

DATE/TIME OF TEST: 03/11/2010 00:14:20

---

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.079	00:08
Air Blank	0.000	00:09
Subject Sample #1	0.027	00:09
Breath Volume	2.460	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample #2	0.027	00:12
Breath Volume	2.769	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

---

Data Download Date/Time 04/22/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DATKA, JOSHUA, L  
BREATH TEST OPERATOR AGENCY: NAS Key West  
ARRESTING OFFICER: HOGAN, CODY  
ARRESTING OFFICER AGENCY: NAS Key West  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: ADMIN  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/11/2010 02:27:14

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:22
Subject Sample #1	0.231	02:22
Breath Volume	1.136	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample #2	0.225	02:25
Breath Volume	1.511	02:25
Air Blank	0.000	02:26
Control Test	0.077	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time 04/22/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONROY, ALEXANDRA, P  
BREATH TEST OPERATOR AGENCY: NAS Key West  
ARRESTING OFFICER: REYES, RICHARD  
ARRESTING OFFICER AGENCY: NAS Key West  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 03/17/2010 01:27:11

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:22
Subject Sample #1	0.150	01:22
Breath Volume	2.308	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample #2	0.149	01:25
Breath Volume	2.484	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

---

Data Download Date/Time 04/22/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUITT, JERRAMI, L  
BREATH TEST OPERATOR AGENCY: NAS Key West  
ARRESTING OFFICER: DIXION, HOWARD  
ARRESTING OFFICER AGENCY: NAS Key West  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: ADMIN  
ARREST TIME: 04:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 03/18/2010 06:01:50

---

Diagnostics Check	OK	05:48	
Air Blank	0.000	05:49	
Control Test	0.079	05:49	
Air Blank	0.000	05:50	
Subject Sample #1	0.216	05:51	
Breath Volume	1.601	05:51	
Air Blank	0.000	05:52	
Air Blank	0.000	05:54	
Subject Sample #2	0.213	05:57	Volume Not Met
Breath Volume	1.011	05:57	
Air Blank	0.000	05:58	
Air Blank	0.000	05:59	
Subject Sample #3	0.221	06:00	
Breath Volume	1.781	06:00	
Air Blank	0.000	06:00	
Control Test	0.077	06:01	
Air Blank	0.000	06:01	
Diagnostics Check	OK	06:01	

---

Data Download Date/Time 04/22/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONROY, ALEXANDRA, P  
BREATH TEST OPERATOR AGENCY: NAS Key West  
ARRESTING OFFICER: TONER, WILLIAM  
ARRESTING OFFICER AGENCY: NAS Key West  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: OTHER  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 03/20/2010 04:53:11

---

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:47
Subject Sample #1	0.136	04:48
Breath Volume	2.746	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample #2	0.121	04:51
Breath Volume	1.308	04:51
Air Blank	0.000	04:52
Control Test	0.076	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

---

Data Download Date/Time 04/22/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONROY, ALEXANDRA, P  
BREATH TEST OPERATOR AGENCY: NAS Key West  
ARRESTING OFFICER: TONER, WILLIAM  
ARRESTING OFFICER AGENCY: NAS Key West  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: OTHER  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 03/20/2010 05:15:00

---

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.078	05:07
Air Blank	0.000	05:07
Subject Sample #1	0.198	05:09
Breath Volume	1.441	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample #2	0.188	05:13
Breath Volume	2.476	05:13
Air Blank	0.000	05:14
Control Test	0.076	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:15

---

Data Download Date/Time 04/22/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONROY, ALEXANDRA, P  
BREATH TEST OPERATOR AGENCY: NAS Key West  
ARRESTING OFFICER: DICKINSON, BRIAN  
ARRESTING OFFICER AGENCY: NAS Key West  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: OTHER  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BERVERLY

DATE/TIME OF TEST: 03/20/2010 05:27:43

---

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.078	05:21
Air Blank	0.000	05:21
Subject Sample #1	0.109	05:23
Breath Volume	3.769	05:23
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample #2	0.103	05:26
Breath Volume	3.449	05:26
Air Blank	0.000	05:26
Control Test	0.077	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

---

Data Download Date/Time 04/22/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DATKA, JOSHUA, L  
BREATH TEST OPERATOR AGENCY: NAS Key West  
ARRESTING OFFICER: GOULDING, MARC  
ARRESTING OFFICER AGENCY: NAS Key West  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: ADMIN  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GODI

DATE/TIME OF TEST: 03/21/2010 01:56:25

---

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Subject Sample #1	0.000	01:52
Breath Volume	2.761	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample #2	0.000	01:55
Breath Volume	2.832	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

---

Data Download Date/Time 04/22/2010 11:01

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: MIRABAL, YERANI

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 01:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/13/2010 01:49:21

---

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.080	01:47	
Air Blank	0.000	01:47	
Subject Sample #1		01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.081	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

---

Data Download Date/Time 03/22/2010 12:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MEDINA, ANTHONY,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: MEDINA, ANTHONY

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 21:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: PRIETO-MOYA

DATE/TIME OF TEST: 03/31/2010 22:27:13

---

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:19
Subject Sample #1	0.239	22:20
Breath Volume	3.726	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample #2	0.238	22:25
Breath Volume	2.382	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

---

Data Download Date/Time 04/29/2010 12:38

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DELIMA, G,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: DELIMA, G

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 21:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: LEIVA

DATE/TIME OF TEST: 03/11/2010 21:54:37

---

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.077	21:52	
Air Blank	0.000	21:53	
Subject Sample #1		21:53	Subject Test Refused
Breath Volume	0.000	21:53	
Air Blank	0.000	21:53	
Control Test	0.078	21:54	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:54	

---

Data Download Date/Time 03/22/2010 12:11

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MEDINA, ANTHONY,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: MEDINA, ANTHONY

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 22:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CHAFRE

DATE/TIME OF TEST: 03/13/2010 22:52:52

---

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:44
Subject Sample #1	0.194	22:47
Breath Volume	1.109	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample #2	0.197	22:51
Breath Volume	1.402	22:51
Air Blank	0.000	22:51
Control Test	0.077	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

---

Data Download Date/Time 03/22/2010 12:11

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PHILPOT, C,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: TILLMON, D

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 03:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GAYTAN

DATE/TIME OF TEST: 03/01/2010 04:34:05

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Subject Sample #1	0.151	04:28
Breath Volume	1.226	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample #2	0.155	04:32
Breath Volume	1.425	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGUEZ, P.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: DOMINGUEZ, P.

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 05:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: HERNANDEZ BARRETO

DATE/TIME OF TEST: 03/01/2010 06:53:00

---

Diagnostics Check	OK	06:46
Air Blank	0.000	06:46
Control Test	0.079	06:46
Air Blank	0.000	06:47
Subject Sample #1	0.190	06:48
Breath Volume	3.121	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:51
Subject Sample #2	0.187	06:51
Breath Volume	2.582	06:51
Air Blank	0.000	06:52
Control Test	0.079	06:52
Air Blank	0.000	06:52
Diagnostics Check	OK	06:53

---

Data Download Date/Time 03/22/2010 11:53

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASTRO, A, J

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: PELAEZ, D

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 04:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: MOJICA

DATE/TIME OF TEST: 03/02/2010 05:41:28

---

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:39	
Control Test	0.080	05:39	
Air Blank	0.000	05:39	
Subject Sample #1		05:40	Subject Test Refused
Breath Volume	0.000	05:40	
Air Blank	0.000	05:40	
Control Test	0.079	05:40	
Air Blank	0.000	05:41	
Diagnostics Check	OK	05:41	

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MEDINA, ANTHONY,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: MEDINA, ANTHONY

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 21:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 03/07/2010 21:28:50

---

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.080	21:26	
Air Blank	0.000	21:26	
Subject Sample #1		21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:27	
Control Test	0.079	21:28	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, L,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: MONIER, J

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 06:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: HERNANDEZ FERNANDEZ

DATE/TIME OF TEST: 03/08/2010 06:40:32

---

Diagnostics Check	OK	06:31	
Air Blank	0.000	06:32	
Control Test	0.080	06:32	
Air Blank	0.000	06:33	
Subject Sample #1	0.167	06:36	Volume Not Met
Breath Volume	0.042	06:36	
Air Blank	0.000	06:36	
Air Blank	0.000	06:38	
Subject Sample #2		06:39	Subject Test Refused
Breath Volume	0.000	06:39	
Air Blank	0.000	06:39	
Control Test	0.077	06:39	
Air Blank	0.000	06:40	
Diagnostics Check	OK	06:40	

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 822101I
BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A	DRY GAS EXPIRY DATE: 08/11/2010
BREATH TEST OPERATOR AGENCY: Hialeah PD	VIOLATION CODE: DUI
ARRESTING OFFICER: TILLMON, DEXTER	ARREST TIME: 03:20
ARRESTING OFFICER AGENCY: Hialeah PD	GENDER: Male
LAST AGENCY INSPECTION: 02/26/2010	OBSERVATION PERIOD BEGAN: 03:40
SUBJECT'S LAST NAME: BARBOSA	DATE/TIME OF TEST: 03/11/2010 04:06:13

---

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.079	04:04	
Air Blank	0.000	04:04	
Subject Sample #1		04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.079	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

---

Data Download Date/Time	03/22/2010	11:53
-------------------------	------------	-------

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JANVIER, FRITZ, G

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: MIGUEL, RUBEN

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 15:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: PEDRAJA MARRERO

DATE/TIME OF TEST: 03/13/2010 16:54:40

---

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:52	
Control Test	0.079	16:52	
Air Blank	0.000	16:53	
Subject Sample #1		16:53	Subject Test Refused
Breath Volume	0.000	16:53	
Air Blank	0.000	16:53	
Control Test	0.079	16:54	
Air Blank	0.000	16:54	
Diagnostics Check	OK	16:54	

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: CORREA, CESAR

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 03:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 03/14/2010 04:26:54

---

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.080	04:24	
Air Blank	0.000	04:25	
Subject Sample #1		04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.000	04:26	RFI Detect
Air Blank	0.000	04:26	

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: CORREA, CESAR

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 03:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 03/14/2010 04:31:11

---

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.079	04:29	
Air Blank	0.000	04:29	
Subject Sample #1		04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:30	
Control Test	0.079	04:30	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: CONSALVO, ESTEBAN

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 01:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: VERDECIA

DATE/TIME OF TEST: 03/15/2010 02:35:41

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29
Subject Sample #1	0.144	02:31
Breath Volume	1.636	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample #2	0.145	02:34
Breath Volume	1.503	02:34
Air Blank	0.000	02:34
Control Test	0.078	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ENRIQUEZ, NELSON,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: YIN, JAMES

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 00:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/17/2010 01:34:12

---

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Subject Sample #1	0.195	01:29
Breath Volume	1.367	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample #2	0.201	01:32
Breath Volume	2.089	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGUEZ, P.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: LARA, P.

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 01:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: RAMIREZ MARTINEZ

DATE/TIME OF TEST: 03/22/2010 02:30:16

---

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.079	02:26	
Air Blank	0.000	02:27	
Subject Sample #1		02:28	RFI Detect
Breath Volume	0.613	02:28	
Air Blank	0.000	02:29	
Control Test	0.080	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGUEZ, P.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: LARA, P.

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 01:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: RAMIREZ MARTINEZ

DATE/TIME OF TEST: 03/22/2010 02:40:06

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.078	02:33
Air Blank	0.000	02:33
Subject Sample #1	0.203	02:34
Breath Volume	1.183	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample #2	0.196	02:38
Breath Volume	1.984	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGUEZ, P.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: GUTIERREZ, G

ARRESTING OFFICER AGENCY: Medley PD

LAST AGENCY INSPECTION: 02/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 02:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: BELTRAN ENRIQUE

DATE/TIME OF TEST: 03/22/2010 03:34:22

---

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.078	03:32	
Air Blank	0.000	03:32	
Subject Sample #1		03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.079	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

---

Data Download Date/Time 03/22/2010 11:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGUEZ, P.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: MONIER, J.

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 02:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LOPEZ-FARHAN

DATE/TIME OF TEST: 03/23/2010 03:23:04

---

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Subject Sample #1	0.154	03:18
Breath Volume	1.449	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample #2	0.165	03:21
Breath Volume	1.535	03:21
Air Blank	0.000	03:22
Control Test	0.078	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:23

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: CONSALVO, ESTEBAN

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 02:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: DIAZ RUIZ

DATE/TIME OF TEST: 03/24/2010 03:54:29

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Subject Sample #1	0.201	03:49
Breath Volume	2.570	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample #2	0.200	03:52
Breath Volume	2.832	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PHILPOT, C,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: TILLMON, D

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 04:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 03/24/2010 05:15:22

---

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.080	05:01	
Air Blank	0.000	05:01	
Subject Sample #1	0.144	05:04	Volume Not Met
Breath Volume	0.878	05:04	
Air Blank	0.000	05:05	
Air Blank	0.000	05:07	
Subject Sample #2	0.144	05:08	
Breath Volume	2.457	05:08	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample #3	0.142	05:13	
Breath Volume	2.234	05:13	
Air Blank	0.000	05:14	
Control Test	0.078	05:14	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MEDINA, ANTHONY,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: MEDINA, ANTHONY

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 20:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 03/24/2010 20:57:33

---

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.080	20:51
Air Blank	0.000	20:51
Subject Sample #1	0.017	20:52
Breath Volume	1.761	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample #2	0.013	20:56
Breath Volume	1.949	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORREA, C.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: CONSALVO, E

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SALINAS-SOTO

DATE/TIME OF TEST: 03/26/2010 02:58:51

---

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.080	02:56	
Air Blank	0.000	02:56	
Subject Sample #1		02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.080	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PALACIOS, JAVIER, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: PALACIOS, JAVIER

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 18:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: REALPOZO

DATE/TIME OF TEST: 03/26/2010 19:15:09

---

Diagnostics Check	OK	19:12	
Air Blank	0.000	19:12	
Control Test	0.079	19:13	
Air Blank	0.000	19:13	
Subject Sample #1		19:13	Subject Test Refused
Breath Volume	0.000	19:13	
Air Blank	0.000	19:14	
Control Test	0.079	19:14	
Air Blank	0.000	19:15	
Diagnostics Check	OK	19:15	

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PHILPOT, C,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: LOPEZ/MIGUEL, N

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 21:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ROCHA

DATE/TIME OF TEST: 03/26/2010 22:34:13

---

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.079	22:32	
Air Blank	0.000	22:32	
Subject Sample #1		22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:33	
Control Test	0.079	22:33	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:34	

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORREA, C.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: PEREZ, F.

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 23:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RONDON

DATE/TIME OF TEST: 03/26/2010 23:53:59

---

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.079	23:51	
Air Blank	0.000	23:52	
Subject Sample #1		23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.079	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGUEZ, P.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: LOPEZ, N.

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 23:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: IBARRA

DATE/TIME OF TEST: 03/27/2010 00:28:06

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:22
Subject Sample #1	0.188	00:22
Breath Volume	2.992	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample #2	0.180	00:26
Breath Volume	3.503	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORREA, C.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: DIEDONNE, K.

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 23:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 03/27/2010 01:11:08

---

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:08	
Control Test	0.079	01:09	
Air Blank	0.000	01:09	
Subject Sample #1		01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:10	
Control Test	0.080	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORREA, C.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: DOMINGUEZ, P.

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PRIETO

DATE/TIME OF TEST: 03/27/2010 02:29:10

---

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.079	02:27	
Air Blank	0.000	02:27	
Subject Sample #1		02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORREA, C.,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: HERRERA, A.

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 01:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 03/27/2010 03:26:55

---

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Subject Sample #1		03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: CONSALVO, ESTEBAN

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 03:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: ZULETA

DATE/TIME OF TEST: 03/27/2010 04:45:43

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Subject Sample #1	0.201	04:40
Breath Volume	2.550	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample #2	0.185	04:44
Breath Volume	1.441	04:44
Air Blank	0.000	04:44
Control Test	0.078	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARDOSO, E,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: JEAN-FRANCOIS, S

ARRESTING OFFICER AGENCY: Opa Locka PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 04:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: VALDIVIA BOUZA

DATE/TIME OF TEST: 03/28/2010 10:44:41

---

Diagnostics Check	OK	10:36
Air Blank	0.000	10:36
Control Test	0.079	10:36
Air Blank	0.000	10:37
Subject Sample #1	0.128	10:39
Breath Volume	1.195	10:39
Air Blank	0.000	10:40
Air Blank	0.000	10:42
Subject Sample #2	0.129	10:43
Breath Volume	1.566	10:43
Air Blank	0.000	10:43
Control Test	0.077	10:44
Air Blank	0.000	10:44
Diagnostics Check	OK	10:44

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: CONSALVO, ESTEBAN

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 03/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 709352

DRY GAS EXPIRY DATE: 12/21/2011

VIOLATION CODE: DUI

ARREST TIME: 03:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: ALARCON-RODRIGUEZ

DATE/TIME OF TEST: 03/31/2010 05:00:33

---

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Subject Sample #1	0.126	04:55
Breath Volume	1.453	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample #2	0.126	04:59
Breath Volume	1.589	04:59
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

---

Data Download Date/Time 04/29/2010 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

INSTRUMENT REGISTERED TO: SURFSIDE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNSCHY, PATRICK, R  
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD  
ARRESTING OFFICER: GENTILE, JOHN  
ARRESTING OFFICER AGENCY: Surfside PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 670754  
DRY GAS EXPIRY DATE: 09/22/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: KRYVIANOK

DATE/TIME OF TEST: 03/07/2010 21:08:02

---

Diagnostics Check	OK	20:54	
Air Blank	0.000	20:54	
Control Test	0.082	20:54	
Air Blank	0.000	20:55	
Subject Sample #1	0.308	20:57	No .020 agreement
Breath Volume	2.371	20:57	
Air Blank	0.000	20:58	
Air Blank	0.000	21:00	
Subject Sample #2	0.276	21:03	Volume Not Met
Breath Volume	0.101	21:03	
Air Blank	0.000	21:04	
Air Blank	0.000	21:05	
Subject Sample #3	0.279	21:06	No .020 agreement
Breath Volume	3.132	21:06	
Air Blank	0.000	21:07	
Control Test	0.076	21:07	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:08	

---

Data Download Date/Time 03/30/2010 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

INSTRUMENT REGISTERED TO: SURFSIDE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNSCHY, PATRICK, R  
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD  
ARRESTING OFFICER: GENTILE, JOHN  
ARRESTING OFFICER AGENCY: Surfside PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 670754  
DRY GAS EXPIRY DATE: 09/22/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 03/08/2010 23:11:15

---

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:06	
Control Test	0.081	23:06	
Air Blank	0.000	23:07	
Subject Sample #1	0.214	23:09	
Breath Volume	3.125	23:09	
Air Blank		23:10	Purge Fail
Air Blank	0.000	23:11	

---

Data Download Date/Time 03/30/2010 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

INSTRUMENT REGISTERED TO: SURFSIDE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNSCHY, PATRICK, R  
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD  
ARRESTING OFFICER: GENTILE, JOHN  
ARRESTING OFFICER AGENCY: Surfside PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 670754  
DRY GAS EXPIRY DATE: 09/22/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: RICHARSON

DATE/TIME OF TEST: 03/08/2010 23:23:55

---

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.079	23:16
Air Blank	0.000	23:16
Subject Sample #1	0.207	23:17
Breath Volume	1.988	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample #2	0.201	23:21
Breath Volume	1.613	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

---

Data Download Date/Time 03/30/2010 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RUIZ, MICHAEL, A

BREATH TEST OPERATOR AGENCY: OTHER Local

ARRESTING OFFICER: BUFORD, THEOPHILUS

ARRESTING OFFICER AGENCY: OTHER Local

LAST AGENCY INSPECTION: 02/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 688217

DRY GAS EXPIRY DATE: 11/09/2011

VIOLATION CODE: DUI

ARREST TIME: 21:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 03/05/2010 22:14:11

---

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:08
Air Blank	0.000	22:09
Subject Sample #1	0.105	22:09
Breath Volume	1.683	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample #2	0.099	22:12
Breath Volume	1.203	22:12
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

---

Data Download Date/Time 03/31/2010 16:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ, LUIS, A  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: RIVERS, CLARENCE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 02/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: MILBURN

DATE/TIME OF TEST: 03/06/2010 07:14:55

---

Diagnostics Check	OK	07:13	
Air Blank	0.000	07:13	
Control Test	0.081	07:14	
Air Blank		07:14	Improper Sample
Air Blank	0.000	07:14	

---

Data Download Date/Time 03/31/2010 16:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ, LUIS, A  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: RIVERS, CLARENCE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 02/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: MILBURN

DATE/TIME OF TEST: 03/06/2010 07:23:30

---

Diagnostics Check	OK	07:17
Air Blank	0.000	07:17
Control Test	0.080	07:18
Air Blank	0.000	07:18
Subject Sample #1	0.198	07:18
Breath Volume	2.933	07:18
Air Blank	0.000	07:19
Air Blank	0.000	07:21
Subject Sample #2	0.197	07:21
Breath Volume	3.976	07:21
Air Blank	0.000	07:22
Control Test	0.079	07:22
Air Blank	0.000	07:23
Diagnostics Check	OK	07:23

---

Data Download Date/Time 03/31/2010 16:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ, LUIS, A  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: SMITH, JESSIE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 02/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: SUTHERLAND

DATE/TIME OF TEST: 03/06/2010 08:02:38

---

Diagnostics Check	OK	07:56
Air Blank	0.000	07:56
Control Test	0.080	07:57
Air Blank	0.000	07:57
Subject Sample #1	0.151	07:58
Breath Volume	3.609	07:58
Air Blank	0.000	07:58
Air Blank	0.000	08:00
Subject Sample #2	0.145	08:01
Breath Volume	3.859	08:01
Air Blank	0.000	08:01
Control Test	0.080	08:02
Air Blank	0.000	08:02
Diagnostics Check	OK	08:02

---

Data Download Date/Time 03/31/2010 16:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ, VERONICA, G  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: LOPEZ, VERONICA  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 02/02/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GUARDO

DATE/TIME OF TEST: 03/14/2010 00:21:22

---

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.081	00:15
Air Blank	0.000	00:16
Subject Sample #1	0.000	00:16
Breath Volume	1.847	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample #2	0.000	00:20
Breath Volume	1.675	00:20
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

---

Data Download Date/Time 03/31/2010 16:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: UCANAN, JAINA, L

BREATH TEST OPERATOR AGENCY: OTHER Local

ARRESTING OFFICER: BARRIOS, JAVIER

ARRESTING OFFICER AGENCY: OTHER Local

LAST AGENCY INSPECTION: 02/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 688217

DRY GAS EXPIRY DATE: 11/09/2011

VIOLATION CODE: DUI

ARREST TIME: 00:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DENNY

DATE/TIME OF TEST: 03/25/2010 01:54:13

---

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.080	01:52	
Air Blank	0.000	01:52	
Subject Sample #1		01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.081	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

---

Data Download Date/Time 03/31/2010 16:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: UCANAN, JAINA, L

BREATH TEST OPERATOR AGENCY: OTHER Local

ARRESTING OFFICER: UCANAN, JAINA

ARRESTING OFFICER AGENCY: OTHER Local

LAST AGENCY INSPECTION: 02/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 688217

DRY GAS EXPIRY DATE: 11/09/2011

VIOLATION CODE: DUI

ARREST TIME: 01:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 03/30/2010 02:31:59

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Subject Sample #1	0.273	02:27
Breath Volume	3.859	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample #2	0.274	02:30
Breath Volume	4.007	02:30
Air Blank	0.000	02:31
Control Test	0.076	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

---

Data Download Date/Time 03/31/2010 16:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 11/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BARTOL

DATE/TIME OF TEST: 03/05/2010 03:17:33

---

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank	0.000	03:15	
Subject Sample #1		03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

---

Data Download Date/Time 03/29/2010 10:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 11/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARCEL

DATE/TIME OF TEST: 03/14/2010 04:31:20

---

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.078	04:27	
Air Blank	0.000	04:28	
Subject Sample #1	0.223	04:29	
Breath Volume	1.191	04:29	
Air Blank		04:30	Purge Fail
Air Blank	0.000	04:31	

---

Data Download Date/Time 03/29/2010 10:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 11/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARCEL

DATE/TIME OF TEST: 03/14/2010 04:34:41

---

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:33	
Control Test	0.073	04:34	Control Outside Tolerance
Air Blank	0.000	04:34	

---

Data Download Date/Time 03/29/2010 10:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 11/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SHIPLEY

DATE/TIME OF TEST: 03/17/2010 23:57:18

---

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.077	23:56	
Air Blank		23:56	RFI Detect
Air Blank	0.000	23:57	

---

Data Download Date/Time 03/29/2010 10:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 11/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SHIPLEY

DATE/TIME OF TEST: 03/18/2010 00:03:28

---

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Subject Sample #1		00:02	RFI Detect
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.078	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

---

Data Download Date/Time 03/29/2010 10:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 11/21/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LICEA

DATE/TIME OF TEST: 03/19/2010 23:27:55

---

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:22
Air Blank	0.000	23:22
Subject Sample #1	0.130	23:23
Breath Volume	1.531	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample #2	0.132	23:26
Breath Volume	1.585	23:26
Air Blank	0.000	23:26
Control Test	0.077	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

---

Data Download Date/Time 03/29/2010 10:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONFESSORE, STEVE,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: BLACIO, GUSTAVO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 03/01/2010 04:06:48

---

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.078	04:04	
Air Blank	0.000	04:04	
Subject Sample #1		04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.079	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HORTON, REGINALD, B  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: WIRTH, PHILIP  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: MAHANY

DATE/TIME OF TEST: 03/04/2010 05:04:50

---

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:51	
Control Test	0.079	04:52	
Air Blank	0.000	04:52	
Subject Sample #1	0.104	04:56	Volume Not Met
Breath Volume	0.019	04:56	
Air Blank	0.000	04:56	
Air Blank	0.000	04:58	
Subject Sample #2	0.099	04:59	
Breath Volume	1.675	04:59	
Air Blank	0.000	05:00	
Air Blank	0.000	05:02	
Subject Sample #3	0.101	05:03	
Breath Volume	2.136	05:03	
Air Blank	0.000	05:03	
Control Test	0.076	05:04	
Air Blank	0.000	05:04	
Diagnostics Check	OK	05:04	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ, ROGER, E  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: CAMPOS, SERGIO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: DA SILVA

DATE/TIME OF TEST: 03/05/2010 00:08:38

---

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:03
Subject Sample #1	0.076	00:04
Breath Volume	2.480	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample #2	0.078	00:07
Breath Volume	4.394	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

---

Data Download Date/Time 03/30/2010 14:46

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELCASTILLO, GUILLERMO, R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DELCASTILLO, G  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CANNATA

DATE/TIME OF TEST: 03/06/2010 02:25:27

---

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample #1	0.011	02:21
Breath Volume	2.613	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample #2	0.009	02:24
Breath Volume	3.894	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELCASTILLO, GUILLERMO, R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: BERRIAN, J  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 03/06/2010 06:52:34

---

Diagnostics Check	OK	06:49	
Air Blank	0.000	06:50	
Control Test	0.079	06:50	
Air Blank	0.000	06:51	
Subject Sample #1		06:51	Subject Test Refused
Breath Volume	0.000	06:51	
Air Blank	0.000	06:51	
Control Test	0.079	06:51	
Air Blank	0.000	06:52	
Diagnostics Check	OK	06:52	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON, IAN, M  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: LESTER, REGINALD  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 17:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/06/2010 18:08:44

---

Diagnostics Check	OK	18:02
Air Blank	0.000	18:02
Control Test	0.078	18:03
Air Blank	0.000	18:03
Subject Sample #1	0.331	18:04
Breath Volume	3.164	18:04
Air Blank	0.000	18:04
Air Blank	0.000	18:06
Subject Sample #2	0.322	18:07
Breath Volume	3.183	18:07
Air Blank	0.000	18:07
Control Test	0.078	18:08
Air Blank	0.000	18:08
Diagnostics Check	OK	18:08

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON, IAN, M  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: MARTINEZ, ROGER  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: HERRICK

DATE/TIME OF TEST: 03/06/2010 19:48:37

---

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.078	19:41
Air Blank	0.000	19:41
Subject Sample #1	0.156	19:43
Breath Volume	1.824	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample #2	0.169	19:46
Breath Volume	3.039	19:46
Air Blank	0.000	19:47
Control Test	0.075	19:48
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELCASTILLO, GUILLERMO, R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DIONNE, M  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 03/07/2010 04:41:13

---

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:36
Subject Sample #1	0.181	04:36
Breath Volume	1.726	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample #2	0.186	04:39
Breath Volume	1.949	04:39
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDER, NEWELL, G  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: JONES, WAYNE  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GOTTLIEB

DATE/TIME OF TEST: 03/08/2010 03:44:14

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:39
Subject Sample #1	0.112	03:39
Breath Volume	1.765	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample #2	0.097	03:42
Breath Volume	2.234	03:42
Air Blank	0.000	03:43
Control Test	0.077	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIONNE, MARTIN,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DIONNE, MARTIN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 03/08/2010 06:06:25

---

Diagnostics Check	OK	05:59
Air Blank	0.000	06:00
Control Test	0.079	06:00
Air Blank	0.000	06:00
Subject Sample #1	0.144	06:01
Breath Volume	1.804	06:01
Air Blank	0.000	06:02
Air Blank	0.000	06:04
Subject Sample #2	0.138	06:04
Breath Volume	1.253	06:04
Air Blank	0.000	06:05
Control Test	0.079	06:05
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELCASTILLO, GUILLERMO, R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DELCASTILLO, GUILLERMO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 03/09/2010 04:49:52

---

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.078	04:44
Air Blank	0.000	04:44
Subject Sample #1	0.149	04:45
Breath Volume	1.261	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:47
Subject Sample #2	0.151	04:48
Breath Volume	1.128	04:48
Air Blank	0.000	04:48
Control Test	0.078	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYEGH, MICHELLE, M  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: VESKI, MICHAEL  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/09/2010 22:21:36

---

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.078	22:15
Air Blank	0.000	22:16
Subject Sample #1	0.000	22:17
Breath Volume	1.785	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample #2	0.000	22:20
Breath Volume	1.773	22:20
Air Blank	0.000	22:20
Control Test	0.078	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVAGNI, R, W  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SILVAGNI, R  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: TEIXEIRA

DATE/TIME OF TEST: 03/10/2010 04:36:07

---

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:31
Subject Sample #1	0.162	04:31
Breath Volume	2.816	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample #2	0.180	04:34
Breath Volume	1.996	04:34
Air Blank	0.000	04:35
Control Test	0.077	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVAGNI, R, W  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SILVAGNI, R  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 03/11/2010 04:18:41

---

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.079	04:16	
Air Blank	0.000	04:16	
Subject Sample #1		04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.078	04:18	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVAGNI, R, W  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: CAMPOS, S  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 03/11/2010 23:58:12

---

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.078	23:52
Air Blank	0.000	23:53
Subject Sample #1	0.000	23:53
Breath Volume	2.687	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample #2	0.000	23:56
Breath Volume	3.421	23:56
Air Blank	0.000	23:57
Control Test	0.077	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIONNE, MARTIN,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DIONNE, MARTIN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DOAR

DATE/TIME OF TEST: 03/12/2010 04:27:20

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Subject Sample #1	0.129	04:22
Breath Volume	1.527	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample #2	0.121	04:25
Breath Volume	1.171	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOCARRAS, PEDRO, A  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SHIMKO, RONALD  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: TIGERTAIL

DATE/TIME OF TEST: 03/12/2010 07:07:02

---

Diagnostics Check	OK	07:00
Air Blank	0.000	07:01
Control Test	0.078	07:01
Air Blank	0.000	07:01
Subject Sample #1	0.090	07:02
Breath Volume	2.835	07:02
Air Blank	0.000	07:03
Air Blank	0.000	07:05
Subject Sample #2	0.090	07:05
Breath Volume	3.039	07:05
Air Blank	0.000	07:06
Control Test	0.078	07:06
Air Blank	0.000	07:06
Diagnostics Check	OK	07:07

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LESTER, REGINALD, D  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: WYATT, PETER  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BOATENG

DATE/TIME OF TEST: 03/12/2010 20:53:09

---

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.078	20:46
Air Blank	0.000	20:47
Subject Sample #1	0.199	20:48
Breath Volume	1.597	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:50
Subject Sample #2	0.201	20:51
Breath Volume	1.605	20:51
Air Blank	0.000	20:52
Control Test	0.077	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOCARRAS, PEDRO, A  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: WIRTH, PHILLIP  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 03/13/2010 05:09:04

---

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:06	
Control Test	0.077	05:06	
Air Blank	0.000	05:07	
Subject Sample #1		05:07	Subject Test Refused
Breath Volume	0.000	05:07	
Air Blank	0.000	05:08	
Control Test	0.078	05:08	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:09	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOCARRAS, PEDRO, A  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DIONNE, MARTIN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: ALONSO GONZALEZ

DATE/TIME OF TEST: 03/13/2010 06:08:20

---

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.078	06:01
Air Blank	0.000	06:01
Subject Sample #1	0.177	06:03
Breath Volume	1.347	06:03
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample #2	0.161	06:06
Breath Volume	1.285	06:06
Air Blank	0.000	06:07
Control Test	0.077	06:07
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELCASTILLO, GUILLERMO, R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DELCASTILLO, G  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 03/14/2010 03:53:40

---

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.078	03:51	
Air Blank	0.000	03:52	
Subject Sample #1		03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.078	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WIRTH, PHILIP, A  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DIONNE, M  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: LINARES

DATE/TIME OF TEST: 03/14/2010 06:12:41

---

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:57	
Control Test	0.079	05:58	
Air Blank	0.000	05:58	
Subject Sample #1	0.116	06:01	Volume Not Met
Breath Volume	0.355	06:01	
Air Blank	0.000	06:02	
Air Blank	0.000	06:04	
Subject Sample #2	0.133	06:05	
Breath Volume	1.960	06:05	
Air Blank	0.000	06:05	
Air Blank	0.000	06:07	
Subject Sample #3	0.126	06:10	
Breath Volume	2.308	06:10	
Air Blank	0.000	06:11	
Control Test	0.079	06:12	
Air Blank	0.000	06:12	
Diagnostics Check	OK	06:12	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYEGH, MICHELLE, M  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SAYEGH, MICHELLE  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: CRAWFORD-ELISTIN

DATE/TIME OF TEST: 03/14/2010 20:46:16

---

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.078	20:40
Air Blank	0.000	20:40
Subject Sample #1	0.189	20:41
Breath Volume	1.523	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample #2	0.182	20:44
Breath Volume	1.390	20:44
Air Blank	0.000	20:45
Control Test	0.078	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LESTER, REGINALD, D  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SANCHEZ, ELIO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 13:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: SCHUPBACK

DATE/TIME OF TEST: 03/15/2010 14:12:50

---

Diagnostics Check	OK	14:09	
Air Blank	0.000	14:10	
Control Test	0.078	14:10	
Air Blank	0.000	14:11	
Subject Sample #1		14:11	Subject Test Refused
Breath Volume	0.000	14:11	
Air Blank	0.000	14:11	
Control Test	0.079	14:12	
Air Blank	0.000	14:12	
Diagnostics Check	OK	14:12	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUNCAS, ANDREW, J  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: CALLAN, TONY  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SAN PEDRO

DATE/TIME OF TEST: 03/16/2010 05:56:55

---

Diagnostics Check	OK	05:50
Air Blank	0.000	05:51
Control Test	0.078	05:51
Air Blank	0.000	05:52
Subject Sample #1	0.129	05:52
Breath Volume	2.746	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample #2	0.124	05:55
Breath Volume	1.746	05:55
Air Blank	0.000	05:56
Control Test	0.076	05:56
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VALDES, MANUEL,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: CHIN, ALAN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 03/17/2010 21:17:36

---

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.077	21:12
Air Blank	0.000	21:12
Subject Sample #1	0.207	21:13
Breath Volume	2.789	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample #2	0.209	21:16
Breath Volume	2.109	21:16
Air Blank	0.000	21:16
Control Test	0.078	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

---

Data Download Date/Time 03/30/2010 14:46

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOCARRAS, PEDRO, A  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SHIMKO, RONALD  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 03/18/2010 03:05:42

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:58
Subject Sample #1	0.000	03:01
Breath Volume	1.726	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample #2	0.000	03:04
Breath Volume	1.160	03:04
Air Blank	0.000	03:04
Control Test	0.078	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ, JOEL,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SOCARRAS, P  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BROSH

DATE/TIME OF TEST: 03/18/2010 05:16:08

---

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.078	05:00	
Air Blank	0.000	05:01	
Subject Sample #1	0.094	05:04	Volume Not Met
Breath Volume	0.273	05:04	
Air Blank	0.000	05:05	
Air Blank	0.000	05:07	
Subject Sample #2	0.110	05:09	
Breath Volume	1.207	05:09	
Air Blank	0.000	05:10	
Air Blank	0.000	05:12	
Subject Sample #3	0.112	05:14	
Breath Volume	2.390	05:14	
Air Blank	0.000	05:15	
Control Test	0.077	05:15	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIONNE, MARTIN,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DIONNE, MARTIN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/19/2010 05:22:55

---

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:20	
Control Test	0.078	05:20	
Air Blank	0.000	05:21	
Subject Sample #1		05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:22	
Control Test	0.078	05:22	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON, IAN, M  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SAN ROMAN, VICENTE  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/19/2010 20:29:16

---

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:26	
Control Test	0.078	20:27	
Air Blank	0.000	20:27	
Subject Sample #1		20:27	Subject Test Refused
Breath Volume	0.000	20:27	
Air Blank	0.000	20:28	
Control Test	0.078	20:28	
Air Blank	0.000	20:29	
Diagnostics Check	OK	20:29	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCVEY, GREGORY, B  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: GONZALEZ, J  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/20/2010 05:09:57

---

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.078	05:04
Air Blank	0.000	05:04
Subject Sample #1	0.133	05:05
Breath Volume	3.175	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:07
Subject Sample #2	0.134	05:08
Breath Volume	3.085	05:08
Air Blank	0.000	05:09
Control Test	0.078	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ, ROGER, E  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: ROBINSON, IAN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 03/20/2010 20:26:50

---

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.078	20:21
Air Blank	0.000	20:22
Subject Sample #1	0.096	20:22
Breath Volume	1.777	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:24
Subject Sample #2	0.092	20:25
Breath Volume	1.484	20:25
Air Blank	0.000	20:25
Control Test	0.076	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYEGH, MICHELLE, M  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SAYEGH, MICHELLE  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/21/2010 18:48:22

---

Diagnostics Check	OK	18:44	
Air Blank	0.000	18:44	
Control Test	0.078	18:45	
Air Blank	0.000	18:45	
Subject Sample #1		18:46	Subject Test Refused
Breath Volume	0.000	18:46	
Air Blank	0.000	18:47	
Control Test	0.078	18:47	
Air Blank	0.000	18:48	
Diagnostics Check	OK	18:48	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELCASTILLO, GUILLERMO, R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: KING, L  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SHACKELFORD

DATE/TIME OF TEST: 03/22/2010 04:51:31

---

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.078	04:46
Air Blank	0.000	04:46
Subject Sample #1	0.207	04:47
Breath Volume	1.863	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample #2	0.202	04:49
Breath Volume	1.550	04:49
Air Blank	0.000	04:50
Control Test	0.078	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, JOAQUIN, R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: BARRETO-VO, L  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 13:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:16

SUBJECT'S LAST NAME: NYE

DATE/TIME OF TEST: 03/23/2010 13:47:43

---

Diagnostics Check	OK	13:41
Air Blank	0.000	13:42
Control Test	0.078	13:42
Air Blank	0.000	13:42
Subject Sample #1	0.000	13:43
Breath Volume	2.398	13:43
Air Blank	0.000	13:43
Air Blank	0.000	13:45
Subject Sample #2	0.000	13:46
Breath Volume	2.343	13:46
Air Blank	0.000	13:46
Control Test	0.078	13:47
Air Blank	0.000	13:47
Diagnostics Check	OK	13:47

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVAGNI, R, W  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: WIRTH, P  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: PERRIN

DATE/TIME OF TEST: 03/25/2010 04:16:38

---

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.078	04:10
Air Blank	0.000	04:11
Subject Sample #1	0.133	04:11
Breath Volume	2.007	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample #2	0.127	04:15
Breath Volume	1.785	04:15
Air Blank	0.000	04:15
Control Test	0.077	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVAGNI, R, W  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: ESTRAVIZ, A  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SEDELBAUER

DATE/TIME OF TEST: 03/25/2010 05:02:49

---

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.078	04:59	
Air Blank	0.000	05:00	
Subject Sample #1		05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:01	
Control Test	0.078	05:02	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOCARRAS, PEDRO, A  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SILVAGNI, ROBERT  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: BAKHRU

DATE/TIME OF TEST: 03/25/2010 06:15:09

---

Diagnostics Check	OK	06:10	
Air Blank	0.000	06:11	
Control Test	0.079	06:11	
Air Blank	0.000	06:12	
Subject Sample #1		06:13	Subject Test Refused
Breath Volume	0.000	06:13	
Air Blank	0.000	06:14	
Control Test	0.079	06:14	
Air Blank	0.000	06:15	
Diagnostics Check	OK	06:15	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LESTER, REGINALD, D  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: VIDAL, ERROL  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 17:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 03/25/2010 18:12:40

---

Diagnostics Check	OK	18:06
Air Blank	0.000	18:06
Control Test	0.078	18:07
Air Blank	0.000	18:07
Subject Sample #1	0.000	18:08
Breath Volume	1.347	18:08
Air Blank	0.000	18:08
Air Blank	0.000	18:10
Subject Sample #2	0.000	18:11
Breath Volume	1.535	18:11
Air Blank	0.000	18:11
Control Test	0.079	18:12
Air Blank	0.000	18:12
Diagnostics Check	OK	18:12

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOMINGUEZ, ERIC,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SOCARRAS, PEDRO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: SORRENTO JR

DATE/TIME OF TEST: 03/26/2010 03:33:28

---

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Subject Sample #1	0.209	03:28
Breath Volume	2.175	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample #2	0.207	03:31
Breath Volume	2.570	03:31
Air Blank	0.000	03:32
Control Test	0.077	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HORTON, REGINALD, B  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: WIRTH, PHILIP  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 03/26/2010 05:18:56

---

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.079	05:12
Air Blank	0.000	05:12
Subject Sample #1	0.131	05:14
Breath Volume	1.824	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample #2	0.125	05:17
Breath Volume	1.675	05:17
Air Blank	0.000	05:18
Control Test	0.078	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOCARRAS, PEDRO, A  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DOMINGUEZ, ERIC  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 03/26/2010 07:12:25

---

Diagnostics Check	OK	07:06
Air Blank	0.000	07:06
Control Test	0.079	07:07
Air Blank	0.000	07:07
Subject Sample #1	0.193	07:08
Breath Volume	1.789	07:08
Air Blank	0.000	07:08
Air Blank	0.000	07:10
Subject Sample #2	0.191	07:10
Breath Volume	1.484	07:10
Air Blank	0.000	07:11
Control Test	0.078	07:11
Air Blank	0.000	07:12
Diagnostics Check	OK	07:12

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVAGNI, R, W  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SILVAGNI, R  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: BUSCH

DATE/TIME OF TEST: 03/27/2010 05:20:26

---

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.078	05:14
Air Blank	0.000	05:14
Subject Sample #1	0.127	05:15
Breath Volume	2.027	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample #2	0.118	05:18
Breath Volume	1.621	05:18
Air Blank	0.000	05:19
Control Test	0.077	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVAGNI, R, W  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: BLACIO, G  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: EISEL

DATE/TIME OF TEST: 03/27/2010 06:13:30

---

Diagnostics Check	OK	06:06
Air Blank	0.000	06:06
Control Test	0.078	06:06
Air Blank	0.000	06:07
Subject Sample #1	0.121	06:08
Breath Volume	1.402	06:08
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample #2	0.121	06:11
Breath Volume	2.539	06:11
Air Blank	0.000	06:12
Control Test	0.077	06:12
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETO-VO, L, O  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: ANDERSON, D  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 07:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/27/2010 08:48:23

---

Diagnostics Check	OK	08:45	
Air Blank	0.000	08:46	
Control Test	0.077	08:46	
Air Blank	0.000	08:46	
Subject Sample #1		08:47	Subject Test Refused
Breath Volume	0.000	08:47	
Air Blank	0.000	08:47	
Control Test	0.077	08:47	
Air Blank	0.000	08:48	
Diagnostics Check	OK	08:48	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ, ROGER, E  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: ROBINSON, IAN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: BINKER

DATE/TIME OF TEST: 03/27/2010 19:15:02

---

Diagnostics Check	OK	19:07
Air Blank	0.000	19:08
Control Test	0.078	19:08
Air Blank	0.000	19:08
Subject Sample #1	0.182	19:10
Breath Volume	2.464	19:10
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample #2	0.175	19:13
Breath Volume	2.621	19:13
Air Blank	0.000	19:14
Control Test	0.077	19:14
Air Blank	0.000	19:14
Diagnostics Check	OK	19:15

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ, ROGER, E  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: ROBINSON, IAN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: STUART

DATE/TIME OF TEST: 03/27/2010 22:59:34

---

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.078	22:54
Air Blank	0.000	22:54
Subject Sample #1	0.272	22:55
Breath Volume	1.402	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample #2	0.263	22:57
Breath Volume	1.289	22:57
Air Blank	0.000	22:58
Control Test	0.078	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONFESSORE, STEVE,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: BLACIO, GUSTAVO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: SPITALIERI

DATE/TIME OF TEST: 03/28/2010 04:24:36

---

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.078	04:22	
Air Blank	0.000	04:22	
Subject Sample #1		04:23	Subject Test Refused
Breath Volume	0.000	04:23	
Air Blank	0.000	04:23	
Control Test	0.078	04:23	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

---

Data Download Date/Time 03/30/2010 14:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUMPHRY, KERRY, R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SILVAGNI, R  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 659357  
DRY GAS EXPIRY DATE: 09/03/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: TOSTA

DATE/TIME OF TEST: 03/31/2010 05:22:08

---

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:19	
Control Test	0.078	05:20	
Air Blank	0.000	05:20	
Subject Sample #1		05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:21	
Control Test	0.078	05:21	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

---

Data Download Date/Time 04/28/2010 14:43

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: METZLER, SHANE, A  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: METZLER, SHANE  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814801G  
DRY GAS EXPIRY DATE: 05/30/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: SQUIRES

DATE/TIME OF TEST: 03/18/2010 03:29:24

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample #1	0.000	03:24
Breath Volume	1.816	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample #2	0.000	03:28
Breath Volume	2.308	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time 04/13/2010 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS JR, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HOLLIS, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814801G  
DRY GAS EXPIRY DATE: 05/30/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 03/19/2010 01:38:30

---

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.078	01:33
Air Blank	0.000	01:33
Subject Sample #1	0.089	01:34
Breath Volume	2.082	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample #2	0.085	01:37
Breath Volume	1.453	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

---

Data Download Date/Time 04/13/2010 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENTON, CHRISTOPHER, S  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: MEYER, D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814801G  
DRY GAS EXPIRY DATE: 05/30/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HAMM

DATE/TIME OF TEST: 03/20/2010 03:37:32

---

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample #1	0.120	03:32
Breath Volume	1.308	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.120	03:36
Breath Volume	1.781	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

---

Data Download Date/Time 04/13/2010 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS, LANE, A  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: WINTERS, LANE  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814801G  
DRY GAS EXPIRY DATE: 05/30/2010  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: LITTLEFIELD

DATE/TIME OF TEST: 03/22/2010 01:45:32

---

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample #1	0.143	01:40
Breath Volume	1.714	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample #2	0.140	01:44
Breath Volume	1.382	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time 04/13/2010 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENTON, CHRISTOPHER, S  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: DENTON, CHRISTOPHER  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814801G  
DRY GAS EXPIRY DATE: 05/30/2010  
VIOLATION CODE: DUI  
ARREST TIME: 19:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 03/24/2010 21:09:43

---

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.078	21:03
Air Blank	0.000	21:04
Subject Sample #1	0.095	21:05
Breath Volume	2.187	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample #2	0.095	21:08
Breath Volume	1.562	21:08
Air Blank	0.000	21:08
Control Test	0.078	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

---

Data Download Date/Time 04/13/2010 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALLO, NICOLE, L  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: MALLO, NICOLE  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814801G  
DRY GAS EXPIRY DATE: 05/30/2010  
VIOLATION CODE: DUI  
ARREST TIME: 11:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: ROBB

DATE/TIME OF TEST: 03/27/2010 16:32:34

---

Diagnostics Check	OK	16:26
Air Blank	0.000	16:26
Control Test	0.079	16:27
Air Blank	0.000	16:27
Subject Sample #1	0.000	16:28
Breath Volume	2.691	16:28
Air Blank	0.000	16:28
Air Blank	0.000	16:30
Subject Sample #2	0.000	16:31
Breath Volume	2.488	16:31
Air Blank	0.000	16:31
Control Test	0.079	16:31
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

---

Data Download Date/Time 04/13/2010 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS, LANE, A  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: RATCLIFF, KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814801G  
DRY GAS EXPIRY DATE: 05/30/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 03/27/2010 23:55:57

---

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.078	23:45	
Air Blank	0.000	23:45	
Subject Sample #1	0.176	23:48	Volume Not Met
Breath Volume	0.246	23:48	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample #2	0.166	23:54	Volume Not Met
Breath Volume	1.011	23:54	
Air Blank	0.000	23:55	
Control Test	0.077	23:55	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

---

Data Download Date/Time 04/13/2010 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS, LANE, A  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: RATCLIFF, KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814801G  
DRY GAS EXPIRY DATE: 05/30/2010  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 03/28/2010 00:10:27

---

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:58	
Control Test	0.077	23:58	
Air Blank	0.000	23:59	
Subject Sample #1	0.186	00:02	Volume Not Met
Breath Volume	0.066	00:02	
Air Blank	0.000	00:02	
Air Blank	0.000	00:04	
Subject Sample #2	0.183	00:05	
Breath Volume	1.265	00:05	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample #3	0.184	00:08	
Breath Volume	1.941	00:08	
Air Blank	0.000	00:09	
Control Test	0.078	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

---

Data Download Date/Time 04/13/2010 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENTON, CHRISTOPHER, S  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: DENTON, CHRISTOPHER  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814801G  
DRY GAS EXPIRY DATE: 05/30/2010  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 03/28/2010 01:19:49

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.077	01:14
Air Blank	0.000	01:14
Subject Sample #1	0.141	01:15
Breath Volume	2.304	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample #2	0.132	01:18
Breath Volume	1.515	01:18
Air Blank	0.000	01:18
Control Test	0.077	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

---

Data Download Date/Time 04/13/2010 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SZELIGA, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: SMITH, GENE  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814801G  
DRY GAS EXPIRY DATE: 05/30/2010  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HETZEL

DATE/TIME OF TEST: 03/30/2010 02:24:07

---

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.080	02:21	
Air Blank	0.000	02:22	
Subject Sample #1		02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.079	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

---

Data Download Date/Time 04/13/2010 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MATAS, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: OSORI BENTACURT

DATE/TIME OF TEST: 03/07/2010 03:35:35

---

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:29
Subject Sample #1	0.185	03:30
Breath Volume	1.730	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample #2	0.192	03:34
Breath Volume	1.511	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

---

Data Download Date/Time 03/10/2010 13:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: POLK, W  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 02/24/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: KRAM

DATE/TIME OF TEST: 03/07/2010 03:46:22

---

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.081	03:44	
Air Blank	0.000	03:44	
Subject Sample #1		03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Control Test	0.081	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

---

Data Download Date/Time 03/10/2010 13:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRUZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: FAUCHER

DATE/TIME OF TEST: 03/10/2010 22:08:10

---

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:56	
Control Test	0.080	21:56	
Air Blank	0.000	21:57	
Subject Sample #1	0.211	21:57	
Breath Volume	1.414	21:57	
Air Blank	0.000	21:58	
Air Blank	0.000	22:00	
Subject Sample #2	0.203	22:03	Slope Not Level
Breath Volume	1.546	22:03	
Air Blank	0.000	22:04	
Air Blank	0.000	22:05	
Subject Sample #3	0.208	22:06	
Breath Volume	1.496	22:06	
Air Blank	0.000	22:07	
Control Test	0.078	22:07	
Air Blank	0.000	22:08	
Diagnostics Check	OK	22:08	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORET, F  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/11/2010 02:12:48

---

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Subject Sample #1	0.098	02:08
Breath Volume	1.113	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample #2	0.092	02:11
Breath Volume	1.191	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBSON, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: DO

DATE/TIME OF TEST: 03/11/2010 23:01:29

---

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.079	22:50	
Air Blank	0.000	22:50	
Subject Sample #1	0.106	22:54	Volume Not Met
Breath Volume	1.093	22:54	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample #2	0.121	22:57	
Breath Volume	1.296	22:57	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample #3	0.124	23:00	
Breath Volume	1.207	23:00	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: PEREZ-JAIMEZ

DATE/TIME OF TEST: 03/11/2010 23:50:05

---

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Subject Sample #1	0.140	23:45
Breath Volume	1.417	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample #2	0.144	23:48
Breath Volume	1.828	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:50

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRUZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 03/12/2010 00:47:26

---

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:42
Subject Sample #1	0.059	00:43
Breath Volume	2.125	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample #2	0.060	00:46
Breath Volume	2.152	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/12/2010 01:47:54

---

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample #1	0.210	01:43
Breath Volume	1.777	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample #2	0.219	01:46
Breath Volume	1.484	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: GOLDBERG

DATE/TIME OF TEST: 03/12/2010 04:50:55

---

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:43
Subject Sample #1	0.183	04:46
Breath Volume	1.953	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample #2	0.189	04:49
Breath Volume	1.800	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: JERRY

DATE/TIME OF TEST: 03/12/2010 23:58:48

---

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample #1	0.211	23:54
Breath Volume	3.011	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:56
Subject Sample #2	0.196	23:57
Breath Volume	2.734	23:57
Air Blank	0.000	23:57
Control Test	0.079	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MATAS, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: PEREZ-LOPEZ

DATE/TIME OF TEST: 03/13/2010 00:50:14

---

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Subject Sample #1	0.261	00:45
Breath Volume	2.152	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample #2	0.267	00:48
Breath Volume	2.406	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARBOSA, F  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: PARAVINJA

DATE/TIME OF TEST: 03/13/2010 02:34:21

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:27
Air Blank	0.000	02:27
Subject Sample #1	0.182	02:29
Breath Volume	1.671	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample #2	0.185	02:32
Breath Volume	1.945	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILSON, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 03/13/2010 02:46:45

---

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:44	
Subject Sample #1		02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.079	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

---

Data Download Date/Time 03/31/2010 09:19

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIVERA, E  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/13/2010 03:09:21

---

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample #1	0.212	03:04
Breath Volume	1.761	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample #2	0.214	03:07
Breath Volume	1.515	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: REIDY

DATE/TIME OF TEST: 03/13/2010 03:29:12

---

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample #1	0.127	03:24
Breath Volume	2.460	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample #2	0.128	03:27
Breath Volume	1.824	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: RAKE

DATE/TIME OF TEST: 03/13/2010 03:37:58

---

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample #1	0.127	03:33
Breath Volume	1.507	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample #2	0.127	03:36
Breath Volume	1.195	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: DURSO

DATE/TIME OF TEST: 03/13/2010 05:37:12

---

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:32
Subject Sample #1	0.131	05:32
Breath Volume	2.609	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample #2	0.132	05:35
Breath Volume	1.394	05:35
Air Blank	0.000	05:36
Control Test	0.081	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SANTIAGO, M  
ARRESTING OFFICER AGENCY: U of So FL Campus PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 03/13/2010 06:09:04

---

Diagnostics Check	OK	06:02
Air Blank	0.000	06:03
Control Test	0.079	06:03
Air Blank	0.000	06:03
Subject Sample #1	0.131	06:04
Breath Volume	2.136	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample #2	0.136	06:07
Breath Volume	1.652	06:07
Air Blank	0.000	06:08
Control Test	0.080	06:08
Air Blank	0.000	06:08
Diagnostics Check	OK	06:09

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KACZMAREK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/14/2010 00:53:40

---

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Subject Sample #1	0.291	00:48
Breath Volume	1.832	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample #2	0.287	00:52
Breath Volume	1.308	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 03/14/2010 04:17:56

---

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.080	04:13	
Air Blank	0.000	04:13	
Subject Sample #1		04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.081	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILSON, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: HOLLINGSWORTH

DATE/TIME OF TEST: 03/14/2010 04:29:06

---

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.080	04:21	
Air Blank	0.000	04:22	
Subject Sample #1	0.146	04:25	Volume Not Met
Breath Volume	0.867	04:25	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample #2		04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:28	
Control Test	0.079	04:28	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILSON, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: HOLLINGSWORTH

DATE/TIME OF TEST: 03/14/2010 04:38:55

---

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:33	
Subject Sample #1	0.145	04:33	
Breath Volume	1.625	04:33	
Air Blank	0.000	04:34	
Air Blank	0.000	04:36	
Subject Sample #2		04:37	Subject Test Refused
Breath Volume	0.265	04:37	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVERAS, N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: HEARING

DATE/TIME OF TEST: 03/14/2010 04:45:49

---

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:43	
Control Test	0.080	04:43	
Air Blank	0.000	04:43	
Subject Sample #1		04:44	Subject Test Refused
Breath Volume	0.000	04:44	
Air Blank	0.000	04:44	
Control Test	0.080	04:45	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVERAS, N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: KU CHABLE

DATE/TIME OF TEST: 03/14/2010 21:30:50

---

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:26	
Control Test	0.080	21:26	
Air Blank	0.000	21:26	
Subject Sample #1		21:29	Subject Test Refused
Breath Volume	0.785	21:29	
Air Blank	0.000	21:29	
Control Test	0.080	21:30	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVERAS, N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: KU CHABLE

DATE/TIME OF TEST: 03/14/2010 21:38:46

---

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.078	21:32
Air Blank	0.000	21:32
Subject Sample #1	0.233	21:33
Breath Volume	1.707	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample #2	0.230	21:37
Breath Volume	1.503	21:37
Air Blank	0.000	21:37
Control Test	0.079	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WHITEHEAD, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 03/14/2010 22:08:43

---

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.079	22:03
Air Blank	0.000	22:03
Subject Sample #1	0.127	22:04
Breath Volume	1.460	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample #2	0.130	22:07
Breath Volume	2.156	22:07
Air Blank	0.000	22:07
Control Test	0.079	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MATAS, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: HELMER

DATE/TIME OF TEST: 03/14/2010 22:25:30

---

Diagnostics Check	OK	22:19
Air Blank	0.000	22:19
Control Test	0.079	22:20
Air Blank	0.000	22:20
Subject Sample #1	0.182	22:21
Breath Volume	1.601	22:21
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample #2	0.176	22:24
Breath Volume	1.722	22:24
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: VILLEGAS

DATE/TIME OF TEST: 03/15/2010 00:37:21

---

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Subject Sample #1	0.172	00:33
Breath Volume	2.496	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample #2	0.171	00:35
Breath Volume	2.316	00:35
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILLAN-CINTRON, E  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/15/2010 01:17:37

---

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample #1	0.134	01:13
Breath Volume	1.207	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample #2	0.137	01:16
Breath Volume	1.761	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KACZMAREK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: CERAK

DATE/TIME OF TEST: 03/15/2010 01:33:44

---

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample #1	0.229	01:28
Breath Volume	2.527	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample #2	0.233	01:32
Breath Volume	2.285	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/15/2010 03:31:30

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Subject Sample #1	0.076	03:27
Breath Volume	3.132	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample #2	0.075	03:30
Breath Volume	3.050	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 03/15/2010 04:18:30

---

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.080	04:04	
Air Blank	0.000	04:05	
Subject Sample #1		04:08	No Sample Provided
Breath Volume	0.000	04:08	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample #2	0.188	04:12	
Breath Volume	1.195	04:12	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample #3	0.185	04:16	
Breath Volume	1.156	04:16	
Air Blank	0.000	04:17	
Control Test	0.079	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MATAS, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GOODRICH

DATE/TIME OF TEST: 03/15/2010 04:23:58

---

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.079	04:20	
Air Blank	0.000	04:21	
Subject Sample #1		04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:23	
Control Test	0.078	04:23	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVERAS, N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: PAULA-NAVEO

DATE/TIME OF TEST: 03/15/2010 05:16:51

---

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Subject Sample #1	0.154	05:12
Breath Volume	1.730	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample #2	0.151	05:15
Breath Volume	1.515	05:15
Air Blank	0.000	05:15
Control Test	0.079	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORREY, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: ARAOZ

DATE/TIME OF TEST: 03/15/2010 05:29:30

---

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Subject Sample #1	0.037	05:25
Breath Volume	3.234	05:25
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample #2	0.036	05:28
Breath Volume	2.957	05:28
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/15/2010 22:23:43

---

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:18
Air Blank	0.000	22:18
Subject Sample #1	0.145	22:19
Breath Volume	1.402	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:21
Subject Sample #2	0.142	22:22
Breath Volume	1.531	22:22
Air Blank	0.000	22:22
Control Test	0.080	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRUZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: COFIELD

DATE/TIME OF TEST: 03/15/2010 23:48:30

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample #1	0.087	23:44
Breath Volume	2.679	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample #2	0.087	23:47
Breath Volume	2.437	23:47
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KENNEDY, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HINES'

DATE/TIME OF TEST: 03/16/2010 00:35:45

---

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Subject Sample #1	0.000	00:30
Breath Volume	1.230	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample #2	0.000	00:34
Breath Volume	1.144	00:34
Air Blank	0.000	00:34
Control Test	0.080	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: STEAD, K  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CASTELLANO-ROMERO

DATE/TIME OF TEST: 03/16/2010 01:35:32

---

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.080	01:31	
Air Blank	0.000	01:32	
Subject Sample #1		01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MELO, J. R.  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 03/16/2010 02:46:34

---

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.081	02:34	
Air Blank	0.000	02:34	
Subject Sample #1	0.219	02:38	Slope Not Level
Breath Volume	1.601	02:38	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample #2	0.221	02:41	
Breath Volume	2.300	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample #3	0.226	02:45	
Breath Volume	2.324	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHANAN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: METTS

DATE/TIME OF TEST: 03/16/2010 04:40:52

---

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample #1	0.088	04:36
Breath Volume	1.953	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample #2	0.089	04:39
Breath Volume	1.164	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: STEIMK, T  
ARRESTING OFFICER AGENCY: Macdill Air Force Ba  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 03/16/2010 20:44:05

---

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.081	20:38
Air Blank	0.000	20:39
Subject Sample #1	0.321	20:39
Breath Volume	2.441	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample #2	0.305	20:42
Breath Volume	1.628	20:42
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKEE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/17/2010 00:12:36

---

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Subject Sample #1	0.091	00:07
Breath Volume	1.218	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample #2	0.085	00:11
Breath Volume	1.410	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILSON, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: KLINIESKI

DATE/TIME OF TEST: 03/17/2010 01:58:36

---

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Subject Sample #1	0.179	01:53
Breath Volume	1.300	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample #2	0.185	01:57
Breath Volume	1.156	01:57
Air Blank	0.000	01:57
Control Test	0.081	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/17/2010 04:11:53

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample #1	0.157	04:07
Breath Volume	2.195	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample #2	0.146	04:10
Breath Volume	1.386	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAVIS, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 03/17/2010 23:08:08

---

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample #1	0.103	23:03
Breath Volume	1.941	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample #2	0.105	23:06
Breath Volume	1.777	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURCHETT, J  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 03/17/2010 23:32:57

---

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample #1	0.126	23:28
Breath Volume	1.968	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample #2	0.120	23:31
Breath Volume	2.562	23:31
Air Blank	0.000	23:32
Control Test	0.078	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORREY, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: DAMRON

DATE/TIME OF TEST: 03/17/2010 23:55:34

---

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:50
Subject Sample #1	0.201	23:51
Breath Volume	2.738	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample #2	0.208	23:54
Breath Volume	2.445	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MAGUIRE III

DATE/TIME OF TEST: 03/18/2010 00:23:27

---

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.080	00:18
Air Blank	0.000	00:18
Subject Sample #1	0.156	00:19
Breath Volume	3.828	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample #2	0.157	00:21
Breath Volume	3.453	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCCALLA, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SUKUP

DATE/TIME OF TEST: 03/18/2010 00:39:25

---

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Subject Sample #1	0.107	00:35
Breath Volume	3.171	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample #2	0.098	00:37
Breath Volume	3.246	00:37
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMS, G  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: GAGE

DATE/TIME OF TEST: 03/18/2010 01:44:07

---

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.081	01:41	
Air Blank	0.000	01:42	
Subject Sample #1		01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CARSON, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/18/2010 01:56:40

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Subject Sample #1	0.205	01:51
Breath Volume	1.953	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample #2	0.210	01:55
Breath Volume	1.914	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VALLEJO, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: VAZUEZ

DATE/TIME OF TEST: 03/18/2010 02:07:48

---

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample #1	0.144	02:03
Breath Volume	2.160	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample #2	0.141	02:06
Breath Volume	2.414	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: COOPER, W  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MENENDEZ

DATE/TIME OF TEST: 03/18/2010 02:59:31

---

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Subject Sample #1	0.154	02:54
Breath Volume	1.167	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.164	02:58
Breath Volume	1.292	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RADU, S  
ARRESTING OFFICER AGENCY: Tampa Airport PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 03/18/2010 03:08:33

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Subject Sample #1	0.181	03:03
Breath Volume	3.238	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample #2	0.181	03:07
Breath Volume	2.332	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: QUICK

DATE/TIME OF TEST: 03/18/2010 04:01:39

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample #1	0.155	03:57
Breath Volume	1.429	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample #2	0.163	04:00
Breath Volume	1.644	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RAMER, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 03/18/2010 04:13:19

---

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:08
Air Blank	0.000	04:08
Subject Sample #1	0.000	04:09
Breath Volume	2.785	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample #2	0.000	04:11
Breath Volume	2.976	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORREY, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 03/18/2010 20:16:39

---

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.080	20:09
Air Blank	0.000	20:09
Subject Sample #1	0.168	20:11
Breath Volume	1.214	20:11
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample #2	0.167	20:15
Breath Volume	1.296	20:15
Air Blank	0.000	20:15
Control Test	0.079	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HANNON, B  
ARRESTING OFFICER AGENCY: Macdill Air Force Ba  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: ARCHIELD

DATE/TIME OF TEST: 03/18/2010 21:39:15

---

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:34
Subject Sample #1	0.095	21:34
Breath Volume	2.675	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample #2	0.097	21:37
Breath Volume	1.992	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KACZMARECK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 03/18/2010 22:33:51

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Subject Sample #1	0.014	22:29
Breath Volume	1.585	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample #2	0.013	22:32
Breath Volume	1.824	22:32
Air Blank	0.000	22:32
Control Test	0.081	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: ORAMA

DATE/TIME OF TEST: 03/19/2010 02:06:31

---

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:01
Subject Sample #1	0.244	02:01
Breath Volume	2.207	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample #2	0.258	02:05
Breath Volume	1.992	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CARSON, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 03/19/2010 03:02:28

---

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:47	
Subject Sample #1	0.115	02:49	
Breath Volume	1.628	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample #2	0.120	02:55	Volume Not Met
Breath Volume	0.261	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:57	
Subject Sample #3	0.129	03:00	
Breath Volume	1.464	03:00	
Air Blank	0.000	03:01	
Control Test	0.079	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: COOPER, W  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 03/19/2010 03:28:03

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Subject Sample #1	0.132	03:23
Breath Volume	1.765	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample #2	0.134	03:26
Breath Volume	1.906	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BORDEN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/19/2010 03:39:16

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Subject Sample #1	0.033	03:35
Breath Volume	1.710	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample #2	0.031	03:37
Breath Volume	1.789	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VALLEJO, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MATEO

DATE/TIME OF TEST: 03/19/2010 04:06:16

---

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample #1	0.194	04:01
Breath Volume	1.964	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample #2	0.193	04:04
Breath Volume	1.675	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARBOSA, F  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 03/19/2010 04:19:00

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample #1	0.150	04:14
Breath Volume	3.515	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.135	04:17
Breath Volume	1.542	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:19

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FIGUEROA, D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: SMITH JR

DATE/TIME OF TEST: 03/19/2010 19:55:24

---

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.081	19:49
Air Blank	0.000	19:49
Subject Sample #1	0.109	19:51
Breath Volume	1.136	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample #2	0.115	19:53
Breath Volume	1.484	19:53
Air Blank	0.000	19:54
Control Test	0.081	19:54
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: CUESTA

DATE/TIME OF TEST: 03/20/2010 00:18:46

---

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.080	00:07	
Air Blank	0.000	00:08	
Subject Sample #1	0.214	00:11	Volume Not Met
Breath Volume	1.070	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample #2	0.233	00:17	Volume Not Met
Breath Volume	0.089	00:17	
Air Blank	0.000	00:17	
Control Test	0.079	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: CUESTA

DATE/TIME OF TEST: 03/20/2010 00:29:19

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Subject Sample #1	0.224	00:23
Breath Volume	1.347	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample #2	0.209	00:27
Breath Volume	1.648	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARBOSA, F  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: MONTES-PANCOATL

DATE/TIME OF TEST: 03/20/2010 01:20:59

---

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample #1	0.232	01:16
Breath Volume	2.746	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample #2	0.228	01:19
Breath Volume	2.808	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 03/20/2010 01:35:23

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Subject Sample #1	0.085	01:30
Breath Volume	5.386	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample #2	0.083	01:33
Breath Volume	4.050	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MATAS, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: LANEY

DATE/TIME OF TEST: 03/20/2010 01:48:04

---

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample #1	0.132	01:43
Breath Volume	2.640	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample #2	0.136	01:46
Breath Volume	2.421	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:48

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: INSIGNARES, P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 03/20/2010 02:27:34

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample #1	0.000	02:21
Breath Volume	1.320	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample #2	0.000	02:26
Breath Volume	1.324	02:26
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KENNEDY, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 03/20/2010 03:10:17

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Subject Sample #1	0.135	03:05
Breath Volume	3.175	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample #2	0.134	03:08
Breath Volume	3.257	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARBOSA, F  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GORSKIE

DATE/TIME OF TEST: 03/20/2010 03:57:46

---

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:52
Air Blank	0.000	03:52
Subject Sample #1	0.178	03:53
Breath Volume	3.468	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample #2	0.179	03:56
Breath Volume	2.750	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: UBILLA

DATE/TIME OF TEST: 03/20/2010 04:22:02

---

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.080	04:08	
Air Blank	0.000	04:08	
Subject Sample #1	0.158	04:12	Volume Not Met
Breath Volume	0.324	04:12	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample #2	0.159	04:17	
Breath Volume	1.257	04:17	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample #3	0.164	04:20	
Breath Volume	1.152	04:20	
Air Blank	0.000	04:21	
Control Test	0.079	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:22	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RICH, D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: TOBON

DATE/TIME OF TEST: 03/20/2010 05:25:42

---

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:22	
Control Test	0.080	05:23	
Air Blank	0.000	05:23	
Subject Sample #1	0.220	05:24	
Breath Volume	1.453	05:24	
Air Blank		05:24	RFI Detect
Air Blank	0.000	05:24	
Control Test	0.080	05:25	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RICH, D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: TOBON

DATE/TIME OF TEST: 03/20/2010 05:38:54

---

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:27	
Control Test	0.080	05:27	
Air Blank	0.000	05:28	
Subject Sample #1	0.244	05:28	No .020 agreement
Breath Volume	2.691	05:28	
Air Blank	0.000	05:29	
Air Blank	0.000	05:31	
Subject Sample #2	0.221	05:31	No .020 agreement
Breath Volume	1.339	05:31	
Air Blank	0.000	05:32	
Air Blank	0.000	05:34	
Subject Sample #3	0.210	05:37	Volume Not Met
Breath Volume	0.390	05:37	
Air Blank	0.000	05:37	
Control Test	0.081	05:38	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RICH, D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: TOBON

DATE/TIME OF TEST: 03/20/2010 05:51:21

---

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.079	05:40
Air Blank	0.000	05:41
Subject Sample #1	0.217	05:42
Breath Volume	2.324	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample #2	0.195	05:45
Breath Volume	1.308	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample #3	0.205	05:49
Breath Volume	1.222	05:49
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: UNO, D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: LAPRADD

DATE/TIME OF TEST: 03/20/2010 19:17:49

---

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.081	19:09
Air Blank	0.000	19:09
Subject Sample #1	0.000	19:10
Breath Volume	1.539	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample #2	0.000	19:16
Breath Volume	1.402	19:16
Air Blank	0.000	19:16
Control Test	0.081	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: INSIGNARES, P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 03/20/2010 21:39:18

---

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:34
Subject Sample #1	0.209	21:34
Breath Volume	3.046	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample #2	0.216	21:37
Breath Volume	2.222	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/20/2010 22:14:35

---

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.080	22:09	
Air Blank	0.000	22:09	
Subject Sample #1	0.191	22:09	
Breath Volume	1.945	22:09	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample #2		22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:13	
Control Test	0.080	22:14	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:14	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/20/2010 22:39:15

---

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.081	22:31
Air Blank	0.000	22:31
Subject Sample #1	0.166	22:32
Breath Volume	1.347	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample #2	0.187	22:34
Breath Volume	1.730	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample #3	0.180	22:37
Breath Volume	1.257	22:37
Air Blank	0.000	22:38
Control Test	0.079	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKEE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/21/2010 00:24:48

---

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:19
Air Blank	0.000	00:19
Subject Sample #1	0.167	00:20
Breath Volume	3.371	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample #2	0.163	00:23
Breath Volume	3.296	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILSON, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MC CORT

DATE/TIME OF TEST: 03/21/2010 01:37:16

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Subject Sample #1	0.161	01:33
Breath Volume	1.781	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample #2	0.165	01:35
Breath Volume	1.843	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRUZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: MITAS

DATE/TIME OF TEST: 03/21/2010 02:00:05

---

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.079	01:57	
Air Blank	0.000	01:58	
Subject Sample #1		01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.080	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVERAS, N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/21/2010 02:10:05

---

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:07	
Subject Sample #1		02:08	Subject Test Refused
Breath Volume	0.269	02:08	
Air Blank	0.000	02:09	
Control Test	0.079	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVERAS, N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/21/2010 02:14:27

---

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:12	
Subject Sample #1		02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORET, F  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BONCHI

DATE/TIME OF TEST: 03/21/2010 02:27:08

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample #1	0.000	02:22
Breath Volume	3.253	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample #2	0.000	02:25
Breath Volume	4.371	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 03/21/2010 03:26:49

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Subject Sample #1	0.177	03:22
Breath Volume	2.238	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample #2	0.175	03:25
Breath Volume	3.281	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: INSIGNARES, P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 03/21/2010 03:40:08

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:33
Air Blank	0.000	03:33
Subject Sample #1	0.110	03:35
Breath Volume	1.652	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample #2	0.105	03:38
Breath Volume	1.937	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DURAN, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: CARTAGENA-MURILLO

DATE/TIME OF TEST: 03/21/2010 04:26:51

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Subject Sample #1	0.117	04:22
Breath Volume	2.261	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.108	04:25
Breath Volume	1.324	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KENNEDY, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: COVINGTON

DATE/TIME OF TEST: 03/21/2010 05:26:59

---

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Subject Sample #1	0.172	05:22
Breath Volume	4.136	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample #2	0.176	05:25
Breath Volume	3.699	05:25
Air Blank	0.000	05:26
Control Test	0.079	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SERRANO, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: MONTANO

DATE/TIME OF TEST: 03/21/2010 06:28:22

---

Diagnostics Check	OK	06:21
Air Blank	0.000	06:21
Control Test	0.080	06:21
Air Blank	0.000	06:22
Subject Sample #1	0.144	06:22
Breath Volume	2.539	06:22
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample #2	0.141	06:26
Breath Volume	1.867	06:26
Air Blank	0.000	06:27
Control Test	0.080	06:27
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKEE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: NOBLES

DATE/TIME OF TEST: 03/21/2010 20:34:01

---

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.079	20:28
Air Blank	0.000	20:28
Subject Sample #1	0.153	20:29
Breath Volume	1.820	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample #2	0.152	20:32
Breath Volume	2.082	20:32
Air Blank	0.000	20:33
Control Test	0.079	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:34

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KRUTHERS, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 03/21/2010 22:33:51

---

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:27
Subject Sample #1	0.226	22:28
Breath Volume	1.390	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample #2	0.229	22:32
Breath Volume	1.781	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KENNEDY, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: FENDERSON

DATE/TIME OF TEST: 03/22/2010 00:44:48

---

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Subject Sample #1	0.149	00:40
Breath Volume	1.367	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample #2	0.164	00:43
Breath Volume	1.675	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: WIMER

DATE/TIME OF TEST: 03/22/2010 01:15:33

---

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Subject Sample #1	0.191	01:11
Breath Volume	1.636	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample #2	0.197	01:14
Breath Volume	1.699	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CARMICIANO

DATE/TIME OF TEST: 03/22/2010 02:15:48

---

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Subject Sample #1		02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:14	
Control Test	0.079	02:15	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 03/22/2010 02:40:33

---

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:35
Air Blank	0.000	02:35
Subject Sample #1	0.206	02:36
Breath Volume	2.093	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample #2	0.202	02:39
Breath Volume	1.550	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILSON, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/22/2010 04:45:50

---

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.080	04:41	
Air Blank	0.000	04:42	
Subject Sample #1		04:44	Subject Test Refused
Breath Volume	0.039	04:44	
Air Blank	0.000	04:44	
Control Test	0.080	04:45	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MATAS, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 03/23/2010 00:00:37

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample #1	0.190	23:56
Breath Volume	1.636	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample #2	0.180	23:59
Breath Volume	1.687	23:59
Air Blank	0.000	23:59
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUSWELL, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 03/23/2010 01:26:23

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample #1	0.183	01:22
Breath Volume	2.093	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample #2	0.184	01:24
Breath Volume	1.718	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUSWELL, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 03/23/2010 05:50:32

---

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.080	05:45
Air Blank	0.000	05:45
Subject Sample #1	0.000	05:46
Breath Volume	2.460	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample #2	0.000	05:49
Breath Volume	2.746	05:49
Air Blank	0.000	05:49
Control Test	0.080	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RAMER, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: PEGG

DATE/TIME OF TEST: 03/23/2010 23:24:42

---

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.080	23:19
Air Blank	0.000	23:19
Subject Sample #1	0.174	23:20
Breath Volume	1.242	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample #2	0.175	23:23
Breath Volume	1.386	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MEDERO-OSORIO

DATE/TIME OF TEST: 03/24/2010 01:22:32

---

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:17
Air Blank	0.000	01:17
Subject Sample #1	0.146	01:18
Breath Volume	1.488	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample #2	0.150	01:21
Breath Volume	1.136	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARR, TALIA, R

BREATH TEST OPERATOR AGENCY: Hillsborough Co SO

ARRESTING OFFICER: STEAD, K

ARRESTING OFFICER AGENCY: Tampa PD

LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 07/30/2011

VIOLATION CODE: DUI

ARREST TIME: 01:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: PITTMAN JR.

DATE/TIME OF TEST: 03/24/2010 02:14:10

---

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample #1	0.000	02:10
Breath Volume	2.917	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample #2	0.000	02:12
Breath Volume	2.878	02:12
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WHITEHEAD, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/24/2010 03:38:41

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample #1	0.079	03:34
Breath Volume	3.296	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample #2	0.078	03:37
Breath Volume	2.847	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURCHETT, J  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/24/2010 05:50:57

---

Diagnostics Check	OK	05:44
Air Blank	0.000	05:45
Control Test	0.081	05:45
Air Blank	0.000	05:46
Subject Sample #1	0.086	05:46
Breath Volume	1.289	05:46
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample #2	0.083	05:49
Breath Volume	1.621	05:49
Air Blank	0.000	05:50
Control Test	0.080	05:50
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: MCCURDY

DATE/TIME OF TEST: 03/24/2010 23:57:35

---

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample #1	0.079	23:53
Breath Volume	2.945	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample #2	0.080	23:56
Breath Volume	3.355	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: STEAD, K  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: SANTOS-SANTIAGO

DATE/TIME OF TEST: 03/25/2010 02:11:01

---

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Subject Sample #1		02:03	No Sample Provided
Breath Volume	0.000	02:03	
Air Blank	0.000	02:04	
Air Blank	0.000	02:06	
Subject Sample #2	0.096	02:09	Volume Not Met
Breath Volume	0.070	02:09	
Air Blank	0.000	02:10	
Control Test	0.080	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:11	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRUZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: GRANHOLM

DATE/TIME OF TEST: 03/25/2010 03:23:05

---

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample #1	0.265	03:17
Breath Volume	2.332	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample #2	0.283	03:21
Breath Volume	2.355	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARBOSA, F  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: DONHAM

DATE/TIME OF TEST: 03/25/2010 03:47:04

---

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.080	03:36	
Air Blank	0.000	03:36	
Subject Sample #1	0.101	03:39	Volume Not Met
Breath Volume	0.714	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample #2	0.128	03:42	
Breath Volume	1.496	03:42	
Air Blank	0.000	03:43	
Air Blank	0.000	03:45	
Subject Sample #3	0.119	03:45	
Breath Volume	1.640	03:45	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:47	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: HINSHAW

DATE/TIME OF TEST: 03/25/2010 03:59:59

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample #1	0.190	03:55
Breath Volume	1.683	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample #2	0.190	03:58
Breath Volume	1.656	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MIGLIORE, C  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/25/2010 04:14:56

---

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample #1	0.190	04:09
Breath Volume	1.613	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample #2	0.190	04:13
Breath Volume	1.859	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKEE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 03/25/2010 04:34:43

---

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:29
Subject Sample #1	0.151	04:30
Breath Volume	1.273	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample #2	0.144	04:33
Breath Volume	2.132	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/25/2010 04:50:42

---

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:45
Subject Sample #1	0.124	04:46
Breath Volume	2.191	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample #2	0.125	04:49
Breath Volume	2.609	04:49
Air Blank	0.000	04:49
Control Test	0.079	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHANAN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: TORAN

DATE/TIME OF TEST: 03/26/2010 06:26:45

---

Diagnostics Check	OK	06:20
Air Blank	0.000	06:21
Control Test	0.079	06:21
Air Blank	0.000	06:22
Subject Sample #1	0.177	06:22
Breath Volume	2.222	06:22
Air Blank	0.000	06:23
Air Blank	0.000	06:24
Subject Sample #2	0.177	06:25
Breath Volume	1.695	06:25
Air Blank	0.000	06:25
Control Test	0.080	06:26
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: JONES, O  
ARRESTING OFFICER AGENCY: Tampa Airport PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SZARAZ

DATE/TIME OF TEST: 03/26/2010 23:17:26

---

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:06	
Control Test	0.080	23:07	
Air Blank	0.000	23:07	
Subject Sample #1	0.287	23:10	
Breath Volume	1.390	23:10	
Air Blank	0.000	23:10	
Air Blank	0.000	23:12	
Subject Sample #2	0.266	23:13	
Breath Volume	1.304	23:13	
Air Blank	0.000	23:13	
Air Blank	0.000	23:15	
Subject Sample #3		23:15	Slope Not Met
Breath Volume	1.019	23:15	
Air Blank	0.000	23:16	
Control Test	0.075	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCMURPHY, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: ISIDRO-FLORES

DATE/TIME OF TEST: 03/26/2010 23:26:19

---

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:21
Subject Sample #1	0.104	23:21
Breath Volume	1.605	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample #2	0.101	23:24
Breath Volume	1.136	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: JONES, O  
ARRESTING OFFICER AGENCY: Tampa Airport PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SZARAZ

DATE/TIME OF TEST: 03/26/2010 23:43:16

---

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.080	23:36
Air Blank	0.000	23:36
Subject Sample #1	0.272	23:38
Breath Volume	1.863	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample #2	0.277	23:41
Breath Volume	1.546	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILSON, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 03/26/2010 23:54:08

---

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:49
Subject Sample #1	0.149	23:49
Breath Volume	1.691	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:51
Subject Sample #2	0.158	23:52
Breath Volume	2.304	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUSWELL, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: GINTHER

DATE/TIME OF TEST: 03/27/2010 01:06:27

---

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample #1	0.142	01:01
Breath Volume	1.656	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample #2	0.147	01:05
Breath Volume	2.007	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BORDEN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/27/2010 01:45:39

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample #1	0.069	01:41
Breath Volume	1.664	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample #2	0.067	01:44
Breath Volume	1.414	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: STEADE, K  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: AVILES-GOROSTIETA

DATE/TIME OF TEST: 03/27/2010 02:17:49

---

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample #1	0.170	02:13
Breath Volume	1.531	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample #2	0.173	02:16
Breath Volume	1.792	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BODAMER, E  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 03/27/2010 03:17:53

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample #1	0.123	03:13
Breath Volume	2.195	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.130	03:16
Breath Volume	3.050	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORRELL, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: GROMALSKI

DATE/TIME OF TEST: 03/27/2010 03:43:25

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Subject Sample #1	0.069	03:39
Breath Volume	4.335	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample #2	0.068	03:41
Breath Volume	4.003	03:41
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KACZMAREK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MARTINEZ-CARRERA

DATE/TIME OF TEST: 03/27/2010 03:55:06

---

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample #1	0.259	03:50
Breath Volume	1.871	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample #2	0.277	03:53
Breath Volume	1.945	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RICH, D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: DISTEFANO

DATE/TIME OF TEST: 03/27/2010 05:34:21

---

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:29
Subject Sample #1	0.216	05:29
Breath Volume	1.460	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample #2	0.221	05:32
Breath Volume	1.433	05:32
Air Blank	0.000	05:33
Control Test	0.079	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAVIS, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TOCCO

DATE/TIME OF TEST: 03/27/2010 19:56:53

---

Diagnostics Check	OK	19:50
Air Blank	0.000	19:51
Control Test	0.080	19:51
Air Blank	0.000	19:51
Subject Sample #1	0.108	19:52
Breath Volume	2.074	19:52
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample #2	0.107	19:55
Breath Volume	2.128	19:55
Air Blank	0.000	19:55
Control Test	0.080	19:56
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: LAMPART

DATE/TIME OF TEST: 03/28/2010 00:50:10

---

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.080	00:48	
Air Blank	0.000	00:48	
Subject Sample #1		00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUSWELL, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 03/28/2010 01:03:48

---

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample #1	0.164	00:59
Breath Volume	1.218	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample #2	0.171	01:02
Breath Volume	1.273	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/28/2010 02:56:52

---

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Subject Sample #1	0.183	02:52
Breath Volume	2.644	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample #2	0.177	02:55
Breath Volume	2.195	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MCNAMEE

DATE/TIME OF TEST: 03/28/2010 03:16:00

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Subject Sample #1	0.156	03:11
Breath Volume	1.773	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample #2	0.165	03:14
Breath Volume	2.265	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:16

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FOSTER, E  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: WELBURN

DATE/TIME OF TEST: 03/28/2010 03:39:52

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:34
Air Blank	0.000	03:34
Subject Sample #1	0.240	03:35
Breath Volume	1.945	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample #2	0.251	03:38
Breath Volume	2.500	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KACZMAREK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: PEREZ-CALLEJAS

DATE/TIME OF TEST: 03/28/2010 04:05:29

---

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample #1	0.064	04:01
Breath Volume	1.781	04:01
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample #2	0.064	04:04
Breath Volume	1.667	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: PADRON

DATE/TIME OF TEST: 03/28/2010 04:31:50

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample #1	0.137	04:27
Breath Volume	3.808	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample #2	0.139	04:30
Breath Volume	3.738	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: CUPP

DATE/TIME OF TEST: 03/28/2010 04:55:51

---

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:49
Subject Sample #1	0.076	04:51
Breath Volume	1.664	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample #2	0.078	04:54
Breath Volume	1.515	04:54
Air Blank	0.000	04:54
Control Test	0.079	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RAMER, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: BENITO

DATE/TIME OF TEST: 03/28/2010 05:01:06

---

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:58	
Control Test	0.079	04:59	
Air Blank	0.000	04:59	
Subject Sample #1		04:59	Subject Test Refused
Breath Volume	0.000	04:59	
Air Blank	0.000	05:00	
Control Test	0.080	05:00	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:01	

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAVIS, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: POSADA

DATE/TIME OF TEST: 03/28/2010 22:56:25

---

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:51
Air Blank	0.000	22:51
Subject Sample #1	0.000	22:52
Breath Volume	3.304	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample #2	0.000	22:55
Breath Volume	2.542	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DOWNEY

DATE/TIME OF TEST: 03/29/2010 00:41:39

---

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:36
Subject Sample #1	0.129	00:37
Breath Volume	2.792	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample #2	0.125	00:40
Breath Volume	2.527	00:40
Air Blank	0.000	00:40
Control Test	0.079	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRUZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: LEPAK

DATE/TIME OF TEST: 03/29/2010 23:15:11

---

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.081	23:08
Air Blank	0.000	23:09
Subject Sample #1	0.188	23:09
Breath Volume	2.257	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample #2	0.176	23:13
Breath Volume	1.785	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCLAIN, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/29/2010 23:56:09

---

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Subject Sample #1	0.182	23:51
Breath Volume	2.054	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample #2	0.183	23:54
Breath Volume	1.527	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 03/30/2010 01:31:44

---

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample #1	0.000	01:27
Breath Volume	3.222	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample #2	0.000	01:30
Breath Volume	2.792	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARBOSA, F  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SANZ JR.

DATE/TIME OF TEST: 03/30/2010 01:40:32

---

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample #1	0.138	01:36
Breath Volume	1.132	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample #2	0.137	01:39
Breath Volume	1.406	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L.D.  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: JARRETT

DATE/TIME OF TEST: 03/30/2010 01:58:23

---

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Subject Sample #1	0.125	01:53
Breath Volume	1.687	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.125	01:56
Breath Volume	1.859	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: BOSWELL

DATE/TIME OF TEST: 03/30/2010 02:31:46

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample #1	0.103	02:27
Breath Volume	2.582	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample #2	0.103	02:30
Breath Volume	2.234	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SAURET

DATE/TIME OF TEST: 03/30/2010 02:59:18

---

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Subject Sample #1	0.151	02:55
Breath Volume	2.628	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.155	02:57
Breath Volume	2.457	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIVERA, E  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CASTILLO JEREZ

DATE/TIME OF TEST: 03/30/2010 03:37:15

---

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample #1	0.156	03:32
Breath Volume	1.449	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.171	03:35
Breath Volume	1.765	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVERAS, N  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 03/30/2010 03:59:15

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample #1	0.086	03:55
Breath Volume	3.046	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample #2	0.083	03:57
Breath Volume	2.441	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARBOSA, F  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: WOODSIDE

DATE/TIME OF TEST: 03/30/2010 04:47:18

---

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample #1	0.117	04:43
Breath Volume	1.785	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample #2	0.114	04:45
Breath Volume	1.187	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

---

Data Download Date/Time 03/31/2010 09:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FOSTER, E  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: KINAL

DATE/TIME OF TEST: 03/10/2010 19:45:09

---

Diagnostics Check	OK	19:38
Air Blank	0.000	19:39
Control Test	0.081	19:39
Air Blank	0.000	19:39
Subject Sample #1	0.254	19:40
Breath Volume	2.562	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample #2	0.264	19:43
Breath Volume	2.707	19:43
Air Blank	0.000	19:44
Control Test	0.080	19:44
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

---

Data Download Date/Time 03/11/2010 13:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KRUTHERS, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: ISIDRO

DATE/TIME OF TEST: 03/10/2010 23:09:31

---

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.082	23:03
Air Blank	0.000	23:04
Subject Sample #1	0.263	23:05
Breath Volume	1.265	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample #2	0.267	23:07
Breath Volume	1.750	23:07
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

---

Data Download Date/Time 03/11/2010 13:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHANAN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/10/2010 23:36:14

---

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:29
Subject Sample #1	0.141	23:30
Breath Volume	1.371	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample #2	0.145	23:34
Breath Volume	1.566	23:34
Air Blank	0.000	23:35
Control Test	0.079	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

---

Data Download Date/Time 03/11/2010 13:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKEE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: BEANS

DATE/TIME OF TEST: 03/11/2010 02:08:58

---

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Subject Sample #1	0.182	02:04
Breath Volume	2.539	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:06
Subject Sample #2	0.177	02:07
Breath Volume	1.777	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

---

Data Download Date/Time 03/11/2010 13:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 03/11/2010 02:40:52

---

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.081	02:38	
Air Blank	0.000	02:39	
Subject Sample #1		02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.081	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

---

Data Download Date/Time 03/11/2010 13:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 03/11/2010 02:48:38

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Subject Sample #1	0.212	02:44
Breath Volume	1.378	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample #2	0.217	02:47
Breath Volume	1.925	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

---

Data Download Date/Time 03/11/2010 13:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CARSON, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: PENDLETON

DATE/TIME OF TEST: 03/11/2010 03:54:41

---

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:49
Subject Sample #1	0.167	03:50
Breath Volume	1.929	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample #2	0.166	03:53
Breath Volume	2.378	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

---

Data Download Date/Time 03/11/2010 13:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBSON, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: KENNEY

DATE/TIME OF TEST: 03/11/2010 04:20:20

---

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Subject Sample #1	0.167	04:15
Breath Volume	1.980	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample #2	0.170	04:18
Breath Volume	1.777	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

---

Data Download Date/Time 03/11/2010 13:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIDGWAY, M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: ADMIN  
ARREST TIME: 15:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SOMOHAND

DATE/TIME OF TEST: 03/11/2010 16:40:21

---

Diagnostics Check	OK	16:32
Air Blank	0.000	16:32
Control Test	0.081	16:33
Air Blank	0.000	16:33
Subject Sample #1	0.447	16:35
Breath Volume	1.433	16:35
Air Blank	0.000	16:36
Air Blank	0.000	16:38
Subject Sample #2	0.445	16:38
Breath Volume	2.523	16:38
Air Blank	0.000	16:39
Control Test	0.079	16:39
Air Blank	0.000	16:40
Diagnostics Check	OK	16:40

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LAROSE, JAMES  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MCCULLAR

DATE/TIME OF TEST: 03/12/2010 00:36:50

---

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Subject Sample #1	0.248	00:32
Breath Volume	3.304	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:34
Subject Sample #2	0.256	00:35
Breath Volume	3.378	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCCALLA, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: RANSOM III

DATE/TIME OF TEST: 03/12/2010 01:03:37

---

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:58
Air Blank	0.000	00:58
Subject Sample #1	0.150	00:59
Breath Volume	1.906	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample #2	0.148	01:02
Breath Volume	2.410	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: INSIGNARES, P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: HERNANDEZ PEREZ

DATE/TIME OF TEST: 03/12/2010 01:47:53

---

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Subject Sample #1	0.126	01:43
Breath Volume	2.273	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample #2	0.128	01:46
Breath Volume	2.839	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FIGUEROA, D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/12/2010 20:11:22

---

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.081	20:06
Air Blank	0.000	20:06
Subject Sample #1	0.000	20:07
Breath Volume	1.824	20:07
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample #2	0.000	20:10
Breath Volume	1.417	20:10
Air Blank	0.000	20:10
Control Test	0.081	20:10
Air Blank	0.000	20:11
Diagnostics Check	OK	20:11

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUSWELL, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 03/12/2010 23:32:12

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Subject Sample #1	0.000	23:27
Breath Volume	4.164	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample #2	0.000	23:30
Breath Volume	4.199	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SORETH

DATE/TIME OF TEST: 03/13/2010 00:01:17

---

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:56
Subject Sample #1	0.189	23:56
Breath Volume	1.949	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample #2	0.189	23:59
Breath Volume	1.753	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DRAKE, J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 03/13/2010 02:06:35

---

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.081	01:52	
Air Blank	0.000	01:53	
Subject Sample #1	0.185	01:56	Volume Not Met
Breath Volume	0.074	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:58	
Subject Sample #2	0.173	01:59	
Breath Volume	1.527	01:59	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample #3	0.177	02:04	
Breath Volume	1.406	02:04	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORREY, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 03/13/2010 02:29:28

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.079	02:24
Air Blank	0.000	02:24
Subject Sample #1	0.185	02:24
Breath Volume	1.589	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample #2	0.198	02:27
Breath Volume	2.222	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MATTIZA, M  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SHAMSUDDIN

DATE/TIME OF TEST: 03/13/2010 02:55:44

---

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Subject Sample #1	0.095	02:50
Breath Volume	1.226	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample #2	0.095	02:54
Breath Volume	1.722	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCCOMBER, S  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: PLUMMER

DATE/TIME OF TEST: 03/13/2010 03:29:54

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.081	03:24
Air Blank	0.000	03:24
Subject Sample #1	0.213	03:25
Breath Volume	2.593	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample #2	0.216	03:28
Breath Volume	2.937	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORREY, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: SINCLAIR

DATE/TIME OF TEST: 03/13/2010 04:25:57

---

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Subject Sample #1	0.148	04:21
Breath Volume	1.703	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.154	04:24
Breath Volume	1.757	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RICH, D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: CLAIBORNE

DATE/TIME OF TEST: 03/13/2010 06:06:34

---

Diagnostics Check	OK	06:00
Air Blank	0.000	06:00
Control Test	0.081	06:01
Air Blank	0.000	06:01
Subject Sample #1	0.136	06:02
Breath Volume	1.554	06:02
Air Blank	0.000	06:02
Air Blank	0.000	06:04
Subject Sample #2	0.136	06:05
Breath Volume	1.824	06:05
Air Blank	0.000	06:05
Control Test	0.081	06:05
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L. D.  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: RETANA-SALAZAR

DATE/TIME OF TEST: 03/13/2010 21:02:59

---

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.081	20:57
Air Blank	0.000	20:57
Subject Sample #1	0.090	20:58
Breath Volume	1.933	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample #2	0.093	21:01
Breath Volume	2.468	21:01
Air Blank	0.000	21:02
Control Test	0.081	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: GAJEWSKI

DATE/TIME OF TEST: 03/14/2010 00:49:02

---

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.081	00:32	
Air Blank	0.000	00:33	
Subject Sample #1	0.201	00:36	
Breath Volume	1.484	00:36	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample #2	0.148	00:41	Volume Not Met
Breath Volume	0.308	00:41	
Air Blank	0.000	00:42	
Air Blank	0.000	00:44	
Subject Sample #3	0.169	00:47	Volume Not Met
Breath Volume	0.187	00:47	
Air Blank	0.000	00:48	
Control Test	0.079	00:48	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:49	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: GAJEWSKI

DATE/TIME OF TEST: 03/14/2010 00:58:27

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Subject Sample #1	0.212	00:52
Breath Volume	2.273	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample #2	0.208	00:56
Breath Volume	1.839	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORRELL, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: MEREDITH

DATE/TIME OF TEST: 03/14/2010 03:35:39

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:30
Subject Sample #1	0.142	03:31
Breath Volume	1.773	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample #2	0.141	03:34
Breath Volume	1.792	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: WOLLENBERG

DATE/TIME OF TEST: 03/14/2010 04:18:35

---

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample #1	0.082	04:14
Breath Volume	2.531	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample #2	0.082	04:17
Breath Volume	2.140	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORREY, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PRAWL

DATE/TIME OF TEST: 03/14/2010 04:41:29

---

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:35
Air Blank	0.000	04:35
Subject Sample #1	0.100	04:36
Breath Volume	1.488	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:38
Subject Sample #2	0.100	04:39
Breath Volume	1.406	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, TALIA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCMURPHY, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: COLBERT

DATE/TIME OF TEST: 03/14/2010 06:25:01

---

Diagnostics Check	OK	06:18
Air Blank	0.000	06:19
Control Test	0.080	06:19
Air Blank	0.000	06:20
Subject Sample #1	0.059	06:20
Breath Volume	1.531	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:22
Subject Sample #2	0.061	06:23
Breath Volume	1.843	06:23
Air Blank	0.000	06:24
Control Test	0.081	06:24
Air Blank	0.000	06:24
Diagnostics Check	OK	06:25

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRAUSS, MARTIN, J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KACZMAREK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 03/14/2010 06:45:57

---

Diagnostics Check	OK	06:39
Air Blank	0.000	06:39
Control Test	0.081	06:40
Air Blank	0.000	06:40
Subject Sample #1	0.168	06:41
Breath Volume	4.941	06:41
Air Blank	0.000	06:41
Air Blank	0.000	06:43
Subject Sample #2	0.168	06:44
Breath Volume	3.398	06:44
Air Blank	0.000	06:45
Control Test	0.081	06:45
Air Blank	0.000	06:45
Diagnostics Check	OK	06:45

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRAUSS, MARTIN, J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ARNOLD, F  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 08:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:03

SUBJECT'S LAST NAME: SIRMANS

DATE/TIME OF TEST: 03/14/2010 09:31:34

---

Diagnostics Check	OK	09:23
Air Blank	0.000	09:24
Control Test	0.081	09:24
Air Blank	0.000	09:24
Subject Sample #1	0.057	09:26
Breath Volume	1.453	09:26
Air Blank	0.000	09:27
Air Blank	0.000	09:28
Subject Sample #2	0.053	09:30
Breath Volume	1.343	09:30
Air Blank	0.000	09:30
Control Test	0.081	09:30
Air Blank	0.000	09:31
Diagnostics Check	OK	09:31

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRAUSS, MARTIN, J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FOOTMAN, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: ADMIN  
ARREST TIME: 13:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/14/2010 15:41:13

---

Diagnostics Check	OK	15:33
Air Blank	0.000	15:34
Control Test	0.081	15:34
Air Blank	0.000	15:35
Subject Sample #1	0.115	15:36
Breath Volume	1.722	15:36
Air Blank	0.000	15:37
Air Blank	0.000	15:39
Subject Sample #2	0.112	15:39
Breath Volume	1.628	15:39
Air Blank	0.000	15:40
Control Test	0.081	15:40
Air Blank	0.000	15:41
Diagnostics Check	OK	15:41

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKEE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/15/2010 19:48:53

---

Diagnostics Check	OK	19:45	
Air Blank	0.000	19:46	
Control Test	0.081	19:46	
Air Blank	0.000	19:47	
Subject Sample #1	0.000	19:47	
Breath Volume	1.703	19:47	
Air Blank		19:48	Purge Fail
Air Blank	0.000	19:48	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKEE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/15/2010 19:56:37

---

Diagnostics Check	OK	19:50
Air Blank	0.000	19:50
Control Test	0.081	19:50
Air Blank	0.000	19:51
Subject Sample #1	0.000	19:52
Breath Volume	1.425	19:52
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample #2	0.000	19:55
Breath Volume	1.300	19:55
Air Blank	0.000	19:55
Control Test	0.081	19:56
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRUZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: COFIELD

DATE/TIME OF TEST: 03/15/2010 23:40:01

---

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:37	
Control Test	0.082	23:37	
Air Blank	0.000	23:38	
Subject Sample #1		23:38	Interferent Detect
Breath Volume	2.945	23:38	
Air Blank	0.000	23:39	
Control Test	0.081	23:39	Interferent Detect
Air Blank	0.000	23:40	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VALENTINE, S  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 03/16/2010 19:19:16

---

Diagnostics Check	OK	19:11
Air Blank	0.000	19:11
Control Test	0.081	19:12
Air Blank	0.000	19:12
Subject Sample #1	0.098	19:14
Breath Volume	1.812	19:14
Air Blank	0.000	19:15
Air Blank	0.000	19:16
Subject Sample #2	0.095	19:17
Breath Volume	1.804	19:17
Air Blank	0.000	19:18
Control Test	0.082	19:18
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: GUT

DATE/TIME OF TEST: 03/16/2010 21:23:28

---

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.082	21:18
Air Blank	0.000	21:18
Subject Sample #1	0.169	21:19
Breath Volume	3.160	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample #2	0.162	21:21
Breath Volume	3.234	21:21
Air Blank	0.000	21:22
Control Test	0.081	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCCALLA, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: CAMARILLO-MARTINEZ

DATE/TIME OF TEST: 03/17/2010 02:03:33

---

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.083	01:55
Air Blank	0.000	01:55
Subject Sample #1	0.154	01:57
Breath Volume	1.683	01:57
Air Blank	0.000	01:58
Air Blank	0.000	01:59
Subject Sample #2	0.150	02:01
Breath Volume	2.085	02:01
Air Blank	0.000	02:02
Control Test	0.081	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BORDEN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 03/17/2010 03:21:08

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.082	03:15
Air Blank	0.000	03:15
Subject Sample #1	0.158	03:16
Breath Volume	1.527	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample #2	0.161	03:19
Breath Volume	1.703	03:19
Air Blank	0.000	03:20
Control Test	0.082	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: STEAD, K  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: LOCKLEAR

DATE/TIME OF TEST: 03/17/2010 04:59:53

---

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Subject Sample #1	0.160	04:54
Breath Volume	1.367	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample #2	0.171	04:58
Breath Volume	1.554	04:58
Air Blank	0.000	04:58
Control Test	0.080	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: STEAD, K  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: MAGEE

DATE/TIME OF TEST: 03/17/2010 23:39:11

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.082	23:31
Air Blank	0.000	23:31
Subject Sample #1	0.139	23:32
Breath Volume	1.152	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample #2	0.124	23:37
Breath Volume	1.304	23:37
Air Blank	0.000	23:38
Control Test	0.081	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBSON, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/18/2010 00:29:49

---

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.082	00:24
Air Blank	0.000	00:24
Subject Sample #1	0.168	00:25
Breath Volume	2.675	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample #2	0.172	00:28
Breath Volume	2.453	00:28
Air Blank	0.000	00:28
Control Test	0.081	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORREY, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: TETIDRICK

DATE/TIME OF TEST: 03/18/2010 02:11:02

---

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.079	01:58	
Air Blank	0.000	01:59	
Subject Sample #1	0.168	02:00	
Breath Volume	1.152	02:00	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample #2	0.155	02:05	Volume Not Met
Breath Volume	0.699	02:05	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample #3	0.166	02:09	
Breath Volume	1.179	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:11	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: LIGGINS

DATE/TIME OF TEST: 03/18/2010 03:08:58

---

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Subject Sample #1	0.182	03:04
Breath Volume	1.773	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample #2	0.173	03:07
Breath Volume	1.289	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHANAN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: WASIELEWSKI

DATE/TIME OF TEST: 03/18/2010 03:59:39

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:54
Subject Sample #1	0.202	03:55
Breath Volume	2.312	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample #2	0.195	03:58
Breath Volume	2.109	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ANGELA, V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAVIS, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: HYMOWITZ

DATE/TIME OF TEST: 03/18/2010 04:37:33

---

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Subject Sample #1	0.189	04:33
Breath Volume	1.695	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample #2	0.178	04:35
Breath Volume	1.636	04:35
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BORDEN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: CHYGRIN

DATE/TIME OF TEST: 03/18/2010 06:09:03

---

Diagnostics Check	OK	06:02
Air Blank	0.000	06:02
Control Test	0.081	06:03
Air Blank	0.000	06:03
Subject Sample #1	0.177	06:04
Breath Volume	2.378	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample #2	0.188	06:07
Breath Volume	1.925	06:07
Air Blank	0.000	06:08
Control Test	0.081	06:08
Air Blank	0.000	06:08
Diagnostics Check	OK	06:09

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, REBECCA, J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TORRES, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 10:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:26

SUBJECT'S LAST NAME: WOOTTEN

DATE/TIME OF TEST: 03/18/2010 12:54:24

---

Diagnostics Check	OK	12:47
Air Blank	0.000	12:48
Control Test	0.082	12:48
Air Blank	0.000	12:49
Subject Sample #1	0.220	12:49
Breath Volume	2.039	12:49
Air Blank	0.000	12:50
Air Blank	0.000	12:52
Subject Sample #2	0.227	12:52
Breath Volume	2.863	12:52
Air Blank	0.000	12:53
Control Test	0.082	12:53
Air Blank	0.000	12:54
Diagnostics Check	OK	12:54

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORET, F  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 03/19/2010 04:25:51

---

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.082	04:20
Air Blank	0.000	04:20
Subject Sample #1	0.142	04:21
Breath Volume	2.417	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.138	04:24
Breath Volume	2.007	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DURAN, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 03/20/2010 03:55:59

---

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.082	03:50
Air Blank	0.000	03:50
Subject Sample #1	0.204	03:51
Breath Volume	2.148	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample #2	0.185	03:54
Breath Volume	1.644	03:54
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBSON, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MONTERO

DATE/TIME OF TEST: 03/20/2010 04:09:38

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Subject Sample #1	0.138	04:05
Breath Volume	1.371	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.141	04:08
Breath Volume	1.804	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KRUTHERS, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: TURPIN

DATE/TIME OF TEST: 03/20/2010 05:21:20

---

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:15
Air Blank	0.000	05:15
Subject Sample #1	0.199	05:16
Breath Volume	3.039	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample #2	0.194	05:19
Breath Volume	3.050	05:19
Air Blank	0.000	05:20
Control Test	0.080	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KENNEDY, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: LARKINS

DATE/TIME OF TEST: 03/20/2010 05:48:02

---

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.082	05:42
Air Blank	0.000	05:43
Subject Sample #1	0.170	05:43
Breath Volume	1.203	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:45
Subject Sample #2	0.177	05:46
Breath Volume	1.511	05:46
Air Blank	0.000	05:47
Control Test	0.081	05:47
Air Blank	0.000	05:47
Diagnostics Check	OK	05:48

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORET, F  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: CAJIGAS

DATE/TIME OF TEST: 03/20/2010 07:02:27

---

Diagnostics Check	OK	06:56
Air Blank	0.000	06:56
Control Test	0.081	06:56
Air Blank	0.000	06:57
Subject Sample #1	0.108	06:57
Breath Volume	3.972	06:57
Air Blank	0.000	06:58
Air Blank	0.000	07:00
Subject Sample #2	0.104	07:00
Breath Volume	3.582	07:00
Air Blank	0.000	07:01
Control Test	0.080	07:01
Air Blank	0.000	07:02
Diagnostics Check	OK	07:02

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 03/21/2010 03:41:56

---

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.082	03:39	
Air Blank	0.000	03:40	
Subject Sample #1		03:40	Subject Test Refused
Breath Volume	0.000	03:40	
Air Blank	0.000	03:41	
Control Test	0.082	03:41	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: IEROME

DATE/TIME OF TEST: 03/21/2010 05:18:26

---

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:04	
Control Test	0.081	05:04	
Air Blank	0.000	05:05	
Subject Sample #1	0.213	05:08	No .020 agreement
Breath Volume	1.496	05:08	
Air Blank	0.000	05:09	
Air Blank	0.000	05:10	
Subject Sample #2	0.239	05:13	No .020 agreement
Breath Volume	1.722	05:13	
Air Blank	0.000	05:14	
Air Blank	0.000	05:16	
Subject Sample #3	0.261	05:16	No .020 agreement
Breath Volume	2.187	05:16	
Air Blank	0.000	05:17	
Control Test	0.078	05:17	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:18	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: IEROME

DATE/TIME OF TEST: 03/21/2010 05:29:05

---

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Subject Sample #1	0.232	05:23
Breath Volume	1.207	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample #2	0.236	05:27
Breath Volume	1.667	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SUSTEK, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 03/22/2010 04:47:54

---

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Subject Sample #1	0.150	04:43
Breath Volume	1.644	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:45
Subject Sample #2	0.151	04:46
Breath Volume	2.019	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRAUSS, MARTIN, J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: STRPHENS, J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 15:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: DONNELL

DATE/TIME OF TEST: 03/22/2010 16:21:45

---

Diagnostics Check	OK	16:15
Air Blank	0.000	16:15
Control Test	0.082	16:16
Air Blank	0.000	16:16
Subject Sample #1	0.424	16:17
Breath Volume	1.816	16:17
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample #2	0.423	16:20
Breath Volume	1.894	16:20
Air Blank	0.000	16:20
Control Test	0.080	16:21
Air Blank	0.000	16:21
Diagnostics Check	OK	16:21

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GLOVER, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/23/2010 23:59:04

---

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.082	23:53
Air Blank	0.000	23:54
Subject Sample #1	0.103	23:54
Breath Volume	3.296	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample #2	0.102	23:57
Breath Volume	1.832	23:57
Air Blank	0.000	23:58
Control Test	0.081	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:59

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RAMER, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: DEHAVEN

DATE/TIME OF TEST: 03/24/2010 01:57:03

---

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample #1	0.188	01:52
Breath Volume	3.292	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample #2	0.176	01:55
Breath Volume	2.363	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:57

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHER, P  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/24/2010 04:33:16

---

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.082	04:27
Air Blank	0.000	04:28
Subject Sample #1	0.092	04:28
Breath Volume	1.957	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample #2	0.091	04:31
Breath Volume	2.214	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MIGLIORE, C  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: OLMEDO

DATE/TIME OF TEST: 03/24/2010 22:37:30

---

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:27	
Subject Sample #1	0.171	22:30	Volume Not Met
Breath Volume	0.097	22:30	
Air Blank	0.000	22:30	
Air Blank	0.000	22:32	
Subject Sample #2	0.183	22:35	Volume Not Met
Breath Volume	0.167	22:35	
Air Blank	0.000	22:36	
Control Test	0.079	22:36	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MIGLIORE, C  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: OLMEDO

DATE/TIME OF TEST: 03/24/2010 22:49:31

---

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:44
Air Blank	0.000	22:44
Subject Sample #1	0.191	22:45
Breath Volume	1.214	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample #2	0.184	22:47
Breath Volume	1.222	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCMURPHY, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: VEGA VEGA

DATE/TIME OF TEST: 03/25/2010 01:08:19

---

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Subject Sample #1	0.091	01:03
Breath Volume	1.707	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample #2	0.091	01:06
Breath Volume	1.898	01:06
Air Blank	0.000	01:07
Control Test	0.082	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, REBECCA, J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KELLELER, P  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 03/25/2010 07:15:21

---

Diagnostics Check	OK	07:08
Air Blank	0.000	07:09
Control Test	0.081	07:09
Air Blank	0.000	07:10
Subject Sample #1	0.131	07:10
Breath Volume	1.390	07:10
Air Blank	0.000	07:11
Air Blank	0.000	07:13
Subject Sample #2	0.134	07:13
Breath Volume	1.976	07:13
Air Blank	0.000	07:14
Control Test	0.081	07:14
Air Blank	0.000	07:15
Diagnostics Check	OK	07:15

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: STEAD, K  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SCHINKER

DATE/TIME OF TEST: 03/26/2010 02:23:12

---

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Subject Sample #1	0.150	02:18
Breath Volume	2.484	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample #2	0.146	02:21
Breath Volume	2.937	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VALLEJO, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: DOMANSKI

DATE/TIME OF TEST: 03/26/2010 02:34:57

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Subject Sample #1	0.200	02:29
Breath Volume	1.773	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample #2	0.197	02:33
Breath Volume	1.664	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHANAN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CHRISTEN

DATE/TIME OF TEST: 03/26/2010 03:04:26

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Subject Sample #1	0.124	02:59
Breath Volume	1.183	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample #2	0.137	03:02
Breath Volume	1.121	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKEE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: PADRON

DATE/TIME OF TEST: 03/26/2010 05:40:10

---

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:32
Subject Sample #1	0.236	05:34
Breath Volume	1.226	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample #2	0.240	05:38
Breath Volume	1.257	05:38
Air Blank	0.000	05:39
Control Test	0.078	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWENS, EDWARD, W  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MONTAGUE, M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: ADMIN  
ARREST TIME: 09:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: BRUNT

DATE/TIME OF TEST: 03/26/2010 10:21:50

---

Diagnostics Check	OK	10:15	
Air Blank	0.000	10:15	
Control Test	0.081	10:16	
Air Blank	0.000	10:16	
Subject Sample #1	0.274	10:17	
Breath Volume	1.105	10:17	
Air Blank	0.000	10:17	
Air Blank	0.000	10:19	
Subject Sample #2		10:20	Subject Test Refused
Breath Volume	0.000	10:20	
Air Blank	0.000	10:20	
Control Test	0.081	10:21	
Air Blank	0.000	10:21	
Diagnostics Check	OK	10:21	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, REBECCA, J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDERSON, R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 10:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:41

SUBJECT'S LAST NAME: SWERDLOFF

DATE/TIME OF TEST: 03/26/2010 11:14:09

---

Diagnostics Check	OK	11:06
Air Blank	0.000	11:07
Control Test	0.081	11:07
Air Blank	0.000	11:08
Subject Sample #1	0.000	11:09
Breath Volume	1.562	11:09
Air Blank	0.000	11:09
Air Blank	0.000	11:11
Subject Sample #2	0.000	11:12
Breath Volume	1.746	11:12
Air Blank	0.000	11:13
Control Test	0.081	11:13
Air Blank	0.000	11:14
Diagnostics Check	OK	11:14

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MELO, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: GONZALEZ ISERN

DATE/TIME OF TEST: 03/26/2010 23:10:37

---

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:04
Subject Sample #1	0.101	23:05
Breath Volume	1.398	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample #2	0.102	23:09
Breath Volume	1.304	23:09
Air Blank	0.000	23:09
Control Test	0.081	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RAMER, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: GESEMYER

DATE/TIME OF TEST: 03/26/2010 23:50:55

---

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:45
Subject Sample #1	0.269	23:46
Breath Volume	1.644	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample #2	0.270	23:49
Breath Volume	1.578	23:49
Air Blank	0.000	23:49
Control Test	0.079	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KACZMAREK, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: LAZARO HERRERA

DATE/TIME OF TEST: 03/27/2010 00:05:48

---

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Subject Sample #1	0.166	00:01
Breath Volume	1.746	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:03
Subject Sample #2	0.176	00:04
Breath Volume	1.921	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHANAN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 03/27/2010 01:05:54

---

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.081	01:00
Air Blank	0.000	01:00
Subject Sample #1	0.147	01:01
Breath Volume	1.390	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample #2	0.144	01:04
Breath Volume	1.937	01:04
Air Blank	0.000	01:04
Control Test	0.081	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBSON, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: PANZO

DATE/TIME OF TEST: 03/27/2010 01:50:17

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:44
Air Blank	0.000	01:44
Subject Sample #1	0.207	01:45
Breath Volume	2.621	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample #2	0.210	01:48
Breath Volume	1.507	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILSON, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: TALLMAN

DATE/TIME OF TEST: 03/27/2010 02:23:20

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Subject Sample #1	0.184	02:18
Breath Volume	1.468	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample #2	0.192	02:21
Breath Volume	2.347	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MCMURPHY, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: LOPEZ-DEJESUS

DATE/TIME OF TEST: 03/27/2010 03:52:45

---

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Subject Sample #1	0.087	03:48
Breath Volume	2.113	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:50
Subject Sample #2	0.092	03:51
Breath Volume	2.632	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORET, F  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/27/2010 04:08:22

---

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample #1	0.072	04:03
Breath Volume	1.687	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample #2	0.076	04:06
Breath Volume	2.898	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RAMER, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GODDEN

DATE/TIME OF TEST: 03/27/2010 04:33:01

---

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:27
Subject Sample #1	0.130	04:28
Breath Volume	2.058	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample #2	0.132	04:31
Breath Volume	2.429	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:33

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FOSTER, E  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: WINKLER

DATE/TIME OF TEST: 03/27/2010 04:45:11

---

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.081	04:36
Air Blank	0.000	04:36
Subject Sample #1	0.221	04:39
Breath Volume	1.765	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample #2	0.226	04:43
Breath Volume	1.937	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MELONIE, E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BUCHANAN, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: SPECTOR

DATE/TIME OF TEST: 03/27/2010 04:54:45

---

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:49
Subject Sample #1	0.000	04:50
Breath Volume	3.621	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample #2	0.000	04:53
Breath Volume	3.519	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WHITEHEAD, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: CHADWICK

DATE/TIME OF TEST: 03/27/2010 21:31:45

---

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.082	21:26
Air Blank	0.000	21:26
Subject Sample #1	0.350	21:27
Breath Volume	2.324	21:27
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample #2	0.331	21:30
Breath Volume	1.406	21:30
Air Blank	0.000	21:30
Control Test	0.080	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RICE, K  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: BUGGS

DATE/TIME OF TEST: 03/28/2010 01:14:42

---

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:06
Air Blank	0.000	01:06
Subject Sample #1	0.205	01:09
Breath Volume	1.906	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample #2	0.188	01:13
Breath Volume	1.398	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MORREY, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 03/28/2010 03:46:10

---

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.081	03:35	
Air Blank	0.000	03:35	
Subject Sample #1	0.139	03:38	Volume Not Met
Breath Volume	0.054	03:38	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample #2	0.143	03:44	Volume Not Met
Breath Volume	0.390	03:44	
Air Blank	0.000	03:45	
Control Test	0.081	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILAN-CINTRON, E  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JIMENEZ-LEON

DATE/TIME OF TEST: 03/28/2010 04:13:40

---

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Subject Sample #1		04:06	No Sample Provided
Breath Volume	0.000	04:06	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample #2		04:12	No Sample Provided
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Control Test	0.080	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILLAN-CINTRON, E  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: JIMENEZ-LEOM

DATE/TIME OF TEST: 03/28/2010 04:26:15

---

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Subject Sample #1	0.177	04:21
Breath Volume	2.714	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample #2	0.172	04:24
Breath Volume	2.472	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILSON, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: AMIDA

DATE/TIME OF TEST: 03/29/2010 00:37:35

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.081	00:32
Air Blank	0.000	00:32
Subject Sample #1	0.202	00:33
Breath Volume	2.425	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample #2	0.202	00:36
Breath Volume	2.093	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILLAN-CINTRON, E  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/29/2010 03:57:00

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Subject Sample #1	0.173	03:52
Breath Volume	1.976	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample #2	0.172	03:55
Breath Volume	1.285	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINTON, JENNA, R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILLAN-CINTRON, E  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 03/30/2010 04:40:08

---

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.082	04:34
Air Blank	0.000	04:34
Subject Sample #1	0.204	04:35
Breath Volume	2.261	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:37
Subject Sample #2	0.213	04:38
Breath Volume	2.183	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FAUSKEE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: FORNEY

DATE/TIME OF TEST: 03/30/2010 19:54:14

---

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.083	19:48
Air Blank	0.000	19:49
Subject Sample #1	0.184	19:49
Breath Volume	1.652	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample #2	0.180	19:52
Breath Volume	1.246	19:52
Air Blank	0.000	19:53
Control Test	0.082	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CARTER, A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: HASBROOK

DATE/TIME OF TEST: 03/31/2010 00:04:46

---

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:02	
Control Test	0.082	00:02	
Air Blank	0.000	00:03	
Subject Sample #1		00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Control Test	0.081	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/31/2010 02:09:28

---

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.082	02:03
Air Blank	0.000	02:04
Subject Sample #1	0.147	02:05
Breath Volume	1.156	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample #2	0.167	02:07
Breath Volume	2.496	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HYDER, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: KROSLAK

DATE/TIME OF TEST: 03/31/2010 04:13:08

---

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.082	04:07
Air Blank	0.000	04:08
Subject Sample #1	0.115	04:08
Breath Volume	1.855	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample #2	0.116	04:11
Breath Volume	2.222	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KENNEDY, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: PEGG

DATE/TIME OF TEST: 03/31/2010 04:55:20

---

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.082	04:49
Air Blank	0.000	04:50
Subject Sample #1	0.203	04:50
Breath Volume	2.113	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample #2	0.205	04:53
Breath Volume	1.640	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THOMPSON, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646831  
DRY GAS EXPIRY DATE: 07/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: CLEMEN

DATE/TIME OF TEST: 03/31/2010 20:51:28

---

Diagnostics Check	OK	20:44
Air Blank	0.000	20:45
Control Test	0.081	20:45
Air Blank	0.000	20:45
Subject Sample #1	0.096	20:46
Breath Volume	1.789	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample #2	0.096	20:49
Breath Volume	1.800	20:49
Air Blank	0.000	20:50
Control Test	0.081	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

---

Data Download Date/Time 04/02/2010 10:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRERETON, PAUL, D

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: MONROE, EUGENE

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 03/08/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 807901I

DRY GAS EXPIRY DATE: 03/21/2010

VIOLATION CODE: DUI

ARREST TIME: 23:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/14/2010 00:03:23

---

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.079	23:52	
Air Blank	0.000	23:52	
Subject Sample #1	0.201	23:56	Volume Not Met
Breath Volume	0.117	23:56	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample #2	0.220	00:01	Volume Not Met
Breath Volume	0.753	00:01	
Air Blank	0.000	00:02	
Control Test	0.077	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

---

Data Download Date/Time 04/01/2010 16:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRERETON, PAUL, D

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: MONROE, EUGENE

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 03/08/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 807901I

DRY GAS EXPIRY DATE: 03/21/2010

VIOLATION CODE: DUI

ARREST TIME: 23:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/14/2010 00:15:15

---

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.076	00:09
Air Blank	0.000	00:09
Subject Sample #1	0.236	00:10
Breath Volume	1.566	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample #2	0.233	00:13
Breath Volume	1.722	00:13
Air Blank	0.000	00:14
Control Test	0.078	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

---

Data Download Date/Time 04/01/2010 16:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MIXSON, ROBERT, E

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: MIXSON, ROBERT

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 03/08/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 812301G

DRY GAS EXPIRY DATE: 05/06/2010

VIOLATION CODE: DUI

ARREST TIME: 11:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: GOODSON

DATE/TIME OF TEST: 03/22/2010 16:31:45

---

Diagnostics Check	OK	16:28	
Air Blank	0.000	16:29	
Control Test	0.078	16:29	
Air Blank	0.000	16:30	
Subject Sample #1		16:30	Subject Test Refused
Breath Volume	0.000	16:30	
Air Blank	0.000	16:30	
Control Test	0.078	16:31	
Air Blank	0.000	16:31	
Diagnostics Check	OK	16:31	

---

Data Download Date/Time 04/01/2010 16:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STUART, JOHN, B

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: STUART, JOHN

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: OSTRANDER

DATE/TIME OF TEST: 03/07/2010 01:41:09

---

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:35
Subject Sample #1	0.181	01:36
Breath Volume	3.750	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample #2	0.179	01:39
Breath Volume	2.558	01:39
Air Blank	0.000	01:40
Control Test	0.078	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

---

Data Download Date/Time 04/07/2010 22:53

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STUART, JOHN, B

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: WALKER, SHELIA

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 23:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 03/11/2010 23:51:50

---

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.078	23:46
Air Blank	0.000	23:46
Subject Sample #1	0.148	23:47
Breath Volume	2.585	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample #2	0.144	23:50
Breath Volume	2.750	23:50
Air Blank	0.000	23:50
Control Test	0.078	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

---

Data Download Date/Time 04/07/2010 22:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALKER, SHELIA,

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: WALKER, SHELIA

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 03:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRINKLEY

DATE/TIME OF TEST: 03/14/2010 03:53:18

---

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.078	03:47
Air Blank	0.000	03:47
Subject Sample #1	0.230	03:48
Breath Volume	1.683	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:50
Subject Sample #2	0.228	03:51
Breath Volume	2.921	03:51
Air Blank	0.000	03:52
Control Test	0.076	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

---

Data Download Date/Time 04/07/2010 22:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALKER, SHELIA,

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: WALKER, SHELIA

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 10:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: RUDDLE

DATE/TIME OF TEST: 03/16/2010 23:09:28

---

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.078	23:03
Air Blank	0.000	23:04
Subject Sample #1	0.177	23:04
Breath Volume	3.781	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample #2	0.175	23:07
Breath Volume	4.488	23:07
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

---

Data Download Date/Time 04/07/2010 22:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALKER, SHELIA,

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: GOLUB, JAY

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 02:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: BENZ

DATE/TIME OF TEST: 03/20/2010 02:54:56

---

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:49
Subject Sample #1	0.092	02:50
Breath Volume	1.984	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample #2	0.087	02:53
Breath Volume	1.632	02:53
Air Blank	0.000	02:54
Control Test	0.078	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

---

Data Download Date/Time 04/07/2010 22:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TILLIE, JOHN, R

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: TILLIE, JOHN

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 10:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 03/26/2010 00:28:25

---

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:23
Subject Sample #1	0.000	00:24
Breath Volume	1.917	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample #2	0.000	00:27
Breath Volume	1.949	00:27
Air Blank	0.000	00:27
Control Test	0.077	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

---

Data Download Date/Time 04/07/2010 22:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALKER, SHELIA,

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: PACE, TRACY

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 14:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: STERLING

DATE/TIME OF TEST: 03/26/2010 14:34:15

---

Diagnostics Check	OK	14:28
Air Blank	0.000	14:28
Control Test	0.078	14:28
Air Blank	0.000	14:29
Subject Sample #1	0.000	14:29
Breath Volume	1.773	14:29
Air Blank	0.000	14:30
Air Blank	0.000	14:32
Subject Sample #2	0.000	14:32
Breath Volume	2.437	14:32
Air Blank	0.000	14:33
Control Test	0.077	14:33
Air Blank	0.000	14:34
Diagnostics Check	OK	14:34

---

Data Download Date/Time 04/07/2010 22:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PUTNEL, THOMAS, K

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: MIDDLETON, HANK

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 12:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 03/27/2010 00:42:37

---

Diagnostics Check	OK	00:41	
Air Blank		00:42	RFI Detect
Air Blank	0.000	00:42	

---

Data Download Date/Time 04/07/2010 22:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PUTNEL, THOMAS, K

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: MIDDLETON, HANK

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 00:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 03/27/2010 00:59:20

---

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Subject Sample #1		00:55	No Sample Provided
Breath Volume	0.000	00:55	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample #2		00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:58	
Control Test	0.078	00:58	
Air Blank	0.000	00:59	
Diagnostics Check	OK	00:59	

---

Data Download Date/Time 04/07/2010 22:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HISLER-PACE, TRACY,

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: WALKER, SHELIA

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 17:19

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 03/29/2010 19:56:30

---

Diagnostics Check	OK	19:50
Air Blank	0.000	19:50
Control Test	0.079	19:51
Air Blank	0.000	19:51
Subject Sample #1	0.000	19:52
Breath Volume	2.656	19:52
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample #2	0.000	19:55
Breath Volume	2.890	19:55
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

---

Data Download Date/Time 04/07/2010 22:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TAYLOR, JAMES, R

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: GOLUB, JAY

ARRESTING OFFICER AGENCY: Lake City PD

LAST AGENCY INSPECTION: 03/04/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 22:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: INGRAHAM

DATE/TIME OF TEST: 03/30/2010 23:27:03

---

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.076	23:22	
Air Blank	0.000	23:23	
Subject Sample #1	0.321	23:23	
Breath Volume	2.660	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:25	
Subject Sample #2		23:26	Sequence Aborted
Breath Volume	0.000	23:26	
Air Blank	0.000	23:27	

---

Data Download Date/Time 04/07/2010 22:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY CORR  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEEMS, JASON, A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: TEEMS, JASON  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: ROECK

DATE/TIME OF TEST: 03/09/2010 22:41:42

---

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.079	22:32	
Air Blank	0.000	22:32	
Subject Sample #1	0.162	22:35	Volume Not Met
Breath Volume	0.015	22:35	
Air Blank	0.000	22:36	
Air Blank	0.000	22:38	
Subject Sample #2		22:40	Subject Test Refused
Breath Volume	0.984	22:40	
Air Blank	0.000	22:40	
Control Test	0.078	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

---

Data Download Date/Time 03/24/2010 11:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY CORR  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEEMS, JASON, A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: TEEMS, JASON  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 622245  
DRY GAS EXPIRY DATE: 05/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SHIPLEY

DATE/TIME OF TEST: 03/14/2010 00:10:01

---

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.079	00:05	
Air Blank	0.000	00:05	
Subject Sample #1		00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:09	
Control Test	0.080	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:10	

---

Data Download Date/Time 03/24/2010 11:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILLEY, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HILLEY, MICHAEL  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/28/2010

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BORRESEN

DATE/TIME OF TEST: 03/03/2010 00:54:19

---

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.081	00:51	
Air Blank	0.000	00:52	
Subject Sample #1	0.237	00:52	
Breath Volume	1.597	00:52	
Air Blank		00:53	Purge Fail
Air Blank	0.000	00:54	

---

Data Download Date/Time 03/24/2010 11:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HILLEY, MICHAEL, E

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: HILLEY, MICHAEL

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 02/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 02:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 03/16/2010 02:48:19

---

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.077	02:43	
Air Blank	0.000	02:43	
Subject Sample #1	0.188	02:44	
Breath Volume	3.460	02:44	
Air Blank	0.000	02:44	
Air Blank	0.000	02:46	
Subject Sample #2	0.181	02:47	
Breath Volume	4.085	02:47	
Air Blank		02:47	Purge Fail
Air Blank	0.000	02:48	

---

Data Download Date/Time 03/24/2010 11:22