

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEAD, C, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MULVEY, K  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: WODRAZKA

DATE/TIME OF TEST: 11/01/2011 00:29:39

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:23 |
| Air Blank         | 0.000 | 00:23 |
| Control Test      | 0.078 | 00:24 |
| Air Blank         | 0.000 | 00:24 |
| Subject Sample #1 | 0.118 | 00:25 |
| Breath Volume     | 2.042 | 00:25 |
| Air Blank         | 0.000 | 00:25 |
| Air Blank         | 0.000 | 00:27 |
| Subject Sample #2 | 0.122 | 00:28 |
| Breath Volume     | 2.257 | 00:28 |
| Air Blank         | 0.000 | 00:28 |
| Control Test      | 0.078 | 00:29 |
| Air Blank         | 0.000 | 00:29 |
| Diagnostics Check | OK    | 00:29 |

---

Data Download Date/Time 11/01/2011 12:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIEIGA, FRANK,  
BREATH TEST OPERATOR AGENCY: North Bay Village PD  
ARRESTING OFFICER: PIEIGA, FRANK  
ARRESTING OFFICER AGENCY: North Bay Village PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 11/01/2011 23:12:51

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:05 |
| Air Blank         | 0.000 | 23:05 |
| Control Test      | 0.080 | 23:06 |
| Air Blank         | 0.000 | 23:06 |
| Subject Sample #1 | 0.179 | 23:07 |
| Breath Volume     | 2.808 | 23:07 |
| Air Blank         | 0.000 | 23:08 |
| Air Blank         | 0.000 | 23:09 |
| Subject Sample #2 | 0.174 | 23:11 |
| Breath Volume     | 2.078 | 23:11 |
| Air Blank         | 0.000 | 23:11 |
| Control Test      | 0.079 | 23:12 |
| Air Blank         | 0.000 | 23:12 |
| Diagnostics Check | OK    | 23:12 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON, ORLARRY, S  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: BACALLAO, RAUL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/03/2011 05:09:26

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:02 |
| Air Blank         | 0.000 | 05:03 |
| Control Test      | 0.080 | 05:03 |
| Air Blank         | 0.000 | 05:04 |
| Subject Sample #1 | 0.271 | 05:04 |
| Breath Volume     | 1.699 | 05:04 |
| Air Blank         | 0.000 | 05:05 |
| Air Blank         | 0.000 | 05:07 |
| Subject Sample #2 | 0.259 | 05:07 |
| Breath Volume     | 2.460 | 05:07 |
| Air Blank         | 0.000 | 05:08 |
| Control Test      | 0.078 | 05:08 |
| Air Blank         | 0.000 | 05:09 |
| Diagnostics Check | OK    | 05:09 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH, FELICIA, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH, FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: SCHAIBLE

DATE/TIME OF TEST: 11/04/2011 00:58:55

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:52 |
| Air Blank         | 0.000 | 00:53 |
| Control Test      | 0.080 | 00:53 |
| Air Blank         | 0.000 | 00:53 |
| Subject Sample #1 | 0.088 | 00:54 |
| Breath Volume     | 1.460 | 00:54 |
| Air Blank         | 0.000 | 00:55 |
| Air Blank         | 0.000 | 00:56 |
| Subject Sample #2 | 0.089 | 00:57 |
| Breath Volume     | 2.105 | 00:57 |
| Air Blank         | 0.000 | 00:57 |
| Control Test      | 0.078 | 00:58 |
| Air Blank         | 0.000 | 00:58 |
| Diagnostics Check | OK    | 00:58 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: JEAN-BAPTISTE

DATE/TIME OF TEST: 11/05/2011 02:41:09

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:32 |
| Air Blank         | 0.000 | 02:33 |
| Control Test      | 0.080 | 02:33 |
| Air Blank         | 0.000 | 02:34 |
| Subject Sample #1 | 0.178 | 02:35 |
| Breath Volume     | 1.937 | 02:35 |
| Air Blank         | 0.000 | 02:36 |
| Air Blank         | 0.000 | 02:38 |
| Subject Sample #2 | 0.176 | 02:39 |
| Breath Volume     | 1.679 | 02:39 |
| Air Blank         | 0.000 | 02:40 |
| Control Test      | 0.078 | 02:40 |
| Air Blank         | 0.000 | 02:41 |
| Diagnostics Check | OK    | 02:41 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH, FELICIA, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GOMEZ, P  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 11/08/2011 02:23:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:20 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Control Test      | 0.081 | 02:20 |                      |
| Air Blank         | 0.000 | 02:21 |                      |
| Subject Sample #1 |       | 02:21 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:21 |                      |
| Air Blank         | 0.000 | 02:22 |                      |
| Control Test      | 0.080 | 02:22 |                      |
| Air Blank         | 0.000 | 02:23 |                      |
| Diagnostics Check | OK    | 02:23 |                      |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL, REGINALD, J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: LOCKE, M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HASKINS

DATE/TIME OF TEST: 11/08/2011 02:42:26

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:35 |
| Air Blank         | 0.000 | 02:35 |
| Control Test      | 0.080 | 02:36 |
| Air Blank         | 0.000 | 02:36 |
| Subject Sample #1 | 0.174 | 02:37 |
| Breath Volume     | 1.445 | 02:37 |
| Air Blank         | 0.000 | 02:38 |
| Air Blank         | 0.000 | 02:40 |
| Subject Sample #2 | 0.165 | 02:40 |
| Breath Volume     | 1.472 | 02:40 |
| Air Blank         | 0.000 | 02:41 |
| Control Test      | 0.078 | 02:41 |
| Air Blank         | 0.000 | 02:42 |
| Diagnostics Check | OK    | 02:42 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: LIQUITAY

DATE/TIME OF TEST: 11/10/2011 00:11:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:07 |                      |
| Air Blank         | 0.000 | 00:07 |                      |
| Control Test      | 0.081 | 00:08 |                      |
| Air Blank         | 0.000 | 00:08 |                      |
| Subject Sample #1 |       | 00:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:10 |                      |
| Air Blank         | 0.000 | 00:10 |                      |
| Control Test      | 0.080 | 00:10 |                      |
| Air Blank         | 0.000 | 00:11 |                      |
| Diagnostics Check | OK    | 00:11 |                      |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 11/10/2011 03:55:36

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 03:40 |                |
| Air Blank         | 0.000 | 03:40 |                |
| Control Test      | 0.080 | 03:41 |                |
| Air Blank         | 0.000 | 03:41 |                |
| Subject Sample #1 | 0.180 | 03:44 | Volume Not Met |
| Breath Volume     | 0.105 | 03:44 |                |
| Air Blank         | 0.000 | 03:45 |                |
| Air Blank         | 0.000 | 03:47 |                |
| Subject Sample #2 | 0.174 | 03:50 |                |
| Breath Volume     | 1.277 | 03:50 |                |
| Air Blank         | 0.000 | 03:51 |                |
| Air Blank         | 0.000 | 03:52 |                |
| Subject Sample #3 | 0.180 | 03:53 |                |
| Breath Volume     | 1.453 | 03:53 |                |
| Air Blank         | 0.000 | 03:54 |                |
| Control Test      | 0.078 | 03:54 |                |
| Air Blank         | 0.000 | 03:55 |                |
| Diagnostics Check | OK    | 03:55 |                |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GOMEZ, PETE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 11/10/2011 05:32:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:20 |                      |
| Air Blank         | 0.000 | 05:21 |                      |
| Control Test      | 0.080 | 05:21 |                      |
| Air Blank         | 0.000 | 05:22 |                      |
| Subject Sample #1 | 0.096 | 05:25 | Volume Not Met       |
| Breath Volume     | 0.453 | 05:25 |                      |
| Air Blank         | 0.000 | 05:25 |                      |
| Air Blank         | 0.000 | 05:27 |                      |
| Subject Sample #2 |       | 05:30 | Subject Test Refused |
| Breath Volume     | 0.105 | 05:30 |                      |
| Air Blank         | 0.000 | 05:31 |                      |
| Control Test      | 0.080 | 05:31 |                      |
| Air Blank         | 0.000 | 05:32 |                      |
| Diagnostics Check | OK    | 05:32 |                      |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILIAN, WANDA,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: TANIS, DAREU  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 15:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:43

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/12/2011 16:16:46

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 16:10 |
| Air Blank         | 0.000 | 16:11 |
| Control Test      | 0.081 | 16:11 |
| Air Blank         | 0.000 | 16:11 |
| Subject Sample #1 | 0.000 | 16:12 |
| Breath Volume     | 4.035 | 16:12 |
| Air Blank         | 0.000 | 16:12 |
| Air Blank         | 0.000 | 16:14 |
| Subject Sample #2 | 0.000 | 16:15 |
| Breath Volume     | 4.097 | 16:15 |
| Air Blank         | 0.000 | 16:15 |
| Control Test      | 0.079 | 16:16 |
| Air Blank         | 0.000 | 16:16 |
| Diagnostics Check | OK    | 16:16 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVILA, DANIEL,  
BREATH TEST OPERATOR AGENCY: Golden Beach PD  
ARRESTING OFFICER: AVILA, DANIEL  
ARRESTING OFFICER AGENCY: Golden Beach PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FARCASIU

DATE/TIME OF TEST: 11/12/2011 23:28:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:25 |                      |
| Air Blank         | 0.000 | 23:26 |                      |
| Control Test      | 0.079 | 23:26 |                      |
| Air Blank         | 0.000 | 23:26 |                      |
| Subject Sample #1 |       | 23:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:27 |                      |
| Air Blank         | 0.000 | 23:27 |                      |
| Control Test      | 0.080 | 23:27 |                      |
| Air Blank         | 0.000 | 23:28 |                      |
| Diagnostics Check | OK    | 23:28 |                      |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COIMIN, JEAN, W  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: JACKSON, SHANNON  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: LOPEZ VALDIVIA

DATE/TIME OF TEST: 11/13/2011 02:12:56

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:06 |
| Air Blank         | 0.000 | 02:06 |
| Control Test      | 0.081 | 02:07 |
| Air Blank         | 0.000 | 02:07 |
| Subject Sample #1 | 0.189 | 02:08 |
| Breath Volume     | 1.558 | 02:08 |
| Air Blank         | 0.000 | 02:08 |
| Air Blank         | 0.000 | 02:10 |
| Subject Sample #2 | 0.197 | 02:11 |
| Breath Volume     | 2.738 | 02:11 |
| Air Blank         | 0.000 | 02:12 |
| Control Test      | 0.081 | 02:12 |
| Air Blank         | 0.000 | 02:12 |
| Diagnostics Check | OK    | 02:12 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOGELGREN, D, S  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: FOGELGREN, D  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 11/13/2011 02:29:34

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:21 |
| Air Blank         | 0.000 | 02:21 |
| Control Test      | 0.080 | 02:22 |
| Air Blank         | 0.000 | 02:22 |
| Subject Sample #1 | 0.177 | 02:24 |
| Breath Volume     | 2.656 | 02:24 |
| Air Blank         | 0.000 | 02:25 |
| Air Blank         | 0.000 | 02:26 |
| Subject Sample #2 | 0.176 | 02:27 |
| Breath Volume     | 2.617 | 02:27 |
| Air Blank         | 0.000 | 02:28 |
| Control Test      | 0.077 | 02:28 |
| Air Blank         | 0.000 | 02:29 |
| Diagnostics Check | OK    | 02:29 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIETO, ANDRES, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PRIETO, ANDRES  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: KOVACS

DATE/TIME OF TEST: 11/13/2011 22:32:39

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:29 |                      |
| Air Blank         | 0.000 | 22:30 |                      |
| Control Test      | 0.080 | 22:30 |                      |
| Air Blank         | 0.000 | 22:31 |                      |
| Subject Sample #1 |       | 22:31 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:31 |                      |
| Air Blank         | 0.000 | 22:31 |                      |
| Control Test      | 0.081 | 22:32 |                      |
| Air Blank         | 0.000 | 22:32 |                      |
| Diagnostics Check | OK    | 22:32 |                      |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNDY, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: GOMEZ, PEDRO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 11/17/2011 01:01:26

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:54 |
| Air Blank         | 0.000 | 00:54 |
| Control Test      | 0.078 | 00:55 |
| Air Blank         | 0.000 | 00:55 |
| Subject Sample #1 | 0.220 | 00:56 |
| Breath Volume     | 3.667 | 00:56 |
| Air Blank         | 0.000 | 00:57 |
| Air Blank         | 0.000 | 00:59 |
| Subject Sample #2 | 0.227 | 00:59 |
| Breath Volume     | 3.457 | 00:59 |
| Air Blank         | 0.000 | 01:00 |
| Control Test      | 0.076 | 01:00 |
| Air Blank         | 0.000 | 01:01 |
| Diagnostics Check | OK    | 01:01 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOEL, LINDSLEY, J  
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD  
ARRESTING OFFICER: NOEL, LINDSLEY  
ARRESTING OFFICER AGENCY: Bay Harbor Island PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 11/19/2011 05:46:43

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:40 |
| Air Blank         | 0.000 | 05:40 |
| Control Test      | 0.081 | 05:41 |
| Air Blank         | 0.000 | 05:41 |
| Subject Sample #1 | 0.138 | 05:42 |
| Breath Volume     | 2.085 | 05:42 |
| Air Blank         | 0.000 | 05:42 |
| Air Blank         | 0.000 | 05:44 |
| Subject Sample #2 | 0.128 | 05:45 |
| Breath Volume     | 1.722 | 05:45 |
| Air Blank         | 0.000 | 05:45 |
| Control Test      | 0.079 | 05:46 |
| Air Blank         | 0.000 | 05:46 |
| Diagnostics Check | OK    | 05:46 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH, FELICIA, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH, FRLICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: LAINE

DATE/TIME OF TEST: 11/20/2011 05:42:05

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:33 |
| Air Blank         | 0.000 | 05:33 |
| Control Test      | 0.081 | 05:34 |
| Air Blank         | 0.000 | 05:34 |
| Subject Sample #1 | 0.158 | 05:37 |
| Breath Volume     | 1.156 | 05:37 |
| Air Blank         | 0.000 | 05:38 |
| Air Blank         | 0.000 | 05:39 |
| Subject Sample #2 | 0.162 | 05:40 |
| Breath Volume     | 1.437 | 05:40 |
| Air Blank         | 0.000 | 05:41 |
| Control Test      | 0.077 | 05:41 |
| Air Blank         | 0.000 | 05:42 |
| Diagnostics Check | OK    | 05:42 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 11/23/2011 04:44:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:41 |                      |
| Air Blank         | 0.000 | 04:41 |                      |
| Control Test      | 0.080 | 04:42 |                      |
| Air Blank         | 0.000 | 04:42 |                      |
| Subject Sample #1 |       | 04:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:43 |                      |
| Air Blank         | 0.000 | 04:43 |                      |
| Control Test      | 0.080 | 04:43 |                      |
| Air Blank         | 0.000 | 04:44 |                      |
| Diagnostics Check | OK    | 04:44 |                      |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KUNCAS, ANDREW, J

BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD

ARRESTING OFFICER: KUNCAS, ANDREW

ARRESTING OFFICER AGENCY: Bal Harbour Vil PD

LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 23:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: POZO

DATE/TIME OF TEST: 11/26/2011 00:52:47

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:49 |                      |
| Air Blank         | 0.000 | 00:50 |                      |
| Control Test      | 0.080 | 00:50 |                      |
| Air Blank         | 0.000 | 00:50 |                      |
| Subject Sample #1 |       | 00:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:51 |                      |
| Air Blank         | 0.000 | 00:51 |                      |
| Control Test      | 0.078 | 00:52 |                      |
| Air Blank         | 0.000 | 00:52 |                      |
| Diagnostics Check | OK    | 00:52 |                      |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH, FELICIA, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: DELGADO, J  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 07:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: GIORDANO

DATE/TIME OF TEST: 11/26/2011 08:30:49

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 08:23 |
| Air Blank         | 0.000 | 08:24 |
| Control Test      | 0.080 | 08:24 |
| Air Blank         | 0.000 | 08:24 |
| Subject Sample #1 | 0.209 | 08:25 |
| Breath Volume     | 1.101 | 08:25 |
| Air Blank         | 0.000 | 08:26 |
| Air Blank         | 0.000 | 08:27 |
| Subject Sample #2 | 0.196 | 08:29 |
| Breath Volume     | 1.371 | 08:29 |
| Air Blank         | 0.000 | 08:29 |
| Control Test      | 0.079 | 08:30 |
| Air Blank         | 0.000 | 08:30 |
| Diagnostics Check | OK    | 08:30 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWANDOWSKA, SABINA, M  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: LEWANDOWSKA, SABINA  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: URIBE

DATE/TIME OF TEST: 11/28/2011 02:36:52

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:30 |
| Air Blank         | 0.000 | 02:30 |
| Control Test      | 0.080 | 02:31 |
| Air Blank         | 0.000 | 02:31 |
| Subject Sample #1 | 0.369 | 02:32 |
| Breath Volume     | 3.042 | 02:32 |
| Air Blank         | 0.000 | 02:33 |
| Air Blank         | 0.000 | 02:34 |
| Subject Sample #2 | 0.365 | 02:35 |
| Breath Volume     | 3.621 | 02:35 |
| Air Blank         | 0.000 | 02:35 |
| Control Test      | 0.077 | 02:36 |
| Air Blank         | 0.000 | 02:36 |
| Diagnostics Check | OK    | 02:36 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILPART, THOMAS, B  
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD  
ARRESTING OFFICER: PHILPART, THOMAS  
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: VELA

DATE/TIME OF TEST: 11/30/2011 01:10:18

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:03 |
| Air Blank         | 0.000 | 01:03 |
| Control Test      | 0.080 | 01:03 |
| Air Blank         | 0.000 | 01:04 |
| Subject Sample #1 | 0.136 | 01:05 |
| Breath Volume     | 2.941 | 01:05 |
| Air Blank         | 0.000 | 01:06 |
| Air Blank         | 0.000 | 01:07 |
| Subject Sample #2 | 0.132 | 01:08 |
| Breath Volume     | 2.207 | 01:08 |
| Air Blank         | 0.000 | 01:09 |
| Control Test      | 0.077 | 01:09 |
| Air Blank         | 0.000 | 01:10 |
| Diagnostics Check | OK    | 01:10 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BIEDERMANN

DATE/TIME OF TEST: 11/30/2011 02:20:00

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:13 |
| Air Blank         | 0.000 | 02:13 |
| Control Test      | 0.080 | 02:13 |
| Air Blank         | 0.000 | 02:14 |
| Subject Sample #1 | 0.149 | 02:15 |
| Breath Volume     | 1.179 | 02:15 |
| Air Blank         | 0.000 | 02:16 |
| Air Blank         | 0.000 | 02:17 |
| Subject Sample #2 | 0.151 | 02:18 |
| Breath Volume     | 2.433 | 02:18 |
| Air Blank         | 0.000 | 02:19 |
| Control Test      | 0.080 | 02:19 |
| Air Blank         | 0.000 | 02:19 |
| Diagnostics Check | OK    | 02:20 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: HANDFIELD

DATE/TIME OF TEST: 11/30/2011 04:39:47

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:33 |
| Air Blank         | 0.000 | 04:33 |
| Control Test      | 0.080 | 04:33 |
| Air Blank         | 0.000 | 04:34 |
| Subject Sample #1 | 0.086 | 04:35 |
| Breath Volume     | 2.535 | 04:35 |
| Air Blank         | 0.000 | 04:35 |
| Air Blank         | 0.000 | 04:37 |
| Subject Sample #2 | 0.081 | 04:38 |
| Breath Volume     | 1.183 | 04:38 |
| Air Blank         | 0.000 | 04:38 |
| Control Test      | 0.080 | 04:39 |
| Air Blank         | 0.000 | 04:39 |
| Diagnostics Check | OK    | 04:39 |

---

Data Download Date/Time 12/04/2011 00:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, KEVIN, W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WOOLDRIDGE, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/05/2011 00:23:20

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:17 |
| Air Blank         | 0.000 | 00:17 |
| Control Test      | 0.080 | 00:18 |
| Air Blank         | 0.000 | 00:18 |
| Subject Sample #1 | 0.000 | 00:19 |
| Breath Volume     | 4.164 | 00:19 |
| Air Blank         | 0.000 | 00:19 |
| Air Blank         | 0.000 | 00:21 |
| Subject Sample #2 | 0.000 | 00:21 |
| Breath Volume     | 4.515 | 00:21 |
| Air Blank         | 0.000 | 00:22 |
| Control Test      | 0.079 | 00:22 |
| Air Blank         | 0.000 | 00:23 |
| Diagnostics Check | OK    | 00:23 |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, KEVIN, W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSTON, C  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HERTEL

DATE/TIME OF TEST: 11/05/2011 03:31:42

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:25 |
| Air Blank         | 0.000 | 03:26 |
| Control Test      | 0.080 | 03:26 |
| Air Blank         | 0.000 | 03:26 |
| Subject Sample #1 | 0.179 | 03:27 |
| Breath Volume     | 1.605 | 03:27 |
| Air Blank         | 0.000 | 03:27 |
| Air Blank         | 0.000 | 03:29 |
| Subject Sample #2 | 0.195 | 03:30 |
| Breath Volume     | 2.039 | 03:30 |
| Air Blank         | 0.000 | 03:30 |
| Control Test      | 0.080 | 03:31 |
| Air Blank         | 0.000 | 03:31 |
| Diagnostics Check | OK    | 03:31 |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BYRNE, J

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 01:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/06/2011 01:48:13

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:45 |                      |
| Air Blank         | 0.000 | 01:45 |                      |
| Control Test      | 0.080 | 01:46 |                      |
| Air Blank         | 0.000 | 01:46 |                      |
| Subject Sample #1 |       | 01:46 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:46 |                      |
| Air Blank         | 0.000 | 01:47 |                      |
| Control Test      | 0.080 | 01:47 |                      |
| Air Blank         | 0.000 | 01:48 |                      |
| Diagnostics Check | OK    | 01:48 |                      |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, KEVIN, W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISSELL, M  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: ORR

DATE/TIME OF TEST: 11/06/2011 04:11:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:09 |                      |
| Air Blank         | 0.000 | 04:09 |                      |
| Control Test      | 0.080 | 04:09 |                      |
| Air Blank         | 0.000 | 04:10 |                      |
| Subject Sample #1 |       | 04:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:10 |                      |
| Air Blank         | 0.000 | 04:11 |                      |
| Control Test      | 0.081 | 04:11 |                      |
| Air Blank         | 0.000 | 04:11 |                      |
| Diagnostics Check | OK    | 04:11 |                      |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, KEVIN, W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ODER, C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 11/07/2011 01:40:41

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 01:27 |                |
| Air Blank         | 0.000 | 01:28 |                |
| Control Test      | 0.081 | 01:28 |                |
| Air Blank         | 0.000 | 01:28 |                |
| Subject Sample #1 | 0.241 | 01:32 | Volume Not Met |
| Breath Volume     | 0.015 | 01:32 |                |
| Air Blank         | 0.000 | 01:32 |                |
| Air Blank         | 0.000 | 01:34 |                |
| Subject Sample #2 | 0.266 | 01:34 |                |
| Breath Volume     | 1.421 | 01:34 |                |
| Air Blank         | 0.000 | 01:35 |                |
| Air Blank         | 0.000 | 01:37 |                |
| Subject Sample #3 | 0.258 | 01:39 |                |
| Breath Volume     | 1.648 | 01:39 |                |
| Air Blank         | 0.000 | 01:39 |                |
| Control Test      | 0.079 | 01:40 |                |
| Air Blank         | 0.000 | 01:40 |                |
| Diagnostics Check | OK    | 01:40 |                |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HINTON, A

ARRESTING OFFICER AGENCY: Jacksonville Beach

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 22:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: OSBORN

DATE/TIME OF TEST: 11/08/2011 00:34:24

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:31 |                      |
| Air Blank         | 0.000 | 00:31 |                      |
| Control Test      | 0.079 | 00:32 |                      |
| Air Blank         | 0.000 | 00:32 |                      |
| Subject Sample #1 |       | 00:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:33 |                      |
| Air Blank         | 0.000 | 00:33 |                      |
| Control Test      | 0.079 | 00:33 |                      |
| Air Blank         | 0.000 | 00:34 |                      |
| Diagnostics Check | OK    | 00:34 |                      |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: MOSLEY, M

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 02:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 11/08/2011 04:43:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:40 |                      |
| Air Blank         | 0.000 | 04:40 |                      |
| Control Test      | 0.080 | 04:41 |                      |
| Air Blank         | 0.000 | 04:41 |                      |
| Subject Sample #1 |       | 04:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:42 |                      |
| Air Blank         | 0.000 | 04:42 |                      |
| Control Test      | 0.080 | 04:42 |                      |
| Air Blank         | 0.000 | 04:43 |                      |
| Diagnostics Check | OK    | 04:43 |                      |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, KEVIN, W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ARCHBOLD, N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: WINEGEART

DATE/TIME OF TEST: 11/08/2011 05:36:05

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:30 |
| Air Blank         | 0.000 | 05:30 |
| Control Test      | 0.080 | 05:30 |
| Air Blank         | 0.000 | 05:31 |
| Subject Sample #1 | 0.185 | 05:31 |
| Breath Volume     | 1.574 | 05:31 |
| Air Blank         | 0.000 | 05:32 |
| Air Blank         | 0.000 | 05:34 |
| Subject Sample #2 | 0.181 | 05:34 |
| Breath Volume     | 1.214 | 05:34 |
| Air Blank         | 0.000 | 05:35 |
| Control Test      | 0.079 | 05:35 |
| Air Blank         | 0.000 | 05:36 |
| Diagnostics Check | OK    | 05:36 |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: MCNAMEE, M

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 22:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 11/11/2011 00:13:45

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:05 |
| Air Blank         | 0.000 | 00:06 |
| Control Test      | 0.080 | 00:06 |
| Air Blank         | 0.000 | 00:07 |
| Subject Sample #1 | 0.000 | 00:08 |
| Breath Volume     | 1.425 | 00:08 |
| Air Blank         | 0.000 | 00:09 |
| Air Blank         | 0.000 | 00:11 |
| Subject Sample #2 | 0.000 | 00:12 |
| Breath Volume     | 1.125 | 00:12 |
| Air Blank         | 0.000 | 00:12 |
| Control Test      | 0.080 | 00:13 |
| Air Blank         | 0.000 | 00:13 |
| Diagnostics Check | OK    | 00:13 |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, KEVIN, W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSTON, C  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: KUHL

DATE/TIME OF TEST: 11/11/2011 03:21:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:19 |                      |
| Air Blank         | 0.000 | 03:19 |                      |
| Control Test      | 0.081 | 03:19 |                      |
| Air Blank         | 0.000 | 03:20 |                      |
| Subject Sample #1 |       | 03:20 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:20 |                      |
| Air Blank         | 0.000 | 03:20 |                      |
| Control Test      | 0.082 | 03:21 |                      |
| Air Blank         | 0.000 | 03:21 |                      |
| Diagnostics Check | OK    | 03:21 |                      |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, KEVIN, W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LEWIS, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 11/11/2011 05:48:20

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:42 |
| Air Blank         | 0.000 | 05:42 |
| Control Test      | 0.081 | 05:42 |
| Air Blank         | 0.000 | 05:43 |
| Subject Sample #1 | 0.136 | 05:43 |
| Breath Volume     | 2.406 | 05:43 |
| Air Blank         | 0.000 | 05:44 |
| Air Blank         | 0.000 | 05:46 |
| Subject Sample #2 | 0.130 | 05:46 |
| Breath Volume     | 1.199 | 05:46 |
| Air Blank         | 0.000 | 05:47 |
| Control Test      | 0.081 | 05:47 |
| Air Blank         | 0.000 | 05:48 |
| Diagnostics Check | OK    | 05:48 |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HEALY, GREG  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: PARVEZ

DATE/TIME OF TEST: 11/11/2011 23:35:46

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:29 |
| Air Blank         | 0.000 | 23:29 |
| Control Test      | 0.080 | 23:29 |
| Air Blank         | 0.000 | 23:30 |
| Subject Sample #1 | 0.153 | 23:31 |
| Breath Volume     | 3.593 | 23:31 |
| Air Blank         | 0.000 | 23:31 |
| Air Blank         | 0.000 | 23:33 |
| Subject Sample #2 | 0.149 | 23:34 |
| Breath Volume     | 1.949 | 23:34 |
| Air Blank         | 0.000 | 23:34 |
| Control Test      | 0.079 | 23:35 |
| Air Blank         | 0.000 | 23:35 |
| Diagnostics Check | OK    | 23:35 |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JARRELL, C E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BOWLUS

DATE/TIME OF TEST: 11/12/2011 03:00:20

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:52 |
| Air Blank         | 0.000 | 02:53 |
| Control Test      | 0.080 | 02:53 |
| Air Blank         | 0.000 | 02:54 |
| Subject Sample #1 | 0.189 | 02:55 |
| Breath Volume     | 1.792 | 02:55 |
| Air Blank         | 0.000 | 02:56 |
| Air Blank         | 0.000 | 02:57 |
| Subject Sample #2 | 0.190 | 02:58 |
| Breath Volume     | 1.996 | 02:58 |
| Air Blank         | 0.000 | 02:59 |
| Control Test      | 0.079 | 02:59 |
| Air Blank         | 0.000 | 03:00 |
| Diagnostics Check | OK    | 03:00 |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: REDFERN, L A  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 11/14/2011 04:34:16

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:27 |
| Air Blank         | 0.000 | 04:28 |
| Control Test      | 0.080 | 04:28 |
| Air Blank         | 0.000 | 04:28 |
| Subject Sample #1 | 0.138 | 04:29 |
| Breath Volume     | 2.191 | 04:29 |
| Air Blank         | 0.000 | 04:30 |
| Air Blank         | 0.000 | 04:32 |
| Subject Sample #2 | 0.145 | 04:32 |
| Breath Volume     | 3.171 | 04:32 |
| Air Blank         | 0.000 | 04:33 |
| Control Test      | 0.080 | 04:33 |
| Air Blank         | 0.000 | 04:34 |
| Diagnostics Check | OK    | 04:34 |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MCNAMEE, MORGAN  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: DUCHAUSSEE

DATE/TIME OF TEST: 11/15/2011 00:52:37

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:46 |
| Air Blank         | 0.000 | 00:46 |
| Control Test      | 0.080 | 00:46 |
| Air Blank         | 0.000 | 00:47 |
| Subject Sample #1 | 0.159 | 00:47 |
| Breath Volume     | 2.929 | 00:47 |
| Air Blank         | 0.000 | 00:48 |
| Air Blank         | 0.000 | 00:50 |
| Subject Sample #2 | 0.157 | 00:51 |
| Breath Volume     | 3.078 | 00:51 |
| Air Blank         | 0.000 | 00:51 |
| Control Test      | 0.079 | 00:52 |
| Air Blank         | 0.000 | 00:52 |
| Diagnostics Check | OK    | 00:52 |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: CHRISTMAS, F

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 23:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 11/16/2011 00:55:40

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:49 |
| Air Blank         | 0.000 | 00:49 |
| Control Test      | 0.080 | 00:50 |
| Air Blank         | 0.000 | 00:50 |
| Subject Sample #1 | 0.044 | 00:51 |
| Breath Volume     | 2.117 | 00:51 |
| Air Blank         | 0.000 | 00:51 |
| Air Blank         | 0.000 | 00:53 |
| Subject Sample #2 | 0.038 | 00:54 |
| Breath Volume     | 1.492 | 00:54 |
| Air Blank         | 0.000 | 00:54 |
| Control Test      | 0.079 | 00:55 |
| Air Blank         | 0.000 | 00:55 |
| Diagnostics Check | OK    | 00:55 |

---

Data Download Date/Time 11/16/2011 12:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BYRNE, J F  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: WHALEN

DATE/TIME OF TEST: 11/18/2011 00:57:45

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:54 |                      |
| Air Blank         | 0.000 | 00:55 |                      |
| Control Test      | 0.081 | 00:55 |                      |
| Air Blank         | 0.000 | 00:56 |                      |
| Subject Sample #1 |       | 00:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:56 |                      |
| Air Blank         | 0.000 | 00:56 |                      |
| Control Test      | 0.082 | 00:57 |                      |
| Air Blank         | 0.000 | 00:57 |                      |
| Diagnostics Check | OK    | 00:57 |                      |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: EANES, J L  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: LAY

DATE/TIME OF TEST: 11/18/2011 04:06:13

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:59 |
| Air Blank         | 0.000 | 04:00 |
| Control Test      | 0.080 | 04:00 |
| Air Blank         | 0.000 | 04:01 |
| Subject Sample #1 | 0.163 | 04:01 |
| Breath Volume     | 2.523 | 04:01 |
| Air Blank         | 0.000 | 04:02 |
| Air Blank         | 0.000 | 04:04 |
| Subject Sample #2 | 0.154 | 04:04 |
| Breath Volume     | 2.714 | 04:04 |
| Air Blank         | 0.000 | 04:05 |
| Control Test      | 0.080 | 04:05 |
| Air Blank         | 0.000 | 04:06 |
| Diagnostics Check | OK    | 04:06 |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISSEL, M J  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 11/18/2011 04:40:40

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:34 |
| Air Blank         | 0.000 | 04:34 |
| Control Test      | 0.081 | 04:35 |
| Air Blank         | 0.000 | 04:35 |
| Subject Sample #1 | 0.120 | 04:36 |
| Breath Volume     | 1.972 | 04:36 |
| Air Blank         | 0.000 | 04:36 |
| Air Blank         | 0.000 | 04:38 |
| Subject Sample #2 | 0.117 | 04:39 |
| Breath Volume     | 1.496 | 04:39 |
| Air Blank         | 0.000 | 04:39 |
| Control Test      | 0.080 | 04:40 |
| Air Blank         | 0.000 | 04:40 |
| Diagnostics Check | OK    | 04:40 |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POTTER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HEALY, GREG

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 00:31

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/19/2011 01:48:52

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:42 |
| Air Blank         | 0.000 | 01:43 |
| Control Test      | 0.079 | 01:43 |
| Air Blank         | 0.000 | 01:43 |
| Subject Sample #1 | 0.130 | 01:44 |
| Breath Volume     | 2.128 | 01:44 |
| Air Blank         | 0.000 | 01:45 |
| Air Blank         | 0.000 | 01:46 |
| Subject Sample #2 | 0.132 | 01:47 |
| Breath Volume     | 2.093 | 01:47 |
| Air Blank         | 0.000 | 01:47 |
| Control Test      | 0.079 | 01:48 |
| Air Blank         | 0.000 | 01:48 |
| Diagnostics Check | OK    | 01:48 |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BIAS, R A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 11/19/2011 04:06:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:04 |                      |
| Air Blank         | 0.000 | 04:04 |                      |
| Control Test      | 0.080 | 04:04 |                      |
| Air Blank         | 0.000 | 04:05 |                      |
| Subject Sample #1 |       | 04:05 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:05 |                      |
| Air Blank         | 0.000 | 04:06 |                      |
| Control Test      | 0.081 | 04:06 |                      |
| Air Blank         | 0.000 | 04:06 |                      |
| Diagnostics Check | OK    | 04:06 |                      |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HEALY, GREG  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SHOCKLEY

DATE/TIME OF TEST: 11/19/2011 23:43:02

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:40 |                      |
| Air Blank         | 0.000 | 23:40 |                      |
| Control Test      | 0.079 | 23:40 |                      |
| Air Blank         | 0.000 | 23:41 |                      |
| Subject Sample #1 |       | 23:41 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:41 |                      |
| Air Blank         | 0.000 | 23:42 |                      |
| Control Test      | 0.079 | 23:42 |                      |
| Air Blank         | 0.000 | 23:42 |                      |
| Diagnostics Check | OK    | 23:43 |                      |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DZIEDZICKI, A S  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ROLAND

DATE/TIME OF TEST: 11/20/2011 02:22:32

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:16 |
| Air Blank         | 0.000 | 02:16 |
| Control Test      | 0.080 | 02:16 |
| Air Blank         | 0.000 | 02:17 |
| Subject Sample #1 | 0.150 | 02:17 |
| Breath Volume     | 2.949 | 02:17 |
| Air Blank         | 0.000 | 02:18 |
| Air Blank         | 0.000 | 02:20 |
| Subject Sample #2 | 0.145 | 02:21 |
| Breath Volume     | 2.492 | 02:21 |
| Air Blank         | 0.000 | 02:21 |
| Control Test      | 0.079 | 02:21 |
| Air Blank         | 0.000 | 02:22 |
| Diagnostics Check | OK    | 02:22 |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POTTER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: ROBERTS, W A

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 02:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/20/2011 04:20:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:17 |                      |
| Air Blank         | 0.000 | 04:18 |                      |
| Control Test      | 0.080 | 04:18 |                      |
| Air Blank         | 0.000 | 04:18 |                      |
| Subject Sample #1 |       | 04:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:19 |                      |
| Air Blank         | 0.000 | 04:19 |                      |
| Control Test      | 0.080 | 04:19 |                      |
| Air Blank         | 0.000 | 04:20 |                      |
| Diagnostics Check | OK    | 04:20 |                      |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HORTON, L, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: MAYS, C

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 08:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: SEAGER

DATE/TIME OF TEST: 11/21/2011 10:14:51

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 10:05 |
| Air Blank         | 0.000 | 10:05 |
| Control Test      | 0.079 | 10:06 |
| Air Blank         | 0.000 | 10:06 |
| Subject Sample #1 | 0.101 | 10:08 |
| Breath Volume     | 2.054 | 10:08 |
| Air Blank         | 0.000 | 10:09 |
| Air Blank         | 0.000 | 10:11 |
| Subject Sample #2 | 0.105 | 10:13 |
| Breath Volume     | 2.550 | 10:13 |
| Air Blank         | 0.000 | 10:13 |
| Control Test      | 0.080 | 10:14 |
| Air Blank         | 0.000 | 10:14 |
| Diagnostics Check | OK    | 10:14 |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LESSARD, M J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 11/22/2011 01:31:34

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:28 |                      |
| Air Blank         | 0.000 | 01:28 |                      |
| Control Test      | 0.080 | 01:29 |                      |
| Air Blank         | 0.000 | 01:29 |                      |
| Subject Sample #1 |       | 01:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:30 |                      |
| Air Blank         | 0.000 | 01:30 |                      |
| Control Test      | 0.079 | 01:30 |                      |
| Air Blank         | 0.000 | 01:31 |                      |
| Diagnostics Check | OK    | 01:31 |                      |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POTTER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HAMLIN, DAN

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 23:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: WIEGMANN

DATE/TIME OF TEST: 11/23/2011 00:58:01

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:55 |                      |
| Air Blank         | 0.000 | 00:55 |                      |
| Control Test      | 0.080 | 00:55 |                      |
| Air Blank         | 0.000 | 00:56 |                      |
| Subject Sample #1 |       | 00:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:56 |                      |
| Air Blank         | 0.000 | 00:57 |                      |
| Control Test      | 0.080 | 00:57 |                      |
| Air Blank         | 0.000 | 00:57 |                      |
| Diagnostics Check | OK    | 00:58 |                      |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POTTER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: JARRELL, C E

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 00:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SAGAN

DATE/TIME OF TEST: 11/24/2011 02:18:03

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:11 |
| Air Blank         | 0.000 | 02:12 |
| Control Test      | 0.080 | 02:12 |
| Air Blank         | 0.000 | 02:13 |
| Subject Sample #1 | 0.251 | 02:13 |
| Breath Volume     | 3.910 | 02:13 |
| Air Blank         | 0.000 | 02:14 |
| Air Blank         | 0.000 | 02:16 |
| Subject Sample #2 | 0.257 | 02:16 |
| Breath Volume     | 3.707 | 02:16 |
| Air Blank         | 0.000 | 02:17 |
| Control Test      | 0.080 | 02:17 |
| Air Blank         | 0.000 | 02:17 |
| Diagnostics Check | OK    | 02:18 |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HORTON, L, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CHAPMAN, S  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 05:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: SCHEXNAYDER

DATE/TIME OF TEST: 11/24/2011 07:30:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 07:27 |                      |
| Air Blank         | 0.000 | 07:27 |                      |
| Control Test      | 0.081 | 07:28 |                      |
| Air Blank         | 0.000 | 07:28 |                      |
| Subject Sample #1 |       | 07:28 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:28 |                      |
| Air Blank         | 0.000 | 07:29 |                      |
| Control Test      | 0.081 | 07:29 |                      |
| Air Blank         | 0.000 | 07:30 |                      |
| Diagnostics Check | OK    | 07:30 |                      |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND, J, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HEALY, G  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: SAENZ

DATE/TIME OF TEST: 11/26/2011 01:55:52

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:48 |
| Air Blank         | 0.000 | 01:49 |
| Control Test      | 0.080 | 01:49 |
| Air Blank         | 0.000 | 01:50 |
| Subject Sample #1 | 0.064 | 01:51 |
| Breath Volume     | 2.746 | 01:51 |
| Air Blank         | 0.000 | 01:51 |
| Air Blank         | 0.000 | 01:53 |
| Subject Sample #2 | 0.064 | 01:54 |
| Breath Volume     | 3.183 | 01:54 |
| Air Blank         | 0.000 | 01:54 |
| Control Test      | 0.081 | 01:55 |
| Air Blank         | 0.000 | 01:55 |
| Diagnostics Check | OK    | 01:55 |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, KEVIN, W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HINTON, A  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: GRAYSON

DATE/TIME OF TEST: 11/27/2011 03:17:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:15 |                      |
| Air Blank         | 0.000 | 03:15 |                      |
| Control Test      | 0.080 | 03:15 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Subject Sample #1 |       | 03:16 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:16 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Control Test      | 0.081 | 03:17 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Diagnostics Check | OK    | 03:17 |                      |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHELETTE, C, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BORGERT, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 06:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 11/27/2011 09:28:45

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 09:22 |
| Air Blank         | 0.000 | 09:23 |
| Control Test      | 0.080 | 09:23 |
| Air Blank         | 0.000 | 09:23 |
| Subject Sample #1 | 0.057 | 09:24 |
| Breath Volume     | 2.664 | 09:24 |
| Air Blank         | 0.000 | 09:24 |
| Air Blank         | 0.000 | 09:26 |
| Subject Sample #2 | 0.053 | 09:27 |
| Breath Volume     | 2.878 | 09:27 |
| Air Blank         | 0.000 | 09:27 |
| Control Test      | 0.079 | 09:28 |
| Air Blank         | 0.000 | 09:28 |
| Diagnostics Check | OK    | 09:28 |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAND, J, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BAI, S

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 01:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: CLIFTON

DATE/TIME OF TEST: 11/28/2011 02:43:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:40 |                      |
| Air Blank         | 0.000 | 02:41 |                      |
| Control Test      | 0.080 | 02:41 |                      |
| Air Blank         | 0.000 | 02:42 |                      |
| Subject Sample #1 |       | 02:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:42 |                      |
| Air Blank         | 0.000 | 02:42 |                      |
| Control Test      | 0.080 | 02:43 |                      |
| Air Blank         | 0.000 | 02:43 |                      |
| Diagnostics Check | OK    | 02:43 |                      |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POTTER, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HOOVER, JOHN  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: CHHEM

DATE/TIME OF TEST: 11/30/2011 02:34:46

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:28 |
| Air Blank         | 0.000 | 02:29 |
| Control Test      | 0.081 | 02:29 |
| Air Blank         | 0.000 | 02:29 |
| Subject Sample #1 | 0.092 | 02:30 |
| Breath Volume     | 2.750 | 02:30 |
| Air Blank         | 0.000 | 02:30 |
| Air Blank         | 0.000 | 02:32 |
| Subject Sample #2 | 0.090 | 02:33 |
| Breath Volume     | 2.710 | 02:33 |
| Air Blank         | 0.000 | 02:33 |
| Control Test      | 0.080 | 02:34 |
| Air Blank         | 0.000 | 02:34 |
| Diagnostics Check | OK    | 02:34 |

---

Data Download Date/Time 11/30/2011 09:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BEESON, S, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: JARRELL, C

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 00:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GATZA

DATE/TIME OF TEST: 11/03/2011 01:51:41

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:45 |
| Air Blank         | 0.000 | 01:45 |
| Control Test      | 0.080 | 01:46 |
| Air Blank         | 0.000 | 01:46 |
| Subject Sample #1 | 0.109 | 01:47 |
| Breath Volume     | 4.835 | 01:47 |
| Air Blank         | 0.000 | 01:47 |
| Air Blank         | 0.000 | 01:49 |
| Subject Sample #2 | 0.114 | 01:50 |
| Breath Volume     | 4.359 | 01:50 |
| Air Blank         | 0.000 | 01:50 |
| Control Test      | 0.080 | 01:51 |
| Air Blank         | 0.000 | 01:51 |
| Diagnostics Check | OK    | 01:51 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, CHRISTOPHER, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HOOVER, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/03/2011 19:52:23

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:44 |
| Air Blank         | 0.000 | 19:45 |
| Control Test      | 0.080 | 19:45 |
| Air Blank         | 0.000 | 19:45 |
| Subject Sample #1 | 0.244 | 19:47 |
| Breath Volume     | 2.980 | 19:47 |
| Air Blank         | 0.000 | 19:48 |
| Air Blank         | 0.000 | 19:49 |
| Subject Sample #2 | 0.232 | 19:50 |
| Breath Volume     | 2.824 | 19:50 |
| Air Blank         | 0.000 | 19:51 |
| Control Test      | 0.079 | 19:51 |
| Air Blank         | 0.000 | 19:52 |
| Diagnostics Check | OK    | 19:52 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHERRY, TYNA, M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSTON, C. A.  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LILLY

DATE/TIME OF TEST: 11/04/2011 03:55:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:53 |                      |
| Air Blank         | 0.000 | 03:53 |                      |
| Control Test      | 0.081 | 03:53 |                      |
| Air Blank         | 0.000 | 03:54 |                      |
| Subject Sample #1 |       | 03:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:54 |                      |
| Air Blank         | 0.000 | 03:55 |                      |
| Control Test      | 0.080 | 03:55 |                      |
| Air Blank         | 0.000 | 03:55 |                      |
| Diagnostics Check | OK    | 03:55 |                      |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHERRY, TYNA, M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JARRELL, C. E.  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/05/2011 03:08:03

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:05 |                      |
| Air Blank         | 0.000 | 03:05 |                      |
| Control Test      | 0.081 | 03:05 |                      |
| Air Blank         | 0.000 | 03:06 |                      |
| Subject Sample #1 |       | 03:06 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:06 |                      |
| Air Blank         | 0.000 | 03:07 |                      |
| Control Test      | 0.081 | 03:07 |                      |
| Air Blank         | 0.000 | 03:07 |                      |
| Diagnostics Check | OK    | 03:08 |                      |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BEESON, S, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: COLEMAN, F

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 01:48

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 11/09/2011 04:01:59

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:55 |
| Air Blank         | 0.000 | 03:55 |
| Control Test      | 0.080 | 03:56 |
| Air Blank         | 0.000 | 03:56 |
| Subject Sample #1 | 0.124 | 03:57 |
| Breath Volume     | 1.375 | 03:57 |
| Air Blank         | 0.000 | 03:57 |
| Air Blank         | 0.000 | 03:59 |
| Subject Sample #2 | 0.127 | 04:00 |
| Breath Volume     | 2.074 | 04:00 |
| Air Blank         | 0.000 | 04:01 |
| Control Test      | 0.079 | 04:01 |
| Air Blank         | 0.000 | 04:01 |
| Diagnostics Check | OK    | 04:01 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FIELDS, WILLIE, M

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: MATTHEWS, E

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 15:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 11/10/2011 17:36:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:33 |                      |
| Air Blank         | 0.000 | 17:34 |                      |
| Control Test      | 0.080 | 17:34 |                      |
| Air Blank         | 0.000 | 17:35 |                      |
| Subject Sample #1 |       | 17:35 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:35 |                      |
| Air Blank         | 0.000 | 17:35 |                      |
| Control Test      | 0.081 | 17:36 |                      |
| Air Blank         | 0.000 | 17:36 |                      |
| Diagnostics Check | OK    | 17:36 |                      |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIELDS, WILLIE, M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HOOVER, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 11/10/2011 20:48:20

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 20:41 |
| Air Blank         | 0.000 | 20:42 |
| Control Test      | 0.081 | 20:42 |
| Air Blank         | 0.000 | 20:43 |
| Subject Sample #1 | 0.106 | 20:43 |
| Breath Volume     | 2.015 | 20:43 |
| Air Blank         | 0.000 | 20:44 |
| Air Blank         | 0.000 | 20:46 |
| Subject Sample #2 | 0.115 | 20:46 |
| Breath Volume     | 1.992 | 20:46 |
| Air Blank         | 0.000 | 20:47 |
| Control Test      | 0.081 | 20:47 |
| Air Blank         | 0.000 | 20:48 |
| Diagnostics Check | OK    | 20:48 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JARRELL, C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MCGRIFF

DATE/TIME OF TEST: 11/11/2011 01:01:31

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:55 |
| Air Blank         | 0.000 | 00:55 |
| Control Test      | 0.081 | 00:56 |
| Air Blank         | 0.000 | 00:56 |
| Subject Sample #1 | 0.155 | 00:57 |
| Breath Volume     | 3.234 | 00:57 |
| Air Blank         | 0.000 | 00:57 |
| Air Blank         | 0.000 | 00:59 |
| Subject Sample #2 | 0.147 | 00:59 |
| Breath Volume     | 1.859 | 00:59 |
| Air Blank         | 0.000 | 01:00 |
| Control Test      | 0.080 | 01:00 |
| Air Blank         | 0.000 | 01:01 |
| Diagnostics Check | OK    | 01:01 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BARTILUCCI, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 11/11/2011 03:59:33

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:53 |
| Air Blank         | 0.000 | 03:53 |
| Control Test      | 0.081 | 03:53 |
| Air Blank         | 0.000 | 03:54 |
| Subject Sample #1 | 0.184 | 03:54 |
| Breath Volume     | 3.886 | 03:54 |
| Air Blank         | 0.000 | 03:55 |
| Air Blank         | 0.000 | 03:57 |
| Subject Sample #2 | 0.186 | 03:57 |
| Breath Volume     | 4.335 | 03:57 |
| Air Blank         | 0.000 | 03:58 |
| Control Test      | 0.080 | 03:58 |
| Air Blank         | 0.000 | 03:59 |
| Diagnostics Check | OK    | 03:59 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JENSEN, R, H

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: MCNAMEE, M

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 21:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 11/11/2011 22:45:04

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:33 |
| Air Blank         | 0.000 | 22:34 |
| Control Test      | 0.082 | 22:34 |
| Air Blank         | 0.000 | 22:34 |
| Subject Sample #1 | 0.094 | 22:37 |
| Breath Volume     | 1.273 | 22:37 |
| Air Blank         | 0.000 | 22:38 |
| Air Blank         | 0.000 | 22:40 |
| Subject Sample #2 | 0.094 | 22:43 |
| Breath Volume     | 1.851 | 22:43 |
| Air Blank         | 0.000 | 22:44 |
| Control Test      | 0.081 | 22:44 |
| Air Blank         | 0.000 | 22:44 |
| Diagnostics Check | OK    | 22:45 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: S, EZELL  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: HOLLIDAY

DATE/TIME OF TEST: 11/12/2011 01:50:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:47 |                      |
| Air Blank         | 0.000 | 01:48 |                      |
| Control Test      | 0.081 | 01:48 |                      |
| Air Blank         | 0.000 | 01:48 |                      |
| Subject Sample #1 |       | 01:49 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:49 |                      |
| Air Blank         | 0.000 | 01:49 |                      |
| Control Test      | 0.081 | 01:49 |                      |
| Air Blank         | 0.000 | 01:50 |                      |
| Diagnostics Check | OK    | 01:50 |                      |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HAMLIN, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BURMEISTER

DATE/TIME OF TEST: 11/12/2011 02:59:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:56 |                      |
| Air Blank         | 0.000 | 02:56 |                      |
| Control Test      | 0.081 | 02:56 |                      |
| Air Blank         | 0.000 | 02:57 |                      |
| Subject Sample #1 |       | 02:57 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:57 |                      |
| Air Blank         | 0.000 | 02:58 |                      |
| Control Test      | 0.081 | 02:58 |                      |
| Air Blank         | 0.000 | 02:59 |                      |
| Diagnostics Check | OK    | 02:59 |                      |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSTON, C  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: SCHROEDER

DATE/TIME OF TEST: 11/13/2011 01:11:09

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:08 |                      |
| Air Blank         | 0.000 | 01:08 |                      |
| Control Test      | 0.081 | 01:09 |                      |
| Air Blank         | 0.000 | 01:09 |                      |
| Subject Sample #1 |       | 01:09 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:09 |                      |
| Air Blank         | 0.000 | 01:10 |                      |
| Control Test      | 0.081 | 01:10 |                      |
| Air Blank         | 0.000 | 01:11 |                      |
| Diagnostics Check | OK    | 01:11 |                      |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BEESON, S, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: ROBERTS, W

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 00:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 11/13/2011 02:21:04

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:17 |                      |
| Air Blank         | 0.000 | 02:18 |                      |
| Control Test      | 0.081 | 02:18 |                      |
| Air Blank         | 0.000 | 02:18 |                      |
| Subject Sample #1 |       | 02:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:19 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Control Test      | 0.080 | 02:20 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Diagnostics Check | OK    | 02:21 |                      |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSTON, C  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MEST

DATE/TIME OF TEST: 11/13/2011 04:48:37

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:45 |                      |
| Air Blank         | 0.000 | 04:46 |                      |
| Control Test      | 0.081 | 04:46 |                      |
| Air Blank         | 0.000 | 04:47 |                      |
| Subject Sample #1 |       | 04:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:47 |                      |
| Air Blank         | 0.000 | 04:47 |                      |
| Control Test      | 0.080 | 04:48 |                      |
| Air Blank         | 0.000 | 04:48 |                      |
| Diagnostics Check | OK    | 04:48 |                      |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: REDGATE, E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: GUARA

DATE/TIME OF TEST: 11/13/2011 05:55:44

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:48 |
| Air Blank         | 0.000 | 05:49 |
| Control Test      | 0.081 | 05:49 |
| Air Blank         | 0.000 | 05:50 |
| Subject Sample #1 | 0.132 | 05:50 |
| Breath Volume     | 3.042 | 05:50 |
| Air Blank         | 0.000 | 05:51 |
| Air Blank         | 0.000 | 05:53 |
| Subject Sample #2 | 0.138 | 05:54 |
| Breath Volume     | 3.800 | 05:54 |
| Air Blank         | 0.000 | 05:54 |
| Control Test      | 0.081 | 05:55 |
| Air Blank         | 0.000 | 05:55 |
| Diagnostics Check | OK    | 05:55 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUCHANAN, J,

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: DOYLE, B

ARRESTING OFFICER AGENCY: Jacksonville Beach

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 19:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SPIVEY

DATE/TIME OF TEST: 11/13/2011 22:32:57

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:26 |
| Air Blank         | 0.000 | 22:27 |
| Control Test      | 0.080 | 22:27 |
| Air Blank         | 0.000 | 22:27 |
| Subject Sample #1 | 0.180 | 22:28 |
| Breath Volume     | 1.500 | 22:28 |
| Air Blank         | 0.000 | 22:28 |
| Air Blank         | 0.000 | 22:30 |
| Subject Sample #2 | 0.196 | 22:31 |
| Breath Volume     | 1.898 | 22:31 |
| Air Blank         | 0.000 | 22:32 |
| Control Test      | 0.079 | 22:32 |
| Air Blank         | 0.000 | 22:32 |
| Diagnostics Check | OK    | 22:32 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLISS, F, C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HEALY, G  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 11/14/2011 22:39:51

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:28 |                    |
| Air Blank         | 0.000 | 22:28 |                    |
| Control Test      | 0.080 | 22:29 |                    |
| Air Blank         | 0.000 | 22:29 |                    |
| Subject Sample #1 |       | 22:32 | No Sample Provided |
| Breath Volume     | 0.000 | 22:32 |                    |
| Air Blank         | 0.000 | 22:33 |                    |
| Air Blank         | 0.000 | 22:35 |                    |
| Subject Sample #2 |       | 22:38 | No Sample Provided |
| Breath Volume     | 0.000 | 22:38 |                    |
| Air Blank         | 0.000 | 22:38 |                    |
| Control Test      | 0.080 | 22:39 |                    |
| Air Blank         | 0.000 | 22:39 |                    |
| Diagnostics Check | OK    | 22:39 |                    |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLISS, F, C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HEALY, G  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 11/14/2011 22:51:00

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:48 |                      |
| Air Blank         | 0.000 | 22:48 |                      |
| Control Test      | 0.081 | 22:48 |                      |
| Air Blank         | 0.000 | 22:49 |                      |
| Subject Sample #1 |       | 22:49 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:49 |                      |
| Air Blank         | 0.000 | 22:50 |                      |
| Control Test      | 0.080 | 22:50 |                      |
| Air Blank         | 0.000 | 22:50 |                      |
| Diagnostics Check | OK    | 22:51 |                      |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FIELDS, WILLIE, M

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: MARRERO, J

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 19:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: BEYER

DATE/TIME OF TEST: 11/15/2011 22:19:49

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:13 |
| Air Blank         | 0.000 | 22:13 |
| Control Test      | 0.081 | 22:14 |
| Air Blank         | 0.000 | 22:14 |
| Subject Sample #1 | 0.000 | 22:15 |
| Breath Volume     | 2.097 | 22:15 |
| Air Blank         | 0.000 | 22:15 |
| Air Blank         | 0.000 | 22:17 |
| Subject Sample #2 | 0.000 | 22:18 |
| Breath Volume     | 2.925 | 22:18 |
| Air Blank         | 0.000 | 22:18 |
| Control Test      | 0.081 | 22:19 |
| Air Blank         | 0.000 | 22:19 |
| Diagnostics Check | OK    | 22:19 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONACO, SCOTT,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: COLEMAN, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: FICHTER

DATE/TIME OF TEST: 11/16/2011 05:15:23

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:08 |
| Air Blank         | 0.000 | 05:08 |
| Control Test      | 0.080 | 05:08 |
| Air Blank         | 0.000 | 05:09 |
| Subject Sample #1 | 0.071 | 05:10 |
| Breath Volume     | 2.785 | 05:10 |
| Air Blank         | 0.000 | 05:10 |
| Air Blank         | 0.000 | 05:12 |
| Subject Sample #2 | 0.067 | 05:13 |
| Breath Volume     | 1.773 | 05:13 |
| Air Blank         | 0.000 | 05:14 |
| Control Test      | 0.080 | 05:14 |
| Air Blank         | 0.000 | 05:15 |
| Diagnostics Check | OK    | 05:15 |

---

Data Download Date/Time 11/16/2011 12:34

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ARCHBOLD, N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 11/18/2011 03:50:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:48 |                      |
| Air Blank         | 0.000 | 03:48 |                      |
| Control Test      | 0.080 | 03:48 |                      |
| Air Blank         | 0.000 | 03:49 |                      |
| Subject Sample #1 |       | 03:49 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:49 |                      |
| Air Blank         | 0.000 | 03:49 |                      |
| Control Test      | 0.081 | 03:50 |                      |
| Air Blank         | 0.000 | 03:50 |                      |
| Diagnostics Check | OK    | 03:50 |                      |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISSEL, M  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: KESSLER

DATE/TIME OF TEST: 11/18/2011 04:26:52

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:19 |
| Air Blank         | 0.000 | 04:20 |
| Control Test      | 0.081 | 04:20 |
| Air Blank         | 0.000 | 04:21 |
| Subject Sample #1 | 0.163 | 04:21 |
| Breath Volume     | 4.140 | 04:21 |
| Air Blank         | 0.000 | 04:22 |
| Air Blank         | 0.000 | 04:24 |
| Subject Sample #2 | 0.165 | 04:25 |
| Breath Volume     | 4.257 | 04:25 |
| Air Blank         | 0.000 | 04:25 |
| Control Test      | 0.082 | 04:26 |
| Air Blank         | 0.000 | 04:26 |
| Diagnostics Check | OK    | 04:26 |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FIELDS, WILLIE, M

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: REINERT, B

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 19:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HAGOOD

DATE/TIME OF TEST: 11/18/2011 21:43:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:41 |                      |
| Air Blank         | 0.000 | 21:41 |                      |
| Control Test      | 0.080 | 21:41 |                      |
| Air Blank         | 0.000 | 21:42 |                      |
| Subject Sample #1 |       | 21:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:42 |                      |
| Air Blank         | 0.000 | 21:42 |                      |
| Control Test      | 0.080 | 21:43 |                      |
| Air Blank         | 0.000 | 21:43 |                      |
| Diagnostics Check | OK    | 21:43 |                      |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HEALY, G  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: GROOMS

DATE/TIME OF TEST: 11/18/2011 23:44:58

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:38 |
| Air Blank         | 0.000 | 23:39 |
| Control Test      | 0.080 | 23:39 |
| Air Blank         | 0.000 | 23:40 |
| Subject Sample #1 | 0.108 | 23:40 |
| Breath Volume     | 2.011 | 23:40 |
| Air Blank         | 0.000 | 23:41 |
| Air Blank         | 0.000 | 23:42 |
| Subject Sample #2 | 0.110 | 23:43 |
| Breath Volume     | 1.898 | 23:43 |
| Air Blank         | 0.000 | 23:44 |
| Control Test      | 0.080 | 23:44 |
| Air Blank         | 0.000 | 23:44 |
| Diagnostics Check | OK    | 23:44 |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HOOVER, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 11/19/2011 02:11:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:09 |                      |
| Air Blank         | 0.000 | 02:09 |                      |
| Control Test      | 0.080 | 02:09 |                      |
| Air Blank         | 0.000 | 02:10 |                      |
| Subject Sample #1 |       | 02:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:10 |                      |
| Air Blank         | 0.000 | 02:11 |                      |
| Control Test      | 0.081 | 02:11 |                      |
| Air Blank         | 0.000 | 02:11 |                      |
| Diagnostics Check | OK    | 02:11 |                      |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITAKER, L  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 11/19/2011 05:18:49

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:16 |                      |
| Air Blank         | 0.000 | 05:16 |                      |
| Control Test      | 0.080 | 05:16 |                      |
| Air Blank         | 0.000 | 05:17 |                      |
| Subject Sample #1 |       | 05:17 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:17 |                      |
| Air Blank         | 0.000 | 05:17 |                      |
| Control Test      | 0.080 | 05:18 |                      |
| Air Blank         | 0.000 | 05:18 |                      |
| Diagnostics Check | OK    | 05:18 |                      |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAY, K, J

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: MCNAMEE, M

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 17:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: MAHONEY

DATE/TIME OF TEST: 11/19/2011 19:11:02

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:04 |
| Air Blank         | 0.000 | 19:04 |
| Control Test      | 0.081 | 19:04 |
| Air Blank         | 0.000 | 19:05 |
| Subject Sample #1 | 0.218 | 19:06 |
| Breath Volume     | 1.621 | 19:06 |
| Air Blank         | 0.000 | 19:07 |
| Air Blank         | 0.000 | 19:08 |
| Subject Sample #2 | 0.212 | 19:09 |
| Breath Volume     | 2.304 | 19:09 |
| Air Blank         | 0.000 | 19:10 |
| Control Test      | 0.081 | 19:10 |
| Air Blank         | 0.000 | 19:10 |
| Diagnostics Check | OK    | 19:11 |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAY, K, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HOOVER, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 11/19/2011 22:25:30

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:18 |
| Air Blank         | 0.000 | 22:19 |
| Control Test      | 0.080 | 22:19 |
| Air Blank         | 0.000 | 22:20 |
| Subject Sample #1 | 0.205 | 22:20 |
| Breath Volume     | 1.285 | 22:20 |
| Air Blank         | 0.000 | 22:21 |
| Air Blank         | 0.000 | 22:23 |
| Subject Sample #2 | 0.200 | 22:23 |
| Breath Volume     | 1.441 | 22:23 |
| Air Blank         | 0.000 | 22:24 |
| Control Test      | 0.079 | 22:24 |
| Air Blank         | 0.000 | 22:25 |
| Diagnostics Check | OK    | 22:25 |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MACK-JACKSON, S  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SKIRM

DATE/TIME OF TEST: 11/21/2011 02:51:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:48 |                      |
| Air Blank         | 0.000 | 02:49 |                      |
| Control Test      | 0.080 | 02:49 |                      |
| Air Blank         | 0.000 | 02:50 |                      |
| Subject Sample #1 |       | 02:50 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:50 |                      |
| Air Blank         | 0.000 | 02:50 |                      |
| Control Test      | 0.080 | 02:51 |                      |
| Air Blank         | 0.000 | 02:51 |                      |
| Diagnostics Check | OK    | 02:51 |                      |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARRIS, C, E

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: WILLIAMS, D

ARRESTING OFFICER AGENCY: Atlantic Beach PD

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 15:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: SABOL

DATE/TIME OF TEST: 11/22/2011 17:32:26

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 17:25 |            |
| Air Blank         | 0.000 | 17:25 |            |
| Control Test      | 0.080 | 17:26 |            |
| Air Blank         | 0.000 | 17:26 |            |
| Subject Sample #1 | 0.147 | 17:28 |            |
| Breath Volume     | 1.718 | 17:28 |            |
| Air Blank         | 0.000 | 17:29 |            |
| Air Blank         | 0.000 | 17:30 |            |
| Subject Sample #2 | 0.138 | 17:31 |            |
| Breath Volume     | 1.605 | 17:31 |            |
| Air Blank         |       | 17:31 | RFI Detect |
| Air Blank         | 0.000 | 17:32 |            |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARRIS, C, E

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: WILLIAMS, D

ARRESTING OFFICER AGENCY: Atlantic Beach PD

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 15:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: SABOL

DATE/TIME OF TEST: 11/22/2011 17:43:28

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 17:35 |
| Air Blank         | 0.000 | 17:36 |
| Control Test      | 0.079 | 17:36 |
| Air Blank         | 0.000 | 17:37 |
| Subject Sample #1 | 0.127 | 17:37 |
| Breath Volume     | 1.507 | 17:37 |
| Air Blank         | 0.000 | 17:38 |
| Air Blank         | 0.000 | 17:40 |
| Subject Sample #2 | 0.136 | 17:41 |
| Breath Volume     | 1.386 | 17:41 |
| Air Blank         | 0.000 | 17:42 |
| Control Test      | 0.079 | 17:42 |
| Air Blank         | 0.000 | 17:43 |
| Diagnostics Check | OK    | 17:43 |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FIELDS, WILLIE, M

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: THOMASON, PATRICK

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 19:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 11/22/2011 21:30:26

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 21:23 |
| Air Blank         | 0.000 | 21:24 |
| Control Test      | 0.080 | 21:24 |
| Air Blank         | 0.000 | 21:24 |
| Subject Sample #1 | 0.090 | 21:25 |
| Breath Volume     | 1.902 | 21:25 |
| Air Blank         | 0.000 | 21:26 |
| Air Blank         | 0.000 | 21:28 |
| Subject Sample #2 | 0.096 | 21:28 |
| Breath Volume     | 2.000 | 21:28 |
| Air Blank         | 0.000 | 21:29 |
| Control Test      | 0.079 | 21:29 |
| Air Blank         | 0.000 | 21:30 |
| Diagnostics Check | OK    | 21:30 |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIELDS, WILLIE, M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MCNAMEE, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: MORANTE

DATE/TIME OF TEST: 11/23/2011 22:31:45

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:25 |
| Air Blank         | 0.000 | 22:25 |
| Control Test      | 0.080 | 22:26 |
| Air Blank         | 0.000 | 22:26 |
| Subject Sample #1 | 0.000 | 22:27 |
| Breath Volume     | 1.945 | 22:27 |
| Air Blank         | 0.000 | 22:27 |
| Air Blank         | 0.000 | 22:29 |
| Subject Sample #2 | 0.000 | 22:30 |
| Breath Volume     | 2.078 | 22:30 |
| Air Blank         | 0.000 | 22:30 |
| Control Test      | 0.080 | 22:31 |
| Air Blank         | 0.000 | 22:31 |
| Diagnostics Check | OK    | 22:31 |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEESON, S, D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WOOLDRIDGE, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/24/2011 03:37:38

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:31 |
| Air Blank         | 0.000 | 03:31 |
| Control Test      | 0.080 | 03:32 |
| Air Blank         | 0.000 | 03:32 |
| Subject Sample #1 | 0.216 | 03:33 |
| Breath Volume     | 1.988 | 03:33 |
| Air Blank         | 0.000 | 03:33 |
| Air Blank         | 0.000 | 03:35 |
| Subject Sample #2 | 0.211 | 03:36 |
| Breath Volume     | 2.007 | 03:36 |
| Air Blank         | 0.000 | 03:36 |
| Control Test      | 0.080 | 03:37 |
| Air Blank         | 0.000 | 03:37 |
| Diagnostics Check | OK    | 03:37 |

---

Data Download Date/Time 11/30/2011 09:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAY, K, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO, E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/30/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: PROUDFIT

DATE/TIME OF TEST: 11/30/2011 22:52:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:49 |                      |
| Air Blank         | 0.000 | 22:50 |                      |
| Control Test      | 0.081 | 22:50 |                      |
| Air Blank         | 0.000 | 22:51 |                      |
| Subject Sample #1 |       | 22:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:51 |                      |
| Air Blank         | 0.000 | 22:51 |                      |
| Control Test      | 0.080 | 22:52 |                      |
| Air Blank         | 0.000 | 22:52 |                      |
| Diagnostics Check | OK    | 22:52 |                      |

---

Data Download Date/Time 12/15/2011 08:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLIVER, J, P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JARRELL, C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: ASWEGAN

DATE/TIME OF TEST: 11/02/2011 00:25:18

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:19 |
| Air Blank         | 0.000 | 00:19 |
| Control Test      | 0.078 | 00:20 |
| Air Blank         | 0.000 | 00:20 |
| Subject Sample #1 | 0.160 | 00:20 |
| Breath Volume     | 2.066 | 00:20 |
| Air Blank         | 0.000 | 00:21 |
| Air Blank         | 0.000 | 00:23 |
| Subject Sample #2 | 0.158 | 00:23 |
| Breath Volume     | 1.726 | 00:23 |
| Air Blank         | 0.000 | 00:24 |
| Control Test      | 0.078 | 00:24 |
| Air Blank         | 0.000 | 00:25 |
| Diagnostics Check | OK    | 00:25 |

---

Data Download Date/Time 11/16/2011 12:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: OLIVER, J, P

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HOOVER, J

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 23:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: SWAINE

DATE/TIME OF TEST: 11/03/2011 00:29:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:26 |                      |
| Air Blank         | 0.000 | 00:27 |                      |
| Control Test      | 0.078 | 00:27 |                      |
| Air Blank         | 0.000 | 00:27 |                      |
| Subject Sample #1 |       | 00:28 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:28 |                      |
| Air Blank         | 0.000 | 00:29 |                      |
| Control Test      | 0.077 | 00:29 |                      |
| Air Blank         | 0.000 | 00:29 |                      |
| Diagnostics Check | OK    | 00:29 |                      |

---

Data Download Date/Time 11/16/2011 12:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLIVER, J, P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WOOLDRIDGE, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 11/03/2011 02:47:55

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:41 |
| Air Blank         | 0.000 | 02:42 |
| Control Test      | 0.078 | 02:42 |
| Air Blank         | 0.000 | 02:42 |
| Subject Sample #1 | 0.190 | 02:43 |
| Breath Volume     | 3.527 | 02:43 |
| Air Blank         | 0.000 | 02:43 |
| Air Blank         | 0.000 | 02:45 |
| Subject Sample #2 | 0.181 | 02:46 |
| Breath Volume     | 3.285 | 02:46 |
| Air Blank         | 0.000 | 02:47 |
| Control Test      | 0.078 | 02:47 |
| Air Blank         | 0.000 | 02:47 |
| Diagnostics Check | OK    | 02:47 |

---

Data Download Date/Time 11/16/2011 12:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTOFFERSEN,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WOOLDRIDGE, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: WARNOCK

DATE/TIME OF TEST: 11/11/2011 23:37:37

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:30 |
| Air Blank         | 0.000 | 23:30 |
| Control Test      | 0.079 | 23:31 |
| Air Blank         | 0.000 | 23:31 |
| Subject Sample #1 | 0.000 | 23:32 |
| Breath Volume     | 2.316 | 23:32 |
| Air Blank         | 0.000 | 23:33 |
| Air Blank         | 0.000 | 23:35 |
| Subject Sample #2 | 0.000 | 23:36 |
| Breath Volume     | 2.371 | 23:36 |
| Air Blank         | 0.000 | 23:36 |
| Control Test      | 0.079 | 23:37 |
| Air Blank         | 0.000 | 23:37 |
| Diagnostics Check | OK    | 23:37 |

---

Data Download Date/Time 11/16/2011 12:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTOFFERSEN, C, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: INGLE, M  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 11/12/2011 02:36:09

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:29 |
| Air Blank         | 0.000 | 02:30 |
| Control Test      | 0.078 | 02:30 |
| Air Blank         | 0.000 | 02:30 |
| Subject Sample #1 | 0.142 | 02:31 |
| Breath Volume     | 2.195 | 02:31 |
| Air Blank         | 0.000 | 02:32 |
| Air Blank         | 0.000 | 02:33 |
| Subject Sample #2 | 0.140 | 02:34 |
| Breath Volume     | 1.679 | 02:34 |
| Air Blank         | 0.000 | 02:35 |
| Control Test      | 0.078 | 02:35 |
| Air Blank         | 0.000 | 02:36 |
| Diagnostics Check | OK    | 02:36 |

---

Data Download Date/Time 11/16/2011 12:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTOFFERSEN, C, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HOOVER, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 11/12/2011 03:52:10

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:44 |
| Air Blank         | 0.000 | 03:44 |
| Control Test      | 0.079 | 03:45 |
| Air Blank         | 0.000 | 03:45 |
| Subject Sample #1 | 0.136 | 03:47 |
| Breath Volume     | 2.089 | 03:47 |
| Air Blank         | 0.000 | 03:48 |
| Air Blank         | 0.000 | 03:50 |
| Subject Sample #2 | 0.134 | 03:50 |
| Breath Volume     | 2.070 | 03:50 |
| Air Blank         | 0.000 | 03:51 |
| Control Test      | 0.077 | 03:51 |
| Air Blank         | 0.000 | 03:52 |
| Diagnostics Check | OK    | 03:52 |

---

Data Download Date/Time 11/16/2011 12:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTOFFERSEN, C, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BLALOCK, R  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23411080A2  
DRY GAS EXPIRY DATE: 10/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 19:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 11/18/2011 00:04:37

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:58 |
| Air Blank         | 0.000 | 23:58 |
| Control Test      | 0.079 | 23:58 |
| Air Blank         | 0.000 | 23:59 |
| Subject Sample #1 | 0.000 | 00:00 |
| Breath Volume     | 1.746 | 00:00 |
| Air Blank         | 0.000 | 00:00 |
| Air Blank         | 0.000 | 00:02 |
| Subject Sample #2 | 0.000 | 00:03 |
| Breath Volume     | 1.785 | 00:03 |
| Air Blank         | 0.000 | 00:03 |
| Control Test      | 0.078 | 00:04 |
| Air Blank         | 0.000 | 00:04 |
| Diagnostics Check | OK    | 00:04 |

---

Data Download Date/Time 11/30/2011 09:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTOFFERSEN, C, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: EZELL, S  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23411080A2  
DRY GAS EXPIRY DATE: 10/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: FLIPSE

DATE/TIME OF TEST: 11/18/2011 03:57:21

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 03:42 |                |
| Air Blank         | 0.000 | 03:43 |                |
| Control Test      | 0.079 | 03:43 |                |
| Air Blank         | 0.000 | 03:43 |                |
| Subject Sample #1 | 0.105 | 03:47 | Volume Not Met |
| Breath Volume     | 0.027 | 03:47 |                |
| Air Blank         | 0.000 | 03:47 |                |
| Air Blank         | 0.000 | 03:49 |                |
| Subject Sample #2 | 0.111 | 03:51 |                |
| Breath Volume     | 1.820 | 03:51 |                |
| Air Blank         | 0.000 | 03:52 |                |
| Air Blank         | 0.000 | 03:54 |                |
| Subject Sample #3 | 0.106 | 03:55 |                |
| Breath Volume     | 1.843 | 03:55 |                |
| Air Blank         | 0.000 | 03:56 |                |
| Control Test      | 0.078 | 03:56 |                |
| Air Blank         | 0.000 | 03:57 |                |
| Diagnostics Check | OK    | 03:57 |                |

---

Data Download Date/Time 11/30/2011 09:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTOFFERSEN, C, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HINTON, A  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23411080A2  
DRY GAS EXPIRY DATE: 10/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: DE MORAES

DATE/TIME OF TEST: 11/18/2011 04:52:16

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:45 |
| Air Blank         | 0.000 | 04:45 |
| Control Test      | 0.079 | 04:46 |
| Air Blank         | 0.000 | 04:46 |
| Subject Sample #1 | 0.116 | 04:47 |
| Breath Volume     | 2.398 | 04:47 |
| Air Blank         | 0.000 | 04:48 |
| Air Blank         | 0.000 | 04:50 |
| Subject Sample #2 | 0.108 | 04:50 |
| Breath Volume     | 2.347 | 04:50 |
| Air Blank         | 0.000 | 04:51 |
| Control Test      | 0.078 | 04:51 |
| Air Blank         | 0.000 | 04:52 |
| Diagnostics Check | OK    | 04:52 |

---

Data Download Date/Time 11/30/2011 09:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NICHOLS, A, B

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HINTON, A

ARRESTING OFFICER AGENCY: Atlantic Beach PD

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23411080A2

DRY GAS EXPIRY DATE: 10/01/2013

VIOLATION CODE: DUI

ARREST TIME: 22:31

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: DEBUTY

DATE/TIME OF TEST: 11/19/2011 02:13:08

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:06 |
| Air Blank         | 0.000 | 02:06 |
| Control Test      | 0.078 | 02:07 |
| Air Blank         | 0.000 | 02:07 |
| Subject Sample #1 | 0.260 | 02:08 |
| Breath Volume     | 2.417 | 02:08 |
| Air Blank         | 0.000 | 02:08 |
| Air Blank         | 0.000 | 02:10 |
| Subject Sample #2 | 0.262 | 02:11 |
| Breath Volume     | 2.593 | 02:11 |
| Air Blank         | 0.000 | 02:12 |
| Control Test      | 0.077 | 02:12 |
| Air Blank         | 0.000 | 02:13 |
| Diagnostics Check | OK    | 02:13 |

---

Data Download Date/Time 11/30/2011 09:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICHOLS, A, B  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BYRNE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23411080A2  
DRY GAS EXPIRY DATE: 10/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 11/19/2011 05:16:16

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:09 |
| Air Blank         | 0.000 | 05:10 |
| Control Test      | 0.078 | 05:10 |
| Air Blank         | 0.000 | 05:10 |
| Subject Sample #1 | 0.219 | 05:11 |
| Breath Volume     | 1.843 | 05:11 |
| Air Blank         | 0.000 | 05:11 |
| Air Blank         | 0.000 | 05:13 |
| Subject Sample #2 | 0.205 | 05:14 |
| Breath Volume     | 1.886 | 05:14 |
| Air Blank         | 0.000 | 05:15 |
| Control Test      | 0.077 | 05:15 |
| Air Blank         | 0.000 | 05:16 |
| Diagnostics Check | OK    | 05:16 |

---

Data Download Date/Time 11/30/2011 09:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HALLIDAY, R, A

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: ARCHBOLD, N

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23411080A2

DRY GAS EXPIRY DATE: 10/01/2013

VIOLATION CODE: DUI

ARREST TIME: 03:02

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LOMBARDI

DATE/TIME OF TEST: 11/26/2011 05:22:55

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:16 |
| Air Blank         | 0.000 | 05:16 |
| Control Test      | 0.078 | 05:16 |
| Air Blank         | 0.000 | 05:17 |
| Subject Sample #1 | 0.133 | 05:18 |
| Breath Volume     | 1.691 | 05:18 |
| Air Blank         | 0.000 | 05:18 |
| Air Blank         | 0.000 | 05:20 |
| Subject Sample #2 | 0.119 | 05:21 |
| Breath Volume     | 1.382 | 05:21 |
| Air Blank         | 0.000 | 05:22 |
| Control Test      | 0.076 | 05:22 |
| Air Blank         | 0.000 | 05:22 |
| Diagnostics Check | OK    | 05:22 |

---

Data Download Date/Time 11/30/2011 09:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD, MELISSA, K  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: CARTER, J  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 11/09/2011 20:56:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:53 |                      |
| Air Blank         | 0.000 | 20:53 |                      |
| Control Test      | 0.081 | 20:54 |                      |
| Air Blank         | 0.000 | 20:54 |                      |
| Subject Sample #1 |       | 20:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:55 |                      |
| Air Blank         | 0.000 | 20:55 |                      |
| Control Test      | 0.081 | 20:55 | Interferent Detect   |
| Air Blank         | 0.000 | 20:56 |                      |

---

Data Download Date/Time 11/11/2011 08:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THIRSK, K, M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: THIRSK, K  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: ADMIN  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: THIRSK

DATE/TIME OF TEST: 11/09/2011 21:22:49

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:21 |                    |
| Air Blank         | 0.000 | 21:21 |                    |
| Control Test      | 0.081 | 21:22 | Interferent Detect |
| Air Blank         | 0.000 | 21:22 |                    |

---

Data Download Date/Time 11/11/2011 08:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON, P, S  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: REYNOLDS, D  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: BROWNING

DATE/TIME OF TEST: 11/12/2011 01:04:40

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:58 |
| Air Blank         | 0.000 | 00:58 |
| Control Test      | 0.082 | 00:59 |
| Air Blank         | 0.000 | 00:59 |
| Subject Sample #1 | 0.246 | 01:00 |
| Breath Volume     | 3.597 | 01:00 |
| Air Blank         | 0.000 | 01:00 |
| Air Blank         | 0.000 | 01:02 |
| Subject Sample #2 | 0.254 | 01:03 |
| Breath Volume     | 2.621 | 01:03 |
| Air Blank         | 0.000 | 01:03 |
| Control Test      | 0.081 | 01:04 |
| Air Blank         | 0.000 | 01:04 |
| Diagnostics Check | OK    | 01:04 |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES, POLLY, E  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: MALLARD, J  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 17:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: NIEMAN

DATE/TIME OF TEST: 11/13/2011 19:29:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:26 |                      |
| Air Blank         | 0.000 | 19:26 |                      |
| Control Test      | 0.081 | 19:27 |                      |
| Air Blank         | 0.000 | 19:27 |                      |
| Subject Sample #1 |       | 19:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:27 |                      |
| Air Blank         | 0.000 | 19:28 |                      |
| Control Test      | 0.081 | 19:28 |                      |
| Air Blank         | 0.000 | 19:29 |                      |
| Diagnostics Check | OK    | 19:29 |                      |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BENNETT, J, K  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: KELLY, W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17811080A1  
DRY GAS EXPIRY DATE: 08/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/14/2011 01:38:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:35 |                      |
| Air Blank         | 0.000 | 01:36 |                      |
| Control Test      | 0.081 | 01:36 |                      |
| Air Blank         | 0.000 | 01:37 |                      |
| Subject Sample #1 |       | 01:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:37 |                      |
| Air Blank         | 0.000 | 01:37 |                      |
| Control Test      | 0.081 | 01:38 |                      |
| Air Blank         | 0.000 | 01:38 |                      |
| Diagnostics Check | OK    | 01:38 |                      |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON, C, C  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: LIVINGSTON, J  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17811080A1  
DRY GAS EXPIRY DATE: 08/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 11/15/2011 01:49:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:46 |                      |
| Air Blank         | 0.000 | 01:46 |                      |
| Control Test      | 0.081 | 01:47 |                      |
| Air Blank         | 0.000 | 01:47 |                      |
| Subject Sample #1 |       | 01:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:47 |                      |
| Air Blank         | 0.000 | 01:48 |                      |
| Control Test      | 0.081 | 01:48 |                      |
| Air Blank         | 0.000 | 01:49 |                      |
| Diagnostics Check | OK    | 01:49 |                      |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD, MELISSA, K  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: ARLINE, C  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17811080A1  
DRY GAS EXPIRY DATE: 08/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 11/18/2011 23:19:07

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 23:12 |                    |
| Air Blank         | 0.000 | 23:13 |                    |
| Control Test      | 0.080 | 23:13 |                    |
| Air Blank         | 0.000 | 23:14 |                    |
| Subject Sample #1 | 0.077 | 23:14 |                    |
| Breath Volume     | 1.179 | 23:14 |                    |
| Air Blank         | 0.000 | 23:15 |                    |
| Air Blank         | 0.000 | 23:16 |                    |
| Subject Sample #2 | 0.080 | 23:17 |                    |
| Breath Volume     | 2.066 | 23:17 |                    |
| Air Blank         | 0.000 | 23:18 |                    |
| Control Test      | 0.079 | 23:18 | Interferent Detect |
| Air Blank         | 0.000 | 23:19 |                    |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON, C, C  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: ARLINE, C  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17811080A1  
DRY GAS EXPIRY DATE: 08/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 11/19/2011 03:55:26

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:47 |
| Air Blank         | 0.000 | 03:47 |
| Control Test      | 0.082 | 03:48 |
| Air Blank         | 0.000 | 03:48 |
| Subject Sample #1 | 0.124 | 03:49 |
| Breath Volume     | 1.968 | 03:49 |
| Air Blank         | 0.000 | 03:49 |
| Air Blank         | 0.000 | 03:51 |
| Subject Sample #2 | 0.116 | 03:53 |
| Breath Volume     | 3.195 | 03:53 |
| Air Blank         | 0.000 | 03:54 |
| Control Test      | 0.080 | 03:54 |
| Air Blank         | 0.000 | 03:55 |
| Diagnostics Check | OK    | 03:55 |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THIRSK, K, M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: ENNIS, J  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17811080A1  
DRY GAS EXPIRY DATE: 08/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 11/19/2011 21:19:13

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 21:07 |                 |
| Air Blank         | 0.000 | 21:07 |                 |
| Control Test      | 0.081 | 21:07 |                 |
| Air Blank         | 0.000 | 21:08 |                 |
| Subject Sample #1 | 0.119 | 21:11 | Slope Not Level |
| Breath Volume     | 1.167 | 21:11 |                 |
| Air Blank         | 0.000 | 21:12 |                 |
| Air Blank         | 0.000 | 21:14 |                 |
| Subject Sample #2 | 0.117 | 21:14 |                 |
| Breath Volume     | 2.144 | 21:14 |                 |
| Air Blank         | 0.000 | 21:15 |                 |
| Air Blank         | 0.000 | 21:17 |                 |
| Subject Sample #3 | 0.117 | 21:17 |                 |
| Breath Volume     | 2.242 | 21:17 |                 |
| Air Blank         | 0.000 | 21:18 |                 |
| Control Test      | 0.081 | 21:18 |                 |
| Air Blank         | 0.000 | 21:19 |                 |
| Diagnostics Check | OK    | 21:19 |                 |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON, C, C  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: ARLINE, C  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17811080A1  
DRY GAS EXPIRY DATE: 08/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 11/20/2011 03:48:05

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:41 |
| Air Blank         | 0.000 | 03:42 |
| Control Test      | 0.082 | 03:42 |
| Air Blank         | 0.000 | 03:42 |
| Subject Sample #1 | 0.069 | 03:43 |
| Breath Volume     | 1.894 | 03:43 |
| Air Blank         | 0.000 | 03:44 |
| Air Blank         | 0.000 | 03:46 |
| Subject Sample #2 | 0.065 | 03:46 |
| Breath Volume     | 1.203 | 03:46 |
| Air Blank         | 0.000 | 03:47 |
| Control Test      | 0.079 | 03:47 |
| Air Blank         | 0.000 | 03:48 |
| Diagnostics Check | OK    | 03:48 |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON, C, C  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: LIVINGSTON, J  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17811080A1  
DRY GAS EXPIRY DATE: 08/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: TOURNAY

DATE/TIME OF TEST: 11/21/2011 03:02:00

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:55 |
| Air Blank         | 0.000 | 02:56 |
| Control Test      | 0.080 | 02:56 |
| Air Blank         | 0.000 | 02:57 |
| Subject Sample #1 | 0.305 | 02:57 |
| Breath Volume     | 1.562 | 02:57 |
| Air Blank         | 0.000 | 02:58 |
| Air Blank         | 0.000 | 02:59 |
| Subject Sample #2 | 0.288 | 03:00 |
| Breath Volume     | 1.117 | 03:00 |
| Air Blank         | 0.000 | 03:01 |
| Control Test      | 0.080 | 03:01 |
| Air Blank         | 0.000 | 03:01 |
| Diagnostics Check | OK    | 03:02 |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THIRSK, K, M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: ARLINE, C  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17811080A1  
DRY GAS EXPIRY DATE: 08/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 11/29/2011 21:35:16

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:31 |                    |
| Air Blank         | 0.000 | 21:32 |                    |
| Control Test      | 0.082 | 21:32 |                    |
| Air Blank         | 0.000 | 21:33 |                    |
| Subject Sample #1 |       | 21:33 | Interferent Detect |
| Breath Volume     | 4.480 | 21:33 |                    |
| Air Blank         | 0.000 | 21:34 |                    |
| Control Test      | 0.081 | 21:34 |                    |
| Air Blank         | 0.000 | 21:35 |                    |
| Diagnostics Check | OK    | 21:35 |                    |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THIRSK, K, M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: ARLINE, C  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 11/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17811080A1  
DRY GAS EXPIRY DATE: 08/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 11/29/2011 21:39:44

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:36 |                    |
| Air Blank         | 0.000 | 21:36 |                    |
| Control Test      | 0.081 | 21:37 |                    |
| Air Blank         | 0.000 | 21:37 |                    |
| Subject Sample #1 |       | 21:38 | Interferent Detect |
| Breath Volume     | 2.460 | 21:38 |                    |
| Air Blank         | 0.000 | 21:38 |                    |
| Control Test      | 0.082 | 21:39 |                    |
| Air Blank         | 0.000 | 21:39 |                    |
| Diagnostics Check | OK    | 21:39 |                    |

---

Data Download Date/Time 12/23/2011 10:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEFFENSE, JEAN-GUY,

BREATH TEST OPERATOR AGENCY: U of FI PD

ARRESTING OFFICER: GENTRY, DARIN

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 03:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BORRERO-RODRIQUEZ

DATE/TIME OF TEST: 11/03/2011 03:55:59

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:49 |
| Air Blank         | 0.000 | 03:49 |
| Control Test      | 0.079 | 03:49 |
| Air Blank         | 0.000 | 03:50 |
| Subject Sample #1 | 0.111 | 03:51 |
| Breath Volume     | 3.039 | 03:51 |
| Air Blank         | 0.000 | 03:52 |
| Air Blank         | 0.000 | 03:53 |
| Subject Sample #2 | 0.111 | 03:54 |
| Breath Volume     | 3.300 | 03:54 |
| Air Blank         | 0.000 | 03:55 |
| Control Test      | 0.078 | 03:55 |
| Air Blank         | 0.000 | 03:55 |
| Diagnostics Check | OK    | 03:55 |

---

Data Download Date/Time 11/23/2011 10:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECK, TIMOTHY, L  
BREATH TEST OPERATOR AGENCY: U of FI PD  
ARRESTING OFFICER: CLARK, WAYNE  
ARRESTING OFFICER AGENCY: U of FI PD  
LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 07:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: BUJALSKI

DATE/TIME OF TEST: 11/04/2011 08:35:37

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 08:33 |            |
| Air Blank         | 0.000 | 08:34 |            |
| Control Test      | 0.078 | 08:34 |            |
| Air Blank         |       | 08:35 | RFI Detect |
| Air Blank         | 0.000 | 08:35 |            |

---

Data Download Date/Time 11/23/2011 10:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECK, TIMOTHY, L  
BREATH TEST OPERATOR AGENCY: U of FI PD  
ARRESTING OFFICER: CLARK, WAYNE  
ARRESTING OFFICER AGENCY: U of FI PD  
LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 07:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: BUJALSKI

DATE/TIME OF TEST: 11/04/2011 08:40:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 08:37 |                      |
| Air Blank         | 0.000 | 08:38 |                      |
| Control Test      | 0.079 | 08:38 |                      |
| Air Blank         | 0.000 | 08:39 |                      |
| Subject Sample #1 |       | 08:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 08:39 |                      |
| Air Blank         | 0.000 | 08:39 |                      |
| Control Test      | 0.079 | 08:40 |                      |
| Air Blank         | 0.000 | 08:40 |                      |
| Diagnostics Check | OK    | 08:40 |                      |

---

Data Download Date/Time 11/23/2011 10:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOLWAY, CHAD, S

BREATH TEST OPERATOR AGENCY: U of FI PD

ARRESTING OFFICER: ROBERTS, BRADLEY

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 02:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: WESTCOTT

DATE/TIME OF TEST: 11/05/2011 03:16:53

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:09 |
| Air Blank         | 0.000 | 03:09 |
| Control Test      | 0.079 | 03:10 |
| Air Blank         | 0.000 | 03:10 |
| Subject Sample #1 | 0.111 | 03:12 |
| Breath Volume     | 1.351 | 03:12 |
| Air Blank         | 0.000 | 03:12 |
| Air Blank         | 0.000 | 03:14 |
| Subject Sample #2 | 0.114 | 03:15 |
| Breath Volume     | 1.671 | 03:15 |
| Air Blank         | 0.000 | 03:15 |
| Control Test      | 0.080 | 03:16 |
| Air Blank         | 0.000 | 03:16 |
| Diagnostics Check | OK    | 03:16 |

---

Data Download Date/Time 11/23/2011 10:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARONI, JUSTIN, S

BREATH TEST OPERATOR AGENCY: U of FI PD

ARRESTING OFFICER: ROBERTS, BRADLEY

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 01:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: ROYCE

DATE/TIME OF TEST: 11/06/2011 01:33:13

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:26 |
| Air Blank         | 0.000 | 01:27 |
| Control Test      | 0.079 | 01:27 |
| Air Blank         | 0.000 | 01:28 |
| Subject Sample #1 | 0.085 | 01:28 |
| Breath Volume     | 3.281 | 01:28 |
| Air Blank         | 0.000 | 01:29 |
| Air Blank         | 0.000 | 01:31 |
| Subject Sample #2 | 0.085 | 01:31 |
| Breath Volume     | 2.824 | 01:31 |
| Air Blank         | 0.000 | 01:32 |
| Control Test      | 0.080 | 01:32 |
| Air Blank         | 0.000 | 01:33 |
| Diagnostics Check | OK    | 01:33 |

---

Data Download Date/Time 11/23/2011 10:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEXTON, SAMUEL, D  
BREATH TEST OPERATOR AGENCY: U of FI PD  
ARRESTING OFFICER: HAGER, JAMES  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: ROBERT

DATE/TIME OF TEST: 11/14/2011 00:40:53

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:32 |
| Air Blank         | 0.000 | 00:32 |
| Control Test      | 0.078 | 00:33 |
| Air Blank         | 0.000 | 00:33 |
| Subject Sample #1 | 0.237 | 00:36 |
| Breath Volume     | 1.273 | 00:36 |
| Air Blank         | 0.000 | 00:36 |
| Air Blank         | 0.000 | 00:38 |
| Subject Sample #2 | 0.227 | 00:39 |
| Breath Volume     | 2.054 | 00:39 |
| Air Blank         | 0.000 | 00:39 |
| Control Test      | 0.077 | 00:40 |
| Air Blank         | 0.000 | 00:40 |
| Diagnostics Check | OK    | 00:40 |

---

Data Download Date/Time 11/23/2011 10:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARONI, JUSTIN, S  
BREATH TEST OPERATOR AGENCY: U of FI PD  
ARRESTING OFFICER: HENDON, BOBBIE  
ARRESTING OFFICER AGENCY: U of FI PD  
LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: LEVY

DATE/TIME OF TEST: 11/16/2011 03:09:10

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:57 |                    |
| Air Blank         | 0.000 | 02:58 |                    |
| Control Test      | 0.078 | 02:58 |                    |
| Air Blank         | 0.000 | 02:59 |                    |
| Subject Sample #1 |       | 03:02 | No Sample Provided |
| Breath Volume     | 0.000 | 03:02 |                    |
| Air Blank         | 0.000 | 03:02 |                    |
| Air Blank         | 0.000 | 03:04 |                    |
| Subject Sample #2 |       | 03:07 | No Sample Provided |
| Breath Volume     | 0.000 | 03:07 |                    |
| Air Blank         | 0.000 | 03:08 |                    |
| Control Test      | 0.080 | 03:08 |                    |
| Air Blank         | 0.000 | 03:09 |                    |
| Diagnostics Check | OK    | 03:09 |                    |

---

Data Download Date/Time 11/23/2011 10:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEFFENSE, JEAN-GUY,

BREATH TEST OPERATOR AGENCY: U of FI PD

ARRESTING OFFICER: GENTRY, DARIN

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 03:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: LONG-JONES

DATE/TIME OF TEST: 11/18/2011 04:09:20

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:59 |                      |
| Air Blank         | 0.000 | 03:59 |                      |
| Control Test      | 0.079 | 04:00 |                      |
| Air Blank         | 0.000 | 04:00 |                      |
| Subject Sample #1 |       | 04:03 | No Sample Provided   |
| Breath Volume     | 0.000 | 04:03 |                      |
| Air Blank         | 0.000 | 04:04 |                      |
| Air Blank         | 0.000 | 04:06 |                      |
| Subject Sample #2 |       | 04:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:07 |                      |
| Air Blank         | 0.000 | 04:08 |                      |
| Control Test      | 0.080 | 04:08 |                      |
| Air Blank         | 0.000 | 04:09 |                      |
| Diagnostics Check | OK    | 04:09 |                      |

---

Data Download Date/Time 11/23/2011 10:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEFFENSE, JEAN-GUY,

BREATH TEST OPERATOR AGENCY: U of FI PD

ARRESTING OFFICER: AUSGOOD, BRIAN

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 02:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MIZELL

DATE/TIME OF TEST: 11/23/2011 03:20:37

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:14 |
| Air Blank         | 0.000 | 03:14 |
| Control Test      | 0.078 | 03:14 |
| Air Blank         | 0.000 | 03:15 |
| Subject Sample #1 | 0.157 | 03:16 |
| Breath Volume     | 1.613 | 03:16 |
| Air Blank         | 0.000 | 03:16 |
| Air Blank         | 0.000 | 03:18 |
| Subject Sample #2 | 0.147 | 03:19 |
| Breath Volume     | 1.585 | 03:19 |
| Air Blank         | 0.000 | 03:19 |
| Control Test      | 0.078 | 03:20 |
| Air Blank         | 0.000 | 03:20 |
| Diagnostics Check | OK    | 03:20 |

---

Data Download Date/Time 11/23/2011 10:19

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELUCCA, SAMUEL, E  
BREATH TEST OPERATOR AGENCY: U of FI PD  
ARRESTING OFFICER: DRAKE, C  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 11/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 08:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: DUPUIS

DATE/TIME OF TEST: 11/25/2011 09:26:45

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 09:21 |                      |
| Air Blank         | 0.000 | 09:22 |                      |
| Control Test      | 0.080 | 09:22 |                      |
| Air Blank         | 0.000 | 09:23 |                      |
| Subject Sample #1 |       | 09:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 09:25 |                      |
| Air Blank         | 0.000 | 09:25 |                      |
| Control Test      | 0.081 | 09:26 |                      |
| Air Blank         | 0.000 | 09:26 |                      |
| Diagnostics Check | OK    | 09:26 |                      |

---

Data Download Date/Time 12/07/2011 14:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SEXTON, SAMUEL, D

BREATH TEST OPERATOR AGENCY: U of FI PD

ARRESTING OFFICER: NGUYEN, RICK

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 11/23/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 01:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MCCANN

DATE/TIME OF TEST: 11/26/2011 02:58:46

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 02:47 |                |
| Air Blank         | 0.000 | 02:47 |                |
| Control Test      | 0.078 | 02:47 |                |
| Air Blank         | 0.000 | 02:48 |                |
| Subject Sample #1 | 0.195 | 02:51 | Volume Not Met |
| Breath Volume     | 0.261 | 02:51 |                |
| Air Blank         | 0.000 | 02:52 |                |
| Air Blank         | 0.000 | 02:54 |                |
| Subject Sample #2 | 0.218 | 02:57 | Volume Not Met |
| Breath Volume     | 0.988 | 02:57 |                |
| Air Blank         | 0.000 | 02:57 |                |
| Control Test      | 0.077 | 02:58 |                |
| Air Blank         | 0.000 | 02:58 |                |
| Diagnostics Check | OK    | 02:58 |                |

---

Data Download Date/Time 12/07/2011 14:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRUITT, JACOB, R  
BREATH TEST OPERATOR AGENCY: U of FI PD  
ARRESTING OFFICER: BUSATTA, THOMAS  
ARRESTING OFFICER AGENCY: U of FI PD  
LAST AGENCY INSPECTION: 11/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 719876  
DRY GAS EXPIRY DATE: 01/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 11/26/2011 21:04:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:55 |                      |
| Air Blank         | 0.000 | 20:56 |                      |
| Control Test      | 0.079 | 20:56 |                      |
| Air Blank         | 0.000 | 20:57 |                      |
| Subject Sample #1 |       | 21:00 | No Sample Provided   |
| Breath Volume     | 0.000 | 21:00 |                      |
| Air Blank         | 0.000 | 21:00 |                      |
| Air Blank         | 0.000 | 21:02 |                      |
| Subject Sample #2 |       | 21:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:02 |                      |
| Air Blank         | 0.000 | 21:03 |                      |
| Control Test      | 0.078 | 21:03 |                      |
| Air Blank         | 0.000 | 21:04 |                      |
| Diagnostics Check | OK    | 21:04 |                      |

---

Data Download Date/Time 12/07/2011 14:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHEPPARD, JOHNNY, A  
BREATH TEST OPERATOR AGENCY: High Springs PD  
ARRESTING OFFICER: SHEPPARD, JOHNNY  
ARRESTING OFFICER AGENCY: High Springs PD  
LAST AGENCY INSPECTION: 10/30/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 817739  
DRY GAS EXPIRY DATE: 08/20/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 11/01/2011 04:50:59

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 04:38 |                |
| Air Blank         | 0.000 | 04:39 |                |
| Control Test      | 0.080 | 04:39 |                |
| Air Blank         | 0.000 | 04:39 |                |
| Subject Sample #1 | 0.129 | 04:43 | Volume Not Met |
| Breath Volume     | 0.625 | 04:43 |                |
| Air Blank         | 0.000 | 04:43 |                |
| Air Blank         | 0.000 | 04:45 |                |
| Subject Sample #2 | 0.126 | 04:46 |                |
| Breath Volume     | 2.007 | 04:46 |                |
| Air Blank         | 0.000 | 04:46 |                |
| Air Blank         | 0.000 | 04:48 |                |
| Subject Sample #3 | 0.123 | 04:49 |                |
| Breath Volume     | 1.136 | 04:49 |                |
| Air Blank         | 0.000 | 04:50 |                |
| Control Test      | 0.078 | 04:50 |                |
| Air Blank         | 0.000 | 04:50 |                |
| Diagnostics Check | OK    | 04:50 |                |

---

Data Download Date/Time 11/25/2011 15:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHEPPARD, JOHNNY, A  
BREATH TEST OPERATOR AGENCY: High Springs PD  
ARRESTING OFFICER: SHEPPARD, JOHNNY  
ARRESTING OFFICER AGENCY: High Springs PD  
LAST AGENCY INSPECTION: 10/30/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 817739  
DRY GAS EXPIRY DATE: 08/20/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/05/2011 22:57:27

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:51 |
| Air Blank         | 0.000 | 22:51 |
| Control Test      | 0.081 | 22:51 |
| Air Blank         | 0.000 | 22:52 |
| Subject Sample #1 | 0.247 | 22:52 |
| Breath Volume     | 1.476 | 22:52 |
| Air Blank         | 0.000 | 22:53 |
| Air Blank         | 0.000 | 22:55 |
| Subject Sample #2 | 0.253 | 22:55 |
| Breath Volume     | 2.210 | 22:55 |
| Air Blank         | 0.000 | 22:56 |
| Control Test      | 0.078 | 22:56 |
| Air Blank         | 0.000 | 22:57 |
| Diagnostics Check | OK    | 22:57 |

---

Data Download Date/Time 11/25/2011 15:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHEPPARD, JOHNNY, A  
BREATH TEST OPERATOR AGENCY: High Springs PD  
ARRESTING OFFICER: JOY, ADAM  
ARRESTING OFFICER AGENCY: High Springs PD  
LAST AGENCY INSPECTION: 10/30/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 817739  
DRY GAS EXPIRY DATE: 08/20/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WASH

DATE/TIME OF TEST: 11/05/2011 23:36:12

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:29 |
| Air Blank         | 0.000 | 23:29 |
| Control Test      | 0.081 | 23:29 |
| Air Blank         | 0.000 | 23:30 |
| Subject Sample #1 | 0.086 | 23:31 |
| Breath Volume     | 1.722 | 23:31 |
| Air Blank         | 0.000 | 23:32 |
| Air Blank         | 0.000 | 23:33 |
| Subject Sample #2 | 0.084 | 23:34 |
| Breath Volume     | 1.324 | 23:34 |
| Air Blank         | 0.000 | 23:35 |
| Control Test      | 0.078 | 23:35 |
| Air Blank         | 0.000 | 23:36 |
| Diagnostics Check | OK    | 23:36 |

---

Data Download Date/Time 11/25/2011 15:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BENNETT, JONATHON, K  
BREATH TEST OPERATOR AGENCY: Alachua PD  
ARRESTING OFFICER: VARGAS, CESAR  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 10/30/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 817739  
DRY GAS EXPIRY DATE: 08/20/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 11/12/2011 21:56:52

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 21:55 |                           |
| Air Blank         | 0.000 | 21:55 |                           |
| Control Test      | 0.195 | 21:56 | Control Outside Tolerance |
| Air Blank         | 0.000 | 21:56 |                           |

---

Data Download Date/Time 11/25/2011 15:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BENNETT, JONATHON, K  
BREATH TEST OPERATOR AGENCY: Alachua PD  
ARRESTING OFFICER: VARGAS, CESAR  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 10/30/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 817739  
DRY GAS EXPIRY DATE: 08/20/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 11/12/2011 22:11:49

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 22:06 |            |
| Air Blank         | 0.000 | 22:06 |            |
| Control Test      | 0.081 | 22:07 |            |
| Air Blank         | 0.000 | 22:07 |            |
| Subject Sample #1 | 0.213 | 22:08 |            |
| Breath Volume     | 2.621 | 22:08 |            |
| Air Blank         | 0.000 | 22:08 |            |
| Air Blank         | 0.000 | 22:10 |            |
| Subject Sample #2 | 0.201 | 22:10 |            |
| Breath Volume     | 1.195 | 22:10 |            |
| Air Blank         |       | 22:11 | RFI Detect |
| Air Blank         | 0.000 | 22:11 |            |

---

Data Download Date/Time 11/25/2011 15:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHEPPARD, JOHNNY, A  
BREATH TEST OPERATOR AGENCY: High Springs PD  
ARRESTING OFFICER: HAMPTON, KENDRICK  
ARRESTING OFFICER AGENCY: High Springs PD  
LAST AGENCY INSPECTION: 10/30/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 817739  
DRY GAS EXPIRY DATE: 08/20/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: BERNARD - STEVENS

DATE/TIME OF TEST: 11/24/2011 22:42:14

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:36 |
| Air Blank         | 0.000 | 22:36 |
| Control Test      | 0.080 | 22:36 |
| Air Blank         | 0.000 | 22:37 |
| Subject Sample #1 | 0.158 | 22:37 |
| Breath Volume     | 3.507 | 22:37 |
| Air Blank         | 0.000 | 22:38 |
| Air Blank         | 0.000 | 22:40 |
| Subject Sample #2 | 0.154 | 22:40 |
| Breath Volume     | 3.542 | 22:40 |
| Air Blank         | 0.000 | 22:41 |
| Control Test      | 0.079 | 22:41 |
| Air Blank         | 0.000 | 22:42 |
| Diagnostics Check | OK    | 22:42 |

---

Data Download Date/Time 11/25/2011 15:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STALVEY, STEPHEN, W  
BREATH TEST OPERATOR AGENCY: Gilchrist County SO  
ARRESTING OFFICER: STALVEY, STEPHEN  
ARRESTING OFFICER AGENCY: Gilchrist County SO  
LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05911080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: RENAUD

DATE/TIME OF TEST: 11/26/2011 03:17:13

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:10 |
| Air Blank         | 0.000 | 03:11 |
| Control Test      | 0.077 | 03:11 |
| Air Blank         | 0.000 | 03:12 |
| Subject Sample #1 | 0.173 | 03:12 |
| Breath Volume     | 2.175 | 03:12 |
| Air Blank         | 0.000 | 03:13 |
| Air Blank         | 0.000 | 03:15 |
| Subject Sample #2 | 0.176 | 03:15 |
| Breath Volume     | 2.992 | 03:15 |
| Air Blank         | 0.000 | 03:16 |
| Control Test      | 0.076 | 03:16 |
| Air Blank         | 0.000 | 03:17 |
| Diagnostics Check | OK    | 03:17 |

---

Data Download Date/Time 11/27/2011 23:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001286

INSTRUMENT REGISTERED TO: BAKER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOWNSEND, CHARLES, W  
BREATH TEST OPERATOR AGENCY: Baker County SO  
ARRESTING OFFICER: BISHARA, SHAWN  
ARRESTING OFFICER AGENCY: Baker County SO  
LAST AGENCY INSPECTION: 11/13/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: PREVATT

DATE/TIME OF TEST: 11/24/2011 00:03:02

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:56 |
| Air Blank         | 0.000 | 23:57 |
| Control Test      | 0.077 | 23:57 |
| Air Blank         | 0.000 | 23:58 |
| Subject Sample #1 | 0.144 | 23:58 |
| Breath Volume     | 1.695 | 23:58 |
| Air Blank         | 0.000 | 23:59 |
| Air Blank         | 0.000 | 00:00 |
| Subject Sample #2 | 0.147 | 00:01 |
| Breath Volume     | 1.761 | 00:01 |
| Air Blank         | 0.000 | 00:02 |
| Control Test      | 0.077 | 00:02 |
| Air Blank         | 0.000 | 00:02 |
| Diagnostics Check | OK    | 00:03 |

---

Data Download Date/Time 12/11/2011 19:11

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILLIARD, KARI, A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: HILLIARD, KARI  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01011080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 08:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 11/17/2011 21:22:34

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:19 |                      |
| Air Blank         | 0.000 | 21:20 |                      |
| Control Test      | 0.080 | 21:20 |                      |
| Air Blank         | 0.000 | 21:21 |                      |
| Subject Sample #1 |       | 21:21 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:21 |                      |
| Air Blank         | 0.000 | 21:21 |                      |
| Control Test      | 0.079 | 21:21 |                      |
| Air Blank         | 0.000 | 21:22 |                      |
| Diagnostics Check | OK    | 21:22 |                      |

---

Data Download Date/Time 12/11/2011 19:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRIGGERS, RODNEY, M  
BREATH TEST OPERATOR AGENCY: Baker County SO  
ARRESTING OFFICER: DRIGGERS, RODNEY  
ARRESTING OFFICER AGENCY: Baker County SO  
LAST AGENCY INSPECTION: 11/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01011080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: FAIRBROTHER

DATE/TIME OF TEST: 11/18/2011 04:05:40

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:56 |
| Air Blank         | 0.000 | 03:56 |
| Control Test      | 0.080 | 03:56 |
| Air Blank         | 0.000 | 03:57 |
| Subject Sample #1 | 0.154 | 03:58 |
| Breath Volume     | 1.792 | 03:58 |
| Air Blank         | 0.000 | 03:59 |
| Air Blank         | 0.000 | 04:01 |
| Subject Sample #2 | 0.151 | 04:03 |
| Breath Volume     | 1.113 | 04:03 |
| Air Blank         | 0.000 | 04:04 |
| Control Test      | 0.077 | 04:05 |
| Air Blank         | 0.000 | 04:05 |
| Diagnostics Check | OK    | 04:05 |

---

Data Download Date/Time 12/11/2011 19:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YARBOROUGH, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Baker County SO  
ARRESTING OFFICER: MOBLEY, DARYL  
ARRESTING OFFICER AGENCY: Baker County SO  
LAST AGENCY INSPECTION: 11/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01011080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: KERVIN

DATE/TIME OF TEST: 11/23/2011 00:12:59

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:04 |
| Air Blank         | 0.000 | 00:04 |
| Control Test      | 0.080 | 00:04 |
| Air Blank         | 0.000 | 00:05 |
| Subject Sample #1 | 0.184 | 00:05 |
| Breath Volume     | 1.265 | 00:05 |
| Air Blank         | 0.000 | 00:06 |
| Air Blank         | 0.000 | 00:08 |
| Subject Sample #2 | 0.189 | 00:11 |
| Breath Volume     | 1.191 | 00:11 |
| Air Blank         | 0.000 | 00:12 |
| Control Test      | 0.079 | 00:12 |
| Air Blank         | 0.000 | 00:12 |
| Diagnostics Check | OK    | 00:12 |

---

Data Download Date/Time 12/11/2011 19:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEALY, G, B

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: HEALY, G

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 11/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 01011080A2

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 00:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 11/27/2011 00:53:03

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:50 |                      |
| Air Blank         | 0.000 | 00:50 |                      |
| Control Test      | 0.080 | 00:51 |                      |
| Air Blank         | 0.000 | 00:51 |                      |
| Subject Sample #1 |       | 00:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:51 |                      |
| Air Blank         | 0.000 | 00:52 |                      |
| Control Test      | 0.079 | 00:52 |                      |
| Air Blank         | 0.000 | 00:52 |                      |
| Diagnostics Check | OK    | 00:53 |                      |

---

Data Download Date/Time 12/11/2011 19:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOTRIGHT JR, KENNETH, M  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: BOATRRIGHT JR, KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01011080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 11:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/27/2011 01:36:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:33 |                      |
| Air Blank         | 0.000 | 01:33 |                      |
| Control Test      | 0.079 | 01:34 |                      |
| Air Blank         | 0.000 | 01:34 |                      |
| Subject Sample #1 |       | 01:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:34 |                      |
| Air Blank         | 0.000 | 01:35 |                      |
| Control Test      | 0.079 | 01:35 |                      |
| Air Blank         | 0.000 | 01:36 |                      |
| Diagnostics Check | OK    | 01:36 |                      |

---

Data Download Date/Time 12/11/2011 19:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOWNSEND, CHARLES, W  
BREATH TEST OPERATOR AGENCY: Union County SO  
ARRESTING OFFICER: TOWNSEND, CHARLES  
ARRESTING OFFICER AGENCY: Union County SO  
LAST AGENCY INSPECTION: 11/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 11/21/2011 03:18:53

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 03:16 |            |
| Air Blank         | 0.000 | 03:17 |            |
| Control Test      | 0.079 | 03:17 |            |
| Air Blank         | 0.000 | 03:18 |            |
| Subject Sample #1 |       | 03:18 | RFI Detect |
| Breath Volume     | 0.000 | 03:18 |            |
| Air Blank         |       | 03:18 | RFI Detect |
| Air Blank         | 0.000 | 03:18 |            |

---

Data Download Date/Time 01/05/2012 20:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOWNSEND, CHARLES, W  
BREATH TEST OPERATOR AGENCY: Union County SO  
ARRESTING OFFICER: TOWNSEND, CHARLES  
ARRESTING OFFICER AGENCY: Union County SO  
LAST AGENCY INSPECTION: 11/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 11/21/2011 03:25:45

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:22 |                      |
| Air Blank         | 0.000 | 03:23 |                      |
| Control Test      | 0.079 | 03:23 |                      |
| Air Blank         | 0.000 | 03:24 |                      |
| Subject Sample #1 |       | 03:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:24 |                      |
| Air Blank         | 0.000 | 03:24 |                      |
| Control Test      | 0.079 | 03:25 |                      |
| Air Blank         | 0.000 | 03:25 |                      |
| Diagnostics Check | OK    | 03:25 |                      |

---

Data Download Date/Time 01/05/2012 20:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KONKEL, SCOTT, M  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: WALLACE, AMANDA  
ARRESTING OFFICER AGENCY: Starke PD  
LAST AGENCY INSPECTION: 10/30/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/06/2011 04:58:13

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:50 |
| Air Blank         | 0.000 | 04:51 |
| Control Test      | 0.081 | 04:51 |
| Air Blank         | 0.000 | 04:52 |
| Subject Sample #1 | 0.203 | 04:53 |
| Breath Volume     | 1.664 | 04:53 |
| Air Blank         | 0.000 | 04:54 |
| Air Blank         | 0.000 | 04:56 |
| Subject Sample #2 | 0.192 | 04:56 |
| Breath Volume     | 1.328 | 04:56 |
| Air Blank         | 0.000 | 04:57 |
| Control Test      | 0.079 | 04:57 |
| Air Blank         | 0.000 | 04:58 |
| Diagnostics Check | OK    | 04:58 |

---

Data Download Date/Time 11/15/2011 16:25

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOATRIGHT JR, KENNETH, M  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: BOATRIGHT JR, KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/15/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 10:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: CHASTAIN

DATE/TIME OF TEST: 11/18/2011 23:14:20

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:08 |
| Air Blank         | 0.000 | 23:09 |
| Control Test      | 0.080 | 23:09 |
| Air Blank         | 0.000 | 23:09 |
| Subject Sample #1 | 0.000 | 23:10 |
| Breath Volume     | 2.960 | 23:10 |
| Air Blank         | 0.000 | 23:10 |
| Air Blank         | 0.000 | 23:12 |
| Subject Sample #2 | 0.000 | 23:12 |
| Breath Volume     | 2.273 | 23:12 |
| Air Blank         | 0.000 | 23:13 |
| Control Test      | 0.079 | 23:13 |
| Air Blank         | 0.000 | 23:14 |
| Diagnostics Check | OK    | 23:14 |

---

Data Download Date/Time 12/13/2011 10:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAGE, TRENT, A  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: PAGE, TRENT  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 11/15/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 11/19/2011 22:23:30

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:12 |                    |
| Air Blank         | 0.000 | 22:12 |                    |
| Control Test      | 0.078 | 22:12 |                    |
| Air Blank         | 0.000 | 22:13 |                    |
| Subject Sample #1 | 0.241 | 22:16 | Volume Not Met     |
| Breath Volume     | 0.308 | 22:16 |                    |
| Air Blank         | 0.000 | 22:17 |                    |
| Air Blank         | 0.000 | 22:19 |                    |
| Subject Sample #2 |       | 22:22 | No Sample Provided |
| Breath Volume     | 0.000 | 22:22 |                    |
| Air Blank         | 0.000 | 22:22 |                    |
| Control Test      | 0.078 | 22:22 |                    |
| Air Blank         | 0.000 | 22:23 |                    |
| Diagnostics Check | OK    | 22:23 |                    |

---

Data Download Date/Time 12/13/2011 10:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE, KELLY, D  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: RODRIGUEZ, JUAN  
ARRESTING OFFICER AGENCY: Hampton PD  
LAST AGENCY INSPECTION: 11/15/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: TURKNETT

DATE/TIME OF TEST: 11/27/2011 00:31:03

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 00:18 |                |
| Air Blank         | 0.000 | 00:18 |                |
| Control Test      | 0.080 | 00:19 |                |
| Air Blank         | 0.000 | 00:19 |                |
| Subject Sample #1 | 0.057 | 00:20 |                |
| Breath Volume     | 1.488 | 00:20 |                |
| Air Blank         | 0.000 | 00:21 |                |
| Air Blank         | 0.000 | 00:23 |                |
| Subject Sample #2 | 0.051 | 00:26 | Volume Not Met |
| Breath Volume     | 0.820 | 00:26 |                |
| Air Blank         | 0.000 | 00:26 |                |
| Air Blank         | 0.000 | 00:28 |                |
| Subject Sample #3 | 0.050 | 00:29 |                |
| Breath Volume     | 2.285 | 00:29 |                |
| Air Blank         | 0.000 | 00:30 |                |
| Control Test      | 0.079 | 00:30 |                |
| Air Blank         | 0.000 | 00:30 |                |
| Diagnostics Check | OK    | 00:31 |                |

---

Data Download Date/Time 12/13/2011 10:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KONKEL, SCOTT, M  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: KONKEL, SCOTT  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 11/15/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/27/2011 02:22:08

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:16 |
| Air Blank         | 0.000 | 02:16 |
| Control Test      | 0.081 | 02:16 |
| Air Blank         | 0.000 | 02:17 |
| Subject Sample #1 | 0.164 | 02:17 |
| Breath Volume     | 4.628 | 02:17 |
| Air Blank         | 0.000 | 02:18 |
| Air Blank         | 0.000 | 02:20 |
| Subject Sample #2 | 0.159 | 02:20 |
| Breath Volume     | 4.644 | 02:20 |
| Air Blank         | 0.000 | 02:21 |
| Control Test      | 0.080 | 02:21 |
| Air Blank         | 0.000 | 02:22 |
| Diagnostics Check | OK    | 02:22 |

---

Data Download Date/Time 12/13/2011 10:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHILLER, JOSEPH, C  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: MURRAY, WILLIAM  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 11/15/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 11/27/2011 04:02:24

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:55 |
| Air Blank         | 0.000 | 03:56 |
| Control Test      | 0.080 | 03:56 |
| Air Blank         | 0.000 | 03:56 |
| Subject Sample #1 | 0.199 | 03:58 |
| Breath Volume     | 2.082 | 03:58 |
| Air Blank         | 0.000 | 03:58 |
| Air Blank         | 0.000 | 04:00 |
| Subject Sample #2 | 0.180 | 04:00 |
| Breath Volume     | 2.929 | 04:00 |
| Air Blank         | 0.000 | 04:01 |
| Control Test      | 0.078 | 04:01 |
| Air Blank         | 0.000 | 04:02 |
| Diagnostics Check | OK    | 04:02 |

---

Data Download Date/Time 12/13/2011 10:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, CHARLES, K  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: HARRIS, CHARLES  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 18:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: MAXFIELD

DATE/TIME OF TEST: 11/09/2011 19:57:26

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:51 |
| Air Blank         | 0.000 | 19:51 |
| Control Test      | 0.081 | 19:51 |
| Air Blank         | 0.000 | 19:52 |
| Subject Sample #1 | 0.163 | 19:52 |
| Breath Volume     | 3.289 | 19:52 |
| Air Blank         | 0.000 | 19:53 |
| Air Blank         | 0.000 | 19:55 |
| Subject Sample #2 | 0.154 | 19:55 |
| Breath Volume     | 2.304 | 19:55 |
| Air Blank         | 0.000 | 19:56 |
| Control Test      | 0.081 | 19:56 |
| Air Blank         | 0.000 | 19:57 |
| Diagnostics Check | OK    | 19:57 |

---

Data Download Date/Time 11/29/2011 16:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COSSEBOOM, M, J  
BREATH TEST OPERATOR AGENCY: Cedar Keys PD  
ARRESTING OFFICER: COSSEBOOM, M  
ARRESTING OFFICER AGENCY: Cedar Keys PD  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: EILAND

DATE/TIME OF TEST: 11/10/2011 02:45:45

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:39 |
| Air Blank         | 0.000 | 02:39 |
| Control Test      | 0.082 | 02:39 |
| Air Blank         | 0.000 | 02:40 |
| Subject Sample #1 | 0.166 | 02:41 |
| Breath Volume     | 1.750 | 02:41 |
| Air Blank         | 0.000 | 02:41 |
| Air Blank         | 0.000 | 02:43 |
| Subject Sample #2 | 0.174 | 02:44 |
| Breath Volume     | 2.562 | 02:44 |
| Air Blank         | 0.000 | 02:44 |
| Control Test      | 0.081 | 02:45 |
| Air Blank         | 0.000 | 02:45 |
| Diagnostics Check | OK    | 02:45 |

---

Data Download Date/Time 11/29/2011 16:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COSSEBOOM, M, J  
BREATH TEST OPERATOR AGENCY: Cedar Keys PD  
ARRESTING OFFICER: COSSEBOOM, M  
ARRESTING OFFICER AGENCY: Cedar Keys PD  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SOMERS

DATE/TIME OF TEST: 11/19/2011 22:32:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:28 |                      |
| Air Blank         | 0.000 | 22:28 |                      |
| Control Test      | 0.081 | 22:29 |                      |
| Air Blank         | 0.000 | 22:29 |                      |
| Subject Sample #1 |       | 22:31 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:31 |                      |
| Air Blank         | 0.000 | 22:32 |                      |
| Control Test      | 0.081 | 22:32 |                      |
| Air Blank         | 0.000 | 22:32 |                      |
| Diagnostics Check | OK    | 22:32 |                      |

---

Data Download Date/Time 11/29/2011 16:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOK-OTT, LINDA, N  
BREATH TEST OPERATOR AGENCY: Levy County SO  
ARRESTING OFFICER: HORNE, ALTON  
ARRESTING OFFICER AGENCY: Chiefland PD  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: WINTERBERG

DATE/TIME OF TEST: 11/20/2011 04:40:47

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 04:29 |                    |
| Air Blank         | 0.000 | 04:29 |                    |
| Control Test      | 0.080 | 04:30 |                    |
| Air Blank         | 0.000 | 04:30 |                    |
| Subject Sample #1 |       | 04:33 | No Sample Provided |
| Breath Volume     | 0.000 | 04:33 |                    |
| Air Blank         | 0.000 | 04:34 |                    |
| Air Blank         | 0.000 | 04:36 |                    |
| Subject Sample #2 |       | 04:39 | No Sample Provided |
| Breath Volume     | 0.000 | 04:39 |                    |
| Air Blank         | 0.000 | 04:39 |                    |
| Control Test      | 0.081 | 04:40 |                    |
| Air Blank         | 0.000 | 04:40 |                    |
| Diagnostics Check | OK    | 04:40 |                    |

---

Data Download Date/Time 11/29/2011 16:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHEPPARD, KEVIN, B

BREATH TEST OPERATOR AGENCY: Williston PD

ARRESTING OFFICER: HILBERER, BRIAN

ARRESTING OFFICER AGENCY: Inglis PD

LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1187042

DRY GAS EXPIRY DATE: 07/22/2014

VIOLATION CODE: DUI

ARREST TIME: 20:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 11/01/2011 21:57:12

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 21:50 |
| Air Blank         | 0.000 | 21:51 |
| Control Test      | 0.080 | 21:51 |
| Air Blank         | 0.000 | 21:52 |
| Subject Sample #1 | 0.102 | 21:52 |
| Breath Volume     | 2.910 | 21:52 |
| Air Blank         | 0.000 | 21:53 |
| Air Blank         | 0.000 | 21:55 |
| Subject Sample #2 | 0.100 | 21:55 |
| Breath Volume     | 3.171 | 21:55 |
| Air Blank         | 0.000 | 21:56 |
| Control Test      | 0.079 | 21:56 |
| Air Blank         | 0.000 | 21:57 |
| Diagnostics Check | OK    | 21:57 |

---

Data Download Date/Time 11/20/2011 08:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHEPPARD, KEVIN, B

BREATH TEST OPERATOR AGENCY: Williston PD

ARRESTING OFFICER: HUTSON, HARRY

ARRESTING OFFICER AGENCY: Chiefland PD

LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1187042

DRY GAS EXPIRY DATE: 07/22/2014

VIOLATION CODE: DUI

ARREST TIME: 22:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 11/06/2011 23:11:22

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:08 |                      |
| Air Blank         | 0.000 | 23:08 |                      |
| Control Test      | 0.079 | 23:09 |                      |
| Air Blank         | 0.000 | 23:09 |                      |
| Subject Sample #1 |       | 23:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:10 |                      |
| Air Blank         | 0.000 | 23:10 |                      |
| Control Test      | 0.079 | 23:10 |                      |
| Air Blank         | 0.000 | 23:11 |                      |
| Diagnostics Check | OK    | 23:11 |                      |

---

Data Download Date/Time 11/20/2011 08:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHEPPARD, KEVIN, B

BREATH TEST OPERATOR AGENCY: Williston PD

ARRESTING OFFICER: KING, MATTHEW

ARRESTING OFFICER AGENCY: Levy County SO

LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1187042

DRY GAS EXPIRY DATE: 07/22/2014

VIOLATION CODE: DUI

ARREST TIME: 22:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: STILL

DATE/TIME OF TEST: 11/09/2011 22:51:42

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:45 |
| Air Blank         | 0.000 | 22:45 |
| Control Test      | 0.079 | 22:46 |
| Air Blank         | 0.000 | 22:46 |
| Subject Sample #1 | 0.169 | 22:47 |
| Breath Volume     | 1.445 | 22:47 |
| Air Blank         | 0.000 | 22:47 |
| Air Blank         | 0.000 | 22:49 |
| Subject Sample #2 | 0.168 | 22:50 |
| Breath Volume     | 1.296 | 22:50 |
| Air Blank         | 0.000 | 22:50 |
| Control Test      | 0.079 | 22:51 |
| Air Blank         | 0.000 | 22:51 |
| Diagnostics Check | OK    | 22:51 |

---

Data Download Date/Time 11/20/2011 08:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHEPPARD, KEVIN, B

BREATH TEST OPERATOR AGENCY: Williston PD

ARRESTING OFFICER: SHEPPARD, KEVIN

ARRESTING OFFICER AGENCY: Williston PD

LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1187042

DRY GAS EXPIRY DATE: 07/22/2014

VIOLATION CODE: DUI

ARREST TIME: 19:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 11/10/2011 19:46:02

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:40 |
| Air Blank         | 0.000 | 19:40 |
| Control Test      | 0.079 | 19:40 |
| Air Blank         | 0.000 | 19:41 |
| Subject Sample #1 | 0.000 | 19:41 |
| Breath Volume     | 2.457 | 19:41 |
| Air Blank         | 0.000 | 19:42 |
| Air Blank         | 0.000 | 19:44 |
| Subject Sample #2 | 0.000 | 19:44 |
| Breath Volume     | 1.664 | 19:44 |
| Air Blank         | 0.000 | 19:45 |
| Control Test      | 0.078 | 19:45 |
| Air Blank         | 0.000 | 19:45 |
| Diagnostics Check | OK    | 19:46 |

---

Data Download Date/Time 11/20/2011 08:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHEPPARD, KEVIN, B

BREATH TEST OPERATOR AGENCY: Williston PD

ARRESTING OFFICER: SHEPPARD, KEVIN

ARRESTING OFFICER AGENCY: Williston PD

LAST AGENCY INSPECTION: 11/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1187042

DRY GAS EXPIRY DATE: 07/22/2014

VIOLATION CODE: DUI

ARREST TIME: 23:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: DYESS

DATE/TIME OF TEST: 11/28/2011 23:37:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:28 |                      |
| Air Blank         | 0.000 | 23:29 |                      |
| Control Test      | 0.078 | 23:29 |                      |
| Air Blank         | 0.000 | 23:30 |                      |
| Subject Sample #1 |       | 23:33 | No Sample Provided   |
| Breath Volume     | 0.000 | 23:33 |                      |
| Air Blank         | 0.000 | 23:33 |                      |
| Air Blank         | 0.000 | 23:35 |                      |
| Subject Sample #2 |       | 23:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:36 |                      |
| Air Blank         | 0.000 | 23:36 |                      |
| Control Test      | 0.079 | 23:36 |                      |
| Air Blank         | 0.000 | 23:37 |                      |
| Diagnostics Check | OK    | 23:37 |                      |

---

Data Download Date/Time 12/14/2011 13:44

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RITTENHOUSE, WILLIE, P  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RITTENHOUSE, WILLIE  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 11/30/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 18:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: KNOLL

DATE/TIME OF TEST: 11/30/2011 19:40:15

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:32 |
| Air Blank         | 0.000 | 19:32 |
| Control Test      | 0.077 | 19:32 |
| Air Blank         | 0.000 | 19:33 |
| Subject Sample #1 | 0.220 | 19:35 |
| Breath Volume     | 1.355 | 19:35 |
| Air Blank         | 0.000 | 19:36 |
| Air Blank         | 0.000 | 19:38 |
| Subject Sample #2 | 0.231 | 19:38 |
| Breath Volume     | 1.257 | 19:38 |
| Air Blank         | 0.000 | 19:39 |
| Control Test      | 0.076 | 19:39 |
| Air Blank         | 0.000 | 19:40 |
| Diagnostics Check | OK    | 19:40 |

---

Data Download Date/Time 12/30/2011 15:43

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GODWIN, AARON, D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: GODWIN, AARON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HOUSE

DATE/TIME OF TEST: 11/01/2011 01:37:30

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:27 |
| Air Blank         | 0.000 | 01:28 |
| Control Test      | 0.080 | 01:28 |
| Air Blank         | 0.000 | 01:29 |
| Subject Sample #1 | 0.201 | 01:29 |
| Breath Volume     | 1.300 | 01:29 |
| Air Blank         | 0.000 | 01:30 |
| Air Blank         | 0.000 | 01:32 |
| Subject Sample #2 | 0.227 | 01:32 |
| Breath Volume     | 1.753 | 01:32 |
| Air Blank         | 0.000 | 01:33 |
| Air Blank         | 0.000 | 01:35 |
| Subject Sample #3 | 0.212 | 01:35 |
| Breath Volume     | 1.484 | 01:35 |
| Air Blank         | 0.000 | 01:36 |
| Control Test      | 0.080 | 01:36 |
| Air Blank         | 0.000 | 01:37 |
| Diagnostics Check | OK    | 01:37 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON, DURON, A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: NELSON, DURON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 17:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: LUSSIER

DATE/TIME OF TEST: 11/01/2011 19:19:05

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 19:07 |                    |
| Air Blank         | 0.000 | 19:08 |                    |
| Control Test      | 0.080 | 19:08 |                    |
| Air Blank         | 0.000 | 19:08 |                    |
| Subject Sample #1 |       | 19:12 | No Sample Provided |
| Breath Volume     | 0.000 | 19:12 |                    |
| Air Blank         | 0.000 | 19:12 |                    |
| Air Blank         | 0.000 | 19:14 |                    |
| Subject Sample #2 |       | 19:17 | No Sample Provided |
| Breath Volume     | 0.000 | 19:17 |                    |
| Air Blank         | 0.000 | 19:18 |                    |
| Control Test      | 0.081 | 19:18 |                    |
| Air Blank         | 0.000 | 19:19 |                    |
| Diagnostics Check | OK    | 19:19 |                    |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOZARD, CHRYSTAL, D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BOZARD, CHRYSTAL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: HENDRIX

DATE/TIME OF TEST: 11/02/2011 23:35:24

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 23:24 |                    |
| Air Blank         | 0.000 | 23:24 |                    |
| Control Test      | 0.080 | 23:24 |                    |
| Air Blank         | 0.000 | 23:25 |                    |
| Subject Sample #1 |       | 23:28 | No Sample Provided |
| Breath Volume     | 0.000 | 23:28 |                    |
| Air Blank         | 0.000 | 23:28 |                    |
| Air Blank         | 0.000 | 23:30 |                    |
| Subject Sample #2 |       | 23:34 | No Sample Provided |
| Breath Volume     | 0.000 | 23:34 |                    |
| Air Blank         | 0.000 | 23:34 |                    |
| Control Test      | 0.080 | 23:34 |                    |
| Air Blank         | 0.000 | 23:35 |                    |
| Diagnostics Check | OK    | 23:35 |                    |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHLEIFER, DION, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SCHLEIFER, DION  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KEOWN

DATE/TIME OF TEST: 11/05/2011 23:18:57

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:12 |
| Air Blank         | 0.000 | 23:12 |
| Control Test      | 0.080 | 23:13 |
| Air Blank         | 0.000 | 23:13 |
| Subject Sample #1 | 0.119 | 23:14 |
| Breath Volume     | 1.878 | 23:14 |
| Air Blank         | 0.000 | 23:15 |
| Air Blank         | 0.000 | 23:16 |
| Subject Sample #2 | 0.115 | 23:17 |
| Breath Volume     | 1.839 | 23:17 |
| Air Blank         | 0.000 | 23:18 |
| Control Test      | 0.080 | 23:18 |
| Air Blank         | 0.000 | 23:18 |
| Diagnostics Check | OK    | 23:18 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILLIPS, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: PHILLIPS, SAMUEL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 11/06/2011 01:59:24

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:53 |
| Air Blank         | 0.000 | 01:53 |
| Control Test      | 0.080 | 01:54 |
| Air Blank         | 0.000 | 01:54 |
| Subject Sample #1 | 0.126 | 01:54 |
| Breath Volume     | 1.285 | 01:54 |
| Air Blank         | 0.000 | 01:55 |
| Air Blank         | 0.000 | 01:57 |
| Subject Sample #2 | 0.133 | 01:57 |
| Breath Volume     | 1.191 | 01:57 |
| Air Blank         | 0.000 | 01:58 |
| Control Test      | 0.080 | 01:58 |
| Air Blank         | 0.000 | 01:59 |
| Diagnostics Check | OK    | 01:59 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PAUL, JAYSON, R

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: PAUL, JAYSON

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 12:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: BONOYER

DATE/TIME OF TEST: 11/08/2011 13:31:37

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 13:22 |
| Air Blank         | 0.000 | 13:23 |
| Control Test      | 0.080 | 13:23 |
| Air Blank         | 0.000 | 13:24 |
| Subject Sample #1 | 0.000 | 13:25 |
| Breath Volume     | 1.714 | 13:25 |
| Air Blank         | 0.000 | 13:26 |
| Air Blank         | 0.000 | 13:28 |
| Subject Sample #2 | 0.000 | 13:30 |
| Breath Volume     | 1.585 | 13:30 |
| Air Blank         | 0.000 | 13:30 |
| Control Test      | 0.079 | 13:31 |
| Air Blank         | 0.000 | 13:31 |
| Diagnostics Check | OK    | 13:31 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEICHNER, ADAM, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: TEICHNER, ADAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 18:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: PURVIS

DATE/TIME OF TEST: 11/08/2011 19:32:12

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:25 |
| Air Blank         | 0.000 | 19:26 |
| Control Test      | 0.078 | 19:26 |
| Air Blank         | 0.000 | 19:26 |
| Subject Sample #1 | 0.214 | 19:27 |
| Breath Volume     | 1.820 | 19:27 |
| Air Blank         | 0.000 | 19:28 |
| Air Blank         | 0.000 | 19:29 |
| Subject Sample #2 | 0.205 | 19:30 |
| Breath Volume     | 1.906 | 19:30 |
| Air Blank         | 0.000 | 19:31 |
| Control Test      | 0.079 | 19:31 |
| Air Blank         | 0.000 | 19:32 |
| Diagnostics Check | OK    | 19:32 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KELLY, DAVID, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KELLY, DAVID

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 03:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: QUACKENBUSH

DATE/TIME OF TEST: 11/12/2011 04:02:32

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 03:58 |                  |
| Air Blank         | 0.000 | 03:59 |                  |
| Control Test      | 0.081 | 03:59 |                  |
| Air Blank         | 0.000 | 03:59 |                  |
| Subject Sample #1 |       | 04:02 | Sequence Aborted |
| Breath Volume     | 0.000 | 04:02 |                  |
| Air Blank         | 0.000 | 04:02 |                  |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLY, DAVID, S  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: KELLY, DAVID  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: QUACKENBUSH

DATE/TIME OF TEST: 11/12/2011 04:14:15

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:06 |
| Air Blank         | 0.000 | 04:07 |
| Control Test      | 0.080 | 04:07 |
| Air Blank         | 0.000 | 04:08 |
| Subject Sample #1 | 0.171 | 04:09 |
| Breath Volume     | 1.316 | 04:09 |
| Air Blank         | 0.000 | 04:09 |
| Air Blank         | 0.000 | 04:11 |
| Subject Sample #2 | 0.168 | 04:12 |
| Breath Volume     | 1.230 | 04:12 |
| Air Blank         | 0.000 | 04:13 |
| Control Test      | 0.081 | 04:13 |
| Air Blank         | 0.000 | 04:14 |
| Diagnostics Check | OK    | 04:14 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON, BRADLEY, D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HARRISON, BRADLEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: NOA

DATE/TIME OF TEST: 11/18/2011 03:27:18

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:19 |
| Air Blank         | 0.000 | 03:19 |
| Control Test      | 0.081 | 03:20 |
| Air Blank         | 0.000 | 03:20 |
| Subject Sample #1 | 0.209 | 03:22 |
| Breath Volume     | 4.078 | 03:22 |
| Air Blank         | 0.000 | 03:22 |
| Air Blank         | 0.000 | 03:24 |
| Subject Sample #2 | 0.198 | 03:25 |
| Breath Volume     | 3.589 | 03:25 |
| Air Blank         | 0.000 | 03:26 |
| Control Test      | 0.081 | 03:26 |
| Air Blank         | 0.000 | 03:27 |
| Diagnostics Check | OK    | 03:27 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILLIPS, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WALL, DAVID  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 11/19/2011 04:41:43

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 04:40 |                           |
| Air Blank         | 0.000 | 04:40 |                           |
| Control Test      | 0.128 | 04:40 | Control Outside Tolerance |
| Air Blank         | 0.000 | 04:41 |                           |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILLIPS, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WALLS, DAVID  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 11/19/2011 04:50:29

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:44 |
| Air Blank         | 0.000 | 04:44 |
| Control Test      | 0.079 | 04:44 |
| Air Blank         | 0.000 | 04:45 |
| Subject Sample #1 | 0.145 | 04:46 |
| Breath Volume     | 2.074 | 04:46 |
| Air Blank         | 0.000 | 04:46 |
| Air Blank         | 0.000 | 04:48 |
| Subject Sample #2 | 0.143 | 04:48 |
| Breath Volume     | 3.445 | 04:48 |
| Air Blank         | 0.000 | 04:49 |
| Control Test      | 0.080 | 04:49 |
| Air Blank         | 0.000 | 04:50 |
| Diagnostics Check | OK    | 04:50 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAUGHTRY, NILES, D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: DAUGHTRY, NILES  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TRAVO

DATE/TIME OF TEST: 11/19/2011 22:39:15

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:33 |
| Air Blank         | 0.000 | 22:33 |
| Control Test      | 0.080 | 22:33 |
| Air Blank         | 0.000 | 22:34 |
| Subject Sample #1 | 0.000 | 22:34 |
| Breath Volume     | 4.183 | 22:34 |
| Air Blank         | 0.000 | 22:35 |
| Air Blank         | 0.000 | 22:37 |
| Subject Sample #2 | 0.000 | 22:37 |
| Breath Volume     | 3.898 | 22:37 |
| Air Blank         | 0.000 | 22:38 |
| Control Test      | 0.078 | 22:38 |
| Air Blank         | 0.000 | 22:39 |
| Diagnostics Check | OK    | 22:39 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE, JOSHUA, J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WHITE, JOSHUA  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: YARBROUGH

DATE/TIME OF TEST: 11/20/2011 21:36:46

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 21:27 |
| Air Blank         | 0.000 | 21:28 |
| Control Test      | 0.080 | 21:28 |
| Air Blank         | 0.000 | 21:29 |
| Subject Sample #1 | 0.196 | 21:31 |
| Breath Volume     | 1.648 | 21:31 |
| Air Blank         | 0.000 | 21:32 |
| Air Blank         | 0.000 | 21:34 |
| Subject Sample #2 | 0.206 | 21:35 |
| Breath Volume     | 1.546 | 21:35 |
| Air Blank         | 0.000 | 21:35 |
| Control Test      | 0.078 | 21:36 |
| Air Blank         | 0.000 | 21:36 |
| Diagnostics Check | OK    | 21:36 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DAUGHTRY, NILES, D

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: DAUGHTRY, NILES

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 21:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: BELCHER

DATE/TIME OF TEST: 11/20/2011 22:36:38

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:30 |
| Air Blank         | 0.000 | 22:30 |
| Control Test      | 0.079 | 22:31 |
| Air Blank         | 0.000 | 22:31 |
| Subject Sample #1 | 0.000 | 22:32 |
| Breath Volume     | 1.738 | 22:32 |
| Air Blank         | 0.000 | 22:32 |
| Air Blank         | 0.000 | 22:34 |
| Subject Sample #2 | 0.000 | 22:35 |
| Breath Volume     | 1.398 | 22:35 |
| Air Blank         | 0.000 | 22:35 |
| Control Test      | 0.078 | 22:36 |
| Air Blank         | 0.000 | 22:36 |
| Diagnostics Check | OK    | 22:36 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIRD, LANCE, R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BIRD, LANCE  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 11/24/2011 23:25:53

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:19 |
| Air Blank         | 0.000 | 23:20 |
| Control Test      | 0.079 | 23:20 |
| Air Blank         | 0.000 | 23:21 |
| Subject Sample #1 | 0.000 | 23:21 |
| Breath Volume     | 3.265 | 23:21 |
| Air Blank         | 0.000 | 23:22 |
| Air Blank         | 0.000 | 23:24 |
| Subject Sample #2 | 0.000 | 23:24 |
| Breath Volume     | 3.335 | 23:24 |
| Air Blank         | 0.000 | 23:24 |
| Control Test      | 0.077 | 23:25 |
| Air Blank         | 0.000 | 23:25 |
| Diagnostics Check | OK    | 23:25 |

---

Data Download Date/Time 11/30/2011 14:56

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON, BRADLEY, D  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WALLS, DAVID  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/26/2011 02:15:12

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:08 |
| Air Blank         | 0.000 | 02:09 |
| Control Test      | 0.080 | 02:09 |
| Air Blank         | 0.000 | 02:09 |
| Subject Sample #1 | 0.000 | 02:10 |
| Breath Volume     | 3.507 | 02:10 |
| Air Blank         | 0.000 | 02:11 |
| Air Blank         | 0.000 | 02:13 |
| Subject Sample #2 | 0.000 | 02:13 |
| Breath Volume     | 3.945 | 02:13 |
| Air Blank         | 0.000 | 02:14 |
| Control Test      | 0.079 | 02:14 |
| Air Blank         | 0.000 | 02:15 |
| Diagnostics Check | OK    | 02:15 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROPER, DAVID, C

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: ROPER, DAVID

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 15:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: HARVILLE

DATE/TIME OF TEST: 11/26/2011 16:38:21

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 16:31 |
| Air Blank         | 0.000 | 16:32 |
| Control Test      | 0.079 | 16:32 |
| Air Blank         | 0.000 | 16:32 |
| Subject Sample #1 | 0.111 | 16:33 |
| Breath Volume     | 3.421 | 16:33 |
| Air Blank         | 0.000 | 16:34 |
| Air Blank         | 0.000 | 16:36 |
| Subject Sample #2 | 0.116 | 16:36 |
| Breath Volume     | 3.316 | 16:36 |
| Air Blank         | 0.000 | 16:37 |
| Control Test      | 0.078 | 16:37 |
| Air Blank         | 0.000 | 16:38 |
| Diagnostics Check | OK    | 16:38 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILLIPS, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: PHILLIPS, SAMUEL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 11/27/2011 01:26:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:23 |                      |
| Air Blank         | 0.000 | 01:23 |                      |
| Control Test      | 0.079 | 01:24 |                      |
| Air Blank         | 0.000 | 01:24 |                      |
| Subject Sample #1 |       | 01:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:25 |                      |
| Air Blank         | 0.000 | 01:25 |                      |
| Control Test      | 0.079 | 01:25 |                      |
| Air Blank         | 0.000 | 01:26 |                      |
| Diagnostics Check | OK    | 01:26 |                      |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, CLAYTON, M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SMITH, CLAYTON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TOMBLIN

DATE/TIME OF TEST: 11/28/2011 02:33:08

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:30 |                      |
| Air Blank         | 0.000 | 02:30 |                      |
| Control Test      | 0.079 | 02:31 |                      |
| Air Blank         | 0.000 | 02:31 |                      |
| Subject Sample #1 |       | 02:31 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:31 |                      |
| Air Blank         | 0.000 | 02:32 |                      |
| Control Test      | 0.080 | 02:32 |                      |
| Air Blank         | 0.000 | 02:33 |                      |
| Diagnostics Check | OK    | 02:33 |                      |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK, STEPHEN, F  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CLARK, STEPHEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 11/29/2011 00:37:15

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:30 |
| Air Blank         | 0.000 | 00:31 |
| Control Test      | 0.080 | 00:31 |
| Air Blank         | 0.000 | 00:32 |
| Subject Sample #1 | 0.201 | 00:32 |
| Breath Volume     | 1.941 | 00:32 |
| Air Blank         | 0.000 | 00:33 |
| Air Blank         | 0.000 | 00:35 |
| Subject Sample #2 | 0.204 | 00:35 |
| Breath Volume     | 2.140 | 00:35 |
| Air Blank         | 0.000 | 00:36 |
| Control Test      | 0.079 | 00:36 |
| Air Blank         | 0.000 | 00:37 |
| Diagnostics Check | OK    | 00:37 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENNELL, DAVID, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: RICHARDS, ANDREW  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GILLMAN

DATE/TIME OF TEST: 11/30/2011 03:12:27

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:06 |
| Air Blank         | 0.000 | 03:06 |
| Control Test      | 0.078 | 03:06 |
| Air Blank         | 0.000 | 03:07 |
| Subject Sample #1 | 0.134 | 03:07 |
| Breath Volume     | 2.093 | 03:07 |
| Air Blank         | 0.000 | 03:08 |
| Air Blank         | 0.000 | 03:10 |
| Subject Sample #2 | 0.132 | 03:10 |
| Breath Volume     | 2.343 | 03:10 |
| Air Blank         | 0.000 | 03:11 |
| Control Test      | 0.077 | 03:11 |
| Air Blank         | 0.000 | 03:12 |
| Diagnostics Check | OK    | 03:12 |

---

Data Download Date/Time 11/30/2011 14:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TARTER, JEFFREY, A

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: TARTER, JEFFREY

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 01:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 11/03/2011 01:45:35

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:39 |
| Air Blank         | 0.000 | 01:39 |
| Control Test      | 0.079 | 01:40 |
| Air Blank         | 0.000 | 01:40 |
| Subject Sample #1 | 0.122 | 01:41 |
| Breath Volume     | 3.585 | 01:41 |
| Air Blank         | 0.000 | 01:41 |
| Air Blank         | 0.000 | 01:43 |
| Subject Sample #2 | 0.121 | 01:44 |
| Breath Volume     | 3.597 | 01:44 |
| Air Blank         | 0.000 | 01:44 |
| Control Test      | 0.078 | 01:44 |
| Air Blank         | 0.000 | 01:45 |
| Diagnostics Check | OK    | 01:45 |

---

Data Download Date/Time 11/20/2011 14:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COWAN, RICHARD, T  
BREATH TEST OPERATOR AGENCY: Jackson County SO  
ARRESTING OFFICER: COWAN, RICHARD  
ARRESTING OFFICER AGENCY: Jackson County SO  
LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG025702  
DRY GAS EXPIRY DATE: 09/14/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: BENFIELD

DATE/TIME OF TEST: 11/05/2011 06:43:04

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 06:36 |
| Air Blank         | 0.000 | 06:37 |
| Control Test      | 0.078 | 06:37 |
| Air Blank         | 0.000 | 06:38 |
| Subject Sample #1 | 0.179 | 06:38 |
| Breath Volume     | 2.089 | 06:38 |
| Air Blank         | 0.000 | 06:39 |
| Air Blank         | 0.000 | 06:41 |
| Subject Sample #2 | 0.180 | 06:41 |
| Breath Volume     | 2.453 | 06:41 |
| Air Blank         | 0.000 | 06:42 |
| Control Test      | 0.078 | 06:42 |
| Air Blank         | 0.000 | 06:42 |
| Diagnostics Check | OK    | 06:43 |

---

Data Download Date/Time 11/20/2011 14:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARNLEY, JARED, K

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: CARNLEY, JARED

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 10:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 11/11/2011 23:57:01

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:50 |
| Air Blank         | 0.000 | 23:51 |
| Control Test      | 0.078 | 23:51 |
| Air Blank         | 0.000 | 23:51 |
| Subject Sample #1 | 0.131 | 23:52 |
| Breath Volume     | 4.304 | 23:52 |
| Air Blank         | 0.000 | 23:53 |
| Air Blank         | 0.000 | 23:55 |
| Subject Sample #2 | 0.128 | 23:55 |
| Breath Volume     | 3.015 | 23:55 |
| Air Blank         | 0.000 | 23:56 |
| Control Test      | 0.079 | 23:56 |
| Air Blank         | 0.000 | 23:56 |
| Diagnostics Check | OK    | 23:57 |

---

Data Download Date/Time 11/20/2011 14:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TARTER, JEFFREY, A

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: TARTER, JEFFREY

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 22:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: RAINGE

DATE/TIME OF TEST: 11/12/2011 23:28:28

---

|                   |       |       |               |
|-------------------|-------|-------|---------------|
| Diagnostics Check | OK    | 23:22 |               |
| Air Blank         | 0.000 | 23:22 |               |
| Control Test      | 0.079 | 23:22 |               |
| Air Blank         | 0.000 | 23:23 |               |
| Subject Sample #1 | 0.187 | 23:23 |               |
| Breath Volume     | 2.937 | 23:23 |               |
| Air Blank         | 0.000 | 23:24 |               |
| Air Blank         | 0.000 | 23:26 |               |
| Subject Sample #2 |       | 23:26 | Slope Not Met |
| Breath Volume     | 1.820 | 23:26 |               |
| Air Blank         | 0.000 | 23:27 |               |
| Control Test      | 0.079 | 23:27 |               |
| Air Blank         | 0.000 | 23:28 |               |
| Diagnostics Check | OK    | 23:28 |               |

---

Data Download Date/Time 11/20/2011 14:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: OWENS, GEORGE, D

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: OWENS, GEORGE

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 21:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 11/17/2011 22:48:48

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:42 |
| Air Blank         | 0.000 | 22:43 |
| Control Test      | 0.079 | 22:43 |
| Air Blank         | 0.000 | 22:43 |
| Subject Sample #1 | 0.159 | 22:44 |
| Breath Volume     | 3.675 | 22:44 |
| Air Blank         | 0.000 | 22:44 |
| Air Blank         | 0.000 | 22:46 |
| Subject Sample #2 | 0.164 | 22:47 |
| Breath Volume     | 3.625 | 22:47 |
| Air Blank         | 0.000 | 22:47 |
| Control Test      | 0.078 | 22:48 |
| Air Blank         | 0.000 | 22:48 |
| Diagnostics Check | OK    | 22:48 |

---

Data Download Date/Time 11/20/2011 14:54

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARNLEY, JASON, S  
BREATH TEST OPERATOR AGENCY: Marianna PD  
ARRESTING OFFICER: SCARBOROUGH, TYLER  
ARRESTING OFFICER AGENCY: Marianna PD  
LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG025702  
DRY GAS EXPIRY DATE: 09/14/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 11/19/2011 02:14:07

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 02:13 |                 |
| Air Blank         |       | 02:13 | Improper Sample |
| Air Blank         | 0.000 | 02:14 |                 |

---

|                         |            |       |
|-------------------------|------------|-------|
| Data Download Date/Time | 11/20/2011 | 14:54 |
|-------------------------|------------|-------|

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARNLEY, JASON, S  
BREATH TEST OPERATOR AGENCY: Marianna PD  
ARRESTING OFFICER: SCARBOROUGH, TYLER  
ARRESTING OFFICER AGENCY: Marianna PD  
LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG025702  
DRY GAS EXPIRY DATE: 09/14/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 11/19/2011 02:24:49

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:18 |
| Air Blank         | 0.000 | 02:19 |
| Control Test      | 0.079 | 02:19 |
| Air Blank         | 0.000 | 02:19 |
| Subject Sample #1 | 0.169 | 02:20 |
| Breath Volume     | 2.765 | 02:20 |
| Air Blank         | 0.000 | 02:20 |
| Air Blank         | 0.000 | 02:22 |
| Subject Sample #2 | 0.172 | 02:23 |
| Breath Volume     | 3.414 | 02:23 |
| Air Blank         | 0.000 | 02:23 |
| Control Test      | 0.079 | 02:24 |
| Air Blank         | 0.000 | 02:24 |
| Diagnostics Check | OK    | 02:24 |

---

Data Download Date/Time 11/20/2011 14:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARNLEY, JASON, S

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: CARNLEY, JASON

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 10/23/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 03:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 11/19/2011 04:24:16

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 04:11 |                |
| Air Blank         | 0.000 | 04:12 |                |
| Control Test      | 0.079 | 04:12 |                |
| Air Blank         | 0.000 | 04:13 |                |
| Subject Sample #1 | 0.285 | 04:16 | Volume Not Met |
| Breath Volume     | 0.054 | 04:16 |                |
| Air Blank         | 0.000 | 04:17 |                |
| Air Blank         | 0.000 | 04:18 |                |
| Subject Sample #2 | 0.304 | 04:19 |                |
| Breath Volume     | 2.855 | 04:19 |                |
| Air Blank         | 0.000 | 04:20 |                |
| Air Blank         | 0.000 | 04:22 |                |
| Subject Sample #3 | 0.304 | 04:22 |                |
| Breath Volume     | 1.687 | 04:22 |                |
| Air Blank         | 0.000 | 04:23 |                |
| Control Test      | 0.076 | 04:23 |                |
| Air Blank         | 0.000 | 04:24 |                |
| Diagnostics Check | OK    | 04:24 |                |

---

Data Download Date/Time 11/20/2011 14:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: OWENS, GEORGE, D

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: STROUD, NICHOLAS

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 11/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 02:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VICTORIANO

DATE/TIME OF TEST: 11/22/2011 02:55:27

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:49 |
| Air Blank         | 0.000 | 02:49 |
| Control Test      | 0.078 | 02:50 |
| Air Blank         | 0.000 | 02:50 |
| Subject Sample #1 | 0.160 | 02:50 |
| Breath Volume     | 3.105 | 02:50 |
| Air Blank         | 0.000 | 02:51 |
| Air Blank         | 0.000 | 02:53 |
| Subject Sample #2 | 0.162 | 02:53 |
| Breath Volume     | 3.855 | 02:53 |
| Air Blank         | 0.000 | 02:54 |
| Control Test      | 0.078 | 02:54 |
| Air Blank         | 0.000 | 02:55 |
| Diagnostics Check | OK    | 02:55 |

---

Data Download Date/Time 11/29/2011 11:40

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ELMORE, EDDIE, W

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: ELMORE, EDDIE

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 11/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 21:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/25/2011 20:50:48

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 20:44 |
| Air Blank         | 0.000 | 20:44 |
| Control Test      | 0.079 | 20:45 |
| Air Blank         | 0.000 | 20:45 |
| Subject Sample #1 | 0.155 | 20:46 |
| Breath Volume     | 4.535 | 20:46 |
| Air Blank         | 0.000 | 20:47 |
| Air Blank         | 0.000 | 20:48 |
| Subject Sample #2 | 0.147 | 20:49 |
| Breath Volume     | 3.730 | 20:49 |
| Air Blank         | 0.000 | 20:49 |
| Control Test      | 0.078 | 20:50 |
| Air Blank         | 0.000 | 20:50 |
| Diagnostics Check | OK    | 20:50 |

---

Data Download Date/Time 11/29/2011 11:40

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARNLEY, JARED, K  
BREATH TEST OPERATOR AGENCY: Marianna PD  
ARRESTING OFFICER: MITCHELL, WESLEY  
ARRESTING OFFICER AGENCY: Jackson County SO  
LAST AGENCY INSPECTION: 11/20/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG025702  
DRY GAS EXPIRY DATE: 09/14/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 11/27/2011 01:03:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:00 |                      |
| Air Blank         | 0.000 | 01:01 |                      |
| Control Test      | 0.079 | 01:01 |                      |
| Air Blank         | 0.000 | 01:02 |                      |
| Subject Sample #1 |       | 01:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:02 |                      |
| Air Blank         | 0.000 | 01:03 |                      |
| Control Test      | 0.077 | 01:03 |                      |
| Air Blank         | 0.000 | 01:03 |                      |
| Diagnostics Check | OK    | 01:03 |                      |

---

Data Download Date/Time 11/29/2011 11:40

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ELMORE, EDDIE, L

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: ELMORE, EDDIE

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 11/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 14:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: JESSICA

DATE/TIME OF TEST: 11/27/2011 15:01:16

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 14:55 |
| Air Blank         | 0.000 | 14:55 |
| Control Test      | 0.078 | 14:55 |
| Air Blank         | 0.000 | 14:56 |
| Subject Sample #1 | 0.314 | 14:56 |
| Breath Volume     | 3.121 | 14:56 |
| Air Blank         | 0.000 | 14:57 |
| Air Blank         | 0.000 | 14:59 |
| Subject Sample #2 | 0.325 | 14:59 |
| Breath Volume     | 2.859 | 14:59 |
| Air Blank         | 0.000 | 15:00 |
| Control Test      | 0.078 | 15:00 |
| Air Blank         | 0.000 | 15:01 |
| Diagnostics Check | OK    | 15:01 |

---

Data Download Date/Time 11/29/2011 11:40

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUBER, PHILLIP, W  
BREATH TEST OPERATOR AGENCY: FDLE  
ARRESTING OFFICER: SMITH, JIM  
ARRESTING OFFICER AGENCY: FDLE  
LAST AGENCY INSPECTION: 11/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG123001  
DRY GAS EXPIRY DATE: 08/18/2013  
VIOLATION CODE: OTHER  
ARREST TIME: 12:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/30/2011 14:25:23

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 14:18 |
| Air Blank         | 0.000 | 14:19 |
| Control Test      | 0.079 | 14:19 |
| Air Blank         | 0.000 | 14:20 |
| Subject Sample #1 | 0.082 | 14:20 |
| Breath Volume     | 1.285 | 14:20 |
| Air Blank         | 0.000 | 14:21 |
| Air Blank         | 0.000 | 14:23 |
| Subject Sample #2 | 0.083 | 14:23 |
| Breath Volume     | 1.265 | 14:23 |
| Air Blank         | 0.000 | 14:24 |
| Control Test      | 0.078 | 14:24 |
| Air Blank         | 0.000 | 14:25 |
| Diagnostics Check | OK    | 14:25 |

---

Data Download Date/Time 12/03/2011 12:39

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SUBER, PHILLIP, W

BREATH TEST OPERATOR AGENCY: FDLE

ARRESTING OFFICER: SMITH, JIM

ARRESTING OFFICER AGENCY: FDLE

LAST AGENCY INSPECTION: 11/29/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG123001

DRY GAS EXPIRY DATE: 08/18/2013

VIOLATION CODE: OTHER

ARREST TIME: 13:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/30/2011 14:43:22

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 14:36 |
| Air Blank         | 0.000 | 14:36 |
| Control Test      | 0.078 | 14:37 |
| Air Blank         | 0.000 | 14:37 |
| Subject Sample #1 | 0.083 | 14:38 |
| Breath Volume     | 1.839 | 14:38 |
| Air Blank         | 0.000 | 14:39 |
| Air Blank         | 0.000 | 14:41 |
| Subject Sample #2 | 0.083 | 14:41 |
| Breath Volume     | 1.683 | 14:41 |
| Air Blank         | 0.000 | 14:42 |
| Control Test      | 0.078 | 14:42 |
| Air Blank         | 0.000 | 14:43 |
| Diagnostics Check | OK    | 14:43 |

---

Data Download Date/Time 12/03/2011 12:39

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS, DAVID, C  
BREATH TEST OPERATOR AGENCY: Madison County SO  
ARRESTING OFFICER: MYERS, DAVID  
ARRESTING OFFICER AGENCY: Madison County SO  
LAST AGENCY INSPECTION: 11/04/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 05911080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 11/08/2011 23:47:48

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:41 |
| Air Blank         | 0.000 | 23:42 |
| Control Test      | 0.078 | 23:42 |
| Air Blank         | 0.000 | 23:42 |
| Subject Sample #1 | 0.129 | 23:43 |
| Breath Volume     | 3.628 | 23:43 |
| Air Blank         | 0.000 | 23:43 |
| Air Blank         | 0.000 | 23:45 |
| Subject Sample #2 | 0.128 | 23:46 |
| Breath Volume     | 3.398 | 23:46 |
| Air Blank         | 0.000 | 23:46 |
| Control Test      | 0.076 | 23:47 |
| Air Blank         | 0.000 | 23:47 |
| Diagnostics Check | OK    | 23:47 |

---

Data Download Date/Time 12/08/2011 10:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS, DAVID, C  
BREATH TEST OPERATOR AGENCY: Madison County SO  
ARRESTING OFFICER: MYERS, DAVID  
ARRESTING OFFICER AGENCY: Madison County SO  
LAST AGENCY INSPECTION: 11/04/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05911080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: OTHER  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: RAWLINS

DATE/TIME OF TEST: 11/17/2011 22:39:12

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:32 |
| Air Blank         | 0.000 | 22:33 |
| Control Test      | 0.077 | 22:33 |
| Air Blank         | 0.000 | 22:33 |
| Subject Sample #1 | 0.201 | 22:34 |
| Breath Volume     | 2.835 | 22:34 |
| Air Blank         | 0.000 | 22:35 |
| Air Blank         | 0.000 | 22:37 |
| Subject Sample #2 | 0.193 | 22:37 |
| Breath Volume     | 3.496 | 22:37 |
| Air Blank         | 0.000 | 22:38 |
| Control Test      | 0.076 | 22:38 |
| Air Blank         | 0.000 | 22:39 |
| Diagnostics Check | OK    | 22:39 |

---

Data Download Date/Time 12/08/2011 10:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SWINDLE, CHARLES, M

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: SHAW, PJ

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 11/04/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 05911080A2

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 21:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: FRYFOGLE

DATE/TIME OF TEST: 11/19/2011 23:09:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:06 |                      |
| Air Blank         | 0.000 | 23:06 |                      |
| Control Test      | 0.076 | 23:07 |                      |
| Air Blank         | 0.000 | 23:07 |                      |
| Subject Sample #1 |       | 23:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:07 |                      |
| Air Blank         | 0.000 | 23:08 |                      |
| Control Test      | 0.075 | 23:08 |                      |
| Air Blank         | 0.000 | 23:09 |                      |
| Diagnostics Check | OK    | 23:09 |                      |

---

Data Download Date/Time 12/08/2011 10:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEST, KATHRINE, B  
BREATH TEST OPERATOR AGENCY: Madison PD  
ARRESTING OFFICER: SMITH, JOEY  
ARRESTING OFFICER AGENCY: Madison PD  
LAST AGENCY INSPECTION: 11/04/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05911080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 19:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 11/24/2011 20:20:08

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 20:08 |                           |
| Air Blank         | 0.000 | 20:08 |                           |
| Control Test      | 0.076 | 20:09 |                           |
| Air Blank         | 0.000 | 20:09 |                           |
| Subject Sample #1 | 0.314 | 20:12 | Volume Not Met            |
| Breath Volume     | 0.457 | 20:12 |                           |
| Air Blank         | 0.000 | 20:13 |                           |
| Air Blank         | 0.000 | 20:15 |                           |
| Subject Sample #2 | 0.306 | 20:18 | Volume Not Met            |
| Breath Volume     | 0.093 | 20:18 |                           |
| Air Blank         | 0.000 | 20:19 |                           |
| Control Test      | 0.074 | 20:19 | Control Outside Tolerance |
| Air Blank         | 0.000 | 20:20 |                           |

---

Data Download Date/Time 12/08/2011 10:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE, CARLON, F  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: SAINT, RODERICK  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 15:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/03/2011 17:21:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:10 |                      |
| Air Blank         | 0.000 | 17:10 |                      |
| Control Test      | 0.080 | 17:11 |                      |
| Air Blank         | 0.000 | 17:11 |                      |
| Subject Sample #1 | 0.198 | 17:14 | Volume Not Met       |
| Breath Volume     | 0.699 | 17:14 |                      |
| Air Blank         | 0.000 | 17:15 |                      |
| Air Blank         | 0.000 | 17:17 |                      |
| Subject Sample #2 |       | 17:19 | Subject Test Refused |
| Breath Volume     | 0.367 | 17:19 |                      |
| Air Blank         | 0.000 | 17:20 |                      |
| Control Test      | 0.077 | 17:20 |                      |
| Air Blank         | 0.000 | 17:21 |                      |
| Diagnostics Check | OK    | 17:21 |                      |

---

Data Download Date/Time 11/29/2011 16:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEWELL, RICKY,  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: KLINENSMITH, J  
ARRESTING OFFICER AGENCY: Parker PD  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 11/04/2011 00:15:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:12 |                      |
| Air Blank         | 0.000 | 00:13 |                      |
| Control Test      | 0.080 | 00:13 |                      |
| Air Blank         | 0.000 | 00:14 |                      |
| Subject Sample #1 |       | 00:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:14 |                      |
| Air Blank         | 0.000 | 00:14 |                      |
| Control Test      | 0.080 | 00:14 |                      |
| Air Blank         | 0.000 | 00:15 |                      |
| Diagnostics Check | OK    | 00:15 |                      |

---

Data Download Date/Time 11/29/2011 16:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOPKINS, DAVID, J  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: WILLIAMS, MURRAY  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: CRILLY

DATE/TIME OF TEST: 11/11/2011 23:00:52

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:52 |
| Air Blank         | 0.000 | 22:52 |
| Control Test      | 0.080 | 22:53 |
| Air Blank         | 0.000 | 22:53 |
| Subject Sample #1 | 0.212 | 22:55 |
| Breath Volume     | 2.386 | 22:55 |
| Air Blank         | 0.000 | 22:56 |
| Air Blank         | 0.000 | 22:58 |
| Subject Sample #2 | 0.199 | 22:59 |
| Breath Volume     | 1.988 | 22:59 |
| Air Blank         | 0.000 | 22:59 |
| Control Test      | 0.080 | 23:00 |
| Air Blank         | 0.000 | 23:00 |
| Diagnostics Check | OK    | 23:00 |

---

Data Download Date/Time 11/29/2011 16:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOPKINS, DAVID, J  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: SUMERALL, JON  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BOONE

DATE/TIME OF TEST: 11/17/2011 04:07:37

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:01 |
| Air Blank         | 0.000 | 04:01 |
| Control Test      | 0.080 | 04:01 |
| Air Blank         | 0.000 | 04:02 |
| Subject Sample #1 | 0.123 | 04:03 |
| Breath Volume     | 4.644 | 04:03 |
| Air Blank         | 0.000 | 04:03 |
| Air Blank         | 0.000 | 04:05 |
| Subject Sample #2 | 0.124 | 04:06 |
| Breath Volume     | 3.824 | 04:06 |
| Air Blank         | 0.000 | 04:06 |
| Control Test      | 0.080 | 04:07 |
| Air Blank         | 0.000 | 04:07 |
| Diagnostics Check | OK    | 04:07 |

---

Data Download Date/Time 11/29/2011 16:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOODWIN, MICHAEL, P  
BREATH TEST OPERATOR AGENCY: OTHER State  
ARRESTING OFFICER: GOODWIN, MICHAEL  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 11/21/2011 23:17:32

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:06 |
| Air Blank         | 0.000 | 23:07 |
| Control Test      | 0.081 | 23:07 |
| Air Blank         | 0.000 | 23:07 |
| Subject Sample #1 | 0.188 | 23:10 |
| Breath Volume     | 1.550 | 23:10 |
| Air Blank         | 0.000 | 23:11 |
| Air Blank         | 0.000 | 23:13 |
| Subject Sample #2 | 0.186 | 23:15 |
| Breath Volume     | 1.445 | 23:15 |
| Air Blank         | 0.000 | 23:16 |
| Control Test      | 0.079 | 23:16 |
| Air Blank         | 0.000 | 23:17 |
| Diagnostics Check | OK    | 23:17 |

---

Data Download Date/Time 11/29/2011 16:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELCH, MARY, E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: ALGER, DEPUTY  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: TERRELL JR

DATE/TIME OF TEST: 11/24/2011 00:43:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:38 |                      |
| Air Blank         | 0.000 | 00:38 |                      |
| Control Test      | 0.080 | 00:39 |                      |
| Air Blank         | 0.000 | 00:39 |                      |
| Subject Sample #1 |       | 00:41 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:41 |                      |
| Air Blank         | 0.000 | 00:42 |                      |
| Control Test      | 0.078 | 00:42 |                      |
| Air Blank         | 0.000 | 00:43 |                      |
| Diagnostics Check | OK    | 00:43 |                      |

---

Data Download Date/Time 11/29/2011 16:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOPKINS, DAVID, J  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: HERNANDEZ, MICHAEL  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 11/26/2011 22:27:52

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:20 |
| Air Blank         | 0.000 | 22:20 |
| Control Test      | 0.081 | 22:20 |
| Air Blank         | 0.000 | 22:21 |
| Subject Sample #1 | 0.033 | 22:23 |
| Breath Volume     | 2.207 | 22:23 |
| Air Blank         | 0.000 | 22:23 |
| Air Blank         | 0.000 | 22:25 |
| Subject Sample #2 | 0.032 | 22:26 |
| Breath Volume     | 1.347 | 22:26 |
| Air Blank         | 0.000 | 22:26 |
| Control Test      | 0.081 | 22:27 |
| Air Blank         | 0.000 | 22:27 |
| Diagnostics Check | OK    | 22:27 |

---

Data Download Date/Time 11/29/2011 16:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEARNES, JOSEPH, R

BREATH TEST OPERATOR AGENCY: Niceville PD

ARRESTING OFFICER: HULION, TIM

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 19:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: BOGDahn

DATE/TIME OF TEST: 11/20/2011 20:15:58

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 20:09 |
| Air Blank         | 0.000 | 20:10 |
| Control Test      | 0.077 | 20:10 |
| Air Blank         | 0.000 | 20:10 |
| Subject Sample #1 | 0.192 | 20:11 |
| Breath Volume     | 3.589 | 20:11 |
| Air Blank         | 0.000 | 20:12 |
| Air Blank         | 0.000 | 20:13 |
| Subject Sample #2 | 0.192 | 20:14 |
| Breath Volume     | 2.910 | 20:14 |
| Air Blank         | 0.000 | 20:15 |
| Control Test      | 0.077 | 20:15 |
| Air Blank         | 0.000 | 20:15 |
| Diagnostics Check | OK    | 20:15 |

---

Data Download Date/Time 11/30/2011 16:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, DONALD, J

BREATH TEST OPERATOR AGENCY: Niceville PD

ARRESTING OFFICER: SMITH, MICHAEL

ARRESTING OFFICER AGENCY: Niceville PD

LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 06:34

GENDER: Female

OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: HADDEN

DATE/TIME OF TEST: 11/26/2011 07:23:40

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 07:20 |                      |
| Air Blank         | 0.000 | 07:21 |                      |
| Control Test      | 0.078 | 07:21 |                      |
| Air Blank         | 0.000 | 07:22 |                      |
| Subject Sample #1 |       | 07:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:22 |                      |
| Air Blank         | 0.000 | 07:22 |                      |
| Control Test      | 0.078 | 07:23 |                      |
| Air Blank         | 0.000 | 07:23 |                      |
| Diagnostics Check | OK    | 07:23 |                      |

---

Data Download Date/Time 11/30/2011 16:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN, NATHAN, J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: HOOKS, RYAN  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/02/2011 01:33:54

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:25 |
| Air Blank         | 0.000 | 01:25 |
| Control Test      | 0.082 | 01:25 |
| Air Blank         | 0.000 | 01:26 |
| Subject Sample #1 | 0.149 | 01:27 |
| Breath Volume     | 1.117 | 01:27 |
| Air Blank         | 0.000 | 01:28 |
| Air Blank         | 0.000 | 01:29 |
| Subject Sample #2 | 0.134 | 01:32 |
| Breath Volume     | 1.281 | 01:32 |
| Air Blank         | 0.000 | 01:32 |
| Control Test      | 0.082 | 01:33 |
| Air Blank         | 0.000 | 01:33 |
| Diagnostics Check | OK    | 01:33 |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AIMANOVICH, MATTHEW, J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: BOBLITT, RICK  
ARRESTING OFFICER AGENCY: DeFuniak Springs PD  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/04/2011 01:19:37

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:13 |
| Air Blank         | 0.000 | 01:13 |
| Control Test      | 0.082 | 01:13 |
| Air Blank         | 0.000 | 01:14 |
| Subject Sample #1 | 0.167 | 01:14 |
| Breath Volume     | 1.660 | 01:14 |
| Air Blank         | 0.000 | 01:15 |
| Air Blank         | 0.000 | 01:17 |
| Subject Sample #2 | 0.166 | 01:17 |
| Breath Volume     | 1.355 | 01:17 |
| Air Blank         | 0.000 | 01:18 |
| Control Test      | 0.081 | 01:19 |
| Air Blank         | 0.000 | 01:19 |
| Diagnostics Check | OK    | 01:19 |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MESSENGER, SARA  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 15:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MINYARD

DATE/TIME OF TEST: 11/05/2011 17:09:21

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 17:02 |
| Air Blank         | 0.000 | 17:02 |
| Control Test      | 0.082 | 17:03 |
| Air Blank         | 0.000 | 17:03 |
| Subject Sample #1 | 0.007 | 17:04 |
| Breath Volume     | 3.679 | 17:04 |
| Air Blank         | 0.000 | 17:05 |
| Air Blank         | 0.000 | 17:07 |
| Subject Sample #2 | 0.000 | 17:07 |
| Breath Volume     | 2.011 | 17:07 |
| Air Blank         | 0.000 | 17:08 |
| Control Test      | 0.082 | 17:08 |
| Air Blank         | 0.000 | 17:09 |
| Diagnostics Check | OK    | 17:09 |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN, NATHAN, J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: COSSON, DUSTIN  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: POCHE

DATE/TIME OF TEST: 11/06/2011 04:49:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:46 |                      |
| Air Blank         | 0.000 | 04:47 |                      |
| Control Test      | 0.083 | 04:47 |                      |
| Air Blank         | 0.000 | 04:47 |                      |
| Subject Sample #1 |       | 04:48 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:48 |                      |
| Air Blank         | 0.000 | 04:48 |                      |
| Control Test      | 0.083 | 04:48 |                      |
| Air Blank         | 0.000 | 04:49 |                      |
| Diagnostics Check | OK    | 04:49 |                      |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAY, JODY, L  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: DAY, JODY  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 11/12/2011 03:13:53

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:06 |
| Air Blank         | 0.000 | 03:07 |
| Control Test      | 0.081 | 03:07 |
| Air Blank         | 0.000 | 03:08 |
| Subject Sample #1 | 0.149 | 03:09 |
| Breath Volume     | 2.125 | 03:09 |
| Air Blank         | 0.000 | 03:09 |
| Air Blank         | 0.000 | 03:11 |
| Subject Sample #2 | 0.152 | 03:12 |
| Breath Volume     | 1.871 | 03:12 |
| Air Blank         | 0.000 | 03:12 |
| Control Test      | 0.082 | 03:13 |
| Air Blank         | 0.000 | 03:13 |
| Diagnostics Check | OK    | 03:13 |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELTON, MINDY, A  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: BIERNACKI, CHAD  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BOODEN

DATE/TIME OF TEST: 11/17/2011 00:22:40

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:15 |
| Air Blank         | 0.000 | 00:15 |
| Control Test      | 0.082 | 00:15 |
| Air Blank         | 0.000 | 00:16 |
| Subject Sample #1 | 0.221 | 00:16 |
| Breath Volume     | 2.246 | 00:16 |
| Air Blank         | 0.000 | 00:17 |
| Air Blank         | 0.000 | 00:19 |
| Subject Sample #2 | 0.219 | 00:20 |
| Breath Volume     | 1.910 | 00:20 |
| Air Blank         | 0.000 | 00:21 |
| Control Test      | 0.080 | 00:22 |
| Air Blank         | 0.000 | 00:22 |
| Diagnostics Check | OK    | 00:22 |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELTON, MINDY, A  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: BIERNACKI, CHAD  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: KODASEET

DATE/TIME OF TEST: 11/17/2011 21:35:31

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:32 |                      |
| Air Blank         | 0.000 | 21:32 |                      |
| Control Test      | 0.082 | 21:33 |                      |
| Air Blank         | 0.000 | 21:33 |                      |
| Subject Sample #1 |       | 21:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:34 |                      |
| Air Blank         | 0.000 | 21:34 |                      |
| Control Test      | 0.082 | 21:34 |                      |
| Air Blank         | 0.000 | 21:35 |                      |
| Diagnostics Check | OK    | 21:35 |                      |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KING, MARJORIE, G  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MALOY, JAMES  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BURROUGH

DATE/TIME OF TEST: 11/17/2011 23:53:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:49 |                      |
| Air Blank         | 0.000 | 23:49 |                      |
| Control Test      | 0.083 | 23:50 |                      |
| Air Blank         | 0.000 | 23:50 |                      |
| Subject Sample #1 |       | 23:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:51 |                      |
| Air Blank         | 0.000 | 23:52 |                      |
| Control Test      | 0.082 | 23:52 |                      |
| Air Blank         | 0.000 | 23:53 |                      |
| Diagnostics Check | OK    | 23:53 |                      |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN, NATHAN, J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MARTIN, NATHAN  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 11/21/2011 03:37:07

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:28 |
| Air Blank         | 0.000 | 03:28 |
| Control Test      | 0.080 | 03:29 |
| Air Blank         | 0.000 | 03:29 |
| Subject Sample #1 | 0.176 | 03:32 |
| Breath Volume     | 1.207 | 03:32 |
| Air Blank         | 0.000 | 03:33 |
| Air Blank         | 0.000 | 03:34 |
| Subject Sample #2 | 0.170 | 03:35 |
| Breath Volume     | 1.289 | 03:35 |
| Air Blank         | 0.000 | 03:36 |
| Control Test      | 0.079 | 03:36 |
| Air Blank         | 0.000 | 03:37 |
| Diagnostics Check | OK    | 03:37 |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUXHOLD, RICHARD, E  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: COSSON, DUSTIN  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 11:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 11/25/2011 01:22:37

---

|                   |       |       |                   |
|-------------------|-------|-------|-------------------|
| Diagnostics Check | OK    | 01:09 |                   |
| Air Blank         | 0.000 | 01:09 |                   |
| Control Test      | 0.081 | 01:10 |                   |
| Air Blank         | 0.000 | 01:10 |                   |
| Subject Sample #1 | 0.202 | 01:11 | No .020 agreement |
| Breath Volume     | 2.375 | 01:11 |                   |
| Air Blank         | 0.000 | 01:12 |                   |
| Air Blank         | 0.000 | 01:13 |                   |
| Subject Sample #2 | 0.227 | 01:15 | No .020 agreement |
| Breath Volume     | 1.566 | 01:15 |                   |
| Air Blank         | 0.000 | 01:16 |                   |
| Air Blank         | 0.000 | 01:17 |                   |
| Subject Sample #3 | 0.187 | 01:20 | Volume Not Met    |
| Breath Volume     | 0.035 | 01:20 |                   |
| Air Blank         | 0.000 | 01:21 |                   |
| Control Test      | 0.079 | 01:22 |                   |
| Air Blank         | 0.000 | 01:22 |                   |
| Diagnostics Check | OK    | 01:22 |                   |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELTON, MINDY, A  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: N/A, N/A  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: FRANKS

DATE/TIME OF TEST: 11/25/2011 20:45:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:42 |                      |
| Air Blank         | 0.000 | 20:43 |                      |
| Control Test      | 0.081 | 20:43 |                      |
| Air Blank         | 0.000 | 20:44 |                      |
| Subject Sample #1 |       | 20:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:44 |                      |
| Air Blank         | 0.000 | 20:44 |                      |
| Control Test      | 0.081 | 20:45 |                      |
| Air Blank         | 0.000 | 20:45 |                      |
| Diagnostics Check | OK    | 20:45 |                      |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELTON, MINDY, A  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: SHELTON, MINDY  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 11/26/2011 03:32:43

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:25 |
| Air Blank         | 0.000 | 03:25 |
| Control Test      | 0.082 | 03:26 |
| Air Blank         | 0.000 | 03:26 |
| Subject Sample #1 | 0.199 | 03:27 |
| Breath Volume     | 2.691 | 03:27 |
| Air Blank         | 0.000 | 03:28 |
| Air Blank         | 0.000 | 03:30 |
| Subject Sample #2 | 0.186 | 03:30 |
| Breath Volume     | 1.601 | 03:30 |
| Air Blank         | 0.000 | 03:31 |
| Control Test      | 0.079 | 03:32 |
| Air Blank         | 0.000 | 03:32 |
| Diagnostics Check | OK    | 03:32 |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELTON, MINDY, A  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: SHELTON, MINDY  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/27/2011 04:04:43

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:57 |
| Air Blank         | 0.000 | 03:58 |
| Control Test      | 0.081 | 03:58 |
| Air Blank         | 0.000 | 03:59 |
| Subject Sample #1 | 0.192 | 04:00 |
| Breath Volume     | 1.871 | 04:00 |
| Air Blank         | 0.000 | 04:00 |
| Air Blank         | 0.000 | 04:02 |
| Subject Sample #2 | 0.183 | 04:02 |
| Breath Volume     | 1.746 | 04:02 |
| Air Blank         | 0.000 | 04:03 |
| Control Test      | 0.079 | 04:04 |
| Air Blank         | 0.000 | 04:04 |
| Diagnostics Check | OK    | 04:04 |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELTON, MINDY, A  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: SHELTON, MINDY  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/27/2011 04:39:53

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:33 |
| Air Blank         | 0.000 | 04:33 |
| Control Test      | 0.079 | 04:33 |
| Air Blank         | 0.000 | 04:34 |
| Subject Sample #1 | 0.171 | 04:34 |
| Breath Volume     | 2.070 | 04:34 |
| Air Blank         | 0.000 | 04:35 |
| Air Blank         | 0.000 | 04:37 |
| Subject Sample #2 | 0.164 | 04:38 |
| Breath Volume     | 1.972 | 04:38 |
| Air Blank         | 0.000 | 04:38 |
| Control Test      | 0.076 | 04:39 |
| Air Blank         | 0.000 | 04:39 |
| Diagnostics Check | OK    | 04:39 |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARRICO, KELLY, D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CARRICO, KELLY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: HARRISON JR

DATE/TIME OF TEST: 11/27/2011 06:25:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 06:22 |                      |
| Air Blank         | 0.000 | 06:23 |                      |
| Control Test      | 0.081 | 06:23 |                      |
| Air Blank         | 0.000 | 06:23 |                      |
| Subject Sample #1 |       | 06:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 06:24 |                      |
| Air Blank         | 0.000 | 06:24 |                      |
| Control Test      | 0.082 | 06:25 |                      |
| Air Blank         | 0.000 | 06:25 |                      |
| Diagnostics Check | OK    | 06:25 |                      |

---

Data Download Date/Time 11/29/2011 10:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUXHOLD, RICHARD, E  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: HUXHOLD, RICHARD  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 11/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/30/2011 01:12:46

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:05 |
| Air Blank         | 0.000 | 01:06 |
| Control Test      | 0.079 | 01:06 |
| Air Blank         | 0.000 | 01:06 |
| Subject Sample #1 | 0.109 | 01:07 |
| Breath Volume     | 1.113 | 01:07 |
| Air Blank         | 0.000 | 01:08 |
| Air Blank         | 0.000 | 01:10 |
| Subject Sample #2 | 0.110 | 01:11 |
| Breath Volume     | 1.960 | 01:11 |
| Air Blank         | 0.000 | 01:11 |
| Control Test      | 0.078 | 01:12 |
| Air Blank         | 0.000 | 01:12 |
| Diagnostics Check | OK    | 01:12 |

---

Data Download Date/Time 12/14/2011 13:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001313

INSTRUMENT REGISTERED TO: SPRINGFIELD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOSTER, JOHN, A  
BREATH TEST OPERATOR AGENCY: Springfield PD  
ARRESTING OFFICER: ANDERS, STEVEN  
ARRESTING OFFICER AGENCY: Springfield PD  
LAST AGENCY INSPECTION: 10/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06611080A2  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 11/25/2011 22:24:10

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:16 |
| Air Blank         | 0.000 | 22:16 |
| Control Test      | 0.077 | 22:16 |
| Air Blank         | 0.000 | 22:17 |
| Subject Sample #1 | 0.217 | 22:18 |
| Breath Volume     | 1.187 | 22:18 |
| Air Blank         | 0.000 | 22:19 |
| Air Blank         | 0.000 | 22:21 |
| Subject Sample #2 | 0.216 | 22:22 |
| Breath Volume     | 1.460 | 22:22 |
| Air Blank         | 0.000 | 22:23 |
| Control Test      | 0.076 | 22:23 |
| Air Blank         | 0.000 | 22:24 |
| Diagnostics Check | OK    | 22:24 |

---

Data Download Date/Time 12/01/2011 07:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWACH, KATHY, R  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: HAZY, JOHN  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 11/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814704  
DRY GAS EXPIRY DATE: 08/18/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ZAHN

DATE/TIME OF TEST: 11/18/2011 00:50:09

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:41 |
| Air Blank         | 0.000 | 00:42 |
| Control Test      | 0.079 | 00:42 |
| Air Blank         | 0.000 | 00:43 |
| Subject Sample #1 | 0.172 | 00:45 |
| Breath Volume     | 2.964 | 00:45 |
| Air Blank         | 0.000 | 00:46 |
| Air Blank         | 0.000 | 00:48 |
| Subject Sample #2 | 0.170 | 00:48 |
| Breath Volume     | 2.855 | 00:48 |
| Air Blank         | 0.000 | 00:49 |
| Control Test      | 0.079 | 00:49 |
| Air Blank         | 0.000 | 00:50 |
| Diagnostics Check | OK    | 00:50 |

---

Data Download Date/Time 12/12/2011 08:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWACH, KATHY, R  
BREATH TEST OPERATOR AGENCY: Okeechobee PD  
ARRESTING OFFICER: HOLROYD, RYAN, RYAN  
ARRESTING OFFICER AGENCY: Okeechobee PD  
LAST AGENCY INSPECTION: 11/10/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 814704  
DRY GAS EXPIRY DATE: 08/18/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: RUSINKO, JEFFREY

DATE/TIME OF TEST: 11/20/2011 23:34:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:30 |                      |
| Air Blank         | 0.000 | 23:30 |                      |
| Control Test      | 0.079 | 23:31 |                      |
| Air Blank         | 0.000 | 23:31 |                      |
| Subject Sample #1 |       | 23:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:33 |                      |
| Air Blank         | 0.000 | 23:33 |                      |
| Control Test      | 0.080 | 23:33 |                      |
| Air Blank         | 0.000 | 23:34 |                      |
| Diagnostics Check | OK    | 23:34 |                      |

---

Data Download Date/Time 12/12/2011 08:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWACH, KATHY, R  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: BURK, TODD  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 11/10/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 814704  
DRY GAS EXPIRY DATE: 08/18/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: VIVAS

DATE/TIME OF TEST: 11/26/2011 00:45:27

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:42 |                      |
| Air Blank         | 0.000 | 00:42 |                      |
| Control Test      | 0.080 | 00:43 |                      |
| Air Blank         | 0.000 | 00:43 |                      |
| Subject Sample #1 |       | 00:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:44 |                      |
| Air Blank         | 0.000 | 00:44 |                      |
| Control Test      | 0.079 | 00:44 |                      |
| Air Blank         | 0.000 | 00:45 |                      |
| Diagnostics Check | OK    | 00:45 |                      |

---

Data Download Date/Time 12/12/2011 08:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWACH, KATHY, R  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: JACKSON, PAUL  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 10/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 814704  
DRY GAS EXPIRY DATE: 08/18/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 11/02/2011 22:38:12

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:31 |
| Air Blank         | 0.000 | 22:32 |
| Control Test      | 0.081 | 22:32 |
| Air Blank         | 0.000 | 22:33 |
| Subject Sample #1 | 0.181 | 22:33 |
| Breath Volume     | 1.375 | 22:33 |
| Air Blank         | 0.000 | 22:34 |
| Air Blank         | 0.000 | 22:36 |
| Subject Sample #2 | 0.183 | 22:36 |
| Breath Volume     | 1.324 | 22:36 |
| Air Blank         | 0.000 | 22:37 |
| Control Test      | 0.081 | 22:37 |
| Air Blank         | 0.000 | 22:38 |
| Diagnostics Check | OK    | 22:38 |

---

Data Download Date/Time 11/10/2011 07:20

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORGES, FRANKIE, J  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: BORGES, FRANKIE  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: NEWHOUSE

DATE/TIME OF TEST: 11/05/2011 01:02:39

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:54 |
| Air Blank         | 0.000 | 00:55 |
| Control Test      | 0.079 | 00:55 |
| Air Blank         | 0.000 | 00:56 |
| Subject Sample #1 | 0.186 | 00:57 |
| Breath Volume     | 2.496 | 00:57 |
| Air Blank         | 0.000 | 00:57 |
| Air Blank         | 0.000 | 00:59 |
| Subject Sample #2 | 0.171 | 01:01 |
| Breath Volume     | 1.777 | 01:01 |
| Air Blank         | 0.000 | 01:01 |
| Control Test      | 0.077 | 01:02 |
| Air Blank         | 0.000 | 01:02 |
| Diagnostics Check | OK    | 01:02 |

---

Data Download Date/Time 11/10/2011 04:30

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON, COURTNEY, B  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: KNAGGS, FRED  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: VENEGAS

DATE/TIME OF TEST: 11/06/2011 03:31:44

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:25 |
| Air Blank         | 0.000 | 03:25 |
| Control Test      | 0.080 | 03:26 |
| Air Blank         | 0.000 | 03:26 |
| Subject Sample #1 | 0.244 | 03:27 |
| Breath Volume     | 1.437 | 03:27 |
| Air Blank         | 0.000 | 03:27 |
| Air Blank         | 0.000 | 03:29 |
| Subject Sample #2 | 0.254 | 03:30 |
| Breath Volume     | 2.011 | 03:30 |
| Air Blank         | 0.000 | 03:30 |
| Control Test      | 0.080 | 03:31 |
| Air Blank         | 0.000 | 03:31 |
| Diagnostics Check | OK    | 03:31 |

---

Data Download Date/Time 11/10/2011 04:30

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, IAN,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: ROSKO, CHRISTINE  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: TIDWELL

DATE/TIME OF TEST: 11/08/2011 21:53:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:49 |                      |
| Air Blank         | 0.000 | 21:50 |                      |
| Control Test      | 0.080 | 21:50 |                      |
| Air Blank         | 0.000 | 21:50 |                      |
| Subject Sample #1 |       | 21:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:52 |                      |
| Air Blank         | 0.000 | 21:52 |                      |
| Control Test      | 0.080 | 21:52 |                      |
| Air Blank         | 0.000 | 21:53 |                      |
| Diagnostics Check | OK    | 21:53 |                      |

---

Data Download Date/Time 11/10/2011 04:30

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, IAN,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: HARRIS, IAN  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CORMIER

DATE/TIME OF TEST: 11/09/2011 03:14:09

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:07 |
| Air Blank         | 0.000 | 03:08 |
| Control Test      | 0.080 | 03:08 |
| Air Blank         | 0.000 | 03:09 |
| Subject Sample #1 | 0.212 | 03:09 |
| Breath Volume     | 1.785 | 03:09 |
| Air Blank         | 0.000 | 03:10 |
| Air Blank         | 0.000 | 03:12 |
| Subject Sample #2 | 0.208 | 03:12 |
| Breath Volume     | 1.503 | 03:12 |
| Air Blank         | 0.000 | 03:13 |
| Control Test      | 0.079 | 03:13 |
| Air Blank         | 0.000 | 03:14 |
| Diagnostics Check | OK    | 03:14 |

---

Data Download Date/Time 11/10/2011 04:30

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, IAN,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: SCHRADER, TODD  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 11/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 698118  
DRY GAS EXPIRY DATE: 11/30/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DOMINGOS

DATE/TIME OF TEST: 11/10/2011 04:11:54

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:03 |
| Air Blank         | 0.000 | 04:04 |
| Control Test      | 0.080 | 04:04 |
| Air Blank         | 0.000 | 04:04 |
| Subject Sample #1 | 0.148 | 04:06 |
| Breath Volume     | 1.863 | 04:06 |
| Air Blank         | 0.000 | 04:07 |
| Air Blank         | 0.000 | 04:08 |
| Subject Sample #2 | 0.143 | 04:10 |
| Breath Volume     | 1.898 | 04:10 |
| Air Blank         | 0.000 | 04:11 |
| Control Test      | 0.078 | 04:11 |
| Air Blank         | 0.000 | 04:11 |
| Diagnostics Check | OK    | 04:11 |

---

Data Download Date/Time 11/10/2011 04:30

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, IAN,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: HARRIS, IAN  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 11/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ROBILLARD

DATE/TIME OF TEST: 11/11/2011 03:27:40

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:21 |
| Air Blank         | 0.000 | 03:21 |
| Control Test      | 0.080 | 03:22 |
| Air Blank         | 0.000 | 03:22 |
| Subject Sample #1 | 0.147 | 03:23 |
| Breath Volume     | 3.179 | 03:23 |
| Air Blank         | 0.000 | 03:23 |
| Air Blank         | 0.000 | 03:25 |
| Subject Sample #2 | 0.143 | 03:26 |
| Breath Volume     | 3.382 | 03:26 |
| Air Blank         | 0.000 | 03:26 |
| Control Test      | 0.078 | 03:27 |
| Air Blank         | 0.000 | 03:27 |
| Diagnostics Check | OK    | 03:27 |

---

Data Download Date/Time 12/22/2011 03:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEINKRAUS, ROBERT, W  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: AUGUSTE, MIAGA  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 11/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 17:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: VENEROS

DATE/TIME OF TEST: 11/13/2011 19:25:50

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:14 |
| Air Blank         | 0.000 | 19:15 |
| Control Test      | 0.080 | 19:15 |
| Air Blank         | 0.000 | 19:16 |
| Subject Sample #1 | 0.304 | 19:17 |
| Breath Volume     | 2.542 | 19:17 |
| Air Blank         | 0.000 | 19:18 |
| Air Blank         | 0.000 | 19:20 |
| Subject Sample #2 | 0.251 | 19:21 |
| Breath Volume     | 2.187 | 19:21 |
| Air Blank         | 0.000 | 19:21 |
| Air Blank         | 0.000 | 19:23 |
| Subject Sample #3 | 0.247 | 19:24 |
| Breath Volume     | 1.882 | 19:24 |
| Air Blank         | 0.000 | 19:24 |
| Control Test      | 0.079 | 19:25 |
| Air Blank         | 0.000 | 19:25 |
| Diagnostics Check | OK    | 19:25 |

---

Data Download Date/Time 12/22/2011 03:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORGES, FRANKIE, J  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: HINES, LAUREN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 11/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WESTCOTT

DATE/TIME OF TEST: 11/26/2011 04:00:19

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:51 |
| Air Blank         | 0.000 | 03:52 |
| Control Test      | 0.080 | 03:52 |
| Air Blank         | 0.000 | 03:53 |
| Subject Sample #1 | 0.120 | 03:55 |
| Breath Volume     | 1.906 | 03:55 |
| Air Blank         | 0.000 | 03:55 |
| Air Blank         | 0.000 | 03:57 |
| Subject Sample #2 | 0.111 | 03:58 |
| Breath Volume     | 1.113 | 03:58 |
| Air Blank         | 0.000 | 03:59 |
| Control Test      | 0.078 | 03:59 |
| Air Blank         | 0.000 | 04:00 |
| Diagnostics Check | OK    | 04:00 |

---

Data Download Date/Time 12/22/2011 03:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHUNN, PETER, E  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: CHUNN, PETER  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 11/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: GHEZZI

DATE/TIME OF TEST: 11/28/2011 00:53:11

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:45 |
| Air Blank         | 0.000 | 00:45 |
| Control Test      | 0.080 | 00:46 |
| Air Blank         | 0.000 | 00:46 |
| Subject Sample #1 | 0.278 | 00:47 |
| Breath Volume     | 1.792 | 00:47 |
| Air Blank         | 0.000 | 00:48 |
| Air Blank         | 0.000 | 00:50 |
| Subject Sample #2 | 0.269 | 00:51 |
| Breath Volume     | 1.253 | 00:51 |
| Air Blank         | 0.000 | 00:52 |
| Control Test      | 0.079 | 00:52 |
| Air Blank         | 0.000 | 00:53 |
| Diagnostics Check | OK    | 00:53 |

---

Data Download Date/Time 12/22/2011 03:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMEL, ANDREW, J  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: KNOTT, JUSTIN  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MCGURREN

DATE/TIME OF TEST: 11/02/2011 23:50:18

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:44 |
| Air Blank         | 0.000 | 23:44 |
| Control Test      | 0.079 | 23:44 |
| Air Blank         | 0.000 | 23:45 |
| Subject Sample #1 | 0.000 | 23:45 |
| Breath Volume     | 2.472 | 23:45 |
| Air Blank         | 0.000 | 23:46 |
| Air Blank         | 0.000 | 23:48 |
| Subject Sample #2 | 0.000 | 23:48 |
| Breath Volume     | 2.097 | 23:48 |
| Air Blank         | 0.000 | 23:49 |
| Control Test      | 0.078 | 23:49 |
| Air Blank         | 0.000 | 23:50 |
| Diagnostics Check | OK    | 23:50 |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS, LUCRECE,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: REEVES, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HEITZ

DATE/TIME OF TEST: 11/05/2011 22:23:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:20 |                      |
| Air Blank         | 0.000 | 22:21 |                      |
| Control Test      | 0.078 | 22:21 |                      |
| Air Blank         | 0.000 | 22:21 |                      |
| Subject Sample #1 |       | 22:21 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:21 |                      |
| Air Blank         | 0.000 | 22:22 |                      |
| Control Test      | 0.078 | 22:22 |                      |
| Air Blank         | 0.000 | 22:23 |                      |
| Diagnostics Check | OK    | 22:23 |                      |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHUMAN, GARRETT, L  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: SHUMAN, GARRETT  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 11:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/07/2011 01:09:46

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:03 |
| Air Blank         | 0.000 | 01:03 |
| Control Test      | 0.078 | 01:04 |
| Air Blank         | 0.000 | 01:04 |
| Subject Sample #1 | 0.077 | 01:05 |
| Breath Volume     | 3.257 | 01:05 |
| Air Blank         | 0.000 | 01:05 |
| Air Blank         | 0.000 | 01:07 |
| Subject Sample #2 | 0.076 | 01:08 |
| Breath Volume     | 2.843 | 01:08 |
| Air Blank         | 0.000 | 01:08 |
| Control Test      | 0.078 | 01:09 |
| Air Blank         | 0.000 | 01:09 |
| Diagnostics Check | OK    | 01:09 |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND, PAMELA, R  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: LANAM, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 11/07/2011 05:13:15

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 04:56 |                |
| Air Blank         | 0.000 | 04:57 |                |
| Control Test      | 0.078 | 04:57 |                |
| Air Blank         | 0.000 | 04:57 |                |
| Subject Sample #1 | 0.161 | 05:01 | Volume Not Met |
| Breath Volume     | 0.062 | 05:01 |                |
| Air Blank         | 0.000 | 05:01 |                |
| Air Blank         | 0.000 | 05:03 |                |
| Subject Sample #2 | 0.183 | 05:05 |                |
| Breath Volume     | 1.195 | 05:05 |                |
| Air Blank         | 0.000 | 05:06 |                |
| Air Blank         | 0.000 | 05:08 |                |
| Subject Sample #3 | 0.159 | 05:11 | Volume Not Met |
| Breath Volume     | 0.101 | 05:11 |                |
| Air Blank         | 0.000 | 05:12 |                |
| Control Test      | 0.077 | 05:12 |                |
| Air Blank         | 0.000 | 05:13 |                |
| Diagnostics Check | OK    | 05:13 |                |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON, TERRANCE, T  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: NELSON, TERRANCE  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/11/2011 02:06:44

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:00 |
| Air Blank         | 0.000 | 02:00 |
| Control Test      | 0.079 | 02:00 |
| Air Blank         | 0.000 | 02:01 |
| Subject Sample #1 | 0.185 | 02:02 |
| Breath Volume     | 3.281 | 02:02 |
| Air Blank         | 0.000 | 02:02 |
| Air Blank         | 0.000 | 02:04 |
| Subject Sample #2 | 0.185 | 02:05 |
| Breath Volume     | 3.019 | 02:05 |
| Air Blank         | 0.000 | 02:05 |
| Control Test      | 0.078 | 02:06 |
| Air Blank         | 0.000 | 02:06 |
| Diagnostics Check | OK    | 02:06 |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, TERRENCE, R  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MCCULLOUGH, LARRY  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: KNOEPEL

DATE/TIME OF TEST: 11/13/2011 22:24:30

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:11 |                    |
| Air Blank         | 0.000 | 22:12 |                    |
| Control Test      | 0.079 | 22:12 |                    |
| Air Blank         | 0.000 | 22:12 |                    |
| Subject Sample #1 |       | 22:16 | No Sample Provided |
| Breath Volume     | 0.000 | 22:16 |                    |
| Air Blank         | 0.000 | 22:16 |                    |
| Air Blank         | 0.000 | 22:18 |                    |
| Subject Sample #2 | 0.097 | 22:19 |                    |
| Breath Volume     | 3.132 | 22:19 |                    |
| Air Blank         | 0.000 | 22:20 |                    |
| Air Blank         | 0.000 | 22:21 |                    |
| Subject Sample #3 | 0.099 | 22:22 |                    |
| Breath Volume     | 2.640 | 22:22 |                    |
| Air Blank         | 0.000 | 22:23 |                    |
| Control Test      | 0.078 | 22:23 |                    |
| Air Blank         | 0.000 | 22:24 |                    |
| Diagnostics Check | OK    | 22:24 |                    |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS, LUCRECE,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: RAHMNG, DERRICK  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MILGRAM

DATE/TIME OF TEST: 11/13/2011 22:47:25

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:39 |                      |
| Air Blank         | 0.000 | 22:39 |                      |
| Control Test      | 0.079 | 22:39 |                      |
| Air Blank         | 0.000 | 22:40 |                      |
| Subject Sample #1 |       | 22:43 | No Sample Provided   |
| Breath Volume     | 0.000 | 22:43 |                      |
| Air Blank         | 0.000 | 22:43 |                      |
| Air Blank         | 0.000 | 22:45 |                      |
| Subject Sample #2 |       | 22:46 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:46 |                      |
| Air Blank         | 0.000 | 22:46 |                      |
| Control Test      | 0.078 | 22:46 |                      |
| Air Blank         | 0.000 | 22:47 |                      |
| Diagnostics Check | OK    | 22:47 |                      |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: URBANCZYK, CRAIG, T  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: URBANCZYK, CRAIG  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 11/14/2011 23:38:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:35 |                      |
| Air Blank         | 0.000 | 23:36 |                      |
| Control Test      | 0.078 | 23:36 |                      |
| Air Blank         | 0.000 | 23:36 |                      |
| Subject Sample #1 |       | 23:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:37 |                      |
| Air Blank         | 0.000 | 23:37 |                      |
| Control Test      | 0.079 | 23:37 |                      |
| Air Blank         | 0.000 | 23:38 |                      |
| Diagnostics Check | OK    | 23:38 |                      |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAPUTO, LAURA, J  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: BIRTEL, MARK  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FAILLACE

DATE/TIME OF TEST: 11/18/2011 00:38:20

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:31 |
| Air Blank         | 0.000 | 00:31 |
| Control Test      | 0.079 | 00:31 |
| Air Blank         | 0.000 | 00:32 |
| Subject Sample #1 | 0.171 | 00:33 |
| Breath Volume     | 2.167 | 00:33 |
| Air Blank         | 0.000 | 00:34 |
| Air Blank         | 0.000 | 00:36 |
| Subject Sample #2 | 0.164 | 00:36 |
| Breath Volume     | 2.253 | 00:36 |
| Air Blank         | 0.000 | 00:37 |
| Control Test      | 0.078 | 00:37 |
| Air Blank         | 0.000 | 00:38 |
| Diagnostics Check | OK    | 00:38 |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEARS, TONY, D  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: BELL, BRIAN  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 11/19/2011 01:34:46

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:25 |
| Air Blank         | 0.000 | 01:26 |
| Control Test      | 0.079 | 01:26 |
| Air Blank         | 0.000 | 01:27 |
| Subject Sample #1 | 0.202 | 01:30 |
| Breath Volume     | 1.718 | 01:30 |
| Air Blank         | 0.000 | 01:30 |
| Air Blank         | 0.000 | 01:32 |
| Subject Sample #2 | 0.201 | 01:33 |
| Breath Volume     | 2.289 | 01:33 |
| Air Blank         | 0.000 | 01:33 |
| Control Test      | 0.078 | 01:34 |
| Air Blank         | 0.000 | 01:34 |
| Diagnostics Check | OK    | 01:34 |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEST, TEST, T  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: TEST, TEST  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 11/19/2011 02:06:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:03 |                      |
| Air Blank         | 0.000 | 02:03 |                      |
| Control Test      | 0.079 | 02:04 |                      |
| Air Blank         | 0.000 | 02:04 |                      |
| Subject Sample #1 |       | 02:04 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:04 |                      |
| Air Blank         | 0.000 | 02:05 |                      |
| Control Test      | 0.078 | 02:05 |                      |
| Air Blank         | 0.000 | 02:06 |                      |
| Diagnostics Check | OK    | 02:06 |                      |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAPUTO, LAURA, J  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: LANAM, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 11/19/2011 02:48:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:45 |                      |
| Air Blank         | 0.000 | 02:46 |                      |
| Control Test      | 0.079 | 02:46 |                      |
| Air Blank         | 0.000 | 02:47 |                      |
| Subject Sample #1 |       | 02:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:47 |                      |
| Air Blank         | 0.000 | 02:47 |                      |
| Control Test      | 0.079 | 02:48 |                      |
| Air Blank         | 0.000 | 02:48 |                      |
| Diagnostics Check | OK    | 02:48 |                      |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIGGOTT, MYRON, J  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: FARQUHARSON, NORMAN  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 09:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 11/19/2011 10:26:27

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 10:11 |                |
| Air Blank         | 0.000 | 10:11 |                |
| Control Test      | 0.079 | 10:12 |                |
| Air Blank         | 0.000 | 10:12 |                |
| Subject Sample #1 | 0.246 | 10:15 | Volume Not Met |
| Breath Volume     | 0.050 | 10:15 |                |
| Air Blank         | 0.000 | 10:16 |                |
| Air Blank         | 0.000 | 10:18 |                |
| Subject Sample #2 | 0.219 | 10:21 |                |
| Breath Volume     | 1.421 | 10:21 |                |
| Air Blank         | 0.000 | 10:22 |                |
| Air Blank         | 0.000 | 10:23 |                |
| Subject Sample #3 | 0.226 | 10:24 |                |
| Breath Volume     | 1.695 | 10:24 |                |
| Air Blank         | 0.000 | 10:25 |                |
| Control Test      | 0.077 | 10:25 |                |
| Air Blank         | 0.000 | 10:26 |                |
| Diagnostics Check | OK    | 10:26 |                |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TANGEL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: TOMKINSON, STEVE  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HOUGHTON

DATE/TIME OF TEST: 11/19/2011 22:37:42

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:31 |
| Air Blank         | 0.000 | 22:31 |
| Control Test      | 0.079 | 22:32 |
| Air Blank         | 0.000 | 22:32 |
| Subject Sample #1 | 0.172 | 22:33 |
| Breath Volume     | 1.507 | 22:33 |
| Air Blank         | 0.000 | 22:33 |
| Air Blank         | 0.000 | 22:35 |
| Subject Sample #2 | 0.182 | 22:36 |
| Breath Volume     | 2.054 | 22:36 |
| Air Blank         | 0.000 | 22:36 |
| Control Test      | 0.079 | 22:37 |
| Air Blank         | 0.000 | 22:37 |
| Diagnostics Check | OK    | 22:37 |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ, LEO,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MILLER, ED  
ARRESTING OFFICER AGENCY: Indian River Shores  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ANDREA

DATE/TIME OF TEST: 11/20/2011 01:27:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:24 |                      |
| Air Blank         | 0.000 | 01:25 |                      |
| Control Test      | 0.079 | 01:25 |                      |
| Air Blank         | 0.000 | 01:25 |                      |
| Subject Sample #1 |       | 01:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:26 |                      |
| Air Blank         | 0.000 | 01:26 |                      |
| Control Test      | 0.079 | 01:26 |                      |
| Air Blank         | 0.000 | 01:27 |                      |
| Diagnostics Check | OK    | 01:27 |                      |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL, ALAN, R  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: FELDMAN, CHARLIE  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CROWTHER

DATE/TIME OF TEST: 11/22/2011 00:24:31

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 00:21 |                    |
| Air Blank         | 0.000 | 00:21 |                    |
| Control Test      | 0.079 | 00:21 |                    |
| Air Blank         | 0.000 | 00:22 |                    |
| Subject Sample #1 |       | 00:23 | Interferent Detect |
| Breath Volume     | 4.250 | 00:23 |                    |
| Air Blank         | 0.000 | 00:23 |                    |
| Control Test      | 0.080 | 00:23 |                    |
| Air Blank         | 0.000 | 00:24 |                    |
| Diagnostics Check | OK    | 00:24 |                    |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHUMAN, GARRETT, L  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: SHUMAN, GARRETT  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: RULE

DATE/TIME OF TEST: 11/26/2011 05:30:16

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:23 |
| Air Blank         | 0.000 | 05:24 |
| Control Test      | 0.078 | 05:24 |
| Air Blank         | 0.000 | 05:24 |
| Subject Sample #1 | 0.133 | 05:25 |
| Breath Volume     | 2.757 | 05:25 |
| Air Blank         | 0.000 | 05:26 |
| Air Blank         | 0.000 | 05:28 |
| Subject Sample #2 | 0.130 | 05:28 |
| Breath Volume     | 1.695 | 05:28 |
| Air Blank         | 0.000 | 05:29 |
| Control Test      | 0.078 | 05:29 |
| Air Blank         | 0.000 | 05:30 |
| Diagnostics Check | OK    | 05:30 |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS, LUCRECE,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: BRANDES, MICHAEL  
ARRESTING OFFICER AGENCY: Sebastian PD  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 17:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: KIEVIT

DATE/TIME OF TEST: 11/26/2011 18:22:26

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 18:16 |
| Air Blank         | 0.000 | 18:16 |
| Control Test      | 0.078 | 18:17 |
| Air Blank         | 0.000 | 18:17 |
| Subject Sample #1 | 0.000 | 18:18 |
| Breath Volume     | 1.988 | 18:18 |
| Air Blank         | 0.000 | 18:18 |
| Air Blank         | 0.000 | 18:20 |
| Subject Sample #2 | 0.000 | 18:21 |
| Breath Volume     | 2.230 | 18:21 |
| Air Blank         | 0.000 | 18:21 |
| Control Test      | 0.079 | 18:21 |
| Air Blank         | 0.000 | 18:22 |
| Diagnostics Check | OK    | 18:22 |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMEL, ANDREW, J  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: DOYLE, JUSTIN  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 850329  
DRY GAS EXPIRY DATE: 09/22/2012  
VIOLATION CODE: DUI  
ARREST TIME: 18:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: KNAPP

DATE/TIME OF TEST: 11/26/2011 19:14:31

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:08 |
| Air Blank         | 0.000 | 19:08 |
| Control Test      | 0.079 | 19:08 |
| Air Blank         | 0.000 | 19:09 |
| Subject Sample #1 | 0.201 | 19:09 |
| Breath Volume     | 2.378 | 19:09 |
| Air Blank         | 0.000 | 19:10 |
| Air Blank         | 0.000 | 19:12 |
| Subject Sample #2 | 0.203 | 19:12 |
| Breath Volume     | 2.777 | 19:12 |
| Air Blank         | 0.000 | 19:13 |
| Control Test      | 0.078 | 19:13 |
| Air Blank         | 0.000 | 19:14 |
| Diagnostics Check | OK    | 19:14 |

---

Data Download Date/Time 12/01/2011 18:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: METCALF, DEBORAH, K  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: SANTOS, JOEY  
ARRESTING OFFICER AGENCY: Arcadia PD  
LAST AGENCY INSPECTION: 10/19/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 12:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:31

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 11/02/2011 13:58:39

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 13:52 |
| Air Blank         | 0.000 | 13:52 |
| Control Test      | 0.079 | 13:53 |
| Air Blank         | 0.000 | 13:53 |
| Subject Sample #1 | 0.204 | 13:54 |
| Breath Volume     | 2.375 | 13:54 |
| Air Blank         | 0.000 | 13:54 |
| Air Blank         | 0.000 | 13:56 |
| Subject Sample #2 | 0.214 | 13:57 |
| Breath Volume     | 2.308 | 13:57 |
| Air Blank         | 0.000 | 13:57 |
| Control Test      | 0.078 | 13:58 |
| Air Blank         | 0.000 | 13:58 |
| Diagnostics Check | OK    | 13:58 |

---

Data Download Date/Time 11/18/2011 16:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARROLL, STEVEN, C  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: HURST, TRAE  
ARRESTING OFFICER AGENCY: DeSoto County SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 11/19/2011 00:32:28

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:26 |
| Air Blank         | 0.000 | 00:26 |
| Control Test      | 0.080 | 00:27 |
| Air Blank         | 0.000 | 00:27 |
| Subject Sample #1 | 0.122 | 00:28 |
| Breath Volume     | 1.265 | 00:28 |
| Air Blank         | 0.000 | 00:28 |
| Air Blank         | 0.000 | 00:30 |
| Subject Sample #2 | 0.121 | 00:30 |
| Breath Volume     | 3.000 | 00:30 |
| Air Blank         | 0.000 | 00:31 |
| Control Test      | 0.079 | 00:31 |
| Air Blank         | 0.000 | 00:32 |
| Diagnostics Check | OK    | 00:32 |

---

Data Download Date/Time 12/05/2011 13:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARROLL, STEVEN, C  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: PARK, RUSSELL  
ARRESTING OFFICER AGENCY: DeSoto County SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 11/20/2011 02:09:57

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:03 |
| Air Blank         | 0.000 | 02:04 |
| Control Test      | 0.081 | 02:04 |
| Air Blank         | 0.000 | 02:04 |
| Subject Sample #1 | 0.205 | 02:05 |
| Breath Volume     | 3.652 | 02:05 |
| Air Blank         | 0.000 | 02:06 |
| Air Blank         | 0.000 | 02:07 |
| Subject Sample #2 | 0.199 | 02:08 |
| Breath Volume     | 3.695 | 02:08 |
| Air Blank         | 0.000 | 02:09 |
| Control Test      | 0.078 | 02:09 |
| Air Blank         | 0.000 | 02:09 |
| Diagnostics Check | OK    | 02:09 |

---

Data Download Date/Time 12/05/2011 13:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTLER, CHRIS, T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: DENIRO, OFFICER  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: ADMIN  
ARREST TIME: 16:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 11/17/2011 16:43:27

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 16:33 |
| Air Blank         | 0.000 | 16:33 |
| Control Test      | 0.078 | 16:33 |
| Air Blank         | 0.000 | 16:34 |
| Subject Sample #1 | 0.329 | 16:35 |
| Breath Volume     | 1.253 | 16:35 |
| Air Blank         | 0.000 | 16:36 |
| Air Blank         | 0.000 | 16:38 |
| Subject Sample #2 | 0.303 | 16:38 |
| Breath Volume     | 1.363 | 16:38 |
| Air Blank         | 0.000 | 16:39 |
| Air Blank         | 0.000 | 16:40 |
| Subject Sample #3 | 0.316 | 16:41 |
| Breath Volume     | 1.480 | 16:41 |
| Air Blank         | 0.000 | 16:42 |
| Control Test      | 0.077 | 16:42 |
| Air Blank         | 0.000 | 16:43 |
| Diagnostics Check | OK    | 16:43 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, RICHARD, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: JOHNSON, RICHARD  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 19:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: RIEDL

DATE/TIME OF TEST: 11/17/2011 20:24:51

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 20:13 |                |
| Air Blank         | 0.000 | 20:13 |                |
| Control Test      | 0.077 | 20:14 |                |
| Air Blank         | 0.000 | 20:14 |                |
| Subject Sample #1 | 0.096 | 20:17 | Volume Not Met |
| Breath Volume     | 0.082 | 20:17 |                |
| Air Blank         | 0.000 | 20:18 |                |
| Air Blank         | 0.000 | 20:20 |                |
| Subject Sample #2 | 0.092 | 20:23 | Volume Not Met |
| Breath Volume     | 0.269 | 20:23 |                |
| Air Blank         | 0.000 | 20:23 |                |
| Control Test      | 0.078 | 20:24 |                |
| Air Blank         | 0.000 | 20:24 |                |
| Diagnostics Check | OK    | 20:24 |                |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, RICHARD, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FORMAL, DOMINIC  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 11/17/2011 21:37:24

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 21:31 |
| Air Blank         | 0.000 | 21:31 |
| Control Test      | 0.077 | 21:31 |
| Air Blank         | 0.000 | 21:32 |
| Subject Sample #1 | 0.000 | 21:32 |
| Breath Volume     | 2.433 | 21:32 |
| Air Blank         | 0.000 | 21:33 |
| Air Blank         | 0.000 | 21:35 |
| Subject Sample #2 | 0.000 | 21:36 |
| Breath Volume     | 1.753 | 21:36 |
| Air Blank         | 0.000 | 21:36 |
| Control Test      | 0.078 | 21:36 |
| Air Blank         | 0.000 | 21:37 |
| Diagnostics Check | OK    | 21:37 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FELTOVIC, DEP, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FELTOVIC, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: FRALEY

DATE/TIME OF TEST: 11/18/2011 00:21:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:15 |                      |
| Air Blank         | 0.000 | 00:16 |                      |
| Control Test      | 0.078 | 00:16 |                      |
| Air Blank         | 0.000 | 00:17 |                      |
| Subject Sample #1 |       | 00:19 | Subject Test Refused |
| Breath Volume     | 0.101 | 00:19 |                      |
| Air Blank         | 0.000 | 00:20 |                      |
| Control Test      | 0.077 | 00:20 |                      |
| Air Blank         | 0.000 | 00:21 |                      |
| Diagnostics Check | OK    | 00:21 |                      |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATHMAN, JAMES, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: KATHMAN, JAMES  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MONTONEY

DATE/TIME OF TEST: 11/18/2011 02:07:41

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:00 |
| Air Blank         | 0.000 | 02:00 |
| Control Test      | 0.079 | 02:00 |
| Air Blank         | 0.000 | 02:01 |
| Subject Sample #1 | 0.000 | 02:02 |
| Breath Volume     | 1.664 | 02:02 |
| Air Blank         | 0.000 | 02:02 |
| Air Blank         | 0.000 | 02:04 |
| Subject Sample #2 | 0.000 | 02:06 |
| Breath Volume     | 2.101 | 02:06 |
| Air Blank         | 0.000 | 02:06 |
| Control Test      | 0.080 | 02:07 |
| Air Blank         | 0.000 | 02:07 |
| Diagnostics Check | OK    | 02:07 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WRIGHT, DEP, L  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: HAUSE, JEFF  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 13:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:47

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 11/18/2011 16:13:50

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 16:07 |
| Air Blank         | 0.000 | 16:08 |
| Control Test      | 0.078 | 16:08 |
| Air Blank         | 0.000 | 16:09 |
| Subject Sample #1 | 0.000 | 16:09 |
| Breath Volume     | 1.675 | 16:09 |
| Air Blank         | 0.000 | 16:10 |
| Air Blank         | 0.000 | 16:12 |
| Subject Sample #2 | 0.000 | 16:12 |
| Breath Volume     | 1.625 | 16:12 |
| Air Blank         | 0.000 | 16:12 |
| Control Test      | 0.078 | 16:13 |
| Air Blank         | 0.000 | 16:13 |
| Diagnostics Check | OK    | 16:13 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE, MARK,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FELTOVIC, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 11/19/2011 02:24:55

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:18 |
| Air Blank         | 0.000 | 02:18 |
| Control Test      | 0.078 | 02:18 |
| Air Blank         | 0.000 | 02:19 |
| Subject Sample #1 | 0.173 | 02:20 |
| Breath Volume     | 2.105 | 02:20 |
| Air Blank         | 0.000 | 02:20 |
| Air Blank         | 0.000 | 02:22 |
| Subject Sample #2 | 0.163 | 02:23 |
| Breath Volume     | 2.027 | 02:23 |
| Air Blank         | 0.000 | 02:24 |
| Control Test      | 0.079 | 02:24 |
| Air Blank         | 0.000 | 02:24 |
| Diagnostics Check | OK    | 02:24 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, DEP, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PIZZO, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: THURMOND

DATE/TIME OF TEST: 11/20/2011 00:36:50

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:30 |
| Air Blank         | 0.000 | 00:31 |
| Control Test      | 0.079 | 00:31 |
| Air Blank         | 0.000 | 00:32 |
| Subject Sample #1 | 0.082 | 00:32 |
| Breath Volume     | 2.511 | 00:32 |
| Air Blank         | 0.000 | 00:32 |
| Air Blank         | 0.000 | 00:34 |
| Subject Sample #2 | 0.089 | 00:35 |
| Breath Volume     | 3.937 | 00:35 |
| Air Blank         | 0.000 | 00:35 |
| Control Test      | 0.079 | 00:36 |
| Air Blank         | 0.000 | 00:36 |
| Diagnostics Check | OK    | 00:36 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, DEP, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: DILLING, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: EARL

DATE/TIME OF TEST: 11/20/2011 04:53:43

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:47 |
| Air Blank         | 0.000 | 04:47 |
| Control Test      | 0.078 | 04:48 |
| Air Blank         | 0.000 | 04:48 |
| Subject Sample #1 | 0.155 | 04:49 |
| Breath Volume     | 5.058 | 04:49 |
| Air Blank         | 0.000 | 04:49 |
| Air Blank         | 0.000 | 04:51 |
| Subject Sample #2 | 0.149 | 04:52 |
| Breath Volume     | 2.585 | 04:52 |
| Air Blank         | 0.000 | 04:52 |
| Control Test      | 0.079 | 04:53 |
| Air Blank         | 0.000 | 04:53 |
| Diagnostics Check | OK    | 04:53 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIANDI, JAMIE, B  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: STEVENS, TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/21/2011 22:19:49

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:13 |
| Air Blank         | 0.000 | 22:13 |
| Control Test      | 0.079 | 22:14 |
| Air Blank         | 0.000 | 22:14 |
| Subject Sample #1 | 0.261 | 22:15 |
| Breath Volume     | 3.300 | 22:15 |
| Air Blank         | 0.000 | 22:16 |
| Air Blank         | 0.000 | 22:17 |
| Subject Sample #2 | 0.261 | 22:18 |
| Breath Volume     | 3.238 | 22:18 |
| Air Blank         | 0.000 | 22:18 |
| Control Test      | 0.077 | 22:19 |
| Air Blank         | 0.000 | 22:19 |
| Diagnostics Check | OK    | 22:19 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHMITT, DEP, K  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PARIS, OFFICER  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 11/22/2011 03:44:54

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:38 |
| Air Blank         | 0.000 | 03:38 |
| Control Test      | 0.078 | 03:39 |
| Air Blank         | 0.000 | 03:39 |
| Subject Sample #1 | 0.169 | 03:40 |
| Breath Volume     | 2.425 | 03:40 |
| Air Blank         | 0.000 | 03:40 |
| Air Blank         | 0.000 | 03:42 |
| Subject Sample #2 | 0.158 | 03:43 |
| Breath Volume     | 1.828 | 03:43 |
| Air Blank         | 0.000 | 03:43 |
| Control Test      | 0.077 | 03:44 |
| Air Blank         | 0.000 | 03:44 |
| Diagnostics Check | OK    | 03:44 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FELTOVIC, DEP, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FELTOVIC, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: PAGONES

DATE/TIME OF TEST: 11/22/2011 21:25:07

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 21:12 |                |
| Air Blank         | 0.000 | 21:13 |                |
| Control Test      | 0.078 | 21:13 |                |
| Air Blank         | 0.000 | 21:13 |                |
| Subject Sample #1 | 0.154 | 21:16 | Volume Not Met |
| Breath Volume     | 0.875 | 21:16 |                |
| Air Blank         | 0.000 | 21:17 |                |
| Air Blank         | 0.000 | 21:19 |                |
| Subject Sample #2 | 0.190 | 21:20 |                |
| Breath Volume     | 1.792 | 21:20 |                |
| Air Blank         | 0.000 | 21:20 |                |
| Air Blank         | 0.000 | 21:22 |                |
| Subject Sample #3 | 0.191 | 21:23 |                |
| Breath Volume     | 1.828 | 21:23 |                |
| Air Blank         | 0.000 | 21:24 |                |
| Control Test      | 0.078 | 21:24 |                |
| Air Blank         | 0.000 | 21:25 |                |
| Diagnostics Check | OK    | 21:25 |                |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FELTOVIC, DEP, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FELTOVIC, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 11/23/2011 02:53:24

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:47 |
| Air Blank         | 0.000 | 02:47 |
| Control Test      | 0.079 | 02:47 |
| Air Blank         | 0.000 | 02:48 |
| Subject Sample #1 | 0.087 | 02:48 |
| Breath Volume     | 2.410 | 02:48 |
| Air Blank         | 0.000 | 02:49 |
| Air Blank         | 0.000 | 02:51 |
| Subject Sample #2 | 0.080 | 02:51 |
| Breath Volume     | 1.234 | 02:51 |
| Air Blank         | 0.000 | 02:52 |
| Control Test      | 0.079 | 02:52 |
| Air Blank         | 0.000 | 02:53 |
| Diagnostics Check | OK    | 02:53 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, DEP, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FORMAL, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: JULIEN

DATE/TIME OF TEST: 11/23/2011 22:11:07

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:04 |
| Air Blank         | 0.000 | 22:05 |
| Control Test      | 0.079 | 22:05 |
| Air Blank         | 0.000 | 22:06 |
| Subject Sample #1 | 0.079 | 22:06 |
| Breath Volume     | 3.382 | 22:06 |
| Air Blank         | 0.000 | 22:07 |
| Air Blank         | 0.000 | 22:09 |
| Subject Sample #2 | 0.080 | 22:09 |
| Breath Volume     | 3.675 | 22:09 |
| Air Blank         | 0.000 | 22:10 |
| Control Test      | 0.079 | 22:10 |
| Air Blank         | 0.000 | 22:11 |
| Diagnostics Check | OK    | 22:11 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEAD, JENNIFER, L

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: HEAD, JENNIFER

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26911080A2

DRY GAS EXPIRY DATE: 11/01/2013

VIOLATION CODE: DUI

ARREST TIME: 03:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEAKER

DATE/TIME OF TEST: 11/25/2011 04:44:47

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:37 |
| Air Blank         | 0.000 | 04:38 |
| Control Test      | 0.079 | 04:38 |
| Air Blank         | 0.000 | 04:39 |
| Subject Sample #1 | 0.132 | 04:39 |
| Breath Volume     | 1.277 | 04:39 |
| Air Blank         | 0.000 | 04:40 |
| Air Blank         | 0.000 | 04:42 |
| Subject Sample #2 | 0.134 | 04:43 |
| Breath Volume     | 1.167 | 04:43 |
| Air Blank         | 0.000 | 04:43 |
| Control Test      | 0.079 | 04:44 |
| Air Blank         | 0.000 | 04:44 |
| Diagnostics Check | OK    | 04:44 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHMITT, DEP, K  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FORMAL, DEPUTY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: OSTERBY

DATE/TIME OF TEST: 11/26/2011 02:41:20

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:34 |
| Air Blank         | 0.000 | 02:35 |
| Control Test      | 0.079 | 02:35 |
| Air Blank         | 0.000 | 02:36 |
| Subject Sample #1 | 0.167 | 02:36 |
| Breath Volume     | 2.160 | 02:36 |
| Air Blank         | 0.000 | 02:37 |
| Air Blank         | 0.000 | 02:39 |
| Subject Sample #2 | 0.155 | 02:39 |
| Breath Volume     | 1.496 | 02:39 |
| Air Blank         | 0.000 | 02:40 |
| Control Test      | 0.078 | 02:40 |
| Air Blank         | 0.000 | 02:41 |
| Diagnostics Check | OK    | 02:41 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHMITT, DEP, K  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: VANIK, OFFICER  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GRABER

DATE/TIME OF TEST: 11/26/2011 02:53:27

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:45 |                      |
| Air Blank         | 0.000 | 02:46 |                      |
| Control Test      | 0.077 | 02:46 |                      |
| Air Blank         | 0.000 | 02:46 |                      |
| Subject Sample #1 | 0.201 | 02:47 |                      |
| Breath Volume     | 3.281 | 02:47 |                      |
| Air Blank         | 0.000 | 02:47 |                      |
| Air Blank         | 0.000 | 02:49 |                      |
| Subject Sample #2 |       | 02:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:52 |                      |
| Air Blank         | 0.000 | 02:52 |                      |
| Control Test      | 0.078 | 02:52 |                      |
| Air Blank         | 0.000 | 02:53 |                      |
| Diagnostics Check | OK    | 02:53 |                      |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHMITT, DEP, K  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: CARDWELL, TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 11/26/2011 03:51:29

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:45 |
| Air Blank         | 0.000 | 03:45 |
| Control Test      | 0.079 | 03:45 |
| Air Blank         | 0.000 | 03:46 |
| Subject Sample #1 | 0.117 | 03:47 |
| Breath Volume     | 2.757 | 03:47 |
| Air Blank         | 0.000 | 03:47 |
| Air Blank         | 0.000 | 03:49 |
| Subject Sample #2 | 0.102 | 03:49 |
| Breath Volume     | 1.289 | 03:49 |
| Air Blank         | 0.000 | 03:50 |
| Control Test      | 0.077 | 03:50 |
| Air Blank         | 0.000 | 03:51 |
| Diagnostics Check | OK    | 03:51 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHMITT, DEP, K  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: LYNCH, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: TOMOVICH

DATE/TIME OF TEST: 11/26/2011 04:23:42

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:17 |
| Air Blank         | 0.000 | 04:17 |
| Control Test      | 0.079 | 04:18 |
| Air Blank         | 0.000 | 04:18 |
| Subject Sample #1 | 0.189 | 04:19 |
| Breath Volume     | 4.023 | 04:19 |
| Air Blank         | 0.000 | 04:19 |
| Air Blank         | 0.000 | 04:21 |
| Subject Sample #2 | 0.193 | 04:22 |
| Breath Volume     | 1.632 | 04:22 |
| Air Blank         | 0.000 | 04:22 |
| Control Test      | 0.077 | 04:23 |
| Air Blank         | 0.000 | 04:23 |
| Diagnostics Check | OK    | 04:23 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHMITT, DEP, K  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: CARRASCO, TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 17:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 11/27/2011 19:06:39

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:00 |
| Air Blank         | 0.000 | 19:01 |
| Control Test      | 0.077 | 19:01 |
| Air Blank         | 0.000 | 19:01 |
| Subject Sample #1 | 0.228 | 19:02 |
| Breath Volume     | 1.902 | 19:02 |
| Air Blank         | 0.000 | 19:02 |
| Air Blank         | 0.000 | 19:04 |
| Subject Sample #2 | 0.224 | 19:05 |
| Breath Volume     | 2.207 | 19:05 |
| Air Blank         | 0.000 | 19:05 |
| Control Test      | 0.078 | 19:06 |
| Air Blank         | 0.000 | 19:06 |
| Diagnostics Check | OK    | 19:06 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATHMAN, JAMES, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: KATHMAN, JAMES  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RODRIGUEZCAMPOS

DATE/TIME OF TEST: 11/30/2011 00:39:21

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:32 |
| Air Blank         | 0.000 | 00:32 |
| Control Test      | 0.079 | 00:33 |
| Air Blank         | 0.000 | 00:33 |
| Subject Sample #1 | 0.119 | 00:34 |
| Breath Volume     | 1.589 | 00:34 |
| Air Blank         | 0.000 | 00:35 |
| Air Blank         | 0.000 | 00:37 |
| Subject Sample #2 | 0.122 | 00:37 |
| Breath Volume     | 1.886 | 00:37 |
| Air Blank         | 0.000 | 00:38 |
| Control Test      | 0.079 | 00:38 |
| Air Blank         | 0.000 | 00:39 |
| Diagnostics Check | OK    | 00:39 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, RICHARD, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: JOHNSON, RICHARD  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: BRILEY

DATE/TIME OF TEST: 11/30/2011 20:01:23

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:55 |
| Air Blank         | 0.000 | 19:55 |
| Control Test      | 0.079 | 19:56 |
| Air Blank         | 0.000 | 19:56 |
| Subject Sample #1 | 0.167 | 19:56 |
| Breath Volume     | 1.761 | 19:56 |
| Air Blank         | 0.000 | 19:57 |
| Air Blank         | 0.000 | 19:59 |
| Subject Sample #2 | 0.168 | 19:59 |
| Breath Volume     | 1.898 | 19:59 |
| Air Blank         | 0.000 | 20:00 |
| Control Test      | 0.083 | 20:00 |
| Air Blank         | 0.000 | 20:01 |
| Diagnostics Check | OK    | 20:01 |

---

Data Download Date/Time 12/20/2011 12:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DILLON, LARRY, G

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: FIGUEROA, HECTOR

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 00:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: METHENEY

DATE/TIME OF TEST: 11/01/2011 01:21:01

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:18 |                      |
| Air Blank         | 0.000 | 01:18 |                      |
| Control Test      | 0.079 | 01:18 |                      |
| Air Blank         | 0.000 | 01:19 |                      |
| Subject Sample #1 |       | 01:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:19 |                      |
| Air Blank         | 0.000 | 01:20 |                      |
| Control Test      | 0.079 | 01:20 |                      |
| Air Blank         | 0.000 | 01:20 |                      |
| Diagnostics Check | OK    | 01:21 |                      |

---

Data Download Date/Time 11/30/2011 07:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEMENY, ANTHONY, M

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: TAYLOR, KEITH

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 04:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 11/05/2011 06:19:35

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 06:13 |
| Air Blank         | 0.000 | 06:13 |
| Control Test      | 0.078 | 06:14 |
| Air Blank         | 0.000 | 06:14 |
| Subject Sample #1 | 0.181 | 06:15 |
| Breath Volume     | 3.988 | 06:15 |
| Air Blank         | 0.000 | 06:15 |
| Air Blank         | 0.000 | 06:17 |
| Subject Sample #2 | 0.178 | 06:18 |
| Breath Volume     | 2.277 | 06:18 |
| Air Blank         | 0.000 | 06:18 |
| Control Test      | 0.078 | 06:19 |
| Air Blank         | 0.000 | 06:19 |
| Diagnostics Check | OK    | 06:19 |

---

Data Download Date/Time 11/30/2011 07:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEMENY, ANTHONY, M

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: ROBINSON, RUSSELL

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 02:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CLOUGH

DATE/TIME OF TEST: 11/10/2011 02:56:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:54 |                      |
| Air Blank         | 0.000 | 02:54 |                      |
| Control Test      | 0.079 | 02:54 |                      |
| Air Blank         | 0.000 | 02:55 |                      |
| Subject Sample #1 |       | 02:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:55 |                      |
| Air Blank         | 0.000 | 02:56 |                      |
| Control Test      | 0.078 | 02:56 |                      |
| Air Blank         | 0.000 | 02:56 |                      |
| Diagnostics Check | OK    | 02:56 |                      |

---

Data Download Date/Time 11/30/2011 07:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DILLON, LARRY, G

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: DILLON, LARRY

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 22:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: JEANTY

DATE/TIME OF TEST: 11/12/2011 23:24:24

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:18 |
| Air Blank         | 0.000 | 23:18 |
| Control Test      | 0.078 | 23:18 |
| Air Blank         | 0.000 | 23:19 |
| Subject Sample #1 | 0.301 | 23:19 |
| Breath Volume     | 3.656 | 23:19 |
| Air Blank         | 0.000 | 23:20 |
| Air Blank         | 0.000 | 23:22 |
| Subject Sample #2 | 0.290 | 23:22 |
| Breath Volume     | 3.503 | 23:22 |
| Air Blank         | 0.000 | 23:23 |
| Control Test      | 0.077 | 23:23 |
| Air Blank         | 0.000 | 23:24 |
| Diagnostics Check | OK    | 23:24 |

---

Data Download Date/Time 11/30/2011 07:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEMENY, ANTHONY, M

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: PARKER, CLAY

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 04:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 11/21/2011 05:10:03

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 04:57 |                |
| Air Blank         | 0.000 | 04:57 |                |
| Control Test      | 0.077 | 04:57 |                |
| Air Blank         | 0.000 | 04:58 |                |
| Subject Sample #1 | 0.212 | 04:58 |                |
| Breath Volume     | 2.503 | 04:58 |                |
| Air Blank         | 0.000 | 04:59 |                |
| Air Blank         | 0.000 | 05:01 |                |
| Subject Sample #2 | 0.177 | 05:04 | Volume Not Met |
| Breath Volume     | 0.101 | 05:04 |                |
| Air Blank         | 0.000 | 05:05 |                |
| Air Blank         | 0.000 | 05:07 |                |
| Subject Sample #3 | 0.215 | 05:08 |                |
| Breath Volume     | 1.511 | 05:08 |                |
| Air Blank         | 0.000 | 05:09 |                |
| Control Test      | 0.077 | 05:09 |                |
| Air Blank         | 0.000 | 05:09 |                |
| Diagnostics Check | OK    | 05:10 |                |

---

Data Download Date/Time 11/30/2011 07:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEMENY, ANTHONY, M

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: FIGUEROA, HECTOR

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 23:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: ABRAHAMSON

DATE/TIME OF TEST: 11/24/2011 00:05:00

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 00:04 |            |
| Air Blank         |       | 00:04 | RFI Detect |
| Air Blank         | 0.000 | 00:05 |            |

---

|                         |            |       |
|-------------------------|------------|-------|
| Data Download Date/Time | 11/30/2011 | 07:33 |
|-------------------------|------------|-------|

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEMENY, ANTHONY, M

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: FIGUEROA, HECTOR

ARRESTING OFFICER AGENCY: Mount Dora PD

LAST AGENCY INSPECTION: 10/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 23:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: ABRAHAMSON

DATE/TIME OF TEST: 11/24/2011 00:16:27

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:06 |
| Air Blank         | 0.000 | 00:07 |
| Control Test      | 0.078 | 00:07 |
| Air Blank         | 0.000 | 00:07 |
| Subject Sample #1 | 0.190 | 00:10 |
| Breath Volume     | 1.718 | 00:10 |
| Air Blank         | 0.000 | 00:11 |
| Air Blank         | 0.000 | 00:13 |
| Subject Sample #2 | 0.183 | 00:14 |
| Breath Volume     | 1.445 | 00:14 |
| Air Blank         | 0.000 | 00:15 |
| Control Test      | 0.077 | 00:15 |
| Air Blank         | 0.000 | 00:16 |
| Diagnostics Check | OK    | 00:16 |

---

Data Download Date/Time 11/30/2011 07:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANLEY, KURTISS, L  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ELLIS, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: PUTNAM

DATE/TIME OF TEST: 11/01/2011 04:42:50

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:36 |
| Air Blank         | 0.000 | 04:37 |
| Control Test      | 0.082 | 04:37 |
| Air Blank         | 0.000 | 04:37 |
| Subject Sample #1 | 0.129 | 04:38 |
| Breath Volume     | 2.988 | 04:38 |
| Air Blank         | 0.000 | 04:38 |
| Air Blank         | 0.000 | 04:40 |
| Subject Sample #2 | 0.118 | 04:41 |
| Breath Volume     | 3.125 | 04:41 |
| Air Blank         | 0.000 | 04:41 |
| Control Test      | 0.081 | 04:42 |
| Air Blank         | 0.000 | 04:42 |
| Diagnostics Check | OK    | 04:42 |

---

Data Download Date/Time 11/17/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTLER, CHRIS, T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BUTLER, CHRIS  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 09:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 11/01/2011 10:24:03

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 10:17 |
| Air Blank         | 0.000 | 10:18 |
| Control Test      | 0.082 | 10:18 |
| Air Blank         | 0.000 | 10:18 |
| Subject Sample #1 | 0.000 | 10:19 |
| Breath Volume     | 4.320 | 10:19 |
| Air Blank         | 0.000 | 10:19 |
| Air Blank         | 0.000 | 10:21 |
| Subject Sample #2 | 0.000 | 10:22 |
| Breath Volume     | 1.121 | 10:22 |
| Air Blank         | 0.000 | 10:23 |
| Control Test      | 0.082 | 10:23 |
| Air Blank         | 0.000 | 10:23 |
| Diagnostics Check | OK    | 10:24 |

---

Data Download Date/Time 11/17/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANKS, SIMON, A  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FRANKS, SIMON  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 11:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 11/01/2011 12:29:54

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 12:23 |
| Air Blank         | 0.000 | 12:23 |
| Control Test      | 0.082 | 12:24 |
| Air Blank         | 0.000 | 12:24 |
| Subject Sample #1 | 0.233 | 12:25 |
| Breath Volume     | 2.585 | 12:25 |
| Air Blank         | 0.000 | 12:26 |
| Air Blank         | 0.000 | 12:27 |
| Subject Sample #2 | 0.233 | 12:28 |
| Breath Volume     | 2.828 | 12:28 |
| Air Blank         | 0.000 | 12:29 |
| Control Test      | 0.080 | 12:29 |
| Air Blank         | 0.000 | 12:29 |
| Diagnostics Check | OK    | 12:29 |

---

Data Download Date/Time 11/17/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAY, JOHN, T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BARTOLOTTA, OFFICER  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: BURGE

DATE/TIME OF TEST: 11/03/2011 22:41:56

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:35 |
| Air Blank         | 0.000 | 22:36 |
| Control Test      | 0.080 | 22:36 |
| Air Blank         | 0.000 | 22:37 |
| Subject Sample #1 | 0.239 | 22:37 |
| Breath Volume     | 2.972 | 22:37 |
| Air Blank         | 0.000 | 22:38 |
| Air Blank         | 0.000 | 22:40 |
| Subject Sample #2 | 0.231 | 22:40 |
| Breath Volume     | 2.152 | 22:40 |
| Air Blank         | 0.000 | 22:41 |
| Control Test      | 0.079 | 22:41 |
| Air Blank         | 0.000 | 22:41 |
| Diagnostics Check | OK    | 22:41 |

---

Data Download Date/Time 11/17/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GELLER, DEBORAH, A  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ELLIS, DEPUTY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: VEROST

DATE/TIME OF TEST: 11/05/2011 23:16:19

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 23:02 |                |
| Air Blank         | 0.000 | 23:03 |                |
| Control Test      | 0.082 | 23:03 |                |
| Air Blank         | 0.000 | 23:03 |                |
| Subject Sample #1 | 0.154 | 23:07 | Volume Not Met |
| Breath Volume     | 0.726 | 23:07 |                |
| Air Blank         | 0.000 | 23:07 |                |
| Air Blank         | 0.000 | 23:09 |                |
| Subject Sample #2 | 0.146 | 23:11 |                |
| Breath Volume     | 2.199 | 23:11 |                |
| Air Blank         | 0.000 | 23:12 |                |
| Air Blank         | 0.000 | 23:14 |                |
| Subject Sample #3 | 0.154 | 23:14 |                |
| Breath Volume     | 1.828 | 23:14 |                |
| Air Blank         | 0.000 | 23:15 |                |
| Control Test      | 0.079 | 23:15 |                |
| Air Blank         | 0.000 | 23:16 |                |
| Diagnostics Check | OK    | 23:16 |                |

---

Data Download Date/Time 11/17/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BYRD, J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BYRD, J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CHU

DATE/TIME OF TEST: 11/06/2011 04:09:57

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:03 |
| Air Blank         | 0.000 | 04:04 |
| Control Test      | 0.081 | 04:04 |
| Air Blank         | 0.000 | 04:04 |
| Subject Sample #1 | 0.191 | 04:05 |
| Breath Volume     | 1.347 | 04:05 |
| Air Blank         | 0.000 | 04:06 |
| Air Blank         | 0.000 | 04:08 |
| Subject Sample #2 | 0.177 | 04:08 |
| Breath Volume     | 1.582 | 04:08 |
| Air Blank         | 0.000 | 04:09 |
| Control Test      | 0.080 | 04:09 |
| Air Blank         | 0.000 | 04:09 |
| Diagnostics Check | OK    | 04:09 |

---

Data Download Date/Time 11/17/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GELLER, DEBORAH, A  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FUSSELL, OFFICER  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 11/07/2011 00:09:00

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:02 |
| Air Blank         | 0.000 | 00:02 |
| Control Test      | 0.082 | 00:03 |
| Air Blank         | 0.000 | 00:03 |
| Subject Sample #1 | 0.191 | 00:04 |
| Breath Volume     | 1.718 | 00:04 |
| Air Blank         | 0.000 | 00:04 |
| Air Blank         | 0.000 | 00:06 |
| Subject Sample #2 | 0.189 | 00:07 |
| Breath Volume     | 2.246 | 00:07 |
| Air Blank         | 0.000 | 00:08 |
| Control Test      | 0.081 | 00:08 |
| Air Blank         | 0.000 | 00:08 |
| Diagnostics Check | OK    | 00:09 |

---

Data Download Date/Time 11/17/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SURRAN, DEPUTY, C  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ELLIS, ERIC  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 11/07/2011 06:47:06

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 06:35 |                |
| Air Blank         | 0.000 | 06:35 |                |
| Control Test      | 0.082 | 06:36 |                |
| Air Blank         | 0.000 | 06:36 |                |
| Subject Sample #1 | 0.158 | 06:39 | Volume Not Met |
| Breath Volume     | 0.097 | 06:39 |                |
| Air Blank         | 0.000 | 06:40 |                |
| Air Blank         | 0.000 | 06:42 |                |
| Subject Sample #2 | 0.174 | 06:45 | Volume Not Met |
| Breath Volume     | 0.410 | 06:45 |                |
| Air Blank         | 0.000 | 06:46 |                |
| Control Test      | 0.080 | 06:46 |                |
| Air Blank         | 0.000 | 06:47 |                |
| Diagnostics Check | OK    | 06:47 |                |

---

Data Download Date/Time 11/17/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SURRAN, DEPUTY, C  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ELLIS, ERIC  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 11/07/2011 06:59:56

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 06:51 |
| Air Blank         | 0.000 | 06:52 |
| Control Test      | 0.081 | 06:52 |
| Air Blank         | 0.000 | 06:52 |
| Subject Sample #1 | 0.182 | 06:53 |
| Breath Volume     | 1.500 | 06:53 |
| Air Blank         | 0.000 | 06:54 |
| Air Blank         | 0.000 | 06:56 |
| Subject Sample #2 | 0.164 | 06:58 |
| Breath Volume     | 1.472 | 06:58 |
| Air Blank         | 0.000 | 06:59 |
| Control Test      | 0.080 | 06:59 |
| Air Blank         | 0.000 | 06:59 |
| Diagnostics Check | OK    | 06:59 |

---

Data Download Date/Time 11/17/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANLEY, KURTISS, L  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: CASTIGLIONE, OFC  
ARRESTING OFFICER AGENCY: Venice PD  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SMALLY

DATE/TIME OF TEST: 11/10/2011 04:00:31

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:54 |
| Air Blank         | 0.000 | 03:54 |
| Control Test      | 0.082 | 03:54 |
| Air Blank         | 0.000 | 03:55 |
| Subject Sample #1 | 0.147 | 03:56 |
| Breath Volume     | 1.527 | 03:56 |
| Air Blank         | 0.000 | 03:56 |
| Air Blank         | 0.000 | 03:58 |
| Subject Sample #2 | 0.146 | 03:59 |
| Breath Volume     | 1.859 | 03:59 |
| Air Blank         | 0.000 | 03:59 |
| Control Test      | 0.082 | 03:59 |
| Air Blank         | 0.000 | 04:00 |
| Diagnostics Check | OK    | 04:00 |

---

Data Download Date/Time 11/17/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BYRD, J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BYRD, J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 11/18/2011 01:03:56

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:53 |
| Air Blank         | 0.000 | 00:54 |
| Control Test      | 0.081 | 00:54 |
| Air Blank         | 0.000 | 00:55 |
| Subject Sample #1 | 0.185 | 00:57 |
| Breath Volume     | 1.101 | 00:57 |
| Air Blank         | 0.000 | 00:58 |
| Air Blank         | 0.000 | 01:00 |
| Subject Sample #2 | 0.179 | 01:02 |
| Breath Volume     | 1.300 | 01:02 |
| Air Blank         | 0.000 | 01:03 |
| Control Test      | 0.077 | 01:03 |
| Air Blank         | 0.000 | 01:03 |
| Diagnostics Check | OK    | 01:03 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANLEY, KURTISS, L  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ELLIS, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 11/19/2011 04:22:57

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:15 |
| Air Blank         | 0.000 | 04:16 |
| Control Test      | 0.081 | 04:16 |
| Air Blank         | 0.000 | 04:17 |
| Subject Sample #1 | 0.163 | 04:17 |
| Breath Volume     | 1.589 | 04:17 |
| Air Blank         | 0.000 | 04:18 |
| Air Blank         | 0.000 | 04:20 |
| Subject Sample #2 | 0.163 | 04:21 |
| Breath Volume     | 3.996 | 04:21 |
| Air Blank         | 0.000 | 04:22 |
| Control Test      | 0.080 | 04:22 |
| Air Blank         | 0.000 | 04:22 |
| Diagnostics Check | OK    | 04:22 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEAD, JENNIFER, L  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: HEAD, JENNIFER  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 11/20/2011 01:47:37

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:34 |                    |
| Air Blank         | 0.000 | 01:35 |                    |
| Control Test      | 0.082 | 01:35 |                    |
| Air Blank         | 0.000 | 01:35 |                    |
| Subject Sample #1 |       | 01:39 | No Sample Provided |
| Breath Volume     | 0.000 | 01:39 |                    |
| Air Blank         | 0.000 | 01:39 |                    |
| Air Blank         | 0.000 | 01:41 |                    |
| Subject Sample #2 | 0.109 | 01:42 |                    |
| Breath Volume     | 1.542 | 01:42 |                    |
| Air Blank         | 0.000 | 01:43 |                    |
| Air Blank         | 0.000 | 01:45 |                    |
| Subject Sample #3 | 0.117 | 01:46 |                    |
| Breath Volume     | 3.582 | 01:46 |                    |
| Air Blank         | 0.000 | 01:46 |                    |
| Control Test      | 0.080 | 01:47 |                    |
| Air Blank         | 0.000 | 01:47 |                    |
| Diagnostics Check | OK    | 01:47 |                    |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEAD, JENNIFER, L  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ELLIS, ERIC  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DEY

DATE/TIME OF TEST: 11/20/2011 02:13:19

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:06 |
| Air Blank         | 0.000 | 02:07 |
| Control Test      | 0.081 | 02:07 |
| Air Blank         | 0.000 | 02:07 |
| Subject Sample #1 | 0.136 | 02:08 |
| Breath Volume     | 4.902 | 02:08 |
| Air Blank         | 0.000 | 02:09 |
| Air Blank         | 0.000 | 02:11 |
| Subject Sample #2 | 0.125 | 02:11 |
| Breath Volume     | 2.328 | 02:11 |
| Air Blank         | 0.000 | 02:12 |
| Control Test      | 0.080 | 02:12 |
| Air Blank         | 0.000 | 02:13 |
| Diagnostics Check | OK    | 02:13 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANLEY, KURTISS, L  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FUSSELL, OFC  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: CHITTESTER

DATE/TIME OF TEST: 11/21/2011 01:56:07

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:49 |
| Air Blank         | 0.000 | 01:49 |
| Control Test      | 0.081 | 01:49 |
| Air Blank         | 0.000 | 01:50 |
| Subject Sample #1 | 0.209 | 01:51 |
| Breath Volume     | 2.203 | 01:51 |
| Air Blank         | 0.000 | 01:52 |
| Air Blank         | 0.000 | 01:53 |
| Subject Sample #2 | 0.201 | 01:54 |
| Breath Volume     | 1.789 | 01:54 |
| Air Blank         | 0.000 | 01:55 |
| Control Test      | 0.080 | 01:55 |
| Air Blank         | 0.000 | 01:56 |
| Diagnostics Check | OK    | 01:56 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GELLER, DEBORAH, A  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FORMAL, DEPUTY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: EWING

DATE/TIME OF TEST: 11/24/2011 05:12:24

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:05 |
| Air Blank         | 0.000 | 05:06 |
| Control Test      | 0.081 | 05:06 |
| Air Blank         | 0.000 | 05:07 |
| Subject Sample #1 | 0.185 | 05:07 |
| Breath Volume     | 3.816 | 05:07 |
| Air Blank         | 0.000 | 05:08 |
| Air Blank         | 0.000 | 05:10 |
| Subject Sample #2 | 0.187 | 05:10 |
| Breath Volume     | 3.320 | 05:10 |
| Air Blank         | 0.000 | 05:11 |
| Control Test      | 0.081 | 05:11 |
| Air Blank         | 0.000 | 05:12 |
| Diagnostics Check | OK    | 05:12 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GELLER, DEBORAH, A  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FORMAL, DEPUTY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: AHRENS

DATE/TIME OF TEST: 11/25/2011 00:06:55

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:00 |
| Air Blank         | 0.000 | 00:00 |
| Control Test      | 0.082 | 00:01 |
| Air Blank         | 0.000 | 00:01 |
| Subject Sample #1 | 0.000 | 00:02 |
| Breath Volume     | 1.718 | 00:02 |
| Air Blank         | 0.000 | 00:02 |
| Air Blank         | 0.000 | 00:04 |
| Subject Sample #2 | 0.000 | 00:05 |
| Breath Volume     | 2.980 | 00:05 |
| Air Blank         | 0.000 | 00:06 |
| Control Test      | 0.081 | 00:06 |
| Air Blank         | 0.000 | 00:06 |
| Diagnostics Check | OK    | 00:06 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECK, DENNIS, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARIER, CHRISTOPHER  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 09:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: STINER

DATE/TIME OF TEST: 11/28/2011 10:52:15

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 10:50 |                           |
| Air Blank         | 0.000 | 10:51 |                           |
| Control Test      | 0.008 | 10:51 | Control Outside Tolerance |
| Air Blank         | 0.000 | 10:52 |                           |

---

Data Download Date/Time 12/20/2011 14:23

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECK, DENNIS, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARIER, CHRISTOPHER  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 09:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: STINER

DATE/TIME OF TEST: 11/28/2011 10:56:47

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 10:54 |                 |
| Air Blank         | 0.000 | 10:55 |                 |
| Control Test      | 0.082 | 10:55 |                 |
| Air Blank         | 0.000 | 10:56 |                 |
|                   |       | 10:56 | Improper Sample |
| Air Blank         | 0.000 | 10:56 |                 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECK, DENNIS, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARIER, CHRISTOPHER  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 09:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: STINER

DATE/TIME OF TEST: 11/28/2011 11:00:40

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 11:00 |                 |
| Air Blank         |       | 11:00 | Improper Sample |
| Air Blank         | 0.000 | 11:00 |                 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECK, DENNIS, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARIER, CHRISTOPHER  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 09:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: STINER

DATE/TIME OF TEST: 11/28/2011 11:08:55

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 11:03 |
| Air Blank         | 0.000 | 11:03 |
| Control Test      | 0.081 | 11:03 |
| Air Blank         | 0.000 | 11:04 |
| Subject Sample #1 | 0.000 | 11:04 |
| Breath Volume     | 3.320 | 11:04 |
| Air Blank         | 0.000 | 11:05 |
| Air Blank         | 0.000 | 11:07 |
| Subject Sample #2 | 0.000 | 11:07 |
| Breath Volume     | 4.070 | 11:07 |
| Air Blank         | 0.000 | 11:08 |
| Control Test      | 0.081 | 11:08 |
| Air Blank         | 0.000 | 11:08 |
| Diagnostics Check | OK    | 11:08 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANLEY, KURTISS, L  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: LAROCK, OFC  
ARRESTING OFFICER AGENCY: Venice PD  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 11/28/2011 23:52:55

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:46 |
| Air Blank         | 0.000 | 23:47 |
| Control Test      | 0.080 | 23:47 |
| Air Blank         | 0.000 | 23:47 |
| Subject Sample #1 | 0.085 | 23:48 |
| Breath Volume     | 4.210 | 23:48 |
| Air Blank         | 0.000 | 23:48 |
| Air Blank         | 0.000 | 23:50 |
| Subject Sample #2 | 0.079 | 23:51 |
| Breath Volume     | 4.113 | 23:51 |
| Air Blank         | 0.000 | 23:52 |
| Control Test      | 0.081 | 23:52 |
| Air Blank         | 0.000 | 23:52 |
| Diagnostics Check | OK    | 23:52 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COSS, MICHAEL, 6  
BREATH TEST OPERATOR AGENCY: North Port PD  
ARRESTING OFFICER: COSS, MICHAEL  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 11/29/2011 19:18:41

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:12 |
| Air Blank         | 0.000 | 19:13 |
| Control Test      | 0.081 | 19:13 |
| Air Blank         | 0.000 | 19:13 |
| Subject Sample #1 | 0.170 | 19:14 |
| Breath Volume     | 4.171 | 19:14 |
| Air Blank         | 0.000 | 19:14 |
| Air Blank         | 0.000 | 19:16 |
| Subject Sample #2 | 0.172 | 19:17 |
| Breath Volume     | 3.367 | 19:17 |
| Air Blank         | 0.000 | 19:17 |
| Control Test      | 0.082 | 19:18 |
| Air Blank         | 0.000 | 19:18 |
| Diagnostics Check | OK    | 19:18 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, RICHARD, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: JOHNSON, RICHARD  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: TOTH

DATE/TIME OF TEST: 11/29/2011 21:26:41

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 21:20 |
| Air Blank         | 0.000 | 21:20 |
| Control Test      | 0.081 | 21:20 |
| Air Blank         | 0.000 | 21:21 |
| Subject Sample #1 | 0.136 | 21:21 |
| Breath Volume     | 1.597 | 21:21 |
| Air Blank         | 0.000 | 21:22 |
| Air Blank         | 0.000 | 21:24 |
| Subject Sample #2 | 0.129 | 21:25 |
| Breath Volume     | 1.875 | 21:25 |
| Air Blank         | 0.000 | 21:25 |
| Control Test      | 0.080 | 21:26 |
| Air Blank         | 0.000 | 21:26 |
| Diagnostics Check | OK    | 21:26 |

---

Data Download Date/Time 12/20/2011 14:23

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANLEY, KURTISS, L  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: TUGGLE, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 11/30/2011 01:49:31

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:42 |
| Air Blank         | 0.000 | 01:42 |
| Control Test      | 0.081 | 01:43 |
| Air Blank         | 0.000 | 01:43 |
| Subject Sample #1 | 0.136 | 01:45 |
| Breath Volume     | 3.402 | 01:45 |
| Air Blank         | 0.000 | 01:45 |
| Air Blank         | 0.000 | 01:47 |
| Subject Sample #2 | 0.126 | 01:48 |
| Breath Volume     | 3.539 | 01:48 |
| Air Blank         | 0.000 | 01:48 |
| Control Test      | 0.078 | 01:48 |
| Air Blank         | 0.000 | 01:49 |
| Diagnostics Check | OK    | 01:49 |

---

Data Download Date/Time 12/20/2011 14:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER, ROGER, G  
BREATH TEST OPERATOR AGENCY: FDLE  
ARRESTING OFFICER: TEST, TEST  
ARRESTING OFFICER AGENCY: FDLE  
LAST AGENCY INSPECTION: 11/02/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG032601  
DRY GAS EXPIRY DATE: 11/22/2012  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 07:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/02/2011 08:55:56

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 08:47 |
| Air Blank         | 0.000 | 08:47 |
| Control Test      | 0.081 | 08:48 |
| Air Blank         | 0.000 | 08:48 |
| Subject Sample #1 | 0.079 | 08:50 |
| Breath Volume     | 13.30 | 08:50 |
| Air Blank         | 0.000 | 08:50 |
| Air Blank         | 0.000 | 08:52 |
| Subject Sample #2 | 0.075 | 08:54 |
| Breath Volume     | 16.02 | 08:54 |
| Air Blank         | 0.000 | 08:55 |
| Control Test      | 0.080 | 08:55 |
| Air Blank         | 0.000 | 08:55 |
| Diagnostics Check | OK    | 08:55 |

---

Data Download Date/Time 11/02/2011 14:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, DEP, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: JOHNSON, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 11/01/2011 23:28:44

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:22 |
| Air Blank         | 0.000 | 23:22 |
| Control Test      | 0.079 | 23:22 |
| Air Blank         | 0.000 | 23:23 |
| Subject Sample #1 | 0.178 | 23:24 |
| Breath Volume     | 2.578 | 23:24 |
| Air Blank         | 0.000 | 23:24 |
| Air Blank         | 0.000 | 23:26 |
| Subject Sample #2 | 0.171 | 23:27 |
| Breath Volume     | 2.761 | 23:27 |
| Air Blank         | 0.000 | 23:27 |
| Control Test      | 0.080 | 23:28 |
| Air Blank         | 0.000 | 23:28 |
| Diagnostics Check | OK    | 23:28 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, DEP, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: JOHNSON, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MURCH

DATE/TIME OF TEST: 11/02/2011 02:58:34

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:52 |
| Air Blank         | 0.000 | 02:52 |
| Control Test      | 0.080 | 02:53 |
| Air Blank         | 0.000 | 02:53 |
| Subject Sample #1 | 0.089 | 02:54 |
| Breath Volume     | 3.062 | 02:54 |
| Air Blank         | 0.000 | 02:54 |
| Air Blank         | 0.000 | 02:56 |
| Subject Sample #2 | 0.084 | 02:57 |
| Breath Volume     | 3.707 | 02:57 |
| Air Blank         | 0.000 | 02:57 |
| Control Test      | 0.080 | 02:57 |
| Air Blank         | 0.000 | 02:58 |
| Diagnostics Check | OK    | 02:58 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHMITT, DEP, K  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: DUNE, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HILLEGAS

DATE/TIME OF TEST: 11/03/2011 04:46:06

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:39 |
| Air Blank         | 0.000 | 04:40 |
| Control Test      | 0.080 | 04:40 |
| Air Blank         | 0.000 | 04:40 |
| Subject Sample #1 | 0.175 | 04:41 |
| Breath Volume     | 1.164 | 04:41 |
| Air Blank         | 0.000 | 04:42 |
| Air Blank         | 0.000 | 04:44 |
| Subject Sample #2 | 0.167 | 04:44 |
| Breath Volume     | 1.554 | 04:44 |
| Air Blank         | 0.000 | 04:45 |
| Control Test      | 0.080 | 04:45 |
| Air Blank         | 0.000 | 04:46 |
| Diagnostics Check | OK    | 04:46 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, DEP, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FORMAL, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CORNICELLI

DATE/TIME OF TEST: 11/06/2011 04:57:36

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:51 |
| Air Blank         | 0.000 | 04:51 |
| Control Test      | 0.081 | 04:52 |
| Air Blank         | 0.000 | 04:52 |
| Subject Sample #1 | 0.153 | 04:53 |
| Breath Volume     | 2.621 | 04:53 |
| Air Blank         | 0.000 | 04:53 |
| Air Blank         | 0.000 | 04:55 |
| Subject Sample #2 | 0.150 | 04:56 |
| Breath Volume     | 3.738 | 04:56 |
| Air Blank         | 0.000 | 04:56 |
| Control Test      | 0.080 | 04:57 |
| Air Blank         | 0.000 | 04:57 |
| Diagnostics Check | OK    | 04:57 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATHMAN, JAMES, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: KATHMAN, JAMES  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/08/2011 22:23:58

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:15 |
| Air Blank         | 0.000 | 22:15 |
| Control Test      | 0.080 | 22:16 |
| Air Blank         | 0.000 | 22:16 |
| Subject Sample #1 | 0.181 | 22:19 |
| Breath Volume     | 1.257 | 22:19 |
| Air Blank         | 0.000 | 22:19 |
| Air Blank         | 0.000 | 22:21 |
| Subject Sample #2 | 0.190 | 22:22 |
| Breath Volume     | 1.398 | 22:22 |
| Air Blank         | 0.000 | 22:23 |
| Control Test      | 0.078 | 22:23 |
| Air Blank         | 0.000 | 22:23 |
| Diagnostics Check | OK    | 22:23 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIANDI, JAMIE, B  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FORMAL, DEPUTY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 11/09/2011 00:59:27

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:50 |
| Air Blank         | 0.000 | 00:50 |
| Control Test      | 0.080 | 00:51 |
| Air Blank         | 0.000 | 00:51 |
| Subject Sample #1 | 0.235 | 00:54 |
| Breath Volume     | 1.453 | 00:54 |
| Air Blank         | 0.000 | 00:54 |
| Air Blank         | 0.000 | 00:56 |
| Subject Sample #2 | 0.240 | 00:57 |
| Breath Volume     | 1.675 | 00:57 |
| Air Blank         | 0.000 | 00:58 |
| Control Test      | 0.078 | 00:58 |
| Air Blank         | 0.000 | 00:59 |
| Diagnostics Check | OK    | 00:59 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATHMAN, JAMES, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: KATHMAN, JAMES  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: LARRANAGA

DATE/TIME OF TEST: 11/09/2011 23:00:42

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:52 |
| Air Blank         | 0.000 | 22:53 |
| Control Test      | 0.080 | 22:53 |
| Air Blank         | 0.000 | 22:53 |
| Subject Sample #1 | 0.116 | 22:55 |
| Breath Volume     | 2.218 | 22:55 |
| Air Blank         | 0.000 | 22:56 |
| Air Blank         | 0.000 | 22:58 |
| Subject Sample #2 | 0.118 | 22:59 |
| Breath Volume     | 2.335 | 22:59 |
| Air Blank         | 0.000 | 22:59 |
| Control Test      | 0.079 | 23:00 |
| Air Blank         | 0.000 | 23:00 |
| Diagnostics Check | OK    | 23:00 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, DEP, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: JOHNSON, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 11/10/2011 00:04:59

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:58 |
| Air Blank         | 0.000 | 23:59 |
| Control Test      | 0.080 | 23:59 |
| Air Blank         | 0.000 | 00:00 |
| Subject Sample #1 | 0.143 | 00:00 |
| Breath Volume     | 4.222 | 00:00 |
| Air Blank         | 0.000 | 00:01 |
| Air Blank         | 0.000 | 00:03 |
| Subject Sample #2 | 0.148 | 00:03 |
| Breath Volume     | 4.035 | 00:03 |
| Air Blank         | 0.000 | 00:04 |
| Control Test      | 0.079 | 00:04 |
| Air Blank         | 0.000 | 00:04 |
| Diagnostics Check | OK    | 00:04 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE, MARK,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FELTOVIC, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: GUAGENTI

DATE/TIME OF TEST: 11/10/2011 03:28:30

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:22 |
| Air Blank         | 0.000 | 03:22 |
| Control Test      | 0.080 | 03:23 |
| Air Blank         | 0.000 | 03:23 |
| Subject Sample #1 | 0.133 | 03:24 |
| Breath Volume     | 1.527 | 03:24 |
| Air Blank         | 0.000 | 03:24 |
| Air Blank         | 0.000 | 03:26 |
| Subject Sample #2 | 0.133 | 03:27 |
| Breath Volume     | 1.167 | 03:27 |
| Air Blank         | 0.000 | 03:27 |
| Control Test      | 0.080 | 03:27 |
| Air Blank         | 0.000 | 03:28 |
| Diagnostics Check | OK    | 03:28 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATHMAN, JAMES, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: KATHMAN, JAMES  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: PAPADANTONAKIS

DATE/TIME OF TEST: 11/11/2011 02:48:41

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:41 |
| Air Blank         | 0.000 | 02:42 |
| Control Test      | 0.080 | 02:42 |
| Air Blank         | 0.000 | 02:42 |
| Subject Sample #1 | 0.134 | 02:43 |
| Breath Volume     | 2.300 | 02:43 |
| Air Blank         | 0.000 | 02:44 |
| Air Blank         | 0.000 | 02:46 |
| Subject Sample #2 | 0.129 | 02:47 |
| Breath Volume     | 2.101 | 02:47 |
| Air Blank         | 0.000 | 02:47 |
| Control Test      | 0.080 | 02:48 |
| Air Blank         | 0.000 | 02:48 |
| Diagnostics Check | OK    | 02:48 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, RICHARD, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: JOHNSON, RICHARD  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/11/2011 03:21:59

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:15 |
| Air Blank         | 0.000 | 03:15 |
| Control Test      | 0.079 | 03:16 |
| Air Blank         | 0.000 | 03:16 |
| Subject Sample #1 | 0.049 | 03:17 |
| Breath Volume     | 2.464 | 03:17 |
| Air Blank         | 0.000 | 03:18 |
| Air Blank         | 0.000 | 03:20 |
| Subject Sample #2 | 0.048 | 03:20 |
| Breath Volume     | 2.500 | 03:20 |
| Air Blank         | 0.000 | 03:21 |
| Control Test      | 0.080 | 03:21 |
| Air Blank         | 0.000 | 03:21 |
| Diagnostics Check | OK    | 03:21 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BYRD, J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BYRD, J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CUDD

DATE/TIME OF TEST: 11/11/2011 03:33:47

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:27 |
| Air Blank         | 0.000 | 03:27 |
| Control Test      | 0.080 | 03:27 |
| Air Blank         | 0.000 | 03:28 |
| Subject Sample #1 | 0.187 | 03:29 |
| Breath Volume     | 1.757 | 03:29 |
| Air Blank         | 0.000 | 03:29 |
| Air Blank         | 0.000 | 03:31 |
| Subject Sample #2 | 0.206 | 03:32 |
| Breath Volume     | 2.894 | 03:32 |
| Air Blank         | 0.000 | 03:32 |
| Control Test      | 0.080 | 03:33 |
| Air Blank         | 0.000 | 03:33 |
| Diagnostics Check | OK    | 03:33 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATHMAN, JAMES, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: KATHMAN, JAMES  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 11/11/2011 21:05:06

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 20:53 |                |
| Air Blank         | 0.000 | 20:54 |                |
| Control Test      | 0.080 | 20:54 |                |
| Air Blank         | 0.000 | 20:54 |                |
| Subject Sample #1 | 0.158 | 20:57 | Volume Not Met |
| Breath Volume     | 0.140 | 20:57 |                |
| Air Blank         | 0.000 | 20:58 |                |
| Air Blank         | 0.000 | 21:00 |                |
| Subject Sample #2 | 0.165 | 21:03 | Volume Not Met |
| Breath Volume     | 0.183 | 21:03 |                |
| Air Blank         | 0.000 | 21:04 |                |
| Control Test      | 0.080 | 21:04 |                |
| Air Blank         | 0.000 | 21:05 |                |
| Diagnostics Check | OK    | 21:05 |                |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATHMAN, JAMES, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: KATHMAN, JAMES  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 11/11/2011 21:18:11

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 21:06 |                |
| Air Blank         | 0.000 | 21:07 |                |
| Control Test      | 0.080 | 21:07 |                |
| Air Blank         | 0.000 | 21:07 |                |
| Subject Sample #1 | 0.175 | 21:11 | Volume Not Met |
| Breath Volume     | 0.789 | 21:11 |                |
| Air Blank         | 0.000 | 21:11 |                |
| Air Blank         | 0.000 | 21:13 |                |
| Subject Sample #2 | 0.165 | 21:16 | Volume Not Met |
| Breath Volume     | 0.171 | 21:16 |                |
| Air Blank         | 0.000 | 21:17 |                |
| Control Test      | 0.080 | 21:17 |                |
| Air Blank         | 0.000 | 21:18 |                |
| Diagnostics Check | OK    | 21:18 |                |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BYRD, J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FORMAL, D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HALLISEY

DATE/TIME OF TEST: 11/12/2011 00:51:48

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:45 |
| Air Blank         | 0.000 | 00:45 |
| Control Test      | 0.080 | 00:46 |
| Air Blank         | 0.000 | 00:46 |
| Subject Sample #1 | 0.193 | 00:47 |
| Breath Volume     | 2.835 | 00:47 |
| Air Blank         | 0.000 | 00:48 |
| Air Blank         | 0.000 | 00:49 |
| Subject Sample #2 | 0.190 | 00:50 |
| Breath Volume     | 2.410 | 00:50 |
| Air Blank         | 0.000 | 00:50 |
| Control Test      | 0.079 | 00:51 |
| Air Blank         | 0.000 | 00:51 |
| Diagnostics Check | OK    | 00:51 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, RICHARD, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: JOHNSON, RICHARD  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 11/12/2011 01:01:57

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:55 |
| Air Blank         | 0.000 | 00:55 |
| Control Test      | 0.080 | 00:56 |
| Air Blank         | 0.000 | 00:56 |
| Subject Sample #1 | 0.113 | 00:57 |
| Breath Volume     | 1.207 | 00:57 |
| Air Blank         | 0.000 | 00:57 |
| Air Blank         | 0.000 | 00:59 |
| Subject Sample #2 | 0.120 | 01:00 |
| Breath Volume     | 1.980 | 01:00 |
| Air Blank         | 0.000 | 01:01 |
| Control Test      | 0.080 | 01:01 |
| Air Blank         | 0.000 | 01:01 |
| Diagnostics Check | OK    | 01:01 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, RICHARD, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: JOHNSON, RICHARD  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: KATZ

DATE/TIME OF TEST: 11/12/2011 22:21:08

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:15 |
| Air Blank         | 0.000 | 22:15 |
| Control Test      | 0.081 | 22:15 |
| Air Blank         | 0.000 | 22:16 |
| Subject Sample #1 | 0.150 | 22:16 |
| Breath Volume     | 1.992 | 22:16 |
| Air Blank         | 0.000 | 22:17 |
| Air Blank         | 0.000 | 22:19 |
| Subject Sample #2 | 0.147 | 22:19 |
| Breath Volume     | 1.398 | 22:19 |
| Air Blank         | 0.000 | 22:20 |
| Control Test      | 0.080 | 22:20 |
| Air Blank         | 0.000 | 22:21 |
| Diagnostics Check | OK    | 22:21 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, RICHARD, M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: VANIK, JON  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: EARLEY

DATE/TIME OF TEST: 11/12/2011 22:34:28

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:27 |
| Air Blank         | 0.000 | 22:28 |
| Control Test      | 0.080 | 22:28 |
| Air Blank         | 0.000 | 22:29 |
| Subject Sample #1 | 0.124 | 22:29 |
| Breath Volume     | 1.519 | 22:29 |
| Air Blank         | 0.000 | 22:30 |
| Air Blank         | 0.000 | 22:32 |
| Subject Sample #2 | 0.125 | 22:33 |
| Breath Volume     | 1.597 | 22:33 |
| Air Blank         | 0.000 | 22:33 |
| Control Test      | 0.080 | 22:33 |
| Air Blank         | 0.000 | 22:34 |
| Diagnostics Check | OK    | 22:34 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JONES, TROOPER, M

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: JONES, TROOPER

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26911080A2

DRY GAS EXPIRY DATE: 11/01/2013

VIOLATION CODE: DUI

ARREST TIME: 02:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 11/15/2011 02:59:17

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:51 |
| Air Blank         | 0.000 | 02:52 |
| Control Test      | 0.079 | 02:52 |
| Air Blank         | 0.000 | 02:53 |
| Subject Sample #1 | 0.134 | 02:54 |
| Breath Volume     | 2.175 | 02:54 |
| Air Blank         | 0.000 | 02:55 |
| Air Blank         | 0.000 | 02:57 |
| Subject Sample #2 | 0.130 | 02:57 |
| Breath Volume     | 1.332 | 02:57 |
| Air Blank         | 0.000 | 02:58 |
| Control Test      | 0.079 | 02:58 |
| Air Blank         | 0.000 | 02:59 |
| Diagnostics Check | OK    | 02:59 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES, TROOPER, M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JONES, TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 11/15/2011 23:32:05

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 23:20 |                |
| Air Blank         | 0.000 | 23:21 |                |
| Control Test      | 0.080 | 23:21 |                |
| Air Blank         | 0.000 | 23:21 |                |
| Subject Sample #1 | 0.132 | 23:24 | Volume Not Met |
| Breath Volume     | 0.617 | 23:24 |                |
| Air Blank         | 0.000 | 23:25 |                |
| Air Blank         | 0.000 | 23:27 |                |
| Subject Sample #2 | 0.116 | 23:30 | Volume Not Met |
| Breath Volume     | 0.503 | 23:30 |                |
| Air Blank         | 0.000 | 23:31 |                |
| Control Test      | 0.079 | 23:31 |                |
| Air Blank         | 0.000 | 23:31 |                |
| Diagnostics Check | OK    | 23:32 |                |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES, TROOPER, M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JONES, TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 11/15/2011 23:45:30

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 23:34 |                |
| Air Blank         | 0.000 | 23:34 |                |
| Control Test      | 0.079 | 23:34 |                |
| Air Blank         | 0.000 | 23:35 |                |
| Subject Sample #1 | 0.125 | 23:38 | Volume Not Met |
| Breath Volume     | 0.613 | 23:38 |                |
| Air Blank         | 0.000 | 23:39 |                |
| Air Blank         | 0.000 | 23:40 |                |
| Subject Sample #2 | 0.123 | 23:44 | Volume Not Met |
| Breath Volume     | 0.292 | 23:44 |                |
| Air Blank         | 0.000 | 23:44 |                |
| Control Test      | 0.078 | 23:44 |                |
| Air Blank         | 0.000 | 23:45 |                |
| Diagnostics Check | OK    | 23:45 |                |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON, DEP, W  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: JOHNSON, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 11/16/2011 02:56:15

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 02:40 |                |
| Air Blank         | 0.000 | 02:41 |                |
| Control Test      | 0.080 | 02:41 |                |
| Air Blank         | 0.000 | 02:42 |                |
| Subject Sample #1 | 0.109 | 02:45 | Volume Not Met |
| Breath Volume     | 0.250 | 02:45 |                |
| Air Blank         | 0.000 | 02:45 |                |
| Air Blank         | 0.000 | 02:47 |                |
| Subject Sample #2 | 0.135 | 02:50 |                |
| Breath Volume     | 2.351 | 02:50 |                |
| Air Blank         | 0.000 | 02:51 |                |
| Air Blank         | 0.000 | 02:53 |                |
| Subject Sample #3 | 0.128 | 02:54 |                |
| Breath Volume     | 1.437 | 02:54 |                |
| Air Blank         | 0.000 | 02:55 |                |
| Control Test      | 0.079 | 02:55 |                |
| Air Blank         | 0.000 | 02:56 |                |
| Diagnostics Check | OK    | 02:56 |                |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE, MARK,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FELTOVIC, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/16/2011 04:31:01

---

|                   |       |       |                   |
|-------------------|-------|-------|-------------------|
| Diagnostics Check | OK    | 04:17 |                   |
| Air Blank         | 0.000 | 04:17 |                   |
| Control Test      | 0.080 | 04:17 |                   |
| Air Blank         | 0.000 | 04:18 |                   |
| Subject Sample #1 | 0.169 | 04:21 | Volume Not Met    |
| Breath Volume     | 0.019 | 04:21 |                   |
| Air Blank         | 0.000 | 04:22 |                   |
| Air Blank         | 0.000 | 04:24 |                   |
| Subject Sample #2 | 0.217 | 04:24 | No .020 agreement |
| Breath Volume     | 1.937 | 04:24 |                   |
| Air Blank         | 0.000 | 04:25 |                   |
| Air Blank         | 0.000 | 04:27 |                   |
| Subject Sample #3 | 0.195 | 04:29 | No .020 agreement |
| Breath Volume     | 2.699 | 04:29 |                   |
| Air Blank         | 0.000 | 04:30 |                   |
| Control Test      | 0.079 | 04:30 |                   |
| Air Blank         | 0.000 | 04:30 |                   |
| Diagnostics Check | OK    | 04:31 |                   |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE, MARK,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FELTOVIC, DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/16/2011 04:40:53

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:33 |
| Air Blank         | 0.000 | 04:34 |
| Control Test      | 0.076 | 04:34 |
| Air Blank         | 0.000 | 04:35 |
| Subject Sample #1 | 0.202 | 04:35 |
| Breath Volume     | 2.156 | 04:35 |
| Air Blank         | 0.000 | 04:36 |
| Air Blank         | 0.000 | 04:37 |
| Subject Sample #2 | 0.205 | 04:39 |
| Breath Volume     | 2.273 | 04:39 |
| Air Blank         | 0.000 | 04:39 |
| Control Test      | 0.081 | 04:40 |
| Air Blank         | 0.000 | 04:40 |
| Diagnostics Check | OK    | 04:40 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTLER, CHRIS, T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BUTLER, CHRIS  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 12:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 11/17/2011 13:26:22

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 13:19 |
| Air Blank         | 0.000 | 13:20 |
| Control Test      | 0.081 | 13:20 |
| Air Blank         | 0.000 | 13:21 |
| Subject Sample #1 | 0.000 | 13:21 |
| Breath Volume     | 3.761 | 13:21 |
| Air Blank         | 0.000 | 13:22 |
| Air Blank         | 0.000 | 13:24 |
| Subject Sample #2 | 0.000 | 13:25 |
| Breath Volume     | 1.578 | 13:25 |
| Air Blank         | 0.000 | 13:25 |
| Control Test      | 0.080 | 13:25 |
| Air Blank         | 0.000 | 13:26 |
| Diagnostics Check | OK    | 13:26 |

---

Data Download Date/Time 11/17/2011 14:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHIPP, KRISTEN, S  
BREATH TEST OPERATOR AGENCY: FDLE  
ARRESTING OFFICER: SMITH, JOHN  
ARRESTING OFFICER AGENCY: FDLE  
LAST AGENCY INSPECTION: 11/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26911080A2  
DRY GAS EXPIRY DATE: 11/01/2013  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 10:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/28/2011 11:56:41

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 11:50 |
| Air Blank         | 0.000 | 11:50 |
| Control Test      | 0.078 | 11:50 |
| Air Blank         | 0.000 | 11:51 |
| Subject Sample #1 | 0.081 | 11:52 |
| Breath Volume     | 1.890 | 11:52 |
| Air Blank         | 0.000 | 11:52 |
| Air Blank         | 0.000 | 11:54 |
| Subject Sample #2 | 0.080 | 11:55 |
| Breath Volume     | 1.328 | 11:55 |
| Air Blank         | 0.000 | 11:55 |
| Control Test      | 0.078 | 11:56 |
| Air Blank         | 0.000 | 11:56 |
| Diagnostics Check | OK    | 11:56 |

---

Data Download Date/Time 01/10/2012 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: FRUITLAND PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRENSHAW, JARED, M  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: CRENSHAW, JARED  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 11/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 11/13/2011 02:56:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:53 |                      |
| Air Blank         | 0.000 | 02:54 |                      |
| Control Test      | 0.080 | 02:54 |                      |
| Air Blank         | 0.000 | 02:54 |                      |
| Subject Sample #1 |       | 02:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:55 |                      |
| Air Blank         | 0.000 | 02:55 |                      |
| Control Test      | 0.080 | 02:55 |                      |
| Air Blank         | 0.000 | 02:56 |                      |
| Diagnostics Check | OK    | 02:56 |                      |

---

Data Download Date/Time 12/14/2011 08:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, GARY, D  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILLIAMS, GARY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG110501  
DRY GAS EXPIRY DATE: 04/15/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: VALERIO

DATE/TIME OF TEST: 11/06/2011 03:41:39

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:35 |
| Air Blank         | 0.000 | 03:35 |
| Control Test      | 0.078 | 03:36 |
| Air Blank         | 0.000 | 03:36 |
| Subject Sample #1 | 0.237 | 03:37 |
| Breath Volume     | 2.933 | 03:37 |
| Air Blank         | 0.000 | 03:37 |
| Air Blank         | 0.000 | 03:39 |
| Subject Sample #2 | 0.235 | 03:40 |
| Breath Volume     | 3.070 | 03:40 |
| Air Blank         | 0.000 | 03:40 |
| Control Test      | 0.079 | 03:41 |
| Air Blank         | 0.000 | 03:41 |
| Diagnostics Check | OK    | 03:41 |

---

Data Download Date/Time 11/25/2011 21:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, GARY, D  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DARROW, MICHAEL  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG110501  
DRY GAS EXPIRY DATE: 04/15/2013  
VIOLATION CODE: DUI  
ARREST TIME: 05:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 11/06/2011 06:53:18

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 06:46 |
| Air Blank         | 0.000 | 06:46 |
| Control Test      | 0.079 | 06:47 |
| Air Blank         | 0.000 | 06:47 |
| Subject Sample #1 | 0.132 | 06:48 |
| Breath Volume     | 1.187 | 06:48 |
| Air Blank         | 0.000 | 06:48 |
| Air Blank         | 0.000 | 06:50 |
| Subject Sample #2 | 0.126 | 06:51 |
| Breath Volume     | 2.718 | 06:51 |
| Air Blank         | 0.000 | 06:52 |
| Control Test      | 0.079 | 06:52 |
| Air Blank         | 0.000 | 06:53 |
| Diagnostics Check | OK    | 06:53 |

---

Data Download Date/Time 11/25/2011 21:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, GARY, D  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILLIAMS, GARY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG110501  
DRY GAS EXPIRY DATE: 04/15/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MENIKHEIM

DATE/TIME OF TEST: 11/07/2011 01:52:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:49 |                      |
| Air Blank         | 0.000 | 01:49 |                      |
| Control Test      | 0.079 | 01:50 |                      |
| Air Blank         | 0.000 | 01:50 |                      |
| Subject Sample #1 |       | 01:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:51 |                      |
| Air Blank         | 0.000 | 01:51 |                      |
| Control Test      | 0.079 | 01:52 |                      |
| Air Blank         | 0.000 | 01:52 |                      |
| Diagnostics Check | OK    | 01:52 |                      |

---

Data Download Date/Time 11/25/2011 21:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROSS, M, S  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CROSS, M  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG110501  
DRY GAS EXPIRY DATE: 04/15/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ULRICH

DATE/TIME OF TEST: 11/18/2011 02:17:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:15 |                      |
| Air Blank         | 0.000 | 02:15 |                      |
| Control Test      | 0.079 | 02:15 |                      |
| Air Blank         | 0.000 | 02:16 |                      |
| Subject Sample #1 |       | 02:17 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:17 |                      |
| Air Blank         |       | 02:17 | RFI Detect           |
| Air Blank         | 0.000 | 02:17 |                      |

---

Data Download Date/Time 11/25/2011 21:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEMBOROSKI, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BOGGESS, OSCAR  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG110501  
DRY GAS EXPIRY DATE: 04/15/2013  
VIOLATION CODE: DUI  
ARREST TIME: 05:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: CORDONNIER

DATE/TIME OF TEST: 11/25/2011 06:48:40

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 06:45 |                 |
| Air Blank         | 0.000 | 06:45 |                 |
| Control Test      | 0.078 | 06:45 |                 |
| Air Blank         | 0.000 | 06:46 |                 |
| Subject Sample #1 | 0.232 | 06:47 |                 |
| Breath Volume     | 2.136 | 06:47 |                 |
| Air Blank         |       | 06:47 | Improper Sample |
| Air Blank         | 0.000 | 06:47 |                 |
| Control Test      | 0.078 | 06:48 |                 |
| Air Blank         | 0.000 | 06:48 |                 |
| Diagnostics Check | OK    | 06:48 |                 |

---

Data Download Date/Time 11/25/2011 21:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SZYMANSKI, JACEK,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SYERS JR, JOHN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG110501  
DRY GAS EXPIRY DATE: 04/15/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 11/29/2011 22:23:03

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:16 |
| Air Blank         | 0.000 | 22:17 |
| Control Test      | 0.079 | 22:17 |
| Air Blank         | 0.000 | 22:18 |
| Subject Sample #1 | 0.126 | 22:18 |
| Breath Volume     | 2.742 | 22:18 |
| Air Blank         | 0.000 | 22:19 |
| Air Blank         | 0.000 | 22:21 |
| Subject Sample #2 | 0.124 | 22:21 |
| Breath Volume     | 2.597 | 22:21 |
| Air Blank         | 0.000 | 22:22 |
| Control Test      | 0.078 | 22:22 |
| Air Blank         | 0.000 | 22:22 |
| Diagnostics Check | OK    | 22:23 |

---

Data Download Date/Time 12/29/2011 20:44

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAHIM, ARAF,  
BREATH TEST OPERATOR AGENCY: Mayport Naval Base  
ARRESTING OFFICER: K E ESANCY, K  
ARRESTING OFFICER AGENCY: Mayport Naval Base  
LAST AGENCY INSPECTION: 10/19/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PELLAND

DATE/TIME OF TEST: 11/05/2011 02:55:49

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:49 |
| Air Blank         | 0.000 | 02:49 |
| Control Test      | 0.079 | 02:50 |
| Air Blank         | 0.000 | 02:50 |
| Subject Sample #1 | 0.221 | 02:51 |
| Breath Volume     | 2.531 | 02:51 |
| Air Blank         | 0.000 | 02:51 |
| Air Blank         | 0.000 | 02:53 |
| Subject Sample #2 | 0.217 | 02:54 |
| Breath Volume     | 2.019 | 02:54 |
| Air Blank         | 0.000 | 02:54 |
| Control Test      | 0.079 | 02:55 |
| Air Blank         | 0.000 | 02:55 |
| Diagnostics Check | OK    | 02:55 |

---

Data Download Date/Time 11/08/2011 10:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAHIM, ARAF,  
BREATH TEST OPERATOR AGENCY: Mayport Naval Base  
ARRESTING OFFICER: PASTRANA, G  
ARRESTING OFFICER AGENCY: Mayport Naval Base  
LAST AGENCY INSPECTION: 10/19/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/05/2011 04:09:40

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:03 |
| Air Blank         | 0.000 | 04:03 |
| Control Test      | 0.079 | 04:04 |
| Air Blank         | 0.000 | 04:04 |
| Subject Sample #1 | 0.126 | 04:05 |
| Breath Volume     | 1.636 | 04:05 |
| Air Blank         | 0.000 | 04:05 |
| Air Blank         | 0.000 | 04:07 |
| Subject Sample #2 | 0.125 | 04:08 |
| Breath Volume     | 3.750 | 04:08 |
| Air Blank         | 0.000 | 04:08 |
| Control Test      | 0.079 | 04:09 |
| Air Blank         | 0.000 | 04:09 |
| Diagnostics Check | OK    | 04:09 |

---

Data Download Date/Time 11/08/2011 10:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOHN, J  
BREATH TEST OPERATOR AGENCY: FDLE  
ARRESTING OFFICER: SMITH, JOHN  
ARRESTING OFFICER AGENCY: FDLE  
LAST AGENCY INSPECTION: 11/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG123001  
DRY GAS EXPIRY DATE: 08/18/2013  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 08:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/15/2011 09:58:54

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 09:52 |
| Air Blank         | 0.000 | 09:53 |
| Control Test      | 0.079 | 09:53 |
| Air Blank         | 0.000 | 09:53 |
| Subject Sample #1 | 0.082 | 09:54 |
| Breath Volume     | 1.417 | 09:54 |
| Air Blank         | 0.000 | 09:54 |
| Air Blank         | 0.000 | 09:56 |
| Subject Sample #2 | 0.082 | 09:57 |
| Breath Volume     | 1.902 | 09:57 |
| Air Blank         | 0.000 | 09:57 |
| Control Test      | 0.078 | 09:58 |
| Air Blank         | 0.000 | 09:58 |
| Diagnostics Check | OK    | 09:58 |

---

Data Download Date/Time 11/28/2011 11:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, JOHN, J

BREATH TEST OPERATOR AGENCY: FDLE

ARRESTING OFFICER: SMITH, JOHN

ARRESTING OFFICER AGENCY: FDLE

LAST AGENCY INSPECTION: 11/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG123001

DRY GAS EXPIRY DATE: 08/18/2013

VIOLATION CODE: SYS CHECK

ARREST TIME: 08:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/15/2011 10:15:06

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 10:08 |
| Air Blank         | 0.000 | 10:09 |
| Control Test      | 0.078 | 10:09 |
| Air Blank         | 0.000 | 10:10 |
| Subject Sample #1 | 0.083 | 10:10 |
| Breath Volume     | 1.437 | 10:10 |
| Air Blank         | 0.000 | 10:11 |
| Air Blank         | 0.000 | 10:13 |
| Subject Sample #2 | 0.082 | 10:13 |
| Breath Volume     | 1.234 | 10:13 |
| Air Blank         | 0.000 | 10:14 |
| Control Test      | 0.079 | 10:14 |
| Air Blank         | 0.000 | 10:15 |
| Diagnostics Check | OK    | 10:15 |

---

Data Download Date/Time 11/28/2011 11:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARMAND, ELVIE, D

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: DANJOU, S

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 19:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 11/16/2011 21:01:27

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 20:55 |
| Air Blank         | 0.000 | 20:55 |
| Control Test      | 0.080 | 20:55 |
| Air Blank         | 0.000 | 20:56 |
| Subject Sample #1 | 0.174 | 20:56 |
| Breath Volume     | 2.886 | 20:56 |
| Air Blank         | 0.000 | 20:57 |
| Air Blank         | 0.000 | 20:59 |
| Subject Sample #2 | 0.174 | 20:59 |
| Breath Volume     | 2.453 | 20:59 |
| Air Blank         | 0.000 | 21:00 |
| Control Test      | 0.081 | 21:00 |
| Air Blank         | 0.000 | 21:01 |
| Diagnostics Check | OK    | 21:01 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRITY, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: IMPARATO, FRANK  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: FREPPON

DATE/TIME OF TEST: 11/16/2011 22:40:29

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:34 |
| Air Blank         | 0.000 | 22:34 |
| Control Test      | 0.082 | 22:35 |
| Air Blank         | 0.000 | 22:35 |
| Subject Sample #1 | 0.155 | 22:36 |
| Breath Volume     | 2.656 | 22:36 |
| Air Blank         | 0.000 | 22:36 |
| Air Blank         | 0.000 | 22:38 |
| Subject Sample #2 | 0.155 | 22:38 |
| Breath Volume     | 2.699 | 22:38 |
| Air Blank         | 0.000 | 22:39 |
| Control Test      | 0.081 | 22:39 |
| Air Blank         | 0.000 | 22:40 |
| Diagnostics Check | OK    | 22:40 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: GAMBRELL, CHRISTOPHER

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 01:51

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: LANDAU

DATE/TIME OF TEST: 11/17/2011 03:39:37

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 03:27 |                |
| Air Blank         | 0.000 | 03:27 |                |
| Control Test      | 0.082 | 03:27 |                |
| Air Blank         | 0.000 | 03:28 |                |
| Subject Sample #1 | 0.232 | 03:29 |                |
| Breath Volume     | 2.343 | 03:29 |                |
| Air Blank         | 0.000 | 03:29 |                |
| Air Blank         | 0.000 | 03:31 |                |
| Subject Sample #2 | 0.187 | 03:34 | Volume Not Met |
| Breath Volume     | 1.078 | 03:34 |                |
| Air Blank         | 0.000 | 03:35 |                |
| Air Blank         | 0.000 | 03:37 |                |
| Subject Sample #3 | 0.225 | 03:38 |                |
| Breath Volume     | 1.410 | 03:38 |                |
| Air Blank         | 0.000 | 03:38 |                |
| Control Test      | 0.080 | 03:39 |                |
| Air Blank         | 0.000 | 03:39 |                |
| Diagnostics Check | OK    | 03:39 |                |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: HUSIC, MARY

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 03:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: CHAPLIN

DATE/TIME OF TEST: 11/17/2011 04:45:34

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:37 |
| Air Blank         | 0.000 | 04:37 |
| Control Test      | 0.081 | 04:37 |
| Air Blank         | 0.000 | 04:38 |
| Subject Sample #1 | 0.225 | 04:40 |
| Breath Volume     | 2.226 | 04:40 |
| Air Blank         | 0.000 | 04:41 |
| Air Blank         | 0.000 | 04:43 |
| Subject Sample #2 | 0.236 | 04:43 |
| Breath Volume     | 1.253 | 04:43 |
| Air Blank         | 0.000 | 04:44 |
| Control Test      | 0.079 | 04:44 |
| Air Blank         | 0.000 | 04:45 |
| Diagnostics Check | OK    | 04:45 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARMAND, ELVIE, D  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: EVANS, J  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: SCHAEFER

DATE/TIME OF TEST: 11/17/2011 20:28:22

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 20:22 |
| Air Blank         | 0.000 | 20:22 |
| Control Test      | 0.082 | 20:23 |
| Air Blank         | 0.000 | 20:23 |
| Subject Sample #1 | 0.000 | 20:24 |
| Breath Volume     | 3.656 | 20:24 |
| Air Blank         | 0.000 | 20:24 |
| Air Blank         | 0.000 | 20:26 |
| Subject Sample #2 | 0.000 | 20:27 |
| Breath Volume     | 3.121 | 20:27 |
| Air Blank         | 0.000 | 20:27 |
| Control Test      | 0.081 | 20:27 |
| Air Blank         | 0.000 | 20:28 |
| Diagnostics Check | OK    | 20:28 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRITY, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GAMBRELL, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 11/18/2011 02:02:33

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 01:48 |                 |
| Air Blank         | 0.000 | 01:48 |                 |
| Control Test      | 0.082 | 01:48 |                 |
| Air Blank         | 0.000 | 01:49 |                 |
| Subject Sample #1 | 0.155 | 01:52 |                 |
| Breath Volume     | 1.699 | 01:52 |                 |
| Air Blank         | 0.000 | 01:53 |                 |
| Air Blank         | 0.000 | 01:54 |                 |
| Subject Sample #2 | 0.140 | 01:58 | Slope Not Level |
| Breath Volume     | 1.140 | 01:58 |                 |
| Air Blank         | 0.000 | 01:58 |                 |
| Air Blank         | 0.000 | 02:00 |                 |
| Subject Sample #3 | 0.153 | 02:00 |                 |
| Breath Volume     | 1.890 | 02:00 |                 |
| Air Blank         | 0.000 | 02:01 |                 |
| Control Test      | 0.078 | 02:01 |                 |
| Air Blank         | 0.000 | 02:02 |                 |
| Diagnostics Check | OK    | 02:02 |                 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: RIVERA, REBECCA

ARRESTING OFFICER AGENCY: Orlando PD

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: DEL RIO RIVERA

DATE/TIME OF TEST: 11/18/2011 03:40:28

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:34 |
| Air Blank         | 0.000 | 03:35 |
| Control Test      | 0.081 | 03:35 |
| Air Blank         | 0.000 | 03:35 |
| Subject Sample #1 | 0.138 | 03:36 |
| Breath Volume     | 1.851 | 03:36 |
| Air Blank         | 0.000 | 03:36 |
| Air Blank         | 0.000 | 03:38 |
| Subject Sample #2 | 0.141 | 03:38 |
| Breath Volume     | 1.921 | 03:38 |
| Air Blank         | 0.000 | 03:39 |
| Control Test      | 0.081 | 03:39 |
| Air Blank         | 0.000 | 03:40 |
| Diagnostics Check | OK    | 03:40 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRITY, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DOUGHERTY, WILLIAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: LEVEY

DATE/TIME OF TEST: 11/19/2011 01:06:12

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 00:51 |                |
| Air Blank         | 0.000 | 00:51 |                |
| Control Test      | 0.081 | 00:51 |                |
| Air Blank         | 0.000 | 00:52 |                |
| Subject Sample #1 | 0.283 | 00:55 | Volume Not Met |
| Breath Volume     | 0.304 | 00:55 |                |
| Air Blank         | 0.000 | 00:56 |                |
| Air Blank         | 0.000 | 00:58 |                |
| Subject Sample #2 | 0.276 | 00:58 |                |
| Breath Volume     | 1.371 | 00:58 |                |
| Air Blank         | 0.000 | 00:59 |                |
| Air Blank         | 0.000 | 01:01 |                |
| Subject Sample #3 | 0.262 | 01:04 | Volume Not Met |
| Breath Volume     | 0.078 | 01:04 |                |
| Air Blank         | 0.000 | 01:05 |                |
| Control Test      | 0.080 | 01:05 |                |
| Air Blank         | 0.000 | 01:06 |                |
| Diagnostics Check | OK    | 01:06 |                |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRITY, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CASTLEBERRY, MARK  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PISCIOTTI

DATE/TIME OF TEST: 11/19/2011 03:46:34

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:39 |
| Air Blank         | 0.000 | 03:40 |
| Control Test      | 0.080 | 03:40 |
| Air Blank         | 0.000 | 03:41 |
| Subject Sample #1 | 0.126 | 03:41 |
| Breath Volume     | 2.183 | 03:41 |
| Air Blank         | 0.000 | 03:42 |
| Air Blank         | 0.000 | 03:44 |
| Subject Sample #2 | 0.131 | 03:45 |
| Breath Volume     | 2.605 | 03:45 |
| Air Blank         | 0.000 | 03:45 |
| Control Test      | 0.081 | 03:45 |
| Air Blank         | 0.000 | 03:46 |
| Diagnostics Check | OK    | 03:46 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GRAHAM, JOSHUA  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: KOWLASAR

DATE/TIME OF TEST: 11/19/2011 22:15:20

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:07 |
| Air Blank         | 0.000 | 22:08 |
| Control Test      | 0.081 | 22:08 |
| Air Blank         | 0.000 | 22:09 |
| Subject Sample #1 | 0.220 | 22:09 |
| Breath Volume     | 1.339 | 22:09 |
| Air Blank         | 0.000 | 22:10 |
| Air Blank         | 0.000 | 22:12 |
| Subject Sample #2 | 0.211 | 22:13 |
| Breath Volume     | 1.121 | 22:13 |
| Air Blank         | 0.000 | 22:14 |
| Control Test      | 0.079 | 22:14 |
| Air Blank         | 0.000 | 22:15 |
| Diagnostics Check | OK    | 22:15 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: CHAPLIN, CHRISTOPHER

ARRESTING OFFICER AGENCY: Orlando PD

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 22:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: KELSTROM

DATE/TIME OF TEST: 11/19/2011 23:17:50

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:09 |
| Air Blank         | 0.000 | 23:09 |
| Control Test      | 0.081 | 23:09 |
| Air Blank         | 0.000 | 23:10 |
| Subject Sample #1 | 0.263 | 23:10 |
| Breath Volume     | 1.867 | 23:10 |
| Air Blank         | 0.000 | 23:11 |
| Air Blank         | 0.000 | 23:13 |
| Subject Sample #2 | 0.246 | 23:16 |
| Breath Volume     | 1.871 | 23:16 |
| Air Blank         | 0.000 | 23:16 |
| Control Test      | 0.081 | 23:17 |
| Air Blank         | 0.000 | 23:17 |
| Diagnostics Check | OK    | 23:17 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LAMB, JAMES  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 11/20/2011 00:38:58

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:32 |
| Air Blank         | 0.000 | 00:33 |
| Control Test      | 0.081 | 00:33 |
| Air Blank         | 0.000 | 00:34 |
| Subject Sample #1 | 0.198 | 00:34 |
| Breath Volume     | 2.515 | 00:34 |
| Air Blank         | 0.000 | 00:35 |
| Air Blank         | 0.000 | 00:37 |
| Subject Sample #2 | 0.200 | 00:37 |
| Breath Volume     | 2.601 | 00:37 |
| Air Blank         | 0.000 | 00:38 |
| Control Test      | 0.080 | 00:38 |
| Air Blank         | 0.000 | 00:38 |
| Diagnostics Check | OK    | 00:38 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WILSON, AARON  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: BISBALL

DATE/TIME OF TEST: 11/20/2011 01:56:38

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:50 |
| Air Blank         | 0.000 | 01:50 |
| Control Test      | 0.080 | 01:50 |
| Air Blank         | 0.000 | 01:51 |
| Subject Sample #1 | 0.101 | 01:51 |
| Breath Volume     | 1.117 | 01:51 |
| Air Blank         | 0.000 | 01:52 |
| Air Blank         | 0.000 | 01:54 |
| Subject Sample #2 | 0.096 | 01:55 |
| Breath Volume     | 1.367 | 01:55 |
| Air Blank         | 0.000 | 01:55 |
| Control Test      | 0.082 | 01:56 |
| Air Blank         | 0.000 | 01:56 |
| Diagnostics Check | OK    | 01:56 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WILSON, LOGAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: HUERTA

DATE/TIME OF TEST: 11/20/2011 04:40:05

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:33 |
| Air Blank         | 0.000 | 04:34 |
| Control Test      | 0.082 | 04:34 |
| Air Blank         | 0.000 | 04:34 |
| Subject Sample #1 | 0.165 | 04:35 |
| Breath Volume     | 2.933 | 04:35 |
| Air Blank         | 0.000 | 04:36 |
| Air Blank         | 0.000 | 04:38 |
| Subject Sample #2 | 0.160 | 04:38 |
| Breath Volume     | 3.289 | 04:38 |
| Air Blank         | 0.000 | 04:39 |
| Control Test      | 0.080 | 04:39 |
| Air Blank         | 0.000 | 04:39 |
| Diagnostics Check | OK    | 04:40 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PECZINKA, ANDREW  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/20/2011 05:16:35

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:10 |
| Air Blank         | 0.000 | 05:10 |
| Control Test      | 0.081 | 05:10 |
| Air Blank         | 0.000 | 05:11 |
| Subject Sample #1 | 0.161 | 05:11 |
| Breath Volume     | 1.441 | 05:11 |
| Air Blank         | 0.000 | 05:12 |
| Air Blank         | 0.000 | 05:14 |
| Subject Sample #2 | 0.161 | 05:15 |
| Breath Volume     | 1.210 | 05:15 |
| Air Blank         | 0.000 | 05:15 |
| Control Test      | 0.082 | 05:15 |
| Air Blank         | 0.000 | 05:16 |
| Diagnostics Check | OK    | 05:16 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAWKINS, DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 11/20/2011 05:38:07

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:30 |
| Air Blank         | 0.000 | 05:30 |
| Control Test      | 0.082 | 05:31 |
| Air Blank         | 0.000 | 05:31 |
| Subject Sample #1 | 0.138 | 05:33 |
| Breath Volume     | 1.269 | 05:33 |
| Air Blank         | 0.000 | 05:34 |
| Air Blank         | 0.000 | 05:36 |
| Subject Sample #2 | 0.135 | 05:36 |
| Breath Volume     | 2.460 | 05:36 |
| Air Blank         | 0.000 | 05:37 |
| Control Test      | 0.080 | 05:37 |
| Air Blank         | 0.000 | 05:38 |
| Diagnostics Check | OK    | 05:38 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: BRIERLY, JOSE

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 22:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: LEYBA

DATE/TIME OF TEST: 11/20/2011 23:31:47

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 23:14 |                |
| Air Blank         | 0.000 | 23:14 |                |
| Control Test      | 0.082 | 23:15 |                |
| Air Blank         | 0.000 | 23:15 |                |
| Subject Sample #1 | 0.150 | 23:18 | Volume Not Met |
| Breath Volume     | 0.097 | 23:18 |                |
| Air Blank         | 0.000 | 23:19 |                |
| Air Blank         | 0.000 | 23:21 |                |
| Subject Sample #2 | 0.194 | 23:24 |                |
| Breath Volume     | 1.152 | 23:24 |                |
| Air Blank         | 0.000 | 23:25 |                |
| Air Blank         | 0.000 | 23:26 |                |
| Subject Sample #3 | 0.187 | 23:30 | Volume Not Met |
| Breath Volume     | 0.410 | 23:30 |                |
| Air Blank         | 0.000 | 23:30 |                |
| Control Test      | 0.078 | 23:31 |                |
| Air Blank         | 0.000 | 23:31 |                |
| Diagnostics Check | OK    | 23:31 |                |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: BRIERLY, JOSE

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 22:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: LEYBA

DATE/TIME OF TEST: 11/20/2011 23:50:05

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 23:32 |                    |
| Air Blank         | 0.000 | 23:33 |                    |
| Control Test      | 0.079 | 23:33 |                    |
| Air Blank         | 0.000 | 23:34 |                    |
| Subject Sample #1 | 0.200 | 23:37 |                    |
| Breath Volume     | 1.554 | 23:37 |                    |
| Air Blank         | 0.000 | 23:38 |                    |
| Air Blank         | 0.000 | 23:39 |                    |
| Subject Sample #2 |       | 23:43 | No Sample Provided |
| Breath Volume     | 0.000 | 23:43 |                    |
| Air Blank         | 0.000 | 23:43 |                    |
| Air Blank         | 0.000 | 23:45 |                    |
| Subject Sample #3 |       | 23:48 | No Sample Provided |
| Breath Volume     | 0.000 | 23:48 |                    |
| Air Blank         | 0.000 | 23:49 |                    |
| Control Test      | 0.080 | 23:49 |                    |
| Air Blank         | 0.000 | 23:50 |                    |
| Diagnostics Check | OK    | 23:50 |                    |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARMAND, ELVIE, D  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VELEZ, J  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: CHRISTIANSEN

DATE/TIME OF TEST: 11/22/2011 04:06:44

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:00 |
| Air Blank         | 0.000 | 04:00 |
| Control Test      | 0.081 | 04:01 |
| Air Blank         | 0.000 | 04:01 |
| Subject Sample #1 | 0.163 | 04:02 |
| Breath Volume     | 3.460 | 04:02 |
| Air Blank         | 0.000 | 04:02 |
| Air Blank         | 0.000 | 04:04 |
| Subject Sample #2 | 0.163 | 04:05 |
| Breath Volume     | 3.121 | 04:05 |
| Air Blank         | 0.000 | 04:05 |
| Control Test      | 0.082 | 04:06 |
| Air Blank         | 0.000 | 04:06 |
| Diagnostics Check | OK    | 04:06 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MELVILLE, KELLY, M

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: WATERS, SYLVESTER

ARRESTING OFFICER AGENCY: Winter Garden PD

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 14:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: MCGLOTHLIN

DATE/TIME OF TEST: 11/22/2011 15:31:40

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 15:20 |                |
| Air Blank         | 0.000 | 15:20 |                |
| Control Test      | 0.080 | 15:20 |                |
| Air Blank         | 0.000 | 15:21 |                |
| Subject Sample #1 | 0.197 | 15:24 | Volume Not Met |
| Breath Volume     | 0.210 | 15:24 |                |
| Air Blank         | 0.000 | 15:25 |                |
| Air Blank         | 0.000 | 15:26 |                |
| Subject Sample #2 | 0.167 | 15:30 | Volume Not Met |
| Breath Volume     | 0.269 | 15:30 |                |
| Air Blank         | 0.000 | 15:30 |                |
| Control Test      | 0.079 | 15:31 |                |
| Air Blank         | 0.000 | 15:31 |                |
| Diagnostics Check | OK    | 15:31 |                |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELVILLE, KELLY, M  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WATERS, SYLVESTER  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 14:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: MCGLOTHLIN

DATE/TIME OF TEST: 11/22/2011 15:44:29

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 15:33 |
| Air Blank         | 0.000 | 15:33 |
| Control Test      | 0.080 | 15:33 |
| Air Blank         | 0.000 | 15:34 |
| Subject Sample #1 | 0.166 | 15:35 |
| Breath Volume     | 1.507 | 15:35 |
| Air Blank         | 0.000 | 15:36 |
| Air Blank         | 0.000 | 15:38 |
| Subject Sample #2 | 0.193 | 15:38 |
| Breath Volume     | 1.105 | 15:38 |
| Air Blank         | 0.000 | 15:39 |
| Air Blank         | 0.000 | 15:41 |
| Subject Sample #3 | 0.182 | 15:42 |
| Breath Volume     | 2.292 | 15:42 |
| Air Blank         | 0.000 | 15:43 |
| Control Test      | 0.078 | 15:43 |
| Air Blank         | 0.000 | 15:44 |
| Diagnostics Check | OK    | 15:44 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ARENAS, DAVID  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: CHAPA ROMERO

DATE/TIME OF TEST: 11/23/2011 03:23:40

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:17 |
| Air Blank         | 0.000 | 03:17 |
| Control Test      | 0.079 | 03:18 |
| Air Blank         | 0.000 | 03:18 |
| Subject Sample #1 | 0.052 | 03:19 |
| Breath Volume     | 2.218 | 03:19 |
| Air Blank         | 0.000 | 03:19 |
| Air Blank         | 0.000 | 03:21 |
| Subject Sample #2 | 0.051 | 03:22 |
| Breath Volume     | 2.195 | 03:22 |
| Air Blank         | 0.000 | 03:22 |
| Control Test      | 0.079 | 03:23 |
| Air Blank         | 0.000 | 03:23 |
| Diagnostics Check | OK    | 03:23 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: BRUNS, DANIEL

ARRESTING OFFICER AGENCY: Orlando PD

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11611080A2

DRY GAS EXPIRY DATE: 06/01/2013

VIOLATION CODE: DUI

ARREST TIME: 02:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BARROSO

DATE/TIME OF TEST: 11/23/2011 04:09:38

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:03 |
| Air Blank         | 0.000 | 04:03 |
| Control Test      | 0.079 | 04:04 |
| Air Blank         | 0.000 | 04:04 |
| Subject Sample #1 | 0.117 | 04:05 |
| Breath Volume     | 2.363 | 04:05 |
| Air Blank         | 0.000 | 04:05 |
| Air Blank         | 0.000 | 04:07 |
| Subject Sample #2 | 0.116 | 04:08 |
| Breath Volume     | 2.964 | 04:08 |
| Air Blank         | 0.000 | 04:08 |
| Control Test      | 0.079 | 04:09 |
| Air Blank         | 0.000 | 04:09 |
| Diagnostics Check | OK    | 04:09 |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRITY, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CASTLEBERRY, MARK  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: TAVERAS

DATE/TIME OF TEST: 11/24/2011 02:18:34

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 02:05 |                |
| Air Blank         | 0.000 | 02:05 |                |
| Control Test      | 0.080 | 02:06 |                |
| Air Blank         | 0.000 | 02:06 |                |
| Subject Sample #1 | 0.098 | 02:09 | Volume Not Met |
| Breath Volume     | 0.550 | 02:09 |                |
| Air Blank         | 0.000 | 02:10 |                |
| Air Blank         | 0.000 | 02:12 |                |
| Subject Sample #2 | 0.129 | 02:13 |                |
| Breath Volume     | 2.351 | 02:13 |                |
| Air Blank         | 0.000 | 02:14 |                |
| Air Blank         | 0.000 | 02:15 |                |
| Subject Sample #3 | 0.116 | 02:17 |                |
| Breath Volume     | 1.886 | 02:17 |                |
| Air Blank         | 0.000 | 02:17 |                |
| Control Test      | 0.078 | 02:17 |                |
| Air Blank         | 0.000 | 02:18 |                |
| Diagnostics Check | OK    | 02:18 |                |

---

Data Download Date/Time 11/24/2011 14:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARMAND, ELVIE, D  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: STRAWN, J  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: CASHMERE

DATE/TIME OF TEST: 11/26/2011 01:07:07

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:57 |
| Air Blank         | 0.000 | 00:58 |
| Control Test      | 0.080 | 00:58 |
| Air Blank         | 0.000 | 00:59 |
| Subject Sample #1 | 0.185 | 01:01 |
| Breath Volume     | 1.996 | 01:01 |
| Air Blank         | 0.000 | 01:01 |
| Air Blank         | 0.000 | 01:03 |
| Subject Sample #2 | 0.177 | 01:05 |
| Breath Volume     | 1.855 | 01:05 |
| Air Blank         | 0.000 | 01:06 |
| Control Test      | 0.077 | 01:06 |
| Air Blank         | 0.000 | 01:07 |
| Diagnostics Check | OK    | 01:07 |

---

Data Download Date/Time 12/01/2011 13:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARMAND, ELVIE, D  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ANDREWS, M  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: TUTTLE

DATE/TIME OF TEST: 11/26/2011 05:07:11

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:01 |
| Air Blank         | 0.000 | 05:01 |
| Control Test      | 0.079 | 05:01 |
| Air Blank         | 0.000 | 05:02 |
| Subject Sample #1 | 0.192 | 05:02 |
| Breath Volume     | 2.601 | 05:02 |
| Air Blank         | 0.000 | 05:03 |
| Air Blank         | 0.000 | 05:05 |
| Subject Sample #2 | 0.177 | 05:05 |
| Breath Volume     | 1.636 | 05:05 |
| Air Blank         | 0.000 | 05:06 |
| Control Test      | 0.079 | 05:06 |
| Air Blank         | 0.000 | 05:07 |
| Diagnostics Check | OK    | 05:07 |

---

Data Download Date/Time 12/01/2011 13:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ALVAREZ, HECTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 05:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: GONZALEZ BARILLAS

DATE/TIME OF TEST: 11/27/2011 06:23:54

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 06:17 |
| Air Blank         | 0.000 | 06:17 |
| Control Test      | 0.080 | 06:18 |
| Air Blank         | 0.000 | 06:18 |
| Subject Sample #1 | 0.233 | 06:19 |
| Breath Volume     | 2.562 | 06:19 |
| Air Blank         | 0.000 | 06:20 |
| Air Blank         | 0.000 | 06:22 |
| Subject Sample #2 | 0.223 | 06:22 |
| Breath Volume     | 1.101 | 06:22 |
| Air Blank         | 0.000 | 06:22 |
| Control Test      | 0.079 | 06:23 |
| Air Blank         | 0.000 | 06:23 |
| Diagnostics Check | OK    | 06:23 |

---

Data Download Date/Time 12/01/2011 13:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WILSON, LOGAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 05:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: FASNACHT

DATE/TIME OF TEST: 11/27/2011 06:44:50

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 06:38 |
| Air Blank         | 0.000 | 06:38 |
| Control Test      | 0.079 | 06:39 |
| Air Blank         | 0.000 | 06:39 |
| Subject Sample #1 | 0.204 | 06:40 |
| Breath Volume     | 2.511 | 06:40 |
| Air Blank         | 0.000 | 06:40 |
| Air Blank         | 0.000 | 06:42 |
| Subject Sample #2 | 0.212 | 06:43 |
| Breath Volume     | 2.753 | 06:43 |
| Air Blank         | 0.000 | 06:43 |
| Control Test      | 0.079 | 06:44 |
| Air Blank         | 0.000 | 06:44 |
| Diagnostics Check | OK    | 06:44 |

---

Data Download Date/Time 12/01/2011 13:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARMAND, ELVIE, D

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: EVANS, J

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11611080A2

DRY GAS EXPIRY DATE: 06/01/2013

VIOLATION CODE: DUI

ARREST TIME: 00:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 11/30/2011 01:18:20

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:12 |
| Air Blank         | 0.000 | 01:12 |
| Control Test      | 0.079 | 01:12 |
| Air Blank         | 0.000 | 01:13 |
| Subject Sample #1 | 0.186 | 01:13 |
| Breath Volume     | 3.906 | 01:13 |
| Air Blank         | 0.000 | 01:14 |
| Air Blank         | 0.000 | 01:16 |
| Subject Sample #2 | 0.176 | 01:16 |
| Breath Volume     | 2.621 | 01:16 |
| Air Blank         | 0.000 | 01:17 |
| Control Test      | 0.080 | 01:17 |
| Air Blank         | 0.000 | 01:18 |
| Diagnostics Check | OK    | 01:18 |

---

Data Download Date/Time 12/01/2011 13:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MATHEWS, MARK  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MAGRUDER

DATE/TIME OF TEST: 11/30/2011 03:23:22

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:16 |
| Air Blank         | 0.000 | 03:16 |
| Control Test      | 0.080 | 03:17 |
| Air Blank         | 0.000 | 03:17 |
| Subject Sample #1 | 0.143 | 03:18 |
| Breath Volume     | 3.316 | 03:18 |
| Air Blank         | 0.000 | 03:19 |
| Air Blank         | 0.000 | 03:21 |
| Subject Sample #2 | 0.148 | 03:21 |
| Breath Volume     | 3.535 | 03:21 |
| Air Blank         | 0.000 | 03:22 |
| Control Test      | 0.079 | 03:22 |
| Air Blank         | 0.000 | 03:23 |
| Diagnostics Check | OK    | 03:23 |

---

Data Download Date/Time 12/01/2011 13:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: YOUNG, MICHAEL

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11611080A2

DRY GAS EXPIRY DATE: 06/01/2013

VIOLATION CODE: DUI

ARREST TIME: 02:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: RAVAL

DATE/TIME OF TEST: 11/30/2011 04:01:54

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:55 |
| Air Blank         | 0.000 | 03:56 |
| Control Test      | 0.080 | 03:56 |
| Air Blank         | 0.000 | 03:56 |
| Subject Sample #1 | 0.130 | 03:57 |
| Breath Volume     | 3.488 | 03:57 |
| Air Blank         | 0.000 | 03:58 |
| Air Blank         | 0.000 | 03:59 |
| Subject Sample #2 | 0.124 | 04:00 |
| Breath Volume     | 2.902 | 04:00 |
| Air Blank         | 0.000 | 04:00 |
| Control Test      | 0.080 | 04:01 |
| Air Blank         | 0.000 | 04:01 |
| Diagnostics Check | OK    | 04:01 |

---

Data Download Date/Time 12/01/2011 13:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: REBELLO, SCOT  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 11/06/2011 04:35:32

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:23 |
| Air Blank         | 0.000 | 04:23 |
| Control Test      | 0.082 | 04:24 |
| Air Blank         | 0.000 | 04:24 |
| Subject Sample #1 | 0.189 | 04:25 |
| Breath Volume     | 1.269 | 04:25 |
| Air Blank         | 0.000 | 04:26 |
| Air Blank         | 0.000 | 04:28 |
| Subject Sample #2 | 0.216 | 04:30 |
| Breath Volume     | 2.039 | 04:30 |
| Air Blank         | 0.000 | 04:31 |
| Air Blank         | 0.000 | 04:33 |
| Subject Sample #3 | 0.214 | 04:33 |
| Breath Volume     | 3.304 | 04:33 |
| Air Blank         | 0.000 | 04:34 |
| Control Test      | 0.080 | 04:34 |
| Air Blank         | 0.000 | 04:35 |
| Diagnostics Check | OK    | 04:35 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: VIGNALI, JAMES

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 07:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:34

SUBJECT'S LAST NAME: BUCZYNSKI

DATE/TIME OF TEST: 11/06/2011 09:05:15

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 08:59 |
| Air Blank         | 0.000 | 08:59 |
| Control Test      | 0.083 | 09:00 |
| Air Blank         | 0.000 | 09:00 |
| Subject Sample #1 | 0.000 | 09:01 |
| Breath Volume     | 4.007 | 09:01 |
| Air Blank         | 0.000 | 09:01 |
| Air Blank         | 0.000 | 09:03 |
| Subject Sample #2 | 0.000 | 09:03 |
| Breath Volume     | 3.359 | 09:03 |
| Air Blank         | 0.000 | 09:04 |
| Control Test      | 0.082 | 09:04 |
| Air Blank         | 0.000 | 09:05 |
| Diagnostics Check | OK    | 09:05 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLEMAN, BRETT, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: SOTTILE, NICHOLAS

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: RIVERA RIVERA

DATE/TIME OF TEST: 11/09/2011 03:02:36

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:56 |
| Air Blank         | 0.000 | 02:57 |
| Control Test      | 0.083 | 02:57 |
| Air Blank         | 0.000 | 02:57 |
| Subject Sample #1 | 0.093 | 02:58 |
| Breath Volume     | 3.683 | 02:58 |
| Air Blank         | 0.000 | 02:58 |
| Air Blank         | 0.000 | 03:00 |
| Subject Sample #2 | 0.094 | 03:01 |
| Breath Volume     | 2.937 | 03:01 |
| Air Blank         | 0.000 | 03:01 |
| Control Test      | 0.083 | 03:02 |
| Air Blank         | 0.000 | 03:02 |
| Diagnostics Check | OK    | 03:02 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLEMAN, BRETT, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: SOTTILE, NICHOLAS

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 23:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 11/10/2011 00:29:20

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:23 |
| Air Blank         | 0.000 | 00:23 |
| Control Test      | 0.083 | 00:24 |
| Air Blank         | 0.000 | 00:24 |
| Subject Sample #1 | 0.055 | 00:25 |
| Breath Volume     | 2.859 | 00:25 |
| Air Blank         | 0.000 | 00:25 |
| Air Blank         | 0.000 | 00:27 |
| Subject Sample #2 | 0.053 | 00:27 |
| Breath Volume     | 3.160 | 00:27 |
| Air Blank         | 0.000 | 00:28 |
| Control Test      | 0.081 | 00:28 |
| Air Blank         | 0.000 | 00:29 |
| Diagnostics Check | OK    | 00:29 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: MATHEWS, MARK

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 04:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: AZUADIAZ

DATE/TIME OF TEST: 11/10/2011 05:04:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:01 |                      |
| Air Blank         | 0.000 | 05:02 |                      |
| Control Test      | 0.083 | 05:02 |                      |
| Air Blank         | 0.000 | 05:02 |                      |
| Subject Sample #1 |       | 05:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:03 |                      |
| Air Blank         | 0.000 | 05:03 |                      |
| Control Test      | 0.082 | 05:03 |                      |
| Air Blank         | 0.000 | 05:04 |                      |
| Diagnostics Check | OK    | 05:04 |                      |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRITY, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MOORE, ANDREW  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 11/11/2011 00:25:04

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 00:13 |                |
| Air Blank         | 0.000 | 00:13 |                |
| Control Test      | 0.082 | 00:14 |                |
| Air Blank         | 0.000 | 00:14 |                |
| Subject Sample #1 | 0.179 | 00:17 | Volume Not Met |
| Breath Volume     | 0.039 | 00:17 |                |
| Air Blank         | 0.000 | 00:18 |                |
| Air Blank         | 0.000 | 00:20 |                |
| Subject Sample #2 | 0.176 | 00:23 | Volume Not Met |
| Breath Volume     | 1.089 | 00:23 |                |
| Air Blank         | 0.000 | 00:24 |                |
| Control Test      | 0.082 | 00:24 |                |
| Air Blank         | 0.000 | 00:24 |                |
| Diagnostics Check | OK    | 00:25 |                |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: MOORE, ANDREW

ARRESTING OFFICER AGENCY: Maitland PD

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 22:46

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 11/11/2011 00:37:41

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 00:26 |                 |
| Air Blank         | 0.000 | 00:26 |                 |
| Control Test      | 0.081 | 00:26 |                 |
| Air Blank         | 0.000 | 00:27 |                 |
| Subject Sample #1 | 0.184 | 00:30 | Slope Not Level |
| Breath Volume     | 1.500 | 00:30 |                 |
| Air Blank         | 0.000 | 00:31 |                 |
| Air Blank         | 0.000 | 00:32 |                 |
| Subject Sample #2 | 0.180 | 00:36 | Volume Not Met  |
| Breath Volume     | 0.347 | 00:36 |                 |
| Air Blank         | 0.000 | 00:36 |                 |
| Control Test      | 0.081 | 00:37 |                 |
| Air Blank         | 0.000 | 00:37 |                 |
| Diagnostics Check | OK    | 00:37 |                 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: DANJOU, SCOTT

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DUSABLON

DATE/TIME OF TEST: 11/11/2011 04:06:55

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:00 |
| Air Blank         | 0.000 | 04:00 |
| Control Test      | 0.082 | 04:01 |
| Air Blank         | 0.000 | 04:01 |
| Subject Sample #1 | 0.224 | 04:02 |
| Breath Volume     | 2.578 | 04:02 |
| Air Blank         | 0.000 | 04:03 |
| Air Blank         | 0.000 | 04:05 |
| Subject Sample #2 | 0.204 | 04:05 |
| Breath Volume     | 1.128 | 04:05 |
| Air Blank         | 0.000 | 04:06 |
| Control Test      | 0.082 | 04:06 |
| Air Blank         | 0.000 | 04:06 |
| Diagnostics Check | OK    | 04:06 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLEMAN, BRETT, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: ROBINSON, DAVID

ARRESTING OFFICER AGENCY: Orlando PD

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 19:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: MASTERS

DATE/TIME OF TEST: 11/11/2011 20:48:22

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 20:42 |
| Air Blank         | 0.000 | 20:42 |
| Control Test      | 0.082 | 20:42 |
| Air Blank         | 0.000 | 20:43 |
| Subject Sample #1 | 0.305 | 20:43 |
| Breath Volume     | 3.675 | 20:43 |
| Air Blank         | 0.000 | 20:44 |
| Air Blank         | 0.000 | 20:46 |
| Subject Sample #2 | 0.301 | 20:46 |
| Breath Volume     | 3.785 | 20:46 |
| Air Blank         | 0.000 | 20:47 |
| Control Test      | 0.082 | 20:47 |
| Air Blank         | 0.000 | 20:48 |
| Diagnostics Check | OK    | 20:48 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: RAMIREZ, JOSE

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 01:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: GOLDENBERG

DATE/TIME OF TEST: 11/12/2011 02:51:42

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:45 |
| Air Blank         | 0.000 | 02:45 |
| Control Test      | 0.083 | 02:46 |
| Air Blank         | 0.000 | 02:46 |
| Subject Sample #1 | 0.082 | 02:47 |
| Breath Volume     | 2.308 | 02:47 |
| Air Blank         | 0.000 | 02:47 |
| Air Blank         | 0.000 | 02:49 |
| Subject Sample #2 | 0.085 | 02:50 |
| Breath Volume     | 2.550 | 02:50 |
| Air Blank         | 0.000 | 02:50 |
| Control Test      | 0.081 | 02:51 |
| Air Blank         | 0.000 | 02:51 |
| Diagnostics Check | OK    | 02:51 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLEMAN, BRETT, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: PORTER, CHARLES

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PRAINO

DATE/TIME OF TEST: 11/12/2011 03:19:21

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:13 |
| Air Blank         | 0.000 | 03:13 |
| Control Test      | 0.081 | 03:13 |
| Air Blank         | 0.000 | 03:14 |
| Subject Sample #1 | 0.187 | 03:14 |
| Breath Volume     | 2.496 | 03:14 |
| Air Blank         | 0.000 | 03:15 |
| Air Blank         | 0.000 | 03:17 |
| Subject Sample #2 | 0.198 | 03:17 |
| Breath Volume     | 2.718 | 03:17 |
| Air Blank         | 0.000 | 03:18 |
| Control Test      | 0.083 | 03:18 |
| Air Blank         | 0.000 | 03:19 |
| Diagnostics Check | OK    | 03:19 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ALVAREZ, HECTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RUIZ RODRIGUEZ

DATE/TIME OF TEST: 11/12/2011 04:21:14

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 04:06 |                |
| Air Blank         | 0.000 | 04:07 |                |
| Control Test      | 0.082 | 04:07 |                |
| Air Blank         | 0.000 | 04:07 |                |
| Subject Sample #1 | 0.173 | 04:11 |                |
| Breath Volume     | 2.167 | 04:11 |                |
| Air Blank         | 0.000 | 04:11 |                |
| Air Blank         | 0.000 | 04:13 |                |
| Subject Sample #2 | 0.137 | 04:16 | Volume Not Met |
| Breath Volume     | 0.589 | 04:16 |                |
| Air Blank         | 0.000 | 04:17 |                |
| Air Blank         | 0.000 | 04:19 |                |
| Subject Sample #3 | 0.163 | 04:19 |                |
| Breath Volume     | 1.687 | 04:19 |                |
| Air Blank         | 0.000 | 04:20 |                |
| Control Test      | 0.080 | 04:20 |                |
| Air Blank         | 0.000 | 04:21 |                |
| Diagnostics Check | OK    | 04:21 |                |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLEMAN, BRETT, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: MOORE, ANDREW

ARRESTING OFFICER AGENCY: Maitland PD

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 03:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: BROOKINGS

DATE/TIME OF TEST: 11/12/2011 05:25:33

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:19 |
| Air Blank         | 0.000 | 05:20 |
| Control Test      | 0.082 | 05:20 |
| Air Blank         | 0.000 | 05:20 |
| Subject Sample #1 | 0.104 | 05:21 |
| Breath Volume     | 1.699 | 05:21 |
| Air Blank         | 0.000 | 05:21 |
| Air Blank         | 0.000 | 05:23 |
| Subject Sample #2 | 0.105 | 05:24 |
| Breath Volume     | 1.867 | 05:24 |
| Air Blank         | 0.000 | 05:24 |
| Control Test      | 0.082 | 05:24 |
| Air Blank         | 0.000 | 05:25 |
| Diagnostics Check | OK    | 05:25 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, BRETT, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ELLIS, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SANBORN

DATE/TIME OF TEST: 11/13/2011 23:16:03

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:08 |
| Air Blank         | 0.000 | 23:09 |
| Control Test      | 0.082 | 23:09 |
| Air Blank         | 0.000 | 23:10 |
| Subject Sample #1 | 0.155 | 23:11 |
| Breath Volume     | 1.750 | 23:11 |
| Air Blank         | 0.000 | 23:12 |
| Air Blank         | 0.000 | 23:14 |
| Subject Sample #2 | 0.161 | 23:14 |
| Breath Volume     | 2.472 | 23:14 |
| Air Blank         | 0.000 | 23:15 |
| Control Test      | 0.082 | 23:15 |
| Air Blank         | 0.000 | 23:15 |
| Diagnostics Check | OK    | 23:16 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLEMAN, BRETT, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: KING, GORDON

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HERRON

DATE/TIME OF TEST: 11/14/2011 03:38:18

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:29 |
| Air Blank         | 0.000 | 03:29 |
| Control Test      | 0.083 | 03:30 |
| Air Blank         | 0.000 | 03:30 |
| Subject Sample #1 | 0.184 | 03:31 |
| Breath Volume     | 1.109 | 03:31 |
| Air Blank         | 0.000 | 03:32 |
| Air Blank         | 0.000 | 03:34 |
| Subject Sample #2 | 0.181 | 03:36 |
| Breath Volume     | 1.804 | 03:36 |
| Air Blank         | 0.000 | 03:37 |
| Control Test      | 0.082 | 03:37 |
| Air Blank         | 0.000 | 03:38 |
| Diagnostics Check | OK    | 03:38 |

---

Data Download Date/Time 11/14/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLEMAN, BRETT, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: BOURQUE, JESSIE

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 10/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 11/15/2011 02:57:30

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:48 |
| Air Blank         | 0.000 | 02:49 |
| Control Test      | 0.083 | 02:49 |
| Air Blank         | 0.000 | 02:50 |
| Subject Sample #1 | 0.211 | 02:51 |
| Breath Volume     | 2.269 | 02:51 |
| Air Blank         | 0.000 | 02:52 |
| Air Blank         | 0.000 | 02:54 |
| Subject Sample #2 | 0.198 | 02:55 |
| Breath Volume     | 1.648 | 02:55 |
| Air Blank         | 0.000 | 02:56 |
| Control Test      | 0.082 | 02:56 |
| Air Blank         | 0.000 | 02:57 |
| Diagnostics Check | OK    | 02:57 |

---

Data Download Date/Time 11/16/2011 11:26

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMKISHUN, PAMELA, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CASTLEBERRY, MARK  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: CAUDLE

DATE/TIME OF TEST: 11/18/2011 02:20:23

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:14 |
| Air Blank         | 0.000 | 02:14 |
| Control Test      | 0.083 | 02:15 |
| Air Blank         | 0.000 | 02:15 |
| Subject Sample #1 | 0.196 | 02:16 |
| Breath Volume     | 2.835 | 02:16 |
| Air Blank         | 0.000 | 02:16 |
| Air Blank         | 0.000 | 02:18 |
| Subject Sample #2 | 0.197 | 02:18 |
| Breath Volume     | 2.640 | 02:18 |
| Air Blank         | 0.000 | 02:19 |
| Control Test      | 0.082 | 02:19 |
| Air Blank         | 0.000 | 02:20 |
| Diagnostics Check | OK    | 02:20 |

---

Data Download Date/Time 11/24/2011 14:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DENSON, TRAVIS, W

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: KING, GORDON

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 11/18/2011 03:30:55

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:24 |
| Air Blank         | 0.000 | 03:24 |
| Control Test      | 0.082 | 03:25 |
| Air Blank         | 0.000 | 03:25 |
| Subject Sample #1 | 0.144 | 03:26 |
| Breath Volume     | 2.445 | 03:26 |
| Air Blank         | 0.000 | 03:26 |
| Air Blank         | 0.000 | 03:28 |
| Subject Sample #2 | 0.152 | 03:29 |
| Breath Volume     | 1.910 | 03:29 |
| Air Blank         | 0.000 | 03:30 |
| Control Test      | 0.082 | 03:30 |
| Air Blank         | 0.000 | 03:30 |
| Diagnostics Check | OK    | 03:30 |

---

Data Download Date/Time 11/24/2011 14:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMKISHUN, PAMELA, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HUSIC, MARY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: SLAYTON

DATE/TIME OF TEST: 11/18/2011 03:47:26

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:40 |
| Air Blank         | 0.000 | 03:41 |
| Control Test      | 0.082 | 03:41 |
| Air Blank         | 0.000 | 03:42 |
| Subject Sample #1 | 0.192 | 03:42 |
| Breath Volume     | 2.550 | 03:42 |
| Air Blank         | 0.000 | 03:43 |
| Air Blank         | 0.000 | 03:45 |
| Subject Sample #2 | 0.184 | 03:45 |
| Breath Volume     | 2.410 | 03:45 |
| Air Blank         | 0.000 | 03:46 |
| Control Test      | 0.081 | 03:46 |
| Air Blank         | 0.000 | 03:47 |
| Diagnostics Check | OK    | 03:47 |

---

Data Download Date/Time 11/24/2011 14:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: FELLOWS, BRENT  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 11/18/2011 05:22:22

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:16 |
| Air Blank         | 0.000 | 05:16 |
| Control Test      | 0.081 | 05:17 |
| Air Blank         | 0.000 | 05:17 |
| Subject Sample #1 | 0.204 | 05:18 |
| Breath Volume     | 1.953 | 05:18 |
| Air Blank         | 0.000 | 05:18 |
| Air Blank         | 0.000 | 05:20 |
| Subject Sample #2 | 0.201 | 05:20 |
| Breath Volume     | 1.457 | 05:20 |
| Air Blank         | 0.000 | 05:21 |
| Control Test      | 0.082 | 05:21 |
| Air Blank         | 0.000 | 05:22 |
| Diagnostics Check | OK    | 05:22 |

---

Data Download Date/Time 11/24/2011 14:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAMKISHUN, PAMELA, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: WAESCO, SEAN

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 00:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: PILLMEAR

DATE/TIME OF TEST: 11/19/2011 01:43:56

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:37 |
| Air Blank         | 0.000 | 01:37 |
| Control Test      | 0.083 | 01:37 |
| Air Blank         | 0.000 | 01:38 |
| Subject Sample #1 | 0.190 | 01:38 |
| Breath Volume     | 1.914 | 01:38 |
| Air Blank         | 0.000 | 01:39 |
| Air Blank         | 0.000 | 01:41 |
| Subject Sample #2 | 0.201 | 01:42 |
| Breath Volume     | 2.832 | 01:42 |
| Air Blank         | 0.000 | 01:43 |
| Control Test      | 0.082 | 01:43 |
| Air Blank         | 0.000 | 01:43 |
| Diagnostics Check | OK    | 01:43 |

---

Data Download Date/Time 11/24/2011 14:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMKISHUN, PAMELA, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: TEJADA, INGRID  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 11/19/2011 04:10:36

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:03 |
| Air Blank         | 0.000 | 04:04 |
| Control Test      | 0.083 | 04:04 |
| Air Blank         | 0.000 | 04:04 |
| Subject Sample #1 | 0.176 | 04:06 |
| Breath Volume     | 1.230 | 04:06 |
| Air Blank         | 0.000 | 04:06 |
| Air Blank         | 0.000 | 04:08 |
| Subject Sample #2 | 0.181 | 04:09 |
| Breath Volume     | 1.339 | 04:09 |
| Air Blank         | 0.000 | 04:09 |
| Control Test      | 0.082 | 04:10 |
| Air Blank         | 0.000 | 04:10 |
| Diagnostics Check | OK    | 04:10 |

---

Data Download Date/Time 11/24/2011 14:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RAMIREZ, JOSE  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WEISFUSS-JOHN S

DATE/TIME OF TEST: 11/24/2011 02:05:19

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:58 |
| Air Blank         | 0.000 | 01:59 |
| Control Test      | 0.083 | 01:59 |
| Air Blank         | 0.000 | 02:00 |
| Subject Sample #1 | 0.104 | 02:00 |
| Breath Volume     | 2.960 | 02:00 |
| Air Blank         | 0.000 | 02:01 |
| Air Blank         | 0.000 | 02:02 |
| Subject Sample #2 | 0.107 | 02:03 |
| Breath Volume     | 2.632 | 02:03 |
| Air Blank         | 0.000 | 02:04 |
| Control Test      | 0.082 | 02:04 |
| Air Blank         | 0.000 | 02:05 |
| Diagnostics Check | OK    | 02:05 |

---

Data Download Date/Time 11/24/2011 14:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAWKINS, DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/24/2011 04:26:14

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:20 |
| Air Blank         | 0.000 | 04:20 |
| Control Test      | 0.082 | 04:21 |
| Air Blank         | 0.000 | 04:21 |
| Subject Sample #1 | 0.142 | 04:21 |
| Breath Volume     | 1.695 | 04:21 |
| Air Blank         | 0.000 | 04:22 |
| Air Blank         | 0.000 | 04:24 |
| Subject Sample #2 | 0.144 | 04:24 |
| Breath Volume     | 1.425 | 04:24 |
| Air Blank         | 0.000 | 04:25 |
| Control Test      | 0.082 | 04:25 |
| Air Blank         | 0.000 | 04:26 |
| Diagnostics Check | OK    | 04:26 |

---

Data Download Date/Time 11/24/2011 14:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BURKE, JIMMY, A

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: HARWOOD, ARMANDO

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 05:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 11/24/2011 06:21:30

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 06:14 |
| Air Blank         | 0.000 | 06:14 |
| Control Test      | 0.083 | 06:15 |
| Air Blank         | 0.000 | 06:15 |
| Subject Sample #1 | 0.134 | 06:17 |
| Breath Volume     | 1.878 | 06:17 |
| Air Blank         | 0.000 | 06:17 |
| Air Blank         | 0.000 | 06:19 |
| Subject Sample #2 | 0.131 | 06:19 |
| Breath Volume     | 1.347 | 06:19 |
| Air Blank         | 0.000 | 06:20 |
| Control Test      | 0.082 | 06:20 |
| Air Blank         | 0.000 | 06:21 |
| Diagnostics Check | OK    | 06:21 |

---

Data Download Date/Time 11/24/2011 14:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BURKE, JIMMY, A

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: FORT, ROBERT

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 06:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: CHMILNITZKY

DATE/TIME OF TEST: 11/24/2011 07:44:01

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 07:35 |
| Air Blank         | 0.000 | 07:35 |
| Control Test      | 0.082 | 07:36 |
| Air Blank         | 0.000 | 07:36 |
| Subject Sample #1 | 0.170 | 07:38 |
| Breath Volume     | 1.253 | 07:38 |
| Air Blank         | 0.000 | 07:39 |
| Air Blank         | 0.000 | 07:41 |
| Subject Sample #2 | 0.173 | 07:42 |
| Breath Volume     | 1.238 | 07:42 |
| Air Blank         | 0.000 | 07:43 |
| Control Test      | 0.081 | 07:43 |
| Air Blank         | 0.000 | 07:43 |
| Diagnostics Check | OK    | 07:44 |

---

Data Download Date/Time 11/24/2011 14:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DENSON, TRAVIS, W

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: RAMIREZ, JOSE

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 03:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: VICENTY

DATE/TIME OF TEST: 11/25/2011 04:05:23

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:59 |
| Air Blank         | 0.000 | 03:59 |
| Control Test      | 0.084 | 04:00 |
| Air Blank         | 0.000 | 04:00 |
| Subject Sample #1 | 0.102 | 04:00 |
| Breath Volume     | 4.550 | 04:00 |
| Air Blank         | 0.000 | 04:01 |
| Air Blank         | 0.000 | 04:03 |
| Subject Sample #2 | 0.100 | 04:03 |
| Breath Volume     | 4.359 | 04:03 |
| Air Blank         | 0.000 | 04:04 |
| Control Test      | 0.083 | 04:04 |
| Air Blank         | 0.000 | 04:05 |
| Diagnostics Check | OK    | 04:05 |

---

Data Download Date/Time 12/01/2011 13:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DENSON, TRAVIS, W

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: CAMPBELL, ROBERT

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:24

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 11/27/2011 04:00:14

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:54 |
| Air Blank         | 0.000 | 03:54 |
| Control Test      | 0.083 | 03:55 |
| Air Blank         | 0.000 | 03:55 |
| Subject Sample #1 | 0.000 | 03:56 |
| Breath Volume     | 2.398 | 03:56 |
| Air Blank         | 0.000 | 03:56 |
| Air Blank         | 0.000 | 03:58 |
| Subject Sample #2 | 0.000 | 03:58 |
| Breath Volume     | 2.421 | 03:58 |
| Air Blank         | 0.000 | 03:59 |
| Control Test      | 0.082 | 03:59 |
| Air Blank         | 0.000 | 04:00 |
| Diagnostics Check | OK    | 04:00 |

---

Data Download Date/Time 12/01/2011 13:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: WOODYARD, ROBERT

ARRESTING OFFICER AGENCY: Orlando PD

LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 03:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BARNO

DATE/TIME OF TEST: 11/30/2011 04:01:49

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:55 |
| Air Blank         | 0.000 | 03:56 |
| Control Test      | 0.083 | 03:56 |
| Air Blank         | 0.000 | 03:56 |
| Subject Sample #1 | 0.183 | 03:57 |
| Breath Volume     | 2.996 | 03:57 |
| Air Blank         | 0.000 | 03:58 |
| Air Blank         | 0.000 | 03:59 |
| Subject Sample #2 | 0.181 | 04:00 |
| Breath Volume     | 2.578 | 04:00 |
| Air Blank         | 0.000 | 04:00 |
| Control Test      | 0.083 | 04:01 |
| Air Blank         | 0.000 | 04:01 |
| Diagnostics Check | OK    | 04:01 |

---

Data Download Date/Time 12/01/2011 13:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: YOUNG, MICHAEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 11/30/2011 21:11:36

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 21:05 |
| Air Blank         | 0.000 | 21:05 |
| Control Test      | 0.082 | 21:06 |
| Air Blank         | 0.000 | 21:06 |
| Subject Sample #1 | 0.171 | 21:07 |
| Breath Volume     | 1.382 | 21:07 |
| Air Blank         | 0.000 | 21:07 |
| Air Blank         | 0.000 | 21:09 |
| Subject Sample #2 | 0.176 | 21:10 |
| Breath Volume     | 2.128 | 21:10 |
| Air Blank         | 0.000 | 21:10 |
| Control Test      | 0.083 | 21:11 |
| Air Blank         | 0.000 | 21:11 |
| Diagnostics Check | OK    | 21:11 |

---

Data Download Date/Time 12/01/2011 13:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES, LAUREN, K  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32210080A1  
DRY GAS EXPIRY DATE: 12/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PEDRO

DATE/TIME OF TEST: 11/24/2011 05:55:04

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 05:48 |
| Air Blank         | 0.000 | 05:48 |
| Control Test      | 0.079 | 05:49 |
| Air Blank         | 0.000 | 05:49 |
| Subject Sample #1 | 0.163 | 05:50 |
| Breath Volume     | 1.562 | 05:50 |
| Air Blank         | 0.000 | 05:51 |
| Air Blank         | 0.000 | 05:53 |
| Subject Sample #2 | 0.154 | 05:53 |
| Breath Volume     | 1.734 | 05:53 |
| Air Blank         | 0.000 | 05:54 |
| Control Test      | 0.077 | 05:54 |
| Air Blank         | 0.000 | 05:54 |
| Diagnostics Check | OK    | 05:55 |

---

Data Download Date/Time 11/29/2011 09:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANTON, JON, E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HANTON, JON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32210080A1  
DRY GAS EXPIRY DATE: 12/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 07:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: PANICHELLA

DATE/TIME OF TEST: 11/26/2011 08:53:54

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 08:44 |
| Air Blank         | 0.000 | 08:45 |
| Control Test      | 0.078 | 08:45 |
| Air Blank         | 0.000 | 08:46 |
| Subject Sample #1 | 0.289 | 08:48 |
| Breath Volume     | 1.218 | 08:48 |
| Air Blank         | 0.000 | 08:49 |
| Air Blank         | 0.000 | 08:51 |
| Subject Sample #2 | 0.296 | 08:52 |
| Breath Volume     | 1.542 | 08:52 |
| Air Blank         | 0.000 | 08:52 |
| Control Test      | 0.077 | 08:53 |
| Air Blank         | 0.000 | 08:53 |
| Diagnostics Check | OK    | 08:53 |

---

Data Download Date/Time 11/29/2011 09:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS, CHRISTOPHER, L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32210080A1  
DRY GAS EXPIRY DATE: 12/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: REECE

DATE/TIME OF TEST: 11/27/2011 22:26:58

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:20 |
| Air Blank         | 0.000 | 22:21 |
| Control Test      | 0.078 | 22:21 |
| Air Blank         | 0.000 | 22:22 |
| Subject Sample #1 | 0.152 | 22:22 |
| Breath Volume     | 2.363 | 22:22 |
| Air Blank         | 0.000 | 22:23 |
| Air Blank         | 0.000 | 22:25 |
| Subject Sample #2 | 0.155 | 22:25 |
| Breath Volume     | 2.042 | 22:25 |
| Air Blank         | 0.000 | 22:26 |
| Control Test      | 0.079 | 22:26 |
| Air Blank         | 0.000 | 22:26 |
| Diagnostics Check | OK    | 22:26 |

---

Data Download Date/Time 11/29/2011 09:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES, LAUREN, K  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HINES, LAUREN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 11/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32210080A1  
DRY GAS EXPIRY DATE: 12/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: POLZIN

DATE/TIME OF TEST: 11/29/2011 20:43:02

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 20:36 |
| Air Blank         | 0.000 | 20:36 |
| Control Test      | 0.079 | 20:37 |
| Air Blank         | 0.000 | 20:37 |
| Subject Sample #1 | 0.000 | 20:38 |
| Breath Volume     | 1.730 | 20:38 |
| Air Blank         | 0.000 | 20:39 |
| Air Blank         | 0.000 | 20:41 |
| Subject Sample #2 | 0.000 | 20:41 |
| Breath Volume     | 2.605 | 20:41 |
| Air Blank         | 0.000 | 20:42 |
| Control Test      | 0.079 | 20:42 |
| Air Blank         | 0.000 | 20:42 |
| Diagnostics Check | OK    | 20:43 |

---

Data Download Date/Time 12/14/2011 11:11

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES, LAUREN, K  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HINES, LAUREN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 11/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32210080A1  
DRY GAS EXPIRY DATE: 12/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DONNICK III

DATE/TIME OF TEST: 11/30/2011 02:07:22

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:01 |
| Air Blank         | 0.000 | 02:01 |
| Control Test      | 0.078 | 02:01 |
| Air Blank         | 0.000 | 02:02 |
| Subject Sample #1 | 0.127 | 02:02 |
| Breath Volume     | 1.339 | 02:02 |
| Air Blank         | 0.000 | 02:03 |
| Air Blank         | 0.000 | 02:05 |
| Subject Sample #2 | 0.131 | 02:05 |
| Breath Volume     | 1.332 | 02:05 |
| Air Blank         | 0.000 | 02:06 |
| Control Test      | 0.078 | 02:06 |
| Air Blank         | 0.000 | 02:07 |
| Diagnostics Check | OK    | 02:07 |

---

Data Download Date/Time 12/14/2011 11:11

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILCKEN, JOHN, J  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: HILCKEN, JOHN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 11/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32210080A1  
DRY GAS EXPIRY DATE: 12/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 18:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: ROBISON

DATE/TIME OF TEST: 11/30/2011 19:11:28

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 19:02 |
| Air Blank         | 0.000 | 19:02 |
| Control Test      | 0.079 | 19:03 |
| Air Blank         | 0.000 | 19:03 |
| Subject Sample #1 | 0.074 | 19:06 |
| Breath Volume     | 1.296 | 19:06 |
| Air Blank         | 0.000 | 19:07 |
| Air Blank         | 0.000 | 19:08 |
| Subject Sample #2 | 0.070 | 19:09 |
| Breath Volume     | 1.746 | 19:09 |
| Air Blank         | 0.000 | 19:10 |
| Control Test      | 0.077 | 19:10 |
| Air Blank         | 0.000 | 19:11 |
| Diagnostics Check | OK    | 19:11 |

---

Data Download Date/Time 12/14/2011 11:11

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001440

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROONQUIST, DON, R  
BREATH TEST OPERATOR AGENCY: Kennedy Space Center  
ARRESTING OFFICER: LEWIS, HAYDEN  
ARRESTING OFFICER AGENCY: Kennedy Space Center  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11111080A1  
DRY GAS EXPIRY DATE: 05/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: MANCINI

DATE/TIME OF TEST: 11/12/2011 16:12:30

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 16:05 |                           |
| Air Blank         | 0.000 | 16:06 |                           |
| Control Test      | 0.077 | 16:06 |                           |
| Air Blank         | 0.000 | 16:07 |                           |
| Subject Sample #1 | 0.164 | 16:07 |                           |
| Breath Volume     | 3.316 | 16:07 |                           |
| Air Blank         | 0.000 | 16:08 |                           |
| Air Blank         | 0.000 | 16:10 |                           |
| Subject Sample #2 | 0.166 | 16:10 |                           |
| Breath Volume     | 1.718 | 16:10 |                           |
| Air Blank         | 0.000 | 16:11 |                           |
| Control Test      | 0.074 | 16:11 | Control Outside Tolerance |
| Air Blank         | 0.000 | 16:12 |                           |

---

Data Download Date/Time 11/23/2011 00:08

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

INSTRUMENT REGISTERED TO: BISCAYNE NAT PARK  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIERLEY, JARED, P  
BREATH TEST OPERATOR AGENCY: Biscayne Nat. Park  
ARRESTING OFFICER: KINGSTON, JUSTIN  
ARRESTING OFFICER AGENCY: Everglades Nat Park  
LAST AGENCY INSPECTION: 11/02/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 16:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 11/13/2011 17:49:32

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 17:43 |
| Air Blank         | 0.000 | 17:43 |
| Control Test      | 0.079 | 17:43 |
| Air Blank         | 0.000 | 17:44 |
| Subject Sample #1 | 0.089 | 17:45 |
| Breath Volume     | 3.738 | 17:45 |
| Air Blank         | 0.000 | 17:45 |
| Air Blank         | 0.000 | 17:47 |
| Subject Sample #2 | 0.091 | 17:48 |
| Breath Volume     | 3.695 | 17:48 |
| Air Blank         | 0.000 | 17:48 |
| Control Test      | 0.079 | 17:48 |
| Air Blank         | 0.000 | 17:49 |
| Diagnostics Check | OK    | 17:49 |

---

Data Download Date/Time 12/01/2011 17:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

INSTRUMENT REGISTERED TO: BISCAYNE NAT PARK  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKFORD, EVAN, L  
BREATH TEST OPERATOR AGENCY: Biscayne Nat. Park  
ARRESTING OFFICER: PICKFORD, EVAN  
ARRESTING OFFICER AGENCY: Biscayne Nat. Park  
LAST AGENCY INSPECTION: 11/02/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 10:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: FERRARO

DATE/TIME OF TEST: 11/26/2011 14:21:05

---

|                   |       |       |                   |
|-------------------|-------|-------|-------------------|
| Diagnostics Check | OK    | 14:09 |                   |
| Air Blank         | 0.000 | 14:09 |                   |
| Control Test      | 0.080 | 14:10 |                   |
| Air Blank         | 0.000 | 14:10 |                   |
| Subject Sample #1 | 0.286 | 14:13 | No .020 agreement |
| Breath Volume     | 1.429 | 14:13 |                   |
| Air Blank         | 0.000 | 14:13 |                   |
| Air Blank         | 0.000 | 14:15 |                   |
| Subject Sample #2 | 0.260 | 14:16 | No .020 agreement |
| Breath Volume     | 1.730 | 14:16 |                   |
| Air Blank         | 0.000 | 14:16 |                   |
| Air Blank         | 0.000 | 14:18 |                   |
| Subject Sample #3 | 0.229 | 14:19 | No .020 agreement |
| Breath Volume     | 2.800 | 14:19 |                   |
| Air Blank         | 0.000 | 14:20 |                   |
| Control Test      | 0.080 | 14:20 |                   |
| Air Blank         | 0.000 | 14:20 |                   |
| Diagnostics Check | OK    | 14:21 |                   |

---

Data Download Date/Time 12/01/2011 17:35

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

INSTRUMENT REGISTERED TO: BISCAYNE NAT PARK  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKFORD, EVAN, L  
BREATH TEST OPERATOR AGENCY: Biscayne Nat. Park  
ARRESTING OFFICER: PICKFORD, EVAN  
ARRESTING OFFICER AGENCY: Biscayne Nat. Park  
LAST AGENCY INSPECTION: 11/02/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 13:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: FERRARO

DATE/TIME OF TEST: 11/26/2011 14:32:24

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 14:25 |
| Air Blank         | 0.000 | 14:26 |
| Control Test      | 0.079 | 14:26 |
| Air Blank         | 0.000 | 14:27 |
| Subject Sample #1 | 0.284 | 14:27 |
| Breath Volume     | 3.191 | 14:27 |
| Air Blank         | 0.000 | 14:28 |
| Air Blank         | 0.000 | 14:30 |
| Subject Sample #2 | 0.271 | 14:30 |
| Breath Volume     | 1.593 | 14:30 |
| Air Blank         | 0.000 | 14:31 |
| Control Test      | 0.079 | 14:31 |
| Air Blank         | 0.000 | 14:32 |
| Diagnostics Check | OK    | 14:32 |

---

Data Download Date/Time 12/01/2011 17:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURCH, ROBRET, M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: SAUNDERS, J  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LANFORD

DATE/TIME OF TEST: 11/05/2011 04:35:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:27 |                      |
| Air Blank         | 0.000 | 04:27 |                      |
| Control Test      | 0.079 | 04:27 |                      |
| Air Blank         | 0.000 | 04:28 |                      |
| Subject Sample #1 |       | 04:31 | No Sample Provided   |
| Breath Volume     | 0.000 | 04:31 |                      |
| Air Blank         | 0.000 | 04:32 |                      |
| Air Blank         | 0.000 | 04:34 |                      |
| Subject Sample #2 |       | 04:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:34 |                      |
| Air Blank         | 0.000 | 04:34 |                      |
| Control Test      | 0.079 | 04:35 |                      |
| Air Blank         | 0.000 | 04:35 |                      |
| Diagnostics Check | OK    | 04:35 |                      |

---

Data Download Date/Time 11/12/2011 11:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURCH, ROBERT, M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GLENN, C  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SOUPENE

DATE/TIME OF TEST: 11/10/2011 04:33:31

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:27 |
| Air Blank         | 0.000 | 04:27 |
| Control Test      | 0.079 | 04:27 |
| Air Blank         | 0.000 | 04:28 |
| Subject Sample #1 | 0.197 | 04:29 |
| Breath Volume     | 3.277 | 04:29 |
| Air Blank         | 0.000 | 04:29 |
| Air Blank         | 0.000 | 04:31 |
| Subject Sample #2 | 0.201 | 04:32 |
| Breath Volume     | 3.312 | 04:32 |
| Air Blank         | 0.000 | 04:32 |
| Control Test      | 0.078 | 04:32 |
| Air Blank         | 0.000 | 04:33 |
| Diagnostics Check | OK    | 04:33 |

---

Data Download Date/Time 11/12/2011 11:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURCH, ROBERT, M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: LINEBERRY, S.  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 10/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DWORSCHAK

DATE/TIME OF TEST: 11/11/2011 03:37:35

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:28 |
| Air Blank         | 0.000 | 03:29 |
| Control Test      | 0.079 | 03:29 |
| Air Blank         | 0.000 | 03:30 |
| Subject Sample #1 | 0.127 | 03:31 |
| Breath Volume     | 1.554 | 03:31 |
| Air Blank         | 0.000 | 03:32 |
| Air Blank         | 0.000 | 03:34 |
| Subject Sample #2 | 0.125 | 03:36 |
| Breath Volume     | 1.460 | 03:36 |
| Air Blank         | 0.000 | 03:36 |
| Control Test      | 0.079 | 03:37 |
| Air Blank         | 0.000 | 03:37 |
| Diagnostics Check | OK    | 03:37 |

---

Data Download Date/Time 11/12/2011 11:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPAULDING, JORDON, S  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: TENNANT, R  
ARRESTING OFFICER AGENCY: Orange Park PD  
LAST AGENCY INSPECTION: 10/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SALVIN

DATE/TIME OF TEST: 11/12/2011 02:16:26

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:10 |
| Air Blank         | 0.000 | 02:10 |
| Control Test      | 0.079 | 02:10 |
| Air Blank         | 0.000 | 02:11 |
| Subject Sample #1 | 0.172 | 02:11 |
| Breath Volume     | 1.628 | 02:11 |
| Air Blank         | 0.000 | 02:12 |
| Air Blank         | 0.000 | 02:14 |
| Subject Sample #2 | 0.159 | 02:14 |
| Breath Volume     | 3.167 | 02:14 |
| Air Blank         | 0.000 | 02:15 |
| Control Test      | 0.079 | 02:15 |
| Air Blank         | 0.000 | 02:16 |
| Diagnostics Check | OK    | 02:16 |

---

Data Download Date/Time 11/12/2011 11:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WENDORFF, DANIEL, E  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: SIMANDL, WAYNE  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/10/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: WILDS

DATE/TIME OF TEST: 11/12/2011 03:52:14

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:45 |
| Air Blank         | 0.000 | 03:46 |
| Control Test      | 0.080 | 03:46 |
| Air Blank         | 0.000 | 03:47 |
| Subject Sample #1 | 0.066 | 03:47 |
| Breath Volume     | 2.757 | 03:47 |
| Air Blank         | 0.000 | 03:48 |
| Air Blank         | 0.000 | 03:50 |
| Subject Sample #2 | 0.064 | 03:50 |
| Breath Volume     | 2.484 | 03:50 |
| Air Blank         | 0.000 | 03:51 |
| Control Test      | 0.080 | 03:51 |
| Air Blank         | 0.000 | 03:52 |
| Diagnostics Check | OK    | 03:52 |

---

Data Download Date/Time 11/12/2011 11:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLUDSWORTH, DEPUTY, J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: SIMANDL, DEPUTY  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 11/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SAVEA

DATE/TIME OF TEST: 11/13/2011 03:15:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:06 |                      |
| Air Blank         | 0.000 | 03:07 |                      |
| Control Test      | 0.079 | 03:07 |                      |
| Air Blank         | 0.000 | 03:08 |                      |
| Subject Sample #1 |       | 03:11 | No Sample Provided   |
| Breath Volume     | 0.000 | 03:11 |                      |
| Air Blank         | 0.000 | 03:11 |                      |
| Air Blank         | 0.000 | 03:13 |                      |
| Subject Sample #2 |       | 03:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:14 |                      |
| Air Blank         | 0.000 | 03:14 |                      |
| Control Test      | 0.078 | 03:14 |                      |
| Air Blank         | 0.000 | 03:15 |                      |
| Diagnostics Check | OK    | 03:15 |                      |

---

Data Download Date/Time 12/13/2011 07:57

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPAULDING, JORDON, S  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: CLARK, J  
ARRESTING OFFICER AGENCY: Green Cove Springs  
LAST AGENCY INSPECTION: 11/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: OBED

DATE/TIME OF TEST: 11/13/2011 03:45:36

---

|                   |       |       |                |
|-------------------|-------|-------|----------------|
| Diagnostics Check | OK    | 03:33 |                |
| Air Blank         | 0.000 | 03:33 |                |
| Control Test      | 0.079 | 03:34 |                |
| Air Blank         | 0.000 | 03:34 |                |
| Subject Sample #1 | 0.155 | 03:37 | Volume Not Met |
| Breath Volume     | 0.527 | 03:37 |                |
| Air Blank         | 0.000 | 03:38 |                |
| Air Blank         | 0.000 | 03:40 |                |
| Subject Sample #2 | 0.148 | 03:41 |                |
| Breath Volume     | 1.156 | 03:41 |                |
| Air Blank         | 0.000 | 03:41 |                |
| Air Blank         | 0.000 | 03:43 |                |
| Subject Sample #3 | 0.159 | 03:44 |                |
| Breath Volume     | 2.394 | 03:44 |                |
| Air Blank         | 0.000 | 03:44 |                |
| Control Test      | 0.078 | 03:45 |                |
| Air Blank         | 0.000 | 03:45 |                |
| Diagnostics Check | OK    | 03:45 |                |

---

Data Download Date/Time 12/13/2011 07:57

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURCH, ROBERT, M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: CORBETT, M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 11/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 11/16/2011 03:09:04

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:02 |
| Air Blank         | 0.000 | 03:03 |
| Control Test      | 0.079 | 03:03 |
| Air Blank         | 0.000 | 03:03 |
| Subject Sample #1 | 0.103 | 03:04 |
| Breath Volume     | 1.519 | 03:04 |
| Air Blank         | 0.000 | 03:05 |
| Air Blank         | 0.000 | 03:06 |
| Subject Sample #2 | 0.101 | 03:07 |
| Breath Volume     | 1.726 | 03:07 |
| Air Blank         | 0.000 | 03:08 |
| Control Test      | 0.078 | 03:08 |
| Air Blank         | 0.000 | 03:08 |
| Diagnostics Check | OK    | 03:09 |

---

Data Download Date/Time 12/13/2011 07:57

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, BRANDON, A  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS, ELISEO  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 11/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: VILLANUEVA

DATE/TIME OF TEST: 11/23/2011 03:34:44

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:28 |
| Air Blank         | 0.000 | 03:28 |
| Control Test      | 0.079 | 03:29 |
| Air Blank         | 0.000 | 03:29 |
| Subject Sample #1 | 0.286 | 03:30 |
| Breath Volume     | 3.609 | 03:30 |
| Air Blank         | 0.000 | 03:30 |
| Air Blank         | 0.000 | 03:32 |
| Subject Sample #2 | 0.284 | 03:33 |
| Breath Volume     | 3.156 | 03:33 |
| Air Blank         | 0.000 | 03:33 |
| Control Test      | 0.079 | 03:34 |
| Air Blank         | 0.000 | 03:34 |
| Diagnostics Check | OK    | 03:34 |

---

Data Download Date/Time 12/13/2011 07:57

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, BRANDON, A  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: HURLEY, C.  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 11/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 11/26/2011 22:38:43

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:31 |
| Air Blank         | 0.000 | 22:32 |
| Control Test      | 0.078 | 22:32 |
| Air Blank         | 0.000 | 22:33 |
| Subject Sample #1 | 0.101 | 22:34 |
| Breath Volume     | 1.527 | 22:34 |
| Air Blank         | 0.000 | 22:34 |
| Air Blank         | 0.000 | 22:36 |
| Subject Sample #2 | 0.095 | 22:37 |
| Breath Volume     | 1.476 | 22:37 |
| Air Blank         | 0.000 | 22:37 |
| Control Test      | 0.078 | 22:38 |
| Air Blank         | 0.000 | 22:38 |
| Diagnostics Check | OK    | 22:38 |

---

Data Download Date/Time 12/13/2011 07:57

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELCH, MARY, E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: STRICKLAND, W  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 10/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 11/20/2011 06:11:21

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 05:58 |                    |
| Air Blank         | 0.000 | 05:59 |                    |
| Control Test      | 0.080 | 05:59 |                    |
| Air Blank         | 0.000 | 06:00 |                    |
| Subject Sample #1 |       | 06:03 | No Sample Provided |
| Breath Volume     | 0.000 | 06:03 |                    |
| Air Blank         | 0.000 | 06:03 |                    |
| Air Blank         | 0.000 | 06:05 |                    |
| Subject Sample #2 | 0.157 | 06:06 |                    |
| Breath Volume     | 1.433 | 06:06 |                    |
| Air Blank         | 0.000 | 06:07 |                    |
| Air Blank         | 0.000 | 06:09 |                    |
| Subject Sample #3 | 0.160 | 06:09 |                    |
| Breath Volume     | 1.832 | 06:09 |                    |
| Air Blank         | 0.000 | 06:10 |                    |
| Control Test      | 0.077 | 06:10 |                    |
| Air Blank         | 0.000 | 06:11 |                    |
| Diagnostics Check | OK    | 06:11 |                    |

---

Data Download Date/Time 11/29/2011 16:26

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEDENBAUGH, JUSTIN, H  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: LEE, COLTON  
ARRESTING OFFICER AGENCY: Palatka PD  
LAST AGENCY INSPECTION: 10/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011802  
DRY GAS EXPIRY DATE: 04/28/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/01/2011 23:33:36

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:27 |
| Air Blank         | 0.000 | 23:27 |
| Control Test      | 0.078 | 23:27 |
| Air Blank         | 0.000 | 23:28 |
| Subject Sample #1 | 0.108 | 23:29 |
| Breath Volume     | 1.578 | 23:29 |
| Air Blank         | 0.000 | 23:29 |
| Air Blank         | 0.000 | 23:31 |
| Subject Sample #2 | 0.105 | 23:32 |
| Breath Volume     | 2.300 | 23:32 |
| Air Blank         | 0.000 | 23:32 |
| Control Test      | 0.078 | 23:32 |
| Air Blank         | 0.000 | 23:33 |
| Diagnostics Check | OK    | 23:33 |

---

Data Download Date/Time 12/01/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VEGA, KELVIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: VEGA, KELVIN  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 10/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011802  
DRY GAS EXPIRY DATE: 04/28/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: HEWES

DATE/TIME OF TEST: 11/14/2011 21:38:58

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 21:31 |
| Air Blank         | 0.000 | 21:31 |
| Control Test      | 0.077 | 21:32 |
| Air Blank         | 0.000 | 21:32 |
| Subject Sample #1 | 0.135 | 21:34 |
| Breath Volume     | 2.941 | 21:34 |
| Air Blank         | 0.000 | 21:35 |
| Air Blank         | 0.000 | 21:36 |
| Subject Sample #2 | 0.139 | 21:37 |
| Breath Volume     | 3.515 | 21:37 |
| Air Blank         | 0.000 | 21:38 |
| Control Test      | 0.076 | 21:38 |
| Air Blank         | 0.000 | 21:38 |
| Diagnostics Check | OK    | 21:38 |

---

Data Download Date/Time 12/01/2011 10:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIELS, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Kenneth City PD  
ARRESTING OFFICER: VIENO, MICHAEL  
ARRESTING OFFICER AGENCY: Kenneth City PD  
LAST AGENCY INSPECTION: 10/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MCCULLOUGH

DATE/TIME OF TEST: 11/09/2011 02:46:38

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:38 |
| Air Blank         | 0.000 | 02:38 |
| Control Test      | 0.079 | 02:38 |
| Air Blank         | 0.000 | 02:39 |
| Subject Sample #1 | 0.100 | 02:41 |
| Breath Volume     | 1.945 | 02:41 |
| Air Blank         | 0.000 | 02:42 |
| Air Blank         | 0.000 | 02:44 |
| Subject Sample #2 | 0.096 | 02:44 |
| Breath Volume     | 2.726 | 02:44 |
| Air Blank         | 0.000 | 02:45 |
| Control Test      | 0.077 | 02:45 |
| Air Blank         | 0.000 | 02:46 |
| Diagnostics Check | OK    | 02:46 |

---

Data Download Date/Time 11/28/2011 23:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIELS, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Kenneth City PD  
ARRESTING OFFICER: VIENO, MICHAEL  
ARRESTING OFFICER AGENCY: Kenneth City PD  
LAST AGENCY INSPECTION: 10/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WESTFALL

DATE/TIME OF TEST: 11/09/2011 23:07:53

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:00 |
| Air Blank         | 0.000 | 23:01 |
| Control Test      | 0.078 | 23:01 |
| Air Blank         | 0.000 | 23:02 |
| Subject Sample #1 | 0.321 | 23:02 |
| Breath Volume     | 3.769 | 23:02 |
| Air Blank         | 0.000 | 23:03 |
| Air Blank         | 0.000 | 23:05 |
| Subject Sample #2 | 0.334 | 23:05 |
| Breath Volume     | 3.335 | 23:05 |
| Air Blank         | 0.000 | 23:06 |
| Control Test      | 0.078 | 23:07 |
| Air Blank         | 0.000 | 23:07 |
| Diagnostics Check | OK    | 23:07 |

---

Data Download Date/Time 11/28/2011 23:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VIENO, MICHEL, J  
BREATH TEST OPERATOR AGENCY: Kenneth City PD  
ARRESTING OFFICER: 0050, WILLIAM  
ARRESTING OFFICER AGENCY: Kenneth City PD  
LAST AGENCY INSPECTION: 10/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MIZENER

DATE/TIME OF TEST: 11/11/2011 01:36:37

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:30 |
| Air Blank         | 0.000 | 01:30 |
| Control Test      | 0.081 | 01:30 |
| Air Blank         | 0.000 | 01:31 |
| Subject Sample #1 | 0.090 | 01:31 |
| Breath Volume     | 2.625 | 01:31 |
| Air Blank         | 0.000 | 01:32 |
| Air Blank         | 0.000 | 01:34 |
| Subject Sample #2 | 0.078 | 01:34 |
| Breath Volume     | 2.617 | 01:34 |
| Air Blank         | 0.000 | 01:35 |
| Control Test      | 0.080 | 01:35 |
| Air Blank         | 0.000 | 01:36 |
| Diagnostics Check | OK    | 01:36 |

---

Data Download Date/Time 11/28/2011 23:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VIENO, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Kenneth City PD  
ARRESTING OFFICER: VIENO, MICHAEL  
ARRESTING OFFICER AGENCY: Kenneth City PD  
LAST AGENCY INSPECTION: 10/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RIDENOUR

DATE/TIME OF TEST: 11/12/2011 02:26:55

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:20 |
| Air Blank         | 0.000 | 02:20 |
| Control Test      | 0.081 | 02:21 |
| Air Blank         | 0.000 | 02:21 |
| Subject Sample #1 | 0.104 | 02:22 |
| Breath Volume     | 2.062 | 02:22 |
| Air Blank         | 0.000 | 02:23 |
| Air Blank         | 0.000 | 02:24 |
| Subject Sample #2 | 0.109 | 02:25 |
| Breath Volume     | 1.195 | 02:25 |
| Air Blank         | 0.000 | 02:25 |
| Control Test      | 0.081 | 02:26 |
| Air Blank         | 0.000 | 02:26 |
| Diagnostics Check | OK    | 02:26 |

---

Data Download Date/Time 11/28/2011 23:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VIENO, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Kenneth City PD  
ARRESTING OFFICER: VIENO, MICHAEL  
ARRESTING OFFICER AGENCY: Kenneth City PD  
LAST AGENCY INSPECTION: 10/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LUBISCO

DATE/TIME OF TEST: 11/15/2011 03:11:21

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:04 |
| Air Blank         | 0.000 | 03:05 |
| Control Test      | 0.079 | 03:05 |
| Air Blank         | 0.000 | 03:06 |
| Subject Sample #1 | 0.077 | 03:06 |
| Breath Volume     | 1.636 | 03:06 |
| Air Blank         | 0.000 | 03:07 |
| Air Blank         | 0.000 | 03:09 |
| Subject Sample #2 | 0.075 | 03:09 |
| Breath Volume     | 2.019 | 03:09 |
| Air Blank         | 0.000 | 03:10 |
| Control Test      | 0.078 | 03:10 |
| Air Blank         | 0.000 | 03:11 |
| Diagnostics Check | OK    | 03:11 |

---

Data Download Date/Time 11/28/2011 23:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VIENO, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Kenneth City PD  
ARRESTING OFFICER: DANIELS, KEVIN  
ARRESTING OFFICER AGENCY: Kenneth City PD  
LAST AGENCY INSPECTION: 10/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: RACICOT

DATE/TIME OF TEST: 11/16/2011 02:26:11

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:19 |
| Air Blank         | 0.000 | 02:20 |
| Control Test      | 0.079 | 02:20 |
| Air Blank         | 0.000 | 02:21 |
| Subject Sample #1 | 0.185 | 02:21 |
| Breath Volume     | 1.386 | 02:21 |
| Air Blank         | 0.000 | 02:22 |
| Air Blank         | 0.000 | 02:24 |
| Subject Sample #2 | 0.193 | 02:24 |
| Breath Volume     | 1.742 | 02:24 |
| Air Blank         | 0.000 | 02:25 |
| Control Test      | 0.078 | 02:25 |
| Air Blank         | 0.000 | 02:26 |
| Diagnostics Check | OK    | 02:26 |

---

Data Download Date/Time 11/28/2011 23:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VIENO, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Kenneth City PD  
ARRESTING OFFICER: VIENO, MICHAEL  
ARRESTING OFFICER AGENCY: Kenneth City PD  
LAST AGENCY INSPECTION: 10/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GILLISPIE

DATE/TIME OF TEST: 11/19/2011 03:45:48

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:39 |
| Air Blank         | 0.000 | 03:39 |
| Control Test      | 0.081 | 03:39 |
| Air Blank         | 0.000 | 03:40 |
| Subject Sample #1 | 0.140 | 03:40 |
| Breath Volume     | 1.191 | 03:40 |
| Air Blank         | 0.000 | 03:41 |
| Air Blank         | 0.000 | 03:43 |
| Subject Sample #2 | 0.140 | 03:44 |
| Breath Volume     | 1.718 | 03:44 |
| Air Blank         | 0.000 | 03:44 |
| Control Test      | 0.079 | 03:45 |
| Air Blank         | 0.000 | 03:45 |
| Diagnostics Check | OK    | 03:45 |

---

Data Download Date/Time 11/28/2011 23:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001650

INSTRUMENT REGISTERED TO: KEY BISCAZYNE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAL, NELIA, M  
BREATH TEST OPERATOR AGENCY: Key Biscayne PD  
ARRESTING OFFICER: ALVAREZ, JORGE  
ARRESTING OFFICER AGENCY: Key Biscayne PD  
LAST AGENCY INSPECTION: 10/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG016502  
DRY GAS EXPIRY DATE: 06/14/2010  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DALMAU

DATE/TIME OF TEST: 11/19/2011 03:57:02

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:48 |                      |
| Air Blank         | 0.000 | 03:49 |                      |
| Control Test      | 0.081 | 03:49 |                      |
| Air Blank         | 0.000 | 03:49 |                      |
| Subject Sample #1 |       | 03:53 | No Sample Provided   |
| Breath Volume     | 0.000 | 03:53 |                      |
| Air Blank         | 0.000 | 03:53 |                      |
| Air Blank         | 0.000 | 03:55 |                      |
| Subject Sample #2 |       | 03:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:55 |                      |
| Air Blank         | 0.000 | 03:56 |                      |
| Control Test      | 0.082 | 03:56 |                      |
| Air Blank         | 0.000 | 03:56 |                      |
| Diagnostics Check | OK    | 03:57 |                      |

---

Data Download Date/Time 11/29/2011 17:26

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001652

INSTRUMENT REGISTERED TO: WEST MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER, ROGER, G  
BREATH TEST OPERATOR AGENCY: FDLE  
ARRESTING OFFICER: TEST, TEST  
ARRESTING OFFICER AGENCY: FDLE  
LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG112601  
DRY GAS EXPIRY DATE: 05/06/2013  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 07:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/02/2011 09:58:16

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 09:51 |
| Air Blank         | 0.000 | 09:52 |
| Control Test      | 0.077 | 09:52 |
| Air Blank         | 0.000 | 09:53 |
| Subject Sample #1 | 0.082 | 09:53 |
| Breath Volume     | 1.812 | 09:53 |
| Air Blank         | 0.000 | 09:54 |
| Air Blank         | 0.000 | 09:56 |
| Subject Sample #2 | 0.081 | 09:56 |
| Breath Volume     | 1.285 | 09:56 |
| Air Blank         | 0.000 | 09:57 |
| Control Test      | 0.077 | 09:57 |
| Air Blank         | 0.000 | 09:58 |
| Diagnostics Check | OK    | 09:58 |

---

Data Download Date/Time 11/02/2011 14:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001652

INSTRUMENT REGISTERED TO: WEST MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER, ROGER, G  
BREATH TEST OPERATOR AGENCY: FDLE  
ARRESTING OFFICER: TEST, TEST  
ARRESTING OFFICER AGENCY: FDLE  
LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG112601  
DRY GAS EXPIRY DATE: 05/06/2013  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 07:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/02/2011 10:12:23

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 10:06 |
| Air Blank         | 0.000 | 10:06 |
| Control Test      | 0.078 | 10:06 |
| Air Blank         | 0.000 | 10:07 |
| Subject Sample #1 | 0.081 | 10:07 |
| Breath Volume     | 1.300 | 10:07 |
| Air Blank         | 0.000 | 10:08 |
| Air Blank         | 0.000 | 10:10 |
| Subject Sample #2 | 0.081 | 10:10 |
| Breath Volume     | 1.429 | 10:10 |
| Air Blank         | 0.000 | 10:11 |
| Control Test      | 0.078 | 10:11 |
| Air Blank         | 0.000 | 10:12 |
| Diagnostics Check | OK    | 10:12 |

---

Data Download Date/Time 11/02/2011 14:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHIPP, KRISTEN, S

BREATH TEST OPERATOR AGENCY: FDLE

ARRESTING OFFICER: SMITH, JOHN

ARRESTING OFFICER AGENCY: FDLE

LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG032601

DRY GAS EXPIRY DATE: 11/22/2012

VIOLATION CODE: SYS CHECK

ARREST TIME: 15:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/02/2011 15:50:15

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 15:46 |                    |
| Air Blank         | 0.000 | 15:46 |                    |
| Control Test      | 0.081 | 15:47 |                    |
| Air Blank         | 0.000 | 15:47 |                    |
| Subject Sample #1 |       | 15:48 | Interferent Detect |
| Breath Volume     | 1.152 | 15:48 |                    |
| Air Blank         | 0.000 | 15:49 |                    |
| Control Test      | 0.084 | 15:49 |                    |
| Air Blank         | 0.000 | 15:50 |                    |
| Diagnostics Check | OK    | 15:50 |                    |

---

Data Download Date/Time 11/03/2011 07:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHIPP, KRISTEN, S

BREATH TEST OPERATOR AGENCY: FDLE

ARRESTING OFFICER: SMITH, JOHN

ARRESTING OFFICER AGENCY: FDLE

LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG032601

DRY GAS EXPIRY DATE: 11/22/2012

VIOLATION CODE: SYS CHECK

ARREST TIME: 15:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/02/2011 16:39:36

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 16:32 |
| Air Blank         | 0.000 | 16:32 |
| Control Test      | 0.081 | 16:32 |
| Air Blank         | 0.000 | 16:33 |
| Subject Sample #1 | 0.081 | 16:34 |
| Breath Volume     | 1.714 | 16:34 |
| Air Blank         | 0.000 | 16:35 |
| Air Blank         | 0.000 | 16:37 |
| Subject Sample #2 | 0.080 | 16:38 |
| Breath Volume     | 1.777 | 16:38 |
| Air Blank         | 0.000 | 16:38 |
| Control Test      | 0.081 | 16:39 |
| Air Blank         | 0.000 | 16:39 |
| Diagnostics Check | OK    | 16:39 |

---

Data Download Date/Time 11/03/2011 07:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EMERT, SHAWN, A  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: EMERT, SHAWN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17910080A3  
DRY GAS EXPIRY DATE: 07/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: PRESTON

DATE/TIME OF TEST: 11/04/2011 03:14:37

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 03:08 |
| Air Blank         | 0.000 | 03:08 |
| Control Test      | 0.080 | 03:09 |
| Air Blank         | 0.000 | 03:09 |
| Subject Sample #1 | 0.261 | 03:10 |
| Breath Volume     | 2.539 | 03:10 |
| Air Blank         | 0.000 | 03:10 |
| Air Blank         | 0.000 | 03:12 |
| Subject Sample #2 | 0.257 | 03:13 |
| Breath Volume     | 1.503 | 03:13 |
| Air Blank         | 0.000 | 03:13 |
| Control Test      | 0.078 | 03:14 |
| Air Blank         | 0.000 | 03:14 |
| Diagnostics Check | OK    | 03:14 |

---

Data Download Date/Time 11/23/2011 14:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKTON, QUENTIN, Q  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: BECKTON, QUENTIN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17910080A3  
DRY GAS EXPIRY DATE: 07/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MERICH

DATE/TIME OF TEST: 11/04/2011 03:30:37

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:21 |                      |
| Air Blank         | 0.000 | 03:21 |                      |
| Control Test      | 0.079 | 03:21 |                      |
| Air Blank         | 0.000 | 03:22 |                      |
| Subject Sample #1 | 0.232 | 03:24 |                      |
| Breath Volume     | 1.460 | 03:24 |                      |
| Air Blank         | 0.000 | 03:25 |                      |
| Air Blank         | 0.000 | 03:26 |                      |
| Subject Sample #2 |       | 03:29 | Subject Test Refused |
| Breath Volume     | 1.683 | 03:29 |                      |
| Air Blank         | 0.000 | 03:29 |                      |
| Control Test      | 0.077 | 03:30 |                      |
| Air Blank         | 0.000 | 03:30 |                      |
| Diagnostics Check | OK    | 03:30 |                      |

---

Data Download Date/Time 11/23/2011 14:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKTON, QUENTIN, Q  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: BECKTON, QUENTIN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17910080A3  
DRY GAS EXPIRY DATE: 07/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: DALTO

DATE/TIME OF TEST: 11/06/2011 23:25:43

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:18 |
| Air Blank         | 0.000 | 23:19 |
| Control Test      | 0.079 | 23:19 |
| Air Blank         | 0.000 | 23:20 |
| Subject Sample #1 | 0.203 | 23:21 |
| Breath Volume     | 1.550 | 23:21 |
| Air Blank         | 0.000 | 23:21 |
| Air Blank         | 0.000 | 23:23 |
| Subject Sample #2 | 0.197 | 23:24 |
| Breath Volume     | 1.570 | 23:24 |
| Air Blank         | 0.000 | 23:24 |
| Control Test      | 0.080 | 23:25 |
| Air Blank         | 0.000 | 23:25 |
| Diagnostics Check | OK    | 23:25 |

---

Data Download Date/Time 11/23/2011 14:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKTON, QUENTIN, Q  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: GAZDICK, STEVE  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17910080A3  
DRY GAS EXPIRY DATE: 07/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 11/13/2011 02:58:33

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:50 |
| Air Blank         | 0.000 | 02:51 |
| Control Test      | 0.080 | 02:51 |
| Air Blank         | 0.000 | 02:52 |
| Subject Sample #1 | 0.165 | 02:53 |
| Breath Volume     | 1.218 | 02:53 |
| Air Blank         | 0.000 | 02:54 |
| Air Blank         | 0.000 | 02:56 |
| Subject Sample #2 | 0.174 | 02:57 |
| Breath Volume     | 1.558 | 02:57 |
| Air Blank         | 0.000 | 02:57 |
| Control Test      | 0.079 | 02:57 |
| Air Blank         | 0.000 | 02:58 |
| Diagnostics Check | OK    | 02:58 |

---

Data Download Date/Time 11/23/2011 14:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELLAMY, MATTHEW, Q  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: LIVINGSTON, K  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17910080A3  
DRY GAS EXPIRY DATE: 07/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/14/2011 01:41:31

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:35 |
| Air Blank         | 0.000 | 01:35 |
| Control Test      | 0.079 | 01:36 |
| Air Blank         | 0.000 | 01:36 |
| Subject Sample #1 | 0.196 | 01:37 |
| Breath Volume     | 3.273 | 01:37 |
| Air Blank         | 0.000 | 01:37 |
| Air Blank         | 0.000 | 01:39 |
| Subject Sample #2 | 0.196 | 01:40 |
| Breath Volume     | 3.273 | 01:40 |
| Air Blank         | 0.000 | 01:40 |
| Control Test      | 0.079 | 01:40 |
| Air Blank         | 0.000 | 01:41 |
| Diagnostics Check | OK    | 01:41 |

---

Data Download Date/Time 11/23/2011 14:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKTON, QUENTIN, Q  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: BECKTON, QUENTIN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17910080A3  
DRY GAS EXPIRY DATE: 07/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: YELTON

DATE/TIME OF TEST: 11/20/2011 01:15:40

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:09 |
| Air Blank         | 0.000 | 01:09 |
| Control Test      | 0.079 | 01:10 |
| Air Blank         | 0.000 | 01:10 |
| Subject Sample #1 | 0.294 | 01:11 |
| Breath Volume     | 1.871 | 01:11 |
| Air Blank         | 0.000 | 01:11 |
| Air Blank         | 0.000 | 01:13 |
| Subject Sample #2 | 0.291 | 01:14 |
| Breath Volume     | 1.445 | 01:14 |
| Air Blank         | 0.000 | 01:14 |
| Control Test      | 0.078 | 01:15 |
| Air Blank         | 0.000 | 01:15 |
| Diagnostics Check | OK    | 01:15 |

---

Data Download Date/Time 11/23/2011 14:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EMERT, SHAWN, A  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: EMERT, SHAWN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17910080A3  
DRY GAS EXPIRY DATE: 07/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 11/20/2011 02:39:08

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:29 |
| Air Blank         | 0.000 | 02:30 |
| Control Test      | 0.079 | 02:30 |
| Air Blank         | 0.000 | 02:30 |
| Subject Sample #1 | 0.210 | 02:31 |
| Breath Volume     | 1.402 | 02:31 |
| Air Blank         | 0.000 | 02:32 |
| Air Blank         | 0.000 | 02:34 |
| Subject Sample #2 | 0.180 | 02:34 |
| Breath Volume     | 1.910 | 02:34 |
| Air Blank         | 0.000 | 02:35 |
| Air Blank         | 0.000 | 02:37 |
| Subject Sample #3 | 0.164 | 02:37 |
| Breath Volume     | 2.191 | 02:37 |
| Air Blank         | 0.000 | 02:38 |
| Control Test      | 0.078 | 02:38 |
| Air Blank         | 0.000 | 02:39 |
| Diagnostics Check | OK    | 02:39 |

---

Data Download Date/Time 11/23/2011 14:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKTON, QUENTIN, Q  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: CARROLL, JEFF  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17910080A3  
DRY GAS EXPIRY DATE: 07/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FREEL

DATE/TIME OF TEST: 11/20/2011 04:06:26

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:00 |
| Air Blank         | 0.000 | 04:00 |
| Control Test      | 0.079 | 04:00 |
| Air Blank         | 0.000 | 04:01 |
| Subject Sample #1 | 0.156 | 04:02 |
| Breath Volume     | 2.449 | 04:02 |
| Air Blank         | 0.000 | 04:02 |
| Air Blank         | 0.000 | 04:04 |
| Subject Sample #2 | 0.155 | 04:04 |
| Breath Volume     | 1.488 | 04:04 |
| Air Blank         | 0.000 | 04:05 |
| Control Test      | 0.078 | 04:05 |
| Air Blank         | 0.000 | 04:06 |
| Diagnostics Check | OK    | 04:06 |

---

Data Download Date/Time 11/23/2011 14:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHORTER, DENNIS, F  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: SHORTER, DENNIS  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17910080A3  
DRY GAS EXPIRY DATE: 07/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: KNOWLES

DATE/TIME OF TEST: 11/26/2011 04:28:52

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 04:22 |
| Air Blank         | 0.000 | 04:22 |
| Control Test      | 0.080 | 04:23 |
| Air Blank         | 0.000 | 04:23 |
| Subject Sample #1 | 0.126 | 04:24 |
| Breath Volume     | 3.820 | 04:24 |
| Air Blank         | 0.000 | 04:25 |
| Air Blank         | 0.000 | 04:26 |
| Subject Sample #2 | 0.119 | 04:27 |
| Breath Volume     | 3.093 | 04:27 |
| Air Blank         | 0.000 | 04:27 |
| Control Test      | 0.078 | 04:28 |
| Air Blank         | 0.000 | 04:28 |
| Diagnostics Check | OK    | 04:28 |

---

Data Download Date/Time 12/21/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHORTER, DENNIS, F  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: STAGGERS, NAPOLIAN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 11/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17910080A3  
DRY GAS EXPIRY DATE: 07/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 11:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FOLLMER

DATE/TIME OF TEST: 11/28/2011 01:37:01

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 01:30 |
| Air Blank         | 0.000 | 01:31 |
| Control Test      | 0.079 | 01:31 |
| Air Blank         | 0.000 | 01:31 |
| Subject Sample #1 | 0.100 | 01:32 |
| Breath Volume     | 1.937 | 01:32 |
| Air Blank         | 0.000 | 01:33 |
| Air Blank         | 0.000 | 01:35 |
| Subject Sample #2 | 0.098 | 01:35 |
| Breath Volume     | 1.601 | 01:35 |
| Air Blank         | 0.000 | 01:36 |
| Control Test      | 0.079 | 01:36 |
| Air Blank         | 0.000 | 01:36 |
| Diagnostics Check | OK    | 01:37 |

---

Data Download Date/Time 12/21/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHANAHAN, JAKE, L

BREATH TEST OPERATOR AGENCY: FDLE

ARRESTING OFFICER: TEST, TEST

ARRESTING OFFICER AGENCY: FDLE

LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG123001

DRY GAS EXPIRY DATE: 08/08/2013

VIOLATION CODE: SYS CHECK

ARREST TIME: 08:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/01/2011 10:07:21

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 10:00 |
| Air Blank         | 0.000 | 10:01 |
| Control Test      | 0.076 | 10:01 |
| Air Blank         | 0.000 | 10:02 |
| Subject Sample #1 | 0.082 | 10:02 |
| Breath Volume     | 1.109 | 10:02 |
| Air Blank         | 0.000 | 10:03 |
| Air Blank         | 0.000 | 10:05 |
| Subject Sample #2 | 0.081 | 10:05 |
| Breath Volume     | 1.324 | 10:05 |
| Air Blank         | 0.000 | 10:06 |
| Control Test      | 0.076 | 10:06 |
| Air Blank         | 0.000 | 10:07 |
| Diagnostics Check | OK    | 10:07 |

---

Data Download Date/Time 11/01/2011 15:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHANAHAN, JAKE, L  
BREATH TEST OPERATOR AGENCY: FDLE  
ARRESTING OFFICER: TEST2, TEST2  
ARRESTING OFFICER AGENCY: FDLE  
LAST AGENCY INSPECTION: 10/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG123001  
DRY GAS EXPIRY DATE: 08/08/2013  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 08:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 11/01/2011 10:22:02

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 10:15 |
| Air Blank         | 0.000 | 10:16 |
| Control Test      | 0.076 | 10:16 |
| Air Blank         | 0.000 | 10:17 |
| Subject Sample #1 | 0.083 | 10:17 |
| Breath Volume     | 1.710 | 10:17 |
| Air Blank         | 0.000 | 10:18 |
| Air Blank         | 0.000 | 10:20 |
| Subject Sample #2 | 0.080 | 10:20 |
| Breath Volume     | 1.386 | 10:20 |
| Air Blank         | 0.000 | 10:21 |
| Control Test      | 0.076 | 10:21 |
| Air Blank         | 0.000 | 10:21 |
| Diagnostics Check | OK    | 10:22 |

---

Data Download Date/Time 11/01/2011 15:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMB, JAMES, A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: LAMB, JAMES  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: ERAZO

DATE/TIME OF TEST: 11/02/2011 22:51:47

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 22:45 |
| Air Blank         | 0.000 | 22:46 |
| Control Test      | 0.078 | 22:46 |
| Air Blank         | 0.000 | 22:46 |
| Subject Sample #1 | 0.183 | 22:47 |
| Breath Volume     | 3.007 | 22:47 |
| Air Blank         | 0.000 | 22:48 |
| Air Blank         | 0.000 | 22:49 |
| Subject Sample #2 | 0.174 | 22:50 |
| Breath Volume     | 2.839 | 22:50 |
| Air Blank         | 0.000 | 22:50 |
| Control Test      | 0.078 | 22:51 |
| Air Blank         | 0.000 | 22:51 |
| Diagnostics Check | OK    | 22:51 |

---

Data Download Date/Time 11/28/2011 10:07

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMB, JAMES, A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: LAMB, JAMES  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 11/08/2011 02:42:18

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 02:40 |            |
| Air Blank         | 0.000 | 02:41 |            |
| Control Test      | 0.078 | 02:41 |            |
| Air Blank         |       | 02:41 | RFI Detect |
| Air Blank         | 0.000 | 02:42 |            |

---

Data Download Date/Time 11/28/2011 10:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAMB, JAMES, A

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: LAMB, JAMES

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 02:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 11/08/2011 02:52:14

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:45 |
| Air Blank         | 0.000 | 02:46 |
| Control Test      | 0.079 | 02:46 |
| Air Blank         | 0.000 | 02:47 |
| Subject Sample #1 | 0.146 | 02:47 |
| Breath Volume     | 1.925 | 02:47 |
| Air Blank         | 0.000 | 02:48 |
| Air Blank         | 0.000 | 02:49 |
| Subject Sample #2 | 0.150 | 02:50 |
| Breath Volume     | 1.921 | 02:50 |
| Air Blank         | 0.000 | 02:51 |
| Control Test      | 0.077 | 02:51 |
| Air Blank         | 0.000 | 02:52 |
| Diagnostics Check | OK    | 02:52 |

---

Data Download Date/Time 11/28/2011 10:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMB, JAMES, A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: LAMB, JAMES  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 11/12/2011 00:25:57

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 00:19 |
| Air Blank         | 0.000 | 00:19 |
| Control Test      | 0.079 | 00:20 |
| Air Blank         | 0.000 | 00:20 |
| Subject Sample #1 | 0.165 | 00:21 |
| Breath Volume     | 1.605 | 00:21 |
| Air Blank         | 0.000 | 00:22 |
| Air Blank         | 0.000 | 00:23 |
| Subject Sample #2 | 0.158 | 00:24 |
| Breath Volume     | 1.937 | 00:24 |
| Air Blank         | 0.000 | 00:25 |
| Control Test      | 0.078 | 00:25 |
| Air Blank         | 0.000 | 00:25 |
| Diagnostics Check | OK    | 00:25 |

---

Data Download Date/Time 11/28/2011 10:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAMB, JAMES, A

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: LAMB, JAMES

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 02:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: FRISSORA

DATE/TIME OF TEST: 11/12/2011 02:55:15

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 02:48 |
| Air Blank         | 0.000 | 02:49 |
| Control Test      | 0.080 | 02:49 |
| Air Blank         | 0.000 | 02:50 |
| Subject Sample #1 | 0.155 | 02:50 |
| Breath Volume     | 1.992 | 02:50 |
| Air Blank         | 0.000 | 02:51 |
| Air Blank         | 0.000 | 02:53 |
| Subject Sample #2 | 0.156 | 02:53 |
| Breath Volume     | 1.695 | 02:53 |
| Air Blank         | 0.000 | 02:54 |
| Control Test      | 0.079 | 02:54 |
| Air Blank         | 0.000 | 02:55 |
| Diagnostics Check | OK    | 02:55 |

---

Data Download Date/Time 11/28/2011 10:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMB, JAMES, A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: LAMB, JAMES  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/28/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 11/22/2011 23:28:29

---

|                   |       |       |
|-------------------|-------|-------|
| Diagnostics Check | OK    | 23:21 |
| Air Blank         | 0.000 | 23:22 |
| Control Test      | 0.078 | 23:22 |
| Air Blank         | 0.000 | 23:23 |
| Subject Sample #1 | 0.144 | 23:23 |
| Breath Volume     | 2.140 | 23:23 |
| Air Blank         | 0.000 | 23:24 |
| Air Blank         | 0.000 | 23:26 |
| Subject Sample #2 | 0.139 | 23:26 |
| Breath Volume     | 2.109 | 23:26 |
| Air Blank         | 0.000 | 23:27 |
| Control Test      | 0.077 | 23:27 |
| Air Blank         | 0.000 | 23:28 |
| Diagnostics Check | OK    | 23:28 |

---

Data Download Date/Time 11/28/2011 10:07