

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DELAUGHTER, MARC, C

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: SMITH, JAMES

ARRESTING OFFICER AGENCY: Cocoa PD

LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/28/2011

VIOLATION CODE: DUI

ARREST TIME: 03:53

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 04/10/2010 04:23:19

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Subject Sample #1	0.168	04:18
Breath Volume	1.933	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample #2	0.187	04:21
Breath Volume	1.527	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time 04/13/2010 22:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS, STEVEN, E
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: OWENS, STEVEN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 595761
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: CARRANO

DATE/TIME OF TEST: 04/09/2010 23:27:45

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.079	23:25	
Air Blank	0.000	23:25	
Subject Sample #1		23:26	RFI Detect
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Control Test	0.079	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time 04/21/2010 04:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS, STEVEN, E
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: OWENS, STEVEN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 595761
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: CARRANO

DATE/TIME OF TEST: 04/09/2010 23:35:48

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Subject Sample #1	0.055	23:31
Breath Volume	3.167	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample #2	0.056	23:34
Breath Volume	3.269	23:34
Air Blank	0.000	23:34
Control Test	0.078	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time 04/21/2010 04:29

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS, STEVEN, E
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: OWENS, STEVEN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 595761
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/19/2010 03:38:52

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Subject Sample #1	0.286	03:34
Breath Volume	1.222	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample #2	0.274	03:37
Breath Volume	1.332	03:37
Air Blank	0.000	03:37
Control Test	0.078	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time 04/21/2010 04:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YUERGENS, AARON, M

BREATH TEST OPERATOR AGENCY: Satellite Beach PD

ARRESTING OFFICER: YUERGENS, AARON

ARRESTING OFFICER AGENCY: Satellite Beach PD

LAST AGENCY INSPECTION: 04/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 595761

DRY GAS EXPIRY DATE: 04/13/2011

VIOLATION CODE: DUI

ARREST TIME: 02:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 04/23/2010 03:11:43

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Subject Sample #1	0.137	03:07
Breath Volume	2.148	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample #2	0.136	03:10
Breath Volume	2.601	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time 05/12/2010 11:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: OWENS, STEVEN, E

BREATH TEST OPERATOR AGENCY: Satellite Beach PD

ARRESTING OFFICER: FROST, DARREN

ARRESTING OFFICER AGENCY: Satellite Beach PD

LAST AGENCY INSPECTION: 04/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 595761

DRY GAS EXPIRY DATE: 04/13/2011

VIOLATION CODE: DUI

ARREST TIME: 03:22

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 04/28/2010 04:29:53

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Subject Sample #1	0.242	04:24
Breath Volume	1.726	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample #2	0.242	04:28
Breath Volume	2.187	04:28
Air Blank	0.000	04:28
Control Test	0.079	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time 05/12/2010 11:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, JOHN, T
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SOUTHER, TODD
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822601I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 04/10/2010 04:55:30

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.078	04:49
Air Blank	0.000	04:50
Subject Sample #1	0.158	04:51
Breath Volume	2.960	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample #2	0.148	04:53
Breath Volume	2.195	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time 04/17/2010 22:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON, MONICA, L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WHITE, KEVIN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822601I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: MUNGUIA

DATE/TIME OF TEST: 04/16/2010 14:03:52

Diagnostics Check	OK	14:01	
Air Blank	0.000	14:01	
Control Test	0.079	14:01	
Air Blank	0.000	14:02	
Subject Sample #1		14:02	RFI Detect
Breath Volume	0.000	14:02	
Air Blank	0.000	14:02	
Control Test	0.078	14:03	
Air Blank	0.000	14:03	
Diagnostics Check	OK	14:03	

Data Download Date/Time 04/17/2010 22:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON, MONICA, L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WHITE, KEVIN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822601I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: MUNGUIA

DATE/TIME OF TEST: 04/16/2010 14:12:29

Diagnostics Check	OK	14:06
Air Blank	0.000	14:06
Control Test	0.078	14:07
Air Blank	0.000	14:07
Subject Sample #1	0.251	14:07
Breath Volume	3.125	14:07
Air Blank	0.000	14:08
Air Blank	0.000	14:10
Subject Sample #2	0.252	14:10
Breath Volume	3.148	14:10
Air Blank	0.000	14:11
Control Test	0.078	14:11
Air Blank	0.000	14:12
Diagnostics Check	OK	14:12

Data Download Date/Time 04/17/2010 22:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, YOLANDA, C
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: DEAN, GREG
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822601I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: PROB/PAROL
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/19/2010 19:56:50

Diagnostics Check	OK	19:48
Air Blank	0.000	19:49
Control Test	0.078	19:49
Air Blank	0.000	19:50
Subject Sample #1	0.395	19:51
Breath Volume	1.230	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample #2	0.395	19:55
Breath Volume	1.312	19:55
Air Blank	0.000	19:55
Control Test	0.076	19:56
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time 05/24/2010 13:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRYANT, RICHARD, W

BREATH TEST OPERATOR AGENCY: Hardee County SO

ARRESTING OFFICER: WARD, SHANE

ARRESTING OFFICER AGENCY: Hardee County SO

LAST AGENCY INSPECTION: 04/17/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822601I

DRY GAS EXPIRY DATE: 08/21/2010

VIOLATION CODE: DUI

ARREST TIME: 20:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/22/2010 21:40:06

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:37	
Control Test	0.079	21:38	
Air Blank	0.000	21:38	
Subject Sample #1		21:38	Subject Test Refused
Breath Volume	0.000	21:38	
Air Blank	0.000	21:39	
Control Test	0.079	21:39	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

Data Download Date/Time 05/24/2010 13:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, YOLANDA, C
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: HART, LYLE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822601I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: DARTY

DATE/TIME OF TEST: 04/25/2010 00:14:40

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:09
Subject Sample #1	0.121	00:10
Breath Volume	2.917	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample #2	0.123	00:13
Breath Volume	2.687	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time 05/24/2010 13:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, M, S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CROSS, M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GOLDBERG

DATE/TIME OF TEST: 04/02/2010 03:09:49

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.078	02:59	
Air Blank	0.000	02:59	
Subject Sample #1	0.151	03:03	Volume Not Met
Breath Volume	0.082	03:03	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample #2		03:08	Subject Test Refused
Breath Volume	0.085	03:08	
Air Blank	0.000	03:08	
Control Test	0.079	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VILLANO, DANIEL, F

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: VILLANO, DANIEL

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 21:29

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CASPOLICH

DATE/TIME OF TEST: 04/05/2010 22:32:53

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.079	22:29	
Air Blank	0.000	22:29	
Subject Sample #1		22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.079	22:32	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACOCK, CHERYL, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYERS, DEPUTY B.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: MILCHAN

DATE/TIME OF TEST: 04/09/2010 23:02:27

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample #1	0.037	22:58
Breath Volume	1.546	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample #2	0.037	23:00
Breath Volume	1.464	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CROSS, M, S

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: CROSS, M

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 03:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MARCIA-CHAVEZ

DATE/TIME OF TEST: 04/11/2010 04:22:58

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:17
Air Blank	0.000	04:17
Subject Sample #1	0.174	04:18
Breath Volume	1.582	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample #2	0.171	04:21
Breath Volume	1.597	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACOCK, CHERYL, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CORLETT, SGT. R.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 04/14/2010 03:57:49

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Subject Sample #1	0.175	03:53
Breath Volume	2.292	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample #2	0.179	03:56
Breath Volume	1.882	03:56
Air Blank	0.000	03:56
Control Test	0.078	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS, P, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GARRETT, L
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:11

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 04/14/2010 08:37:34

Diagnostics Check	OK	08:31
Air Blank	0.000	08:32
Control Test	0.079	08:32
Air Blank	0.000	08:32
Subject Sample #1	0.000	08:33
Breath Volume	1.640	08:33
Air Blank	0.000	08:33
Air Blank	0.000	08:35
Subject Sample #2	0.000	08:36
Breath Volume	1.960	08:36
Air Blank	0.000	08:36
Control Test	0.080	08:36
Air Blank	0.000	08:37
Diagnostics Check	OK	08:37

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CROSS, M, S

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: CROSS, M

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 01:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DIPIAZZA

DATE/TIME OF TEST: 04/18/2010 02:14:18

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.078	02:12	
Air Blank	0.000	02:12	
Subject Sample #1		02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.078	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY, ROBERT, S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAY, ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: SAYLOR

DATE/TIME OF TEST: 04/18/2010 23:54:56

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.078	23:52	
Air Blank	0.000	23:53	
Subject Sample #1		23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:54	
Control Test	0.078	23:54	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACOCK, CHERYL, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CIOFFI, DEPUTY D.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: STOCKS

DATE/TIME OF TEST: 04/21/2010 21:02:32

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Subject Sample #1	0.164	20:57
Breath Volume	1.183	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample #2	0.175	21:01
Breath Volume	1.304	21:01
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAY, ROBERT, S

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: RAY, ROBERT

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 19:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: WEIN

DATE/TIME OF TEST: 04/23/2010 20:35:01

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:25	
Control Test	0.080	20:26	
Air Blank	0.000	20:26	
Subject Sample #1	0.118	20:29	Volume Not Met
Breath Volume	0.046	20:29	
Air Blank	0.000	20:30	
Air Blank	0.000	20:32	
Subject Sample #2		20:33	Subject Test Refused
Breath Volume	0.050	20:33	
Air Blank	0.000	20:34	
Control Test	0.078	20:34	
Air Blank	0.000	20:34	
Diagnostics Check	OK	20:35	

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VILLANO, DANIEL, F

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: VILLANO, DANIEL

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 03:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/24/2010 04:20:57

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:14
Subject Sample #1	0.302	04:16
Breath Volume	1.957	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample #2	0.309	04:19
Breath Volume	2.394	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIEGLO, DEBORAH, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CIEGLO, DEBORAH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 10:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 04/24/2010 22:56:51

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.079	22:52	
Air Blank	0.000	22:53	
Subject Sample #1		22:56	Sequence Aborted
Breath Volume	0.000	22:56	
Air Blank	0.000	22:56	

Data Download Date/Time 04/28/2010 02:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACOCK, CHERYL, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUBBARD, DEPUTY J.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: LOPEZ-HERNANDEZ

DATE/TIME OF TEST: 04/28/2010 04:26:46

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.079	04:13	
Air Blank	0.000	04:14	
Subject Sample #1	0.151	04:17	Volume Not Met
Breath Volume	0.128	04:17	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample #2	0.155	04:21	
Breath Volume	2.023	04:21	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample #3	0.162	04:25	
Breath Volume	1.828	04:25	
Air Blank	0.000	04:25	
Control Test	0.079	04:26	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time 05/21/2010 01:53

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACOCK, CHERYL, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NAULT, DEPUTY C.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 04/30/2010 02:49:11

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:41
Air Blank	0.000	02:41
Subject Sample #1	0.115	02:43
Breath Volume	1.230	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample #2	0.107	02:47
Breath Volume	1.273	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time 05/21/2010 01:53

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CROSS, M, S

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: CROSS, M

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 04/28/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 01:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/30/2010 03:16:55

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample #1	0.175	03:12
Breath Volume	2.546	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample #2	0.179	03:15
Breath Volume	2.292	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time 05/21/2010 01:53

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS, P, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGLAIS, P
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 09:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/30/2010 10:32:07

Diagnostics Check	OK	10:25
Air Blank	0.000	10:26
Control Test	0.079	10:26
Air Blank	0.000	10:26
Subject Sample #1	0.273	10:27
Breath Volume	2.621	10:27
Air Blank	0.000	10:28
Air Blank	0.000	10:30
Subject Sample #2	0.261	10:30
Breath Volume	2.835	10:30
Air Blank	0.000	10:31
Control Test	0.079	10:31
Air Blank	0.000	10:32
Diagnostics Check	OK	10:32

Data Download Date/Time 05/21/2010 01:53

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASSOLINI, CHAD, E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MASSOLINI, CHAD
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG004701
DRY GAS EXPIRY DATE: 02/16/2012
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HICKEY

DATE/TIME OF TEST: 04/08/2010 23:12:51

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:01	
Control Test	0.079	23:02	
Air Blank	0.000	23:02	
Subject Sample #1	0.128	23:05	Volume Not Met
Breath Volume	0.332	23:05	
Air Blank	0.000	23:06	
Air Blank	0.000	23:08	
Subject Sample #2	0.138	23:11	Volume Not Met
Breath Volume	0.089	23:11	
Air Blank	0.000	23:11	
Control Test	0.080	23:12	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time 04/21/2010 20:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARTMAN, JAMES, R

BREATH TEST OPERATOR AGENCY: Melbourne PD

ARRESTING OFFICER: HARTMAN, JAMES

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 03/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 22:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: CHILDRRESS

DATE/TIME OF TEST: 04/02/2010 23:28:20

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Subject Sample #1	0.141	23:23
Breath Volume	1.234	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample #2	0.134	23:26
Breath Volume	1.167	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time 04/25/2010 16:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLER, MATTHEW, G

BREATH TEST OPERATOR AGENCY: Melbourne PD

ARRESTING OFFICER: MILLER, MATTHEW

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 03/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 02:31

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 04/03/2010 03:14:39

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample #1	0.150	03:09
Breath Volume	2.062	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample #2	0.165	03:13
Breath Volume	1.882	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time 04/25/2010 16:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EDER, BRENDON, W

BREATH TEST OPERATOR AGENCY: Indialantic PD

ARRESTING OFFICER: EDER, BRENDON

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 03/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 02:08

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: YARCHAK

DATE/TIME OF TEST: 04/08/2010 02:43:41

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample #1	0.054	02:39
Breath Volume	1.996	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample #2	0.051	02:42
Breath Volume	1.863	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time 04/25/2010 16:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLER, MATTHEW, G

BREATH TEST OPERATOR AGENCY: Melbourne PD

ARRESTING OFFICER: MILLER, MATTHEW

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 03/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: DOWNEY

DATE/TIME OF TEST: 04/09/2010 01:07:26

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample #1	0.322	01:02
Breath Volume	1.207	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample #2	0.318	01:05
Breath Volume	1.878	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time 04/25/2010 16:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, B, D

BREATH TEST OPERATOR AGENCY: Melbourne PD

ARRESTING OFFICER: SCHMID, M

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 03/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 10:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/10/2010 11:03:12

Diagnostics Check	OK	10:57
Air Blank	0.000	10:57
Control Test	0.080	10:58
Air Blank	0.000	10:58
Subject Sample #1	0.000	10:59
Breath Volume	4.656	10:59
Air Blank	0.000	10:59
Air Blank	0.000	11:01
Subject Sample #2	0.000	11:01
Breath Volume	2.949	11:01
Air Blank	0.000	11:02
Control Test	0.080	11:02
Air Blank	0.000	11:03
Diagnostics Check	OK	11:03

Data Download Date/Time 04/25/2010 16:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARTMAN, JAMES, R

BREATH TEST OPERATOR AGENCY: Melbourne PD

ARRESTING OFFICER: HARTMAN, JAMES

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 03/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 23:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SALTZMAN

DATE/TIME OF TEST: 04/13/2010 00:20:05

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.081	00:15
Air Blank	0.000	00:15
Subject Sample #1	0.000	00:15
Breath Volume	3.082	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample #2	0.000	00:18
Breath Volume	2.375	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time 04/25/2010 16:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARTMAN, JAMES, R

BREATH TEST OPERATOR AGENCY: Melbourne PD

ARRESTING OFFICER: HARTMAN, JAMES

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 03/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 02:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: EVERINGTON

DATE/TIME OF TEST: 04/18/2010 03:27:12

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample #1	0.129	03:22
Breath Volume	2.453	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample #2	0.135	03:25
Breath Volume	1.597	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time 04/25/2010 16:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLER, MATTHEW, G

BREATH TEST OPERATOR AGENCY: Melbourne PD

ARRESTING OFFICER: MILLER, MATTHEW

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 03/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 21:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GUNN

DATE/TIME OF TEST: 04/18/2010 22:22:55

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:18
Subject Sample #1	0.113	22:18
Breath Volume	3.777	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:20
Subject Sample #2	0.110	22:21
Breath Volume	3.773	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time 04/25/2010 16:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER, TIMOTHY, C
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: WEBER, TIMOTHY
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 03/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 581913
DRY GAS EXPIRY DATE: 02/23/2011
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/21/2010 02:06:12

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.081	01:55	
Air Blank	0.000	01:56	
Subject Sample #1		01:59	No Sample Provided
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample #2		02:04	No Sample Provided
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time 04/25/2010 16:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EDER, BRENDON, W

BREATH TEST OPERATOR AGENCY: Indialantic PD

ARRESTING OFFICER: EDER, BRENDON

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 03/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 20:29

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: CAPODILUPO

DATE/TIME OF TEST: 04/22/2010 21:09:45

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.080	21:04
Air Blank	0.000	21:04
Subject Sample #1	0.000	21:05
Breath Volume	2.156	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample #2	0.000	21:08
Breath Volume	2.082	21:08
Air Blank	0.000	21:08
Control Test	0.082	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time 04/25/2010 16:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAGUE, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TEAGUE, MICHAEL
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SOLER-MARQUEZ

DATE/TIME OF TEST: 04/10/2010 03:52:46

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.079	03:47
Air Blank	0.000	03:47
Subject Sample #1	0.290	03:48
Breath Volume	1.511	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample #2	0.280	03:51
Breath Volume	1.566	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time 04/15/2010 16:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAGUE, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TEAGUE, MICHAEL
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ADELFO-ESCOBAR

DATE/TIME OF TEST: 04/11/2010 04:35:35

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:29
Air Blank	0.000	04:30
Subject Sample #1	0.179	04:30
Breath Volume	2.316	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample #2	0.164	04:34
Breath Volume	1.304	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time 04/15/2010 16:49

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS, RONALD, G
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ADAMS, RON
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: HAMRICK

DATE/TIME OF TEST: 04/13/2010 12:10:10

Diagnostics Check	OK	12:03
Air Blank	0.000	12:03
Control Test	0.080	12:04
Air Blank	0.000	12:04
Subject Sample #1	0.154	12:05
Breath Volume	1.234	12:05
Air Blank	0.000	12:06
Air Blank	0.000	12:08
Subject Sample #2	0.150	12:08
Breath Volume	1.550	12:08
Air Blank	0.000	12:09
Control Test	0.080	12:09
Air Blank	0.000	12:10
Diagnostics Check	OK	12:10

Data Download Date/Time 04/15/2010 16:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS, RON, G
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: NA, NA
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 04/13/2010 12:21:07

Diagnostics Check	OK	12:15
Air Blank	0.000	12:15
Control Test	0.080	12:16
Air Blank	0.000	12:16
Subject Sample #1	0.000	12:16
Breath Volume	1.687	12:16
Air Blank	0.000	12:17
Air Blank	0.000	12:19
Subject Sample #2	0.000	12:19
Breath Volume	2.144	12:19
Air Blank	0.000	12:20
Control Test	0.078	12:20
Air Blank	0.000	12:21
Diagnostics Check	OK	12:21

Data Download Date/Time 04/15/2010 16:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAGUE, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TEAGUE, MICHAEL
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GOODE

DATE/TIME OF TEST: 04/18/2010 05:13:07

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:10	
Control Test	0.080	05:10	
Air Blank	0.000	05:11	
Subject Sample #1		05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:12	
Control Test	0.080	05:12	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time 05/16/2010 15:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS, RONALD, G
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ADAMS, RONALD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:23

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 04/26/2010 12:55:18

Diagnostics Check	OK	12:43	
Air Blank	0.000	12:44	
Control Test	0.076	12:44	
Air Blank	0.000	12:44	
Subject Sample #1	0.277	12:48	Volume Not Met
Breath Volume	0.222	12:48	
Air Blank	0.000	12:48	
Air Blank	0.000	12:50	
Subject Sample #2	0.234	12:53	Volume Not Met
Breath Volume	0.347	12:53	
Air Blank	0.000	12:54	
Control Test	0.078	12:54	
Air Blank	0.000	12:55	
Diagnostics Check	OK	12:55	

Data Download Date/Time 05/16/2010 15:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAGUE, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: PAYNE, TISHELLE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 08:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: JORGE

DATE/TIME OF TEST: 04/30/2010 10:06:55

Diagnostics Check	OK	10:00
Air Blank	0.000	10:01
Control Test	0.080	10:01
Air Blank	0.000	10:02
Subject Sample #1	0.000	10:02
Breath Volume	1.890	10:02
Air Blank	0.000	10:03
Air Blank	0.000	10:05
Subject Sample #2	0.000	10:05
Breath Volume	2.160	10:05
Air Blank	0.000	10:06
Control Test	0.081	10:06
Air Blank	0.000	10:06
Diagnostics Check	OK	10:06

Data Download Date/Time 05/16/2010 15:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COOK, JEFF, A

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: COOK, JEFF

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 03/27/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 02:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: KISTLER

DATE/TIME OF TEST: 04/03/2010 03:12:02

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample #1	0.197	03:07
Breath Volume	1.742	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample #2	0.193	03:10
Breath Volume	1.519	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:12

Data Download Date/Time 04/22/2010 15:52

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GRENIER, D, K

BREATH TEST OPERATOR AGENCY: West Melbourne PD

ARRESTING OFFICER: BRACKETT, M

ARRESTING OFFICER AGENCY: West Melbourne PD

LAST AGENCY INSPECTION: 03/27/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 01:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 04/04/2010 01:54:41

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.081	01:44	
Air Blank	0.000	01:44	
Subject Sample #1		01:47	No Sample Provided
Breath Volume	0.000	01:47	
Air Blank	0.000	01:48	
Air Blank	0.000	01:50	
Subject Sample #2		01:53	No Sample Provided
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.081	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time 04/22/2010 15:52

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GRENIER, D, K

BREATH TEST OPERATOR AGENCY: West Melbourne PD

ARRESTING OFFICER: CRONIN, S

ARRESTING OFFICER AGENCY: West Melbourne PD

LAST AGENCY INSPECTION: 03/27/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 03:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: PRUE

DATE/TIME OF TEST: 04/10/2010 03:59:28

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Subject Sample #1	0.098	03:54
Breath Volume	2.140	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample #2	0.101	03:57
Breath Volume	2.234	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time 04/22/2010 15:52

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRENIER, D, K
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CRONIN, S
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 03/27/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 827601G
DRY GAS EXPIRY DATE: 10/06/2010
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BARBOUR

DATE/TIME OF TEST: 04/12/2010 02:35:41

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:30
Subject Sample #1	0.000	02:31
Breath Volume	2.390	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample #2	0.000	02:34
Breath Volume	1.648	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time 04/22/2010 15:52

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENNY, PATRICK, M

BREATH TEST OPERATOR AGENCY: West Melbourne PD

ARRESTING OFFICER: CUBAS, N

ARRESTING OFFICER AGENCY: West Melbourne PD

LAST AGENCY INSPECTION: 03/27/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 02:07

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: CRAFT

DATE/TIME OF TEST: 04/13/2010 02:52:06

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Subject Sample #1	0.079	02:47
Breath Volume	2.242	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample #2	0.081	02:50
Breath Volume	2.312	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time 04/22/2010 15:52

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GRENIER, D, K

BREATH TEST OPERATOR AGENCY: West Melbourne PD

ARRESTING OFFICER: CRONIN, S

ARRESTING OFFICER AGENCY: West Melbourne PD

LAST AGENCY INSPECTION: 03/27/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 02:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 04/17/2010 02:58:04

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:52
Air Blank	0.000	02:52
Subject Sample #1	0.093	02:53
Breath Volume	2.894	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample #2	0.091	02:56
Breath Volume	2.898	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:58

Data Download Date/Time 04/22/2010 15:52

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GRENIER, D, K

BREATH TEST OPERATOR AGENCY: West Melbourne PD

ARRESTING OFFICER: CRONIN, S

ARRESTING OFFICER AGENCY: West Melbourne PD

LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 02:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ZAYAS-OQUENDO

DATE/TIME OF TEST: 04/25/2010 02:57:02

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Subject Sample #1	0.174	02:51
Breath Volume	1.437	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample #2	0.174	02:55
Breath Volume	2.332	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:57

Data Download Date/Time 05/22/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GRENIER, D, K

BREATH TEST OPERATOR AGENCY: West Melbourne PD

ARRESTING OFFICER: CRONIN, S

ARRESTING OFFICER AGENCY: West Melbourne PD

LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 827601G

DRY GAS EXPIRY DATE: 10/06/2010

VIOLATION CODE: DUI

ARREST TIME: 03:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 04/30/2010 04:10:25

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:04
Air Blank	0.000	04:04
Subject Sample #1	0.133	04:05
Breath Volume	1.347	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.141	04:08
Breath Volume	2.507	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time 05/22/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 675059
BREATH TEST OPERATOR: COWAN, URSULA, E	DRY GAS EXPIRY DATE: 09/25/2011
BREATH TEST OPERATOR AGENCY: Palm Bay PD	VIOLATION CODE: DUI
ARRESTING OFFICER: COWAN, URSULA	ARREST TIME: 21:12
ARRESTING OFFICER AGENCY: Palm Bay PD	GENDER: Male
LAST AGENCY INSPECTION: 03/30/2010	OBSERVATION PERIOD BEGAN: 20:50
SUBJECT'S LAST NAME: JEAN-GILLES	DATE/TIME OF TEST: 04/08/2010 21:29:17

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:18	
Control Test	0.076	21:18	
Air Blank	0.000	21:18	
Subject Sample #1	0.281	21:22	Volume Not Met
Breath Volume	0.035	21:22	
Air Blank	0.000	21:22	
Air Blank	0.000	21:24	
Subject Sample #2	0.297	21:27	Volume Not Met
Breath Volume	0.277	21:27	
Air Blank	0.000	21:28	
Control Test	0.075	21:28	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

Data Download Date/Time	05/01/2010	03:12
-------------------------	------------	-------

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COWAN, URSULA, E

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: COWAN, URSULA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 21:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: JEAN-GILLES

DATE/TIME OF TEST: 04/08/2010 21:35:56

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.076	21:31	
Air Blank	0.000	21:31	
Subject Sample #1		21:34	Subject Test Refused
Breath Volume	0.652	21:34	
Air Blank	0.000	21:35	
Control Test	0.077	21:35	
Air Blank		21:35	RFI Detect
Air Blank		21:35	RFI Detect

Data Download Date/Time 05/01/2010 03:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COWAN, URSULA, E

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: COWAN, URSULA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 00:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MACKIE

DATE/TIME OF TEST: 04/13/2010 00:33:40

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample #1	0.155	00:25
Breath Volume	1.625	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:27
Subject Sample #2	0.184	00:29
Breath Volume	1.542	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample #3	0.180	00:32
Breath Volume	2.093	00:32
Air Blank	0.000	00:32
Control Test	0.078	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time 05/01/2010 03:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COWAN, URSULA, E

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: COWAN, URSULA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 21:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: LOVAS

DATE/TIME OF TEST: 04/13/2010 21:24:28

Diagnostics Check	OK	21:17
Air Blank	0.000	21:18
Control Test	0.078	21:18
Air Blank	0.000	21:19
Subject Sample #1	0.164	21:19
Breath Volume	1.707	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample #2	0.164	21:22
Breath Volume	1.660	21:22
Air Blank	0.000	21:23
Control Test	0.078	21:23
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time 05/01/2010 03:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COWAN, URSULA, E

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: COWAN, URSULA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 18:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/15/2010 18:40:43

Diagnostics Check	OK	18:33
Air Blank	0.000	18:34
Control Test	0.078	18:34
Air Blank	0.000	18:34
Subject Sample #1	0.124	18:36
Breath Volume	2.019	18:36
Air Blank	0.000	18:36
Air Blank	0.000	18:38
Subject Sample #2	0.126	18:39
Breath Volume	1.726	18:39
Air Blank	0.000	18:39
Control Test	0.079	18:40
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time 05/01/2010 03:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COWAN, URSULA, E

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: COWAN, URSULA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 04:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 04/21/2010 05:05:57

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	04:59
Subject Sample #1	0.215	05:00
Breath Volume	1.558	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample #2	0.220	05:04
Breath Volume	1.828	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time 05/01/2010 03:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COWAN, URSULA, E

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: COWAN, URSULA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 01:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 04/26/2010 01:40:14

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Subject Sample #1	0.140	01:34	
Breath Volume	1.187	01:34	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample #2	0.142	01:38	
Breath Volume	1.734	01:38	
Air Blank	0.000	01:39	
Control Test	0.074	01:39	RFI Detect
Air Blank	0.000	01:40	

Data Download Date/Time 05/01/2010 03:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRENT, DOUGLAS, A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HERNANDEZ, MANNY
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG004701
DRY GAS EXPIRY DATE: 02/16/2012
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 04/01/2010 02:34:33

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.217	02:29
Breath Volume	1.781	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample #2	0.216	02:32
Breath Volume	2.718	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time 04/21/2010 20:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES, JASON, C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MAZZONE, MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG004701
DRY GAS EXPIRY DATE: 02/16/2012
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SLOTNIK

DATE/TIME OF TEST: 04/19/2010 00:14:09

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:04	
Air Blank	0.000	00:04	
Subject Sample #1		00:07	No Sample Provided
Breath Volume	0.000	00:07	
Air Blank	0.000	00:08	
Air Blank	0.000	00:10	
Subject Sample #2		00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:13	
Control Test	0.080	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time 04/21/2010 20:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZI, SHANNON, R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA, JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 04/02/2010 04:49:21

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.082	04:44
Air Blank	0.000	04:44
Subject Sample #1	0.000	04:45
Breath Volume	2.664	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample #2	0.000	04:48
Breath Volume	3.320	04:48
Air Blank	0.000	04:48
Control Test	0.083	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time 04/20/2010 11:58

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS, OFC, B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SIMPKINS, B
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 04/08/2010 22:22:24

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.082	22:16
Air Blank	0.000	22:16
Subject Sample #1	0.215	22:17
Breath Volume	3.894	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample #2	0.202	22:20
Breath Volume	3.394	22:20
Air Blank	0.000	22:21
Control Test	0.082	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time 04/20/2010 11:58

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS, OFC, B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA, J
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: ACUNA

DATE/TIME OF TEST: 04/11/2010 01:06:06

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.082	01:03	
Air Blank	0.000	01:03	
Subject Sample #1		01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.083	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time 04/20/2010 11:58

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS, OFC, B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA, J
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: ACUNA

DATE/TIME OF TEST: 04/11/2010 01:21:11

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.082	01:15
Air Blank	0.000	01:16
Subject Sample #1	0.152	01:16
Breath Volume	2.898	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample #2	0.165	01:19
Breath Volume	3.046	01:19
Air Blank	0.000	01:20
Control Test	0.084	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time 04/20/2010 11:58

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCHE, MELINDA,
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BELMONTE, M
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/13/2010 04:31:27

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.082	04:28	
Air Blank	0.000	04:29	
Subject Sample #1		04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:30	
Control Test	0.082	04:30	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time 04/20/2010 11:58

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS, OFC, B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JACOBSEN, S
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: SCHOONOVER

DATE/TIME OF TEST: 04/16/2010 03:05:04

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.082	03:02	
Air Blank	0.000	03:03	
Subject Sample #1		03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.083	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:05	

Data Download Date/Time 04/20/2010 11:58

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFLENKO, KENNETH, J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: AFLENKO, KENNETH
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: ISAAC

DATE/TIME OF TEST: 04/17/2010 08:33:36

Diagnostics Check	OK	08:18	
Air Blank	0.000	08:18	
Control Test	0.083	08:19	
Air Blank	0.000	08:19	
Subject Sample #1	0.212	08:21	No .020 agreement
Breath Volume	1.113	08:21	
Air Blank	0.000	08:22	
Air Blank	0.000	08:24	
Subject Sample #2	0.186	08:26	No .020 agreement
Breath Volume	1.195	08:26	
Air Blank	0.000	08:27	
Air Blank	0.000	08:28	
Subject Sample #3	0.191	08:32	Volume Not Met
Breath Volume	0.332	08:32	
Air Blank	0.000	08:32	
Control Test	0.082	08:33	
Air Blank	0.000	08:33	
Diagnostics Check	OK	08:33	

Data Download Date/Time 04/20/2010 11:58

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIACENZA, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: PIACENZA, M
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 04/19/2010 18:17:25

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.082	18:12
Air Blank	0.000	18:12
Subject Sample #1	0.000	18:13
Breath Volume	4.097	18:13
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample #2	0.000	18:16
Breath Volume	3.382	18:16
Air Blank	0.000	18:16
Control Test	0.082	18:16
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time 04/20/2010 11:58

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER, GERALD, D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JACOBSEN, S
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/22/2010 03:40:29

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.083	03:34
Air Blank	0.000	03:35
Subject Sample #1	0.000	03:36
Breath Volume	4.292	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample #2	0.000	03:39
Breath Volume	4.152	03:39
Air Blank	0.000	03:39
Control Test	0.083	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, R, W
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: MCKENZIE, R
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 09:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/23/2010 10:30:17

Diagnostics Check	OK	10:24
Air Blank	0.000	10:24
Control Test	0.082	10:24
Air Blank	0.000	10:25
Subject Sample #1	0.000	10:25
Breath Volume	3.773	10:25
Air Blank	0.000	10:26
Air Blank	0.000	10:28
Subject Sample #2	0.000	10:28
Breath Volume	2.667	10:28
Air Blank	0.000	10:29
Control Test	0.083	10:29
Air Blank	0.000	10:30
Diagnostics Check	OK	10:30

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLEY, CPL, S
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GALLEY, CPL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: LLANO

DATE/TIME OF TEST: 04/24/2010 02:35:33

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.082	02:33	
Air Blank	0.000	02:33	
Subject Sample #1		02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.082	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS, OFC, M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ZIMA, OFC
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: TURNIPSEED

DATE/TIME OF TEST: 04/24/2010 04:55:06

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:52	
Control Test	0.083	04:52	
Air Blank	0.000	04:53	
Subject Sample #1		04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:54	
Control Test	0.082	04:54	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTSON, A,
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BUTSON, A
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: STREBLER

DATE/TIME OF TEST: 04/25/2010 18:44:35

Diagnostics Check	OK	18:37
Air Blank	0.000	18:38
Control Test	0.083	18:38
Air Blank	0.000	18:38
Subject Sample #1	0.000	18:39
Breath Volume	4.683	18:39
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample #2	0.000	18:43
Breath Volume	5.816	18:43
Air Blank	0.000	18:43
Control Test	0.084	18:44
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS, OFC, M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JANDOLA, OFC
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 04/25/2010 23:49:14

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:41	
Control Test	0.082	23:41	
Air Blank	0.000	23:41	
Subject Sample #1		23:45	No Sample Provided
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample #2		23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:48	
Control Test	0.083	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS, OFC, M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JANDOLA, OFC
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 04/26/2010 00:01:14

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.083	23:54
Air Blank	0.000	23:55
Subject Sample #1	0.178	23:56
Breath Volume	5.578	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample #2	0.164	23:59
Breath Volume	4.675	23:59
Air Blank	0.000	00:00
Control Test	0.082	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIACENZA, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: PIACENZA, MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: BRIGHT

DATE/TIME OF TEST: 04/27/2010 20:05:12

Diagnostics Check	OK	19:59
Air Blank	0.000	19:59
Control Test	0.083	19:59
Air Blank	0.000	20:00
Subject Sample #1	0.000	20:00
Breath Volume	3.242	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample #2	0.000	20:03
Breath Volume	3.796	20:03
Air Blank	0.000	20:04
Control Test	0.083	20:04
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZI, SHANNON, R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROZZI, SHANNON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: BUENO

DATE/TIME OF TEST: 04/27/2010 23:40:22

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.082	23:34
Air Blank	0.000	23:35
Subject Sample #1	0.085	23:35
Breath Volume	4.730	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample #2	0.080	23:38
Breath Volume	3.761	23:38
Air Blank	0.000	23:39
Control Test	0.083	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLEY, CPL, S
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GALLEY, CPL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:57

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 04/29/2010 09:24:44

Diagnostics Check	OK	09:18
Air Blank	0.000	09:18
Control Test	0.083	09:19
Air Blank	0.000	09:19
Subject Sample #1	0.000	09:20
Breath Volume	3.042	09:20
Air Blank	0.000	09:20
Air Blank	0.000	09:22
Subject Sample #2	0.000	09:23
Breath Volume	2.632	09:23
Air Blank	0.000	09:23
Control Test	0.083	09:24
Air Blank	0.000	09:24
Diagnostics Check	OK	09:24

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS, OFC, B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KASIC, I
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831802I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 04/30/2010 17:47:15

Diagnostics Check	OK	17:41
Air Blank	0.000	17:41
Control Test	0.083	17:41
Air Blank	0.000	17:42
Subject Sample #1	0.000	17:42
Breath Volume	3.152	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample #2	0.000	17:45
Breath Volume	3.050	17:45
Air Blank	0.000	17:46
Control Test	0.083	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time 05/12/2010 14:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH, KEVIN, J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: KILBORN, R
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HESTON

DATE/TIME OF TEST: 04/01/2010 04:33:46

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample #1	0.187	04:29
Breath Volume	2.296	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample #2	0.173	04:32
Breath Volume	1.996	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT, JOHN, E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BERMUDEZ, PABLO
ARRESTING OFFICER AGENCY: Avon Park PD
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/03/2010 17:02:47

Diagnostics Check	OK	16:54
Air Blank	0.000	16:54
Control Test	0.081	16:55
Air Blank	0.000	16:55
Subject Sample #1	0.241	16:58
Breath Volume	1.496	16:58
Air Blank	0.000	16:59
Air Blank	0.000	17:00
Subject Sample #2	0.254	17:01
Breath Volume	1.871	17:01
Air Blank	0.000	17:01
Control Test	0.079	17:02
Air Blank	0.000	17:02
Diagnostics Check	OK	17:02

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES, KENNETH, L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: POLLARD, B
ARRESTING OFFICER AGENCY: Avon Park PD
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/04/2010 00:42:13

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample #1	0.183	00:37
Breath Volume	1.917	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample #2	0.184	00:40
Breath Volume	2.140	00:40
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORSE, MICHAEL, P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GONZALEZ, MANUEL
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: KENT

DATE/TIME OF TEST: 04/05/2010 12:12:11

Diagnostics Check	OK	12:09	
Air Blank	0.000	12:09	
Control Test	0.080	12:10	
Air Blank	0.000	12:10	
Subject Sample #1		12:10	Subject Test Refused
Breath Volume	0.000	12:10	
Air Blank	0.000	12:11	
Control Test	0.081	12:11	
Air Blank	0.000	12:12	
Diagnostics Check	OK	12:12	

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLEY, BEVERLY, A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BUEFORD, S
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 04/05/2010 23:22:39

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:20	
Control Test	0.080	23:20	
Air Blank	0.000	23:20	
Subject Sample #1		23:21	Subject Test Refused
Breath Volume	0.000	23:21	
Air Blank	0.000	23:21	
Control Test	0.080	23:22	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLEY, BEVERLY, A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BUEFORD, S
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 04/06/2010 19:00:21

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.080	18:54
Air Blank	0.000	18:55
Subject Sample #1	0.095	18:55
Breath Volume	2.648	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample #2	0.094	18:58
Breath Volume	1.835	18:58
Air Blank	0.000	18:59
Control Test	0.080	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY, JAMES,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WILLIAMS, S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:02

SUBJECT'S LAST NAME: DUFFEY

DATE/TIME OF TEST: 04/07/2010 13:29:16

Diagnostics Check	OK	13:22
Air Blank	0.000	13:23
Control Test	0.081	13:23
Air Blank	0.000	13:23
Subject Sample #1	0.094	13:24
Breath Volume	1.433	13:24
Air Blank	0.000	13:25
Air Blank	0.000	13:27
Subject Sample #2	0.089	13:27
Breath Volume	1.125	13:27
Air Blank	0.000	13:28
Control Test	0.081	13:28
Air Blank	0.000	13:29
Diagnostics Check	OK	13:29

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY, JAMES,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GONZALEZ, SANDRA
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: COURT ORD
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 04/07/2010 15:32:04

Diagnostics Check	OK	15:26
Air Blank	0.000	15:26
Control Test	0.080	15:27
Air Blank	0.000	15:27
Subject Sample #1	0.000	15:27
Breath Volume	3.281	15:27
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample #2	0.000	15:30
Breath Volume	2.960	15:30
Air Blank	0.000	15:31
Control Test	0.081	15:31
Air Blank	0.000	15:31
Diagnostics Check	OK	15:32

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLIAM, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARCUS, M
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 04/10/2010 04:21:01

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:18	
Control Test	0.080	04:18	
Air Blank	0.000	04:19	
Subject Sample #1		04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Control Test	0.080	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:21	

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLIAM, M, J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARKHAM, A
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 04/11/2010 03:48:35

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample #1	0.193	03:44
Breath Volume	1.480	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample #2	0.185	03:47
Breath Volume	1.976	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLIAM, M, J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON, J
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: LAYPORT

DATE/TIME OF TEST: 04/16/2010 02:21:29

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:15
Air Blank	0.000	02:15
Subject Sample #1	0.120	02:16
Breath Volume	1.800	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.117	02:20
Breath Volume	1.835	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLEY, BEVERLY, A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FIORITO, J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: CRUICKSHANK

DATE/TIME OF TEST: 04/16/2010 02:48:06

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Subject Sample #1	0.198	02:43
Breath Volume	1.382	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample #2	0.202	02:46
Breath Volume	1.949	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time 04/17/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR, ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR, ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 04/09/2010 03:05:56

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Subject Sample #1	0.148	03:00
Breath Volume	1.214	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample #2	0.158	03:04
Breath Volume	1.410	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time 05/02/2010 16:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR, ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR, ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 04/10/2010 01:17:54

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample #1	0.199	01:13
Breath Volume	1.250	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample #2	0.195	01:16
Breath Volume	1.292	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time 05/02/2010 16:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR, ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: MARTS, ADAM
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 04/10/2010 04:41:24

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.079	04:36
Air Blank	0.000	04:36
Subject Sample #1	0.094	04:37
Breath Volume	1.488	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample #2	0.091	04:39
Breath Volume	1.347	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time 05/02/2010 16:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR, ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR, ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ARISMENDI

DATE/TIME OF TEST: 04/18/2010 02:58:09

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Subject Sample #1	0.221	02:53
Breath Volume	2.121	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.211	02:56
Breath Volume	1.566	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time 05/02/2010 16:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR, ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR, ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 04/18/2010 06:11:17

Diagnostics Check	OK	06:05
Air Blank	0.000	06:05
Control Test	0.079	06:06
Air Blank	0.000	06:06
Subject Sample #1	0.114	06:06
Breath Volume	2.207	06:06
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample #2	0.114	06:09
Breath Volume	1.718	06:09
Air Blank	0.000	06:10
Control Test	0.081	06:10
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

Data Download Date/Time 05/02/2010 16:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR, ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR, ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: NINAN

DATE/TIME OF TEST: 04/20/2010 03:22:19

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample #1	0.129	03:17
Breath Volume	1.500	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample #2	0.122	03:20
Breath Volume	1.898	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time 05/02/2010 16:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIN, BRIAN, R
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: MULLIN, BRIAN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/21/2010 03:28:37

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:23
Air Blank	0.000	03:23
Subject Sample #1	0.076	03:24
Breath Volume	1.207	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample #2	0.076	03:27
Breath Volume	1.953	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time 05/02/2010 16:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELTON, FRED, H
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: CAPONE, RALPH
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/07/2010

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: TOBIS

DATE/TIME OF TEST: 04/24/2010 16:46:31

Diagnostics Check	OK	16:40
Air Blank	0.000	16:40
Control Test	0.079	16:41
Air Blank	0.000	16:41
Subject Sample #1	0.000	16:42
Breath Volume	2.558	16:42
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample #2	0.000	16:45
Breath Volume	2.277	16:45
Air Blank	0.000	16:45
Control Test	0.079	16:45
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time 05/02/2010 16:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR, ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR, ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DEBEVITS

DATE/TIME OF TEST: 04/28/2010 03:55:14

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Subject Sample #1		03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time 05/02/2010 16:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TAWFIK, H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: TOLBERT

DATE/TIME OF TEST: 04/01/2010 01:24:01

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:18
Subject Sample #1	0.161	01:19
Breath Volume	1.812	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample #2	0.180	01:22
Breath Volume	1.304	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:24

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POTTER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: ROBBINS, J N

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 23:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SCHUPOLSKY

DATE/TIME OF TEST: 04/01/2010 03:14:31

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:09
Air Blank	0.000	03:09
Subject Sample #1	0.108	03:10
Breath Volume	1.285	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample #2	0.107	03:13
Breath Volume	1.804	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POTTER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BARTILUCCI, D E

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 02:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CORNEJO

DATE/TIME OF TEST: 04/02/2010 03:54:29

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Subject Sample #1	0.135	03:49
Breath Volume	1.128	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample #2	0.142	03:53
Breath Volume	1.648	03:53
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EVERETT, C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: DEL CAPRIO

DATE/TIME OF TEST: 04/04/2010 01:49:03

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:47	
Subject Sample #1		01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:48	
Control Test	0.079	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:49	

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POTTER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HOOVER, JOHN

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 01:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/04/2010 03:07:30

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Subject Sample #1	0.060	03:03
Breath Volume	3.714	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample #2	0.059	03:06
Breath Volume	3.789	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EVERETT, C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BALICKI

DATE/TIME OF TEST: 04/06/2010 01:40:54

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.079	01:38	
Air Blank	0.000	01:38	
Subject Sample #1	0.170	01:39	
Breath Volume	1.394	01:39	
Air Blank		01:39	RFI Detect
Air Blank	0.000	01:40	
Control Test	0.077	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EVERETT, C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BALICKI

DATE/TIME OF TEST: 04/06/2010 01:48:25

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.078	01:43
Air Blank	0.000	01:43
Subject Sample #1	0.169	01:44
Breath Volume	1.328	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample #2	0.172	01:46
Breath Volume	1.953	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time 04/15/2010 09:59

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEVIN, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 04/08/2010 01:11:18

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:08	
Control Test	0.079	01:09	
Air Blank	0.000	01:09	
Subject Sample #1		01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:10	
Control Test	0.079	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEVIN, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSEL, M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: NICOL

DATE/TIME OF TEST: 04/08/2010 05:06:01

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.080	05:03	
Air Blank	0.000	05:04	
Subject Sample #1		05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:05	
Control Test	0.078	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:06	

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRON, K O
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/08/2010 23:42:10

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample #1	0.141	23:37
Breath Volume	1.812	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample #2	0.142	23:40
Breath Volume	2.445	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BITTNER

DATE/TIME OF TEST: 04/09/2010 03:54:27

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Subject Sample #1	0.104	03:49
Breath Volume	1.304	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample #2	0.102	03:52
Breath Volume	1.308	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, KEN
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 04/09/2010 05:43:34

Diagnostics Check	OK	05:39	
Air Blank	0.000	05:40	
Control Test	0.080	05:40	
Air Blank	0.000	05:40	
Subject Sample #1		05:42	RFI Detect
Breath Volume	0.050	05:42	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, KEN
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 04/09/2010 05:49:31

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:46	
Control Test	0.078	05:47	
Air Blank	0.000	05:47	
Subject Sample #1		05:48	Subject Test Refused
Breath Volume	0.000	05:48	
Air Blank	0.000	05:48	
Control Test	0.079	05:48	
Air Blank	0.000	05:49	
Diagnostics Check	OK	05:49	

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EVERETT, C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: HAMMONDS

DATE/TIME OF TEST: 04/10/2010 01:58:23

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:53
Subject Sample #1	0.165	01:53
Breath Volume	2.425	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.172	01:56
Breath Volume	1.386	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSTON, C A
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: CANARTE

DATE/TIME OF TEST: 04/10/2010 03:35:25

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:28
Air Blank	0.000	03:28
Subject Sample #1	0.105	03:30
Breath Volume	1.148	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample #2	0.106	03:33
Breath Volume	1.167	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JACKSON, RON
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/10/2010 04:20:47

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.078	04:15
Air Blank	0.000	04:15
Subject Sample #1	0.098	04:16
Breath Volume	1.101	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample #2	0.095	04:19
Breath Volume	2.527	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY, MIKE
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: MADDOX

DATE/TIME OF TEST: 04/10/2010 04:59:36

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Subject Sample #1	0.147	04:54
Breath Volume	1.164	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample #2	0.153	04:58
Breath Volume	1.699	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRON, K O
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 04/10/2010 05:41:01

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.080	05:35
Air Blank	0.000	05:35
Subject Sample #1	0.106	05:36
Breath Volume	1.949	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample #2	0.103	05:39
Breath Volume	1.453	05:39
Air Blank	0.000	05:40
Control Test	0.078	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:41

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, KEN
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: COLLETT

DATE/TIME OF TEST: 04/10/2010 23:46:21

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.079	23:44	
Air Blank	0.000	23:44	
Subject Sample #1		23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Control Test	0.079	23:45	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EVERETT, C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 04/11/2010 01:34:22

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Subject Sample #1		01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, KEN
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 04/11/2010 03:54:45

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample #1	0.170	03:50
Breath Volume	1.660	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample #2	0.174	03:53
Breath Volume	1.929	03:53
Air Blank	0.000	03:53
Control Test	0.079	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POTTER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BARTILUCCI, D E

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 22:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: KOVALENKO

DATE/TIME OF TEST: 04/12/2010 00:07:45

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.079	00:05	
Air Blank	0.000	00:06	
Subject Sample #1		00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Control Test	0.080	00:07	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER, JOHN
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: TOMKINS

DATE/TIME OF TEST: 04/12/2010 03:14:47

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:13	
Subject Sample #1		03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EVERETT, C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: EHRISMAN

DATE/TIME OF TEST: 04/13/2010 00:01:10

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Subject Sample #1	0.000	23:56
Breath Volume	2.511	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample #2	0.000	23:59
Breath Volume	2.593	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WOOLDRIDGE, D C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: NOWAK

DATE/TIME OF TEST: 04/13/2010 04:27:02

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:18
Subject Sample #1	0.144	04:19
Breath Volume	4.175	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample #2	0.118	04:22
Breath Volume	2.355	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample #3	0.135	04:25
Breath Volume	3.468	04:25
Air Blank	0.000	04:26
Control Test	0.078	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:27

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE, I J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 04/14/2010 03:48:09

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Subject Sample #1	0.065	03:43
Breath Volume	1.144	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample #2	0.063	03:46
Breath Volume	1.933	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HORTON, L, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: MASON, C

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 05:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: AKAM

DATE/TIME OF TEST: 04/14/2010 07:25:39

Diagnostics Check	OK	07:18
Air Blank	0.000	07:19
Control Test	0.079	07:19
Air Blank	0.000	07:20
Subject Sample #1	0.162	07:20
Breath Volume	2.542	07:20
Air Blank	0.000	07:21
Air Blank	0.000	07:23
Subject Sample #2	0.166	07:24
Breath Volume	2.851	07:24
Air Blank	0.000	07:24
Control Test	0.078	07:25
Air Blank	0.000	07:25
Diagnostics Check	OK	07:25

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ASHMORE, R W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 04/14/2010 23:17:14

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample #1	0.282	23:12
Breath Volume	2.031	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample #2	0.279	23:15
Breath Volume	1.640	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RIVERA, A M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BOICE

DATE/TIME OF TEST: 04/15/2010 01:59:54

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.079	01:54
Air Blank	0.000	01:54
Subject Sample #1	0.179	01:55
Breath Volume	1.566	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample #2	0.181	01:58
Breath Volume	2.734	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POTTER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HAMLIN, DAN

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 02:35

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 04/15/2010 04:19:34

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:14
Air Blank	0.000	04:14
Subject Sample #1	0.080	04:15
Breath Volume	1.921	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.078	04:18
Breath Volume	2.195	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time 04/15/2010 09:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEVIN, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/15/2010 23:30:02

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.080	23:27	
Air Blank	0.000	23:28	
Subject Sample #1		23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:29	
Control Test	0.079	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:30	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEVIN, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ODER, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: WILBOURN

DATE/TIME OF TEST: 04/16/2010 03:19:58

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample #1	0.117	03:15
Breath Volume	1.632	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample #2	0.117	03:18
Breath Volume	1.417	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEVIN, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, K
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: ERNDL

DATE/TIME OF TEST: 04/16/2010 04:40:46

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample #1	0.137	04:36
Breath Volume	2.140	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample #2	0.131	04:39
Breath Volume	1.890	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: PARIS, J

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 04:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:09

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 04/16/2010 06:35:53

Diagnostics Check	OK	06:29
Air Blank	0.000	06:29
Control Test	0.079	06:30
Air Blank	0.000	06:30
Subject Sample #1	0.076	06:31
Breath Volume	2.406	06:31
Air Blank	0.000	06:32
Air Blank	0.000	06:34
Subject Sample #2	0.075	06:34
Breath Volume	1.566	06:34
Air Blank	0.000	06:34
Control Test	0.080	06:35
Air Blank	0.000	06:35
Diagnostics Check	OK	06:35

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: WALDREP, B

ARRESTING OFFICER AGENCY: Neptune Beach PD

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 23:13

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: INSALACO

DATE/TIME OF TEST: 04/17/2010 01:06:37

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Subject Sample #1	0.135	01:02
Breath Volume	2.117	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample #2	0.133	01:05
Breath Volume	2.371	01:05
Air Blank	0.000	01:05
Control Test	0.079	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEVIN, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO, E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ALVIN

DATE/TIME OF TEST: 04/17/2010 01:52:37

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample #1	0.141	01:48
Breath Volume	3.828	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample #2	0.151	01:51
Breath Volume	2.320	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: MOSLEY, M

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 00:54

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: DEAL

DATE/TIME OF TEST: 04/17/2010 03:04:39

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Subject Sample #1	0.000	03:00
Breath Volume	1.871	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample #2	0.000	03:03
Breath Volume	2.109	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: VALERIO, E

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 21:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: LANGLEY

DATE/TIME OF TEST: 04/19/2010 23:51:49

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample #1	0.009	23:47
Breath Volume	2.828	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample #2	0.010	23:50
Breath Volume	1.843	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEVIN, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHRISTMAS, F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: CARBALLOSA

DATE/TIME OF TEST: 04/20/2010 01:19:46

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:14
Subject Sample #1	0.097	01:15
Breath Volume	3.726	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample #2	0.095	01:18
Breath Volume	3.515	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRY, TYNA, M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSTON, C. A.
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CROWDER

DATE/TIME OF TEST: 04/20/2010 03:54:33

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.079	03:52	
Air Blank	0.000	03:53	
Subject Sample #1		03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.078	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEVIN, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ODER, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/20/2010 04:48:41

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:43	
Control Test	0.079	04:43	
Air Blank	0.000	04:44	
Subject Sample #1	0.000	04:44	
Breath Volume	2.105	04:44	
Air Blank	0.000	04:44	
Air Blank	0.000	04:46	
Subject Sample #2	0.000	04:47	
Breath Volume	1.691	04:47	
Air Blank	0.000	04:47	
Control Test	0.079	04:48	
Air Blank		04:48	RFI Detect
Air Blank	0.000	04:48	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: ODER, C

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 03:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/20/2010 04:56:14

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.078	04:51
Air Blank	0.000	04:51
Subject Sample #1	0.000	04:52
Breath Volume	2.351	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample #2	0.000	04:54
Breath Volume	1.832	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEVIN, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WOOLDRIDGE, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BATHALTER

DATE/TIME OF TEST: 04/23/2010 02:32:24

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Subject Sample #1	0.179	02:27
Breath Volume	2.531	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample #2	0.171	02:30
Breath Volume	2.742	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, KEN
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/25/2010 00:02:46

Diagnostics Check	OK	23:59	
Air Blank	0.000	00:00	
Control Test	0.080	00:00	
Air Blank	0.000	00:01	
Subject Sample #1		00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Control Test	0.079	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRON, K O
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/25/2010 00:37:50

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Subject Sample #1		00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:36	
Control Test	0.078	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS, S N
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 04/25/2010 02:38:31

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:33
Air Blank	0.000	02:33
Subject Sample #1	0.000	02:34
Breath Volume	2.941	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample #2	0.000	02:37
Breath Volume	2.687	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 04/25/2010 03:45:21

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.079	03:43	
Air Blank	0.000	03:43	
Subject Sample #1		03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.079	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORGERT, MATTHEW
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HENDRYX

DATE/TIME OF TEST: 04/25/2010 04:24:27

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:22	
Control Test	0.080	04:22	
Air Blank	0.000	04:22	
Subject Sample #1		04:23	Subject Test Refused
Breath Volume	0.000	04:23	
Air Blank	0.000	04:23	
Control Test	0.079	04:23	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSEL, M J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: GAMMON

DATE/TIME OF TEST: 04/25/2010 05:05:10

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:02	
Control Test	0.078	05:02	
Air Blank	0.000	05:03	
Subject Sample #1		05:03	Subject Test Refused
Breath Volume	0.000	05:03	
Air Blank	0.000	05:04	
Control Test	0.080	05:04	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO, J L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: ARAGON

DATE/TIME OF TEST: 04/25/2010 05:54:32

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:52	
Control Test	0.079	05:52	
Air Blank	0.000	05:52	
Subject Sample #1		05:53	Subject Test Refused
Breath Volume	0.000	05:53	
Air Blank	0.000	05:53	
Control Test	0.080	05:53	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 04/25/2010 23:35:06

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Subject Sample #1	0.137	23:30
Breath Volume	1.398	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample #2	0.135	23:33
Breath Volume	1.816	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time 04/29/2010 14:39

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSTON, C A
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/26/2010 02:21:09

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample #1	0.250	02:16
Breath Volume	2.378	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.249	02:19
Breath Volume	2.515	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORGERT, MATTHEW
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/26/2010 23:37:21

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.079	23:36	
Air Blank	0.000	23:36	
Subject Sample #1		23:36	Sequence Aborted
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORGERT, MATTHEW
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/26/2010 23:58:10

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.078	23:51
Air Blank	0.000	23:52
Subject Sample #1	0.060	23:53
Breath Volume	1.441	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample #2	0.062	23:56
Breath Volume	1.750	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POTTER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: MELEAR, D B

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 01:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CORBETT

DATE/TIME OF TEST: 04/27/2010 03:09:54

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Subject Sample #1	0.169	03:05
Breath Volume	2.531	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:07
Subject Sample #2	0.169	03:08
Breath Volume	1.898	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSEL, M J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/27/2010 05:30:42

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:24
Subject Sample #1	0.048	05:26
Breath Volume	1.328	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample #2	0.050	05:29
Breath Volume	2.589	05:29
Air Blank	0.000	05:29
Control Test	0.079	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BORGERT, M

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: CATO

DATE/TIME OF TEST: 04/29/2010 02:15:24

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample #1	0.149	02:11
Breath Volume	1.996	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample #2	0.142	02:13
Breath Volume	1.894	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEVIN, W

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: JACKSON, R

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 02:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: FRAZEE

DATE/TIME OF TEST: 04/29/2010 03:40:03

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.080	03:37	
Air Blank	0.000	03:38	
Subject Sample #1		03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Control Test	0.079	03:39	
Air Blank		03:39	RFI Detect
Air Blank	0.000	03:40	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEVIN, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JACKSON, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: FRAZEE

DATE/TIME OF TEST: 04/29/2010 03:47:25

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.079	03:45	
Air Blank	0.000	03:45	
Subject Sample #1		03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time 04/29/2010 14:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RIVERA, A M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 04/30/2010 01:08:02

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:06	
Subject Sample #1		01:06	Subject Test Refused
Breath Volume	0.000	01:06	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:08	

Data Download Date/Time 05/13/2010 13:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCHMIDT, D J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 04/30/2010 04:56:58

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:47
Subject Sample #1	0.159	04:50
Breath Volume	1.648	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:52
Subject Sample #2	0.145	04:55
Breath Volume	1.574	04:55
Air Blank	0.000	04:56
Control Test	0.078	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time 05/13/2010 13:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, C, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GARRISON, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823901G
DRY GAS EXPIRY DATE: 08/29/2010
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: HUTCHESON

DATE/TIME OF TEST: 04/01/2010 19:39:04

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.078	19:33
Air Blank	0.000	19:34
Subject Sample #1	0.143	19:34
Breath Volume	5.367	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample #2	0.143	19:37
Breath Volume	5.476	19:37
Air Blank	0.000	19:38
Control Test	0.078	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:39

Data Download Date/Time 04/15/2010 03:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, C, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GLENN, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823901G
DRY GAS EXPIRY DATE: 08/29/2010
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/02/2010 03:08:46

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:03
Subject Sample #1	0.126	03:04
Breath Volume	1.597	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample #2	0.122	03:07
Breath Volume	1.882	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time 04/15/2010 03:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPAULDING, JORDON, S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HURLEY, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823901G
DRY GAS EXPIRY DATE: 08/29/2010
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 04/02/2010 03:35:22

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:30
Subject Sample #1	0.147	03:30
Breath Volume	1.832	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample #2	0.134	03:33
Breath Volume	2.085	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time 04/15/2010 03:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: VITELLARO, P
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823901G
DRY GAS EXPIRY DATE: 08/29/2010
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/03/2010 09:48:48

Diagnostics Check	OK	09:41
Air Blank	0.000	09:42
Control Test	0.078	09:42
Air Blank	0.000	09:43
Subject Sample #1	0.194	09:44
Breath Volume	3.570	09:44
Air Blank	0.000	09:44
Air Blank	0.000	09:46
Subject Sample #2	0.198	09:47
Breath Volume	3.304	09:47
Air Blank	0.000	09:47
Control Test	0.078	09:48
Air Blank	0.000	09:48
Diagnostics Check	OK	09:48

Data Download Date/Time 04/15/2010 03:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUEFLOHN, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MCDOWELL, M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823901G
DRY GAS EXPIRY DATE: 08/29/2010
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 04/09/2010 02:08:32

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:03
Air Blank	0.000	02:03
Subject Sample #1	0.000	02:04
Breath Volume	1.937	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample #2	0.000	02:07
Breath Volume	1.847	02:07
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time 04/15/2010 03:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT, JAMES, L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON, M
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823901G
DRY GAS EXPIRY DATE: 08/29/2010
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HERRIN

DATE/TIME OF TEST: 04/10/2010 02:37:52

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:33
Subject Sample #1	0.000	02:33
Breath Volume	3.695	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample #2	0.000	02:36
Breath Volume	4.023	02:36
Air Blank	0.000	02:36
Control Test	0.078	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time 04/15/2010 03:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT, JAMES, L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GARRISON, G
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823901G
DRY GAS EXPIRY DATE: 08/29/2010
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 04/15/2010 04:17:21

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:11
Subject Sample #1	0.133	04:12
Breath Volume	2.460	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample #2	0.122	04:15
Breath Volume	1.296	04:15
Air Blank	0.000	04:16
Control Test	0.078	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time 05/18/2010 14:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, C, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GARRISON, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823901G
DRY GAS EXPIRY DATE: 08/29/2010
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: HENLEY

DATE/TIME OF TEST: 04/19/2010 19:08:17

Diagnostics Check	OK	18:59	
Air Blank	0.000	19:00	
Control Test	0.079	19:00	
Air Blank	0.000	19:01	
Subject Sample #1		19:04	No Sample Provided
Breath Volume	0.000	19:04	
Air Blank	0.000	19:04	
Air Blank	0.000	19:06	
Subject Sample #2		19:06	Subject Test Refused
Breath Volume	0.000	19:06	
Air Blank	0.000	19:07	
Control Test	0.078	19:07	
Air Blank	0.000	19:08	
Diagnostics Check	OK	19:08	

Data Download Date/Time 05/18/2010 14:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, C, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FRAZHO, T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823901G
DRY GAS EXPIRY DATE: 08/29/2010
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FLANNAGAN

DATE/TIME OF TEST: 04/24/2010 04:04:49

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Subject Sample #1	0.162	03:54
Breath Volume	1.355	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample #2	0.141	03:57
Breath Volume	1.230	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample #3	0.139	04:03
Breath Volume	1.910	04:03
Air Blank	0.000	04:03
Control Test	0.078	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time 05/18/2010 14:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NANZ, PATRICIA, M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TINDALL, ANDREA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HARKINS

DATE/TIME OF TEST: 04/01/2010 00:32:34

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.080	00:20	
Air Blank	0.000	00:20	
Subject Sample #1	0.104	00:22	
Breath Volume	1.347	00:22	
Air Blank	0.000	00:23	
Air Blank	0.000	00:24	
Subject Sample #2	0.093	00:28	Volume Not Met
Breath Volume	0.906	00:28	
Air Blank	0.000	00:28	
Air Blank	0.000	00:30	
Subject Sample #3	0.106	00:31	
Breath Volume	1.523	00:31	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/01/2010 04:30:00

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.081	04:24
Air Blank	0.000	04:24
Subject Sample #1	0.223	04:25
Breath Volume	1.562	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample #2	0.230	04:28
Breath Volume	1.277	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:30

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ANDRE

DATE/TIME OF TEST: 04/02/2010 00:58:02

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.081	00:52
Air Blank	0.000	00:52
Subject Sample #1	0.109	00:53
Breath Volume	1.960	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample #2	0.112	00:56
Breath Volume	1.945	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:58

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DICKS

DATE/TIME OF TEST: 04/02/2010 04:02:03

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Subject Sample #1	0.156	03:54	Volume Not Met
Breath Volume	0.171	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample #2	0.164	03:57	
Breath Volume	1.285	03:57	
Air Blank	0.000	03:58	
Air Blank	0.000	04:00	
Subject Sample #3	0.174	04:00	
Breath Volume	1.636	04:00	
Air Blank	0.000	04:01	
Control Test	0.078	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:02	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERBERT, JAMES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: COLLISON

DATE/TIME OF TEST: 04/03/2010 02:01:31

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:56
Subject Sample #1	0.000	01:57
Breath Volume	2.320	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample #2	0.000	02:00
Breath Volume	2.011	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLANDO, IRENE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERBERT, JAMES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 04/03/2010 04:57:38

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.080	04:52
Air Blank	0.000	04:52
Subject Sample #1	0.074	04:53
Breath Volume	2.468	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample #2	0.074	04:56
Breath Volume	2.519	04:56
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERMAN, LEIF
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: BLUM

DATE/TIME OF TEST: 04/03/2010 19:28:09

Diagnostics Check	OK	19:21
Air Blank	0.000	19:22
Control Test	0.081	19:22
Air Blank	0.000	19:23
Subject Sample #1	0.179	19:23
Breath Volume	2.296	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample #2	0.187	19:26
Breath Volume	2.652	19:26
Air Blank	0.000	19:27
Control Test	0.081	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER, MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/04/2010 04:43:54

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample #1	0.221	04:39
Breath Volume	3.039	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample #2	0.211	04:42
Breath Volume	2.214	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN, TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER, MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CHRISTY

DATE/TIME OF TEST: 04/05/2010 02:20:17

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Subject Sample #1	0.174	02:15
Breath Volume	1.839	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample #2	0.172	02:18
Breath Volume	1.578	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLANDO, IRENE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JOHNSON, LOCKSNY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: DIGIOVANNI

DATE/TIME OF TEST: 04/05/2010 04:45:28

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Subject Sample #1	0.152	04:41
Breath Volume	2.125	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample #2	0.156	04:44
Breath Volume	2.203	04:44
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLANDO, IRENE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HANSON, JON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/06/2010 02:10:25

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.080	01:58	
Air Blank	0.000	01:58	
Subject Sample #1	0.131	02:01	Volume Not Met
Breath Volume	0.996	02:01	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample #2	0.137	02:05	
Breath Volume	1.742	02:05	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample #3	0.132	02:08	
Breath Volume	1.523	02:08	
Air Blank	0.000	02:09	
Control Test	0.079	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN, TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JOHNSON, LOCKSLEY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CLAYTON JR

DATE/TIME OF TEST: 04/06/2010 04:45:30

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample #1	0.000	04:41
Breath Volume	2.578	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample #2	0.000	04:44
Breath Volume	1.828	04:44
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DODOHUE, T
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: HERNANDES

DATE/TIME OF TEST: 04/06/2010 20:51:51

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:44
Subject Sample #1	0.249	20:47
Breath Volume	1.500	20:47
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample #2	0.240	20:50
Breath Volume	1.902	20:50
Air Blank	0.000	20:50
Control Test	0.079	20:51
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STEEL, ERWIN
ARRESTING OFFICER AGENCY: Hillsboro Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PAEZ

DATE/TIME OF TEST: 04/07/2010 03:58:11

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Subject Sample #1	0.123	03:50	Volume Not Met
Breath Volume	0.503	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:52	
Subject Sample #2	0.143	03:53	
Breath Volume	1.738	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample #3	0.129	03:56	
Breath Volume	1.355	03:56	
Air Blank	0.000	03:57	
Control Test	0.079	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERRING, CHARLES
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 04/08/2010 04:52:57

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Subject Sample #1	0.115	04:48
Breath Volume	1.222	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:50
Subject Sample #2	0.111	04:51
Breath Volume	1.144	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERRING, CHARLES
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: TROMBLEY

DATE/TIME OF TEST: 04/08/2010 23:35:18

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Subject Sample #1	0.097	23:31
Breath Volume	1.875	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample #2	0.092	23:33
Breath Volume	1.582	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SMITH, AARON
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: RADHAY

DATE/TIME OF TEST: 04/09/2010 03:57:01

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Subject Sample #1	0.155	03:52
Breath Volume	1.910	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample #2	0.151	03:55
Breath Volume	1.957	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER, MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HEINRICH

DATE/TIME OF TEST: 04/10/2010 00:21:51

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:07	
Air Blank	0.000	00:08	
Subject Sample #1	0.184	00:11	Volume Not Met
Breath Volume	0.109	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample #2	0.228	00:16	
Breath Volume	1.816	00:16	
Air Blank	0.000	00:16	
Air Blank	0.000	00:18	
Subject Sample #3	0.230	00:20	
Breath Volume	1.441	00:20	
Air Blank	0.000	00:20	
Control Test	0.079	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAGER, ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GOLSTON

DATE/TIME OF TEST: 04/10/2010 00:59:46

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:53
Subject Sample #1	0.278	00:55
Breath Volume	1.660	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample #2	0.282	00:58
Breath Volume	1.476	00:58
Air Blank	0.000	00:58
Control Test	0.078	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERBERT, JAMES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: OTHER
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 04/10/2010 02:46:50

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Subject Sample #1	0.287	02:42
Breath Volume	2.679	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample #2	0.292	02:45
Breath Volume	2.738	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS, CLINT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: WERDLOW

DATE/TIME OF TEST: 04/10/2010 04:29:19

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:23
Air Blank	0.000	04:23
Subject Sample #1	0.000	04:24
Breath Volume	1.476	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample #2	0.000	04:27
Breath Volume	1.835	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAGER, ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: PETKOVIC

DATE/TIME OF TEST: 04/10/2010 04:56:28

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:54	
Control Test	0.080	04:54	
Air Blank	0.000	04:54	
Subject Sample #1		04:55	Subject Test Refused
Breath Volume	0.000	04:55	
Air Blank	0.000	04:55	
Control Test	0.080	04:55	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SMITH, AARON
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: ESTRIN

DATE/TIME OF TEST: 04/10/2010 22:05:36

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	21:59
Subject Sample #1	0.000	22:01
Breath Volume	1.886	22:01
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample #2	0.000	22:04
Breath Volume	1.949	22:04
Air Blank	0.000	22:04
Control Test	0.081	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER, MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MAIMONE

DATE/TIME OF TEST: 04/11/2010 02:50:36

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample #1	0.120	02:46
Breath Volume	3.257	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample #2	0.119	02:49
Breath Volume	2.492	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN, TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARTIN, JOSE
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CORADO-ORDONEZ

DATE/TIME OF TEST: 04/11/2010 04:56:56

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Subject Sample #1	0.232	04:52
Breath Volume	1.855	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample #2	0.242	04:55
Breath Volume	1.585	04:55
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN, TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAGER, ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: FLORIL

DATE/TIME OF TEST: 04/11/2010 06:17:42

Diagnostics Check	OK	06:14	
Air Blank	0.000	06:14	
Control Test	0.081	06:14	
Air Blank	0.000	06:15	
Subject Sample #1		06:16	Subject Test Refused
Breath Volume	0.000	06:16	
Air Blank	0.000	06:16	
Control Test	0.081	06:17	
Air Blank	0.000	06:17	
Diagnostics Check	OK	06:17	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN, MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: RICHMOND JR

DATE/TIME OF TEST: 04/11/2010 20:22:36

Diagnostics Check	OK	20:16
Air Blank	0.000	20:17
Control Test	0.080	20:17
Air Blank	0.000	20:17
Subject Sample #1	0.349	20:18
Breath Volume	1.800	20:18
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample #2	0.329	20:21
Breath Volume	1.476	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:22
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN, TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER, MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 04/11/2010 23:55:13

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Subject Sample #1	0.077	23:50
Breath Volume	1.988	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample #2	0.077	23:53
Breath Volume	2.378	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAGER, ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BRINKMANN

DATE/TIME OF TEST: 04/12/2010 00:10:56

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample #1	0.041	00:06
Breath Volume	1.183	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample #2	0.040	00:09
Breath Volume	1.109	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN, TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PATTERSON, JONATHAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ARELLANO-RODRIGUEZ

DATE/TIME OF TEST: 04/12/2010 04:11:55

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Subject Sample #1	0.207	04:07
Breath Volume	1.781	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample #2	0.204	04:10
Breath Volume	1.753	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAGER, ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: RIVERA-VILLANUEVA

DATE/TIME OF TEST: 04/12/2010 04:40:55

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample #1	0.078	04:36
Breath Volume	2.183	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample #2	0.076	04:39
Breath Volume	2.269	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLANDO, IRENE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FAHE, L
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: GAJEWSKI

DATE/TIME OF TEST: 04/12/2010 04:56:12

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.081	04:49
Air Blank	0.000	04:49
Subject Sample #1	0.296	04:51
Breath Volume	1.503	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:53
Subject Sample #2	0.296	04:54
Breath Volume	1.480	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAGER, ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 04/12/2010 23:57:56

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:53
Subject Sample #1	0.192	23:53
Breath Volume	2.027	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample #2	0.187	23:56
Breath Volume	2.207	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN, TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM, JOHN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 04/13/2010 03:25:53

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Subject Sample #1	0.168	03:14	Volume Not Met
Breath Volume	0.464	03:14	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample #2	0.214	03:19	
Breath Volume	1.281	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample #3	0.233	03:24	
Breath Volume	1.472	03:24	
Air Blank	0.000	03:24	
Control Test	0.079	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/15/2010 01:11:40

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:05
Air Blank	0.000	01:05
Subject Sample #1	0.201	01:07
Breath Volume	2.164	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample #2	0.187	01:10
Breath Volume	1.136	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STEEL, ERWIN
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: KAPPNER

DATE/TIME OF TEST: 04/15/2010 04:06:31

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:53	
Control Test	0.081	03:53	
Air Blank	0.000	03:54	
Subject Sample #1	0.208	03:57	Volume Not Met
Breath Volume	1.000	03:57	
Air Blank	0.000	03:58	
Air Blank	0.000	03:59	
Subject Sample #2	0.207	04:00	
Breath Volume	1.277	04:00	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample #3	0.203	04:04	
Breath Volume	1.453	04:04	
Air Blank	0.000	04:05	
Control Test	0.081	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARTIN, TRISHA
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 04/16/2010 00:30:07

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:21
Subject Sample #1	0.159	00:24
Breath Volume	1.343	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample #2	0.163	00:28
Breath Volume	1.953	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 04/16/2010 01:36:24

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.080	01:23	
Air Blank	0.000	01:23	
Subject Sample #1	0.193	01:25	
Breath Volume	1.343	01:25	
Air Blank	0.000	01:26	
Air Blank	0.000	01:27	
Subject Sample #2	0.155	01:31	Volume Not Met
Breath Volume	0.660	01:31	
Air Blank	0.000	01:31	
Air Blank	0.000	01:33	
Subject Sample #3	0.208	01:34	
Breath Volume	2.097	01:34	
Air Blank	0.000	01:35	
Control Test	0.079	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER, NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/16/2010 02:46:46

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:38
Air Blank	0.000	02:38
Subject Sample #1	0.118	02:40
Breath Volume	1.746	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample #2	0.115	02:45
Breath Volume	1.410	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DERBIN

DATE/TIME OF TEST: 04/16/2010 03:05:39

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample #1	0.180	02:54
Breath Volume	1.542	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.158	03:00
Breath Volume	1.882	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample #3	0.168	03:04
Breath Volume	1.296	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SLAYTON, GARY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SIERRA-CUERVO

DATE/TIME OF TEST: 04/16/2010 03:37:01

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample #1	0.184	03:32
Breath Volume	2.675	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample #2	0.188	03:35
Breath Volume	2.390	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PATTERSON, JONATHAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 04/16/2010 04:28:58

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:21
Air Blank	0.000	04:21
Subject Sample #1	0.092	04:23
Breath Volume	1.347	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.083	04:27
Breath Volume	1.316	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEDEJ, MEDJON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: DELBUONO

DATE/TIME OF TEST: 04/16/2010 21:27:22

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.081	21:22
Air Blank	0.000	21:22
Subject Sample #1	0.000	21:23
Breath Volume	1.742	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample #2	0.000	21:26
Breath Volume	1.339	21:26
Air Blank	0.000	21:26
Control Test	0.081	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SMITH, AARON
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 04/17/2010 03:44:08

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:38
Air Blank	0.000	03:38
Subject Sample #1	0.000	03:39
Breath Volume	1.171	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample #2	0.000	03:42
Breath Volume	1.687	03:42
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERRING, CHARLES
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: TORGERSEN

DATE/TIME OF TEST: 04/17/2010 22:50:11

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Subject Sample #1	0.193	22:45
Breath Volume	2.187	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample #2	0.197	22:48
Breath Volume	2.515	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS, CLINT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GIFFORD

DATE/TIME OF TEST: 04/18/2010 02:00:52

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.079	01:50	
Air Blank	0.000	01:50	
Subject Sample #1	0.076	01:53	Volume Not Met
Breath Volume	0.054	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample #2	0.132	01:59	Volume Not Met
Breath Volume	0.109	01:59	
Air Blank	0.000	01:59	
Control Test	0.079	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLANDO, IRENE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER, MIKE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SANASEE

DATE/TIME OF TEST: 04/18/2010 04:16:35

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:11
Subject Sample #1	0.270	04:12
Breath Volume	2.191	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample #2	0.272	04:15
Breath Volume	2.328	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERRING, CHARLES
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: REYES JR

DATE/TIME OF TEST: 04/18/2010 05:00:03

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.081	04:54
Air Blank	0.000	04:55
Subject Sample #1	0.275	04:55
Breath Volume	2.867	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample #2	0.276	04:58
Breath Volume	2.371	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	05:00

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARTIN, JENNIFER
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 04/20/2010 00:22:14

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Subject Sample #1	0.196	00:17
Breath Volume	2.250	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample #2	0.191	00:20
Breath Volume	2.484	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLANDO, IRENE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER, MIKE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ATLAN

DATE/TIME OF TEST: 04/20/2010 02:23:25

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Subject Sample #1	0.173	02:19
Breath Volume	1.882	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample #2	0.170	02:21
Breath Volume	1.355	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAGER, ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 04/20/2010 04:27:48

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.074	04:27	Control Outside Tolerance
Air Blank	0.000	04:27	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAGER, ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 04/20/2010 04:35:47

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.075	04:30
Air Blank	0.000	04:30
Subject Sample #1	0.250	04:31
Breath Volume	1.582	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample #2	0.254	04:34
Breath Volume	1.292	04:34
Air Blank	0.000	04:34
Control Test	0.077	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLANDO, IRENE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JOHNSON, LOCKSLEY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: KIRSCHBAUM

DATE/TIME OF TEST: 04/20/2010 05:21:19

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.080	05:16
Air Blank	0.000	05:16
Subject Sample #1	0.175	05:16
Breath Volume	3.230	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample #2	0.178	05:19
Breath Volume	3.062	05:19
Air Blank	0.000	05:20
Control Test	0.080	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: IRELAN, ANDREW
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/21/2010 23:25:19

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:20
Subject Sample #1	0.204	23:20
Breath Volume	1.855	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample #2	0.200	23:23
Breath Volume	1.484	23:23
Air Blank	0.000	23:24
Control Test	0.081	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 04/22/2010 01:43:48

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.081	01:39	
Air Blank	0.000	01:40	
Subject Sample #1		01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.081	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI, DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: RIZZITELLO

DATE/TIME OF TEST: 04/22/2010 03:36:41

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:30
Subject Sample #1	0.187	03:32
Breath Volume	1.625	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample #2	0.186	03:35
Breath Volume	1.226	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SLAYTON, GARY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/22/2010 04:32:06

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:29	
Control Test	0.081	04:30	
Air Blank	0.000	04:30	
Subject Sample #1		04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.081	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time 04/22/2010 15:37

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PARRY

DATE/TIME OF TEST: 04/23/2010 00:27:47

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:21
Subject Sample #1	0.162	00:22
Breath Volume	1.566	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample #2	0.163	00:26
Breath Volume	2.140	00:26
Air Blank	0.000	00:26
Control Test	0.082	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAPOTA, SAM
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: MONTALBANO

DATE/TIME OF TEST: 04/23/2010 05:39:36

Diagnostics Check	OK	05:33
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:34
Subject Sample #1	0.202	05:34
Breath Volume	1.269	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample #2	0.203	05:38
Breath Volume	2.207	05:38
Air Blank	0.000	05:38
Control Test	0.079	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKINGER, LANI, A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRETON, DAMARIS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: OLMEDA

DATE/TIME OF TEST: 04/23/2010 06:35:17

Diagnostics Check	OK	06:29
Air Blank	0.000	06:29
Control Test	0.080	06:29
Air Blank	0.000	06:30
Subject Sample #1	0.280	06:30
Breath Volume	1.382	06:30
Air Blank	0.000	06:31
Air Blank	0.000	06:33
Subject Sample #2	0.276	06:33
Breath Volume	1.910	06:33
Air Blank	0.000	06:34
Control Test	0.079	06:34
Air Blank	0.000	06:35
Diagnostics Check	OK	06:35

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SMITH, AARON
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 04/23/2010 19:37:55

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.081	19:32
Air Blank	0.000	19:33
Subject Sample #1	0.000	19:33
Breath Volume	1.640	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample #2	0.000	19:36
Breath Volume	2.414	19:36
Air Blank	0.000	19:37
Control Test	0.081	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER, MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 04/23/2010 22:14:38

Diagnostics Check	OK	22:07
Air Blank	0.000	22:08
Control Test	0.080	22:08
Air Blank	0.000	22:08
Subject Sample #1	0.280	22:09
Breath Volume	1.382	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:11
Subject Sample #2	0.288	22:13
Breath Volume	1.886	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:14
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NONE, NONE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: COLA

DATE/TIME OF TEST: 04/23/2010 23:09:25

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample #1	0.000	23:05
Breath Volume	1.789	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample #2	0.000	23:08
Breath Volume	1.535	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLANDO, IRENE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER, N
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CALLOWAY

DATE/TIME OF TEST: 04/24/2010 18:02:50

Diagnostics Check	OK	17:56
Air Blank	0.000	17:57
Control Test	0.080	17:57
Air Blank	0.000	17:57
Subject Sample #1	0.102	17:58
Breath Volume	1.371	17:58
Air Blank	0.000	17:59
Air Blank	0.000	18:00
Subject Sample #2	0.099	18:01
Breath Volume	1.316	18:01
Air Blank	0.000	18:01
Control Test	0.080	18:02
Air Blank	0.000	18:02
Diagnostics Check	OK	18:02

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER, MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 04/24/2010 23:16:16

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Subject Sample #1	0.185	23:11
Breath Volume	1.847	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample #2	0.168	23:14
Breath Volume	1.746	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TINDALL, ANDREA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/25/2010 02:03:47

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Subject Sample #1	0.210	01:56
Breath Volume	2.437	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample #2	0.187	01:59
Breath Volume	1.281	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample #3	0.188	02:02
Breath Volume	1.531	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLANDO, IRENE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: CAMPOSECO

DATE/TIME OF TEST: 04/25/2010 05:43:18

Diagnostics Check	OK	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:38
Subject Sample #1	0.175	05:38
Breath Volume	1.507	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample #2	0.176	05:41
Breath Volume	2.074	05:41
Air Blank	0.000	05:42
Control Test	0.081	05:42
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLANDO, IRENE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ELLIS, A
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 07:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: ACOSTA - CACERES

DATE/TIME OF TEST: 04/25/2010 08:04:05

Diagnostics Check	OK	07:56
Air Blank	0.000	07:57
Control Test	0.081	07:57
Air Blank	0.000	07:58
Subject Sample #1	0.247	07:59
Breath Volume	1.308	07:59
Air Blank	0.000	07:59
Air Blank	0.000	08:01
Subject Sample #2	0.234	08:02
Breath Volume	1.414	08:02
Air Blank	0.000	08:03
Control Test	0.080	08:03
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KLIMA, RICHARD
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: CASCIANO

DATE/TIME OF TEST: 04/27/2010 18:32:59

Diagnostics Check	OK	18:26
Air Blank	0.000	18:27
Control Test	0.081	18:27
Air Blank	0.000	18:28
Subject Sample #1	0.000	18:28
Breath Volume	1.546	18:28
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample #2	0.000	18:31
Breath Volume	1.464	18:31
Air Blank	0.000	18:32
Control Test	0.081	18:32
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: YUHAS

DATE/TIME OF TEST: 04/27/2010 23:43:13

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:37
Subject Sample #1	0.236	23:38
Breath Volume	1.144	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample #2	0.231	23:41
Breath Volume	1.269	23:41
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN, MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: OSULLIVAN

DATE/TIME OF TEST: 04/28/2010 00:38:43

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample #1	0.184	00:34
Breath Volume	1.222	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample #2	0.181	00:37
Breath Volume	1.601	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ST MARTIN, JOSEPH
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 04/28/2010 22:19:14

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.080	22:13
Air Blank	0.000	22:13
Subject Sample #1	0.190	22:14
Breath Volume	1.105	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample #2	0.196	22:17
Breath Volume	1.921	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CONLEY, BRIAN
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KOLEFF

DATE/TIME OF TEST: 04/29/2010 00:39:54

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Subject Sample #1	0.227	00:33
Breath Volume	1.457	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample #2	0.225	00:38
Breath Volume	1.457	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LUCERO

DATE/TIME OF TEST: 04/29/2010 01:42:10

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.081	01:36
Air Blank	0.000	01:36
Subject Sample #1	0.251	01:37
Breath Volume	1.496	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample #2	0.270	01:40
Breath Volume	1.890	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTERO, ANGELA, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARO, CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: WINSKI

DATE/TIME OF TEST: 04/29/2010 22:40:13

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.081	22:34
Air Blank	0.000	22:34
Subject Sample #1	0.176	22:35
Breath Volume	2.347	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample #2	0.168	22:38
Breath Volume	2.101	22:38
Air Blank	0.000	22:39
Control Test	0.081	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI, DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MORELAND

DATE/TIME OF TEST: 04/30/2010 01:12:03

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample #1	0.126	01:07
Breath Volume	1.589	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample #2	0.125	01:10
Breath Volume	1.875	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:12

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY, SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUTTERY, CRAIG
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BENDUHN

DATE/TIME OF TEST: 04/30/2010 03:01:41

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Subject Sample #1	0.182	02:57
Breath Volume	2.343	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.179	03:00
Breath Volume	2.394	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI, DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 659357
DRY GAS EXPIRY DATE: 09/02/2011
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BRECKENRIDGE

DATE/TIME OF TEST: 04/30/2010 03:44:25

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample #1	0.142	03:39
Breath Volume	1.304	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample #2	0.144	03:42
Breath Volume	1.800	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time 05/14/2010 15:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

INSTRUMENT REGISTERED TO: LAKE ALFRED PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, BRENTON, L
BREATH TEST OPERATOR AGENCY: Lake Alfred PD
ARRESTING OFFICER: TRIMBLE, BRENTON
ARRESTING OFFICER AGENCY: Lake Alfred PD
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 828901G
DRY GAS EXPIRY DATE: 10/21/2010
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: SHIRAH

DATE/TIME OF TEST: 04/11/2010 05:44:51

Diagnostics Check	OK	05:36
Air Blank	0.000	05:37
Control Test	0.078	05:37
Air Blank	0.000	05:38
Subject Sample #1	0.145	05:39
Breath Volume	4.500	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample #2	0.138	05:43
Breath Volume	3.460	05:43
Air Blank	0.000	05:43
Control Test	0.079	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time 04/18/2010 14:56

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORNDORFF, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: ORNDORFF, WILLIAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: NUTTER JR

DATE/TIME OF TEST: 04/15/2010 23:22:37

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:17
Subject Sample #1	0.184	23:18
Breath Volume	2.011	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample #2	0.188	23:21
Breath Volume	1.125	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS, LANE, A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WINTERS, LANE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MULLIKEN

DATE/TIME OF TEST: 04/16/2010 00:02:24

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Subject Sample #1	0.000	23:57
Breath Volume	3.125	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample #2	0.000	00:01
Breath Volume	2.378	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEMAN, DANIEL, C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STAVRO, JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 04/20/2010 19:46:48

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.077	19:41
Air Blank	0.000	19:41
Subject Sample #1	0.262	19:42
Breath Volume	2.207	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:44
Subject Sample #2	0.265	19:45
Breath Volume	2.101	19:45
Air Blank	0.000	19:45
Control Test	0.078	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR, ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MONSEGUR, ROBERT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SOLER

DATE/TIME OF TEST: 04/22/2010 04:12:03

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Subject Sample #1	0.000	04:07
Breath Volume	2.566	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.000	04:10
Breath Volume	2.324	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:12

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR, ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MEYER, DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: RANKIN

DATE/TIME OF TEST: 04/22/2010 05:37:17

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.080	05:31
Air Blank	0.000	05:31
Subject Sample #1	0.147	05:32
Breath Volume	1.988	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample #2	0.150	05:35
Breath Volume	2.531	05:35
Air Blank	0.000	05:36
Control Test	0.080	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, ADAM, M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORRIS, ADAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: MAISONET

DATE/TIME OF TEST: 04/24/2010 06:25:14

Diagnostics Check	OK	06:15
Air Blank	0.000	06:16
Control Test	0.080	06:16
Air Blank	0.000	06:17
Subject Sample #1	0.171	06:17
Breath Volume	2.375	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample #2	0.136	06:20
Breath Volume	1.882	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample #3	0.154	06:23
Breath Volume	2.703	06:23
Air Blank	0.000	06:24
Control Test	0.079	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MORRIS, ADAM, M

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: MORRIS, ADAM

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 22:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CHACHE

DATE/TIME OF TEST: 04/25/2010 00:00:27

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.077	23:54
Air Blank	0.000	23:54
Subject Sample #1	0.202	23:55
Breath Volume	1.269	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample #2	0.199	23:58
Breath Volume	1.710	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS, R, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EVANS, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SWINFORD

DATE/TIME OF TEST: 04/29/2010 00:52:01

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.079	00:45
Air Blank	0.000	00:45
Subject Sample #1	0.100	00:47
Breath Volume	4.062	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample #2	0.099	00:50
Breath Volume	4.000	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:52

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR, ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MONSEGUR, ROBERT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 04/29/2010 05:30:10

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:25
Subject Sample #1	0.158	05:25
Breath Volume	3.019	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample #2	0.156	05:28
Breath Volume	2.613	05:28
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEMAN, DANIEL, C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JARABEK, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 11:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:57

SUBJECT'S LAST NAME: UTERSTAEDT

DATE/TIME OF TEST: 04/29/2010 13:27:19

Diagnostics Check	OK	13:21
Air Blank	0.000	13:21
Control Test	0.080	13:22
Air Blank	0.000	13:22
Subject Sample #1	0.070	13:22
Breath Volume	1.289	13:22
Air Blank	0.000	13:23
Air Blank	0.000	13:25
Subject Sample #2	0.067	13:25
Breath Volume	1.796	13:25
Air Blank	0.000	13:26
Control Test	0.079	13:26
Air Blank	0.000	13:27
Diagnostics Check	OK	13:27

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCUNE, R, J

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: EVANS, R

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 21:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MCCUNE

DATE/TIME OF TEST: 04/30/2010 00:17:44

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample #1	0.142	00:09
Breath Volume	3.988	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample #2	0.121	00:12
Breath Volume	1.199	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample #3	0.143	00:16
Breath Volume	3.835	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS, R, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EVANS, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FLESSAS

DATE/TIME OF TEST: 04/30/2010 04:10:04

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Subject Sample #1	0.177	04:04
Breath Volume	4.070	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.166	04:08
Breath Volume	3.054	04:08
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:10

Data Download Date/Time 05/11/2010 14:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALVIN, JOHN,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: COUPE, ROBERT
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/10/2011
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: NEVES

DATE/TIME OF TEST: 04/10/2010 06:11:13

Diagnostics Check	OK	06:08	
Air Blank	0.000	06:08	
Control Test	0.078	06:09	
Air Blank	0.000	06:09	
Subject Sample #1		06:09	Subject Test Refused
Breath Volume	0.000	06:09	
Air Blank	0.000	06:10	
Control Test	0.079	06:10	
Air Blank	0.000	06:11	
Diagnostics Check	OK	06:11	

Data Download Date/Time 04/20/2010 09:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MALVIN, JOHN, W

BREATH TEST OPERATOR AGENCY: Coral Springs PD

ARRESTING OFFICER: CASSINI, ANTHONY

ARRESTING OFFICER AGENCY: Coral Springs PD

LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/10/2011

VIOLATION CODE: DUI

ARREST TIME: 16:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: NOWRAN

DATE/TIME OF TEST: 04/11/2010 17:07:42

Diagnostics Check	OK	17:00
Air Blank	0.000	17:01
Control Test	0.078	17:01
Air Blank	0.000	17:02
Subject Sample #1	0.279	17:03
Breath Volume	3.660	17:03
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample #2	0.273	17:06
Breath Volume	2.914	17:06
Air Blank	0.000	17:06
Control Test	0.078	17:07
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time 04/20/2010 09:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALVIN, JOHN, W
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: FERNANDES, DEREK
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/10/2011
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 04/17/2010 02:11:09

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.078	02:01
Air Blank	0.000	02:01
Subject Sample #1	0.188	02:02
Breath Volume	1.761	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample #2	0.214	02:06
Breath Volume	2.921	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #3	0.201	02:09
Breath Volume	2.976	02:09
Air Blank	0.000	02:10
Control Test	0.076	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time 04/20/2010 09:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALVIN, JOHN, W
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: FERNANDES, DERICK
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/10/2011
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MCCARTIN

DATE/TIME OF TEST: 04/18/2010 01:01:54

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.078	00:59	
Air Blank	0.000	01:00	
Subject Sample #1		01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Control Test	0.078	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time 04/20/2010 09:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SCHUSTER, ERIC, D

BREATH TEST OPERATOR AGENCY: Coral Springs PD

ARRESTING OFFICER: EASON, KATHARINE

ARRESTING OFFICER AGENCY: Coral Springs PD

LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/10/2011

VIOLATION CODE: DUI

ARREST TIME: 10:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: WILLNER

DATE/TIME OF TEST: 04/19/2010 11:07:57

Diagnostics Check	OK	11:02
Air Blank	0.000	11:02
Control Test	0.078	11:02
Air Blank	0.000	11:03
Subject Sample #1	0.000	11:03
Breath Volume	2.339	11:03
Air Blank	0.000	11:04
Air Blank	0.000	11:06
Subject Sample #2	0.000	11:06
Breath Volume	2.906	11:06
Air Blank	0.000	11:07
Control Test	0.078	11:07
Air Blank	0.000	11:07
Diagnostics Check	OK	11:07

Data Download Date/Time 04/20/2010 09:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISANTO, FRANK,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KUCZYNSKI, STAN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 04/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/10/2011
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: NERENBERG

DATE/TIME OF TEST: 04/22/2010 11:26:43

Diagnostics Check	OK	11:20
Air Blank	0.000	11:21
Control Test	0.078	11:21
Air Blank	0.000	11:21
Subject Sample #1	0.000	11:22
Breath Volume	2.699	11:22
Air Blank	0.000	11:22
Air Blank	0.000	11:24
Subject Sample #2	0.000	11:25
Breath Volume	1.968	11:25
Air Blank	0.000	11:25
Control Test	0.078	11:26
Air Blank	0.000	11:26
Diagnostics Check	OK	11:26

Data Download Date/Time 05/21/2010 13:47

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE, JOHN, V
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: RHODEN, GARLAND
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 04/07/2010 20:40:33

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:37	
Control Test	0.079	20:38	
Air Blank	0.000	20:38	
Subject Sample #1		20:39	Interferent Detect
Breath Volume	5.722	20:39	
Air Blank	0.000	20:39	
Control Test	0.078	20:39	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:40	

Data Download Date/Time 05/01/2010 09:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE, JOHN, V
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: RHODEN, GARLAND
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 04/07/2010 20:46:01

Diagnostics Check	OK	20:42	
Air Blank	0.000	20:43	
Control Test	0.078	20:43	
Air Blank	0.000	20:44	
Subject Sample #1		20:44	Interferent Detect
Breath Volume	4.976	20:44	
Air Blank	0.000	20:45	
Control Test	0.078	20:45	
Air Blank	0.000	20:45	
Diagnostics Check	OK	20:46	

Data Download Date/Time 05/01/2010 09:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE, JOHN, V
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: COLE, ERNIE
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 04/13/2010 21:25:36

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.078	21:13
Air Blank	0.000	21:14
Subject Sample #1	0.174	21:14
Breath Volume	3.664	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample #2	0.179	21:17
Breath Volume	3.023	21:17
Air Blank	0.000	21:18
Control Test	0.077	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:25

Data Download Date/Time 05/01/2010 09:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK, K, M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WALKER, S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: ELMORE

DATE/TIME OF TEST: 04/15/2010 21:02:45

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.078	20:57
Air Blank	0.000	20:57
Subject Sample #1	0.161	20:58
Breath Volume	3.132	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample #2	0.160	21:01
Breath Volume	3.898	21:01
Air Blank	0.000	21:01
Control Test	0.078	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time 05/01/2010 09:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE, JOHN, V
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MALLARD, JARED
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 04/17/2010 03:48:46

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.079	03:46	
Air Blank	0.000	03:46	
Subject Sample #1		03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.079	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time 05/01/2010 09:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE, JOHN, V
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LUCAS, CHARLIE
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/18/2010 01:21:28

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:19	
Control Test	0.079	01:19	
Air Blank	0.000	01:19	
Subject Sample #1		01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:20	
Control Test	0.079	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time 05/01/2010 09:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK, K, M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CARTER, JD
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/25/2010 02:52:24

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.078	02:47
Air Blank	0.000	02:47
Subject Sample #1	0.164	02:48
Breath Volume	1.898	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample #2	0.160	02:50
Breath Volume	2.242	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time 05/01/2010 09:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NYE, YERRY, L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BLACKWELL, B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: LEMKE

DATE/TIME OF TEST: 04/26/2010 14:40:18

Diagnostics Check	OK	14:36	
Air Blank	0.000	14:37	
Control Test	0.078	14:37	
Air Blank	0.000	14:38	
Subject Sample #1		14:38	Subject Test Refused
Breath Volume	0.000	14:38	
Air Blank	0.000	14:39	
Control Test	0.078	14:39	
Air Blank	0.000	14:40	
Diagnostics Check	OK	14:40	

Data Download Date/Time 05/01/2010 09:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH, KEVIN, J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON, J
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 04/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 04/20/2010 23:47:09

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.077	23:40
Air Blank	0.000	23:40
Subject Sample #1	0.052	23:42
Breath Volume	2.253	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample #2	0.051	23:45
Breath Volume	1.714	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time 05/15/2010 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEET, DAVID, N
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON, JOHN
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 04/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 04/24/2010 02:23:56

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.078	02:18
Air Blank	0.000	02:18
Subject Sample #1	0.130	02:19
Breath Volume	2.300	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample #2	0.137	02:22
Breath Volume	2.636	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time 05/15/2010 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUKKEN, TOVAH, N
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WILLIAMS, D
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: GOPHER

DATE/TIME OF TEST: 04/26/2010 23:37:31

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:35	
Control Test	0.078	23:35	
Air Blank	0.000	23:35	
Subject Sample #1		23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:36	
Control Test	0.077	23:36	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time 05/15/2010 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEET, DAVID, N
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON, JOHN
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 04/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GALAN

DATE/TIME OF TEST: 04/30/2010 03:11:31

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.078	03:05
Air Blank	0.000	03:05
Subject Sample #1	0.160	03:06
Breath Volume	1.187	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample #2	0.153	03:09
Breath Volume	1.832	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time 05/15/2010 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 03:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 04/01/2010 03:51:15

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.083	03:46
Air Blank	0.000	03:46
Subject Sample #1	0.149	03:46
Breath Volume	1.632	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample #2	0.146	03:49
Breath Volume	3.246	03:49
Air Blank	0.000	03:50
Control Test	0.083	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 02:21

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: WASCHEK

DATE/TIME OF TEST: 04/02/2010 03:09:46

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:04
Subject Sample #1	0.162	03:05
Breath Volume	2.875	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample #2	0.164	03:08
Breath Volume	1.382	03:08
Air Blank	0.000	03:08
Control Test	0.082	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: QUACKENBUSH, BRET, C

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: QUACKENBUSH, BRET

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 02:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: REKO

DATE/TIME OF TEST: 04/03/2010 03:27:06

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Subject Sample #1	0.253	03:22
Breath Volume	1.453	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample #2	0.254	03:25
Breath Volume	2.187	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 02:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RIDENHOWER

DATE/TIME OF TEST: 04/06/2010 03:13:43

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.083	03:02	
Air Blank	0.000	03:02	
Subject Sample #1		03:05	No Sample Provided
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample #2	0.222	03:09	
Breath Volume	2.585	03:09	
Air Blank	0.000	03:09	
Air Blank	0.000	03:11	
Subject Sample #3	0.213	03:12	
Breath Volume	2.371	03:12	
Air Blank	0.000	03:12	
Control Test	0.081	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 20:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/06/2010 21:06:29

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:04	
Control Test	0.083	21:04	
Air Blank	0.000	21:04	
Subject Sample #1		21:05	Subject Test Refused
Breath Volume	0.000	21:05	
Air Blank	0.000	21:05	
Control Test	0.082	21:05	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 22:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 04/06/2010 23:27:39

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.083	23:16	
Air Blank	0.000	23:17	
Subject Sample #1	0.199	23:20	Volume Not Met
Breath Volume	0.128	23:20	
Air Blank	0.000	23:21	
Air Blank	0.000	23:23	
Subject Sample #2	0.000	23:26	Volume Not Met
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Control Test	0.081	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 22:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 04/07/2010 01:04:16

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.083	00:56	
Air Blank	0.000	00:57	
Subject Sample #1	0.168	01:00	Volume Not Met
Breath Volume	0.214	01:00	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample #2		01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:03	
Control Test	0.082	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LLOYD, RAYMOND, W

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: LLOYD, RAYMOND

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 16:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: WHORTON

DATE/TIME OF TEST: 04/07/2010 16:46:26

Diagnostics Check	OK	16:40
Air Blank	0.000	16:40
Control Test	0.081	16:40
Air Blank	0.000	16:41
Subject Sample #1	0.234	16:41
Breath Volume	2.394	16:41
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample #2	0.232	16:44
Breath Volume	1.574	16:44
Air Blank	0.000	16:45
Control Test	0.082	16:45
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 23:21

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WILLCOCKSON

DATE/TIME OF TEST: 04/09/2010 23:56:52

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.082	23:51
Air Blank	0.000	23:52
Subject Sample #1	0.222	23:52
Breath Volume	2.214	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample #2	0.216	23:55
Breath Volume	2.140	23:55
Air Blank	0.000	23:55
Control Test	0.082	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 01:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: CONNOR

DATE/TIME OF TEST: 04/10/2010 02:14:54

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.083	02:12	
Air Blank	0.000	02:12	
Subject Sample #1		02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.083	02:14	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 02:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VALDEZ-RODRIGUEZ

DATE/TIME OF TEST: 04/10/2010 03:17:04

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.083	03:11
Air Blank	0.000	03:11
Subject Sample #1	0.189	03:12
Breath Volume	1.800	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample #2	0.181	03:15
Breath Volume	1.937	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:17

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 03:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CRUZ-LANDIN

DATE/TIME OF TEST: 04/10/2010 04:02:03

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.082	03:56
Air Blank	0.000	03:57
Subject Sample #1	0.130	03:57
Breath Volume	1.824	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample #2	0.130	04:00
Breath Volume	2.363	04:00
Air Blank	0.000	04:01
Control Test	0.083	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:02

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 01:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/11/2010 01:48:51

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.083	01:46	
Air Blank	0.000	01:47	
Subject Sample #1		01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.083	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FITZGERALD, F, F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FITZGERALD, F

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 02:01

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TEEHAN

DATE/TIME OF TEST: 04/16/2010 02:53:56

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.083	02:46	
Air Blank	0.000	02:46	
Subject Sample #1		02:49	No Sample Provided
Breath Volume	0.000	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample #2		02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.084	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: QUACKENBUSH, BRET, C

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: QUACKENBUSH, BRET

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 23:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: GLASS

DATE/TIME OF TEST: 04/18/2010 00:05:41

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.083	00:00
Air Blank	0.000	00:00
Subject Sample #1	0.216	00:01
Breath Volume	2.972	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample #2	0.218	00:04
Breath Volume	2.875	00:04
Air Blank	0.000	00:04
Control Test	0.083	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time 04/18/2010 08:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FITZGERALD, F, F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FITZGERALD, F

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 04/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 02:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: RICKS

DATE/TIME OF TEST: 04/22/2010 03:32:13

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.083	03:28	
Air Blank	0.000	03:29	
Subject Sample #1		03:30	Subject Test Refused
Breath Volume	0.480	03:30	
Air Blank	0.000	03:31	
Control Test	0.082	03:31	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time 05/02/2010 09:19

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FITZGERALD, F, F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FITZGERALD, F

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 04/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 03:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 04/23/2010 04:22:08

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.083	04:16
Air Blank	0.000	04:17
Subject Sample #1	0.183	04:17
Breath Volume	2.601	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample #2	0.183	04:20
Breath Volume	2.472	04:20
Air Blank	0.000	04:21
Control Test	0.083	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time 05/02/2010 09:19

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 04/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 03:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 04/24/2010 03:47:16

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.083	03:41
Air Blank	0.000	03:41
Subject Sample #1	0.219	03:42
Breath Volume	1.785	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample #2	0.217	03:45
Breath Volume	1.886	03:45
Air Blank	0.000	03:46
Control Test	0.083	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time 05/02/2010 09:19

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 04/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 00:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BEBEE

DATE/TIME OF TEST: 04/25/2010 00:57:57

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.083	00:52
Air Blank	0.000	00:52
Subject Sample #1	0.096	00:53
Breath Volume	1.355	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample #2	0.090	00:56
Breath Volume	1.253	00:56
Air Blank	0.000	00:57
Control Test	0.083	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time 05/02/2010 09:19

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 04/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 04:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: LEVISTER

DATE/TIME OF TEST: 04/25/2010 04:39:14

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.083	04:33
Air Blank	0.000	04:33
Subject Sample #1	0.159	04:34
Breath Volume	2.890	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:36
Subject Sample #2	0.159	04:37
Breath Volume	3.308	04:37
Air Blank	0.000	04:38
Control Test	0.083	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time 05/02/2010 09:19

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERSON, RHONDA, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MAZURYK, MICHAEL
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 04/03/2010 02:42:10

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Subject Sample #1	0.135	02:37
Breath Volume	1.523	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample #2	0.134	02:40
Breath Volume	1.515	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time 04/29/2010 07:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERSON, RHONDA, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: SMITH, JASON
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/04/2010 03:34:03

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Subject Sample #1		03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.079	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:34	

Data Download Date/Time 04/29/2010 07:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE, FRANK, M
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: SIMMONS, JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: WILKES

DATE/TIME OF TEST: 04/11/2010 03:07:00

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:04	
Control Test	0.080	03:04	
Air Blank	0.000	03:05	
Subject Sample #1		03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:07	

Data Download Date/Time 04/29/2010 07:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE, FRANK, M
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: SIMMONS, JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 704221
DRY GAS EXPIRY DATE: 12/11/2011
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HOLZENDORF

DATE/TIME OF TEST: 04/30/2010 04:11:20

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:07	
Control Test	0.077	04:07	
Air Blank	0.000	04:07	
Subject Sample #1		04:09	Subject Test Refused
Breath Volume	0.042	04:09	
Air Blank	0.000	04:10	
Control Test	0.079	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time 05/31/2010 01:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DITTY, ALLEN, S
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOCK, PATRICK
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: GURDIAN

DATE/TIME OF TEST: 04/07/2010 02:12:59

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.079	02:11	
Air Blank	0.000	02:12	
Subject Sample #1		02:12	Sequence Aborted
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	

Data Download Date/Time 04/22/2010 01:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DITTY, ALLEN, S
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOCK, PATRICK
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 03/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: GURDIAN

DATE/TIME OF TEST: 04/07/2010 02:20:49

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.078	02:18	
Air Blank	0.000	02:19	
Subject Sample #1		02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Control Test	0.078	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time 04/22/2010 01:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLS, ZACHARY, B

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: MILLS, ZACHARY

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 828901G

DRY GAS EXPIRY DATE: 10/21/2010

VIOLATION CODE: DUI

ARREST TIME: 01:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HURTER

DATE/TIME OF TEST: 04/02/2010 01:52:31

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:42
Subject Sample #1	0.202	01:44
Breath Volume	1.492	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample #2	0.251	01:47
Breath Volume	3.269	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample #3	0.197	01:51
Breath Volume	2.449	01:51
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time 04/30/2010 05:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLS, ZACHARY, B

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: MILLS, ZACHARY

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 828901G

DRY GAS EXPIRY DATE: 10/21/2010

VIOLATION CODE: DUI

ARREST TIME: 01:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HURTER

DATE/TIME OF TEST: 04/02/2010 02:02:59

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:57
Subject Sample #1	0.220	01:58
Breath Volume	3.007	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample #2	0.216	02:01
Breath Volume	2.777	02:01
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time 04/30/2010 05:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOTTE, JASON, J

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: MOTTE, JASON

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 828901G

DRY GAS EXPIRY DATE: 10/21/2010

VIOLATION CODE: DUI

ARREST TIME: 09:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 04/09/2010 09:58:28

Diagnostics Check	OK	09:52
Air Blank	0.000	09:52
Control Test	0.079	09:53
Air Blank	0.000	09:53
Subject Sample #1	0.000	09:54
Breath Volume	2.230	09:54
Air Blank	0.000	09:54
Air Blank	0.000	09:56
Subject Sample #2	0.000	09:57
Breath Volume	1.890	09:57
Air Blank	0.000	09:57
Control Test	0.079	09:57
Air Blank	0.000	09:58
Diagnostics Check	OK	09:58

Data Download Date/Time 04/30/2010 05:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HORNING, PETER, T

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: ASHCOM, TIFFANY

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 828901G

DRY GAS EXPIRY DATE: 10/21/2010

VIOLATION CODE: DUI

ARREST TIME: 10:48

GENDER: Female

OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: AMATO

DATE/TIME OF TEST: 04/10/2010 11:45:40

Diagnostics Check	OK	11:39
Air Blank	0.000	11:39
Control Test	0.078	11:40
Air Blank	0.000	11:40
Subject Sample #1	0.224	11:41
Breath Volume	3.042	11:41
Air Blank	0.000	11:41
Air Blank	0.000	11:43
Subject Sample #2	0.230	11:44
Breath Volume	3.675	11:44
Air Blank	0.000	11:44
Control Test	0.077	11:45
Air Blank	0.000	11:45
Diagnostics Check	OK	11:45

Data Download Date/Time 04/30/2010 05:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODMAN, THOMAS, B
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: APOSTOLU, JODY
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 828901G
DRY GAS EXPIRY DATE: 10/21/2010
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: HELM

DATE/TIME OF TEST: 04/11/2010 22:51:06

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Subject Sample #1	0.250	22:46
Breath Volume	2.445	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample #2	0.233	22:49
Breath Volume	2.089	22:49
Air Blank	0.000	22:50
Control Test	0.076	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time 04/30/2010 05:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WOODMAN, THOMAS, B

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: MCLAUGHLIN, ROBERT

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 828901G

DRY GAS EXPIRY DATE: 10/21/2010

VIOLATION CODE: DUI

ARREST TIME: 15:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 04/16/2010 16:09:23

Diagnostics Check	OK	16:06	
Air Blank	0.000	16:07	
Control Test	0.079	16:07	
Air Blank	0.000	16:07	
Subject Sample #1		16:08	Subject Test Refused
Breath Volume	0.000	16:08	
Air Blank	0.000	16:08	
Control Test	0.079	16:08	
Air Blank	0.000	16:09	
Diagnostics Check	OK	16:09	

Data Download Date/Time 04/30/2010 05:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLS, ZACHARY, B

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: MILLS, ZACHARY

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 828901G

DRY GAS EXPIRY DATE: 10/21/2010

VIOLATION CODE: DUI

ARREST TIME: 21:22

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/19/2010 22:03:39

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Subject Sample #1	0.211	21:59
Breath Volume	2.496	21:59
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample #2	0.206	22:02
Breath Volume	2.246	22:02
Air Blank	0.000	22:02
Control Test	0.078	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time 04/30/2010 05:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WOODMAN, THOMAS, B

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: PARKS, MATTHEW

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 828901G

DRY GAS EXPIRY DATE: 10/21/2010

VIOLATION CODE: DUI

ARREST TIME: 23:29

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: ISSE-DURNEIKA

DATE/TIME OF TEST: 04/25/2010 23:59:38

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:53
Air Blank	0.000	23:53
Subject Sample #1	0.227	23:54
Breath Volume	2.000	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:56
Subject Sample #2	0.235	23:58
Breath Volume	1.867	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time 04/30/2010 05:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS, ZACHARY, B
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: MILLS, ZACHARY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 828901G
DRY GAS EXPIRY DATE: 10/21/2010
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: KUDRYASHOVA

DATE/TIME OF TEST: 04/28/2010 01:13:16

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:10	
Control Test	0.078	01:10	
Air Blank	0.000	01:11	
Subject Sample #1		01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:12	
Control Test	0.078	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time 04/30/2010 05:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLS, ZACHARY, B

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: MILLS, ZACHARY

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 828901G

DRY GAS EXPIRY DATE: 10/21/2010

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/29/2010 03:25:00

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Subject Sample #1	0.222	03:20
Breath Volume	4.855	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample #2	0.206	03:23
Breath Volume	3.496	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:25

Data Download Date/Time 04/30/2010 05:20

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFIELD, ADAM, G
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: BURTON, BRANT
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SURDICH

DATE/TIME OF TEST: 04/02/2010 23:12:08

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.082	23:09	
Air Blank	0.000	23:10	
Subject Sample #1		23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:11	
Control Test	0.082	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOCK, PHILIP, L
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: BOCK, PHILIP
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: GILEAU

DATE/TIME OF TEST: 04/04/2010 02:21:11

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:09	
Control Test	0.082	02:09	
Air Blank	0.000	02:10	
Subject Sample #1	0.265	02:10	
Breath Volume	1.234	02:10	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample #2	0.219	02:16	Volume Not Met
Breath Volume	0.042	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample #3		02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.078	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFIELD, ADAM, G
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: BEELER, B
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: BEAVER

DATE/TIME OF TEST: 04/10/2010 03:11:01

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.082	03:05
Air Blank	0.000	03:05
Subject Sample #1	0.160	03:06
Breath Volume	1.277	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample #2	0.145	03:09
Breath Volume	1.644	03:09
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:11

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFIELD, ADAM, G
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: BURTON, BRANDT
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/17/2010 18:08:16

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:05	
Control Test	0.082	18:06	
Air Blank	0.000	18:06	
Subject Sample #1		18:06	Subject Test Refused
Breath Volume	0.000	18:06	
Air Blank	0.000	18:07	
Control Test	0.081	18:07	
Air Blank	0.000	18:08	
Diagnostics Check	OK	18:08	

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFIELD, ADAM, G
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: LIGHTFIELD, ADAM
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: LIU

DATE/TIME OF TEST: 04/18/2010 20:28:46

Diagnostics Check	OK	20:19
Air Blank	0.000	20:20
Control Test	0.082	20:20
Air Blank	0.000	20:21
Subject Sample #1	0.315	20:21
Breath Volume	2.031	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:23
Subject Sample #2	0.285	20:24
Breath Volume	1.476	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:26
Subject Sample #3	0.304	20:27
Breath Volume	1.835	20:27
Air Blank	0.000	20:27
Control Test	0.081	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFIELD, ADAM, G
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: BURTON, BRANDT
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 04/19/2010 01:38:31

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:32
Subject Sample #1	0.227	01:34
Breath Volume	1.601	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample #2	0.229	01:36
Breath Volume	1.609	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWETNICH, JARED, A
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: SWETNICH, JARED
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/20/2010 03:28:28

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.082	03:27	
Air Blank		03:28	RFI Detect
Air Blank	0.000	03:28	

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWETNICH, JARED, A
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: SWETNICH, JARED
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/20/2010 03:38:22

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.082	03:30
Air Blank	0.000	03:31
Subject Sample #1	0.179	03:33
Breath Volume	1.445	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample #2	0.179	03:36
Breath Volume	1.382	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDDY, GERALD, C
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: TIMLIN, STEVEN
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: REDDY

DATE/TIME OF TEST: 04/24/2010 16:46:04

Diagnostics Check	OK	16:45	
Air Blank		16:45	Improper Sample
Air Blank	0.000	16:46	

Data Download Date/Time	05/07/2010	09:51
-------------------------	------------	-------

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDDY, GERALD, C
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: TIMLIN, STEVEN
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: REDDY

DATE/TIME OF TEST: 04/24/2010 16:51:13

Diagnostics Check	OK	16:49	
Air Blank	0.000	16:50	
Control Test	0.081	16:50	
Air Blank		16:50	Improper Sample
Air Blank	0.000	16:51	

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDDY, GERALD, C
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: TIMLIN, STEVE
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: REDDY

DATE/TIME OF TEST: 04/24/2010 17:09:24

Diagnostics Check	OK	17:03
Air Blank	0.000	17:03
Control Test	0.082	17:04
Air Blank	0.000	17:04
Subject Sample #1	0.274	17:05
Breath Volume	1.980	17:05
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample #2	0.283	17:07
Breath Volume	3.281	17:07
Air Blank	0.000	17:08
Control Test	0.081	17:08
Air Blank	0.000	17:09
Diagnostics Check	OK	17:09

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOCK, PHILIP, L
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: BOCK, PHILIP
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KLEM

DATE/TIME OF TEST: 04/25/2010 00:45:00

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Subject Sample #1	0.263	00:40
Breath Volume	2.921	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample #2	0.265	00:43
Breath Volume	3.457	00:43
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:45

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ, HECTOR,
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: KENNEDY, JAMES
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: KILBURN

DATE/TIME OF TEST: 04/27/2010 18:21:53

Diagnostics Check	OK	18:12	
Air Blank	0.000	18:12	
Control Test	0.082	18:13	
Air Blank	0.000	18:13	
Subject Sample #1	0.227	18:16	Volume Not Met
Breath Volume	0.816	18:16	
Air Blank	0.000	18:17	
Air Blank	0.000	18:19	
Subject Sample #2		18:20	Subject Test Refused
Breath Volume	0.000	18:20	
Air Blank	0.000	18:20	
Control Test	0.081	18:21	
Air Blank	0.000	18:21	
Diagnostics Check	OK	18:21	

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ, HECTOR,
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: SEVERIN, TAMMY
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 04/28/2010 04:18:53

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:14	
Control Test	0.082	04:15	
Air Blank	0.000	04:15	
Subject Sample #1		04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.083	04:18	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time 05/07/2010 09:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET, R,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET, R
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/21/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814301G
DRY GAS EXPIRY DATE: 05/30/2010
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 04/04/2010 00:42:02

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.077	00:36
Air Blank	0.000	00:37
Subject Sample #1	0.186	00:37
Breath Volume	2.226	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample #2	0.181	00:40
Breath Volume	2.386	00:40
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:42

Data Download Date/Time 04/18/2010 18:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET, R,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: COOKE, J
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 04/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814301G
DRY GAS EXPIRY DATE: 05/30/2010
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: COUTRIER

DATE/TIME OF TEST: 04/24/2010 03:49:59

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.077	03:44
Air Blank	0.000	03:45
Subject Sample #1	0.125	03:45
Breath Volume	1.367	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample #2	0.126	03:48
Breath Volume	1.468	03:48
Air Blank	0.000	03:49
Control Test	0.077	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time 05/23/2010 19:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET, R,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: CHANG, A
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 04/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814301G
DRY GAS EXPIRY DATE: 05/30/2010
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/25/2010 05:37:26

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.077	05:31
Air Blank	0.000	05:32
Subject Sample #1	0.164	05:32
Breath Volume	1.863	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample #2	0.160	05:35
Breath Volume	1.808	05:35
Air Blank	0.000	05:36
Control Test	0.078	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time 05/23/2010 19:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSON, DONALD, I
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: FLORES, ANTHONY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 04/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814301G
DRY GAS EXPIRY DATE: 05/30/2010
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:19

SUBJECT'S LAST NAME: OSTROWSKI

DATE/TIME OF TEST: 04/25/2010 15:49:18

Diagnostics Check	OK	15:42
Air Blank	0.000	15:42
Control Test	0.077	15:43
Air Blank	0.000	15:43
Subject Sample #1	0.316	15:44
Breath Volume	2.433	15:44
Air Blank	0.000	15:45
Air Blank	0.000	15:46
Subject Sample #2	0.332	15:47
Breath Volume	2.906	15:47
Air Blank	0.000	15:48
Control Test	0.076	15:48
Air Blank	0.000	15:49
Diagnostics Check	OK	15:49

Data Download Date/Time 05/23/2010 19:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON, CHRISTOPHER, S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: VALDES, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/03/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: MESNARD

DATE/TIME OF TEST: 04/10/2010 18:51:02

Diagnostics Check	OK	18:44
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:45
Subject Sample #1	0.146	18:46
Breath Volume	1.769	18:46
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample #2	0.155	18:49
Breath Volume	1.832	18:49
Air Blank	0.000	18:50
Control Test	0.079	18:50
Air Blank	0.000	18:50
Diagnostics Check	OK	18:51

Data Download Date/Time 06/03/2010 06:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON, CHRISTOPHER, S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: DENTON, CHRISTOPHER
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/03/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: FELDKAMP

DATE/TIME OF TEST: 04/11/2010 04:57:33

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.080	04:52
Air Blank	0.000	04:52
Subject Sample #1	0.000	04:53
Breath Volume	2.281	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample #2	0.000	04:56
Breath Volume	1.796	04:56
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time 06/03/2010 06:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON, CHRISTOPHER, S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: DENTON, CHRISTOPHER
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/03/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/22/2010 02:44:46

Diagnostics Check	Fail	02:44	Temperature Regulation Test
Air Blank	0.000	02:44	

Data Download Date/Time 06/03/2010 06:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON, CHRISTOPHER, S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: DENTON, CHRISTOPHER
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/03/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/22/2010 02:55:44

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Subject Sample #1	0.141	02:47
Breath Volume	2.238	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample #2	0.117	02:50
Breath Volume	1.921	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample #3	0.136	02:54
Breath Volume	2.351	02:54
Air Blank	0.000	02:54
Control Test	0.078	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time 06/03/2010 06:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASCALLI, JOSEPH, M
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PASCALLI, JOSEPH
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/03/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: FORTE

DATE/TIME OF TEST: 04/27/2010 15:23:22

Diagnostics Check	OK	15:15
Air Blank	0.000	15:16
Control Test	0.079	15:16
Air Blank	0.000	15:16
Subject Sample #1	0.000	15:17
Breath Volume	2.558	15:17
Air Blank	0.000	15:18
Air Blank	0.000	15:20
Subject Sample #2	0.000	15:22
Breath Volume	2.167	15:22
Air Blank	0.000	15:22
Control Test	0.079	15:22
Air Blank	0.000	15:23
Diagnostics Check	OK	15:23

Data Download Date/Time 06/03/2010 06:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON, CHRISTOPHER, S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: DENTON, CHRISTOPHER
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/03/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: PROCTOR

DATE/TIME OF TEST: 04/29/2010 03:44:17

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Subject Sample #1	0.137	03:39
Breath Volume	1.500	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample #2	0.134	03:42
Breath Volume	2.449	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time 06/03/2010 06:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VELSOR, DONALD, R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: VELSOR, DONALD
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 04/03/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MAZUR

DATE/TIME OF TEST: 04/30/2010 00:06:55

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:01
Subject Sample #1	0.089	00:02
Breath Volume	3.406	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample #2	0.092	00:05
Breath Volume	2.078	00:05
Air Blank	0.000	00:06
Control Test	0.078	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time 06/03/2010 06:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONDEK, CHARLES, R
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: OCHOA, LIONEL
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822601I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HEISHMAN

DATE/TIME OF TEST: 04/04/2010 06:08:19

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:05	
Control Test	0.078	06:06	
Air Blank	0.000	06:06	
Subject Sample #1		06:06	Subject Test Refused
Breath Volume	0.000	06:06	
Air Blank	0.000	06:07	
Control Test	0.079	06:07	
Air Blank	0.000	06:08	
Diagnostics Check	OK	06:08	

Data Download Date/Time 05/07/2010 08:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL, ROBERT, A
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: WEIL, ROBERT
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: HELMAN

DATE/TIME OF TEST: 04/03/2010 02:58:34

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:53
Subject Sample #1	0.000	02:54
Breath Volume	3.781	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.000	02:57
Breath Volume	3.035	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL, ROBERT, A
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: WEIL, ROBERT
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 04/03/2010 23:26:02

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:22	
Control Test	0.079	23:22	
Air Blank	0.000	23:23	
Subject Sample #1		23:24	Slope Not Met
Breath Volume	0.691	23:24	
Air Blank	0.000	23:25	
Control Test	0.075	23:25	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:26	

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL, ROBERT, A
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: WEIL, ROBERT
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 04/03/2010 23:29:39

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:28	
Control Test	0.074	23:29	Control Outside Tolerance
Air Blank	0.000	23:29	

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL, ROBERT, A
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: WEIL, ROBERT
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 04/03/2010 23:41:07

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.080	23:31	
Air Blank	0.000	23:32	
Subject Sample #1	0.231	23:33	
Breath Volume	1.183	23:33	
Air Blank	0.000	23:34	
Air Blank	0.000	23:35	
Subject Sample #2	0.279	23:36	
Breath Volume	1.414	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:38	
Subject Sample #3		23:39	Slope Not Met
Breath Volume	0.574	23:39	
Air Blank	0.000	23:40	
Control Test	0.074	23:40	Control Outside Tolerance
Air Blank	0.000	23:41	

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL, ROBERT, A
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: WEIL, ROBERT
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 04/03/2010 23:49:21

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.076	23:44
Air Blank	0.000	23:44
Subject Sample #1	0.282	23:44
Breath Volume	1.878	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample #2	0.298	23:47
Breath Volume	1.773	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAMBACH, BRETT, M
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: APOSTOLU, JODY
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 04/04/2010 03:33:34

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Subject Sample #1	0.185	03:29
Breath Volume	1.167	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample #2	0.188	03:31
Breath Volume	1.515	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RALEIGH, A,
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: RALEIGH, A
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: BEYETTE

DATE/TIME OF TEST: 04/08/2010 19:06:51

Diagnostics Check	OK	19:03	
Air Blank	0.000	19:04	
Control Test	0.082	19:04	
Air Blank	0.000	19:05	
Subject Sample #1		19:05	Subject Test Refused
Breath Volume	0.000	19:05	
Air Blank	0.000	19:05	
Control Test	0.081	19:06	
Air Blank	0.000	19:06	
Diagnostics Check	OK	19:06	

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RALEIGH, A,
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: RALEIGH, A
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 04/10/2010 01:48:03

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Subject Sample #1	0.118	01:42
Breath Volume	1.304	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample #2	0.115	01:46
Breath Volume	1.289	01:46
Air Blank	0.000	01:47
Control Test	0.081	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:48

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL, ROBERT, A
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: SEVERIN, TAMMY
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/13/2010 23:50:16

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Subject Sample #1	0.187	23:45
Breath Volume	2.109	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample #2	0.168	23:48
Breath Volume	1.648	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RALEIGH, A,
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: RALEIGH, A
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BRODERICK

DATE/TIME OF TEST: 04/16/2010 03:01:53

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.081	02:53
Air Blank	0.000	02:53
Subject Sample #1	0.140	02:54
Breath Volume	1.925	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.161	02:57
Breath Volume	2.179	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample #3	0.122	03:00
Breath Volume	1.566	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAMBACH, BRETT, M
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: APOSTOLU, JODY
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 04/18/2010 06:34:28

Diagnostics Check	OK	06:30	
Air Blank	0.000	06:30	
Control Test	0.081	06:31	
Air Blank	0.000	06:31	
Subject Sample #1		06:33	Subject Test Refused
Breath Volume	0.000	06:33	
Air Blank	0.000	06:33	
Control Test	0.080	06:33	
Air Blank	0.000	06:34	
Diagnostics Check	OK	06:34	

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RALEIGH, A,
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: RALEIGH, A
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: PANUTHOS

DATE/TIME OF TEST: 04/23/2010 22:09:00

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:04
Subject Sample #1	0.216	22:04
Breath Volume	2.351	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample #2	0.214	22:07
Breath Volume	2.195	22:07
Air Blank	0.000	22:08
Control Test	0.080	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:09

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RALEIGH, A,
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: RALEIGH, A
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/24/2010 23:37:21

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.081	23:34	
Air Blank	0.000	23:35	
Subject Sample #1		23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:36	
Control Test	0.081	23:36	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RALEIGH, A,
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: RALEIGH, A
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 04/25/2010 01:04:35

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:59
Subject Sample #1	0.179	01:00
Breath Volume	1.691	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample #2	0.182	01:03
Breath Volume	1.875	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time 04/28/2010 10:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK, R, A
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: LUNDY, B
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 04/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: DUFFEY

DATE/TIME OF TEST: 04/28/2010 22:37:53

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Subject Sample #1	0.219	22:32
Breath Volume	1.769	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample #2	0.218	22:36
Breath Volume	1.546	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time 05/28/2010 15:57

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: ST PETE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIN, DENNIS, E
BREATH TEST OPERATOR AGENCY: St. Pete Beach PD
ARRESTING OFFICER: LUNDY, BRANDON
ARRESTING OFFICER AGENCY: St. Pete Beach PD
LAST AGENCY INSPECTION: 04/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/25/2011
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 04/30/2010 19:05:42

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.079	19:00
Air Blank	0.000	19:00
Subject Sample #1	0.179	19:01
Breath Volume	3.613	19:01
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample #2	0.181	19:04
Breath Volume	3.179	19:04
Air Blank	0.000	19:04
Control Test	0.080	19:05
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time 05/28/2010 15:57

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESLER, OFC.K., G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RESLER, OFC.K.
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 581913
DRY GAS EXPIRY DATE: 02/23/2011
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CHUNN

DATE/TIME OF TEST: 04/05/2010 03:52:57

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample #1	0.157	03:48
Breath Volume	1.605	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:50
Subject Sample #2	0.145	03:51
Breath Volume	1.707	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time 04/30/2010 07:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURPACK, T,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: TURPACK, T
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 581913
DRY GAS EXPIRY DATE: 02/23/2011
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: LARCOMB

DATE/TIME OF TEST: 04/09/2010 20:34:17

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.080	20:28
Air Blank	0.000	20:29
Subject Sample #1	0.000	20:29
Breath Volume	3.933	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample #2	0.000	20:32
Breath Volume	3.828	20:32
Air Blank	0.000	20:33
Control Test	0.081	20:33
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time 04/30/2010 07:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLER, THEODORE, J

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: MILLER, THEODORE

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 581913

DRY GAS EXPIRY DATE: 02/23/2011

VIOLATION CODE: DUI

ARREST TIME: 14:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:33

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 04/10/2010 14:59:49

Diagnostics Check	OK	14:53
Air Blank	0.000	14:54
Control Test	0.081	14:54
Air Blank	0.000	14:55
Subject Sample #1	0.000	14:55
Breath Volume	2.609	14:55
Air Blank	0.000	14:55
Air Blank	0.000	14:57
Subject Sample #2	0.000	14:58
Breath Volume	1.730	14:58
Air Blank	0.000	14:58
Control Test	0.081	14:59
Air Blank	0.000	14:59
Diagnostics Check	OK	14:59

Data Download Date/Time 04/30/2010 07:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESLER, OFC.K., G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RESLER, OFC.K.
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 581913
DRY GAS EXPIRY DATE: 02/23/2011
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/13/2010 23:41:46

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Subject Sample #1	0.050	23:37
Breath Volume	2.000	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample #2	0.051	23:40
Breath Volume	1.800	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time 04/30/2010 07:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIM, TYRONE, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CRIM, TRPR T.
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 581913
DRY GAS EXPIRY DATE: 02/23/2011
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SELF

DATE/TIME OF TEST: 04/21/2010 05:38:02

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:35	
Control Test	0.080	05:35	
Air Blank	0.000	05:36	
Subject Sample #1		05:36	Subject Test Refused
Breath Volume	0.000	05:36	
Air Blank	0.000	05:37	
Control Test	0.079	05:37	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:38	

Data Download Date/Time 04/30/2010 07:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THEODORE, J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILLER, THEODORE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 581913
DRY GAS EXPIRY DATE: 02/23/2011
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: LANEY

DATE/TIME OF TEST: 04/22/2010 01:29:15

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample #1	0.144	01:24
Breath Volume	4.304	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample #2	0.141	01:27
Breath Volume	4.605	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time 04/30/2010 07:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOWNS, T, J

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: WALDREP, B

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 22:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: TORRES-HIPINA

DATE/TIME OF TEST: 04/01/2010 00:07:05

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.082	00:01
Air Blank	0.000	00:02
Subject Sample #1	0.155	00:02
Breath Volume	2.968	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample #2	0.158	00:05
Breath Volume	1.671	00:05
Air Blank	0.000	00:06
Control Test	0.083	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNS, T, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, K
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MULLANEY

DATE/TIME OF TEST: 04/01/2010 03:13:17

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.082	03:07
Air Blank	0.000	03:08
Subject Sample #1	0.108	03:08
Breath Volume	3.984	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample #2	0.102	03:11
Breath Volume	3.054	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARRIDINE, W, L

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: JONES, C

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 16:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 04/01/2010 19:12:02

Diagnostics Check	OK	19:04
Air Blank	0.000	19:05
Control Test	0.082	19:05
Air Blank	0.000	19:06
Subject Sample #1	0.173	19:07
Breath Volume	1.546	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample #2	0.172	19:10
Breath Volume	1.988	19:10
Air Blank	0.000	19:11
Control Test	0.082	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:12

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOWNS, T, J

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: KEY, M

ARRESTING OFFICER AGENCY: Neptune Beach PD

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 21:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: TIPPINS

DATE/TIME OF TEST: 04/02/2010 00:20:09

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:09	
Subject Sample #1	0.157	00:13	Volume Not Met
Breath Volume	0.132	00:13	
Air Blank	0.000	00:13	
Air Blank	0.000	00:15	
Subject Sample #2	0.155	00:18	Volume Not Met
Breath Volume	0.726	00:18	
Air Blank	0.000	00:19	
Control Test	0.082	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNS, T, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEY, M
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: TIPPINS

DATE/TIME OF TEST: 04/02/2010 01:03:39

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.082	00:54	
Air Blank	0.000	00:54	
Subject Sample #1	0.137	00:58	Volume Not Met
Breath Volume	0.019	00:58	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample #2		01:02	Subject Test Refused
Breath Volume	1.179	01:02	
Air Blank	0.000	01:02	
Control Test	0.081	01:03	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNS, T, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, K
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SMITHA

DATE/TIME OF TEST: 04/02/2010 04:43:57

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:31	
Control Test	0.081	04:31	
Air Blank	0.000	04:32	
Subject Sample #1		04:35	No Sample Provided
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample #2	0.143	04:39	
Breath Volume	2.449	04:39	
Air Blank	0.000	04:39	
Air Blank	0.000	04:41	
Subject Sample #3	0.141	04:42	
Breath Volume	1.632	04:42	
Air Blank	0.000	04:43	
Control Test	0.081	04:43	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FULLER, ELWIN, E

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HURT, S

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 05:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: MANLEY

DATE/TIME OF TEST: 04/02/2010 07:12:24

Diagnostics Check	OK	07:05
Air Blank	0.000	07:06
Control Test	0.081	07:06
Air Blank	0.000	07:06
Subject Sample #1	0.157	07:07
Breath Volume	4.539	07:07
Air Blank	0.000	07:08
Air Blank	0.000	07:10
Subject Sample #2	0.155	07:10
Breath Volume	4.230	07:10
Air Blank	0.000	07:11
Control Test	0.082	07:11
Air Blank	0.000	07:12
Diagnostics Check	OK	07:12

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARRIDINE, W, L

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HOUSEND, B

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 20:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: MCKINZIE

DATE/TIME OF TEST: 04/02/2010 21:53:47

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.082	21:51	
Air Blank	0.000	21:51	
Subject Sample #1		21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Control Test	0.082	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NICHOLS, A, B

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: KEY, M

ARRESTING OFFICER AGENCY: Neptune Beach PD

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 21:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/02/2010 23:24:12

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.082	23:21	
Air Blank	0.000	23:22	
Subject Sample #1		23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:23	
Control Test	0.083	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, C, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAMLIN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: MACIAS

DATE/TIME OF TEST: 04/03/2010 00:20:38

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:18	
Control Test	0.082	00:18	
Air Blank	0.000	00:18	
Subject Sample #1		00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:19	
Control Test	0.082	00:20	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARRIDINE, W, L

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: STRAHLER, F

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 15:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:49

SUBJECT'S LAST NAME: RAPHAEL

DATE/TIME OF TEST: 04/04/2010 17:13:11

Diagnostics Check	OK	17:10	
Air Blank	0.000	17:10	
Control Test	0.083	17:11	
Air Blank	0.000	17:11	
Subject Sample #1		17:11	Subject Test Refused
Breath Volume	0.000	17:11	
Air Blank	0.000	17:12	
Control Test	0.081	17:12	
Air Blank	0.000	17:13	
Diagnostics Check	OK	17:13	

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIDINE, W, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRAY, B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/04/2010 20:12:16

Diagnostics Check	OK	20:05
Air Blank	0.000	20:06
Control Test	0.082	20:06
Air Blank	0.000	20:06
Subject Sample #1	0.157	20:07
Breath Volume	4.000	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample #2	0.160	20:10
Breath Volume	4.230	20:10
Air Blank	0.000	20:11
Control Test	0.083	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEITH, C, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: CHAPMAN, S

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 01:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: SLOTT

DATE/TIME OF TEST: 04/05/2010 02:41:35

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.082	02:38	
Air Blank	0.000	02:38	
Subject Sample #1		02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Control Test	0.082	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEITH, C, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: SIRMONS, M

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 03:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MANDAJADO

DATE/TIME OF TEST: 04/05/2010 04:44:28

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.082	04:38
Air Blank	0.000	04:39
Subject Sample #1	0.166	04:39
Breath Volume	3.410	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample #2	0.164	04:43
Breath Volume	3.609	04:43
Air Blank	0.000	04:43
Control Test	0.083	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T.A.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: NEELEY

DATE/TIME OF TEST: 04/06/2010 23:25:11

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.082	23:19
Air Blank	0.000	23:19
Subject Sample #1	0.220	23:20
Breath Volume	1.289	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample #2	0.224	23:23
Breath Volume	1.515	23:23
Air Blank	0.000	23:24
Control Test	0.081	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS, S.N.
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 04/07/2010 02:35:51

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample #1	0.266	02:31
Breath Volume	2.136	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample #2	0.270	02:34
Breath Volume	1.261	02:34
Air Blank	0.000	02:34
Control Test	0.082	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JACKSON, RON
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/07/2010 23:44:58

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.082	23:39
Air Blank	0.000	23:39
Subject Sample #1	0.000	23:40
Breath Volume	2.292	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample #2	0.000	23:43
Breath Volume	1.148	23:43
Air Blank	0.000	23:44
Control Test	0.083	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARRIDINE, W, L

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BARRON, K

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 17:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/09/2010 18:47:35

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.083	18:42
Air Blank	0.000	18:42
Subject Sample #1	0.227	18:43
Breath Volume	2.394	18:43
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample #2	0.223	18:46
Breath Volume	1.468	18:46
Air Blank	0.000	18:46
Control Test	0.083	18:47
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEITH, C, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: ODER, C

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 22:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: RASOOL

DATE/TIME OF TEST: 04/10/2010 00:17:46

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.082	00:12
Air Blank	0.000	00:12
Subject Sample #1	0.100	00:13
Breath Volume	2.386	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample #2	0.100	00:16
Breath Volume	2.351	00:16
Air Blank	0.000	00:16
Control Test	0.083	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEITH, C, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS, S
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 04/10/2010 03:32:11

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.083	03:26
Air Blank	0.000	03:27
Subject Sample #1	0.038	03:27
Breath Volume	4.906	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample #2	0.037	03:30
Breath Volume	4.746	03:30
Air Blank	0.000	03:31
Control Test	0.083	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEITH, C, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TOLMAN, J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/10/2010 04:35:02

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.083	04:29
Air Blank	0.000	04:29
Subject Sample #1	0.132	04:30
Breath Volume	2.410	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample #2	0.135	04:33
Breath Volume	1.773	04:33
Air Blank	0.000	04:34
Control Test	0.083	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:35

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEITH, C, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: FISSEL, M

ARRESTING OFFICER AGENCY: Atlantic Beach PD

LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 03:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: KARISH

DATE/TIME OF TEST: 04/10/2010 05:42:55

Diagnostics Check	OK	05:39	
Air Blank	0.000	05:40	
Control Test	0.083	05:40	
Air Blank	0.000	05:41	
Subject Sample #1		05:41	Subject Test Refused
Breath Volume	0.000	05:41	
Air Blank	0.000	05:42	
Control Test	0.083	05:42	
Air Blank	0.000	05:42	
Diagnostics Check	OK	05:42	

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS, WILLIE, M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAMLIN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 04/13/2010 20:36:39

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.082	20:30
Air Blank	0.000	20:30
Subject Sample #1	0.000	20:32
Breath Volume	1.523	20:32
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample #2	0.000	20:35
Breath Volume	1.625	20:35
Air Blank	0.000	20:35
Control Test	0.082	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time 04/15/2010 09:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRON, KEITH
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 04/17/2010 01:20:13

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.082	01:14
Air Blank	0.000	01:15
Subject Sample #1	0.147	01:15
Breath Volume	3.410	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample #2	0.148	01:18
Breath Volume	3.281	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORGERT, MATT
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: WINGFIELD

DATE/TIME OF TEST: 04/17/2010 02:47:43

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:42
Subject Sample #1	0.170	02:43
Breath Volume	3.761	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample #2	0.174	02:46
Breath Volume	3.382	02:46
Air Blank	0.000	02:46
Control Test	0.082	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JAMISON, C.S.
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 04/17/2010 03:40:02

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Subject Sample #1	0.136	03:35
Breath Volume	2.285	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample #2	0.135	03:38
Breath Volume	1.683	03:38
Air Blank	0.000	03:39
Control Test	0.082	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:40

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRON, KEITH
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: HENDRICKS

DATE/TIME OF TEST: 04/17/2010 04:25:21

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:20
Subject Sample #1	0.093	04:20
Breath Volume	2.128	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.094	04:23
Breath Volume	1.304	04:23
Air Blank	0.000	04:24
Control Test	0.082	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ODER, C.D.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HEAD

DATE/TIME OF TEST: 04/17/2010 05:22:30

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.081	05:16
Air Blank	0.000	05:17
Subject Sample #1	0.187	05:18
Breath Volume	2.296	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample #2	0.198	05:21
Breath Volume	2.187	05:21
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO, E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/17/2010 19:19:41

Diagnostics Check	OK	19:13
Air Blank	0.000	19:13
Control Test	0.082	19:13
Air Blank	0.000	19:14
Subject Sample #1	0.134	19:15
Breath Volume	2.296	19:15
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample #2	0.129	19:18
Breath Volume	1.550	19:18
Air Blank	0.000	19:18
Control Test	0.082	19:19
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: PIKULSKI

DATE/TIME OF TEST: 04/17/2010 20:43:17

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:40	
Control Test	0.082	20:41	
Air Blank	0.000	20:41	
Subject Sample #1		20:41	Subject Test Refused
Breath Volume	0.000	20:41	
Air Blank	0.000	20:42	
Control Test	0.082	20:42	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:43	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRON, K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/17/2010 22:38:09

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:35	
Control Test	0.082	22:35	
Air Blank	0.000	22:35	
Subject Sample #1		22:36	Subject Test Refused
Breath Volume	0.000	22:36	
Air Blank	0.000	22:37	
Control Test	0.082	22:37	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:38	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT, RICHARD, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: WEISE

DATE/TIME OF TEST: 04/18/2010 00:07:33

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.081	00:01
Air Blank	0.000	00:01
Subject Sample #1	0.086	00:02
Breath Volume	1.792	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample #2	0.086	00:06
Breath Volume	1.156	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT, RICHARD, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRON, K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: JEMENEZ

DATE/TIME OF TEST: 04/18/2010 01:47:51

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.082	01:41
Air Blank	0.000	01:41
Subject Sample #1	0.155	01:43
Breath Volume	1.332	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample #2	0.165	01:46
Breath Volume	2.167	01:46
Air Blank	0.000	01:46
Control Test	0.082	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT, RICHARD, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSTON, C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: CASEY

DATE/TIME OF TEST: 04/18/2010 03:20:22

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.082	03:18	
Air Blank	0.000	03:18	
Subject Sample #1		03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.081	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT, RICHARD, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, K
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CHEEVERS

DATE/TIME OF TEST: 04/18/2010 03:46:44

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.083	03:40
Air Blank	0.000	03:41
Subject Sample #1	0.152	03:42
Breath Volume	1.703	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample #2	0.156	03:45
Breath Volume	2.285	03:45
Air Blank	0.000	03:45
Control Test	0.083	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORTER, GARY, W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ODER, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: PRAK

DATE/TIME OF TEST: 04/18/2010 04:23:46

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Subject Sample #1	0.189	04:19
Breath Volume	3.085	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample #2	0.185	04:22
Breath Volume	3.332	04:22
Air Blank	0.000	04:22
Control Test	0.082	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT, RICHARD, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRINER, C
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 04/18/2010 05:43:22

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:40	
Control Test	0.082	05:41	
Air Blank	0.000	05:41	
Subject Sample #1		05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:42	
Control Test	0.082	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT, RICHARD, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRINER, C
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: ADMIN
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 04/18/2010 05:52:16

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.082	05:46
Air Blank	0.000	05:47
Subject Sample #1	0.113	05:47
Breath Volume	3.289	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample #2	0.113	05:50
Breath Volume	3.644	05:50
Air Blank	0.000	05:51
Control Test	0.083	05:51
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JENNINGS, A, S

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BARRON, K

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 15:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:49

SUBJECT'S LAST NAME: BENITO

DATE/TIME OF TEST: 04/18/2010 17:19:02

Diagnostics Check	OK	17:10
Air Blank	0.000	17:10
Control Test	0.084	17:10
Air Blank	0.000	17:11
Subject Sample #1	0.187	17:14
Breath Volume	1.328	17:14
Air Blank	0.000	17:14
Air Blank	0.000	17:16
Subject Sample #2	0.198	17:17
Breath Volume	2.492	17:17
Air Blank	0.000	17:18
Control Test	0.081	17:18
Air Blank	0.000	17:18
Diagnostics Check	OK	17:19

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRIDGES, E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 04/18/2010 19:25:00

Diagnostics Check	OK	19:17
Air Blank	0.000	19:18
Control Test	0.083	19:18
Air Blank	0.000	19:18
Subject Sample #1	0.123	19:20
Breath Volume	2.175	19:20
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample #2	0.124	19:23
Breath Volume	2.183	19:23
Air Blank	0.000	19:24
Control Test	0.083	19:24
Air Blank	0.000	19:24
Diagnostics Check	OK	19:25

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: OSBURN

DATE/TIME OF TEST: 04/18/2010 21:55:19

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.082	21:49
Air Blank	0.000	21:50
Subject Sample #1	0.147	21:50
Breath Volume	4.472	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample #2	0.144	21:53
Breath Volume	4.625	21:53
Air Blank	0.000	21:54
Control Test	0.082	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JACKSON, RON
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 04/19/2010 01:29:33

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:27	
Control Test	0.081	01:27	
Air Blank	0.000	01:27	
Subject Sample #1		01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.082	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T.A.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/19/2010 02:35:59

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Subject Sample #1	0.141	02:30
Breath Volume	1.820	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample #2	0.149	02:34
Breath Volume	2.691	02:34
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALDREP, B.A.
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 04/19/2010 03:48:11

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.082	03:46	
Air Blank	0.000	03:46	
Subject Sample #1		03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.082	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LYON, G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: LIKE

DATE/TIME OF TEST: 04/20/2010 21:31:13

Diagnostics Check	OK	21:24
Air Blank	0.000	21:25
Control Test	0.081	21:25
Air Blank	0.000	21:25
Subject Sample #1	0.122	21:26
Breath Volume	2.187	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:28
Subject Sample #2	0.111	21:29
Breath Volume	1.871	21:29
Air Blank	0.000	21:30
Control Test	0.082	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEITH, C, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WOOLDRIDGE, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/22/2010 00:27:36

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.082	00:22
Air Blank	0.000	00:22
Subject Sample #1	0.161	00:23
Breath Volume	2.816	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample #2	0.154	00:26
Breath Volume	1.820	00:26
Air Blank	0.000	00:26
Control Test	0.082	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIDINE, W, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAMLIN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 04/22/2010 19:39:41

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.082	19:34
Air Blank	0.000	19:35
Subject Sample #1	0.000	19:35
Breath Volume	4.234	19:35
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample #2	0.000	19:38
Breath Volume	4.222	19:38
Air Blank	0.000	19:38
Control Test	0.082	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNS, T, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTILUCCI, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BUTTERMAN

DATE/TIME OF TEST: 04/23/2010 00:51:56

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.082	00:46
Air Blank	0.000	00:47
Subject Sample #1	0.000	00:47
Breath Volume	2.234	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample #2	0.000	00:50
Breath Volume	2.332	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNS, T, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, K
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/23/2010 04:23:13

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.081	04:21	
Air Blank	0.000	04:21	
Subject Sample #1		04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Control Test	0.081	04:22	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEITH, C, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSTON, C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 04/24/2010 02:22:15

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.082	02:16
Air Blank	0.000	02:17
Subject Sample #1	0.085	02:17
Breath Volume	3.566	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample #2	0.083	02:20
Breath Volume	3.664	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEITH, C, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: EZELL, S

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 01:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/24/2010 03:11:08

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.081	03:08	
Air Blank	0.000	03:09	
Subject Sample #1		03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEITH, C, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: ROBERTS, W

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 02:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/24/2010 03:58:20

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.082	03:56	
Air Blank	0.000	03:56	
Subject Sample #1		03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:57	
Control Test	0.081	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FIELDS, WILLIE, M

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BROWN, T

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 20:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 04/24/2010 22:39:21

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.082	22:37	
Air Blank	0.000	22:37	
Subject Sample #1		22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.082	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FIELDS, WILLIE, M

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BARRON, K

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 19:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: HELTON

DATE/TIME OF TEST: 04/26/2010 20:45:11

Diagnostics Check	OK	20:42	
Air Blank	0.000	20:42	
Control Test	0.083	20:43	
Air Blank	0.000	20:43	
Subject Sample #1		20:43	Subject Test Refused
Breath Volume	0.000	20:43	
Air Blank	0.000	20:44	
Control Test	0.083	20:44	
Air Blank	0.000	20:45	
Diagnostics Check	OK	20:45	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FIELDS, WILLIE, M

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BRIDGES, E

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 19:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/26/2010 23:12:03

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:06
Subject Sample #1	0.000	23:07
Breath Volume	1.625	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample #2	0.000	23:10
Breath Volume	1.539	23:10
Air Blank	0.000	23:11
Control Test	0.082	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:12

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAMPBELL, C, J

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: CHRISTMAS, F

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 22:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: PUTNAM

DATE/TIME OF TEST: 04/27/2010 00:22:34

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:17
Air Blank	0.000	00:17
Subject Sample #1	0.185	00:18
Breath Volume	3.906	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample #2	0.170	00:21
Breath Volume	2.179	00:21
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHRISTMAS, FOLIN
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/28/2010 02:14:02

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.083	02:11	
Air Blank	0.000	02:12	
Subject Sample #1		02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.083	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:14	

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY, MIKE
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: FALLIN

DATE/TIME OF TEST: 04/28/2010 03:52:37

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.083	03:46
Air Blank	0.000	03:47
Subject Sample #1	0.133	03:48
Breath Volume	1.343	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample #2	0.136	03:51
Breath Volume	1.671	03:51
Air Blank	0.000	03:51
Control Test	0.082	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT, RICHARD, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALDREP, B
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: LAND

DATE/TIME OF TEST: 04/28/2010 05:36:26

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.082	05:30
Air Blank	0.000	05:31
Subject Sample #1	0.117	05:32
Breath Volume	1.750	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample #2	0.112	05:34
Breath Volume	1.429	05:34
Air Blank	0.000	05:35
Control Test	0.083	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, JAMES
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 04/29/2010 02:21:21

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.081	02:16
Air Blank	0.000	02:16
Subject Sample #1	0.083	02:17
Breath Volume	3.308	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.088	02:19
Breath Volume	3.039	02:19
Air Blank	0.000	02:20
Control Test	0.082	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, KENNETH, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORGERT, MATT
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: DONNELLY

DATE/TIME OF TEST: 04/29/2010 04:34:51

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.081	04:29
Air Blank	0.000	04:29
Subject Sample #1	0.193	04:30
Breath Volume	3.988	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample #2	0.178	04:33
Breath Volume	4.097	04:33
Air Blank	0.000	04:33
Control Test	0.083	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time 04/29/2010 14:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PRUETT, RICHAD, L

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: WOOLDRIDGE, D

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 04/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 00:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BYRNES

DATE/TIME OF TEST: 04/30/2010 02:09:59

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.082	02:07	
Air Blank	0.000	02:07	
Subject Sample #1		02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:09	
Control Test	0.082	02:09	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time 05/13/2010 13:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT, RICHARD, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRIFFIS, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: MCWHERTER

DATE/TIME OF TEST: 04/30/2010 06:36:10

Diagnostics Check	OK	06:30
Air Blank	0.000	06:30
Control Test	0.081	06:30
Air Blank	0.000	06:31
Subject Sample #1	0.139	06:31
Breath Volume	2.539	06:31
Air Blank	0.000	06:32
Air Blank	0.000	06:34
Subject Sample #2	0.135	06:34
Breath Volume	2.257	06:34
Air Blank	0.000	06:35
Control Test	0.082	06:35
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time 05/13/2010 13:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

INSTRUMENT REGISTERED TO: CEDAR KEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMULLEN, JOHN, L
BREATH TEST OPERATOR AGENCY: Cedar Keys PD
ARRESTING OFFICER: ROBERTS, BRANDON
ARRESTING OFFICER AGENCY: Cedar Keys PD
LAST AGENCY INSPECTION: 03/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 581913
DRY GAS EXPIRY DATE: 02/26/2011
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/04/2010 02:33:51

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.296	02:29
Breath Volume	1.558	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample #2	0.292	02:32
Breath Volume	2.183	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time 04/29/2010 17:16

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUENN, O, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLS, D
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: WHITENER

DATE/TIME OF TEST: 04/01/2010 22:52:03

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:41	
Control Test	0.078	22:41	
Air Blank	0.000	22:41	
Subject Sample #1		22:45	No Sample Provided
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample #2		22:50	No Sample Provided
Breath Volume	0.000	22:50	
Air Blank	0.000	22:51	
Control Test	0.079	22:51	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:52	

Data Download Date/Time 04/15/2010 02:56

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRAY, B, A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MURRAY, B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: COURT ORD
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/04/2010 15:39:44

Diagnostics Check	OK	15:33
Air Blank	0.000	15:34
Control Test	0.078	15:34
Air Blank	0.000	15:35
Subject Sample #1	0.000	15:35
Breath Volume	2.035	15:35
Air Blank	0.000	15:35
Air Blank	0.000	15:37
Subject Sample #2	0.000	15:38
Breath Volume	1.851	15:38
Air Blank	0.000	15:38
Control Test	0.079	15:39
Air Blank	0.000	15:39
Diagnostics Check	OK	15:39

Data Download Date/Time 04/15/2010 02:56

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPAULDING, JORDON, S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GARRISON, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: HANGEN

DATE/TIME OF TEST: 04/05/2010 20:46:02

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.078	20:40
Air Blank	0.000	20:41
Subject Sample #1	0.199	20:41
Breath Volume	1.390	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample #2	0.204	20:44
Breath Volume	1.476	20:44
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:46

Data Download Date/Time 04/15/2010 02:56

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH, ROBERT, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MINYARD, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MINARD

DATE/TIME OF TEST: 04/07/2010 03:35:09

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.078	03:29
Air Blank	0.000	03:29
Subject Sample #1	0.158	03:30
Breath Volume	1.218	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample #2	0.153	03:33
Breath Volume	2.933	03:33
Air Blank	0.000	03:34
Control Test	0.077	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time 04/15/2010 02:56

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY, DALE, V
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ELDER, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 04/07/2010 07:25:11

Diagnostics Check	OK	07:18
Air Blank	0.000	07:19
Control Test	0.079	07:19
Air Blank	0.000	07:20
Subject Sample #1	0.156	07:20
Breath Volume	3.125	07:20
Air Blank	0.000	07:21
Air Blank	0.000	07:23
Subject Sample #2	0.160	07:23
Breath Volume	3.207	07:23
Air Blank	0.000	07:24
Control Test	0.078	07:24
Air Blank	0.000	07:25
Diagnostics Check	OK	07:25

Data Download Date/Time 04/15/2010 02:56

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUENN, O, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GARRISON, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: BATTEN

DATE/TIME OF TEST: 04/10/2010 00:02:27

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:54	
Control Test	0.080	23:54	
Air Blank	0.000	23:55	
Subject Sample #1		23:58	No Sample Provided
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Air Blank	0.000	00:00	
Subject Sample #2		00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time 04/15/2010 02:56

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUENN, O, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GARRISON, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/14/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: SOSA CEREZO

DATE/TIME OF TEST: 04/11/2010 23:10:01

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:04
Subject Sample #1	0.108	23:05
Breath Volume	1.296	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample #2	0.120	23:08
Breath Volume	2.054	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:10

Data Download Date/Time 04/15/2010 02:56

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH, ROBERT, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GARRISON, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 04/16/2010 03:29:46

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.079	03:23
Air Blank	0.000	03:23
Subject Sample #1	0.081	03:24
Breath Volume	1.292	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample #2	0.083	03:28
Breath Volume	1.488	03:28
Air Blank	0.000	03:28
Control Test	0.077	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, S, E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CAMPBELL, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: SHULTS

DATE/TIME OF TEST: 04/16/2010 23:50:36

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.079	23:44
Air Blank	0.000	23:44
Subject Sample #1	0.110	23:45
Breath Volume	1.113	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample #2	0.110	23:49
Breath Volume	1.941	23:49
Air Blank	0.000	23:49
Control Test	0.078	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY, DALE, V
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: KACZKOWSKI, ED
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 04/18/2010 02:21:26

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Subject Sample #1		02:17	No Sample Provided
Breath Volume	0.000	02:17	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample #2		02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.079	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS, DEPUTY, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RUSSELL, DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 04/18/2010 21:32:38

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Subject Sample #1	0.152	21:28
Breath Volume	2.824	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample #2	0.154	21:31
Breath Volume	2.312	21:31
Air Blank	0.000	21:31
Control Test	0.077	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS, DEPUTY, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RUSSELL, DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 11:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: RIGNEY

DATE/TIME OF TEST: 04/23/2010 01:28:59

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.079	01:21	
Air Blank	0.000	01:21	
Subject Sample #1		01:24	No Sample Provided
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample #2		01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Control Test	0.078	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH, K, J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TAYLOR, S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: ACKMAN

DATE/TIME OF TEST: 04/24/2010 13:25:45

Diagnostics Check	OK	13:17
Air Blank	0.000	13:17
Control Test	0.079	13:17
Air Blank	0.000	13:18
Subject Sample #1	0.235	13:20
Breath Volume	1.765	13:20
Air Blank	0.000	13:20
Air Blank	0.000	13:22
Subject Sample #2	0.236	13:24
Breath Volume	2.171	13:24
Air Blank	0.000	13:24
Control Test	0.078	13:25
Air Blank	0.000	13:25
Diagnostics Check	OK	13:25

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPAULDING, JORDON, S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: EZELL, S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: GANN

DATE/TIME OF TEST: 04/25/2010 02:01:12

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample #1	0.082	01:56
Breath Volume	3.339	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample #2	0.080	01:59
Breath Volume	3.363	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH, ROBERT, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TAYLOR, S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: STANFORD

DATE/TIME OF TEST: 04/25/2010 18:07:05

Diagnostics Check	OK	18:00
Air Blank	0.000	18:01
Control Test	0.079	18:01
Air Blank	0.000	18:02
Subject Sample #1	0.000	18:02
Breath Volume	2.328	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample #2	0.000	18:05
Breath Volume	2.867	18:05
Air Blank	0.000	18:06
Control Test	0.077	18:06
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY, DALE, V
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RUSSELL, ROBERT
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/26/2010 20:45:35

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.078	20:39
Air Blank	0.000	20:39
Subject Sample #1	0.202	20:40
Breath Volume	2.546	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample #2	0.205	20:44
Breath Volume	2.640	20:44
Air Blank	0.000	20:44
Control Test	0.077	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, S, E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RUSSELL, R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: DILMUTH

DATE/TIME OF TEST: 04/27/2010 01:26:39

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample #1	0.048	01:22
Breath Volume	2.886	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample #2	0.043	01:25
Breath Volume	4.167	01:25
Air Blank	0.000	01:25
Control Test	0.078	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURRETTE, T, K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LEWIS, D
ARRESTING OFFICER AGENCY: FDOT MCC
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 04/27/2010 17:44:58

Diagnostics Check	OK	17:38
Air Blank	0.000	17:39
Control Test	0.079	17:39
Air Blank	0.000	17:40
Subject Sample #1	0.109	17:40
Breath Volume	1.964	17:40
Air Blank	0.000	17:41
Air Blank	0.000	17:43
Subject Sample #2	0.104	17:43
Breath Volume	1.917	17:43
Air Blank	0.000	17:44
Control Test	0.078	17:44
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY, DALE, V
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RUSSELL, ROBERT
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 04/27/2010 23:37:07

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.081	23:25	
Air Blank	0.000	23:25	
Subject Sample #1		23:28	No Sample Provided
Breath Volume	0.000	23:28	
Air Blank	0.000	23:29	
Air Blank	0.000	23:31	
Subject Sample #2	0.193	23:31	
Breath Volume	3.136	23:31	
Air Blank	0.000	23:32	
Air Blank	0.000	23:34	
Subject Sample #3	0.195	23:35	
Breath Volume	2.710	23:35	
Air Blank	0.000	23:36	
Control Test	0.079	23:36	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH, ROBERT, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RUBY, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: SHCICKLER

DATE/TIME OF TEST: 04/30/2010 03:53:50

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:47
Subject Sample #1	0.130	03:48
Breath Volume	2.515	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:50
Subject Sample #2	0.131	03:52
Breath Volume	1.976	03:52
Air Blank	0.000	03:52
Control Test	0.078	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time 05/18/2010 14:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STONE, OFC., C

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: NAPOLEON, DEP.

ARRESTING OFFICER AGENCY: Pasco County SO

LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 812201G

DRY GAS EXPIRY DATE: 05/06/2010

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: UNGERER

DATE/TIME OF TEST: 04/03/2010 02:51:09

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.077	02:45
Air Blank	0.000	02:45
Subject Sample #1	0.180	02:46
Breath Volume	2.238	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample #2	0.174	02:49
Breath Volume	1.343	02:49
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time 04/21/2010 16:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BURNS, KEVIN, T

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: BURNS, KEVIN

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 812201G

DRY GAS EXPIRY DATE: 05/06/2010

VIOLATION CODE: DUI

ARREST TIME: 23:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 04/04/2010 00:08:11

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.077	00:02
Air Blank	0.000	00:02
Subject Sample #1	0.140	00:03
Breath Volume	2.066	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample #2	0.141	00:06
Breath Volume	1.644	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time 04/21/2010 16:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE, DAVID, S
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: ROQUE, DAVID
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 812201G
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HAMBRICK

DATE/TIME OF TEST: 04/04/2010 03:49:26

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.076	03:41	
Air Blank	0.000	03:41	
Subject Sample #1	0.280	03:43	
Breath Volume	2.027	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:45	
Subject Sample #2		03:48	Sequence Aborted
Breath Volume	0.117	03:48	
Air Blank	0.000	03:49	

Data Download Date/Time 04/21/2010 16:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STONE, OFC., C

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: MECHLER, OFC.

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 812201G

DRY GAS EXPIRY DATE: 05/06/2010

VIOLATION CODE: DUI

ARREST TIME: 15:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: EARNEST

DATE/TIME OF TEST: 04/06/2010 16:04:01

Diagnostics Check	OK	15:57
Air Blank	0.000	15:58
Control Test	0.078	15:58
Air Blank	0.000	15:59
Subject Sample #1	0.113	15:59
Breath Volume	2.996	15:59
Air Blank	0.000	16:00
Air Blank	0.000	16:01
Subject Sample #2	0.110	16:02
Breath Volume	2.394	16:02
Air Blank	0.000	16:03
Control Test	0.079	16:03
Air Blank	0.000	16:03
Diagnostics Check	OK	16:04

Data Download Date/Time 04/21/2010 16:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BALDWIN, DEAN,

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: JOHNSON, JOSH

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 812201G

DRY GAS EXPIRY DATE: 05/06/2010

VIOLATION CODE: DUI

ARREST TIME: 20:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/10/2010 21:49:34

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.078	21:43
Air Blank	0.000	21:44
Subject Sample #1	0.174	21:44
Breath Volume	1.113	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample #2	0.175	21:47
Breath Volume	1.148	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time 04/21/2010 16:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS, KEVIN, T
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: PETERSEN, ANTHONY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/19/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 812201G
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 04/12/2010 23:36:53

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:31
Subject Sample #1	0.179	23:32
Breath Volume	2.871	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample #2	0.181	23:35
Breath Volume	3.777	23:35
Air Blank	0.000	23:35
Control Test	0.078	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time 04/21/2010 16:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VELSOR, DONALD, R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: VELSOR, DONALD
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/01/2010 04:05:44

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample #1		04:04	Interferent Detect
Breath Volume	1.554	04:04	
Air Blank		04:04	Purge Fail
Air Blank		04:05	Purge Fail

Data Download Date/Time 04/13/2010 13:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VELSOR, DONALD, R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: VELSOR, DONALD
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/01/2010 04:15:50

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.079	04:09	
Air Blank	0.000	04:09	
Subject Sample #1	0.142	04:10	
Breath Volume	1.769	04:10	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample #2	0.148	04:14	
Breath Volume	2.054	04:14	
Air Blank		04:14	Purge Fail
Air Blank	0.000	04:15	

Data Download Date/Time 04/13/2010 13:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK, SCOTT, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHUCK, SCOTT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 04/02/2010 04:06:48

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Subject Sample #1	0.104	04:05	
Breath Volume	1.804	04:05	
Air Blank		04:06	Purge Fail
Air Blank	0.000	04:06	

Data Download Date/Time 04/13/2010 13:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS, R, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EVANS, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 04/03/2010 05:03:34

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.080	04:59	
Air Blank	0.000	05:00	
Subject Sample #1		05:02	Subject Test Refused
Breath Volume	0.000	05:02	
Air Blank	0.000	05:02	
Control Test	0.080	05:02	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time 04/13/2010 13:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR, ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MONSEGUR, ROBERT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: ERD

DATE/TIME OF TEST: 04/03/2010 22:43:29

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.080	22:41	
Air Blank	0.000	22:41	
Subject Sample #1	0.205	22:42	
Breath Volume	1.953	22:42	
Air Blank		22:43	Purge Fail
Air Blank	0.000	22:43	

Data Download Date/Time 04/13/2010 13:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR, ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MONSEGUR, ROBERT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: ERD

DATE/TIME OF TEST: 04/03/2010 23:11:24

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Subject Sample #1	0.202	23:06
Breath Volume	2.359	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample #2	0.198	23:09
Breath Volume	2.652	23:09
Air Blank	0.000	23:10
Control Test	0.082	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time 04/13/2010 13:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VELSOR, DONALD, R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: VELSOR, DONALD
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: CHOLOPRAY

DATE/TIME OF TEST: 04/06/2010 03:02:43

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.082	02:57
Air Blank	0.000	02:57
Subject Sample #1	0.000	02:58
Breath Volume	2.433	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample #2	0.000	03:01
Breath Volume	1.949	03:01
Air Blank	0.000	03:01
Control Test	0.083	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time 04/13/2010 13:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWTHORNE, RICHARD, S
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: MCCLINTOCH, CHRIS
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/10/2011
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 04/01/2010 04:08:27

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample #1	0.232	04:03
Breath Volume	1.867	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample #2	0.231	04:06
Breath Volume	1.101	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time 04/18/2010 23:16

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, JOHN, T
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: MARCUS, MARK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822601I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GARET

DATE/TIME OF TEST: 04/12/2010 03:26:49

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.076	03:19
Air Blank	0.000	03:20
Subject Sample #1	0.030	03:22
Breath Volume	1.386	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample #2	0.029	03:25
Breath Volume	1.511	03:25
Air Blank	0.000	03:25
Control Test	0.075	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time 04/17/2010 22:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSNELL, JEFFREY, K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/01/2010 20:38:30

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.082	20:25	
Air Blank	0.000	20:26	
Subject Sample #1	0.237	20:26	
Breath Volume	1.125	20:26	
Air Blank	0.000	20:27	
Air Blank	0.000	20:29	
Subject Sample #2	0.179	20:32	Volume Not Met
Breath Volume	0.390	20:32	
Air Blank	0.000	20:33	
Air Blank	0.000	20:35	
Subject Sample #3	0.234	20:36	
Breath Volume	1.355	20:36	
Air Blank	0.000	20:37	
Control Test	0.078	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSNELL, JEFFREY, K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: PEPPERLING

DATE/TIME OF TEST: 04/01/2010 23:12:06

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:07
Subject Sample #1	0.000	23:07
Breath Volume	3.019	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample #2	0.000	23:10
Breath Volume	2.429	23:10
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX, LENARD, E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SADIKI, NAYO
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BAUL

DATE/TIME OF TEST: 04/02/2010 00:24:40

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:22	
Control Test	0.080	00:22	
Air Blank	0.000	00:23	
Subject Sample #1		00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.079	00:24	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX, LENARD, E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISOPF, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: BARNEY

DATE/TIME OF TEST: 04/02/2010 02:32:32

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:20	
Control Test	0.080	02:20	
Air Blank	0.000	02:21	
Subject Sample #1	0.180	02:24	Volume Not Met
Breath Volume	0.179	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:26	
Subject Sample #2	0.190	02:27	
Breath Volume	1.699	02:27	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample #3	0.188	02:31	
Breath Volume	2.082	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSNELL, JEFFREY, K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GOSNELL, JEFFREY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 04/03/2010 02:02:52

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Subject Sample #1	0.159	01:58
Breath Volume	1.914	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample #2	0.156	02:01
Breath Volume	1.687	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SESSIONS, C, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SESSIONS, C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 08:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 04/03/2010 09:37:04

Diagnostics Check	OK	09:22	
Air Blank	0.000	09:22	
Control Test	0.080	09:22	
Air Blank	0.000	09:23	
Subject Sample #1		09:26	No Sample Provided
Breath Volume	0.000	09:26	
Air Blank	0.000	09:27	
Air Blank	0.000	09:29	
Subject Sample #2	0.187	09:29	
Breath Volume	1.339	09:29	
Air Blank	0.000	09:30	
Air Blank	0.000	09:32	
Subject Sample #3	0.182	09:35	Volume Not Met
Breath Volume	0.203	09:35	
Air Blank	0.000	09:36	
Control Test	0.079	09:36	
Air Blank	0.000	09:36	
Diagnostics Check	OK	09:37	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX, LENARD, E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SADIKI, NAYO
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/03/2010 23:19:04

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.075	23:17	
Air Blank	0.000	23:17	
Subject Sample #1		23:17	Subject Test Refused
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Control Test	0.080	23:18	
Air Blank	0.000	23:18	
Diagnostics Check	OK	23:19	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GOSNELL, JEFFREY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/04/2010 02:25:35

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.080	02:11	
Air Blank	0.000	02:11	
Subject Sample #1	0.213	02:14	Volume Not Met
Breath Volume	0.058	02:14	
Air Blank	0.000	02:15	
Air Blank	0.000	02:17	
Subject Sample #2	0.252	02:18	
Breath Volume	1.113	02:18	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample #3	0.184	02:23	Volume Not Met
Breath Volume	0.660	02:23	
Air Blank	0.000	02:24	
Control Test	0.079	02:24	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GOSNELL, JEFFREY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/04/2010 02:42:35

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.079	02:28	
Air Blank	0.000	02:28	
Subject Sample #1	0.254	02:29	
Breath Volume	1.812	02:29	
Air Blank	0.000	02:30	
Air Blank	0.000	02:32	
Subject Sample #2	0.255	02:35	Volume Not Met
Breath Volume	0.042	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:38	
Subject Sample #3		02:41	No Sample Provided
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.078	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESHUR, MICHAEL, F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WESTON

DATE/TIME OF TEST: 04/04/2010 03:22:34

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.079	03:11	
Air Blank	0.000	03:12	
Subject Sample #1		03:15	No Sample Provided
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Air Blank	0.000	03:17	
Subject Sample #2	0.161	03:21	Volume Not Met
Breath Volume	0.023	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORIGLIO, BRIAN, C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: INNES, JOSH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 04/06/2010 05:36:32

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Subject Sample #1	0.165	05:30
Breath Volume	2.343	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample #2	0.161	05:35
Breath Volume	1.800	05:35
Air Blank	0.000	05:35
Control Test	0.081	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GOSNELL, JEFFREY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PRUNA

DATE/TIME OF TEST: 04/08/2010 00:27:01

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:22
Subject Sample #1	0.103	00:22
Breath Volume	1.738	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample #2	0.099	00:25
Breath Volume	1.691	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:27

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSNELL, JEFFREY, K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PLEUNE

DATE/TIME OF TEST: 04/08/2010 01:00:38

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Subject Sample #1	0.175	00:55
Breath Volume	1.980	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample #2	0.163	00:59
Breath Volume	1.476	00:59
Air Blank	0.000	00:59
Control Test	0.081	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX, LENARD, E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/08/2010 03:12:16

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample #1	0.212	03:07
Breath Volume	1.339	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample #2	0.206	03:10
Breath Volume	1.269	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX, LENARD, E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 04/08/2010 04:49:01

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:40
Subject Sample #1	0.112	04:41
Breath Volume	1.914	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample #2	0.135	04:44
Breath Volume	3.441	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample #3	0.132	04:47
Breath Volume	4.625	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:49

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUGLAS, JOHN, J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: FARELL, AUBRIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:43

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 04/08/2010 10:15:45

Diagnostics Check	OK	10:04	
Air Blank	0.000	10:04	
Control Test	0.080	10:04	
Air Blank	0.000	10:05	
Subject Sample #1	0.000	10:08	Volume Not Met
Breath Volume	0.894	10:08	
Air Blank	0.000	10:08	
Air Blank	0.000	10:10	
Subject Sample #2	0.000	10:11	
Breath Volume	2.320	10:11	
Air Blank	0.000	10:11	
Air Blank	0.000	10:13	
Subject Sample #3	0.000	10:14	
Breath Volume	1.851	10:14	
Air Blank	0.000	10:14	
Control Test	0.080	10:15	
Air Blank	0.000	10:15	
Diagnostics Check	OK	10:15	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX, LENARD, E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 04/08/2010 19:01:49

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.078	18:55
Air Blank	0.000	18:55
Subject Sample #1	0.321	18:56
Breath Volume	1.937	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample #2	0.323	19:00
Breath Volume	2.425	19:00
Air Blank	0.000	19:00
Control Test	0.077	19:01
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESHUR, MICHAEL, F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: FELTON

DATE/TIME OF TEST: 04/08/2010 23:15:02

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Subject Sample #1	0.159	23:10
Breath Volume	2.246	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:12
Subject Sample #2	0.156	23:13
Breath Volume	1.414	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:15

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LARSON, RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: PERSECHINO

DATE/TIME OF TEST: 04/09/2010 01:25:55

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Subject Sample #1	0.108	01:21
Breath Volume	1.339	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample #2	0.106	01:24
Breath Volume	1.582	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESHUR, MICHAEL, F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRIM, TYRONE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: AHERN

DATE/TIME OF TEST: 04/09/2010 02:23:58

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.080	02:21	
Air Blank	0.000	02:22	
Subject Sample #1		02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: WOLFE III

DATE/TIME OF TEST: 04/09/2010 03:11:03

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:09	
Subject Sample #1		03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:11	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/09/2010 18:51:54

Diagnostics Check	OK	18:48	
Air Blank	0.000	18:49	
Control Test	0.080	18:49	
Air Blank	0.000	18:50	
Subject Sample #1		18:50	Subject Test Refused
Breath Volume	0.000	18:50	
Air Blank	0.000	18:51	
Control Test	0.081	18:51	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/09/2010 19:12:17

Diagnostics Check	OK	19:05	
Air Blank	0.000	19:06	
Control Test	0.080	19:06	
Air Blank	0.000	19:07	
Subject Sample #1	0.299	19:08	
Breath Volume	2.082	19:08	
Air Blank	0.000	19:09	
Air Blank	0.000	19:11	
Subject Sample #2	0.290	19:11	
Breath Volume	1.394	19:11	
Air Blank		19:11	RFI Detect
Air Blank	0.000	19:12	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/09/2010 19:23:33

Diagnostics Check	OK	19:13
Air Blank	0.000	19:13
Control Test	0.080	19:14
Air Blank	0.000	19:14
Subject Sample #1	0.295	19:15
Breath Volume	1.417	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample #2	0.323	19:18
Breath Volume	1.714	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample #3	0.297	19:22
Breath Volume	2.820	19:22
Air Blank	0.000	19:22
Control Test	0.079	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESHUR, MICHAEL, F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HOBBY

DATE/TIME OF TEST: 04/10/2010 02:42:17

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample #1	0.030	02:37
Breath Volume	3.085	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.030	02:40
Breath Volume	3.140	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON, MATTHEW, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DEMARK, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 04/10/2010 04:30:33

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Subject Sample #1		04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:29	
Control Test	0.080	04:29	
Air Blank		04:30	RFI Detect
Air Blank	0.000	04:30	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON, MATTHEW, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DEMARK, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 04/10/2010 04:35:54

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:33	
Control Test	0.079	04:33	
Air Blank	0.000	04:34	
Subject Sample #1		04:34	Subject Test Refused
Breath Volume	0.000	04:34	
Air Blank	0.000	04:35	
Control Test	0.079	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON, MATTHEW, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: NICHOLS, JEFFREY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: MCARTHUR

DATE/TIME OF TEST: 04/10/2010 06:08:37

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:05	
Control Test	0.081	06:06	
Air Blank	0.000	06:06	
Subject Sample #1		06:07	Subject Test Refused
Breath Volume	0.000	06:07	
Air Blank	0.000	06:07	
Control Test	0.080	06:08	
Air Blank	0.000	06:08	
Diagnostics Check	OK	06:08	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESHUR, MICHAEL, F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HARRIS, JASON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/10/2010 18:24:37

Diagnostics Check	OK	18:18
Air Blank	0.000	18:18
Control Test	0.080	18:19
Air Blank	0.000	18:19
Subject Sample #1	0.222	18:20
Breath Volume	1.632	18:20
Air Blank	0.000	18:20
Air Blank	0.000	18:22
Subject Sample #2	0.227	18:23
Breath Volume	2.652	18:23
Air Blank	0.000	18:23
Control Test	0.078	18:24
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

Data Download Date/Time 04/28/2010 19:02

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GOSNELL, JEFFREY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 04/10/2010 19:57:45

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.082	19:52
Air Blank	0.000	19:52
Subject Sample #1	0.000	19:53
Breath Volume	1.644	19:53
Air Blank	0.000	19:53
Air Blank	0.000	19:56
Subject Sample #2	0.000	19:56
Breath Volume	1.539	19:56
Air Blank	0.000	19:56
Control Test	0.081	19:57
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESHUR, MICHAEL, F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SCRIMO

DATE/TIME OF TEST: 04/10/2010 22:51:35

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:43	
Control Test	0.080	22:44	
Air Blank	0.000	22:44	
Subject Sample #1	0.238	22:45	
Breath Volume	2.164	22:45	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample #2		22:50	Subject Test Refused
Breath Volume	0.000	22:50	
Air Blank	0.000	22:50	
Control Test	0.080	22:51	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESHUR, MICHAEL, F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 04/11/2010 03:03:18

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.080	03:01	
Air Blank	0.000	03:01	
Subject Sample #1		03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.080	03:02	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEWITT, AMY, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALKER, CALVIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: SCHOEN

DATE/TIME OF TEST: 04/12/2010 20:50:19

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:37	
Control Test	0.080	20:38	
Air Blank	0.000	20:38	
Subject Sample #1	0.221	20:41	Volume Not Met
Breath Volume	0.656	20:41	
Air Blank	0.000	20:42	
Air Blank	0.000	20:44	
Subject Sample #2	0.246	20:45	No .020 agreement
Breath Volume	2.570	20:45	
Air Blank	0.000	20:45	
Air Blank	0.000	20:47	
Subject Sample #3	0.225	20:48	No .020 agreement
Breath Volume	1.703	20:48	
Air Blank	0.000	20:49	
Control Test	0.080	20:49	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEWITT, AMY, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALKER, C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: SCHOEN

DATE/TIME OF TEST: 04/12/2010 21:08:08

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.079	20:59	
Air Blank	0.000	21:00	
Subject Sample #1	0.231	21:01	
Breath Volume	1.746	21:01	
Air Blank	0.000	21:02	
Air Blank	0.000	21:04	
Subject Sample #2		21:06	RFI Detect
Breath Volume	0.000	21:06	
Air Blank	0.000	21:07	
Control Test	0.078	21:07	
Air Blank	0.000	21:08	
Diagnostics Check	OK	21:08	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORIGLIO, BRIAN, C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: INNES, JOSH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: CHENG

DATE/TIME OF TEST: 04/13/2010 03:11:29

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Subject Sample #1	0.098	03:07
Breath Volume	2.792	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample #2	0.106	03:10
Breath Volume	1.250	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORIGLIO, BRIAN, C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: INNES, JOSH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 04/14/2010 05:31:03

Diagnostics Check	OK	05:27	
Air Blank	0.000	05:27	
Control Test	0.080	05:28	
Air Blank	0.000	05:28	
Subject Sample #1		05:29	Subject Test Refused
Breath Volume	0.000	05:29	
Air Blank	0.000	05:30	
Control Test	0.080	05:30	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:31	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSNELL, JEFFREY, K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: OUELLET

DATE/TIME OF TEST: 04/14/2010 23:58:43

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.081	23:56	
Air Blank	0.000	23:57	
Subject Sample #1		23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSNELL, JEFFREY, K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: HASTLER

DATE/TIME OF TEST: 04/16/2010 01:50:24

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Subject Sample #1	0.222	01:44
Breath Volume	2.101	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample #2	0.215	01:48
Breath Volume	2.179	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSNELL, JEFFREY, K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BAHENA

DATE/TIME OF TEST: 04/16/2010 03:45:36

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.081	03:33	
Air Blank	0.000	03:33	
Subject Sample #1	0.171	03:37	Volume Not Met
Breath Volume	0.593	03:37	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample #2	0.168	03:40	
Breath Volume	2.042	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample #3	0.167	03:44	
Breath Volume	1.175	03:44	
Air Blank	0.000	03:44	
Control Test	0.079	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSNELL, JEFFREY, K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ROOD

DATE/TIME OF TEST: 04/16/2010 23:52:27

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:50	
Subject Sample #1		23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Control Test	0.081	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 04/17/2010 03:27:10

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Subject Sample #1	0.086	03:21
Breath Volume	1.457	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample #2	0.087	03:25
Breath Volume	1.957	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time 04/28/2010 19:02

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: DEEL

DATE/TIME OF TEST: 04/18/2010 02:58:41

Diagnostics Check	OK	02:58	
Air Blank		02:58	RFI Detect
Air Blank	0.000	02:58	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: DEEL

DATE/TIME OF TEST: 04/18/2010 03:08:41

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample #1	0.150	03:04
Breath Volume	2.042	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample #2	0.155	03:07
Breath Volume	2.074	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEWITT, AMY, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DEWITT, AMY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: LAWTON

DATE/TIME OF TEST: 04/18/2010 19:55:03

Diagnostics Check	OK	19:52	
Air Blank	0.000	19:52	
Control Test	0.079	19:52	
Air Blank	0.000	19:53	
Subject Sample #1		19:53	Subject Test Refused
Breath Volume	0.000	19:53	
Air Blank	0.000	19:54	
Control Test	0.079	19:54	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:55	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON, MATTHEW, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: FORTIN, JASON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 04/19/2010 04:13:38

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:10	
Control Test	0.079	04:10	
Air Blank	0.000	04:10	
Subject Sample #1		04:12	RFI Detect
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Control Test	0.079	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON, MATTHEW, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: FORTIN, JASON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 04/19/2010 04:18:41

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.079	04:16	
Air Blank	0.000	04:17	
Subject Sample #1		04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.079	04:18	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER, MATTHEW, R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER, MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 04/19/2010 18:17:25

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:06	
Control Test	0.080	18:06	
Air Blank	0.000	18:07	
Subject Sample #1		18:10	No Sample Provided
Breath Volume	0.000	18:10	
Air Blank	0.000	18:10	
Air Blank	0.000	18:12	
Subject Sample #2		18:16	No Sample Provided
Breath Volume	0.000	18:16	
Air Blank	0.000	18:16	
Control Test	0.080	18:16	
Air Blank	0.000	18:17	
Diagnostics Check	OK	18:17	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TEST SUBJECT, TEST SUBJECT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: SYS CHECK
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: TEST SUBJECT

DATE/TIME OF TEST: 04/21/2010 14:21:59

Diagnostics Check	OK	14:18	
Air Blank	0.000	14:18	
Control Test	0.081	14:19	
Air Blank	0.000	14:19	
Subject Sample #1		14:20	Slope Not Met
Breath Volume	1.398	14:20	
Air Blank	0.000	14:21	
Control Test	0.080	14:21	
Air Blank		14:21	RFI Detect
Air Blank		14:21	RFI Detect

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TEST SUBJECT, TEST SUBJECT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: SYS CHECK
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: TEST SUBJECT

DATE/TIME OF TEST: 04/21/2010 14:26:57

Diagnostics Check	OK	14:23	
Air Blank	0.000	14:24	
Control Test	0.079	14:24	
Air Blank	0.000	14:24	
Subject Sample #1		14:25	Slope Not Met
Breath Volume	3.000	14:25	
Air Blank	0.000	14:26	
Control Test	0.080	14:26	
Air Blank	0.000	14:26	
Diagnostics Check	OK	14:26	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TEST SUBJECT, TEST SUBJECT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: SYS CHECK
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: TEST SUBJECT

DATE/TIME OF TEST: 04/21/2010 14:31:59

Diagnostics Check	OK	14:28	
Air Blank	0.000	14:28	
Control Test	0.080	14:29	
Air Blank	0.000	14:29	
Subject Sample #1		14:30	Range Exceeded
Breath Volume	1.687	14:30	
Air Blank	0.000	14:31	
Control Test	0.079	14:31	
Air Blank	0.000	14:31	
Diagnostics Check	OK	14:31	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: RODERICK

DATE/TIME OF TEST: 04/22/2010 03:26:03

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.080	03:24	
Air Blank	0.000	03:24	
Subject Sample #1		03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:26	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON, MATTHEW, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: RAMBARAN, NEIL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: MAHIN

DATE/TIME OF TEST: 04/22/2010 05:52:37

Diagnostics Check	OK	05:48	
Air Blank	0.000	05:49	
Control Test	0.080	05:49	
Air Blank	0.000	05:49	
Subject Sample #1		05:51	Subject Test Refused
Breath Volume	0.000	05:51	
Air Blank	0.000	05:51	
Control Test	0.081	05:52	
Air Blank	0.000	05:52	
Diagnostics Check	OK	05:52	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESHUR, MICHAEL, F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: CORBETT

DATE/TIME OF TEST: 04/22/2010 18:24:55

Diagnostics Check	OK	18:18
Air Blank	0.000	18:19
Control Test	0.081	18:19
Air Blank	0.000	18:20
Subject Sample #1	0.000	18:20
Breath Volume	2.765	18:20
Air Blank	0.000	18:21
Air Blank	0.000	18:23
Subject Sample #2	0.000	18:23
Breath Volume	1.925	18:23
Air Blank	0.000	18:24
Control Test	0.081	18:24
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 04/23/2010 02:09:37

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:04
Air Blank	0.000	02:04
Subject Sample #1	0.181	02:05
Breath Volume	2.535	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample #2	0.192	02:08
Breath Volume	3.054	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAIG, TAMMY, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CORDIVIOLA, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:42

SUBJECT'S LAST NAME: GRIMMER

DATE/TIME OF TEST: 04/23/2010 15:08:17

Diagnostics Check	OK	15:02	
Air Blank	0.000	15:03	
Control Test	0.080	15:03	
Air Blank	0.000	15:04	
Subject Sample #1		15:06	Subject Test Refused
Breath Volume	0.000	15:06	
Air Blank	0.000	15:07	
Control Test	0.080	15:07	
Air Blank	0.000	15:08	
Diagnostics Check	OK	15:08	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA, RICHARD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MIRANDA, RICHARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: KECK

DATE/TIME OF TEST: 04/24/2010 00:16:47

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.081	00:11
Air Blank	0.000	00:11
Subject Sample #1	0.000	00:12
Breath Volume	2.261	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample #2	0.000	00:15
Breath Volume	1.855	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA, RICHARD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: INNES, JOSH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 04/24/2010 03:05:44

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.080	03:03	
Air Blank	0.000	03:04	
Subject Sample #1		03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.080	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSNELL, JEFFREY, K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GOSNELL, JEFFREY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: GUESS

DATE/TIME OF TEST: 04/25/2010 00:24:21

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:19
Subject Sample #1	0.212	00:19
Breath Volume	1.402	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample #2	0.212	00:22
Breath Volume	2.074	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: BALLOS

DATE/TIME OF TEST: 04/25/2010 02:33:18

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Subject Sample #1		02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Control Test	0.079	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GOSNELL, JEFFREY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: HOLEMAN

DATE/TIME OF TEST: 04/25/2010 03:20:14

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.081	03:18	
Air Blank	0.000	03:18	
Subject Sample #1		03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.078	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUGLAS, JOHN, J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: INESS, JOSH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: FLEISCHER

DATE/TIME OF TEST: 04/25/2010 08:14:59

Diagnostics Check	OK	08:09
Air Blank	0.000	08:09
Control Test	0.081	08:09
Air Blank	0.000	08:10
Subject Sample #1	0.000	08:10
Breath Volume	1.402	08:10
Air Blank	0.000	08:11
Air Blank	0.000	08:13
Subject Sample #2	0.000	08:13
Breath Volume	3.609	08:13
Air Blank	0.000	08:14
Control Test	0.080	08:14
Air Blank	0.000	08:14
Diagnostics Check	OK	08:14

Data Download Date/Time 04/28/2010 19:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: LUDLOW

DATE/TIME OF TEST: 04/29/2010 01:14:19

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.082	01:12	
Air Blank	0.000	01:12	
Subject Sample #1		01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:13	
Control Test	0.082	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time 05/26/2010 15:11

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON, MATTHEW, L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 04/29/2010 02:51:44

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:49	
Control Test	0.081	02:49	
Air Blank	0.000	02:50	
Subject Sample #1		02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Control Test	0.081	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time 05/26/2010 15:11

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GARVIN

DATE/TIME OF TEST: 04/29/2010 20:23:11

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:20	
Control Test	0.081	20:21	
Air Blank	0.000	20:21	
Subject Sample #1		20:21	Subject Test Refused
Breath Volume	0.000	20:21	
Air Blank	0.000	20:22	
Control Test	0.081	20:22	
Air Blank	0.000	20:23	
Diagnostics Check	OK	20:23	

Data Download Date/Time 05/26/2010 15:11

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX, LENARD, E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLECKNER, MATT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/18/2011
VIOLATION CODE: BUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BRULINSKI

DATE/TIME OF TEST: 04/30/2010 00:56:20

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.082	00:50
Air Blank	0.000	00:51
Subject Sample #1	0.235	00:51
Breath Volume	3.503	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample #2	0.230	00:54
Breath Volume	2.609	00:54
Air Blank	0.000	00:55
Control Test	0.081	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time 05/26/2010 15:11

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALAMEDA, JORGE, L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALAMEDA, JORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 02:40

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: NEEDHAM

DATE/TIME OF TEST: 04/03/2010 03:34:05

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.082	03:28
Air Blank	0.000	03:29
Subject Sample #1	0.175	03:29
Breath Volume	2.511	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample #2	0.180	03:32
Breath Volume	2.507	03:32
Air Blank	0.000	03:33
Control Test	0.082	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALAMEDA, JORGE, L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALAMEDA, JORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 02:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 04/04/2010 03:16:06

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:10
Subject Sample #1	0.000	03:11
Breath Volume	2.949	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample #2	0.000	03:14
Breath Volume	1.152	03:14
Air Blank	0.000	03:15
Control Test	0.082	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILSON, CHRIS, R

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: VEGENSKI, MICHAEL

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 22:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 04/05/2010 22:55:07

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.082	22:46	
Air Blank	0.000	22:46	
Subject Sample #1	0.170	22:49	Volume Not Met
Breath Volume	0.035	22:49	
Air Blank	0.000	22:50	
Air Blank	0.000	22:52	
Subject Sample #2		22:53	Subject Test Refused
Breath Volume	0.097	22:53	
Air Blank	0.000	22:54	
Control Test	0.081	22:54	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUGHES, CORY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HUGHES, CORY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 21:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: HARRIGAN

DATE/TIME OF TEST: 04/06/2010 21:59:20

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:56	
Control Test	0.083	21:56	
Air Blank	0.000	21:57	
Subject Sample #1		21:57	Subject Test Refused
Breath Volume	0.000	21:57	
Air Blank	0.000	21:58	
Control Test	0.083	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CITRANO, ANTHONY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: CITRANO, ANTHONY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 01:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/08/2010 02:26:29

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.083	02:24	
Air Blank	0.000	02:24	
Subject Sample #1		02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Control Test	0.082	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CITRANO, ANTHONY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: CITRANO, ANTHONY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 21:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MCNALLY

DATE/TIME OF TEST: 04/08/2010 22:06:25

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.082	22:00
Air Blank	0.000	22:00
Subject Sample #1	0.040	22:01
Breath Volume	1.808	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample #2	0.039	22:04
Breath Volume	1.460	22:04
Air Blank	0.000	22:05
Control Test	0.083	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALAMEDA, JORGE, L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALAMEDA, JORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 18:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LONAS

DATE/TIME OF TEST: 04/09/2010 19:44:43

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:42	
Control Test	0.082	19:42	
Air Blank	0.000	19:43	
Subject Sample #1		19:43	Subject Test Refused
Breath Volume	0.000	19:43	
Air Blank	0.000	19:43	
Control Test	0.083	19:44	
Air Blank	0.000	19:44	
Diagnostics Check	OK	19:44	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALAMEDA, JORGE, L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALAMEDA, JORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 21:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: FISHBURN

DATE/TIME OF TEST: 04/09/2010 21:13:48

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.083	21:08
Air Blank	0.000	21:08
Subject Sample #1	0.000	21:09
Breath Volume	3.199	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample #2	0.000	21:12
Breath Volume	1.980	21:12
Air Blank	0.000	21:12
Control Test	0.081	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUGHES, CORY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HUGHES, CORY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 16:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/10/2010 17:07:57

Diagnostics Check	OK	16:58
Air Blank	0.000	16:58
Control Test	0.083	16:59
Air Blank	0.000	16:59
Subject Sample #1	0.332	17:00
Breath Volume	2.308	17:00
Air Blank	0.000	17:01
Air Blank	0.000	17:03
Subject Sample #2	0.306	17:03
Breath Volume	1.375	17:03
Air Blank	0.000	17:04
Air Blank	0.000	17:06
Subject Sample #3	0.316	17:06
Breath Volume	1.105	17:06
Air Blank	0.000	17:07
Control Test	0.081	17:07
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALAMEDA, JORGE, L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALAMEDA, JORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 04/11/2010 03:29:12

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.083	03:26	
Air Blank	0.000	03:27	
Subject Sample #1		03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.082	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALTERS, GEORGE, L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: WALTERS, GEORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 07:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: MACATUGOB

DATE/TIME OF TEST: 04/12/2010 08:34:38

Diagnostics Check	OK	08:28
Air Blank	0.000	08:29
Control Test	0.082	08:29
Air Blank	0.000	08:29
Subject Sample #1	0.000	08:30
Breath Volume	2.335	08:30
Air Blank	0.000	08:30
Air Blank	0.000	08:32
Subject Sample #2	0.000	08:33
Breath Volume	2.183	08:33
Air Blank	0.000	08:33
Control Test	0.082	08:34
Air Blank	0.000	08:34
Diagnostics Check	OK	08:34

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALAMEDA, JORGE, L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALAMEDA, JORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 00:49

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SMIRL

DATE/TIME OF TEST: 04/15/2010 02:17:22

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.083	02:14	
Air Blank	0.000	02:15	
Subject Sample #1		02:16	Subject Test Refused
Breath Volume	0.000	02:16	
Air Blank	0.000	02:16	
Control Test	0.082	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CITRANO, ANTHONY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: CITRANO, ANTHONY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DONNELLY

DATE/TIME OF TEST: 04/16/2010 02:01:31

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.082	01:48	
Air Blank	0.000	01:48	
Subject Sample #1	0.024	01:51	
Breath Volume	1.457	01:51	
Air Blank	0.000	01:52	
Air Blank	0.000	01:54	
Subject Sample #2	0.019	01:57	Volume Not Met
Breath Volume	0.539	01:57	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample #3		02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.082	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CITRANO, ANTHONY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: CITRANO, ANTHONY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 03:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: BUILDER

DATE/TIME OF TEST: 04/16/2010 04:39:30

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.083	04:33
Air Blank	0.000	04:34
Subject Sample #1	0.120	04:35
Breath Volume	1.859	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample #2	0.131	04:38
Breath Volume	1.625	04:38
Air Blank	0.000	04:38
Control Test	0.083	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUGHES, CORY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: WEDIN, KEITH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 17:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: MUTH

DATE/TIME OF TEST: 04/16/2010 18:18:35

Diagnostics Check	OK	18:11
Air Blank	0.000	18:12
Control Test	0.083	18:12
Air Blank	0.000	18:13
Subject Sample #1	0.057	18:13
Breath Volume	1.667	18:13
Air Blank	0.000	18:14
Air Blank	0.000	18:16
Subject Sample #2	0.059	18:17
Breath Volume	2.011	18:17
Air Blank	0.000	18:17
Control Test	0.083	18:17
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CITRANO, ANTHONY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: CITRANO, ANTONY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 18:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: BELCHER

DATE/TIME OF TEST: 04/17/2010 19:22:16

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:16	
Control Test	0.082	19:16	
Air Blank	0.000	19:16	
Subject Sample #1	0.165	19:17	
Breath Volume	1.378	19:17	
Air Blank	0.000	19:18	
Air Blank	0.000	19:20	
Subject Sample #2		19:20	Subject Test Refused
Breath Volume	0.000	19:20	
Air Blank	0.000	19:21	
Control Test	0.081	19:21	
Air Blank	0.000	19:22	
Diagnostics Check	OK	19:22	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RECLA, THOMAS,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LOMONACO, C
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831801I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: GILSON

DATE/TIME OF TEST: 04/18/2010 18:49:46

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.082	18:44
Air Blank	0.000	18:44
Subject Sample #1	0.294	18:45
Breath Volume	2.593	18:45
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample #2	0.299	18:48
Breath Volume	3.789	18:48
Air Blank	0.000	18:48
Control Test	0.078	18:49
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALAMEDA, JORGE, L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALAMEDA, JORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 750980

DRY GAS EXPIRY DATE: 04/07/2012

VIOLATION CODE: DUI

ARREST TIME: 22:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 04/21/2010 23:47:33

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:45	
Control Test	0.082	23:45	
Air Blank	0.000	23:46	
Subject Sample #1		23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Control Test	0.082	23:46	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CITRANO, ANTHONY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: CITRANO, ANTHONY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 750980

DRY GAS EXPIRY DATE: 04/07/2012

VIOLATION CODE: DUI

ARREST TIME: 02:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PERRELLA

DATE/TIME OF TEST: 04/23/2010 03:53:11

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.082	03:51	
Air Blank	0.000	03:51	
Subject Sample #1		03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.082	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALAMEDA, JORGE, L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALAMEDA, JORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 750980

DRY GAS EXPIRY DATE: 04/07/2012

VIOLATION CODE: DUI

ARREST TIME: 02:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MCKEAGUE

DATE/TIME OF TEST: 04/24/2010 03:26:49

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.082	03:24	
Air Blank	0.000	03:25	
Subject Sample #1		03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.082	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALAMEDA, JORGE, L

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALAMEDA, JORGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 750980

DRY GAS EXPIRY DATE: 04/07/2012

VIOLATION CODE: DUI

ARREST TIME: 23:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: COIT

DATE/TIME OF TEST: 04/25/2010 00:18:21

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.082	00:13
Air Blank	0.000	00:13
Subject Sample #1	0.000	00:14
Breath Volume	2.757	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample #2	0.000	00:16
Breath Volume	1.968	00:16
Air Blank	0.000	00:17
Control Test	0.082	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SOTO, ALBIN, E

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HERNANDEZ, BRUCE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 750980

DRY GAS EXPIRY DATE: 04/07/2012

VIOLATION CODE: DUI

ARREST TIME: 01:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/28/2010 02:49:20

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.082	02:43
Air Blank	0.000	02:44
Subject Sample #1	0.000	02:45
Breath Volume	1.496	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample #2	0.000	02:47
Breath Volume	1.789	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUGHES, CORY,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HUGHES, CORY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 750980

DRY GAS EXPIRY DATE: 04/07/2012

VIOLATION CODE: DUI

ARREST TIME: 17:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: MCCUNE

DATE/TIME OF TEST: 04/28/2010 17:47:43

Diagnostics Check	OK	17:44	
Air Blank	0.000	17:45	
Control Test	0.083	17:45	
Air Blank	0.000	17:46	
Subject Sample #1		17:46	Subject Test Refused
Breath Volume	0.000	17:46	
Air Blank	0.000	17:46	
Control Test	0.083	17:47	
Air Blank	0.000	17:47	
Diagnostics Check	OK	17:47	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES, CORY,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: VIEZI, ALBAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750980
DRY GAS EXPIRY DATE: 04/07/2012
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: O BRIEN

DATE/TIME OF TEST: 04/28/2010 18:33:53

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:31	
Control Test	0.082	18:31	
Air Blank	0.000	18:32	
Subject Sample #1		18:32	RFI Detect
Breath Volume	0.000	18:32	
Air Blank	0.000	18:32	
Control Test	0.082	18:33	
Air Blank	0.000	18:33	
Diagnostics Check	OK	18:33	

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES, CORY,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: VIEZI, ALBAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750980
DRY GAS EXPIRY DATE: 04/07/2012
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: O BRIEN

DATE/TIME OF TEST: 04/28/2010 18:41:35

Diagnostics Check	OK	18:35
Air Blank	0.000	18:35
Control Test	0.082	18:36
Air Blank	0.000	18:36
Subject Sample #1	0.234	18:37
Breath Volume	2.394	18:37
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample #2	0.236	18:40
Breath Volume	2.554	18:40
Air Blank	0.000	18:40
Control Test	0.081	18:40
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

Data Download Date/Time 05/04/2010 17:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH, PATTY, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DOYON, LUC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 04/01/2010 01:12:59

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.078	01:07
Air Blank	0.000	01:07
Subject Sample #1	0.155	01:08
Breath Volume	4.675	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample #2	0.152	01:11
Breath Volume	2.820	01:11
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE, ANTHONY, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALLAIRE, ANTHONY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: KEELEY

DATE/TIME OF TEST: 04/02/2010 02:59:11

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Subject Sample #1	0.269	02:52
Breath Volume	1.605	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample #2	0.268	02:57
Breath Volume	1.234	02:57
Air Blank	0.000	02:58
Control Test	0.076	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGG, DAVID,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OGG, DAVID
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: CALLIHAN

DATE/TIME OF TEST: 04/02/2010 04:54:17

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:46
Subject Sample #1	0.160	04:49
Breath Volume	1.121	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample #2	0.156	04:52
Breath Volume	1.347	04:52
Air Blank	0.000	04:53
Control Test	0.077	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN, DAVE, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DENMARK, HAROLD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: BEACH

DATE/TIME OF TEST: 04/02/2010 22:32:21

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample #1	0.229	22:27
Breath Volume	1.187	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample #2	0.225	22:30
Breath Volume	2.511	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLOWAN, WILLIAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAGER, KEVIN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ELLINGTON

DATE/TIME OF TEST: 04/02/2010 23:18:38

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:10
Subject Sample #1	0.181	23:11
Breath Volume	1.503	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample #2	0.184	23:17
Breath Volume	1.316	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULVERSON, REGINA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HULVERSON, REGINA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: BELLO

DATE/TIME OF TEST: 04/03/2010 00:53:48

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.079	00:48
Air Blank	0.000	00:49
Subject Sample #1	0.193	00:49
Breath Volume	1.464	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:51
Subject Sample #2	0.197	00:52
Breath Volume	1.519	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULVERSON, REGINA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOODBURY, DONALD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DANSER

DATE/TIME OF TEST: 04/03/2010 01:50:08

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Subject Sample #1	0.140	01:45
Breath Volume	2.703	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample #2	0.137	01:48
Breath Volume	2.441	01:48
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLIETT, JEREMY, L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CLIETT, JEREMY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/03/2010 02:19:42

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:13
Subject Sample #1	0.202	02:14
Breath Volume	1.910	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample #2	0.207	02:18
Breath Volume	3.089	02:18
Air Blank	0.000	02:18
Control Test	0.078	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULVERSON, REGINA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HULVERSON, REGINA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 04/03/2010 03:46:35

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Subject Sample #1	0.228	03:41
Breath Volume	2.058	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample #2	0.221	03:45
Breath Volume	1.964	03:45
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULVERSON, REGINA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HULVERSON, REGINA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LOPEZ-ALICEA

DATE/TIME OF TEST: 04/04/2010 03:30:47

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample #1	0.165	03:26
Breath Volume	1.890	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample #2	0.160	03:29
Breath Volume	2.007	03:29
Air Blank	0.000	03:29
Control Test	0.078	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS, DERYCK, P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HAGER, KEVIN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CASTELLANOS MEJIA

DATE/TIME OF TEST: 04/05/2010 02:33:14

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Subject Sample #1	0.000	02:28
Breath Volume	2.507	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample #2	0.000	02:31
Breath Volume	1.820	02:31
Air Blank	0.000	02:32
Control Test	0.077	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH, PATTY, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MARMANIDES, JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 04/05/2010 22:39:33

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.079	22:37	
Air Blank	0.000	22:38	
Subject Sample #1		22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.079	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE, JAMES, M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAGAR, KEVIN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LEMMA

DATE/TIME OF TEST: 04/09/2010 05:32:15

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:29	
Control Test	0.079	05:30	
Air Blank	0.000	05:30	
Subject Sample #1		05:30	Subject Test Refused
Breath Volume	0.000	05:30	
Air Blank	0.000	05:31	
Control Test	0.080	05:31	
Air Blank	0.000	05:32	
Diagnostics Check	OK	05:32	

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE, ANTHONY, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALLAIRE, ANTHONY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/10/2010 03:47:40

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:40
Subject Sample #1	0.242	03:41
Breath Volume	2.230	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample #2	0.250	03:46
Breath Volume	2.421	03:46
Air Blank	0.000	03:46
Control Test	0.079	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTALBANO, JONATHAN, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GRIFFIN, DAVID
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CLAY

DATE/TIME OF TEST: 04/10/2010 04:22:44

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Subject Sample #1	0.215	04:18
Breath Volume	1.804	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample #2	0.208	04:21
Breath Volume	2.753	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH, PATTY, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: QUINN, PATRICK
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 04/11/2010 21:29:14

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Subject Sample #1	0.264	21:24
Breath Volume	1.429	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample #2	0.262	21:27
Breath Volume	1.890	21:27
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN, DAVE, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RICE, WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 04/13/2010 01:34:44

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:29
Air Blank	0.000	01:29
Subject Sample #1	0.106	01:30
Breath Volume	1.187	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample #2	0.108	01:33
Breath Volume	1.855	01:33
Air Blank	0.000	01:33
Control Test	0.078	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN, DAVE, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAGER, KEVIN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: TEMPLE

DATE/TIME OF TEST: 04/13/2010 22:44:53

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.079	22:42	
Air Blank	0.000	22:43	
Subject Sample #1		22:43	Subject Test Refused
Breath Volume	0.000	22:43	
Air Blank	0.000	22:43	
Control Test	0.079	22:44	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLOMAN, WILLIAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOODBERRY, DONALD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/17/2010 04:13:16

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Subject Sample #1	0.170	04:08
Breath Volume	3.074	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample #2	0.164	04:11
Breath Volume	3.027	04:11
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLOMAN, WILLIAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMSON, DIRK
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/17/2010 23:05:35

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:03	
Control Test	0.079	23:03	
Air Blank	0.000	23:03	
Subject Sample #1		23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:04	
Control Test	0.079	23:04	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULVERSON, REGINA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RICE, WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 04/18/2010 03:27:23

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Subject Sample #1		03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULVERSON, REGINA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GORECK, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BARDELE

DATE/TIME OF TEST: 04/18/2010 03:57:58

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.079	03:52	
Air Blank	0.000	03:53	
Subject Sample #1	0.129	03:54	
Breath Volume	1.476	03:54	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample #2	0.138	03:56	
Breath Volume	1.328	03:56	
Air Blank		03:57	RFI Detect
Air Blank	0.000	03:57	

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULVERSON, REGINA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GORECK, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BARDELE

DATE/TIME OF TEST: 04/18/2010 04:06:01

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Subject Sample #1	0.132	04:01
Breath Volume	1.222	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample #2	0.131	04:04
Breath Volume	1.531	04:04
Air Blank	0.000	04:05
Control Test	0.078	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:06

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS, ROBERT, D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WILLIAMSON, DIRK
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 04/18/2010 05:00:53

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.079	05:00	
Air Blank		05:00	RFI Detect
Air Blank	0.000	05:00	

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS, ROBERT, D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WILLIAMSON, DIRK
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 04/18/2010 05:09:23

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.078	05:03
Air Blank	0.000	05:04
Subject Sample #1	0.161	05:04
Breath Volume	1.136	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample #2	0.159	05:07
Breath Volume	2.378	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLOWAN, WILLIAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAGER, KEVIN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 04/19/2010 00:41:05

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Subject Sample #1	0.000	00:36
Breath Volume	3.582	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample #2	0.000	00:39
Breath Volume	3.589	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:41

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULVERSON, REGINA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HULVERSON, REGINA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609611
DRY GAS EXPIRY DATE: 04/27/2011
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: KAHN

DATE/TIME OF TEST: 04/23/2010 20:52:02

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.083	20:46
Air Blank	0.000	20:47
Subject Sample #1	0.000	20:47
Breath Volume	2.519	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample #2	0.000	20:50
Breath Volume	2.640	20:50
Air Blank	0.000	20:51
Control Test	0.083	20:51
Air Blank	0.000	20:51
Diagnostics Check	OK	20:52

Data Download Date/Time 04/26/2010 10:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WISE, DENISE, L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOODBURY, D
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 04/26/2010 22:30:23

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:25
Subject Sample #1	0.218	22:25
Breath Volume	2.183	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample #2	0.233	22:28
Breath Volume	2.894	22:28
Air Blank	0.000	22:29
Control Test	0.079	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time 05/19/2010 11:22

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULVERSON, REGINA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HULVERSON, REGINA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RADLOFF

DATE/TIME OF TEST: 04/27/2010 00:37:47

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:29
Subject Sample #1	0.123	00:31
Breath Volume	1.156	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample #2	0.119	00:36
Breath Volume	1.304	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time 05/19/2010 11:22

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANEY, AMY, L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RICE, W
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: JANSEN

DATE/TIME OF TEST: 04/27/2010 21:22:35

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.079	21:16
Air Blank	0.000	21:17
Subject Sample #1	0.175	21:18
Breath Volume	2.082	21:18
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample #2	0.178	21:21
Breath Volume	1.757	21:21
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time 05/19/2010 11:22

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANEY, AMY, L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RICE, W
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: VALBUENA

DATE/TIME OF TEST: 04/28/2010 03:19:42

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Subject Sample #1	0.140	03:15
Breath Volume	1.226	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample #2	0.144	03:18
Breath Volume	1.207	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time 05/19/2010 11:22

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO, IONI,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: EDISON, MICHAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: DEASE

DATE/TIME OF TEST: 04/28/2010 23:38:10

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.080	23:36	
Air Blank	0.000	23:36	
Subject Sample #1		23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time 05/19/2010 11:22

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO, IONI,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SNELL, JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: REYES-MARTINEZ

DATE/TIME OF TEST: 04/30/2010 00:32:07

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample #1	0.181	00:27
Breath Volume	1.449	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample #2	0.195	00:30
Breath Volume	1.589	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time 05/19/2010 11:22

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGG, DAVID,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OGG, DAVID
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 603632
DRY GAS EXPIRY DATE: 04/13/2011
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/30/2010 03:49:19

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample #1	0.208	03:44
Breath Volume	1.324	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample #2	0.216	03:47
Breath Volume	2.839	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time 05/19/2010 11:22

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK, SCOTT, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHUCK, SCOTT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: STEIER

DATE/TIME OF TEST: 04/02/2010 20:52:01

Diagnostics Check	Fail	20:51	Voltage/Current Test
Air Blank	0.000	20:52	

Data Download Date/Time	04/13/2010	13:41
-------------------------	------------	-------

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK, SCOTT, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHUCK, SCOTT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: STEIER

DATE/TIME OF TEST: 04/02/2010 21:00:12

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:55
Subject Sample #1	0.000	20:55
Breath Volume	2.281	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample #2	0.000	20:58
Breath Volume	2.355	20:58
Air Blank	0.000	20:59
Control Test	0.080	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time 04/13/2010 13:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI, ADAM, D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: METZLER, SHANE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MILOSCIA

DATE/TIME OF TEST: 04/03/2010 01:59:08

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Subject Sample #1	0.037	01:54
Breath Volume	2.226	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample #2	0.038	01:57
Breath Volume	2.488	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time 04/13/2010 13:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK, SCOTT, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHUCK, SCOTT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/08/2010 00:09:15

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.079	00:05	
Air Blank	0.000	00:06	
Subject Sample #1		00:07	Subject Test Refused
Breath Volume	0.371	00:07	
Air Blank	0.000	00:08	
Control Test	0.080	00:08	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time 04/13/2010 13:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON, CHRISTOPHER, S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: DENTON, CHRISTOPHER
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 04/08/2010 02:38:46

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:33
Air Blank	0.000	02:33
Subject Sample #1	0.185	02:34
Breath Volume	1.761	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample #2	0.187	02:37
Breath Volume	2.707	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time 04/13/2010 13:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR, ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MONSEGUR, ROBERT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: HALLORAN

DATE/TIME OF TEST: 04/09/2010 02:14:23

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Subject Sample #1	0.034	02:09
Breath Volume	3.921	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample #2	0.035	02:12
Breath Volume	4.304	02:12
Air Blank	0.000	02:13
Control Test	0.078	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time 04/13/2010 13:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK, SCOTT, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHUCK, SCOTT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 04/10/2010 02:31:43

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.079	02:23	
Air Blank	0.000	02:24	
Subject Sample #1		02:27	No Sample Provided
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample #2		02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time 04/13/2010 13:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI, ADAM, D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI, ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LINE

DATE/TIME OF TEST: 04/10/2010 03:22:51

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Subject Sample #1	0.211	03:18
Breath Volume	3.652	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample #2	0.208	03:21
Breath Volume	3.113	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time 04/13/2010 13:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK, SCOTT, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHUCK, SCOTT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: NISSLEY

DATE/TIME OF TEST: 04/11/2010 03:23:34

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:18
Subject Sample #1	0.134	03:18
Breath Volume	2.656	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample #2	0.136	03:22
Breath Volume	2.679	03:22
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time 04/13/2010 13:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI, ADAM, D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI, ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: JVODIC

DATE/TIME OF TEST: 04/11/2010 03:43:32

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample #1	0.000	03:39
Breath Volume	1.636	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample #2	0.000	03:42
Breath Volume	1.304	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time 04/13/2010 13:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK, SCOTT, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHUCK, SCOTT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 04/17/2010 02:12:36

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:06
Subject Sample #1	0.186	02:07
Breath Volume	2.160	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample #2	0.184	02:11
Breath Volume	2.015	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time 05/11/2010 14:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI, ADAM, D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI, ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: TASHKANDI

DATE/TIME OF TEST: 04/18/2010 03:07:00

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample #1	0.274	02:59
Breath Volume	1.636	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample #2	0.260	03:05
Breath Volume	1.316	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:07

Data Download Date/Time 05/11/2010 14:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METZLER, SHANE, A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: METZLER, SHANE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FELDHEIM

DATE/TIME OF TEST: 04/18/2010 03:42:43

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Subject Sample #1	0.000	03:38
Breath Volume	3.800	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample #2	0.000	03:41
Breath Volume	3.714	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time 05/11/2010 14:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METZLER, SHANE, A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: METZLER, SHANE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 04/22/2010 22:30:29

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.078	22:24
Air Blank	0.000	22:25
Subject Sample #1	0.016	22:26
Breath Volume	2.515	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample #2	0.014	22:29
Breath Volume	2.359	22:29
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time 05/11/2010 14:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI, ADAM, D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: METZLER, SHANE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: BOYLAN

DATE/TIME OF TEST: 04/24/2010 23:09:01

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Subject Sample #1	0.015	23:04
Breath Volume	2.449	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample #2	0.013	23:07
Breath Volume	1.714	23:07
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:09

Data Download Date/Time 05/11/2010 14:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR, ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MONSEGUR, ROBERT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/13/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HESLER

DATE/TIME OF TEST: 04/30/2010 23:04:06

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:58
Subject Sample #1	0.205	22:59
Breath Volume	1.539	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample #2	0.208	23:02
Breath Volume	2.406	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time 05/11/2010 14:23

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MENDEZ-MILLER, JOSEPHINE, M

BREATH TEST OPERATOR AGENCY: Plantation PD

ARRESTING OFFICER: BEAREGARD, J

ARRESTING OFFICER AGENCY: Plantation PD

LAST AGENCY INSPECTION: 03/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 20:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: AGUDELO

DATE/TIME OF TEST: 04/11/2010 21:21:45

Diagnostics Check	OK	21:21	
Air Blank		21:21	RFI Detect
Air Blank	0.000	21:21	

Data Download Date/Time	04/20/2010	20:08
-------------------------	------------	-------

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MENDEZ-MILLER, JOSEPHINE, M

BREATH TEST OPERATOR AGENCY: Plantation PD

ARRESTING OFFICER: BEAREGARD, J

ARRESTING OFFICER AGENCY: Plantation PD

LAST AGENCY INSPECTION: 03/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 20:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: AGUDELO

DATE/TIME OF TEST: 04/11/2010 21:29:57

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.081	21:24
Air Blank	0.000	21:25
Subject Sample #1	0.121	21:25
Breath Volume	1.957	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample #2	0.122	21:28
Breath Volume	1.464	21:28
Air Blank	0.000	21:29
Control Test	0.081	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time 04/20/2010 20:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MENDEZ-MILLER, JOSEPHINE, M

BREATH TEST OPERATOR AGENCY: Plantation PD

ARRESTING OFFICER: PASCARELLA, T

ARRESTING OFFICER AGENCY: Davie PD

LAST AGENCY INSPECTION: 03/23/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 15:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: BOOTS

DATE/TIME OF TEST: 04/16/2010 16:18:13

Diagnostics Check	OK	16:06	
Air Blank	0.000	16:06	
Control Test	0.080	16:07	
Air Blank	0.000	16:07	
Subject Sample #1	0.116	16:08	No .020 agreement
Breath Volume	1.363	16:08	
Air Blank	0.000	16:08	
Air Blank	0.000	16:10	
Subject Sample #2	0.140	16:10	No .020 agreement
Breath Volume	1.433	16:10	
Air Blank	0.000	16:11	
Air Blank	0.000	16:13	
Subject Sample #3	0.138	16:16	Volume Not Met
Breath Volume	0.691	16:16	
Air Blank	0.000	16:17	
Control Test	0.080	16:17	
Air Blank	0.000	16:18	
Diagnostics Check	OK	16:18	

Data Download Date/Time 04/20/2010 20:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUAN, OFC,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: MASON, OFC
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 03/23/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/21/2011
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PORCARO

DATE/TIME OF TEST: 04/17/2010 00:54:55

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:53	
Subject Sample #1		00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:54	
Control Test	0.080	00:54	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time 04/20/2010 20:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAILEY, B, A
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BEAUREGARD, J
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 03/23/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/21/2011
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: LAUNDREE

DATE/TIME OF TEST: 04/19/2010 18:16:52

Diagnostics Check	OK	18:14	
Air Blank	0.000	18:14	
Control Test	0.080	18:14	
Air Blank	0.000	18:15	
Subject Sample #1		18:15	RFI Detect
Breath Volume	0.000	18:15	
Air Blank	0.000	18:15	
Control Test	0.080	18:16	
Air Blank	0.000	18:16	
Diagnostics Check	OK	18:16	

Data Download Date/Time 04/20/2010 20:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAILEY, B, A
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BEAUREGARD, J
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 03/23/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 08/21/2011
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: LAUNDREE

DATE/TIME OF TEST: 04/19/2010 18:21:44

Diagnostics Check	OK	18:18	
Air Blank	0.000	18:19	
Control Test	0.080	18:19	
Air Blank	0.000	18:20	
Subject Sample #1		18:20	Subject Test Refused
Breath Volume	0.000	18:20	
Air Blank	0.000	18:20	
Control Test	0.080	18:21	
Air Blank	0.000	18:21	
Diagnostics Check	OK	18:21	

Data Download Date/Time 04/20/2010 20:08

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROUTON, KYLE, P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STAVRO, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 812301G
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: MURICKEN

DATE/TIME OF TEST: 04/04/2010 20:04:59

Diagnostics Check	OK	20:00	
Air Blank	0.000	20:00	
Control Test	0.075	20:00	
Air Blank	0.000	20:01	
Subject Sample #1		20:03	RFI Detect
Breath Volume	0.675	20:03	
Air Blank	0.000	20:04	
Control Test	0.077	20:04	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time 04/30/2010 12:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROUTON, KYLE, P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STAVRO, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 812301G
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: MURICKEN

DATE/TIME OF TEST: 04/04/2010 20:14:46

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.075	20:07	
Air Blank	0.000	20:07	
Subject Sample #1	0.261	20:08	
Breath Volume	1.507	20:08	
Air Blank	0.000	20:08	
Air Blank	0.000	20:10	
Subject Sample #2		20:13	Subject Test Refused
Breath Volume	0.257	20:13	
Air Blank	0.000	20:13	
Control Test	0.076	20:14	
Air Blank	0.000	20:14	
Diagnostics Check	OK	20:14	

Data Download Date/Time 04/30/2010 12:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TROUTON, KYLE, P

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: STAVRO, J

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 812301G

DRY GAS EXPIRY DATE: 05/06/2010

VIOLATION CODE: DUI

ARREST TIME: 19:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 04/05/2010 21:01:29

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.075	20:56
Air Blank	0.000	20:56
Subject Sample #1	0.000	20:57
Breath Volume	1.253	20:57
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample #2	0.000	21:00
Breath Volume	1.125	21:00
Air Blank	0.000	21:00
Control Test	0.077	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time 04/30/2010 12:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MILLER, G
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 812301G
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: PRENTZLER

DATE/TIME OF TEST: 04/16/2010 22:33:55

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:27	
Control Test	0.076	22:27	
Air Blank	0.000	22:28	
Subject Sample #1	0.271	22:29	
Breath Volume	2.007	22:29	
Air Blank	0.000	22:29	
Air Blank	0.000	22:31	
Subject Sample #2	0.260	22:32	
Breath Volume	1.890	22:32	
Air Blank	0.000	22:33	
Control Test	0.074	22:33	Control Outside Tolerance
Air Blank	0.000	22:33	

Data Download Date/Time 04/30/2010 12:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ABARE, T
ARRESTING OFFICER AGENCY: FDOT MCC
LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 812301G
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 04/16/2010 22:49:05

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.076	22:46	
Air Blank	0.000	22:47	
Subject Sample #1		22:47	Subject Test Refused
Breath Volume	0.000	22:47	
Air Blank	0.000	22:48	
Control Test	0.075	22:48	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:49	

Data Download Date/Time 04/30/2010 12:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JACKSON, GREG, P

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: JACKSON, GREG

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 823901G

DRY GAS EXPIRY DATE: 08/29/2010

VIOLATION CODE: DUI

ARREST TIME: 16:09

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: DUMAS

DATE/TIME OF TEST: 04/01/2010 18:10:58

Diagnostics Check	OK	18:04
Air Blank	0.000	18:05
Control Test	0.079	18:05
Air Blank	0.000	18:05
Subject Sample #1	0.199	18:06
Breath Volume	1.207	18:06
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample #2	0.198	18:09
Breath Volume	1.441	18:09
Air Blank	0.000	18:10
Control Test	0.080	18:10
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time 04/29/2010 12:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MATTINGLEY, MARK, P

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: MATTINGLEY, MARK

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 823901G

DRY GAS EXPIRY DATE: 08/29/2010

VIOLATION CODE: DUI

ARREST TIME: 04:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: DILSAVER

DATE/TIME OF TEST: 04/05/2010 05:12:20

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.081	05:06
Air Blank	0.000	05:07
Subject Sample #1	0.082	05:07
Breath Volume	3.281	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample #2	0.077	05:10
Breath Volume	2.449	05:10
Air Blank	0.000	05:11
Control Test	0.080	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time 04/29/2010 12:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BAILEY, RAYMOND, F

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: BAILEY, RAYMOND

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 823901G

DRY GAS EXPIRY DATE: 08/29/2010

VIOLATION CODE: DUI

ARREST TIME: 17:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: YOUNGBLOOD

DATE/TIME OF TEST: 04/07/2010 18:27:01

Diagnostics Check	OK	18:20
Air Blank	0.000	18:20
Control Test	0.079	18:21
Air Blank	0.000	18:21
Subject Sample #1	0.132	18:22
Breath Volume	4.238	18:22
Air Blank	0.000	18:23
Air Blank	0.000	18:24
Subject Sample #2	0.141	18:25
Breath Volume	3.566	18:25
Air Blank	0.000	18:26
Control Test	0.080	18:26
Air Blank	0.000	18:26
Diagnostics Check	OK	18:27

Data Download Date/Time 04/29/2010 12:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WAGNER, BRETT, M

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: WAGNER, BRETT

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 823901G

DRY GAS EXPIRY DATE: 08/29/2010

VIOLATION CODE: DUI

ARREST TIME: 03:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: SPRINGER

DATE/TIME OF TEST: 04/08/2010 06:23:59

Diagnostics Check	OK	06:19	
Air Blank	0.000	06:20	
Control Test	0.080	06:20	
Air Blank	0.000	06:21	
Subject Sample #1		06:22	Subject Test Refused
Breath Volume	0.000	06:22	
Air Blank	0.000	06:23	
Control Test	0.080	06:23	
Air Blank	0.000	06:23	
Diagnostics Check	OK	06:23	

Data Download Date/Time 04/29/2010 12:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTINGLEY, MARK, P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MATTINGLEY, MARK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 823901G
DRY GAS EXPIRY DATE: 08/29/2010
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: EDMONSON

DATE/TIME OF TEST: 04/10/2010 00:07:16

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:01
Subject Sample #1	0.072	00:02
Breath Volume	1.757	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample #2	0.069	00:05
Breath Volume	2.980	00:05
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time 04/29/2010 12:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MATHERS, WILLIAM, A

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: MATHERS, WILLIAM

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 823901G

DRY GAS EXPIRY DATE: 08/29/2010

VIOLATION CODE: DUI

ARREST TIME: 01:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/15/2010 02:08:23

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Subject Sample #1	0.182	02:03
Breath Volume	1.660	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample #2	0.179	02:06
Breath Volume	2.503	02:06
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time 04/29/2010 12:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WAGNER, BRETT, M

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: WAGNER, BRETT

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 823901G

DRY GAS EXPIRY DATE: 08/29/2010

VIOLATION CODE: DUI

ARREST TIME: 03:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 04/24/2010 04:26:21

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.080	04:18	
Air Blank	0.000	04:19	
Subject Sample #1		04:22	No Sample Provided
Breath Volume	0.000	04:22	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample #2		04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time 04/29/2010 12:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCNABB, KENNETH, E

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: MCNABB, KENNETH

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 823901G

DRY GAS EXPIRY DATE: 08/29/2010

VIOLATION CODE: DUI

ARREST TIME: 15:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/26/2010 18:00:58

Diagnostics Check	OK	17:52
Air Blank	0.000	17:52
Control Test	0.079	17:52
Air Blank	0.000	17:53
Subject Sample #1	0.341	17:55
Breath Volume	1.324	17:55
Air Blank	0.000	17:56
Air Blank	0.000	17:58
Subject Sample #2	0.343	17:59
Breath Volume	2.269	17:59
Air Blank	0.000	18:00
Control Test	0.079	18:00
Air Blank	0.000	18:00
Diagnostics Check	OK	18:00

Data Download Date/Time 04/29/2010 12:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WAGNER, BRETT, M

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: WAGNER, BRETT

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 04/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 823901G

DRY GAS EXPIRY DATE: 08/29/2010

VIOLATION CODE: DUI

ARREST TIME: 02:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: NUMMY

DATE/TIME OF TEST: 04/30/2010 03:28:05

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:19	
Air Blank	0.000	03:19	
Subject Sample #1	0.163	03:22	Volume Not Met
Breath Volume	0.550	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample #2		03:26	Subject Test Refused
Breath Volume	0.527	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time 05/30/2010 11:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FHP TROOP D - COCOA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GEORGE, JACKI, L

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: CARPENTER, RODGER

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 03/08/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 585711

DRY GAS EXPIRY DATE: 03/16/2011

VIOLATION CODE: DUI

ARREST TIME: 21:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GLAAB

DATE/TIME OF TEST: 04/06/2010 22:23:06

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.077	22:17
Air Blank	0.000	22:17
Subject Sample #1	0.176	22:18
Breath Volume	3.316	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample #2	0.185	22:21
Breath Volume	3.496	22:21
Air Blank	0.000	22:22
Control Test	0.077	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time 04/10/2010 10:47

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FHP TROOP D - COCOA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DONKER, MICHAEL, R

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: BARROW, ERIC

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 03/08/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 585711

DRY GAS EXPIRY DATE: 03/16/2011

VIOLATION CODE: DUI

ARREST TIME: 15:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: BOUTTE

DATE/TIME OF TEST: 04/09/2010 16:51:57

Diagnostics Check	OK	16:41	
Air Blank	0.000	16:42	
Control Test	0.078	16:42	
Air Blank	0.000	16:42	
Subject Sample #1	0.196	16:46	Volume Not Met
Breath Volume	0.203	16:46	
Air Blank	0.000	16:46	
Air Blank	0.000	16:48	
Subject Sample #2		16:50	Subject Test Refused
Breath Volume	0.097	16:50	
Air Blank	0.000	16:51	
Control Test	0.076	16:51	
Air Blank	0.000	16:51	
Diagnostics Check	OK	16:51	

Data Download Date/Time 04/10/2010 10:47

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FHP TROOP D - COCOA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEELY, STEVE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HILDRETH, KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 585711
DRY GAS EXPIRY DATE: 03/16/2011
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 04/25/2010 04:42:18

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.076	04:36
Air Blank	0.000	04:36
Subject Sample #1	0.194	04:37
Breath Volume	2.218	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample #2	0.194	04:40
Breath Volume	1.476	04:40
Air Blank	0.000	04:41
Control Test	0.077	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time 05/07/2010 11:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FHP TROOP H	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 643837
BREATH TEST OPERATOR: YARBOROUGH, CARLTON, B	DRY GAS EXPIRY DATE: 07/17/2011
BREATH TEST OPERATOR AGENCY: FHP Troop H	VIOLATION CODE: DUI
ARRESTING OFFICER: YARBOROUGH, CARLTON	ARREST TIME: 20:21
ARRESTING OFFICER AGENCY: FHP Troop H	GENDER: Female
LAST AGENCY INSPECTION: 04/01/2010	OBSERVATION PERIOD BEGAN: 20:59
SUBJECT'S LAST NAME: FLEMING	DATE/TIME OF TEST: 04/30/2010 21:34:48

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.078	21:29
Air Blank	0.000	21:29
Subject Sample #1	0.097	21:30
Breath Volume	1.964	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample #2	0.091	21:33
Breath Volume	1.980	21:33
Air Blank	0.000	21:33
Control Test	0.078	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time 05/05/2010 13:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NAVARRO, CHRISTIAN, O

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: NAVARRO, CHRIS

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 102597

DRY GAS EXPIRY DATE: 02/28/2012

VIOLATION CODE: DUI

ARREST TIME: 20:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:58

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 04/06/2010 21:05:45

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.077	21:00
Air Blank	0.000	21:01
Subject Sample #1	0.000	21:01
Breath Volume	2.062	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample #2	0.000	21:04
Breath Volume	2.386	21:04
Air Blank	0.000	21:04
Control Test	0.077	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time 04/19/2010 02:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NAVARRO, CHRISTIAN, O

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: ROSS, LAKEISHA

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 102597

DRY GAS EXPIRY DATE: 02/28/2012

VIOLATION CODE: DUI

ARREST TIME: 20:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: AMUSO

DATE/TIME OF TEST: 04/06/2010 21:15:54

Diagnostics Check	OK	21:09	
Air Blank	0.000	21:10	
Control Test	0.076	21:10	
Air Blank	0.000	21:10	
Subject Sample #1	0.254	21:11	
Breath Volume	5.113	21:11	
Air Blank	0.000	21:12	
Air Blank	0.000	21:13	
Subject Sample #2	0.258	21:14	
Breath Volume	5.042	21:14	
Air Blank	0.000	21:15	
Control Test	0.073	21:15	Control Outside Tolerance
Air Blank	0.000	21:15	

Data Download Date/Time 04/19/2010 02:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BONHAM, HOWARD, C

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: RUHBUSCH, PAUL

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 03/16/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 102592

DRY GAS EXPIRY DATE: 02/28/2012

VIOLATION CODE: DUI

ARREST TIME: 08:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: FUGITT

DATE/TIME OF TEST: 04/09/2010 09:04:39

Diagnostics Check	OK	08:58
Air Blank	0.000	08:58
Control Test	0.077	08:58
Air Blank	0.000	08:59
Subject Sample #1	0.000	09:00
Breath Volume	1.476	09:00
Air Blank	0.000	09:00
Air Blank	0.000	09:02
Subject Sample #2	0.000	09:03
Breath Volume	1.980	09:03
Air Blank	0.000	09:03
Control Test	0.077	09:04
Air Blank	0.000	09:04
Diagnostics Check	OK	09:04

Data Download Date/Time 04/19/2010 02:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, JASON, C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS, JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/03/2010 04:02:09

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample #1	0.131	03:57
Breath Volume	1.753	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample #2	0.135	04:00
Breath Volume	2.707	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time 04/06/2010 15:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, JASON, C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS, JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: VANDERBRINK

DATE/TIME OF TEST: 04/04/2010 04:11:59

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Subject Sample #1	0.036	04:07
Breath Volume	2.445	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.033	04:10
Breath Volume	1.398	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time 04/06/2010 15:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKELVIN, R, G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCKELVIN, RONALD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CARRILLO

DATE/TIME OF TEST: 04/11/2010 20:31:21

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:28	
Control Test	0.079	20:28	
Air Blank	0.000	20:28	
Subject Sample #1		20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Control Test	0.079	20:30	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT, ANDREW, M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BERGSTRESSER, PETER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: CABELLO

DATE/TIME OF TEST: 04/11/2010 20:49:50

Diagnostics Check	OK	20:42	
Air Blank	0.000	20:42	
Control Test	0.079	20:42	
Air Blank	0.000	20:43	
Subject Sample #1	0.051	20:44	
Breath Volume	1.343	20:44	
Air Blank	0.000	20:44	
Air Blank	0.000	20:46	
Subject Sample #2		20:48	Subject Test Refused
Breath Volume	0.695	20:48	
Air Blank	0.000	20:48	
Control Test	0.079	20:49	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT, ANDREW, M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BERGSTRESSER, PETER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: CABELLO

DATE/TIME OF TEST: 04/11/2010 21:20:14

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:15
Subject Sample #1	0.044	21:15
Breath Volume	1.757	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample #2	0.045	21:18
Breath Volume	1.558	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, JASON, C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS, JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: COSSUM

DATE/TIME OF TEST: 04/12/2010 04:03:36

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.078	03:58
Air Blank	0.000	03:58
Subject Sample #1	0.190	03:59
Breath Volume	3.914	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample #2	0.189	04:02
Breath Volume	3.984	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, JASON, C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS, JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: KILLINGBECK

DATE/TIME OF TEST: 04/16/2010 02:30:40

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Subject Sample #1	0.089	02:26
Breath Volume	2.253	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample #2	0.089	02:29
Breath Volume	1.652	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, JASON, C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS, JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: KONIECZKI

DATE/TIME OF TEST: 04/18/2010 02:58:17

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:53
Subject Sample #1	0.160	02:53
Breath Volume	3.484	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.165	02:56
Breath Volume	3.445	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, JASON, C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS, JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: SHADICK

DATE/TIME OF TEST: 04/20/2010 04:18:05

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.077	04:06	
Air Blank	0.000	04:06	
Subject Sample #1	0.146	04:07	No .020 agreement
Breath Volume	1.730	04:07	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample #2	0.124	04:11	No .020 agreement
Breath Volume	2.277	04:11	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample #3		04:16	No Sample Provided
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.078	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKELVIN, R, G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCKELVIN, R
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/23/2010 17:47:42

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.077	17:41
Air Blank	0.000	17:41
Subject Sample #1	0.265	17:42
Breath Volume	2.285	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample #2	0.258	17:46
Breath Volume	2.574	17:46
Air Blank	0.000	17:46
Control Test	0.078	17:47
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, JASON, C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS, JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PFLUEGER

DATE/TIME OF TEST: 04/24/2010 03:20:09

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:13
Subject Sample #1	0.059	03:15
Breath Volume	2.359	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample #2	0.064	03:18
Breath Volume	4.117	03:18
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, JASON, C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS, JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JEANTY

DATE/TIME OF TEST: 04/25/2010 02:27:08

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.078	02:10	
Air Blank	0.000	02:11	
Subject Sample #1	0.056	02:14	
Breath Volume	1.542	02:14	
Air Blank	0.000	02:15	
Air Blank	0.000	02:16	
Subject Sample #2		02:20	No Sample Provided
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample #3	0.056	02:25	
Breath Volume	1.382	02:25	
Air Blank	0.000	02:26	
Control Test	0.079	02:26	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FHP TROOP D - DELAND
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER, RICKIE, R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER, RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 811301I
DRY GAS EXPIRY DATE: 05/06/2010
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 04/29/2010 04:57:37

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:52
Air Blank	0.000	04:52
Subject Sample #1	0.218	04:53
Breath Volume	2.527	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample #2	0.207	04:56
Breath Volume	1.382	04:56
Air Blank	0.000	04:56
Control Test	0.079	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time 05/09/2010 17:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALVAREZ, NORGE, C

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: CASILLAS, DAVID

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 22:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 04/03/2010 00:09:27

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:04
Subject Sample #1	0.161	00:04
Breath Volume	3.472	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample #2	0.151	00:07
Breath Volume	2.234	00:07
Air Blank	0.000	00:08
Control Test	0.077	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time 04/30/2010 20:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ADDISON, RAYMOND, E

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: MILNES, KENNETH

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PASOS RICARTE

DATE/TIME OF TEST: 04/11/2010 03:00:41

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.080	02:57	
Air Blank	0.000	02:58	
Subject Sample #1		02:59	Slope Not Met
Breath Volume	0.789	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time 04/30/2010 20:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ADDISON, RAYMOND, E

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: MILNES, KENNETH

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PASOS RICARTE

DATE/TIME OF TEST: 04/11/2010 03:11:56

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Subject Sample #1	0.270	03:04
Breath Volume	3.597	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample #2	0.220	03:07
Breath Volume	1.304	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample #3	0.228	03:10
Breath Volume	1.296	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time 04/30/2010 20:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ADDISON, RAYMOND, E

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: MILNES, KENNETH

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PASOS RICARTE

DATE/TIME OF TEST: 04/11/2010 03:25:06

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:22	
Subject Sample #1		03:23	RFI Detect
Breath Volume	2.648	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time 04/30/2010 20:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ADDISON, RAYMOND, E

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: MILNES, KENNETH

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PASOS RICARTE

DATE/TIME OF TEST: 04/11/2010 03:33:42

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Subject Sample #1	0.211	03:29
Breath Volume	1.613	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample #2	0.208	03:32
Breath Volume	2.812	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time 04/30/2010 20:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARIAS, ELOY,

BREATH TEST OPERATOR AGENCY: FHP Troop K

ARRESTING OFFICER: WILLIAMS, LAURIE

ARRESTING OFFICER AGENCY: FHP Troop K

LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 17:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/17/2010 18:31:55

Diagnostics Check	OK	18:25
Air Blank	0.000	18:25
Control Test	0.081	18:25
Air Blank	0.000	18:26
Subject Sample #1	0.234	18:27
Breath Volume	4.929	18:27
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample #2	0.219	18:30
Breath Volume	3.023	18:30
Air Blank	0.000	18:30
Control Test	0.078	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time 04/30/2010 20:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARIAS, ELOY,

BREATH TEST OPERATOR AGENCY: FHP Troop K

ARRESTING OFFICER: ARIAS, ELOY

ARRESTING OFFICER AGENCY: FHP Troop K

LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 21:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GERMAN

DATE/TIME OF TEST: 04/17/2010 22:37:45

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.080	22:27	
Air Blank	0.000	22:27	
Subject Sample #1		22:30	No Sample Provided
Breath Volume	0.000	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:33	
Subject Sample #2		22:36	No Sample Provided
Breath Volume	0.000	22:36	
Air Blank	0.000	22:36	
Control Test	0.081	22:37	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time 04/30/2010 20:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARIAS, MANUEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: GASCON, CARLOS

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 12:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/24/2010 01:45:06

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:39
Air Blank	0.000	01:39
Subject Sample #1	0.233	01:40
Breath Volume	3.597	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample #2	0.219	01:43
Breath Volume	2.496	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time 04/30/2010 20:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROSARIO, JUAN, C

BREATH TEST OPERATOR AGENCY: FHP Troop K

ARRESTING OFFICER: SERRANO, MARILYN

ARRESTING OFFICER AGENCY: FHP Troop K

LAST AGENCY INSPECTION: 03/29/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 01:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LARRAURI

DATE/TIME OF TEST: 04/24/2010 02:19:04

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Subject Sample #1	0.078	02:14
Breath Volume	3.199	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample #2	0.065	02:17
Breath Volume	1.878	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:19

Data Download Date/Time 04/30/2010 20:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE, DAVID, L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: ASHLEY, CHRIS
ARRESTING OFFICER AGENCY: Palatka PD
LAST AGENCY INSPECTION: 03/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: KENNER

DATE/TIME OF TEST: 04/09/2010 03:53:47

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample #1	0.286	03:49
Breath Volume	1.527	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample #2	0.286	03:52
Breath Volume	1.714	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time 04/20/2010 16:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS, RONALD, D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LEWIS, RONALD
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HENSON

DATE/TIME OF TEST: 04/03/2010 02:44:41

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.077	02:38
Air Blank	0.000	02:39
Subject Sample #1	0.110	02:40
Breath Volume	2.441	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample #2	0.103	02:43
Breath Volume	2.250	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time 04/20/2010 16:18

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE, DAVID, L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: TAYLOR, DANIEL
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 03/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: COSTON

DATE/TIME OF TEST: 04/03/2010 05:23:57

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.079	05:18
Air Blank	0.000	05:18
Subject Sample #1	0.168	05:19
Breath Volume	3.292	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:21
Subject Sample #2	0.170	05:22
Breath Volume	2.992	05:22
Air Blank	0.000	05:23
Control Test	0.078	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time 04/20/2010 16:18

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEDENBAUGH, JUSTIN, H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BEDENBAUGH, JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: FELTMATE

DATE/TIME OF TEST: 04/03/2010 18:06:44

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:01	
Control Test	0.079	18:01	
Air Blank	0.000	18:02	
Subject Sample #1		18:05	Subject Test Refused
Breath Volume	0.000	18:05	
Air Blank	0.000	18:05	
Control Test	0.081	18:06	
Air Blank	0.000	18:06	
Diagnostics Check	OK	18:06	

Data Download Date/Time 04/20/2010 16:18

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE, DAVID, L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: ROE, DAVID
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 03/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SUIT

DATE/TIME OF TEST: 04/04/2010 04:38:04

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.078	04:29	
Air Blank	0.000	04:29	
Subject Sample #1		04:32	No Sample Provided
Breath Volume	0.000	04:32	
Air Blank	0.000	04:33	
Air Blank	0.000	04:35	
Subject Sample #2		04:36	Subject Test Refused
Breath Volume	0.000	04:36	
Air Blank	0.000	04:37	
Control Test	0.079	04:37	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:38	

Data Download Date/Time 04/20/2010 16:18

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEDENBAUGH, JUSTIN, H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BEDENBAUGH, JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/04/2010 17:40:12

Diagnostics Check	OK	17:29
Air Blank	0.000	17:29
Control Test	0.079	17:30
Air Blank	0.000	17:30
Subject Sample #1	0.256	17:32
Breath Volume	2.414	17:32
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample #2	0.235	17:35
Breath Volume	2.175	17:35
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample #3	0.246	17:38
Breath Volume	1.578	17:38
Air Blank	0.000	17:39
Control Test	0.077	17:39
Air Blank	0.000	17:40
Diagnostics Check	OK	17:40

Data Download Date/Time 04/20/2010 16:18

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE, DAVID, L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: ROE, DAVID
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 03/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: SHELL

DATE/TIME OF TEST: 04/16/2010 04:27:59

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21
Subject Sample #1	0.143	04:23
Breath Volume	2.093	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.143	04:26
Breath Volume	1.625	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time 04/20/2010 16:18

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAXWELL, WILLIAM, D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MARTIR, JUSTIN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831801I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BURROUGHS

DATE/TIME OF TEST: 04/02/2010 00:55:43

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Subject Sample #1	0.126	00:51
Breath Volume	1.339	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample #2	0.131	00:54
Breath Volume	1.316	00:54
Air Blank	0.000	00:54
Control Test	0.078	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time 04/22/2010 18:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DWAN, S, P

BREATH TEST OPERATOR AGENCY: Volusia County SO

ARRESTING OFFICER: DWAN, S

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 03/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 13:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:01

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 04/06/2010 14:34:13

Diagnostics Check	OK	14:27
Air Blank	0.000	14:28
Control Test	0.080	14:28
Air Blank	0.000	14:28
Subject Sample #1	0.219	14:29
Breath Volume	2.851	14:29
Air Blank	0.000	14:30
Air Blank	0.000	14:32
Subject Sample #2	0.220	14:32
Breath Volume	2.085	14:32
Air Blank	0.000	14:33
Control Test	0.078	14:33
Air Blank	0.000	14:34
Diagnostics Check	OK	14:34

Data Download Date/Time 04/22/2010 18:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEW, TIMOTHY, C
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: TEW, TIMOTHY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/15/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 831801I
DRY GAS EXPIRY DATE: 11/20/2010
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: VOURIOTIS

DATE/TIME OF TEST: 04/08/2010 20:09:34

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:07	
Control Test	0.079	20:07	
Air Blank	0.000	20:08	
Subject Sample #1		20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:08	
Control Test	0.080	20:08	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:09	

Data Download Date/Time 04/22/2010 18:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MAXWELL, WILLIAM, D

BREATH TEST OPERATOR AGENCY: Volusia County SO

ARRESTING OFFICER: GAETA, KOREY

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 03/15/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 23:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: VITELLO

DATE/TIME OF TEST: 04/20/2010 00:15:08

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Subject Sample #1	0.205	00:10
Breath Volume	1.671	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample #2	0.207	00:13
Breath Volume	2.152	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time 04/22/2010 18:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BOWMAN, EDWARD,

BREATH TEST OPERATOR AGENCY: Volusia County SO

ARRESTING OFFICER: WHITTET, JAMES

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 831801I

DRY GAS EXPIRY DATE: 11/20/2010

VIOLATION CODE: DUI

ARREST TIME: 22:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/23/2010 23:17:48

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:12
Subject Sample #1	0.083	23:13
Breath Volume	2.082	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample #2	0.086	23:16
Breath Volume	1.925	23:16
Air Blank	0.000	23:16
Control Test	0.078	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time 05/23/2010 22:45

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN, EDWARD,
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BOWMAN, EDWARD
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: WHITTAKER

DATE/TIME OF TEST: 04/01/2010 19:13:41

Diagnostics Check	OK	19:13	
Air Blank		19:13	RFI Detect
Air Blank		19:13	RFI Detect

Data Download Date/Time 04/22/2010 16:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BOWMAN, EDWARD,

BREATH TEST OPERATOR AGENCY: Volusia County SO

ARRESTING OFFICER: BOWMAN, EDWARD

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 03/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 18:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: WHITTAKER

DATE/TIME OF TEST: 04/01/2010 19:22:26

Diagnostics Check	OK	19:16
Air Blank	0.000	19:16
Control Test	0.080	19:16
Air Blank	0.000	19:17
Subject Sample #1	0.105	19:17
Breath Volume	1.488	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample #2	0.107	19:20
Breath Volume	1.488	19:20
Air Blank	0.000	19:21
Control Test	0.080	19:21
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time 04/22/2010 16:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BOWMAN, EDWARD,

BREATH TEST OPERATOR AGENCY: Volusia County SO

ARRESTING OFFICER: BOWMAN, EDWARD

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 03/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 18:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BARNLUND

DATE/TIME OF TEST: 04/02/2010 19:44:22

Diagnostics Check	OK	19:39	
Air Blank	0.000	19:40	
Control Test	0.080	19:40	
Air Blank	0.000	19:41	
Subject Sample #1		19:43	Subject Test Refused
Breath Volume	0.000	19:43	
Air Blank	0.000	19:43	
Control Test	0.079	19:43	
Air Blank	0.000	19:44	
Diagnostics Check	OK	19:44	

Data Download Date/Time 04/22/2010 16:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN, EDWARD,
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BOWMAN, EDWARD
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: BROW

DATE/TIME OF TEST: 04/04/2010 16:56:33

Diagnostics Check	OK	16:50
Air Blank	0.000	16:50
Control Test	0.080	16:51
Air Blank	0.000	16:51
Subject Sample #1	0.140	16:52
Breath Volume	1.515	16:52
Air Blank	0.000	16:52
Air Blank	0.000	16:54
Subject Sample #2	0.139	16:55
Breath Volume	1.179	16:55
Air Blank	0.000	16:55
Control Test	0.079	16:55
Air Blank	0.000	16:56
Diagnostics Check	OK	16:56

Data Download Date/Time 04/22/2010 16:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BOWMAN, EDWARD,

BREATH TEST OPERATOR AGENCY: Volusia County SO

ARRESTING OFFICER: DEARBORN, DENNIS

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 03/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 22:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FAY

DATE/TIME OF TEST: 04/04/2010 23:42:41

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample #1	0.201	23:38
Breath Volume	3.562	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample #2	0.200	23:41
Breath Volume	2.609	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time 04/22/2010 16:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST, JOSEPH, W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BARRETT, ANDREW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/09/2010 20:53:31

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:41	
Control Test	0.080	20:41	
Air Blank	0.000	20:41	
Subject Sample #1	0.119	20:44	Volume Not Met
Breath Volume	0.699	20:44	
Air Blank	0.000	20:45	
Air Blank	0.000	20:47	
Subject Sample #2	0.149	20:48	
Breath Volume	1.445	20:48	
Air Blank	0.000	20:49	
Air Blank	0.000	20:51	
Subject Sample #3	0.144	20:52	
Breath Volume	1.312	20:52	
Air Blank	0.000	20:52	
Control Test	0.080	20:52	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

Data Download Date/Time 04/22/2010 16:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FROST, JOSEPH, W

BREATH TEST OPERATOR AGENCY: Volusia County SO

ARRESTING OFFICER: CARLSON, RANDALL

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 03/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 23:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MORGAL

DATE/TIME OF TEST: 04/10/2010 00:04:39

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:02	
Control Test	0.079	00:02	
Air Blank	0.000	00:03	
Subject Sample #1		00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Control Test	0.079	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time 04/22/2010 16:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST, JOSEPH, W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: FROST, JOSEPH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/10/2010 02:53:42

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.078	02:43	
Air Blank	0.000	02:44	
Subject Sample #1	0.148	02:47	Slope Not Level
Breath Volume	1.140	02:47	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample #2		02:52	Subject Test Refused
Breath Volume	1.046	02:52	
Air Blank	0.000	02:52	
Control Test	0.078	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time 04/22/2010 16:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN, EDWARD,
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: LUMA, BRIAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/02/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 04/19/2010 11:47:13

Diagnostics Check	OK	11:44	
Air Blank	0.000	11:44	
Control Test	0.079	11:45	
Air Blank	0.000	11:45	
Subject Sample #1		11:45	Subject Test Refused
Breath Volume	0.000	11:45	
Air Blank	0.000	11:46	
Control Test	0.079	11:46	
Air Blank	0.000	11:47	
Diagnostics Check	OK	11:47	

Data Download Date/Time 04/22/2010 16:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MACDONALD, DAVID, W

BREATH TEST OPERATOR AGENCY: Volusia County SO

ARRESTING OFFICER: ELMAZI, JULIAN

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 03/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 18:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DENNEY

DATE/TIME OF TEST: 04/21/2010 19:27:05

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.079	19:21
Air Blank	0.000	19:22
Subject Sample #1	0.208	19:22
Breath Volume	3.882	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample #2	0.207	19:25
Breath Volume	3.113	19:25
Air Blank	0.000	19:26
Control Test	0.079	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:27

Data Download Date/Time 04/22/2010 16:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TEW, TIMOTHY, C

BREATH TEST OPERATOR AGENCY: Volusia County SO

ARRESTING OFFICER: ECKBERG, CARL

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 22:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 04/22/2010 23:31:23

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:25
Subject Sample #1	0.256	23:26
Breath Volume	3.343	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample #2	0.254	23:29
Breath Volume	2.558	23:29
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time 05/25/2010 13:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAXWELL, WILLIAM, D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HENDERSON, B
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 04/26/2010 02:07:35

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Subject Sample #1	0.148	02:03
Breath Volume	1.671	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample #2	0.144	02:06
Breath Volume	2.570	02:06
Air Blank	0.000	02:06
Control Test	0.078	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time 05/25/2010 13:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IHNKEN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: IHNKEN, M
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/22/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 832201I
DRY GAS EXPIRY DATE: 11/22/2010
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/29/2010 19:56:33

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.079	19:50
Air Blank	0.000	19:50
Subject Sample #1	0.249	19:51
Breath Volume	1.531	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample #2	0.239	19:55
Breath Volume	1.347	19:55
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time 05/25/2010 13:12