

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: FREY, ROBERT, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: LAVANTURE, K

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 03:45

Gender: Male

Observation Period Began: 03:45

Subject's Last Name: HOOTS

Date/Time of Test: 04/01/2007 04:25

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.078	04:19
Air Blank	0.000	04:20
Subject Sample #1	0.147	04:21
Breath Volume	1.242	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.142	04:24
Breath Volume	3.296	04:24
Air Blank	0.000	04:24
Control Test	0.075	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORGAN, DARIN

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 23:08

Gender: Male

Observation Period Began: 22:48

Subject's Last Name: URQUIZA

Date/Time of Test: 04/01/2007 23:36

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:31
Subject Sample #1	0.167	23:31
Breath Volume	1.503	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample #2	0.165	23:34
Breath Volume	2.035	23:34
Air Blank	0.000	23:35
Control Test	0.077	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MARCINIK, DAVID

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 02:06

Gender: Female

Observation Period Began: 02:01

Subject's Last Name: JACKSON

Date/Time of Test: 04/06/2007 02:57

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.077	02:48
Air Blank	0.000	02:49
Subject Sample #1	0.137	02:52
Breath Volume	1.589	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:54
Subject Sample #2	0.143	02:55
Breath Volume	1.515	02:55
Air Blank	0.000	02:56
Control Test	0.075	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:57

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: FREY, ROBERT, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: RICHARD, MARK

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 20:03

Gender: Female

Observation Period Began: 20:03

Subject's Last Name: ROBBINS

Date/Time of Test: 04/08/2007 20:47

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.078	20:40
Air Blank	0.000	20:41
Subject Sample #1	0.193	20:41
Breath Volume	1.562	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample #2	0.183	20:45
Breath Volume	1.800	20:45
Air Blank	0.000	20:46
Control Test	0.078	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: FREY, ROBERT, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: HOLLOMAN, J

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 00:20

Gender: Male

Observation Period Began: 00:20

Subject's Last Name: CACIOPPO

Date/Time of Test: 04/09/2007 01:04

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.078	00:58
Air Blank	0.000	00:59
Subject Sample #1	0.177	00:59
Breath Volume	1.945	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample #2	0.180	01:02
Breath Volume	2.078	01:02
Air Blank	0.000	01:03
Control Test	0.077	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MARCINIK, DAVID

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 01:00

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: PODKULSKI

Date/Time of Test: 04/12/2007 02:44

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.077	02:40	
Air Blank	0.000	02:41	
Subject Sample #1		02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Control Test	0.078	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORGAN, DARIN

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 03:10

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: SCOTT

Date/Time of Test: 04/14/2007 03:46

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.078	03:35	
Air Blank	0.000	03:36	
Subject Sample #1	0.044	03:39	Volume Not Met
Breath Volume	0.089	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:41	
Subject Sample #2	0.000	03:45	Volume Not Met
Breath Volume	0.082	03:45	
Air Blank	0.000	03:45	
Control Test	0.078	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORGAN, D

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 03:10

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: SCOTT

Date/Time of Test: 04/14/2007 03:55

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.077	03:50	
Air Blank	0.000	03:50	
Subject Sample #1		03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Control Test	0.078	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORGAN, DARIN

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 00:40

Gender: Female

Observation Period Began: 00:30

Subject's Last Name: CONTI

Date/Time of Test: 04/15/2007 01:03

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.077	00:58
Air Blank	0.000	00:58
Subject Sample #1	0.125	00:59
Breath Volume	1.988	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample #2	0.126	01:02
Breath Volume	1.820	01:02
Air Blank	0.000	01:02
Control Test	0.076	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORGAN, DARIN

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 03:10

Gender: Male

Observation Period Began: 02:50

Subject's Last Name: MCGEE

Date/Time of Test: 04/15/2007 03:28

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.077	03:23
Air Blank	0.000	03:23
Subject Sample #1	0.149	03:24
Breath Volume	2.496	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample #2	0.138	03:27
Breath Volume	1.425	03:27
Air Blank	0.000	03:27
Control Test	0.076	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: BADMAN, LIN, M

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: BAUMAN, DONALD

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 12:50

Gender: Male

Observation Period Began: 13:00

Subject's Last Name: FRANKLIN

Date/Time of Test: 04/17/2007 13:38

Diagnostics Check	OK	13:32
Air Blank	0.000	13:32
Control Test	0.078	13:32
Air Blank	0.000	13:33
Subject Sample #1	0.000	13:34
Breath Volume	2.585	13:34
Air Blank	0.000	13:35
Air Blank	0.000	13:37
Subject Sample #2	0.000	13:37
Breath Volume	2.332	13:37
Air Blank	0.000	13:38
Control Test	0.078	13:38
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: BADMAN, LIN, M

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORRIS, KEVIN

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 510902I

Dry Gas Expiry Date: 04/21/2007

Violation Code: DUI

Arrest Time: 22:24

Gender: Male

Observation Period Began: 22:30

Subject's Last Name: CAMPBELL

Date/Time of Test: 04/17/2007 23:07

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:02
Subject Sample #1	0.158	23:03
Breath Volume	3.257	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample #2	0.156	23:06
Breath Volume	3.199	23:06
Air Blank	0.000	23:06
Control Test	0.077	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MARCINIK, DAVID

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 00:35

Gender: Male

Observation Period Began: 00:35

Subject's Last Name: COFFIE

Date/Time of Test: 04/24/2007 01:29

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:24
Air Blank	0.000	01:24
Subject Sample #1	0.076	01:25
Breath Volume	1.847	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample #2	0.074	01:28
Breath Volume	2.378	01:28
Air Blank	0.000	01:28
Control Test	0.078	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: VALENTI, NEAL, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MENDEZ, WANDA

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 10:02

Gender: Male

Observation Period Began: 10:00

Subject's Last Name: GARCIA

Date/Time of Test: 04/24/2007 10:57

Diagnostics Check	OK	10:49
Air Blank	0.000	10:49
Control Test	0.079	10:49
Air Blank	0.000	10:50
Subject Sample #1	0.172	10:53
Breath Volume	2.101	10:53
Air Blank	0.000	10:53
Air Blank	0.000	10:55
Subject Sample #2	0.164	10:55
Breath Volume	2.242	10:55
Air Blank	0.000	10:56
Control Test	0.077	10:56
Air Blank	0.000	10:57
Diagnostics Check	OK	10:57

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MARCINIK, DAVID

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 02:20

Gender: Female

Observation Period Began: 02:20

Subject's Last Name: POPE

Date/Time of Test: 04/25/2007 03:19

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank	0.000	03:15	
Subject Sample #1		03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: VALENTI, NEAL, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: GARCEAU, GLENN

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 07:35

Gender: Male

Observation Period Began: 07:50

Subject's Last Name: WILLIAMS

Date/Time of Test: 04/25/2007 09:29

Diagnostics Check	OK	09:21
Air Blank	0.000	09:22
Control Test	0.079	09:22
Air Blank	0.000	09:23
Subject Sample #1	0.179	09:24
Breath Volume	2.546	09:24
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample #2	0.183	09:27
Breath Volume	3.195	09:27
Air Blank	0.000	09:28
Control Test	0.078	09:28
Air Blank	0.000	09:29
Diagnostics Check	OK	09:29

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MARCINIK, MARCINIK

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 02:26

Gender: Male

Observation Period Began: 02:26

Subject's Last Name: LITTLES

Date/Time of Test: 04/26/2007 03:26

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:21
Air Blank	0.000	03:21
Subject Sample #1	0.133	03:22
Breath Volume	1.191	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample #2	0.136	03:25
Breath Volume	1.835	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 04/26/2007 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: WHEELER, W

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 00:36

Gender: Female

Observation Period Began: 00:36

Subject's Last Name: ORTNER

Date/Time of Test: 04/28/2007 01:30

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Subject Sample #1	0.172	01:25
Breath Volume	1.265	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample #2	0.157	01:28
Breath Volume	1.730	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 05/30/2007 19:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: AUSTIN, R

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 01:28

Gender: Male

Observation Period Began: 01:28

Subject's Last Name: DIGEORGIO

Date/Time of Test: 04/28/2007 02:11

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Subject Sample #1	0.147	02:06
Breath Volume	3.855	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #2	0.146	02:09
Breath Volume	3.867	02:09
Air Blank	0.000	02:10
Control Test	0.076	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 05/30/2007 19:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MARCINIK, DAVID

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 04:53

Gender: Female

Observation Period Began: 05:01

Subject's Last Name: JEPSON

Date/Time of Test: 04/29/2007 05:42

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.080	05:34
Air Blank	0.000	05:35
Subject Sample #1	0.109	05:36
Breath Volume	2.710	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample #2	0.099	05:40
Breath Volume	2.000	05:40
Air Blank	0.000	05:41
Control Test	0.078	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 05/30/2007 19:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: BADMAN, LIN, M

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: PIDGEON, ANDREW

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 20:35

Gender: Male

Observation Period Began: 20:30

Subject's Last Name: JACOBS

Date/Time of Test: 04/29/2007 21:22

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.080	21:15
Air Blank	0.000	21:15
Subject Sample #1	0.184	21:16
Breath Volume	1.812	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample #2	0.177	21:20
Breath Volume	2.246	21:20
Air Blank	0.000	21:21
Control Test	0.078	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 05/30/2007 19:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MARCINIK, DAVID

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 02:08

Gender: Male

Observation Period Began: 02:08

Subject's Last Name: LYNCH

Date/Time of Test: 04/30/2007 03:14

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample #1	0.214	03:09
Breath Volume	4.183	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample #2	0.218	03:12
Breath Volume	4.277	03:12
Air Blank	0.000	03:13
Control Test	0.077	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 05/30/2007 19:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: FITHIAN, RACHEL

Arresting Officer Agency: Sanibel PD

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:09

Gender: Male

Observation Period Began: 01:09

Subject's Last Name: RIVERA RIOS

Date/Time of Test: 04/14/2007 01:39

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:35	
Air Blank	0.000	01:35	
Subject Sample #1		01:37	Subject Test Refused
Breath Volume	0.082	01:37	
Air Blank	0.000	01:38	
Control Test	0.079	01:38	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: CELA, MIKE

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:39

Gender: Male

Observation Period Began: 01:45

Subject's Last Name: WHITE

Date/Time of Test: 04/14/2007 02:25

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample #1	0.118	02:20
Breath Volume	2.460	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample #2	0.119	02:23
Breath Volume	2.136	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: LEE, SELENA

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: PEREZ

Date/Time of Test: 04/14/2007 02:38

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample #1	0.174	02:34
Breath Volume	1.332	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample #2	0.166	02:37
Breath Volume	1.460	02:37
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: CHECO, JOSE

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:52

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: MORRISON

Date/Time of Test: 04/14/2007 03:10

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Subject Sample #1	0.130	03:06
Breath Volume	3.199	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:08
Subject Sample #2	0.117	03:09
Breath Volume	1.597	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: KONIECZKI, DOMINIC

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:53

Gender: Male

Observation Period Began: 02:58

Subject's Last Name: MONTEROSA

Date/Time of Test: 04/14/2007 03:29

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.078	03:23
Air Blank	0.000	03:23
Subject Sample #1	0.178	03:24
Breath Volume	2.316	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample #2	0.182	03:27
Breath Volume	1.968	03:27
Air Blank	0.000	03:28
Control Test	0.077	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: CASTRO, JESUS

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 03:14

Gender: Male

Observation Period Began: 03:07

Subject's Last Name: ARGUELLES

Date/Time of Test: 04/14/2007 03:46

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Subject Sample #1	0.241	03:38
Breath Volume	2.199	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample #2	0.219	03:41
Breath Volume	2.437	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample #3	0.228	03:44
Breath Volume	2.101	03:44
Air Blank	0.000	03:45
Control Test	0.077	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:46

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: URRRA, ALEJANDRO

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 03:18

Gender: Male

Observation Period Began: 03:18

Subject's Last Name: ANDRES

Date/Time of Test: 04/14/2007 03:59

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:53
Subject Sample #1	0.120	03:54
Breath Volume	3.425	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample #2	0.110	03:57
Breath Volume	2.472	03:57
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: CORNELIUS, TIM

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:07

Gender: Male

Observation Period Began: 23:07

Subject's Last Name: SYRACUSE

Date/Time of Test: 04/20/2007 23:44

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Subject Sample #1	0.014	23:39
Breath Volume	2.320	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample #2	0.016	23:42
Breath Volume	3.410	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: ROODY, HEATH

Arresting Officer Agency: Cape Coral PD

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:16

Gender: Female

Observation Period Began: 23:36

Subject's Last Name: SIANO

Date/Time of Test: 04/21/2007 00:07

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:01
Subject Sample #1	0.102	00:02
Breath Volume	2.472	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample #2	0.100	00:05
Breath Volume	1.593	00:05
Air Blank	0.000	00:06
Control Test	0.078	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: TRACEY, ALAN

Arresting Officer Agency: Cape Coral PD

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:45

Gender: Female

Observation Period Began: 23:45

Subject's Last Name: SNIDER

Date/Time of Test: 04/21/2007 00:21

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:12
Subject Sample #1	0.084	00:15
Breath Volume	2.316	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample #2	0.084	00:19
Breath Volume	1.691	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: BOLIEK, P

Arresting Officer Agency: Collier County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 01:30

Subject's Last Name: WELCH

Date/Time of Test: 04/21/2007 01:56

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.080	01:52	
Air Blank	0.000	01:52	
Subject Sample #1		01:54	Subject Test Refused
Breath Volume	0.070	01:54	
Air Blank	0.000	01:55	
Control Test	0.080	01:55	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: TEJEDA, ORLANDO

Arresting Officer Agency: Cape Coral PD

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:25

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: DIAGO

Date/Time of Test: 04/21/2007 02:21

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:16
Subject Sample #1	0.073	02:16
Breath Volume	1.320	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample #2	0.074	02:20
Breath Volume	1.265	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: TESTT, TTSET

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 20:59

Gender: Male

Observation Period Began: 21:00

Subject's Last Name: TESTT

Date/Time of Test: 04/27/2007 21:28

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.080	21:23
Air Blank	0.000	21:24
Subject Sample #1	0.000	21:24
Breath Volume	1.812	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample #2	0.000	21:27
Breath Volume	1.710	21:27
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: GOULD, LEONARD

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:25

Gender: Female

Observation Period Began: 23:30

Subject's Last Name: ULINO

Date/Time of Test: 04/28/2007 00:00

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.078	23:52
Air Blank	0.000	23:52
Subject Sample #1	0.021	23:54
Breath Volume	1.617	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample #2	0.024	23:58
Breath Volume	1.203	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: FITHIAN, RACHEL

Arresting Officer Agency: Sanibel PD

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:40

Gender: Male

Observation Period Began: 23:40

Subject's Last Name: HOFFMAN

Date/Time of Test: 04/28/2007 00:12

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Subject Sample #1	0.099	00:07
Breath Volume	2.093	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample #2	0.099	00:10
Breath Volume	1.343	00:10
Air Blank	0.000	00:11
Control Test	0.078	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: SOMMERS, EDWARD

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 23:44

Gender: Male

Observation Period Began: 23:57

Subject's Last Name: HITCHCOCK

Date/Time of Test: 04/28/2007 00:34

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.078	00:28
Air Blank	0.000	00:28
Subject Sample #1	0.108	00:29
Breath Volume	3.156	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample #2	0.112	00:32
Breath Volume	3.937	00:32
Air Blank	0.000	00:33
Control Test	0.078	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:34

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: DEKLE, GREG

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 00:25

Gender: Female

Observation Period Began: 00:25

Subject's Last Name: REDMOND

Date/Time of Test: 04/28/2007 00:58

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.078	00:52
Air Blank	0.000	00:52
Subject Sample #1	0.166	00:53
Breath Volume	1.187	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample #2	0.180	00:56
Breath Volume	3.390	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: PETRACCA, O

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 00:50

Gender: Female

Observation Period Began: 00:54

Subject's Last Name: WRIGHT

Date/Time of Test: 04/28/2007 01:21

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:15
Air Blank	0.000	01:15
Subject Sample #1	0.123	01:16
Breath Volume	1.886	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample #2	0.115	01:19
Breath Volume	2.046	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: PETRACCA, DENNIS

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 00:48

Gender: Male

Observation Period Began: 00:48

Subject's Last Name: SASHER

Date/Time of Test: 04/28/2007 01:32

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Subject Sample #1	0.202	01:28
Breath Volume	2.746	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample #2	0.195	01:30
Breath Volume	2.527	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: CHIAPPETTA, LOU

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:01

Gender: Male

Observation Period Began: 02:01

Subject's Last Name: SOLIZ

Date/Time of Test: 04/28/2007 02:36

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample #1	0.239	02:32
Breath Volume	2.113	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample #2	0.240	02:35
Breath Volume	2.316	02:35
Air Blank	0.000	02:35
Control Test	0.078	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: KONIECZKI, DOMINIC

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 04/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:05

Gender: Male

Observation Period Began: 02:24

Subject's Last Name: OLIVO

Date/Time of Test: 04/28/2007 02:58

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:53
Subject Sample #1	0.166	02:53
Breath Volume	2.121	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.165	02:57
Breath Volume	2.863	02:57
Air Blank	0.000	02:57
Control Test	0.078	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 05/02/2007 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: COLDING, LINDA, D

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: MORREY, S

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 03/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 22:00

Gender: Male

Observation Period Began: 22:17

Subject's Last Name: BANKS

Date/Time of Test: 04/04/2007 22:43

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:42	
Control Test	0.074	22:42	Control Outside Tolerance
Air Blank	0.000	22:43	

Data Download Date/Time: 04/11/2007 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: COLDING, LINDA, D

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: MORREY, S

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 03/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 22:00

Gender: Male

Observation Period Began: 22:17

Subject's Last Name: BANKS

Date/Time of Test: 04/04/2007 22:50

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:45
Subject Sample #1	0.175	22:46
Breath Volume	1.777	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample #2	0.166	22:49
Breath Volume	1.539	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 04/11/2007 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: COLDING, LINDA, D
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: MCCALLA, C
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 03/20/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 02:52
Gender: Male
Observation Period Began: 03:22

Subject's Last Name: LEWIS

Date/Time of Test: 04/05/2007 03:49

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.078	03:43
Air Blank	0.000	03:44
Subject Sample #1	0.022	03:44
Breath Volume	3.796	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample #2	0.019	03:47
Breath Volume	2.757	03:47
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 04/11/2007 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: COLDING, LINDA, D

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: MORREY, S

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 03/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 03:14

Gender: Male

Observation Period Began: 03:31

Subject's Last Name: SHEPHARD

Date/Time of Test: 04/05/2007 04:03

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample #1	0.157	03:56
Breath Volume	2.679	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample #2	0.156	04:02
Breath Volume	2.019	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/11/2007 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: COLDING, LINDA, D
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: STABINS, K
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 03/20/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 04:00
Gender: Male
Observation Period Began: 04:19

Subject's Last Name: TANGARIFE

Date/Time of Test: 04/05/2007 04:48

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:40
Subject Sample #1	0.189	04:42
Breath Volume	1.125	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample #2	0.170	04:46
Breath Volume	1.320	04:46
Air Blank	0.000	04:47
Control Test	0.077	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 04/11/2007 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: GRAY, BEVERLY, A
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: MORRELL, L
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 03/20/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 00:57
Gender: Male
Observation Period Began: 01:14

Subject's Last Name: LEWIS

Date/Time of Test: 04/07/2007 01:45

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Subject Sample #1	0.106	01:40
Breath Volume	2.164	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample #2	0.110	01:43
Breath Volume	2.210	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 04/11/2007 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: COLDING, LINDA, D
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: MCMURPHY, M
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 03/20/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 01:26
Gender: Male
Observation Period Began: 01:49

Subject's Last Name: JOHNSON

Date/Time of Test: 04/07/2007 02:18

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Subject Sample #1	0.105	02:13
Breath Volume	2.062	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample #2	0.104	02:16
Breath Volume	2.566	02:16
Air Blank	0.000	02:17
Control Test	0.078	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 04/11/2007 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: WILLIAMS, G

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 03/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 02:13

Gender: Male

Observation Period Began: 02:32

Subject's Last Name: SHEPPARD

Date/Time of Test: 04/07/2007 03:07

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:02
Subject Sample #1	0.120	03:03
Breath Volume	1.910	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample #2	0.111	03:05
Breath Volume	2.453	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 04/11/2007 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: COLDING, LINDA, D

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: MCCALLA, C

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 03/20/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 02:36

Gender: Male

Observation Period Began: 03:02

Subject's Last Name: WETHERS

Date/Time of Test: 04/07/2007 03:28

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Subject Sample #1	0.209	03:24
Breath Volume	2.167	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample #2	0.217	03:26
Breath Volume	1.839	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 04/11/2007 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: SNYDER, MELONIE, E
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: DOBSON, B
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 04/11/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 02:52
Gender: Male
Observation Period Began: 03:32

Subject's Last Name: JACKSON JR

Date/Time of Test: 04/12/2007 04:01

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:55
Subject Sample #1	0.065	03:56
Breath Volume	2.472	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample #2	0.065	04:00
Breath Volume	2.945	04:00
Air Blank	0.000	04:00
Control Test	0.078	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: SNYDER, MELONIE, E
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: MORRELL, L
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 04/11/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 03:30
Gender: Male
Observation Period Began: 03:54

Subject's Last Name: FRADLEY

Date/Time of Test: 04/12/2007 04:27

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:22
Subject Sample #1	0.205	04:22
Breath Volume	1.875	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.203	04:26
Breath Volume	1.894	04:26
Air Blank	0.000	04:27
Control Test	0.077	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: SNYDER, MELONIE, E
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: WILLIAMS, G
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 04/11/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 04:10
Gender: Male
Observation Period Began: 04:27

Subject's Last Name: TAVLAN

Date/Time of Test: 04/12/2007 04:55

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.078	04:49
Air Blank	0.000	04:50
Subject Sample #1	0.118	04:50
Breath Volume	2.437	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample #2	0.123	04:53
Breath Volume	2.210	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: GRAY, BEVERLY, A
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: GLOVER, J
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 04/11/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 23:18
Gender: Male
Observation Period Began: 23:31

Subject's Last Name: RODRIQUEZ-GARCIA

Date/Time of Test: 04/14/2007 00:03

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Subject Sample #1	0.122	23:59
Breath Volume	1.757	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample #2	0.127	00:02
Breath Volume	1.910	00:02
Air Blank	0.000	00:02
Control Test	0.078	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: OWENS, EDWARD, W

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: SPEAKS, J

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 23:36

Gender: Male

Observation Period Began: 23:48

Subject's Last Name: TORRES-GARCIA

Date/Time of Test: 04/14/2007 00:15

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.079	00:10
Air Blank	0.000	00:10
Subject Sample #1	0.129	00:11
Breath Volume	1.570	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample #2	0.130	00:14
Breath Volume	1.253	00:14
Air Blank	0.000	00:14
Control Test	0.077	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: COLDING, LINDA, D
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: GLOVER, J
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 04/11/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 518602I
Dry Gas Expiry Date: 07/07/2007
Violation Code: DUI
Arrest Time: 21:16
Gender: Male
Observation Period Began: 22:06

Subject's Last Name: PATTERSON

Date/Time of Test: 04/20/2007 22:34

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Subject Sample #1	0.103	22:29
Breath Volume	1.820	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample #2	0.105	22:32
Breath Volume	1.820	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: COLDING, LINDA, D
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: DEPIANTA, R
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 04/11/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 518602I
Dry Gas Expiry Date: 07/07/2007
Violation Code: DUI
Arrest Time: 23:08
Gender: Male
Observation Period Began: 23:35

Subject's Last Name: PEREZ-PEREZ

Date/Time of Test: 04/21/2007 00:02

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample #1	0.174	23:58
Breath Volume	2.457	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample #2	0.171	00:01
Breath Volume	2.878	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: COLDING, LINDA, D
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: DOBSON, B
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 04/11/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 518602I
Dry Gas Expiry Date: 07/07/2007
Violation Code: DUI
Arrest Time: 01:33
Gender: Male
Observation Period Began: 02:01

Subject's Last Name: MILLER

Date/Time of Test: 04/21/2007 02:30

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Subject Sample #1	0.029	02:25
Breath Volume	3.457	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample #2	0.028	02:28
Breath Volume	4.464	02:28
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: COLDING, LINDA, D

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: MCGINNIS, T

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 518602I

Dry Gas Expiry Date: 07/07/2007

Violation Code: DUI

Arrest Time: 02:15

Gender: Female

Observation Period Began: 02:32

Subject's Last Name: HOWELL

Date/Time of Test: 04/21/2007 02:59

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Subject Sample #1	0.103	02:54
Breath Volume	2.511	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.101	02:57
Breath Volume	2.933	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:59

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: COLDING, LINDA, D
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: GLOVER, J
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 04/11/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 518602I
Dry Gas Expiry Date: 07/07/2007
Violation Code: DUI
Arrest Time: 02:46
Gender: Female
Observation Period Began: 03:40

Subject's Last Name: ORTIZ

Date/Time of Test: 04/21/2007 04:10

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Subject Sample #1	0.182	04:06
Breath Volume	2.058	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample #2	0.186	04:09
Breath Volume	2.054	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: OWENS, EDWARD, W

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: GROVES, D

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 518602I

Dry Gas Expiry Date: 07/07/2007

Violation Code: DUI

Arrest Time: 01:53

Gender: Male

Observation Period Began: 02:18

Subject's Last Name: GUARDIOLA

Date/Time of Test: 04/28/2007 02:45

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Subject Sample #1	0.117	02:40
Breath Volume	3.355	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample #2	0.124	02:43
Breath Volume	2.281	02:43
Air Blank	0.000	02:44
Control Test	0.078	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 05/25/2007 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: NUGENT, DAVID, A

Breath Test Operator Agency: Clearwater PD

Arresting Officer: NUGENT, DAVID

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 00:37

Gender: Female

Observation Period Began: 01:41

Subject's Last Name: FATE

Date/Time of Test: 04/01/2007 02:09

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:04
Air Blank	0.000	02:04
Subject Sample #1	0.076	02:04
Breath Volume	1.492	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample #2	0.081	02:07
Breath Volume	1.730	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 04/10/2007 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILLER, THEODORE, J

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MILLER, THEODORE

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 12:05

Gender: Male

Observation Period Began: 12:37

Subject's Last Name: MEZQUITE

Date/Time of Test: 04/05/2007 13:09

Diagnostics Check	OK	12:57
Air Blank	0.000	12:57
Control Test	0.081	12:58
Air Blank	0.000	12:58
Subject Sample #1	0.265	12:59
Breath Volume	2.082	12:59
Air Blank	0.000	13:00
Air Blank	0.000	13:02
Subject Sample #2	0.244	13:04
Breath Volume	2.093	13:04
Air Blank	0.000	13:05
Air Blank	0.000	13:06
Subject Sample #3	0.259	13:07
Breath Volume	1.996	13:07
Air Blank	0.000	13:08
Control Test	0.078	13:08
Air Blank	0.000	13:09
Diagnostics Check	OK	13:09

Data Download Date/Time: 04/10/2007 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: TURPACK, T,

Breath Test Operator Agency: Clearwater PD

Arresting Officer: TURPACK, T

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 03/23/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 14:07

Gender: Male

Observation Period Began: 14:55

Subject's Last Name: STEWART

Date/Time of Test: 04/06/2007 15:26

Diagnostics Check	OK	15:22	
Air Blank	0.000	15:22	
Control Test	0.081	15:23	
Air Blank	0.000	15:23	
Subject Sample #1		15:25	Subject Test Refused
Breath Volume	0.000	15:25	
Air Blank	0.000	15:25	
Control Test	0.080	15:25	
Air Blank	0.000	15:26	
Diagnostics Check	OK	15:26	

Data Download Date/Time: 04/10/2007 19:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: BRUNEAU, DAVID, N

Breath Test Operator Agency: Clearwater PD

Arresting Officer: BRUNEAU, DAVID

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 21:27

Gender: Female

Observation Period Began: 21:55

Subject's Last Name: HERNANDEZ

Date/Time of Test: 04/11/2007 22:29

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Subject Sample #1	0.241	22:24
Breath Volume	1.488	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample #2	0.254	22:27
Breath Volume	2.218	22:27
Air Blank	0.000	22:28
Control Test	0.076	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 05/25/2007 04:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: NUGENT, DAVID, A

Breath Test Operator Agency: Clearwater PD

Arresting Officer: NUGENT, DAVID

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 00:49

Gender: Male

Observation Period Began: 01:32

Subject's Last Name: PHIPPS

Date/Time of Test: 04/12/2007 02:01

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.080	01:59	
Air Blank	0.000	02:00	
Subject Sample #1		02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 05/25/2007 04:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILNE, THOMAS, P

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MILNE, THOMAS

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 00:16

Gender: Female

Observation Period Began: 00:40

Subject's Last Name: COLOMBO

Date/Time of Test: 04/14/2007 01:06

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.082	01:03	
Air Blank	0.000	01:04	
Subject Sample #1		01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 05/25/2007 04:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: HOLE, S., A

Breath Test Operator Agency: Clearwater PD

Arresting Officer: HOLE, S.

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 20:22

Gender: Male

Observation Period Began: 21:08

Subject's Last Name: BANDE

Date/Time of Test: 04/17/2007 21:38

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.081	21:33
Air Blank	0.000	21:33
Subject Sample #1	0.101	21:34
Breath Volume	3.496	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:36
Subject Sample #2	0.101	21:37
Breath Volume	3.351	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 05/25/2007 04:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: HOLE, S., A

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MEDLIN, N.

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 21:27

Gender: Male

Observation Period Began: 21:50

Subject's Last Name: PAYNE

Date/Time of Test: 04/17/2007 22:21

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Subject Sample #1	0.179	22:16
Breath Volume	1.402	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample #2	0.165	22:20
Breath Volume	1.265	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 05/25/2007 04:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: NUGENT, DAVID, A

Breath Test Operator Agency: Clearwater PD

Arresting Officer: NUGENT, DAVID

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 21:00

Gender: Male

Observation Period Began: 21:04

Subject's Last Name: COOPER

Date/Time of Test: 04/20/2007 21:27

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.081	21:25	
Air Blank	0.000	21:26	
Subject Sample #1		21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:27	
Control Test	0.080	21:27	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:27	

Data Download Date/Time: 05/25/2007 04:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: FAULK, BRETT, A

Breath Test Operator Agency: Clearwater PD

Arresting Officer: LAMBE, JASON

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 02:32

Gender: Male

Observation Period Began: 03:32

Subject's Last Name: CLARKE

Date/Time of Test: 04/21/2007 04:03

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample #1	0.229	03:58
Breath Volume	4.710	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample #2	0.226	04:01
Breath Volume	4.710	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 05/25/2007 04:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILNE, THOMAS, P

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MILNE, THOMAS

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 00:20

Gender: Male

Observation Period Began: 00:47

Subject's Last Name: MUESSIG

Date/Time of Test: 04/22/2007 01:16

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.081	01:10
Air Blank	0.000	01:11
Subject Sample #1	0.108	01:11
Breath Volume	3.105	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample #2	0.109	01:14
Breath Volume	3.523	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 05/25/2007 04:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: YUEN, JONATHON, K

Breath Test Operator Agency: Clearwater PD

Arresting Officer: YUEN, JONATHON

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 04/10/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 23:42

Gender: Female

Observation Period Began: 23:50

Subject's Last Name: MCKENNEY

Date/Time of Test: 04/29/2007 00:13

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.081	00:12	
Air Blank	0.000	00:12	
Subject Sample #1		00:12	Sequence Aborted
Breath Volume	0.000	00:12	
Air Blank	0.000	00:13	

Data Download Date/Time: 05/25/2007 04:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FAZZINO, V

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 03:10

Gender: Male

Observation Period Began: 03:50

Subject's Last Name: MORO

Date/Time of Test: 04/01/2007 04:26

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:20
Air Blank	0.000	04:20
Subject Sample #1	0.140	04:21
Breath Volume	2.515	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.135	04:24
Breath Volume	2.554	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: O'NEAL, SHARI, L

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: ALTMAN, E

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 12:20

Gender: Female

Observation Period Began: 13:06

Subject's Last Name: WILLIAMS

Date/Time of Test: 04/03/2007 13:43

Diagnostics Check	OK	13:36
Air Blank	0.000	13:37
Control Test	0.080	13:37
Air Blank	0.000	13:38
Subject Sample #1	0.275	13:38
Breath Volume	1.507	13:38
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample #2	0.276	13:41
Breath Volume	1.132	13:41
Air Blank	0.000	13:42
Control Test	0.079	13:42
Air Blank	0.000	13:43
Diagnostics Check	OK	13:43

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: DORSEY, PAMELA, F

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: KEENE, T

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: BUI

Arrest Time: 00:20

Gender: Male

Observation Period Began: 00:49

Subject's Last Name: BROWN

Date/Time of Test: 04/04/2007 01:25

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample #1	0.206	01:20
Breath Volume	2.566	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample #2	0.206	01:23
Breath Volume	2.441	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: NEWMAN, C

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 00:19

Gender: Male

Observation Period Began: 01:10

Subject's Last Name: ELVIR MOLINA

Date/Time of Test: 04/07/2007 01:44

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample #1	0.162	01:40
Breath Volume	1.867	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample #2	0.158	01:43
Breath Volume	1.511	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: WALDRON, SHELLEY, H

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: OTT, M

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:10

Gender: Male

Observation Period Began: 01:59

Subject's Last Name: TREJO

Date/Time of Test: 04/07/2007 02:31

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Subject Sample #1	0.208	02:27
Breath Volume	1.730	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample #2	0.196	02:30
Breath Volume	1.714	02:30
Air Blank	0.000	02:30
Control Test	0.078	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FEOLA, A

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:16

Gender: Male

Observation Period Began: 02:20

Subject's Last Name: ALSTON

Date/Time of Test: 04/07/2007 02:57

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample #1	0.119	02:52
Breath Volume	1.164	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample #2	0.125	02:55
Breath Volume	1.730	02:55
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: PARENT, GARY, J

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: KATES-PAULUS, D

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 10:14

Gender: Male

Observation Period Began: 11:35

Subject's Last Name: MOULTON

Date/Time of Test: 04/07/2007 12:34

Diagnostics Check	OK	12:24
Air Blank	0.000	12:24
Control Test	0.079	12:25
Air Blank	0.000	12:25
Subject Sample #1	0.181	12:26
Breath Volume	3.011	12:26
Air Blank	0.000	12:26
Air Blank	0.000	12:28
Subject Sample #2	0.160	12:29
Breath Volume	1.445	12:29
Air Blank	0.000	12:29
Air Blank	0.000	12:31
Subject Sample #3	0.178	12:32
Breath Volume	2.742	12:32
Air Blank	0.000	12:33
Control Test	0.078	12:33
Air Blank	0.000	12:33
Diagnostics Check	OK	12:34

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FAZZINO, V

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 03:30

Gender: Male

Observation Period Began: 04:00

Subject's Last Name: SANCHEZ

Date/Time of Test: 04/08/2007 04:46

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:40
Subject Sample #1	0.186	04:41
Breath Volume	1.214	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample #2	0.186	04:45
Breath Volume	1.781	04:45
Air Blank	0.000	04:45
Control Test	0.078	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: JHONSON, A

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 06:50

Gender: Male

Observation Period Began: 07:30

Subject's Last Name: VELASCO MENDEZ

Date/Time of Test: 04/08/2007 08:04

Diagnostics Check	OK	07:58
Air Blank	0.000	07:58
Control Test	0.080	07:58
Air Blank	0.000	07:59
Subject Sample #1	0.170	07:59
Breath Volume	1.687	07:59
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample #2	0.172	08:02
Breath Volume	1.523	08:02
Air Blank	0.000	08:03
Control Test	0.079	08:03
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: ALTMAN, E

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 07:50

Gender: Male

Observation Period Began: 09:00

Subject's Last Name: HAROON

Date/Time of Test: 04/08/2007 09:39

Diagnostics Check	OK	09:30
Air Blank	0.000	09:31
Control Test	0.079	09:31
Air Blank	0.000	09:32
Subject Sample #1	0.155	09:34
Breath Volume	1.972	09:34
Air Blank	0.000	09:35
Air Blank	0.000	09:37
Subject Sample #2	0.148	09:37
Breath Volume	1.921	09:37
Air Blank	0.000	09:38
Control Test	0.077	09:38
Air Blank	0.000	09:38
Diagnostics Check	OK	09:39

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: DORSEY, PAMELA, F

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: DOERR, C

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 20:56

Gender: Female

Observation Period Began: 21:20

Subject's Last Name: AQUINO

Date/Time of Test: 04/09/2007 21:56

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:49
Subject Sample #1	0.182	21:51
Breath Volume	2.445	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:53
Subject Sample #2	0.183	21:55
Breath Volume	1.957	21:55
Air Blank	0.000	21:55
Control Test	0.078	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: WALDRON, SHELLEY, H

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FLOWERS, GARY

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:09

Gender: Male

Observation Period Began: 01:59

Subject's Last Name: ROACH

Date/Time of Test: 04/11/2007 02:39

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.078	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.127	02:31
Breath Volume	1.292	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample #2	0.158	02:34
Breath Volume	1.843	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample #3	0.131	02:37
Breath Volume	1.132	02:37
Air Blank	0.000	02:38
Control Test	0.077	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: WALDRON, SHELLEY, H

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: SUSSMAN, S

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:08

Gender: Male

Observation Period Began: 01:27

Subject's Last Name: BROWN

Date/Time of Test: 04/12/2007 02:05

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.079	01:54	
Air Blank	0.000	01:55	
Subject Sample #1	0.239	01:58	Volume Not Met
Breath Volume	0.031	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:00	
Subject Sample #2	0.165	02:04	Volume Not Met
Breath Volume	0.164	02:04	
Air Blank	0.000	02:04	
Control Test	0.076	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: WALDRON, SHELLEY, H

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: SUSSMAN, S

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:08

Gender: Male

Observation Period Began: 01:27

Subject's Last Name: BROWN

Date/Time of Test: 04/12/2007 02:14

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.077	02:08
Air Blank	0.000	02:08
Subject Sample #1	0.239	02:09
Breath Volume	2.027	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample #2	0.233	02:12
Breath Volume	1.390	02:12
Air Blank	0.000	02:13
Control Test	0.077	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: DORSEY, PAMELA, F

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: SUSSMAN, S

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 00:05

Gender: Male

Observation Period Began: 00:20

Subject's Last Name: ORTEGA

Date/Time of Test: 04/13/2007 00:56

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:45
Subject Sample #1	0.199	00:48
Breath Volume	2.468	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample #2	0.175	00:51
Breath Volume	1.421	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:53
Subject Sample #3	0.184	00:54
Breath Volume	1.656	00:54
Air Blank	0.000	00:55
Control Test	0.078	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: TALLENTIRE, M

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:25

Gender: Female

Observation Period Began: 02:45

Subject's Last Name: HENDRIX

Date/Time of Test: 04/13/2007 03:16

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.079	03:14	
Air Blank	0.000	03:14	
Subject Sample #1		03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.079	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: MORGAN, C

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 03/16/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 03:22

Gender: Female

Observation Period Began: 04:20

Subject's Last Name: MASSEY

Date/Time of Test: 04/13/2007 04:54

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:47
Subject Sample #1	0.169	04:49
Breath Volume	1.460	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample #2	0.149	04:53
Breath Volume	1.425	04:53
Air Blank	0.000	04:53
Control Test	0.078	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 04/13/2007 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: BLOESER, JR., R

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 18:30

Gender: Male

Observation Period Began: 19:30

Subject's Last Name: LUIS

Date/Time of Test: 04/13/2007 20:05

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.079	19:57
Air Blank	0.000	19:57
Subject Sample #1	0.183	20:00
Breath Volume	1.523	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample #2	0.175	20:04
Breath Volume	1.382	20:04
Air Blank	0.000	20:04
Control Test	0.077	20:05
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, 5

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: JONES, R

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:50

Gender: Male

Observation Period Began: 06:20

Subject's Last Name: CORNETT

Date/Time of Test: 04/14/2007 07:02

Diagnostics Check	OK	06:56
Air Blank	0.000	06:56
Control Test	0.079	06:56
Air Blank	0.000	06:57
Subject Sample #1	0.000	06:57
Breath Volume	2.960	06:57
Air Blank	0.000	06:58
Air Blank	0.000	07:00
Subject Sample #2	0.000	07:00
Breath Volume	2.664	07:00
Air Blank	0.000	07:01
Control Test	0.080	07:01
Air Blank	0.000	07:02
Diagnostics Check	OK	07:02

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: OWEN, SUE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: JUNTUNEN, H

Arresting Officer Agency: FFWCC (West Palm)

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 22:52

Gender: Male

Observation Period Began: 00:12

Subject's Last Name: FOLEY

Date/Time of Test: 04/15/2007 00:42

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample #1	0.149	00:38
Breath Volume	2.167	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample #2	0.135	00:41
Breath Volume	2.031	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: RIZZO, V

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 00:45

Gender: Male

Observation Period Began: 01:50

Subject's Last Name: GREER

Date/Time of Test: 04/15/2007 02:30

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:20	
Control Test	0.079	02:20	
Air Blank	0.000	02:21	
Subject Sample #1	0.216	02:24	Volume Not Met
Breath Volume	0.472	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:26	
Subject Sample #2		02:28	Subject Test Refused
Breath Volume	1.277	02:28	
Air Blank	0.000	02:29	
Control Test	0.076	02:30	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: OWEN, SUE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: DEVAUL, E

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 02:27

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: ALMENDARES

Date/Time of Test: 04/15/2007 03:45

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.078	03:34	
Air Blank	0.000	03:34	
Subject Sample #1	0.193	03:38	Volume Not Met
Breath Volume	0.121	03:38	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample #2	0.172	03:43	Volume Not Met
Breath Volume	0.105	03:43	
Air Blank	0.000	03:44	
Control Test	0.078	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: JOHNSON, G

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 02:55

Gender: Male

Observation Period Began: 03:40

Subject's Last Name: WOLFE, JR.

Date/Time of Test: 04/15/2007 04:14

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.077	04:07
Air Blank	0.000	04:07
Subject Sample #1	0.135	04:09
Breath Volume	1.265	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample #2	0.139	04:12
Breath Volume	2.210	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: CAMPBELL, G

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 03:59

Gender: Male

Observation Period Began: 04:52

Subject's Last Name: PEREZ

Date/Time of Test: 04/15/2007 05:30

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:25
Subject Sample #1	0.148	05:26
Breath Volume	1.488	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample #2	0.148	05:29
Breath Volume	1.332	05:29
Air Blank	0.000	05:29
Control Test	0.079	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: OLACIREGUI, M

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 04:40

Gender: Male

Observation Period Began: 05:21

Subject's Last Name: LUCAS

Date/Time of Test: 04/15/2007 06:17

Diagnostics Check	OK	06:06
Air Blank	0.000	06:07
Control Test	0.078	06:07
Air Blank	0.000	06:08
Subject Sample #1	0.256	06:09
Breath Volume	2.750	06:09
Air Blank	0.000	06:09
Air Blank	0.000	06:11
Subject Sample #2	0.226	06:12
Breath Volume	1.144	06:12
Air Blank	0.000	06:13
Air Blank	0.000	06:14
Subject Sample #3	0.255	06:15
Breath Volume	3.359	06:15
Air Blank	0.000	06:16
Control Test	0.078	06:16
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: WELLS, M

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 20:00

Gender: Male

Observation Period Began: 21:05

Subject's Last Name: SMITH

Date/Time of Test: 04/15/2007 21:44

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:38
Air Blank	0.000	21:38
Subject Sample #1	0.000	21:39
Breath Volume	1.425	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample #2	0.000	21:42
Breath Volume	1.226	21:42
Air Blank	0.000	21:43
Control Test	0.081	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: STCLOUD, F

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 20:50

Gender: Male

Observation Period Began: 22:05

Subject's Last Name: GONZALEZ

Date/Time of Test: 04/15/2007 22:46

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:32	
Control Test	0.079	22:32	
Air Blank	0.000	22:33	
Subject Sample #1	0.108	22:33	
Breath Volume	1.753	22:33	
Air Blank	0.000	22:34	
Air Blank	0.000	22:35	
Subject Sample #2	0.102	22:39	Volume Not Met
Breath Volume	0.191	22:39	
Air Blank	0.000	22:39	
Air Blank	0.000	22:41	
Subject Sample #3	0.114	22:44	Volume Not Met
Breath Volume	0.167	22:44	
Air Blank	0.000	22:45	
Control Test	0.077	22:45	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: MITIAL, J

Arresting Officer Agency: Delray Beach PD

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 00:22

Gender: Male

Observation Period Began: 01:05

Subject's Last Name: HAVERSTICK

Date/Time of Test: 04/17/2007 01:38

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample #1	0.210	01:33
Breath Volume	2.753	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample #2	0.204	01:36
Breath Volume	2.039	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: OWEN, SUE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FAZZINO, V

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 05:15

Gender: Female

Observation Period Began: 05:55

Subject's Last Name: DE JESUS

Date/Time of Test: 04/17/2007 06:24

Diagnostics Check	OK	06:17
Air Blank	0.000	06:17
Control Test	0.080	06:18
Air Blank	0.000	06:18
Subject Sample #1	0.219	06:19
Breath Volume	1.628	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample #2	0.213	06:22
Breath Volume	2.222	06:22
Air Blank	0.000	06:23
Control Test	0.079	06:23
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: DROSEY, PAMELA, F

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: RODRIGUEZ, T

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 18:15

Gender: Male

Observation Period Began: 18:35

Subject's Last Name: THOMPSON

Date/Time of Test: 04/17/2007 19:05

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.080	19:00
Air Blank	0.000	19:00
Subject Sample #1	0.173	19:01
Breath Volume	2.238	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:03
Subject Sample #2	0.182	19:04
Breath Volume	3.019	19:04
Air Blank	0.000	19:04
Control Test	0.079	19:05
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: PARENT, GARY, J

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FOLMAR, KEVIN

Arresting Officer Agency: Palm Beach Gardens

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 15:14

Gender: Male

Observation Period Began: 17:20

Subject's Last Name: ALEXANDER

Date/Time of Test: 04/18/2007 17:58

Diagnostics Check	OK	17:51
Air Blank	0.000	17:51
Control Test	0.080	17:52
Air Blank	0.000	17:52
Subject Sample #1	0.288	17:53
Breath Volume	3.796	17:53
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample #2	0.277	17:56
Breath Volume	2.839	17:56
Air Blank	0.000	17:57
Control Test	0.079	17:57
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: DORSEY, PAMELA, F

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: SUSSMAN, S

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:11

Gender: Male

Observation Period Began: 01:35

Subject's Last Name: CAMACHO

Date/Time of Test: 04/19/2007 02:13

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:58	
Subject Sample #1	0.152	02:00	
Breath Volume	1.335	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample #2	0.171	02:06	Volume Not Met
Breath Volume	0.015	02:06	
Air Blank	0.000	02:07	
Air Blank	0.000	02:09	
Subject Sample #3	0.157	02:12	Volume Not Met
Breath Volume	0.265	02:12	
Air Blank	0.000	02:12	
Control Test	0.078	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: DORSEY, PAMELA, F

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: SUSSMAN, S

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:11

Gender: Male

Observation Period Began: 01:35

Subject's Last Name: CAMACHO

Date/Time of Test: 04/19/2007 02:19

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:15	
Control Test	0.078	02:15	
Air Blank	0.000	02:16	
Subject Sample #1		02:18	Subject Test Refused
Breath Volume	0.500	02:18	
Air Blank	0.000	02:18	
Control Test	0.078	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: DORSEY, PAMELA, F

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: CARDASCIA, J

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:58

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: STICHTER

Date/Time of Test: 04/19/2007 03:04

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample #1	0.086	02:59
Breath Volume	3.382	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample #2	0.083	03:02
Breath Volume	2.535	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: COUNCIL, R

Arresting Officer Agency: Palm Beach Gardens

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 00:22

Gender: Female

Observation Period Began: 01:00

Subject's Last Name: WADLEY

Date/Time of Test: 04/20/2007 01:33

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:25
Subject Sample #1	0.174	01:28
Breath Volume	2.085	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:30
Subject Sample #2	0.177	01:31
Breath Volume	2.562	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:33

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: WENTZ, R

Arresting Officer Agency: Delray Beach PD

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 20:49

Gender: Male

Observation Period Began: 21:24

Subject's Last Name: AVILA

Date/Time of Test: 04/20/2007 22:07

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:01
Subject Sample #1	0.221	22:02
Breath Volume	1.511	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample #2	0.232	22:05
Breath Volume	2.355	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SEN, ARNOLD, N

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: SUSSMAN, S

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:15

Gender: Female

Observation Period Began: 01:50

Subject's Last Name: RODRIGUEZ BELTRE

Date/Time of Test: 04/21/2007 02:22

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample #1	0.155	02:17
Breath Volume	1.175	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample #2	0.165	02:21
Breath Volume	2.226	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: WALDRON, SHELLEY, H

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: ROSE, L

Arresting Officer Agency: Riviera Beach PD

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 05:45

Gender: Male

Observation Period Began: 06:50

Subject's Last Name: GARNER

Date/Time of Test: 04/21/2007 07:26

Diagnostics Check	OK	07:18
Air Blank	0.000	07:19
Control Test	0.079	07:19
Air Blank	0.000	07:20
Subject Sample #1	0.110	07:20
Breath Volume	1.281	07:20
Air Blank	0.000	07:21
Air Blank	0.000	07:23
Subject Sample #2	0.109	07:24
Breath Volume	2.867	07:24
Air Blank	0.000	07:25
Control Test	0.080	07:25
Air Blank	0.000	07:26
Diagnostics Check	OK	07:26

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: RODRIGUEZ, T

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 17:38

Gender: Female

Observation Period Began: 17:47

Subject's Last Name: YOUMANS

Date/Time of Test: 04/21/2007 18:23

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.080	18:15
Air Blank	0.000	18:15
Subject Sample #1	0.000	18:17
Breath Volume	1.902	18:17
Air Blank	0.000	18:18
Air Blank	0.000	18:20
Subject Sample #2	0.000	18:21
Breath Volume	1.367	18:21
Air Blank	0.000	18:22
Control Test	0.079	18:22
Air Blank	0.000	18:22
Diagnostics Check	OK	18:23

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: CROWLEY, P

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 20:39

Gender: Female

Observation Period Began: 21:02

Subject's Last Name: CONELLY

Date/Time of Test: 04/21/2007 21:41

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:29	
Control Test	0.079	21:29	
Air Blank	0.000	21:30	
Subject Sample #1	0.325	21:32	
Breath Volume	1.953	21:32	
Air Blank	0.000	21:33	
Air Blank	0.000	21:35	
Subject Sample #2	0.301	21:36	
Breath Volume	1.800	21:36	
Air Blank	0.000	21:37	
Air Blank	0.000	21:39	
Subject Sample #3	0.330	21:39	
Breath Volume	2.140	21:39	
Air Blank	0.000	21:40	
Control Test	0.074	21:40	Control Outside Tolerance
Air Blank	0.000	21:41	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: CROWLEY, P

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 20:39

Gender: Female

Observation Period Began: 21:02

Subject's Last Name: CONELLY

Date/Time of Test: 04/21/2007 21:58

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:45	
Control Test	0.077	21:45	
Air Blank	0.000	21:46	
Subject Sample #1	0.306	21:47	
Breath Volume	1.585	21:47	
Air Blank	0.000	21:48	
Air Blank	0.000	21:49	
Subject Sample #2	0.278	21:52	Volume Not Met
Breath Volume	0.000	21:52	
Air Blank	0.000	21:53	
Air Blank	0.000	21:55	
Subject Sample #3	0.311	21:56	
Breath Volume	1.441	21:56	
Air Blank	0.000	21:57	
Control Test	0.076	21:57	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: OWEN, SUE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: CROWLEY, P

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 05:16

Gender: Male

Observation Period Began: 05:50

Subject's Last Name: PAREDES

Date/Time of Test: 04/22/2007 06:25

Diagnostics Check	OK	06:14	
Air Blank	0.000	06:14	
Control Test	0.079	06:15	
Air Blank	0.000	06:15	
Subject Sample #1	0.136	06:18	Volume Not Met
Breath Volume	0.097	06:18	
Air Blank	0.000	06:19	
Air Blank	0.000	06:21	
Subject Sample #2		06:24	Subject Test Refused
Breath Volume	0.113	06:24	
Air Blank	0.000	06:24	
Control Test	0.078	06:25	
Air Blank	0.000	06:25	
Diagnostics Check	OK	06:25	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: EDELSTEIN, J

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 19:30

Gender: Male

Observation Period Began: 20:02

Subject's Last Name: MENDOZA GOMEZ

Date/Time of Test: 04/22/2007 20:44

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Subject Sample #1	0.194	20:39
Breath Volume	2.238	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample #2	0.193	20:42
Breath Volume	1.859	20:42
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:44

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: CROWLEY, P

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 20:04

Gender: Male

Observation Period Began: 20:35

Subject's Last Name: HEARRELL

Date/Time of Test: 04/22/2007 21:29

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Subject Sample #1	0.194	21:24
Breath Volume	2.675	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample #2	0.194	21:27
Breath Volume	2.914	21:27
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: DORSEY, PAMELA, F

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FULKES, C

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 17:45

Gender: Male

Observation Period Began: 18:40

Subject's Last Name: SAMAYAO

Date/Time of Test: 04/24/2007 19:15

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.081	19:09
Air Blank	0.000	19:10
Subject Sample #1	0.272	19:11
Breath Volume	2.109	19:11
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample #2	0.272	19:14
Breath Volume	1.285	19:14
Air Blank	0.000	19:14
Control Test	0.079	19:15
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: WALDRON, SHELLEY, H

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: PIOTROWSKI, M

Arresting Officer Agency: North Palm Beach

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 00:20

Gender: Male

Observation Period Began: 00:55

Subject's Last Name: DAVIS

Date/Time of Test: 04/25/2007 01:27

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Subject Sample #1	0.228	01:22
Breath Volume	1.660	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample #2	0.212	01:25
Breath Volume	1.175	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: WALDRON, SHELLEY, H

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: NEUMAN, C

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 03:00

Gender: Female

Observation Period Began: 03:56

Subject's Last Name: O LEARY

Date/Time of Test: 04/25/2007 04:29

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Subject Sample #1	0.211	04:20
Breath Volume	1.746	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.179	04:23
Breath Volume	1.902	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample #3	0.192	04:27
Breath Volume	1.394	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:29

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: WALDRON, SHELLEY, H

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: JENNE, T

Arresting Officer Agency: Jupiter PD

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 02:52

Gender: Male

Observation Period Began: 04:44

Subject's Last Name: INGLECIAS

Date/Time of Test: 04/25/2007 05:22

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.079	05:13
Air Blank	0.000	05:13
Subject Sample #1	0.249	05:14
Breath Volume	1.964	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:16
Subject Sample #2	0.279	05:17
Breath Volume	2.066	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample #3	0.273	05:20
Breath Volume	1.167	05:20
Air Blank	0.000	05:21
Control Test	0.078	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: WALDRON, SHELLEY, H

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: DOERR, C

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 03:09

Gender: Male

Observation Period Began: 03:35

Subject's Last Name: PETERS

Date/Time of Test: 04/26/2007 04:05

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Subject Sample #1	0.155	04:00
Breath Volume	2.242	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample #2	0.145	04:03
Breath Volume	2.824	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: O'NEAL, SHARI, L

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: TURNER, J

Arresting Officer Agency: Jupiter PD

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 01:38

Gender: Male

Observation Period Began: 02:27

Subject's Last Name: KALBERER

Date/Time of Test: 04/27/2007 03:00

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:54
Subject Sample #1	0.160	02:55
Breath Volume	1.898	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.159	02:58
Breath Volume	2.199	02:58
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	03:00

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: O'NEAL, SHARI, L

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: PRIDE, A

Arresting Officer Agency: FFWCC (West Palm)

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 02:49

Gender: Male

Observation Period Began: 03:24

Subject's Last Name: SEELEY

Date/Time of Test: 04/27/2007 03:56

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Subject Sample #1	0.229	03:52
Breath Volume	1.562	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample #2	0.229	03:55
Breath Volume	2.273	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FAZZINO, V

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 20:40

Gender: Male

Observation Period Began: 21:10

Subject's Last Name: PEREZ

Date/Time of Test: 04/27/2007 21:44

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.079	21:37
Air Blank	0.000	21:37
Subject Sample #1	0.121	21:38
Breath Volume	1.433	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample #2	0.117	21:42
Breath Volume	1.835	21:42
Air Blank	0.000	21:43
Control Test	0.079	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:44

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: COUNCIL, R

Arresting Officer Agency: Palm Beach Gardens

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 22:43

Gender: Male

Observation Period Began: 23:20

Subject's Last Name: JARRETT

Date/Time of Test: 04/28/2007 00:00

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Subject Sample #1	0.158	23:56
Breath Volume	1.890	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample #2	0.154	23:59
Breath Volume	2.109	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: OWEN, SUE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: TALLENTIRE, M

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 00:44

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: MONTOYA

Date/Time of Test: 04/28/2007 02:11

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Subject Sample #1	0.090	02:07
Breath Volume	1.730	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #2	0.090	02:09
Breath Volume	1.941	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: CAIN, JEANETTE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: SUSSMAN, S

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 03:15

Gender: Male

Observation Period Began: 03:37

Subject's Last Name: JIMENEZ

Date/Time of Test: 04/28/2007 04:12

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:05
Air Blank	0.000	04:05
Subject Sample #1	0.208	04:07
Breath Volume	2.414	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample #2	0.204	04:10
Breath Volume	3.308	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: OWEN, SUE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: MANCINO, D

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 07:03

Gender: Male

Observation Period Began: 07:35

Subject's Last Name: NAVAS

Date/Time of Test: 04/28/2007 08:09

Diagnostics Check	OK	08:02
Air Blank	0.000	08:02
Control Test	0.079	08:03
Air Blank	0.000	08:03
Subject Sample #1	0.164	08:04
Breath Volume	1.675	08:04
Air Blank	0.000	08:05
Air Blank	0.000	08:07
Subject Sample #2	0.163	08:07
Breath Volume	2.437	08:07
Air Blank	0.000	08:08
Control Test	0.079	08:08
Air Blank	0.000	08:09
Diagnostics Check	OK	08:09

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: DOSEY, PAMELA, F

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: BARBIRE, M

Arresting Officer Agency: Delray Beach PD

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 12:33

Gender: Male

Observation Period Began: 13:00

Subject's Last Name: RODOCKER

Date/Time of Test: 04/28/2007 13:49

Diagnostics Check	OK	13:42
Air Blank	0.000	13:42
Control Test	0.079	13:43
Air Blank	0.000	13:43
Subject Sample #1	0.222	13:44
Breath Volume	1.121	13:44
Air Blank	0.000	13:45
Air Blank	0.000	13:47
Subject Sample #2	0.216	13:47
Breath Volume	1.703	13:47
Air Blank	0.000	13:48
Control Test	0.078	13:48
Air Blank	0.000	13:49
Diagnostics Check	OK	13:49

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO
Software Version: 8100.27
Breath Test Operator: CAIN, JEANETTE,
Breath Test Operator Agency: Palm Beach County SO
Arresting Officer: BROTZ, K
Arresting Officer Agency: Delray Beach PD
Last Agency Inspection: 04/13/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 632102I
Dry Gas Expiry Date: 11/20/2008
Violation Code: DUI
Arrest Time: 20:06
Gender: Male
Observation Period Began: 20:35

Subject's Last Name: CRUDUP

Date/Time of Test: 04/28/2007 21:32

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:18	
Control Test	0.079	21:18	
Air Blank	0.000	21:18	
Subject Sample #1	0.084	21:22	Slope Not Level
Breath Volume	1.851	21:22	
Air Blank	0.000	21:22	
Air Blank	0.000	21:24	
Subject Sample #2	0.089	21:25	
Breath Volume	3.179	21:25	
Air Blank	0.000	21:26	
Air Blank	0.000	21:28	
Subject Sample #3		21:31	No Sample Provided
Breath Volume	0.000	21:31	
Air Blank	0.000	21:31	
Control Test	0.079	21:31	
Air Blank	0.000	21:32	
Diagnostics Check	OK	21:32	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: OWEN, SUE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: ZEITLER, J

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 04:30

Gender: Female

Observation Period Began: 05:46

Subject's Last Name: FREDRICKSON

Date/Time of Test: 04/29/2007 06:17

Diagnostics Check	OK	06:10
Air Blank	0.000	06:11
Control Test	0.079	06:11
Air Blank	0.000	06:12
Subject Sample #1	0.207	06:12
Breath Volume	1.582	06:12
Air Blank	0.000	06:13
Air Blank	0.000	06:15
Subject Sample #2	0.195	06:15
Breath Volume	1.703	06:15
Air Blank	0.000	06:16
Control Test	0.078	06:16
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: OWEN, SUE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: OKEEFE, J

Arresting Officer Agency: Jupiter PD

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 04:43

Gender: Female

Observation Period Began: 05:35

Subject's Last Name: TREVINO

Date/Time of Test: 04/30/2007 06:13

Diagnostics Check	OK	06:02	
Air Blank	0.000	06:02	
Control Test	0.079	06:02	
Air Blank	0.000	06:03	
Subject Sample #1	0.217	06:06	Volume Not Met
Breath Volume	0.433	06:06	
Air Blank	0.000	06:07	
Air Blank	0.000	06:08	
Subject Sample #2		06:12	No Sample Provided
Breath Volume	0.000	06:12	
Air Blank	0.000	06:12	
Control Test	0.077	06:12	
Air Blank	0.000	06:13	
Diagnostics Check	OK	06:13	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: OWEN, SUE,

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: OKEEFE, J

Arresting Officer Agency: Jupiter PD

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 04:43

Gender: Female

Observation Period Began: 06:06

Subject's Last Name: TREVINO

Date/Time of Test: 04/30/2007 06:38

Diagnostics Check	OK	06:31
Air Blank	0.000	06:32
Control Test	0.079	06:32
Air Blank	0.000	06:33
Subject Sample #1	0.243	06:33
Breath Volume	1.156	06:33
Air Blank	0.000	06:34
Air Blank	0.000	06:36
Subject Sample #2	0.240	06:37
Breath Volume	1.410	06:37
Air Blank	0.000	06:37
Control Test	0.078	06:38
Air Blank	0.000	06:38
Diagnostics Check	OK	06:38

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: O'NEAL, SHARI, L

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: RIEGER, M

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 12:51

Gender: Female

Observation Period Began: 14:07

Subject's Last Name: DIXON

Date/Time of Test: 04/30/2007 14:38

Diagnostics Check	OK	14:32
Air Blank	0.000	14:32
Control Test	0.079	14:33
Air Blank	0.000	14:33
Subject Sample #1	0.000	14:34
Breath Volume	1.335	14:34
Air Blank	0.000	14:34
Air Blank	0.000	14:36
Subject Sample #2	0.000	14:37
Breath Volume	1.347	14:37
Air Blank	0.000	14:37
Control Test	0.079	14:38
Air Blank	0.000	14:38
Diagnostics Check	OK	14:38

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: O'NEAL, SHARI, L

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: JEZEWSKI, B

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 04/13/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 15:15

Gender: Male

Observation Period Began: 15:38

Subject's Last Name: FLASTERSTEIN

Date/Time of Test: 04/30/2007 16:35

Diagnostics Check	OK	16:20	
Air Blank	0.000	16:21	
Control Test	0.078	16:21	
Air Blank	0.000	16:22	
Subject Sample #1	0.221	16:25	Volume Not Met
Breath Volume	0.000	16:25	
Air Blank	0.000	16:26	
Air Blank	0.000	16:27	
Subject Sample #2	0.225	16:29	
Breath Volume	1.195	16:29	
Air Blank	0.000	16:29	
Air Blank	0.000	16:31	
Subject Sample #3	0.225	16:33	
Breath Volume	1.226	16:33	
Air Blank	0.000	16:34	
Control Test	0.077	16:34	
Air Blank	0.000	16:34	
Diagnostics Check	OK	16:35	

Data Download Date/Time: 05/04/2007 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001865

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: FITZGERALD, F, F

Breath Test Operator Agency: Lakeland PD

Arresting Officer: HENNIS, G

Arresting Officer Agency: FHP Troop C

Last Agency Inspection: 03/14/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6191011

Dry Gas Expiry Date: 07/12/2008

Violation Code: DUI

Arrest Time: 01:36

Gender: Male

Observation Period Began: 02:05

Subject's Last Name: HELMLY

Date/Time of Test: 04/09/2007 02:41

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.082	02:35
Air Blank	0.000	02:36
Subject Sample #1	0.132	02:36
Breath Volume	1.945	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample #2	0.139	02:39
Breath Volume	2.476	02:39
Air Blank	0.000	02:40
Control Test	0.081	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 04/27/2007 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BUTTERWORTH, DAVID, S

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: THIERY, ROBERT

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 03/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 03:34

Gender: Male

Observation Period Began: 03:40

Subject's Last Name: CHEEK

Date/Time of Test: 04/01/2007 04:21

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.077	04:16
Air Blank	0.000	04:17
Subject Sample #1	0.234	04:17
Breath Volume	2.714	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample #2	0.250	04:20
Breath Volume	1.359	04:20
Air Blank	0.000	04:20
Control Test	0.078	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 04/11/2007 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: BRADY, KEVIN, E
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: BRADY, KEVIN
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 03/26/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 607401G
Dry Gas Expiry Date: 03/23/2008
Violation Code: DUI
Arrest Time: 03:17
Gender: Female
Observation Period Began: 03:45

Subject's Last Name: CEESAY

Date/Time of Test: 04/11/2007 04:19

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.077	04:13
Air Blank	0.000	04:14
Subject Sample #1	0.035	04:14
Breath Volume	1.332	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.035	04:18
Breath Volume	1.460	04:18
Air Blank	0.000	04:18
Control Test	0.076	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 04/11/2007 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BRADY, KEVIN, E

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: FAZIO, JOHN

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 04:38

Gender: Male

Observation Period Began: 04:50

Subject's Last Name: MORRELL

Date/Time of Test: 04/20/2007 05:33

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.077	05:22
Air Blank	0.000	05:23
Subject Sample #1	0.212	05:25
Breath Volume	1.781	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:27
Subject Sample #2	0.235	05:28
Breath Volume	1.261	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:30
Subject Sample #3	0.232	05:32
Breath Volume	2.984	05:32
Air Blank	0.000	05:32
Control Test	0.075	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 05/31/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BRADY, KEVIN, E

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: BELL, DAVID

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 04:58

Gender: Male

Observation Period Began: 05:15

Subject's Last Name: FICK

Date/Time of Test: 04/20/2007 05:52

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:40	
Control Test	0.077	05:40	
Air Blank	0.000	05:41	
Subject Sample #1	0.148	05:44	Volume Not Met
Breath Volume	0.296	05:44	
Air Blank	0.000	05:45	
Air Blank	0.000	05:46	
Subject Sample #2	0.120	05:47	
Breath Volume	1.789	05:47	
Air Blank	0.000	05:48	
Air Blank	0.000	05:50	
Subject Sample #3	0.127	05:51	
Breath Volume	1.925	05:51	
Air Blank	0.000	05:51	
Control Test	0.073	05:52	Control Outside Tolerance
Air Blank	0.000	05:52	

Data Download Date/Time: 05/31/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BRADY, KEVIN, E

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: BRADY, KEVIN

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 04/11/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 18:47

Gender: Male

Observation Period Began: 19:05

Subject's Last Name: CUMMINGS

Date/Time of Test: 04/24/2007 19:32

Diagnostics Check	OK	19:31	
Air Blank	0.000	19:32	
Control Test	0.074	19:32	Control Outside Tolerance
Air Blank	0.000	19:32	

Data Download Date/Time: 05/31/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: BRADY, KEVIN, E
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: BRADY, KEVIN
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 04/11/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 607401G
Dry Gas Expiry Date: 03/23/2008
Violation Code: DUI
Arrest Time: 18:43
Gender: Male
Observation Period Began: 19:05

Subject's Last Name: CUMMINGS

Date/Time of Test: 04/24/2007 19:40

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.078	19:34
Air Blank	0.000	19:35
Subject Sample #1	0.265	19:35
Breath Volume	1.726	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample #2	0.266	19:38
Breath Volume	3.593	19:38
Air Blank	0.000	19:39
Control Test	0.075	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 05/31/2007 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE

Software Version: 8100.27

Breath Test Operator: ROBERTS, OLIVE, M

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: PATTERSON, CORY

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 03/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 612501G

Dry Gas Expiry Date: 05/08/2008

Violation Code: DUI

Arrest Time: 00:19

Gender: Male

Observation Period Began: 00:45

Subject's Last Name: LOOKFONG

Date/Time of Test: 04/07/2007 01:37

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample #1	0.075	01:32
Breath Volume	2.082	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample #2	0.076	01:35
Breath Volume	2.250	01:35
Air Blank	0.000	01:36
Control Test	0.078	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/25/2007 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE

Software Version: 8100.27

Breath Test Operator: WHITE, SAMUEL, M

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: DOE, JOHN

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 03/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 612501G

Dry Gas Expiry Date: 05/08/2008

Violation Code: OTHER

Arrest Time: 21:08

Gender: Male

Observation Period Began: 13:20

Subject's Last Name: WHITE

Date/Time of Test: 04/17/2007 13:48

Diagnostics Check	OK	13:42
Air Blank	0.000	13:43
Control Test	0.080	13:43
Air Blank	0.000	13:44
Subject Sample #1	0.000	13:44
Breath Volume	6.617	13:44
Air Blank	0.000	13:44
Air Blank	0.000	13:46
Subject Sample #2	0.000	13:47
Breath Volume	5.839	13:47
Air Blank	0.000	13:47
Control Test	0.079	13:48
Air Blank	0.000	13:48
Diagnostics Check	OK	13:48

Data Download Date/Time: 04/25/2007 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE

Software Version: 8100.27

Breath Test Operator: CLARK, TEREZ, R

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: BRANCH, CLEOFUS

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 03/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 612501G

Dry Gas Expiry Date: 05/08/2008

Violation Code: OTHER

Arrest Time: 12:00

Gender: Female

Observation Period Began: 13:07

Subject's Last Name: GADZALLA

Date/Time of Test: 04/17/2007 13:58

Diagnostics Check	OK	13:54	
Air Blank	0.000	13:55	
Control Test	0.079	13:55	
Air Blank	0.000	13:55	
Subject Sample #1		13:57	Sequence Aborted
Breath Volume	0.035	13:57	
Air Blank	0.000	13:58	

Data Download Date/Time: 04/25/2007 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE

Software Version: 8100.27

Breath Test Operator: WISER, BUD, Q

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: MIGHTY, MOUSE

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 03/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 612501G

Dry Gas Expiry Date: 05/08/2008

Violation Code: OTHER

Arrest Time: 12:34

Gender: Female

Observation Period Began: 13:40

Subject's Last Name: MILLER

Date/Time of Test: 04/17/2007 14:07

Diagnostics Check	OK	14:04	
Air Blank	0.000	14:04	
Control Test	0.077	14:04	
Air Blank	0.000	14:05	
Subject Sample #1	0.000	14:05	
Breath Volume	1.460	14:05	
Air Blank		14:06	Improper Sample
Air Blank	0.000	14:06	
Control Test	0.078	14:06	
Air Blank		14:07	Sequence Aborted
Air Blank	0.000	14:07	

Data Download Date/Time: 04/25/2007 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE

Software Version: 8100.27

Breath Test Operator: ZITO, WILLIAM, A

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: CASERTA, CLIFFORD

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 03/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 612501G

Dry Gas Expiry Date: 05/08/2008

Violation Code: OTHER

Arrest Time: 13:00

Gender: Male

Observation Period Began: 13:53

Subject's Last Name: MISSEY

Date/Time of Test: 04/17/2007 14:17

Diagnostics Check	OK	14:13	
Air Blank	0.000	14:14	
Control Test	0.079	14:14	
Air Blank	0.000	14:15	
Subject Sample #1		14:15	Slope Not Met
Breath Volume	1.054	14:15	
Air Blank	0.000	14:16	
Control Test	0.071	14:16	Control Outside Tolerance
Air Blank	0.000	14:17	

Data Download Date/Time: 04/25/2007 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE

Software Version: 8100.27

Breath Test Operator: YOUNG, CHARLES, L

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: YOUNG, CHARLES

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 03/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 612501G

Dry Gas Expiry Date: 05/08/2008

Violation Code: OTHER

Arrest Time: 13:20

Gender: Male

Observation Period Began: 13:50

Subject's Last Name: MISSEY

Date/Time of Test: 04/17/2007 14:22

Diagnostics Check	OK	14:19	
Air Blank	0.000	14:19	
Control Test	0.079	14:20	
Air Blank	0.000	14:20	
Subject Sample #1		14:20	Slope Not Met
Breath Volume	2.867	14:20	
Air Blank	0.000	14:21	
Control Test	0.080	14:21	
Air Blank	0.000	14:22	
Diagnostics Check	OK	14:22	

Data Download Date/Time: 04/25/2007 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE

Software Version: 8100.27

Breath Test Operator: WRIGHT, ANDREW, E

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: GARDZALLA, RPBERT

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 03/15/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 612501G

Dry Gas Expiry Date: 05/08/2008

Violation Code: OTHER

Arrest Time: 22:45

Gender: Male

Observation Period Began: 23:50

Subject's Last Name: CURE

Date/Time of Test: 04/21/2007 00:31

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:25
Air Blank	0.000	00:25
Subject Sample #1	0.132	00:26
Breath Volume	3.183	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample #2	0.133	00:29
Breath Volume	3.179	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:31

Data Download Date/Time: 04/25/2007 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE

Software Version: 8100.27

Breath Test Operator: ROBERTS, OLIVE, M

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: SCOTT, ANDREW

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 04/25/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 612501G

Dry Gas Expiry Date: 05/08/2008

Violation Code: ZERO TOL

Arrest Time: 02:04

Gender: Male

Observation Period Began: 02:45

Subject's Last Name: SWARTZBERG

Date/Time of Test: 04/27/2007 03:14

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Subject Sample #1	0.183	03:10
Breath Volume	5.734	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample #2	0.184	03:13
Breath Volume	4.253	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 05/25/2007 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE
Software Version: 8100.27
Breath Test Operator: ROBERTS, OLIVE, M
Breath Test Operator Agency: NAS Jacksonville PD
Arresting Officer: SCOTT, ANDREW
Arresting Officer Agency: NAS Jacksonville PD
Last Agency Inspection: 04/25/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 612501G
Dry Gas Expiry Date: 05/08/2008
Violation Code: ZERO TOL
Arrest Time: 02:04
Gender: Female
Observation Period Began: 02:50

Subject's Last Name: SCHECKEL

Date/Time of Test: 04/27/2007 03:37

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:32	
Subject Sample #1	0.088	03:32	
Breath Volume	3.316	03:32	
Air Blank	0.000	03:33	
Air Blank	0.000	03:35	
Subject Sample #2	0.083	03:35	
Breath Volume	3.398	03:35	
Air Blank	0.000	03:36	
Control Test	0.073	03:36	Control Outside Tolerance
Air Blank	0.000	03:37	

Data Download Date/Time: 05/25/2007 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: HOLCOMB, BRUCE, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: KIRKLAND, RICHARD

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 03:18

Gender: Male

Observation Period Began: 03:35

Subject's Last Name: WATSON

Date/Time of Test: 04/02/2007 03:58

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.081	03:56	
Air Blank	0.000	03:57	
Subject Sample #1		03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.081	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 04/26/2007 12:29

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: PAGE, BRIAN, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: BASS, DANIEL

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 01:40

Gender: Male

Observation Period Began: 01:58

Subject's Last Name: REED

Date/Time of Test: 04/05/2007 02:27

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.080	02:22	
Air Blank	0.000	02:23	
Subject Sample #1		02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Control Test	0.080	02:26	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: BASS, DAN

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 01:11

Gender: Male

Observation Period Began: 01:25

Subject's Last Name: KING

Date/Time of Test: 04/06/2007 01:48

Diagnostics Check	OK	01:47	
Air Blank		01:48	RFI Detect
Air Blank	0.000	01:48	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: BASS, DAN

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 01:11

Gender: Male

Observation Period Began: 01:25

Subject's Last Name: KING

Date/Time of Test: 04/06/2007 01:57

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.081	01:54	
Air Blank	0.000	01:55	
Subject Sample #1		01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:56	
Control Test	0.081	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: HOLCOMB, BRUCE, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: KIRKLAND, RICHARD

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 02:10

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: BAKER

Date/Time of Test: 04/06/2007 02:58

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.081	02:56	
Air Blank	0.000	02:57	
Subject Sample #1		02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: REVE, JASON, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: REVE, JASON

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 02:29

Gender: Female

Observation Period Began: 02:39

Subject's Last Name: WHITTINGTON

Date/Time of Test: 04/09/2007 03:03

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.081	03:00	
Air Blank	0.000	03:01	
Subject Sample #1		03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.081	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: FREEMAN, DAVID

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 20:13

Gender: Male

Observation Period Began: 20:20

Subject's Last Name: MONNETTE

Date/Time of Test: 04/14/2007 20:48

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:43
Subject Sample #1	0.250	20:44
Breath Volume	2.445	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:46
Subject Sample #2	0.236	20:47
Breath Volume	1.917	20:47
Air Blank	0.000	20:47
Control Test	0.080	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: PAGE, BRAIN

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 21:00

Gender: Male

Observation Period Began: 22:00

Subject's Last Name: HERRERA

Date/Time of Test: 04/14/2007 22:26

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.079	22:21
Air Blank	0.000	22:21
Subject Sample #1	0.161	22:22
Breath Volume	3.011	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample #2	0.158	22:24
Breath Volume	3.394	22:24
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: PAGE, BRIAN, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: KIRKLAND, RICHARD

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 23:20

Gender: Male

Observation Period Began: 23:32

Subject's Last Name: SIMKANICH

Date/Time of Test: 04/14/2007 23:57

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.080	23:54	
Air Blank	0.000	23:54	
Subject Sample #1		23:56	Subject Test Refused
Breath Volume	0.000	23:56	
Air Blank	0.000	23:56	
Control Test	0.080	23:57	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: FREEMAN, DAVID

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 02:01

Gender: Male

Observation Period Began: 02:10

Subject's Last Name: CUTRONE

Date/Time of Test: 04/15/2007 02:34

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Subject Sample #1		02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.081	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: DELAHANTY, MICHAEL, T

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: CLANCY, KELLY

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 20:48

Gender: Male

Observation Period Began: 21:04

Subject's Last Name: MCGARVEY

Date/Time of Test: 04/17/2007 21:30

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.081	21:25
Air Blank	0.000	21:25
Subject Sample #1	0.029	21:26
Breath Volume	2.183	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample #2	0.028	21:29
Breath Volume	2.921	21:29
Air Blank	0.000	21:29
Control Test	0.081	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: KIRKLAND, RICHARD

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 21:19

Gender: Female

Observation Period Began: 21:25

Subject's Last Name: PLUMMER

Date/Time of Test: 04/18/2007 21:52

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.081	21:46
Air Blank	0.000	21:46
Subject Sample #1	0.215	21:47
Breath Volume	1.312	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample #2	0.215	21:50
Breath Volume	1.289	21:50
Air Blank	0.000	21:51
Control Test	0.079	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: PAGE, BRIAN, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: BASS, DANIEL

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 00:59

Gender: Male

Observation Period Began: 01:23

Subject's Last Name: MICHELBRINK

Date/Time of Test: 04/19/2007 01:49

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:44
Subject Sample #1	0.160	01:45
Breath Volume	2.742	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:47
Subject Sample #2	0.164	01:48
Breath Volume	3.007	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: HOLCOMB, BRUCE, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: KIRKLAND, RICHARD

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 04:10

Gender: Male

Observation Period Began: 04:25

Subject's Last Name: SWAN

Date/Time of Test: 04/19/2007 04:46

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:45	
Control Test	0.058	04:46	RFI Detect
Air Blank	0.000	04:46	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: HOLCOMB, BRUCE, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: KIRKLAND, RICHARD

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 04:10

Gender: Male

Observation Period Began: 04:25

Subject's Last Name: SWAN

Date/Time of Test: 04/19/2007 04:56

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:53	
Control Test	0.081	04:54	
Air Blank	0.000	04:54	
Subject Sample #1		04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:55	
Control Test	0.081	04:55	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: BASS, DAN

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 22:20

Gender: Male

Observation Period Began: 22:40

Subject's Last Name: BUMSTEAD

Date/Time of Test: 04/19/2007 23:03

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.080	23:01	
Air Blank	0.000	23:01	
Subject Sample #1		23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.079	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: DELAHANTY, MICHAEL, T

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: CLANCY, KELLY

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 19:44

Gender: Female

Observation Period Began: 19:59

Subject's Last Name: MEZA

Date/Time of Test: 04/20/2007 20:22

Diagnostics Check	OK	20:19	
Air Blank	0.000	20:20	
Control Test	0.081	20:20	
Air Blank	0.000	20:20	
Subject Sample #1		20:21	Subject Test Refused
Breath Volume	0.000	20:21	
Air Blank	0.000	20:21	
Control Test	0.081	20:21	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: DEAN, JENNIFER, D

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: DEAN, JENNIFER

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 18:42

Gender: Female

Observation Period Began: 18:55

Subject's Last Name: MAJKICH

Date/Time of Test: 04/21/2007 19:24

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.078	19:18
Air Blank	0.000	19:18
Subject Sample #1	0.158	19:19
Breath Volume	1.992	19:19
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample #2	0.156	19:22
Breath Volume	2.703	19:22
Air Blank	0.000	19:23
Control Test	0.079	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: PAGE, BRIAN, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: HOLCOMB, BRUCE

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 02:07

Gender: Male

Observation Period Began: 02:42

Subject's Last Name: GARCIA

Date/Time of Test: 04/24/2007 03:09

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:04
Subject Sample #1	0.023	03:04
Breath Volume	3.316	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample #2	0.022	03:07
Breath Volume	3.585	03:07
Air Blank	0.000	03:08
Control Test	0.081	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: HOLCOMB, BRUCE, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: KIRKLAND, RICHARD

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 03/28/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 22:05

Gender: Male

Observation Period Began: 22:20

Subject's Last Name: BARKER

Date/Time of Test: 04/24/2007 22:44

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:41	
Control Test	0.080	22:41	
Air Blank	0.000	22:42	
Subject Sample #1		22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.080	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 04/26/2007 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: REVE, JASON, R

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: REVE, JASON

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: BUI

Arrest Time: 00:42

Gender: Female

Observation Period Began: 00:49

Subject's Last Name: MCKENNEY

Date/Time of Test: 04/27/2007 01:26

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.081	01:20
Air Blank	0.000	01:20
Subject Sample #1	0.197	01:21
Breath Volume	1.449	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample #2	0.191	01:24
Breath Volume	2.003	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 05/30/2007 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: BASS, DANIEL

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: Yes

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 18:02

Gender: Male

Observation Period Began: 19:00

Subject's Last Name: HOOVER

Date/Time of Test: 04/27/2007 19:26

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.081	19:21
Air Blank	0.000	19:21
Subject Sample #1	0.170	19:22
Breath Volume	3.582	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample #2	0.167	19:25
Breath Volume	3.050	19:25
Air Blank	0.000	19:25
Control Test	0.079	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 05/30/2007 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: INKHEN, MATT

Arresting Officer Agency: Volusia County SO

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 02:23

Gender: Male

Observation Period Began: 02:45

Subject's Last Name: MILLER

Date/Time of Test: 04/28/2007 03:11

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Subject Sample #1	0.163	03:07
Breath Volume	2.882	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample #2	0.163	03:10
Breath Volume	2.246	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 05/30/2007 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: BASS, DANIEL

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 18:55

Gender: Male

Observation Period Began: 19:10

Subject's Last Name: GUEST

Date/Time of Test: 04/28/2007 19:38

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.081	19:33
Air Blank	0.000	19:33
Subject Sample #1	0.241	19:34
Breath Volume	2.050	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:36
Subject Sample #2	0.245	19:37
Breath Volume	2.414	19:37
Air Blank	0.000	19:38
Control Test	0.080	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 05/30/2007 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: KIRKLAND, RICHARD

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 01:26

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: MCCANN

Date/Time of Test: 04/29/2007 02:27

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample #1	0.272	02:22
Breath Volume	3.535	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample #2	0.258	02:25
Breath Volume	2.332	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 05/30/2007 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002231

Instrument Registered To: NEW SYMRNA BEACH PD

Software Version: 8100.27

Breath Test Operator: FREEMAN, DAVID, A

Breath Test Operator Agency: New Smyrna Beach PD

Arresting Officer: BASS, DANIEL

Arresting Officer Agency: New Smyrna Beach PD

Last Agency Inspection: 04/26/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 6209011

Dry Gas Expiry Date: 08/02/2008

Violation Code: DUI

Arrest Time: 02:52

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: PRICE

Date/Time of Test: 04/29/2007 03:23

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:21	
Subject Sample #1		03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 05/30/2007 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SAYEGH, MICHELLE, M

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: VESKI, MICHAEL

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 02:37

Gender: Male

Observation Period Began: 03:03

Subject's Last Name: WITKOWSKI

Date/Time of Test: 04/03/2007 03:27

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:26	
Subject Sample #1		03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SAYEGH, MICHELLE, M

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SAYEGH, MICHELLE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 03:39

Gender: Female

Observation Period Began: 03:56

Subject's Last Name: ORTIZ

Date/Time of Test: 04/04/2007 04:21

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:18	
Control Test	0.079	04:18	
Air Blank	0.000	04:19	
Subject Sample #1		04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Control Test	0.080	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 03:04

Gender: Male

Observation Period Began: 03:10

Subject's Last Name: FAJARDO

Date/Time of Test: 04/07/2007 03:34

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Subject Sample #1		03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.079	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MCVEY, GREGORY, B

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MCVEY, GREGORY

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 04:30

Gender: Male

Observation Period Began: 05:07

Subject's Last Name: BAEZ

Date/Time of Test: 04/07/2007 05:38

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.078	05:32
Air Blank	0.000	05:32
Subject Sample #1	0.256	05:33
Breath Volume	2.113	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample #2	0.241	05:36
Breath Volume	1.210	05:36
Air Blank	0.000	05:37
Control Test	0.077	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 00:50

Gender: Female

Observation Period Began: 01:10

Subject's Last Name: TASHMAN

Date/Time of Test: 04/09/2007 01:38

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.078	01:31
Air Blank	0.000	01:31
Subject Sample #1	0.116	01:32
Breath Volume	1.140	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample #2	0.124	01:36
Breath Volume	3.527	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:38

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MILLAN, KEVIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 04:13

Gender: Male

Observation Period Began: 04:22

Subject's Last Name: POSEY

Date/Time of Test: 04/09/2007 04:52

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:50	
Control Test	0.079	04:50	
Air Blank	0.000	04:51	
Subject Sample #1		04:51	Subject Test Refused
Breath Volume	0.000	04:51	
Air Blank	0.000	04:52	
Control Test	0.079	04:52	
Air Blank	0.000	04:52	
Diagnostics Check	OK	04:52	

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: COSNER, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 04:10

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: BELISSMELIS

Date/Time of Test: 04/13/2007 04:59

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.078	04:54
Air Blank	0.000	04:54
Subject Sample #1	0.128	04:54
Breath Volume	2.171	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample #2	0.128	04:57
Breath Volume	1.726	04:57
Air Blank	0.000	04:58
Control Test	0.077	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GARCIA, EDUARDO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 05:40

Gender: Male

Observation Period Began: 05:45

Subject's Last Name: ELIAS

Date/Time of Test: 04/14/2007 06:11

Diagnostics Check	OK	06:05
Air Blank	0.000	06:05
Control Test	0.079	06:06
Air Blank	0.000	06:06
Subject Sample #1	0.123	06:07
Breath Volume	4.800	06:07
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample #2	0.111	06:10
Breath Volume	2.875	06:10
Air Blank	0.000	06:10
Control Test	0.079	06:11
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SAYEGH, MICHELLE, M

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: VESKI, MICHAEL

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 23:47

Gender: Male

Observation Period Began: 00:04

Subject's Last Name: KUAN

Date/Time of Test: 04/15/2007 00:28

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.078	00:25	
Air Blank	0.000	00:26	
Subject Sample #1		00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:27	
Control Test	0.079	00:27	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GARCIA, EDUARDO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: GARCIA, EDUARDO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 04:40

Gender: Male

Observation Period Began: 04:48

Subject's Last Name: BIDDLE

Date/Time of Test: 04/15/2007 05:15

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:13	
Control Test	0.078	05:13	
Air Blank	0.000	05:14	
Subject Sample #1		05:14	Subject Test Refused
Breath Volume	0.000	05:14	
Air Blank	0.000	05:14	
Control Test	0.078	05:15	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: PEREZ, LUIS, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 05:46

Gender: Male

Observation Period Began: 06:10

Subject's Last Name: VEGA SALGADO

Date/Time of Test: 04/18/2007 06:39

Diagnostics Check	OK	06:31
Air Blank	0.000	06:32
Control Test	0.080	06:32
Air Blank	0.000	06:33
Subject Sample #1	0.164	06:34
Breath Volume	1.730	06:34
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample #2	0.162	06:38
Breath Volume	1.839	06:38
Air Blank	0.000	06:38
Control Test	0.076	06:39
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: COSNER, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 01:21

Gender: Male

Observation Period Began: 01:50

Subject's Last Name: RICCIO

Date/Time of Test: 04/21/2007 02:26

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.079	02:23	
Air Blank	0.000	02:24	
Subject Sample #1		02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.080	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ROJO, MARIO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 02:30

Gender: Female

Observation Period Began: 03:00

Subject's Last Name: SWEENEY

Date/Time of Test: 04/21/2007 03:28

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:26	
Subject Sample #1		03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.079	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GONZALEZ, JOEL,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: GARCIA, EDUARDO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 03/27/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 03:45

Gender: Male

Observation Period Began: 03:59

Subject's Last Name: BOENINGHAUS

Date/Time of Test: 04/22/2007 04:27

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.076	04:21	
Air Blank	0.000	04:21	
Subject Sample #1	0.248	04:22	
Breath Volume	1.664	04:22	
Air Blank	0.000	04:23	
Air Blank	0.000	04:24	
Subject Sample #2	0.264	04:25	
Breath Volume	1.921	04:25	
Air Blank	0.000	04:26	
Control Test	0.074	04:26	Control Outside Tolerance
Air Blank	0.000	04:27	

Data Download Date/Time: 04/24/2007 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 04/24/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 05:31

Gender: Male

Observation Period Began: 05:46

Subject's Last Name: JUSTILIEN

Date/Time of Test: 04/27/2007 06:25

Diagnostics Check	OK	06:19
Air Blank	0.000	06:19
Control Test	0.079	06:19
Air Blank	0.000	06:20
Subject Sample #1	0.172	06:20
Breath Volume	2.933	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample #2	0.162	06:24
Breath Volume	2.261	06:24
Air Blank	0.000	06:25
Control Test	0.077	06:25
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 05/30/2007 07:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: RIVERO, MARLEN, M

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MORALES, A

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 04/24/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 634801G

Dry Gas Expiry Date: 12/18/2008

Violation Code: DUI

Arrest Time: 05:07

Gender: Female

Observation Period Began: 05:15

Subject's Last Name: RAMOS

Date/Time of Test: 04/29/2007 05:43

Diagnostics Check	OK	05:35
Air Blank	0.000	05:36
Control Test	0.079	05:36
Air Blank	0.000	05:37
Subject Sample #1	0.125	05:38
Breath Volume	1.601	05:38
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample #2	0.131	05:41
Breath Volume	2.140	05:41
Air Blank	0.000	05:42
Control Test	0.079	05:42
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 05/30/2007 07:36