

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SANCHEZ, LUCIANO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MORALES ARGUETA

DATE/TIME OF TEST: 06/03/2011 23:27:40

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:22
Air Blank	0.000	23:22
Subject Sample #1	0.203	23:23
Breath Volume	2.109	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample #2	0.197	23:26
Breath Volume	1.437	23:26
Air Blank	0.000	23:26
Control Test	0.078	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time 06/09/2011 11:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LEIGHTON, FRANCISCO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: OLIVA

DATE/TIME OF TEST: 06/03/2011 23:53:10

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:47
Air Blank	0.000	23:47
Subject Sample #1	0.000	23:48
Breath Volume	2.011	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample #2	0.000	23:51
Breath Volume	1.972	23:51
Air Blank	0.000	23:52
Control Test	0.077	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time 06/09/2011 11:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ENRIQUEZ, JUAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: AHILON

DATE/TIME OF TEST: 06/04/2011 01:48:10

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:45	
Control Test	0.078	01:45	
Air Blank	0.000	01:46	
Subject Sample #1		01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Control Test	0.078	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time 06/09/2011 11:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CELLA, DOMENICO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 06/04/2011 02:01:09

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.079	01:58	
Air Blank	0.000	01:59	
Subject Sample #1		01:59	Slope Not Met
Breath Volume	1.003	01:59	
Air Blank	0.000	02:00	
Control Test	0.078	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time 06/09/2011 11:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VAN GILS, PATRICK
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: PINTO

DATE/TIME OF TEST: 06/04/2011 02:16:21

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Subject Sample #1	0.109	02:11
Breath Volume	1.804	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample #2	0.100	02:14
Breath Volume	1.218	02:14
Air Blank	0.000	02:15
Control Test	0.076	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time 06/09/2011 11:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MCGRATH, ROBERT
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: VALLEJO

DATE/TIME OF TEST: 06/04/2011 02:32:09

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.078	02:26
Air Blank	0.000	02:26
Subject Sample #1	0.165	02:27
Breath Volume	1.156	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:29
Subject Sample #2	0.153	02:30
Breath Volume	1.445	02:30
Air Blank	0.000	02:31
Control Test	0.077	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time 06/09/2011 11:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CELLA, DOMENICO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 06/04/2011 02:49:25

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.077	02:43
Air Blank	0.000	02:44
Subject Sample #1	0.221	02:44
Breath Volume	3.777	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample #2	0.224	02:47
Breath Volume	3.484	02:47
Air Blank	0.000	02:48
Control Test	0.076	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time 06/09/2011 11:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ, CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/04/2011 04:02:39

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.077	03:56	
Air Blank	0.000	03:56	
Subject Sample #1	0.239	03:58	
Breath Volume	1.730	03:58	
Air Blank	0.000	03:58	
Air Blank	0.000	04:00	
Subject Sample #2	0.243	04:01	
Breath Volume	1.882	04:01	
Air Blank	0.000	04:01	
Control Test	0.074	04:02	Control Outside Tolerance
Air Blank	0.000	04:02	

Data Download Date/Time 06/09/2011 11:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ, CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/04/2011 04:10:10

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.076	04:04
Air Blank	0.000	04:05
Subject Sample #1	0.239	04:05
Breath Volume	1.742	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample #2	0.235	04:08
Breath Volume	1.535	04:08
Air Blank	0.000	04:09
Control Test	0.076	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time 06/09/2011 11:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE ROSA, MARCOS, A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DE ROSA, MARCOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: RUBAL

DATE/TIME OF TEST: 06/11/2011 01:08:23

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.078	01:06	
Air Blank	0.000	01:06	
Subject Sample #1		01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CELLA, DOMENICO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/11/2011 01:28:10

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample #1	0.036	01:22
Breath Volume	1.867	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample #2	0.032	01:26
Breath Volume	1.246	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: STRINGER, ROBERT
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 06/11/2011 01:43:40

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample #1	0.020	01:37
Breath Volume	1.804	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample #2	0.017	01:42
Breath Volume	1.906	01:42
Air Blank	0.000	01:42
Control Test	0.077	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ, CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: BORDENAVE

DATE/TIME OF TEST: 06/11/2011 02:00:11

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Subject Sample #1	0.000	01:55
Breath Volume	2.136	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample #2	0.000	01:58
Breath Volume	1.453	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS, RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MIYAR

DATE/TIME OF TEST: 06/11/2011 02:22:27

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Subject Sample #1	0.053	02:17
Breath Volume	3.812	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample #2	0.052	02:20
Breath Volume	3.753	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TORRES, DAVID
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/11/2011 02:36:11

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:30
Subject Sample #1	0.135	02:31
Breath Volume	1.125	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample #2	0.124	02:34
Breath Volume	1.125	02:34
Air Blank	0.000	02:35
Control Test	0.077	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SANCHEZ, BRIDGET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CALVAIRE

DATE/TIME OF TEST: 06/11/2011 03:02:25

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.078	02:52
Air Blank	0.000	02:52
Subject Sample #1	0.266	02:54
Breath Volume	1.449	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:56
Subject Sample #2	0.239	02:57
Breath Volume	2.542	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample #3	0.266	03:00
Breath Volume	1.843	03:00
Air Blank	0.000	03:01
Control Test	0.077	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS, RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: ADMIN
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MIYAR

DATE/TIME OF TEST: 06/11/2011 03:28:44

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:24
Subject Sample #1	0.028	03:24
Breath Volume	1.875	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample #2	0.025	03:27
Breath Volume	2.113	03:27
Air Blank	0.000	03:27
Control Test	0.078	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ, MICHEAL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: LAJES

DATE/TIME OF TEST: 06/11/2011 03:58:02

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.077	03:55	
Air Blank	0.000	03:56	
Subject Sample #1		03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:57	
Control Test	0.078	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:58	

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AGUIAR, JORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/15/2011 21:46:38

Diagnostics Check	OK	21:33	
Air Blank	0.000	21:34	
Control Test	0.078	21:34	
Air Blank	0.000	21:34	
Subject Sample #1	0.216	21:38	Volume Not Met
Breath Volume	0.398	21:38	
Air Blank	0.000	21:38	
Air Blank	0.000	21:40	
Subject Sample #2	0.295	21:41	
Breath Volume	3.179	21:41	
Air Blank	0.000	21:42	
Air Blank	0.000	21:43	
Subject Sample #3	0.274	21:45	
Breath Volume	1.859	21:45	
Air Blank	0.000	21:45	
Control Test	0.074	21:46	Control Outside Tolerance
Air Blank	0.000	21:46	

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AGUIAR, JORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/15/2011 22:07:42

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:54	
Control Test	0.076	21:54	
Air Blank	0.000	21:55	
Subject Sample #1	0.273	21:55	
Breath Volume	1.515	21:55	
Air Blank	0.000	21:56	
Air Blank	0.000	21:58	
Subject Sample #2	0.208	22:01	Volume Not Met
Breath Volume	0.734	22:01	
Air Blank	0.000	22:01	
Air Blank	0.000	22:03	
Subject Sample #3		22:06	Subject Test Refused
Breath Volume	0.027	22:06	
Air Blank	0.000	22:06	
Control Test	0.076	22:07	
Air Blank	0.000	22:07	
Diagnostics Check	OK	22:07	

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ACOSTA, ALBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/15/2011 23:32:41

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:30	
Control Test	0.078	23:30	
Air Blank	0.000	23:31	
Subject Sample #1		23:31	Subject Test Refused
Breath Volume	0.000	23:31	
Air Blank	0.000	23:31	
Control Test	0.078	23:32	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO, GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: MONTERO

DATE/TIME OF TEST: 06/16/2011 00:19:43

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Subject Sample #1	0.000	00:15
Breath Volume	2.281	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample #2	0.000	00:18
Breath Volume	1.542	00:18
Air Blank	0.000	00:18
Control Test	0.078	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MCGRATH, ROBERT
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: CORTINA

DATE/TIME OF TEST: 06/16/2011 00:46:29

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.078	00:41
Air Blank	0.000	00:41
Subject Sample #1	0.000	00:42
Breath Volume	2.636	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample #2	0.000	00:45
Breath Volume	2.300	00:45
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GONZALEZ, JORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 06/16/2011 02:45:55

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.076	02:43	
Air Blank	0.000	02:44	
Subject Sample #1		02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Control Test	0.077	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS, RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/16/2011 02:59:42

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.077	02:53
Air Blank	0.000	02:54
Subject Sample #1	0.144	02:55
Breath Volume	4.183	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.145	02:58
Breath Volume	4.699	02:58
Air Blank	0.000	02:58
Control Test	0.077	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time 07/13/2011 13:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ADAMS, STEVEN, L

BREATH TEST OPERATOR AGENCY: Miami-Dade PD

ARRESTING OFFICER: ADAMS, STEVEN

ARRESTING OFFICER AGENCY: Miami-Dade PD

LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 688217

DRY GAS EXPIRY DATE: 11/09/2011

VIOLATION CODE: DUI

ARREST TIME: 22:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: CABAL

DATE/TIME OF TEST: 06/01/2011 23:15:17

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:09
Subject Sample #1	0.126	23:10
Breath Volume	1.257	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample #2	0.136	23:13
Breath Volume	3.667	23:13
Air Blank	0.000	23:14
Control Test	0.078	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time 06/10/2011 14:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS, STEVEN, L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS, STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 06/03/2011 01:37:52

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample #1	0.181	01:33
Breath Volume	1.292	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample #2	0.185	01:36
Breath Volume	1.421	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time 06/10/2011 14:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO, GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO, GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MERLAN

DATE/TIME OF TEST: 06/04/2011 03:21:47

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample #1	0.128	03:16
Breath Volume	1.417	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample #2	0.128	03:20
Breath Volume	2.332	03:20
Air Blank	0.000	03:20
Control Test	0.077	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time 06/10/2011 14:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS, STEVEN, L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS, STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GUARDIA-MARTINEC

DATE/TIME OF TEST: 06/07/2011 04:51:16

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:45
Air Blank	0.000	04:45
Subject Sample #1	0.239	04:46
Breath Volume	2.000	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample #2	0.232	04:49
Breath Volume	2.003	04:49
Air Blank	0.000	04:50
Control Test	0.078	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time 06/10/2011 14:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEY, SHANNON, C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JARRELL, C.E.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/16/2011 03:01:27

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Subject Sample #1	0.154	02:56
Breath Volume	3.015	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.151	02:59
Breath Volume	3.316	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIDINE, W, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOUSEND, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 06/17/2011 19:43:43

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.078	19:37
Air Blank	0.000	19:37
Subject Sample #1	0.268	19:39
Breath Volume	2.675	19:39
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample #2	0.274	19:42
Breath Volume	2.792	19:42
Air Blank	0.000	19:42
Control Test	0.077	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIDINE, W, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ARCHBOLD, N.O.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: KOON

DATE/TIME OF TEST: 06/18/2011 21:30:03

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.079	21:24
Air Blank	0.000	21:25
Subject Sample #1	0.136	21:25
Breath Volume	2.156	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample #2	0.132	21:28
Breath Volume	1.972	21:28
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:30

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIDINE, W, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ELLIS, G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: LOTSEY

DATE/TIME OF TEST: 06/20/2011 17:06:54

Diagnostics Check	OK	17:00
Air Blank	0.000	17:01
Control Test	0.077	17:01
Air Blank	0.000	17:02
Subject Sample #1	0.336	17:02
Breath Volume	3.789	17:02
Air Blank	0.000	17:03
Air Blank	0.000	17:04
Subject Sample #2	0.337	17:05
Breath Volume	3.664	17:05
Air Blank	0.000	17:06
Control Test	0.077	17:06
Air Blank	0.000	17:06
Diagnostics Check	OK	17:06

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARRIDINE, W, L

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: ODER, C.D.

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 17:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: SRYBNYKH

DATE/TIME OF TEST: 06/20/2011 20:41:25

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:37	
Control Test	0.078	20:37	
Air Blank	0.000	20:38	
Subject Sample #1		20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:40	
Control Test	0.078	20:40	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:41	

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN, C, M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBERTS, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 06/22/2011 20:19:09

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.079	20:13
Air Blank	0.000	20:14
Subject Sample #1	0.000	20:14
Breath Volume	3.480	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample #2	0.000	20:17
Breath Volume	3.890	20:17
Air Blank	0.000	20:18
Control Test	0.078	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HALLIDAY, R, A

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: JOHNSTON, C

ARRESTING OFFICER AGENCY: Atlantic Beach PD

LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 00:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: HARRIES

DATE/TIME OF TEST: 06/23/2011 01:56:59

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Subject Sample #1	0.185	01:51
Breath Volume	1.492	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample #2	0.182	01:55
Breath Volume	1.207	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HALLIDAY, R, A

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: JARRELL, C

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 22:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 06/23/2011 23:33:16

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.078	23:27
Air Blank	0.000	23:27
Subject Sample #1	0.085	23:28
Breath Volume	3.335	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample #2	0.088	23:31
Breath Volume	2.875	23:31
Air Blank	0.000	23:32
Control Test	0.078	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUCHANAN, J, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HOOVER, J

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 17:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: VAUGHAN

DATE/TIME OF TEST: 06/24/2011 18:59:14

Diagnostics Check	OK	18:55	
Air Blank	0.000	18:56	
Control Test	0.078	18:56	
Air Blank	0.000	18:57	
Subject Sample #1		18:57	Subject Test Refused
Breath Volume	0.000	18:57	
Air Blank	0.000	18:58	
Control Test	0.078	18:58	
Air Blank	0.000	18:59	
Diagnostics Check	OK	18:59	

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCHANAN, J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCNAMEE, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 06/24/2011 22:27:35

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.079	22:18
Air Blank	0.000	22:19
Subject Sample #1	0.389	22:19
Breath Volume	1.242	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:21
Subject Sample #2	0.418	22:22
Breath Volume	2.593	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample #3	0.416	22:26
Breath Volume	2.031	22:26
Air Blank	0.000	22:26
Control Test	0.078	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEY, SHANNON, C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER, J.K.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 06/25/2011 00:54:47

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.078	00:49
Air Blank	0.000	00:49
Subject Sample #1	0.207	00:50
Breath Volume	2.484	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample #2	0.200	00:53
Breath Volume	2.003	00:53
Air Blank	0.000	00:53
Control Test	0.079	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULLER, ELWIN, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PROTONOTARIOUS, N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: PANDO

DATE/TIME OF TEST: 06/25/2011 07:36:50

Diagnostics Check	OK	07:30
Air Blank	0.000	07:31
Control Test	0.078	07:31
Air Blank	0.000	07:31
Subject Sample #1	0.097	07:32
Breath Volume	2.644	07:32
Air Blank	0.000	07:33
Air Blank	0.000	07:34
Subject Sample #2	0.095	07:35
Breath Volume	2.550	07:35
Air Blank	0.000	07:35
Control Test	0.078	07:36
Air Blank	0.000	07:36
Diagnostics Check	OK	07:36

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAMPBELL, C, J

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: JARRELL, C

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 00:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MYHAND

DATE/TIME OF TEST: 06/26/2011 02:32:09

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Subject Sample #1	0.038	02:27
Breath Volume	1.726	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample #2	0.038	02:30
Breath Volume	2.085	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAMPBELL, C, J

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HEALY, G

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 02:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 06/26/2011 04:12:15

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Subject Sample #1	0.058	04:07
Breath Volume	1.269	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.064	04:10
Breath Volume	2.496	04:10
Air Blank	0.000	04:11
Control Test	0.078	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAMPBELL, C, J

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: REDGATE, E

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 03:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: BACKES

DATE/TIME OF TEST: 06/26/2011 04:56:54

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.079	04:54	
Air Blank	0.000	04:55	
Subject Sample #1		04:55	Subject Test Refused
Breath Volume	0.000	04:55	
Air Blank	0.000	04:56	
Control Test	0.077	04:56	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARRIDINE, W, L

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: ODER, C

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 18:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 06/28/2011 20:37:33

Diagnostics Check	OK	20:31
Air Blank	0.000	20:32
Control Test	0.079	20:32
Air Blank	0.000	20:32
Subject Sample #1	0.201	20:33
Breath Volume	3.238	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample #2	0.184	20:36
Breath Volume	1.476	20:36
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time 06/30/2011 08:48

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN, RHONDA, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:03

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/30/2011 12:30:24

Diagnostics Check	OK	12:24
Air Blank	0.000	12:24
Control Test	0.079	12:25
Air Blank	0.000	12:25
Subject Sample #1	0.000	12:26
Breath Volume	1.925	12:26
Air Blank	0.000	12:26
Air Blank	0.000	12:28
Subject Sample #2	0.000	12:29
Breath Volume	4.109	12:29
Air Blank	0.000	12:29
Control Test	0.079	12:29
Air Blank	0.000	12:30
Diagnostics Check	OK	12:30

Data Download Date/Time 07/15/2011 12:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCHANAN, J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DUPUY, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: MCLENDON

DATE/TIME OF TEST: 06/30/2011 21:12:59

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.078	21:10	
Air Blank	0.000	21:11	
Subject Sample #1		21:11	Subject Test Refused
Breath Volume	0.000	21:11	
Air Blank	0.000	21:12	
Control Test	0.078	21:12	
Air Blank	0.000	21:12	
Diagnostics Check	OK	21:12	

Data Download Date/Time 07/15/2011 12:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BOSTON, D, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BAI, S

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 01011080A2

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 02:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ARENSDORF

DATE/TIME OF TEST: 06/01/2011 03:51:37

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample #1	0.164	03:46
Breath Volume	1.859	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample #2	0.160	03:50
Breath Volume	2.082	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS, T, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TAWFIK, H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01011080A2
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 06:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: BENESI

DATE/TIME OF TEST: 06/01/2011 09:27:13

Diagnostics Check	OK	09:21
Air Blank	0.000	09:21
Control Test	0.080	09:21
Air Blank	0.000	09:22
Subject Sample #1	0.060	09:22
Breath Volume	3.640	09:22
Air Blank	0.000	09:23
Air Blank	0.000	09:25
Subject Sample #2	0.054	09:25
Breath Volume	1.984	09:25
Air Blank	0.000	09:26
Control Test	0.081	09:26
Air Blank	0.000	09:27
Diagnostics Check	OK	09:27

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PRUETT, RICHARD, L

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: CHRISTMAS, F

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 01011080A2

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 23:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RIGDON

DATE/TIME OF TEST: 06/02/2011 00:51:48

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:46
Subject Sample #1	0.000	00:47
Breath Volume	1.589	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample #2	0.000	00:50
Breath Volume	1.757	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GIVENS, J, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: YARBROUGH, J

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 01011080A2

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 17:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: MOULTRIE

DATE/TIME OF TEST: 06/02/2011 20:03:36

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.080	19:58
Air Blank	0.000	19:58
Subject Sample #1	0.291	19:59
Breath Volume	1.292	19:59
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample #2	0.283	20:02
Breath Volume	1.761	20:02
Air Blank	0.000	20:02
Control Test	0.080	20:03
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEY, SHANNON, C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HEALEY, G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01011080A2
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: RALPH

DATE/TIME OF TEST: 06/03/2011 04:06:31

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:04	
Control Test	0.081	04:04	
Air Blank	0.000	04:04	
Subject Sample #1		04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.080	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEY, SHANNON, C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ARCHBOLD, N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01011080A2
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 06/03/2011 05:43:21

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:41	
Control Test	0.081	05:41	
Air Blank	0.000	05:41	
Subject Sample #1		05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA, K, M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALDREP, B
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/03/2011 19:07:12

Diagnostics Check	OK	18:57
Air Blank	0.000	18:57
Control Test	0.080	18:58
Air Blank	0.000	18:58
Subject Sample #1	0.174	18:59
Breath Volume	1.343	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:01
Subject Sample #2	0.200	19:02
Breath Volume	2.000	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample #3	0.191	19:05
Breath Volume	1.609	19:05
Air Blank	0.000	19:06
Control Test	0.079	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEITH, C, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: ABBOTT, B

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 21:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DENMARK

DATE/TIME OF TEST: 06/03/2011 23:30:02

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:21
Air Blank	0.000	23:21
Subject Sample #1	0.184	23:23
Breath Volume	1.468	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample #2	0.180	23:28
Breath Volume	1.296	23:28
Air Blank	0.000	23:29
Control Test	0.078	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:30

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEITH, C, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BROWN, T

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 02:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/04/2011 04:19:39

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:17	
Subject Sample #1	0.145	04:17	
Breath Volume	2.589	04:17	
Air Blank		04:18	RFI Detect
Air Blank	0.000	04:18	
Control Test	0.080	04:19	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEITH, C, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: BROWN, T

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 02:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/04/2011 04:28:24

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Subject Sample #1	0.140	04:24
Breath Volume	2.648	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample #2	0.142	04:26
Breath Volume	2.519	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEY, SHANNON, C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAMLIN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/05/2011 03:21:01

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:19	
Subject Sample #1		03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:21	

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEY, SHANNON, C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DANZIE, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 06/05/2011 05:47:05

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.079	05:41
Air Blank	0.000	05:41
Subject Sample #1	0.177	05:42
Breath Volume	2.234	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:44
Subject Sample #2	0.174	05:45
Breath Volume	2.960	05:45
Air Blank	0.000	05:46
Control Test	0.080	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:47

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUCHANAN, J, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: SUTEK, S

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 21:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/05/2011 23:02:50

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.079	22:52	
Air Blank	0.000	22:53	
Subject Sample #1	0.036	22:56	
Breath Volume	2.566	22:56	
Air Blank	0.000	22:56	
Air Blank	0.000	22:58	
Subject Sample #2		23:01	Slope Not Met
Breath Volume	1.109	23:01	
Air Blank	0.000	23:01	
Control Test	0.080	23:02	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCHANAN, J, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SUTEK, S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/05/2011 23:33:30

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:27
Subject Sample #1	0.028	23:28
Breath Volume	2.261	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample #2	0.029	23:32
Breath Volume	2.179	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEY, SHANNON, C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VEATCH, K.C.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/08/2011 03:22:50

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:20	
Control Test	0.079	03:20	
Air Blank	0.000	03:21	
Subject Sample #1		03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HALLIDAY, R, A

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: HEALY, G

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 23:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: EASTERLING

DATE/TIME OF TEST: 06/11/2011 01:06:01

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:01
Subject Sample #1	0.229	01:01
Breath Volume	3.976	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:03
Subject Sample #2	0.231	01:04
Breath Volume	3.839	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:06

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND, J, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HEALY, G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/11/2011 02:51:53

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample #1	0.104	02:47
Breath Volume	1.679	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample #2	0.111	02:50
Breath Volume	2.906	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND, J, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 06/11/2011 04:09:41

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample #1	0.206	04:03
Breath Volume	1.972	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample #2	0.193	04:08
Breath Volume	2.558	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAND, J, D

BREATH TEST OPERATOR AGENCY: Jacksonville SO

ARRESTING OFFICER: O'BRITIS, A

ARRESTING OFFICER AGENCY: Jacksonville SO

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 03:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: KLINGER

DATE/TIME OF TEST: 06/12/2011 05:17:41

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:15	
Control Test	0.080	05:15	
Air Blank	0.000	05:15	
Subject Sample #1		05:16	Subject Test Refused
Breath Volume	0.000	05:16	
Air Blank	0.000	05:16	
Control Test	0.080	05:17	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN, C, M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHRISTMAS, F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 06/12/2011 19:19:05

Diagnostics Check	OK	19:12
Air Blank	0.000	19:13
Control Test	0.079	19:13
Air Blank	0.000	19:14
Subject Sample #1	0.111	19:14
Breath Volume	1.222	19:14
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample #2	0.113	19:17
Breath Volume	1.218	19:17
Air Blank	0.000	19:18
Control Test	0.080	19:18
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN, C, M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06111080A1
DRY GAS EXPIRY DATE: 03/01/2013
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: MORALES-LUCAS

DATE/TIME OF TEST: 06/14/2011 19:03:50

Diagnostics Check	OK	18:51	
Air Blank	0.000	18:52	
Control Test	0.079	18:52	
Air Blank	0.000	18:52	
Subject Sample #1	0.236	18:53	
Breath Volume	1.484	18:53	
Air Blank	0.000	18:53	
Air Blank	0.000	18:55	
Subject Sample #2	0.214	18:57	
Breath Volume	1.578	18:57	
Air Blank	0.000	18:58	
Air Blank	0.000	19:00	
Subject Sample #3		19:02	Subject Test Refused
Breath Volume	0.058	19:02	
Air Blank	0.000	19:02	
Control Test	0.078	19:03	
Air Blank	0.000	19:03	
Diagnostics Check	OK	19:03	

Data Download Date/Time 06/15/2011 09:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES, POLLY, E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CARTER, D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770940
DRY GAS EXPIRY DATE: 06/02/2012
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 06/01/2011 21:00:41

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.082	20:54
Air Blank	0.000	20:55
Subject Sample #1	0.000	20:56
Breath Volume	1.503	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample #2	0.000	20:59
Breath Volume	1.406	20:59
Air Blank	0.000	20:59
Control Test	0.082	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time 06/15/2011 12:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES, POLLY, E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MALLARD, J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770940
DRY GAS EXPIRY DATE: 06/02/2012
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: DORAN

DATE/TIME OF TEST: 06/02/2011 03:24:56

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.082	03:14	
Air Blank	0.000	03:14	
Subject Sample #1		03:17	No Sample Provided
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample #2		03:23	Subject Test Refused
Breath Volume	0.109	03:23	
Air Blank	0.000	03:24	
Control Test	0.082	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time 06/15/2011 12:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUSTIN, JEREMY, R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BASS, J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770940
DRY GAS EXPIRY DATE: 06/02/2012
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 06/04/2011 02:20:26

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.082	02:14
Air Blank	0.000	02:15
Subject Sample #1	0.175	02:16
Breath Volume	1.886	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample #2	0.174	02:18
Breath Volume	2.480	02:18
Air Blank	0.000	02:19
Control Test	0.082	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time 06/15/2011 12:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUSTIN, JEREMY, R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HEALY, G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770940
DRY GAS EXPIRY DATE: 06/02/2012
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BRIGMAN

DATE/TIME OF TEST: 06/04/2011 03:48:55

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.082	03:42
Air Blank	0.000	03:43
Subject Sample #1	0.110	03:44
Breath Volume	2.046	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample #2	0.113	03:47
Breath Volume	3.148	03:47
Air Blank	0.000	03:47
Control Test	0.082	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time 06/15/2011 12:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK, K, M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HEALY, G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770940
DRY GAS EXPIRY DATE: 06/02/2012
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: KEELING

DATE/TIME OF TEST: 06/05/2011 03:13:09

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.082	03:06
Air Blank	0.000	03:07
Subject Sample #1	0.113	03:08
Breath Volume	3.335	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample #2	0.100	03:11
Breath Volume	1.507	03:11
Air Blank	0.000	03:12
Control Test	0.082	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time 06/15/2011 12:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK, K, M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BASS, J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770940
DRY GAS EXPIRY DATE: 06/02/2012
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 06/06/2011 00:51:20

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.083	00:49	
Air Blank	0.000	00:49	
Subject Sample #1		00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:50	
Control Test	0.082	00:50	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time 06/15/2011 12:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYD, MELISSA, K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BASS, J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770940
DRY GAS EXPIRY DATE: 06/02/2012
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 06/06/2011 04:34:50

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Subject Sample #1	0.218	04:26
Breath Volume	1.503	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample #2	0.180	04:30
Breath Volume	2.070	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample #3	0.207	04:33
Breath Volume	3.304	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time 06/15/2011 12:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES, POLLY, E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HODGES, POLLY
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770940
DRY GAS EXPIRY DATE: 06/02/2012
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 06/10/2011 18:46:17

Diagnostics Check	OK	18:39
Air Blank	0.000	18:40
Control Test	0.083	18:40
Air Blank	0.000	18:41
Subject Sample #1	0.130	18:41
Breath Volume	1.617	18:41
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample #2	0.144	18:44
Breath Volume	1.468	18:44
Air Blank	0.000	18:45
Control Test	0.081	18:45
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time 06/15/2011 12:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT, J, K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HEALY, G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770940
DRY GAS EXPIRY DATE: 06/02/2012
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/12/2011 02:11:42

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.083	02:05
Air Blank	0.000	02:06
Subject Sample #1	0.201	02:06
Breath Volume	1.410	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #2	0.213	02:10
Breath Volume	1.777	02:10
Air Blank	0.000	02:10
Control Test	0.081	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time 06/15/2011 12:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, KEITH, J
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: SMITH, KEITH
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 05/14/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 719876
DRY GAS EXPIRY DATE: 01/22/2012
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: UNSAL

DATE/TIME OF TEST: 06/02/2011 03:27:21

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:17	
Subject Sample #1		03:20	No Sample Provided
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample #2		03:25	No Sample Provided
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time 06/16/2011 16:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO, HERSLYN, A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: ROMERO, HERSLYN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/14/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 719876
DRY GAS EXPIRY DATE: 01/22/2012
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: HASKELL

DATE/TIME OF TEST: 06/05/2011 05:00:42

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Subject Sample #1	0.109	04:56
Breath Volume	1.339	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample #2	0.125	04:59
Breath Volume	1.433	04:59
Air Blank	0.000	04:59
Control Test	0.076	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time 06/16/2011 16:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEXTON, SAMUEL, D
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: STEMEN, AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/14/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 719876
DRY GAS EXPIRY DATE: 01/22/2012
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SCHEEL

DATE/TIME OF TEST: 06/07/2011 04:01:05

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.079	03:50	
Air Blank	0.000	03:50	
Subject Sample #1		03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample #2		03:59	No Sample Provided
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.080	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time 06/16/2011 16:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBERTS, BRADLEY, A

BREATH TEST OPERATOR AGENCY: U of FI PD

ARRESTING OFFICER: ROBERTS, BRADLEY

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 05/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 01:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HATTON

DATE/TIME OF TEST: 06/09/2011 02:04:29

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:59
Subject Sample #1	0.180	01:59
Breath Volume	1.394	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample #2	0.180	02:02
Breath Volume	1.777	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time 06/16/2011 16:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, KEITH, J

BREATH TEST OPERATOR AGENCY: U of FI PD

ARRESTING OFFICER: SMITH, KEITH

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 05/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 05:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 06/12/2011 05:45:14

Diagnostics Check	OK	05:38
Air Blank	0.000	05:39
Control Test	0.078	05:39
Air Blank	0.000	05:40
Subject Sample #1	0.187	05:40
Breath Volume	3.777	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample #2	0.195	05:43
Breath Volume	4.105	05:43
Air Blank	0.000	05:44
Control Test	0.079	05:44
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

Data Download Date/Time 06/16/2011 16:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS, JESSICA, A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: ROBERTS, BRADLEY
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 05/14/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 719876
DRY GAS EXPIRY DATE: 01/22/2012
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 06/15/2011 02:19:03

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Subject Sample #1	0.069	02:14
Breath Volume	2.023	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample #2	0.071	02:17
Breath Volume	2.742	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:19

Data Download Date/Time 06/16/2011 16:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBERTS, BRADLEY, A

BREATH TEST OPERATOR AGENCY: U of FI PD

ARRESTING OFFICER: ROBERTS, BRADLEY

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 06/19/2011 02:19:09

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Subject Sample #1	0.148	02:14
Breath Volume	1.714	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample #2	0.146	02:17
Breath Volume	1.234	02:17
Air Blank	0.000	02:18
Control Test	0.077	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time 07/26/2011 10:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS, BRADLEY, A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: ROBERTS, BRADLEY
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 719876
DRY GAS EXPIRY DATE: 01/22/2012
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CLEMMONS

DATE/TIME OF TEST: 06/23/2011 03:07:57

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:57	
Subject Sample #1		03:00	No Sample Provided
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample #2		03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.081	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time 07/26/2011 10:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEFFENSE, JEAN-GUY,

BREATH TEST OPERATOR AGENCY: U of FI PD

ARRESTING OFFICER: ROBINSON, TAKYDRA

ARRESTING OFFICER AGENCY: U of FI PD

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 719876

DRY GAS EXPIRY DATE: 01/22/2012

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GUASCH

DATE/TIME OF TEST: 06/26/2011 03:23:46

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Subject Sample #1	0.253	03:19
Breath Volume	1.984	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample #2	0.244	03:22
Breath Volume	2.296	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time 07/26/2011 10:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001286

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD, KARI, A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD, KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01011080A2
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: PEEK

DATE/TIME OF TEST: 06/11/2011 18:39:53

Diagnostics Check	OK	18:33
Air Blank	0.000	18:33
Control Test	0.079	18:33
Air Blank	0.000	18:34
Subject Sample #1	0.053	18:35
Breath Volume	4.035	18:35
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample #2	0.054	18:38
Breath Volume	4.496	18:38
Air Blank	0.000	18:38
Control Test	0.078	18:39
Air Blank	0.000	18:39
Diagnostics Check	OK	18:39

Data Download Date/Time 06/22/2011 20:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS, DANNY, L
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MYERS, DANNY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/22/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/28/2011 19:30:51

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:24	
Control Test	0.078	19:24	
Air Blank	0.000	19:24	
Subject Sample #1	0.012	19:25	
Breath Volume	1.300	19:25	
Air Blank	0.000	19:26	
Air Blank	0.000	19:27	
Subject Sample #2		19:29	Subject Test Refused
Breath Volume	0.218	19:29	
Air Blank	0.000	19:29	
Control Test	0.078	19:30	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

Data Download Date/Time 07/26/2011 06:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMICHAEL, REGINALD, L
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CARMICHAEL, REGINALD
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/08/2011

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 770941
DRY GAS EXPIRY DATE: 05/27/2012
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: ELIXSON

DATE/TIME OF TEST: 06/14/2011 18:52:00

Diagnostics Check	OK	18:46
Air Blank	0.000	18:46
Control Test	0.079	18:46
Air Blank	0.000	18:47
Subject Sample #1	0.178	18:47
Breath Volume	3.285	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample #2	0.165	18:50
Breath Volume	3.539	18:50
Air Blank	0.000	18:51
Control Test	0.079	18:51
Air Blank	0.000	18:51
Diagnostics Check	OK	18:52

Data Download Date/Time 07/20/2011 19:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX, CHARLES, G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: PAGE, T
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HOSKINS

DATE/TIME OF TEST: 06/11/2011 01:37:47

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Subject Sample #1	0.000	01:33
Breath Volume	2.660	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample #2	0.000	01:36
Breath Volume	2.800	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time 06/23/2011 10:56

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX, CHARLES, G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: PAGE, T
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 06/11/2011 02:50:08

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:44
Subject Sample #1	0.218	02:45
Breath Volume	4.742	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample #2	0.219	02:48
Breath Volume	4.039	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time 06/23/2011 10:56

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYES, CARON, N
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: SMITH, JERED
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BAUGESS

DATE/TIME OF TEST: 06/04/2011 20:10:01

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.079	20:01
Air Blank	0.000	20:02
Subject Sample #1	0.306	20:02
Breath Volume	2.949	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample #2	0.285	20:05
Breath Volume	1.640	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample #3	0.300	20:08
Breath Volume	3.359	20:08
Air Blank	0.000	20:09
Control Test	0.077	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:10

Data Download Date/Time 06/23/2011 10:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYES, CARON, N
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KONKEL, SCOTT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 06/05/2011 04:49:29

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.080	04:41	
Air Blank	0.000	04:42	
Subject Sample #1		04:45	No Sample Provided
Breath Volume	0.000	04:45	
Air Blank	0.000	04:45	
Air Blank	0.000	04:47	
Subject Sample #2		04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.080	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time 06/23/2011 10:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYES, CARON, N
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: BENNETT, CHRIS
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 06/06/2011 03:26:40

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Subject Sample #1	0.182	03:22
Breath Volume	2.269	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample #2	0.182	03:25
Breath Volume	2.355	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time 06/23/2011 10:04

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX, CHARLES, G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: PAGE, T
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HOSKINS

DATE/TIME OF TEST: 06/11/2011 01:30:58

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.080	01:29	
Air Blank		01:30	Purge Fail
Air Blank		01:30	Purge Fail

Data Download Date/Time 06/23/2011 10:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYES, CARON, N
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: SMITH, JERED
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ROSIER

DATE/TIME OF TEST: 06/18/2011 03:36:13

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample #1	0.156	03:30
Breath Volume	2.074	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample #2	0.158	03:34
Breath Volume	2.195	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time 06/23/2011 10:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYES, CARON, N
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KONKEL, SCOTT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/19/2011 04:08:35

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample #1	0.173	04:04
Breath Volume	1.359	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample #2	0.178	04:07
Breath Volume	2.011	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time 06/23/2011 10:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYES, CARON, N
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: SMITH, JERED
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: REITER

DATE/TIME OF TEST: 06/23/2011 02:59:42

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample #1	0.134	02:55
Breath Volume	1.566	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.142	02:58
Breath Volume	1.515	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time 06/23/2011 10:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY, JASON, D
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KONKEL, SCOTT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 06/23/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: MINCEY

DATE/TIME OF TEST: 06/25/2011 23:50:40

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.081	23:42	
Air Blank	0.000	23:43	
Subject Sample #1		23:46	No Sample Provided
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Air Blank	0.000	23:48	
Subject Sample #2		23:49	Subject Test Refused
Breath Volume	0.000	23:49	
Air Blank	0.000	23:49	
Control Test	0.080	23:50	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time 07/21/2011 15:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY, JASON, D
BREATH TEST OPERATOR AGENCY: Starke PD
ARRESTING OFFICER: KONKEL, SCOTT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 06/23/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: DOWDY

DATE/TIME OF TEST: 06/26/2011 03:37:17

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.079	03:23	
Air Blank	0.000	03:23	
Subject Sample #1	0.125	03:26	Volume Not Met
Breath Volume	0.746	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample #2	0.122	03:30	
Breath Volume	1.687	03:30	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample #3	0.119	03:35	
Breath Volume	1.464	03:35	
Air Blank	0.000	03:36	
Control Test	0.079	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time 07/21/2011 15:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL, GEORGE, L
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: SAPP, TOMMY
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 06/23/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 885290
DRY GAS EXPIRY DATE: 10/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 06/26/2011 04:06:04

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:00
Subject Sample #1	0.172	04:01
Breath Volume	1.324	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample #2	0.188	04:04
Breath Volume	2.605	04:04
Air Blank	0.000	04:05
Control Test	0.077	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:06

Data Download Date/Time 07/21/2011 15:30

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK-OTT, LINDA, N
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: HORNE, ALTON
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 05/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 698118
DRY GAS EXPIRY DATE: 11/30/2011
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DEFFEE

DATE/TIME OF TEST: 06/04/2011 19:39:31

Diagnostics Check	OK	19:38	
Air Blank	0.000	19:38	
Control Test	0.080	19:39	Interferent Detect
Air Blank	0.000	19:39	

Data Download Date/Time 06/28/2011 19:05

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK-OTT, LINDA, N
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: HORNE, ALTON
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 05/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 698118
DRY GAS EXPIRY DATE: 11/30/2011
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DEFEE

DATE/TIME OF TEST: 06/04/2011 19:53:17

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:42	
Control Test	0.081	19:42	
Air Blank	0.000	19:43	
Subject Sample #1		19:46	No Sample Provided
Breath Volume	0.000	19:46	
Air Blank	0.000	19:46	
Air Blank	0.000	19:48	
Subject Sample #2		19:51	No Sample Provided
Breath Volume	0.000	19:51	
Air Blank	0.000	19:52	
Control Test	0.080	19:52	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:53	

Data Download Date/Time 06/28/2011 19:05

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK-OTT, LINDA, N
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: HILBERER, BRIAN
ARRESTING OFFICER AGENCY: Inglis PD
LAST AGENCY INSPECTION: 05/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 698118
DRY GAS EXPIRY DATE: 11/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/05/2011 01:35:32

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:25	
Subject Sample #1		01:28	No Sample Provided
Breath Volume	0.000	01:28	
Air Blank	0.000	01:29	
Air Blank	0.000	01:31	
Subject Sample #2		01:34	No Sample Provided
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time 06/28/2011 19:05

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS, RICHARD, M
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: HARRISON, ALAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 698118
DRY GAS EXPIRY DATE: 11/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BARLEY

DATE/TIME OF TEST: 06/07/2011 01:39:36

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Subject Sample #1	0.073	01:35
Breath Volume	1.320	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample #2	0.073	01:38
Breath Volume	1.410	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time 06/28/2011 19:05

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DOUGLAS, JUSTIN
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 05/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 698118
DRY GAS EXPIRY DATE: 11/30/2011
VIOLATION CODE: ZERO TOL
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/10/2011 02:49:47

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.081	02:49	Interferent Detect
Air Blank	0.000	02:49	

Data Download Date/Time 06/28/2011 19:05

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DOUGLAS, JUSTIN
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 05/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 698118
DRY GAS EXPIRY DATE: 11/30/2011
VIOLATION CODE: ZERO TOL
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/10/2011 03:01:23

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Subject Sample #1	0.179	02:56
Breath Volume	1.726	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.173	02:59
Breath Volume	3.003	02:59
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time 06/28/2011 19:05

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOKOTT, LINDA, N
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: HILBERER, BRIAN
ARRESTING OFFICER AGENCY: Inglis PD
LAST AGENCY INSPECTION: 05/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 698118
DRY GAS EXPIRY DATE: 11/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 06/15/2011 04:07:21

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:06	
Control Test	0.081	04:06	Interferent Detect
Air Blank	0.000	04:07	

Data Download Date/Time 06/28/2011 19:05

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOKOTT, LINDA, N
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: HILBERER, BRIAN
ARRESTING OFFICER AGENCY: Inglis PD
LAST AGENCY INSPECTION: 05/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 698118
DRY GAS EXPIRY DATE: 11/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 06/15/2011 04:17:11

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.081	04:10	
Air Blank	0.000	04:11	
Subject Sample #1	0.000	04:12	
Breath Volume	1.476	04:12	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample #2	0.000	04:15	
Breath Volume	1.125	04:15	
Air Blank	0.000	04:16	
Control Test	0.081	04:16	Interferent Detect
Air Blank	0.000	04:17	

Data Download Date/Time 06/28/2011 19:05

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOOLE, MONICA, N
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TOOLE, MONICA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 06/02/2011 00:34:33

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Subject Sample #1	0.000	00:30
Breath Volume	3.066	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample #2	0.000	00:33
Breath Volume	3.578	00:33
Air Blank	0.000	00:33
Control Test	0.077	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, BRADLEY, D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARRISON, BRADLEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: REGISTER

DATE/TIME OF TEST: 06/02/2011 04:09:30

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Subject Sample #1	0.248	04:04
Breath Volume	3.320	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.239	04:07
Breath Volume	3.386	04:07
Air Blank	0.000	04:08
Control Test	0.077	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN, MORRIS, J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROSS, JENNIFER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/04/2011 01:55:02

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.079	01:44	
Air Blank	0.000	01:44	
Subject Sample #1		01:48	No Sample Provided
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Air Blank	0.000	01:50	
Subject Sample #2		01:53	No Sample Provided
Breath Volume	0.000	01:53	
Air Blank	0.000	01:54	
Control Test	0.078	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:55	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEYES, GABRIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KEYES, GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32210080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 17:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: LOCKLEAR

DATE/TIME OF TEST: 06/04/2011 18:58:20

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.079	18:52
Air Blank	0.000	18:52
Subject Sample #1	0.111	18:53
Breath Volume	2.164	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample #2	0.122	18:56
Breath Volume	1.210	18:56
Air Blank	0.000	18:57
Control Test	0.077	18:57
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE, JOSHUA, J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WHITE, JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 06/05/2011 21:06:53

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:03	
Control Test	0.078	21:04	
Air Blank	0.000	21:04	
Subject Sample #1		21:05	Subject Test Refused
Breath Volume	0.000	21:05	
Air Blank	0.000	21:05	
Control Test	0.078	21:06	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, HENDERSON, S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CAMPBELL, HENDERSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/07/2011 19:18:56

Diagnostics Check	OK	19:12
Air Blank	0.000	19:13
Control Test	0.078	19:13
Air Blank	0.000	19:13
Subject Sample #1	0.000	19:14
Breath Volume	3.320	19:14
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample #2	0.000	19:17
Breath Volume	2.785	19:17
Air Blank	0.000	19:18
Control Test	0.078	19:18
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEYES, GABRIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KEYES, GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32210080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 22:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 06/07/2011 23:28:31

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.077	23:23
Air Blank	0.000	23:23
Subject Sample #1	0.136	23:24
Breath Volume	2.816	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample #2	0.134	23:27
Breath Volume	3.000	23:27
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON, JOSH, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DIXON, JOSH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: FORST

DATE/TIME OF TEST: 06/08/2011 02:12:24

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.078	02:06
Air Blank	0.000	02:07
Subject Sample #1	0.000	02:08
Breath Volume	3.148	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample #2	0.000	02:11
Breath Volume	2.695	02:11
Air Blank	0.000	02:11
Control Test	0.077	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUL, JAYSON, R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PAUL, JAYSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 06/08/2011 17:51:32

Diagnostics Check	OK	17:45
Air Blank	0.000	17:45
Control Test	0.077	17:46
Air Blank	0.000	17:46
Subject Sample #1	0.066	17:47
Breath Volume	3.042	17:47
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample #2	0.060	17:50
Breath Volume	2.046	17:50
Air Blank	0.000	17:50
Control Test	0.078	17:50
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOOLE, MONICA, N
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TOOLE, MONICA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 06/10/2011 00:37:48

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.077	00:35	
Air Blank	0.000	00:36	
Subject Sample #1		00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:36	
Control Test	0.078	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KEYES, GABRIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: KEYES, GABRIEL

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32210080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 18:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 06/10/2011 20:07:49

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.078	20:02
Air Blank	0.000	20:02
Subject Sample #1	0.092	20:03
Breath Volume	1.929	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample #2	0.092	20:06
Breath Volume	2.296	20:06
Air Blank	0.000	20:06
Control Test	0.077	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, BRADLEY, D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COUNTS, RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 06/11/2011 02:10:41

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.078	02:08	
Air Blank	0.000	02:09	
Subject Sample #1		02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.078	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAWAY, KELLY, C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BERRIAN, A
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/11/2011 03:43:17

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.078	03:41	
Air Blank	0.000	03:41	
Subject Sample #1		03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.078	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, CLAYTON, M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SMITH, CLAYTON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 06/11/2011 17:39:15

Diagnostics Check	OK	17:36	
Air Blank	0.000	17:36	
Control Test	0.078	17:37	
Air Blank	0.000	17:37	
Subject Sample #1		17:37	Subject Test Refused
Breath Volume	0.000	17:37	
Air Blank	0.000	17:38	
Control Test	0.078	17:38	
Air Blank	0.000	17:39	
Diagnostics Check	OK	17:39	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAWAY, KELLY, C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: PERKINS, J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 06/12/2011 03:05:44

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.078	02:57
Air Blank	0.000	02:57
Subject Sample #1	0.231	02:59
Breath Volume	2.062	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample #2	0.231	03:04
Breath Volume	1.699	03:04
Air Blank	0.000	03:04
Control Test	0.077	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, BRADLEY, D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARRISON, BRADLEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 06/15/2011 02:45:30

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.078	02:39
Air Blank	0.000	02:39
Subject Sample #1	0.295	02:40
Breath Volume	3.242	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample #2	0.286	02:43
Breath Volume	2.785	02:43
Air Blank	0.000	02:44
Control Test	0.077	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, BRADLEY, D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARRISON, BRADLEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: CITARELLA

DATE/TIME OF TEST: 06/16/2011 04:06:21

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.077	03:58	
Air Blank	0.000	03:59	
Subject Sample #1		04:02	No Sample Provided
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample #2		04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.078	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PAUL, JAYSON, R

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: PAUL, JAYSON

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32210080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 12:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: CRANFORD

DATE/TIME OF TEST: 06/16/2011 13:28:31

Diagnostics Check	OK	13:20	
Air Blank	0.000	13:21	
Control Test	0.078	13:21	
Air Blank	0.000	13:22	
Subject Sample #1	0.000	13:23	
Breath Volume	1.417	13:23	
Air Blank	0.000	13:23	
Air Blank	0.000	13:25	
Subject Sample #2		13:27	Subject Test Refused
Breath Volume	0.128	13:27	
Air Blank	0.000	13:27	
Control Test	0.077	13:27	
Air Blank	0.000	13:28	
Diagnostics Check	OK	13:28	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEICHNER, ADAM, J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TEICHNER, ADAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 06/17/2011 00:22:39

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.078	00:11	
Air Blank	0.000	00:11	
Subject Sample #1	0.205	00:12	
Breath Volume	1.121	00:12	
Air Blank	0.000	00:13	
Air Blank	0.000	00:14	
Subject Sample #2	0.186	00:18	Volume Not Met
Breath Volume	0.535	00:18	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample #3	0.208	00:21	
Breath Volume	2.460	00:21	
Air Blank	0.000	00:21	
Control Test	0.076	00:22	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENNELL, DAVID, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROSS, JAMIESON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CAPPELLA

DATE/TIME OF TEST: 06/17/2011 03:16:22

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.078	03:11	
Air Blank	0.000	03:12	
Subject Sample #1		03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.078	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON, JOSH, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DIXON, JOSH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 06/19/2011 03:13:54

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Subject Sample #1	0.268	03:08
Breath Volume	1.277	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample #2	0.249	03:12
Breath Volume	1.117	03:12
Air Blank	0.000	03:12
Control Test	0.077	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENSWEIG, ERIC,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROSENSWEIG, ERIC
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HAUN

DATE/TIME OF TEST: 06/19/2011 04:27:13

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Subject Sample #1	0.104	04:22
Breath Volume	2.136	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.098	04:25
Breath Volume	2.023	04:25
Air Blank	0.000	04:26
Control Test	0.077	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUL, JAYSON, R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PAUL, JAYSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LOVITT

DATE/TIME OF TEST: 06/19/2011 22:27:23

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.078	22:21
Air Blank	0.000	22:21
Subject Sample #1	0.124	22:22
Breath Volume	1.488	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample #2	0.124	22:25
Breath Volume	1.929	22:25
Air Blank	0.000	22:26
Control Test	0.077	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAWAWI, JAMAL, S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALLEN, BENJAMIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: WESSON

DATE/TIME OF TEST: 06/21/2011 00:39:08

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.078	00:33
Air Blank	0.000	00:34
Subject Sample #1	0.147	00:34
Breath Volume	3.113	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample #2	0.149	00:37
Breath Volume	2.808	00:37
Air Blank	0.000	00:38
Control Test	0.077	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, BRADLEY, D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GILLMAN, KYLE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ELSCOTT

DATE/TIME OF TEST: 06/21/2011 02:19:07

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:15	
Control Test	0.077	02:15	
Air Blank	0.000	02:15	
Subject Sample #1		02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.078	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAWAY, KELLY, C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: PERKINS, JEFFERY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 06/23/2011 03:37:17

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.078	03:35	
Air Blank	0.000	03:35	
Subject Sample #1		03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:36	
Control Test	0.078	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, JAMES, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MILLER, JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: OTHER
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PRESLEY

DATE/TIME OF TEST: 06/24/2011 03:16:53

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:11
Air Blank	0.000	03:11
Subject Sample #1	0.148	03:12
Breath Volume	3.351	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample #2	0.143	03:15
Breath Volume	3.277	03:15
Air Blank	0.000	03:15
Control Test	0.077	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time 06/24/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENHOUSE, WILLIE, P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RITTENHOUSE, WILLIE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32210080A1
DRY GAS EXPIRY DATE: 12/01/2012
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: RADCLIFF

DATE/TIME OF TEST: 06/17/2011 00:34:43

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample #1	0.118	00:29
Breath Volume	1.441	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample #2	0.115	00:33
Breath Volume	2.648	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time 06/24/2011 12:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VENNEL, JOHN, B

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: VENNEL, JOHN

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 20:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 06/04/2011 20:58:50

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:53
Subject Sample #1	0.187	20:54
Breath Volume	3.363	20:54
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample #2	0.185	20:57
Breath Volume	3.218	20:57
Air Blank	0.000	20:57
Control Test	0.078	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time 06/20/2011 05:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VENNEL, JOHN, B

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: VENNEL, JOHN

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 19:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 06/07/2011 22:40:50

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.077	22:35
Air Blank	0.000	22:35
Subject Sample #1	0.253	22:36
Breath Volume	2.804	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample #2	0.242	22:39
Breath Volume	2.882	22:39
Air Blank	0.000	22:39
Control Test	0.078	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time 06/20/2011 05:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLER, MICHAEL, P

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: N/A, N/A

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: ADMIN

ARREST TIME: 00:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: WINGET

DATE/TIME OF TEST: 06/09/2011 11:15:36

Diagnostics Check	OK	11:09
Air Blank	0.000	11:10
Control Test	0.078	11:10
Air Blank	0.000	11:10
Subject Sample #1	0.000	11:11
Breath Volume	5.035	11:11
Air Blank	0.000	11:11
Air Blank	0.000	11:13
Subject Sample #2	0.000	11:14
Breath Volume	4.093	11:14
Air Blank	0.000	11:14
Control Test	0.077	11:15
Air Blank	0.000	11:15
Diagnostics Check	OK	11:15

Data Download Date/Time 06/20/2011 05:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLER, MICHAEL, P

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: N/A, N/A

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: ADMIN

ARREST TIME: 00:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: VONIER

DATE/TIME OF TEST: 06/09/2011 11:24:14

Diagnostics Check	OK	11:18
Air Blank	0.000	11:18
Control Test	0.078	11:18
Air Blank	0.000	11:19
Subject Sample #1	0.000	11:19
Breath Volume	4.324	11:19
Air Blank	0.000	11:20
Air Blank	0.000	11:22
Subject Sample #2	0.000	11:22
Breath Volume	4.371	11:22
Air Blank	0.000	11:23
Control Test	0.078	11:23
Air Blank	0.000	11:24
Diagnostics Check	OK	11:24

Data Download Date/Time 06/20/2011 05:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: OWENS, GEORGE, D

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: TARTER, JEFFREY

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 16:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/10/2011 17:38:39

Diagnostics Check	OK	17:32
Air Blank	0.000	17:32
Control Test	0.078	17:33
Air Blank	0.000	17:33
Subject Sample #1	0.203	17:34
Breath Volume	2.921	17:34
Air Blank	0.000	17:34
Air Blank	0.000	17:36
Subject Sample #2	0.210	17:37
Breath Volume	2.785	17:37
Air Blank	0.000	17:37
Control Test	0.078	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time 06/20/2011 05:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FRENCH, JASON, K

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: FRENCH, JASON

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 16:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/12/2011 17:28:48

Diagnostics Check	OK	17:22
Air Blank	0.000	17:22
Control Test	0.078	17:23
Air Blank	0.000	17:23
Subject Sample #1	0.159	17:24
Breath Volume	2.171	17:24
Air Blank	0.000	17:25
Air Blank	0.000	17:26
Subject Sample #2	0.171	17:27
Breath Volume	2.414	17:27
Air Blank	0.000	17:27
Control Test	0.077	17:28
Air Blank	0.000	17:28
Diagnostics Check	OK	17:28

Data Download Date/Time 06/20/2011 05:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RHODES, PIERRE, J

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: DANIELS, GARY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 20:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: PEBBLES JR

DATE/TIME OF TEST: 06/20/2011 21:45:28

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:40
Subject Sample #1	0.148	21:40
Breath Volume	3.425	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample #2	0.133	21:43
Breath Volume	1.144	21:43
Air Blank	0.000	21:44
Control Test	0.078	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time 07/13/2011 21:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCNEAL, RUSSELL, E

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: MCNEAL, RUSSELL

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 06/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 10:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:28

SUBJECT'S LAST NAME: FAIRCLOTH

DATE/TIME OF TEST: 06/27/2011 11:09:16

Diagnostics Check	OK	11:03
Air Blank	0.000	11:03
Control Test	0.076	11:03
Air Blank	0.000	11:04
Subject Sample #1	0.278	11:04
Breath Volume	4.078	11:04
Air Blank	0.000	11:05
Air Blank	0.000	11:07
Subject Sample #2	0.268	11:07
Breath Volume	3.636	11:07
Air Blank	0.000	11:08
Control Test	0.077	11:08
Air Blank	0.000	11:09
Diagnostics Check	OK	11:09

Data Download Date/Time 07/13/2011 21:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCNEAL, RUSSELL, E

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: SCARBOROUGH, TYLER

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 06/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG025702

DRY GAS EXPIRY DATE: 09/14/2012

VIOLATION CODE: DUI

ARREST TIME: 21:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/28/2011 22:09:22

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.078	22:03
Air Blank	0.000	22:03
Subject Sample #1	0.116	22:04
Breath Volume	4.820	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample #2	0.109	22:07
Breath Volume	3.941	22:07
Air Blank	0.000	22:08
Control Test	0.078	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time 07/13/2011 21:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JIMMY, L
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: HUFF, CLAY
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 05/19/2011

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 643837
DRY GAS EXPIRY DATE: 07/17/2011
VIOLATION CODE: BUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 06/07/2011 22:52:44

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.075	22:47
Air Blank	0.000	22:47
Subject Sample #1	0.126	22:48
Breath Volume	1.566	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample #2	0.125	22:51
Breath Volume	2.253	22:51
Air Blank	0.000	22:51
Control Test	0.076	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time 06/16/2011 09:24

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENTLEY, TIMOTHY, A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HERNANDEZ, MICHAEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670753
DRY GAS EXPIRY DATE: 09/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: RAWUS

DATE/TIME OF TEST: 06/11/2011 03:55:50

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample #1	0.141	03:51
Breath Volume	3.089	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample #2	0.135	03:54
Breath Volume	1.839	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time 06/30/2011 14:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER, JOHN, A
BREATH TEST OPERATOR AGENCY: Springfield PD
ARRESTING OFFICER: FOSTER, ALEX
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670753
DRY GAS EXPIRY DATE: 09/30/2011
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 06/20/2011 20:09:09

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.081	20:03
Air Blank	0.000	20:03
Subject Sample #1	0.047	20:04
Breath Volume	1.875	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample #2	0.044	20:07
Breath Volume	1.984	20:07
Air Blank	0.000	20:08
Control Test	0.081	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time 06/30/2011 14:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPKINS, DAVID, J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PERRY, TRAVIS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670753
DRY GAS EXPIRY DATE: 09/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HIGDON

DATE/TIME OF TEST: 06/22/2011 22:56:28

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:43	
Control Test	0.080	22:44	
Air Blank	0.000	22:44	
Subject Sample #1	0.092	22:47	Volume Not Met
Breath Volume	0.015	22:47	
Air Blank	0.000	22:48	
Air Blank	0.000	22:50	
Subject Sample #2	0.204	22:51	
Breath Volume	1.527	22:51	
Air Blank	0.000	22:52	
Air Blank	0.000	22:54	
Subject Sample #3	0.219	22:54	
Breath Volume	2.738	22:54	
Air Blank	0.000	22:55	
Control Test	0.080	22:55	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time 06/30/2011 14:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDRINE, CHAD, M
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: VIDRINE, CHAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670753
DRY GAS EXPIRY DATE: 09/30/2011
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: ELMORE

DATE/TIME OF TEST: 06/24/2011 06:32:33

Diagnostics Check	OK	06:26
Air Blank	0.000	06:26
Control Test	0.081	06:26
Air Blank	0.000	06:27
Subject Sample #1	0.222	06:27
Breath Volume	4.140	06:27
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample #2	0.223	06:31
Breath Volume	4.574	06:31
Air Blank	0.000	06:31
Control Test	0.081	06:31
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

Data Download Date/Time 06/30/2011 14:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SEWELL, RICKY,

BREATH TEST OPERATOR AGENCY: Parker PD

ARRESTING OFFICER: SHEPPARD, M

ARRESTING OFFICER AGENCY: Parker PD

LAST AGENCY INSPECTION: 06/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670753

DRY GAS EXPIRY DATE: 09/30/2011

VIOLATION CODE: DUI

ARREST TIME: 12:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: BURDESHAW

DATE/TIME OF TEST: 06/28/2011 13:53:25

Diagnostics Check	OK	13:50	
Air Blank	0.000	13:51	
Control Test	0.081	13:51	
Air Blank	0.000	13:51	
Subject Sample #1		13:52	Subject Test Refused
Breath Volume	0.000	13:52	
Air Blank	0.000	13:52	
Control Test	0.081	13:52	
Air Blank	0.000	13:53	
Diagnostics Check	OK	13:53	

Data Download Date/Time 06/30/2011 14:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WARD, JOSEPH, G

BREATH TEST OPERATOR AGENCY: Niceville PD

ARRESTING OFFICER: WARD, JOSEPH

ARRESTING OFFICER AGENCY: Niceville PD

LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 21:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: VANDERSCHAAF

DATE/TIME OF TEST: 06/16/2011 23:12:12

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Subject Sample #1	0.000	23:07
Breath Volume	1.921	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample #2	0.000	23:10
Breath Volume	1.789	23:10
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time 06/19/2011 11:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WARD, JOSEPH, G

BREATH TEST OPERATOR AGENCY: Niceville PD

ARRESTING OFFICER: GILL, WAYLON

ARRESTING OFFICER AGENCY: Niceville PD

LAST AGENCY INSPECTION: 06/19/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646831

DRY GAS EXPIRY DATE: 08/21/2011

VIOLATION CODE: DUI

ARREST TIME: 21:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/30/2011 22:16:58

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.079	22:11
Air Blank	0.000	22:11
Subject Sample #1	0.000	22:12
Breath Volume	1.101	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample #2	0.000	22:15
Breath Volume	3.289	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time 07/16/2011 15:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUXHOLD, RICHARD, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAY, JODY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JARRETT

DATE/TIME OF TEST: 06/04/2011 04:17:02

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:11
Subject Sample #1	0.137	04:12
Breath Volume	3.449	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample #2	0.137	04:15
Breath Volume	3.156	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:17

Data Download Date/Time 06/20/2011 11:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, DAVID, C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PAUL, JOSEPH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 06/05/2011 02:03:12

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.081	02:00	
Air Blank	0.000	02:01	
Subject Sample #1		02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	
Control Test	0.082	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time 06/20/2011 11:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIMANOVICH, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ALEXANDER, C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: WOFFORD

DATE/TIME OF TEST: 06/11/2011 01:00:41

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.081	00:58	
Air Blank	0.000	00:59	
Subject Sample #1		00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Control Test	0.082	01:00	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time 06/20/2011 11:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDAGE, WENDY, M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MURSCH, TED
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 06/13/2011 20:41:58

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:40	
Control Test	0.081	20:41	
Air Blank		20:41	RFI Detect
Air Blank	0.000	20:41	

Data Download Date/Time 06/20/2011 11:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDAGE, WENDY, M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MURSCH, TED
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 06/13/2011 20:52:22

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.082	20:45
Air Blank	0.000	20:46
Subject Sample #1	0.175	20:47
Breath Volume	1.460	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample #2	0.178	20:50
Breath Volume	1.789	20:50
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time 06/20/2011 11:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUXHOLD, RICHARD, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: HUXHOLD, RICHARD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/14/2011 22:42:05

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.081	22:36
Air Blank	0.000	22:37
Subject Sample #1	0.263	22:37
Breath Volume	2.914	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:39
Subject Sample #2	0.258	22:40
Breath Volume	2.726	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time 06/20/2011 11:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIEB, JOHN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LIEB, JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 06/16/2011 00:59:01

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:52
Subject Sample #1	0.157	00:53
Breath Volume	1.230	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample #2	0.160	00:57
Breath Volume	1.996	00:57
Air Blank	0.000	00:58
Control Test	0.082	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:59

Data Download Date/Time 06/20/2011 11:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLINETS, THOMAS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GATHER, PHILLIP
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 06/16/2011 02:04:45

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.081	01:58
Air Blank	0.000	01:58
Subject Sample #1	0.202	02:00
Breath Volume	1.421	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample #2	0.214	02:03
Breath Volume	2.390	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time 06/20/2011 11:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUXHOLD, RICHARD, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: YATES, HAMPTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: CUMBIE

DATE/TIME OF TEST: 06/19/2011 00:58:19

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:53
Subject Sample #1	0.000	00:54
Breath Volume	1.710	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample #2	0.000	00:56
Breath Volume	1.789	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time 06/20/2011 11:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER, DERRAL, W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: LEWIS, PAUL
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 06/20/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 06/24/2011 12:27:26

Diagnostics Check	OK	12:18	
Air Blank	0.000	12:19	
Control Test	0.082	12:19	
Air Blank	0.000	12:20	
Subject Sample #1		12:23	No Sample Provided
Breath Volume	0.000	12:23	
Air Blank	0.000	12:23	
Air Blank	0.000	12:25	
Subject Sample #2		12:26	Subject Test Refused
Breath Volume	0.000	12:26	
Air Blank	0.000	12:26	
Control Test	0.082	12:26	
Air Blank	0.000	12:27	
Diagnostics Check	OK	12:27	

Data Download Date/Time 07/27/2011 08:25

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALEXANDER, C, L

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: ALEXANDER, C

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 06/20/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 761073

DRY GAS EXPIRY DATE: 05/06/2012

VIOLATION CODE: DUI

ARREST TIME: 02:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GUNN

DATE/TIME OF TEST: 06/25/2011 04:37:47

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:32
Air Blank	0.000	04:32
Subject Sample #1	0.139	04:33
Breath Volume	1.738	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample #2	0.148	04:36
Breath Volume	1.554	04:36
Air Blank	0.000	04:36
Control Test	0.081	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time 07/27/2011 08:25

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIMANOVICH, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: STEPHENS, JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/20/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GENGERKE

DATE/TIME OF TEST: 06/26/2011 03:39:07

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.081	03:37	
Air Blank	0.000	03:37	
Subject Sample #1		03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.080	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time 07/27/2011 08:25

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHATMOUGH, CHRIS, J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WHATMOUGH, CHRIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/20/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ROPER

DATE/TIME OF TEST: 06/28/2011 04:19:40

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.082	04:10
Air Blank	0.000	04:11
Subject Sample #1	0.143	04:11
Breath Volume	3.875	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample #2	0.122	04:14
Breath Volume	2.582	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #3	0.144	04:18
Breath Volume	3.875	04:18
Air Blank	0.000	04:18
Control Test	0.082	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time 07/27/2011 08:25

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARK, STEVEN, T
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: STEPHENS, JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/03/2011 02:44:14

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample #1	0.186	02:39
Breath Volume	1.148	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample #2	0.177	02:42
Breath Volume	2.171	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time 06/20/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK, LEONARD, C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: STEPHENS, JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 06/08/2011 06:08:56

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:01	
Control Test	0.079	06:01	
Air Blank	0.000	06:02	
Subject Sample #1	0.138	06:03	
Breath Volume	1.664	06:03	
Air Blank	0.000	06:04	
Air Blank	0.000	06:06	
Subject Sample #2		06:08	Sequence Aborted
Breath Volume	0.000	06:08	
Air Blank	0.000	06:08	

Data Download Date/Time 06/20/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK, LEONARD, C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: STEPHENS, JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 06/08/2011 06:33:28

Diagnostics Check	OK	06:25
Air Blank	0.000	06:25
Control Test	0.078	06:26
Air Blank	0.000	06:26
Subject Sample #1	0.131	06:28
Breath Volume	2.121	06:28
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample #2	0.127	06:31
Breath Volume	1.914	06:31
Air Blank	0.000	06:32
Control Test	0.076	06:32
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

Data Download Date/Time 06/20/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING, MARJORIE, G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KING, MARJORIE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 761073
DRY GAS EXPIRY DATE: 05/06/2012
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: FAOUTAS

DATE/TIME OF TEST: 06/12/2011 00:36:42

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Subject Sample #1	0.105	00:32
Breath Volume	3.925	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample #2	0.105	00:35
Breath Volume	3.808	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time 06/20/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALEXANDER, C, L

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: ALEXANDER, C

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 05/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 761073

DRY GAS EXPIRY DATE: 05/06/2012

VIOLATION CODE: DUI

ARREST TIME: 02:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/12/2011 03:43:51

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Subject Sample #1		03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.078	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time 06/20/2011 11:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULANDA, FREDERICK, J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: HURST, MATTHEW
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 814704
DRY GAS EXPIRY DATE: 08/18/2012
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MCCOIN

DATE/TIME OF TEST: 06/03/2011 00:40:34

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.079	00:35
Air Blank	0.000	00:35
Subject Sample #1	0.000	00:36
Breath Volume	1.789	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample #2	0.000	00:39
Breath Volume	2.062	00:39
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time 06/22/2011 20:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULANDA, FREDERICK, J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ASHBY, JOHN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 814704
DRY GAS EXPIRY DATE: 08/18/2012
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SALLETTE

DATE/TIME OF TEST: 06/16/2011 21:11:25

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:01	
Control Test	0.080	21:02	
Air Blank	0.000	21:02	
Subject Sample #1	0.138	21:05	Volume Not Met
Breath Volume	0.253	21:05	
Air Blank	0.000	21:06	
Air Blank	0.000	21:08	
Subject Sample #2		21:09	Subject Test Refused
Breath Volume	0.519	21:09	
Air Blank	0.000	21:10	
Control Test	0.076	21:10	
Air Blank	0.000	21:11	
Diagnostics Check	OK	21:11	

Data Download Date/Time 06/22/2011 20:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRELL, PAUL, E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MCKINLEY, STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/22/2011

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 814704
DRY GAS EXPIRY DATE: 08/18/2012
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/19/2011 01:49:23

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:43
Air Blank	0.000	01:43
Subject Sample #1	0.249	01:44
Breath Volume	1.214	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample #2	0.247	01:47
Breath Volume	1.261	01:47
Air Blank	0.000	01:48
Control Test	0.077	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time 06/22/2011 20:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLROYD, RYAN, D
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: HOLROYD, RYAN
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 759279
DRY GAS EXPIRY DATE: 04/20/2012
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/05/2011 01:25:23

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:20
Subject Sample #1	0.174	01:20
Breath Volume	1.687	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample #2	0.173	01:23
Breath Volume	2.167	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time 06/30/2011 11:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROUTMAN, CHAD, A
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: TROUTMAN, CHAD
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 759279
DRY GAS EXPIRY DATE: 04/20/2012
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 06/12/2011 03:19:25

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Subject Sample #1		03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.080	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time 06/30/2011 11:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLROYD, RYAN, D
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: HOLROYD, RYAN
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 759279
DRY GAS EXPIRY DATE: 04/20/2012
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 06/14/2011 01:40:43

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.078	01:34
Air Blank	0.000	01:34
Subject Sample #1	0.206	01:35
Breath Volume	2.218	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample #2	0.192	01:39
Breath Volume	3.343	01:39
Air Blank	0.000	01:39
Control Test	0.077	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time 06/30/2011 11:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLROYD, RYAN, D
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: HOLROYD, RYAN
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 759279
DRY GAS EXPIRY DATE: 04/20/2012
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MCMASTERS

DATE/TIME OF TEST: 06/19/2011 02:47:35

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Subject Sample #1	0.094	02:40	Volume Not Met
Breath Volume	0.296	02:40	
Air Blank	0.000	02:41	
Air Blank	0.000	02:42	
Subject Sample #2	0.100	02:45	Slope Not Level
Breath Volume	1.148	02:45	
Air Blank	0.000	02:46	
Control Test	0.077	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time 06/30/2011 11:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLROYD, RYAN, D
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: HOLROYD, RYAN
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 759279
DRY GAS EXPIRY DATE: 04/20/2012
VIOLATION CODE: SYS CHECK
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/19/2011 17:07:42

Diagnostics Check	OK	17:00
Air Blank	0.000	17:01
Control Test	0.080	17:01
Air Blank	0.000	17:02
Subject Sample #1	0.000	17:02
Breath Volume	1.265	17:02
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample #2	0.000	17:06
Breath Volume	1.253	17:06
Air Blank	0.000	17:06
Control Test	0.079	17:07
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time 06/30/2011 11:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOON, JACK, D
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: BOON, JACK
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 759279
DRY GAS EXPIRY DATE: 04/20/2012
VIOLATION CODE: COURT ORD
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 06/23/2011 15:12:30

Diagnostics Check	OK	15:00	
Air Blank	0.000	15:01	
Control Test	0.079	15:01	
Air Blank	0.000	15:01	
Subject Sample #1	0.204	15:05	Volume Not Met
Breath Volume	0.660	15:05	
Air Blank	0.000	15:05	
Air Blank	0.000	15:07	
Subject Sample #2	0.192	15:10	Volume Not Met
Breath Volume	0.488	15:10	
Air Blank	0.000	15:11	
Control Test	0.076	15:11	
Air Blank	0.000	15:12	
Diagnostics Check	OK	15:12	

Data Download Date/Time 06/30/2011 11:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOON, JACK, D
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: BOON, JACK
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 759279
DRY GAS EXPIRY DATE: 04/20/2012
VIOLATION CODE: COURT ORD
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 06/23/2011 15:30:06

Diagnostics Check	OK	15:15	
Air Blank	0.000	15:16	
Control Test	0.077	15:16	
Air Blank	0.000	15:17	
Subject Sample #1	0.000	15:19	
Breath Volume	3.085	15:19	
Air Blank	0.000	15:20	
Air Blank	0.000	15:22	
Subject Sample #2	0.193	15:25	Volume Not Met
Breath Volume	0.917	15:25	
Air Blank	0.000	15:26	
Air Blank	0.000	15:27	
Subject Sample #3	0.000	15:28	
Breath Volume	1.968	15:28	
Air Blank	0.000	15:29	
Control Test	0.078	15:29	
Air Blank	0.000	15:30	
Diagnostics Check	OK	15:30	

Data Download Date/Time 06/30/2011 11:10

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GHOBRIAL, MARTIN, A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LEVASSEUR, ERIK
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35710080A3
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 13:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: MOYLES

DATE/TIME OF TEST: 06/09/2011 14:47:06

Diagnostics Check	OK	14:44	
Air Blank	0.000	14:44	
Control Test	0.079	14:45	
Air Blank	0.000	14:45	
Subject Sample #1	0.168	14:46	
Breath Volume	4.011	14:46	
Air Blank		14:46	RFI Detect
Air Blank		14:46	RFI Detect
Air Blank	0.000	14:47	

Data Download Date/Time 06/29/2011 18:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GHOBRIAL, MARTIN, A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LEVASSEUR, ERIK
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35710080A3
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 13:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: MOYLES

DATE/TIME OF TEST: 06/09/2011 14:55:59

Diagnostics Check	OK	14:50	
Air Blank	0.000	14:50	
Control Test	0.078	14:50	
Air Blank	0.000	14:51	
Subject Sample #1	0.164	14:52	
Breath Volume	4.562	14:52	
Air Blank	0.000	14:52	
Air Blank	0.000	14:54	
Subject Sample #2	0.160	14:55	
Breath Volume	2.996	14:55	
Air Blank		14:55	RFI Detect
Air Blank	0.000	14:55	

Data Download Date/Time 06/29/2011 18:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES, FRANKIE, J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ZAMOYSKI, A
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35710080A3
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: PUJOLS

DATE/TIME OF TEST: 06/17/2011 03:31:34

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Subject Sample #1	0.175	03:26
Breath Volume	3.019	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample #2	0.174	03:30
Breath Volume	2.296	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time 06/29/2011 18:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS, IAN,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HARRIS, IAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35710080A3
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 06/24/2011 00:43:48

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:37
Air Blank	0.000	00:37
Subject Sample #1	0.163	00:38
Breath Volume	2.062	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample #2	0.154	00:42
Breath Volume	2.332	00:42
Air Blank	0.000	00:42
Control Test	0.078	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time 06/29/2011 18:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS, IAN,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HARRIS, IAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35710080A3
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: STANFORD

DATE/TIME OF TEST: 06/24/2011 03:23:13

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:15
Subject Sample #1	0.245	03:16
Breath Volume	1.554	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample #2	0.239	03:21
Breath Volume	2.425	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time 06/29/2011 18:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES, FRANKIE, J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: BORGES, FRANKIE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35710080A3
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 06/24/2011 03:58:18

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Subject Sample #1	0.192	03:53
Breath Volume	2.113	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample #2	0.183	03:56
Breath Volume	1.972	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time 06/29/2011 18:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES, FRANKIE, J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: BORGES, FRANKIE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/26/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35710080A3
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: WENZLER

DATE/TIME OF TEST: 06/25/2011 03:07:56

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Subject Sample #1	0.128	03:01	
Breath Volume	1.632	03:01	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample #2	0.096	03:06	Volume Not Met
Breath Volume	0.488	03:06	
Air Blank	0.000	03:07	
Air Blank	0.000	03:07	

Data Download Date/Time 06/29/2011 18:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON, ANDI, L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ANDERSON, CHRISTOPHER
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/06/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: PEARCE

DATE/TIME OF TEST: 06/01/2011 21:56:47

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:48	
Control Test	0.080	21:49	
Air Blank	0.000	21:49	
Subject Sample #1	0.301	21:52	Volume Not Met
Breath Volume	0.546	21:52	
Air Blank	0.000	21:53	
Air Blank	0.000	21:55	
Subject Sample #2		21:55	Subject Test Refused
Breath Volume	0.000	21:55	
Air Blank	0.000	21:55	
Control Test	0.077	21:56	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time 06/02/2011 11:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND, PAMELA, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HAWKINS, ERIK
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: CYPRESS

DATE/TIME OF TEST: 06/04/2011 05:02:49

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:56
Subject Sample #1	0.158	04:57
Breath Volume	1.179	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample #2	0.161	05:01
Breath Volume	1.828	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUMAN, GARRETT, L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: SHUMAN, GARRETT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: VELEA

DATE/TIME OF TEST: 06/05/2011 02:53:20

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample #1	0.136	02:45
Breath Volume	2.035	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample #2	0.164	02:48
Breath Volume	4.269	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample #3	0.150	02:51
Breath Volume	1.832	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIBERATORE, SCOTT, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HAWKINS, ERIK
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SCHWEITZER

DATE/TIME OF TEST: 06/10/2011 04:30:56

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:28	
Control Test	0.080	04:28	
Air Blank	0.000	04:29	
Subject Sample #1		04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:30	
Control Test	0.080	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE, MELINDA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SAPUTO, JEFF
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: BIEZUNSKI

DATE/TIME OF TEST: 06/11/2011 17:36:00

Diagnostics Check	OK	17:33	
Air Blank	0.000	17:33	
Control Test	0.080	17:33	
Air Blank	0.000	17:34	
Subject Sample #1		17:34	Subject Test Refused
Breath Volume	0.000	17:34	
Air Blank	0.000	17:35	
Control Test	0.080	17:35	
Air Blank	0.000	17:35	
Diagnostics Check	OK	17:36	

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND, PAMELA, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WRIGHT, DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CESPEDES

DATE/TIME OF TEST: 06/12/2011 02:02:25

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Subject Sample #1	0.122	01:57
Breath Volume	1.644	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample #2	0.130	02:00
Breath Volume	1.867	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIRINO, ARTHUR, D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CIRINO, ARTHUR
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: OTHER
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:22

SUBJECT'S LAST NAME: CHRISTIANO

DATE/TIME OF TEST: 06/12/2011 10:53:20

Diagnostics Check	OK	10:47
Air Blank	0.000	10:47
Control Test	0.080	10:47
Air Blank	0.000	10:48
Subject Sample #1	0.097	10:48
Breath Volume	1.558	10:48
Air Blank	0.000	10:49
Air Blank	0.000	10:51
Subject Sample #2	0.098	10:51
Breath Volume	1.285	10:51
Air Blank	0.000	10:52
Control Test	0.080	10:52
Air Blank	0.000	10:53
Diagnostics Check	OK	10:53

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIRINO, ARTHUR, D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEEK, STEVE
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: LAURIE

DATE/TIME OF TEST: 06/12/2011 11:45:30

Diagnostics Check	OK	11:39
Air Blank	0.000	11:39
Control Test	0.080	11:40
Air Blank	0.000	11:40
Subject Sample #1	0.017	11:41
Breath Volume	2.281	11:41
Air Blank	0.000	11:41
Air Blank	0.000	11:43
Subject Sample #2	0.016	11:44
Breath Volume	2.320	11:44
Air Blank	0.000	11:44
Control Test	0.080	11:44
Air Blank	0.000	11:45
Diagnostics Check	OK	11:45

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIRINO, ARTHUR, D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CIRINO, ARTHUR
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: OTHER
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: CHRISTIANO

DATE/TIME OF TEST: 06/12/2011 12:33:08

Diagnostics Check	OK	12:27
Air Blank	0.000	12:27
Control Test	0.080	12:27
Air Blank	0.000	12:28
Subject Sample #1	0.070	12:28
Breath Volume	2.140	12:28
Air Blank	0.000	12:29
Air Blank	0.000	12:31
Subject Sample #2	0.071	12:31
Breath Volume	1.609	12:31
Air Blank	0.000	12:32
Control Test	0.080	12:32
Air Blank	0.000	12:33
Diagnostics Check	OK	12:33

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIRINO, ARTHUR, D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CIRINO, ARTHUR
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: OTHER
ARREST TIME: 13:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:24

SUBJECT'S LAST NAME: CHRISTIANO

DATE/TIME OF TEST: 06/12/2011 13:54:04

Diagnostics Check	OK	13:48
Air Blank	0.000	13:48
Control Test	0.080	13:48
Air Blank	0.000	13:49
Subject Sample #1	0.044	13:49
Breath Volume	1.320	13:49
Air Blank	0.000	13:50
Air Blank	0.000	13:52
Subject Sample #2	0.042	13:52
Breath Volume	1.164	13:52
Air Blank	0.000	13:53
Control Test	0.080	13:53
Air Blank	0.000	13:53
Diagnostics Check	OK	13:54

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIRINO, ARTHUR, D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CIRINO, ARTHUR
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: OTHER
ARREST TIME: 13:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:37

SUBJECT'S LAST NAME: BIEZUNSKI

DATE/TIME OF TEST: 06/12/2011 14:07:03

Diagnostics Check	OK	14:00
Air Blank	0.000	14:00
Control Test	0.081	14:00
Air Blank	0.000	14:01
Subject Sample #1	0.000	14:02
Breath Volume	1.789	14:02
Air Blank	0.000	14:02
Air Blank	0.000	14:04
Subject Sample #2	0.000	14:05
Breath Volume	1.273	14:05
Air Blank	0.000	14:06
Control Test	0.080	14:06
Air Blank	0.000	14:06
Diagnostics Check	OK	14:07

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMEL, ANDREW, J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SCRIBNER, CHRISTOPHER
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: AKEY

DATE/TIME OF TEST: 06/28/2011 22:22:44

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:19	
Control Test	0.079	22:19	
Air Blank	0.000	22:20	
Subject Sample #1		22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.079	22:22	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMONS, NORA, J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SMITH, K
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:03

SUBJECT'S LAST NAME: KRESCH

DATE/TIME OF TEST: 06/29/2011 14:32:37

Diagnostics Check	OK	14:26
Air Blank	0.000	14:26
Control Test	0.079	14:26
Air Blank	0.000	14:27
Subject Sample #1	0.000	14:28
Breath Volume	1.871	14:28
Air Blank	0.000	14:28
Air Blank	0.000	14:30
Subject Sample #2	0.000	14:31
Breath Volume	2.128	14:31
Air Blank	0.000	14:31
Control Test	0.078	14:32
Air Blank	0.000	14:32
Diagnostics Check	OK	14:32

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS, TONY, D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARR, STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 06/30/2011 05:21:50

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:16	
Control Test	0.080	05:17	
Air Blank	0.000	05:17	
Subject Sample #1		05:20	Subject Test Refused
Breath Volume	0.164	05:20	
Air Blank	0.000	05:20	
Control Test	0.080	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS, TONY, D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARR, STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 850329
DRY GAS EXPIRY DATE: 09/22/2012
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 06/30/2011 05:31:00

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.080	05:26	
Air Blank	0.000	05:26	
Subject Sample #1		05:29	Subject Test Refused
Breath Volume	0.167	05:29	
Air Blank	0.000	05:30	
Control Test	0.081	05:30	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:31	

Data Download Date/Time 07/29/2011 10:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, CLETE, J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LACELLE, RACHAEL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 06/03/2011 22:34:25

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Subject Sample #1	0.298	22:29
Breath Volume	3.441	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample #2	0.299	22:32
Breath Volume	3.656	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time 07/29/2011 10:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVENDER, RANDALL, W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LAVENDER, RANDALL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: OTHER
ARREST TIME: 11:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: CESPEDES

DATE/TIME OF TEST: 06/12/2011 12:17:52

Diagnostics Check	OK	12:11
Air Blank	0.000	12:11
Control Test	0.080	12:12
Air Blank	0.000	12:12
Subject Sample #1	0.000	12:13
Breath Volume	2.480	12:13
Air Blank	0.000	12:13
Air Blank	0.000	12:15
Subject Sample #2	0.000	12:16
Breath Volume	2.507	12:16
Air Blank	0.000	12:16
Control Test	0.078	12:17
Air Blank	0.000	12:17
Diagnostics Check	OK	12:17

Data Download Date/Time 07/29/2011 10:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON, ANDI, L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RODRIGUEZ, DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MAHLSTEDT

DATE/TIME OF TEST: 06/15/2011 17:21:00

Diagnostics Check	OK	17:06	
Air Blank	0.000	17:06	
Control Test	0.079	17:07	
Air Blank	0.000	17:07	
Subject Sample #1	0.109	17:10	Volume Not Met
Breath Volume	0.621	17:10	
Air Blank	0.000	17:11	
Air Blank	0.000	17:13	
Subject Sample #2	0.143	17:13	
Breath Volume	2.015	17:13	
Air Blank	0.000	17:14	
Air Blank	0.000	17:16	
Subject Sample #3	0.143	17:19	
Breath Volume	1.867	17:19	
Air Blank	0.000	17:20	
Control Test	0.079	17:20	
Air Blank	0.000	17:20	
Diagnostics Check	OK	17:21	

Data Download Date/Time 07/29/2011 10:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIRINO, ARTHUR, D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MEWBORN, PAUL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 06/25/2011 14:04:14

Diagnostics Check	OK	14:01	
Air Blank	0.000	14:01	
Control Test	0.080	14:02	
Air Blank	0.000	14:02	
Subject Sample #1		14:02	Subject Test Refused
Breath Volume	0.000	14:02	
Air Blank	0.000	14:03	
Control Test	0.080	14:03	
Air Blank	0.000	14:04	
Diagnostics Check	OK	14:04	

Data Download Date/Time 07/29/2011 10:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND, PAMELA, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YANCHULA, CHRISTIAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/26/2011 01:21:59

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Subject Sample #1	0.228	01:14
Breath Volume	1.242	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:16
Subject Sample #2	0.206	01:17
Breath Volume	1.101	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample #3	0.230	01:20
Breath Volume	2.175	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time 07/29/2011 10:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIBERATORE, SCOTT, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROBINSON, TAYO
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MIGUELES

DATE/TIME OF TEST: 06/26/2011 01:50:55

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:45
Subject Sample #1	0.119	01:46
Breath Volume	2.132	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:48
Subject Sample #2	0.115	01:49
Breath Volume	2.808	01:49
Air Blank	0.000	01:50
Control Test	0.078	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time 07/29/2011 10:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVENDER, RANDALL, W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LAVENDER, RANDALL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: OTHER
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: MIGUELES

DATE/TIME OF TEST: 06/26/2011 12:51:49

Diagnostics Check	OK	12:45
Air Blank	0.000	12:46
Control Test	0.079	12:46
Air Blank	0.000	12:47
Subject Sample #1	0.000	12:47
Breath Volume	3.164	12:47
Air Blank	0.000	12:48
Air Blank	0.000	12:50
Subject Sample #2	0.000	12:50
Breath Volume	3.265	12:50
Air Blank	0.000	12:50
Control Test	0.077	12:51
Air Blank	0.000	12:51
Diagnostics Check	OK	12:51

Data Download Date/Time 07/29/2011 10:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVENDER, RANDALL, W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LAVENDER, RANDALL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: OTHER
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/26/2011 13:00:17

Diagnostics Check	OK	12:53
Air Blank	0.000	12:54
Control Test	0.078	12:54
Air Blank	0.000	12:54
Subject Sample #1	0.028	12:55
Breath Volume	3.761	12:55
Air Blank	0.000	12:56
Air Blank	0.000	12:58
Subject Sample #2	0.025	12:58
Breath Volume	3.175	12:58
Air Blank	0.000	12:59
Control Test	0.077	12:59
Air Blank	0.000	13:00
Diagnostics Check	OK	13:00

Data Download Date/Time 07/29/2011 10:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANGEL, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CULVER, KRIS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/02/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 06/26/2011 21:13:34

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.079	21:11	
Air Blank	0.000	21:11	
Subject Sample #1		21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:12	
Control Test	0.080	21:12	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time 07/29/2011 10:41

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLEOCKLER, LARRY, G
BREATH TEST OPERATOR AGENCY: Arcadia PD
ARRESTING OFFICER: GLEOCKLER, LARRY
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: SIEGMAN

DATE/TIME OF TEST: 06/01/2011 23:37:22

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:34	
Control Test	0.079	23:34	
Air Blank	0.000	23:35	
Subject Sample #1		23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:36	
Control Test	0.080	23:36	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time 06/29/2011 20:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLEOCKLER, LARRY, G
BREATH TEST OPERATOR AGENCY: Arcadia PD
ARRESTING OFFICER: GLEOCKLER, LARRY
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FELDER

DATE/TIME OF TEST: 06/03/2011 01:54:39

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.079	01:52	
Air Blank	0.000	01:53	
Subject Sample #1		01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.080	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time 06/29/2011 20:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILIPOWICZ, KEVIN, M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: STEYER, MATTHEW
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 06/11/2011 02:01:14

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:55
Subject Sample #1	0.145	01:56
Breath Volume	1.878	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample #2	0.144	01:59
Breath Volume	2.968	01:59
Air Blank	0.000	02:00
Control Test	0.078	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time 06/29/2011 20:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILIPOWICZ, KEVIN, M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: JOHNSON, BILL
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: LIGON

DATE/TIME OF TEST: 06/11/2011 04:12:32

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Subject Sample #1	0.224	04:08
Breath Volume	2.363	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.210	04:10
Breath Volume	2.390	04:10
Air Blank	0.000	04:11
Control Test	0.078	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time 06/29/2011 20:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZA, STEVE, J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BIERMAN, BRANDON
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/16/2011 17:03:49

Diagnostics Check	OK	16:57
Air Blank	0.000	16:57
Control Test	0.078	16:57
Air Blank	0.000	16:58
Subject Sample #1	0.133	16:59
Breath Volume	2.273	16:59
Air Blank	0.000	16:59
Air Blank	0.000	17:01
Subject Sample #2	0.132	17:02
Breath Volume	2.460	17:02
Air Blank	0.000	17:02
Control Test	0.078	17:03
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

Data Download Date/Time 06/29/2011 20:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL, STEVEN, C
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: SCOTT, DALE
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 06/19/2011 00:20:08

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:14
Subject Sample #1	0.197	00:15
Breath Volume	1.707	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample #2	0.192	00:18
Breath Volume	1.851	00:18
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time 06/29/2011 20:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL, STEVEN, C
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BATES, JAMES
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: MCMANUS

DATE/TIME OF TEST: 06/22/2011 19:37:56

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.080	19:31
Air Blank	0.000	19:32
Subject Sample #1	0.221	19:33
Breath Volume	2.355	19:33
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample #2	0.227	19:36
Breath Volume	1.816	19:36
Air Blank	0.000	19:37
Control Test	0.078	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time 06/29/2011 20:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILIPOWICZ, KEVIN, M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: SIMMONS, BRUCE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 06/25/2011 21:45:48

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.079	21:40
Air Blank	0.000	21:40
Subject Sample #1	0.126	21:41
Breath Volume	3.500	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:43
Subject Sample #2	0.129	21:44
Breath Volume	3.390	21:44
Air Blank	0.000	21:44
Control Test	0.078	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time 06/29/2011 20:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILIPOWICZ, KEVIN, M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BAIROS, JOSEPH
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 06/26/2011 00:30:28

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.079	00:28	
Air Blank	0.000	00:28	
Subject Sample #1		00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time 06/29/2011 20:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HORST, CHRIS, A

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: DAY, DAVID

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 23:18

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MULHOLLAND

DATE/TIME OF TEST: 06/02/2011 00:03:24

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.078	00:01	
Air Blank	0.000	00:01	
Subject Sample #1		00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.078	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time 06/30/2011 23:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEVINGSTON, PHILIP, A

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: DAY, DAVID

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 22:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 06/11/2011 23:48:17

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.078	23:42
Air Blank	0.000	23:42
Subject Sample #1	0.149	23:43
Breath Volume	2.507	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample #2	0.157	23:46
Breath Volume	2.234	23:46
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time 06/30/2011 23:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HORST, CHRIS, A

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: RICHARDS, ALEX

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 04:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: MAISONET

DATE/TIME OF TEST: 06/12/2011 04:47:59

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.078	04:42
Air Blank	0.000	04:42
Subject Sample #1	0.040	04:43
Breath Volume	2.515	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:45
Subject Sample #2	0.038	04:46
Breath Volume	3.484	04:46
Air Blank	0.000	04:47
Control Test	0.077	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time 06/30/2011 23:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BANNER, JEAN, A

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: CASON, JOHN

ARRESTING OFFICER AGENCY: Lake County SO

LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30510080A1

DRY GAS EXPIRY DATE: 12/01/2012

VIOLATION CODE: DUI

ARREST TIME: 07:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: GOODBREAD

DATE/TIME OF TEST: 06/19/2011 08:55:51

Diagnostics Check	OK	08:49
Air Blank	0.000	08:49
Control Test	0.078	08:50
Air Blank	0.000	08:50
Subject Sample #1	0.127	08:51
Breath Volume	1.332	08:51
Air Blank	0.000	08:51
Air Blank	0.000	08:53
Subject Sample #2	0.127	08:54
Breath Volume	1.804	08:54
Air Blank	0.000	08:54
Control Test	0.079	08:55
Air Blank	0.000	08:55
Diagnostics Check	OK	08:55

Data Download Date/Time 06/30/2011 23:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, RICHARD, M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LUTZ, SAM
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: FINNEGAN

DATE/TIME OF TEST: 06/02/2011 04:12:11

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:06
Air Blank	0.000	04:07
Subject Sample #1	0.145	04:07
Breath Volume	1.488	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.149	04:10
Breath Volume	1.593	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/03/2011 01:58:29

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Subject Sample #1	0.141	01:53
Breath Volume	1.152	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.157	01:56
Breath Volume	1.511	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 06/03/2011 20:33:11

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.082	20:27
Air Blank	0.000	20:28
Subject Sample #1	0.157	20:28
Breath Volume	1.902	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample #2	0.153	20:31
Breath Volume	1.929	20:31
Air Blank	0.000	20:32
Control Test	0.081	20:32
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE, MARK,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FELTOVIC, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/04/2011 03:30:55

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample #1	0.167	03:26
Breath Volume	2.582	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample #2	0.157	03:29
Breath Volume	1.296	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE, MARK,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BYRD, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GERONIMO ESCOBAR

DATE/TIME OF TEST: 06/04/2011 03:47:17

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.082	03:41
Air Blank	0.000	03:42
Subject Sample #1	0.160	03:42
Breath Volume	1.707	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample #2	0.157	03:45
Breath Volume	1.531	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONEBRAKER, TROY, D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: STONEBRAKER, TROY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: PIAZZA

DATE/TIME OF TEST: 06/04/2011 19:36:32

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.082	19:31
Air Blank	0.000	19:31
Subject Sample #1	0.000	19:32
Breath Volume	1.585	19:32
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample #2	0.000	19:35
Breath Volume	2.074	19:35
Air Blank	0.000	19:35
Control Test	0.081	19:35
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, DEPUTY, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WORTHINGTON, OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: ADMIN
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: VILANTI

DATE/TIME OF TEST: 06/04/2011 22:37:43

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.082	22:30	
Air Blank	0.000	22:31	
Subject Sample #1	0.187	22:31	
Breath Volume	2.117	22:31	
Air Blank	0.000	22:32	
Air Blank	0.000	22:33	
Subject Sample #2		22:36	RFI Detect
Breath Volume	0.000	22:36	
Air Blank	0.000	22:36	
Control Test	0.081	22:37	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE, MARK,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SELL, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: FLISS

DATE/TIME OF TEST: 06/05/2011 19:31:11

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.082	19:23
Air Blank	0.000	19:24
Subject Sample #1	0.153	19:24
Breath Volume	1.390	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample #2	0.148	19:29
Breath Volume	1.449	19:29
Air Blank	0.000	19:30
Control Test	0.082	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIANDI, JAMIE, B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GNESIN, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: CHWOJKO

DATE/TIME OF TEST: 06/06/2011 21:35:43

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:24	
Control Test	0.082	21:24	
Air Blank	0.000	21:25	
Subject Sample #1		21:28	No Sample Provided
Breath Volume	0.000	21:28	
Air Blank	0.000	21:28	
Air Blank	0.000	21:30	
Subject Sample #2	0.244	21:34	Volume Not Met
Breath Volume	0.148	21:34	
Air Blank	0.000	21:34	
Control Test	0.082	21:35	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIANDI, JAMIE, B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PARIS, OFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CENATIS

DATE/TIME OF TEST: 06/07/2011 04:02:39

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.083	03:57
Air Blank	0.000	03:57
Subject Sample #1	0.000	03:58
Breath Volume	3.523	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample #2	0.000	04:01
Breath Volume	1.960	04:01
Air Blank	0.000	04:01
Control Test	0.081	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, DEPUTY, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FELTOVIC, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: FITCH

DATE/TIME OF TEST: 06/09/2011 03:01:25

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Subject Sample #1	0.086	02:57
Breath Volume	1.531	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.090	02:59
Breath Volume	1.769	02:59
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: FEHRING

DATE/TIME OF TEST: 06/11/2011 00:34:37

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.082	00:29
Air Blank	0.000	00:29
Subject Sample #1	0.151	00:30
Breath Volume	2.300	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample #2	0.150	00:33
Breath Volume	2.550	00:33
Air Blank	0.000	00:33
Control Test	0.082	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 06/12/2011 02:55:15

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.082	02:49
Air Blank	0.000	02:50
Subject Sample #1	0.104	02:50
Breath Volume	2.414	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample #2	0.101	02:53
Breath Volume	2.179	02:53
Air Blank	0.000	02:54
Control Test	0.082	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIANDI, JAMIE, B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GURNEY, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: DEXTER

DATE/TIME OF TEST: 06/13/2011 01:24:33

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.082	01:19
Air Blank	0.000	01:19
Subject Sample #1	0.196	01:20
Breath Volume	1.484	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample #2	0.211	01:23
Breath Volume	2.082	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JONES, TROOPER, M

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: JONES, TROOPER

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: DUI

ARREST TIME: 01:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 06/16/2011 03:06:46

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.082	03:01
Air Blank	0.000	03:01
Subject Sample #1	0.201	03:02
Breath Volume	2.664	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample #2	0.199	03:05
Breath Volume	2.343	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BYRD, J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 06/16/2011 04:14:50

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.082	04:02	
Air Blank	0.000	04:03	
Subject Sample #1	0.127	04:06	Volume Not Met
Breath Volume	0.269	04:06	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample #2	0.109	04:09	
Breath Volume	1.222	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample #3	0.123	04:13	
Breath Volume	1.421	04:13	
Air Blank	0.000	04:13	
Control Test	0.080	04:14	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FUSSEL, J
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: DURRANCE

DATE/TIME OF TEST: 06/16/2011 04:40:15

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.082	04:33
Air Blank	0.000	04:34
Subject Sample #1	0.117	04:34
Breath Volume	1.308	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample #2	0.125	04:38
Breath Volume	1.613	04:38
Air Blank	0.000	04:39
Control Test	0.082	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELTOVIC, DEP, M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FELTOVIC, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: SYLVESTER

DATE/TIME OF TEST: 06/17/2011 00:51:37

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:46
Subject Sample #1	0.101	00:47
Breath Volume	2.699	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample #2	0.107	00:50
Breath Volume	3.253	00:50
Air Blank	0.000	00:50
Control Test	0.082	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BYRD, J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: FAREY

DATE/TIME OF TEST: 06/17/2011 01:23:08

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Subject Sample #1	0.123	01:18
Breath Volume	2.015	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample #2	0.129	01:21
Breath Volume	2.777	01:21
Air Blank	0.000	01:22
Control Test	0.081	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FORMAL, D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WISHON

DATE/TIME OF TEST: 06/17/2011 01:42:09

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:36
Subject Sample #1	0.142	01:37
Breath Volume	3.179	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample #2	0.142	01:40
Breath Volume	2.894	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BYRD, J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SHULL

DATE/TIME OF TEST: 06/17/2011 04:06:32

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample #1	0.197	04:01
Breath Volume	4.730	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample #2	0.205	04:05
Breath Volume	4.515	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE, MARK,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FELTOVIC, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 06/17/2011 21:03:31

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Subject Sample #1	0.030	20:59
Breath Volume	1.957	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample #2	0.028	21:02
Breath Volume	1.652	21:02
Air Blank	0.000	21:02
Control Test	0.082	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE, MARK,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FELTOVIC, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MAUTNER

DATE/TIME OF TEST: 06/18/2011 02:24:44

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.082	02:19
Air Blank	0.000	02:20
Subject Sample #1	0.218	02:20
Breath Volume	2.140	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample #2	0.223	02:23
Breath Volume	2.398	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE, MARK,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FORMAL, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SCHALLER

DATE/TIME OF TEST: 06/18/2011 22:36:30

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.082	22:31
Air Blank	0.000	22:31
Subject Sample #1	0.000	22:32
Breath Volume	2.785	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample #2	0.000	22:35
Breath Volume	1.804	22:35
Air Blank	0.000	22:35
Control Test	0.082	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE, MARK,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ARTHUR, TPR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MAYERS

DATE/TIME OF TEST: 06/18/2011 23:10:43

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:05
Air Blank	0.000	23:05
Subject Sample #1	0.171	23:06
Breath Volume	2.546	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample #2	0.179	23:09
Breath Volume	2.972	23:09
Air Blank	0.000	23:09
Control Test	0.081	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YOCK, GEORGE, B

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: YOCK, GEORGE

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: DUI

ARREST TIME: 22:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/18/2011 23:54:14

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.082	23:48
Air Blank	0.000	23:48
Subject Sample #1	0.253	23:49
Breath Volume	1.890	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample #2	0.248	23:52
Breath Volume	2.414	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE, MARK,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ARTHUR, TPR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: BHAYAT

DATE/TIME OF TEST: 06/19/2011 23:23:00

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.081	23:17
Air Blank	0.000	23:17
Subject Sample #1	0.179	23:18
Breath Volume	1.871	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample #2	0.177	23:21
Breath Volume	1.582	23:21
Air Blank	0.000	23:22
Control Test	0.081	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:23

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARTHUR, MELVIN, J

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: ARTHUR, MELVIN

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: DUI

ARREST TIME: 22:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: DENISAR

DATE/TIME OF TEST: 06/21/2011 00:14:23

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.081	00:09
Air Blank	0.000	00:09
Subject Sample #1	0.215	00:09
Breath Volume	1.855	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample #2	0.215	00:12
Breath Volume	2.558	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIANDI, JAMIE, B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CONLEY, OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 06/21/2011 02:36:15

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.082	02:34	
Air Blank	0.000	02:34	
Subject Sample #1		02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:35	
Control Test	0.082	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 06/23/2011 05:03:09

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.081	04:54
Air Blank	0.000	04:54
Subject Sample #1	0.116	04:57
Breath Volume	1.238	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample #2	0.115	05:01
Breath Volume	1.187	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: RATHFELDER

DATE/TIME OF TEST: 06/23/2011 21:39:55

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.081	21:33
Air Blank	0.000	21:34
Subject Sample #1	0.182	21:35
Breath Volume	1.660	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample #2	0.176	21:38
Breath Volume	2.054	21:38
Air Blank	0.000	21:39
Control Test	0.081	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAJOR, DEPUTY, T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DILLING, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: FLIPPO

DATE/TIME OF TEST: 06/24/2011 06:31:35

Diagnostics Check	OK	06:25
Air Blank	0.000	06:25
Control Test	0.082	06:26
Air Blank	0.000	06:26
Subject Sample #1	0.180	06:27
Breath Volume	1.972	06:27
Air Blank	0.000	06:27
Air Blank	0.000	06:29
Subject Sample #2	0.169	06:30
Breath Volume	1.503	06:30
Air Blank	0.000	06:30
Control Test	0.080	06:31
Air Blank	0.000	06:31
Diagnostics Check	OK	06:31

Data Download Date/Time 06/24/2011 13:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/25/2011 00:24:15

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.082	00:18
Air Blank	0.000	00:19
Subject Sample #1	0.094	00:19
Breath Volume	1.542	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample #2	0.096	00:22
Breath Volume	1.589	00:22
Air Blank	0.000	00:23
Control Test	0.082	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BYRD, J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MIHAILOVIC

DATE/TIME OF TEST: 06/25/2011 02:18:38

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.082	02:13
Air Blank	0.000	02:13
Subject Sample #1	0.135	02:14
Breath Volume	3.480	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample #2	0.133	02:17
Breath Volume	3.031	02:17
Air Blank	0.000	02:17
Control Test	0.082	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MACWHORTER, B
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 06/25/2011 02:43:55

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.082	02:38
Air Blank	0.000	02:39
Subject Sample #1	0.094	02:39
Breath Volume	4.164	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample #2	0.093	02:42
Breath Volume	3.675	02:42
Air Blank	0.000	02:43
Control Test	0.082	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VANIK, J
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/25/2011 04:41:31

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Subject Sample #1	0.260	04:36
Breath Volume	1.121	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample #2	0.268	04:39
Breath Volume	1.546	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARTHUR, MELVIN, J

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: ARTHUR, MELVIN

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: DUI

ARREST TIME: 21:33

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: GRAVELL

DATE/TIME OF TEST: 06/26/2011 22:49:56

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.082	22:44
Air Blank	0.000	22:44
Subject Sample #1	0.000	22:45
Breath Volume	3.144	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample #2	0.000	22:48
Breath Volume	1.839	22:48
Air Blank	0.000	22:49
Control Test	0.082	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORSI, SHERRI, A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VAN ARSDALE, JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: BACHUSS

DATE/TIME OF TEST: 06/28/2011 11:07:35

Diagnostics Check	OK	11:01	
Air Blank	0.000	11:01	
Control Test	0.082	11:02	
Air Blank	0.000	11:02	
Subject Sample #1	0.078	11:03	
Breath Volume	2.457	11:03	
Air Blank	0.000	11:03	
Air Blank	0.000	11:05	
Subject Sample #2	0.076	11:06	
Breath Volume	1.539	11:06	
Air Blank	0.000	11:06	
Control Test	0.081	11:06	
Air Blank		11:07	RFI Detect
Air Blank	0.000	11:07	

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, RICHARD, M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JOHNSON, RICHARD
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 06/28/2011 23:55:31

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Subject Sample #1	0.205	23:47
Breath Volume	1.468	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample #2	0.183	23:50
Breath Volume	1.332	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample #3	0.181	23:54
Breath Volume	1.574	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE, MARK,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JOHNSON, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: METCALFE

DATE/TIME OF TEST: 06/29/2011 03:26:21

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Subject Sample #1	0.197	03:21
Breath Volume	2.023	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample #2	0.194	03:24
Breath Volume	2.128	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIANDI, JAMIE, B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FELTOVIC, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: BAGNI

DATE/TIME OF TEST: 06/30/2011 00:08:21

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.082	00:03
Air Blank	0.000	00:03
Subject Sample #1	0.000	00:04
Breath Volume	3.343	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample #2	0.000	00:07
Breath Volume	2.964	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEAD, JENNIFER, L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HEAD, JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: SUMMERS-STEELE

DATE/TIME OF TEST: 06/30/2011 00:30:34

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.081	00:22
Air Blank	0.000	00:23
Subject Sample #1	0.124	00:26
Breath Volume	1.984	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample #2	0.127	00:29
Breath Volume	1.710	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIANDI, JAMIE, B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VANIK, OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/30/2011 02:32:40

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.082	02:27
Air Blank	0.000	02:27
Subject Sample #1	0.170	02:28
Breath Volume	4.148	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample #2	0.161	02:31
Breath Volume	2.851	02:31
Air Blank	0.000	02:31
Control Test	0.078	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time 07/15/2011 12:29

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANLEY, KURTISS, L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MURREY, TPR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HYLAND

DATE/TIME OF TEST: 06/04/2011 04:47:20

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample #1	0.198	04:42
Breath Volume	7.363	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample #2	0.194	04:45
Breath Volume	7.140	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANLEY, KURTISS, L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FORMAL, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HERBER

DATE/TIME OF TEST: 06/05/2011 00:11:48

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample #1	0.108	00:07
Breath Volume	4.449	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample #2	0.108	00:10
Breath Volume	5.312	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG, CODY, M
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: LONG, CODY
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LOPES

DATE/TIME OF TEST: 06/05/2011 04:27:51

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Subject Sample #1	0.140	04:23
Breath Volume	4.609	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.135	04:26
Breath Volume	4.496	04:26
Air Blank	0.000	04:26
Control Test	0.079	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANLEY, KURTISS, L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FORMAL, DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PETTIGREW

DATE/TIME OF TEST: 06/10/2011 01:30:13

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample #1	0.185	01:25
Breath Volume	5.472	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample #2	0.166	01:28
Breath Volume	4.703	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG, CODY, M
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: LONG, CODY
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SMINK

DATE/TIME OF TEST: 06/10/2011 01:57:22

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample #1	0.000	01:52
Breath Volume	4.843	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample #2	0.000	01:56
Breath Volume	5.015	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: PICAZIO

DATE/TIME OF TEST: 06/10/2011 03:31:07

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:25	
Subject Sample #1	0.159	03:26	
Breath Volume	3.253	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample #2		03:29	Slope Not Met
Breath Volume	2.546	03:29	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: PICAZIO

DATE/TIME OF TEST: 06/10/2011 03:33:41

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.078	03:32	
Air Blank		03:33	Sequence Aborted
Air Blank	0.000	03:33	

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: PICAZIO

DATE/TIME OF TEST: 06/10/2011 03:58:03

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.078	03:52	
Air Blank	0.000	03:53	
Subject Sample #1	0.144	03:53	
Breath Volume	2.871	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample #2		03:56	Slope Not Met
Breath Volume	2.117	03:56	
Air Blank	0.000	03:57	
Control Test	0.078	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:58	

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATHMAN, JAMES, W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KATHMAN, JAMES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: PICAZIO

DATE/TIME OF TEST: 06/10/2011 04:23:42

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:18
Subject Sample #1	0.120	04:19
Breath Volume	1.269	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample #2	0.137	04:22
Breath Volume	2.183	04:22
Air Blank	0.000	04:22
Control Test	0.077	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY, JOHN, T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WHITAKER, OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FAETT

DATE/TIME OF TEST: 06/16/2011 00:08:07

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:01
Subject Sample #1	0.000	00:03
Breath Volume	2.933	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample #2	0.000	00:06
Breath Volume	2.812	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY, JOHN, T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GUINART, OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: YODER

DATE/TIME OF TEST: 06/17/2011 01:26:19

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample #1	0.066	01:22
Breath Volume	1.550	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample #2	0.063	01:24
Breath Volume	2.652	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BYRD, J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SANDOVAL VIDELA

DATE/TIME OF TEST: 06/18/2011 01:48:34

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.081	01:47	
Air Blank	0.000	01:47	
Subject Sample #1		01:47	RFI Detect
Breath Volume	0.000	01:47	
Air Blank		01:48	RFI Detect
Air Blank	0.000	01:48	

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BYRD, J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SANDOVAL VIDELA

DATE/TIME OF TEST: 06/18/2011 01:56:41

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.080	01:51	
Air Blank	0.000	01:51	
Subject Sample #1	0.245	01:52	
Breath Volume	4.406	01:52	
Air Blank	0.000	01:52	
Air Blank	0.000	01:54	
Subject Sample #2		01:55	Slope Not Met
Breath Volume	0.886	01:55	
Air Blank	0.000	01:55	
Control Test	0.080	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD, J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BYRD, J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SANDOVAL VIDELA

DATE/TIME OF TEST: 06/18/2011 02:29:49

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample #1	0.238	02:21
Breath Volume	3.156	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample #2	0.207	02:25
Breath Volume	1.589	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample #3	0.226	02:28
Breath Volume	2.496	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANLEY, KURTISS, L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: URBINA, OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 06/18/2011 02:47:22

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Subject Sample #1	0.214	02:43
Breath Volume	1.929	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample #2	0.216	02:45
Breath Volume	2.765	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time 06/24/2011 10:07

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY, JOHN, T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TARASEWICH, OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 06/26/2011 00:45:20

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Subject Sample #1	0.038	00:40
Breath Volume	2.828	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample #2	0.037	00:43
Breath Volume	2.925	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time 07/15/2011 13:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY, JOHN, T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 06/26/2011 01:11:30

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Subject Sample #1	0.111	01:06
Breath Volume	1.921	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample #2	0.119	01:10
Breath Volume	2.566	01:10
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time 07/15/2011 13:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITCHARD, DEP, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LEVESQUE, OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: ARDEN

DATE/TIME OF TEST: 06/29/2011 18:03:34

Diagnostics Check	OK	17:57
Air Blank	0.000	17:58
Control Test	0.079	17:58
Air Blank	0.000	17:58
Subject Sample #1	0.237	17:59
Breath Volume	1.289	17:59
Air Blank	0.000	17:59
Air Blank	0.000	18:01
Subject Sample #2	0.222	18:02
Breath Volume	1.171	18:02
Air Blank	0.000	18:02
Control Test	0.080	18:02
Air Blank	0.000	18:03
Diagnostics Check	OK	18:03

Data Download Date/Time 07/15/2011 13:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY, JOHN, T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FORMAL, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SAYLES

DATE/TIME OF TEST: 06/30/2011 01:56:42

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Subject Sample #1	0.110	01:52
Breath Volume	2.937	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample #2	0.102	01:55
Breath Volume	2.589	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time 07/15/2011 13:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSS, MICHAEL, G
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK, AARON
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 06/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: CARINI

DATE/TIME OF TEST: 06/30/2011 19:12:36

Diagnostics Check	OK	19:06
Air Blank	0.000	19:07
Control Test	0.080	19:07
Air Blank	0.000	19:08
Subject Sample #1	0.000	19:08
Breath Volume	3.335	19:08
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample #2	0.000	19:11
Breath Volume	2.945	19:11
Air Blank	0.000	19:11
Control Test	0.081	19:12
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time 07/15/2011 13:12

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: FRUITLAND PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATEY, JOHN, P
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: MATEY, JOHN
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 05/07/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: ZIMMER

DATE/TIME OF TEST: 06/04/2011 23:47:31

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.079	23:45	
Air Blank	0.000	23:45	
Subject Sample #1		23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Control Test	0.079	23:46	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time 06/12/2011 14:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: FRUITLAND PARK P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SIROLLI, BECKIE,

BREATH TEST OPERATOR AGENCY: Fruitland Park PD

ARRESTING OFFICER: SIROLLI, BECKIE

ARRESTING OFFICER AGENCY: Fruitland Park PD

LAST AGENCY INSPECTION: 05/07/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 781787

DRY GAS EXPIRY DATE: 06/10/2012

VIOLATION CODE: DUI

ARREST TIME: 21:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: BRUCKE

DATE/TIME OF TEST: 06/05/2011 22:58:48

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:56	
Control Test	0.080	22:56	
Air Blank	0.000	22:57	
Subject Sample #1		22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:57	
Control Test	0.079	22:58	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time 06/12/2011 14:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: FRUITLAND PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER, STEVEN, M
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: FOSTER, STEVEN
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 06/12/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 781787
DRY GAS EXPIRY DATE: 06/10/2012
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: SWEET

DATE/TIME OF TEST: 06/30/2011 20:09:23

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:07	
Control Test	0.078	20:07	
Air Blank	0.000	20:07	
Subject Sample #1		20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:08	
Control Test	0.077	20:08	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:09	

Data Download Date/Time 07/09/2011 09:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FHP TROOP D - COCOA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TAYLOR, CHANNING, L

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: THOMAS, CHARLES

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 05911080A2

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 10:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 11:04

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/06/2011 11:41:27

Diagnostics Check	OK	11:34
Air Blank	0.000	11:34
Control Test	0.080	11:35
Air Blank	0.000	11:35
Subject Sample #1	0.000	11:36
Breath Volume	1.363	11:36
Air Blank	0.000	11:36
Air Blank	0.000	11:38
Subject Sample #2	0.000	11:40
Breath Volume	1.406	11:40
Air Blank	0.000	11:40
Control Test	0.079	11:40
Air Blank	0.000	11:41
Diagnostics Check	OK	11:41

Data Download Date/Time 06/12/2011 17:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FHP TROOP D - COCOA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER, RODGER, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HILDRETH, KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05911080A2
DRY GAS EXPIRY DATE: 02/01/2013
VIOLATION CODE: DUI
ARREST TIME: 13:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:51

SUBJECT'S LAST NAME: PAGLIARO

DATE/TIME OF TEST: 06/10/2011 15:07:36

Diagnostics Check	OK	15:01
Air Blank	0.000	15:02
Control Test	0.080	15:02
Air Blank	0.000	15:02
Subject Sample #1	0.000	15:03
Breath Volume	1.570	15:03
Air Blank	0.000	15:03
Air Blank	0.000	15:05
Subject Sample #2	0.000	15:06
Breath Volume	2.074	15:06
Air Blank	0.000	15:06
Control Test	0.080	15:06
Air Blank	0.000	15:07
Diagnostics Check	OK	15:07

Data Download Date/Time 06/12/2011 17:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE, M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FELLICETTA, F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 06/01/2011 23:34:33

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:32	
Subject Sample #1		23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.078	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARTZIG, BOBBY, L

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: HARTZIG, BOBBY

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG110501

DRY GAS EXPIRY DATE: 04/15/2013

VIOLATION CODE: DUI

ARREST TIME: 04:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 06/03/2011 05:48:37

Diagnostics Check	OK	05:41
Air Blank	0.000	05:41
Control Test	0.078	05:42
Air Blank	0.000	05:42
Subject Sample #1	0.214	05:43
Breath Volume	1.996	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample #2	0.220	05:47
Breath Volume	2.527	05:47
Air Blank	0.000	05:47
Control Test	0.078	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE, M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RAY, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BILLITTERI

DATE/TIME OF TEST: 06/05/2011 00:12:03

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:04	
Control Test	0.078	00:04	
Air Blank	0.000	00:04	
Subject Sample #1		00:08	No Sample Provided
Breath Volume	0.000	00:08	
Air Blank	0.000	00:08	
Air Blank	0.000	00:10	
Subject Sample #2		00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:11	
Control Test	0.079	00:11	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:12	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE, M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SAVALOX, D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 06/05/2011 02:58:09

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Subject Sample #1	0.203	02:53
Breath Volume	2.285	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample #2	0.213	02:56
Breath Volume	2.496	02:56
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE, M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARTWICK, M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 06/05/2011 04:02:10

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.078	03:51	
Air Blank	0.000	03:51	
Subject Sample #1	0.203	03:54	Volume Not Met
Breath Volume	0.050	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample #2	0.257	04:00	Volume Not Met
Breath Volume	0.203	04:00	
Air Blank	0.000	04:01	
Control Test	0.077	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, BRIDGETT, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORRIS, BRIDGETT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 06/07/2011 05:48:55

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:46	
Control Test	0.078	05:46	
Air Blank	0.000	05:47	
Subject Sample #1		05:47	Subject Test Refused
Breath Volume	0.000	05:47	
Air Blank	0.000	05:48	
Control Test	0.078	05:48	
Air Blank	0.000	05:48	
Diagnostics Check	OK	05:48	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE, M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOVER, R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: KINCHEN

DATE/TIME OF TEST: 06/08/2011 22:47:50

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.079	22:32	
Air Blank	0.000	22:32	
Subject Sample #1		22:35	No Sample Provided
Breath Volume	0.000	22:35	
Air Blank	0.000	22:36	
Air Blank	0.000	22:38	
Subject Sample #2	0.224	22:40	
Breath Volume	1.550	22:40	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample #3	0.216	22:46	
Breath Volume	2.023	22:46	
Air Blank	0.000	22:46	
Control Test	0.076	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE, M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TRUMP, R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BIXLER

DATE/TIME OF TEST: 06/10/2011 03:53:14

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:47
Subject Sample #1	0.192	03:48
Breath Volume	2.843	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample #2	0.186	03:51
Breath Volume	2.269	03:51
Air Blank	0.000	03:52
Control Test	0.077	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, GARY, D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILLIAMS, GARY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HATTON

DATE/TIME OF TEST: 06/10/2011 04:28:29

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.077	04:22
Air Blank	0.000	04:23
Subject Sample #1	0.139	04:23
Breath Volume	2.968	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample #2	0.138	04:27
Breath Volume	3.000	04:27
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKINS, CHRISTOPHER, L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PARKINS, CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 06/10/2011 05:10:50

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.078	05:05
Air Blank	0.000	05:05
Subject Sample #1	0.087	05:06
Breath Volume	1.906	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample #2	0.095	05:09
Breath Volume	2.640	05:09
Air Blank	0.000	05:09
Control Test	0.078	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MIDDLETON, DANIEL, S

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: MIDDLETON, DANIEL

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG110501

DRY GAS EXPIRY DATE: 04/15/2013

VIOLATION CODE: DUI

ARREST TIME: 17:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: LOCKHART

DATE/TIME OF TEST: 06/10/2011 19:24:03

Diagnostics Check	OK	19:18	
Air Blank	0.000	19:19	
Control Test	0.078	19:19	
Air Blank	0.000	19:20	
Subject Sample #1		19:22	Subject Test Refused
Breath Volume	0.000	19:22	
Air Blank	0.000	19:23	
Control Test	0.077	19:23	
Air Blank	0.000	19:23	
Diagnostics Check	OK	19:24	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE, M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RAY, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/12/2011 04:29:45

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:16	
Control Test	0.078	04:16	
Air Blank	0.000	04:17	
Subject Sample #1		04:20	No Sample Provided
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Air Blank	0.000	04:22	
Subject Sample #2	0.128	04:25	
Breath Volume	1.238	04:25	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample #3	0.130	04:28	
Breath Volume	1.246	04:28	
Air Blank	0.000	04:28	
Control Test	0.078	04:29	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE, M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARTWICK, M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: RODRIGUEZ CASTILLO

DATE/TIME OF TEST: 06/12/2011 05:24:50

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:20	
Control Test	0.078	05:21	
Air Blank	0.000	05:21	
Subject Sample #1		05:23	Subject Test Refused
Breath Volume	0.000	05:23	
Air Blank	0.000	05:23	
Control Test	0.077	05:24	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORT, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMALIOS, GEORGEOS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/12/2011 21:47:46

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.077	21:42
Air Blank	0.000	21:42
Subject Sample #1	0.132	21:43
Breath Volume	2.925	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:45
Subject Sample #2	0.122	21:46
Breath Volume	1.148	21:46
Air Blank	0.000	21:46
Control Test	0.077	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, BRIDGETT, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DICICCO, DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: MCGARRY

DATE/TIME OF TEST: 06/14/2011 07:45:29

Diagnostics Check	OK	07:38
Air Blank	0.000	07:39
Control Test	0.078	07:39
Air Blank	0.000	07:39
Subject Sample #1	0.168	07:40
Breath Volume	1.261	07:40
Air Blank	0.000	07:41
Air Blank	0.000	07:43
Subject Sample #2	0.164	07:43
Breath Volume	1.878	07:43
Air Blank	0.000	07:44
Control Test	0.078	07:44
Air Blank	0.000	07:45
Diagnostics Check	OK	07:45

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE, M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EULER, K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: GEARY

DATE/TIME OF TEST: 06/15/2011 23:06:27

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.078	22:59
Air Blank	0.000	23:00
Subject Sample #1	0.035	23:01
Breath Volume	1.109	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample #2	0.034	23:04
Breath Volume	1.203	23:04
Air Blank	0.000	23:05
Control Test	0.078	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLLEY, CPL, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: OLLEY, CPL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: KANGANIS

DATE/TIME OF TEST: 06/16/2011 01:47:02

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:42
Subject Sample #1	0.261	01:42
Breath Volume	2.582	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample #2	0.252	01:45
Breath Volume	2.558	01:45
Air Blank	0.000	01:46
Control Test	0.078	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:47

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE, M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TSANAKALIOS, R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MOUNDROS

DATE/TIME OF TEST: 06/16/2011 04:03:11

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Subject Sample #1	0.182	03:57
Breath Volume	1.875	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample #2	0.176	04:01
Breath Volume	1.972	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, GARY, D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOGGESS, OSCAR
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: STEORTS

DATE/TIME OF TEST: 06/19/2011 04:52:40

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:50	
Control Test	0.078	04:50	
Air Blank	0.000	04:51	
Subject Sample #1		04:51	Subject Test Refused
Breath Volume	0.000	04:51	
Air Blank	0.000	04:51	
Control Test	0.079	04:52	
Air Blank	0.000	04:52	
Diagnostics Check	OK	04:52	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKINS, CHRISTOPHER, L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TSANAKALIOS, REGINA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 06/22/2011 05:25:24

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.078	05:19
Air Blank	0.000	05:20
Subject Sample #1	0.221	05:20
Breath Volume	2.394	05:20
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample #2	0.218	05:23
Breath Volume	2.449	05:23
Air Blank	0.000	05:24
Control Test	0.078	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARL, CHAD, M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEGERSMITH, DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 09:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: TINKHAM

DATE/TIME OF TEST: 06/23/2011 11:14:39

Diagnostics Check	OK	11:11	
Air Blank	0.000	11:12	
Control Test	0.077	11:12	
Air Blank	0.000	11:13	
Subject Sample #1		11:13	Subject Test Refused
Breath Volume	0.000	11:13	
Air Blank	0.000	11:13	
Control Test	0.078	11:14	
Air Blank	0.000	11:14	
Diagnostics Check	OK	11:14	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARL, CHAD, M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUBBARD, JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:42

SUBJECT'S LAST NAME: WINNING

DATE/TIME OF TEST: 06/23/2011 15:09:03

Diagnostics Check	OK	15:02
Air Blank	0.000	15:03
Control Test	0.078	15:03
Air Blank	0.000	15:04
Subject Sample #1	0.000	15:04
Breath Volume	1.785	15:04
Air Blank	0.000	15:05
Air Blank	0.000	15:07
Subject Sample #2	0.000	15:07
Breath Volume	1.632	15:07
Air Blank	0.000	15:08
Control Test	0.078	15:08
Air Blank	0.000	15:08
Diagnostics Check	OK	15:09

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS, P, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGLAIS, P
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BIANCO

DATE/TIME OF TEST: 06/27/2011 04:15:08

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Subject Sample #1	0.231	04:10
Breath Volume	3.136	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample #2	0.226	04:13
Breath Volume	1.953	04:13
Air Blank	0.000	04:14
Control Test	0.078	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, DEPUTY W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUBBARD, DEPUTY J.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: WHITT

DATE/TIME OF TEST: 06/27/2011 16:23:39

Diagnostics Check	OK	16:15
Air Blank	0.000	16:16
Control Test	0.078	16:16
Air Blank	0.000	16:17
Subject Sample #1	0.000	16:18
Breath Volume	1.285	16:18
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample #2	0.000	16:22
Breath Volume	1.699	16:22
Air Blank	0.000	16:22
Control Test	0.078	16:23
Air Blank	0.000	16:23
Diagnostics Check	OK	16:23

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLETON, DANIEL, S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MIDDLETON, DANIEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: RAGER

DATE/TIME OF TEST: 06/28/2011 17:00:46

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:51	
Control Test	0.078	16:52	
Air Blank	0.000	16:52	
Subject Sample #1		16:55	No Sample Provided
Breath Volume	0.000	16:55	
Air Blank	0.000	16:56	
Air Blank	0.000	16:58	
Subject Sample #2		16:59	Subject Test Refused
Breath Volume	0.000	16:59	
Air Blank	0.000	16:59	
Control Test	0.079	17:00	
Air Blank	0.000	17:00	
Diagnostics Check	OK	17:00	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON, GREG, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORTON, GREG
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: FINNEY

DATE/TIME OF TEST: 06/29/2011 00:29:34

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.078	00:27	
Air Blank	0.000	00:28	
Subject Sample #1		00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.078	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON, GREG, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DICICCO, DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BRANDENBORG

DATE/TIME OF TEST: 06/29/2011 04:05:09

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	03:59
Subject Sample #1	0.015	04:00
Breath Volume	1.671	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample #2	0.014	04:03
Breath Volume	1.945	04:03
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON, GREGORY, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORTON, GREGORY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/29/2011 05:16:36

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.078	05:14	
Air Blank	0.000	05:15	
Subject Sample #1		05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:15	
Control Test	0.078	05:16	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON, GREGORY, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORTON, GREGORY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/29/2011 05:32:56

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.078	05:26
Air Blank	0.000	05:27
Subject Sample #1	0.108	05:28
Breath Volume	1.125	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample #2	0.112	05:31
Breath Volume	1.757	05:31
Air Blank	0.000	05:32
Control Test	0.078	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ, ALVIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ACKERSON, MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: JAEGER

DATE/TIME OF TEST: 06/29/2011 18:30:40

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.078	18:24
Air Blank	0.000	18:25
Subject Sample #1	0.144	18:26
Breath Volume	1.574	18:26
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample #2	0.141	18:29
Breath Volume	1.808	18:29
Air Blank	0.000	18:29
Control Test	0.078	18:30
Air Blank	0.000	18:30
Diagnostics Check	OK	18:30

Data Download Date/Time 06/29/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM, JAMES, D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILHELM, JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG110501
DRY GAS EXPIRY DATE: 04/15/2013
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: PLEASANT

DATE/TIME OF TEST: 06/30/2011 05:27:00

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:23	
Control Test	0.078	05:23	
Air Blank	0.000	05:24	
Subject Sample #1		05:25	Subject Test Refused
Breath Volume	0.000	05:25	
Air Blank	0.000	05:26	
Control Test	0.078	05:26	
Air Blank	0.000	05:26	
Diagnostics Check	OK	05:27	

Data Download Date/Time 07/23/2011 20:59

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERRITY, SAMUEL, S
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KING, GORGON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: SHIPMAN

DATE/TIME OF TEST: 06/16/2011 20:05:11

Diagnostics Check	OK	19:59
Air Blank	0.000	19:59
Control Test	0.081	20:00
Air Blank	0.000	20:00
Subject Sample #1	0.000	20:01
Breath Volume	1.625	20:01
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample #2	0.000	20:03
Breath Volume	1.429	20:03
Air Blank	0.000	20:04
Control Test	0.080	20:04
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: MATHEWS, MARK

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 00:51

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 06/17/2011 02:08:16

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Subject Sample #1	0.146	02:03
Breath Volume	2.335	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample #2	0.147	02:06
Breath Volume	1.707	02:06
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: LINK, RAYMOND

ARRESTING OFFICER AGENCY: Maitland PD

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: HERBERT

DATE/TIME OF TEST: 06/17/2011 03:38:48

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample #1	0.228	03:34
Breath Volume	1.738	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample #2	0.216	03:37
Breath Volume	2.468	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: PAPPAS, WILLIAM

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 03:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SPICER

DATE/TIME OF TEST: 06/17/2011 04:07:00

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Subject Sample #1	0.136	04:02
Breath Volume	3.429	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample #2	0.145	04:05
Breath Volume	4.511	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:07

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DENSON, TRAVIS, W

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: MUNDEN, CHRIS

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 00:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/18/2011 01:34:27

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Subject Sample #1	0.209	01:29
Breath Volume	3.757	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample #2	0.202	01:32
Breath Volume	4.070	01:32
Air Blank	0.000	01:33
Control Test	0.078	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DENSON, TRAVIS, W

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: LINK, RAYMOND

ARRESTING OFFICER AGENCY: Maitland PD

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 06/18/2011 03:38:36

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Subject Sample #1		03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BURKE, JIMMY, A

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: JONAS, DANIEL

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 03:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: BITTENCOURT

DATE/TIME OF TEST: 06/18/2011 04:32:03

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Subject Sample #1	0.113	04:27
Breath Volume	3.082	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample #2	0.108	04:30
Breath Volume	1.765	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:32

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DENSON, TRAVIS, W

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: AVILA, JOHNNY

ARRESTING OFFICER AGENCY: Winter Garden PD

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 23:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: WINSETT

DATE/TIME OF TEST: 06/20/2011 00:57:08

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample #1	0.185	00:52
Breath Volume	2.070	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample #2	0.193	00:55
Breath Volume	1.910	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOPKINS, SHELBY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 06/20/2011 21:16:57

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:08
Subject Sample #1	0.196	21:08
Breath Volume	1.457	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample #2	0.218	21:11
Breath Volume	1.871	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample #3	0.211	21:15
Breath Volume	1.503	21:15
Air Blank	0.000	21:16
Control Test	0.079	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARMAND, ELVIE, D

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: BLONDIN, D

ARRESTING OFFICER AGENCY: Orlando PD

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MANGAN

DATE/TIME OF TEST: 06/21/2011 03:41:40

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Subject Sample #1	0.188	03:36
Breath Volume	1.601	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample #2	0.197	03:40
Breath Volume	2.742	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLEMAN, BRETT, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: RAMIREZ, JOSE

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BELLUCCIO

DATE/TIME OF TEST: 06/23/2011 04:04:51

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:59
Air Blank	0.000	03:59
Subject Sample #1	0.112	04:00
Breath Volume	3.511	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample #2	0.115	04:03
Breath Volume	3.285	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time 06/23/2011 15:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN, JAMES, E
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHAPLIN, CHRISTOPHER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 06/24/2011 03:14:30

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.081	02:59	
Air Blank	0.000	02:59	
Subject Sample #1	0.153	03:02	Volume Not Met
Breath Volume	0.621	03:02	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample #2	0.150	03:07	
Breath Volume	1.332	03:07	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample #3	0.154	03:12	Volume Not Met
Breath Volume	0.093	03:12	
Air Blank	0.000	03:13	
Control Test	0.078	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN, JAMES, E
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BONNER, CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: FONTANEZ ORTIZ

DATE/TIME OF TEST: 06/25/2011 02:49:24

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample #1	0.140	02:44
Breath Volume	3.101	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample #2	0.145	02:47
Breath Volume	4.335	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN, JAMES, E
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BELLOCHI, JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: SCOVILLE

DATE/TIME OF TEST: 06/25/2011 21:00:21

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.081	20:50
Air Blank	0.000	20:51
Subject Sample #1	0.318	20:52
Breath Volume	1.250	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample #2	0.355	20:55
Breath Volume	2.667	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample #3	0.336	20:58
Breath Volume	2.386	20:58
Air Blank	0.000	20:59
Control Test	0.078	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARMAND, ELVIE, D

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: MAESTRE, F

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 01:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/26/2011 02:12:03

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:06
Subject Sample #1	0.171	02:07
Breath Volume	3.511	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample #2	0.169	02:10
Breath Volume	3.199	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:12

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERRITY, SAMUEL, S
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LINK, RAYMOND
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 06/26/2011 03:34:27

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.081	03:29
Air Blank	0.000	03:29
Subject Sample #1	0.144	03:30
Breath Volume	3.277	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample #2	0.127	03:32
Breath Volume	1.320	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: WYNN, GREGG

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: FOULK

DATE/TIME OF TEST: 06/26/2011 04:06:56

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:01
Subject Sample #1	0.206	04:02
Breath Volume	3.132	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample #2	0.197	04:05
Breath Volume	1.636	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMAND, ELVIE, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BOYADJIEV, D
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: REYES GUTIERREZ

DATE/TIME OF TEST: 06/27/2011 03:08:14

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample #1	0.189	03:02
Breath Volume	1.300	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample #2	0.183	03:06
Breath Volume	1.699	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERRITY, SAMUEL, S
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GIARDIELLO, AL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 06/29/2011 02:49:56

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.077	02:44
Air Blank	0.000	02:45
Subject Sample #1	0.196	02:45
Breath Volume	2.664	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample #2	0.195	02:48
Breath Volume	2.804	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAMKISHUN, PAMELA, J

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: YOUNG, MICHAEL

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 01:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 06/30/2011 02:39:50

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample #1	0.164	02:35
Breath Volume	4.371	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample #2	0.167	02:38
Breath Volume	3.378	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN, JAMES, E
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ, JOSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: KAZAROS

DATE/TIME OF TEST: 06/30/2011 03:19:16

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.081	03:13
Air Blank	0.000	03:14
Subject Sample #1	0.000	03:14
Breath Volume	2.855	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample #2	0.000	03:17
Breath Volume	3.132	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMKISHUN, PAMELA, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GAMBRELL, CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: GETSEE

DATE/TIME OF TEST: 06/30/2011 03:51:11

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Subject Sample #1	0.189	03:46
Breath Volume	3.960	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample #2	0.191	03:49
Breath Volume	3.917	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time 06/30/2011 16:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDEN, WILLIAM, L

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: GIARDIELLO, AL

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 00:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: WONG

DATE/TIME OF TEST: 06/17/2011 01:24:48

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:19
Air Blank	0.000	01:19
Subject Sample #1	0.035	01:20
Breath Volume	3.148	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample #2	0.032	01:23
Breath Volume	1.394	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time 06/23/2011 15:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDEN, WILLIAM, L

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: DANJOU, SCOTT

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: LAM

DATE/TIME OF TEST: 06/17/2011 03:38:30

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample #1	0.179	03:33
Breath Volume	2.007	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample #2	0.175	03:37
Breath Volume	1.464	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time 06/23/2011 15:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WENNER, MICA
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SILVEIRA

DATE/TIME OF TEST: 06/17/2011 04:19:32

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample #1	0.163	04:15
Breath Volume	2.074	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.165	04:18
Breath Volume	1.828	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time 06/23/2011 15:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDEN, WILLIAM, L

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: YOUNG, MICHAEL

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SAFFORD

DATE/TIME OF TEST: 06/18/2011 03:33:22

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Subject Sample #1	0.114	03:28
Breath Volume	1.523	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample #2	0.123	03:31
Breath Volume	1.312	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time 06/23/2011 15:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARDINAL, TIM
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 904241
DRY GAS EXPIRY DATE: 11/12/2012
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: SEGOVIA ACEVEDO

DATE/TIME OF TEST: 06/18/2011 06:22:03

Diagnostics Check	OK	06:12
Air Blank	0.000	06:12
Control Test	0.080	06:12
Air Blank	0.000	06:13
Subject Sample #1	0.176	06:14
Breath Volume	3.027	06:14
Air Blank	0.000	06:15
Air Blank	0.000	06:17
Subject Sample #2	0.151	06:17
Breath Volume	1.101	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample #3	0.170	06:20
Breath Volume	2.585	06:20
Air Blank	0.000	06:21
Control Test	0.079	06:21
Air Blank	0.000	06:21
Diagnostics Check	OK	06:22

Data Download Date/Time 06/23/2011 15:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDEN, WILLIAM, L

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: WENNER, MICA

ARRESTING OFFICER AGENCY: Central FL Univ PD

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/21/2011 03:37:21

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample #1	0.126	03:32
Breath Volume	2.261	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.129	03:35
Breath Volume	2.687	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time 06/23/2011 15:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDEN, WILLIAM, L

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: ALVAREZ, HECTOR

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 22:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: KIAH

DATE/TIME OF TEST: 06/21/2011 23:49:10

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.081	23:47	
Air Blank	0.000	23:47	
Subject Sample #1		23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:48	
Control Test	0.081	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time 06/23/2011 15:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDEN, WILLIAM, L

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: KING, GORDON

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 04:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: MALDONADO CAMACHO

DATE/TIME OF TEST: 06/22/2011 05:55:01

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.081	05:47
Air Blank	0.000	05:48
Subject Sample #1	0.191	05:50
Breath Volume	1.843	05:50
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample #2	0.183	05:53
Breath Volume	1.902	05:53
Air Blank	0.000	05:54
Control Test	0.079	05:54
Air Blank	0.000	05:54
Diagnostics Check	OK	05:55

Data Download Date/Time 06/23/2011 15:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDEN, WILLIAM, L

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: BOURQUE, JESSIE

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 01:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 06/23/2011 03:49:46

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:42
Subject Sample #1	0.188	03:45
Breath Volume	2.574	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample #2	0.194	03:48
Breath Volume	1.355	03:48
Air Blank	0.000	03:48
Control Test	0.078	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time 06/23/2011 15:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDEN, WILLIAM, L

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: DOBBINS, SEAN

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 05:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/23/2011 06:06:06

Diagnostics Check	OK	05:59
Air Blank	0.000	06:00
Control Test	0.080	06:00
Air Blank	0.000	06:00
Subject Sample #1	0.123	06:01
Breath Volume	4.468	06:01
Air Blank	0.000	06:02
Air Blank	0.000	06:04
Subject Sample #2	0.121	06:04
Breath Volume	4.515	06:04
Air Blank	0.000	06:05
Control Test	0.080	06:05
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time 06/23/2011 15:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDEN, WILLIAM, L

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: DANJOU, SCOTT

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HAMASHIN

DATE/TIME OF TEST: 06/24/2011 04:25:26

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:19
Subject Sample #1	0.169	04:20
Breath Volume	3.406	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.171	04:23
Breath Volume	3.593	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time 06/30/2011 16:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARDEN, WILLIAM, L

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: YOUNG, MICHAEL

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 06/16/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 01:53

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/29/2011 03:56:31

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Subject Sample #1	0.107	03:52
Breath Volume	2.460	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample #2	0.114	03:55
Breath Volume	2.921	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time 06/30/2011 16:02

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS, CHRISTOPHER, L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS, CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 06/04/2011 01:21:13

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:15
Air Blank	0.000	01:15
Subject Sample #1	0.115	01:16
Breath Volume	2.656	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample #2	0.114	01:19
Breath Volume	3.253	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY, J, K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY, J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DE NOVA-NUNEZ

DATE/TIME OF TEST: 06/04/2011 03:00:58

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Subject Sample #1	0.194	02:55
Breath Volume	1.128	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample #2	0.197	02:59
Breath Volume	2.402	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, BRIAN, T
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PATTERSON, BRIAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: JUAN-JOSE

DATE/TIME OF TEST: 06/04/2011 19:15:53

Diagnostics Check	OK	19:09
Air Blank	0.000	19:10
Control Test	0.079	19:10
Air Blank	0.000	19:10
Subject Sample #1	0.280	19:11
Breath Volume	2.179	19:11
Air Blank	0.000	19:12
Air Blank	0.000	19:13
Subject Sample #2	0.265	19:14
Breath Volume	1.402	19:14
Air Blank	0.000	19:14
Control Test	0.078	19:15
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDHOFF, ANDREAS, W
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SUDHOFF, ANDREAS
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 06/05/2011 03:03:54

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Subject Sample #1	0.228	02:53
Breath Volume	1.832	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.201	02:58
Breath Volume	2.375	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample #3	0.204	03:02
Breath Volume	2.265	03:02
Air Blank	0.000	03:02
Control Test	0.078	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY, J, K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY, J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 06/05/2011 23:25:50

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:20
Subject Sample #1	0.141	23:21
Breath Volume	2.238	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample #2	0.142	23:24
Breath Volume	1.562	23:24
Air Blank	0.000	23:24
Control Test	0.078	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BULLOCK, JOSEPH, J

BREATH TEST OPERATOR AGENCY: FHP Troop L

ARRESTING OFFICER: BULLOCK, JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop L

LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 11:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GANZ-LABOMBARD

DATE/TIME OF TEST: 06/08/2011 00:58:49

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:53
Subject Sample #1	0.000	00:54
Breath Volume	2.515	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample #2	0.000	00:57
Breath Volume	1.984	00:57
Air Blank	0.000	00:57
Control Test	0.079	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNOLDS, RICHARD, A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: COKER, HARRY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/11/2011 03:20:21

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample #1	0.244	03:15
Breath Volume	2.929	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample #2	0.226	03:18
Breath Volume	1.437	03:18
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DURAN, DANIEL, M

BREATH TEST OPERATOR AGENCY: Stuart PD

ARRESTING OFFICER: DURAN, DANIEL

ARRESTING OFFICER AGENCY: Stuart PD

LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 00:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MONETTE

DATE/TIME OF TEST: 06/12/2011 01:17:31

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:06	
Subject Sample #1	0.082	01:09	Volume Not Met
Breath Volume	0.542	01:09	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample #2	0.078	01:12	
Breath Volume	2.371	01:12	
Air Blank	0.000	01:13	
Air Blank	0.000	01:15	
Subject Sample #3	0.069	01:15	
Breath Volume	2.109	01:15	
Air Blank	0.000	01:16	
Control Test	0.078	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURAN, DANIEL, M
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: DURAN, DANIEL
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MONETTE

DATE/TIME OF TEST: 06/12/2011 01:30:23

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample #1	0.065	01:25
Breath Volume	2.171	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample #2	0.068	01:28
Breath Volume	1.753	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: CERNUTO, MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 06/12/2011 22:59:17

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Subject Sample #1	0.258	22:54
Breath Volume	2.632	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample #2	0.240	22:57
Breath Volume	1.863	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDHOFF, ANDREAS, W
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SUDHOFF, ANDREAS
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: VOGEL

DATE/TIME OF TEST: 06/13/2011 00:32:22

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.080	00:21	
Air Blank	0.000	00:21	
Subject Sample #1	0.300	00:25	Volume Not Met
Breath Volume	0.097	00:25	
Air Blank	0.000	00:25	
Air Blank	0.000	00:27	
Subject Sample #2	0.247	00:30	Volume Not Met
Breath Volume	0.230	00:30	
Air Blank	0.000	00:31	
Control Test	0.078	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLOCK, JOSEPH, J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BULLOCK, JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/16/2011 02:37:39

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.079	02:26	
Air Blank	0.000	02:27	
Subject Sample #1	0.123	02:30	Volume Not Met
Breath Volume	0.050	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:32	
Subject Sample #2	0.100	02:36	Volume Not Met
Breath Volume	0.292	02:36	
Air Blank	0.000	02:36	
Control Test	0.079	02:37	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLOCK, JOSEPH, J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BULLOCK, JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/16/2011 02:48:40

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.078	02:40	
Air Blank	0.000	02:41	
Subject Sample #1	0.109	02:44	Volume Not Met
Breath Volume	0.476	02:44	
Air Blank	0.000	02:44	
Air Blank	0.000	02:46	
Subject Sample #2		02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.078	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, BRADFORD, B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER, BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LEFFEL

DATE/TIME OF TEST: 06/16/2011 04:02:50

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:56
Air Blank	0.000	03:56
Subject Sample #1	0.189	03:57
Breath Volume	1.160	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample #2	0.193	04:01
Breath Volume	1.523	04:01
Air Blank	0.000	04:01
Control Test	0.077	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS, CHRISTOPHER, L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS, CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: LINES

DATE/TIME OF TEST: 06/17/2011 21:56:39

Diagnostics Check	OK	21:49
Air Blank	0.000	21:50
Control Test	0.079	21:50
Air Blank	0.000	21:51
Subject Sample #1	0.201	21:51
Breath Volume	1.804	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample #2	0.191	21:55
Breath Volume	1.601	21:55
Air Blank	0.000	21:55
Control Test	0.079	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY, J, K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY, J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/17/2011 23:10:30

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:05
Air Blank	0.000	23:05
Subject Sample #1	0.164	23:05
Breath Volume	1.765	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample #2	0.152	23:08
Breath Volume	2.847	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY, J, K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY, J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: PENDARVIS

DATE/TIME OF TEST: 06/18/2011 01:43:43

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Subject Sample #1	0.199	01:39
Breath Volume	2.183	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample #2	0.205	01:42
Breath Volume	3.203	01:42
Air Blank	0.000	01:42
Control Test	0.078	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DURAN, DANIEL, M

BREATH TEST OPERATOR AGENCY: Stuart PD

ARRESTING OFFICER: DURAN, DANIEL

ARRESTING OFFICER AGENCY: Stuart PD

LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 01:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: COTTO

DATE/TIME OF TEST: 06/19/2011 02:13:09

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Subject Sample #1	0.221	02:08
Breath Volume	3.640	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample #2	0.208	02:11
Breath Volume	3.671	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS, CHRISTOHER, L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS, CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: AHLSTROM

DATE/TIME OF TEST: 06/19/2011 23:30:22

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.080	23:25
Air Blank	0.000	23:25
Subject Sample #1	0.131	23:26
Breath Volume	1.578	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample #2	0.130	23:28
Breath Volume	1.707	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY, J, K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY, J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MIANO

DATE/TIME OF TEST: 06/20/2011 02:28:08

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Subject Sample #1	0.215	02:23
Breath Volume	2.742	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample #2	0.224	02:26
Breath Volume	2.421	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, BRADFORD, B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER, BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 06/21/2011 04:54:54

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:50
Subject Sample #1	0.137	04:50
Breath Volume	2.312	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample #2	0.134	04:53
Breath Volume	2.156	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, BRADFORD, B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER, BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: PELLEGRINI

DATE/TIME OF TEST: 06/21/2011 20:50:34

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.079	20:45
Air Blank	0.000	20:45
Subject Sample #1	0.140	20:46
Breath Volume	2.378	20:46
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample #2	0.137	20:49
Breath Volume	2.726	20:49
Air Blank	0.000	20:49
Control Test	0.077	20:49
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS, CHRISTOPHER, L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS, CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PEREZ-GARCIA

DATE/TIME OF TEST: 06/25/2011 03:42:49

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample #1	0.177	03:38
Breath Volume	1.800	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample #2	0.167	03:41
Breath Volume	2.304	03:41
Air Blank	0.000	03:41
Control Test	0.078	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time 06/25/2011 09:49

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIATAK, JAMES, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PIATAK, JAMES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 09:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:13

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 06/26/2011 11:52:19

Diagnostics Check	OK	11:46
Air Blank	0.000	11:46
Control Test	0.078	11:46
Air Blank	0.000	11:47
Subject Sample #1	0.233	11:47
Breath Volume	2.039	11:47
Air Blank	0.000	11:48
Air Blank	0.000	11:50
Subject Sample #2	0.232	11:50
Breath Volume	1.421	11:50
Air Blank	0.000	11:51
Control Test	0.078	11:51
Air Blank	0.000	11:52
Diagnostics Check	OK	11:52

Data Download Date/Time 07/15/2011 10:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, BRADFORD, B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER, BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: POINTER

DATE/TIME OF TEST: 06/29/2011 02:22:44

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:17
Subject Sample #1	0.182	02:18
Breath Volume	2.933	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample #2	0.174	02:21
Breath Volume	2.417	02:21
Air Blank	0.000	02:21
Control Test	0.078	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time 07/15/2011 10:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDHOFF, ANDREAS, W
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SUDHOFF, ANDREAS
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/25/2011

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 06/30/2011 05:05:12

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.079	05:01	
Air Blank	0.000	05:01	
Subject Sample #1	0.124	05:03	
Breath Volume	1.890	05:03	
Air Blank		05:04	Purge Fail
Air Blank		05:05	Purge Fail

Data Download Date/Time 07/15/2011 10:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDHOFF, ANDREAS, W
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SUDHOFF, ANDREAS
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 06/30/2011 05:10:58

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:10	
Control Test	0.061	05:10	Interferent Detect
Air Blank	0.000	05:10	

Data Download Date/Time 07/15/2011 10:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDHOFF, ANDREAS, W
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SUDHOFF, ANDREAS
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 06/30/2011 05:17:54

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:16	
Control Test	0.068	05:17	Control Outside Tolerance
Air Blank	0.000	05:17	

Data Download Date/Time 07/15/2011 10:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILCKEN, JOHN, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILCKEN, JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: BAUM

DATE/TIME OF TEST: 06/30/2011 16:40:54

Diagnostics Check	OK	16:33
Air Blank	0.000	16:33
Control Test	0.077	16:34
Air Blank	0.000	16:34
Subject Sample #1	0.123	16:35
Breath Volume	1.171	16:35
Air Blank	0.000	16:36
Air Blank	0.000	16:37
Subject Sample #2	0.117	16:39
Breath Volume	1.156	16:39
Air Blank	0.000	16:39
Control Test	0.075	16:40
Air Blank	0.000	16:40
Diagnostics Check	OK	16:40

Data Download Date/Time 07/15/2011 10:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, BRADFORD, B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER, BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/25/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: CARLETTI

DATE/TIME OF TEST: 06/30/2011 22:31:10

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.078	22:20	
Air Blank	0.000	22:21	
Subject Sample #1	0.211	22:24	Volume Not Met
Breath Volume	0.144	22:24	
Air Blank	0.000	22:24	
Air Blank	0.000	22:26	
Subject Sample #2		22:29	No Sample Provided
Breath Volume	0.000	22:29	
Air Blank	0.000	22:30	
Control Test	0.077	22:30	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time 07/15/2011 10:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH, DEPUTY, J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CLARK, OFFICER
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/12/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 06/03/2011 04:24:13

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.081	04:12	
Air Blank	0.000	04:12	
Subject Sample #1		04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample #2	0.238	04:19	
Breath Volume	2.953	04:19	
Air Blank	0.000	04:20	
Air Blank	0.000	04:21	
Subject Sample #3	0.229	04:22	
Breath Volume	1.718	04:22	
Air Blank	0.000	04:23	
Control Test	0.080	04:23	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time 06/10/2011 19:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH, ROBERT, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PINTER, J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 05/12/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 817739
DRY GAS EXPIRY DATE: 08/20/2012
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ROBB

DATE/TIME OF TEST: 06/04/2011 01:54:53

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:42	
Air Blank	0.000	01:42	
Subject Sample #1	0.158	01:44	
Breath Volume	1.679	01:44	
Air Blank	0.000	01:44	
Air Blank	0.000	01:46	
Subject Sample #2	0.132	01:49	Volume Not Met
Breath Volume	0.515	01:49	
Air Blank	0.000	01:50	
Air Blank	0.000	01:52	
Subject Sample #3		01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.077	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time 06/10/2011 19:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH, ROBERT, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS, JACOB
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/12/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 817739
DRY GAS EXPIRY DATE: 08/20/2012
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: WOODALL

DATE/TIME OF TEST: 06/09/2011 00:49:30

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample #1	0.151	00:44
Breath Volume	1.539	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample #2	0.152	00:47
Breath Volume	1.164	00:47
Air Blank	0.000	00:48
Control Test	0.077	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time 06/10/2011 19:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH, ROBERT, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GLENN, CHRIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/12/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 817739
DRY GAS EXPIRY DATE: 08/20/2012
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: COLITZAS

DATE/TIME OF TEST: 06/10/2011 04:00:55

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.080	03:52	
Air Blank	0.000	03:52	
Subject Sample #1		03:55	No Sample Provided
Breath Volume	0.000	03:55	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample #2		03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.080	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time 06/10/2011 19:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSE, A, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RUSSELL, R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 06/11/2011 00:51:34

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:46
Subject Sample #1	0.000	00:47
Breath Volume	1.878	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample #2	0.000	00:50
Breath Volume	2.222	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time 06/20/2011 21:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSE, A, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HYATT, R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 06/11/2011 01:39:45

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.079	01:29	
Air Blank	0.000	01:29	
Subject Sample #1		01:32	No Sample Provided
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Air Blank	0.000	01:35	
Subject Sample #2		01:38	No Sample Provided
Breath Volume	0.000	01:38	
Air Blank	0.000	01:38	
Control Test	0.080	01:39	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time 06/20/2011 21:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSE, A, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HYATT, R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HARSEY

DATE/TIME OF TEST: 06/12/2011 02:56:56

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:51
Air Blank	0.000	02:51
Subject Sample #1	0.161	02:52
Breath Volume	1.898	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample #2	0.154	02:55
Breath Volume	1.964	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time 06/20/2011 21:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUEFLOHN, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CLARK, L
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 06/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: BEDFORD

DATE/TIME OF TEST: 06/16/2011 20:49:54

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Subject Sample #1	0.194	20:45
Breath Volume	1.933	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample #2	0.197	20:48
Breath Volume	1.605	20:48
Air Blank	0.000	20:49
Control Test	0.078	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time 06/20/2011 21:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH, ROBERT, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RUBY, CHRIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 817739
DRY GAS EXPIRY DATE: 08/20/2012
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 06/20/2011 00:10:30

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:08	
Air Blank	0.000	00:08	
Subject Sample #1		00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time 06/20/2011 21:15

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LILLEY, KATE, A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: LILLEY, KATE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670753
DRY GAS EXPIRY DATE: 09/30/2011
VIOLATION CODE: DUI
ARREST TIME: 08:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:14

SUBJECT'S LAST NAME: NEWCOMB

DATE/TIME OF TEST: 06/06/2011 09:51:34

Diagnostics Check	OK	09:44
Air Blank	0.000	09:44
Control Test	0.080	09:45
Air Blank	0.000	09:45
Subject Sample #1	0.262	09:46
Breath Volume	1.550	09:46
Air Blank	0.000	09:47
Air Blank	0.000	09:49
Subject Sample #2	0.263	09:49
Breath Volume	1.710	09:49
Air Blank	0.000	09:50
Control Test	0.077	09:50
Air Blank	0.000	09:51
Diagnostics Check	OK	09:51

Data Download Date/Time 06/30/2011 14:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOUTH, TIMOTHY, R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CAPPS, JOSH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670753
DRY GAS EXPIRY DATE: 09/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/08/2011 04:05:48

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.079	04:03	
Air Blank	0.000	04:03	
Subject Sample #1		04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.080	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time 06/30/2011 14:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENTLEY, TIMOTHY, A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HERNANDEZ, MICHAEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670753
DRY GAS EXPIRY DATE: 09/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RAWUS

DATE/TIME OF TEST: 06/11/2011 03:46:13

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.080	03:35	
Air Blank	0.000	03:35	
Subject Sample #1	0.116	03:38	Volume Not Met
Breath Volume	0.367	03:38	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample #2	0.106	03:44	Volume Not Met
Breath Volume	0.929	03:44	
Air Blank	0.000	03:45	
Control Test	0.077	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time 06/30/2011 14:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON, LARRY, P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MISTRIC, B
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670753
DRY GAS EXPIRY DATE: 09/30/2011
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/18/2011 18:19:34

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:06	
Control Test	0.078	18:06	
Air Blank	0.000	18:07	
Subject Sample #1	0.138	18:10	Volume Not Met
Breath Volume	0.691	18:10	
Air Blank	0.000	18:11	
Air Blank	0.000	18:13	
Subject Sample #2	0.133	18:14	
Breath Volume	1.312	18:14	
Air Blank	0.000	18:15	
Air Blank	0.000	18:16	
Subject Sample #3	0.142	18:17	
Breath Volume	2.128	18:17	
Air Blank	0.000	18:18	
Control Test	0.076	18:18	
Air Blank	0.000	18:19	
Diagnostics Check	OK	18:19	

Data Download Date/Time 06/30/2011 14:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPKINS, DAVID, J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PERRY, TRAVIS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/24/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670753
DRY GAS EXPIRY DATE: 09/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HIGDON

DATE/TIME OF TEST: 06/22/2011 22:41:09

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.079	22:31	
Air Blank	0.000	22:31	
Subject Sample #1	0.196	22:34	Volume Not Met
Breath Volume	0.035	22:34	
Air Blank	0.000	22:35	
Air Blank	0.000	22:37	
Subject Sample #2		22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:40	
Control Test	0.076	22:40	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time 06/30/2011 14:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCSWINEY, RICHARD, A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCSWINEY, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670753
DRY GAS EXPIRY DATE: 09/30/2011
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: KALAMA

DATE/TIME OF TEST: 06/29/2011 23:47:16

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample #1	0.202	23:42
Breath Volume	2.390	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample #2	0.200	23:45
Breath Volume	1.675	23:45
Air Blank	0.000	23:46
Control Test	0.077	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time 06/30/2011 14:06

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHAEL, STEPHEN, A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SMITH, BERNARD
ARRESTING OFFICER AGENCY: Crescent City PD
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AGO11802
DRY GAS EXPIRY DATE: 04/28/2012
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 06/17/2011 04:35:46

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Subject Sample #1	0.086	04:30
Breath Volume	1.507	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample #2	0.086	04:34
Breath Volume	1.750	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time 07/25/2011 19:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONOVER, RICHARD, T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CONOVER, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG032601
DRY GAS EXPIRY DATE: 11/22/2012
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/18/2011 01:07:28

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Subject Sample #1	0.208	01:02
Breath Volume	2.578	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample #2	0.213	01:05
Breath Volume	3.207	01:05
Air Blank	0.000	01:06
Control Test	0.077	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time 07/25/2011 19:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHAEL, STEPHEN, A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MICHAEL, STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011802
DRY GAS EXPIRY DATE: 04/28/2012
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 06/18/2011 06:31:09

Diagnostics Check	OK	06:24
Air Blank	0.000	06:25
Control Test	0.078	06:25
Air Blank	0.000	06:25
Subject Sample #1	0.128	06:26
Breath Volume	1.878	06:26
Air Blank	0.000	06:27
Air Blank	0.000	06:29
Subject Sample #2	0.124	06:29
Breath Volume	1.550	06:29
Air Blank	0.000	06:30
Control Test	0.079	06:30
Air Blank	0.000	06:31
Diagnostics Check	OK	06:31

Data Download Date/Time 07/25/2011 19:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONOVER, RICHARD, T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CONOVER, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG032601
DRY GAS EXPIRY DATE: 11/22/2012
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 06/27/2011 00:32:55

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.079	00:17	
Air Blank	0.000	00:18	
Subject Sample #1	0.144	00:21	Volume Not Met
Breath Volume	0.804	00:21	
Air Blank	0.000	00:22	
Air Blank	0.000	00:23	
Subject Sample #2	0.126	00:25	
Breath Volume	1.273	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:28	
Subject Sample #3	0.122	00:30	
Breath Volume	1.843	00:30	
Air Blank	0.000	00:31	
Control Test	0.074	00:32	Control Outside Tolerance
Air Blank	0.000	00:32	

Data Download Date/Time 07/25/2011 19:31

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001650

INSTRUMENT REGISTERED TO: KEY BISCAZYNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAIME, RAUL,
BREATH TEST OPERATOR AGENCY: Key Biscayne PD
ARRESTING OFFICER: HUERTAS, FELIX
ARRESTING OFFICER AGENCY: Key Biscayne PD
LAST AGENCY INSPECTION: 06/09/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG016502
DRY GAS EXPIRY DATE: 06/14/2012
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: MERCHANT

DATE/TIME OF TEST: 06/14/2011 10:08:29

Diagnostics Check	OK	10:02
Air Blank	0.000	10:02
Control Test	0.081	10:02
Air Blank	0.000	10:03
Subject Sample #1	0.118	10:04
Breath Volume	1.691	10:04
Air Blank	0.000	10:04
Air Blank	0.000	10:06
Subject Sample #2	0.114	10:06
Breath Volume	1.933	10:06
Air Blank	0.000	10:07
Control Test	0.081	10:07
Air Blank	0.000	10:08
Diagnostics Check	OK	10:08

Data Download Date/Time 08/31/2011 12:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALLS, DANIEL, S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SALLS, DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/01/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 731636
DRY GAS EXPIRY DATE: 03/08/2012
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: CRUZ LUNA

DATE/TIME OF TEST: 06/24/2011 18:59:10

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.081	18:53
Air Blank	0.000	18:53
Subject Sample #1	0.230	18:54
Breath Volume	1.625	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:56
Subject Sample #2	0.222	18:57
Breath Volume	2.417	18:57
Air Blank	0.000	18:58
Control Test	0.080	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time 07/04/2011 16:52

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALLS, DANIEL, S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SALLS, DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/01/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 731636
DRY GAS EXPIRY DATE: 03/08/2012
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: LOZANO GONZALEZ

DATE/TIME OF TEST: 06/25/2011 16:31:39

Diagnostics Check	OK	16:24
Air Blank	0.000	16:25
Control Test	0.081	16:25
Air Blank	0.000	16:26
Subject Sample #1	0.231	16:27
Breath Volume	1.523	16:27
Air Blank	0.000	16:27
Air Blank	0.000	16:29
Subject Sample #2	0.229	16:29
Breath Volume	1.761	16:29
Air Blank	0.000	16:30
Control Test	0.081	16:31
Air Blank	0.000	16:31
Diagnostics Check	OK	16:31

Data Download Date/Time 07/04/2011 16:52

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALLS, DANIEL, S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SALLS, DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/01/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 731636
DRY GAS EXPIRY DATE: 03/08/2012
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MAGLOIRE

DATE/TIME OF TEST: 06/26/2011 03:06:58

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample #1	0.240	02:57
Breath Volume	1.160	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample #2	0.219	03:02
Breath Volume	1.703	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample #3	0.211	03:05
Breath Volume	1.210	03:05
Air Blank	0.000	03:06
Control Test	0.077	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time 07/04/2011 16:52

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSNELL, JEFFREY, K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: CALLAN

DATE/TIME OF TEST: 06/01/2011 20:00:35

Diagnostics Check	OK	19:57	
Air Blank	0.000	19:58	
Control Test	0.080	19:58	
Air Blank	0.000	19:58	
Subject Sample #1		19:59	Subject Test Refused
Breath Volume	0.000	19:59	
Air Blank	0.000	19:59	
Control Test	0.081	20:00	
Air Blank	0.000	20:00	
Diagnostics Check	OK	20:00	

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE JR, SCOTT, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: PAULTON

DATE/TIME OF TEST: 06/04/2011 04:59:19

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.081	04:47	
Air Blank	0.000	04:48	
Subject Sample #1	0.154	04:51	Volume Not Met
Breath Volume	0.171	04:51	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample #2	0.162	04:54	
Breath Volume	1.328	04:54	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample #3	0.160	04:57	
Breath Volume	1.109	04:57	
Air Blank	0.000	04:58	
Control Test	0.080	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE JR, SCOTT, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIENIK, KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 06/04/2011 05:27:24

Diagnostics Check	OK	05:24	
Air Blank	0.000	05:24	
Control Test	0.080	05:25	
Air Blank	0.000	05:25	
Subject Sample #1		05:26	Subject Test Refused
Breath Volume	0.000	05:26	
Air Blank	0.000	05:26	
Control Test	0.080	05:26	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE, SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 06/04/2011 23:53:33

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.081	23:51	
Air Blank	0.000	23:51	
Subject Sample #1		23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.081	23:52	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIENIK, KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ORTEGA VALLEJO

DATE/TIME OF TEST: 06/05/2011 04:44:07

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.079	04:41	
Air Blank	0.000	04:41	
Subject Sample #1	0.162	04:42	
Breath Volume	2.062	04:42	
Air Blank		04:42	RFI Detect
Air Blank	0.000	04:43	
Control Test	0.080	04:43	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIENIK, KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ORTEGA VALLEJO

DATE/TIME OF TEST: 06/05/2011 04:52:11

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Subject Sample #1	0.161	04:47
Breath Volume	1.707	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample #2	0.157	04:50
Breath Volume	2.199	04:50
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIENIK, KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: SHIFFMAN

DATE/TIME OF TEST: 06/06/2011 22:57:18

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.081	22:53	
Air Blank	0.000	22:53	
Subject Sample #1		22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:56	
Control Test	0.080	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE, SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: PINSKER

DATE/TIME OF TEST: 06/07/2011 00:49:51

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.081	00:43	
Air Blank	0.000	00:43	
Subject Sample #1	0.209	00:44	
Breath Volume	1.210	00:44	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample #2	0.220	00:48	
Breath Volume	1.324	00:48	
Air Blank	0.000	00:48	
Control Test	0.079	00:48	
Air Blank		00:49	RFI Detect
Air Blank	0.000	00:49	

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE, SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: PINSKER

DATE/TIME OF TEST: 06/07/2011 01:01:10

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:52
Subject Sample #1	0.210	00:55
Breath Volume	1.398	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample #2	0.216	00:59
Breath Volume	1.441	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX, LENARD, E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH, ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/07/2011 22:36:32

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.080	22:30
Air Blank	0.000	22:30
Subject Sample #1	0.130	22:31
Breath Volume	2.164	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample #2	0.132	22:35
Breath Volume	1.542	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE JR, SCOTT, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: BECKNER

DATE/TIME OF TEST: 06/08/2011 23:11:44

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:05
Subject Sample #1	0.000	23:06
Breath Volume	1.253	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample #2	0.000	23:10
Breath Volume	1.414	23:10
Air Blank	0.000	23:10
Control Test	0.082	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE, SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: EVENS

DATE/TIME OF TEST: 06/14/2011 22:38:20

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:35	
Control Test	0.080	22:35	
Air Blank	0.000	22:36	
Subject Sample #1		22:36	RFI Detect
Breath Volume	0.312	22:36	
Air Blank	0.000	22:37	
Control Test	0.079	22:37	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:38	

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE, SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: EVENS

DATE/TIME OF TEST: 06/14/2011 22:46:24

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Subject Sample #1	0.199	22:41
Breath Volume	1.550	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample #2	0.197	22:44
Breath Volume	1.789	22:44
Air Blank	0.000	22:45
Control Test	0.080	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIENIK, KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 06/16/2011 03:50:24

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.077	03:41
Air Blank	0.000	03:42
Subject Sample #1	0.198	03:43
Breath Volume	1.722	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample #2	0.177	03:46
Breath Volume	2.105	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample #3	0.169	03:48
Breath Volume	1.378	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 06/16/2011 22:34:29

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Subject Sample #1	0.200	22:29
Breath Volume	1.179	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample #2	0.194	22:33
Breath Volume	1.234	22:33
Air Blank	0.000	22:33
Control Test	0.081	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIENIK, KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MCDEVITT

DATE/TIME OF TEST: 06/17/2011 02:35:36

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample #1	0.248	02:31
Breath Volume	2.703	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample #2	0.235	02:34
Breath Volume	2.300	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX, LENARD, E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GOSNELL, JEFFREY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: LAMBRECHT

DATE/TIME OF TEST: 06/17/2011 22:43:34

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:37
Subject Sample #1	0.286	22:39
Breath Volume	2.230	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample #2	0.267	22:42
Breath Volume	1.238	22:42
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX, LENARD, E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/29/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: SCHARVER

DATE/TIME OF TEST: 06/18/2011 23:22:26

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:20	
Control Test	0.078	23:20	
Air Blank	0.000	23:20	
Subject Sample #1		23:21	Subject Test Refused
Breath Volume	0.000	23:21	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time 06/21/2011 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIENIK, KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: WITTY

DATE/TIME OF TEST: 06/22/2011 01:16:43

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.081	01:02	
Air Blank	0.000	01:02	
Subject Sample #1	0.115	01:05	Volume Not Met
Breath Volume	0.253	01:05	
Air Blank	0.000	01:06	
Air Blank	0.000	01:08	
Subject Sample #2	0.109	01:10	
Breath Volume	1.609	01:10	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample #3		01:15	Subject Test Refused
Breath Volume	0.890	01:15	
Air Blank	0.000	01:15	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time 07/21/2011 20:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIENIK, KENNETH, A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE, SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: HEBRON

DATE/TIME OF TEST: 06/24/2011 02:20:04

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.081	02:17	
Air Blank	0.000	02:18	
Subject Sample #1		02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:19	
Control Test	0.080	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:20	

Data Download Date/Time 07/21/2011 20:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER, DAVID, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WHETSTON, B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: SMIT

DATE/TIME OF TEST: 06/25/2011 15:58:12

Diagnostics Check	OK	15:55	
Air Blank	0.000	15:55	
Control Test	0.082	15:56	
Air Blank	0.000	15:56	
Subject Sample #1		15:56	Subject Test Refused
Breath Volume	0.000	15:56	
Air Blank	0.000	15:57	
Control Test	0.081	15:57	
Air Blank	0.000	15:58	
Diagnostics Check	OK	15:58	

Data Download Date/Time 07/21/2011 20:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER, DAVID, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMALLEN, J
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 06/25/2011 18:49:24

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:46	
Control Test	0.082	18:46	
Air Blank	0.000	18:47	
Subject Sample #1		18:47	Subject Test Refused
Breath Volume	0.000	18:47	
Air Blank	0.000	18:48	
Control Test	0.082	18:48	
Air Blank		18:48	RFI Detect
Air Blank	0.000	18:49	

Data Download Date/Time 07/21/2011 20:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER, MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GOWER

DATE/TIME OF TEST: 06/30/2011 03:06:33

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:01	
Subject Sample #1	0.160	03:02	
Breath Volume	2.273	03:02	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample #2		03:05	RFI Detect
Breath Volume	0.000	03:05	
Air Blank		03:06	RFI Detect
Air Blank	0.000	03:06	

Data Download Date/Time 07/21/2011 20:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH, ROBBIE, S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER, MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/21/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 896383
DRY GAS EXPIRY DATE: 11/19/2012
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GOWER

DATE/TIME OF TEST: 06/30/2011 03:14:25

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.077	03:08
Air Blank	0.000	03:08
Subject Sample #1	0.169	03:09
Breath Volume	2.101	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample #2	0.164	03:12
Breath Volume	2.214	03:12
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time 07/21/2011 20:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKTON, QUENTIN, Q
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BECKTON, QUENTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: GILBERTSEN

DATE/TIME OF TEST: 06/10/2011 22:55:07

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.078	22:48
Air Blank	0.000	22:49
Subject Sample #1	0.148	22:50
Breath Volume	2.101	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample #2	0.147	22:53
Breath Volume	2.007	22:53
Air Blank	0.000	22:54
Control Test	0.079	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEASLEY, PAUL, M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BECKTON, Q
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: SULIKOWSKI

DATE/TIME OF TEST: 06/11/2011 22:31:11

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:26
Subject Sample #1	0.000	22:26
Breath Volume	1.925	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample #2	0.000	22:29
Breath Volume	1.628	22:29
Air Blank	0.000	22:30
Control Test	0.078	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORTER, DENNIS, F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SHORTER, DENNIS
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BRAUN

DATE/TIME OF TEST: 06/15/2011 21:47:44

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:36	
Control Test	0.079	21:36	
Air Blank	0.000	21:37	
Subject Sample #1	0.228	21:40	Volume Not Met
Breath Volume	0.613	21:40	
Air Blank	0.000	21:41	
Air Blank	0.000	21:43	
Subject Sample #2	0.211	21:46	Volume Not Met
Breath Volume	0.394	21:46	
Air Blank	0.000	21:46	
Control Test	0.077	21:47	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKTON, QUENTIN, Q
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BECKTON, QUENTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HORMANN

DATE/TIME OF TEST: 06/16/2011 22:52:08

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.078	22:45
Air Blank	0.000	22:46
Subject Sample #1	0.349	22:47
Breath Volume	1.417	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample #2	0.348	22:50
Breath Volume	2.082	22:50
Air Blank	0.000	22:51
Control Test	0.078	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKTON, QUENTIN, Q
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BECKTON, QUENTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BARRA

DATE/TIME OF TEST: 06/17/2011 03:37:19

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Subject Sample #1	0.028	03:32
Breath Volume	1.902	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.026	03:35
Breath Volume	1.773	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCHKIE, MICHAL, F
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: LINSKY, MICHAEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 06/19/2011 00:49:07

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:42
Subject Sample #1	0.234	00:44
Breath Volume	1.847	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample #2	0.244	00:47
Breath Volume	1.277	00:47
Air Blank	0.000	00:48
Control Test	0.077	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, BRYAN, S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SEAY, BRYAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 06/19/2011 02:25:04

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:18
Subject Sample #1	0.086	02:20
Breath Volume	1.437	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample #2	0.077	02:23
Breath Volume	1.535	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:25

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLES, RACHEL, M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRACE, DOUGLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: SHACKELFORD

DATE/TIME OF TEST: 06/19/2011 19:05:07

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:54	
Control Test	0.079	18:54	
Air Blank	0.000	18:55	
Subject Sample #1		18:58	No Sample Provided
Breath Volume	0.000	18:58	
Air Blank	0.000	18:58	
Air Blank	0.000	19:00	
Subject Sample #2		19:03	No Sample Provided
Breath Volume	0.000	19:03	
Air Blank	0.000	19:04	
Control Test	0.077	19:04	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:05	

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKTON, QUENTIN, Q
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BECKTON, QUENTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 11:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/23/2011 00:29:04

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:24
Subject Sample #1	0.155	00:24
Breath Volume	4.121	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample #2	0.154	00:27
Breath Volume	2.585	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:29

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, BRYAN, S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BECKTON, QUENTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PACCHIANA

DATE/TIME OF TEST: 06/24/2011 22:57:39

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:52
Subject Sample #1	0.048	22:53
Breath Volume	1.445	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample #2	0.047	22:56
Breath Volume	2.015	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHN'S COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKTON, QUENTIN, Q
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BECKTON, QUENTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/30/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17910080A3
DRY GAS EXPIRY DATE: 07/01/2012
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 06/25/2011 22:05:30

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Subject Sample #1	0.000	22:01
Breath Volume	2.062	22:01
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample #2	0.000	22:04
Breath Volume	2.140	22:04
Air Blank	0.000	22:04
Control Test	0.078	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time 06/29/2011 12:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, GARY, L
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: JONES, GARY
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 731635
DRY GAS EXPIRY DATE: 03/04/2012
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/03/2011 05:20:27

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:18	
Control Test	0.079	05:18	
Air Blank	0.000	05:18	
Subject Sample #1		05:19	Subject Test Refused
Breath Volume	0.000	05:19	
Air Blank	0.000	05:19	
Control Test	0.077	05:19	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time 06/08/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLANT, EMILE, C
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: HOLLANT, EMILE
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 731635
DRY GAS EXPIRY DATE: 03/04/2012
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: MONEY

DATE/TIME OF TEST: 06/05/2011 17:14:22

Diagnostics Check	OK	17:05	
Air Blank	0.000	17:06	
Control Test	0.079	17:06	
Air Blank	0.000	17:07	
Subject Sample #1		17:10	No Sample Provided
Breath Volume	0.000	17:10	
Air Blank	0.000	17:10	
Air Blank	0.000	17:12	
Subject Sample #2		17:12	Subject Test Refused
Breath Volume	0.000	17:12	
Air Blank	0.000	17:13	
Control Test	0.078	17:13	
Air Blank	0.000	17:14	
Diagnostics Check	OK	17:14	

Data Download Date/Time 06/08/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLANT, EMILE, C
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: HOLLANT, EMILE
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 05/27/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 731635
DRY GAS EXPIRY DATE: 03/04/2012
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 06/05/2011 18:40:50

Diagnostics Check	OK	18:34
Air Blank	0.000	18:35
Control Test	0.078	18:35
Air Blank	0.000	18:35
Subject Sample #1	0.172	18:36
Breath Volume	2.156	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:38
Subject Sample #2	0.167	18:39
Breath Volume	2.175	18:39
Air Blank	0.000	18:39
Control Test	0.078	18:40
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time 06/08/2011 21:33

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, GARY, L
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: JONES, GARY
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 06/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 731635
DRY GAS EXPIRY DATE: 03/04/2012
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 06/10/2011 05:35:07

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.078	05:29
Air Blank	0.000	05:30
Subject Sample #1	0.152	05:30
Breath Volume	2.558	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:32
Subject Sample #2	0.150	05:33
Breath Volume	2.550	05:33
Air Blank	0.000	05:34
Control Test	0.078	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time 07/19/2011 15:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, GARY, L
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: JONES, GARY
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 06/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 731635
DRY GAS EXPIRY DATE: 03/04/2012
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RIASCOS ARCOZ

DATE/TIME OF TEST: 06/21/2011 04:25:25

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.075	04:18
Air Blank	0.000	04:18
Subject Sample #1	0.204	04:20
Breath Volume	1.906	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.224	04:23
Breath Volume	2.695	04:23
Air Blank	0.000	04:24
Control Test	0.077	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time 07/19/2011 15:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEANITE, DUHAMEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: JEANITE, DUHAMEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 06/08/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 731635
DRY GAS EXPIRY DATE: 03/04/2012
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: AZELIN

DATE/TIME OF TEST: 06/21/2011 17:21:04

Diagnostics Check	OK	17:14
Air Blank	0.000	17:14
Control Test	0.078	17:15
Air Blank	0.000	17:15
Subject Sample #1	0.060	17:16
Breath Volume	1.960	17:16
Air Blank	0.000	17:17
Air Blank	0.000	17:18
Subject Sample #2	0.057	17:19
Breath Volume	1.839	17:19
Air Blank	0.000	17:20
Control Test	0.078	17:20
Air Blank	0.000	17:20
Diagnostics Check	OK	17:21

Data Download Date/Time 07/19/2011 15:43

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS, CARLOS, A
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: RAMOS, CARLOS
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 05/10/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 731636
DRY GAS EXPIRY DATE: 03/08/2012
VIOLATION CODE: DUI
ARREST TIME: 09:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: VIDAL\$^330 N WEST DR

DATE/TIME OF TEST: 06/10/2011 11:24:16

Diagnostics Check	OK	11:21	
Air Blank	0.000	11:21	
Control Test	0.079	11:22	
Air Blank	0.000	11:22	
Subject Sample #1		11:22	Subject Test Refused
Breath Volume	0.000	11:22	
Air Blank	0.000	11:23	
Control Test	0.078	11:23	
Air Blank	0.000	11:24	
Diagnostics Check	OK	11:24	

Data Download Date/Time 06/21/2011 15:03

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLZER, JACQUELINE,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MOLZER, JACQUELINE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 06/04/2011 03:30:11

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.077	03:26	
Air Blank	0.000	03:27	
Subject Sample #1		03:28	Subject Test Refused
Breath Volume	0.000	03:28	
Air Blank	0.000	03:29	
Control Test	0.078	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time 06/28/2011 07:50

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOLZER, JACQUELINE,

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: MOLZER, JACQUELINE

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: DUI

ARREST TIME: 01:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SANCHEZ FIGUEROA

DATE/TIME OF TEST: 06/14/2011 01:48:15

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:42
Subject Sample #1	0.107	01:43
Breath Volume	1.921	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample #2	0.106	01:46
Breath Volume	1.582	01:46
Air Blank	0.000	01:47
Control Test	0.077	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time 06/28/2011 07:50

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLZER, JACQUELINE,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MOLZER, JACQUELINE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: PROCTOR

DATE/TIME OF TEST: 06/19/2011 04:40:40

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.078	04:37	
Air Blank	0.000	04:37	
Subject Sample #1	0.175	04:38	
Breath Volume	1.519	04:38	
Air Blank		04:39	RFI Detect
Air Blank	0.000	04:39	
Control Test	0.078	04:40	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time 06/28/2011 07:50

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLZER, JACQUELINE,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MOLZER, JACQUELINE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/31/2011

Date or Time Changed: No
DRY GAS STANDARD LOT #: 670754
DRY GAS EXPIRY DATE: 09/22/2011
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: PROCTOR

DATE/TIME OF TEST: 06/19/2011 04:49:43

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.077	04:42
Air Blank	0.000	04:43
Subject Sample #1	0.176	04:44
Breath Volume	1.265	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample #2	0.178	04:47
Breath Volume	1.437	04:47
Air Blank	0.000	04:48
Control Test	0.075	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time 06/28/2011 07:50