

FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

20-001347  
LDB

RECEIVED

SEP 06 2011

FDLE  
Alcohol Testing Program

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: SARASOTA COUNTY SO  
Instrument Serial Number: 80-001347 Software: 8100.27  
Date of Test: 09/02/2011

Date of Last Agency Inspection: 08/16/2011  
Observation Period Began: 16:00  
Subject's Name: TEST T TEST

DOB: 01/01/1971 Sex: M

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	16:50
	Air Blank	0.000	16:50
	Control Test	0.080	16:50
	Air Blank	0.000	16:51
	Subject Sample #1	0.201	16:52
	Air Blank	0.000	16:52
	Air Blank	0.000	16:54
	Subject Sample #2	0.199	16:54
	Air Blank	0.000	16:55
	Control Test	0.079	16:55
	Air Blank	0.000	16:56
	Diagnostics Check	OK	16:56

*ARS 20 used through breath tube to check if instrument providing accurate results. Check due to volumes reporting high on several test subjects.*

Cylinder Lot: 904241  
Exp: 11/12/2012

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I CHRIS T BUTLER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

20-001347  
LAB

RECEIVED

SEP 06 2011

FDLE  
Alcohol Testing Program

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: SARASOTA COUNTY SO  
Instrument Serial Number: 80-001347 Software: 8100.27  
Date of Test: 09/02/2011

Date of Last Agency Inspection: 08/16/2011  
Observation Period Began: 16:00  
Subject's Name: TEST T TEST

DOB: 01/01/1971 Sex: M

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	16:39
	Air Blank	0.000	16:40
	Control Test	0.080	16:40
	Air Blank	0.000	16:40
	Subject Sample #1	0.080	16:41
	Air Blank	0.000	16:41
	Air Blank	0.000	16:43
	Subject Sample #2	0.080	16:44
	Air Blank	0.000	16:44
	Control Test	0.080	16:45
	Air Blank	0.000	16:45
	Diagnostics Check	OK	16:45

I hooked up ARES 08 to breath hose and blew through simulator. Drop so 2x. This is due to questions about volumes reporting high on subject tests.

Cylinder Lot: 904241  
Exp: 11/12/2012

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I CHRIS T BUTLER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

80-001347  
JOB  
RECEIVED  
SEP 06 2011  
FLDLE  
Alcohol Testing Program

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: SARASOTA COUNTY SO  
Instrument Serial Number: 80-001347 Software: 8100.27  
Date of Test: 09/02/2011

Date of Last Agency Inspection: 08/16/2011  
Observation Period Began: 16:00  
Subject's Name: TEST T TEST

DOB: 01/01/1971 Sex: M

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	16:35
	Air Blank	0.000	16:35
	Control Test	0.079	16:36
	Air Blank	0.000	16:36
	Subject Sample #1	0.049	16:37
	Air Blank	IPS*	16:37
	Air Blank	0.000	16:37
	Control Test	0.080	16:38
	Air Blank	0.000	16:38
	Diagnostics Check	OK	16:38

\*Improper Sample

System check @ .05 level ARS. I hooked up simulator to mouthpiece to nose and blew through simulator to pass over pressure sensor.

Cylinder Lot: 904241  
Exp: 11/12/2012

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I CHRIS T BUTLER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.